Artificial Intelligence for context-aware surgical guidance in complex robot-assisted oncological procedures: An exploratory feasibility study

Fiona R. Kolbinger[a,b,c], Sebastian Bodenstedt[d], Matthias Carstens[a], Stefan Leger[c,d], Stefanie Krell[d], Franziska M. Rinner[a], Thomas P. Nielen[a], Johanna Kirchberg[a,b], Johannes Fritzmann[a,b], Jürgen Weitz[a,b,c], Marius Distler[a,b,§], Stefanie Speidel[c,d,§,✉]

[a] Department of Visceral, Thoracic and Vascular Surgery, University Hospital and Faculty of Medicine Carl Gustav Carus, Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany

[b] National Center for Tumor Diseases (NCT/UCC), Dresden, Germany: German Cancer Research Center (DKFZ), Heidelberg, Germany; Faculty of Medicine and University Hospital Carl Gustav Carus, Technische Universität Dresden, Dresden, Germany; Helmholtz-Zentrum Dresden - Rossendorf (HZDR), Fetscherstraße 74, 01307 Dresden, Germany

[c] Else Kröner Fresenius Center for Digital Health (EKFZ), Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany

[d] Division of Translational Surgical Oncology, National Center for Tumor Diseases (NCT), Partner Site Dresden, Fetscherstraße 74, 01307 Dresden, Germany

§   These authors jointly supervised this work

✉ Corresponding author: Prof. Dr. Stefanie Speidel, Division of Translational Surgical Oncology, National Center for Tumor Diseases (NCT/UCC), Partner Site Dresden, Fetscherstrasse 74, 01307 Dresden, Germany
☏ +49 (0) 351 5413
✉ stefanie.speidel@nct-dresden.de

KEYWORDS

Surgical data science; rectal resection; surgical oncology; robot-assisted surgery; dissection plane; mesorectal fascia

ABBREVIATIONS

AI: Artificial Intelligence, CNN: Convolutional neural network, F1: F1 score, IoU: Intersection-over-Union, LM: Lateral mobilization (phase), ME: Mesorectal excision (phase), MM: Medial mobilization (phase), PME: Partial mesorectal excision, SD: Standard deviation, TME: Total mesorectal excision, VD: Vascular dissection (phase)

ABSTRACT

Introduction: Complex oncological procedures pose various surgical challenges including

dissection in distinct tissue planes and preservation of vulnerable anatomical structures

throughout different surgical phases. In rectal surgery, violation of dissection planes increases the

risk of local recurrence and autonomous nerve damage resulting in incontinence and sexual

dysfunction. This work explores the feasibility of phase recognition and target structure

segmentation in robot-assisted rectal resection (RARR) using machine learning.

Materials and Methods: A total of 57 RARR were recorded and annotated with respect to surgical

phases and exact locations of target structures (anatomical structures, tissue types, static

structures, and dissection areas). For surgical phase recognition, three machine learning models

were trained: LSTM, MSTCN, and TransSVNet. Based on pixel-wise annotations of target

structures in 9037 images, individual segmentation models based on DeepLabV3 were trained.

Model performance was evaluated using F1 score, Intersection-over-Union (IoU), accuracy,

precision, recall, and specificity.

Results: The best results for phase recognition were achieved with the MSTCN model (F1 score:

0.82 ± 0.01, accuracy: 0.84 ± 0.03). Mean IoUs for target structure segmentation ranged from 0.14

± 0.22 to 0.80 ± 0.14 for organs and tissue types and from 0.11 ± 0.11 to 0.44 ± 0.30 for dissection

areas. Image quality, distorting factors (i.e. blood, smoke), and technical challenges (i.e. lack of

depth perception) considerably impacted segmentation performance.

Conclusion: Machine learning-based phase recognition and segmentation of selected target

structures are feasible in RARR. In the future, such functionalities could be integrated into a

context-aware surgical guidance system for rectal surgery.

1 INTRODUCTION

Despite advances in nonsurgical management of rectal cancer, the majority of curative treatment approaches require oncologically radical surgical removal of the primary tumor and locoregional lymph nodes within the mesorectal envelope (total mesorectal excision, TME). This surgical step requires delicate preparation along the mesorectal fascia to ensure complete resection and minimize the risk of local tumor recurrence [1,2]. In parallel, autonomous nerves running in close proximity to the resection area must be protected to avoid postoperative incontinence and sexual dysfunction. With highly varying numbers depending on surgeon expertise and center volume, these functional sequelae affect up to 85% of patients after oncological rectal resection and can be a direct result of suboptimal surgical preparation outside of anatomical planes, leading to nerve injury [3,4].

Surgical robots offer mechanical enhancements such as improved instrument dexterity, hand-eye coordination, and a three-dimensional view, however in many types of surgery including rectal resection, no robust clinical benefit over conventional laparoscopic surgery could be demonstrated to date [5]. Based on the hypothesis that an integration of surgical guidance (i.e. segmentation overlays of specific structures in specific phases of a procedure) into minimally-invasive surgical systems could markedly improve surgical quality and outcome and diminish quality variation between surgeons, extensive research on the development of image recognition algorithms has been conducted in recent years using laparoscopic imaging data. Most of these endeavors, however, have focused on tasks with indirect surgical benefit, such as automated instrument detection [6,7]. Thus far, translational Artificial Intelligence (AI)-based success stories in the field of surgery are lacking and clinical applications are mostly limited to orthopedic, neurosurgical, and hepatic surgical procedures [8,9]. With regard to approaches with high translational potential in laparoscopic surgery, deep learning-based algorithms have recently been shown to identify relevant anatomical areas during cholecystectomy [10,11] and organs during laparoscopy [12,13]. Of note, such methods have primarily been established for less complex surgical procedures, for which – in part – open-access datasets exist for the purpose of algorithm development and validation [14,15]. In these operations, however, the benefit of AI-based assistance is marginal.

In this project, we explore the potential of deep learning-based phase recognition – a prerequisite for context awareness – and target structure segmentation using algorithms based on state-of-the-art convolutional neural networks (CNNs) on a dataset of spatio-temporal data in the form of temporally annotated robot-assisted rectal resection (RARR) videos and semantically segmented video frames. In the future, these components could serve as a basis for a context-aware surgical guidance system.

2 MATERIAL AND METHODS

2.1 Patient population and dataset

Between February 2019 and January 2021, video data from 57 robot-assisted anterior rectal resections, intersphincteric resections, or abdominoperineal resections were collected at the University Hospital Carl Gustav Carus Dresden, Germany (Appendix A). All patients had a clinical indication for the surgical procedure, recommended by an interdisciplinary tumor board. Procedures were performed using the da Vinci® Xi system (Intuitive Surgical, Sunnyvale, CA, USA). Surgeries were recorded using the CAST system (Orpheus Medical GmBH, Frankfurt a.M., Germany) and saved at a resolution of 1920 x 1080 pixels in MPEG-4 format.

This study was performed in accord with the ethical standards of the Helsinki declaration and its later amendments. The local Institutional Review Board (ethics committee at the Technical University Dresden) reviewed and approved this study (approval number: BO-EK-140032021). Written informed consent was waived. The trial was registered on clinicaltrials.gov (trial registration ID: NCT05268432).

2.2 Annotations

The dataset of 57 RARR recordings was randomly split into a pre-training dataset (24 temporally non-annotated videos) and an annotated training dataset (33 videos).

According to a previously created protocol (Appendix B), the surgical process was temporally annotated using b<>com *Surgery Workflow Toolbox* [Annotate] version 2.2.0 (b<>com, Cesson-Sévigné, France) (Figure 1). The definition and order of the five phases – preparation and intraabdominal orientation, medial mobilization of descending colon, lateral mobilization of descending colon, mesorectal excision, and extraabdominal preparation – was based on institutional standards and international recommendations for RARR [16,17]. Splenic flexure and descending colon mobilization was carried out from medial to lateral as described previously by Ahmed *et al.* [16].

For training and validation of phase-specific semantic segmentation algorithms, equidistant frames from at least ten selected video recordings for each phase were selected based on criteria of target structure visibility and exposition (Appendix C) and annotated with regard to visibility of anatomical structures and dissection planes (Figure 1, Appendix D) using the open-source software 3D Slicer with the SlicerRT program extension [18]. Annotations were performed by two members of the research team (MC, FMR) after thorough education and several rounds of training annotations with two experts in robot-assisted rectal surgery (JW, MD), and reviewed by a surgeon in training with considerable experience in robot-assisted rectal surgery (FRK).

Figure 1: Overview of the data collection, data processing, and machine learning process in this study. Robot-assisted rectal resections were recorded and video data was annotated with respect to surgical phases and exact location of target structures. Phase recognition and structure segmentation models were based on a ResNet50 as backbone and combined with temporal models or segmentation models, respectively. Model performance was assessed using 4-fold cross validation and classical computer vision metrics. Abbreviations: Intersection over union (IoU), Lateral mobilization (LM), Medial mobilization (MM), Mesorectal excision (ME), Vascular dissection (VD).

2.3   Model development and evaluation: Phase recognition

For identification of the five surgical phases, we applied a typical two-stage pipeline. In the first stage, a residual neural network (ResNet50) [19] was used to learn image features, having demonstrated a high performance on the Cholec80 dataset [20]. In the second stage, the features of the ResNet50 were used to learn the temporal context with the following models: standard long short-term memory network (LSTM) [21], MSTCN [20], and Trans-SVNet [22] (Figure 1). Trans-SVNet has a third step, in which the features of ResNet50 and the features of MSTCN are used to train a Transformer model.

Pre-training of the feature extractor was performed in two different configurations: (a) the ResNet50 was pre-trained on the ImageNet database [23] only and (b) the pre-trained feature extractor was fine-tuned on 24 randomly selected temporally non-annotated RARR videos using self-supervised training. The self-supervised learning approach described by Funke *et al.* [24] was adapted, so that a ResNet50 was trained based on temporally coherent video frame embedding using 1st and 2nd order contrastive loss [25,26].

Afterwards, the two networks were trained frame-wise with the label information of the current phase. Data augmentation consisted of image rotation, flipping and noise. The network was trained for 30 epochs, each consisting of 3000 random chosen batches with a size of 64. As optimizer, AdamW [27] with an initial learning rate of 0.0001 was used and the weighted cross entropy loss was applied to handle the class-imbalance problem.

As self-supervised pre-training did not result in any performance benefits (Appendix E), the ResNet50 pre-trained on ImageNet was selected as feature extractor and combined with the temporal models. The 33 temporally annotated RARR videos were used for training and evaluated within a 4-fold cross validation scheme. Performance was evaluated on the test folds using F1 score and accuracy.

2.4   Model development and evaluation: Semantic segmentation

For training of the segmentation algorithms for each individual phase, at least ten video recordings were selected per phase with regard to visibility of anatomical structures and dissection planes (Appendix C) and annotated as described above. For each surgery phase, a separate convolutional neural network was trained for the phase-specific segmentation of target structures and tissue planes. For this purpose, the DeepLabv3 [28] framework was used with a ResNet50 backbone with default PyTorch pretraining on the COCO dataset [29]. The model was trained for 100 epochs with a batch size of 16 and evaluated within a 4-fold cross validation scheme (Figure 1). Data augmentation consisting of image resizing; image shifts and rotations were used to enhance the training process. For optimization of the network, the AdamW algorithm [27] was

173 used with a learning rate of 0.0001. Segmentation performance was assessed on the test folds

174 using F1 score, IoU, precision, recall, and specificity [30], commonly used technical measures of

175 prediction exactness, ranging from 0 (least exact prediction) to 1 (entirely correct prediction).

3   RESULTS

3.1   Dataset

In total, 57 RARR with an average duration of 414 ± 112 minutes performed by nine experts in

robot-assisted oncological surgery were recorded and analyzed. The majority of patients (n = 43,

75.4%) were male, the predominant indication for the surgical procedure was colorectal cancer

(n = 54, 94.7%). Twenty-five patients (43.9%) had undergone neoadjuvant irradiation to the

surgical field. Overall, the sample represents a wide variety of rectal cancer manifestations,

including a considerable number of patients (n = 27, 47.4%) with locally advanced tumors

(Figure 1, Appendix A).

3.2   Surgical phase recognition

Thirty-three videos were temporally annotated with regard to the sequence of five standardized

surgical phases (Figure 1). The most accurate phase recognition results were achieved with

MSTCN [20] (F1 score: 0.82 ± 0.01), outperforming the standard LSTM [21] (F1 score: 0.79 ±

0.02) and Trans-SVNet [22] (F1 score: 0.79 ± 0.01) (Table 1).

Temporal Model	F1 score	Accuracy
LSTM	0.79 ± 0.02	0.82 ± 0.01
MSTCN	**0.82 ± 0.01**	**0.84 ± 0.03**
Trans-SVNet	0.79 ± 0.01	0.83 ± 0.02

193 **Table 1: Summary of performance metrics for surgical phase recognition using a standard LSTM, MSTCN, and**
194 **Trans-SVNet.** For each metric, mean and standard deviation are displayed.
195

3.3   Segmentation of anatomical structures

For semantic segmentation, individual models based on DeepLabv3 [28] were trained to identify

16 distinct target structures, encompassing nine anatomical structures (Inferior Mesenteric Artery,

Inferior Mesenteric Vein, Gerota's fascia, the mesocolon, the abdominal wall, the colon, the small

intestine, the rectum, and seminal vesicles), two additional tissue types (adhesion and fat), two

static structures (plastic clip and metal clip), and three dissection areas (the dissection line during

medial mobilization of the colon, the avascular dissection plane as well as the exact dissection

line during mesorectal excision). In total, 9037 annotated frames were used for training of the

semantic segmentation algorithms (Figure 1, Appendix C, Appendix D). Table 2 displays mean

segmentation metrics for individual segments predicted by the trained phase-specific algorithms.

Phase	Target Structure	F1 score	IoU	Precision	Recall	Specificity
VD	Inferior Mesenteric Artery	0.29 ± 0.30	0.21 ± 0.24	0.48 ± 0.41	0.36 ± 0.30	1.00 ± 0.01
	Inferior Mesenteric Vein	0.33 ± 0.34	0.26 ± 0.28	0.47 ± 0.39	0.48 ± 0.37	1.00 ± 0.00
	Plastic Clip	0.35 ± 0.38	0.28 ± 0.32	0.54 ± 0.33	0.47 ± 0.45	1.00 ± 0.00
	Metal Clip	0.11 ± 0.21	0.07 ± 0.15	0.12 ± 0.21	0.47 ± 0.42	1.00 ± 0.01
MM	Gerota's Fascia	0.84 ± 0.18	0.75 ± 0.21	0.82 ± 0.21	0.90 ± 0.12	0.94 ± 0.06
	Mesocolon	0.79 ± 0.20	0.69 ± 0.23	0.79 ± 0.24	0.84 ± 0.16	0.94 ± 0.07
	Dissection Line (MM)	0.21 ± 0.21	0.13 ± 0.14	0.21 ± 0.22	0.39 ± 0.29	1.00 ± 0.00
LM	Abdominal Wall	0.88 ± 0.10	0.80 ± 0.14	0.88 ± 0.13	0.90 ± 0.10	0.96 ± 0.04
	Adhesion	0.33 ± 0.27	0.23 ± 0.20	0.37 ± 0.31	0.47 ± 0.27	0.99 ± 0.01
	Colon	0.62 ± 0.35	0.53 ± 0.33	0.67 ± 0.36	0.74 ± 0.27	0.98 ± 0.02
	Fat	0.79 ± 0.22	0.69 ± 0.23	0.78 ± 0.24	0.87 ± 0.14	0.95 ± 0.05
	Small Intestine	0.26 ± 0.38	0.22 ± 0.33	0.33 ± 0.43	0.57 ± 0.38	1.00 ± 0.01
ME	Rectum	0.51 ± 0.27	0.38 ± 0.23	0.54 ± 0.29	0.64 ± 0.26	0.99 ± 0.03
	Dissection Plane (ME)	0.54 ± 0.33	0.44 ± 0.30	0.59 ± 0.34	0.63 ± 0.32	0.94 ± 0.06
	Dissection Line (ME)	0.17 ± 0.17	0.11 ± 0.11	0.17 ± 0.17	0.25 ± 0.23	1.00 ± 0.00
	Seminal Vesicles	0.18 ± 0.29	0.14 ± 0.22	0.30 ± 0.38	0.33 ± 0.35	1.00 ± 0.00

207 **Table 2: Summary of performance metrics for phase-specific semantic segmentation of target structures and**
208 **dissection areas.** For each metric, mean and standard deviation are displayed. Abbreviations: Intersection over union
209 (IoU), Lateral mobilization (LM), Medial mobilization (MM), Mesorectal excision (ME), Vascular dissection (VD).
210

Overall, segmentation model performance varied strongly both between structures (mirrored by

variations in mean performance metrics) and between individual images (mirrored by large

standard deviations and uneven metric distributions within one structure) (Table 2, Figure 2,

Figure 3). Among anatomical structures, segmentation performance was highest for the

abdominal wall (F1 score: 0.88 ± 0.10), Gerota's fascia (F1 score: 0.84 ± 0.18), and the mesocolon

(F1 score: 0.79 ± 0.20), whereas segmentation performance was weaker for seminal vesicles (F1

score: 0.18 ± 0.29), the small intestine (F1 score: 0.26 ± 0.38), and intestinal vessels (F1 score:

0.29 ± 0.30 for the Inferior Mesenteric Artery and 0.33 ± 0.34 for the Inferior Mesenteric Vein)

(Table 2, Figure 2).

Figure 2: Performance of machine learning models based on DeepLabv3 for segmentation of anatomical structures. Violin plot illustrations display the distributions of performance metrics, with the median and quartiles illustrated as solid and dashed lines, respectively. Example images with the highest and weakest segmentation performance (by IoU) are displayed for each target structure. Ground truth is displayed as blue line (left image), model segmentations are displayed as white overlay (right image). Abbreviations: F1 score (F1), Intersection over union (IoU).

3.4   Segmentation of tissue types, static structures and dissection planes

To explore the capability to provide spatial information exceeding anatomical structures,
segmentation models based on DeepLabv3 were trained to delineate tissue types with relevance
in the context of surgical challenges (fat and adhesion), surgical clips that often require accurate
positioning, and dissection areas that are of prognostic relevance in rectal surgery.

Fat was among the best-predicted structures overall (F1 score: 0.79 ± 0.22), while prediction
quality of tissue adhesion was variable and overall weaker than several organs (F1 score: 0.33 ±
0.27) (Table 2, Figure 2, Figure 3 a). The plastic and the metal clip, both small but distinct-
appearing structures, were falsely not predicted in several images displaying them, resulting in a
large proportion of unpredicted images (Figure 3 b). During mesorectal excision, the avascular
dissection plane could be predicted with an F1 score of 0.54 ± 0.33, whereas prediction of the
exact dissection lines both during medial mobilization of the mesocolon and mesorectal dissection
was largely unsuccessful and, when predicted, highly variable at F1 scores of 0.21 ± 0.21 and
0.17 ± 0.17, respectively (Table 2, Figure 3 c).

3.5   Challenges of segmentation in laparoscopy

Review of all segmentation results, including images with high and weak segmentation
performance for each target structure, revealed challenges related to segmentation algorithms
(Figure 4). First, segment size substantially impacted segmentation performance, with a trend
towards better prediction of large, uniform-appearing organs such as the abdominal wall or the
mesocolon as compared to small or tubular and variable-appearing structures such as seminal
vesicles and dissection planes. In turn, some larger structures such as Gerota's fascia, the
mesocolon, or fat were prone to unspecific false positive prediction (Figure 2, Figure 3). Second,
images distorted by excessive amounts of blood, smoke, soiling of the endoscope lens, or blurring
caused by camera shake tended to result in inferior segmentation performance (Figure 2,
Figure 3). Third, there were specific prediction errors for selected target structures; for example,
the metal clip was occasionally predicted as part of surgical instruments and the mesenteric
vessels (inferior mesenteric artery and vein) were frequently mixed up (Figure 2, Figure 3).

Figure 3: Performance of machine learning models based on DeepLabv3 for segmentation of tissue types (a), static structures (b), and dissection areas (c). Violin plot illustrations display the distributions of performance metrics, with the median and quartiles illustrated as solid and dashed lines, respectively. Example images with the highest and weakest segmentation performance (by IoU) are displayed for each target structure. Ground truth is displayed as blue line (left image), model segmentations are displayed as white overlay (right image). Abbreviations: F1 score (F1), Intersection over union (IoU), Medial Mobilization phase (MM), Mesorectal Excision phase (ME).

4 DISCUSSION

Correct recognition of visual cues is a key challenge in minimally-invasive surgery, as the majority of currently available laparoscopic and robotic instruments and devices do not provide any haptic feedback or additional imaging to visualize hidden structures. In rectal surgery, correct identification of anatomical structures and their boundaries and preparation in thin dissection planes is crucial given the strong association with both local recurrence risk and the risk of incontinence and impaired sexual function [1–4]. In this work, we explored two major computer vision tasks in the context of robot-assisted rectal surgery that could serve as a basis for future context-aware guidance: recognition of surgical phases and exact delineation of target structures including organs and dissection areas.

In summary, differentiation of five surgical phases was feasible at an average F1 score of 0.82 of the best-performing model, MSTCN [20]. Segmentation of anatomical structures and dissection areas demonstrated promising results for selected target structures using models based on DeepLabv3 [28]. However, several challenges related to training data, model development and validation, and clinical translation limit the potential to leverage the potential of machine learning-based segmentation in laparoscopy (Figure 4).

Figure 4: Challenges of segmentation in laparoscopy. This figure highlights existing challenges of target structure segmentation in laparoscopic imaging related to training data, model development, model evaluation, and clinical translation.

In general, the presented observations are in line with previous studies on AI-based temporal and
spatial analysis of laparoscopic images [31]. With regard to surgical phase recognition, similar
accuracies around 0.8 have been reported for other laparoscopic procedures including
cholecystectomy [20] and sigmoidectomy [32]. Previous studies investigating segmentation of
anatomical structures have mostly focused on classification of the presence of certain organs [15]
or rough localization of the detected organs using bounding boxes [33]. A recent preprint [12]
describes semantic segmentation of organs based on a large-scale dataset of organ
segmentations [13]. This preprint covers six organs that were also analyzed in the present work.
Comparing the results of both works, similarly accurate segmentation performances were
observed for the abdominal wall, while considerably inferior segmentation performances were
observed for the colon, inferior mesenteric vessels, the small intestine, and seminal vesicles. In
part, this inferior performance is likely related to a smaller amount of training data. For instance,
algorithms for identification of seminal vesicles and the small intestine were trained on comparably
small datasets of 336 and 104 annotated images, respectively (Appendix D), while over 1000
annotated images were used to train models in the aforementioned work. Integration of more and
diverse – ideally multicentric – training data would likely improve segmentation performance and
increase clinical relevance of respective models [34]. One of the main reasons for scarcity of
annotated laparoscopic datasets is the high annotation effort required for an integration of clinically
meaningful knowledge into a dataset [35]. In this context, standardization of (video) data recording
and adaptability of the annotation process to differences in surgical techniques and approaches
pose important challenges to consider for future studies using annotated laparoscopic data. Of
note, the Society of American Gastrointestinal and Endoscopic Surgeons recently published
recommendations for surgical video annotation to increase the amount of publicly available
datasets and enable the creation of benchmark AI algorithms with a clinical impact [36].
Besides organs, this work also explores feasibility of segmentation of dissection areas. Here,
segmentation of the avascular dissection plane during mesorectal excision was promising at an
average F1 score of 0.54, which is comparable to previous findings on segmentation of loose
connective tissue fibers in gastric surgery [37]. In contrast, exact dissection lines could not be
successfully segmented. Especially during mesorectal excision, the appearance of embryonal
tissue planes and the closely related dissection line at the mesorectal fascia can vary considerably
based on neoadjuvant (radio-)therapy and individual factors such as body composition, making
identification particularly challenging [38]. Performance variations of segmentation algorithms
were moreover related to common limitations of classical computer vision metrics measuring
overlap between ground truth and model prediction [39,40]. For example, such metrics are skewed
by segment size and shape, which could further contribute to poor segmentation performance in
small and tubular structures such as clips and dissection lines.

The limitations of this work are mostly related to the outlined challenges of AI methods in
laparoscopy (Figure 4). First, the monocentric study design and single-rater annotation of training
data considerably restrict generalizability [41]. Second, segmentation of dissection lines and small
structures such as metal clips was largely unsuccessful, likely related to diverse appearance and
challenges in existing metrics that are considerably influenced by segment shape and size. To
pertain real-world image challenges, images with suboptimal image quality because of camera
shake and presence of significant amounts of blood or smoke were not excluded; these factors
negatively impacted segmentation performance. Last, the exploratory study design does not allow
for estimation of the clinical value of the described phase or target structure recognition models.
To enhance applicability and clinical value of the system [42], future works will focus on an
integration of recognition uncertainty into a graphical user interface displaying target structure
segmentations, for instance through Bayesian computation methods and heatmap-like
visualizations [43]. For clinical implementation, such a system will need to meet real-time
requirements for clinical usability, necessitating exact determination and subsequent minimization
of the computation-associated delay to the range of tenths of a second. The existing limitations
notwithstanding, this work represents an important addition to the growing body of research on
image-based guidance in abdominal surgery, as it significantly expands the horizon of data
science-assisted surgery to procedures of higher complexity and technical difficulty.

5   CONCLUSIONS

In conclusion, the results suggest that surgical phase recognition and segmentation of selected
target structures can satisfactorily be implemented in complex minimally-invasive surgical
procedures. These capabilities represent essential components of potential future context-aware
guidance systems in complex minimally-invasive oncological surgery. On a general level, this work
highlights challenges in the implementation of clinically relevant surgical guidance functions, in
particular related to training data, model training and evaluation, and clinical translation. This
encourages inter-institutional collaboration and improved data accessibility to facilitate data
science-based approaches in a diverse range of surgical procedures with varying complexity.

REFERENCES

[1]   Heald RJ, Husband EM, Ryall RDH. The mesorectum in rectal cancer surgery—the clue to pelvic recurrence? Br J Surg 1982. https://doi.org/10.1002/bjs.1800691019.

[2]   Quirke P, Steele R, Monson J, Grieve R, Khanna S, Couture J, et al. Effect of the plane of surgery achieved on local recurrence in patients with operable rectal cancer: a prospective study using data from the MRC CR07 and NCIC-CTG CO16 randomised clinical trial. Lancet 2009;373. https://doi.org/10.1016/S0140-6736(09)60485-2.

[3]   Chew MH, Yeh YT, Lim E, Seow-Choen F. Pelvic autonomic nerve preservation in radical rectal cancer surgery: Changes in the past 3 decades. Gastroenterol Rep 2016. https://doi.org/10.1093/gastro/gow023.

[4]   Sturiale A, Martellucci J, Zurli L, Vaccaro C, Brusciano L, Limongelli P, et al. Long-term functional follow-up after anterior rectal resection for cancer. Int J Colorectal Dis 2017. https://doi.org/10.1007/s00384-016-2659-6.

[5]   Jayne D, Pigazzi A, Marshall H, Croft J, Corrigan N, Copeland J, et al. Effect of robotic-assisted vs conventional laparoscopic surgery on risk of conversion to open laparotomy among patients undergoing resection for rectal cancer the rolarr randomized clinical trial. JAMA - J Am Med Assoc 2017;318:1569–80. https://doi.org/10.1001/jama.2017.7219.

[6]   Alsheakhali M, Eslami A, Roodaki H, Navab N. CRF-Based Model for Instrument Detection and Pose Estimation in Retinal Microsurgery. Comput Math Methods Med 2016;2016. https://doi.org/10.1155/2016/1067509.

[7]   Jin A, Yeung S, Jopling J, Krause J, Azagury D, Milstein A, et al. Tool Detection and Operative Skill Assessment in Surgical Videos Using Region-Based Convolutional Neural Networks. Proc - 2018 IEEE Winter Conf Appl Comput Vision, WACV 2018 2018;2018-January:691–9.

[8]   Burström G, Balicki M, Patriciu A, Kyne S, Popovic A, Holthuizen R, et al. Feasibility and accuracy of a robotic guidance system for navigated spine surgery in a hybrid operating room: a cadaver study. Sci Rep 2020;10. https://doi.org/10.1038/s41598-020-64462-x.

[9]   Hu Z, Fang C, Li B, Zhang Z, Cao C, Cai M, et al. First-in-human liver-tumour surgery guided by multispectral fluorescence imaging in the visible and near-infrared-I/II windows. Nat Biomed Eng 2020;4:259–71. https://doi.org/10.1038/s41551-019-0494-0.

[10]   Madani A, Namazi B, Altieri MS, Hashimoto DA, Rivera AM, Pucher PH, et al. Artificial Intelligence for Intraoperative Guidance. Ann Surg 2020. https://doi.org/10.1097/sla.0000000000004594.

[11]   Mascagni P, Vardazaryan A, Alapatt D, Urade T, Emre T, Fiorillo C, et al. Artificial Intelligence for Surgical Safety: Automatic Assessment of the Critical View of Safety in Laparoscopic Cholecystectomy Using Deep Learning. Ann Surg 2020.

386   https://doi.org/10.1097/SLA.0000000000004351.

387   [12]   Kolbinger FR, Rinner FM, Jenke AC, Carstens M, Leger S, Distler M, et al. Better than
388   humans? Machine learning-based anatomy recognition in minimally-invasive abdominal
389   surgery. MedRxiv 2022:2022.11.11.22282215.
390   https://doi.org/10.1101/2022.11.11.22282215.

391   [13]   Carstens M, Rinner FM, Bodenstedt S, Jenke AC, Weitz J, Distler M, et al. The Dresden
392   Surgical Anatomy Dataset for Abdominal Organ Segmentation in Surgical Data Science.
393   Sci Data 2023;10:1–8. https://doi.org/10.1038/s41597-022-01719-2.

394   [14]   Twinanda AP, Shehata S, Mutter D, Marescaux J, de Mathelin M, Padoy N. EndoNet: A
395   Deep Architecture for Recognition Tasks on Laparoscopic Videos. IEEE Trans Med
396   Imaging 2016;36:86–97.

397   [15]   Leibetseder A, Petscharnig S, Primus MJ, Kletz S, Münzer B, Schoeffmann K, et al.
398   LapGyn4: A Dataset for 4 Automatic Content Analysis Problems in the Domain of
399   Laparoscopic Gynecology. Proc 9th ACM Multimed Syst Conf 2018;18.
400   https://doi.org/10.1145/3204949.

401   [16]   Ahmed J, Kuzu MA, Figueiredo N, Khan J, Parvaiz A. Three-step standardized approach
402   for complete mobilization of the splenic flexure during robotic rectal cancer surgery. Color
403   Dis 2016;18:O171–4. https://doi.org/10.1111/CODI.13313.

404   [17]   Panteleimonitis S, Harper M, Hall S, Figueiredo N, Tahseen Qureshi ·, Parvaiz A. Precision
405   in robotic rectal surgery using the da Vinci Xi system and integrated table motion, a
406   technical note. J Robot Surg 2018;12:433–6. https://doi.org/10.1007/s11701-017-0752-7.

407   [18]   Kikinis R, Pieper SD, Vosburgh KG. 3D Slicer: A Platform for Subject-Specific Image
408   Analysis, Visualization, and Clinical Support. Intraoperative Imaging Image-Guided Ther.,
409   Springer New York; 2014, p. 277–89. https://doi.org/10.1007/978-1-4614-7657-3_19.

410   [19]   He K, Zhang X, Ren S, Sun J. Deep Residual Learning for Image Recognition. Proc IEEE
411   Comput Soc Conf Comput Vis Pattern Recognit 2015;2016-December:770–8.

412   [20]   Czempiel T, Paschali M, Keicher M, Simson W, Feussner H, Kim ST, et al. TeCNO: Surgical
413   Phase Recognition with Multi-Stage Temporal Convolutional Networks n.d.

414   [21]   Hochreiter S, Schmidhuber J. Long Short-Term Memory. Neural Comput 1997;9:1735–80.
415   https://doi.org/10.1162/NECO.1997.9.8.1735.

416   [22]   Gao X, Jin Y, Long Y, Dou Q, Heng PA. Trans-SVNet: Accurate Phase Recognition from
417   Surgical Videos via Hybrid Embedding Aggregation Transformer. Lect Notes Comput Sci
418   (Including Subser Lect Notes Artif Intell Lect Notes Bioinformatics) 2021;12904 LNCS:593–
419   603. https://doi.org/10.1007/978-3-030-87202-1_57/TABLES/3.

420   [23]   Deng J, Dong W, Socher R, Li L-J, Kai Li, Li Fei-Fei. ImageNet: A large-scale hierarchical
421   image database 2010:248–55. https://doi.org/10.1109/CVPR.2009.5206848.

[24] Funke I, Jenke A, Mees ST, Weitz J, Speidel S, Bodenstedt S. Temporal coherence-based
self-supervised learning for laparoscopic workflow analysis. Lect Notes Comput Sci
(Including Subser Lect Notes Artif Intell Lect Notes Bioinformatics) 2018;11041 LNCS:85–
93. https://doi.org/10.1007/978-3-030-01201-4_11.

[25] Goroshin R, Bruna J, Tompson J, Eigen D, LeCun Y. Unsupervised Learning of
Spatiotemporally Coherent Metrics. Proc IEEE Int Conf Comput Vis 2015;2015 Inter:4086–
93.

[26] Jayaraman D, Grauman K. Slow and steady feature analysis: higher order temporal
coherence in video. Proc IEEE Comput Soc Conf Comput Vis Pattern Recognit 2015;2016-
December:3852–61.

[27] Loshchilov I, Hutter F. Decoupled Weight Decay Regularization. 7th Int Conf Learn
Represent ICLR 2019 2017. https://doi.org/10.48550/arxiv.1711.05101.

[28] Chen L-C, Papandreou G, Schroff F, Adam H. Rethinking Atrous Convolution for Semantic
Image Segmentation. ArXiv 2017. https://doi.org/10.48550/arxiv.1706.05587.

[29] Lin TY, Maire M, Belongie S, Hays J, Perona P, Ramanan D, et al. Microsoft COCO:
Common Objects in Context. Lect Notes Comput Sci (Including Subser Lect Notes Artif
Intell   Lect   Notes   Bioinformatics)   2014;8693   LNCS:740–55.
https://doi.org/10.48550/arxiv.1405.0312.

[30] Leger S, Zwanenburg A, Pilz K, Lohaus F, Linge A, Zöphel K, et al. A comparative study of
machine learning methods for time-to-event survival data for radiomics risk modelling. Sci
Rep 2017;7:11. https://doi.org/10.1038/s41598-017-13448-3.

[31] Anteby R, Horesh N, Soffer S, Zager Y, Barash Y, Amiel J, et al. Deep learning visual
analysis in laparoscopic surgery: a systematic review and diagnostic test accuracy meta-
analysis. Surg Endosc 2021;35:1521–33. https://doi.org/10.1007/S00464-020-08168-1.

[32] Garrow CR, Kowalewski KF, Li L, Wagner M, Schmidt MW, Engelhardt S, et al. Machine
Learning for Surgical Phase Recognition: A Systematic Review. Ann Surg 2021;273:684–
93. https://doi.org/10.1097/SLA.0000000000004425.

[33] Tokuyasu T, Iwashita Y, Matsunobu Y, Kamiyama T, Ishikake M, Sakaguchi S, et al.
Development of an artificial intelligence system using deep learning to indicate anatomical
landmarks during laparoscopic cholecystectomy. Surg Endosc 2020 354 2020;35:1651–8.
https://doi.org/10.1007/S00464-020-07548-X.

[34] Mascagni P, Alapatt D, Sestini L, Altieri MS, Madani A, Watanabe Y, et al. Computer vision
in surgery: from potential to clinical value. Npj Digit Med 2022 51 2022;5:1–9.
https://doi.org/10.1038/s41746-022-00707-5.

[35] Maier-Hein L, Vedula SS, Speidel S, Navab N, Kikinis R, Park A, et al. Surgical data science
for next-generation interventions. Nat Biomed Eng 2017. https://doi.org/10.1038/s41551-

017-0132-7.

[36]   Meireles OR, Rosman G, Altieri MS, Carin L, Hager G, Madani A, et al. SAGES consensus
460   recommendations on an annotation framework for surgical video. Surg Endosc 2021
461   2021;1:1–12. https://doi.org/10.1007/S00464-021-08578-9.

462   [37]   Kumazu Y, Kobayashi N, Kitamura N, Rayan E, Neculoiu P, Misumi T, et al. Automated
463   segmentation by deep learning of loose connective tissue fibers to define safe dissection
464   planes   in   robot-assisted   gastrectomy.   Sci   Rep   2021;11:1–10.
465   https://doi.org/10.1038/s41598-021-00557-3.

466   [38]   Ward TM, Fer DM, Ban Y, Rosman G, Meireles OR, Hashimoto DA. Challenges in surgical
467   video   annotation.   Comput   Assist   Surg   2021;26:58–68.
468   https://doi.org/10.1080/24699322.2021.1937320.

469   [39]   Reinke A, Tizabi MD, Sudre CH, Eisenmann M, Rädsch T, Baumgartner M, et al. Common
470   Limitations   of   Image   Processing   Metrics: A   Picture   Story.   ArXiv   2021.
471   https://doi.org/10.48550/arxiv.2104.05642.

472   [40]   Zhang Y, Mehta S, Caspi A. Rethinking Semantic Segmentation Evaluation for
473   Explainability and Model Selection 2021.

474   [41]   Sylolypavan A, Sleeman D, Wu H, Sim M. The impact of inconsistent human annotations
475   on   AI   driven   clinical   decision   making.   Npj   Digit   Med   2023;6:1–13.
476   https://doi.org/10.1038/s41746-023-00773-3.

477   [42]   Maier-Hein L, Eisenmann M, Sarikaya D, März K, Collins T, Malpani A, et al. Surgical data
478   science – from concepts toward clinical translation. Med Image Anal 2022;76:102306.
479   https://doi.org/10.1016/J.MEDIA.2021.102306.

480   [43]   Kwon Y, Won JH, Kim BJ, Paik MC. Uncertainty quantification using Bayesian neural
481   networks in classification: Application to biomedical image segmentation. Comput Stat Data
482   Anal 2020;142:106816. https://doi.org/10.1016/J.CSDA.2019.106816.

483   [44]   Paszke A, Gross S, Massa F, Lerer A, Bradbury Google J, Chanan G, et al. PyTorch: An
484   Imperative Style, High-Performance Deep Learning Library 2019.

ACKNOWLEDGMENTS AND FUNDING

486
FRK, SL, JW, and SS were supported through project funding within the Else Kröner Fresenius
Center for Digital Health (EKFZ), Dresden, Germany (project "CoBot"). FRK received funding from
the Medical Faculty of the Technical University Dresden within the MedDrive Start program (grant
number 60487) and from the Joachim Herz Foundation (Add-On Fellowship for Interdisciplinary
Life Science). FMR received a doctoral student scholarship from the *Carus Promotionskolleg*
Dresden. Figures were created with BioRender.com. The authors acknowledge excellent project
coordination by Dr. Elisabeth Fischermeier and Dr. Grit Krause-Jüttler.
494

AUTHOR CONTRIBUTIONS

FRK, SL, JW, MD, and SS conceptualized the study. FRK, MC, FMR, and TN collected and
annotated clinical and video data and contributed to data analysis. SL, SK, and SB implemented
and trained the neural networks and contributed to data analysis. JK gave substantial clinical input.
JW, MD, JF, and SS supervised the project, provided infrastructure and gave important scientific
input. FRK drafted the initial manuscript text. All authors reviewed, edited, and approved the final
manuscript.
502

COMPETING INTERESTS

504 The authors declare no competing interests.
505

DATA AVAILABILITY

507 The datasets generated and analysed during the current study are available from the
508 corresponding author (Prof. Stefanie Speidel, stefanie.speidel@nct-dresden.de) on reasonable
509 request. To gain access, data requestors will need to sign a data access agreement.
510

CODE AVAILABILITY

512 The developed framework in this current study based on the DeepLabv3 [28] using the PyTorch
513 library [44]. The source code for surgical phase recognition and the source code for image
514 segmentation are publicly available: https://gitlab.com/nct_tso_public/cobot_phase_recognition
515 (phase recognition) and https://gitlab.com/nct_tso_public/cobot-multi-instance-segmentation
516 (structure segmentation).
517