Comparison of RT-qPCR and Digital PCR Methods for Wastewater-Based Testing of SARS-CoV-2

3

4 Adrian Hinkle¹, Hannah D. Greenwald¹, Matthew Metzger¹, Melissa Thornton¹, Lauren C.

5 Kennedy¹, Kristin Loomis², Monica B Herrera², Raymond-John Abayan², Kara L. Nelson¹, and

- 6 Rose S. Kantor¹*
- 7

8 1 Department of Civil and Environmental Engineering, University of California, Berkeley, CA,
9 USA

- 10 2 Bio-Rad Laboratories, Pleasanton, California, USA
- 11 *Corresponding author: <u>rkantor@berkeley.edu</u>

12 Abstract

13 Wastewater-based epidemiology is an important tool for monitoring SARS-CoV-2 and other 14 molecular targets in populations, using wastewater as a pooled sample. We compared the 15 sensitivity, susceptibility to inhibition, and quantification of reverse transcription quantitative PCR 16 (RT-qPCR), microfluidic well digital RT-PCR (RT-dPCR), and droplet digital RT-PCR (RT-17 ddPCR) measurements of SARS-CoV-2 (N1 gene target) and Pepper Mild Mottle Virus (PMMoV) RNA in 40 wastewater RNA extracts. All three methods were highly sensitive, but 18 19 appeared less accurate at very low concentrations. Lower inhibition was observed for RT-20 ddPCR than RT-qPCR with both SARS-CoV-2 and PMMoV targets, but inhibition appeared to 21 be mitigated by dilution of template RNA. The concentrations of N1 and PMMoV from all three 22 methods were significantly correlated (Pearson's r=0.97-0.98 for N1 and r=0.89-0.93 for 23 PMMoV), although RT-qPCR reported higher concentrations than digital methods. Taken 24 together, this study provides support for the application of all three methods in wastewater-25 based epidemiology, with additional guidelines for the use of RT-qPCR.

26 Impact Statement

PCR-based assays are the current standard for sensitive, specific, rapid pathogen quantification in environmental samples, including wastewater. The increased availability of multiple digital PCR technologies necessitates side-by-side comparison between platforms, including traditional qPCR, to guide the application of these methods. Specifically, this work can inform interpretation of wastewater SARS-CoV-2 PCR data, as reported to public health agencies for pandemic response.

33 **1. Introduction**

34 SARS-CoV-2 wastewater-based epidemiology (WBE) is now widely applied as a means 35 to monitor the spread of COVID-19 in communities (1,2). WBE may become more critical for 36 providing accurate COVID-19 prevalence information as vaccines have become widely available 37 in the U.S. and at-home testing has replaced clinical testing. WBE for human pathogens

38 typically involves the extraction of nucleic acids from raw wastewater and PCR-based 39 quantification of an RNA or DNA target. While a national database has been established for 40 SARS-CoV-2 WBE in the United States (CDC-NWSS), there remains no standardized protocol, 41 and laboratories are often limited by the equipment already available to them. Interlaboratory 42 comparisons have shown one or more orders of magnitude differences in results reported by 43 groups using different methods (3-5). While much work has focused on the comparison of viral 44 concentration and RNA extraction methods from wastewater, there is limited data to specifically 45 address differences introduced by the choice of quantification platform.

46 Three commonly-used reverse transcription polymerase chain reaction (RT-PCR) 47 methods that have been applied for WBE are quantitative PCR (qPCR), fixed array-based digital 48 PCR (dPCR), and droplet digital PCR (ddPCR). While qPCR is based on whole-sample real-49 time quantification relative to a standard curve, digital PCR methods rely on sample digitization 50 or partitioning and use end-point detection to determine the number of positive microchambers 51 or partitions. A Poisson distribution is used to calculate the initial concentration of the target in 52 the sample based on the fraction of positive microchambers or partitions. Commonly cited 53 advantages of dPCR and ddPCR include robustness to inhibition, high sensitivity, high reproducibility, and no need for a standard curve. However, droplet digital and digital PCR have 54 55 somewhat more limited ability to measure high-concentration samples, meaning a priori 56 knowledge is required for adequate dilution (5), and manual thresholding is sometimes required 57 to determine separation between positive and negative droplets (for ddPCR) (6). While digital PCR is becoming more widespread, it is still out of reach for many laboratories due to the high 58 59 initial cost of equipment and longer turnaround time from sample to results. Fixed array platebased dPCR is technically more straightforward and faster to perform because it removes the 60 61 droplet formation step, but it has lower throughput compared to ddPCR, and few studies have 62 directly compared this technique to ddPCR and gPCR.

63 With the increased adoption of digital PCR platforms, there have been several extensive 64 comparisons of ddPCR and qPCR applied for sensitive quantification of microorganisms in the 65 environment (3-5). As SARS-CoV-2 WBE developed, many researchers performed 66 methodological comparisons that included side-by-side tests of ddPCR and gPCR for SARS-67 CoV-2 gene targets (7–11). Overall, higher quantities were determined by qPCR than ddPCR, 68 and, when measured, the correlation between results from the two platforms was reasonable 69 (Table S1). However, conclusions were often limited by the small numbers of samples tested. 70 Several investigations found that ddPCR was more sensitive and less likely to experience 71 inhibition than qPCR (7,9,12,13), while one study suggested that inhibition of reverse 72 transcription was greater in ddPCR than qPCR (11). In the most extensive of these studies, 73 Ciesielski et al. (7) quantified SARS-CoV-2 in 63 wastewater samples with ddPCR and gPCR 74 using the US CDC N2 assay. Fewer than half of the samples were above the defined limits of 75 detection, and while inhibition was observed, the extent of inhibition in each platform was not 76 determined. Just one study compared quantification of SARS-CoV-2 in wastewater using qPCR 77 and fixed array-based dPCR (8). Notably, the N2 assay was previously found to be more 78 sensitive in ddPCR (14), while the CDC N1 assay was more sensitive for qPCR (15,16).

Given the importance of accurate, reproducible WBE data despite differences in
 laboratory equipment, we sought to determine the comparability of three quantification
 methods-RT-qPCR, RT-ddPCR, and RT-dPCR—with RNA from northern California wastewater

82 samples. We applied SARS-CoV-2 (CDC N1 diagnostic assay) and Pepper Mild Mottle Virus 83 (PMMoV; fecal strength indicator (17–19)) assays on all three platforms. This study includes a 84 direct comparison of 40 samples across three PCR platforms and a summary of the 85 comparisons performed thus far in the literature. Sampling locations ranged from large 86 treatment plants to residential buildings, allowing testing of a wide range of concentrations, 87 sensitivity near the limit of detection, and the impact of inhibition.

88 2. Methods

89 **2.1 Sample collection and RNA extraction**

90 Forty raw wastewater samples were collected from locations across the San Francisco Bay Area between December 3rd and December 10th, 2020. In total, 30 unique locations were 91 92 represented in the sample set, including wastewater treatment plant influents, subsewersheds, and residential buildings or campuses. For each sample, 40 mL of raw wastewater from a 24-93 94 hour composite sample was transferred to a 50 mL tube containing 9.35 g sodium chloride and 95 400 µL TE buffer (1 M TRIS, 100 mM EDTA). The sample tubes were shipped overnight with an 96 ice pack to the University of California, Berkeley (UC Berkeley), and total nucleic acids were 97 extracted within 48 hours of sampling. Nucleic acid extraction followed the 4S method (20). 98 Briefly, the samples were heated at 70°C for 45 minutes, then filtered through 5 µm PVDF filters. mixed 1:1 with 70% vol/vol ethanol, bound to Zymo III-P columns (ZymoResearch), 99 100 rinsed with wash buffers, and eluted with 200 µL ZymoPURE elution buffer. The resulting 200 101 µL of nucleic acid eluate for each sample was divided into four 50 µL aliguots and stored in 102 LoBind tubes (Eppendorf) at -80°C.

103 Sample collection and extraction complied with the Environmental Microbiology 104 Minimum Information (EMMI) guidelines (21). Collection was part of routine weekly monitoring 105 conducted by the UC Berkeley COVID-WEB wastewater surveillance project. There was no 106 evidence of contamination during sample collection in 10 months of sampling prior to this study. 107 Therefore, environmental sampling controls were appropriately deemed unnecessary (21). Each 108 sample was spiked with one of two amounts of the enveloped ssRNA virus Murine Hepatitis Virus (MHV, ATCC) (to a total spike amount of either 7.08 x 10⁴ gene copies or 7.08 x 10⁵ gene 109 110 copies per 40 mL of raw wastewater) prior to heat inactivation. MHV served as a matrix 111 recovery proxy to enable estimation of extraction efficiency of SARS-CoV-2. The post-extraction 112 concentration was quantified using ddPCR to generate an extraction efficiency. A 40-mL 113 phosphate buffered saline (1x PBS) negative extraction control was included with the batch of 114 extractions on December 8th, 2020, and was quantified on one N1 plate and one PMMoV plate 115 using RT-gPCR. No amplification of N1 was detected in the negative control and amplification of 116 PMMoV was minimal (Cg=39.7).

117 2.2 RT-qPCR

After overnight storage at -80 °C, samples were thawed on ice and processed with RT-qPCR on
a QuantStudio3 Real-Time PCR System qPCR machine (Thermo Fisher) at the University of
California, Berkeley. The One-Step RT-qPCR for SARS-CoV-2 Wastewater Surveillance
protocol (22) was followed. Briefly, each reaction contained 15 μL of reaction mix and 5 μL of

template. The template consisted of undiluted RNA or five-fold diluted RNA in PCR water. The
reaction mix contained TaqMan Fast Virus 1-Step Master Mix (Thermo Fisher Catalog
#4444434) and primers and probes (Integrated DNA Technologies) in PCR water (see
Greenwald et al. for sequences and concentrations (22). Methods complied with the Minimum
Information for Publication of Quantitative Real-Time PCR Experiments (MIQE) guidelines (23).

127 The CDC N1 diagnostic assay was used for SARS-CoV-2 (24) and the coat protein gene 128 was used for PMMoV (25). Both were singleplex assays with FAM-labeled probes. The N1 129 standard was made using the nucleocapsid (N) gene plasmid (Integrated DNA Technologies) 130 grown in E. coli, digested to make a linearized plasmid stock, and guantified via Qubit, and 131 Absolute Q (dPCR; see below). The PMMoV standard used on the first four plates of this study 132 was an RNA ultramer (IDT), but it was found to have degraded (Figure S1). The standard used 133 on the fifth plate was a linear DNA fragment (gBlock). For analysis, all PMMoV data were 134 analyzed by applying the standard curve equation derived from the plate with the DNA standard 135 (plate 1138; see Data Analysis).

136 Each of the six N1 and five PMMoV RT-gPCR plates contained samples, no-template 137 controls (PCR water), and a standard curve, in triplicate wells (Table S2). All no-template 138 controls were negative for all plates. The N1 standard plasmid was diluted in TE buffer to a stock concentration of approximately 1×10^7 gene copies per microliter (gc/µL), later quantified 139 140 by dPCR (see below). Serial dilutions in PCR water were performed to generate a seven point 141 standard curve (5, 10, 20, 100, 1000, 10000, and 10000 gc/well, before adjusting based on the dPCR quantification of the standard). For N1 plates, standard curve efficiency ranged from 85% 142 to 93% and R² ranged from 0.90 to 1.0. A seven point standard curve was also used for PMMoV 143 (10² to 10⁸ gc/well). For PMMoV plates, the standard curve efficiency ranged from 84% to 122% 144 and R² ranged from 0.92 to 1.0. The concentration for each sample was determined by 145 146 averaging the Cq values from all technical replicates that amplified and calculating a 147 concentration based on the standard curve from each plate. Before averaging, outlier Cg values 148 were removed from groups of technical triplicates using Grubb's test (alpha = 0.05).

The limit of detection (LOD) for the N1 assay was determined to be 0.49 gc/uL RNA (2.43 gc per well), the point at which \ge 95% of technical replicates were positive on 6 standard curves each run in triplicate (**Table S3**). The limit of detection for other assays was not investigated. Samples were called positive if at least 2 of 3 replicate wells amplified and the average concentration was above the LOD.

154 2.3 RT-dPCR

Extracted RNA samples were stored at -80 °C at UC Berkeley for one day past the last day of 155 156 sampling, then hand-carried on dry ice to Combinati (Palo Alto, California) and stored at -80 °C 157 for 50 additional days. Samples were thawed on ice and processed using RT-dPCR by 158 Combinati. Steps complied with the digital MIQE guidelines (26.27). Each 9 µL reaction consisted of 1 µL of RNA, 2.25 µL of 4X Combinati 1-step RT-dPCR MasterMix, 0.45 µL of 159 SARS-CoV-2 Wastewater Surveillance 4-plex assay (Combinati), and 5.3 µL water. Each plate 160 161 contained negative controls (PCR water) and a positive control. The positive 4-target PCR 162 control was made of synthetic single-stranded DNA containing the target N1, N2, and PMMoV 163 sequences plus extracted Bovine Coronavirus RNA (PBS Animal Health).

After preparing the dPCR mix, 9 μ L of the reaction mixture was loaded into the MAP16 plate followed by an overlay of 15 μ L of isolation buffer. The prepared MAP16 plate was then loaded on the Absolute Q instrument (Applied Biosystems QuantStudio Absolute Q Digital PCR system). Thermocycling was performed using Absolute Q with the following program: reverse transcription at 50 °C for 10 minutes, preheating (enzyme activation) at 95 °C for 10 minutes, and 45 cycles of denaturation at 95 °C for 5 seconds and annealing/extension at 55 °C for 30 seconds.

171 The limit of detection and limit of quantification used for the assay were both two positive 172 microchambers per reaction ($\sim 0.2 \text{ gc/}\mu\text{L}$). These values are the same because digital PCR 173 provides absolute quantification without the need for a standard curve. Data analysis was 174 performed on the Absolute Q Analysis Software (v4.2.1), which reports the concentration of 175 each reaction in gene copies per microliter ($gc/\mu L$) (**Table S4**). The initial sample concentration 176 was manually calculated by adjusting the reported dPCR concentration based on the input 177 sample volume for each sample. The arithmetic mean concentration of the duplicate 178 measurements was calculated and reported for each sample. Samples were determined to be 179 positive if at least 3 total microchambers were positive across duplicate reactions.

180 2.4 RT-ddPCR

181 Extracted RNA samples were stored at -80 °C at UC Berkeley for 41 days past the last day of 182 sampling, then hand-carried on dry ice to Bio-Rad (Pleasanton, California) and stored at -80 °C 183 for two additional days. Samples were thawed on ice and processed using the One-Step RT-184 ddPCR Advanced Kit for Probes (Bio-Rad); 9 µL was used for a 20 µL reaction volume and 185 samples were analyzed without technical replicates. Each droplet was 0.85 nL. Steps complied 186 with the digital MIQE guidelines (26,27). The primer and probe concentration in the reaction was 187 900 nM and 250 nM, respectively. Assays were purchased from IDT and manufactured under 188 SARS-CoV-2 template-free conditions. The CDC N1 target for SARS-CoV-2 (HEX-labeled 189 probe) (24) and PMMoV target (FAM-labeled probe) were run in singleplex reactions. The CDC 190 N2 (24), E (28), and MHV targets (FAM-, HEX- and FAM- mixture, and HEX-labeled probes, 191 respectively) were run as a triplex reaction (Table S5). Each reaction was dropletized using the 192 QX200 Auto DG (Bio-Rad). Thermal cycling was performed using a C1000 Touch (Bio-Rad) 193 with the following program: reverse transcription at 50 °C for 60 minutes, enzyme activation at 194 95 °C for 10 minutes, 40 cycles of denaturation at 94 °C for 30 seconds and 195 annealing/extension at 55 °C for 1 minute, enzyme deactivation at 98 °C for 10 minutes, and a 196 final hold at 4 °C. The ramp rate was set to 2 °C/second.

Samples were analyzed without dilution and at a 1:5 dilution (7 µL diluted into a final volume of 35 µl with nuclease free water). Synthetic SARS-CoV-2 RNA (Exact Diagnostics) was used as a positive control on all runs. Nuclease free water was used as a no-template control on all runs (1-2 wells per run) and all negative controls were negative.

Data were analyzed using the QXManager v1.2. All wells were thresholded manually and were confirmed to have appropriate droplet counts (**Table S6**). A sample was interpreted as

positive if the ddPCR concentration for N1 was ≥ 0.1 gc/µL (RNA sample concentration of 0.22

204 gc/µL) and there were two or more positive droplets per sample (no technical replicates were

performed). This limit of detection was determined using a serially diluted heat-inactivated virus
 to establish the lowest detectable concentration of SARS-CoV-2 at which at least 95% of true
 positive replicates tested positive.

208 **2.5 Quantification of qPCR standards by dPCR**

209 Two aliquots of the RT-qPCR standards for N1 and PMMoV were hand carried on dry ice to 210 Combinati for quantification. The standards were diluted 1000-fold, 10,000-fold, and 100,000-211 fold before quantification using the Absolute Q digital PCR platform. Each dPCR reaction 212 consisted of 1 µL of diluted standard material, 2.25 µL of 4X Combinati 1-step RT-dPCR 213 MasterMix, 0.45 µL of 4-plex Wastewater assay, and 5.3 µL water. Each dilution point was 214 quantified in duplicate. After correcting the reported concentration based on the appropriate 215 dilution factor, the absolute concentration of the standard material was calculated using the 216 average values of both tubes across the dilution series (Table S7). The N1 (linearized plasmid) 217 and PMMoV (dsDNA gBlock) standard concentrations used in data analysis were based on 218 quantification of the standards via Absolute Q dPCR (Table S2).

219 2.6 Data Analysis

220 Data analysis was performed using Python (v3.9.5) using modules Pandas (v1.2.5), NumPy 221 (v1.21.0), SciPy (v1.7.0), and Plotnine (v0.8.0). RT-qPCR analysis of standard curves and 222 unknowns performed was using custom code as previously described (29) 223 (https://github.com/wastewaterlab/data analysis). Pearson correlation coefficients were 224 determined using SciPy, and linear regressions were conducted using NumPy polyfit. Paired 225 concentrations of diluted and undiluted samples were compared using the Mann-Whitney U test 226 (SciPy) to determine whether there was significant inhibition during RT-PCR. All analysis code 227 can be found at

228 https://github.com/wastewaterlab/data_analysis/blob/master/notebooks/pcr_comparison.ipynb.

229 3. Results

230 A total of 40 raw wastewater samples from 30 distinct residential, subsewershed, and treatment 231 plant influent locations in the San Francisco Bay Area underwent large-volume nucleic acid 232 extraction at UC Berkeley. Quantification of SARS-CoV-2 (CDC N1 diagnostic assay) and 233 PMMoV (coat protein gene) RNA was performed by three laboratories using qPCR, dPCR, and 234 ddPCR, respectively (Table S8). Undiluted RNA was quantified for all samples, while five-fold 235 diluted RNA was also quantified with qPCR and ddPCR to assess inhibition. All three methods 236 produced similar trends across samples that ranged three orders of magnitude in concentration 237 (Figures 1 and S2).

238



239

Figure 1. Comparison of SARS-CoV-2 and PMMoV concentrations in RNA from 40 wastewater samples from the San Francisco Bay Area. Samples are ordered by average N1 concentration rank for each method across all methods. Each sample was measured undiluted via dPCR and both undiluted and fivefold diluted via ddPCR and qPCR. The N1 limit of detection is shown for each method with a dashed line; all non-detects are plotted as open points at the LOD for each method. Concentrations shown for diluted qPCR and ddPCR samples are corrected for the dilution factor.

246 3.1 Sensitivity

247 The limits of detection for the N1 assay with qPCR, dPCR and ddPCR were found to be 0.5, 248 0.2, and 0.22 gene copies per microliter (qc/µL) RNA, respectively (see Methods). The 249 sensitivity of each platform is also impacted by PCR inhibition (see below) and by the volume of 250 template RNA included in the PCR reaction. For a single reaction well, qPCR used 5 µL, dPCR 251 used 1 µL, and ddPCR used 9 µL. Additionally, measuring the same sample in technical 252 replicate wells may allow for improved sensitivity. Here, gPCR utilized 3 wells, dPCR utilized 2 253 wells, and ddPCR utilized 1 well (no replicates). Of 40 samples compared with the N1 assay 254 (undiluted), there was total agreement on 2 negatives and 29 positives across all platforms 255 (Table S8). ddPCR and dPCR jointly reported 1 additional negative, gPCR reported 1 additional 256 negative, and dPCR reported 7 unique negatives (Figure 2). All but one of the negatives were 257 samples from residential sewersheds (buildings or campuses). Notably, the Murine Hepatitis 258 Virus (MHV) matrix recovery control had >50% calculated recovery efficiency for all but five 259 samples. Of those five, two were non-detects with N1 in dPCR only. This suggests that, overall, 260 non-detects were not due to poor extractions.

261



262

Figure 2. Venn diagram of the samples with non-detects in one or more PCR method. Samples are depicted as squares (a single WWTP influent site) or circles (residential sites).

265 **3.2 Inhibition**

Next, we assessed inhibition, which can lead to under-quantification of the molecular target or 266 267 false negative results (15.30). Inhibition is especially a concern in WBE because of the inhibitory 268 substances found in wastewater and because inhibition impedes comparison between 269 laboratories that use different molecular methods (31). To test for inhibition, we compared the 270 results of qPCR and ddPCR with and without prior dilution of the RNA template (21,32). If 271 inhibitors were present, dilution would decrease their concentration in the PCR reaction, 272 resulting in higher reported concentrations after accounting for the dilution factor. Five-fold dilution was chosen after prior qPCR testing of 5-, and 10-fold dilutions revealed that 10-fold 273 274 often led to loss of signal (17). We quantified N1 and PMMoV in undiluted and five-fold diluted 275 samples using qPCR and ddPCR (Figure 1) and calculated the ratio between diluted and 276 undiluted concentrations (Figure 3). Median diluted-to-undiluted ratios for ddPCR and qPCR were 1.15 vs. 1.82 for N1 and 1.17 vs. 1.88 for PMMoV, indicating that qPCR results were 277 278 significantly more affected by inhibition (two-sided Mann-Whitney U test p=1.2x10⁻⁸ for n=31 comparisons of N1 and $p=3.6x10^{-10}$ for n=33 comparisons of PMMoV). 279



280

Figure 3. Ratio of target quantification in five-fold diluted sample relative to undiluted sample for ddPCR and qPCR. The concentrations for the diluted samples were corrected for the dilution factor (multiplied by 5) prior to calculating the ratio. Samples with non-detects are not shown. The boxes show the first and third quartiles and the whiskers extend to the largest value within 1.5 times the interquartile range. Individual data points are shown in gray and samples are connected by gray lines.

286 3.3 Quantification

287 We found that N1 and PMMoV results from all three methods were linearly correlated after 288 log10 scaling with significant Pearson correlation coefficients for each pair of methods (Table 1, 289 Figures S3 and S4). Correlations were determined using the data from 5-fold dilutions of gPCR 290 and ddPCR due to the presence of inhibition in some samples (see above). Samples that were 291 non-detect with any method were removed prior to calculation. Despite the strong correlation, 292 undiluted qPCR concentrations were higher than dPCR concentrations (median fold difference 293 = 1.7) and ddPCR concentrations (median fold difference = 1.8). Notably, qPCR standards were 294 quantified by dPCR, which reported concentrations for N1 and PMMoV that were, respectively, 295 60.8% and 39.2% lower than those found with Qubit (Table S7). This difference was accounted 296 for during data analysis.

297

298 Table 1. Correlation statistics and linear regressions for log10-scaled data from each pair of PCR 299 methods.

Target	Independent	Dependent	n	Slope	Intercept	Pearson's r*	Spearman r*
N1	qPCR	dPCR	27	1.06	-0.60	0.98	0.97
	qPCR	ddPCR	27	0.93	-0.27	0.97	0.95
	dPCR	qPCR	27	0.90	0.63	0.98	0.97
	dPCR	ddPCR	27	0.86	0.29	0.97	0.97
	ddPCR	qPCR	27	1.02	0.38	0.97	0.95
	ddPCR	dPCR	27	1.11	-0.24	0.97	0.97
PMMoV	qPCR	dPCR	40	1.04	-0.46	0.90	0.90
	qPCR	ddPCR	40	1.03	-0.28	0.93	0.91
	dPCR	qPCR	40	0.78	1.13	0.90	0.90
	dPCR	ddPCR	40	0.85	0.68	0.89	0.94
	ddPCR	qPCR	40	0.84	0.81	0.93	0.91
	ddPCR	dPCR	40	0.93	0.16	0.89	0.94

300

* All correlations were significant ($p < 1.5 \times 10^{-12}$)

301 4. Discussion

All three PCR methods produced quantitative measurements of N1 and PMMoV RNA 302 303 concentrations across three orders of magnitude (Figure 1), and concentrations were strongly 304 correlated between methods (**Table 1**). Despite quantification of the qPCR standards with dPCR 305 prior to analysis, measured qPCR concentrations were consistently higher than dPCR and 306 ddPCR concentrations, which were in closer agreement. This trend has also been observed in 307 previous studies (Table S1), indicating a need to normalize concentrations before comparisons 308 can be made between measurements using different methods. One possible explanation for this 309 discrepancy is that qPCR standard curves may be affected by DNA adherence to the walls of 310 plastic tubing or degradation of DNA samples, leading to overestimation of the DNA standard, 311 and subsequent over-quantification of the wastewater sample. Conversely, these same factors 312 should affect the sample nucleic acids as well, resulting in potential under-quantification of 313 samples by ddPCR and dPCR. Another possible explanation is that the reverse transcription 314 step in RT-qPCR could have produced more than one cDNA molecule per RNA template, 315 resulting in over-quantification of the sample material (RNA) but not the DNA standards. 316 Meanwhile, the reverse transcription step in digital PCR methods was conducted within the 317 partitions/microchambers, and a proportional increase in cDNA would not affect the readout of 318 these endpoint PCR assays.

Also consistent with prior work (**Table S1**), inhibition was significantly lower in ddPCR than qPCR based on the lower diluted-to-undiluted ratio in ddPCR (**Figure 3**). Nonetheless, inhibition in qPCR does not appear to have led to many false negatives relative to digital PCR methods (**Figure 2**). Additionally, while inhibition was not measured with dPCR, similar results

would be expected given that both digital PCR methods measure the presence or absence of
 DNA template in individual microchambers and are therefore less impacted by the decrease in
 amplification efficiency caused by PCR inhibitors (12,33).

Lastly, limits of detection measurements are a function of the total sample volume included in the reaction and for digital PCR, the number partitions or microchambers measured, but these are not standardized across platforms. Additionally, the method for defining the limit of detection was different for each platform. Thus, conclusions comparing the sensitivity of each platform based on the defined LoDs should be taken with care.

331 5. Conclusion

332 All three methods are highly sensitive and quantitative, making them appropriate tools for use in 333 WBE. However, based on our findings, additional guidelines for RT-qPCR should be followed 334 when this method is used for WBE. First, to increase the accuracy of RT-qPCR, dilutions should 335 be performed when there are signs of inhibition. Undiluted samples should also be run in 336 parallel in case dilution causes the signal to drop below detectable levels. The same should be 337 performed for digital PCR, if sample volumes and logistical constraints allow. Second, the 338 differences between Qubit and fixed array-based digital PCR measurements of the DNA 339 standard in this study (see Methods) point to the importance of the method of quantification for 340 gPCR standards. We suggest that WBE databases such as CDC-NWSS should track how 341 qPCR standards are quantified, and this should be reported in all qPCR-based WBE studies. 342 For internal consistency, long-term monitoring projects should limit changes to standards and 343 carefully compare new batches of standards as part of quality control.

344 Third, we have found that results of long term-monitoring using RT-qPCR may be 345 subject to plate-to-plate fluctuations due to differences in standard curve serial dilutions (29). While standard curve efficiency and R² are key components of quality control, the v-intercept 346 347 should also be monitored as an indicator of standard degradation or batch differences. Finally, 348 we note that all methods are less accurate at low concentrations due to stochasticity (whether 349 the RNA molecules are captured in the subsample that is taken for PCR analysis). Here, testing 350 more technical replicates may improve accuracy, and replication may be most important for 351 presence/absence determinations in residential facilities. We also expect that including more 352 RNA template in the PCR reaction would improve accuracy. The sample volume included in the 353 PCR reactions was a major difference between the platforms, as ddPCR utilized 9 µL, RR-354 qPCR utilized 5 μL, and dPCR utilized 1 μL. This may explain why SARS-CoV-2 was detected 355 in more samples using RT-gPCR and ddPCR compared to dPCR.

356 Digital PCR methods are advantageous for their ability to rapidly bring online multiplexed 357 assays for new SARS-CoV-2 mutation detection or other pathogens. Additionally, the absolute 358 quantification and increased tolerance to inhibition of dPCR and ddPCR allows higher 359 confidence in the quantitative comparison of assays for multiple targets from the same organism 360 (e.g. to calculate the percentage of total SARS-CoV-2 corresponding to a variant lineage). Despite these advantages, qPCR is a familiar, more widely available method for many 361 362 researchers globally, and it can be fast, cost-effective, and high-throughput. Overall, the data 363 presented in this study and summarized from previous work provide evidence that, when used 364 appropriately, qPCR, dPCR, and ddPCR are all suitable WBE methods that generate highly 365 correlated results for a wide range of wastewater sources and target concentrations.

366 **5. Acknowledgements**

We thank Christina Bouwens and Robert Lin (Thermo Fisher) for their collaboration to run 367 368 samples on the Absolute Q. We thank our partners at the following agencies and facilities for 369 sample collection: Sanitary District No. 5 of Marin County; Central Marin Sanitation Agency; 370 East Bay Municipal Utility District; San Francisco Public Utilities Commission; Las Gallinas 371 Sanitary District: Sewerage Agency of Southern Marin: Novato Sanitary District: Central Contra 372 Costa Sanitary District: Delta Diablo: West County Water District: Yountville: American Canvon: California Medical Facility; and University of California, Berkeley. We also thank volunteers in 373 374 the UC Berkeley COVID-WEB wastewater testing laboratory. We gratefully acknowledge 375 funding from the Catena Foundation. A.W.H., H.D.G., and L.C.K. were supported by the 376 National Science Foundation (NSF) Graduate Research Fellowship [grant number DGE-377 1752814].

378

379 6. Conflict of interest

Raymond John Abayan, Kristin Loomis, and Monica Herrera are employees of Bio-Rad
 Laboratories, which commercializes equipment and assays for ddPCR.

- 382
- 383
- 384
- 385
- 386

387 References

- Bivins A, North D, Ahmad A, Ahmed W, Alm E, Been F, et al. Wastewater-Based
 Epidemiology: Global Collaborative to Maximize Contributions in the Fight Against COVID Environ Sci Technol. 2020 Jul 7;54(13):7754–7.
- Medema G, Heijnen L, Elsinga G, Italiaander R, Brouwer A. Presence of SARS Coronavirus-2 RNA in Sewage and Correlation with Reported COVID-19 Prevalence in the
 Early Stage of the Epidemic in The Netherlands. Environ Sci Technol Lett. 2020 Jul
 14;7(7):511–6.
- 395 3. Nshimyimana JP, Cruz MC, Wuertz S, Thompson JR. Variably improved microbial source 396 tracking with digital droplet PCR. Water Res. 2019 Aug 1;159:192–202.
- Crain C, Kezer K, Steele S, Owiti J, Rao S, Victorio M, et al. Application of ddPCR for
 detection of Enterococcus spp. in coastal water quality monitoring. J Microbiol Methods.
 2021 May;184:106206.
- Zhao Y, Xia Q, Yin Y, Wang Z. Comparison of Droplet Digital PCR and Quantitative PCR
 Assays for Quantitative Detection of Xanthomonas citri Subsp. citri. PLoS ONE. 2016 Jul
 18;11(7):e0159004.
- 403 6. Kokkoris V, Vukicevich E, Richards A, Thomsen C, Hart MM. Challenges Using Droplet
 404 Digital PCR for Environmental Samples. Appl Microbiol. 2021 Jun;1(1):74–88.
- 405 7. Ciesielski M, Blackwood D, Clerkin T, Gonzalez R, Thompson H, Larson A, et al.
 406 Assessing sensitivity and reproducibility of RT-ddPCR and RT-qPCR for the quantification
 407 of SARS-CoV-2 in wastewater. J Virol Methods. 2021 Jul 9;114230.
- Boogaerts T, Jacobs L, De Roeck N, Van den Bogaert S, Aertgeerts B, Lahousse L, et al.
 An alternative approach for bioanalytical assay optimization for wastewater-based
 epidemiology of SARS-CoV-2. Sci Total Environ. 2021 Oct 1;789:148043.
- 411 9. Graham KE, Loeb SK, Wolfe MK, Catoe D, Sinnott-Armstrong N, Kim S, et al. SARS-CoV412 2 RNA in Wastewater Settled Solids Is Associated with COVID-19 Cases in a Large Urban
 413 Sewershed. Env Sci Technol. 2021;55(1):488–98.
- Steele JA, Zimmer-Faust AG, Griffith JF, Weisberg SB. Sources of variability in methods
 for processing, storing, and concentrating SARS-CoV-2 in influent from urban wastewater
 treatment plants. medRxiv. 2021 Jun 21;2021.06.16.21259063.
- 11. D'Aoust PM, Mercier E, Montpetit D, Jia JJ, Alexandrov I, Neault N, et al. Quantitative
 analysis of SARS-CoV-2 RNA from wastewater solids in communities with low COVID-19
 incidence and prevalence. Water Res. 2021 Jan 1;188:116560.
- 420 12. Dingle TC, Sedlak RH, Cook L, Jerome KR. Tolerance of Droplet-Digital PCR vs Real 421 Time Quantitative PCR to Inhibitory Substances. Clin Chem. 2013 Nov 1;59(11):1670–2.
- Rački N, Dreo T, Gutierrez-Aguirre I, Blejec A, Ravnikar M. Reverse transcriptase droplet
 digital PCR shows high resilience to PCR inhibitors from plant, soil and water samples.
 Plant Methods. 2014 Dec 31;10:42.
- 425 14. Gonzalez R, Curtis K, Bivins A, Bibby K, Weir MH, Yetka K, et al. COVID-19 surveillance in
 426 Southeastern Virginia using wastewater-based epidemiology. Water Res. 2020 Nov
 427 1;186:116296.
- Ahmed W, Simpson SL, Bertsch PM, Bibby K, Bivins A, Blackall LL, et al. Minimizing errors
 in RT-PCR detection and quantification of SARS-CoV-2 RNA for wastewater surveillance.
 Sci Total Environ. 2022 Jan 20;805:149877.
- 431 16. Kaya D, Niemeier D, Ahmed W, Kjellerup BV. Evaluation of multiple analytical methods for
 432 SARS-CoV-2 surveillance in wastewater samples. Sci Total Environ. 2021 Dec 6;152033.
- 433 17. Greenwald HD, Kennedy LC, Hinkle A, Whitney ON, Fan VB, Crits-Christoph A, et al.
 434 Tools for interpretation of wastewater SARS-CoV-2 temporal and spatial trends
 435 demonstrated with data collected in the San Francisco Bay Area. Water Res X. 2021 Aug
 436 1;12:100111.

- 437 18. Feng S, Roguet A, McClary-Gutierrez JS, Newton RJ, Kloczko N, Meiman JG, et al.
 438 Evaluation of Sampling, Analysis, and Normalization Methods for SARS-CoV-2
- Evaluation of Sampling, Analysis, and Normalization Methods for SARS-CoV-2
 Concentrations in Wastewater to Assess COVID-19 Burdens in Wisconsin Communities.
 ACS EST Water. 2021 Aug 13;1(8):1955–65.
- 19. Greaves J, Stone D, Wu Z, Bibby K. Persistence of emerging viral fecal indicators in largescale freshwater mesocosms. Water Res X. 2020 Dec 1;9:100067.
- Borchardt MA, Boehm AB, Salit M, Spencer SK, Wigginton KR, Noble RT. The
 Environmental Microbiology Minimum Information (EMMI) Guidelines: qPCR and dPCR
 Quality and Reporting for Environmental Microbiology. Environ Sci Technol. 2021 Aug
 3;55(15):10210–23.
- 451 22. Greenwald H, Kennedy LC, Fan VB, Kantor R, Nelson KL. One-Step RT-qPCR for SARS452 CoV-2 Wastewater Surveillance: N1, PMMoV, BCoV, SOC. 2020 Dec 4 [cited 2021 Jul
 453 27]; Available from: https://www.protocols.io/view/one-step-rt-qpcr-for-sars-cov-2454 wastewater-surveil-bpk3mkyn
- 455
 456
 456
 456
 456
 457
 457
 458
 459
 459
 450
 450
 450
 450
 451
 451
 452
 453
 454
 455
 455
 455
 455
 455
 455
 455
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
 457
- 458 24. Lu X, Wang L, Sakthivel SK, Whitaker B, Murray J, Kamili S, et al. US CDC Real-Time
 459 Reverse Transcription PCR Panel for Detection of Severe Acute Respiratory Syndrome
 460 Coronavirus 2. Emerg Infect Dis. 2020 Aug;26(8):1654–65.
- 461 25. Haramoto E, Kitajima M, Kishida N, Konno Y, Katayama H, Asami M, et al. Occurrence of
 462 Pepper Mild Mottle Virus in Drinking Water Sources in Japan. Appl Environ Microbiol. 2013
 463 Dec 1;79(23):7413–8.
- 464 26. Huggett JF, Foy CA, Benes V, Emslie K, Garson JA, Haynes R, et al. The Digital MIQE
 465 Guidelines: Minimum Information for Publication of Quantitative Digital PCR Experiments.
 466 Clin Chem. 2013 Jun 1;59(6):892–902.
- 467 27. The dMIQE Group, Huggett JF. The Digital MIQE Guidelines Update: Minimum Information
 468 for Publication of Quantitative Digital PCR Experiments for 2020. Clin Chem. 2020 Aug
 469 1;66(8):1012–29.
- 28. Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DK, et al. Detection of 2019
 novel coronavirus (2019-nCoV) by real-time RT-PCR. Eurosurveillance. 2020 Jan
 23;25(3):2000045.
- 473 29. Kantor RS, Greenwald HD, Kennedy LC, Hinkle A, Harris-Lovett S, Metzger M, et al.
 474 Operationalizing a routine wastewater monitoring laboratory for SARS-CoV-2. PLOS
 475 Water. 2022 Feb 15;1(2):e0000007.
- 476 30. Gibson KE, Schwab KJ, Spencer SK, Borchardt MA. Measuring and mitigating inhibition
 477 during quantitative real time PCR analysis of viral nucleic acid extracts from large-volume
 478 environmental water samples. Water Res. 2012;46(13):4281–91.
- 479 31. Mao K, Zhang K, Du W, Ali W, Feng X, Zhang H. The potential of wastewater-based
 480 epidemiology as surveillance and early warning of infectious disease outbreaks. Curr Opin
 481 Environ Sci Health. 2020;17:1–7.
- 482 32. Schrader C, Schielke A, Ellerbroek L, Johne R. PCR inhibitors occurrence, properties
 483 and removal. J Appl Microbiol. 2012;113(5):1014–26.
- 484 33. Sidstedt M, Rådström P, Hedman J. PCR inhibition in qPCR, dPCR and MPS—
- 485 mechanisms and solutions. Anal Bioanal Chem. 2020 Apr 1;412(9):2009–23.