All rights reserved. No reuse allowed without permission.

Development and Clinical Validation of a Miniaturized Finger Probe for Bedside Hemodynamic Monitoring

Tuukka Panula^{*1}, Jukka-Pekka Sirkiä¹, Tero Koivisto¹, Mikko Pänkäälä¹, Teemu Niiranen^{2,3}, Ilkka Kantola³, and Matti Kaisti^{*1}

¹Department of Computing, University of Turku, Finland ²Department of Internal Medicine, University of Turku, Finland ³Division of Turku University Hospital, Finland

August 23, 2022

$_{1}$ Abstract

Background: Hypertension, or elevated blood 2 pressure (BP), is a marker for many cardiovas-3 cular diseases and can lead to life-threatening conditions such as heart failure, coronary artery 5 disease, and stroke. BP monitoring using a tra-6 ditional arm cuff device is often inconvenient and possibly painful in long-term use, i.e. during sleep. Methods: We present a miniature cuf-9 fless tonometric finger probe system, that uses 10 the oscillometric method to measure blood pres-11 sure (BP). Our approach uses a motorized press 12 that is used to apply pressure to the finger tip to 13 measure BP. *Results*: We verified the functional-14 ity of the device in a clinical trial (n=43) resulting 15 in systolic (SBP) and diastolic (DBP) pressures 16 $((mean \pm SD) \text{ mmHg})$ of $(-3.5 \pm 8.4) \text{ mmHg}$ and 17 (-4.0 ± 4.4) mmHg, respectively. Comparison 18 was made with manual auscultation (n=26) and 19 automated cuff oscillometry (n=18). In addition 20 to BP, we demonstrated the ability of the de-21 vice to assess arterial stiffness (n=18) and detect 22 atrial fibrillation (n=6). Conclusions: We were 23 able to introduce a sufficiently small device that 24 could be used for convenient ambulatory mea-25 surements and worn during sleep with minimal 26 discomfort. Lastly, we evaluate methods to fur-27 ther develop the concept and discuss future di-28 rections. 29

Plain language summary

Traditional cuff-based automated BP instruments are mainly suitable for individual measure-

ments and are often inconvenient to use. Patients 33 in a hospital ward are typically monitored with 34 many systems and devices. Our aim is to de-35 velop a blood pressure measurement technology 36 that could be integrated into a finger-worn pulse 37 oximeter. This way the need for a bulky and in-38 convenient arm cuff could be eliminated and the 39 number of needed monitoring devices could be 40 reduced. The proposed technology is especially 41 suitable for pseudo-continuous BP measurement, 42 that is, taking automated measurements periodi-43 cally, enabling BP trend tracking and analysis. 44 In this study, we clinically validate our tech-45 nology for blood pressure monitoring and fur-46 ther demonstrate its potential to measure arterial 47 stiffness and detect atrial fibrillation. 48

1 Introduction

Untreated hypertension is the greatest single contributor for cardiovascular disease and sudden cardiac death. Uncontrolled high blood pressure (BP) can lead to disability, poor quality of life, and fatal heart attack or stroke [31]. With changing global lifestyle, hypertension is affecting ever49

50

51

52

53

54

55

65

younger individuals and its prevalence is increas-56 ing. [8]. Currently, the trend in BP monitoring 57 devices is going towards wearable technologies 58 and ease-of-use [25]. Current hypertension guide-59 lines also recommend that blood pressure is mea-60 sured regularly and frequently (instead of ran-61 dom spot measurement)[17], despite traditional 62 BP devices with an inflatable arm cuff being un-63 comfortable, especially in long-term use. 64

Most wearable BP measurement devices use

All rights reserved. No reuse allowed without permission.

pulse wave propagation or pulse wave analysis 66 methods to measure BP [18, 19, 29]. These de-67 vices often require initial arm cuff calibration us-68 ing a brachial cuff device. Moreover, these meth-69 ods typically lack clinical validation and spe-70 cial situations and populations are seldom con-71 sidered. Our device relies on the oscillometric 72 method to acquire spot BP measurements and 73 it does not require any calibration [9]. Similar 74 techniques have been introduced before, but they 75 require a bulky external pneumatic unit for pres-76 sure generation, or are based on user actuation 77 [20, 26, 6].78

In our previous study we presented a table-79 top tonometric instrument for measuring blood 80 pressure from the fingertip[23]. Operation of 81 the device was validated (n=33) against an au-82 tomated arm cuff device (Omron M3) resulting 83 in systolic (SBP) and diastolic (DBP) pressures 84 $((mean \pm SD) \text{ mmHg})$ of $(-0.9 \pm 7.3) \text{ mmHg}$ and 85 (-3.3 ± 6.6) mmHg, respectively. We have also 86 proposed a miniaturized version of the device in 87 another study [24]. 88

In this study, our main objective was to vali-89 date the proposed device according to the inter-90 national BP standards. This involves compar-91 ing it to manual auscultation, the gold standard 92 method for non-invasive monitoring of BP[3]. 93 The standards used for most BP devices are 94 maintained by the US Association for the Ad-95 vancement of Medical Instrumentation (AAMI), 96 the European Society of Hypertension Working 97 Group on Blood Pressure Monitoring (ESH), 98 British Hypertension Society (BHS) and the 99 International Organization for Standardization 100 (ISO) [30, 4, 21]. 101

Our secondary objective was to miniaturize the 102 technology. The main motivation for miniatur-103 ization is to be able to integrate the technol-104 ogy to a pulse oximeter, eliminating the need for 105 an inconvenient brachial cuff and acquiring both 106 the blood pressure and blood oxygenation from 107 a sinle finger probe. In addition, we assessed the 108 device's capability to measure arterial stiffness 109 and atrial fibrillation. 110

Thirdly, we wanted to demonstrate the de-111 vice's capability to recognize cardiac arrhyth-112 mias. Atrial fibrillation (AF) is known to affect 113 the accuracy of the oscillometric method[22]. In 114 the case of AF, the BP results should be exam-115 ined critically. Inaccurate BP values can lead 116 to incorrect treatment or even lack of it. Con-117 sequently, built-in AF detection algorithms are 118 a requirement of a well performing BP monitor 119

 $\mathbf{2}$

Methods

2.1 Device development

2.1.1 Electronics

The device uses the nRF52840 (Nordic Semi-124 conductor, Norway) microcontroller unit (MCU) 125 dongle. The MCU was selected because of its 126 high performance ARM Cortex M4 core and 127 built-in USB capabilities. A barometric pres-128 sure sensor (BMP180, Bosch Sensortec, Ger-129 many) is connected to the MCU via I^2C connec-130 tion. A DC motor is used to apply pressure to 131 the finger. The MCU is interfaced to the motor 132 via DRV8837 motor driver (Texas Instruments, 133 USA) H-bridge that enables bidirectional motor 134 control. The MCU also monitors the two limit 135 switches (SSCQ110100, Alps Alpine, Japan) lo-136 cated on a separate PCB. The MCU is connected 137 to a laptop computer via USB. 138

120

121

122

123

139

2.1.2 Mechanics

The mechanics of the instrument house the elec-140 tronic components, provide a mechanism for ap-141 plying controlled pressure to the finger as well 142 as a structural base for the tonometric pressure 143 sensor. The enclosure was 3D printed using a se-144 lective laser sintering process. A standard pres-145 sure sensor is modified by prying off its protec-146 tive metal lid, and placing an air-filled cushion on 147 top of it. A cylindrical piston is placed onto the 148 cushion and it acts as the tonometer probe. The 149 piston consists of two pieces attached together 150 from the ends of a threaded shaft, enabling pis-151 ton extension and distension in case the protru-152 sion height of the piston has to be adjusted. A 153 DC motor (generic N20 type) followed by a re-154 duction gear train is placed in the inner support-155 ing enclosure. A worm gear assembly further re-156 duces the rotational speed and changes the angle 157 of rotation. The whole transmission is encased 158 in an inner support structure. A type of folding 159 press system is used to apply pressure to the fin-160 ger. The portion making contact to the finger 161 is made from a flexible 3D printing filament. In 162 order to make sure the pressing device does not 163 move out of its range, the hinge is equipped with 164 limit switches for the lower and upper bound-165 aries. This also prevents possible injuries that 166



Fig. 1. Instrument overview. a) Photograph showing the device in use. b) Cross-sectional view of the sensor setup. c) 3D model of the device. d) Oscillometric method. The pressure ramp signal is filtered with a pass-band filter (1 - 10 Hz) and Hilbert transformed. The peaks, or oscillations, are detected, and a polynomial curve is fitted to the peaks. The resulting bell-shaped curve is called an oscillometric waveform envelope (OMWE). Mean arterial pressure (MAP) is found at the point on the pressure curve that corresponds with the maximum of OMWE. DBP is found on the left side of the OMWE at 80% of maximum pressure. SBP is computed from DBP and MAP.



Fig. 2. Description of the system. a) Block diagram of the system. An ARM Cortex M4 based MCU is used as the controller unit. The MCU is connected to a laptop running a Python GUI for data logging. b) Inner mechanics illustrated in a 3D view. A DC motor is connected to a worm gear assembly through a reduction gear cassette. This rotates a hinge-type press that is used to apply pressure to the finger. A pair of micro limit switches assembled on a custom circuit board is used to limit the motion.

All rights reserved. No reuse allowed without permission.

could be caused by excessive pressure to the fin-ger.

¹⁶⁹ **2.1.3 Software**

The device firmware is written in the C programming language using the nRF5 software development kit (SDK). The main functionalities of the firmware are sampling the pressure sensor with sampling frequency of 128 Hz, and controlling the pressure-generating motor.

The measurement is controlled with a graphi-176 cal user interface (GUI) written in Python. The 177 application connects to the device through a USB 178 serial connection, which is used to send com-179 mands and receive data. The application plots 180 the incoming data in real-time. The user inter-181 face allows the user to change the settings of the 182 device, such as the speed of the motor and the 183 maximum allowed pressure, as well as start and 184 stop the measurement. Once a measurement has 185 been completed, the received data together with 186 the used device settings are stored in a comma-187 separated values (CSV) file. 188

189 2.2 Data analysis

Data analysis was performed on MATLAB
R2019a (Mathworks, USA). Each measurement
was stored in a separate CSV file and processed
individually.

Blood pressure was computed by first subtract-194 ing the atmospheric pressure from the signal to 195 set the base level for the measurement. The data 196 are filtered with a pass-band of 1 to 10 Hz and 197 the filtered signal is Hilbert transformed. A peak 198 detection algorithm followed by 8th order polyno-199 mial fitting is applied. The resulting bell-shaped 200 curve is called an oscillometric waveform enve-201 lope (OMWE). Mean arterial pressure (MAP) is 202 found at the point on the pressure curve that 203 corresponds with the maximum of OMWE. This 204 also represents zero transmural pressure, $P_t = 0$. 205 Transmural pressure is defined as the difference 206 between intra-arterial and external applied pres-207 sure. DBP is found at the pressure level corre-208 sponding to 80% of maximum oscillations at the 209 left side of OMWE. [23] 210

Systolic pressure is calculated from DBP andMAP using the following equation:

$$SBP = \frac{MAP - (1 - k)DBP}{k} \tag{1}$$

We used two methods to calculate the parameter k, resulting in two different values for SBP. In the standard method, k was defined as k = 0.4 for all subjects. For a patient-specific approach, we used ASI as an additional parameter in k: k = 0.4 + l, where $l = 5(\tilde{x}_{ASI} - ASI)$. $\tilde{x}_{ASI} = 0.03$ is the median of the calculated ASIs in the subject group with measured PWV (n = 220 18).

Arterial stiffness was calculated using the oscillometric waveform envelope. The OMWE was plotted against P_t . The signal was then numerically integrated, resulting in an S-shaped volumetransmural pressure $(V-P_t)$ plot. To fit the data into a descriptive function, we used nonlinear regression: 228

$$y \sim f(x,\beta) \tag{2}$$

where x is an independent variable, y its associated dependent variable and β is a vector of parameters of a nonlinear function f. We fit the measured volume-pressure data to an inverse trigonometric function: 233

$$f(x) = \beta_1 \arctan(\beta_2 + \beta_3 x) + \beta_4 \tag{3}$$

where $\beta_{1...4}$ are the parameters computed using non-linear fitting and variable x is transmural pressure P_t . Since β_2 describes the slope of the curve, it is used as the arterial stiffness index (ASI).[16] 238

We used the oscillometric waveform to detect 239 AF. This was done during BP computation af-240 ter fitting the 8th order polynomial curve to the 241 found peaks. The absolute difference, or error, 242 between the found peak and the polynomial fit 243 curve was calculated for each peak. The mean 244 error was calculated for each measurement and 245 compared with a threshold to decide between AF 246 and sinus rhythm. The principle of this is shown 247 in Figure 7. 248

2.3 Human studies

We recruited 32 volunteers (age: $\mu = 47$ years 250 [range: 24 to 83 years], 5 women) for measure-251 ments using the proposed device and manual aus-252 cultation. One subject was excluded due to se-253 vere nail clubbing. Six subjects having persistent 254 atrial fibrillation were excluded from the BP vali-255 dation data set, but were used to validate the de-256 tection of AF. The validation study was carried 257 out at Turku University Hospital. Subjects were 258 placed in a supine position to ensure that both 259 the finger device and the arm cuff were at the 260 same level. Subjects were also asked to relax and 261

249

All rights reserved. No reuse allowed without permission.



Fig. 3. a) Hydrostatic challenge. The subjects wore the finger device and a reference wrist BP monitor on the same hand. Three measurements were taken with the hand raised above and lowered below heart level, causing a change in BP. b) Clinical trial flow chart.

wait 10 minutes before the first measurement to 262 ensure that BP was stable. The finger probe was 263 placed on the index finger and an arm cuff con-264 nected to a mercury column sphygmomanome-265 ter was placed around the upper arm. Manual 266 auscultation was performed by two trained ob-267 servers, and the maximum acceptable difference 268 between the readings from the observers was 4 269 mmHg for both SBP and DBP. After an ini-270 tial test measurement, three measurements were 271 taken in a cyclical manner one device after an-272 other, both with the finger device and by aus-273 cultation. Data is displayed in Bland-Altman 274 and correlation plots for SBP, MBP and DBP 275 individually[1]. 276

In order to broaden the BP range, we recruited 277 9 individuals (age: $\mu = 48$ [range: 25 to 78 years], 278 4 females) to perform a hydrostatic challenge. 279 Four of the subjects were on BP lowering med-280 ication. We measured BP using our device and 281 a clinically validated wrist-worn BP cuff (Omron 282 R2, Japan) [27]. The wrist cuff uses the oscillo-283 metric method for the acquisition of BP. Both in-284 struments were worn in the same arm and held 10 285 cm below and above heart level resulting in a to-286 tal of 18 data points of SBP, MAP and DBP. This 287 results in a change of approximately 15 mmHg 288 between the two levels. Three consecutive mea-289 surements were taken for both hydrostatic levels. 290 Three of the subjects measured in this group were 291 also in the auscultated group, but the measure-292 ments were taken over 6 months apart. 293

In addition, we took a series of subsequent measurements to study the device's capability to take repeated measurements. The study was performed on one subject over approximately 20 minutes while comparing our device to a continuous reference. We used a volume clamp device CNAP 500 (CNSystems, Austria) as a reference

[11]. CNAP 500 is capable of outputting con-301 tinuous BP data including SBP, MAP and DBP 302 as well as the full cardiac waveform. The ref-303 erence device's cuffs were placed on the middle 304 and ring fingers ipsilaterally and our device was 305 placed on the index finger. The reference device 306 was calibrated using the built-in brachial sphyg-307 momanometer at the beginning of the measure-308 ment and again at 15 minutes. A total of 9 mea-309 surements were taken with our device, approx-310 imately once every two minutes. A timestamp 311 was set on the reference device after each oscillo-312 metric measurement taken on our device. As the 313 oscillometric sweep takes approximately 20 s, the 314 reference BP values were averaged over 20 s for 315 each measurement. We computed $((Mean \pm SD))$ 316 mmHg) for the measurements. 317

PWV was recorded from 18 subjects using 318 SphygmoCor XCEL (Atcor Medical, Australia) 319 and XCEL 1.3 data acquisition software [7]. 320 The patients remained in supine position while 321 a femoral artery cuff was wrapped around the 322 thigh. A pen-like tonometer was used to acquire 323 the carotid waveform. The distance between the 324 cuff and the carotid artery via jugular notch was 325 measured and used to calculate the PWV. An 326 average of three consecutive measurements was 327 used for data analysis. 328

The measurements were conducted according to the Declaration of Helsinki guidelines with the permission of Ethical Committee of the Hospital District of Southwest Finland and National Supervisory Authority for Welfare and Health[2].



Fig. 4. Bland-Altman and correlation plots for SBP, MBP and DBP. The two data sets, auscultated and hydrostatic challenge, are plotted with different markers.

All rights reserved. No reuse allowed without permission.

334 **3** Results

335 3.1 Blood pressure

Accuracy of the blood pressure measurement was 336 verified by combining the two data sets collected: 337 the auscultatory set (n=25) and the hydrostatic 338 challenge set (n=18). A total of 43 pairs of BP 339 values were analyzed. The blood pressures mea-340 sured ranged from 96 to 152 mmHg and from 341 50 to 96 mmHg for SBP and DBP respectively. 342 We calculated $((mean \pm SD) \text{ mmHg})$ for SBP, 343 MBP and DBP, resulting in $((-4.9\pm9.5) \text{ mmHg})$, 344 $((-1.2 \pm 3.9) \text{ mmHg})$ and $((-4.0 \pm 4.4) \text{ mmHg})$ 345 respectively. The patient-specific method for 346 computing SBP performed better $((-3.5 \pm 8.4))$ 347 mmHg) than the standard method. Mean and 348 diastolic pressure values comply with the inter-349 national standards for non-invasive BP monitors. 350 Standard deviation for SBP (8.4 mmHg) was 0.4351 mmHg higher than the AAMI standard thresh-352 old (8 mmHg) [30]. However, the method for 353 computing SBP was very simple and we believe 354 that using more sophisticated computation tech-355 niques, the standard could be met even for SBP. 356 Moreover, the accuracy of MBP is suitable for 357 longitudinal bedside BP monitoring. 358

359 3.2 Repeatability

We wanted to evaluate the device's capability 360 for long term monitoring. We took 9 consecu-361 tive measurements from the same subject over a 362 20 min period. Consecutive measurements taken 363 with our device were very close to the reference 364 device and followed the changes in BP correctly. 365 Comparison of the repeated measurements to the 366 reference device resulted in SBP of (-6.9 ± 4.7) 367 mmHg, MBP of (0.6 ± 4.0) mmHg, and DBP of 368 (1.6 ± 3.4) mmHg. The entire measurement is 369 shown in Fig. 5. 370

371 3.3 Arterial stiffness

Arterial stiffness was evaluated by analyzing the 372 shape of the oscillogram acquired by the fin-373 ger device, resulting in an arterial stiffness in-374 dex (ASI). The ASI describes the compressibility 375 of the artery when external pressure is applied 376 to it. This is characterized by the slope of the 377 curve in the V- P_t plot. The computed ASI was 378 compared with various other parameters. PWV 379 and ASI showed a negative correlation (r=-0.4), 380 which is comparable to the results presented in 381 the literature [16]. This suggests that individuals 382



Fig. 5. Repeatability study. Oscillometric measurements (shown as black vertical lines and markers) were taken with the finger device approx. every 2 minutes. The red and light red lines represent the continuous BP taken with the reference device CNAP 500.

with high PWV readings have lower ASI, thus 383 less compliant arteries. Individuals with lower 384 PWV seem to have more variation in ASI with 385 high and low ASI values. Similar phenomenon 386 was noticed when we compared ASI and age. 387 Younger subjects (<40 y) tend to have higher 388 variation in ASI than older ones (>60 y). How-389 ever, the younger group shows a 40% higher mean 390 value than the older group. In interpreting the 391 results between PWV and ASI, it is important 392 to understand that PWV is dependent on two 393 intrinsic variables: arterial stiffness and BP. Due 394 to this, PWV and ASI do not have a linear re-395 lationship [10]. As can be concluded from the 396 results shown in Subsection 3.1, ASI does indeed 397 hold BP information and can be used as an ad-398 ditional parameter for computing BP. 399

3.4 Atrial fibrillation detection

400

We wanted to both demonstrate the device's ca-401 pability of detecting AF and exclude these pa-402 tients from the validation data set, since AF is 403 known to affect oscillometric BP measurements 404 significantly. AF was found on six subjects. The 405 mean error from OMWE in mmHg was computed 406 for each subject. In AF, the pulse pressure varies 407 randomly, resulting in deformation of the oscil-408 lometric envelope. Single pulses can be higher 409 or lower than the OMWE, increasing the mean 410 error. A 0.8 mmHg threshold was used to differ-411

All rights reserved. No reuse allowed without permission.



Fig. 6. a) Example $V-P_t$ plots estimating arterial compliance for a young subject with low carotidto-femoral PWV (6.3 m/s) and an elderly subject with high PWV (10.6 m/s). The slope of the curve in the young subject's case is higher, resulting in higher ASI. b) Figure showing the negative correlation between ASI and PWV.

entiate AF from sinus rhythm. Figure 7 shows 412 that AF cases can be classified using a simple 413 threshold. 414

4 Discussion 415

We presented a miniaturized finger probe for 416 measuring BP, assessing arterial stiffness and de-417 tecting atrial fibrillation. In this study, the sen-418 sor technology introduced in our previous article 419 was integrated into a pulse-oximeter-style wear-420 able probe [23]. This allows the device to be dis-421 crete and unobtrusive while making integration 422 to existing patient monitors possible. We vali-423 dated the the device in a clinical trial, compar-424 ing the proposed device with auscultation - the 425 gold standard method for non-invasive BP mea-426 surement. In an attempt to improve the credi-427 bility of the validation, we introduced additional 428 high- and low-BP values into the data set via hy-429 drostatic challenge. This way we were able to 430 record more high and low pressures without hav-431 ing to recruit subjects having such BP values. 432 The results for MBP and DBP fulfill the bias 433 and precision criteria set by AAMI, but the stan-434 dard deviation for the SBP are slightly above the 435 standard threshold (0.4 mmHg above the limit). 436 The accuracy could be improved by more sophis-437 ticated algorithms and better positioning of the 438 finger. One possibility is either to use a person-439 alized molded insert or an adjustable back piece 440 to make sure the device geometry is optimal for 441 each individual. This would be viable especially 442 for long-term use. The study differs from stan-443 dard protocol in population size and BP range. 444 Our study size (n = 43) satisfies the criteria for 445 the number of subjects set by ESH (n=33) but 446 not by AAMI (n=85). In addition, we used the 447 hydrostatic challenge to increase the data set, 448 which is not a method recommended in any of 449 the standards. Not all of the study population 450 requirements were met either. The subjects re-451 cruited in the study were either normotensive or 452 had their hypertension under control. We were 453 able to broaden the BP range with hydrostatic 454 challenge, but we still miss the very high BP of 455 >160 mmHg for SBP and >100 mmHg for DBP. 456

Arterial stiffness was evaluated by analyzing the shape of the oscillogram. Our measurements suggest that ASI can be used as an additional parameter for computing BP. We found a negative correlation between ASI and carotid-to-femoral PWV, similar to the findings in the literature[16].

All rights reserved. No reuse allowed without permission.



Fig. 7. Atrial fibrillation detection. The difference between the oscillometric envelope and the amplitude of each peak is used to calculate the mean error. The algorithm was able to differentiate AF from sinus rhythm using the error between oscillometric envelope and pulse peaks.

Age and ASI were also found to correlate. ASI
could possibly be used as a feature in machine
learning applications for more pervasive hemodynamic monitoring. We were also able to distinguish AF cases from sinus rhythm cases using
simple metrics derived from the oscillometric envelope.

The finger-pressing system has a clear advan-470 tage over cuff-based systems: One device is suit-471 able for all finger sizes. Brachial and finger 472 cuff devices often have different cuffs for small, 473 medium, and large anatomies. The study sub-474 jects also noted that the finger device produced 475 much less discomfort than a traditional brachial 476 cuff. Especially in the case of hypertensive pa-477 tients, the pressure required for complete brachial 478 occlusion can cause significant pain. Compres-479 sion applied on the fingertip causes less distal tis-480 sue occlusion compared to brachial, radial, and 481 proximal phalanx cuff occlusion. Furthermore, 482 taking short oscillometric measurements period-483 ically is much less uncomfortable than having to 484 maintain partial occlusion continuously, which 485 is the case in volume-clamp methods [12]. An-486 other significant advantage to the volume-clamp 487 method and most wearable devices is that our 488 approach does not need any form of calibration, 489 which is usually done by brachial cuff oscillome-490 trv. 491

There are also limitations to the technology. 492 In several cases, the subject had cold fingers, 493 which significantly affected the quality of the sig-494 nal. One subject had undergone surgery on his 495 left arm, making it difficult for him to hold his 496 hand in the correct position. Additionally, for 497 several subjects, the subject's finger had to be 498 repositioned several times to ensure good signal 499 quality and coupling of arterial pulsation to the 500 sensor. This suggests that the sensor is quite sen-501 sitive to misplacement. However, when correctly 502 positioned, the measurements had high repeata-503 bility, as demonstrated in Subsection 3.2, which 504 is crucial for long-term bedside monitoring. 505

Additionally, future directions could include a 506 more exhaustive study on pseudo-continuous BP 507 monitoring. An ideal setup would include an 508 invasive reference e.g. during a surgical oper-509 ation. In our previous publication, we demon-510 strated that using fingertip tonometry, we were 511 able to acquire high resolution BP waveform sim-512 ilar to invasive BP [13]. The combination of peri-513 odic oscillometric measurements and pulse wave-514 form could be of clinical use. Another interesting 515 topic is to study how the device performs in noc-516 turnal BP monitoring, possibly including opti-517 cal oxygen saturation sensing. Nocturnal hypox-518 emia, that is, a decrease in peripheral oxygen sat-519 uration during sleep is a marker of adverse con-520

All rights reserved. No reuse allowed without permission.

ditions, such as chronic obstructive pulmonary 521 disease [5] and obstructive sleep apnea. Further-522 more, it has been reported to inversely correlate 523 with morning surge of BP, which, on the other 524 hand, is an important diagnostic tool in the man-525 agement of hypertension [14, 15]. Making the de-526 vice wireless and battery-operated would enable 527 convenient data collection in nursing homes or 528 even in a home environment. The current enclo-529 sure still has room for a small lithium-polymer 530 battery as well as a charging circuit, and the mi-531 crocontroller used has Bluetooth radio commu-532 nication capabilities. However, this is more of 533 a novel engineering challenge. Compared to a 534 bulky BP device, a small portable device would 535 be easier for the healthcare professional to carry 536 on the hospital ward. 537

Author Contributions

⁵³⁹ Conception and design: T.P., T.N., I.K. and
⁵⁴⁰ M.K., System development: T.P. and J.S., Col⁵⁴¹ lection and assembly of data: T.P., J.S., T.N.,
⁵⁴² I.K. and M.K., Analysis and interpretation of the
⁵⁴³ data: T.P., Drafting of the article: T.P., J.S. and
⁵⁴⁴ M.K., Obtaining funding: M.P., T.K. and M.K.

545 References

- [1] A. Aalok. Bland-altman and correlation
 plot, 2022.
- W. M. Association et al. World medical association declaration of helsinki: ethical principles for medical research involving human subjects. Jama, 310(20):2191–2194, 2013.
- J. Booth. A short history of blood pressure
 measurement. *Proc roy Soc Med*, 70:793– 799, 1977.
- [4] E. Brien, J. Petrie, W. Littler, M. de Swiet,
 P. Padfield, D. Altman, M. Bland, A. Coats,
 and N. Atkins. The british hypertension society protocol for the evaluation of blood
 pressure devices. *J Hypertens*, 11(Suppl.
 2):S43–S62, 1993.
- J. Buekers, J.-M. Aerts, J. Theunis, A. Vaes,
 E. Wouters, M. Spruit, and P. De Boever.
 Multi-night continuous nocturnal oxygen
 saturation measurements in patients with
 copd, 2018.

- [6] A. Chandrasekhar, C.-S. Kim, M. Naji, 566
 K. Natarajan, J.-O. Hahn, and R. Mukka- 567
 mala. Smartphone-based blood pressure 568
 monitoring via the oscillometric finger- 569
 pressing method. Science translational 570
 medicine, 10(431):eaap8674, 2018. 571
- J. M. Davies, M. A. Bailey, K. J. Griffin, 572
 and D. J. A. Scott. Pulse wave velocity and the non-invasive methods used to assess it: Complior, sphygmocor, arteriograph and vicorder. Vascular, 20(6):342–349, 2012. 576
- [8] J. M. Flack and B. Adekola. Blood pressure and the new acc/aha hypertension guidelines. Trends in Cardiovascular Medicine, 30(3):160–164, 2020.
- M. Forouzanfar, H. R. Dajani, V. Z. Groza, M. Bolic, S. Rajan, and I. Batkin. Oscillometric blood pressure estimation: past, present, and future. *IEEE reviews in 584 biomedical engineering*, 8:44–63, 2015.
- [10] A. Giudici, C. Palombo, M. Kozakova, C. Morizzo, J. K. Cruickshank, and A. W. Khir. Subject-specific pressure normalization of local pulse wave velocity: Separating intrinsic from acute load-dependent stiffening in hypertensive patients. *Frontiers in Physiology*, 12:783457–783457, 2021.
- [11] R. Hahn, H. Rinösl, M. Neuner, and S. Kettner. Clinical validation of a continuous non-invasive haemodynamic monitor (cnap[™] 500) during general anaesthesia. British journal of anaesthesia, 108(4):581-585, 2012.
- B. P. Imholz, W. Wieling, G. A. van Mont frans, and K. H. Wesseling. Fifteen years
 experience with finger arterial pressure mon itoring: assessment of the technology. Car diovascular research, 38(3):605-616, 1998.
- [13] M. Kaisti, T. Panula, J. Leppänen, 604 R. Punkkinen, M. Jafari Tadi, T. Vasankari, 605 S. Jaakkola, T. Kiviniemi, J. Airaksinen, 606 P. Kostiainen, et al. Clinical assessment of a 607 non-invasive wearable mems pressure sensor 608 array for monitoring of arterial pulse wave-609 form, heart rate and detection of atrial fib-610 rillation. NPJ digital medicine, 2(1):1-10, 611 2019.612
- K. Kario, H. Kanegae, N. Tomitani, 613
 Y. Okawara, T. Fujiwara, Y. Yano, 614

S. Hoshide, and J.-H. S. Group. Nighttime
blood pressure measured by home blood
pressure monitoring as an independent predictor of cardiovascular events in general
practice: The j-hop nocturnal blood pressure study. *Hypertension*, 73(6):1240–1248,
2019.

- [15] A. Kishimoto, O. Tochikubo, and
 K. Ohshige. Relation between nocturnal
 arterial oxygen desaturation and morning
 blood pressure. *Clinical and Experimental Hypertension*, 29(1):51–60, 2007.
- [16] H. Komine, Y. Asai, T. Yokoi, and
 M. Yoshizawa. Non-invasive assessment of
 arterial stiffness using oscillometric blood
 pressure measurement. *Biomedical engineer- ing online*, 11(1):1–12, 2012.
- [17] P. Muntner, D. Shimbo, R. M. Carey, J. B.
 Charleston, T. Gaillard, S. Misra, M. G.
 Myers, G. Ogedegbe, J. E. Schwartz, R. R.
 Townsend, et al. Measurement of blood pressure in humans: a scientific statement from
 the american heart association. *Hypertension*, 73(5):e35-e66, 2019.
- [18] D. Nachman, Y. Gepner, N. Goldstein,
 E. Kabakov, A. B. Ishay, R. Littman, Y. Azmon, E. Jaffe, and A. Eisenkraft. Comparing blood pressure measurements between
 a photoplethysmography-based and a standard cuff-based manometry device. Scien-*tific reports*, 10(1):1–9, 2020.
- [19] D. Nair, S. Tan, H. Gan, S. Lim, J. Tan, 646 M. Zhu, H. Gao, N. Chua, W. Peh, and 647 K. Mak. The use of ambulatory tonomet-648 ric radial arterial wave capture to measure 649 ambulatory blood pressure: the validation of 650 a novel wrist-bound device in adults. Jour-651 nal of human hypertension, 22(3):220–222, 652 2008.653
- R. Narasimhan, T. Parlikar, G. Verghese,
 and M. V. McConnell. Finger-wearable
 blood pressure monitor. In 2018 40th Annual International Conference of the IEEE
 Engineering in Medicine and Biology Society (EMBC), pages 3792–3795. IEEE, 2018.
- E. O'Brien, T. Pickering, R. Asmar, M. Myers, G. Parati, J. Staessen, T. Mengden,
 Y. Imai, B. Waeber, P. Palatini, et al. Working group on blood pressure monitoring of

the european society of hypertension international protocol for validation of blood pressure measuring devices in adults. *Blood pressure monitoring*, 7(1):3–17, 2002.

- [22] N. Pagonas, S. Schmidt, J. Eysel, F. Compton, C. Hoffmann, F. Seibert, J. Hilpert, C. Tschöpe, W. Zidek, and T. H. Westhoff.
 Impact of atrial fibrillation on the accuracy of oscillometric blood pressure monitoring. *Hypertension*, 62(3):579–584, 2013.
- T. Panula, T. Koivisto, M. Pänkäälä, T. Ni iranen, I. Kantola, and M. Kaisti. An
 instrument for measuring blood pressure
 and assessing cardiovascular health from the
 fingertip. Biosensors and Bioelectronics,
 167:112483, 2020.
- T. Panula, J.-P. Sirkiä, and M. Kaisti. Miniaturization of a finger-worn blood pressure instrument. In 2021 43rd Annual International Conference of the IEEE Engineering in Medicine & Biology Society (EMBC), pages 7441–7444. IEEE, 2021.
- [25] T. Panula, J.-P. Sirkia, D. Wong, and M. Kaisti. Advances in non-invasive blood pressure measurement techniques. *IEEE Reviews in Biomedical Engineering*, 2022.
- [26] L.-Y. Shyu, Y.-L. Kao, W.-Y. Tsai, and W. Hu. Development of a cuffless blood pressure measurement system. In 2012 Annual International Conference of the IEEE Engineering in Medicine and Biology Society, pages 2040–2043. IEEE, 2012.
- [27] J. Topouchian, D. Agnoletti, J. Blacher, 696 A. Youssef, I. Ibanez, J. Khabouth, 697 S. Khawaja, L. Beaino, and R. Asmar. Val-698 idation of four automatic devices for self-699 measurement of blood pressure according to 700 the international protocol of the european 701 society of hypertension. Vascular health and 702 risk management, 7:709, 2011. 703
- W. J. Verberk, S. Omboni, A. Kollias, and G. S. Stergiou. Screening for atrial fibrillation with automated blood pressure measurement: research evidence and practice recommendations. *International journal of cardiology*, 203:465–473, 2016.
- [29] A. Vybornova, E. Polychronopoulou, 710
 A. Wurzner-Ghajarzadeh, S. Fallet, J. Sola, 711
 and G. Wuerzner. Blood pressure from the 712

713	optical aktiia bracelet: a 1-month validation
714	study using an extended iso 81060-2 protocol
715	adapted for a cuffless wrist device. Blood
716	pressure monitoring, 26(4):305-311, Aug
717	2021.

- [30] W. B. White, A. S. Berson, C. Robbins, 718 M. J. Jamieson, L. M. Prisant, E. Roccella, 719 and S. G. Sheps. National standard for mea-720 surement of resting and ambulatory blood 721 pressures with automated sphygmomanome-722 ters. Hypertension, 21(4):504-509, 1993. 723
- [31] World health organization, hypertension. 724 https://www.who.int/news-room/fact-725 sheets/detail/hypertension. Accessed: 726
- 2021-01-04.727