Antigen concentration, viral load, and test performance for SARS-CoV-2 in multiple

specimen types

Allison Golden,[a#] Michelle Oliveira-Silva,[b] Hannah Slater,[a] Alexia Martines Vieira,[b] Pooja

Bansil,[a] Emily Gerth-Guyette,[a] Brandon T. Leader,[a] Stephanie Zobrist,[a] Alan Kennedy Braga

Ferreira,[b] Erika Crhistina Santos de Araujo,[b] Catherine Duran de Lucena Cruz,[b] Eduardo

Garbin,[b] Greg T. Bizilj,[a] Sean J. Carlson,[a] Mariana Sagalovsky,[a] Sampa Pal,[a] Vin Gupta,[c] Leo

Wolansky,[d] David S. Boyle,[a] Deusilene Souza Vieira Dall'Acqua,[e] Felipe Gomes Naveca,[f]

Valdinete Alves do Nascimento,[f] Juan Miguel Villalobos Salcedo,[e] Paul K. Drain,[g,h] Alexandre

Dias Costa,[i] Dhélio Pereira,[b] Gonzalo J Domingo[a]

[a]Diagnostics, PATH, Seattle, Washington, USA

[b]Centro de Pesquisa em Medicina Tropical (CEPEM), Porto Velho, Rondônia, Brazil

[c]Amazon.com, Seattle, Washington, USA

[d]The Rockefeller Foundation, Pandemic Prevention Institute, New York City, New York, USA

[e]Fundação Oswaldo Cruz (FIOCRUZ), Porto Velho, Rondônia, Brazil

[f]Instituto Leônidas e Maria Deane (ILMD), Fundação Oswaldo Cruz (FIOCRUZ), Manaus,

Amazonas, Brazil

[g]Department of Global Health, University of Washington, Seattle, Washington, USA

[h]Department of Medicine, University of Washington, Seattle, Washington, USA

[i]Instituto Carlos Chagas (ICC), Fundação Oswaldo Cruz (FIOCRUZ), Curitiba, Paraná, Brazil

Running head: Antigen concentration and performance study

#Address correspondence to Allison Golden, algolden@path.org.

Allison Golden, Michelle Oliveira-Silva, and Hannah Slater contributed equally to this work and

order of authors was alphabetically listed by last name.

Abstract word count: 244

Abstract

The relationship between N-antigen concentration and viral load within a specimen and across

different specimens is essential for interpretation of rapid diagnostic tests (RDT) clinical

performance in different use cases. A prospective study was conducted in Porto Velho, Brazil, to

investigate RDT performance in different specimen types as a function of the correlation

between antigen concentration and viral load. The study included 214 close contacts with recent

exposures to confirmed cases, aged 12 years and older and with various levels of vaccination.

Antigen concentration was measured in nasopharyngeal swab (NPS), anterior nares swab (ANS),

and saliva specimens. Reverse transcriptase (RT)–PCR was conducted on the NPS and saliva

specimens, and two RDTs were conducted on ANS and one on saliva. Antigen concentration

correlated with viral load when measured in the same specimen type but not across specimen

types. Antigen levels were higher in symptomatic cases compared to

asymptomatic/oligosymptomatic cases and lower in saliva compared to NPS and ANS samples.

Discordant results between the RDTs conducted on ANS and the RT-PCR on NPS were resolved

by antigen concentration values. The analytical limit-of-detection of RDTs can be used to predict

the performance of the tests in populations for which the antigen concentration is known. The

antigen dynamics across different sample types observed in SARS-CoV-2 disease progression

support use of RDTs in nasal samples. Given lower antigen concentrations in saliva, tests using

saliva is expected to require improved analytical sensitivity to achieve clinical sensitivity similar

to testing of nasal samples.

Keywords: LumiraDx assay SARS-CoV-2 Ag Test, STANDARD Q COVID-19 Ag Saliva and

Antigen Tests, SalivaDirect™

Introduction

Diagnostic testing has a critical role in the public health response to the severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) pandemic.[1] Early detection of the virus can help to limit transmission, inform infection-control measures, and guide appropriate clinical management of patients. The reference standard for SARS-CoV-2 testing is molecular detection of viral RNA with reverse transcriptase (RT)–PCR assays.[2] However, this laboratory-based method is technically complex, costly, and requires robust specimen transport and reporting systems – all of which limit its utility and availability in many settings.[3] Since the onset of the pandemic, an unprecedented number of rapid diagnostic tests (RDTs) have been developed to address this gap, including many designed to detect viral antigens.[1] Such tests play important roles in increasing access to SARS-CoV-2 testing globally.

Viral load varies across different specimen types throughout the progression of COVID-19 disease and between individuals.[4,5] The variant and vaccination status may also influence the relative viral loads.[6,7] However, performance of RDTs has often been compared to RT-PCR results conducted on nasopharyngeal swab (NPS) specimens as a gold standard, and was initially the regulatory reference, regardless of the type of specimen that the RDT is conducted on. Understanding RDT performance relative to RT-PCR may be limited by the relationship between the antigen concentration in the sampled oral or nasal cavity and viral copies in the NPS. Understanding antigen concentration dynamics in different specimen types can inform how RDTs can be expected to perform in different stages of infection and use-case scenarios.

In 2021, a prospective diagnostic accuracy study was conducted among close contacts of

COVID-19-positive index cases in Porto Velho, Brazil.[8] This study evaluated the performance of

four tests for SARS-CoV-2—including three rapid antigen tests (STANDARD Q COVID-19 Ag

Nasal and Saliva tests [SD Biosensor, Republic of Korea] and the SARS-CoV-2 Ag Test

[LumiraDx™ Limited, United Kingdom]), as well as one dual-plexed RT-PCR method that uses

minimally processed saliva[9,10] (SalivaDirect™ protocol [Yale University, United States])—

against a regulatory-approved RT-PCR assay run on NPS specimens. Here, we present the

results of reference antigen concentration measurements on various specimen types collected

from this study, and determine the correlations between antigen concentration and viral load.

Materials and methods

Study design and clinical procedures

This design of this clinical study has been described previously.[8] Briefly, between July and

September 2021, a prospective diagnostic accuracy study was conducted among close contacts of

50 confirmed COVID-19-positive index cases in Porto Velho, Brazil. Close contacts (aged 12

years and older) were identified through contact elicitation interviews administered to index

cases. All eligible close contacts were invited to participate in the study. A subset of those who

shared a primary residence with the index case were followed longitudinally for clinical

evaluations and testing every other day for up to five visits total (up to eight additional days after

the initial visit), or until the participant obtained a positive result on the rapid test administered

during the visit. In such cases, only one additional visit was performed, but an NPS was not

collected to minimize staff exposure.

Information on medical history, participant demographics, and symptoms were collected at

enrollment, with the latter also collected at each study visit. Additionally, at each visit, two

paired ANS, one NPS, and one saliva sample, from passive saliva that pools naturally in the

mouth, were collected. The details of specimen collection have been previously published

(supplementary materials in Zobrist et al., 2022[8]). The STANDARD Q COVID-19 Ag Nasal

Test was run on one ANS during the visit. The remaining lysis buffer from this test was retained

for antigen testing and remaining samples transferred to the laboratory to be stored frozen.

Tests evaluated

Three antigen tests were evaluated: the STANDARD Q COVID-19 Ag Nasal Test and

LumiraDx SARS-CoV-2 Ag Test on ANS samples and the STANDARD Q COVID-19 Ag

Saliva Test on saliva samples. One molecular test was also evaluated: the SalivaDirect protocol,

a dual-plexed RT-PCR method for SARS-CoV-2 detection from minimally processed saliva.[10]

Specimen characterization summary

RT-PCR and SalivaDirect RT-PCR were conducted on the NPS and saliva specimens,

respectively. Quantitative viral load values are available for the NPS sample, while only Ct

values are available for the saliva sample. Rapid antigen tests were conducted on ANS and saliva

samples. The antigen concentration was determined in all three specimen types: for NPS it was

measured from VTM, for ANS from the same STANDARD Q COVID-19 Ag lysis buffer used

for the corresponding antigen test, and for saliva directly from saliva.

Laboratory procedures

The saliva samples and the second ANS, extracted with LumiraDx buffer, were stored on ice

following collection and were frozen immediately on return to the laboratory within five hours of

collection. The LumiraDx SARS-CoV-2 Ag Test and the STANDARD Q COVID-19 Ag Saliva

Test were both run using thawed specimens within five days after initial freezing of the samples.

The SalivaDirect assay was run in batches. All tests were run by operators blinded to point-of-

care and reference results.

Reference testing

Reference testing was conducted on NPS with a multiplex real-time PCR assay (Allplex™

SARS-CoV-2 Assay [Seegene Inc., Republic of Korea]). A CFX96 real-time PCR machine (Bio-

Rad, United States)[11] was used, and an automated RNA extraction was conducted using the

Extracta kit (Loccus, Brazil). All SARS-CoV-2-positive specimens were repeated on the same

assay for quantitative estimation of viral load using a dilution series of Amplirun® Coronavirus

RNA Control (Vircell S.L., Spain) for viral load quantification on each plate. The limit of

quantification was 125 copies/mL.

Genomic sequencing

Specimens with cycle threshold (Ct) values <30 underwent genomic sequencing.[8] Genomic

sequencing for this study was conducted using Illumina COVIDSeq® Test (Illumina Inc., United

States), with some modifications,[12] on Illumina's MiSeq® or NextSeq 1000® sequencers. Reads

were assembled using DRAGEN COVID Lineage App, v. 3.5.4, at Illumina BaseSpace

(http://basespace.illumina.com) or BBMap 38.84 embedded in Geneious Prime 2022.1

(https://www.geneious.com). Consensus sequences were analyzed for quality issues with

Nextclade (https://clades.nextstrain.org),[13] and lineages were identified using the pangolin tool

on its most up-to-date version.[14]

All SARS-CoV-2 genomes generated and analyzed in this study are available at the EpiCoV

database in GISAID (https://www.gisaid.org) [15].

Antigen testing

SARS-CoV-2 N concentration was measured using the Meso Scale Discovery platform (Meso

Scale Diagnostics, United States), which uses electrochemiluminescence for detection. The assay

details have been described previously[16]. Specimens were shipped to PATH (Seattle, WA, USA)

where the antigen concentration assay was conducted. The SARS-CoV-2 nucleocapsid protein

quantitative immunoassay on the platform was run using a 10-point recombinant protein standard

curve ranging from 0.128 pg/mL to 100 ng/mL for quantification. Briefly, 25 uL of neat or

diluted viral transport media (VTM) from the eluted NPS, STANDARD Q COVID-19 Ag lysis

buffer ANS, or saliva sample were diluted in an equal volume of 2x running buffer for samples

for a final sample dilution of 2-fold and final detergent concentration of 1%. Neat samples were

used for those testing negative on cognate assays, or with Ct values on the NPS of greater than

25. For NPS specimens with Ct values between 21 and 25 and presumed positives without Ct

information, a sample five-fold dilution was performed, and for those with Ct values of 20 or

less, a sample twenty-fold dilution was performed to ensure the results were within the dynamic

range of the assay. Sample dilution of presumed antigen-positive ANS were five or twenty-fold

depending on combination of assay results and NPS results. Saliva specimens were diluted ten-

fold if both SalivaDirect and the STANDARD Q COVID-19 Ag Saliva Test were positive.

Positive specimens were run in duplicate and discrepant results of coefficient of variation over

20%, or those outside the assay detection range, were repeated. Samples exceeding the dynamic

range of the assay were repeated with increased dilution. Samples which were diluted based on

expected concentration and had no antigen detected were rerun at the lowest dilution of two-fold.

Samples which tested positive for antigen but were expected to be negative were run again to

confirm. Duplicate well quality control samples consisting of previously aliquoted, diluted

recombinant protein at 3 different concentrations were run with each plate. Across 21 plates run,

all controls were within the median +/- 20% for concentration and no controls had CVs over

20%.

Statistical analysis

Sensitivity, specificity, and positive and negative predictive values were calculated using

standard formulas and presented with 95% confidence intervals. Correlations between antigen

values, viral load (viral genome equivalents) values and SalivaDirect ct values were assessed

using the R-squared value generated from a linear model fitted to log10 transformed data (apart

from the SalivaDirect ct data which were not transformed). Differences between distributions of

antigen concentrations based on symptom status were calculated using a two-sided non-paired t-

test. The relationships between antigen concentration or viral load and the probability of

positivity by each test were modeled using a logistic regression model fitted using the R package

brms.[17] Gaussian priors were used for the parameters and models were run for 10,000 iterations

to ensure convergence. Data were collected and managed using REDCap electronic data capture

tools.[18] Statistical analyses were conducted using R 4.0.3 (R Foundation for Statistical

Computing, Austria).

Ethical considerations

191 WCG Institutional Review Board (1301165), the CEPEM ethics committee, and Brazil's

192 National Research Ethics Commission approved this study (44351421.0.0000.0011). Written

193 informed consent was obtained for all participants.

Results

Participant characteristics

In total, 214 close contacts were enrolled from 50 symptomatic COVID-19-confirmed index

cases. Of these 214 close contacts, 64 shared a primary household with the index cases and were

followed longitudinally. Among all the close contacts, 65 tested as SARS-CoV-2 positive by the

reference assay at least once. Confirmed cases were classified as either symptomatic or

oligosymptomatic/asymptomatic, as described in Supplement B. Full details of the study

participant demographics, including vaccination status, has been previously published.[8]

Antigen concentration determination

Antigen concentration was measured in all three available specimen types for each participant

and timepoint combination selected from the total: those participants with at least one positive

test result on any of the molecular or antigen detection tests (n=166) at that time point,

participant time points which were negative for all results for at a time point only but for

participants which at some point had a positive test result (n=33), and participant time points

which were from participants who had never had any positive result on any test for any time

point (n=136). In total, antigen concentration was determined for 955 samples which were

available from the selected participant time points, along with two additional saliva samples from

negative participant time points (see Supplement A).

Relationship between antigen concentration and symptoms

Antigen concentration in specimens from symptomatic individuals ($n \geq 84$) was significantly

higher than in specimens from asymptomatic or oligosymptomatic study participants ($n \geq 39$)

across all clinical specimen types ($P<0.01$) (Figure 1). Antigen concentration was also

significantly lower in saliva specimens as compared to ANS and NPS specimens ($P<0.01$), with

the mean values for symptomatic saliva specimens being almost two orders of magnitude lower

than those for the ANS and NPS specimens.

Relationship between viral load and antigen concentration

A good correlation between antigen concentration and viral load (or Ct value on the SalivaDirect

assay) was observed between measurements conducted on cognate specimens $[virus]_{NPS}$/ [N-

antigen]$_{NPS}$ ($R^2 = 0.721$) and $[Ct]_{saliva}$/[N-antigen]$_{saliva}$ ($R^2 = 0.725$) (Figure 2). Antigen

concentration measurements in NPS were positive for detectable antigen in 77.5% of the paired

RT-PCR positives, with undetectable antigen only in lower viral load samples. In contrast, poor

correlations were observed for the same measurements across specimen types relative to the NPS

viral load: $[virus]_{NPS}/[N-antigen]_{ANS}$ ($R^2 = 0.471$) and $[virus]_{NPS}/[N-antigen]_{saliva}$ ($R^2 = 0.061$).

Relationship between antigen concentration and antigen detection test result

An overall dose response relationship is observed when relating antigen concentration to test

signal intensity on both STANDARD Q COVID-19 Ag tests and on the LumiraDx SARS-CoV-2

Ag Test (Figure 3). For the STANDARD Q COVID-19 Ag tests, line signal intensity defined by

a predefined scale (Supplement C) rose with antigen concentration measured in the same

specimen. Likewise on the LumiraDx SARS-CoV-2 Ag Test, signal intensity as measured by the

LumiraDx instrument (and provided by LumiraDx) rose with antigen concentration about signal

saturation.

The probability of test positivity plotted against antigen concentration in the ANS specimen for

LumiraDx SARS-CoV-2 Ag Test and the STANDARD Q COVID-19 Ag tests conducted on

ANS and against saliva antigen concentration for the STANDARD Q COVID-19 Ag Saliva Test

show overlapping profiles indicative of similar analytical performance of the tests in their

respective specimens (Figure 4A). The probability of positivity for the three tests against NPS

viral load has broader 95% credible intervals and more gradual probability increases with

increasing viral load concentrations, particularly for the STANDARD Q COVID-19 Ag Saliva

Test (Figure 4B).

A subset (N=9) of participant had NPS specimens with Ct values below 28, yet had negative

antigen rapid test results. Genome sequencing was available for 8 of these specimens.

Sequencing did identify sense mutations on the N antigen on some of these discordant

specimens, but the same mutations were also found on specimens where there were no

discordant results (Supplement D). Including also participant samples for which the NPS PCR

result was Ct below 34 yet had one or more negative rapid test result (n=14), antigen

concentrations in the ANS specimen extracted in STANDARD Q COVID-19 Ag lysis buffer for

these participants were found to be low or undetectable. Without clear mutation association and

with overall low antigen concentrations found in the discordant ANS the results suggest that

false negatives to PCR by rapid test were most likely due to low antigen availability in the ANS

specimens.

Validation of predictive performance modeling from analytical sensitivity results

Both the STANDARD Q COVID-19 Ag and LumiraDx SARS-CoV-2 Ag tests had previously

been evaluated for their analytical performance against antigen concentration using a clinical

specimen pool in a laboratory.[16] The antigen limits of detection, as defined by antigen

concentrations at which tests were expected to be positive 90% of the time, were applied as

thresholds in the measured antigen concentration in the ANS specimens and categorize the

predicted test results as detected or not by each rapid test. Clinical performance of the two tests

among the close contacts in the study was then predicted based on comparison of these results to

the gold standard of NPS RT-PCR positive cases.

Table 1 shows the actual observed sensitivity and the predicted sensitivity in this study

population composed of symptomatic and asymptomatic/oligosymptomatic cases based on

analytical performance and antigen concentrations.

Discussion

This study investigated the relationship between antigen concentration and diagnostic test

performance on different types of samples collected in a study conducted in Porto Velho, Brazil,

in 2021.[8] The study shows that antigen concentration is closely related to viral load when

measured from the same specimen type. The NPS antigen concentration corresponded to NPS

viral load, and the saliva antigen concentration measured in saliva samples correlated with Ct

values from the SalivaDirect assay, which has been previously observed.[19,20] However, the

antigen concentrations in ANS and saliva specimens did not correlate well with the viral load

measured in the same study participants' NPS specimens. N-antigen concentrations are reflective

of virus levels in the corresponding oral/nasal cavities from which they are collected but do not

reflect viral levels in other oral/nasal cavities.

The antigen concentration corresponded very well with both the signal intensity and test

positivity of the rapid antigen detection tests within the same sample type. As demonstrated by

overlapping curves, no significant difference in the probability of test positivity as a function of

cognate sample type antigen concentration was observed across the three tests included in this

study: the STANDARD Q COVID-19 Ag conducted on ANS and saliva and the LumiraDx

SARS-CoV-2 Ag Test conducted on ANS. In contrast, when test positivity was modeled as a

function of viral copy number in the NPS, the performance of the saliva antigen test was

significantly lower; this is expected based on the poor correlation between saliva antigen

concentration and NPS viral load. Additionally, "discordant" specimens (with high NPS viral

loads defined by Ct values of <34, but negative antigen test results on ANS) were resolved by

confirming that, in the ANS specimen, there was indeed low or undetectable N-antigen levels.

The false negative results could not be attributed to variations in the N gene and test failure, as

the N gene mutations detected in the study were shared by other samples with concordant results.

These observations are important when considering study designs for antigen detection test

evaluations and assessing their performance. From a regulatory perspective, NPS RT-PCR has

been considered the gold standard for SARS-CoV-2 diagnostic test evaluation, regardless of the

specimen source for the index test. This study shows that discordant results can and should be

anticipated if the specimen source for the RDT under evaluation is different to the specimen used

to measure viral load. Within a specimen type, antigen concentration, and therefore antigen

detection test performance, can be expected to relate to viral load. These results may also explain

some of the differences in performance observed for SARS-CoV-2 tests across different sample

types. [21–23]

The study did confirm higher antigen concentrations in symptomatic cases compared to

asymptomatic and oligosymptomatic cases across all three specimen types, as anticipated given

the association between viral carriage and symptom onset. While recent studies suggest that early

onset of SARS-CoV-2 can be first observed in the saliva by RT-PCR,[4,5,20] the overall lower

antigen levels observed in saliva in this study suggests that current rapid antigen tests are not

sensitive enough to leverage this opportunity for earlier infection detection.

Combining the laboratory analytical limits-of-detection determined from a clinical specimen

pooled sample with the distribution of antigen concentrations in this study population, it was

possible to estimate the observed performance of these antigen detection tests in this study

population This result confirms that benchmarking analytical performance can be related to

clinical performance, albeit with caution. Larger data sets of paired viral load and antigen

concentration measurements across different sample types, across variants, and in different study

populations may improve the ability to predict the performance of antigen detection tests in

different scenarios and use cases. Of note, the population included both symptomatic cases and

asymptomatic/oligosymptomatic cases, including samples with low viral load counts and

undetectable antigen concentration.

Some limitations to the findings of this study are that (a) RT-PCR was not conducted on the

ANS samples, (b) the SalivaDirect test was not conducted as a quantitative test, and (c) antigen

concentration was measured across different matrices from samples collected using different

swabs, potentially leading to inherent biases or systematic errors in the antigen quantification. In

particular, the ANS antigen concentration was determined from the same extracted specimen

from which the STANDARD Q COVID-19 Ag Test was conducted. In contrast, the LumiraDx

SARS-CoV-2 Ag Test was conducted on a second ANS sample taken in parallel that had been

extracted in its corresponding buffer and subsequently frozen according to instructions for use,

prior to testing. Dilutional differences from extraction buffer volumes were accounted for in the

predicted sensitivity, but matrix, swab, and sampling handling differences remain.

In conclusion, this study indicates the value of understanding the underlying antigen

concentration dynamics and its relationship to viral load across different sample types to inform

and predict RDT performance in different settings, variants, and use cases.

Tables

Table 1. Observed sensitivity of the antigen detection tests performed on ANS compared to their

predicted performance based on analytical limit-of-detection. The study population were close

contacts and include symptomatic, and asymptomatic/oligosymptomatic cases. Confirmed

positive cases by RT-PCR on NPS was used to determine a true positive.

Performance vs. PCR (in nasopharyngeal sample)	ANS sample (all)		ANS sample (Ct below 34)	
	STANDARD Q COVID-19 Ag	**SARS-CoV-2 Ag Test [LumiraDx]**	**STANDARD Q COVID-19 Ag**	**SARS-CoV-2 Ag Test [LumiraDx]**
Observed clinical performance in close contacts	55.0 (44.1–66.9) N = 79	50.6 (39.1–62.1) N = 78	84.0 (70.9–92.8) N = 49	79.6 (65.7–89.8) N = 50
Predicted performance in close contacts based on antigen concentration and analytical limits of detection[16]	50.0 (38.5–61.5) N = 78	50.0 (38.5–61.5) N = 78	79.6 (65.7–89.8) N = 49	79.6 (65.7–89.8) N = 49

Figures

Figure 1. **Box and whisker plots showing antigen concentration distributions among**

symptomatic cases and oligo/asymptomatic cases among the close contacts in anterior nares

swab (ANS), nasopharyngeal swab (NPS), and saliva specimens. The number (n) of

specimens per category is listed above each category.

Figure 2. Relationship between viral load and antigen concentration. Viral load was

measured from NPS stored in VTM. The Saliva Direct assay provides Ct values in the saliva

specimen. Antigen concentration was measured from the same NPS-VTM specimen, from ANS

specimen stored in the STANDARD Q lysis buffer and from saliva. Correlations are shown

between (A) NPS viral load and NPS antigen concentration, (B) saliva viral load (in Ct values)

and saliva antigen concentration, (C) NPS viral load and ANS antigen concentration, and (D)

NPS viral load and saliva concentration.

Figure 3. Relationship between N-antigen signal and rapid antigen diagnostic signal.

Antigen concentration was measured from an anterior nasal swab specimen stored in the

STANDARD Q lysis buffer and from saliva. The relationships are shown for Panel (A) the

ANS STANDARD Q test signal and the ANS N-antigen concentration, Panel (B) the saliva

STANDARD Q assigned test signal and the saliva N-antigen concentration, and Panel (C) the

ANS Lumira signal and ANS N-antigen concentration.

Figure 4. **Probability of positivity for the rapid antigen tests as a function of antigen**

concentration on their cognate specimen. Panel (A): Probability for positive test result for

Lumira (blue) and the STANDARD Q test (red) conducted on the ANS specimen as a function

of the antigen concentration in the ANS specimen, and the STANDARD Q test (green)

conducted on the saliva specimen as a function of the antigen concentration in the saliva

specimen. Panel (B): Probability of positive test for the same three tests as a function of viral

load in NPS specimens. The antigen concentration or viral load at which there is greater than

90% probability of a positive test result is indicated.   The shaded areas show the 95% credible

intervals for the probability functions.

Acknowledgements

This work was supported by grants from The Rockefeller Foundation [2020 HTH 039] and

Amazon.com [2D-04020007] to Gonzalo J Domingo. Felipe Gomes Naveca and Valdinete

Alves do Nascimento were supported by the National Council for Scientific and Technological

Development [grant 403276/2020-9] and Inova Fiocruz / Fundação Oswaldo Cruz [grant

VPPCB-007-FIO-18-2-30 - Knowledge generation]. Felipe Gomes Naveca is a National Council

for Scientific and Technological Development (CNPq) fellow. Benchmarking work cited and

that was used in analysis was funded by the Bill & Melinda Gates Foundation

(https://www.gatesfoundation.org/) via grant INV-016821. The funder did not have any role in

the study design, data collection and analysis, decision to publish, or preparation of the

manuscript.

The authors would like to thank all study participants, as well as the clinical and laboratory staff

at CEPEM involved with this study. We also thank the team at Global Health Strategies for their

contributions to this study, as well as SD Biosensor and LumiraDx for facilitating the availability

of their tests for this study. Finally, we thank Amanda Tsang and Mara Lavery for editorial

support with the manuscript.

References

1. Peeling RW, Heymann DL, Teo YY, Garcia PJ. Diagnostics for COVID-19: moving from
pandemic response to control. *Lancet*. 2022;399(10326):757-768. doi:10.1016/S0140-
6736(21)02346-1

2. Mardian Y, Kosasih H, Karyana M, Neal A, Lau CY. Review of Current COVID-19
Diagnostics and Opportunities for Further Development. *Front Med (Lausanne)*.
2021;8:615099. doi:10.3389/fmed.2021.615099

3. Drain PK. Rapid Diagnostic Testing for SARS-CoV-2. *N Engl J Med*. 2022;386(3):264-272.
doi:10.1056/NEJMcp2117115

4. Savela ES, Viloria Winnett A, Romano AE, et al. Quantitative SARS-CoV-2 Viral-Load
Curves in Paired Saliva Samples and Nasal Swabs Inform Appropriate Respiratory Sampling
Site and Analytical Test Sensitivity Required for Earliest Viral Detection. *J Clin Microbiol*.
2022;60(2):e0178521. doi:10.1128/JCM.01785-21

5. Lai J, German J, Hong F, et al. Comparison of Saliva and Midturbinate Swabs for Detection
of SARS-CoV-2. *Microbiol Spectr*. 2022;10(2):e0012822. doi:10.1128/spectrum.00128-22

6. King KL, Wilson S, Napolitano JM, et al. SARS-CoV-2 variants of concern Alpha and Delta
show increased viral load in saliva. *PLoS One*. 2022;17(5):e0267750.
doi:10.1371/journal.pone.0267750

7. Puhach O, Adea K, Hulo N, et al. Infectious viral load in unvaccinated and vaccinated
individuals infected with ancestral, Delta or Omicron SARS-CoV-2. *Nat Med*.
2022;28(7):1491-1500. doi:10.1038/s41591-022-01816-0

8. Zobrist S, Oliveira-Silva M, Vieira AM, et al. Screening for SARS-CoV-2 in close contacts
of individuals with confirmed infection: performance and operational considerations. *J Infect*
Dis. Published online May 20, 2022:jiac204. doi:10.1093/infdis/jiac204

9. Rodríguez Flores SN, Rodríguez-Martínez LM, Reyes-Berrones BL, Fernández-Santos NA,
Sierra-Moncada EJ, Rodríguez-Pérez MA. Comparison Between a Standard and SalivaDirect
RNA Extraction Protocol for Molecular Diagnosis of SARS-CoV-2 Using Nasopharyngeal
Swab and Saliva Clinical Samples. *Front Bioeng Biotechnol*. 2021;9:638902.
doi:10.3389/fbioe.2021.638902

10. Vogels CBF, Watkins AE, Harden CA, et al. SalivaDirect: A simplified and flexible
platform to enhance SARS-CoV-2 testing capacity. *Med (N Y)*. 2021;2(3):263-280.e6.
doi:10.1016/j.medj.2020.12.010

11. Freppel W, Merindol N, Rallu F, Bergevin M. Efficient SARS-CoV-2 detection in
unextracted oro-nasopharyngeal specimens by rRT-PCR with the Seegene Allplex[TM] 2019-
nCoV assay. *Virol J*. 2020;17(1):196. doi:10.1186/s12985-020-01468-x

12. Naveca FG, Nascimento V, de Souza VC, et al. COVID-19 in Amazonas, Brazil, was driven
by the persistence of endemic lineages and P.1 emergence. *Nat Med*. 2021;27(7):1230-1238.
doi:10.1038/s41591-021-01378-7

13. Aksamentov I, Roemer C, Hodcroft EB, Neher RA. Nextclade: clade assignment, mutation
calling and quality control for viral genomes. *Journal of Open Source Software*.
2021;6(67):3773. doi:10.21105/joss.03773

14. O'Toole Á, Scher E, Underwood A, et al. Assignment of epidemiological lineages in an
emerging pandemic using the pangolin tool. *Virus Evol*. 2021;7(2):veab064.
doi:10.1093/ve/veab064

15. Elbe S, Buckland-Merrett G. Data, disease and diplomacy: GISAID's innovative
contribution to global health. *Glob Chall*. 2017;1(1):33-46. doi:10.1002/gch2.1018

16. Golden A, Cantera JL, Lillis L, et al. A Reagent and Virus Benchmarking Panel for a
Uniform Analytical Performance Assessment of N Antigen–Based Diagnostic Tests for
COVID-19. *medRxiv*. Published online January 1, 2022:2022.08.03.22278351.
doi:10.1101/2022.08.03.22278351

17. Bürkner PC. brms: An R Package for Bayesian Multilevel Models Using Stan. *Journal of
Statistical Software*. 2017;80:1-28. doi:10.18637/jss.v080.i01

18. Harris PA, Taylor R, Thielke R, Payne J, Gonzalez N, Conde JG. Research electronic data
capture (REDCap)--a metadata-driven methodology and workflow process for providing
translational research informatics support. *J Biomed Inform*. 2009;42(2):377-381.
doi:10.1016/j.jbi.2008.08.010

19. Pollock NR, Savage TJ, Wardell H, et al. Correlation of SARS-CoV-2 Nucleocapsid Antigen
and RNA Concentrations in Nasopharyngeal Samples from Children and Adults Using an
Ultrasensitive and Quantitative Antigen Assay. *Journal of Clinical Microbiology*.
2021;59(4). doi:10.1128/JCM.03077-20

20. Ren A, Sohaei D, Ulndreaj A, et al. Ultrasensitive assay for saliva-based SARS-CoV-2
antigen detection. *Clin Chem Lab Med*. 2022;60(5):771-777. doi:10.1515/cclm-2021-1142

21. Zhou Y, O'Leary TJ. Relative sensitivity of anterior nares and nasopharyngeal swabs for
initial detection of SARS-CoV-2 in ambulatory patients: Rapid review and meta-analysis.
PLoS One. 2021;16(7):e0254559. doi:10.1371/journal.pone.0254559

22. Wölfl-Duchek M, Bergmann F, Jorda A, et al. Sensitivity and Specificity of SARS-CoV-2
Rapid Antigen Detection Tests Using Oral, Anterior Nasal, and Nasopharyngeal Swabs: a
Diagnostic Accuracy Study. *Microbiol Spectr*. 2022;10(1):e0202921.
doi:10.1128/spectrum.02029-21

23. Lee S, Widyasari K, Yang HR, Jang J, Kang T, Kim S. Evaluation of the Diagnostic
Accuracy of Nasal Cavity and Nasopharyngeal Swab Specimens for SARS-CoV-2 Detection
via Rapid Antigen Test According to Specimen Collection Timing and Viral Load.
Diagnostics (Basel). 2022;12(3):710. doi:10.3390/diagnostics12030710

FIGURE 1

FIGURE 2

FIGURE 3

FIGURE 4