Anatomy Segmentation in Laparoscopic Surgery: Comparison of Machine Learning and Human Expertise

Fiona R. Kolbinger, MD[1,2,3,✉], Franziska M. Rinner[1], Alexander C. Jenke[4], Matthias Carstens[1], Stefanie Krell[4], Stefan Leger, PhD[3,4], Marius Distler, MD[1,2], Jürgen Weitz, MD[1,2,3,5], Stefanie Speidel, PhD[3,4,5], Sebastian Bodenstedt, PhD[4,5,✉]

[1] Department of Visceral, Thoracic and Vascular Surgery, University Hospital and Faculty of Medicine Carl Gustav Carus, Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany
[2] National Center for Tumor Diseases (NCT/UCC), Dresden, Germany: German Cancer Research Center (DKFZ), Heidelberg, Germany; Faculty of Medicine and University Hospital Carl Gustav Carus, Technische Universität Dresden, Dresden, Germany; Helmholtz-Zentrum Dresden-Rossendorf (HZDR), Fetscherstraße 74, 01307 Dresden, Germany
[3] Else Kröner Fresenius Center for Digital Health (EKFZ), Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany
[4] Division of Translational Surgical Oncology, National Center for Tumor Diseases (NCT), Partner Site Dresden, Fetscherstraße 74, 01307 Dresden, Germany
[5] Cluster of Excellence "Centre for Tactile Internet with Human-in-the-Loop" (CeTI), Technische Universität Dresden, 01062, Dresden, Germany

✉ Corresponding authors:
Dr. Fiona Kolbinger, Department of Visceral, Thoracic and Vascular Surgery, University Hospital and Faculty of Medicine Carl Gustav Carus, Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany
☎ +49 (0) 351 458 19624
✉ fiona.kolbinger@uniklinikum-dresden.de

Dr. Sebastian Bodenstedt, Division of Translational Surgical Oncology, National Center for Tumor Diseases (NCT/UCC), Partner Site Dresden, Fetscherstrasse 74, 01307 Dresden, Germany
☎ +49 (0) 351 5413
✉ sebastian.bodenstedt@nct-dresden.de

STRUCTURED ABSTRACT

Background: Lack of anatomy recognition represents a clinically relevant risk in abdominal surgery. Machine learning (ML) methods can help identify visible patterns and risk structures, however, their practical value remains largely unclear.

Materials and Methods: Based on a novel dataset of 13195 laparoscopic images with pixel-wise segmentations of eleven anatomical structures, we developed specialized segmentation models for each structure and combined models for all anatomical structures using two state-of-the-art model architectures (DeepLabv3 and SegFormer), and compared segmentation performance of algorithms to a cohort of 28 physicians, medical students, and medical laypersons using the example of pancreas segmentation.

Results: Mean Intersection-over-Union for semantic segmentation of intraabdominal structures ranged from 0.28 to 0.83 and from 0.23 to 0.77 for the DeepLabv3-based structure-specific and

combined models, and from 0.31 to 0.85 and from 0.26 to 0.67 for the SegFormer-based structure-

specific and combined models, respectively. Both the structure-specific and the combined

DeepLabv3-based models are capable of near-real-time operation, while the SegFormer-based

models are not. All four models outperformed at least 26 out of 28 human participants in pancreas

segmentation.

Conclusions: These results demonstrate that ML methods have the potential to provide relevant

assistance in anatomy recognition in minimally-invasive surgery in near-real-time. Future research

should investigate the educational value and subsequent clinical impact of respective assistance

systems.

HIGHLIGHTS

• Machine learning models to reduce surgical risks that precisely identify 11 anatomical

structures: abdominal wall, colon, intestinal vessels (inferior mesenteric artery and inferior

mesenteric vein with their subsidiary vessels), liver, pancreas, small intestine, spleen,

stomach, ureter and vesicular glands

• Large training dataset of 13195 real-world laparoscopic images with high-quality anatomy

annotations

• Similar performance of individual segmentation models for each structure and combined

segmentation models in identifying intraabdominal structures, and similar segmentation

performance of DeepLabv3-based and SegFormer-based models

• DeepLabv3-based models are capable of near-real-time operation while SegFormer-

based models are not, but SegFormer-based models outperform DeepLabv3-based

models in terms of accuracy and generalizability

• All models outperformed at least 26 out of 28 human participants in pancreas

segmentation, demonstrating their potential for real-time assistance in recognizing

anatomical landmarks during minimally-invasive surgery.

KEYWORDS

Minimally-invasive surgery, laparoscopy, surgical data science, surgical anatomy, surgical

innovation, artificial intelligence

INTRODUCTION

Computer vision describes the computerized analysis of digital images aiming at the automation of human visual capabilities, most commonly using machine learning methods, in particular deep learning. This approach has transformed medicine in recent years, with successful applications including computer-aided diagnosis of colonic polyp dignity in endoscopy (1,2), detection of clinically actionable genetic alterations in histopathology (3), and melanoma detection in dermatology (4). Availability of large amounts of training data is the defining prerequisite for successful application of deep learning methods. With the establishment of laparoscopy as the gold standard for a variety of surgical procedures (5–8) and the increasing availability of computing resources, these concepts have gradually been applied to abdominal surgery. The overwhelming majority of research efforts in the field of Artificial Intelligence (AI)-based analysis of intraoperative surgical imaging data (i.e. video data from laparoscopic or open surgeries) has focused on classifying images with respect to the presence and/or location of previously annotated surgical instruments or anatomical structures (9–13) or on analysis of surgical proficiency (14–16) based on recorded procedures. However, almost all research endeavors in the field of computer vision in laparoscopic surgery have concentrated on preclinical stages and to date, no AI model based on intraoperative surgical imaging data could demonstrate a palpable clinical benefit (17,18). Among the studies closest to clinical application are recent works on identification of instruments and hepatobiliary anatomy during cholecystectomy for automated assessment of the critical view of safety (13), and on the automated segmentation of safe and unsafe preparation zones during cholecystectomy (19).

In surgery, patient outcome heavily depends on experience and performance of the surgical team (20,21). In a recent analysis of Human Performance Deficiencies in major cardiothoracic, vascular, abdominal transplant, surgical oncology, acute care, and general surgical operations, more than half of the cases with postoperative complications were associated with identifiable human error. Among these errors, lack of recognition (including misidentified anatomy) accounted for 18.8%, making it the most common Human Performance Deficiency overall (22). Examples of complications directly related to anatomical misperception are iatrogenic lesions to the ureter in gynecologic procedures (23) and pancreatic injuries during splenic flexure mobilization in colorectal surgery (24). While AI-based systems identifying anatomical risk and target structures would theoretically have the potential to alleviate this risk, limited availability and diversity of (annotated) laparoscopic image data drastically restrict the clinical potential of such applications in practice.

To advance and diversify the applications of computer vision in laparoscopic surgery, we have
recently published the Dresden Surgical Anatomy Dataset (25), providing 13195 laparoscopic
images with high-quality annotations of the presence and exact location of eleven intraabdominal
anatomical structures: abdominal wall, colon, intestinal vessels (inferior mesenteric artery and
inferior mesenteric vein with their subsidiary vessels), liver, pancreas, small intestine, spleen,
stomach, ureter and vesicular glands. Here, we present the first study based on this dataset and
present machine learning models to assist in precisely delineating anatomical structures, aiming
to reduce surgical risks. Specifically, we evaluate automated detection and localization of organs
and anatomical structures in laparoscopic view using two state-of-the-art model architectures:
DeepLabv3 and SegFormer. To assess the clinical value of the presented machine learning
models, we compare algorithm segmentation performance to that of humans using the example
of delineation of the pancreas.

METHODS

Patient cohort

Video data from 32 robot-assisted anterior rectal resections or rectal extirpations were gathered
at the University Hospital Carl Gustav Carus Dresden between February 2019 and February 2021.
All included patients had a clinical indication for the surgical procedure, recommended by an
interdisciplinary tumor board. The procedures were performed using the da Vinci® Xi system
(Intuitive Surgical, Sunnyvale, CA, USA) with a standard Da Vinci® Xi/X Endoscope with Camera
(8 mm diameter, 30˚ angle, Intuitive Surgical, Sunnyvale, CA, USA, Item code 470057). Surgeries
were recorded using the CAST system (Orpheus Medical GmBH, Frankfurt a.M., Germany). Each
record was saved at a resolution of 1920 x 1080 pixels in MPEG-4 format.
All experiments were performed in accordance with the ethical standards of the Declaration of
Helsinki and its later amendments. The local Institutional Review Board (ethics committee at the
Technical University Dresden) reviewed and approved this study (approval number: BO-EK-
140032021). The trial was registered on clinicaltrials.gov (trial registration ID: NCT05268432).
Written informed consent to laparoscopic image data acquisition, data annotation, data analysis,
and anonymized data publication was obtained from all participants. Before publication, all data
was anonymized according to the general data protection regulation of the European Union.

Dataset

Based on the full-length surgery recordings and respective temporal annotations of organ visibility,
individual image frames were extracted and annotated as described previously (25). In brief, three

independent annotators with substantial experience in robot-assisted rectal surgery created pixel-
wise annotations, which were subsequently reviewed by a surgeon with 4 years of experience in
robot-assisted rectal surgery. A detailed description of the annotation process including analyses
of annotator agreement and technical parameters has been published previously (25). To
guarantee real-world applicability of machine learning models trained on the dataset, images with
perturbations such as blurring due to camera movements, soiling of the lens, and presence of
blood or smoke were not specifically excluded. The resulting Dresden Surgical Anatomy Dataset
comprises 13195 distinct images with pixel-wise segmentations of eleven anatomical structures:
abdominal wall, colon, intestinal vessels (inferior mesenteric artery and inferior mesenteric vein
with their subsidiary vessels), liver, pancreas, small intestine, spleen, stomach, ureter and
vesicular glands. Moreover, the dataset comprises binary annotations of the presence of each of
these organs for each image. The dataset is publicly available via the following link:
https://doi.org/10.6084/m9.figshare.21702600.
155
For machine learning purposes, the Dresden Surgical Anatomy Dataset was split into training,
validation, and test data as follows (Figure 1):
— Training set (at least 12 surgeries per anatomical structure): surgeries 1, 4, 5, 6, 8, 9, 10,
12, 15, 16, 17, 19, 22, 23, 24, 25, 27, 28, 29, 30, 31.
— Validation set (3 surgeries per anatomical structure): surgeries 3, 21, 26.
— Test set (5 surgeries per anatomical structure): surgeries 2, 7, 11, 13, 14, 18, 20, 32.
This split is proposed for future works using the Dresden Surgical Anatomy Dataset to reproduce
the variance of the entire dataset within each subset, and to ensure comparability regarding clinical
variables between the training, the validation, and the test set. Surgeries for the test set were
selected to minimize variance regarding the number of frames over the segmented classes. Out
of the remaining surgeries, the validation set was separated from the training set using the same
criterion.
168

Figure 1: Schematic illustration of the structure-specific and combined machine learning models used for semantic segmentation. The Dresden Surgical Anatomy Dataset was split into a training, a validation, and a test set. For spatial segmentation, two sets of machine learning models – a structure-specific model with individual encoders and decoders, and a combined model with a common encoder and structure-specific decoders – were trained for DeepLabv3-based and SegFormer-based model architectures.

Structure-specific semantic segmentation models

To segment each anatomical structure, a separate convolutional neural network for segmentation of individual structures was trained. Specifically, we trained and compared two different

Anatomy Segmentation in Laparoscopic Surgery

architectures: a Deeplabv3 (26) model with a ResNet50 backbone with default PyTorch pretraining

on the COCO dataset (27), and a SegFormer (28) model pretrained on the Cityscapes dataset

(29). The networks were trained using cross-entropy loss and the AdamW optimizer (30) for 100

epochs with a starting learning rate of 10^{-4} and a linear learning rate scheduler decreasing the

learning rate by 0.9 every 10 epochs. For data augmentation, we applied random scaling and

rotation, as well as brightness and contrast adjustments. The final model for each organ was

selected via the Intersection-over-Union (IoU) on the validation dataset and evaluated using the

Dresden Surgical Anatomy Dataset with the abovementioned training-validation-test split (Figure

1).

Segmentation performance was assessed using F1 score, IoU, precision, recall, and specificity

on the test folds. These parameters are commonly used technical measures of prediction

exactness, ranging from 0 (least exact prediction) to 1 (entirely correct prediction without any

misprediction).

Combined semantic segmentation models

A convolutional neural network with a common encoder and eleven decoders for combined

segmentation of the eleven anatomical structures was trained. As for the structure-specific

models, DeepLabv3-based (26) and SegFormer-based (28) models were used. For DeepLabv3,

a shared ResNet50 backbone with default PyTorch pretraining on the COCO dataset (27) was

used. For each class, a DeepLabv3 decoder was then run on the features extracted from a given

image by the backbone. Similarly, for SegFormer, an encoder, pretrained on the Cityscapes

dataset (29), was combined with eleven decoders.

As the images are only annotated for binary classes, the loss is only calculated for every pixel in

images, in which the structure associated with the current decoder is annotated. For images, in

which the associated class is not annotated, only the pixels that are annotated as belonging to

another class are included in the loss, e.g., pixels that were annotated as belonging to the class

"liver" can be used as negative examples for the class "pancreas". The remaining training

procedure was identical to the structure-specific model. The models were trained and evaluated

using the Dresden Surgical Anatomy Dataset with the abovementioned training-validation-test

split (Figure 1).

Segmentation performance was assessed using F1 score, IoU, precision, recall, and specificity

on the test folds.

Evaluation of the semantic segmentation models on an external dataset

To explore generalizability, structure-specific and combined models based on both architectures

(DeepLabv3 and SegFormer) were deployed to laparoscopic image data from the publicly

available LapGyn4 dataset (31). Models were separately deployed for full-scene segmentations

and their performance was visually compared.

Comparative evaluation of algorithmic and human performance

To determine the clinical potential of automated segmentation of anatomical risk structures, the

segmentation performance of 28 humans was compared to that of the structure-specific and the

combined semantic segmentation models using the example of the pancreas. The local

Institutional Review Board (ethics committee at the Technical University Dresden) reviewed and

approved this study (approval number: BO-EK-566122021). All participants provided written

informed consent to anonymous study participation, data acquisition and analysis, and publication.

In total, 28 participants (physician and non-physician medical staff, medical students, and medical

laypersons) marked the pancreas in 35 images from the Dresden Surgical Anatomy Dataset (25)

with bounding boxes. These images originated from 26 different surgeries, and the pancreas was

visible in 16 of the 35 images. Each of the previously selected 35 images was shown once, the

order being arbitrarily chosen but identical for all participants. The open-source annotation

software Computer Vision Annotation Tool (CVAT) was used for annotations. In cases where the

pancreas was seen in multiple, non-connected locations in the image, participants were asked to

create separate bounding boxes for each area.

Based on the structure-specific and the combined semantic segmentation models, axis-aligned

bounding boxes marking the pancreas were generated in the 35 images from the pixel-wise

segmentation. To guarantee that the respective images were not part of the training data, four-

fold cross validation was used, i.e., the origin surgeries were split into four equal-sized batches,

and algorithms were trained on three batches that did not contain the respective origin image

before being applied to segmentation.

To compare human and algorithm performance, the bounding boxes created by each participant

and the structure-specific as well as the combined semantic segmentation models were compared

to bounding boxes derived from the Dresden Surgical Anatomy Dataset, which were defined as

ground truth. IoU between the manual or automatic bounding box and the ground truth was used

to compare segmentation accuracy.

Anatomy Segmentation in Laparoscopic Surgery

DATA AND CODE AVAILABILITY

Data Availability

The Dresden Surgical Anatomy Dataset is publicly available via the following link: https://doi.org/10.6084/m9.figshare.21702600. All other data generated and analyzed during the current study are available from the corresponding authors on reasonable request. To gain access, data requestors will need to sign a data access agreement.

Code Availability

The most relevant scripts used for dataset compilation are publicly available via the following link: https://zenodo.org/record/6958337#.YzsBdnZBzOg. The code used for segmentation algorithms is available at https://gitlab.com/nct_tso_public/anatomy-recognition-dsad.

RESULTS

Machine Learning-based anatomical structure segmentation in structure-specific models

Structure-specific multi-layer convolutional neural networks (Figure 1) were trained to segment the abdominal wall, the colon, intestinal vessels (inferior mesenteric artery and inferior mesenteric vein with their subsidiary vessels), the liver, the pancreas, the small intestine, the spleen, the stomach, the ureter, and vesicular glands (Supplementary Table 1). Table 1 displays mean F1 score, IoU, precision, recall, and specificity for individual anatomical structures as predicted by the structure-specific algorithms on the test data.

Out of the analyzed segmentation models based on DeepLabv3, performance was lowest for vesicular glands (mean IoU: 0.28 ± 0.21), the pancreas (mean IoU: 0.28 ± 0.27), and the ureter (mean IoU: 0.36 ± 0.25), while excellent predictions were achieved for the abdominal wall (mean IoU: 0.83 ± 0.14) and the small intestine (mean IoU: 0.80 ± 0.18). In segmentation of the pancreas, the ureter, vesicular glands and intestinal vessel structures, there was a relevant proportion of images with no detection or no overlap between prediction and ground truth, while for all remaining anatomical structures, this proportion was minimal (Figure 2 a). While the images, in which the highest IoUs were observed, mostly displayed large organ segments that were clearly visible (Figure 2 b), the images with the lowest IoU were of variable quality with confounding factors such as blood, smoke, soiling of the endoscope lens, or pictures blurred by camera shake (Figure 2 c). While overall segmentation performance of both architectures was similar for structure-specific

models, SegFormer-based models showed a trend towards better performance than DeepLabv3-

based models in segmentation of the pancreas, the spleen, and the ureter (Table 1, Figure 2,

Supplementary Figure 1).

For the DeepLabv3-based structure-specific models, inference on a single image with a resolution

of 640 x 512 pixels required, on average, 28 ms on an Nvidia A5000, resulting in a frame rate of

almost 36 frames per second. In contrast, the SegFormer-based structure-specific semantic

segmentation models operated considerably slower at an inference time of 53 ms per image,

resulting in a frame rate of 18 frames per second. This runtime includes one decoder, meaning

that only the segmentation for one anatomical class is included.

Table 1: Summary of performance metrics for anatomical structure segmentation using DeepLabv3-based (a)
and SegFormer-based (b) structure-specific models on the test dataset. For each metric, mean and standard
deviation are displayed.
289

a	Anatomical structure	F1 score	IoU	Precision	Recall	Specificity
	Abdominal wall	0.90 ± 0.10	0.83 ± 0.14	0.89 ± 0.14	0.93 ± 0.07	0.97 ± 0.04
	Colon	0.79 ± 0.20	0.69 ± 0.22	0.80 ± 0.21	0.82 ± 0.21	0.97 ± 0.05
	Inferior mesenteric artery	0.54 ± 0.26	0.41 ± 0.22	0.55 ± 0.25	0.67 ± 0.33	0.99 ± 0.01
	Intestinal veins	0.54 ± 0.33	0.44 ± 0.29	0.70 ± 0.26	0.56 ± 0.36	1.00 ± 0.00
DeepLabv3 (Test)	Liver	0.80 ± 0.23	0.71 ± 0.25	0.85 ± 0.21	0.81 ± 0.24	0.98 ± 0.03
	Pancreas	0.37 ± 0.32	0.28 ± 0.27	0.59 ± 0.37	0.37 ± 0.36	1.00 ± 0.01
	Small intestine	0.87 ± 0.14	0.80 ± 0.18	0.87 ± 0.16	0.91 ± 0.15	0.97 ± 0.04
	Spleen	0.79 ± 0.23	0.69 ± 0.24	0.74 ± 0.22	0.90 ± 0.24	0.99 ± 0.01
	Stomach	0.71 ± 0.24	0.60 ± 0.25	0.65 ± 0.25	0.89 ± 0.21	0.98 ± 0.02
	Ureter	0.47 ± 0.30	0.36 ± 0.25	0.53 ± 0.28	0.57 ± 0.39	1.00 ± 0.00
	Vesicular glands	0.40 ± 0.25	0.28 ± 0.21	0.37 ± 0.28	0.62 ± 0.35	0.97 ± 0.03

b	Anatomical structure	F1 score	IoU	Precision	Recall	Specificity
	Abdominal wall	0.91 ± 0.11	0.85 ± 0.15	0.90 ± 0.14	0.94 ± 0.09	0.98 ± 0.03
	Colon	0.77 ± 0.21	0.66 ± 0.22	0.73 ± 0.22	0.87 ± 0.22	0.95 ± 0.07
	Inferior mesenteric artery	0.60 ± 0.23	0.46 ± 0.21	0.58 ± 0.25	0.73 ± 0.29	0.99 ± 0.01
	Intestinal veins	0.65 ± 0.25	0.52 ± 0.24	0.62 ± 0.27	0.76 ± 0.27	1.00 ± 0.00
SegFormer (Test)	Liver	0.83 ± 0.21	0.75 ± 0.24	0.82 ± 0.23	0.88 ± 0.18	0.98 ± 0.03
	Pancreas	0.47 ± 0.32	0.37 ± 0.28	0.61 ± 0.36	0.48 ± 0.36	0.99 ± 0.01
	Small intestine	0.89 ± 0.13	0.83 ± 0.17	0.87 ± 0.16	0.95 ± 0.10	0.97 ± 0.04
	Spleen	0.85 ± 0.19	0.78 ± 0.21	0.80 ± 0.19	0.95 ± 0.16	1.00 ± 0.01
	Stomach	0.75 ± 0.27	0.66 ± 0.28	0.76 ± 0.25	0.82 ± 0.29	0.99 ± 0.01
	Ureter	0.58 ± 0.27	0.46 ± 0.24	0.53 ± 0.26	0.74 ± 0.32	0.99 ± 0.01
	Vesicular glands	0.43 ± 0.26	0.31 ± 0.22	0.40 ± 0.28	0.63 ± 0.35	0.97 ± 0.03

Anatomy Segmentation in Laparoscopic Surgery

Figure 2: Pixel-wise organ segmentation with DeepLabv3-based structure-specific models trained on the respective organ subsets of the Dresden Surgical Anatomy Dataset. (a) Violin plot illustrations of performance metrics for DeepLabv3-based structure-specific segmentation models on the test dataset. The median and quartiles are illustrated as solid and dashed lines, respectively. **(b)** Example images from the test dataset with the highest IoUs for liver, pancreas, stomach, and ureter segmentation with DeepLabv3-based structure-specific segmentation models. Ground truth is displayed as blue line (upper panel), model segmentations are displayed as white overlay (lower panel). **(c)** Example images from the test dataset with the lowest IoUs for liver, pancreas, stomach, and ureter segmentation with DeepLabv3-based structure-specific segmentation models. Ground truth is displayed as blue line (upper panel), model segmentations are displayed as white overlay (lower panel).

Machine Learning-based anatomical structure segmentation in a combined model

Using annotated images from the Dresden Surgical Anatomy Dataset across anatomical structure classes, a combined model with a mutual encoder and organ-specific decoders was trained (Figure 1, Supplementary Table 2). Table 2 displays mean F1 score, IoU, precision, recall, and specificity for anatomical structure segmentation in the combined model.

The performance of the combined model based on DeepLabv3 was overall similar to that of structure-specific models (Table 1), with highest segmentation performance for the abdominal wall (mean IoU: 0.77 ± 0.15) and the small intestine (mean IoU: 0.72 ± 0.21), and the lowest performance for the pancreas (mean IoU: 0.23 ± 0.29), the ureter (IoU: 0.29 ± 0.22) and vesicular glands (IoU: 0.30 ± 0.23). In comparison to the respective structure-specific models, the combined model performed notably weaker in liver segmentation, while performance for the other anatomical structures was similar. The proportion of un- or entirely mispredicted images was largest in the ureter, the pancreas, the stomach, the abdominal vessel structures, and the vesicular glands (Figure 3 a). Similar to the DeepLabv3-based structure-specific models, trends towards an impact of segment size, uncommon angles of vision, endoscope lens soiling, blurry images, and presence of blood or smoke were seen when comparing image quality of well-predicted images (Figure 3 b) to images with poor or no prediction (Figure 3 c). Similar to the structure-specific models, segmentation performance of the SegFormer-based combined segmentation model was, overall, similar to that of DeepLabv3-based models. For segmentation of the spleen, there was a trend towards weaker performance of SegFormer-based models than DeepLabv3-based models (Table 2, Figure 3, Supplementary Figure 2).

For the DeepLabv3-based structure-specific models, inference on a single image with a resolution of 640 x 512 pixels required, on average, 71 ms on an Nvidia A5000, resulting in a frame rate of about 14 frames per second. As for structure-specific models of both architectures, SegFormer-based structure-specific semantic segmentation models operated considerably slower at an inference time of 102 ms per image, resulting in a frame rate of about 10 frames per second. This runtime includes all 11 decoders, meaning that segmentations for all organ classes are included.

Anatomy Segmentation in Laparoscopic Surgery

Table 2: Summary of performance metrics for anatomical structure segmentation using the DeepLabv3-based (a) and SegFormer-based (b) combined models (common encoder with structure-specific decoders) on the test dataset. For each metric, mean and standard deviation are displayed.

a	Anatomical structure	F1 score	IoU	Precision	Recall	Specificity
	Abdominal wall	0.86 ± 0.11	0.77 ± 0.15	0.81 ± 0.15	0.95 ± 0.09	0.95 ± 0.04
	Colon	0.75 ± 0.19	0.63 ± 0.21	0.71 ± 0.18	0.84 ± 0.23	0.95 ± 0.04
	Inferior mesenteric artery	0.53 ± 0.25	0.40 ± 0.21	0.52 ± 0.22	0.68 ± 0.32	0.99 ± 0.01
	Intestinal veins	0.46 ± 0.32	0.35 ± 0.27	0.70 ± 0.23	0.48 ± 0.36	1.00 ± 0.00
DeepLabv3 (Test)	Liver	0.65 ± 0.34	0.57 ± 0.33	0.76 ± 0.23	0.69 ± 0.38	0.98 ± 0.03
	Pancreas	0.32 ± 0.30	0.23 ± 0.24	0.61 ± 0.33	0.32 ± 0.35	0.99 ± 0.01
	Small intestine	0.81 ± 0.19	0.72 ± 0.21	0.81 ± 0.17	0.87 ± 0.23	0.96 ± 0.03
	Spleen	0.78 ± 0.24	0.69 ± 0.24	0.76 ± 0.18	0.89 ± 0.26	0.99 ± 0.01
	Stomach	0.63 ± 0.32	0.53 ± 0.29	0.68 ± 0.23	0.74 ± 0.37	0.98 ± 0.02
	Ureter	0.40 ± 0.28	0.29 ± 0.22	0.44 ± 0.27	0.56 ± 0.40	0.99 ± 0.01
	Vesicular glands	0.42 ± 0.27	0.30 ± 0.23	0.41 ± 0.30	0.56 ± 0.36	0.98 ± 0.02

b	Anatomical structure	F1 score	IoU	Precision	Recall	Specificity
	Abdominal wall	0.76 ± 0.24	0.66 ± 0.24	0.78 ± 0.15	0.86 ± 0.28	0.92 ± 0.07
	Colon	0.64 ± 0.27	0.52 ± 0.24	0.66 ± 0.19	0.78 ± 0.34	0.93 ± 0.06
	Inferior mesenteric artery	0.40 ± 0.24	0.28 ± 0.18	0.37 ± 0.21	0.68 ± 0.38	0.97 ± 0.02
	Intestinal veins	0.43 ± 0.33	0.33 ± 0.27	0.63 ± 0.24	0.52 ± 0.41	0.99 ± 0.01
SegFormer (Test)	Liver	0.62 ± 0.35	0.53 ± 0.32	0.76 ± 0.19	0.71 ± 0.40	0.95 ± 0.07
	Pancreas	0.35 ± 0.32	0.26 ± 0.25	0.56 ± 0.28	0.43 ± 0.41	0.99 ± 0.01
	Small intestine	0.78 ± 0.19	0.67 ± 0.20	0.74 ± 0.14	0.90 ± 0.23	0.94 ± 0.06
	Spleen	0.71 ± 0.24	0.59 ± 0.23	0.65 ± 0.21	0.89 ± 0.25	0.99 ± 0.01
	Stomach	0.65 ± 0.32	0.55 ± 0.29	0.71 ± 0.22	0.75 ± 0.36	0.98 ± 0.02
	Ureter	0.38 ± 0.29	0.27 ± 0.23	0.43 ± 0.27	0.55 ± 0.40	0.99 ± 0.01
	Vesicular glands	0.38 ± 0.25	0.26 ± 0.20	0.32 ± 0.25	0.66 ± 0.35	0.96 ± 0.03

Anatomy Segmentation in Laparoscopic Surgery

Figure 3: Pixel-wise organ segmentation with the DeepLabv3-based combined model trained on the Dresden Surgical Anatomy Dataset across anatomical structure classes with a common encoder and structure-specific decoders. (a) Violin plot illustrations of performance metrics for the DeepLabv3-based combined segmentation model on the test dataset. The median and quartiles are illustrated as solid and dashed lines, respectively. **(b)** Example images from the test dataset with the highest IoUs for liver, pancreas, stomach, and ureter segmentation with the DeepLabv3-based combined segmentation model. Ground truth is displayed as blue line (upper panel), model segmentations are displayed as white overlay (lower panel). **(c)** Example images from the test dataset with the lowest IoUs for liver, pancreas, stomach, and ureter segmentation with the DeepLabv3-based combined segmentation model. Ground truth is displayed as blue line (upper panel), model segmentations are displayed as white overlay (lower panel).

Performance of machine learning models on an external laparoscopic image dataset

To evaluate model robustness on an external dataset, we deployed the different organ segmentation models onto the publicly available LapGyn4 dataset (31) and qualitatively compared their performance. Overall, the combined models better reflected true anatomical constellations than the structure-specific models that generally lacked specificity. With respect to model architecture, the SegFormer-based segmentations were considerably more robust than the DeepLabv3-based models (Supplementary Figure 3).

Common mispredictions included confusion of liver and spleen, misinterpretation of organs that were not part of the training dataset (i.e. the gallbladder), and poor segmentation performance on less common images (i.e. extreme close-ups) (Supplementary Figure 3).

In summary, the SegFormer-based combined semantic segmentation model resulted in robust segmentations reproducing the true underlying anatomy. The remaining segmentation models provided substantially less specific and less robust segmentation outputs on the external dataset.

Performance of machine learning models in relation to human performance

To approximate the clinical value of the previously described algorithms for anatomical structure segmentation, the performances of the DeepLabv3-based and Transformer-based structure-specific and the combined models were compared to that of a cohort of 28 physicians, medical students, and persons with no medical background (Figure 4 a), and different degrees of experience in laparoscopic surgery (Figure 4 b). A vulnerable anatomical structure (24) with – measured by classical metrics of overlap (Tables 1 and 2) – comparably weak segmentation performance of the trained algorithms, the pancreas was selected as an example.

Comparing bounding box segmentations of the pancreas of human annotators, the medical and laparoscopy-specific experience of participants was mirrored by the respective IoUs describing the overlap between the pancreas annotation and the ground truth. The pancreas-specific segmentation models based on DeepLabv3 (IoU: 0.29) and SegFormer (IoU: 0.37) as well as the combined segmentation models based on DeepLabv3 (IoU: 0.21) and SegFormer (IoU: 0.32) outperformed at least 26 out of the 28 human participants (Figures 4 c and d). Overall, these

results demonstrate that the developed models have clinical potential to improve the recognition of vulnerable anatomical structures in laparoscopy.

Figure 4: Comparison of pancreas segmentation performance of the structure-specific and the combined semantic segmentation models with a cohort of 28 human participants. (a) Distribution of medical and non-medical professions among human participants. **(b)** Distribution of laparoscopy experience among human participants. **(c)** Waterfall chart displaying the average pancreas segmentation IoUs of participants with different professions as compared to the IoU generated by the structure-specific and the combined semantic segmentation models. **(d)** Waterfall chart displaying the average pancreas segmentation IoUs of participants with varying laparoscopy experience as compared to the IoU generated by the structure-specific and the combined semantic segmentation models.

DISCUSSION

In surgery, misinterpretation of visual cues can result in objectifiable errors with serious consequences (22). Machine learning models could augment identification of anatomical structures during minimally-invasive surgery and thereby contribute to a reduction of surgical risks. However, data scarcity and suboptimal dataset quality, among other factors, drastically restrict the clinical impact of applications in the field of surgical data science (17,32–36). Based on a robust public dataset providing 13195 laparoscopic images with segmentations of eleven intra-abdominal anatomical structures, this study explores the potential of machine learning for automated segmentation of these organs, and compares algorithmic segmentation quality to that of humans with varying experience in minimally-invasive abdominal surgery.

In summary, the presented findings suggest that machine learning-based segmentation of intraabdominal organs and anatomical structures is possible and has the potential to provide clinically valuable information. At an average runtime of 71 ms per image, corresponding to a frame rate of 14 frames per second, the combined DeepLabv3-based model would facilitate near-real-time identification of eleven anatomical structures. In contrast, the SegFormer-based model

is further from real-time performance at a runtime of 102 ms per image, resulting in a frame rate
of less than 10 frames per second. These runtimes mirror the performances of non-optimized
versions of the models, which can be significantly improved using methods such as TensorRT
from Nvidia. However, with respect to generalizability and robustness, we observed substantially
more accurate segmentation performance of the SegFormer-based models as compared to the
DeepLabv3-based models when deployed to an external conventional laparoscopic dataset.
Moreover, the structure-specific models exhibited a lack in accuracy and anatomical coherence,
which can be explained by their organ-specific training process.
Measured by classical metrics of overlap between segmentation and ground truth, predictions
were, overall, better for large and similar-appearing organs such as the abdominal wall, the liver,
the stomach, and the spleen as compared to smaller and more diverse-appearing organs such as
the pancreas, the ureter, or vesicular glands. Furthermore, poor image quality (i.e., images blurred
by camera movements, presence of blood or smoke in images) was linked to lower accuracy of
machine learning-based segmentations. Consequently, it is likely that a better nominal
performance of the machine learning models could be achieved through selection of images from
early phases of the surgery, in which such perturbations are not present. However, we purposely
did not exclude images with suboptimal image quality, as selection on image level would introduce
bias and thereby limit applicability. In this context, selection on patient level and on image level is
a common challenge in computer vision (37) that can lead to skewed reporting of outcomes and
poor performance on real-world data (33). Overall, our findings on the influence of image quality
on segmentation performance imply that computer vision studies in laparoscopy should be
carefully interpreted taking representativity and potential selection of underlying training and
validation data into consideration.
Measured by classical metrics of overlap (e.g., IoU, F1 score, precision, recall, specificity) that are
commonly used to evaluate segmentation performance, the structure-specific models and the
combined models provided comparable segmentation performances on the internal test dataset.
Interpretation of such metrics of overlap, however, represents a major challenge in computer
vision applications in medical domains such as dermatology and endoscopy (38–40) as well as
non-medical domains such as autonomous driving (41). In the specific use case of laparoscopic
surgery, evidence suggests that such technical metrics alone are not sufficient to characterize the
clinical potential and utility of segmentation algorithms (36,42). In this context, the subjective
clinical utility of a bounding box-based detection system recognizing the common bile duct and
the cystic duct at average precisions of 0.32 and 0.07, respectively, demonstrated by Tokuyasu
et al., supports this hypothesis (12). In colorectal surgery, anatomical misinterpretation during

splenic flexure mobilization can result in iatrogenic lesions to the pancreas (24). In the presented

analysis, the trained structure-specific and combined machine learning algorithms outperformed

all human participants in the specific task of bounding box segmentation of the pancreas except

for two surgical specialists with over 10 years of experience. This suggests that even for structures

such as the pancreas with seemingly poor segmentation quality (segmentation IoU of the best-

performing model: 0.37 ± 0.28 in the test set) have the potential to provide clinically valuable help

in anatomy recognition. Notably, the best average IoUs achieved in this comparative study were

0.37 (for the SegFormer-based structure-specific model) and 0.36 (for the best human participant),

which would both be considered less reliable segmentation quality measures on paper. This

encourages further discussion about metrics for segmentation quality assessment in clinical AI. In

the future, the potential of the described dataset (25) and organ segmentation algorithms could

be exploited for educational purposes (43,44), for guidance systems facilitating real-time detection

of risk and target structures (19,42,45,46), or as an auxiliary function integrated in more complex

surgical assistance systems, such as guidance systems relying on automated liver registration

(47).

The limitations of this work are mostly related to the dataset and general limitations of machine

learning-based segmentation: First, the Dresden Surgical Anatomy Dataset is a monocentric

dataset based on 32 robot-assisted rectal surgeries. Therefore, the images used for algorithm

training and validation originate from one set of hardware and display organs from specific angles.

As a consequence, given the lack of a laparoscopic image dataset with similarly rigorous organ

annotations, generalizability and transferability of the presented findings to other centers and other

minimally-invasive abdominal surgeries, particularly non-robotic procedures, could only be

qualitatively investigated. Second, annotations were required for training of machine learning

algorithms, potentially inducing some bias towards the way that organs were annotated in the

algorithms, which may differ from individual healthcare professionals' way of recognizing an organ.

This is particularly relevant for organs such as the ureters or the pancreas, which often appear

covered by layers of tissue. Here, computer vision-based algorithms that solely consider the

laparoscopic images provided by the Dresden Surgical Anatomy Dataset for identification of risk

structures will only be able to identify an organ once it is visible. For an earlier recognition of such

hidden risk structures, more training data with meaningful annotations would be necessary.

Importantly, the presented comparison to human performance focused on segmentation of visible

anatomy as well, neglecting that humans (and possibly computers, too) could already identify a

risk structure hidden underneath layers of tissue. Third, the dataset only includes individual

annotated images. In some structures such as the ureter, video data offers considerably more

Anatomy Segmentation in Laparoscopic Surgery

information than still image data. In this context, it is conceivable that an incorporation of temporal

aspects could result in major improvements of both human and algorithm recognition performance.

While the presented machine learning models show promise in improving the identification of

anatomical structures in laparoscopy, their clinical utility still needs to be explored. Successful

adoption of new technologies in surgery depends on factors beyond segmentation performance,

runtime and generalizability, such as visualization of intraoperative decision support (48), human-

machine interaction (49), and interface design. Therefore, interdisciplinary collaboration is critical

to better understand respective surgeon needs. Moreover, prospective trials are needed to

determine the impact of these factors on clinical outcomes. The existing limitations

notwithstanding, the presented study represents an important addition to the growing body of

research on medical image analysis in laparoscopic surgery, particularly by linking technical

metrics to human performance.

In conclusion, this study demonstrates that machine learning methods have the potential to

provide clinically relevant near-real-time assistance in anatomy recognition in minimally-invasive

surgery. This study is the first to use the recently published Dresden Surgical Anatomy Dataset,

providing baseline algorithms for organ segmentation and evaluating the clinical relevance of such

algorithms by introducing more clinically meaningful comparators beyond classical computer

vision metrics. Future research should investigate other segmentation methods, the potential to

integrate high-level anatomical knowledge into segmentation models (37), the transferability of

these results to other surgical procedures, and the clinical impact of real-time surgical assistance

systems and didactic applications based on automated segmentation algorithms. Furthermore,

seeing the DeepLabv3-based models outperform the SegFormer-based models in terms of run-

time, but lacking in accuracy and generalizability, future research could focus on combining the

two, in order to harness the best of both worlds.

ACKNOWLEDGMENTS

Assistance with the study

The authors gratefully acknowledge excellent project coordination by Dr. Elisabeth Fischermeier and Dr. Grit Krause-Jüttler.

Financial support and sponsorship

This work has been funded by the Else Kröner Fresenius Center for Digital Health (EKFZ), Dresden, Germany (project "CoBot"), by the German Research Foundation DFG within the Cluster of Excellence EXC 2050: "Center for Tactile Internet with Human-in-the-Loop (CeTI)" (project number 390696704) and by the German Federal Ministry of Health (BMG) within the "Surgomics" project (grant number BMG 2520DAT82). Furthermore, FRK received funding from the Medical Faculty of the Technical University Dresden within the MedDrive Start program (grant number 60487) and from the Joachim Herz Foundation (Add-On Fellowship for Interdisciplinary Life Science). FMR received a doctoral student scholarship from the *Carus Promotionskolleg* Dresden.

Conflicts of interest

The authors declare no conflicts of interest.

Presentation

None.

Anatomy Segmentation in Laparoscopic Surgery

REFERENCES

1. Wang P, Liu X, Berzin TM, Glissen Brown JR, Liu P, Zhou C, et al. Effect of a deep-learning computer-aided detection system on adenoma detection during colonoscopy (CADe-DB trial): a double-blind randomised study. Lancet Gastroenterol Hepatol [Internet]. 2020 Apr 1 [cited 2022 Oct 3];5(4):343–51. Available from: http://www.thelancet.com/article/S246812531930411X/fulltext

2. Wang P, Berzin TM, Glissen Brown JR, Bharadwaj S, Becq A, Xiao X, et al. Real-time automatic detection system increases colonoscopic polyp and adenoma detection rates: a prospective randomised controlled study. Gut [Internet]. 2019 Oct 1 [cited 2022 Oct 3];68(10):1813–9. Available from: https://gut.bmj.com/content/68/10/1813

3. Kather JN, Heij LR, Grabsch HI, Loeffler C, Echle A, Muti HS, et al. Pan-cancer image-based detection of clinically actionable genetic alterations. Nat Cancer 2020 18 [Internet]. 2020 Jul 27 [cited 2022 Oct 3];1(8):789–99. Available from: https://www.nature.com/articles/s43018-020-0087-6

4. Esteva A, Kuprel B, Novoa RA, Ko J, Swetter SM, Blau HM, et al. Dermatologist-level classification of skin cancer with deep neural networks. Nature [Internet]. 2017 Feb 2 [cited 2021 Feb 15];542(7639):115–8. Available from: https://pubmed.ncbi.nlm.nih.gov/28117445/

5. Simillis C, Lal N, Thoukididou SN, Kontovounisios C, Smith JJ, Hompes R, et al. Open Versus Laparoscopic Versus Robotic Versus Transanal Mesorectal Excision for Rectal Cancer: A Systematic Review and Network Meta-analysis. Ann Surg [Internet]. 2019 Jul 1 [cited 2022 Oct 6];270(1):59–68. Available from: https://pubmed.ncbi.nlm.nih.gov/30720507/

6. Zhao JJ, Syn NL, Chong C, Tan HL, Ng JYX, Yap A, et al. Comparative outcomes of needlescopic, single-incision laparoscopic, standard laparoscopic, mini-laparotomy, and open cholecystectomy: A systematic review and network meta-analysis of 96 randomized controlled trials with 11,083 patients. Surgery [Internet]. 2021 Oct 1 [cited 2022 Oct 6];170(4):994–1003. Available from: https://pubmed.ncbi.nlm.nih.gov/34023139/

7. Luketich JD, Pennathur A, Awais O, Levy RM, Keeley S, Shende M, et al. Outcomes after minimally invasive esophagectomy: review of over 1000 patients. Ann Surg [Internet]. 2012 Jul [cited 2022 Oct 6];256(1):95–103. Available from: https://pubmed.ncbi.nlm.nih.gov/22668811/

8. Thomson JE, Kruger D, Jann-Kruger C, Kiss A, Omoshoro-Jones JAO, Luvhengo T, et al. Laparoscopic versus open surgery for complicated appendicitis: a randomized controlled

trial to prove safety. Surg Endosc [Internet]. 2015 Jul 19 [cited 2022 Oct 6];29(7):2027–32.

Available from: https://pubmed.ncbi.nlm.nih.gov/25318368/

9.   Islam M, Atputharuban DA, Ramesh R, Ren H. Real-time instrument segmentation in

robotic surgery using auxiliary supervised deep adversarial learning. IEEE Robot Autom

Lett. 2019 Apr 1;4(2):2188–95.

10.   Roß T, Reinke A, Full PM, Wagner M, Kenngott H, Apitz M, et al. Comparative validation

of multi-instance instrument segmentation in endoscopy: results of the ROBUST-MIS 2019

challenge. Med Image Anal [Internet]. 2020 Nov [cited 2021 Mar 21];70:101920. Available

from: https://pubmed.ncbi.nlm.nih.gov/33676097/

11.   Shvets AA, Rakhlin A, Kalinin AA, Iglovikov VI. Automatic Instrument Segmentation in

Robot-Assisted Surgery using Deep Learning. In: Proceedings - 17th IEEE International

Conference on Machine Learning and Applications, ICMLA 2018. Institute of Electrical and

Electronics Engineers Inc.; 2019. p. 624–8.

12.   Tokuyasu T, Iwashita Y, Matsunobu Y, Kamiyama T, Ishikake M, Sakaguchi S, et al.

Development of an artificial intelligence system using deep learning to indicate anatomical

landmarks during laparoscopic cholecystectomy. Surg Endosc 2020 354 [Internet]. 2020

Apr   18   [cited   2021   Jul   16];35(4):1651–8.   Available   from:

https://link.springer.com/article/10.1007/s00464-020-07548-x

13.   Mascagni P, Vardazaryan A, Alapatt D, Urade T, Emre T, Fiorillo C, et al. Artificial

Intelligence for Surgical Safety: Automatic Assessment of the Critical View of Safety in

Laparoscopic Cholecystectomy Using Deep Learning. Ann Surg. 2020;

14.   Jin A, Yeung S, Jopling J, Krause J, Azagury D, Milstein A, et al. Tool Detection and

Operative Skill Assessment in Surgical Videos Using Region-Based Convolutional Neural

Networks. Proc - 2018 IEEE Winter Conf Appl Comput Vision, WACV 2018 [Internet]. 2018

Feb   24   [cited   2021   Jul   14];2018-January:691–9.   Available   from:

https://arxiv.org/abs/1802.08774v2

15.   Funke I, Bodenstedt S, Oehme F, von Bechtolsheim F, Weitz J, Speidel S. Using 3D

Convolutional Neural Networks to Learn Spatiotemporal Features for Automatic Surgical

Gesture Recognition in Video. Med Image Comput Comput Assist Interv – MICCAI 2019

Lect Notes Comput Sci [Internet]. 2019 [cited 2022 Aug 19];11768:467–75. Available from:

https://link.springer.com/chapter/10.1007/978-3-030-32254-0_52

16.   Lavanchy JL, Zindel J, Kirtac K, Twick I, Hosgor E, Candinas D, et al. Automation of surgical

skill assessment using a three-stage machine learning algorithm. Sci Reports 2021 111

[Internet].   2021   Mar   4   [cited   2022   Oct   6];11(1):1–9.   Available   from:

https://www.nature.com/articles/s41598-021-84295-6

17.   Maier-Hein L, Eisenmann M, Sarikaya D, März K, Collins T, Malpani A, et al. Surgical data
science – from concepts toward clinical translation. Med Image Anal. 2022 Feb
1;76:102306.

18.   Kolbinger FR, Leger S, Carstens M, Rinner FM, Krell S, Chernykh A, et al. Artificial
Intelligence for context-aware surgical guidance in complex robot-assisted oncological
procedures: an exploratory feasibility study. medRxiv [Internet]. 2022 May 3 [cited 2022
May   13];2022.05.02.22274561.   Available   from:
https://www.medrxiv.org/content/10.1101/2022.05.02.22274561v1

19.   Madani A, Namazi B, Altieri MS, Hashimoto DA, Rivera AM, Pucher PH, et al. Artificial
Intelligence for Intraoperative Guidance. Ann Surg. 2020;

20.   Fecso AB, Szasz P, Kerezov G, Grantcharov TP. The effect of technical performance on
patient outcomes in surgery. Ann Surg [Internet]. 2017 Mar 1 [cited 2022 Oct 6];265(3):492–
501.   Available   from:
https://journals.lww.com/annalsofsurgery/Fulltext/2017/03000/The_Effect_of_Technical_P
erformance_on_Patient.11.aspx

21.   Mazzocco K, Petitti DB, Fong KT, Bonacum D, Brookey J, Graham S, et al. Surgical team
behaviors and patient outcomes. Am J Surg. 2009 May 1;197(5):678–85.

22.   Suliburk JW, Buck QM, Pirko CJ, Massarweh NN, Barshes NR, Singh H, et al. Analysis of
Human Performance Deficiencies Associated With Surgical Adverse Events. JAMA Netw
Open [Internet]. 2019 Jul 3 [cited 2022 Oct 3];2(7):e198067–e198067. Available from:
https://jamanetwork.com/journals/jamanetworkopen/fullarticle/2740065

23.   Adelman MR, Bardsley TR, Sharp HT. Urinary Tract Injuries in Laparoscopic Hysterectomy:
A Systematic Review. J Minim Invasive Gynecol. 2014 Jul 1;21(4):558–66.

24.   Freund MR, Kent I, Horesh N, Smith T, Emile SH, Wexner SD. Pancreatic injuries following
laparoscopic splenic flexure mobilization. Int J Colorectal Dis [Internet]. 2022 Apr 1 [cited
2023   Mar   20];37(4):967–71.   Available   from:
https://link.springer.com/article/10.1007/s00384-022-04112-y

25.   Carstens M, Rinner FM, Bodenstedt S, Jenke AC, Weitz J, Distler M, et al. The Dresden
Surgical Anatomy Dataset for Abdominal Organ Segmentation in Surgical Data Science.
Sci Data [Internet]. 2023 Jan 12 [cited 2023 Jan 12];10(1):1–8. Available from:
https://www.nature.com/articles/s41597-022-01719-2

26.   Chen L-C, Papandreou G, Schroff F, Adam H. Rethinking Atrous Convolution for Semantic
Image Segmentation. arXiv [Internet]. 2017 Jun 17 [cited 2022 Oct 10]; Available from:

620    https://arxiv.org/abs/1706.05587v3

621 27. Lin TY, Maire M, Belongie S, Hays J, Perona P, Ramanan D, et al. Microsoft COCO:
622    Common Objects in Context. Lect Notes Comput Sci (including Subser Lect Notes Artif
623    Intell Lect Notes Bioinformatics) [Internet]. 2014 May 1 [cited 2022 Nov 11];8693
624    LNCS(PART 5):740–55. Available from: https://arxiv.org/abs/1405.0312v3

625 28. Xie E, Wang W, Yu Z, Anandkumar A, Alvarez JM, Luo P. SegFormer: Simple and Efficient
626    Design for Semantic Segmentation with Transformers. Adv Neural Inf Process Syst. 2021
627    Dec 6;34:12077–90.

628 29. Cordts M, Omran M, Ramos S, Rehfeld T, Enzweiler M, Benenson R, et al. The Cityscapes
629    Dataset for Semantic Urban Scene Understanding. Proc IEEE Conf Comput Vis Pattern
630    Recognit. 2016;

631 30. Loshchilov I, Hutter F. Decoupled Weight Decay Regularization. 7th Int Conf Learn
632    Represent ICLR 2019 [Internet]. 2017 Nov 14 [cited 2022 Nov 9]; Available from:
633    https://arxiv.org/abs/1711.05101v3

634 31. Leibetseder A, Petscharnig S, Primus MJ, Kletz S, Münzer B, Schoeffmann K, et al.
635    LapGyn4: A Dataset for 4 Automatic Content Analysis Problems in the Domain of
636    Laparoscopic Gynecology. Proc 9th ACM Multimed Syst Conf [Internet]. 2018 [cited 2021
637    Jul 19];18. Available from: https://doi.org/10.1145/3204949.3208127

638 32. Reddy CL, Mitra S, Meara JG, Atun R, Afshar S. Artificial Intelligence and its role in surgical
639    care in low-income and middle-income countries. Lancet Digit Heal. 2019 Dec 1;1(8):e384–
640    6.

641 33. Moglia A, Georgiou K, Georgiou E, Satava RM, Cuschieri A. A systematic review on artificial
642    intelligence in robot-assisted surgery. Int J Surg. 2021 Nov 1;95:106151.

643 34. Anteby R, Horesh N, Soffer S, Zager Y, Barash Y, Amiel J, et al. Deep learning visual
644    analysis in laparoscopic surgery: a systematic review and diagnostic test accuracy meta-
645    analysis. Surg Endosc [Internet]. 2021 Apr 1 [cited 2021 Jul 16];35(4):1521–33. Available
646    from: https://pubmed.ncbi.nlm.nih.gov/33398560/

647 35. Kuo RYL, Harrison CJ, Jones BE, Geoghegan L, Furniss D. Perspectives: A surgeon's
648    guide to machine learning. Int J Surg. 2021 Oct 1;94:106133.

649 36. Reinke A, Tizabi MD, Sudre CH, Eisenmann M, Rädsch T, Baumgartner M, et al. Common
650    Limitations of Image Processing Metrics: A Picture Story. arXiv [Internet]. 2021 Apr 12 [cited
651    2022 May 13]; Available from: https://arxiv.org/abs/2104.05642v4

652 37. Jin C, Udupa JK, Zhao L, Tong Y, Odhner D, Pednekar G, et al. Object recognition in
653    medical images via anatomy-guided deep learning. Med Image Anal. 2022 Oct

1;81:102527.

38.   Renard F, Guedria S, Palma N De, Vuillerme N. Variability and reproducibility in deep
learning for medical image segmentation. Sci Rep [Internet]. 2020 Aug 13 [cited 2022 Oct
16];10(1):1–16. Available from: https://www.nature.com/articles/s41598-020-69920-0

39.   Powers DMW, Ailab. Evaluation: from precision, recall and F-measure to ROC,
informedness, markedness and correlation. arXiv [Internet]. 2020 Oct 11 [cited 2022 Oct
16]; Available from: https://arxiv.org/abs/2010.16061v1

40.   Parikh RB, Teeple S, Navathe AS. Addressing Bias in Artificial Intelligence in Health Care.
JAMA [Internet]. 2019 Dec 24 [cited 2022 Oct 16];322(24):2377–8. Available from:
https://jamanetwork.com/journals/jama/fullarticle/2756196

41.   Zhang Y, Mehta S, Caspi A. Rethinking Semantic Segmentation Evaluation for
Explainability and Model Selection. 2021 Jan 21 [cited 2021 Jul 27]; Available from:
https://arxiv.org/abs/2101.08418v1

42.   Hashimoto DA, Rosman G, Witkowski ER, Stafford C, Navarette-Welton AJ, Rattner DW,
et al. Computer Vision Analysis of Intraoperative Video: Automated Recognition of
Operative Steps in Laparoscopic Sleeve Gastrectomy. Ann Surg [Internet]. 2019 Sep 1
[cited   2021   Jul   16];270(3):414–21.   Available   from:
https://journals.lww.com/annalsofsurgery/Fulltext/2019/09000/Computer_Vision_Analysis
_of_Intraoperative_Video_.3.aspx

43.   Hu YY, Mazer LM, Yule SJ, Arriaga AF, Greenberg CC, Lipsitz SR, et al. Complementing
Operating Room Teaching With Video-Based Coaching. JAMA Surg [Internet]. 2017 Apr 1
[cited   2022   Oct   27];152(4):318–25.   Available   from:
https://jamanetwork.com/journals/jamasurgery/fullarticle/2593311

44.   Mizota T, Anton NE, Stefanidis D. Surgeons see anatomical structures faster and more
accurately compared to novices: Development of a pattern recognition skill assessment
platform. Am J Surg. 2019 Feb 1;217(2):222–7.

45.   Ward TM, Mascagni P, Ban Y, Rosman G, Padoy N, Meireles O, et al. Computer vision in
surgery. Surgery [Internet]. 2021 May 1 [cited 2022 Oct 27];169(5):1253–6. Available from:
https://pubmed.ncbi.nlm.nih.gov/33272610/

46.   Chopra H, Baig AA, Arora S, Singh I, Kaur A, Emran T Bin. Artificial intelligence in surgery:
Modern trends. Int J Surg. 2022 Oct 1;106:106883.

47.   Docea R, Pfeiffer M, Bodenstedt S, Kolbinger FR, Höller L, Wittig I, et al. Simultaneous
localisation and mapping for laparoscopic liver navigation : a comparative evaluation study.
In: Linte CA, Siewerdsen JH, editors. Medical Imaging 2021: Image-Guided Procedures,

Robotic Interventions, and Modeling [Internet]. SPIE; 2021 [cited 2021 Feb 25]. p. 8.

Available from: https://www.spiedigitallibrary.org/conference-proceedings-of-

spie/11598/2582121/Simultaneous-localisation-and-mapping-for-laparoscopic-liver-

navigation/10.1117/12.2582121.full

48.   Shaalan D, Jusoh S. Visualization in Medical System Interfaces: UX Guidelines. Proc 12th

Int Conf Electron Comput Artif Intell ECAI 2020. 2020 Jun 1;

49.   Henry KE, Kornfield R, Sridharan A, Linton RC, Groh C, Wang T, et al. Human–machine

teaming is key to AI adoption: clinicians' experiences with a deployed machine learning

system. npj Digit Med [Internet]. 2022 Jul 21 [cited 2023 Mar 30];5(1):1–6. Available from:

https://www.nature.com/articles/s41746-022-00597-7

ABBREVIATIONS

700 AI   Artificial Intelligence

701 IoU   Intersection-over-Union

702 SD   Standard deviation

AUTHOR CONTRIBUTIONS

705 FRK, JW, MD, SS, and SB conceptualized the study. FRK, FMR, and MC collected and annotated
706 clinical and video data and contributed to data analysis. ACJ, SK, SL, and SB implemented and
707 trained the neural networks and contributed to data analysis. JW, MD, and SS supervised the
708 project, provided infrastructure and gave important scientific input. FRK drafted the initial
709 manuscript text. All authors reviewed, edited, and approved the final manuscript.