Better than humans? Machine learning-based anatomy recognition in minimally-invasive abdominal surgery

Fiona R. Kolbinger[1,2,3,✉], Franziska M. Rinner[1], Alexander C. Jenke[4], Matthias Carstens[1],
Stefan Leger[3,4], Marius Distler[1,2], Jürgen Weitz[1,2,3,5], Stefanie Speidel[3,4,5],
Sebastian Bodenstedt[4,5,✉]

[1] Department of Visceral, Thoracic and Vascular Surgery, University Hospital and Faculty of Medicine Carl Gustav Carus, Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany

[2] National Center for Tumor Diseases (NCT/UCC), Dresden, Germany: German Cancer Research Center (DKFZ), Heidelberg, Germany; Faculty of Medicine and University Hospital Carl Gustav Carus, Technische Universität Dresden, Dresden, Germany; Helmholtz-Zentrum Dresden-Rossendorf (HZDR), Fetscherstraße 74, 01307 Dresden, Germany

[3] Else Kröner Fresenius Center for Digital Health (EKFZ), Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany

[4] Division of Translational Surgical Oncology, National Center for Tumor Diseases (NCT), Partner Site Dresden, Fetscherstraße 74, 01307 Dresden, Germany

[5] Cluster of Excellence "Centre for Tactile Internet with Human-in-the-Loop" (CeTI), Technische Universität Dresden, 01062, Dresden, Germany

✉ Corresponding authors:

Dr. Fiona Kolbinger, Department of Visceral, Thoracic and Vascular Surgery, University Hospital and Faculty of Medicine Carl Gustav Carus, Technische Universität Dresden, Fetscherstraße 74, 01307 Dresden, Germany

☎ +49 (0) 351 458 19624

✉ fiona.kolbinger@uniklinikum-dresden.de

Dr. Sebastian Bodenstedt, Division of Translational Surgical Oncology, National Center for Tumor Diseases (NCT/UCC), Partner Site Dresden, Fetscherstrasse 74, 01307 Dresden, Germany

☎ +49 (0) 351 5413

✉ sebastian.bodenstedt@nct-dresden.de

Abstract

Background: Lack of anatomy recognition represents a clinically relevant risk factor in abdominal surgery. While machine learning methods have the potential to aid in recognition of visible patterns and structures, limited availability and diversity of (annotated) laparoscopic image data restrict the clinical potential of such applications in practice. This study explores the potential of machine learning algorithms to identify and delineate abdominal organs and anatomical structures using a robust and comprehensive dataset, and compares algorithm performance to that of humans.

Methods: Based on the Dresden Surgical Anatomy Dataset providing 13195 laparoscopic images with pixel-wise segmentations of eleven anatomical structures, two machine learning algorithms were developed: individual segmentation algorithms for each structure, and a combined algorithm with a common encoder and structure-specific decoders. Performance was assessed using F1 score, Intersection-over-Union (IoU), precision, recall, and specificity. Using the example of pancreas segmentation on a sample dataset of 35 images, algorithm performance was compared to that of a cohort of 28 physicians, medical students, and medical laypersons.

Results: Mean IoU for segmentation of intraabdominal structures ranged from 0.28 to 0.83 and from 0.22 to 0.78 for the structure-specific and the combined semantic segmentation model, respectively. Average inference for the structure-specific (one anatomical structure) and the combined model (eleven anatomical structures) took 28 ms and 71 ms, respectively. Both models outperformed 26 out of 28 human participants in pancreas segmentation.

Conclusions: Machine learning methods have the potential to provide relevant assistance in anatomy recognition in minimally-invasive surgery in near-real-time. Future research should investigate the educational value and subsequent clinical impact of respective assistance systems.

Introduction

Computer vision describes the computerized analysis of digital images aiming at the automation of human visual capabilities, most commonly using machine learning methods, in particular deep learning. This approach has transformed medicine in recent years, with successful applications including computer-aided diagnosis of colonic polyp dignity in endoscopy[1,2], detection of clinically actionable genetic alterations in histopathology[3], and melanoma detection in dermatology[4]. Availability of large amounts of training data is the defining prerequisite for successful application of deep learning methods. With the establishment of laparoscopy as the gold standard for a variety of surgical procedures[5–8] and the increasing availability of computing resources, these concepts have gradually been applied to abdominal surgery. The overwhelming majority of research efforts in the field of Artificial Intelligence (AI)-based analysis of intraoperative surgical imaging data (i.e. video data from laparoscopic or open surgeries) has focused on classifying images with respect to the presence and/or location of previously annotated surgical instruments or anatomical structures[9–13] or on analysis of surgical proficiency[14–16] based on recorded procedures. However, almost all research endeavors in the field of computer vision in laparoscopic surgery have concentrated on preclinical stages and to date, no AI model based on intraoperative surgical imaging data could demonstrate a palpable clinical benefit.[17] Among the studies closest to clinical application are recent works on identification of instruments and hepatobiliary anatomy during cholecystectomy for automated assessment of the critical view of safety[13], and on the automated segmentation of safe and unsafe preparation zones during cholecystectomy[18].

In surgery, patient outcome heavily depends on experience and performance of the surgical team.[19,20] In a recent analysis of Human Performance Deficiencies in major cardiothoracic, vascular, abdominal transplant, surgical oncology, acute care, and general surgical operations, more than half of the cases with postoperative complications were associated with identifiable human error. Among these errors, lack of recognition (including misidentified anatomy) accounted for 18.8%, making it the most common Human Performance Deficiency overall.[21] While AI-based systems identifying anatomical risk and target structures would theoretically have the potential to alleviate this risk, limited availability and diversity of (annotated) laparoscopic image data drastically restrict the clinical potential of such applications in practice.

To advance and diversify the applications of computer vision in laparoscopic surgery, we have recently published the Dresden Surgical Anatomy Dataset[22], providing 13195 laparoscopic images with high-quality annotations of the presence and exact location of eleven intraabdominal anatomical structures: abdominal wall, colon, intestinal vessels (inferior mesenteric artery and inferior mesenteric vein with their subsidiary vessels), liver, pancreas, small intestine, spleen,

92 stomach, ureter and vesicular glands. Here, we present the first study evaluating automated
93 detection and localization of organs and anatomical structures in laparoscopic view based on this
94 dataset, and, using the example of delineation of the pancreas, compare algorithm performance
95 to that of humans.
96

Methods

Patient cohort

Video data from 32 robot-assisted anterior rectal resections or rectal extirpations were gathered
at the University Hospital Carl Gustav Carus Dresden between February 2019 and February 2021.
All included patients had a clinical indication for the surgical procedure, recommended by an
interdisciplinary tumor board. The procedures were performed using the da Vinci® Xi system
(Intuitive Surgical, Sunnyvale, CA, USA) with a standard Da Vinci® Xi/X Endoscope with Camera
(8 mm diameter, 30˚ angle, Intuitive Surgical, Sunnyvale, CA, USA, Item code 470057). Surgeries
were recorded using the CAST system (Orpheus Medical GmBH, Frankfurt a.M., Germany). Each
record was saved at a resolution of 1920 x 1080 pixels in MPEG-4 format.
All experiments were performed in accordance with the ethical standards of the Declaration of
Helsinki and its later amendments. The local Institutional Review Board (ethics committee at the
Technical University Dresden) reviewed and approved this study (approval number: BO-EK-
137042018). The trial was registered on clinicaltrials.gov (trial registration ID: NCT05268432).
Written informed consent to laparoscopic image data acquisition, data annotation, data analysis,
and anonymized data publication was obtained from all participants. Before publication, all data
was anonymized according to the general data protection regulation of the European Union.
114

Dataset

Based on the full-length surgery recordings and respective temporal annotations of organ visibility,
individual image frames were extracted and annotated as described previously.[22] The resulting
Dresden Surgical Anatomy Dataset comprises 13195 distinct images with pixel-wise
segmentations of eleven anatomical structures: abdominal wall, colon, intestinal vessels (inferior
mesenteric artery and inferior mesenteric vein with their subsidiary vessels), liver, pancreas, small
intestine, spleen, stomach, ureter and vesicular glands. Moreover, the dataset comprises binary
annotations of the presence of each of these organs for each image. The dataset is publicly
available via the following link: https://www.nature.com/articles/s41597-022-01719-2.
124

For machine learning purposes, the Dresden Surgical Anatomy Dataset was split into training,
validation, and test data as follows (Figure 1):
— Training set (at least 12 surgeries per anatomical structure): surgeries 1, 4, 5, 6, 8, 9, 10,
12, 15, 16, 17, 19, 22, 23, 24, 25, 27, 28, 29, 30, 31.
— Validation set (3 surgeries per anatomical structure): surgeries 3, 21, 26.
— Test set (5 surgeries per anatomical structure): surgeries 2, 7, 11, 13, 14, 18, 20, 32.
This split is proposed for future works using the Dresden Surgical Anatomy Dataset to reproduce
the variance of the entire dataset within each subset, and to ensure comparability regarding clinical
variables between the training, the validation, and the test set. Surgeries for the test set were
selected to minimize variance regarding the number of frames over the segmented classes. Out
of the remaining surgeries, the validation set was separated from the training set using the same
criterion.
137

Structure-specific semantic segmentation model
To segment each anatomical structure, a separate convolutional neural network for segmentation
a DeeplabV3[23] model with a ResNet50 backbone with default PyTorch pretraining on the COCO
dataset[24], was used. The networks were trained using cross-entropy loss and the AdamW
optimizer[25] for 100 epochs with a starting learning rate of 10^{-4} and a linear learning rate scheduler
decreasing the learning rate by 0.9 every 10 epochs. For data augmentation, we applied random
scaling and rotation, as well as brightness adjustments. The final model for each organ was
selected via the Intersection-over-Union (IoU) on the validation dataset and evaluated using the
Dresden Surgical Anatomy Dataset with the abovementioned training-validation-test split (Figure
1).
Segmentation performance was assessed using F1 score, IoU, precision, recall, and specificity
on the test folds. These parameters are commonly used technical measures of prediction
exactness, ranging from 0 (least exact prediction) to 1 (entirely correct prediction without any
misprediction).
152

Combined semantic segmentation model
A convolutional neural network with a common encoder and eleven decoders for combined
segmentation of the eleven anatomical structures was trained. The used architecture is an
extension of DeepLabV3[23]. A shared ResNet50 backbone with default PyTorch pretraining on the
COCO dataset[24], was used. For each class, a DeepLabV3 decoder was then run on the features

extracted from a given image by the backbone. As the images are only annotated for binary
classes, the loss is only calculated for the decoder associated with the class annotated in a given
image. The remaining training procedure was identical to the structure-specific model. The model
was trained and evaluated using the Dresden Surgical Anatomy Dataset with the abovementioned
training-validation-test split (Figure 1).
Segmentation performance was assessed using F1 score, IoU, precision, recall, and specificity
on the test folds.[26]
165

Comparative evaluation of algorithmic and human performance
To determine the clinical potential of automated segmentation of anatomical risk structures, the
segmentation performance of 28 humans was compared to that of the structure-specific and the
combined semantic segmentation models using the example of the pancreas. The local
Institutional Review Board (ethics committee at the Technical University Dresden) reviewed and
approved this study (approval number: BO-EK-566122021). All participants provided written
informed consent to anonymous study participation, data acquisition and analysis, and publication.
In total, 28 participants (physician and non-physician medical staff, medical students, and medical
laypersons) marked the pancreas in 35 images from the Dresden Surgical Anatomy Dataset[22] with
bounding boxes. These images originated from 26 different surgeries, and the pancreas was
visible in 16 of the 35 images. Each of the previously selected 35 images was shown once, the
order being arbitrarily chosen but identical for all participants. The open-source annotation
software Computer Vision Annotation Tool (CVAT) was used for annotations. In cases where the
pancreas was seen in multiple, non-connected locations in the image, participants were asked to
create separate bounding boxes for each area.
Based on the structure-specific and the combined semantic segmentation models, axis-aligned
bounding boxes marking the pancreas were generated in the 35 images from the pixel-wise
segmentation. To guarantee that the respective images were not part of the training data, four-
fold cross validation was used, i.e. the origin surgeries were split into four equal-sized batches,
and algorithms were trained on three batches that did not contain the respective origin image
before being applied to segmentation.
To compare human and algorithm performance, the bounding boxes created by each participant
and the structure-specific as well as the combined semantic segmentation models were compared
to bounding boxes derived from the Dresden Surgical Anatomy Dataset, which were defined as
ground truth. IoU between the manual or algorithmical bounding box and the ground truth was
used to compare segmentation accuracy.

Data Availability

The Dresden Surgical Anatomy Dataset is publicly available via the following link:

https://www.nature.com/articles/s41597-022-01719-2. All other data generated and analyzed

during the current study are available from the corresponding authors on reasonable request. To

gain access, data requestors will need to sign a data access agreement.

Code Availability

The most relevant scripts used for dataset compilation are publicly available via the following link:

https://zenodo.org/record/6958337#.YzsBdnZBzOg.

Results

Machine Learning-based anatomical structure segmentation in structure-specific models

Structure-specific multi-layer convolutional neural networks (Figure 1) were trained to segment

the abdominal wall, the colon, intestinal vessels (inferior mesenteric artery and inferior mesenteric

vein with their subsidiary vessels), the liver, the pancreas, the small intestine, the spleen, the

stomach, the ureter, and vesicular glands. Table 1 displays mean F1 score, IoU, precision, recall,

and specificity for individual anatomical structures as predicted by the structure-specific

algorithms. Out of the analyzed segmentation models, performance was lowest for the vesicular

glands (mean IoU: 0.28 ± 0.21) and the pancreas (mean IoU: 0.28 ± 0.27), while excellent

predictions were achieved for the abdominal wall (mean IoU: 0.83 ± 0.14) and the small intestine

(mean IoU: 0.80 ± 0.18). In segmentation of the ureter, the vesicular glands and the intestinal

vessel structures, there was a relevant proportion of images with no detection or no overlap

between ground truth, while for all remaining anatomical structures, this proportion was minimal

(Figure 2 a). While the images, in which the highest IoUs were observed, mostly displayed large

organ segments that were clearly visible (Figure 2 b), the images with the lowest IoU were of

variable quality with confounding factors such as blood, smoke, soiling of the endoscope lens, or

pictures blurred by camera shake (Figure 2 c).

Inference on a single image with a resolution of 640 x 512 pixels required, on average, 28 ms on

an Nvidia A5000, resulting in a frame rate of almost 36 frames per second. This runtime includes

one decoder, meaning that only the segmentation for one anatomical class is included.

Table 1: Summary of performance metrics for anatomical structure segmentation using structure-specific models based on the DeepLabv3 architecture. For each metric, mean and standard deviation are displayed.

Anatomical structure	F1 score	IoU	Precision	Recall	Specificity
Abdominal wall	0.90 ± 0.10	0.83 ± 0.14	0.89 ± 0.14	0.93 ± 0.07	0.97 ± 0.04
Colon	0.79 ± 0.20	0.69 ± 0.22	0.80 ± 0.21	0.82 ± 0.21	0.97 ± 0.05
Inferior mesenteric artery	0.54 ± 0.26	0.41 ± 0.22	0.55 ± 0.25	0.67 ± 0.33	0.99 ± 0.01
Intestinal veins	0.54 ± 0.33	0.44 ± 0.29	0.70 ± 0.26	0.56 ± 0.36	1.00 ± 0.00
Liver	0.80 ± 0.23	0.71 ± 0.25	0.85 ± 0.21	0.81 ± 0.24	0.98 ± 0.03
Pancreas	0.37 ± 0.32	0.28 ± 0.27	0.59 ± 0.37	0.37 ± 0.36	1.00 ± 0.01
Small intestine	0.87 ± 0.14	0.80 ± 0.18	0.87 ± 0.16	0.91 ± 0.15	0.97 ± 0.04
Spleen	0.79 ± 0.23	0.69 ± 0.24	0.74 ± 0.22	0.90 ± 0.24	0.99 ± 0.01
Stomach	0.71 ± 0.24	0.60 ± 0.25	0.65 ± 0.25	0.89 ± 0.21	0.98 ± 0.02
Ureter	0.47 ± 0.30	0.36 ± 0.25	0.53 ± 0.28	0.57 ± 0.39	1.00 ± 0.00
Vesicular glands	0.40 ± 0.25	0.28 ± 0.21	0.37 ± 0.28	0.62 ± 0.35	0.97 ± 0.03

Machine Learning-based anatomical structure segmentation in a combined model

Using all annotated images from the Dresden Surgical Anatomy Dataset, a combined model with a mutual encoder and organ-specific decoders was trained (Figure 1). Table 2 displays mean F1 score, IoU, precision, recall, and specificity for anatomical structure segmentation in the combined model. The performance of the combined model was overall similar to that of structure-specific models (Table 1), with highest segmentation performance for the abdominal wall (IoU: 0.78 ± 0.14) and the small intestine (IoU: 0.74 ± 0.18), and the lowest performance for the ureter (IoU: 0.22 ± 0.24) and the vesicular glands (IoU: 0.32 ± 0.20). In comparison to the respective structure-specific models, the combined model performed notably weaker in liver, stomach, and ureter segmentation, while performance for the other anatomical structures was similar. The proportion of un- or entirely mispredicted images was largest in the ureter, the stomach, the abdominal vessel structures, and the vesicular glands (Figure 3 a). Similar to the structure-specific models, trends towards an impact of segment size, endoscope lens soiling, blurry images, and presence of blood or smoke were seen when comparing image quality of well-predicted images (Figure 3 b) to images with poor or no prediction (Figure 3 c).

Inference on a single image with a resolution of 640 x 512 pixels required, on average, 71 ms on an Nvidia A5000, resulting in a frame rate of about 14 frames per second. This runtime includes all 11 decoders, meaning that segmentations for all organ classes are included.

Table 2: Summary of performance metrics for anatomical structure segmentation using the combined model (common encoder with structure-specific decoders). For each metric, mean and standard deviation are displayed.

Anatomical structure	F1 score	IoU	Precision	Recall	Specificity
Abdominal wall	0.87 ± 0.10	0.78 ± 0.14	0.85 ± 0.12	0.92 ± 0.13	0.94 ± 0.05
Colon	0.75 ± 0.17	0.62 ± 0.19	0.71 ± 0.18	0.87 ± 0.20	0.96 ± 0.03
Inferior mesenteric artery	0.52 ± 0.29	0.40 ± 0.25	0.57 ± 0.26	0.61 ± 0.35	0.99 ± 0.01
Intestinal veins	0.61 ± 0.25	0.48 ± 0.23	0.61 ± 0.22	0.72 ± 0.31	1.00 ± 0.00
Liver	0.59 ± 0.33	0.49 ± 0.32	0.70 ± 0.30	0.59 ± 0.37	0.97 ± 0.03
Pancreas	0.63 ± 0.21	0.49 ± 0.19	0.65 ± 0.23	0.70 ± 0.23	0.99 ± 0.01
Small intestine	0.84 ± 0.15	0.74 ± 0.18	0.80 ± 0.16	0.91 ± 0.17	0.96 ± 0.04
Spleen	0.68 ± 0.28	0.57 ± 0.28	0.69 ± 0.24	0.78 ± 0.32	0.99 ± 0.01
Stomach	0.46 ± 0.35	0.37 ± 0.31	0.62 ± 0.30	0.48 ± 0.39	0.99 ± 0.02
Ureter	0.30 ± 0.31	0.22 ± 0.24	0.52 ± 0.30	0.36 ± 0.39	1.00 ± 0.01
Vesicular glands	0.45 ± 0.24	0.32 ± 0.20	0.53 ± 0.29	0.51 ± 0.31	0.98 ± 0.02

Performance of machine learning models in relation to human performance

To approximate the clinical value of the previously described algorithms for anatomical structure segmentation, the performances of the structure-specific and the combined model were compared to that of of a cohort of 28 physicians, medical students, and persons with no medical background (Figure 4 a), and different degrees of experience in laparoscopic surgery (Figure 4 b). A vulnerable anatomical structure with – measured by classical metrics of overlap (Tables 1 and 2) – comparably weak segmentation performance of the trained algorithms, the pancreas was selected as an example.

Comparing bounding box segmentations of the pancreas of human annotators, the medical and laparoscopy-specific experience of participants was mirrored by the respective IoUs describing the overlap between the pancreas annotation and the ground truth. The pancreas-specific segmentation model (IoU: 0.29) and the combined segmentation model (IoU: 0.21) outperformed 26 out of the 28 human participants (Figures 4 c and d). Overall, these results demonstrate that the developed models have clinical potential to improve the recognition of vulnerable anatomical structures in laparoscopy.

Discussion

In surgery, misinterpretation of visual cues can result in objectifiable errors with serious consequences.[21] Based on a robust public dataset providing 13195 laparoscopic images with segmentations of eleven intra-abdominal anatomical structures, this study explores the potential

of machine learning for automated segmentation of these organs, and compares algorithmic

segmentation quality to that of humans with varying experience in minimally-invasive abdominal

surgery.

In summary, the presented findings suggest that machine learning-based segmentation of

intraabdominal organs and anatomical structures is possible and has the potential to provide

clinically valuable information. At an average runtime of 71 ms per image, corresponding to a

frame rate of 14 frames per second, the combined model would facilitate near-real-time

identification of eleven anatomical structures. These runtimes mirror the performance of a non-

optimized version of the model, which can be significantly improved using methods such as

TensorRT from Nvidia. Measured by classical metrics of overlap between segmentation and

ground truth, predictions were, overall, better for large and similar-appearing organs such as the

abdominal wall, the liver, the stomach, and the spleen as compared to smaller and more diverse-

appearing organs such as the pancreas, the ureter, or vesicular glands. Furthermore, poor image

quality (i.e. images blurred by camera movements, presence of blood or smoke in images) was

linked to lower accuracy of machine learning-based segmentations. These findings imply that

computer vision studies in laparoscopy should be carefully interpreted taking representativity and

potential selection of underlying training and validation data into consideration.

Measured by classical metrics of overlap (e.g. IoU, F1 score, precision, recall, specificity) that are

commonly used to evaluate segmentation performance, the structure-specific models and the

combined model performed similarly for most anatomical structures with average IoUs ranging

from 0.28 to 0.83 and from 0.22 to 0.78, respectively. The structure-specific model was superior

to the combined segmentation model with respect to liver, stomach, and ureter segmentation.

Interpretation of such metrics of overlap, however, represents a major challenge in computer

vision applications in medical domains such as dermatology and endoscopy[27–29] as well as non-

medical domains such as autonomous driving[30]. In the specific use case of laparoscopic surgery,

evidence suggests that such technical metrics alone are not sufficient to characterize the clinical

potential and utility of segmentation algorithms.[31,32] In this context, the subjective clinical utility of

a bounding box-based detection system recognizing the common bile duct and the cystic duct at

average precisions of 0.32 and 0.07, respectively, demonstrated by Tokuyasu *et al.*, supports this

hypothesis.[12] In the presented analysis, the trained structure-specific and combined machine

learning algorithms outperformed all human participants in the specific task of bounding box

segmentation of the pancreas except for two surgical specialists with over 10 years of experience.

This suggests that even for structures such as the pancreas with seemingly poor segmentation

quality (segmentation IoU of the best-performing model: 0.49 ± 0.19 in the test set) have the

potential to provide clinically valuable help in anatomy recognition. Notably, the best average IoUs
achieved in this comparative study were 0.29 (for the structure-specific model) and 0.36 (for the
best human participant), which would both be considered less reliable segmentation quality
measures on paper. This encourages further discussion about metrics for segmentation quality
assessment in clinical AI. In the future, the potential of the described dataset[22] and organ
segmentation algorithms could be exploited for educational purposes[33,34], for guidance systems
facilitating real-time detection of risk and target structures[18,32,35], or as an auxiliary function
integrated in more complex surgical assistance systems, such as guidance systems relying on
automated liver registration[36].
The limitations of this work are mostly related to the dataset and general limitations of machine
learning-based segmentation: First, the Dresden Surgical Anatomy Dataset is a monocentric
dataset based on 32 robot-assisted rectal surgeries. Therefore, the images used for algorithm
training and validation display organs from specific angles, which could limit generalizability and
transferability of the presented findings to other minimally-invasive abdominal surgeries,
particularly non-robotic procedures. Second, annotations were required for training of machine
learning algorithms, potentially inducing some bias towards the way that organs were annotated
in the algorithms, which may differ from individual healthcare professionals' way of recognizing an
organ. This is particularly relevant for organs such as the ureters or the pancreas, which often
appear covered by layers of tissue. Here, computer vision-based algorithms that solely consider
the laparoscopic images provided by the Dresden Surgical Anatomy Dataset for identification of
risk structures will only be able to identify an organ once it is visible. For an earlier recognition of
such hidden risk structures, more training data with meaningful annotations would be necessary.
Importantly, the presented comparison to human performance focused on segmentation of visible
anatomy as well, neglecting that humans (and possibly computers, too) could already identify a
risk structure hidden underneath tissue layers. The existing limitations notwithstanding, the
presented study represents an important addition to the growing body of research on medical
image analysis in laparoscopic surgery, particularly by linking technical metrics to human
performance.
In conclusion, this study demonstrates that machine learning methods have the potential to
provide clinically relevant near-real-time assistance in anatomy recognition in minimally-invasive
surgery. This study is the first to use the recently published Dresden Surgical Anatomy Dataset,
providing baseline algorithms for organ segmentation and evaluating the clinical relevance of such
algorithms. Future research should investigate other segmentation methods, the transferability of

these results to other surgical procedures, and the clinical impact of real-time surgical assistance

systems and didactic applications based on automated segmentation algorithms.

References

1. Wang, P. *et al.* Effect of a deep-learning computer-aided detection system on adenoma detection during colonoscopy (CADe-DB trial): a double-blind randomised study. *Lancet Gastroenterol. Hepatol.* **5**, 343–351 (2020).

2. Wang, P. *et al.* Real-time automatic detection system increases colonoscopic polyp and adenoma detection rates: a prospective randomised controlled study. *Gut* **68**, 1813–1819 (2019).

3. Kather, J. N. *et al.* Pan-cancer image-based detection of clinically actionable genetic alterations. *Nat. Cancer 2020 18* **1**, 789–799 (2020).

4. Esteva, A. *et al.* Dermatologist-level classification of skin cancer with deep neural networks. *Nature* **542**, 115–118 (2017).

5. Simillis, C. *et al.* Open Versus Laparoscopic Versus Robotic Versus Transanal Mesorectal Excision for Rectal Cancer: A Systematic Review and Network Meta-analysis. *Ann. Surg.* **270**, 59–68 (2019).

6. Zhao, J. J. *et al.* Comparative outcomes of needlescopic, single-incision laparoscopic, standard laparoscopic, mini-laparotomy, and open cholecystectomy: A systematic review and network meta-analysis of 96 randomized controlled trials with 11,083 patients. *Surgery* **170**, 994–1003 (2021).

7. Luketich, J. D. *et al.* Outcomes after minimally invasive esophagectomy: review of over 1000 patients. *Ann. Surg.* **256**, 95–103 (2012).

8. Thomson, J. E. *et al.* Laparoscopic versus open surgery for complicated appendicitis: a randomized controlled trial to prove safety. *Surg. Endosc.* **29**, 2027–2032 (2015).

9. Islam, M., Atputharuban, D. A., Ramesh, R. & Ren, H. Real-time instrument segmentation in robotic surgery using auxiliary supervised deep adversarial learning. *IEEE Robot. Autom. Lett.* **4**, 2188–2195 (2019).

10. Roß, T. *et al.* Comparative validation of multi-instance instrument segmentation in endoscopy: results of the ROBUST-MIS 2019 challenge. *Med. Image Anal.* **70**, 101920 (2020).

11. Shvets, A. A., Rakhlin, A., Kalinin, A. A. & Iglovikov, V. I. Automatic Instrument Segmentation in Robot-Assisted Surgery using Deep Learning. in *Proceedings - 17th IEEE International Conference on Machine Learning and Applications, ICMLA 2018* 624–628 (Institute of Electrical and Electronics Engineers Inc., 2019). doi:10.1109/ICMLA.2018.00100.

12. Tokuyasu, T. *et al.* Development of an artificial intelligence system using deep learning to

indicate anatomical landmarks during laparoscopic cholecystectomy. *Surg. Endosc. 2020*
354 **35**, 1651–1658 (2020).

13. Mascagni, P. *et al.* Artificial Intelligence for Surgical Safety: Automatic Assessment of the
Critical View of Safety in Laparoscopic Cholecystectomy Using Deep Learning. *Ann. Surg.*
(2020) doi:10.1097/SLA.0000000000004351.

14. Jin, A. *et al.* Tool Detection and Operative Skill Assessment in Surgical Videos Using
Region-Based Convolutional Neural Networks. *Proc. - 2018 IEEE Winter Conf. Appl.*
Comput. Vision, WACV 2018 **2018**-**January**, 691–699 (2018).

15. Funke, I. *et al.* Using 3D Convolutional Neural Networks to Learn Spatiotemporal Features
for Automatic Surgical Gesture Recognition in Video. *Med. Image Comput. Comput. Assist.*
Interv. – MICCAI 2019. Lect. Notes Comput. Sci. **11768**, 467–475 (2019).

16. Lavanchy, J. L. *et al.* Automation of surgical skill assessment using a three-stage machine
learning algorithm. *Sci. Reports 2021 111* **11**, 1–9 (2021).

17. Maier-Hein, L. *et al.* Surgical data science – from concepts toward clinical translation. *Med.*
Image Anal. **76**, 102306 (2022).

18. Madani, A. *et al.* Artificial Intelligence for Intraoperative Guidance. *Ann. Surg.* (2020)
doi:10.1097/sla.0000000000004594.

19. Fecso, A. B., Szasz, P., Kerezov, G. & Grantcharov, T. P. The effect of technical
performance on patient outcomes in surgery. *Ann. Surg.* **265**, 492–501 (2017).

20. Mazzocco, K. *et al.* Surgical team behaviors and patient outcomes. *Am. J. Surg.* **197**, 678–
685 (2009).

21. Suliburk, J. W. *et al.* Analysis of Human Performance Deficiencies Associated With Surgical
Adverse Events. *JAMA Netw. Open* **2**, e198067–e198067 (2019).

22. Carstens, M. *et al.* The Dresden Surgical Anatomy Dataset for abdominal organ
segmentation in surgical data science. *Sci. Data* (2023) doi:10.1038/s41597-022-01719-2.

23. Chen, L.-C., Papandreou, G., Schroff, F. & Adam, H. Rethinking Atrous Convolution for
Semantic Image Segmentation. *arXiv* (2017) doi:10.48550/arxiv.1706.05587.

24. Lin, T. Y. *et al.* Microsoft COCO: Common Objects in Context. *Lect. Notes Comput. Sci.*
(including Subser. Lect. Notes Artif. Intell. Lect. Notes Bioinformatics) **8693** **LNCS**, 740–
755 (2014).

25. Loshchilov, I. & Hutter, F. Decoupled Weight Decay Regularization. *7th Int. Conf. Learn.*
Represent. ICLR 2019 (2017) doi:10.48550/arxiv.1711.05101.

26. Leger, S. *et al.* A comparative study of machine learning methods for time-to-event survival
data for radiomics risk modelling. *Sci. Rep.* **7**, 11 (2017).

27. Renard, F., Guedria, S., Palma, N. De & Vuillerme, N. Variability and reproducibility in deep
learning for medical image segmentation. *Sci. Rep.* **10**, 1–16 (2020).

28. Powers, D. M. W. & Ailab. Evaluation: from precision, recall and F-measure to ROC,
informedness, markedness and correlation. *arXiv* (2020) doi:10.48550/arxiv.2010.16061.

29. Parikh, R. B., Teeple, S. & Navathe, A. S. Addressing Bias in Artificial Intelligence in Health
Care. *JAMA* **322**, 2377–2378 (2019).

30. Zhang, Y., Mehta, S. & Caspi, A. Rethinking Semantic Segmentation Evaluation for
Explainability and Model Selection. (2021).

31. Reinke, A. *et al.* Common Limitations of Image Processing Metrics: A Picture Story. *arXiv*
(2021) doi:10.48550/arxiv.2104.05642.

32. Hashimoto, D. A. *et al.* Computer Vision Analysis of Intraoperative Video: Automated
Recognition of Operative Steps in Laparoscopic Sleeve Gastrectomy. *Ann. Surg.* **270**, 414–
421 (2019).

33. Hu, Y. Y. *et al.* Complementing Operating Room Teaching With Video-Based Coaching.
JAMA Surg. **152**, 318–325 (2017).

34. Mizota, T., Anton, N. E. & Stefanidis, D. Surgeons see anatomical structures faster and
more accurately compared to novices: Development of a pattern recognition skill
assessment platform. *Am. J. Surg.* **217**, 222–227 (2019).

35. Ward, T. M. *et al.* Computer vision in surgery. *Surgery* **169**, 1253–1256 (2021).

36. Docea, R. *et al.* Simultaneous localisation and mapping for laparoscopic liver navigation : a
comparative evaluation study. in *Medical Imaging 2021: Image-Guided Procedures,*
Robotic Interventions, and Modeling (eds. Linte, C. A. & Siewerdsen, J. H.) vol. 11598 8
(SPIE, 2021).

434 Figures and Figure Captions

Figure 1

Fig. 1 Schematic illustration of the structure-specific and combined machine learning

models used for semantic segmentation. The Dresden Surgical Anatomy Dataset was split into

a training, a validation, and a test set. For spatial segmentation, two machine learning models

were trained: A structure-specific model with individual encoders and decoders, and a combined

model with a common encoder and structure-specific decoders.

none

Figure 2

Fig. 2 Pixel-wise organ segmentation with structure-specific models trained on the

respective organ subsets of the Dresden Surgical Anatomy Dataset. (a) Violin plot

illustrations of performance metrics for structure-specific segmentation models. The median and

quartiles are illustrated as solid and dashed lines, respectively. **(b)** Example images with the

highest IoUs for liver, pancreas, stomach, and ureter segmentation with structure-specific

segmentation models. Ground truth is displayed as blue line (upper panel), model segmentations

are displayed as white overlay (lower panel). **(c)** Example images with the lowest IoUs for liver,

pancreas, stomach, and ureter segmentation with structure-specific segmentation models.

Ground truth is displayed as blue line (upper panel), model segmentations are displayed as white

overlay (lower panel).

Figure 3

Fig. 3 Pixel-wise organ segmentation with the combined model trained on the entire

Dresden Surgical Anatomy Dataset with a common encoder and structure-specific

decoders. (a) Violin plot illustrations of performance metrics for the combined segmentation

model. The median and quartiles are illustrated as solid and dashed lines, respectively. **(b)**

Example images with the highest IoUs for liver, pancreas, stomach, and ureter segmentation with

the combined segmentation model. Ground truth is displayed as blue line (upper panel), model

segmentations are displayed as white overlay (lower panel). **(c)** Example images with the lowest

IoUs for liver, pancreas, stomach, and ureter segmentation with the combined segmentation

model. Ground truth is displayed as blue line (upper panel), model segmentations are displayed

as white overlay (lower panel).

Figure 4

Fig. 4 Comparison of pancreas segmentation performance of the structure-specific and the

combined semantic segmentation models with a cohort of 28 human participants. (a)

Distribution of medical and non-medical professions among human participants. **(b)** Distribution

of laparoscopy experience among human participants. **(c)** Waterfall chart displaying the average

pancreas segmentation IoUs of participants with different professions as compared to the IoU

generated by the structure-specific and the combined semantic segmentation models. **(d)**

Waterfall chart displaying the average pancreas segmentation IoUs of participants with varying

laparoscopy experience as compared to the IoU generated by the structure-specific and the

combined semantic segmentation models.

Abbreviations

AI   Artificial Intelligence

IoU   Intersection-over-Union

SD   Standard deviation

Acknowledgements and Funding

This work has been funded by the project funding within the Else Kröner Fresenius Center for Digital Health (EKFZ), Dresden, Germany (project "CoBot") and by the German Research Foundation DFG within the Cluster of Excellence EXC 2050: "Center for Tactile Internet with Human-in-the-Loop (CeTI)" (project number 390696704). FRK received funding from the Medical Faculty of the Technical University Dresden within the MedDrive Start program (grant number 60487) and from the Joachim Herz Foundation (Add-On Fellowship for Interdisciplinary Life Science). FMR received a doctoral student scholarship from the *Carus Promotionskolleg* Dresden. The authors gratefully acknowledge excellent project coordination by Dr. Elisabeth Fischermeier and Dr. Grit Krause-Jüttler.

Author contributions

FRK, JW, MD, SS, and SB conceptualized the study. FRK, FMR, and MC collected and annotated clinical and video data and contributed to data analysis. ACJ, SL, and SB implemented and trained the neural networks and contributed to data analysis. JW, MD, and SS supervised the project, provided infrastructure and gave important scientific input. FRK drafted the initial manuscript text. All authors reviewed, edited, and approved the final manuscript.

Competing interests

The authors declare no conflicts of interest.