Structural Elucidation of β_1- and β_2-Transferrin Using Microprobe-Capture In-Emitter Elution

and High-Resolution Mass Spectrometry

Ruben Yiqi Luo[1,2], Christopher Pfaffroth[2], Samuel Yang[1], Kevin Hoang[2], Priscilla S.-W. Yeung[1,2], James L. Zehnder[1,2], Run-Zhang Shi[1,2]

[1]Department of Pathology, Stanford University, Stanford, CA, USA

[2]Clinical Laboratories, Stanford Health Care, Palo Alto, CA, USA

Running Title: Structural Elucidation of β_1- and β_2-Tf Using MPIE and HR-MS

Corresponding Author: Ruben Yiqi Luo

Address: 3375 Hillview Ave, Palo Alto, CA 94304

Tel: 650-724-1318

Email: rubenluo@stanford.edu

Key Words: β_1-Transferrin, β_2-Transferrin, MPIE-ESI-MS, N-Glycan

Abstract

Background: Cerebrospinal fluid (CSF) leak is typically diagnosed by detecting a protein marker β_2-transferrin (β_2-Tf) in secretion samples. β_2-Tf and β_1-transferrin (β_1-Tf) are glycoforms of human transferrin (Tf). A novel affinity capture technique for sample preparation, called microprobe-capture in-emitter elution (MPIE), was incorporated with high-resolution mass spectrometry (HR-MS) to analyze the Tf glycoforms and elucidate the structures of β_1-Tf and β_2-Tf.

Methods: To implement MPIE, an analyte is first captured on the surface of a microprobe, and subsequently eluted from the microprobe inside an electrospray emitter. The capture process is monitored in real-time via next-generation biolayer interferometry (BLI). When electrospray is established from the emitter to a mass spectrometer, the analyte is immediately ionized via electrospray ionization (ESI) for HR-MS analysis. Serum, CSF, and secretion samples were analyzed using MPIE-ESI-MS.

Results: Based on the MPIE-ESI-MS results, the structures of β_1-Tf and β_2-Tf were solved. As Tf glycoforms, β_1-Tf and β_2-Tf share the amino acid sequence but have varying N-glycans. β_1-Tf, the major serum-type Tf, has two G2S2 N-glycans on Asn413 and Asn611. β_2-Tf, the major brain-type Tf, has an M5 N-glycan on Asn413 and a G0FB N-glycan on Asn611.

Conclusions: The demonstration of MPIE-ESI-MS in detection of β_2-Tf paved a way to establish a MS-based clinical assay for β_2-Tf. Through the structural elucidation of β_1-Tf and β_2-Tf, the resolving power of the novel MPIE-ESI-MS method was demonstrated. Moreover, knowing the N-glycan structures on β_2-Tf allows for the design of other novel test methods for β_2-Tf in the future.

Introduction

Cerebrospinal fluid (CSF) leak can occur as a result of laceration, blunt trauma, or surgery. It may lead to potentially life-threatening meningitis if left untreated (1,2). Because β_2-transferrin (β_2-Tf), a proteoform of human transferrin (Tf), is mainly present in CSF and barely detectable in other body fluids (3,4), CSF leak can be diagnosed by detecting β_2-Tf in any body fluid – most commonly in rhinorrhea or otorrhea secretion samples. The clinical utility and diagnostic value of β_2-Tf in CSF leak have been demonstrated (5,6). β_2-Tf, together with the typical Tf proteoform in serum β_1-transferrin (β_1-Tf), were named after their electrophoretic mobility in gel electrophoresis (7). However, the structures of β_1-Tf and β_2-Tf have not been elucidated.

There has been extensive basic research on human Tf since it is a high-abundance protein in human blood with a major role in iron metabolism. The amino acid sequence of Tf precursor was determined through protein cleavage and cDNA characterization, showing a full sequence of 698 amino acids; after the removal of an N-terminal signal peptide, the mature form of Tf contains 679 amino acids, 19 intramolecular disulfide bonds formed between cysteine residues, and two N-glycans attached to the amino groups of the side chains of Asn419 and Asn611 (Asn432 and Asn630 of Tf precursor) (8–11). Tf proteoforms typically vary by the N-glycan structures (11–13), and the Tf proteoforms of interest so far are Tf glycoforms, including β_1-Tf and β_2-Tf (7).

Although it is known that β_2-Tf has desialylated N-glycans while β_1-Tf has fully sialylated N-glycans (3,14,15), the structures of β_1-Tf and β_2-Tf have not been clarified. On the other hand,

the N-glycans on Tf glycoforms in serum and in CSF were characterized using gel

electrophoresis, liquid chromatography, and enzymatic digestion-based mass spectrometry (MS)

by neurobiologists (13,16–19). It was reported that a group of Tf glycoforms were present in

serum, namely serum-type Tf or sTf, among which a specific Tf glycoform (major serum-type

Tf) predominated (12,13); the serum-type Tf also existed in CSF, and the major serum-type Tf

used to be named Tf-2 in CSF (18); an additional group of Tf glycoforms were present in CSF,

namely brain-type Tf or Tf-1 in CSF, among which a specific Tf glycoform (major brain-type

Tf) was more abundant than the rest (15,20). The brain-type Tf was at least partly synthesized in

the CSF-producing tissue choroid plexus rather than being produced by glycosidase digestion of

serum-type Tf (16,20). It was found that the N-glycans on the major serum-type Tf consisted of

bi-antennary oligosaccharide chains with sialylated terminals, and those on the brain-type Tf are

desialylated, or more specifically, unsialylated (asialotransferrin) (15,16,18,20). It was

hypothesized that the major brain-type Tf was β_2-Tf (21), but this hypothesis has not been

proved.

In clinical cases where CSF leak is suspected, secretion samples can be collected from patients

and sent to clinical laboratories to test for the presence of β_2-Tf. The conventional method to test

β_2-Tf as well as β_1-Tf is agarose gel immunofixation electrophoresis (IFE) (2–4). Although it is

widely used in clinical laboratories, it does not provide structural information of the analytes.

Thus, the structures of β_1-Tf and β_2-Tf, particularly the N-glycan structures on these Tf

glycoforms, remained an unanswered question with this method.

As an emerging technology in clinical diagnostics, high-resolution mass spectrometry (HR-MS),

particularly top-down HR-MS, can be used to analyze a protein target in its intact state and

elucidate the post-translational modifications and amino acid variations in its proteoforms (22–

24). While HR-MS is an ideal tool to analyze the Tf glycoforms, the quality of data acquired

during HR-MS analysis depends on sample preparation (25,26). In this article, a novel affinity

capture technique for sample preparation, called microprobe-capture in-emitter elution (MPIE),

was incorporated with HR-MS to study the Tf glycoforms (27). MPIE can directly couple a

label-free optical sensing technology with MS. The label-free optical sensing technology is next-

generation biolayer interferometry (BLI, also named as thin-layer interferometry, TFI), which

senses optical thickness changes on the sensing surface of a microprobe caused by biomolecular

interactions, achieving real-time measurement without employing a reporter molecule (enzyme,

fluorophore, etc.) (28,29). To implement MPIE, an analyte is first captured on the surface of a

microprobe, and subsequently eluted from the microprobe inside an electrospray emitter. The

capture process is monitored in real-time via BLI. When electrospray is established from the

emitter to a mass spectrometer, the analyte is immediately ionized via electrospray ionization

(ESI) for HR-MS analysis. By this means, BLI and HR-MS are directly coupled in the form of

MPIE-ESI-MS, which is readily deployed to analyze the Tf glycoforms and elucidate the

structures of β_1-Tf and β_2-Tf. The study can pave a way for the development of novel clinical

assays for β_2-Tf.

Materials and Methods

Materials and Specimens

LC-MS grade water, acetonitrile, formic acid, and 0.2 µm PVDF syringe filters were purchased

from Thermo Fisher Scientific (Waltham, MA). Tf standard (purified sTf) was purchased from

Sigma-Aldrich (St. Louis, MI). A mouse monoclonal anti-transferrin IgG antibody (anti-Tf Ab)

was obtained from Sinobiological (Wayne, PA), and biotinylated using an EZ-Link HPDP-Biotin

reagent kit (Waltham, MA). Remnant CSF and serum samples from patients, and secretion

samples from patients suspected of CSF leak were obtained from Stanford Health Care and

Stanford Children's Health, following approved institutional review board protocols for the use

of remnant patient specimens.

Sample Preparation

The biotinylated anti-Tf Ab was diluted in phosphate-buffered saline at pH 7.4 with 0.02%

Tween 20 and 0.2% BSA (PBST-B) to 10 µg/ml for use. A pooled CSF sample was made by

mixing 9 CSF samples from patients to explore the analytical sensitivity of MPIE-ESI-MS for

β_2-Tf in CSF. The pooled CSF sample was mixed with water to make a dilution series. All CSF

samples were 1:1 diluted in phosphate-buffered saline at pH 7.4 with 0.02% Tween 20 (PBST)

and all serum samples were 1:19 diluted in PBST-B. Secretion samples from patients were first

mixed with an equal amount of water and filtered using a 0.2 µm PVDF syringe filter, and then

1:1 diluted in PBST.

To study the gel electrophoresis-separated Tf glycoforms, gel electrophoresis of CSF samples

was carried out using a Hydragel 6 β_2 Transferrin kit (Sebia, Lisses, France) following the

manufacturer's protocol. In brief, a CSF sample was first 1:1 mixed with an iron-saturating

solution, and then 10 µl of the sample was loaded to each of the 6 wells on an agarose gel. Gel

electrophoresis was implemented in a Hydrasys 2 instrument (Sebia, Lisses, France), and the

agarose gel was removed from the instrument after the gel electrophoresis was completed,

without running the immunofixation steps. The agarose gel was placed on a paper template

marked with the β_1-Tf and β_2-Tf band regions, and the gel stripes of the band regions were cut

out using a scalpel. Each gel stripe was placed in a 1.5 ml sample tube, 200 µl PBST was added,

and the sample was rocked for 2 hr at room temperature to extract the analyte from the gel stripe.

After extraction, the supernatant was filtered using a 0.2 µm PVDF syringe filter.

<u>MPIE-ESI-MS Instrumentation and Experiment</u>

An MPIE-ESI-MS experiment consists of two parts: BLI-based affinity capture and in-emitter

elution ESI-MS, with the details including instrumentation described elsewhere (27). The BLI-

based affinity capture was implemented in a Gator Plus analyzer (Gator Bio, Palo Alto, CA): a

BLI microprobe pre-coated with streptavidin was first dipped into the biotinylated anti-Tf Ab

solution for 10 min to load the anti-Tf Ab, then dipped into a sample for 10 min to capture Tf

molecules, and rinsed in PBST for 1 min to remove non-specifically bound molecules. The in-

emitter elution ESI-MS was implemented in an EMASS-II ESI ion source which coupled an

ECE-001 capillary electrophoresis instrument (CMP Scientific, Brooklyn, NY) with an Orbitrap

Q-Exactive Plus mass spectrometer (Thermo Scientific, San Jose, CA): an electrospray emitter

with a regular open end and a tapered open end (tip orifice diameter 20-30 µm) was filled with a

sheath liquid (10 mM ammonium formate in water); after affinity capture, the microprobe was

rinsed in the sheath liquid for 10 s, inserted into the emitter through the regular open end, and

settled in the tapered end by gravity; the emitter was mounted to the ESI ion source, and a

capillary was inserted into the emitter through the regular open end and positioned right behind

the microprobe to deliver an elution liquid (80% acetonitrile and 2% formic acid in water); once

electrospray was established by applying a positive voltage to the sheath liquid in the emitter,

HR-MS data acquisition was initiated, and injection of the elution liquid was started

subsequently. The emitter was placed ~2 mm away from the mass spectrometer inlet with the

electrospray voltage set at 2.2 kV. The injection of the elution liquid was driven by 5 psi

pneumatic pressure. The following MS parameters were used: ion-transfer capillary temperature

350°C, S-lens RF level 50, and number of microscans 10. Primary mass spectra were acquired in

positive polarity at resolution 17.5K.

Data Analysis

In HR-MS analysis of proteins, it is necessary to deconvolute raw MS data to merge the multiple

charge states and isotopic peaks of an analyte to obtain its accurate molecular mass. The acquired

data in each MPIE-ESI-MS experiment was viewed as a time trace of MS responses, and the

elution time window of an analyte was identified by checking the molecular ions of the analyte at

each time point. The data in the elution time window were selected for deconvolution using

Biopharma Finder 4.1 (Thermo Fisher Scientific, San Jose, CA) with the ReSpect algorithm. MS

peaks of analytes were displayed in deconvoluted mass spectra at uncharged state showing

average molecular masses.

Results

The performance of MPIE-ESI-MS for Tf analysis was demonstrated and reported elsewhere
(27). MPIE-ESI-MS had a limit of detection for the Tf standard at 0.063 μg/ml, which was
translated to no more than 7 fmol Tf molecules captured on a microprobe (27). The high
analytical sensitivity and specificity provided by affinity capture made MPIE-ESI-MS an ideal
method to study the Tf molecules in serum, CSF, and secretion samples. The results of a set of
samples are shown in Figure 1. The deconvoluted mass spectrum of the serum sample showed a
group of MS peaks around 79554 Da; they are mainly serum-type Tf glycoforms and the
predominant MS peak at 79554 Da should be the major serum-type Tf (N-glycan structures
shown in Figure 1, see details in Discussion). The deconvoluted mass spectrum of the CSF
sample showed the major serum-type Tf and a group of MS peaks around 78008 Da; they are
mainly brain-type Tf glycoforms and the most abundant MS peak at 78008 Da should be the
major brain-type Tf (N-glycan structures shown in Figure 1, see details in Discussion). The
deconvoluted mass spectrum of the secretion sample showed both serum-type Tf and brain-type
Tf glycoforms, meaning that CSF was present in the sample. This finding was consistent with the
fact that the secretion sample was obtained from a patient diagnosed of CSF leak. In addition, the
BLI sensorgram (a time trace of label-free optical sensing responses) of each affinity capture
process (capture of Tf by anti-Tf Ab on a microprobe) is shown in Figure S1, allowing for real-
time monitoring of the capture process.

To elucidate the structures of the Tf glycoforms β_1-Tf and β_2-Tf, after gel electrophoresis of CSF
samples, the extracts from the gel stripes of the β_1-Tf and β_2-Tf band regions were analyzed

using MPIE-ESI-MS. The results of a set of samples are shown in Figure 2. The deconvoluted

mass spectra showed the MS peak of only the major serum-type Tf in the extract from the β_1-Tf

band region and the MS peak of only the major brain-type Tf in the extract from the β_2-Tf band

region. The measured accurate molecular masses were consistent with the experiments in Figure

1 and matched the theoretical molecular masses of the major serum-type Tf and major brain-type

Tf (N-glycan structures shown in Figure 2, see details in Discussion). This observation

confirmed that β_1-Tf and β_2-Tf were actually the major serum-type Tf and major brain-type Tf,

respectively. In addition, the gel area between the β_1-Tf and β_2-Tf band regions was also

analyzed and minor Tf glycoforms in serum and CSF were found, which could be due to the low

abundance of the minor Tf glycoforms and limited analyte quantities in a gel stripe.

A set of 12 secretion samples from patients suspected of CSF leak were analyzed using the

MPIE-ESI-MS method, among which 5 samples were positive for β_2-Tf and the rest were

negative as measured by IFE test and clinical manifestations. As shown in Table 1, the MS peak

at 78008 Da was observed in the MPIE-ESI-MS results of the 5 positive samples but not found

in those of the 7 negative samples, which further verified that β_2-Tf was indeed the major brain-

type Tf. In addition, the analytical sensitivity of the MPIE-ESI-MS method for CSF and

secretion samples was explored. A pooled CSF sample was mixed with water at 1:1, 1:4, 1:9, and

1:19 ratios to prepare a dilution series for analysis. Table S1 summarizes the results, listing the

MS peak intensities of β_1-Tf and β_2-Tf in the deconvoluted mass spectra (the entire time window

of Tf elution selected for deconvolution). The peak intensities decreased with the pooled CSF

sample dilution, and it was demonstrated that the MPIE-ESI-MS method was able to detect β_2-Tf

in at least 10-fold diluted CSF (1:9 pooled CSF : water mixture).

Discussion

The amino acid sequence of Tf was previously reported (8,9) and the primary structure of a

mature Tf molecule is illustrated in Figure 3A. As Tf glycoforms, β_1-Tf and β_2-Tf share the

amino acid backbone but have varying N-glycans. The N-glycans on proteins typically include

bi-, tri-, or tetra-antennary oligosaccharide chains, resulted from sequential action of

glycosyltransferases (30,31). An N-glycan can be named according to its sugar composition and

branching structure using a few nomenclature systems (traditional nomenclature used in this

article) (32). With designated N-glycan types, the accurate molecular mass of a Tf glycoform can

be calculated, and the correctness of the N-glycan types can be confirmed by comparing the

theoretical molecular mass of the Tf glycoform to the measured accurate molecular mass.

Regarding the N-glycans on Tf, the previous enzymatic digestion-based MS analysis of the Tf

glycoforms in CSF revealed explicitly that the most abundant N-glycan type was G2S2 (17,19).

By measuring the trypsin-digested peptides encompassing the glycosylation sites, G2S2 was

found consisting of roughly 65% of all N-glycans on Asn413 and 73% of all N-glycans on

Asn611 (17). Thus, the Tf glycoform with two G2S2 N-glycans on Asn413 and Asn611 is the

major serum-type Tf because (1) it is the most abundant Tf glycoform in serum and CSF, (2) it is

fully sialylated (N-acetylneuraminic acid as the sialic acid), and (3) its theoretical molecular

mass 79554.71 Da exactly matches the measured molecular mass of the major serum-type Tf

(Figure 3B). As the second most abundant Tf glycoform and the most abundant unsialylated Tf

glycoform in CSF, the major brain-type Tf should be associated with N-glycan types M5 and

G0FB (Figure 3B), which were found as the most abundant unsialylated N-glycans (19).

Particularly, M5 was found consisting of roughly 20% of all N-glycans on Asn413 and G0FB

was found consisting of 26% of all N-glycans on Asn611 (17). Thus, the major brain-type Tf has

an M5 N-glycan on Asn413 and a G0FB N-glycan on Asn611. Its theoretical molecular mass

78008.35 Da exactly matches the measured molecular mass of the major brain-type Tf. The

glycosylation sites of the M5 and G0FB N-glycans were confirmed with another study of

trypsin-digested glycopeptides from Tf glycoforms in CSF (18). In addition, the presence of a

bisecting N-acetylglucosamine (GlcNAc) in the G0FB N-glycan was proved by the previously

reported lectin-binding experiments (16), supporting that G0FB is the more favorable N-glycan

type on Asn611 than the other isomeric N-glycan type without a bisecting GlcNAc (19). In

addition, the MPIE-ESI-MS analysis of the extracts from the gel bands proved that β_1-Tf and the

major serum-type Tf were identical, so were β_2-Tf and the major brain-type Tf.

Besides the two major Tf glycoforms in CSF (β_1-Tf and β_2-Tf), it is known that other minor Tf

glycoforms may also exist in CSF, blood, and other body fluids (4,33). As illustrated in the

previous literature, there are a variety of unsialylated N-glycans on Tf molecules, and the

sialylated N-glycan G2S2 also has diversity such as the more branched form G3S3 (13,17,34).

Thus, a number of fully sialylated, partially sialylated, and unsialylated Tf glycoforms can be

formed by the various combinations of the two N-glycans on a Tf molecule. In gel

electrophoresis, sialic acids bring negative charges to a Tf molecule under neutral or alkaline pH

conditions, influencing its electrophoretic mobility. As such, Tf glycoforms migrate in the order

of fully sialylated, disialylated, and unsialylated Tf glycoforms, with regard to the number of

sialic acids on the N-glycans. Thus, the minor Tf glycoforms can migrate within or between the

β_1-Tf and β_2-Tf band regions. For instance, the product insert of the Hydragel 6 β_2 Transferrin kit

states that a band of disialylated Tf glycoforms (disialotransferrin) may exist above the β_2-Tf

band (35). On the other hand, the unsialylated Tf glycoforms besides β_2-Tf can migrate within

the β_2-Tf band region and interfere with the β_2-Tf detection. The product insert suggests that the

ratio of unsialylated to disialylated bands can be used to confirm the presence of β_2-Tf. The

reason of this practice can be explained as follows: when CSF is present in a sample, the

abundance of the major unsialylated Tf glycoform β_2-Tf significantly exceeds the minor

unsialylated and disialylated Tf glycoforms, resulting in a high ratio of the β_2-Tf band to the

disialotransferrin band. This practice in IFE is derived from the lack of molecular structure

information in gel electrophoresis. However, when MS is employed, such as in the use of MPIE-

ESI-MS, it is unnecessary to compare unsialylated and disialylated Tf glycoforms because β_2-Tf

can be specifically detected by its accurate molecular mass, definitively differentiating it from

the other Tf glycoforms with different molecular masses.

The demonstration of MPIE-ESI-MS in detection of β_2-Tf paved a way to establish a MS-based

clinical assay for β_2-Tf. However, more factors need to be taken into consideration for this

purpose. When implementing the accurate molecular mass-based detection of β_2-Tf, Tf variants

resulted from genetic polymorphism should be taken into consideration (11,36–39). In theory,

the amino acid mutations in Tf variants affect the molecular masses of β_1-Tf and β_2-Tf, but they

should not change the molecular mass difference between the two Tf glycoforms (1546 Da),

provided the two glycosylation sites are not modified. This hypothesis is supported by the MPIE-

ESI-MS results of a few Tf variant-containing CSF samples, as shown in Figure S2, but it needs

to be proved with a larger number of Tf variant samples. In addition, the glycoforms in serum-

type Tf can vary under specific pathophysiological conditions and β_1-Tf might not be found in

those samples due to aberrant glycosylation (12,34). In such cases the detection and

identification of β_2-Tf can be impacted. These topics are beyond the scope of this article and can

be investigated in the future when a MS-based clinical assay for β_2-Tf is being developed.

As an innovative affinity capture technique, MPIE facilitates real-time monitoring of the affinity

capture process to overcome the lack of process monitoring in conventional affinity capture

techniques (27). In MPIE-ESI-MS analysis, the real-time monitoring feature can be used for

verifying affinity capture performance and troubleshooting when MS responses are weak or

absent, as well as effective selection of capture agents and step-by-step optimization of affinity

capture conditions. Through the structural elucidation of β_1-Tf and β_2-Tf, the resolving power of

the novel MPIE-ESI-MS method was demonstrated. Moreover, knowing the N-glycan structures

on β_2-Tf allows for the design of other novel test methods for β_2-Tf in the future. For instance, it

is possible to employ carbohydrate-binding reagents to selectively capture β_2-Tf for MS analysis

or to establish antibody-lectin Sandwich immunoassays.

Acknowledgment

The authors thank Gator Bio (Palo Alto, CA) and CMP Scientific (Brooklyn, NY) for kindly

providing equipment and consumables, and C. Wong, S. Ferolino, L. Calayag, and R. Rieta at

Stanford Health Care for collecting samples for this research.

References

1. Warnecke A, Averbeck T, Wurster U, Harmening M, Lenarz T, Stöver T. Diagnostic Relevance of β2-Transferrin for the Detection of Cerebrospinal Fluid Fistulas. Arch Otolaryngol Neck Surg. 2004 Oct 1;130(10):1178.

2. McCudden CR, Senior BA, Hainsworth S, Oliveira W, Silverman LM, Bruns DE, et al. Evaluation of high resolution gel β_2-transferrin for detection of cerebrospinal fluid leak. Clin Chem Lab Med CCLM. 2013 Feb 1;51(2):311–5.

3. Papadea C, Schlosser RJ. Rapid Method for β2-Transferrin in Cerebrospinal Fluid Leakage Using an Automated Immunofixation Electrophoresis System. Clin Chem. 2005 Feb 1;51(2):464–70.

4. Görögh T, Rudolph P, Meyer JE, Werner JA, Lippert BM, Maune S. Separation of β2-Transferrin by Denaturing Gel Electrophoresis to Detect Cerebrospinal Fluid in Ear and Nasal Fluids. Clin Chem. 2005 Sep 1;51(9):1704–10.

5. Zaret DL, Morrison N, Gulbranson R, Keren DF. Immunofixation to Quantify β2-Transferrin in Cerebrospinal Fluid to Detect Leakage of Cerebrospinal Fluid from Skull Injury. Clin Chem. 1992 Sep 1;38(9):1909–12.

6. Nandapalan V, Watson ID, Swift AC. Beta-2-transferrin and cerebrospinal fluid rhinorrhoea. Clin Otolaryngol. 1996 Jun;21(3):259–64.

7. Gallo P, Bracco F, Morara S, Battistin L, Tavolato B. The cerebrospinal fluid transferrin/Tau proteins. J Neurol Sci. 1985 Aug;70(1):81–92.

8. Yang F, Lum JB, McGill JR, Moore CM, Naylor SL, van Bragt PH, et al. Human transferrin: cDNA characterization and chromosomal localization. Proc Natl Acad Sci. 1984 May;81(9):2752–6.

9. MacGillivray RT, Mendez E, Sinha SK, Sutton MR, Lineback-Zins J, Brew K. The complete amino acid sequence of human serum transferrin. Proc Natl Acad Sci. 1982 Apr;79(8):2504–8.

10. Wang S, Kaltashov IA. Identification of Reduction-Susceptible Disulfide Bonds in Transferrin by Differential Alkylation Using O^{16}/O^{18} Labeled Iodoacetic Acid. J Am Soc Mass Spectrom. 2015 May 1;26(5):800–7.

11. de Jong G, van Dijk JP, van Eijk HG. The biology of transferrin. Clin Chim Acta. 1990 Sep;190(1–2):1–46.

12. de Jong G, van Eijk HG. Microheterogeneity of human serum transferrin: A biological phenomenon studied by isoelectric focusing in immobilized pH gradients. Electrophoresis. 1988;9(9):589–98.

13. de Jong G, van Noort WL, van Eijk HG. Carbohydrate analysis of transferrin subfractions
isolated by preparative isoelectric focusing in immobilized pH gradients. Electrophoresis.
1992;13(1):225–8.

14. Delaroche O, Bordureb P, Lippert E, Sagnieza M. Perilymph detection by β2-transferrin
immunoblotting assay. Application to the diagnosis of perilymphatic fistulae. Clin Chim
Acta. 1996 Feb;245(1):93–104.

15. Hoffmann A, Nimtz M, Getzlaff R, Conradt HS. 'Brain-type' *N* -glycosylation of asialo-
transferrin from human cerebrospinal fluid. FEBS Lett. 1995 Feb 13;359(2–3):164–8.

16. Futakawa S, Nara K, Miyajima M, Kuno A, Ito H, Kaji H, et al. A unique N-glycan on
human transferrin in CSF: a possible biomarker for iNPH. Neurobiol Aging. 2012
Aug;33(8):1807–15.

17. Brown KJ, Vanderver A, Hoffman EP, Schiffmann R, Hathout Y. Characterization of
transferrin glycopeptide structures in human cerebrospinal fluid. Int J Mass Spectrom. 2012
Feb;312:97–106.

18. Nagae M, Morita-Matsumoto K, Arai S, Wada I, Matsumoto Y, Saito K, et al. Structural
change of N-glycan exposes hydrophobic surface of human transferrin. Glycobiology. 2014
Aug 28;24(8):693–702.

19. Hoshi K, Ito H, Abe E, Fuwa TJ, Kanno M, Murakami Y, et al. Transferrin Biosynthesized
in the Brain Is a Novel Biomarker for Alzheimer's Disease. Metabolites. 2021 Sep
10;11(9):616.

20. Hoshi K, Matsumoto Y, Ito H, Saito K, Honda T, Yamaguchi Y, et al. A unique glycan-
isoform of transferrin in cerebrospinal fluid: A potential diagnostic marker for neurological
diseases. Biochim Biophys Acta BBA - Gen Subj. 2017 Oct;1861(10):2473–8.

21. Caslavska J, Schild C, Thormann W. High□resolution capillary zone electrophoresis and
mass spectrometry for distinction of undersialylated and hypoglycosylated transferrin
glycoforms in body fluids. J Sep Sci. 2020 Jan;43(1):241–57.

22. Siuti N, Kelleher NL. Decoding protein modifications using top-down mass spectrometry.
Nat Methods. 2007 Oct;4(10):817–21.

23. Brown KA, Melby JA, Roberts DS, Ge Y. Top-down proteomics: challenges, innovations,
and applications in basic and clinical research. Expert Rev Proteomics. 2020 Oct
2;17(10):719–33.

24. Luo RY, Wong C, Xia JQ, Glader BE, Shi RZ, Zehnder JL. Neutral-Coating Capillary
Electrophoresis Coupled with High-Resolution Mass Spectrometry for Top-Down
Identification of Hemoglobin Variants. Clin Chem. 2022 Oct 29;hvac171.

25. Donnelly DP, Rawlins CM, DeHart CJ, Fornelli L, Schachner LF, Lin Z, et al. Best practices
and benchmarks for intact protein analysis for top-down mass spectrometry. Nat Methods.
2019 Jul;16(7):587–94.

26. Padula M, Berry I, O'Rourke M, Raymond B, Santos J, Djordjevic SP. A Comprehensive
Guide for Performing Sample Preparation and Top-Down Protein Analysis. Proteomes. 2017
Apr 7;5(4):11.

27. Luo RY, Yang S. Using Microprobe-Capture In-Emitter Elution to Directly Couple Label-
Free Optical Sensing Technology with Mass Spectrometry for Top-Down Protein Analysis
[Internet]. Chemistry; 2022 Oct [cited 2022 Oct 25]. Available from:
https://chemrxiv.org/engage/chemrxiv/article-details/6352339055a0816c28bee4cb

28. Luo YR, Chakraborty I, Lazar-Molnar E, Wu AHB, Lynch KL. Development of Label-Free
Immunoassays as Novel Solutions for the Measurement of Monoclonal Antibody Drugs and
Antidrug Antibodies. Clin Chem. 2020 Oct 1;66(10):1319–28.

29. Luo YR, Yun C, Chakraborty I, Wu AHB, Lynch KL. A SARS-CoV-2 Label-Free Surrogate
Virus Neutralization Test and a Longitudinal Study of Antibody Characteristics in COVID-
19 Patients. Tang YW, editor. J Clin Microbiol [Internet]. 2021 Jun 18 [cited 2021 Nov
4];59(7). Available from: https://journals.asm.org/doi/10.1128/JCM.00193-21

30. Bieberich E. Synthesis, Processing, and Function of N-glycans in N-glycoproteins. In: Yu
RK, Schengrund CL, editors. Glycobiology of the Nervous System [Internet]. New York,
NY: Springer New York; 2014 [cited 2021 Dec 6]. p. 47–70. (Advances in Neurobiology;
vol. 9). Available from: http://link.springer.com/10.1007/978-1-4939-1154-7_3

31. Cao L, Diedrich JK, Ma Y, Wang N, Pauthner M, Park SKR, et al. Global site-specific
analysis of glycoprotein N-glycan processing. Nat Protoc. 2018 Jun;13(6):1196–212.

32. Maier M, Reusch D, Bruggink C, Bulau P, Wuhrer M, Mølhøj M. Applying mini-bore
HPAEC-MS/MS for the characterization and quantification of Fc N -glycans from
heterogeneously glycosylated IgGs. J Chromatogr B. 2016 Oct;1033–1034:342–52.

33. Kleinert P, Kuster T, Durka S, Ballhausen D, Bosshard NU, Steinmann B, et al. Mass
Spectrometric Analysis of Human Transferrin in Different Body Fluids. Clin Chem Lab
Med. 2003 Jan 4;41(12):1580–8.

34. Coddeville B, Carchon H, Jaeken J, Briand G, Spik G. Determination of glycan structures
and molecular masses of the glycovariants of serum transferrin from a patient with
carbohydrate deficient syndrome type II. Glycoconj J. 1998;15(3):265–73.

35. HYDRAGEL 6 ß2 TRANSFERRIN(E) Kit Instruction Manual. Sebia; 2011.

36. Caslavska J, Lanz C, Burda P, Tobler M, Thormann W. Analysis of genetic variants of
transferrin in human serum after desialylation by capillary zone electrophoresis and capillary
isoelectric focusing. J Sep Sci. 2017 Jun;40(11):2488–97.

37. Lee PL, Halloran C, Trevino R, Felitti V, Beutler E. Human transferrin G277S mutation: a
risk factor for iron deficiency anaemia: Transferrin G277S Mutation and Anaemia. Br J
Haematol. 2001 Nov;115(2):329–33.

38. Pang H, Koda Y, Soejima M, Kimura H. Identification of a mutation (A1879G) of transferrin
from cDNA prepared from peripheral blood cells. Ann Hum Genet. 1998 May;62(3):271–4.

39. Evans RW, Crawley JB, Garratt RC, Grossmann JG, Neu M, Aitken A, et al.
Characterization and Structural Analysis of a Functional Human Serum Transferrin Variant
and Implications for Receptor Recognition. Biochemistry. 1994 Oct 18;33(41):12512–20.

Tables

Table 1. The MPIE-ESI-MS results of 12 secretion samples from patients suspected of CSF leak.

Secretion Sample	β1-Tf		β2-Tf	
	Detected?	Measured Molecular Mass	Detected?	Measured Molecular Mass
Positive Sample 1	Yes	79554.50 Da	Yes	78008.42 Da
Positive Sample 2	Yes	79554.79 Da	Yes	78008.60 Da
Positive Sample 3	Yes	79554.75 Da	Yes	78008.83 Da
Positive Sample 4	Yes	79554.04 Da	Yes	78008.68 Da
Positive Sample 5	Yes	79554.77 Da	Yes	78008.03 Da
Negative Sample 1	Yes	79554.29 Da	No	/
Negative Sample 2	Yes	79554.80 Da	No	/
Negative Sample 3	Yes	79554.81 Da	No	/
Negative Sample 4	No	/	No	/
Negative Sample 5	No	/	No	/
Negative Sample 6	Yes	79554.49 Da	No	/
Negative Sample 7	Yes	79554.73 Da	No	/

 Figure Captions

Figure 1. (A) Experiment workflow of MPIE-ESI-MS for Tf analysis. The MPIE-ESI-MS results

of (B) a serum sample, (C) a CSF sample, and (D) a secretion sample from a patient diagnosed of

CSF leak: HR-MS raw mass spectra (left) and deconvoluted mass spectra (right) of captured Tf

molecules, showing serum-type Tf in (B), and both serum-type Tf and brain-type Tf in (C) and

(D).

Figure 2. The MPIE-ESI-MS results of (A) the extract from the gel stripe of the β_1-Tf band

region and (B) the extract from the gel stripe of the β_2-Tf band region: HR-MS raw mass spectra

(left) and deconvoluted mass spectra (right) of captured Tf molecules, showing β_1-Tf in (B) and

β_2-Tf in (C).

Figure 3. (A) Primary structure of human transferrin, showing the sequence of 679 amino acids,

19 disulfide bonds, and 2 N-glycosylation sites. (B) N-glycan structures on β_1-Tf (major serum-

type Tf) and β_2-Tf (major brain-type Tf), confirmed by comparing the theoretical molecular

masses of the Tf glycoforms with the measured molecular masses in Figure 1, Figure 2, and

Table 1.

Figure 1

Figure 2

(A)

(B)

Figure 3

(A)

VPDKTVRWCA VSEHEATKCQ SFRDHMKSVI PSDGPSVACV KKASYLDCIR AIAANEADAV TLDAGLVYDA YLAPNNLKPV VAEFYGSKED
PQTFYYAVAV VKKDSGFQMN QLRGKKSCHT GLGRSAGWNI PIGLLYCDLP EPRKPLEKAV ANFFSGSCAP CADGTDFPQL CQLCPGCGCS
TLNQYFGYSG AFKCLKDGAG DVAFVKHSTI FENLANKADR DQYELLCLDN TRKPVDEYKD CHLAQVPSHT VVARSMGGKE DLIWELLNQA
QEHFGKDKSK EFQLFSSPHG KDLLFKDSAH GFLKVPPRMD AKMYLGYEYV TAIRNLREGT CPEAPTDECK PVKWCALSHH ERLKCDEWSV
NSVGKIECVS AETTEDCIAK IMNGEADAMS LDGGFVYIAG KCGLVPVLAE NYNKSDNCED TPEAGYFAVA VVKKSASDLT WDNLKGKKSC
HTAVGRTAGW NIPMGLLYNK INHCRFDEFF SEGCAPGSKK DSSLCKLCMG SGLNLCEPNN KEGYYGYTGA FRCLVEKGDV AFVKHQTVPQ
NTGGKNPDPW AKNLNEKDYE LLCLDGTRKP VEEYANCHLA RAPNHAVVTR KDKEACVHKI LRQQQHLFGS NVTDCSGNFC LFRSETKDLL
FRDDTVCLAK LHDRNTYEKY LGEEYVKAVG NLRKCSTSSL LEACTFRRP

Primary Structure of Human Transferrin: 679 AA, 19 Disulfide Bonds (Orange Label), 2 N-Glycosylation Sites (Blue Label)

(B)

■ N-Acetylglucosamine
● Mannose
● Galactose
◆ N-Acetylneuraminic Acid
▲ Fucose

Major Serum-Type Tf
Asn413: G2S2
Asn611: G2S2

Major Brain-Type Tf
Asn413: M5
Asn611: G0FB

Theoretical Molecular Mass: 79554.71 Da

Theoretical Molecular Mass: 78008.35 Da

Measured Molecular Mass: 79554.04 Da – 79554.81 Da

Measured Molecular Mass: 78008.03 Da – 78008.83 Da