Performance of different nebulizers in clinical use for Pressurized Intraperitoneal Aerosol Chemotherapy (PIPAC)

Short title: Technical characterisation of PIPAC nebulizers

Daniel Göhler, Dipl.-Ing.[1,2], Kathrin Oelschlägel, Dipl.-Ing[1], Mehdi Ouaissi, Prof, MD, PhD[3,4], Urs Giger-Pabst, Prof, MD[3,5]

[1] Topas GmbH, Dresden, Germany

[2] Research Group Mechanical Process Engineering, Institute of Process Engineering and Environmental Technology, Technische Universität Dresden, Dresden, Germany

[3] EA4245 Transplantation, Immunology, Inflammation, Université de Tours, France

[4] Department of Digestive, Oncological, Endocrine, Hepato-Biliary, Pancreatic and Liver Transplant Surgery, University Hospital of Tours, France

[5] Fliedner Fachhochschule, University of Applied Science Düsseldorf, Düsseldorf, Germany

Corresponding author:

Urs Giger-Pabst, MD

University of Applied Science

Fliedner Fachhochschule

Geschwister-Aufricht-Straße 9

40489 Düsseldorf, Germany

E-Mail: ursgiger@gmx.net

Phone: +49 163 729 44 07

Key words

Pressurized Intraperitoneal Aerosol Chemotherapy, PIPAC, aerosol droplet size, gravimetry, Nebulizer, MCR-4 Topol®, HurriChem™, QuattroJet

Author Contributions

Daniel Göhler and Urs Giger-Pabst: study design, experiments, data acquisition, data interpretation and drafting of manuscript.

Kathrin Oelschlägel and Mehdi Ouaissi: critical revision for important intellectual content of the manuscript according to their field of research.

Disclosure

Strictly academic study supported by institutional funds. All authors have no conflicts of interest or financial ties to declare.

Data Availability Statement

All relevant data are within the manuscript. Raw data will be provided by the corresponding author upon request.

Legal background

Purely technical analyses without the use of biological material or patients requiring no specific legal authorization or ethics vote.

Acknowledgments

The authors thank Professor Marc Pocard, Hepato-Biliary-Pancreatic Gastrointestinal Surgery and Liver Transplantation, Pitié Salpêtrière Hospital, AP-HP, F-75013 Paris, France, for providing a HurriChem™ (ThermaSolutions, White Bear Lake, MN, USA) nozzle.

Abstract

Objective: Technical ex-vivo comparison of commercial nebulizer nozzles used for Pressurized Intraperitoneal Aerosol Chemotherapy (PIPAC).

Methods: The performance of four different commercial nebulizer nozzles (Nebulizer; HurriChem^{TM}; MCR-4 TOPOL$^{®}$; QuattroJet) was analysed concerning: i) technical design and principle of operation, ii) operational pressure as function of the liquid flow rate, iii) droplet size distribution via laser diffraction spectrometry, iv) spray cone angle, spray cone form as well as horizontal drug deposition by image-metric analyses and v) chemical resistance via exposing to a cytostatic solution and chemical composition by means of spark optical emission spectral analysis.

Results: The Nebulizer shows quasi an identical technical design and thus also a similar performance (e.g., mass median droplet size of 29 µm) as the original PIPAC nozzles (MIP/ CapnoPen). All other nozzles show more or less a performance deviation to the original PIPAC nozzles. The HurriChem^{TM} has a similar design and principle of operation as the Nebulizer, but provides a finer aerosol (22 µm). The principle of operation of MCR-4 TOPOL$^{®}$ and QuattroJet differ significantly from that of the original PIPAC nozzle technology. The MCR-4 TOPOL$^{®}$ offers a hollow spray cone with significantly larger droplets (50 µm) than the original PIPAC nozzles. The QuattroJet generates an aerosol (22 µm) similar to that of the HurriChem^{TM} but with improved spatial drug distribution.

Conclusion: The availability of new PIPAC nozzles is encouraging but can also have a negative impact if their performance and efficacy is unknown. It is recommended that PIPAC nozzles that deviate from the current standard should be subject to bioequivalence testing and implementation in accordance with the IDEAL-D framework prior to routine clinical use.

1   Introduction

More than a decade ago, Pressurized Intraperitoneal Aerosol Chemotherapy (PIPAC) was introduced clinically as a new approach to deliver intraperitoneal chemotherapy to patients suffering from end-stage peritoneal surface malignancies. Using a high-pressure injector connected to a specially designed PIPAC nozzle, liquid chemotherapeutic drugs are aerosolised during laparoscopic surgery within the capnoperitoneum. This approach is expected to have a better spatial distribution pattern, greater depth of penetration, and higher drug concentration in the tissue than conventional liquid intraperitoneal chemotherapy [1, 2]. Clinical data from phase I/II and larger mono- and multicentre case series regarding safety, feasibility, and oncologic efficacy are encouraging. While the therapeutic role of PIPAC is still unclear [3], prospective randomized PIPAC trials are underway and their results are eagerly awaited [4, 5].

For over a decade, only the original PIPAC nozzle was available for clinical use, with more than 18`000 documented clinical applications worldwide were projected by the end of 2022 [3]. To ensure comparability of the outcome data, much efforts was spent to standardize PIPAC therapy worldwide [6, 7]. But more recently, new nebulizer devices are also in clinical use. While the technical and clinical performance data of the original PIPAC nozzle has been extensively studied in pre- and clinical settings [3, 8, 9], no or only very limited comparative data are available for the new PIPAC nozzles. Oncological surgeons around the world are now faced with the question of whether these newer nozzles are equivalent to the original nozzle technology or perhaps even have technical/functional advantages with a potentially better oncological outcome?

On the basis of the methodological findings regarding the technical characterisation of the original PIPAC nozzle [8], the current study deals with the comparative performance characterisation of four commercial nebulizers as nowadays used for PIPAC.

2   Materials and Methods

2.1   Examined PIPAC nozzles

Four commercial single-substance PIPAC nozzles for intraperitoneal drug aerosolization were examined, i.e.,

- Nebulizer, Model 770-12, REGER Medizintechnik, Villingendorf, Germany (A),

• HurriChem™, ThermaSolutions, White Bear Lake, MN, United States of America (B),

• MCR-4 TOPOL®, SKALA-Medica, Soběslav, Czech Republic (C),

• QuattroJet, Model 770-14, REGER Medizintechnik, Villingendorf, Germany (D).

After the experiments, all nozzles were cut-open longitudinally in the middle in a 180° angle
by means of a computerized numerical control milling machine to study their principles of
operation. In addition, also the dimensions of the nozzle outlet orifices were examined by light
microscopy (SMZ1500, Nikon, Tokyo, Japan).

2.2   Barometric characterisation of operational pressure as function of liquid flow rate

To characterise the operational pressure over the volumetric liquid flow rate, the nozzles were
connected via high-pressure hose lines with a high-pressure injector (ACCUTRON® HP-D,
MEDTRON AG, Saarbrücken, Germany) to push the test liquid (Glucosterile 5%, Fresenius
Kabi GmbH, Germany) through the nozzles. The operational pressure induced by the liquid
flow rate was determined by means of a glycerine-filled bourdon gauge (MA7U-25, JRA
Maschinenteile und Geräte GmbH, Reichenbach, Germany), which was implemented in the
high-pressure line. For the analyses, the volumetric liquid flow rate was increased stepwise
either by 0.1 ml/s (for nozzles A, B and D) or by 0.2 ml/s (for nozzle C) until the maximum
permitted pressure of 21 bar of the high-pressure injector was reached. For nozzle D, only the
axial nozzle was tested - the horizontal nozzles were sealed watertight. Analogous to [8],
measurement values were taken at steady state conditions of the aerosolization process and all
analyses were repeated three times.

2.3   Granulometric characterisation of droplet size distributions

The droplet size distributions of the aerosols generated from the test liquid (Glucosterile 5%,
Fresenius Kabi GmbH, Germany) were characterised by laser diffraction spectrometry
(PW180-C spray particle size analyser, Jinan K-Ring Technology Co., Ltd, Shandong, China)
over a size range of (0.57 - 780) µm. The outlets of the PIPAC nozzles were arranged via a
tripod in a distance of 5 mm perpendicular to the centre of the free-accessible red laser beam.
To characterise the aerosolization performance, all analyses were performed contemporaneous
with the barometric characterisation of the operational pressure for various liquid flow rates.
Analogous to [8], measurement values were taken at steady state conditions of the
aerosolization process and all analyses were repeated three times.

2.4 Image-metric characterisation of spray cone angles, form and horizontal drug
deposition areas

The spray cone angles, the form of the spray cones and the horizontal drug deposition area were
characterised with different test liquids at nozzle-specific operation conditions as recommended
by the manufacturers, i.e., at a volumetric liquid flow rate of 0.5 ml/s for the nozzles A, B, at
2.0 ml/s for nozzle C and 1.5 ml/s for nozzle D. The former two characteristics were evaluated
on the base of a 5 wt.-% aqueous glucose solution (Glucosterile 5%, Fresenius Kabi GmbH,
Germany), while the latter characteristic was assessed by operating the nozzles with undiluted
royal blue ink (Pelikan Tinte 4001[®], Hannover, Germany).

For the spray cone angle analyses, the nozzles were fixed on a tripod and vertically aligned.
Photographic images were taken with a camera that was perpendicular positioned to the nozzle
direction. The images were in-silico processed by overlaying with a digital 360° full-circle
protractor for determining the spray cone angles.

The form of the spray cones was visualized by means of a line laser (GCL 2-15, Robert Bosch
Power Tools GmbH, Leinfelden-Echterdingen, Germany) positioned in distance of 60 mm from
the nozzle orifice at right angle into the spray cone. Fully evaluated spray cone forms were
finally documented photographically.

The horizontal drug deposition on a level-aligned blotting paper was examined by operating the
vertically aligned nozzles with a distance of 60 mm between the blotting paper and the nozzle
orifice. The blotting paper was exposed for 3 s to the fully-developed spray jet. To achieve this,
a mechanical diaphragm was placed in front of the spray jet. The diaphragm was opened
automatically within 0.1 s, when the aerosol jet showed steady state nebulisation condition.

2.5 Assessing of chemical resistance and chemical composition

To assess the chemical resistance of the nozzle material against chemotherapeutic drugs, the
nozzles were at first exposed to a cytostatic solution for 12 hours and afterwards stored in the
dark at room temperature for 12 days within petri dishes. The chosen cytostatic solution was
prepared in accordance to the mixture of high pressure/high dose PIPAC (HP/HD-PIPAC) [10],
i.e., 6 mg of doxorubicin (Accord 2 mg/ml, Accord Healthcare GmbH, Munich, Germany) was
admixed with 50 ml of a 0.9 wt.-% aqueous sodium chloride solution (Ecolav[®] 100, B. Braun,
Melsungen, Germany). Finally, the nozzles were milled open in a laminar flow workbench and
macroscopic changes were documented photographically.

Moreover, the chemical composition of the nozzles pipes was characterised for the elements C,

Si, Mn, P, S, Cr, Ni, Mo, Cu, W and N by means of spark optical emission spectral analysis

(SPECTROMAXx, SPECTRO Analytical Instruments GmbH, Kleve, Germany) via an

accredited laboratory (WS Material Service GmbH, Essen, Germany).

3   Results

3.1   Technical design and principle of operation

The 90° sectional views of the head regions in Figure 1 show technical destails of the examined

nozzles.

Figure 1: 90° sectional views of the head regions of the nozzles. Legend: O = outlet orifice; H = nozzle

head; I = bar inlay with distal transverse borehole; M = double metal grid; N = fixed needle; S = shaft;

T = twist body.

Externally, all nozzles consist of a stainless steel shaft (S) with a more or less pronounced nozzle

head (H) on the lower part and a Luer lock thread on the upper part (not shown in Figure 1).

The Luer lock threads serve for the connection of the nozzles with high-pressure injectors via

high-pressure hose lines. Internally, the nozzles show partly considerable differences.

Interestingly, nozzle A and B are quasi identical in construction and their principle of operation,

while the nozzles C and D differ significantly from them and from each other.

In the case of the nozzles A, B and D, the liquid drug is supplied internally from the Luer lock

connector to the nozzle head via an annular gap between the outer shaft (S) and a bar inlay with

distal transverse borehole (I). In contrast, the internal liquid drug supply of nozzle C occurs

directly via the hollow cavity of the shaft (S). Moreover, nozzle C is equipped with a double

metal grid (M) with two different mesh sizes that serve as particle filter.

While the nozzles A and B contain one twist body (T), nozzle D is equipped with four twist

bodies (i.e., with one axial and three lateral twist bodies in 120° arrangement) to improve the

spatial drug distribution within the abdominal cavity. The twist bodies (T) of the nozzles A, B

and D contain longitudinally superfically milled grooves at 180° intervals. As the liquid drug

flow rate passes along the twist bodies (T) they were set into rotation that improves the

aerosolisation prior leaving the nozzle via the outlet orifice (O). In the case of nozzle C, the

twist body is replaced by an fixed metal needle (N). This needle contains also laterally located,

spirally milled axial grooves that induce a whirlwind effect for aerosolisation when passed by

the liquid flow before leaving the nozzle via the outlet orifice (O).

Light microscopic images of the oulet orifices (O) with determined orifice diameters of the

examined nozzles are shown in Figure 2.

Figure 2: Light microscopic images of the outlet orifices with determined orifice diameters of the
examined nozzles; scaling in mm.

3.2   *Operational parameters based on barometric and granulometric analyses*

Figure 3 a depicts at first the determined operational pressure over the liquid flow rate of the

examined nozzles, while in Figure 3 b the mass median diameter of the of the droplet size

distribution over the operational pressure is given. To avoid artefacts due clouding of the optics

of the laser diffraction spectrometer, the lateral nozzles of nozzle D (QuattroJet) were taped off

for the granulometric analyses and a flow rate of 0.5 ml/s was chosen (manufacturer-

recommended flow rate of 1.5 ml/s) Note that the shown data are determined at steady state

conditions of the aerosolization process analogous to [8].

It can be observed in Figure 3 a that the determined operational pressure data for all examined

nozzles fit well with the fluid dynamic theory, i.e., the dynamic pressure (or the dynamic

pressure drop) of an incompressible fluid increases with the fluid velocity by the power of two.

According to the equation of continuity, the fluid velocity of an incompressible fluid is in turn

directly proportional to the volumetric liquid flow rate. The nozzles A, B and D show a similar

performance regarding operational pressure and liquid flow rate, while nozzle C (MCR-4

TOPOL®) has a significantly lower pressure drop and thus a considerable higher volumetric

liquid flow rate at a specific operational pressure.

Beside the whole operational spectrum, also the manufacturer-recommended operational

conditions were separately examined, i.e., at a volumetric liquid flow rate of 0.5 ml/s for nozzle

A (Nebulizer), nozzle B (HurriChem™), (1.3 - 2.0) ml/s for nozzle C (MCR-4 TOPOL®) and

1.5 ml/s for nozzle D (QuattroJet). Under these preconditions nozzle C showed with (18 - 26) s

the shortest initiation time to reach the corresponding steady state pressure of (7.4 – 18.1) bar,

followed by nozzle A with 52 s (15.7 bar) and nozzle D with 94 s (16.0 bar). Note that with

taped-off lateral nozzles, nozzle D shows a higher operational pressure of 19.3 bar (as shown

in Figure 3).

Figure 3: Operational pressure as function of the liquid flow rate from barometric analyses (a), mass median diameter as function of the operational pressure from granulometric analyses (b) and volume-weighted distributions density (c) and cumulative distribution (d) of droplets at certain manufacturer-recommended operational condition; black cycles/ellipses indicate manufacturer-recommended operation condition.

Figure 3 b shows that the mass median diameter of the generated droplet aerosols depends for each nozzle significantly on the operational pressure. With increasing operational pressure, the mass median diameter decreases. For operational pressures of ≤ 4 bar, quasi no significant difference between the different nozzles can be observed. This is attributed to a non-fully developed aerosolization of the supplied liquid. For operation pressures of ≥ 5 bar stable aerosol

generation is reached and differences between the nozzles can be observed. For operational
pressures ≥ 5 bar, nozzle C (MCR-4 TOPOL®) shows the coarsest mass median diameters,
followed by nozzle A (Nebulizer). The finest mass median diameters were determined for
nozzle B (HurriChem™) and D (QuattroJet).

This ranking can also be deduced by the volume-weighted droplet size distributions of the
aerosols as generated by the nozzles at the manufacturer-recommended operation conditions
(Figure 3 c, Figure 3 d). Moreover, it can be observed in Figure 3 c and Figure 3 d, that each
aerosol has a polydisperse and bimodal droplet size distribution.

3.3 Operational parameters based on image-metric analyses

247 Figure 4 shows photographic images for the spray cone angle (upper panel), the spray cone
248 form (mid panel) and the horizontal drug deposition area (lower panel) of each examined nozzle
249 as determined at manufacturer-recommended operational conditions.

Figure 4: Photographic images of spray cone angle (upper panel), of spray cone form (middle panel) and horizontal drug deposition area (lower panel, scale in cm).

According to the upper panel of Figure 4, the widest single spray cone angle was determined with 79° for nozzle C (MCR-4 TOPOL®), followed with 72° for nozzle A (Nebulizer) and with 71° for nozzle B (HurriChem™). Nozzle D (QuattroJet) shows with 67° the smallest single spray cone angel, but it has to keep in mind that nozzle D contains in contrast to the other nozzles of four spray cones. Moreover, it can be observed from the middle panel of Figure 4 that nozzle A (Nebulizer), nozzle B (HurriChem™) and nozzle D (QuattroJet) generate a full spray cone, whereas nozzle C (MCR-4 TOPOL®) produces a hollow spray cone. The full spray cones of the nozzles A, B and C lead also to complete filled circular areas of horizontal drug deposition beneath the nozzles as shown in the lower panel of Figure 4. In the case of the nozzles A and B, a circular deposition area of approx. 38.5 cm² (outer diameter of approx. 7 cm) was determined. The lateral outlets of nozzle D showed beside the axial circle (outer diameter of approx. 7 cm) also 3 additional deposition areas of (13×20) cm that cumulates to an overall horizontal deposition area of approx. 679 cm².

3.5 *Chemical resistance and chemical composition*

Photographic images of the nozzle parts after the exposure to the cytostatic solution are shown in Figure 5.

Figure 5: Photographic images of the nozzle parts after exposure to the cytostatic solution.

Figure 5 shows that in the case of nozzle C (MCR-4 TOPOL®) the exposure to the cytostatic solution led to the formation of iron oxide. These are particularly pronounced on the fine-mesh particle filter, the nozzle needle and the nozzle head housing. No changes were observed for the nozzles A, B and D either visually or by light microscopic analyses.

The nozzles A, B and D fulfil all requirements according to EN 10088-3:2014 [11] on the chemical composition of stainless steel 1.4301 that is typically used for surgical instruments.

Beside a twelve times higher quantity of sulphur (0.012 wt.-% vs. 0.001wt.-%), also quantities

of molybdenum (0.183 wt.-%), copper (0.220 wt.-%) and tungsten (0.134wt.-%) were

identified by spark optical emission spectrometry for nozzle C (MCR-4 TOPOL®).

4   Discussion

Due to the current lack of knowledge, four clinically-operated nebulising nozzles for PIPAC

were comparatively tested regarding their performance. The most important determined

technical characteristics of these nozzles are summarised in Table 1**.**

Table 1:   Overview on technical and functional characteristics of the examined nozzles; * =

manufacturer-recommended operational conditions.

parameter	unit	A (Nebulizer)	B (HurriChem™)	C (MCR-4 TOPOL®)	D (QuattroJet)
liquid flow rate (Q_L)	ml/s	0.5	0.5	1.3 - 2.0	1.5
operational pressure for *Q_L	bar	15.7	14.9	7.4 – 18.1	16.0
pressure initiation time for *Q_L	s	52	100	18 - 26	94
nozzle orifice diameter	µm	200	190	370	170
mass median diameter for 15 bar	µm	28.95	20.99	52.17	24.18
max. spray angle for *Q_L	°	≈ 72	≈ 71	≈ 79	≈ 67
number of nozzles	-	1 × axial	1 × axial	1 × axial	1 × axial, 3 × lateral
kind of spray cone	-	full cone	full cone	hollow cone	full cone
drug deposition area for *Q_L	cm²	≈ 38.5	≈ 38.5	≈ 66	≈ 679

Nozzle A (Nebulizer) shows after an initiation time of 52 s an operational pressure of 15.7 bar

at the manufacturer-recommended operational liquid flow rate of 0.5 ml/s. Thereby, a full spray

jet cone (71°) composed of droplets with a mass median diameter of 29 µm is formed. The

determined data of this study reveal that nozzle A is identical in design and performance to the

primary for PIPAC developed predecessor, i.e., the microinjection pump MIP [8], which was

also distributed under the tradename CapnoPen.

Nozzle B (HurriChem™) is another launched nebulizer for PIPAC. Examinations on design

and principle of operation show a high similarity with nozzle A (Nebulizer) and thus also with

the initial PIPAC nozzle technology. At the manufacturer-recommended operational liquid flow

rate of 0.5 ml/s, nozzle B shows after an initiation time of 100 s an operational pressure of

14.9 bar. The mass median diameter of the droplets in the formed full spray jet cone (73°) was

determined to be 21 µm.

Nozzle C (MCR-4 TOPOL®) differs in technical design, principle of operation, operational

parameter and aerosol characteristics significantly from all other investigated nozzles. The

operation of nozzle C is accompanied by the formation of a hollow spray cone jet (79°). At the

manufacturer-recommended operational liquid flow rate range of (1.3 – 2.0) ml/s, operational

pressures of (7.4 – 18.1) bar were reached within short initiation times of (18 – 26) s. The mass

median droplet size decreases with increasing liquid flow rate, but was found to be in each case

larger than 50 µm.

Nozzle D (QuattroJet) is a further PIPAC nebulizing nozzle that was introduced by the

manufacturer of nozzle A. To optimize the spatial drug distribution pattern and higher

intraabdominal aerosol particle concentration, the conventional axial nozzle is supplemented in

nozzle D by three further nozzles, which are arranged lateral at the nozzle head with an angular

distance of 120°. Nozzle D is based on the same technology as nozzle A. At the manufacturer-

recommended flow rate of 1.5 ml/s, nozzle D shows an operational pressure of 16.0 bar after

an initiation time of 92 s and provides four full spray cone jets (67°) composed of droplets with

a mass median diameter of 21 µm.

Recently, a first attempt regarding recommendations on the minimum technical requirements

on nozzles suitable for PIPAC treatment was published. A minimum requirement for the spray

angle of at least 70° was defined [12] by implying that the spray cone angle corresponds to the

achievable drug deposition area. But there are two limitations found in this study that contradict

this requirement. The present study shows that the requirement is matched directly by nozzle

A, B and C. But despite of a slightly lower spray cone angle of 67° than required, nozzle D

consists of four spatially-displaced spray jets cumulating in a total spray angle of 268°. On the

other hand, Nozzle C, unlike all other nozzles examined, had a hollow spray cone, resulting in

a ring-shaped drug deposition area that was smaller than that of a full spray cone jet at the same

spray cone angle. With regard on the nozzle performance, the drug deposition area seems to be

an even better technical parameter than the spray cone angle.

The examined nozzles C (MCR-4 TOPOL®) and D (QuattroJet) of this study show in contrast

to A (Nebulizer) and B (HurriChem™) and thus in contrast to the primary PIPAC nozzle

technology significant differences in their principle of operation and performance. Nozzle C

offers the largest spray cone angle of all examined nozzles, but also a hollow spray cone. It is
not evident if such a spray jet improves drug distribution and drug penetration, since so far as
known there are no preclinical studies comparing a hollow with a full spray cone. Nozzle D
provides multiple spray cones that can significantly improve the spatial drug distribution by
reduction of high local deposition and thus high local tissue toxicity. Nonetheless, these
potential benefits of multi-nozzle systems need to be confirmed by further research.

A worrying result of this study was, that nozzle C (MCR-4 TOPOL$^®$) shows in contrast to the
other nozzles a macroscopically visible formation of iron oxide after long-term exposure in a
sodium chloride containing cytostatic solution. Based on spark optical emission spectrometry,
it was found that this nozzle was made from a steel that, in addition to a 12-fold higher sulphur
content (0.012 wt.-% vs. 0.001 wt.-%), also contained amounts of molybdenum (0.183 wt.-%),
copper (0.220 wt.-%) and tungsten (0.134 wt.-%). Short-term exposure of the nebulizers to
cytostatic solution shows no immediate corrosion, but a possible risk for the patient cannot be
completely ruled out. This leads to the conclusion that manufacturer and regulatory authorities
for medical devices accreditation must also critically assess the suitability of specific steel
alloys for the administration of cytostatic drugs in their risk assessment.

Currently, there are only limited preclinical data to suggest that there is an optimal technique
for the generation and delivery of PIPAC aerosols that could improve clinical outcome.
However, it is clear that, contrary to claims made by one manufacturer [13], larger aerosol
droplets injected into the peritoneal cavity at higher velocities with a hollow spray cone do not
improve either the spatial distribution pattern [8, 14] or tissue penetration depth per se.
Nebulizers differing from the present standard technology in its design, especially in its
spraying characteristics, cannot automatically be considered equivalent by the clinical user.
Therefore, before their broad clinical use, the individual innovation phases should be
systematically tested, ideally following the recommendation of the IDEAL-D framework for
the introduction of medical devices [15, 16]. While the original nozzle technology has
completed phase I - IIb [17] and phase III trials are ongoing [4, 5], only limited phase I clinical
user data has been published for nozzle C (MCR-4 Topol$^®$) [18]. Such data are lacking for
nozzle B and D.

In the near future, clinical users should be able to rely on technical testing and reporting being
based on scientific standards and generally applicable global standards, such as ISO standards.
For PIPAC nebulizers, such standards should ideally be set by a panel of experts through

consensus conferences. Moreover, nebulizers with significant technical and granulometric

differences from the standard technology should undergo first ex- and in-vivo animal testing

before their first clinical use. Manufacturers should be obliged to have the bioequivalence of

the cytostatic drugs administered independently certified in comparison to standard nebulizer

systems, analogous to drugs and their generics. Relevant outcome measures are aerosol

characteristics, spatial drug distribution, depth of penetration, tissue concentration and the peak

concentration and the area under the curve describing the extent of peritoneal passage [8, 19 -

28]. The ratio between the individual properties of the generic nebulizer and the reference

product would ideally be 1:1 in case of bioequivalence. Since this is unlikely to be achieved,

the US Food and Drug Administration (FDA), for example, requires that the 90% confidence

interval should be between 0.80 and 1.25 [29]. Similar to the FDA specifications, new PIPAC

nozzle technologies could be tested comparatively in the future. Such preclinical testing, ideally

using standardized models, could prevent the use of such devices from compromising clinical

outcomes and/or harming healthcare professionals/patients. Finally, it would be helpful for the

comparability of clinical results if the nebulizer type used in each case will be also recorded in

the PIPAC database (https://isspp.org/professionals/pipac-database/).

5   Conclusion

Four clinically-used nozzles to aerosolise chemotherapeutic drugs in the context of Pressurized

Intraperitoneal Aerosol Chemotherapy (PIPAC), i.e., the Nebulizer, the HurriChem[TM], the

MCR-4 TOPOL[®] and the QuattroJet were examined comparative to determine their

performance.

It could be confirmed that the Nebulizer shows quasi an identical technical design and thus also

a similar performance as the original PIPAC nozzles MIP/CapnoPen. The PIPAC nozzle

HurriChem[TM] is based on a similar technical design as the Nebulizer nozzle, but provides a

finer aerosol due to a smaller nozzle orifice opening. Both, the MCR-4 TOPOL[®] and the

QuattroJet deviate in the principles of operations to that of the Nebulizer and thus to the original

PIPAC technology. While the MCR-4 TOPOL[®] provides the coarsest aerosol of all examined

nozzles, the QuattroJet delivers an aerosol similar to that of the HurriChem[TM]. In contrast to the

HurriChem[TM], the QuattroJet[®] comes with the feature of four spray cones (one axial, three

lateral) to improve the spatial drug distribution and a higher aerosol particle number

concentration.

The availability of new PIPAC nozzles with special features is encouraging, but can also have

a negative impact for establishment of the promising PIPAC approach for the treatment of

peritoneal carcinomatosis, if their performance and efficacy is unknown. It is therefore

recommended that nozzles for which the technical/granulometric characteristics differ from the

current standard technology must be subjected to preclinical proof of equivalence in terms of

spatial drug distribution, tissue penetration and concentration before routine clinical use. It is

expected that representative in-silico models like the ones used in [23, 24] will be available in

the near future. New nebulizers should be investigated and introduced for clinical use in

accordance with the IDEAL-D framework.

References

1. Solaß W, Hetzel A, Nadiradze G, Sagynaliev E, Reymond MA Description of a novel approach for intraperitoneal drug delivery and the related device. Surg. Endosc. 2012;26(7):1849–1855. Doi: 10.1007/s00464-012-2148-0.

2. Solaß W, Kerb R, Mürdter T, et al. Intraperitoneal chemotherapy of peritoneal carcinomatosis using pressurized aerosol as an alternative to liquid solution: first evidence for efficacy. Ann. Surg. Oncol. 2014;21(2):553–559. Doi: 10.1245/s10434-013-3213-1.

3. Alyami M, Hübner M, Grass F, et al. Pressurised intraperitoneal aerosol chemotherapy: rationale, evidence, and potential indications. Lancet Oncol. 2019;20(7):e368–e377. Doi: 10.1016/S1470-2045(19)30318-3

4. Somashekhar SP, Ashwin KR, Rauthan A, Rohit KC Pressurized IntraPeritoneal Aerosol Chemotherapy vs. intravenous chemotherapy for unresectable peritoneal metastases secondary to platinum resistant ovarian cancer – study protocol for a randomized control trial. Pleura and Peritoneum 2019;4(1). Doi: 10.1515/pp-2018-0111

5. Casella F, Bencivenga M, Rosati R, et al. Pressurized intraperitoneal aerosol chemotherapy (PIPAC) in multimodal therapy for patients with oligometastatic peritoneal gastric cancer: a randomized multicenter phase III trial PIPAC VEROne. Pleura and Peritoneum 2022;7(3):135–141. Doi: 10.1515/pp-2022-0111

6. Sgarbura O, Villeneuve L, Alyami M, et al. Current practice of pressurized intraperitoneal aerosol chemotherapy (PIPAC): Still standardized or on the verge of diversification? European Journal of Surgical Oncology 2021;47(1):149–156. Doi: 10.1016/j.ejso.2020.08.020

7. Sgarbura O, Eveno C, Alyami M, et al. Consensus statement for treatment protocols in pressurized intraperitoneal aerosol chemotherapy (PIPAC). Pleura and Peritoneum 2022;7(1):1–7. Doi: 10.1515/pp-2022-0102

8. Göhler D, Khosrawipour V, Khosrawipour T, et al. Technical description of the micro injection pump (MIP®) and granulometric characterization of the aerosol applied for Pressurized IntraPeritoneal Aerosol Chemotherapy (PIPAC). Surg. Endosc. 2017;31(4):1778–1784. Doi: 10.1007/s00464-016-5174-5

9. Nadiradze G, Horvath P, Sautkin Y, et al. Overcoming Drug Resistance by Taking Advantage of Physical Principles: Pressurized Intraperitoneal Aerosol Chemotherapy (PIPAC). Cancers 2020;12(1):34. Doi: 10.3390/cancers12010034.

10. Arias GR, Sindayigaya R, Ouaissi M, et al. Safety and Feasibility of High-Pressure/High-
Dose Pressurized Intraperitoneal Aerosol Chemotherapy (HP/HD-PIPAC) for Primary and
Metastatic Peritoneal Surface Malignancies. Ann. Surg. Oncol. 2022. Doi: 10.1245/s10434-
022-12698-4.

11. EN 10088-3:2014. Stainless steels - Part 3: Technical delivery conditions for semi-finished
products, bars, rods, wire, sections and bright products of corrosion resisting steels for
general purpose.

12. Pocard M, So JBY, Huchon C, et al. PIPAC nebulizer: How to test the new devices in the
market, expert recommendations. Journal of Visceral Surgery 2022. Doi:
10.1016/j.jviscsurg.2022.10.001.

13. PIPAC – Pressurized Intraperitoneal Aerosol Chemotherapy; https://www.skala.cz/en/mcr-
4-topol-en/ (access: 5[th] March 2023 at 11:02 CET)

14. Hinds, William C. (1999). Aerosol technology: properties, behaviour, and measurement of
airborne particles. New York: Wiley. ISBN 978-0-471-19410-1.

15. Sedrakyan A, Campbell B, Merino JG, Kuntz R, Hirst A, McCulloch P IDEAL-D: a rational
framework for evaluating and regulating the use of medical devices. Br. Med. J. 2016:i2372.
Doi: 10.1136/bmj.i2372.

16. Páez A, Rovers M, Hutchison K, Rogers W, Vasey B, McCulloch P Beyond the RCT: When
are Randomized Trials Unnecessary for New Therapeutic Devices, and What Should We
Do Instead? Ann. Surg. 2021;275(2):324–331. Doi: 10.1097/SLA.0000000000005053.

17. Baggaley AE, Lafaurie GBRC, Tate SJ, et al. Pressurized intraperitoneal aerosol
chemotherapy (PIPAC): updated systematic review using the IDEAL framework. Br. J.
Surg. 2022;110(1):10–18. Doi: 10.1093/bjs/znac284.

18. Hoskovec D, Dytrych P, Vocka M, Krska Z, Skala R MCR-4 TOPOL® new device for
PIPAC – first clinical use. Eur. J. Surg. Oncol. 2022;48(2):e155. Doi:
10.1016/j.ejso.2021.12.308.

19. Rahimi-Gorji M, Debbaut C, Ghorbaniasl G, Cosyns S, Willaert W, Ceelen W Optimization
of Intraperitoneal Aerosolized Drug Delivery: A Computational Fluid Dynamics (CFD) and
Experimental Study. Scientific Reports 2021;12(1). Doi: 10.1038/s41598-022-10369-8.

20. Göhler D, Geldner A, Gritzki R, et al. Development of a rat capnoperitoneum phantom
(RCP) to study drug aerosol deposition in the context of anticancer research on peritoneal
carcinomatosis. Sci. Rep. 2021;11:21843. Doi: 10.1038/s41598-021-01332-0.

21. Göhler D, Große S, Bellendorf A, et al. Hyperthermic intracavitary nano-aerosol therapy
(HINAT) as improved approach for pressurised intraperitoneal aerosol chemotherapy
(PIPAC): Technical description, experimental validation and first proof of concept.
Beilstein J. Nanotechnol. 2017;8:2729–2740. Doi: 10.3762/bjnano.8.272.

22. Buggisch JR, Göhler D, Sobilo J, et al. Development and technical validation of an
ultrasound nebulizer to deliver intraperitoneal pressurized aerosols in a rat colon cancer
peritoneal metastases model. BMC Cancer 2022;22(570). Doi: 10.1186/s12885-022-09668-
0.

23. Rezniczek GA, Buggisch J, Sobilo J, et al. Establishment of a mouse ovarian cancer and
peritoneal metastasis model to study intraperitoneal chemotherapy. Cancers 2020;12(12).
Doi: 10.3390/cancers12123818.

24. Höltzcke P, Sautkin I, Clere S, Castagna A, Königsrainer A, Pott PP, Reymond MA.
Feasibility of pressurized intra peritoneal aerosol chemotherapy using an ultrasound aerosol
generator (usPIPAC). Surg Endosc. 2022 Oct;36(10):7848-7858. doi: 10.1007/s00464-022-
09525-y. Epub 2022 Aug 29. PMID: 36038646; PMCID: PMC9485099.

25. Toussaint L, Sautkin Y, Illing B, Weinreich FJ, Nadiradze G, Königsrainer A, Wichmann
D. Correction to: Comparison between microcatheter and nebulizer for generating
Pressurized IntraPeritoneal Aerosol Chemotherapy (PIPAC). Surg Endosc. 2021
Aug;35(8):4901. doi: 10.1007/s00464-021-08577-w. Erratum for: Surg Endosc. 2021
Apr;35(4):1636-1643. PMID: 34106308; PMCID: PMC8263410.

26. Mun J, Park SJ, Kim HS. Rotational intraperitoneal pressurized aerosol chemotherapy in a
porcine model. Gland Surg. 2021 Mar;10(3):1271-1275. doi: 10.21037/gs-2019-ursoc-11.
PMID: 33842275; PMCID: PMC8033043.

27. Mimouni, M., Richard, C., Adenot, P. et al. Pressurized intra-peritoneal aerosol
chemotherapy (PIPAC): increased intraperitoneal pressure does not affect distribution
patterns but leads to deeper penetration depth of doxorubicin in a sheep model. BMC Cancer
21, 461 (2021). https://doi.org/10.1186/s12885-021-07955-w

28. Giger-Pabst, U., Bucur, P., Roger, S. et al. Comparison of Tissue and Blood Concentrations
of Oxaliplatin Administered by Different Modalities of Intraperitoneal Chemotherapy. Ann
Surg Oncol 26, 4445–4451 (2019). https://doi.org/10.1245/s10434-019-07695-z

29. Andrade C Bioequivalence of Generic Drugs: A Simple Explanation for a US Food and

Drug Administration Requirement. Journal of Clinical Psychiatry 2015;76(06):e742–e744.

Doi: 10.4088/JCP.15f10094.