Comparison of a custom Photogrammetry for Anatomical CarE (PHACE) system with

other Low- Cost Facial Scanning Devices

Josiah K. To[1], Anderson N. Vu[1], Lilangi S. Ediriwickrema[1,3] Andrew W. Browne[1,2,3]

[1]Gavin Herbert Eye Institute, Department of Ophthalmology, University of California Irvine,

Irvine California

[2]Department of Biomedical Engineering, University of California Irvine, Irvine California

[3]Institute for Clinical and Translational Science, University of California Irvine, Irvine California

Financial Support: Arnold and Mabel Beckman Foundation Fellowship, Research to Prevent

Blindness unrestricted grant to UCI Department of Ophthalmology, and a UCI ICTS NIH KL2

Grant number KL2 TR001416.

Proprietary Interest Statement: No authors have any financial interest related to the manuscript.

Corresponding Author: Josiah To, MD, MASc, Gavin Herbert Eye Institute, Department of

Ophthalmology, University of California Irvine, Irvine California, 850 Health Sciences Rd, Irvine

CA 92697 (josiah.t2@gmail.com)

Precis: We demonstrate a custom system for facial photogrammetry (Photogrammetry for

Anatomical CarE -PHACE) to produce 3D renderings of facial volume and morphology which

compares with more expensive alternative 3D scanning technologies.

Abstract

<u>Purpose:</u> To compare a custom Photogrammetry for Anatomical CarE (PHACE) system with other

cost-effective 3-dimensional (3D) facial scanning systems to objectively characterize morphology

and volume of periorbital and adnexal anatomy.

<u>Methods:</u> The imaging systems evaluated include the low-cost custom PHACE system and

commercial software product for the iPhone called Scandy Pro (iScandy) application (Scandy,

USA), and the mid-priced Einscan Pro 2X (Shining3D Technologies, China) device and Array of

Reconstructed Cameras 7 (ARC7) facial scanner (Bellus3D, USA). Imaging was performed on a

manikin facemask and humans with various Fitzpatrick scores. Scanner attributes were assessed

using mesh density, reproducibility, surface deviation, and emulation of 3D printed phantom

lesions affixed above the superciliary arch (brow line).

<u>Results:</u> The Einscan served as a reference for lower cost imaging systems because it qualitatively

and quantitatively renders facial morphology with the highest mesh density, reproducibility (0.13

$\pm$ 0.10 mm), and volume recapitulation (approximately 2% of 33.5 μL). Compared to the Einscan,

the PHACE system (0.35 $\pm$ 0.03 mm, 0.33 $\pm$ 0.16 mm) demonstrated non-inferior mean accuracy

and reproducibility root mean square (RMS) compared to the iScandy (0.42 $\pm$ 0.13 mm, 0.58 $\pm$

0.09 mm), and significantly more expensive ARC7 (0.42 $\pm$ 0.03 mm, 0.26 $\pm$ 0.09 mm). Similarly,

the PHACE system showed non-inferior volumetric modeling when rendering a 124 μL phantom

lesion compared to the iScandy and more costly ARC7 (mean percent difference from the Einscan:

4.68 $\pm$ 3.73%, 9.09 $\pm$ 0.94%, and 21.99 $\pm$ 17.91% respectively).

<u>Conclusions:</u> The affordable PHACE system accurately measures periorbital soft tissue as well as

other established mid-cost facial scanning systems. Additionally, the portability, affordability, and

adaptability of PHACE can facilitate widespread adoption of 3D facial anthropometric technology

as an objective measurement tool in ophthalmology.

INTRODUCTION

Three dimensional (3D) digital surface reconstruction technologies and analyses have been increasingly applied to the fields of architecture, surveying, agriculture, sports, and medicine. In ophthalmology, two-dimensional photogrammetry has previously been used to evaluate and monitor tumor pathology, as well as follow a patient during the peri-operative period.[1–5] The advantage of 3D facial anthropometric measurements is the ability to non-invasively measure, reconstruct, and analyze diverse pathologies that could affect clinical decision making. For example, continued advancements in 3D facial anthropometry have assisted in craniofacial flap planning and creation for cleft lip and palate repair, as well as helped to accurately quantify facial skin wrinkles and scars before and after laser resurfacing.[6–8]

These technologies typically fall under two categories: photogrammetry and structured light sensing. Photogrammetry is the 3D spatial reconstruction of shapes and features derived from the mathematical analyses of photographs acquired from multiple vergence angles.[9] Objects in photographs are computationally aligned in 3D digital space to create object feature coordinates that are subsequently used to generate 3D surfaces.[9] Structured light technologies, in contrast, project a known pattern of light onto a subject, which are then captured by cameras from two different angles of known deviation. Each camera captures differences in the projected pattern as the light becomes distorted while conforming to an object's surface. Differences in pattern distortion are stored in reference matrices that correlate the objects' camera coordinates to sets of relative spatial coordinates and are ultimately used to render a 3D surface model of the object.[10]

Available facial photogrammetry technologies include the 3dMD Face System (3dMD, Atlanta, GA), Vectra H1 or XT (Canfield Scientific Inc., Fairfield, NJ), and Eva (Artec 3D, Luxembourg). 3dMD products use structured light sensors and are considered "active" scanners

as these devices contain a light source for optimized data capture. However, portability is limited

due to large, multi-camera systems. The Vectra H1 and the Eva use photogrammetry and are

"passive" scanners since subjects are illuminated using ambient light. Though less expensive than

the 3dMD suite, these passive scanners have been less accurate due to sequential imaging and total

duration of data acquisition.[11–13] The 3dMD Face System costs approximately $25,000 USD, the

Vectra H1 $13,000 USD and the Eva $15,000 USD.[14] Although 3D anthropometry can augment

standard clinical care and further inform medical decision making, technologies are not widely

implemented due to cost, lack of portability, and training required for complicated scanning

systems. Thus, we sought to compare and evaluate alternative low-cost 3D facial modeling systems

for widespread clinical use.

METHODS

This study received Institutional Review Board approval from the University of California,

Irvine and was conducted in accordance with the Declaration of Helsinki. Studies performed were

HIPAA-compliant and all participants enrolled provided written informed consent.

Study Design. We compared a customized low-cost automated photogrammetric facial scanning

system, termed the Photogrammetry for Anatomical CarE (PHACE) system, with other low- and

mid-priced 3D scanning technologies that employed structured light to acquire scan data. 3D

scanners that cost less than $1000 were considered low-cost and scanners more than $1000 but

below $10,000 were considered mid-priced. PHACE and the iPhone Scandy Pro smartphone

application (version 1.9.10, Scandy, USA) are low-cost systems, whereas the Einscan Pro 2X

(Shining 3D Technologies, China) and Array of Reconstructed Cameras 7 (ARC7) facial scanner

(Bellus 3D, USA) are two mid-priced structured light sensing scanners. Four tests were performed

to characterize the utility and limitations of each technique for clinical use. To evaluate precision

we quantified 3D model mesh density and reproducibility using both human subject models and a

mannikin face mask. To evaluate scan accuracy we considered the Einscan 3D to be the reference

test and compared the 3D model surface deviation and volume emulation from the other mid-price

scanner and the two low-cost scanners. The mannikin was scanned in triplicate, and human

subjects were scanned once with each of the 4 scanning modalities. Healthy adults of various

Fitzpatrick skin pigmentation scores were included (mean age 30 ± 5 years; three males; one

European, one East Asian, and one African descent). The manikin facemask was used to assess

3D scanning accuracy and reproducibility since it eliminated facial movement as a source of error

and incorporated facial contours. All faces were scanned in the same environment, location, and

by a single trained researcher.

Facial Scanning Systems. The PHACE system has been previously described by To et al.[15]

Briefly, the PHACE system used off-the-shelf motorized turntables to rotate two Google Pixel 3

smartphones (Android 11 operating system) 360 degrees. Photographs were imported into

Metashape (Agisoft, St. Petersburg, Russia) and an in-house computer script was used to construct

and store photogrammetric 3D renderings on a local computer. The constructed photogrammetric

3D renders were exported as a wavefront object (.OBJ) into the 3D point cloud open source

software CloudCompare (CC) where models were scaled and analysed[16].

The handheld Einscan captured data using structured white light and the company's

proprietary software (EXScan Pro v3.4.05). Resolution was set to maximum (0.2 mm) and the

scanner held approximately 18 inches from the subject. The scanner was rotated around the

subject's face approximately 225 degrees horizontally (ear to ear) and 140 degrees vertically

(forehead to chin).

The iPhone Scandy Pro application utilized the built-in infrared (IR) structured light *True*

Depth Camera system found on iPhone X models and newer versions. Scans for this study were

acquired with an iPhone 11 with 4GB of RAM (iOS 14). Within the Scandy Pro application,

resolution was maximized (0.5mm) and held approximately 10 inches from the face while

scanning across the face in a path similar to the Einscan.

The Bellus ARC7 facial scanner captured data using an array of seven IR structured light

cameras and processed the model within its proprietary software (ARC Scan App v1.8.12). Image

acquisition and processing occured on a dedicated Windows Surface Pro 6 laptop. During imaging,

the software instructed subjects to rotate their head left and right from shoulder to shoulder

approximately 120 degrees horizontally. Rendered 3D models were exported using maximum

resolution and minimal mesh smoothing.

3D Printed Phantom Lesions. Phantom lesions consisted of four custom 3D printed hemispheres

used to simulate variable sized ocular and adnexal volumetric pathology. Each hemispheric lesion

was printed using black polylactic acid filament (Hatchbox, California, USA) on a Prusa i3MKS

(Prusa Research, Prague, Czech Republic) 3D printer with a layer accuracy of 150 μm. Subjects'

faces were scanned without the phantom lesions to establish a baseline facial mesh model. Faces

were scanned a second time with the phantoms affixed with double-sided tape approximately 1 cm

above the superciliary arch .

Measurements and Analysis. 3D models from each scanning technique were imported as a

stereolithography (.STL) file into CC, where models were aligned, cropped, registered, and

analyzed. Data rendering and analyses was performed on a Gigabyte Z390 Aorus Ultra Gaming

PC running Windows 10 with an Intel Core i9-9900k 8-core CPU @ 3.6GHz, 48 GB of RAM, and

a Nvidia GeForce RTX 2080 graphics card.

To analyze the mesh density, 3D scans of the manikin face acquired with the PHACE

system, iScandy, and ARC7 were aligned to one of the Einscan models. Facial scans were further

cropped to the same dimensions (1 mm x 1 mm) at the midline of the glabella. The mesh face

density was subsequently calculated.

The reproducibility of each scanning technique was evaluated by measuring the model

deviation between triplicate scans acquired on the same device of the mannikin face. The triplicate

whole facial models were aligned, cropped, registered to each other, and then the distance

deviation between each scan was measured. A quantitative color-coded depth map was used to

represent the deviation between each model for both the periorbita and the entire face. Analysis of

all models was performed for the entire face and for cropped regions, which only included the

periorbital tissues.

3D Model accuracy was calculated by analyzing the surface deviation of the manikin 3D

models produced by each scanning modality to a reference model rendered by the Einscan. Prior

studies investigated the Einscan's clinical accuracy by comparing rendered facial measurements

with measurements from a vernier calipers, which showed no statistical differences.[17] 3D models

from each scanning modality were aligned and registered to the Einscan reference model and

cropped to have identical boundaries. The absolute maximum deviation, root mean square (RMS),

and distance mean in addition to standard deviation were calculated. A quantitative color-coded

depth map represented each scanning technique's regional accuracy of the mannikin's face.

To quantify the volumetric accuracy of the phantom lesions, facial models with and without

3D printed phantom lesions were exported to CC. Models were manually aligned, automatically

registered to each other using the iterative closest point function, and then exported into

Meshmixer. Facial models without phantom lesions were Boolean subtracted from facial models

with phantom lesions. Hemisphere reference diameters were measured using a digital caliper.

Volumes were calculated for large, medium, small, and mini phantoms with known dimensions of

19.75 ± 0.04 mm (2157 μL), 9.78 ± 0.05 mm (260 μL), 4.72 ± 0.05 (124 μL), and 2.80 ± 0.04 mm

(33.5 μL), respectively. Each rendered hemisphere's volume was measured using Meshmixer's

analysis stability tool and compared with reference dimensions.

Statistical Analysis. A Kruskal-Wallis analysis with multiple comparisons and a one-way

ANOVA with Tukey's multiple comparison test (IBM SPSS Statistics 27 (IBM, Armonk)) were

used to compare RMS, distance deviation means, maximum mean absolute deviations, and mean

volume percent difference to evaluate model reproducibility and accuracy. *p*-values less than 0.05

were considered to be significant.

RESULTS

The advantages and disadvantages of each 3D reconstruction technology is presented in

Table 1.

Table 1. Cost and Benefits - Different 3D imaging technologies applied to facial scanning

	EINSCAN 3D SCANNER	PHACE	iScandy	BELLUS ARC7
Scanning Type	Structured white light	Photographs	Structured infrared Light	Photos and structured infrared light
Manufacturer Precision and Accuracy (mm)	0.2 – 0.3	1	1	0.4
Render Time (min)	7 - 20	25 – 40	Immediately	1.5
Cost ($USD)	5000	500	700	8000
Portability	Moderate	Moderate	Easy	Difficult
Safety	Uncomfortable bright light	Safe	Safe	Safe
Accessibility	Proprietary hardware and software	Uses any mobile phone	Only on Apple iPhone	Proprietary hardware and software
Customizable	Low	High	Low	Low
Imaging Surface	Any surface (not black)	Any non-reflective surface	Any surface	Face only

We compared mesh density (Table 2) of each scanning technology by acquiring triplicate scans of a manikin mask using each imaging modality. The Einscan created the most consistent models (lowest standard deviation between triplicate measurements) with the highest mesh face density. The PHACE system produced the fewest mesh faces with 6.5 times fewer mesh faces than the Einscan. Both the iScandy and ARC7 produced models with meshes about half the density of the Einscan. A statistical significance was seen with fewer mesh faces produced by each modality - PHACE, iScandy, and ARC7 compared to Einscan.

TABLE 2. Scanning Precision and Model Reproducibility of Mannikin Face Mask

Scanning Modality	Mean Mesh Density	Whole Face Scans		Periorbital Anatomy Only	
		Mean Absolute Deviation (mm)	Mean Maximum Absolute Deviation (mm)	Mean Absolute Deviation (mm)	Mean Maximum Absolute Deviation (mm)
Einscan	90.33 ± 3.51*	0.01 ± 0.00, NS	0.23 ± 0.04*	0.01 ± 0.00, NS	0.20 ± 0.03*
PHACE System	13.33 ± 1.15*	0.01 ± 0.01, NS	1.76 ± 0.41*	0.07 ± 0.03*	0.86 ± 0.22*
iScandy	42.67 ± 4.93*	0.04 ± 0.04, NS	1.71 ± 0.70*	0.02 ± 0.01, NS	1.06 ± 0.23*
ARC7	37.33 ± 1.15*	0.01 ± 0.01, NS	1.13 ± 0.25*	0.04 ± 0.02, NS	1.11 ± 0.25*

* p-value ≤ 0.05; NS, not significant

The reproducibility of each scanning modality is shown in Figure 1 and Table 2. The Einscan renders models with the highest reproducibility (lowest RMS, mean absolute deviation, and mean maximum absolute deviation) and precision (lowest standard deviation). The Einscan

demonstrated better reproducibility compared to PHACE ($P = .007$) and iScandy ($P = .001$) for

an inanimate face when analyzing either the whole face or periorbita alone but not compared to

the ARC7 ($P = .475$). The iScandy demonstrated the lowest reproducibility when scanning the

whole manikin face (highest mean absolute deviation, standard deviation, and RMS). The PHACE

system demonstrated lowest reliability when modelling the nose. No statistically significant

differences in reproducibility between PHACE and all other imaging modalities was observed

when rendering the whole face.

FIG. 1. Comparison of the interscan RMS of three independently acquired scans of the same
model to evaluate scanner reproducibility.

3D models of a manikin facemask were produced with each scanning technology (Fig. 2)

where the accuracy and precision were subsequently compared using the Einscan model as a

reference (Table 3). Displacement of 3D renders from the reference model are represented as

quantitative deviation color maps (Fig. 2E-H). In the deviation maps, the color blue indicates

inward deviation toward the center of the model, red indicates outward deviation from the model's

surface, and the color scale bar indicates the degree of deviation in millimeters. The color map

produced for the Einscan model showed no deviation from itself (Fig. 2E) as it is the reference for

the other modalities. The PHACE system modeled facial features similar to the iScandy except for

the inward deviation along the nose (Fig. 2F) in the PHACE model and along the right parietal

ridge (Fig. 2F) in the iScandy model. The iScandy and PHACE consistently created whole face

3D models with lower mean deviation and RMS (Table 3), which was also reflected by the

deviation color maps (Fig. 2F – 2H). Models rendered from the ARC7 device often create facial

models with significant outward deviation along the top of the head (Fig. 2H) and create false

eyelid contours (2D white arrows) in the setting of absent globes in the manikin model (arrows in

Fig. 2D). There were no statistically significant differences in periorbital and adnexal soft tissue

mean deviation or RMS between all three modalities (Table 3) when compared with Einscan.

TABLE 3. Scanning accuracy using the Einscan model as a reference

Scanning Modality	Whole Face Scans		Periorbital Anatomy Only Soft Tissue	
	Mean Deviation (mm)	Mean RMS (mm)	Mean Deviation (mm)	Mean RMS (mm)
PHACE System	0.01 ± 0.14*	0.58 ± 0.10*	0.03 ± 0.05, NS	0.35 ± 0.03, NS
iScandy	-0.23 ± 0.15*	0.69 ± 0.20*	0.00 ± 0.01, NS	0.42 ± 0.13, NS
ARC7	1.49 ± 0.50*	3.04 ± 1.00*	0.06 ± 0.03, NS	0.42 ± 0.03, NS

* p-value ≤ 0.05; NS, not significant

FIG. 2. Comparison of 3D models of a manikin facemask produced with each scanning technology using the Einscan methodology as a reference. (A-D) Rendered mesh models of the manikin facemask. (E-H) Rendered models where color maps indicating amount of deviation from Einscan reference scan (A). Scale bar is in millimeters.

Digitally measured volumes of rendered dome-shaped phantom lesions affixed to the superciliary area (brow line) were compared with calculated volumes based on the diameters measured with a digital caliper. The Einscan qualitatively and quantitatively yielded models with the highest precision and accuracy (Fig. 3A and E). The quantitative comparison for volumetric measuremens of phantoms using Einscan, PHACE, iScandy, and ARC7 systems are shown in Figure 3E. There were no statistically significant differences between the PHACE, Einscan, and ARC7 systems for large, medium and small phantom lesions (Fig. 3E). The iScandy scanner

232 rendered larger volumes for medium sized phantoms but all modalities overestimated the true

233 volume of mini phantom lesions. The ARC7 produced models that partially recapitulated the mini

234 phantoms (Fig. 3D).

FIG. 3. Worm's eye-view of scanned and rendered 3D phantom lesions on participant 3 using
236 the Einscan Pro 2X (A), PHACE system (B), iPhone Scandy Pro (C), and ARC7 Bellus 3D
(D) of participant 2. (E), Comparison of each scanning technique's average percent deviation
237 when measuringwhite phantoms of various sizes placed along the superciliary arch on human
subjects. *V* represents volume of each hemisphere.

239 The Einscan was unable to render models of human faces having different skin tones. The

240 Einscan could only render the chin and portion of the cheeks of the subject with Fitzpatrick score

241 6 (Fig. 4A), as compared to the individual with Fitzpatrick score 2 (Fig. 4b). The model with

242 Fitzpatrick score 6 (Fig. 4A) was missing facial data for for the ears, nose, orbit, periorbital adnexa,

243 forehead and dark T-shirt worn by the subject.

FIG. 4, Comparison of 3D reconstructed models using the Einscan. Generated 3D model of a male with Fitzpatrick score 6 (A) and a male with Fitzpatrick socre 2 (B). Scale bar is in millimeters.

DISCUSSION

3D anthropometry has been extensively applied to craniomaxillofacial surgery to evaluate surgical and clinical outcomes but its application to the periorbital region has been limited. [18–29] Recent technical advances have driven down the cost and access to powerful and portable 3D scanning systems.[14,30–33] Systematic reviews have emphasized that although smart phone-based scanning systems remain less accurate, their variability remain within a clinically acceptable range for facial measurements as previously defined by an RMS value less than 2 mm from a standard reference.[31–34]

In this study, we experimentally compared low and mid-cost 3D scanning systems and discuss their clinical significance and applications. The Einscan Pro 2X was chosen to be the reference scanning device for accuracy analysis given prior validation of its clinical accuracy and precision.[17] The accuracy of the PHACE system in our investigation performed better than a similar smartphone photogrammetry protocol presented by Nightingale et al.[30] The iScandy results

demonstrated less accuracy than Rudy et al's iPhone X Scandy App study but still well within the

clinically acceptable range (RMS < 2 mm).[14] The limitation in Nightingale et al's study was its

exclusion of rendered data and analysis for the eyes and mouth, which, as our data from our human

studies have demonstrated, are areas with the most variability due to small microexpressions. The

reduced reproducibility of the PHACE system is likely related to perspective distortion from

multiple camera vergence angles and magnitude of the depth difference between the nose and the

remainder of the face. This may be overcome with telecentric lenses and greater distances between

the subject and the camera. To the authors' knowledge, the accuracy and efficacy of the ARC7

device has never been compared with other devices in a similar price range.

While published validation methods rely on displacement analysis and color depth maps

to evaluate 3D model accuracy and precision of periorbital soft tissue, few studies used volumetry

to investigate the accuracy of 3D anthropometric devices.[21] Other than To et al, there are no studies

to our knowledge that use volumetry to evaluate low-cost 3D facial scanning systems.[15]

Displacement analysis focuses on distances between mesh surfaces of a model and reference,

which less clearly represent morphological changes seen more plainly with volume analysis. For

example, volume analysis can more clearly depict quantiative changes in planar facial morphology

that do not have depth changes such as a growing facial lesion. Therefore, it is important to not

only measure the linear distance changes between analyzed and reference models but also measure

the total *in vivo* volumetric changes. Our approach to assessing digital facial reconstruction

techniques used phantom lesions made of 3D printed phantoms attached to the brow. Because the

phantoms were printed with 100 μm resolution and measured with a digital caliper with 20 μm

resolution, accurate volume calculations are possible since the error propagation due to

measurement uncertainties remains lower than the resolution of each scanning modality. Accurate

volume calculations permit the phantom lesions to be used as a known reference to compare

measured volumes. Both the Bellus and iScandy had significant average percent differences from

the calculated hemispheric volumes, which could possibly be due to a resolution limitation of

structured IR light – as opposed to shorter wavelengths included in structured white light (Einscan)

that discriminates facial details such as pigmented cutaneous features, rytids and larger contours.

When a 3D model is generated, the mesh face density determines the smallest

morphological features detectable by each scanning technique. The mesh face density has the

potential to limit the model's precision and accuracy. Differences in the number of projected

structured light patterns from the Einscan, iScandy, and Bellus likely contributed to differences in

model mesh density, with the Einscan projecting the highest density of structured light points. The

PHACE system uses photogrammetric computation to infer mesh points from 80-120 images. The

fewer number of data points compared to the density of structure light data may explain the low

mesh density derived from PHACE models. The higher the mesh face density the lower the

likelihood that it will be a limiting factor for model precision and accuracy.

The accuracy of each modality was assessed by analyzing the deviation between the

rendered manikin face generated from the PHACE system, iScandy, and ARC7 compared to the

Einscan reference model. Accuracy based on the RMS calculations indicated that the PHACE

system and iScandy can statistically render a whole face model more accurately than the ARC7.

However, color-coded depth maps more clearly indicated areas of inaccuracies, which gives users

the opportunity to assess the utility and applicability of particular facial compartments. The ARC7

discrepancy in RMS between the whole face and periorbital tissue alone was due to the isolated

but significant deviation along the superior coronal suture of the models. This deviation was

possibly due to a combination of the IR light reflecting off the head of the manikin face as well as

the need to turn the head during facial scanning acquisition, which may have led to software

tracking inconsistencies. In isolating the periorbital soft tissue, both the RMS and color depth map

indicated a high degree of accuracy and no statistically significant difference between the PHACE

system, iScandy, and ARC7. Moreover, when adding both the quantitative and qualitative

volumetry data to evaluate accuracy, the results indicated there are no statistically significant

differences in accuracy between all three scanning modalities and each system could correctly

render volumes as small as 124 μL. Therefore, the RMS, quantitative color-coded depth map, and

volumetry data all indicated similarly clinically significant, accurate performance between the

low-cost facial scanning systems and the ARC7.

The whole face quantitative color-coded depth map was useful to indicate inaccuracies of

specific regions of facial models. The PHACE system inconsistently rendered prominent nose

models, which was likely due to the limited image perspective data restricted due to the constrained

angles that the cameras have with respect to the face. Moreover, in a prior characterization study

of the PHACE system, the nose was inaccurately rendered in only two out of fifteen subjects.[To

et. al] Therefore, in subjects with prominent noses cameras in the PHACE system could be

positioned further apart to increase the perspective data for wider vergence angles.

Clinical Applications. The Einscan 3D scanner was shown to most accurately and precisely

recapitulate inanimate objects and human faces; however, it was notably limited to lighter skin

tones, with Fitzpatrick scores 1-4. Because the Einscan Pro 2X used white structured light, an

absence of reflected light from dark surfaces impeded data acquisition and model rendering. This

limitation may be addressed with next generation systems produced by Shinging3D.

The PHACE system recapitulates facial morphology and volumetry as accurately as other

low and mid-cost options and is independent of Fitzpatrick score. The PHACE system was the

most cost-effective and accessible modality since it can use any smartphone. The total cost of the

system is a few hundred dollars as opposed to a few thousand dollars for competing proprietary

hardware and software. Additionally, minimal training was required as the data acquisition and

model processing was nearly completely automated. The PHACE technique also includes the

added benefit of producing a large photographic dataset from multiple angles for subsequent 2D

image analysis if needed. However, the PHACE system required the most time to render a precise

and accurate 3D model and occasionally modeled noses with prominent bridges inaccurately. In

the clinic, the PHACE system is best suited for modeling facial morphology, depth, and volumetric

changes of the orbit and adnexal tissues.

The iPhone Scandy Pro application was an affordable/cost-effective option that used high

resolution IR structured light via the *True Depth Camera* found in newer generation iPhones

(iPhone X and later). Furthermore, the iOS Scandy Pro app allowed all data acquisition, model

processing, and storage to occur completely offline on the iPhone device without the need to send

personally identifiable facial data to the cloud, unlike the Bellus3D iOS FaceApp. Additionally,

the Scandy Pro app immediately generated models after scanning and consistently recapitulated

most regions of the face, including the noses with prominent bridges. The iPhone Scandy Pro

application is best suited for capturing whole face models in a situation that required a portable

system such as at the slit lamp, bedside, and before/after surgery.

The ARC7 facial scanner acquired data and rendered a model in the shortest time between

the four evaluated scanning modalities. Although the documentation was unclear as to how

Bellus3D renders the final facial model, the company's developer overview and API

documentation indicate a facial tracking function using software.[35,36] Therefore, it was possible

that significantly atypical facial morphology that was not recognized by the software could reduce

model accuracy. Moreover, the ARC7 is limited to imaging faces only. Therefore, the Bellus3D

ARC7 facial scanner is suitable for medical contexts that require facial models to be created

quickly, have grossly normal anatomical features, and only need anthropometric measurements

from the forehead to the chin.

Overall, the PHACE system performed similarly to other low and mid-cost scanners and it

is the most economical, accessible, and easy to use 3D anthropometric scanning tool. The iPhone

Scandy Pro modeling technique is the most portable system that can consistently and quickly

capture and render full face 3D models. The ARC7 is the fastest facial scanning system, whereas

the Einscan is the most accurate and precise scanning system but exhibits limited use.

Limitations. Limitations to this study can narrow the scope of clinical applicability. Subtle micro

expressions drastically reduce the ability for standardized comparisons between models and

ultimately lower the sensitivity of the analysis. In theory, after reconstructed models are registered

to each other in 3D space, all changes detected are due to external factors altering the region of

interest. In the clinical setting, changes between models will ideally be due to systemic or facial

pathology, such as thyroid eye disease, trauma, burn, or infiltrative diseases that alter tissue volume

and depth. However, facial micro-expression can alter the 3D model. Furthermore, when

reconstructed models are registered in 3D space, subtle non-region of interest changes can alter

the efficacy of the alignment and registration of the models to each other. Although the

reconstructed models may be misaligned on the order of millimeters, the smaller the features being

measured (e.g., a millimeter sized lesion) the greater the artificial change contributes to the error.

An additional limitation of this study was that all human subjects were imaged with their

eyes closed because each of the scanning systems struggled to render transparent/translucent

surfaces, such as the cornea. To obviate imaging the cornea directly, scans were taken with eyes

closed so that the skin covering the orbit could be imaged to represent depth and volume of the

cornea. Currently all light based photogrammetric and structured light systems also struggle with

accurately capturing and rending periorbital tissue with eyes open. The next advances in 3D facial

anthropometry development for periorbital soft tissue may benefit from methods to accurately

represent transparent tissue such as the cornea.

In conclusion, we have qualitatively and quantitatively compared 3D facial models

generated from affordable 3D reconstruction technologies. Implementing these techniques to study

changes in the midface and orbital adnexa warrants further investigation to gain understanding of

the value of these methods for objective facial anthropometry in a clinical setting.

References

1. Takamoto T, Schwartz B. Photogrammetric Measurement of Nerve Fiver Layer Thickness. *Ophthalmology*. 1989;96(9):1315-1319. doi:10.1016/S0161-6420(89)32720-5

2. Kaderli A, Katircioglu Y, Singar Ozdemir E, Kaderli ST. Long-term comparison of the efficacies of internal and external browpexy combined with blepharoplasty Comparação a longo prazo da eficácia de browpexy interna ou externa associada à blefaroplastia. *Arq Bras Oftalmol*. 2020;83(3):185-194. doi:10.5935/0004-2749.20200033

3. Romano PE. Simple photogrammetric diagnosis of optic nerve hypoplasia. *Arch Ophthalmol*. 1989;107(6):824-826. doi:10.1001/archopht.1989.01070010846026

4. Azuara-Blanco A, Spaeth GL. Methods to objectify reversibility of glaucomatous cupping. *Curr Opin Ophthalmol*. 1997;8(2):50-54. doi:10.1097/00055735-199704000-00009

5. Choi HKY, Lin W, Loon SC, et al. Facial Scanning with a Digital Camera: A Novel Way of Screening for Primary Angle Closure. *J Glaucoma*. 2015;24(7):522-526. doi:10.1097/IJG.0000000000000017

6. Markiewicz MR, Bell RB. The Use of 3D Imaging Tools in Facial Plastic Surgery. *Facial Plast Surg Clin North Am*. 2011;19(4):655-682. doi:10.1016/j.fsc.2011.07.009

7. Ey-Chmielewska H, Chruściel-Nogalska M, Frączak B. Photogrammetry and its potential application in medical science on the basis of selected literature. *Adv Clin Exp Med*. 2015;24(4):737-741. doi:10.17219/acem/58951

8. Machado BHB, Ivy ID, Pautrat WM, Frame J, Najlah M. Scientific validation of three-dimensional stereophotogrammetry compared to the IGAIS clinical scale for assessing wrinkles and scars after laser treatment. *Sci Rep*. 2021;11(1):1-11. doi:10.1038/s41598-

409 021-91922-9

410 9. Heike CL, Upson K, Stuhaug E, Weinberg SM. 3D digital stereophotogrammetry: A

411  practical guide to facial image acquisition. *Head Face Med*. 2010;6(1):1-11.

412  doi:10.1186/1746-160X-6-18

413 10. Lekakis G, Claes P, Hamilton GS, Hellings PW. Three-Dimensional Surface Imaging and

414  the Continuous Evolution of Preoperative and Postoperative Assessment in Rhinoplasty.

415  *Facial Plast Surg*. 2016;32(1):88-94. doi:10.1055/s-0035-1570122

416 11. Verhulst A, Hol M, Vreeken R, Becking A, Ulrich D, Maal T. Three-Dimensional

417  Imaging of the Face: A Comparison between Three Different Imaging Modalities.

418  *Aesthetic Surg J*. 2018;38(6):579-585. doi:10.1093/asj/sjx227

419 12. L C, M B, WW L, et al. Validation of the Vectra H1 portable three-dimensional

420  photogrammetry system for facial imaging. *Int J Oral Maxillofac Surg*. 2018;47(3):403-

421  410. doi:10.1016/J.IJOM.2017.08.008

422 13. CH T, NM A, I P, et al. Comparison of three-dimensional surface-imaging systems. *J*

423  *Plast Reconstr Aesthet Surg*. 2014;67(4):489-497. doi:10.1016/J.BJPS.2014.01.003

424 14. Rudy HL, Wake N, Yee J, Garfein ES, Tepper OM. Three-Dimensional Facial Scanning

425  at the Fingertips of Patients and Surgeons: Accuracy and Precision Testing of iPhone X

426  Three-Dimensional Scanner. *Plast Reconstr Surg*. 2020:1407-1417.

427  doi:10.1097/PRS.0000000000007387

428 15. To JK, Vu AN, Pai A, Ediriwickrema LS, Browne AW. Optimization of a

429  PHotogrammetry for Automated Anatomical CarE. 2021.

430 16. Girardeau-Montaut D. Cloudcompare-open source project. 2011.

431  https://www.danielgm.net/cc/.

17.   Amornvit P, Sanohkan S. The accuracy of digital face scans obtained from 3D scanners:

An in vitro study. *Int J Environ Res Public Health*. 2019;16(24).

doi:10.3390/ijerph16245061

18.   Knoops PGM, Beaumont CAA, Borghi A, et al. Comparison of three-dimensional scanner

systems for craniomaxillofacial imaging. *J Plast Reconstr Aesthetic Surg*. 2017;70(4):441-

449. doi:10.1016/j.bjps.2016.12.015

19.   Honrado CP, Larrabee WF. Update in three-dimensional imaging in facial plastic surgery.

Curr Opin Otolaryngol Head Neck Surg. 2004;12(4):327-331.

doi:10.1097/01.moo.0000130578.12441.99

20.   Guo Y, Rokohl AC, Lin M, Heindl LM. Three-dimensional anthropometry in periorbital

region. *Ann Eye Sci*. 2021;6:8-8. doi:10.21037/aes-20-99

21.   Hyer JN, Murta F, Juniat VAR, Ezra DG. Validating three-dimensional imaging for

volumetric assessment of periorbital soft tissue. *Orbit (London)*. 2021;40(1):9-17.

doi:10.1080/01676830.2020.1711780

22.   Da Silveira AC, Daw JL, Kusnoto B, Evans C, Cohen M. Craniofacial Applications of

Three-Dimensional Laser Surface Scanning. *J Craniofac Surg*. 2003;14(4):449-456.

doi:10.1097/00001665-200307000-00009

23.   Hammond P. The use of 3D face shape modelling in dysmorphology. *Arch Dis Child*.

2007;92(12):1120-1126. doi:10.1136/adc.2006.103507

24.   Krimmel M, Schuck N, Bacher M, Reinert S. Facial surface changes after cleft alveolar

bone grafting. *J Oral Maxillofac Surg*. 2011;69(1):80-83. doi:10.1016/j.joms.2010.03.009

25.   van Loon B, Maal TJ, Plooij JM, et al. 3D Stereophotogrammetric assessment of pre- and

postoperative volumetric changes in the cleft lip and palate nose. *Int J Oral Maxillofac*

Surg. 2010;39(6):534-540. doi:10.1016/j.ijom.2010.03.022

26. Van Loon B, Van Heerbeek N, Bierenbroodspot F, et al. Three-dimensional changes in nose and upper lip volume after orthognathic surgery. *Int J Oral Maxillofac Surg*. 2015;44(1):83-89. doi:10.1016/j.ijom.2014.08.001

27. Tenhagen M, Bruse JL, Rodriguez-Florez N, et al. Three-Dimensional Handheld Scanning to Quantify Head-Shape Changes in Spring-Assisted Surgery for Sagittal Craniosynostosis. *J Craniofac Surg*. 2016;27(8):2117-2123. doi:10.1097/SCS.0000000000003108

28. Wong JY, Oh AK, Ohta E, et al. Validity and reliability of craniofacial anthropometric measurement of 3D digital photogrammetric images. *Cleft Palate-Craniofacial J*. 2008;45(3):232-239. doi:10.1597/06-175

29. Guo Y, Rokohl AC, Schaub F, et al. Reliability of periocular anthropometry using three-dimensional digital stereophotogrammetry. *Graefe's Arch Clin Exp Ophthalmol*. 2019;257(11):2517-2531. doi:10.1007/s00417-019-04428-6

30. Nightingale RC, Ross MT, Allenby MC, Woodruff MA, Powell SK. A Method for Economical Smartphone-Based Clinical 3D Facial Scanning. *J Prosthodont*. 2020;29(9):818-825. doi:10.1111/jopr.13274

31. Mai HN, Kim J, Choi YH, Lee DH. Accuracy of portable face-scanning devices for obtaining three-dimensional face models: A systematic review and meta-analysis. *Int J Environ Res Public Health*. 2021;18(1):1-15. doi:10.3390/ijerph18010094

32. Mai HN, Lee DH. The effect of perioral scan and artificial skin markers on the accuracy of virtual dentofacial integration: Stereophotogrammetry versus smartphone three-dimensional face-scanning. *Int J Environ Res Public Health*. 2021;18(1):1-12.

doi:10.3390/ijerph18010229

33.   Mai HN, Lee DH. Accuracy of Mobile Device–Compatible 3D Scanners for Facial

Digitization: Systematic Review and Meta-Analysis. *J Med Internet Res*. 2020;22(10).

doi:10.2196/22228

34.   Aung SC, Ngim RCK, Lee ST. Evaluation of the laser scanner as a surface measuring tool

and its accuracy compared with direct facial anthropometric measurements. *Br J Plast*

Surg. 1995;48(8):551-558. doi:10.1016/0007-1226(95)90043-8

35.   Bellus3D. ARC Face Scanning System Overview for Developers ARC : Array

Reconstruction Cameras.

https://www.bellus3d.com/_assets/downloads/arc/B3D_ARC_Overview_for_Developers_

v2.pdf.

36.   Bellus3D. Bellus3D ARC Developer APIs. :1-30.

https://b3ddownload.azureedge.net/arcsdkdoc/ARC_Developer_APIs_1.12.2.pdf.