An Explainable Advanced Electrocardiography Score for Diastolic Dysfunction – Derivation, Validation and Prognostic Performance

Hayden McColl[1,2], Abhinayeni Kuhasri[1,3], Sabbab Chowdhury[1,3], Yalu Wen[4],

Daniel E Loewenstein[5], Todd T Schlegel[5,6], Maren Maanja[5], Kevin X Yang[1,3],

Zaidon Al-Falahi[1], Thomas Lindow[1,7], Michelle Price[8], Elias Fulthorp[8], Sean Lal[2,3,8,9],

Michele McGrady[2,8], Patrick Gladding[4,10], Rebecca Kozor[1], Martin Ugander[1,3,5]

[1]Kolling Institute, Royal North Shore Hospital, University of Sydney, Sydney, Australia

[2]Royal Prince Alfred Hospital, Sydney, Australia

[3]University of New South Wales, Sydney, Australia

[4]University of Auckland, Auckland, New Zealand

[5]Department of Clinical Physiology, Karolinska University Hospital, and Karolinska Institutet, Stockholm, Sweden

[6]Nicollier-Schlegel SARL, Trelex, Switzerland

[7] Respiratory Medicine, Allergology, and Palliative Medicine,, Clinical Sciences, Lund University, Lund, Sweden

[8]Central Sydney Cardiology, Sydney, New South Wales

[9]University of Sydney, Sydney, Australia

[10]North Shore Hospital, Waitemata District Health Board, Auckland, New Zealand

Corresponding Address:

Martin Ugander, MD, PhD
Professor of Cardiac Imaging, University of Sydney
Kolling Building, Level 12, Room 612017
Royal North Shore Hospital
St Leonards, Sydney NSW 2065, Australia
martin.ugander@gmail.com

Abstract

BACKGROUND AND AIMS: Diastolic dysfunction is a precursor to heart failure with

preserved ejection fraction (HFpEF), and early detection by electrocardiography (ECG) would

be valuable. We hypothesised that an explainable advanced ECG (A-ECG) score could

accurately detect diastolic dysfunction with clinically meaningful diagnostic and prognostic

performance. **METHODS**: A derivation cohort was included after standard 12-lead ECG and

echocardiography demonstrating normal systolic function, and either the presence or absence

of diastolic dysfunction (≥grade 2). A multivariable machine learning A-ECG diastolic

dysfunction score (0-100%, positive score defined as >50%) was derived using elastic net

with nested resampling. Performance for identifying diastolic dysfunction was assessed in a

separate external validation cohort. Prognostic performance for predicting cardiovascular

events was evaluated in the UK Biobank. **RESULTS**: A 7-measure A-ECG diastolic

dysfunction score derived in the derivation cohort (n=414) performed excellently in the

validation cohort (n=3,248, area under the receiver operating characteristics curve [95%

confidence interval] 0.88 [0.87-0.90]). Survival analysis in the UK Biobank (n=27,239, 966

events, follow-up 1.9 [0.7–4.4] years, age 66±8 years, 50% female), showed that compared to

a score of ≤25%, a score of >75% had a hazard ratio (HR) of 5.7 [4.5-7.2] (p<0.001), which

persisted when adjusting for age, sex, and cardiovascular risk factors (HR 2.0 [1.4-2.7],

p<0.001). Spline modelling demonstrated a linear relationship between increasing diastolic

dysfunction score and risk of events. **CONCLUSIONS**: Explainable A-ECG analysis of the

standard 12-lead ECG can be used to accurately identify diastolic dysfunction and has a

strong adjusted prognostic association with cardiovascular events.

1 Introduction

2   Left ventricular (LV) diastolic dysfunction describes abnormalities in myocardial

3 relaxation, compliance, and filling, and is fundamental to the development of heart failure

4 with preserved ejection fraction (HFpEF)[1, 2]. Diastolic dysfunction may be initially

5 asymptomatic before progressing to symptomatic HFpEF, and may be present for many years

6 before manifesting as overt heart failure[3, 4]. Aggressive management of cardiovascular risk

7 factors including hypertension, diabetes, and obesity has been shown to prevent or reduce

8 disease progression[5]. Due to diastolic dysfunction commonly being asymptomatic before

9 manifesting as HFpEF, earlier identification to better enable risk factor modification is

10 desirable.

11   Diastolic dysfunction is most commonly assessed using transthoracic

12 echocardiography. However, echocardiographic screening is not feasible on a population

13 scale due to resource-related constraints[6]. As such, a low-cost screening method for diastolic

14 dysfunction would be clinically valuable. Attempts to use the electrocardiogram (ECG) as a

15 screening tool for detecting diastolic dysfunction have shown promise. However, previous

16 studies have relied heavily upon clinical data beyond the ECG, included methods that are

17 either difficult to implement on a largescale due to requirements for dedicated hardware that

18 may not be widely available, or that require the use of artificial intelligence models that can

19 be limited by not being readily explainable[7-9].

20   Advanced electrocardiography (A-ECG) is a term describing the application of

21 multivariable statistics to a large number of measures obtained from a combination of digital

22 signal processing analysis methods applied to the standard resting 12-lead ECG[10, 11]. Notably,

23 A-ECG analysis can be performed on the digital ECG raw data recordings stored by any

24 modern ECG vendor. A-ECG analysis of the standard resting 10-sec ECG enables improved

25 diagnostic performance by incorporating not only conventional ECG measurements such as

waveform amplitudes, axes and durations, but also measures of derived three-dimensional

vectorcardiography (VCG) and waveform complexity via singular value decomposition[11-13].

A-ECG has been able to discriminate between health and disease in several conditions such as

coronary artery disease, cardiomyopathy, LV hypertrophy and LV systolic dysfunction, with

notably higher diagnostic accuracy compared to conventional ECG[11, 14, 15]. However, there

have not yet been studies examining the utility of A-ECG in the detection of diastolic

dysfunction. A-ECG may have use for screening at-risk asymptomatic patients for diastolic

dysfunction in the setting of constrained resources. Such patients might benefit from further

investigation and aggressive risk factor management to prevent progression of disease to

HFpEF and reduce its globally increasing burden.

Therefore, the aim of this study was to derive and validate an A-ECG score for

detecting diastolic dysfunction using the standard resting 12-lead ECG. We hypothesised that

combined measures of A-ECG analysis derived from the ECG could be used to accurately

identify the presence of diastolic dysfunction from echocardiography, and that this score

would have a prognostic association with cardiovascular events in a large outpatient cohort.

Methods

Study Design

This was a retrospective cohort study to derive and validate an A-ECG score for the

detection of diastolic dysfunction as determined by transthoracic echocardiography (TTE).

The derivation cohort consisted of patients who were referred for assessment at an academic

outpatient cardiology practice between January, 2018 and March, 2022. External validation

was subsequently performed in an independent cohort drawn from a separate healthcare

institution. The study received ethics approval from the Sydney Local Health District Ethics

Board (X19-0286 & 2019/ETH12064) as well as the Waitemata District Health Board (2023/93783).

Echocardiographic examinations were evaluated and patients classified as having either normal or abnormal diastolic function defined as meeting criteria for either grade 2 or 3 diastolic dysfunction per the 2016 American Society of Echocardiography guidelines. Normal diastolic function was defined as the absence of any required criteria for diastolic dysfunction[3].

Derivation cohort. The derivation cohort inclusion criteria were age >50 years, sinus rhythm, left ventricular ejection fraction (LVEF) ≥55% and a recorded 12-lead ECG within 1 month of echocardiography. Patients were excluded if they had moderate or greater valvular disease, LVEF <55%, hypertrophic cardiomyopathy, previous valvular replacement, coronary artery bypass surgery, or any other cardiac procedures that had required entering the pericardium. Exclusion criteria were ECG abnormalities that would limit A-ECG analysis such as atrial arrhythmias, predominant ventricular ectopic beats or QRS duration >120ms.

Validation cohort. The diagnostic validation cohort had a broad inclusion criterion of all patients with either normal diastolic function, or grade 2-3 diastolic dysfunction. Again, patients were excluded for ECG abnormalities which would limit A-ECG analysis. Sensitivity analysis of our score's performance was then included by the addition of grade 1 and indeterminate diastolic function to our validation cohort. The inclusion flowchart for the validation and derivation cohorts is outlined in the supplementary material.

Prognosis cohort. The prognostic performance of the A-ECG score for diastolic dysfunction was evaluated in a cohort from the UK biobank, which is an outpatient population recruited from the general public in the UK. Prognosis was assessed as a composite of time to first cardiovascular event or all-cause mortality following the date of the index ECG. Diagnostic codes (ICD-10) and procedural codes (OPCS-4) classified as events are detailed in the

supplementary material. Included subjects had undergone a resting 12-lead ECG as well as a

baseline questionnaire detailing their medical, family and psychosocial history. Included

subjects were divided into those without pre-existing cardiovascular disease and those with

either self-reported cardiovascular disease (stroke, myocardial infarction or angina) or a pre-

existing diagnosis of cardiac conditions as detailed by diagnostic and procedural codes

(Appendix C). Subjects were excluded if they had ECG abnormalities which would limit

analysis, such as significant atrial or ventricular arrhythmia. A subject inclusion flowchart is

detailed in the supplementary material.

Advanced electrocardiography analysis

All ECG recordings were performed by a commercially available ECG system that

acquired a standard resting 12-lead ECG of 10 second duration (Mortara Instruments Inc,

USA; General Electric, USA). A-ECG analysis and its extensive methods have been detailed

previously[11, 16]. In brief, conventional ECG parameters, including all major waveform

intervals, axes, and voltages were derived automatically from the ECG recording software. A-

ECG parameters were obtained after signal averaging the ECG recording using in-house

developed software[11, 13, 16]. Vectorcardiographic measures were obtained from the Frank X, Y,

and Z leads derived from the 12-lead ECG using Kors' regression transformation[17]. QRS- and

T-wave complexity measures were quantified from eigenvector lead waveforms calculated

using singular value decomposition of the 12-lead ECG[11, 16].

Statistical Analysis

Statistical analyses including feature selection procedures were conducted using R

version 4.2.0 (R Core Team, R Foundation for Statistical Computing, Vienna, Austria) with

key packages tidymodels, glmnet, survival, plotRCS used for data analysis and visualisation

[18-21]. A multivariable A-ECG score for assessing diastolic dysfunction was derived using

elastic net logistic regression[22]. Nested resampling was applied to determine cross-validated

diagnostic accuracy and obtain 95% confidence intervals (CI). Nested resampling is a method

of cross validation and model evaluation that uses two levels of repeated sampling to

minimise the risk of overfitting by optimisation bias in a predictive model[23]. The A-ECG

score was then used to estimate the probability of a given subject in the study cohort having

diastolic dysfunction, with a cut-off predicted probability of >50%. The score calculated the

likelihood of having diastolic dysfunction from the elastic net regression as:

$Likelihood\ of\ diastolic\ dysfunction$

$$= \frac{1}{1 + e^{-(coefficient_1 \times parameter_1 + coefficient_2 \times parameter_2 \ldots + adjusted\ constant)}} \times 100$$

Time-to-event analysis was performed using Kaplan–Meier curves with censoring at the

conclusion of follow-up, with subjects grouped into quartiles of their A-ECG score. Restricted

cubic spline plots were generated to demonstrate changes in hazard ratio with increasing

diastolic dysfunction score as a continuous variable. Multivariable spline analysis was also

undertaken to assess the impact of increasing diastolic dysfunction score whilst adjusting for

age and other cardiovascular risk factors including hypertension, diabetes mellitus,

dyslipidaemia, family history, smoking, and elevated body mass index.

Results

Study Population

The derivation cohort consisted of 414 patients, 264 (64%) with normal diastolic

function and 150 (36%) with diastolic dysfunction (133 [32%] with grade 2 and 17 [4%] with

grade 3 diastolic dysfunction). In the normal diastolic function group, 45% were males and

mean age was 62±9 years. In the diastolic dysfunction group, 35% were males and mean age

was 77±9 years. Detailed demographic information of the derivation cohort is presented in the

supplementary material. The diagnostic validation cohort consisted of 3248 patients, 2649

(82%) with normal diastolic function and 599 (18%) with diastolic dysfunction (518 [16%]

with grade 2 and 81 [2%] with grade 3 diastolic dysfunction). Detailed demographic

information of the validation cohort is presented in Table 1. Sensitivity analysis was

performed in an expanded population (total n=4229) by the addition of patients with grade 1

diastolic dysfunction (n=123 [3%]) and indeterminate diastolic function (n=858 [20%]).

Baseline characteristics of this expanded sensitivity analysis cohort are outlined in the

supplementary material.

The UK biobank population for prognostic validation consisted of 27239 subjects, mean age

66±8 years, 50% male. The mean follow-up duration was 1.9 [0.7-4.4] years with 966

cardiovascular events occurring over this period. 23051 (85%) of these subjects had no

previous cardiac history at the time of their inclusion. The baseline characteristics of the UK

biobank cohort are presented in Table 2.

Advanced-ECG Score

ECG feature selection identified an optimal A-ECG score containing 7 ECG measures.

These ECG measures are detailed in Table 3 including descriptive statistics for these ECG

measures on a group level between the normal and diastolic dysfunction groups. This 7-

measure score provided excellent discrimination in the derivation cohort (AUC 0.89 [0.86-

0.92], as well as in the validation cohort (AUC 0.88 [0.87-0.90], see Table 4 for detailed

performance markers). Figure 1 shows representative ECGs from two patients with high

(71%) and low (25%) A-ECG scores for diastolic dysfunction, respectively. Sensitivity

analyses were undertaken by the addition of patients with grade 1 diastolic dysfunction and

indeterminate function, respectively, to the validation cohort. The A-ECG score retained its

ability to identify grade 2 and 3 diastolic dysfunction from patients with grade 1 diastolic

dysfunction and normal diastolic function (AUC 0.87[0.85-0.88]). Even with the addition of

patients with indeterminate diastolic function to the validation cohort, the score was able to

identify patients with grade 2-3 diastolic dysfunction with a robust albeit slightly attenuated

discriminative ability (AUC 0.83 [0.81-0.84]).

Prognosis in the UK Biobank

Prognostic validation of the score was conducted in a cohort selected from the UK Biobank. Subjects divided into quartiles based on their A-ECG score demonstrated increasing risk of cardiovascular events with increasing score (log-rank 338, p<0.001, Figure 2). When compared to those with scores ≤25%, subjects with a score >75% had a hazard ratio (HR) of 5.7 [4.5-7.2] (p<0.001). This elevated risk persisted when patients were divided into subgroups with no prior diagnosis of cardiovascular disease and those with known cardiovascular disease (HR 5.5 [4.0-7.5], p<0.001, and HR 3.3 [2.3-4.5], p<0.001, respectively for diastolic dysfunction scores >75%). A comparison of hazard ratios by quartile of A-ECG score is outlined in Table 5. Restricted cubic spline modelling demonstrated an essentially linear relationship between increasing diastolic dysfunction score and cardiovascular events (p<0.001). This relationship persisted, albeit attenuated even when adjusting for age and co-morbidities (p<0.001, Figure 3).

Discussion

The major finding of this study is that an explainable 7-measure A-ECG score derived from a standard 12-lead ECG can identify diastolic dysfunction with excellent diagnostic and strong prognostic performance. The score's reported likelihood ratios demonstrate the utility of A-ECG analysis on an individual level independent of the prevalence of diastolic dysfunction in the population being studied. For example, the score's inverse negative likelihood ratio of 2.7 suggests that following a negative A-ECG result, a patient is 2.7 times less likely to be diagnosed with diastolic dysfunction on TTE. Furthermore, the score's prognostic performance in the UK biobank population suggests that in addition to diagnostic performance for identifying diastolic dysfunction, the score associates with cardiovascular events and all-cause mortality even after adjusting for age, sex, and cardiovascular risk factors. This further strengthens the utility of A-ECG as a low-cost screening tool for

identifying diastolic dysfunction in at-risk populations potentially amenable to targeted

prevention before development of HFpEF.

Validity and feasibility of A-ECG analyses

The introduction of the ECG in 1902 provided, for the first time, objective information

about the electrical activity and function of the heart that could subsequently be used to

diagnose a number of cardiac pathologies such as atrial fibrillation and myocardial

infarction[24,25, 26]. However, it was only in the 1960s that a multivariable approach to the

conventional ECG analysis was applied to improve diagnostic detection of diseases such as

LV hypertrophy and prior myocardial infarction[27, 28]. A-ECG analysis is an extension of this

arc of development, and is able to extract diagnostic and prognostically valuable information

from a standard resting 12-lead ECG for the detection of a large range of cardiac

pathologies[11]. It has been shown to improve diagnostic accuracy when compared to

conventional ECG in coronary artery disease, cardiomyopathies, LV hypertrophy and LV

systolic dysfunction[11, 14, 15].

To date, there is no widely accepted ECG-based method for detecting diastolic

dysfunction. A machine learning model using clinical and ECG variables has shown potential

to detect diastolic dysfunction with an AUC of 0.80 in an external test cohort[7]. However, that

model relied on inclusion of a range of information including clinical co-morbidities such as

hypertension and dyslipidaemia that may limit its applicability as a screening tool when such

information is not readily available. Signal-processed ECG analysis has also demonstrated

utility in in detecting reduced mitral annular velocities, however data evaluating that method

have been limited to feasibility assessments with small patient cohorts[29]. A single-lead ECG

recorded for 3 minutes using specialised dedicated hardware and a sampling rate far

exceeding routine clinical ECG has also been able to achieve high diagnostic accuracy for

detecting diastolic dysfunction[8]. However, a major barrier for widespread use of that method

as a routine screening tool in large populations is the need for dedicated hardware. Artificial

intelligence (AI) derived models have also shown promise for both diagnosing diastolic

dysfunction and assessing prognostic performance[9, 30]. However, a key limitation of AI-

derived models is their 'black box' nature, providing limited insight into the ECG features

that underpin their diagnostic performance. By comparison, A-ECG analysis can be

performed on a digital recording from a standard, 10-second, resting 12-lead ECG recorded

using existing clinical standard-fidelity ECG devices[11]. In contrast to AI models, the A-ECG

features contributing to the score are transparently known and readily explainable, which may

not only improve generalizability but also help drive both clinical acceptance as well as an

understanding of the underlying mechanisms of diastolic dysfunction.

The Cornell voltage product was the single strongest performing A-ECG feature in the

diastolic dysfunction score. This is consistent with prior research which suggests that an

increased Cornell voltage product could be used to help diagnose diastolic dysfunction and

was a marker of poor prognosis[31]. The score also includes different vectorcardiographic

measures of the spatial representation of the QRS complex as a loop in three-dimensional

space. Indeed, such measures have been shown to have diagnostic and prognostic power in

cardiovascular disease states including other heart failure phenotypes[32, 33].

Two of the six A-ECG features included in the score examined the T wave

morphology. Differences in T wave characteristics and the QT duration have also been

associated with the presence of diastolic dysfunction. It is believed that the progressive

myocardial structural changes seen in diastolic dysfunction impact calcium cycling during

ventricular relaxation which can have subtle changes detectable in the 12-lead ECG[34, 35].

As such, the current study shows that the use of an A-ECG score for detecting

diastolic dysfunction achieves excellent diagnostic performance with methods that are

economically and logistically feasible for widespread clinical use. In addition to its diagnostic

utility, our validation demonstrates that an elevated A-ECG diastolic score carries prognostic information independent of traditional risk factors for cardiovascular disease in a large out-patient population with low prevalence of prior known cardiovascular disease. Furthermore, A-ECG analysis can be performed automatically, rapidly, and inexpensively using a standard resting 12-lead ECG while at the same time also diagnosing or screening for other pathologies [36]. Moreover, it is likely that many patients at risk of having or developing diastolic dysfunction would have pre-existing resting 12-lead ECGs from the past for convenient retrospective analysis.

Practical applications of A-ECG scores

In the future, A-ECG may have application in a range of different clinical settings, in particular in primary care where this novel information may allow for more accurate risk stratification for cardiac disease and assist in identifying only those patients who might most benefit from further cardiac imaging. In addition, the low skill and cost required to perform this analysis could provide substantial value in population-based settings wherein relatively technically complex and expensive methods such as echocardiography might not be feasible. Another beneficial aspect of A-ECG is that it can simultaneously provide probability scores for a wide range of cardiac disease states [36]. Upon knowing the results of transparently explainable numeric probability scores available at the time of consultation, physicians may be better equipped to engage with patients regarding preventative therapy and lifestyle changes to reduce the risk of developing cardiovascular disease. Diastolic dysfunction is under-diagnosed and a precursor to clinically overt HFpEF, the leading cause of heart failure[37]. As such, a convenient and low-resource screening tool for the diagnosis of diastolic dysfunction would be particularly valuable to aid early identification of individuals who might benefit from further cardiac assessment and risk factor modification to prevent or delay the development of HFpEF. Additional, fully prospective follow-on studies using the proposed optimal A-ECG score are therefore justified.

Limitations

The database used to train and validate this model had limited demographic information on patients, including a lack of information on the prevalence of cardiovascular risk factors such as diabetes and hypertension. Given that the control group was recruited from patients referred for outpatient cardiac assessment, it is also likely that this population contains a higher prevalence of cardiac co-morbidities than what would be expected from asymptomatic, healthy controls. It is also likely that even age-matched healthy controls would display an increased likelihood of diastolic dysfunction due to normal healthy ageing. Despite these potential shortcomings, the score demonstrated both robust diagnostic performance and prognostic performance in independent datasets. The diagnostic performance validation was performed in an external cohort selected from another institution in a different country, and among patients with broad inclusion criteria, supporting its utility in other populations. The prognostic performance in the UK Biobank underscores utility in an outpatient population that was largely without prior known cardiovascular disease, highlighting its utility in such populations moving forward.

The strength of any predictive tool is dependent on the population tested. The utility of the score might also be limited in patients underrepresented in our study cohort, in particular patients younger than 50 years of age. However, it is worth noting that younger patients are less likely to be a target population for screening for diastolic dysfunction. Furthermore, only patients in sinus rhythm and without significant conduction disease were included in the study. Patients with atrial arrhythmia and/or ventricular conduction disease, both prevalent in patients with diastolic dysfunction and heart failure, would therefore not be screened using the proposed A-ECG score. It would therefore be beneficial to develop additional, dedicated scores for patients with atrial fibrillation or advanced conduction disease in the future. Whilst this study did not include any patients in atrial fibrillation or flutter, our 7-measure score did not use any P-wave derived measures and demonstrated excellent diagnostic accuracy.

Whether this score might therefore have diagnostic utility in patients with atrial arrhythmias remains to be ascertained. Age is a feature of the score, and diastolic dysfunction prevalence is known to increase with age[5]. By comparison, sex did not make it into the score. Despite age being included in the score, the score retained a robust association with prognosis in the UK Biobank population when adjusted for age, sex, and cardiovascular risk factors. This indicates that age is not the sole driver of the multivariable score, and the ECG features provide unique information. Finally, a substantial proportion of patients undergoing diastolic function assessment are classified as having indeterminate diastolic function. The score demonstrated attenuated performance in sensitivity analyses that included this subgroup. As our understanding of diastolic dysfunction evolves, more accurate classification of these patients may be possible, potentially enhancing the performance of predictive models in this subgroup.

Conclusions

An inherently explainable and transparent score derived from A-ECG analysis of the standard resting 12-lead ECG is able to detect diastolic dysfunction with high diagnostic accuracy and clinically meaningful prognostic performance. A-ECG shows potential as a screening tool for diastolic dysfunction, to potentially reduce the number of unnecessary echocardiographic examinations while facilitating early management of cardiovascular risk factors to help mitigate progression to HFpEF.

Disclosures

TTS is owner and founder of Nicollier-Schlegel SARL, which performs ECG interpretation consultancy using software that can quantify the advanced ECG measures used in the current study. TTS and MU are founders of Advanced ECG Systems, a company that is developing commercial applications of advanced ECG technology used in the current study. RK has a financial interest in Advanced ECG Systems through being married to MU.

References

1.   Nagueh SF. Left Ventricular Diastolic Function: Understanding Pathophysiology, Diagnosis, and Prognosis With Echocardiography. JACC Cardiovasc Imaging. 2020;13(1, Part 2):228-44.

2.   Borlaug BA, Kass DA. Invasive hemodynamic assessment in heart failure. Cardiol Clin. 2011;29(2):269-80.

3.   Nagueh SF, Smiseth OA, Appleton CP, Byrd BF, 3rd, Dokainish H, Edvardsen T, et al. Recommendations for the Evaluation of Left Ventricular Diastolic Function by Echocardiography: An Update from the American Society of Echocardiography and the European Association of Cardiovascular Imaging. J Am Soc Echocardiogr. 2016;29(4):277-314.

4.   Heidenreich PA, Bozkurt B, Aguilar D, Allen LA, Byun JJ, Colvin MM, et al. 2022 AHA/ACC/HFSA Guideline for the Management of Heart Failure: A Report of the American College of Cardiology/American Heart Association Joint Committee on Clinical Practice Guidelines. Circulation. 2022;145(18):e895-e1032.

5.   Satpathy C, Mishra TK, Satpathy R, Satpathy HK, Barone EJ. Diagnosis and management of diastolic dysfunction and heart failure. Am Fam Physician. 2006;73(5):841-6.

6.   Borlaug BA, Paulus WJ. Heart failure with preserved ejection fraction: pathophysiology, diagnosis, and treatment. Eur Heart J. 2010;32(6):670-9.

7.   Kagiyama N, Piccirilli M, Yanamala N, Shrestha S, Farjo PD, Casaclang-Verzosa G, et al. Machine Learning Assessment of Left Ventricular Diastolic Function Based on Electrocardiographic Features. J Am Coll Cardiol. 2020;76(8):930-41.

8.   Kuznetsova N, Gubina A, Sagirova Z, Dhif I, Gognieva D, Melnichuk A, et al. Left Ventricular Diastolic Dysfunction Screening by a Smartphone-Case Based on Single Lead ECG. Clin Med Insights: Cardiol. 2022;16:11795468221120088.

9.   Lee E, Ito S, Miranda WR, Lopez-Jimenez F, Kane GC, Asirvatham SJ, et al. Artificial intelligence-enabled ECG for left ventricular diastolic function and filling pressure. npj Digital Medicine. 2024;7(1):4.

10.   Maanja M. Advanced Electrocardiography in Myocardial Electrical Remodeling-Insights from Cardiovascular Magnetic Resonance Imaging: Karolinska Institutet (Sweden); 2020.

11.   Schlegel TT, Kulecz WB, Feiveson AH, Greco EC, DePalma JL, Starc V, et al. Accuracy of advanced versus strictly conventional 12-lead ECG for detection and screening of coronary artery disease, left ventricular hypertrophy and left ventricular systolic dysfunction. BMC Cardiovasc Disord. 2010;10(1):1-11.

12.   Lindow T, Palencia-Lamela I, Schlegel TT, Ugander M. Heart age estimated using explainable advanced electrocardiography. Sci Rep. 2022;12(1):1-10.

13.   Schlegel TT, Kulecz WB, DePalma JL, Feiveson AH, Wilson JS, Rahman MA, et al. Real-time 12-lead high-frequency QRS electrocardiography for enhanced detection of myocardial ischemia and coronary artery disease. Mayo Clin Proc. 2004;79:339-50.

14.   Maanja M, Schlegel TT, Kozor R, Lundin M, Wieslander B, Wong TC, et al. The electrical determinants of increased wall thickness and mass in left ventricular hypertrophy. J Electrocardiol. 2020;58:80-6.

15.   Johnson K, Neilson S, To A, Amir N, Cave A, Scott T, et al. Advanced electrocardiography identifies left ventricular systolic dysfunction in non-ischemic cardiomyopathy and tracks serial change over time. J Cardiovasc Dev Dis. 2015;2(2):93-107.

16.   Batdorf BH, Feiveson AH, Schlegel TT. The effect of signal averaging on the reproducibility and reliability of measures of T-wave morphology. J Electrocardiol. 2006;39(3):266-70.

17.   KORS JA, VAN HERPEN G, SITTIG AC, VAN BEMMEL JH. Reconstruction of the Frank vectorcardiogram from standard electrocardiographic leads: diagnostic comparison of different methods. European Heart Journal. 1990;11(12):1083-92.

18.   Tay JK, Narasimhan B, Hastie T. Elastic Net Regularization Paths for All Generalized Linear Models. Journal of Statistical Software. 2023;106(1):1 - 31.

19.   Kuhn MW, Hadley. Tidymodels: a collection of packages for modeling and machine learning using tidyverse principles. 2020.

20.   Therneau TM. A Package for Survival Analysis in R. 2020.

21.   Huo R. plotRCS: Plot Restricted Cubic Splines Curves. R package version 0.1.4 ed2023.

22.   Zou H, Hastie T. Regularization and Variable Selection Via the Elastic Net. Journal of the Royal Statistical Society Series B: Statistical Methodology. 2005;67(2):301-20.

23.   Bischl B, Mersmann O, Trautmann H, Weihs C. Resampling Methods for Meta-Model Validation with Recommendations for Evolutionary Computation. Evolutionary Computation. 2012;20(2):249-75.

24.   Howell JD. Diagnostic Technologies: X-Rays, Electrocardiograms, and CAT Scans. S Cal L Rev. 1991;65:529.

25.   Lewis T. Report CXIX. Auricular fibrillation: a common clinical condition. BMJ. 1909;2(2552):1528.

26.   Levine SA. Coronary thrombosis: its various clinical features. Medicine. 1929;8(3):245-418.

27.   Pipberger HV, McCaughan D, Littmann D, Pipberger HA, Cornfield J, Dunn RA, et al. Clinical application of a second generation electrocardiographic computer program. Am J Cardiol. 1975;35(5):597-608.

28.   Kyle MC, Klingeman JD, Conrad JD, Freis ED, Pipberger HV. A new microcomputer-based ECG analysis system. Clin Cardiol. 1983;6(9):447-55.

29.   Sengupta PP, Kulkarni H, Narula J. Prediction of Abnormal Myocardial Relaxation From Signal Processed Surface ECG. Journal of the American College of Cardiology. 2018;71(15):1650-60.

30.   Jentzer JC, Lee E, Attia Z, Hillerson D, Kane GC, Lopez-Jimenez F, et al. Artificial Intelligence ECG Diastolic Dysfunction and Survival in Cardiac Intensive Care Unit Patients. J Am Heart Assoc. 2025;14(5):e037839.

31.   Tan ES, Chan SP, Xu CF, Yap J, Richards AM, Ling LH, et al. Cornell product is an ECG marker of heart failure with preserved ejection fraction. Heart Asia. 2019;11(1):e011108.

32.   Perkins Sidney J, Monovoukas D, Chopra Z, Kucharski K, Powell C, Vejalla A, et al. Vectorcardiography Predicts Heart Failure in Patients Following ST Elevation Myocardial Infarction. Annals of Noninvasive Electrocardiology. 2024;29(5):e70013.

33.   Richman JL, Wolff L. The spatial vectorcardiogram in congenital heart disease and right ventricular hypertrophy. American Heart Journal. 1955;50(1):85-99.

34.   Prenner SB, Shah SJ, Goldberger JJ, Sauer AJ. Repolarization Heterogeneity: Beyond the QT Interval. Journal of the American Heart Association.5(5):e003607.

35.   Vyas H, O'Leary PW, Earing MG, Cetta F, Ackerman MJ. Mechanical Dysfunction in Extreme QT Prolongation. Journal of the American Society of Echocardiography. 2008;21(5):511.e15-.e17.

36.   Hughes RK, Thornton GD, Malcolmson JW, Pierce I, Khoury S, Hornell A, et al. Accurate diagnosis of apical hypertrophic cardiomyopathy using explainable advanced electrocardiogram analysis. EP Europace. 2024;26(4).

37.   van Riet EE, Hoes AW, Limburg A, Landman MA, van der Hoeven H, Rutten FH. Prevalence of unrecognized heart failure in older persons with shortness of breath on exertion. European journal of heart failure. 2014;16(7):772-7.

1 **Tables**

2 **Table 1. Validation Cohort Patient characteristics**

Variable	Normal diastolic function	Grade 2-3 Diastolic Dysfunction
Number of patients, n	2649	599
Male sex, n (%)	1237 (47)	202 (34)
Age, years	51 ± 18	75 ± 13
Body surface area, m^2	1.9 ± 0.3	1.8 ± 0.3
Echocardiographic data		
Ejection fraction (%)	63 ± 5	56 ± 14
LA volume index, ml/m^2	30 ± 7	49 ± 14
LV septal wall thickness, mm	9 ± 2	11 ± 3
LV posterior wall thickness, mm	8 ± 2	10 ± 3
E/A ratio	1.3 ± 0.6	1.7 ± 6.2
Average E/e' ratio	7.5 ± 2.2	16.0 ± 7.1
TR velocity, m/s	2.3 ± 0.3	3.0 ± 0.5
Peak E-Wave velocity, cm/s	71 ± 18	93 ± 25
Septal e' velocity, cm/s	8.9 ± 2.9	5.1 ± 1.8
Lateral e' velocity, cm/s	11.5 ± 4.0	6.4 ± 2.3

4 All data reported as mean ± standard deviation. All other values represent total number

5 (percent) of affected individuals. Abbreviations: E/e' = E to septal e' velocity ratio; LA = left

6 atrial; LV = left ventricular; TR = Tricuspid regurgitation.

1 **Table 2. UK biobank characteristics**

Characteristics	UK Biobank Population, all subjects	Group with no prior history of cardiovascular disease	Group with prior history of cardiovascular disease
n (%)	27,239 (100)	23,051 (85)	4,188 (15)
Age, years	66 ± 8	66 ± 8	69 ± 7
Height, cm	170 ± 9	170 ± 9	171 ± 9
Weight, kg	76 ± 15	75 ± 15	78 ± 15
Body Surface Area, m²	1.89 ± 0.22	1.88 ± 0.22	1.92 ± 0.22
Male sex, n (%)	13,528 (50)	11,015 (48)	2,513 (60)
Family History of Cardiovascular Disease, n (%)	11,760 (43)	9,755 (42)	2,005 (48)
Body Mass Index, kg/m²	26.5 ± 4.5	26.4 ± 4.4	27.2 ± 4.6
Physically Inactive, n (%)	2,994 (11.0)	2,458 (10.7)	536 (12.8)
Diabetes, n (%)	1,667 (6.1)	1,188 (5.2)	479 (11.4)
Hypertension, n (%)	8,030 (29.5)	5,833 (25.3)	2,197 (52.5)
Hyperlipidaemia, n (%)	5,548 (20.4)	3,650 (15.8)	1,898 (45.3)
Medications, n (%)	-	-	-
Lipid-lowering	4,590 (16.9)	3,109 (13.5)	1,481 (35.4)
Antihypertensives	4,120 (15.1)	2,871 (12.5)	1,249 (29.8)
Insulin	134 (0.5)	94 (0.4)	40 (1.0)

Any of the three	6,120 (22.5)	4,383 (19.0)	1,737 (41.5)
Smoking, n (%)			
Current	857 (3.1)	713 (3.1)	144 (3.4)
Previous	9,185 (33.7)	7,575 (32.9)	1,610 (38.4)
Never	17,098 (62.8)	14,681 (63.7)	2,417 (57.7)
Any alcohol, n (%)	-	-	-
Current	25,351 (93.1)	21,531 (93.4)	3,820 (91.2)
Previous	1,025 (3.8)	818 (3.5)	207 (4.9)
Never	846 (3.1)	687 (3.0)	159 (3.8)

All data reported as mean ± standard deviation. All other values represent total number
(percent) of affected individuals.

Table 3. A-ECG measures involved in the respective scores.

Coefficient estimates derived from elastic net logistic regression, with standardized estimates to allow for approximation of relative contribution to the overall score. Median [IQR] values for each coefficient in the normal diastolic function and diastolic dysfunction cohorts.

Feature	Coefficient	Standardized Coefficient	Values for normal diastolic function group, n=2649	Values for grade 2-3 diastolic dysfunction group, n=599
Age, years	0.1186817	1.38	51 [37–65]	75 [69–85]
Cornell voltage product, mV*ms	0.0057002	0.31	119 [77–153]	158 [95–199]
Change in the VCG QRS azimuth from four eighths to five eighths of the duration of the spatial QRS loop, degrees	0.0041626	0.13	-47 [-66 – -24]]	-36 [-56–-15]
The net area subtended by the T wave in derived VCG lead X, mV/ms	-0.0081	-0.12	25.7 [15.5–34.4]	13.8 [2.0–27.1]
The mean difference between the VCG frontal plane T-wave elevation and the azimuth of all T-wave loop samples, degrees	0.0023645	0.07	39 [20–50]	67 [37–93]
Minimum VCG azimuth of the spatial QRS loop, sine radians	0.4874078	0.06	-0.92 [-1 – -0.92]	-0.92 [-1–-0.91]
Magnitude of the QRS in Eigenvector lead 4, microV	0.0002731	0.05	442 [297–523]	517 [338–625]
(Intercept)	-8.8852560	-	-	-

Table 4. Diagnostic Accuracies and Predictive Values of calculated A-ECG Score to identify grade 2+3 diastolic function

Performance Marker	Derivation cohort	Validation cohort and sensitivity analysis cohorts		
	Normal Grade 2-3 DD	Normal Grade 2-3 DD	Normal Grade 1-3 DD	Normal Grade 1-3 and indeterminate DD
AUC [95% confidence interval]	0.89 [0.86-0.92]	0.88 [0.87-0.90]	0.87 [0.85-0.88]	0.83 [0.81-0.84]
Sensitivity, %	73	67	67	67
Specificity, %	89	89	87	81
Positive predictive value, %	80	59	54	37
Negative predictive value, %	85	92	93	94
Positive likelihood ratio	6.85	6.37	5.34	3.53
Inverse negative likelihood ratio	3.27	2.73	2.67	2.48

Abbreviations: DD = diastolic dysfunction, AUC = Area under the receiver operating characteristics curve

Table 5. Comparison of hazard ratio by A-ECG diastolic dysfunction score quartile.

Full cohort, n=27239

A-ECG Diastolic Dysfunction Score, %	Hazard ratio [95%CI] vs score <25%	p value
25–50	2.5 [2.1–3.0]	< 0.001
50–75	4.0 [3.3–4.7]	< 0.001
≥75	5.7 [4.5–7.2]	< 0.001

No prior known cardiovascular disease, n=23051

A-ECG Diastolic Dysfunction Score, %	Hazard ratio [95%CI] vs score <25%	p value
25–50	2.6 [2.1–3.2]	< 0.001
50–75	4.0 [3.2–5.1]	< 0.001
≥75	5.5 [4.0–7.5]	< 0.001

Prior known cardiovascular disease, n=4188

A-ECG Diastolic Dysfunction Score, %	Hazard ratio [95%CI] vs score <25%	p value
25–50	1.8 [1.4–2.4]	< 0.001
50–75	2.4 [1.8–3.2]	< 0.001
≥75	3.3 [2.3–4.7]	< 0.001

Echocardiography measure	Patient A	Patient B
Diastolic dysfunction grade	*Normal*	*Grade II*
LAVI, ml/m^2	26	72
E/A	0.9	1.4
Average e', cm/s	9.3	7.5
E/e'	8	11
TR Vmax, m/s	2.1	3.3

ECG Measure	Patient A	Patient B
Cornell Voltage Product, mV*ms	89	197
Minimum VCG azimuth of the spatial QRS loop, sine radians	-0.954	-0.893
The mean difference between the VCG frontal plane T-wave elevation and the azimuth of all T-wave loop samples, degrees	15.8	64.3
The net area subtended by the T wave in derived VCG lead X, mV/ms	61.5	17.9
Change of the VCG QRS azimuth between five eighths to four eighths of the duration of the spatial QRS loop, degrees	-88	-24
Magnitude of the QRS in Eigenvector lead 4, microV	500	342
A-ECG Score for diastolic dysfunction	25%	71%

Figure 1. Comparison of A-ECG results from the 12-lead ECGs of two representative patients. Patient A has normal diastolic function with a corresponding low likelihood for diastolic dysfunction (25%) by A-ECG scoring. Patient B has diagnosed diastolic dysfunction with a high likelihood for diastolic dysfunction (71%) by A-ECG scoring. Vectorcardiographic (VCG) measures and waveform complexity measures were measured in VCG leads and eigenvector leads, respectively, that were derived from the standard 12-lead ECG.

Figure 2. Cumulative event curves per quartile of A-ECG diastolic dysfunction score (left). Comparison of restricted cubic splines for adjusted and unadjusted models (right).