Comparison of the analytical and clinical sensitivity of thirty-four rapid antigen tests with prevalent

SARS-CoV-2 variants of concern during the COVID-19 pandemic in the UK

Rachel L. Byrne[1*], Rachel S. Owen[1*], Ghaith Aljayyoussi[1], Caitlin Greenland-Bews[1], Konstantina

Kontogianni[1], Anushri Somasundaran[1], Dominic Wooding[1], Christopher T. Williams, LSTM Diagnostics

Group[1], Falcon Steering Group, Margaretha de Vos[2], Richard Body[3], Emily R. Adams[4,5], Camille

Escadafal[2], Thomas Edwards[1], Ana I. Cubas-Atienzar[1#]

[1]Centre for Drugs and Diagnostics, Liverpool School of Tropical Medicine, Liverpool, U.K.

[2]FIND, Foundation for Innovative New Diagnostics, Geneva, Switzerland

[3]Manchester University NHS Foundation Trust, UK

[4]Nuffield Department of Medicine, University of Oxford, Oxford UK

[5]Global Access Diagnostics, Bedfordshire, UK

#Address correspondence to Ana I. Cubas Atienzar: Ana.cubasatienzar@lstmed.ac.uk

*Rachel L. Byrne and Rachel S. Owen contributed equally to this work. Author order was determined

alphabetically.

Keywords: rapid diagnostic test (RDT); antigen detection; COVID-19; SARS-CoV-2 variant of concern

(VOC), B.1.617.2; Delta; B.1.1.529; Omicron; limit of detection (LOD)

Running title: Variant of concern COVID-19 RDT evaluation (max 50 characters)

Conflicts of Interest: Emily R Adams (ERA) is a Director of Epidemics and Neglected Tropical Diseases at

Global Access Diagnostics. Camille Escadafal (CE) and Margaretha de Vos (MDV) are employees of FIND.

ERA, CE and MDV had no role in data collection and analysis.

Word count: Abstract, 221; Importance, 133; Manuscript (Excluding Materials and Methods, Reference

and Figure legends), 2429

Authors' contributions: The study was conceived by AICA and CE. The study design was developed by

AICA, CE, TE and RLB. Data extraction was conducted KK, RLB, DW and CGB. Participant recruitment and

sample processing was performed by LSTMDG with oversight by FSG, RB and AICA. Data analysis and

interpretation were conducted by RSO, RLB, GA, AICA and TE. The initial manuscript was prepared by AICA,

RLB and RSO. All authors edited and approved the final manuscript. Funding acquisition was performed

by AICA, ERA, FSG, RB.

Abstract (250 words max)

Antigen-detection rapid diagnostic tests (Ag-RDTs) have become a central pillar for the management of

COVID-19 worldwide due to their speed and ease of use, and are now being developed for use in other

emerging outbreaks. Like other viruses, SARS-CoV-2 is subject to rapid mutation as it spreads, and new

Variants of Concern (VOCs) emerge frequently, posing a significant challenge for the reliability of Ag-RDTs

designed against the original strain to detect newer, highly mutated variants. It is therefore important

that the performance of Ag-RDTs is regularly evaluated, particularly in outbreak scenarios where rapid

diagnostics are key to limiting disease spread. Here, we present a comprehensive evaluation of the

analytical and clinical sensitivity of 34 commercially available Ag-RDTs with five VOCs, all of which were

highly prevalent in the UK at various times between 2019 and 2023. This evaluation highlights the

importance of regular evaluation of Ag-RDT performance, with several Ag-RDTs demonstrating reduced

performance with some VOCs. This study also highlights the challenges that arise when different

organisations lack a consensus approval threshold, causing confusion over Ag-RDT performance and

making it difficult to properly evaluate individual tests. We conclude that regular performance evaluation

through our proposed pipeline combined with a broad consensus approval threshold across global

organisations is essential to maintaining the effectiveness of Ag-RDTs as a disease management tool

during outbreaks.

Importance (150 words max)

Antigen-detection rapid diagnostic tests (Ag-RDTs) came to global prominence during the COVID-19

pandemic, where they offered a quick and simple at home diagnostic which could be used to manage

disease spread. A major ongoing challenge for the broad use of Ag-RDTs is the speed at which new SARS-

CoV-2 variants emerge, each of which has the potential to reduce the performance of the available Ag-

RDTs. As Ag-RDTs are explored for use in other viral disease outbreaks, pipelines for the regular evaluation

of test performance are essential for ensuring that Ag-RDTs can be employed effectively. Here, we have

developed a robust pipeline for the large-scale evaluation of commercially available Ag-RDTs against

several major SARS-CoV-2 variants, which can be adapted and applied to other emerging outbreaks to

ensure test performance is maintained as a virus evolves.

Introduction

Antigen-detection rapid diagnostic tests (Ag-RDTs) offer quick, inexpensive, laboratory-independent

diagnostics that can be performed at home by lay individuals (1,2). As such they have been a central

pillar of control strategies for SARS-CoV-2 and are now the first line diagnostic in many countries (3).

Whilst the development and deployment of COVID 19 Ag-RDTs was rapid, the frequent emergence and

spread of new variants of concern (VOCs) combined with the global nature of the pandemic response

posed a continuous challenge to disease monitoring. The majority of Ag-RDTs were developed and

evaluated early in the COVID-19 pandemic, utilising wild type (WT) SARS-CoV-2's nucleocapsid (N)

protein as a target, due to its high abundance within the virion (4). However, further VOCs rapidly

emerged, including Alpha, Beta, Gamma and Delta, and by the time of the emergence of the first

Omicron VOC in late 2021 the viral genome had accumulated a significant number of mutations (5,6).

These mutations included several within the N protein, where the Omicron lineage has 3 unique

mutations, making it difficult to predict the performance of Ag-RDTs with emerging VOCs (7).

Initial data on the performance of Ag-RDTs for different COVID-19 VOCs remain contradictory in both

analytical and clinical evaluations. Whilst early reports evaluating a small number of brands found

comparable sensitivities between WT, Delta and Omicron (B.1.1.529) VOCs (8–11), later studies

demonstrated that Ag-RDTs showed loss of sensitivity to Omicron when compared to other VOCs (12),

and clinical evaluations of Ag-RDTs against emerging SARS-CoV-2 lineages have reported inconsistent

results (12–14).

In order for RDTs to be used to their full potential in managing large scale disease outbreaks, there must

be confidence that these tests maintain good clinical specificity even as a pathogen evolves, and regular

clinical and analytical evaluation is essential to ensuring that RDTs are performing to the required

standards. Given the comparatively recent emergence of large scale RDT testing in the context of

infectious disease, relatively few evaluations of these devices on a large scale over time have been

conducted. Here, we present a robust pipeline for the evaluation of commercially available RDTs against

an evolving pathogen, SARS-CoV-2, using the most prevalent VOCs circulating in the United Kingdom

between 2019-2023.

Results

Analytical sensitivity using cultured SARS-CoV-2 virus

For Omicron sub lineage BA.5, all of the 31 Ag-RDTs evaluated had an LOD $\leq 5.0 \times 10^2$ pfu/mL, fulfilling

the criteria set by the British Department of Health and Social Care (DHSC) (Fig. 1) and all except two

brands (RespiStrip and GeneFinder) had an LOD of $\leq 1.0 \times 10^6$ RNA copies/mL thus fulfilling the WHO

Target Product Profile (TPP) recommendations for SARS-CoV-2 Ag-RDTs (15). Whereas for Omicron BA.1,

only 23 of the 34 Ag-RDTs evaluated had an analytical LOD of $\leq 5.0 \times 10^2$ pfu/mL and 32 out of 34

(including Biocredit, Core, Covios, Hotgen, Innova, LumiraDx, PerkinElmer and SureStatus, that all fell

below the DHSC recommendations) had an LOD of $\leq 1.0 \times 10^6$ RNA copies/mL. The more sensitive Ag-

RDTs for detecting Omicron BA.1 were AllTest, Bioperfectus, Flowflex, Fortress, Joysbio, Nadal, Onsite,

RightSign, Roche, StrongStep, Standard Q, Tingsun and Wondfo (Fig. 1) with an LOD $\leq 2.5 \times 10^2$ pfu/mL

and 4.4×10^4 RNA copies/mL. For both Omicron BA.1 and BA.5 the Ag-RDT brand with the lowest

sensitivity was RespiStrip with an LOD of 5.0×10^4 pfu/mL and 9.2×10^6 RNA copies/mL (BA.1) and LOD

of 1.0×10^2 pfu/mL and 3.5×10^6 RNA copies/mL (BA.5) respectively.

For the Delta VOC, 33 out of the 34 Ag-RDTs evaluated had an LOD of $\leq 5.0 \times 10^2$ pfu/mL and 31 of the

34 reported an LOD $\leq 1.0 \times 10^6$ RNA copies/mL (Fig. 1), with Genefinder falling to meet either the DHSC

or WHO recommended LOD.

For the Alpha VOC, 7 of the 34 Ag-RDTs evaluated (Biocredit, Genedia, Hotgen, Onsite, RespiStrip, Tigsun

and Wondfo) had an analytical LOD of $> 5.0 \times 10^2$ pfu/mL thus failing to meet the DHSC minimum

requirements, with Biocredit and Genedia also falling below the WHO requirement.

For the Gamma VOC, 5 Ag-RDT brands (Flowflex, Hotgen, Innova, Onsite and RespiStrip) failed to meet

either the DHSC or WHO requirements with a further 7 Ag-RDT brands falling below the DHSC

recommendations. The Ag-RDTs with the greatest sensitivity for a sample positive for Gamma VOC

where AllTest, Core Test, InTec, Standard-F, Standard-Q, StrongStep and Surestatus (Fig. 1).

For the WT, the target for which all Ag-RDT brands were originally developed, only 19 and 22 met the

DHSC and WHO requirements, respectively.

When comparing only the pfu/mL, we found that tests had significantly higher LODs with Omicron BA.1

compared to Delta (p = 0.000) and significantly lower LODs with Omicron BA.5 compared to all other

VOCs tested (p = 0.001). When comparing RNA copies/mL, the Ag-RDTs detected more sensitively Alpha

(p = 0.000) than the other VOCs (p = 0.000).

FIG 1. Heatmap comparing the LODs of 34 Ag-RDT using the Ancestral (WT), Alpha (B.1.1.7), Gamma (P.1), Delta
(B.1.617.2) Omicron (BA.1) and Omicron (BA.5) variants on pfu/mL (right) and RNA copies/mL (left). Data of the
Ancestral, Alpha and Gamma have been taken from our previously published work (16). Blue colours indicate LODs
fulfilling the DHSC (for pfu/mL) and WHO criteria (for RNA copies/mL). * = p ≤ 0.05 between VOCs.

Retrospective samples: SARS-CoV-2 Ag-RDT clinical sensitivity

The clinical sensitivity of five Ag-RDTs brands (Covios, Flowflex, Hotgen, Onsite and SureStatus) was

evaluated utilising SARS-CoV-2 Alpha (n = 30), Delta (n = 56) and Omicron (n = 49) positive VTM swabs

stored at -80 °C, as previously described. Statistically higher viral loads, determined by RT-qPCR, were

recorded among individuals positive for Alpha and Omicron infection compared to Delta (p = 0.001 and

p = 0.009 respectively) measured by Kruskal–Wallis (Fig. 2).

FIG 2: Boxplot of the SARS-CoV-2 viral load (RNA copies/mL) distribution of RT-qPCR NP swabs collected from
participants recruited to FALCON between January 2021 and March 2022. The whiskers show the 95 %
confidence intervals and the horizontal line the median. Asterisks indicate statistical significance between different
VOCs (ns = non-significant, * = p ≤ 0.05, ** = p ≤ 0.01).

We determined the 50 % and 95 % LODs with Alpha, Delta, and Omicron SARS-CoV-2 positive swab

samples for five Ag-RDT brands based on a logistic regression model (Table 1, Fig. 3). Overall, the lowest

LOD for the Alpha VOC was recorded with Flowflex Ag-RDT (50 % LOD 1.58 x 10^4 RNA copies/mL and 95

% LOD 2.14 x 10^4 RNA copies/mL), for Delta variant with Onsite Ag-RDT (50 % LOD 3.31 x 10^1 RNA

copies/mL and 95 % LOD 3.80 x 10^4 RNA copies/mL) and for Omicron with SureStatus Ag-RDT (50 % LOD

1.78 x 10^3 RNA copies/mL and 95 % LOD 7.41 x 10^4 RNA copies/mL), which were all statistically similar to

the true analytical LOD reported. The Delta VOC exhibited the greatest variability between predicted

LODs from different Ag-RDTs (Table 1) whereas for the true analytical LOD the Alpha VOC showed the

greatest variability (Fig. 1).

The sensitivity of the Ag-RDTs varied from 70 % to 93.3 % (Hotgen and Flowflex) for the Alpha VOC, as

shown in Table 2. For Delta VOC, the sensitivity ranged from 77.6 % to 94.5 % (Covios and Onsite) for four

of the five Ag-RDTs, with Hotgen reporting a significantly lower sensitivity of 44.2 % than Omicron (p <

0.001) and Alpha (p = 0.026). The observed sensitivities for Omicron were consistent across all brands of

Ag-RDTs from 84.4 % and 97.9 % (Hotgen and Onsite). For samples with low Ct values (< 25), the

sensitivities were statistically similar for the Covios, Flowflex, Onsite and SureStatus Ag-RDTs but not for

the Hotgen Ag-RDT. For the Delta VOC, the sensitivity of the Hotgen Ag-RDT was significantly lower

compared to Omicron (p < 0.001) and Alpha (p = 0.011). Whereas samples with high Ct values (Ct > 25)

resulted in reduced test sensitivities across all variants, with the greatest sensitivity reported in samples

positive with the Omicron VOC. A three-way factorial ANOVA assessing the effects of RDT result (value:

positive vs. negative), VOC, and RDT brand, along with their interactions, revealed a significant main effect

of RDT result ($F_{(1, 593)}$ = 98.21, p < 0.001, $\eta^2 p$ = 0.14), indicating a large effect size (Fig. 4). The effect of

VOC was also significant ($F_{(2, 593)}$ = 27.83, p < 0.001, $\eta^2 p$ = 0.09), while RDT brand had a small and non-

significant effect ($F_{(4, 593)}$ = 1.58, p = 0.18, $\eta^2 p$ = 0.01). The interaction between RDT result and VOC was

significant ($F_{(2, 593)}$ = 14.20, p < 0.001, $\eta^2 p$ = 0.05), suggesting that the effect of Ct value on RDT result

differs across VOCs. Other interactions, including the three-way interaction, showed negligible effect sizes

($\eta^2 p$ < 0.01) and were not significant.

TABLE 1: The 50 % and 95 % limit of detection (LOD) (RNA copies/mL) for five Ag-RDT brands (Covios, Flowflex,
Hotgen, Onsite and SureStatus) from 122 clinical NP samples positive for Alpha, Delta and Omicron SARS-CoV-2
VOCs.

	LOD	Ag-RDT brand (RNA copies/mL)				
		Covios	**Flowflex**	**Hotgen**	**Onsite**	**SureStatus**
Alpha	50 %	4.79E+03	1.58E+04	1.82E+05	7.76E+03	4.68E+04
	95 %	2.24E+06	2.14E+04	3.55E+08	1.70E+06	1.62E+07
Delta	50 %	3.47E+03	4.37E+00	1.26E+06	3.31E+01	3.02E+04
	95 %	2.19E+08	1.00E+06	1.00E+12	3.80E+04	2.34E+06
Omicron	50 %	1.02E+04	4.47E+03	3.55E+04	0	1.78E+03
	95 %	8.71E+05	2.45E+04	1.35E+06	0	7.41E+04

165
FIG 3: Limit of detection analyses of upper-respiratory samples positive by RT-qPCR for five SARS-CoV-2 Ag-RDT
tests (Covios, Hotgen, Onsite, Flowflex and SureStatus) using NP swabs. The log10 RNA copies on the x axis were
plotted against a positive (1.0) or negative (0.0) Ag-RDT result on the y axis. Fill curves show logistic regressions of
the viral load on the Ag-RDT result; vertical dashed lines indicate log10 RNA copies subjected to the test at which
50 % and 95 % LOD of the samples are expected positive based on the regression results. No significant differences
were observed.

TABLE 2: The comparison of the sensitivity of five antigen rapid diagnostic tests (Ag-RDTs) between the Alpha, Delta and Omicron variants. Calculated
sensitivities, sample size (in italics) and 95 % confidence intervals (CI95 %) are given.

Population	Variant	Covios	Flowflex	Onsite	Hotgen	Surestatus
	Alpha	90%, *30*	93.3%, *30*	90%, *30*	70%, *30*	86.7%, *15*
	(CI95%)	73.5-97.9%	77.9-99.2%	73.5-97.9%	50.6-85.3%	59.5-98.3%
	Delta	77.6%, *49*	92.9%, *56*	94.5%, *55*	44.2%, *52*	81.6%, *38*
	(CI95%)	63.4-88.2%	82.7-98%	84.9-98.9%	30.5-58.7%	65.7-92.3%
Total cohort	Omicron	89.8%, *49*	95.8%, *48*	97.9%, *47*	84.4%, *45*	95.9%, *49*
	(CI95%)	77.8-96.6%	85.7-99.5%	88.7-99.9%	70.5-93.5%	86-99.5%
	All	85.2%, *128*	94%, *134*	94.7%, *132*	64.6%, *127*	89.2%, *102*
	(CI95%)	77.8-90.8%	88.6-97.4%	89.4-97.8%	55.6-72.8%	81.5-94.5%
	Alpha	95.8%, *24*	100%, *24*	95.8%, *24*	83.3%, *24*	92.3%, *13*
	(CI95%)	78.9-99.9%	85.8-100%	78.9-99.9%	62.6-95.3%	64-99.8%
	Delta	81.8%, *33*	94.6%, *37*	100%, *36*	45.5%, *33*	88%, *25*
	(CI95%)	64.5-93%	81.8-99.3%	90.3-100%	28.1-63.6%	68.8-97.5%
≤ Ct 25	Omicron	97.6%, *42*	100%, *41*	100%, *40*	97.4%, *38*	100%, *42*
	(CI95%)	87.4-99.9%	91.4-100%	91.2-100%	86.2-99.9%	91.6-100%
	All	91.9%, *99*	98%, *102*	99%, *100*	75.8%, *95*	95%, *80*
	(CI95%)	84.7-96.4%	93.1-99.8%	94.6-100%	65.9-84%	87.7-98.6%
	Alpha	66.7%, *6*	66.7%, *6*	66.7%, *6*	16.7%, *6*	50%, *2*
	(CI95%)	22.3-95.7%	22.3-95.7%	22.3-95.7%	0.4-64.1%	1.3-98.7%
	Delta	68.8%, *16*	89.5%, *19*	84.2%, *19*	42.1%, *19*	69.2%, *13*
	(CI95%)	41.3-89%	66.9-98.7%	60.4-96.6%	20.3-66.5%	38.6-90.9%
> Ct 25	Omicron	42.9%, *7*	71.4%, *7*	85.7%, *7*	14.3%, *7*	71.4%, *7*
	(CI95%)	9.9-81.6%	29-96.3%	42.1-99.6%	0.4-57.9%	29-96.3%
	All	62.1%, *29*	81.2%, *32*	81.2%, *32*	31.2%, *32*	68.2%, *22*
	(CI95%)	42.3-79.3%	63.6-92.8%	63.6-92.8%	16.1-50%	45.1-86.1%

Ct = cycle threshold

176 **FIG 4: Positive and negative Ag-RDT results according to RT-qPCR cycle threshold (Ct) values.**
177 P values indicate the results of Bonferroni-adjusted post-hoc tests performed following three-way factorial ANOVA
178 analysis to compare RDT+ and RDT− within each VOC and RDT brand. Asterisks indicate comparisons where
179 Cohen's d ≥ 0.5 and p < 0.05, indicating a significant difference between the Genomic Ct values of RT-PCR+/RDT+
180 samples and RT-PCR+/RDT- samples.
181

Prospective samples: SARS-CoV-2 Ag-RDT clinical sensitivity

During the prospective evaluation, 122 participants tested positive by RT-qPCR for SARS-CoV-2. Of them,

32 were the Delta VOC and 90 the Omicron (BA.1) VOC. Samples were immediately tested using the

Biocredit (RapiGEN, South Korea) Ag-RDT with 99 yielding a positive Biocredit Ag-RDT result (Sens 81.1

%, CI95 % 73.1-87.6 %). When separated by VOC the sensitivity was lower, but statistically similar, for

the Delta VOC 71.9 % (CI 95 % 53.3-86.6 %) compared to the Omicron VOC 84.4 % (CI95 % 75.3-91.2 %).

The difference in Ct value between positive and negative samples by Biocredit was significant (Delta p =

0.038, Omicron p = 0.00007) across both VOCs. Additionally, the intensity of the test band of positive

Ag-RDTs was recorded for 119 samples (three excluded due to line intensity not recorded) with no

significance obtained between the two VOCs (Delta and Omicron BA.1).

Discussion

We present a comprehensive assessment of the analytical sensitivity of 34 commercially available

COVID-19 Ag-RDTs for the detection of the Omicron (BA.1 and BA.5), Alpha, Gamma and Delta VOCs and

WT, the major variants circulating in the UK during the initial five years following the emergence of the

pandemic(5). The rapid emergence of new VOCs throughout the course of the Covid-19 pandemic posed

a significant challenge for reliable rapid diagnostic testing(13). New VOCs often contained mutations in

the N gene, increasing the likelihood that RDT analytical sensitivity could be compromised. Indeed,

studies have identified mutations in the N gene which affected the sensitivity of Ag-RDTs, including the

T135I mutation in the Alpha VOC (17), and the A376T and M241I mutations (18), thus it is important to

maintain regular assessment of Ag-RDT performance over time.

Analytically, the majority of Ag-RDTs successfully met the WHO criteria outlined in the TPP for SARS-

COV-2 Ag-RDTs when detecting the Omicron viral lineages (94.1 %, BA.1 and 93.5 %, BA.5)(15). When

analysing RNA copies/mL, as outlined in the WHO recommendations, the LOD of Omicron VOC BA.5 was

statistically lower than the WT and the Alpha VOC but similar to the Gamma, Delta and Omicron BA.1

VOCs. However, when comparing the PFU/mL, recommended by the UK DHSC, tests have significantly

lower LODs with Omicron BA.5 than all other VOCs and WT.

The observed discrepancies between RNA copies/mL and PFU/mL are likely attributable to difference in

the virus' ability to form plaques and varying ratios of infectious particles to RNA copies present (19).

Whilst the PFU/mL measures infectious virus present, RNA copies/mL encompasses all SARS-CoV-2 RNA

present including non-infectious or dead virus. For Ag-RDT evaluation, RNA copies/mL is recommended

by the WHO and the UK DHSC recommends PFU/mL. It is counterintuitive for organisations to utilise

different units of measurements for comparison of Ag-RDTS and a consensus would allow more

appropriate evaluation, something that is essential for effectively managing responses to new disease

outbreaks.

For 5 of the 34 Ag-RDT brands we present clinical accuracy data utilising clinical respiratory samples

positive for SARS-CoV-2 Alpha, Delta and Omicron (BA.1) VOCs. Hotgen consistently performed poorer

than other Ag-RDTs, typically falling below the WHO TPP guidance with Onsite the best performing

brand overall. For all brands, the clinical sensitivity values obtained with retrospective samples were the

highest for Omicron samples. This is in line with other published works comparing Omicron BA.1 to

other non-Omicron VOCs (20). A similar trend was observed when using only prospective samples

collected when the UK was experiencing the Delta and Omicron waves of infection. A lower sensitivity

was recorded for samples positive for the Delta VOC compared to those positive with an Omicron VOC

infection on the Biocredit Ag-RDT, as shown in cultured samples by Stanley *et al.,* 2022 (21). These

differences in sensitivity of Ag-RDTs between VOCs, which are independent of the temporal emergence

of each variant, highlight the importance of continual and thorough assessment of RDT performance as

new lineages of a pathogen emerge.

For all brands in the retrospective cohort the predicted clinical LODs were statistically similar to the

analytical LODs obtained with spiked laboratory samples. Clinical evaluation is expensive, and often

surplus samples are not available for comprehensive testing, hindering large scale assessment of Ag-RDT

performance. The data presented here demonstrates the efficacy of analytical samples, indicating that

analytical evaluation is a robust alternative for evaluating Ag-RDT performance in the absence of

abundant clinical samples.

This study has several strengths; we have carried out an extensive evaluation of the analytical sensitivity

of 34 commercially available Ag-RDT brands. This list is inclusive of most WHO-EUL recommended tests

and five awaiting approvals at the time of the study thus of high global high public health relevance.

Additionally, we included both viral isolates and clinical specimens to evaluate the Ag-RDT sensitivity.

The clinical specimens used in this study are attributed to three different lineages; Alpha, Delta and

Omicron, which were all at some point a dominant VOC, providing a comprehensive evaluation of assay

performance across a significant temporal and mutational timeframe.

Limitations of this study include the use of retrospective frozen specimens instead of fresh swabs as

recommended by most Ag-RDT manufacturers. However, prospective clinical evaluation studies rarely

include multiple VOCs as their prevalence depends on their time period, and prospective evaluation of

multiple RDT brands simultaneously is complicated by the need for a single swab per test. To correct for

the potential degradation of RNA after a freeze-thaw cycle, viral RNA was re-tested by RT-PCR at the

time of Ag-RDT evaluation and these values were used for comparison. Another limitation of this study

is that we did not repeat the LOD experiments for the WT, Alpha and Gamma viral isolates and used the

data from previous studies performed in our laboratory with the same protocol and compared it to the

newly produced data for the Delta and Omicron VOC isolates.

To conclude, we present a robust pipeline for the large scale evaluation of commercially available Ag-

RDTs against multiple SARS-CoV-2 VOCs, which can be easily adapted to assess further variants or for

different disease outbreaks. We report similar if not superior LODs for Omicron compared to other non-

Omicron VOCs across a wide range of commercially available RDTs. However, we also report decreased

detection of the Delta VOC in both analytical and clinical samples, highlighting the need for continuous

assessment of Ag-RDTs especially those recommended for at home testing. We additionally report

inconsistencies between product fulfilment criteria for WHO and UK DHSC, where several tests show

significantly different performance depending on the guidelines, highlighting the importance of

standardised evaluation criteria for Ag-RDTs, particularly during a global response.

Materials and Methods

Evaluated Ag-RDTs

Thirty-four Ag-RDT brands were evaluated in this study; all were lateral flow assays (LFA), 31 using

colorimetric gold nanoparticle detection, two fluorescence and one based on microfluidic

immunofluorescence technology (Table 3). The selection of the Ag-RDT resulted from an expression of

interest launched by FIND (www.finddx.org) and a scoring process based on defined criteria. This list

includes eight Ag-RDTs on the WHO Emergency Use Listing (WHO-EUL) and six tests that are on the

waiting list for WHO-EUL approval (22). Analytical testing was performed on all Ag-RDT brands (Table 3)

and a small subset of these were further used for the clinical evaluation on retrospective samples based

on brands that showed the best results on clinical diagnostic evaluations under the FIND programme

(Covios, Hotgen, Onsite, SureStatus) and widely used in the UK for mass testing (Flowflex). Results on

prospectively collected samples are only provided with Biocredit.

TABLE 3: An overview of the RDT brands used in the study for analytical performance.

RDT brand	Test/Company/Country	Target Ag	Principle	Approval
ActiveXpress	ActivXpress+ COVID-19 Ag Complete Kit/Edinburgh Genetics Ltd./UK	N	G	CE
AllTest	SARS-CoV-2 Antigen Rapid Test/Hangzhou AllTest Biotech Ltd./China	N	G	CE
Biocredit	Biocredit COVID-19 Ag/Rapidgen Inc./Rep. Korea	N	G	CE/ WHO EUL
Bioperfectus	SARS-CoV-2 Ag Rapid Test/ Jiangsu Bioperfectus Tech. Ltd./China	N	G	CE
Core	COVID-19 Ag Test/Core Technology Ltd./China	N	G	CE
Covi-Go	Covigo/Mologic Ltd./UK	N	G	CE/ UA
Covios	Covios COVID-19 Ag Test device/ Mologic Ltd./UK	N	G	CE
Espline	ESPLINE SARS-CoV-2/Fujirebio Diagnostics Inc./Japan	N	G	CE
Excalibur	Rapid SARS-CoV-2 Antigen test card/ Excalibur Healthcare Services/UK	N	G	CE
Flowflex	Flowflex SARS-CoV-2 Ag Rapid Test/Acon Biotech, Ltd./China	N	G	CE/ WHO EUL
Fortress	CoV Ag Nasal Swab Rapid test/Zhejian Orient Gene Biotech/ China	N	G	CE
Genedia	GENEDIA W COVID-19 Ag/ Green Cross Medical Sciences/Rep. Korea	N	G	CE
GeneFinder	COVID-19 rapid test GeneFinder/OSANG Health Care/ South Korea	N	G	CE/UA
Hotgen	2019-nCoV Antigen test/ Beijin Hotgen Biotech Ltd./China	N	G	CE
iChroma	iChroma COVID-19 Ag Test/ Boditech Medical Inc./Rep. Korea	N	F	CE
Innova	Innova SARS-CoV-2 Antigen Rapid/ Innova Medical Group Ltd./UK	N	G	CE/UA
Intec	Rapid SARS-CoV-2 Antigen test/Intec Products Inc./China	N	G	CE
Joysbio	SARS-CoV-2 Antigen Rapid Test Kit/ Joysbio Biotechnology Ltd./China	N	G	CE/UA
LumiraDx	LumiraDx SARS-CoV-2 antigen test/ Lumira Dx Ltd./US	N	M	CE/ WHO EUL
Nadal	Nadal COVID-19 Ag Test/Nal von minden GmbH/Germany	N	G	CE
NowCheck	NowCheck COVID-19 Ag test/ Bionote Inc./ Rep. Korea	N	G	CE
Onsite	Onsite COVID-19 Ag Rapid Test/CTKBiotech Inc./USA	N	G	CE/ WHO EUL
PanBio	Panbio COVID-19 Ag Rapid Test/Abbott Rapid Diagnostics/Rep. Korea	N	G	CE/WHO EUL
PerkinElmer	PerkinElmer COVID-19 Antigen Test/PerkinElmer/ Switzerland	N	G	CE/UA
RespiStrip	Respi-Strip COVID-19 Ag/Coris Bioconcept/Belgium	N	G	CE
RighSign	COVID-19 Ag Rapid Test Cassette/Hangzhou Biotets Biotech Ltd./China	N	G	CE
Roche	SARS-CoV-2 Rapid Ag Test/ Roche Diagnostics/Switzerland	N	G	CE
Standard-F	Standard F COVID-19 Ag FIA., SD Biosensor Inc./Rep. Korea	N	F	CE
Standard-Q	Standard Q COVID-19, SD Biosensor Inc./Rep. Korea	N	G	CE/WHO EUL
StrongStep	StrongStep SARS-CoV-2 Ag Rapid Test/Nanjing Liming Bio-Products/US	N	G	CE
SureStatus	Sure-Status COVID-19 Antigen Card Test, Premier Medical Corp./ India	N	G	WHO EUL
Tigsun	Tingsun COVID-19 Ag Rapid test/ Beijin Tigsun Diagnostics Ltd./China	N	G	CE
Wantai	Rapid SARS-CoV-2 Antigen test/ Wantai Biological Pharmacy Ltd./China	N	G	CE
Wondfo	Wondfo 2019-nCoV Antigen Test/ Guangzhou Wondfo Biotech/China	N	G	CE/WHO EUL

272   RDT: Rapid diagnostic test. Target N: nucleoprotein; Principle G: LFA using gold (colourimetric detection), F: LFA
273   fluoresce detection, M: microfluidic fluorescent technology. CE: CE making as per European Conformity, WHO EUL:
274   WHO Emergency Use Listing, UA: under assessment for WHO Emergency Use Listing.
275

SARS-CoV-2 viral culture and Ag-RDT limit of detection

The SARS-CoV-2 isolates were grown in Vero E6 cells (C1008; African green monkey kidney cells) and

maintained in culture media (Dulbecco's modified eagle membrane (DMEM) with 2% fetal bovine serum

and 0.05 mg/mL gentamycin) as previously described (23,24). The isolates for Alpha (Genbank accession

number: MW980115), Delta (SARS-CoV-2/human/GBR/Liv_273/2021), Gamma (hCoV-19/Japan/TY7-

503/2021), Omicron (BA.1) (SARS-CoV-2/human/GBR/Liv_1326/2021) and Omicron (BA.5)(SARS-CoV-

2/South Africa/CERI-KRISP-K040013/2022) were used evaluate the analytical limit of detection (LOD) of

the 34 Ag-RDTs using live virus.

Plaque forming units per millilitre (pfu/mL) of the viral stocks were counted using viral plaque assay as

previously described (16) and ten-fold serial dilutions of the viral stock were made starting from 1.0×10^6

pfu/mL using DMEM as a diluent. Two-fold dilutions were made below the ten-fold LOD dilution to

determine the LOD. The LOD was defined as the lowest dilution at which all three replicates were positive

by Ag-RDT. The LODs for WT, Alpha and Gamma VOCs, obtained as part of our previous work utilising the

same protocol (16,24), were used here for practicality to compare to the Delta and Omicron lineages

LODs.

Retrospective and prospective clinical samples

Clinical samples were collected as part of the 'Facilitating Accelerated Clinical Evaluation of Novel

Diagnostic Tests for COVID-19' (FALCON) study (25). Ethical approval was obtained from the National

Research Ethics Service and the Health Research Authority (IRAS ID:28422, clinical trial ID: NCT04408170).

Nasopharyngeal (NP) swabs in vital transport media (VTM) were collected from consenting symptomatic

adults attending the community drive-through COVID-19 test center located in Liverpool John Lennon

Airport, UK between January 2021 and March 2022. Clinical specimens were transported to the Liverpool

School of Tropical Medicine (LSTM) biosafety level 3 (BSL3) laboratories in insulated UN7737 transport

bags and aliquoted and stored at -80°C until further testing.

Prospective clinical samples were collected as part of a diagnostic evaluation of the Biocredit Ag-RDT

(Table 3) (26). Participants were enrolled from December 2021 and March 2022 coinciding with the

emergence of Omicron. NP swabs in VTM were collected for RT-qPCR followed by another NP swab in the

alternate nostril for the BioCredit Ag-RDT evaluation. Specimens were transported to the LSTM BSL3

laboratories as described above and, in this case, processed immediately for Ag-RDT testing.

Samples were confirmed SARS-CoV-2 RNA-positive using the TaqPath™ COVID19 CE-IVD RT-PCR Kit

(Thermo Fisher Scientific, USA). Based upon epidemiological data in the UK at the time of enrolment and

S gene amplification in the PCR assay, consecutive SARS-CoV-2 positive RT-PCR samples were selected as

presumed Alpha if collected between January and March 2021, presumed Delta if collected between June

and August 2021, and presumed Omicron if collected between December 2021 and March 2022. The

variant type was later confirmed by whole genome sequencing. A panel of ten SARS-CoV-2 RNA-negative

VTM samples were also included as negative controls.

Ag-RDTs testing protocol

All Ag-RDTs were performed as specified by the test specific instructions for use (IFU). For the

determination of the LOD using live virus, a specific volume of the serial dilutions was added directly to

the extraction buffers at a 1:10 ratio as previously described (23). For the clinical samples, VTM was

mixed by pipetting at 1:1 ratio with the extraction buffer of the Ag-RDTs. The dilution factor introduced

by the swabs diluted in buffer was accounted for when calculating the viral copy numbers of the tested

swab samples. A negative control of only VTM was incorporated to account for any non-specific reaction

as previously reported for some Ag-RDT brands when using VTM (23). Results were read by two

operators, blinded to each other and if a discrepant result occurred, a third operator acted as a

tiebreaker. The visual read out of the Ag-RDT test band was scored on a quantitative scale from 1 (weak

positive) - 10 (strong positive). Ag-RDT results were classified as invalid when the control line was

absent.

Quantification of viral loads

For quantification of the RNA copy numbers per mL (RNA copies/mL) viral RNA was extracted using QIAmp

Viral RNA mini kit (Qiagen, Germany) according to the manufacturer's instructions. The RNA copies/mL

were established using the COVID-19 Genesig RT-qPCR kit (PrimerDesign, UK). RT-qPCR testing was carried

out using the Rotor-Gene Q (Qiagen, Germany), with a ten-fold serial dilution of quantified in vitro-

transcribed RNA incorporated for each PCR run (27). A total of five replicates were tested for each

standard curve point, and extracted RNA from each culture dilution was tested in triplicate. The RNA

copies/mL for samples was then calculated from the mean Ct value of these replicates.

Whole genome sequencing

Clinical samples underwent whole genome sequencing to confirm the SARS-CoV-2 variant. Sequencing

was performed using the ARTIC V3 (LoCost) (28)sequencing protocol on the MinION R.9.4.1 flow cell

(Oxford Nanopore Technology, UK). RT-PCR was performed with a two- step PCR initially the Arctic RT

PCR 5X LunaScript® RT SuperMix (New England Biolabs, USA) with 8 µL of RNA sample, and a thermal

cycling profile of 2 minutes at 25 °C followed by 10 minutes at 55 °C and then one minute at 95 °C. This

was then followed by the Q5® Hot Start High-Fidelity 2X Master Mix (New England Biolabs, USA) using 10

µM of the ARTIC V4.1 primer pools (Integrated DNA Technologies, USA), and a thermal cycling profile of

30 second at 98 °C for heat inactivation, followed by 25 cycles of a 15 second denaturation at 98 °C and

a five-minute annealing/extension at 65 °C. Library preparation was carried out using the Ligation

Sequencing Kit (SQK-LSK109) and Native Barcoding Expansion Kits (EXP-NBD104 and EXP-NBD114)

(Oxford Nanopore Technologies). Basecalling was carried out via MinKnow (v4.2.8), with demultiplexing

and read filtering using Guppy (v5.0.7.). The ARTIC pipeline was then used to assemble a consensus

genome, BAM files, and variant calling file with *--normalise 200 --threads 4.* Variant calling was carried

out using EPI2ME Desktop Agent v3.3.0 with the ARTIC+NextStrain analysis pipeline.

Statistical analysis

Statistical analyses were performed using SPSS V.28.0, Epi Info V3.01 and R scripts. Binomial confidence

intervals for sensitivities and specificities were computed using the Wilson score interval. Differences in

the analytical LODs of VOCs were compared using Kruskal Wallis with Bonferroni correction for multiple

tests. To further analyse analytical sensitivities, we used logistic regression, with RNA copy number as the

independent and test outcomes as the dependent variable, yielding detection probabilities for each viral

load level. A three-way factorial ANOVA was performed on log-transformed Genomic Ct values to assess

the effects of RDT result (value: positive vs. negative), VOC, and RDT brand, along with their interactions.

Bonferroni-adjusted post-hoc tests were performed to compare RDT+ and RDT– within each VOC and RDT

brand.

References

1.   Peeling RW, Olliaro PL, Boeras DI, Fongwen N. Scaling up COVID-19 rapid antigen tests: promises
and challenges. Lancet Infect Dis [Internet]. 2021 Sep 1;21(9):e290–5. Available from:
https://doi.org/10.1016/S1473-3099(21)00048-7
2.   Cassuto NG, Gravier A, Colin M, Theillay A, Pires-Roteira D, Pallay S, et al. Evaluation of a SARS-
CoV-2 antigen-detecting rapid diagnostic test as a self-test: Diagnostic performance and usability.
J Med Virol [Internet]. 2021 Dec 1;93(12):6686–92. Available from:
https://doi.org/10.1002/jmv.27249
3.   World Health Organisation (WHO). Antigen-detection in the diagnosis of SARS-CoV-2 infection.
2021 Oct.
4.   FIND. https://www.finddx.org/tools-and-resources/dxconnect/test-directory/. DxConnect
Directory.
5.   Wiegand T, Nemudryi A, Nemudraia A, McVey A, Little A, Taylor DN, et al. The Rise and Fall of
SARS-CoV-2 Variants and Ongoing Diversification of Omicron. Viruses. 2022 Sep 10;14(9).
6.   Quaglia F, Salladini E, Carraro M, Minervini G, Tosatto SCE, Le Mercier P. SARS-CoV-2 variants
preferentially emerge at intrinsically disordered protein sites helping immune evasion. FEBS J
[Internet]. 2022 Jul 1;289(14):4240–50. Available from: https://doi.org/10.1111/febs.16379
7.   FDA. SARS-CoV-2 Viral Mutations: Impact on COVID-19 Tests. 2021.

8.   Goderski G, Han W, Stanoeva K, Meijer A. Technical evaluation of SARS-CoV-2 antigen self-tests
with Omicron variant: Evaluation Report. 2021.
9.   UK Health Security Agency. SARS-CoV-2 variants of concern and variants under investigation in
England: Technical briefing 32. 2021 Dec.
10.   Molenkamp R, Igloi Z. Evaluation of Antigen rapid test and PCR test to Omicron variant Detection
of Omicron by VirSNP WHO collaborating centre for Arbovirus and Hemorrhagic Fever Reference
and Research.
11.   Deerain J, Druce J, Thomas T, Batty M, Yoga Y, Fennell M, et al. Assessment of the Analytical
Sensitivity of 10 Lateral Flow Devices against the SARS-CoV-2 Omicron Variant. J Clin Microbiol
[Internet]. 2022 Feb 16;60(2):e02479-21. Available from: https://doi.org/10.1128/jcm.02479-21
12.   Bekliz M, Adea K, Puhach O, Perez-Rodriguez F, Marques Melancia S, Baggio S, et al. Analytical
Sensitivity of Eight Different SARS-CoV-2 Antigen-Detecting Rapid Tests for Omicron-BA.1 Variant.
Microbiol Spectr [Internet]. 2022 Aug 8;10(4):e00853-22. Available from:
https://doi.org/10.1128/spectrum.00853-22
13.   Osterman A, Badell I, Basara E, Stern M, Kriesel F, Eletreby M, et al. Impaired detection of
omicron by SARS-CoV-2 rapid antigen tests. Med Microbiol Immunol. 2022 Jun;211(2–3):105–17.
14.   Cocherie T, Bastide M, Sakhi S, Zafilaza K, Flandre P, Leducq V, et al. Decreased Sensitivity of
Rapid Antigen Test Is Associated with a Lower Viral Load of Omicron than Delta SARS-CoV-2
Variant. Microbiol Spectr [Internet]. 2022 Sep 20;10(5):e01922-22. Available from:
https://doi.org/10.1128/spectrum.01922-22
15.   World Health Organisation. Target product profiles for priority diagnostics to support response to
the COVID-19 pandemic v.1.0. 2020.
16.   Kontogianni K, Bengey D, Wooding D, Buist K, Greenland-Bews C, Williams CT, et al. Analytical
evaluation of thirty-two severe acute respiratory syndrome 2 lateral flow antigen tests
demonstrates sensitivity remains with the SARS-CoV-2 Gamma lineage. Rev Soc Bras Med Trop.
2022 Jun;55:e0016.
17.   Mohammadie ZE, Akhlaghi S, Samaeinasab S, Shaterzadeh-Bojd S, Jamialahmadi T, Sahebkar A.
Clinical performance of rapid antigen tests in comparison to RT-PCR for SARS-COV-2 diagnosis in
Omicron variant: A systematic review and meta-analysis. Rev Med Virol. 2023 Mar;33(2):e2428.
18.   Del Vecchio C, Brancaccio G, Brazzale AR, Lavezzo E, Onelia F, Franchin E, et al. Emergence of N
antigen SARS-CoV-2 genetic variants escaping detection of antigenic tests. medRxiv [Internet].
2021 Jan 1;2021.03.25.21253802. Available from:
http://medrxiv.org/content/early/2021/03/26/2021.03.25.21253802.abstract
19.   Despres HW, Mills MG, Shirley DJ, Schmidt MM, Huang ML, Roychoudhury P, et al. Measuring
infectious SARS-CoV-2 in clinical samples reveals a higher viral titer:RNA ratio for Delta and
Epsilon vs. Alpha variants. Proceedings of the National Academy of Sciences [Internet]. 2022 Feb
1;119(5):e2116518119. Available from: https://doi.org/10.1073/pnas.2116518119
20.   de Michelena P, Torres I, Ramos-García Á, Gozalbes V, Ruiz N, Sanmartín A, et al. Real-life
performance of a COVID-19 rapid antigen detection test targeting the SARS-CoV-2 nucleoprotein
for diagnosis of COVID-19 due to the Omicron variant. Journal of Infection [Internet].
2022;84(5):e64–6. Available from:
https://www.sciencedirect.com/science/article/pii/S0163445322001104

21.   Sydney S, J HD, D WI, Stefan R, Sanjucta D, Elisa C, et al. Limit of Detection for Rapid Antigen
420   Testing of the SARS-CoV-2 Omicron and Delta Variants of Concern Using Live-Virus Culture. J Clin
421   Microbiol [Internet]. 2022 Apr 20;60(5):e00140-22. Available from:
422   https://doi.org/10.1128/jcm.00140-22
423   22.   World Health Organisation. Coronavirus disease (COVID-19) Pandemic — Emergency Use Listing
424   Procedure (EUL) open for IVDs. Prequalification of Medical Products (IVDs, Medicines, Vaccines
425   and Immunization Devices, Vector Control).
426   23.   Cubas-Atienzar AI, Kontogianni K, Edwards T, Wooding D, Buist K, Thompson CR, et al. Limit of
427   detection in different matrices of 19 commercially available rapid antigen tests for the detection
428   of SARS-CoV-2. Sci Rep [Internet]. 2021;11(1):18313. Available from:
429   https://doi.org/10.1038/s41598-021-97489-9
430   24.   Kontogianni K, Cubas-Atienzar AI, Wooding D, Buist K, Thompson CR, Williams CT, et al. Lateral
431   flow antigen tests can sensitively detect live cultured virus of the SARS-CoV-2 B1.1.7 lineage. J
432   Infect. 2021 Jul;83(1):e1–4.
433   25.   Body R, FALCON Team. Facilitating AcceLerated Clinical validation Of Novel diagnostics for
434   COVID-19: FALCON-C19. 2023.
435   26.   Byrne RL, Aljayyoussi G, Kontogianni K, Clerkin K, McIntyre M, Wardale J, et al. Head-to head
436   comparison of anterior nares and nasopharyngeal swabs for SARS-CoV-2 antigen detection in a
437   community drive-through test centre in the UK. medRxiv [Internet]. 2023 Apr 4 [cited 2025 Jan
438   29];2022.09.06.22279637. Available from:
439   https://www.medrxiv.org/content/10.1101/2022.09.06.22279637v2
440   27.   Byrne R, Kay G, Kontogianni K, Aljayyoussi G, Brown L, Collins A, et al. Saliva Alternative to Upper
441   Respiratory Swabs for SARS-CoV-2 Diagnosis. Emerging Infectious Disease journal [Internet].
442   2020;26(11):2769. Available from: https://wwwnc.cdc.gov/eid/article/26/11/20-3283_article
443   28.   Tyson JR, James P, Stoddart D, Sparks N, Wickenhagen A, Hall G, et al. Improvements to the
444   ARTIC multiplex PCR method for SARS-CoV-2 genome sequencing using nanopore. bioRxiv
445   [Internet]. 2020 Jan 1;2020.09.04.283077. Available from:
446   http://biorxiv.org/content/early/2020/09/04/2020.09.04.283077.abstract

Acknowledgements.

449   This work was funded as part of FIND's work as co-convener of the diagnostics pillar of the Access to

450   COVID-19 Tools (ACT) Accelerator, including support from Unitaid [grant number: 2019-32-FIND MDR],

451   the governments of the Netherlands [grant number: MINBUZA-2020.961444] and from UK Department

452   for International Development [grant number 300341-102]. The FALCON study was funded by the

453   National Institute for Health Research, Asthma UK and the British Lung Foundation. This study received

454 funding by a UKRI-Medical Research Council (MR/R015678/1) MRC/CASE scholarship to RLB and CGB. The

455 funders had no role in study design, data collection and interpretation, or the decision to submit the work

456 for publication. We thank all participants who volunteered to take part in the study. In the United

457 Kingdom, special thanks go to the NIHR Clinical Research Network (CRN) for their support with the

458 recruitment, specially to Sue Dowling, to the LSTM Diagnostic group—Thomas Edwards, Christopher T

459 Williams, Kate Buist, Karina Clerking, Lorna Finch, Helen Savage, Caitlin R Thompson and Rachel Watkins

460 for recruitment, sample collection, and processing, and to the CONDOR steering group—A. Joy Allen,

461 Julian Braybrook, Peter Buckle, Paul Dark, Kerrie Davis, Adam Gordon, Colette Inkson, Dan Lasserson, Clare

462 Lendrem, Andrew Lewington, Mary Logan, Massimo Micocci, Brian Nicholson, Rafael Perera-Salazar,

463 Graham Prestwich, D. Ashley Price, Charles Reynard, John Simpson, Valerie Tate, Philip Turner and Mark

464 Wilcox—for oversight of the trial in the United Kingdom. We would like to thank the facilitators of the

465 following SARS-CoV-2 isolates obtained through BEI Resources: Isolate hCoV-19/South Africa/CERI-KRISP-

466 K040013/2022 (Lineage BA.5; Omicron Variant), NR-56798, deposited by Dr. Alex Sigal.