A High-Throughput Broad Neutralizing Antibody Assay for Detecting SARS-CoV-2 Variant Immunity in Population

Xiaohan Zhang[1,2#], Yajie Wang[3#], Mansheng Li[1#], Haolong Li[4#], Xiaomei Zhang[1], Xingming Xu[1], Qingqing Ma[5], Di Hu[6], Yan Jia[6], Te Liang[1], Yunping Zhu[1], Surbhi Bihani[7], Sanjeeva Srivastava[7], Manuel Fuentes[8], Yongzhe Li[3*], Xiaoxu Han[5*], Bingwei Wang[2*] and Xiaobo Yu[1,9-11*]

Affiliations

[1]State Key Laboratory of Medical Proteomics, Beijing Proteome Research Center, National Center for Protein Sciences-Beijing (PHOENIX Center), Beijing Institute of Lifeomics, Beijing, 102206, China.

[2]School of Medicine, Nanjing University of Chinese Medicine, Nanjing 210023, China.

[3] Department of Clinical Laboratory, Beijing Ditan Hospital, Capital Medical University, Beijing 100015, China.

[4] Department of Clinical Laboratory, Peking Union Medical College Hospital, Chinese Academy of Medical Science & Peking Union Medical College, Beijing 100730, China.

[5]State Key Laboratory for Diagnosis and Treatment of Infectious Diseases, National Clinical Research Center for Laboratory Medicine,

The First Hospital of China Medical University, China Medical

University, Shenyang, 110001, China

[6]ProteomicsEra Medical Co., Ltd., Beijing, 102206, China

[7]Department of Biosciences and Bioengineering, Indian Institute of

Technology Bombay, Mumbai, Maharashtra, 400076, India.

[8]Department of Medicine and General Cytometry Service-Nucleus,

CIBERONC   ISCIII,   Cancer   Research   Center

(IBMCC/CSIC/USAL/IBSAL), Salamanca, 37007, Spain.

[9]School of Basic Medicine Sciences, Anhui Medical University, Hefei,

Anhui 230031, China.

[10]College of Life Sciences, Hebei University, Baoding, 071002, China.

[11]School of Basic Medicine, Qingdao University, Qingdao 266071 ,

China

[#]These authors contributed equally to this work.

[*]Correspondence to xiaobo.yu@hotmail.com (X.Y.),

bingweiwang@njucm.edu.cn (B.W.), xxhan@yeah.net (X.H.) and

yongzhelipumch@126.com (Y. L).

42 ABSTRACT

Detecting neutralizing antibodies (NAbs) to SARS-CoV-2 variants that

are evolved is crucial to know the escape of host immunity to the newly

arising variants. To address this need, we developed a high-throughput

broad neutralizing antibody (bNAb) assay using flow cytometry with

magnetic-fluorescent microspheres for detecting NAbs against diverse

SARS-CoV-2 variants. The assay is rapid, reliable, 35-fold more

sensitive than Luminex technology. Our results highly correlated with

IgG serological assay (R = 0.90) , the FDA-approved cPass sVNT

assay (R = 0.92), pseudovirus-based neutralizing assay (R = 0.96,

0.66, 0.65) and live virus based neutralization assay (R = 0.79, 0.64).

When applied to 56 healthy individuals receiving third-dose vaccines

(18 CoronaVac; 38 ZF2001) and 35 HIV patients with breakthrough

infection of COVID-19 (16 with CD4+ < 350 cells/μL; 19 with CD4+ >

500 cells/μL), results showed that the Omicron BA.1-BA.5 variants

exhibited significant resistance to inactivated vaccines in healthy

individuals. In HIV patients, the breakthrough infection of Omicron

BA.5.2 or BF.7 variants can induce broad neutralizing activity to

non-Omicron and Omicron variants together with vaccination. Notably,

the levels of NAbs against most of SARS-CoV-2 variants are much

lower in the decreased immunity of HIV patients (CD4+ < 350 cells/μL)

compared to the recovered immunity (CD4+ > 500 cells/μL), indicating

that maintenance of the immune system is crucial for NAb production.

Altogether, our high-throughput proteomics platform represents a

powerful tool for the detection of bNAbs in the population and may

inform the development of more effective COVID-19 vaccines and

vaccination strategies in the future.

KEYWORDS

COVID-19, SARS-CoV-2, flow cytometry, neutralizing antibody, variant

INTRODUCTION

The emergence of severe acute respiratory syndrome coronavirus 2

(SARS-CoV-2) and the ensuing coronavirus 2019 (COVID-19)

pandemic have profoundly impacted global health[1]. Although the acute

crisis phase has subsided, persistent viral transmission and continuous

SARS-CoV-2 evolution remain significant challenges as evidenced by

emerging variants including EG.5.1, JN.1, and KP.2 that continue to

threaten public health[2]. According to reports, 17.13% (95%CI,

7.55–26.71%) of asymptomatic infected individuals may have

long-term health consequences[3], which may include symptoms such

as fatigue, brain fog, dizziness, gastrointestinal symptoms, as well as

others[4]. Therefore, it would remain significant to judge the ability of

newly arising variants to escape the immunity in population defensed

by previous vaccination and/or infections[5].

Neutralizing antibodies (NAbs) of SARS-CoV-2 are a subset of

antibodies that block viral entry by inhibiting the interaction between

the SARS-CoV-2's surface protein, Spike, and the host cell's receptor, angiotensin-converting enzyme 2 (ACE2). Such NAbs can be produced following COVID-19 vaccination or SARS-CoV-2 infection. Identifying SAR-CoV-2 variant specific NAbs is critical for assessing vaccine efficacy, evaluating individual and population immunity, monitoring variant susceptibility, and information public health measures to control the spread of COVID-19[6,7]. However, traditional methods of NAb detection, such those that use a live virus or a pseudovirus-based neutralization assay, are slow, expensive, carry a potential risk of infection, and may require specialized biosafety level 3 (BSL3) facilities[8-11].

As an alternative approach, enzyme-linked immunosorbent assay (ELISA) can perform the *in vitro* neutralizing testing using purified Spike, ACE2, or Spike's receptor-binding domain (RBD) that interacts directly with ACE2 during viral entry. Several commercial ELISA-based surrogate virus neutralization tests (sVNT) have been developed, including TECO sVNT (TECO Medical) and cPass sVNT (GenScript), which have been approved by the United States' (U.S.) Food and Drug Administration (FDA) or received Conformite Europeenne (CE) marking for use in the European Union[12,13]. Unfortunately, ELISA-based methods only detect a single variant at a time, making

them a time-consuming option when multiple SARS-CoV-2 variants are

analyzed[9].

To address this concern, Fenwick *et al.* developed a multiplexed

in-vitro quantitative neutralization assay using Luminex technology in

2021. The assay enables the detection of NAbs targeting various Spike

mutations, including D614G, D614G plus M153T, N439K, S477N,

S477R, E484K, S459Y, N501T, N501Y, K417N, 60-70, P681H, Y453F,

and their combinations. The results obtained from this assay correlated

with pseudovirus neutralization (R^2=0.65) and live virus infection (R^2 =

0.825) assays[10].

In a similar vein, Lynch et al. developed a multiplexed surrogate

virus neutralization test (plex-sVNT) for detecting NAbs to seven

SARS-CoV-2 variants, namely, wild type (WT), Alpha, Beta, Gamma,

Delta, Kappa, and Epsilon. The results highly correlated (> 96%) with

those obtained from a plaque reduction neutralization test (PRNT)[11].

However, it should be noted that these technologies have

limitations as they require access to specific Luminex and Bio-Rad

instruments. This requirement restricts their adoption primarily to

specialized laboratories equipped with these systems, limiting broader

accessibility and high-throughput screening potential. Moreover, some

recent Omicron variants have not been accessible for analysis using

these methods to date[14].

To overcome these limitations, specifically the need for a

multiplexed, high-throughput assay that is accessible to a wider range

of laboratories using standard equipment and capable of covering

emerging variants, we developed a high-throughput broad NAb (bNAb)

assay. This approach leverages the widespread availability and

standardization of flow cytometers in research and clinical settings. Our

assay enables the systematic and simultaneous detection of NAbs to

eleven SARS-CoV-2 variants of concern (VOCs), including D614G,

Alpha, Beta, Gamma, Delta, Kappa, Omicron BA.1-BA.5, in a format

designed to be reliable, cost-effective, sensitive, and crucially, utilizable

with a standard flow cytometer[15]. This addresses the core accessibility

barrier of the Luminex-based methods.

We conducted a comparative analysis of NAb detection using our

SARS-CoV-2 bNAb assay with commercial ELISA-based serology, U.S.

FDA-approved cPass sVNT, pseudovirus-based neutralization and live

virus based neutralization assays. This validation demonstrates the

performance and reliability of our flow cytometry platform as a robust

alternative or complement to existing technologies. Furthermore, we

explored the applications of the SARS-CoV-2 bNAb assay for

screening therapeutic antibodies, assessing their neutralizing activity

SARS-CoV-2. Additionally, we analyzed the broad responses of

serological NAbs to SARS-CoV-2 variants in individuals who received

the third dose of either the inactivated vaccine (Sinovac-CoronaVac) or

the recombinant subunit vaccine (ZF2001) for COVID-19. Finally, we

assessed the serum NAb inhibition rates (%) against SARS-CoV-2

variants in the decreased immunity (CD4+ < 350 cells/μL) and

recovered immunity (CD4+ > 500 cells/μL) of HIV patients with

breakthrough infection of COVID-19 (Figure S1).

RESULTS

Schematic illustration of SARS-CoV-2 bNAb assay

In the SARS-COV-2 bNAb assay, illustrated schematically in Figure 1A,

the trimerized Spike proteins of six non-Omicron variants (D614G,

Alpha, Beta, Kappa, Gamma, Delta) and five Omicron variants (BA.1,

BA.2, BA.3, BA.4, BA.5) were expressed from human embryonic

kidney 293 (HEK293) cells, purified, and coupled to the

magnetic-fluorescent beads, as previously reported[10,16]. During the

NAbs detection process, all coupled beads were mixed together and

incubated with samples containing NAbs. The NAbs bound to the

Spike trimer proteins, preventing their interaction with the biotinylated

ACE2 receptors, which were coupled to streptavidin-phycoerythrin

(SA-PE). The inhibition rate (%) was calculated as follows: Inhibition

rate (%) = (1 - signal of NAb inhibition on Spike-ACE2 binding / signal

of Spike-ACE2 binding) × 100%.

The sensitive detection of Spike-ACE2 interactions is essential for

assessing the presence of NAbs in clinical samples. To address this

requirement, we coupled the Spike protein from six SARS-CoV-2

variants (D614G, Alpha, Beta, Kappa, Delta, Omicron BA.1) onto

magnetic-fluorescent beads from Luminex and Shenzhen Wellgrow

Technology Co., Ltd. ("Wellgrow"), and then incubated with different

concentrations of ACE2. Magnetic-fluorescent beads with the

SARS-CoV-2 nucleocapsid (N) protein, rather than the Spike protein,

was employed as the negative control. The binding signals were

measured using Luminex-200 and Wellgrow flow cytometry. Notably,

the fluorescent signal intensity and signal to noise ratio (SNR) were

much higher with the Wellgrow platform than the Luminex platform

when ACE2 concentrations ranged from 0.01 µg/mL to 100 µg/mL

(Figure 1B, Figure S2 and Table S6).

Furthermore, we calculated the sensitivity or the lowest detection

limit (LOD) of each method, in which the LOD was equal to the signal

of the buffer control plus ten standard deviations (Figure 1C). The data

showed that the average LOD of our assay using the Wellgrow

platform was 35.89 ± 57.8 [-21.91 ~ 93.69]-fold higher than the

Luminex platform for all variants (D614G, Alpha, Beta, Kappa, Delta,

Omicron BA.1).

At last, we compared the difference of NAb titers obtained by

detecting the inhibition of antibody #26 to Spike-ACE2 interaction. The

results showed that the IC50 obtained by two platforms are highly

consistence with the r correlation of 0.9988 (Figure S2D). Therefore,

we chose the Wellgrow platform as the proteomics platform for the

SARS-CoV-2 bNAb assay in this study.

Mapping Spike-ACE2 interactions of SARS-CoV-2 variants

Using our platform, we conducted a systematic investigation of the

interactions between ACE2 and Spike trimer proteins from eleven

SARS-CoV-2 variants. This was achieved by incubating a bead array

with different concentrations of ACE2 (Figure 2). The results

demonstrated that the fluorescent signals from the Spike trimer

proteins increased with increasing concentrations of ACE2, indicating

the formation of Spike-ACE2 complexes on the beads (Figure 2A).

Spike mutations were found to influence the interaction dynamics.

At the maximal concentration of 10 µg/mL ACE2, Alpha and Delta

variants exhibited the highest signals, followed by Beta, BA.4, BA.5,

D614G, Gamma, BA.1, BA.2, and Kappa, with BA.3 showing the

lowest signal. In contrast, the N protein, used as a negative control,

consistently displayed the lowest signal regardless of ACE2

concentration.

To further assess the binding affinity, we calculated the

half-maximal effective concentration (EC50) using the standard curve

for each SARS-CoV-2 variant and expressed values relative to

Omicron BA.1 (1.00×) (Figure 2B). The results showed that Omicron

BA.1 has the highest binding affinity with an EC50 of 0.2913 μg/mL,

followed by Beta (0.2915 μg/mL, 1.00×), Delta (0.3415 μg/mL, 1.17×),

Omicron BA.4 (0.3592 μg/mL, 1.23×), Gamma (0.3833 μg/mL, 1.32×),

Alpha (0.3940 μg/mL, 1.35×), Kappa (0.3983 μg/mL, 1.37×), Omicron

BA.2 (0.4307 μg/mL, 1.48×), Omicron BA.5 (0.4359 μg/mL, 1.50×),

D614G (0.5729 μg/mL, 1.97×), and Omicron BA.3 (1.2830

μg/mL, 4.40×). These datas demonstrate significant quantitative

differences in the binding affinity of Spike-ACE2 interactions across the

tested SARS-CoV-2 variants. The results are in accordance with

surface plasmon resonance data in which BA.1 has the strongest

interaction with ACE2 with a dissociation constant (K_D) of 2.10 nM,

followed by BA.2 (2.21 nM) and D614G (5.20 nM)[17]. Similar results

were obtained by Mahalingam et al., who used ELISA to determine the

EC50 of ACE2 to Spike Omicron (0.38 nM), Delta (0.48 nM), and WT

(2.28 nM)[18]. Collectively, our data provide a comprehensive landscape

of Spike trimer-ACE2 interactions for SARS-CoV-2 VOCs, as classified

by the U.S. Centers for Disease Control and Prevention (CDC)

(https://www.cdc.gov/coronavirus/2019-ncov/variants/variant-classifica

tions.html). It should be noted that the Epsilon variant was not included

in this work due to the unavailability of purified Spike trimer protein at

the time of executing experiments.

In addition, to demonstrate the specificity of our assay, we tested

the binding ability of ACE2 to the Spike proteins from MERS,

SARS-CoV and six SARS-CoV-2 variants. The results showed that the

binding signal of ACE2 to six SARS-CoV-2 variant Spike proteins was

much higher than that of SARS-CoV and MERS. The binding signal of

SARS-CoV was 3.38-5.74 times lower than that of SARS-CoV-2

variants, due to the homology of Spike sequence. There was no

binding signal observed between ACE2 and MERS, similar to the

negative control, N protein (Figure S3). All these results are in

accordance with the reported results[19], and demonstrate our assay is

specific.

Comparison of SARS-CoV-2 bNAb assay to ELISA , cPass sVNT,

pseudovirus-based neutralization and live virus based

neutralization assays.

To assess the reliability of the SARS-CoV-2 bNAb assay, we compared

the SARS-CoV-2 bNAb assay to an ELISA that was used to detect

anti-Spike IgG antibodies[20]. The results showed a high correlation

between the levels of D614G NAbs detected using SARS-CoV-2 bNAb

assay and the levels of anti-Spike IgG antibodies obtained by the

ELISA (R = 0.90, p < 2.2e-16) in a cohort of 75 individuals who

provided clinical serum samples approximately 180-270 days after the

third dose of the COVID-19 vaccine (Figure 3A). ROC analysis

demonstrated excellent diagnostic performance of the SARS-CoV-2

NAb assay using ELISA as reference standard, yielding an AUC of

0.894 (95% CI: 0.813-0.975) (Figure S4A).

Next, within the same cohort, we compared the NAb results

obtained from the SARS-CoV-2 bNAb assay to those obtained from the

FDA-approved cPass sVNT assay[9,12]. The data obtained using the

SARS-CoV-2 bNAb assay and cPass sVNT assay demonstrated a

higher correlation to each other (R = 0.92, p < 2.2 e-16) compared to

the assay using Luminex technology (R = 0.85), as previously reported

by Marien et al.[21]. The positive percent agreement (PPA) was 92.7%

(80.1% - 98.5%) and the negative percent agreement (NPA) was 94.1%

(80.3% - 99.3%) (Figure 3B). ROC analysis confirmed high accuracy of

the SARS-CoV-2 NAb assay using cPass sVNT as reference standard,

with an AUC of 0.964 (95% CI: 0.923-1.000) (Figure S4B). The findings

suggest that some of the anti-Spike IgG antibodies detected by the

ELISA may lack neutralizing activity and therefore were not detected by

SARS-CoV-2 bNAb assay[22]. Overall, these results demonstrate the

reliability of the SARS-CoV-2 bNAb assay as an *in vitro* platform for

detecting serum NAbs.

Additionally, we compared the inhibition rate (%) of NAbs obtained

by SARS-CoV-2 bNAb assay to the NAb titers obtained by

pseudovirus-based neutralization assay or live virus based

neutralization assay. For 10 serum samples from mice administered

with WT mRNA vaccines and 10 serum samples from mice

administered BA.1 mRNA vaccines, a high correlation was found

between the D614G NAbs inhibition rate (%) detected using

SARS-CoV-2 bNAb assay and WT NAb titers detected by

pseudovirus-based neutralization assay (R = 0.96, p < 0.0001) and live

virus based neutralization assay (R = 0.79, p < 0.0001) (Figures 3C,

3E). Similarly, notable correlations were observed between the levels

of BA.1 NAbs detected using SARS-CoV-2 bNAb assay and

pseudovirus-based neutralization (R = 0.66, p = 0.0017) and live virus

based neutralization assays (R = 0.64, p = 0.0022) (Figures 3D, 3F).

Then, we compared the inhibition rate (%) of NAbs obtained by

SARS-CoV-2 bNAb assay with the NAb titers obtained by

pseudovirus-based neutralization assay in 135 clinical serum samples

collected from individuals 28 days after receiving the second dose of

the COVID-19 vaccine. We detected serum NAbs against the D614G

variant using both assays. The results showed a correlation coefficient

(R) of 0.65, demonstrating the consistency between the results

obtained through inhibition and those obtained with NAb titers (Figure

S5).

Identification of broadly neutralizing monoclonal/polyclonal

antibodies against SARS-CoV-2 variants

To demonstrate the application of SARS-CoV-2 bNAb assay in

identifying monoclonal/polyclonal antibodies for potential treatment

purposes, we tested six anti-Spike antibodies that are available in the

laboratory[23] (Figure 4A). The results revealed that three antibodies

(#26, #20, #21) have neutralizing activity. Antibodies #20 and #21

inhibited ACE2 binding to D614G, Alpha, Beta, Kappa and Delta

variants, but were not effective in inhibiting the ACE2 binding to the

Omicron BA.1 variant. Antibody #26 exhibited the ability to inhibit

Spike-ACE2 interactions for D614G, Alpha, Beta, Kappa,Delta and

Omicron BA.1 variants.

Subsequently, the half-maximal inhibitory concentration (IC50)

values were calculated and represented in a heatmap (Figure 4B). The

results showed that antibody #26 possessed broad neutralizing activity

to all tested SARS-CoV-2 variants (Kappa, D614G, Alpha, Delta, Beta,

Omicron BA.1). The IC50 to D614G (0.47 µg/mL) was confirmed with

ELISA (0.35 µg/mL) and pseudovirus-based neutralization assay (0.21

μg/mL) in our previous work[23]. While the IC50 of antibody #26 to the

Omicron variant was ten-fold lower than non-Omicron variants. These

findings underscore the ability of our assay to rapidly screen for

monoclonal/polyclonal   antibodies   with   potential   therapeutic

applications.

Evaluation of serum NAbs in healthy individuals receiving the

third dose vaccination and HIV patients with breakthrough

infection of COVID-19

We tested the serum NAbs from 56 healthy individuals who received

two doses of the inactivated vaccine (Sinovac-CoronaVac) and

boosted with a third dose with either an inactivated vaccine or

recombinant subunit vaccine (ZF2001) (Table 1)[24]. The inhibition rate

(%) was calculated by deducting the background noise using 147

serum samples collected prior to the COVID-19 pandemic[25]. The

results showed that the third dose vaccination elicited a strong

protective immune response against the non-Omicron variants with the

highest inhibition rate (%) of D614G (reference = 1.00×), followed by

Gamma, Alpha, Delta, and Kappa variants (all > 0.80×). Omicron

variants exhibited substantial immune escape, with BA.4 showing the

greatest escape (0.33×), followed by BA.5 (0.35×), BA.1 (0.43×), BA.3

(0.41×), and BA.2 (0.44×), indicating the resistance of these variants to

the Sinovac-CoronaVac vaccine and the ZF2001 vaccine (Figure 5A).

Next, we compared the inhibition rate (%) of serum NAbs in

individuals who received either a third dose of inactivated vaccine or

recombinant vaccine; no significant difference was observed (Figure

S6). These findings highlight the immunological resistance of different

SARS-CoV-2 variants to the Sinovac-CoronaVac and ZF2001 vaccines,

and suggest the need for developing a new vaccine that provides

broad protection against SARS-CoV-2 variants.

Furthermore, in order to demonstrate the specificity of our

SARS-CoV-2 bNAb assay, we detected the serum NAbs from the

individuals before vaccination (n = 147) and after three-dose

vaccination (n = 56). The results showed that the serum NAb levels

were significantly higher in the vaccination group than un-vaccination

group (p < 0.0001) (Figure S7), indicating our SARS-CoV-2 bNAb

assay is highly specific for the detection of serum NAbs.

Finally, we tested the serum NAbs from 35 HIV patients who

received three-dose inactivated vaccination and experienced Omicron

BA.5.2 or BF.7 breakthrough infections (Table 2)[26]. Subsequently, we

assessed the inhibition rate (%) on 20 SARS-CoV-2 variants, which

included Omicron EG.5.1 and JN.1. The results showed that Omicron

EG.5.1 and JN.1 exhibited strong immune escape with low inhibition

rates. Notably, individuals with recovered immunity (CD4+ > 500 cells/μL) exhibited significantly higher serum NAb inhibition rates against most of SARS-CoV-2 variants compared to individuals with decreased immunity (CD4+ < 350 cells/μL) ($p < 0.05$) (Figure 5B). This demonstrates that while breakthrough infection enhances variant coverage, optimal immune reconstitution, rather than mere infection exposure, is essential for generating cross-protective immunity against highly evasive variants like EG.5.1 and JN.1 in HIV populations.

DISCUSSION

Even though the World Health Organization has announced that COVID-19 is no longer a public health emergency of international concern, the genome sequence of SARS-CoV-2 is continually evolving and the new variants (e.g., EG.5.1, JN.1, and KP.2) may have the potential to escape the immunity and threat the people's health, especially among immunocompromised individuals[27,28]. Therefore, there is still a need for effective COVID-19 vaccines to prevent and control viral transmission. Critically, achieving durable longitudinal immunity — defined as sustained, broad protection against evolving variants over time — remains an unmet challenge. Our data underscore that current vaccine-induced or infection-acquired immunity often wanes rapidly and exhibits limited cross-reactivity

against emerging variants. The development and deployment of

vaccines for COVID-19, as well as the evaluation of population

immunity, continue to be important for public health[29].

In this work, we developed a high-throughput SARS-CoV-2 bNAb

assay to detect a wide range of NAbs to non-Omicron and Omicron

variants. Compared to the assay using Luminex technology[10,11], our

SARS-CoV-2 bNAb assay is 35-fold more sensitive, 2.5-fold less costly,

and can be accessible to any laboratory that has a standard flow

cytometer with the 532 nm and 635 nm lasers. Moreover, the

simultaneous detection of NAbs to multiple SARS-CoV-2 variants can

be achieved within a single experiment, and the assay can be

periodically updated to include new variants. The reliability of NAb

detection using our SARS-CoV-2 bNAb assay is demonstrated through

comparisons with ELISA-based IgG serology, the cPass sVNT assay,

pseudovirus-based neutralization assay and live virus-based

neutralization assay (Figure 3).

Using this platform, we constructed a comprehensive landscape of

Spike trimer and ACE2 interactions for eleven SARS-CoV-2 variants

(Figure 2). The results provide evidence that Omicron BA.1 showed the

strongest binding affinity to the ACE2 receptor[30,31]. Notably, the

transmission fitness of SARS-CoV-2 variants is governed by

multifactorial determinants, including viral binding affinity to host

receptors, immune evasion capability against NAbs, intrinsic viral

replication efficiency, and host population immunity. These collective

factors shape variant-specific epidemic trajectories and underscore the

need for continuous surveillance of emerging variants[32].

Furthermore, we determined the bNAbs in 822 serum samples

from human and mouse vaccinated with different SARS-CoV-2

vaccines (WT, D614G, Delta, BA.1, BA.4/5) as well as 890 serum

samples from individuals infected with SARS-CoV-2 variants (WT,

D614G, Alpha, Delta, BA.1, BA.2, BA.2.76, BA.5.2, BF.7, XBB, and

EG.5). The results showed the adaptive changes of population

immunity variations induced by vaccination and infection[33,34].

Importantly, while NAbs serve as critical correlates of protection,

emerging evidence highlights the complementary role of T cell effector

functions in antiviral immunity. Previous studies, including work from

this group evaluating an RBD-nucleocapsid fusion protein as a

COVID-19 booster candidate, demonstrate that protein-based

vaccines can elicit protective immunity even in the absence of

detectable NAbs[35]. CD8+ T cells and CD4+ T cells contribute to viral

clearance, memory responses, and protection against variant escape,

particularly in immunocompromised populations where antibody

responses may be suboptimal. This underscores the value of

integrating T cell immunity assessments into vaccine evaluation frameworks.

Finally, we found that in HIV patients with breakthrough infection of COVID-19, individuals with recovered immunity (CD4+ > 500 cells/μL) exhibited significantly higher serum NAb inhibition rates (%) against most of SARS-CoV-2 variants compared to individuals with decreased immunity (CD4+ < 350 cells/μL)[36]. These results demonstrate the significance of our proteomics platform in evaluating the broad protective activity of COVID-19 vaccines and COVID-19 infection in population, which can provide valuable insights for future vaccine development and public health strategies.

The imperative for vaccines inducing longitudinal immunity is further emphasized by our population-level findings. The adaptive changes in humoral immunity we observed across vaccinated and infected cohorts reflect the dynamic interplay between viral evolution and immune erosion. This underscores that narrow-variant targeting in vaccines may offer transient protection at best. To establish persistent population immunity, next-generation vaccines must prioritize conserved epitopes capable of eliciting cross-reactive bNAbs against both current and future variants. Our bNAb platform enables rapid immunogenicity benchmarking against emerging variants (e.g., KP.2),

providing essential tools for developing broad-spectrum vaccines that

confer variant-resilient longitudinal immunity.

It is important to acknowledge some limitations of our study. First,

the *in vitro* detection of NAbs may not represent viral neutralization *in*

vivo, although previously reported evidence has shown a good

correlation between NAbs detected *in vitro* and *vivo*[8-11]. Second, the

relationship between the inhibition levels obtained with the

SARS-CoV-2 bNAb assay and the SARS-CoV-2 infection remains

unclear. Third, the limited sample size and cohort design constrained

our ability to conduct granular sub-analyses, particularly in

immunocompromised subgroups (e.g., HIV cohort), to differentiate

immune profiles primarily induced by vaccination versus natural

infection, and to comprehensively evaluate NAb levels across diverse

vaccine types and SARS-CoV-2 variants in larger populations. Fourth,

while the assay currently includes 20 SARS-CoV-2 variants, it does not

encompass all circulating or emerging lineages (e.g., KP.2). This

reflects a practical limitation at the time of study execution. Crucially,

the modular design of our bead array platform inherently allows for

straightforward extension to incorporate new variants in future

iterations. Fifth, the Wellgrow platform utilizes a high-resolution 20-bit

analog-to-digital converter (ADC), resulting in raw fluorescence

intensity values approximately 100-fold higher than systems using

14-bit ADCs (e.g., Luminex). While this instrumentation difference

explains the apparent higher background observed with negative

controls (e.g., N protein) and does not reflect reduced specificity, it is

an important technical consideration when interpreting raw MFI data.

Finally, the SARS-CoV-2 WT was not employed due to the

unavailability of this protein during executing experiments. Instead, we

employed the D614G variant, which shares high homology with the

wild type. Previous studies have demonstrated comparable

neutralization potency between the D614G variant and the WT[37].

Additionally, the D614G substitution does not significantly alter

SARS-CoV-2 morphology and binding to the ACE2 receptor[38].

CONCLUSION

We have successfully developed and validated a high-throughput

proteomics platform, enabling the comprehensive evaluation of NAbs

against a wide range of SARS-CoV-2 variants. This platform offers an

efficient, sensitive, and cost-effective approach for detecting bNAbs

within the population, providing an opportunity to guide the design and

evaluation of vaccines with enhanced protective efficacy against

evolving SARS-CoV-2 variants. Overall, our high-throughput

proteomics platform represents a powerful tool for the detection of

bNAbs in the population and may inform the development of more

effective COVID-19 vaccines and vaccination strategies in the future.

MATERIALS AND METHODS

Collection of experimental samples

56 clinical serum samples, collected from healthy individuals at Beijing

Ditan Hospital 28 days after administration of the third dose of the

COVID-19 vaccine in September 2021 (Table 1). The study received

approval from the Beijing Ditan Hospital Ethics Committee (No.

2021-010-01), and all participants provided written informed consent[39].

35 clinical serum samples were collected from HIV patients with

breakthrough infection of COVID-19 at The First Affiliated Hospital of

China Medical University from March 1, 2023 to July 1, 2023 (Table 2).

By the end of 2022, the predominant variants of the COVID-19 in China

were the Omicron BA.5.2 and BF.7. The subjects were HIV patients

who were infected with the SARS-CoV-2 during the COVID-19

pandemic in China in 2022[26]. HIV patients the following criteria were

included in the study: (1) HIV-1 antibody positive; (2) aged ≥ 18; (3)

having COVID-19 infection information; (4) having records of

COVID-19 vaccination; (5) being willing to provide blood samples for

laboratory testing, such as COVID-19 antibody testing.

A total of 75 clinical serum samples, collected from Peking Union

Medical College Hospital (PUMCH) between May and September 2022

and obtained from individuals approximately 180 or 270 days after receiving the third dose of the COVID-19 vaccine, were utilized for correlation analysis with the results from the commercial ELISA-based serology and cPass sVNT assay. The study was approved by the institutional review board of PUMCH (No. K1965-K22C0433), and informed consent was obtained from all participants.

147 clinical serum samples from unvaccinated healthy individuals, along with 135 clinical serum samples from healthy individuals used for correlation analysis with pseudovirus-based neutralization assays (collected 28 days after the second dose of the COVID-19 vaccine), were obtained from Beijing Chaoyang Hospital, in January 2021. The study was approved by the Beijing Chaoyang Hospital Ethics Committee (No. 2020-ke-491), and informed consent was obtained from all participants.

20 serum samples of mouse were collected from 10 mice immunized with WT mRNA vaccines and 10 mice immunized with Omicron BA.1 mRNA vaccines 28 days post-vaccination, obtained under laboratory conditions.

Preparation of reagents for SARS-CoV-2 bNAb assay

The SARS-CoV-2 bNAb assay was developed in ProteomicsEra Medical Co., Ltd (Beijing, China). Briefly, 3 μg SARS-CoV-2 Spike

trimer proteins (Sino Biological., Novoprotein and ACROBiosystems;

China) (Table S1) were coupled to 1×10^6 magnetic-fluorescent beads

(Wellgrow Technology Co., Ltd.) as previously described[40]. First, 12

differently encoded magnetic-fluorescent beads, numbered 102-108

and 201-205 (Table S2), were chosen. Then 125 µL containing

1×10^6 beads were washed with 100 µL ddH2O and activated with 80

µL of "activation buffer" (50 mM 2-(N-morpholino) ethanesulfonic acid

(MES) pH 5.0) (Sigma-Aldrich, St. Louis, MO). Then the activated

beads were mixed with 10 µL of 50 mg/mL sulfo-N-hydroxysuccinimide

(NHS) (Thermo Scientific, USA) and 10 µL of 50 mg/mL

1-Ethyl-3-[3-dimethylaminopropyl]-carbodiimide hydrochloride (EDC)

(Solarbio, Beijing, China). After incubation for 20 min at room

temperature with gentle mixing, the beads were washed twice with 250

µL "coupling buffer" (50 mM MES pH 5.0) and resuspended in 200 µL

coupling buffer. Then 3 µg of Spike trimer proteins were added to the

bead solution and incubated for 2 h at room temperature. After washing,

the beads underwent blocking with "blocking buffer" (PBST-B, PBS pH

7.4, 0.05% Tween-20, 1% BSA), and the coupled beads were

subsequently stored at 2-8 °C in "storage buffer" (PBST-BN, PBS pH

7.4, 0.05% Tween-20, 0.1% BSA, 0.05% NaN3).

Detection of Spike-ACE2 interactions using SARS-CoV-2 bNAb

assay

First, a 50 μL mixed solution containing 12 different SARS-CoV-2 Spike

trimer protein-coupled beads (2500 beads per type) (Table S3) was

added to each well of a 96-well plate, followed by 50 μL of different

concentrations (0, 0.014, 0.041, 0.123, 0.370, 1.111, 3.333, 10 μg/mL)

of biotinylated ACE2 (Sino Biological). After incubation for 1 h at room

temperature on a shaker, the beads were magnetically separated using

a magnetic separator (Wellgrow Technology Co., Ltd.) and washed

three times with 100 μL PBST-B (PBS pH7.4，0.05% Tween-20, 0.1%

BSA). After washing, the Spike-ACE2 interaction was detected using

50 μL SA-PE (Thermo Scientific, USA) (2 μg/mL) for 30 minutes at

room temperature. Finally, after washing with PBST-B twice, the beads

were resuspended in a 200 μL PBST-B solution. The fluorescent signal

was then detected at 200 beads/region using the EasyCell flow

cytometry (Wellgrow Technology Co., Ltd.) with excitation wavelengths

of 532 nm and 635 nm.

Detection of monoclonal/polyclonal antibodies using SARS-CoV-2

bNAb assay

To detect monoclonal/polyclonal antibodies, a 50 μL mixed solution

containing 7 different SARS-CoV-2 Spike trimer protein-coupled beads

(2500 beads per type) (Table S4) was added to each well of a 96-well

plate. Following that, 50 µL of each antibody solution at different

concentrations (0, 0.014, 0.041, 0.123, 0.370, 1.111, 3.333, 10 µg/mL)

for anti-Spike antibodies (#26, #20, #21, #22, #23, #73) (Table S5) was

added to individual wells of a 96-well plate. After incubation for 2 h at

room temperature on a shaker, the beads were washed with PBST-B

three times, then the beads were incubated with 50 µL biotinylated

ACE2 (0.5 µg/mL) for 1 hour at room temperature. After washing, the

binding of ACE2 to Spike trimer proteins were detected using 50 µL

SA-PE (2 µg/mL) for 30 minutes at room temperature. Finally, after

washing with PBST-B twice, the beads were resuspended in a 200 µL

PBST-B solution. The fluorescent signal was then detected at 200

beads/region using the EasyCell flow cytometry (Wellgrow Technology

Co., Ltd.) with excitation wavelengths of 532 nm and 635 nm.

Detection of Serum NAbs using SARS-CoV-2 bNAb assay

First, a 50 µL mixed solution containing 12 different SARS-CoV-2 Spike

trimer protein-coupled beads (2500 beads per type) (Table S2) was

added to each well of a 96-well plate, followed by 50 µL of the serum

samples which were diluted to 1:20 with PBST-B. After incubation for 2

h at room temperature on a shaker, the beads were washed with

PBST-B three times. Then the beads were incubated with 50 µL

biotinylated ACE2 (0.5 µg/mL) for 1 hour at room temperature. After

washing, the binding of ACE2 to Spike trimer proteins were detected

using 50 µL SA-PE (2 µg/mL) and incubated the mixture for 30 minutes

at room temperature. Finally, after washing with PBST-B twice, the

beads were resuspended in a 200 µL PBST-B solution. The fluorescent

signal was then detected at 200 beads/region using the EasyCell flow

cytometry (Wellgrow Technology Co., Ltd.) with excitation wavelengths

of 532 nm and 635 nm.

Detection of Serum anti-SARS-CoV-2 Spike RBD IgG antibodies

using ELISA

The assay was conducted as described previously[20]. A capture

sandwich ELISA detection kit (PROPRIUM, Hangzhou, China) was

used to detect SARS-CoV-2 antibodies against the Spike protein RBD.

The SARS-CoV-2 Spike RBD protein was pre-coated onto the solid

phase, allowing it to form an antigen-antibody complex with IgG

anti-RBD antibodies from 100 µL of diluted serum samples (1:30

dilution ratio) or standards. After washing, HRP-conjugated anti-human

IgG was added to form an antigen-antibody-HRP complex. The

substrate solution TMB was then introduced, and the resulting color

intensity was proportional to the level of SARS-CoV-2 NAbs. The

optical density (OD) at 450 nm was measured, and a threshold of 10

BAU/ml was used to determine sero-positive and negative samples for

anti-Spike RBD IgG.

Detection of Serum NAbs using cPass sVNT assay

The assay was conducted as described previously[20]. Using competitive

ELISA, the SARS-CoV-2 sVNT assay (Genscript, Nanjing, China)

identified NAbs present in the bloodstream that obstruct the interaction

between the viral Spike glycoprotein RBD and human ACE2. Following

the manufacturer's instructions, diluted serum samples (1:10 dilution

ratio) and controls were pre-incubated at 37□ for 15 minutes to allow

binding of NAb and HRP-RBD (antigen derived from D614G variant),

using a 1:1 volume ratio. Subsequently, the mixture was transferred to

a capture plate, pre-coated with hACE2 protein, where NAb-unbound

HRP-RBD was captured on the plate, while NAb-bound HRP-RBD

remained in the supernatant and was washed away. Next, 100 μL of

TMB and 50 μL of stop solution were added, and the plate was read at

450 nm. The sample's absorbance was inversely correlated with the

titers of anti-SARS-CoV-2-NAbs. To ensure result validity, the OD450

values of positive controls (> 1.0) and negative controls (< 0.3) had to

fall within specific ranges. An inhibition rate of ≥ 30% was considered

positive for SARS-CoV-2 NAb determination. The inhibition rate (%)

was calculated as follows: Inhibition□rate (%) =□(1 - OD value of

sample/OD value of negative control)□×□100%.

Pseudovirus-based neutralization assay

The SARS-CoV-2 pseudovirus was prepared using the VSV-ΔG system in which the glycoprotein (G) gene was replaced with the firefly luciferase (Fluc) reporter gene[41,42]. The S protein was overexpressed and displayed on the VSV pseudovirus. Following S-ACE2 interaction, the pseudovirus entered the host cell where the Fluc gene was transcribed and translated. The addition of luciferase substrate resulted in luminescence where the amount of luminescence is proportional to the level of pseudoviral entry. The preparation process of SARS-CoV-2 pseudovirus begins with the cloning of the S gene encoding the S protein of the virus. Subsequently, this gene is inserted into a modified VSV backbone, replacing the original G gene. The resulting construct, containing the S protein gene in lieu of the G gene and complemented by the Fluc reporter gene, forms the foundation of the pseudovirus. The pseudovirus equipped with the S protein can interact with the ACE2 receptor on the surface of host cells, thereby facilitating the entry of the pseudovirus into the cells. Once inside the cells, the Fluc gene will be transcribed and translated to produce luciferase.

The assay was conducted following previously described methods[23,43]. Huh7 cells were seeded at a concentration of 2×10^4

cells per well in 96-well plates and incubated until reaching 90-100%

confluency, typically around 24 hours. Serum samples, subjected to

serial 3-fold dilutions commencing at 1:10, were then incubated with

650 TCID50 of the pseudovirus for a precisely controlled duration of 1

hour at 37 °C. Dulbecco's modified eagle medium (DMEM) served as

the negative control. As a crucial negative control, DMEM was

employed. Following the incubation period, the supernatant was

carefully aspirated, making way for the addition of luciferase substrate

to each well. A subsequent 2-minute incubation in darkness at room

temperature facilitated optimal enzymatic reactions. Luciferase activity,

indicative of NAb presence, was quantified using the GloMax® 9633

Microplate Luminometer (Promega, Madison, USA).

Live virus based neutralization assay

The assay was conducted as described previously[43]. To assess the

SARS-CoV-2-specific NAb titer in serum, we employed a cytopathic

effect (CPE)-based microneutralization assay, utilizing the

SARS-CoV-2 virus strain BetaCoV/Beijing/IME-BJ01/2020 (Accession

No. GWHACAX01000000) and Vero cells (ATCC, CCL81). Serum

samples underwent heat inactivation for 30 minutes at 56 °C and were

subsequently two-fold serially diluted (ranging from 1:4 to 1:2048)

using DMEM from Thermo Fisher Scientific. These dilutions were then

mixed with an equivalent volume of the virus solution to achieve a 50%

tissue culture infectious dose (TCID50) of 100 in each well.

The serum-virus mixtures were incubated for 1 hour at 37 °C to

allow for sufficient neutralization reactions to occur. Following

incubation, the mixtures were gently added to 96-well plates containing

semi-confluent Vero cells with a density exceeding 80%. The plates

were then incubated for a further 3 days at 37 °C to allow for the

development of CPEs on the Vero cells. Using an inverted microscope,

the CPEs on the Vero cells were carefully observed and recorded. The

neutralizing titer was determined as the reciprocal of the highest

sample dilution that successfully protected at least 50% of the cells

from CPE. In cases where no neutralization reaction was observable

even at the initial serum dilution of 1:4, an arbitrary titer of 2 (half of the

limit of quantification) was assigned.

Data analysis

To compare the inhibition rate (%) of serum NAbs against the D614G

variant and other SARS-CoV-2 variants, the raw inhibition rates of

different variants were first normalized using minimum-maximum

normalization. In order to estimate dose-response relationships, curve

fitting and visualization were performed using GraphPad Prism

software (version 8.3.0). The Wilcoxon rank sum test was performed

using stat_compare_means function implemented in R package

ggpubr (version 0.6.1). Inhibition rate (%) of serum NAbs with a p <

0.05 considered as statistically significant. Correlations between the

SARS-CoV-2 bNAb assay, ELISA, cPass sVNT assay,

pseudovirus-based neutralization assay, and live virus based

neutralization assay were determined using Spearman correlation

coefficients. Spearman correlation analysis was performed using the

stat_cor function implemented in R package ggpubr (version 0.6.1).

erum NAbs with a p < 0.05 considered as statistically significant. To

assess the reliability of the NAb results obtained from the SARS-CoV-2

bNAb assay, ROC analysis was conducted using roc function

implemented in R package pROC (v1.18.2) based on the positive and

negative judgment results from ELISA and cPass sVNT assay[44].

Youden index was used in conjunction with the ROC curve to

determine the optimal threshold through coords function in pROC

(v1.18.2) package. Also, the PPA and NPA between SARS-CoV-2

bNAb assay and cPass sVNT assay were calculated based on the

ideal cutoffs from ROC curves. All data analyses were carried out in R

(v4.2.3) under the R Studio (v19.1.3) environment.

SUPPORTING INFORMATION

The following is the Supporting Information related to this article:

Schematic workflow of the SARS-CoV-2 bNAb assay; Spike

trimer-ACE2 interactions across six SARS-CoV-2 VOCs and ACE2

concentrations on different platforms; ACE2 binding to SARS-CoV-2

variants, SARS-CoV, MERS Spike; ROC curves for SARS-CoV-2

bNAb assay, ELISA and cPass sVNT assay; comparison of

SARS-CoV-2 bNAb assay to other assay; comparison of NAbs from

inactivated vs. recombinant subunit vaccines; serum NAb level

difference between vaccinated and unvaccinated groups; lists of:

SARS-CoV-2 variant Spike/N proteins, EasyMagPlex beads, coupled

proteins/bead codes for Spike-ACE2 & serum NAbs detection, coupled

proteins/bead codes for monoclonal/polyclonal NAbs detection,

SARS-CoV-2 RBD antibodies, SNR for ACE2 detection on

Luminex/Wellgrow (PDF).

REFERENCES

(1)   The, L. The COVID-19 pandemic in 2023: far from over. *Lancet* **2023**, *401*, 79.

(2)   Chung, Y.-S.; Lam, C.-Y.; Tan, P.-H.; Tsang, H.-F.; Wong, S.-C. C. Comprehensive Review of
COVID-19: Epidemiology, Pathogenesis, Advancement in Diagnostic and Detection
Techniques, and Post-Pandemic Treatment Strategies. *International Journal of Molecular*
Sciences **2024**, *25*.

(3)   Ma, Y.; Deng, J.; Liu, Q.; Du, M.; Liu, M.; Liu, J. Long-Term Consequences of
Asymptomatic SARS-CoV-2 Infection: A Systematic Review and Meta-Analysis. *Int J Environ*
Res Public Health **2023**, *20*.

(4)   Thaweethai, T.; Jolley, S. E.; Karlson, E. W.; Levitan, E. B.; Levy, B.; McComsey, G. A.;
McCorkell, L.; Nadkarni, G. N.; Parthasarathy, S.; Singh, U.; Walker, T. A.; Selvaggi, C. A.;
Shinnick, D. J.; Schulte, C. C. M.; Atchley-Challenner, R.; Horwitz, L. I.; Foulkes, A. S.; Aberg, J.
A.; Adolphi, N. L.; Ahirwar, S.; Ahmed, S.; Ahuja, N.; Aikawa, M.; Akerlundh, A.; Akintonwa, T.
T.; Al-Jadiri, A.; Alekhina, N.; Algren, H. A.; Alshawabkeh, A. N.; Ammar, N.; Anand, A.;
Anderson, B. R.; Aponte-Soto, L.; Aschner, J. L.; Atha, M. M.; Atz, A. M.; Aupperle, R. L.;

Ayache, M.; Azziz-Baumgartner, E.; Bailey, L. C.; Baker, F. C.; Balaraman, V.; Bandy, J. A.; Banerjee, D.; Barch, D. M.; Bardes, J. M.; Barlocker, J.; Barr, R. G.; Baskin-Sommers, A.; Basu, S.; Battaglia, T. A.; Baucom, L.; Beamon, C. J.; Beaty, C. L.; Beckmann, N. D.; Berry, J. A.; Bhadelia, N.; Bhargava, D.; Bhuiyan, S.; Bian, J.; Bime, C.; Bjork, J. M.; Black, L. J.; Blish, C. A.; Block, J. P.; Bogie, A.; Bolliger, D.; Bonaventura, W.; Bose-Brill, S.; Bower, M. B.; Boyd, A. D.; Boyineni, J.; Bradfute, S. B.; Bramante, C. T.; Brannock, M. D.; Bremner, J. D.; Brosnahan, S. B.; Buchbinder, N. C.; Bueler, E.; Buhimschi, I. A.; Bukulmez, H.; Bunnell, H. T.; Buse, J. B.; Calhoun, E. A.; Cao, T.; Carrithers, M. D.; Carton, T. W.; Case, A.; Casey, B. J.; Casimero, F. V. C.; Castro, L.; Cato, T.; Ceger, P.; Cerullo, C. L.; Chang, L.; Chang, A. A.; Chebrolu, P.; Chen, Y.; Chen, L. Q.; Chen, B. K. Development of a Definition of Postacute Sequelae of SARS-CoV-2 Infection. *Jama* **2023**, *329*.

(5)   Mühlemann, B.; Wilks, S. H.; Baracco, L.; Bekliz, M.; Carreño, J. M.; Corman, V. M.; Davis-Gardner, M. E.; Dejnirattisai, W.; Diamond, M. S.; Douek, D. C.; Drosten, C.; Eckerle, I.; Edara, V.-V.; Ellis, M.; Fouchier, R. A. M.; Frieman, M.; Godbole, S.; Haagmans, B.; Halfmann, P. J.; Henry, A. R.; Jones, T. C.; Katzelnick, L. C.; Kawaoka, Y.; Kimpel, J.; Krammer, F.; Lai, L.; Liu, C.; Lusvarghi, S.; Meyer, B.; Mongkolsapaya, J.; Montefiori, D. C.; Mykytyn, A.; Netzl, A.; Pollett, S.; Rössler, A.; Screaton, G. R.; Shen, X.; Sigal, A.; Simon, V.; Subramanian, R.; Supasa, P.; Suthar, M. S.; Türeli, S.; Wang, W.; Weiss, C. D.; Smith, D. J. Comparative analysis of SARS-CoV-2 neutralization titers reveals consistency between human and animal model serum and across assays. *Science Translational Medicine* **2024**, *16*.

(6)   Sun, K.; Bhiman, J. N.; Tempia, S.; Kleynhans, J.; Madzorera, V. S.; Mkhize, Q.; Kaldine, H.; McMorrow, M. L.; Wolter, N.; Moyes, J.; Carrim, M.; Martinson, N. A.; Kahn, K.; Lebina, L.; du Toit, J. D.; Mkhencele, T.; von Gottberg, A.; Viboud, C.; Moore, P. L.; Cohen, C.; Buys, A.; Carrim, M.; de Gouveia, L.; du Plessis, M.; du Toit, J.; Gómez-Olivé, F. X.; Kgasago, K. P.; Kotane, R.; McMorrow, M. L.; Moloantoa, T.; Tollman, S.; von Gottberg, A.; Wafawanaka, F.; Wolter, N. SARS-CoV-2 correlates of protection from infection against variants of concern. *Nature Medicine* **2024**, *30*, 2805-2812.

(7)   Russell, T. W.; Townsley, H.; Hellewell, J.; Gahir, J.; Shawe-Taylor, M.; Greenwood, D.; Hodgson, D.; Hobbs, A.; Dowgier, G.; Penn, R.; Sanderson, T.; Stevenson-Leggett, P.; Bazire, J.; Harvey, R.; Fowler, A. S.; Miah, M.; Smith, C.; Miranda, M.; Bawumia, P.; Mears, H. V.; Adams, L.; Hatipoglu, E.; O'Reilly, N.; Warchal, S.; Ambrose, K.; Strange, A.; Kelly, G.; Kjar, S.; Papineni, P.; Corrah, T.; Gilson, R.; Libri, V.; Kassiotis, G.; Gamblin, S.; Lewis, N. S.; Williams, B.; Swanton, C.; Gandhi, S.; Beale, R.; Wu, M. Y.; Bauer, D. L. V.; Carr, E. J.; Wall, E. C.; Kucharski, A. J. Real-time estimation of immunological responses against emerging SARS-CoV-2 variants in the UK: a mathematical modelling study. *The Lancet Infectious Diseases* **2025**, *25*, 80-93.

(8)   Nie, J.; Li, Q.; Wu, J.; Zhao, C.; Hao, H.; Liu, H.; Zhang, L.; Nie, L.; Qin, H.; Wang, M.; Lu, Q.; Li, X.; Sun, Q.; Liu, J.; Fan, C.; Huang, W.; Xu, M.; Wang, Y. Quantification of SARS-CoV-2 neutralizing antibody by a pseudotyped virus-based assay. *Nature Protocols* **2020**, *15*, 3699-3715.

(9)   Tan, C. W.; Chia, W. N.; Qin, X.; Liu, P.; Chen, M. I. C.; Tiu, C.; Hu, Z.; Chen, V. C.-W.; Young, B. E.; Sia, W. R.; Tan, Y.-J.; Foo, R.; Yi, Y.; Lye, D. C.; Anderson, D. E.; Wang, L.-F. A SARS-CoV-2 surrogate virus neutralization test based on antibody-mediated blockage of ACE2–spike protein–protein interaction. *Nature Biotechnology* **2020**, *38*, 1073-1078.

(10) Fenwick, C.; Turelli, P.; Pellaton, C.; Farina, A.; Campos, J.; Raclot, C.; Pojer, F.; Cagno, V.;

Nusslé, S. G.; D'Acremont, V.; Fehr, J.; Puhan, M.; Pantaleo, G.; Trono, D. A high-throughput cell- and virus-free assay shows reduced neutralization of SARS-CoV-2 variants by COVID-19 convalescent plasma. *Science Translational Medicine* **2021**, *13*.

(11) Lynch, K. L.; Zhou, S.; Kaul, R.; Walker, R.; Wu, A. H. Evaluation of Neutralizing Antibodies against SARS-CoV-2 Variants after Infection and Vaccination Using a Multiplexed Surrogate Virus Neutralization Test. *Clin Chem* **2022**, *68*, 702-712.

(12) Perkmann, T.; Mucher, P.; Ösze, D.; Müller, A.; Perkmann-Nagele, N.; Koller, T.; Radakovics, A.; Flieder, I.; Repl, M.; Marculescu, R.; Wolzt, M.; Wagner, O. F.; Binder, C. J.; Haslacher, H. Comparison of five Anti-SARS-CoV-2 antibody assays across three doses of BNT162b2 reveals insufficient standardization of SARS-CoV-2 serology. *Journal of Clinical Virology* **2023**, *158*.

(13) Bal, A.; Pozzetto, B.; Trabaud, M.-A.; Escuret, V.; Rabilloud, M.; Langlois-Jacques, C.; Paul, A.; Guibert, N.; D'Aubarède-Frieh, C.; Massardier-Pilonchery, A.; Fabien, N.; Goncalves, D.; Boibieux, A.; Morfin-Sherpa, F.; Pitiot, V.; Gueyffier, F.; Lina, B.; Fassier, J.-B.; Trouillet-Assant, S. Evaluation of High-Throughput SARS-CoV-2 Serological Assays in a Longitudinal Cohort of Patients with Mild COVID-19: Clinical Sensitivity, Specificity, and Association with Virus Neutralization Test. *Clinical Chemistry* **2021**, *67*, 742-752.

(14) Grant, R.; Sacks, J.; Abraham, P.; Chunsuttiwat, S.; Cohen, C.; Figueroa, J. When to update COVID-19 vaccine composition. *Nature Medicine* **2023**, 776–780.

(15) Chao, Z.; Han, Y.; Jiao, Z.; You, Z.; Zhao, J. Prism Design for Spectral Flow Cytometry. *Micromachines* **2023**, *14*.

(16) Hoheisel, J.; Yu, X.; Hartmann, M.; Wang, Q.; Poetz, O.; Schneiderhan-Marra, N.; Stoll, D.; Kazmaier, C.; Joos, T. O. µFBI: A Microfluidic Bead-Based Immunoassay for Multiplexed Detection of Proteins from a µL Sample Volume. *PLoS ONE* **2010**, *5*.

(17) Wang, Q.; Guo, Y.; Iketani, S.; Nair, M. S.; Li, Z.; Mohri, H.; Wang, M.; Yu, J.; Bowen, A. D.; Chang, J. Y.; Shah, J. G.; Nguyen, N.; Chen, Z.; Meyers, K.; Yin, M. T.; Sobieszczyk, M. E.; Sheng, Z.; Huang, Y.; Liu, L.; Ho, D. D. Antibody evasion by SARS-CoV-2 Omicron subvariants BA.2.12.1, BA.4 and BA.5. *Nature* **2022**, *608*, 603-608.

(18) Mahalingam, G.; Arjunan, P.; Periyasami, Y.; Dhyani, A. K.; Devaraju, N.; Rajendiran, V.; Christopher, A. C.; Kt, R. D.; Dhanasingh, I.; Thangavel, S.; Murugesan, M.; Moorthy, M.; Srivastava, A.; Marepally, S. Correlating the differences in the receptor binding domain of SARS-CoV-2 spike variants on their interactions with human ACE2 receptor. *Scientific Reports* **2023**, *13*.

(19) Gorbalenya, A. E.; Baker, S. C.; Baric, R. S.; de Groot, R. J.; Drosten, C.; Gulyaeva, A. A.; Haagmans, B. L.; Lauber, C.; Leontovich, A. M.; Neuman, B. W.; Penzar, D.; Perlman, S.; Poon, L. L. M.; Samborskiy, D. V.; Sidorov, I. A.; Sola, I.; Ziebuhr, J. The species Severe acute respiratory syndrome-related coronavirus: classifying 2019-nCoV and naming it SARS-CoV-2. *Nature Microbiology* **2020**, *5*, 536-544.

(20) Zhan, H.; Gao, H.; Liu, Y.; Zhang, X.; Li, H.; Li, X.; Wang, L.; Li, C.; Li, B.; Wang, Y.; Dai, E.; Li, Y. Booster shot of inactivated SARS-CoV-2 vaccine induces potent immune responses in people living with HIV. *Journal of Medical Virology* **2023**, *95*.

(21) Mariën, J.; Michiels, J.; Heyndrickx, L.; Nkuba-Ndaye, A.; Ceulemans, A.; Bartholomeeusen, K.; Madinga, J.; Mbala-Kingebeni, P.; Vanlerberghe, V.; Ahuka-Mundeke,

S.; Wang, L.-F.; Ariën, K. K. Evaluation of a surrogate virus neutralization test for high-throughput serosurveillance of SARS-CoV-2. *Journal of Virological Methods* **2021**, *297*.

(22) Chen, Y.; Zhao, X.; Zhou, H.; Zhu, H.; Jiang, S.; Wang, P. Broadly neutralizing antibodies to SARS-CoV-2 and other human coronaviruses. *Nature Reviews Immunology* **2022**, *23*, 189-199.

(23) Liang, T.; Cheng, M.; Teng, F.; Wang, H.; Deng, Y.; Zhang, J.; Qin, C.; Guo, S.; Zhao, H.; Yu, X. Proteome-wide epitope mapping identifies a resource of antibodies for SARS-CoV-2 detection and neutralization. *Signal Transduction and Targeted Therapy* **2021**, *6*.

(24) Zhao, X.; Zhang, R.; Qiao, S.; Wang, X.; Zhang, W.; Ruan, W.; Dai, L.; Han, P.; Gao, G. F. Omicron SARS-CoV-2 Neutralization from Inactivated and ZF2001 Vaccines. *New England Journal of Medicine* **2022**, *387*, 277-280.

(25) Zhang, J.; Teng, F.; Zhang, X.; Wang, H.; Liang, T.; Guo, S.; Yu, X. Down-regulation of SARS-CoV-2 neutralizing antibodies in vaccinated smokers. *MedComm (2020)* **2022**, *3*, e166.

(26) Pan, Y.; Wang, L.; Feng, Z.; Xu, H.; Li, F.; Shen, Y.; Zhang, D.; Liu, W. J.; Gao, G. F.; Wang, Q. Characterisation of SARS-CoV-2 variants in Beijing during 2022: an epidemiological and phylogenetic analysis. *The Lancet* **2023**, *401*, 664-672.

(27) Faraone, J. N.; Qu, P.; Goodarzi, N.; Zheng, Y.-M.; Carlin, C.; Saif, L. J.; Oltz, E. M.; Xu, K.; Jones, D.; Gumina, R. J.; Liu, S.-L. Immune evasion and membrane fusion of SARS-CoV-2 XBB subvariants EG.5.1 and XBB.2.3. *Emerging Microbes & Infections* **2023**, *12*.

(28) Antinori, A.; Bausch-Jurken, M. The Burden of COVID-19 in the Immunocompromised Patient: Implications for Vaccination and Needs for the Future. *The Journal of Infectious Diseases* **2023**, *228*, S4-S12.

(29) Lotfi, H.; Mazar, M. G.; Ei, N. M. H.; Fahim, M.; Yazdi, N. S. Vaccination is the most effective and best way to avoid the disease of COVID-19. *Immunity, Inflammation and Disease* **2023**, *11*.

(30) Yin, W.; Xu, Y.; Xu, P.; Cao, X.; Wu, C.; Gu, C.; He, X.; Wang, X.; Huang, S.; Yuan, Q.; Wu, K.; Hu, W.; Huang, Z.; Liu, J.; Wang, Z.; Jia, F.; Xia, K.; Liu, P.; Wang, X.; Song, B.; Zheng, J.; Jiang, H.; Cheng, X.; Jiang, Y.; Deng, S.-J.; Xu, H. E. Structures of the Omicron spike trimer with ACE2 and an anti-Omicron antibody. *Science* **2022**, *375*, 1048-1053.

(31) Ma, W.; Fu, H.; Jian, F.; Cao, Y.; Li, M. Immune evasion and ACE2 binding affinity contribute to SARS-CoV-2 evolution. *Nature Ecology & Evolution* **2023**, *7*, 1457-1466.

(32) Markov, P. V.; Katzourakis, A.; Stilianakis, N. I. Antigenic evolution will lead to new SARS-CoV-2 variants with unpredictable severity. *Nat Rev Microbiol* **2022**, *20*, 251-252.

(33) Zhang, X.; Li, M.; Zhang, N.; Li, Y.; Teng, F.; Li, Y.; Zhang, X.; Xu, X.; Li, H.; Zhu, Y.; Wang, Y.; Jia, Y.; Qin, C.; Wang, B.; Guo, S.; Wang, Y.; Yu, X. SARS-CoV-2 Evolution: Immune Dynamics, Omicron Specificity, and Predictive Modeling in Vaccinated Populations. *Advanced Science* **2024**, *11*.

(34) Li, Y.; Zhang, X.; Yi, J.; Chen, Y.; Liang, J.; Wang, L.; Ma, J.; Zhu, R.; Zhang, X.; Hu, D.; Jia, Y.; Yu, X.; Wang, Y. Synergistic evolution: The dynamic adaptation of SARS-CoV-2 and human protective immunity in the real world. *Journal of Infection* **2024**, *89*.

(35) Valiate, B. V. S.; Castro, J. T. d.; Marçal, T. G.; Andrade, L. A. F.; Oliveira, L. I. d.; Maia, G. B. F.; Faustino, L. P.; Hojo-Souza, N. S.; Reis, M. A. A. D.; Bagno, F. F.; Salazar, N.; Teixeira, S. R.;

Almeida, G. G.; Gazzinelli, R. T. Evaluation of an RBD-nucleocapsid fusion protein as a
booster candidate for COVID-19 vaccine. *iScience* **2024**, *27*.
(36) Liu, Y.; Han, J.; Li, X.; Chen, D.; Zhao, X.; Qiu, Y.; Zhang, L.; Xiao, J.; Li, B.; Zhao, H.
COVID-19 Vaccination in People Living with HIV (PLWH) in China: A Cross Sectional Study of
Vaccine Hesitancy, Safety, and Immunogenicity. *Vaccines* **2021**, *9*.
(37) Yurkovetskiy, L.; Wang, X.; Pascal, K. E.; Tomkins-Tinch, C.; Nyalile, T. P.; Wang, Y.; Baum,
A.; Diehl, W. E.; Dauphin, A.; Carbone, C.; Veinotte, K.; Egri, S. B.; Schaffner, S. F.; Lemieux, J.
E.; Munro, J. B.; Rafique, A.; Barve, A.; Sabeti, P. C.; Kyratsous, C. A.; Dudkina, N. V.; Shen, K.;
Luban, J. Structural and Functional Analysis of the D614G SARS-CoV-2 Spike Protein Variant.
Cell **2020**, *183*, 739-751.e738.
(38) Hou, Y. J.; Chiba, S.; Halfmann, P.; Ehre, C.; Kuroda, M.; Dinnon, K. H.; Leist, S. R.;
Schäfer, A.; Nakajima, N.; Takahashi, K.; Lee, R. E.; Mascenik, T. M.; Graham, R.; Edwards, C.
E.; Tse, L. V.; Okuda, K.; Markmann, A. J.; Bartelt, L.; de Silva, A.; Margolis, D. M.; Boucher, R.
C.; Randell, S. H.; Suzuki, T.; Gralinski, L. E.; Kawaoka, Y.; Baric, R. S. SARS-CoV-2 D614G
variant exhibits efficient replication ex vivo and transmission in vivo. *Science* **2020**, *370*,
1464-1468.
(39) Zhang, X.; Zheng, M.; Wang, H.; Zhou, H.; Liang, T.; Zhang, J.; Ren, J.; Peng, H.; Li, S.;
Bian, H.; Wei, C.; Yin, S.; He, C.; Han, Y.; Li, M.; Hou, X.; Zhang, J.; Xie, L.; Lv, J.; Kan, B.; Wang,
Y.; Yu, X. Inhibitor screening using microarray identifies the high capacity of neutralizing
antibodies to Spike variants in SARS-CoV-2 infection and vaccination. *Theranostics* **2022**, *12*,
2519-2534.
(40) Yang, L.; Yu, Y.; Ma, C.; Wang, H.; Dai, J.; Duan, H.; Fu, Z.; Wu, P.; Wang, D.; Yu, X.
Development of RBC Membrane Antigen Arrays for Validating Blood Grouping Reagents.
Journal of Proteome Research **2018**, *17*, 3237-3245.
(41) Salazar-García, M.; Acosta-Contreras, S.; Rodríguez-Martínez, G.; Cruz-Rangel, A.;
Flores-Alanis, A.; Patiño-López, G.; Luna-Pineda, V. M. Pseudotyped Vesicular Stomatitis
Virus-Severe Acute Respiratory Syndrome-Coronavirus-2 Spike for the Study of Variants,
Vaccines, and Therapeutics Against Coronavirus Disease 2019. *Frontiers in Microbiology*
2022, *12*.
(42) Nie, J.; Li, Q.; Wu, J.; Zhao, C.; Hao, H.; Liu, H.; Zhang, L.; Nie, L.; Qin, H.; Wang, M.; Lu,
Q.; Li, X.; Sun, Q.; Liu, J.; Fan, C.; Huang, W.; Xu, M.; Wang, Y. Establishment and validation of
a pseudovirus neutralization assay for SARS-CoV-2. *Emerging Microbes & Infections* **2020**, *9*,
680-686.
(43) Chen, G.-L.; Li, X.-F.; Dai, X.-H.; Li, N.; Cheng, M.-L.; Huang, Z.; Shen, J.; Ge, Y.-H.; Shen,
Z.-W.; Deng, Y.-Q.; Yang, S.-Y.; Zhao, H.; Zhang, N.-N.; Zhang, Y.-F.; Wei, L.; Wu, K.-Q.; Zhu,
M.-F.; Peng, C.-G.; Jiang, Q.; Cao, S.-C.; Li, Y.-H.; Zhao, D.-H.; Wu, X.-H.; Ni, L.; Shen, H.-H.;
Dong, C.; Ying, B.; Sheng, G.-P.; Qin, C.-F.; Gao, H.-N.; Li, L.-J. Safety and immunogenicity of
the SARS-CoV-2 ARCoV mRNA vaccine in Chinese adults: a randomised, double-blind,
placebo-controlled, phase 1 trial. *The Lancet Microbe* **2022**, *3*, e193-e202.
(44) Robin, X.; Turck, N.; Hainard, A.; Tiberti, N.; Lisacek, F.; Sanchez, J.-C.; Müller, M. pROC:
an open-source package for R and S+ to analyze and compare ROC curves. *BMC*
Bioinformatics **2011**, *12*.

CONTRIBUTIONS

X.Z., and D.H. executed the experiments; M.L., X. Z., X.Z., X.X., T. L., Y. J., and Y.Z. performed the data analysis; H.L., Y.W., X.H., and Q.M. provided the clinical samples. X.Y., Y. L., B.W., X.Z., S.B., S.S., and M.F. drafted, wrote and revised the paper.

ACKNOWLEDGMENT

This work was supported by the National Key R&D Program of China (2024YFA1307600, 2024YFA1307601, 2022YFE0210400, 2021YFA1301604), Beijing Municipal International Cooperation and Exchange Project (Z231100002723004), Beijing Municipal Natural Science Foundation (M23010, L234034), Capital Health Development Research Project (2024-1-1201 and 2024-1-4093), State Key Laboratory of Proteomics (SKLP-O202205, SKLP-O202007), Guangdong Province Science and Technology Planning Project (2020B1111100006), Guangdong-PHOENIX Center Joint Lab on Chinese Medicine (YN2021DB06), Innovation Team and Talents Cultivation Program of National Administration of Traditional Chinese Medicine (ZYYCXTD-C-202204). We would like to thank the bioinformatics platform of the National Center for Protein Sciences (Beijing) for their support in the data analysis of this project. We also thank Dr. Brianne Petritis for her critical review and editing of this

966   manuscript.

DECLARATION OF INTEREST STATEMENT

969   The authors have declared no conflict of interest.

TABLES

972   Table 1. Clinical serum samples of 56 healthy individuals receiving

973   third-dose vaccines employed in this study.

	3rd dose vaccination cohort （n=56）
Gender	
Male	16(28.6%)
Female	40(71.4%)
Age (years)	
Range	21-70
≤50	50(89.3%)
>50	6(10.7%)
Vaccination type	
Recombinant RBD-subunit vaccine (ZF2001)	38(83.9%)
Inactivated vaccine (Sinovac-Coron aVac)	18(16.1%)
Days of serum collection after 3rd dose vaccination	28

975   Table 2. Clinical serum samples of 35 HIV patients with breakthrough

976   infection of COVID-19 employed in this study.

	CD4+ > 500 cells/µL （ n = 19 ）	CD4+ < 350 cells/µL （ n = 16 ）

Gender			
	Male	19 (100%)	16 (100%)
	Female	0	0
Age (years)			
	Range	22-50	25-72
	≤50	19 (100%)	12 (75.0%)
	>50	0	4 (25.0%)
Vaccination type			
Recombinant RBD-sub unit vaccine		0	0
Inactivated vaccine		19 (100%)	19 (100%)
Number of vaccinations		3	3
Range of infection to blood collection time (days)		107-182	94-176
Range of CD4+ T cell counts (cells/uL)		511-975	124-347

978 **FIGURES AND LEGENDS**

Figure 1

A

B

C

LOD[μg/mL]	D614G	Alpha	Beta	Delta	Kappa	Omicron BA.1	Average
Luminex	0.19	0.12	0.16	0.23	0.14	0.12	0.16
Wellgrow	0.006	0.003	0.005	0.003	0.007	0.008	0.005
Fold change [L/W] (×)	31.67	40	32	76.67	20	15	35.89±57.80 [-21.91~93.69]

Figure 1. Schematic illustration of the high-throughput SARS-CoV-2 bNAb assay. (A) Workflow of the SARS-CoV-2 bNAb assay. The Spike-ACE2 interaction is a direct binding reaction between Spike trimer proteins and biotinylated ACE2. Upon the addition of the NAbs, the beads coupled with SARS-CoV-2 variants Spike trimer

proteins are incubated with NAbs. The NAbs with neutralizing activity

will block the Spike interaction with biotinylated ACE2 and SA-PE. The

signal intensities of the Spike trimer-ACE2 interaction are inversely

proportional to the level of NAbs. During the NAbs detection process,

all coupled beads were mixed together and incubated with samples

containing NAbs. The NAbs bound to the Spike trimer proteins,

preventing their interaction with the biotinylated ACE2 receptors, which

were coupled to SA-PE. (B) Dose relationship curve of Spike-ACE2

interactions using Wellgrow and Luminex platforms. The x-axis

represents the concentrations of ACE2. The y-axis represents the MFI

of the SARS-CoV-2 Spike trimer-ACE2 interaction. (C) A comparison of

the sensitivity to detect Spike-ACE2 interactions with the Wellgrow and

Luminex platforms. N protein, nucleocapsid protein. bNAb, broad

neutralizing antibody. ACE2, angiotensin-converting enzyme 2. SA-PE,

streptavidin-phycoerythrin. MFI, median fluorescence intensity. LOD,

lowest detection limit.

1001

Figure 2

Figure 2. Spike trimer-ACE2 interactions across different SARS-CoV-2 variants and ACE2 concentrations. (A) Dose relationship curve of the interactions between ACE2 and six SARS-CoV-2 non-Omicron variants and five Omicron variants. The x-axis represents the concentrations of the host ACE2 receptor. The

y-axis represents the MFI of the SARS-CoV-2 Spike trimer-ACE2

interaction. (B) The level of binding Affinity of SARS-CoV-2 variant

Spike-ACE2 interactions. The x-axis represents the SARS-CoV-2

variants. The y-axis represents the EC50 calculated from SARS-CoV-2

variant Spike-ACE2 interactions. N protein, nucleocapsid protein.

angiotensin-converting enzyme 2. MFI, median fluorescence intensity.

EC50, half-maximal effective concentration.

Figure 3

Figure 3. Comparison of SARS-CoV-2 bNAb assay to ELISA , cPass sVNT, pseudovirus-based neutralization and live virus based neutralization assays. (A) Correlation between NAbs inhibition rate (%) from the SARS-CoV-2 bNAb assay and the levels of anti-Spike IgG antibodies obtained by the ELISA in 75 serum samples obtained

approximately 180-270 days after the third dose of COVID-19 vaccine.

(B) Correlation between the levels of NAbs obtained by SARS-CoV-2 bNAb assay and the FDA-approved cPass sVNT assay in 75 serum samples obtained approximately 180-270 days after the third dose of COVID-19 vaccine. The concordance of NAb detection using the ELISA and sVNT assays was calculated as the PPA and NPA, respectively. The blue and red dots represent positive (neutralizing activity) and negative (non-neutralizing) cases as determined by the comparative analysis, respectively. (C-D) Correlation between NAbs inhibition rate (%) from the SARS-CoV-2 bNAb assay and NAb titers from pseudovirus-based neutralization assays in serum samples from mRNA-vaccinated mice (n = 20). (E-F) Correlation between NAbs inhibition rate (%) from the SARS-CoV-2 bNAb assay and NAb titers from live virus based neutralization assays in serum samples from mRNA-vaccinated mice (n = 20). N protein, nucleocapsid protein. bNAb, broad neutralizing antibody. WT, wild type. RBD, receptor-binding domain. PPA, positive percent agreement. NPA, negative percent agreement.

Figure 4

1039

Figure 4. Identification of monoclonal/polyclonal antibodies with broad neutralizing capability. (A) Dose relationship curve demonstrating the ability of the NAb to inhibit SARS-CoV-2 variant Spike ACE2 interactions. The x-axis represents the concentrations of the NAb. The y-axis represents the percent activity (%) of the NAb. (B)

1045 Heatmap showing the IC50 of NAbs against different SARS-CoV-2

1046 variants. The detail of monoclonal/polyclonal **antibodies** are provided

1047 in Table S5. N protein, nucleocapsid protein. NAb, neutralizing antibody.

1048 IC50, half-maximal inhibitory concentration.

1049

1050

Figure 5. Detection of serum NAbs in vaccinated healthy

individuals and HIV patients with breakthrough infection of

COVID-19. (A) Boxplot showing the distribution of inhibition rate (%) of

serum NAbs against SARS-CoV-2 variants in 56 vaccinated individuals.

The significance was performed using the Student's *t*-test. (B) Boxplot

showing the distribution of inhibition rate (%) of serum NAbs against

SARS-CoV-2 variants in 35 HIV patients infected with Omicron BA.5.2

and BF.7 variants, incluing 19 immunological responders (CD4+ T cell

counts > 500 cells/μL) and 16 poor immunological responders (CD4+ T

cell counts < 350 cells/μL). NAb, neutralizing antibody. Asterisks

denote statistical differences (*p < 0.05, **p < 0.01, ***p < 0.001, ****p

<0.0001, ns: not signiffcant).