**Title: Foundation versus Domain-Specific Models for Cardiac Ultrasound Segmentation**

**Authors:** Chieh-Ju Chao, MD[1,2], Yunqi Richard Gu[2], Wasan Kumar, Tiange Xiang, BS[2], Lalith Appari, BS[3,4], Justin Wu, BS[2], Juan M. Farina, MD[5], Rachael Wraith, RDCS[5], Jiwoon Jeong, MS[3,4], Reza Arsanjani, MD[5], Garvan C. Kane, MD[1], Jae K. Oh, MD[1], Curtis P. Langlotz, MD, PhD[2], Imon Banerjee, PhD[3,4] , Li Fei-Fei[†], Ph.D.[2], Ehsan Adeli[†], Ph.D.[2]

[†]Li Fei-Fei and Ehsan Adeli are co-corresponding authors on this manuscript.

**Affiliations**:

1. Department of Cardiovascular Medicine, Mayo Clinic Rochester, Rochester, MN
2. Stanford Institute for Human-Centeral Artificial Intelligence, Stanford University, Stanford, CA
3. Department of Radiology, Mayo Clinic Arizona, Scottsdale, AZ
4. School of Computing and Augmented Intelligence, Arizona State University, Phoenix, AZ
5. Department of Cardiovascular Medicine, Mayo Clinic Arizona, Scottsdale, AZ

**Abbreviations:**
AI:  Artificial intelligence
A2C: apical 2 chamber, echocardiography view
A4C: apical 4 chamber, echocardiography view
CAMUS: Cardiac Acquisitions for Multi-structure Ultrasound Segmentation
IoU: Intersection over the union
DSC: Dice similarity coefficient
LV: Left ventricle
POCUS: Point-of-care ultrasound
SAM: Segment anything model
TTE: Transthoracic echocardiography
Vision Transformer: ViT

**Disclosure:** The authors have no conflict of interest to disclose.

**Corresponding Author**:
Ehsan Adeli, Ph.D., Assistant Professor
Department of Psychiatry and Behavioral Sciences
Email: eadeli@stanford.edu
Address: MC#5590, 1070 Arastradero Rd., Palo Alto, CA 94304

**Abstract**

**Background**

Vision foundation model, "Segment Anything (SAM)," promises to segment any objects in images. However, the performance of SAM on clinical echocardiography images has yet to be investigated and compared against state-of-the-art models.

**Method**

SAM was fine-tuned on the training set of EchoNet-Dynamic (Stanford) and then evaluated on external datasets containing transthoracic echocardiography (TTE) and Point-of-Care Ultrasound (POCUS) images, including CAMUS (University Hospital of St Etienne), and the Mayo Clinic dataset (a sample of 99 non-duplicated patients with 58 TTE and 41 POCUS). Fine-tuned SAM was evaluated against the EchoNet and MedSAM models using the Dice similarity coefficient (DSC). We further conducted an annotator study to evaluate the effectiveness of SAM in assisting clinical segmentation tasks.

**Results**

Fine-tuned SAM was superior to EchoNet and MedSAM in most of the datasets. We observed a strong generalization capacity of the fine-tuned SAM model against EchoNet, especially on apical 2 chamber (A2C) images (CAMUS-A2C: DSC $0.891 \pm 0.040$ vs. $0.752 \pm 0.196$, p<0.0001) and POCUS (DSC $0.857 \pm 0.047$ vs. $0.667 \pm 0.279$, p<0.0001). SAM also reduced the annotation time by 50% ($11.6 \pm 4.5$ sec vs. $5.7 \pm 1.7$ sec, p< 0.0001) while maintaining the segmentation quality.

**Conclusions**

Our approach demonstrates an effective strategy for fine-tuning a vision foundation model, enhancing clinical workflow efficiency through human-artificial intelligence (AI) collaboration, and leveraging its generalization capacity across ultrasound modalities.

**Introduction**

Echocardiography provides a comprehensive anatomy and physiology assessment of the heart and is one of the most widely available imaging modalities in the field of Cardiology given its non-radiative, safe, and low-cost nature[1–3]. Cardiac chamber quantification is one of the fundamental tasks of echocardiography studies in the current practice[4], and the results can have direct effects on clinical decisions such as the management of heart failure, valvular heart diseases, and chemotherapy-induced cardiomyopathy [5–8]. Although left ventricular (LV) chamber quantification tasks are performed by trained sonographers or physicians, it is known to be subject to intra- and inter-observer variance which can be up to 7%-13% across studies[9–12]. Many artificial intelligence (AI) applications have been applied to address this essential task and to minimize variations[3,13–18].

While established AI systems are potential solutions, the training of segmentation AI models requires large amounts of training data and their corresponding expert-defined annotations, making them challenging and costly to implement[19]. In recent years, transformers, a type of neural network architecture with self-attention, have revolutionized the field from natural language processing to computer vision[20–22]. Vision transformers (ViT) [20] are a type of transformer specifically designed for images, which have shown impressive performance with simple image patches and have become a popular choice for the building of foundation models that can be fine-tuned for various applications[23]. Building on this success, Meta AI introduced the "Segment Anything Model" (SAM), a foundation large vision model that was trained on diverse datasets and that can adapt to specific tasks. This model achieves "zero-shot" segmentation, i.e., segments user-specified objects at different data resources without needing any training data[24].

However, while the zero-shot performance of SAM on natural image datasets has been promising[24], its performance on complex image datasets, such as medical images, has not been fully investigated. While not specifically including echocardiography images, one study tested SAM on different medical image datasets including ultrasound, and the zero-shot performance was not optimal[25]. Recently, MedSAM was introduced as a universal tool for medical image segmentation, however, ultrasound or echocardiography were less represented in the training set[26]. In this context, we aim to study the zero-shot and fine-tuned performance of SAM in echocardiography images and compare them with MedSAM and a state-of-the-art segmentation model trained with a domain-specific dataset (EchoNet)[3].

We hypothesized that fine-tuning SAM can adapt it to the domain of echocardiography for better segmentation performance. We further designed a human annotation study on a subset of images to investigate whether the model can enhance the efficiency of echocardiography segmentation. Within the echocardiography domain, images obtained from different institutions and modalities (with different image qualities) were used to evaluate the generalization capability of SAM performance on external evaluation sets.

**Method**

*Population and Data Curation*

The EchoNet-Dynamic dataset is publicly available (https://echonet.github.io/dynamic/); details of the dataset have been described previously [3]. In brief, the dataset contains 10,030 apical-4-chamber (A4C) TTE videos at Stanford Health Care in the period of 2016-2018. The raw videos were preprocessed to remove patient identifiers and downsampled by cubic interpolation into standardized $112 \times 112$-pixel videos. Videos were randomly split into 7,465, 1,277, and 1,288 patients, respectively, for the training, validation, and test sets[3]. In this study, the cases without ground truth labels were excluded from this analysis (5 from the train set, 1 from the test set) (**Supplemental Table 1**).

A dataset from Mayo Clinic (Rochester, MN) that includes 99 randomly selected patients (58 TTE in 2017-2018 and 41 point-of-care ultrasound studies (POCUS) in 2022) was used as the external validation dataset. The A4C videos were reviewed by a clinical sonographer (RW) and a cardiologist (JMF). 52 (33 TTE and 19 POCUS) out of the total 99 cases were traced by both of the annotators to select and segment the end-diastolic and end-systolic frames. The tracings were done manually on commercially available software (MD.ai, Inc., NY). The Cardiac Acquisitions for Multi-structure Ultrasound Segmentation (CAMUS) dataset contains 500 cases from the University Hospital of St Etienne (France) with detailed tracings on both A4C and A2C views [27].

*Segment Anything Model (SAM) and MedSAM*

The Segment Anything Model (SAM) is an image segmentation foundation model trained on a dataset of 11 million images and 1.1 billion masks[28]. It can generate object masks from input prompts like points or boxes. SAM's promptable design enables zero-shot transfer to new image distributions and tasks, achieving competitive or superior performance compared to fully supervised methods [28]. In brief, the model comprises a VisionEncoder, PromptEncoder, MaskDecoder, and Neck module, which collectively process image embeddings, point embeddings, and contextualized masks to predict accurate segmentation masks [28]. MedSAM is a fine-tuned version of the base SAM, trained on over 1.5 million medical image-mask pairs across 10 modalities and 30 cancer types, with trained weight available [26].

*Data Preprocessing*

Each EchoNet-Dynamic video ($112 \times 112$ pixels, in avi format) was exported into individual frames without further resizing. End-diastolic and end-systolic frames of each case were extracted, which corresponded to the human expert traced, frame-level ground truth segmentation coordinates in the dataset. Ground truth segmentation masks were generated according to the coordinates and saved in the same size ($112 \times 112$ pixels). The labeled frames of Mayo TTE and POCUS images were exported from the MD.ai platform, and followed a similar preprocessing method to remove patient identifiers, then horizontally flipped and resized to $112 \times 112$ pixels[3]. The CAMUS images were rotated for 270 degrees and resized to be consistent with the EchoNet format. When importing to the SAM model, all the raw images were resized with the built-in function "ResizeLongestSide."[28]

*Zero-shot Evaluation*

The original SAM ViT-base model (model type "vit_b", checkpoint "sam_vit_b_01ec64.pth ") without modification or fine-tuning was used to segment the dataset, referred to as zero-shot learning in machine learning literature[26,28]. The larger versions of SAM (ViT Large and ViT Huge) were not used as they did not offer significant performance improvement despite higher computational demands[28]. Bounding box coordinates of each left ventricle segmentation tracing were generated from the ground truth segmentations and used as the prompt for SAM[24].

*Model Fine-tuning*

The same ViT-base model was used for fine-tuning with a procedure described by Ma et al. [29]. We used the training set cases (n=7,460) of the EchoNet-Dynamic as our customized dataset without data augmentation. The same bounding box was used as the prompt, as described above. We used a customized loss function, which is the unweighted sum of Dice loss and cross-entropy loss[29,30]. Adam optimizer[31] was used (weight decay= 0), with an initial learning rate of 2e-5 (decreased to 3e-6 over 27 epochs). The batch size was 8. The model was fine-tuned on a node on the Stanford AI Lab cluster with a 24 GB NVIDIA RTX TiTAN GPU.

*Validation and Generalization*

The fine-tuned SAM was tested on the test set of the EchoNet-Dynamic dataset. To test the generalization capacity of SAM, we used external validation samples from the CAMUS dataset (both A2C and A4C) [27] and a Mayo Clinic dataset including the A4C view of cases of TTE and POCUS devices.

*Human Annotation Study*

We conducted a human annotation study to evaluate the enhancement of segmentation efficiency and quality when utilizing a fine-tuned SAM. Three annotators (two experts and one medical student) independently annotated 100 echocardiography images, randomly selected from the EchoNet-Dynamic test set. Each one annotated 100 images with and without SAM assistance. In the SAM-assisted workflow, annotators were instructed to draw minimal bounding boxes to prompt the model. In the regular workflow, annotators manually trace the endocardium contour, as per standard clinical practice. Annotators were allowed to refine their annotations until they were satisfactory. We recorded and compared the time required to complete annotations, and the quality of annotations (Dice score against ground truth) using both workflows.

*Statistical Model Performance Evaluation*

The model segmentation performance was directly evaluated by Intersection over Union (IoU) and the Dice similarity coefficient (DSC) against human ground truth labels[32]. Depending on the normality of distribution, two-tailed, paired t-tests or Wilcoxon tests were conducted to assess the statistical significance of the differences in accuracy between models (zero-shot vs. fine-tuned SAM, EchoNet vs. fine-tuned SAM, and MedSAM vs. fine-tuned SAM), as well as the annotation time and quality metrics in human annotation study (AI-assisted vs. non AI-assisted), with $p<0.05$ as significant.

**Results**

*Patient Characteristics*

The EchoNet-Dynamic patient characteristics have been described in detail [3]. The mean left ventricular ejection fraction of the EchoNet-dynamic dataset was 55.8±12.4%, 55.8±12.3%, and 55.5±12.2%, for the training, validation, and test set, respectively[3]. In the Mayo Clinic data set (n=99), the mean age was 47.5 ± 17.8 years, 58 (58.6%) were male, and coronary artery disease, hypertension, and diabetes were found in 20 (20.2%), 42 (42.4%), and 15 (15.2%) of patients, respectively. The dataset contains an apical four-chamber view of 58 TTE cases and 41 POCUS cases, the LVEF was 61.7 ± 7.5% for TTE and 63.2 ±

11.9% for POCUS cases. The CAMUS cohort had a mean age of 65.1 ± 14.4 years, 66% male, and a mean LVEF of 44.4 ± 11.9%.

*SAM's zero-shot performance on echocardiography and POCUS*

Overall, the zero-shot performance on the EchoNet-dynamic test set had a mean DSC of 0.863 ± 0.053. In terms of individual cardiac phase performance, end-diastolic frames were better than end-systolic frames (mean DSC 0.878 ± 0.040 vs. 0.849 ± 0.060). On the Mayo Clinic dataset, the mean DSC was 0.882 ± 0.036 and 0.861 ± 0.043 for TTE and POCUS, respectively. On the CAMUS dataset, we observed mean DSC of 0.866 ± 0.039 and 0.852 ± 0.048 on A4C and A2C views, respectively (**Table 1** and **Supplemental Table 2**). When compared to the ground truth LVEF, the calculated LVEF had an MAE of 11.67%, 6.28%, and 6.38%, on EchoNet, Mayo-TTE, and Mayo-POCUS data, respectively (**Supplemental Table 3**).

**Table 1. Comparison of zero-shot SAM, fine-tuned SAM, MedSAM, and EchoNet model**

| Dataset | Phase | DSC | | | | | | | IoU | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| | | SAM (zero-shot) | SAM (fine-tuned) | MedSAM | EchoNet | p-value * | p-value ** | p-value *** | SAM (zero-shot) | SAM (fine-tuned) | MedSAM | EchoNet | p-value * | p-value ** | p-value *** |
| EchoNet-test | Overall | 0.863 ± 0.053 | 0.911 ± 0.045 | 0.872 ± 0.056 | 0.915 ± 0.047 | <0.0001 | 0.0012 | <0.0001 | 0.763 ± 0.077 | 0.840 ± 0.071 | 0.777 ± 0.083 | 0.847 ± 0.072 | <0.0001 | 0.0003 | <0.0001 |
| | ED | 0.878 ± 0.040 | 0.929 ± 0.030 | 0.893 ± 0.040 | 0.903 ± 0.052 | <0.0001 | <0.0001 | <0.0001 | 0.784 ± 0.062 | 0.868 ± 0.050 | 0.809 ± 0.062 | 0.826 ± 0.078 | <0.0001 | <0.0001 | <0.0001 |
| | ES | 0.849 ± 0.060 | 0.894 ± 0.050 | 0.850 ± 0.061 | 0.928 ± 0.038 | <0.0001 | <0.0001 | <0.0001 | 0.742 ± 0.084 | 0.812 ± 0.077 | 0.744 ± 0.088 | 0.868 ± 0.059 | <0.0001 | <0.0001 | <0.0001 |
| Mayo-TTE | Overall | 0.882 ± 0.036 | 0.902 ± 0.032 | 0.883 ± 0.039 | 0.893 ± 0.090 | <0.0001 | 0.3167 | <0.0001 | 0.790 ± 0.056 | 0.822 ± 0.051 | 0.792 ± 0.061 | 0.814 ± 0.093 | <0.0001 | 0.3720 | <0.0001 |
| | ED | 0.889 ± 0.037 | 0.916 ± 0.024 | 0.903 ± 0.027 | 0.916 ± 0.031 | <0.0001 | 0.9409 | 0.0010 | 0.802 ± 0.058 | 0.846 ± 0.040 | 0.825 ± 0.044 | 0.846 ± 0.051 | <0.0001 | 0.9991 | 0.0010 |
| | ES | 0.875 ± 0.033 | 0.887 ± 0.032 | 0.862 ± 0.039 | 0.870 ± 0.119 | 0.0039 | 0.3055 | <0.0001 | 0.779 ± 0.052 | 0.799 ± 0.050 | 0.760 ± 0.059 | 0.782 ± 0.113 | 0.0039 | 0.3114 | <0.0001 |
| Mayo-POCUS | Overall | 0.861 ± 0.043 | 0.857 ± 0.047 | 0.846 ± 0.072 | 0.667 ± 0.279 | 0.4469 | <0.0001 | 0.7375 | 0.758 ± 0.066 | 0.753 ± 0.070 | 0.740 ± 0.102 | 0.554 ± 0.265 | 0.4695 | <0.0001 | 0.7868 |
| | ED | 0.876 ± 0.032 | 0.878 ± 0.036 | 0.868 ± 0.064 | 0.717 ± 0.255 | 0.7355 | 0.0002 | 0.7411 | 0.781 ± 0.051 | 0.785 ± 0.056 | 0.772 ± 0.092 | 0.607 ± 0.253 | 0.7097 | <0.0001 | 0.7805 |
| | ES | 0.846 ± 0.048 | 0.836 ± 0.047 | 0.825 ± 0.074 | 0.617 ± 0.295 | 0.2080 | <0.0001 | 0.9020 | 0.735 ± 0.072 | 0.720 ± 0.069 | 0.708 ± 0.103 | 0.501 ± 0.270 | 0.1994 | <0.0001 | 0.4171 |
| CAMUS- A2C | Overall | 0.852 ± 0.048 | 0.891 ± 0.040 | 0.857 ± 0.057 | 0.752 ± 0.196 | <0.0001 | <0.0001 | <0.0001 | 0.745 ± 0.069 | 0.805 ± 0.062 | 0.754 ± 0.081 | 0.633 ± 0.196 | <0.0001 | <0.0001 | <0.0001 |
| | ED | 0.860 ± 0.042 | 0.897 ± 0.037 | 0.877 ± 0.040 | 0.754 ± 0.196 | <0.0001 | <0.0001 | <0.0001 | 0.756 ± 0.062 | 0.815 ± 0.059 | 0.783 ± 0.060 | 0.635 ± 0.197 | <0.0001 | <0.0001 | <0.0001 |
| | ES | 0.845 ± 0.052 | 0.885 ± 0.041 | 0.838 ± 0.065 | 0.751 ± 0.196 | <0.0001 | <0.0001 | <0.0001 | 0.734 ± 0.073 | 0.795 ± 0.064 | 0.725 ± 0.088 | 0.632 ± 0.196 | <0.0001 | <0.0001 | <0.0001 |
| CAMUS- A4C | Overall | 0.866 ± 0.039 | 0.897 ± 0.036 | 0.877 ± 0.044 | 0.850 ± 0.097 | <0.0001 | <0.0001 | <0.0001 | 0.766 ± 0.059 | 0.815 ± 0.058 | 0.784 ± 0.066 | 0.749 ± 0.117 | <0.0001 | <0.0001 | <0.0001 |
| | ED | 0.873 ± 0.037 | 0.904 ± 0.033 | 0.890 ± 0.035 | 0.850 ± 0.098 | <0.0001 | <0.0001 | <0.0001 | 0.776 ± 0.056 | 0.827 ± 0.054 | 0.804 ± 0.055 | 0.749 ± 0.119 | <0.0001 | <0.0001 | <0.0001 |
| | ES | 0.860 ± 0.041 | 0.889 ± 0.038 | 0.864 ± 0.048 | 0.850 ± 0.096 | <0.0001 | <0.0001 | <0.0001 | 0.756 ± 0.061 | 0.803 ± 0.060 | 0.763 ± 0.070 | 0.749 ± 0.115 | <0.0001 | <0.0001 | <0.0001 |

*Zero-shot vs. fine-tuned SAM. **Fine-tuned SAM vs. EchoNet model. ***Fine-tuned SAM vs. MedSAM. A2C: apical 2 chamber view, A4C: apical 4 chamber view, CAMUS: Cardiac Acquisitions for Multi-structure Ultrasound Segmentation, DSC: Dice Similarity Score, ED: end-diastolic, ES: end-systolic, IoU: Intersection over Union, SAM: segment anything model, TTE: transthoracic echocardiography, POCUS: point-of-care ultrasound. Data expressed as mean± standard deviation.

*SAM's fine-tuned performance on echocardiography and POCUS*

Fine-tuning generally improved the performance of SAM, with a mean DSC of 0.911 ± 0.045 on the EchoNet-dynamic test set (**Table 1** and **Supplemental Table 2**). Similar improvement was also observed in Mayo TTE data, with an overall mean DSC of 0.902 ± 0.032. In contrast, no significant improvement was observed on the POCUS data (DSC 0.857 ± 0.047), while the performance was numerically improved when compared with the ground truth of the second observer (DSC 0.876 ± 0.038), as summarized in **Table 2.** The EchoNet model had a significant performance drop, especially on Mayo-POCUS and CAMUS A2C datasets (**Table 1, Figure 1**). When compared to the ground truth LVEF, the calculated LVEF had an MAE of 7.52%, 5.47%, and 6.70%, on the EchoNet test, Mayo-TTE, and Mayo-POCUS data, respectively (**Supplemental Table 3**).

**Table 2. Zero-shot vs. Fine-tuned SAM performance on TTE and POCUS (against the second observer).**

| | TTE (n=33) | | | POCUS (n=19) | | |
|---|---|---|---|---|---|---|
| | **Zero-shot** | **Fine-tuned** | **p-value** | **Zero-shot** | **Fine-tuned** | **p-value** |
| **Mean IoU (overall)** | 0.776 ± 0.063 | 0.828 ± 0.061 | <0.0001 | 0.755 ± 0.067 | 0.781 ± 0.059 | 0.0551 |
| **Mean DSC (overall)** | 0.873 ± 0.040 | 0.905 ± 0.038 | <0.0001 | 0.859 ± 0.046 | 0.876 ± 0.038 | 0.0591 |
| **Mean IoU (ED)** | 0.781 ± 0.057 | 0.864 ± 0.033 | <0.0001 | 0.773 ± 0.051 | 0.799 ± 0.052 | 0.1964 |
| **Mean DSC (ED)** | 0.876 ± 0.037 | 0.927 ± 0.019 | <0.0001 | 0.871 ± 0.033 | 0.887 ± 0.032 | 0.2037 |
| **Mean IoU (ES)** | 0.771 ± 0.068 | 0.793 ± 0.062 | 0.1065 | 0.738 ± 0.078 | 0.763 ± 0.061 | 0.1721 |
| **Mean DSC (ES)** | 0.869 ± 0.043 | 0.883 ± 0.039 | 0.1044 | 0.847 ± 0.055 | 0.864 ± 0.041 | 0.1794 |

IoU: intersection over union, DSC: Dice similarity score, ED: end-diastolic, ES: end-systolic. Data expressed as mean± standard deviation.

*Comparison of SAM, MedSAM, and EchoNet models*

Comparisons were made between fine-tuned SAM, MedSAM, and EchoNet. EchoNet demonstrated a similar level of performance compared to fine-tuned SAM on the EchoNet test set (SAM vs. EchoNet: DSC 0.911 ± 0.045 vs. 0.915 ± 0.047, p<0.0001) and the Mayo TTE dataset (DSC 0.902 ± 0.032 vs. 0.893 ± 0.090, p<0.0001). However, EchoNet significantly underperformed on the Mayo POCUS and CAMUS datasets, with a performance drop ranging from 5-25% (**Table 1**). MedSAM also showed about 2-5% worse performance than fine-tuned SAM across all datasets (all p<0.0001) (**Table 1**).

*Human Annotation Study*

In the human annotation study, we observed that AI assistance significantly improved efficiency by decreasing 50% annotation time (p <0.0001; **Figure 2A**), which remains true for both expert-level and medical student-level annotators (**Figure 2B and 2C**). The AI-assisted workflow maintained the quality of left ventricle segmentation, as measured by the Dice Similarity Coefficient (DSC) (**Figure 2D, E, and F**). Qualitative assessments indicated an improvement in performance for inexperienced annotators, which may not be completely captured by the DSC scores (**Supplemental Figure 1**).

**Discussion**

The major contributions of this work include: 1) presenting a data-efficient strategy for training a segmentation model for echocardiography images based on SAM and leveraging its generalization

capability for POCUS images and 2) showing the potential of optimizing clinical workflows through a human-in-the-loop approach, where SAM significantly improved efficiency while maintaining annotation quality.

*SAM's performance on echocardiography/POCUS*
Echocardiography, like other ultrasound modalities, is generally considered an imaging modality with more challenges due to its operator dependency and low signal-to-noise ratio[3,33–35]. Additionally, objects could often have weak border linings or be obstructed by artifacts on ultrasound/echocardiography images, which posed specific challenges for echocardiography segmentation tasks[3,25]. Compared to other non-cardiac ultrasound images[25], SAM seems to have a relatively better performance on echocardiography with DSC above 0.85 across datasets.

The zero-shot performance of SAM on echocardiography images was good (DSC $0.863 \pm 0.053$) and was close to the performance of the EchoNet model[3]. On the EchoNet-dynamic dataset, both zero-shot and fine-tuned SAM had a slightly better performance on end-diastolic frames, which was opposite to the original EchoNet model. This could be a result of the fact that end-diastolic frames usually have better visualized left ventricular endocardial borders, and SAM does not depend on the intermediate frames as the video-based EchoNet-Dynamic model[3]. While the frame-based approach does not consider consecutive inter-frame changes and spatio-temporal features of echocardiogram, fine-tuned SAM achieved superior frame-level segmentation performance on non-EchoNet datasets compared to the original video-based EchoNet model[3]. Importantly, there were cases with suboptimal image quality and imperfect human labels (**Figure 3**) in the EchoNet-dynamic data set[3,36], which can limit the model performance.

The fine-tuned SAM model demonstrated strong generalization capability, with about a 1% and 5% drop in performance on unseen TTE and POCUS data (**Table 1** and **Figure 1**), respectively. The comparable performance on CAMUS A2C and A4C views also indicated SAM's generalization capability, as it was only fine-tuned on A4C images. Interestingly, the fine-tuned model did not demonstrate an overall superior performance on POCUS images compared to zero-shot performance. This is likely due to inter-observer variation as we observed numerical improvement of the model performance on the other observer's tracings (**Table 2**). Furthermore, in qualitative analysis, the fine-tuned model usually predicts a mask that is more consistent with the anticipated LV geometry on POCUS (**Figure 4**).

Importantly, fine-tuned SAM outperformed MedSAM across all datasets (**Table 1**). This is likely due to the fact that cardiac ultrasound images were less represented in MedSAM's training set of over 1 million images. Our observation suggests that, while training on a large dataset can provide generally robust performance, specific tasks involving a single modality require representative training examples, even if they may be smaller. Approaches with representative training examples could be more data-efficient for medical subspecialties to build segmentation models for specific purposes.

*Integrating SAM and Future Foundation Models into Clinical Research and Practice*
One of the major limitations in building machine learning models for healthcare is the collection of high-quality training datasets. In addition, the models' generalizability is frequently questionable[19,37]. Annotators with medical expertise are expensive and limited; while EchoNet is currently the largest

public echo segmentation dataset[3], its size is still far from the web-scale data used to train SAM[28]. Our study demonstrated the potential of a human-in-loop workflow for echocardiography segmentation, significantly reducing the annotation time and maintaining the segmentation quality across users with different experience levels (**Figure 2**). While SAM-assisted workflow did not further improve segmentation quality quantitatively, we believe it is secondary to the inter-observer variability and imperfect ground truth segmentation from the EchoNet dataset. In qualitative examples, we observed substantial improvement for the inexperienced user (**Supplemental Figure 1**).

SAM's capability allows the potential integration of SAM into research or clinical workflows in echocardiography labs to create high-quality segmentation masks at scale [23,25]. Compared to the fully automatic models[3], an advantage of having an interactive, human-in-the-loop approach with SAM is that the segmentations can be directly prompted or modified to the level of interpreters' satisfaction[25,28]. This process also has the potential to gain the trust of human users to facilitate integration[38]. While future studies are needed to assess its real impact on clinical practice, the interactive functionality of SAM could be especially helpful in challenging cases or when fully automatic models fail to predict segmentations accurately [25].

We observed a significant drop in EchoNet's performance on the POCUS dataset (**Table 1**), with completely failed segmentation in 4 (9.8%) of cases. In contrast, the comparable performance of SAM on TTE and POCUS images demonstrated the advantage of leveraging the generalization capabilities of foundation models [23] in building AI solutions for the rapidly growing use of POCUS in cardiac imaging[39]. While evaluation on a larger POCUS dataset is required to better evaluate inter-observer variations, we demonstrated a strategy to fine-tune foundation models using readily available and relatively high-quality TTE results knowing that POCUS images usually come with larger variations in operator skill levels, image quality, and scanning modalities[39–42]. We observed that fine-tuned SAM, similar to MedSAM[29], can accurately segment cases with weak or missing boundaries, which is more common in POCUS images. This can be especially useful in assisting the estimation of cardiac function for POCUS users, who are often not trained for cardiac ultrasound segmentation. SAM can also be used to facilitate the creation of high-quality clinical echocardiography segmentation datasets by reducing the workload of annotation[25,28,29].

Finally, while the EchoNet-dynamic dataset does not include detailed segmentation other than LV endocardium, a potential future step is to include an object detection model that can detect cardiac chambers, and provide corresponding bounding boxes as prompts to SAM to enable a fully automized AI-segmentation framework.

**Conclusion**
Fine-tuning SAM on a relatively small yet representative echocardiography dataset resulted in superior performance compared to the state-of-the-art echocardiography segmentation model, EchoNet, and SAM's variant MedSAM, which was trained on over 1 million medical images. We demonstrated a data-efficient approach for handling complex echocardiography images, and SAM's potential to enhance current echocardiography workflows by improving the efficiency with a human-in-the-loop design.

**Limitations**

This study is limited by its retrospective nature and could be subject to selection bias. Since SAM is an image-based model, its performance was evaluated on pre-selected end-diastolic and end-systolic frames rather than the beat-to-beat assessment of the video as proposed by the EchoNet-Dynamic model. However, we still demonstrated superior frame-level performance with fine-tuned SAM. While adapters are another potential direction to use SAM for echocardiography segmentation[43], we have not specifically explored this approach in the current paper. How SAM can be integrated into the current echocardiography lab workflow and its real-world effects will need to be validated in a prospective setting.

**Ethical review and approval**: EchoNet-Dynamic dataset contains a publicly available, de-identified dataset. The use of the Mayo Clinic dataset was approved by the institutional review board (protocol#22-010944).

**Author Contributions:**

| Author | Contributions |
| --- | --- |
| Chieh-Ju Chao | Conceptualization, methodology, software, validation, data curation, formal analysis, writing - original draft, visualization |
| Yunqi Richard Gu | Methodology, software |
| Wasan Kumar | Software, data curation |
| Tiange Xiang | Methodology, software |
| Lalith Appari | Methodology, software |
| Justin Wu | Methodology, software |
| Juan M. Farina | Data curation |
| Rachael Wraith | Data curation |
| Jiwoon Jeong | Data curation, software |
| Reza Arsanjani | Data curation, resources, writing- review and editing |
| Garvan C. Kane | Data curation, resources, writing- review and editing |
| Jae K. Oh | Data curation, resources, writing- review and editing |
| Curtis P. Langlotz | Resources, writing- review and editing |
| Imon Banerjee | Conceptualization, resources, writing- review and editing, supervision |
| Fei-Fei Li | Conceptualization, methodology, resources, writing- review and editing, supervision |
| Ehsan Adeli | Conceptualization, methodology, resources, writing- review and editing, supervision |

**Data Availability:**

The EchoNet and CAMUS datasets are publicly available. The Mayo Clinic dataset can not be make publicly avaliable due to patient privacy regulations; it is available from the corresponding author on reasonable request.

**Code Availability:**

We released a checkpoint of the fine-tuned SAM model, along with the fine-tuning, inference, and statistical analysis code. The code and checkpoint are available on GitHub:
https://github.com/chiehjuchao/SAM-Echo.git

**Figure 1. Comparison of the overall Dice Similarity Scores (DSC) between EchoNet and fine-tuned SAM models.** This figure illustrates the distribution of overall DSC scores for EchoNet (in light blue) and Fine-tuned SAM (in orange), across multiple datasets. The box plots depict the median, quartiles, and 95% confidence intervals of the overall DSC scores. Fine-tuned SAM demonstrated consistent performance across different datasets and was significantly superior to the EchoNet model on the Mayo-TTE, Mayo-POCUS, and CAMUS datasets (all p<0.0001).

**Figure 2.** SAM-assisted workflow decreased the time for annotation while maintaining the quality of echocardiography image segmentation. **Panel A.** With SAM's assistance, the average annotation time significantly decreased by 50%. **Panel B.** Among experienced annotators, the workflow decreased annotation time by 39.2%; **Panel C**. It also reduced 66.0% annotation time for an inexperienced annotator. **Panel D, E, and F.** No significant difference was observed in segmentation quality when measured by the Dice Similarity Score. ***p<0.0001.

**Figure 3.** Qualitative performance of fine-tuned SAM on representative cases against ground truth on the EchoNet-dynamic test dataset. From top to bottom: 97.5th to 2.5th percentile of DSC. Panels A, B, and C are end-diastolic frames, and Panels D, E, and F are end-systolic frames. We observed that many of the poor-performance cases had suboptimal image qualities, such as weak LV endocardial borders or off-axis views (Panels C and F), suggesting the importance of good input image quality on model performance. Additionally, end-diastolic frames usually have a better delineation of borders than end-systolic frames, which is consistent with the model performance (end-diastolic slightly better than end-systolic).

**Figure 4.** Zero-shot and fine-tuned SAM performance on a representative POCUS case. **Panel A.** end-diastolic frame, **Panel B**. end-systolic frame. From left to right are the ground truth, zero-shot, and fine-tuned mask, with an overlay of bounding boxes (green-colored) and mask (blue-colored), on the original POCUS image. Fine-tuned masks were more consistent with anticipated left ventricular geometry on visualization. Note that POCUS images generally had worse quality compared to transthoracic echocardiography images.

<div align="center">
Original Image     Ground Truth Mask     Mask (SAM)     Mask (FreeHand)
</div>

**Supplemental Figure 1.** This representative qualitative example illustrates the SAM-assisted workflow's impact on improving segmentation quality for an inexperienced user. The panels, from left to right, display the original image, the ground truth mask overlay, the SAM-assisted mask overlay, and the freehand mask overlay. This example highlights an imperfect ground truth mask, showing that the SAM-assisted annotation is substantially better than the freehand annotation. Notably, the freehand annotation incorrectly included the septal and lateral wall, which is inappropriate for left ventricular segmentation.

**Supplemental Table 1. Excluded EchoNet-dynamic dataset cases due to missing ground truth**

| Filename | Split |
|---|---|
| 0X234005774F4CB5CD.avi | TRAIN |
| 0X2DC68261CBCC04AE.avi | TRAIN |
| 0X35291BE9AB90FB89.avi | TRAIN |
| 0X6C435C1B417FDE8A.avi | TRAIN |
| 0X5515B0BD077BE68A.avi | TRAIN |
| 0X5DD5283AC43CCDD1.avi | TEST |

**Supplemental Table 2. Zero-shot and Fine-tuned SAM performance on EchoNet-dynamic Train and Validation set**

| | Zero-shot | | Fine-tuned | |
|---|---|---|---|---|
| | EchoNet-Dynamic-Train | EchoNet-Dynamic-Validation | EchoNet-Dynamic-Train | EchoNet-Dynamic-Validation |
| **Mean IoU (overall)** | 0.761 ± 0.080 | 0.762 ± 0.078 | 0.842 ± 0.070 | 0.841 ± 0.069 |
| **Mean DSC (overall)** | 0.862 ± 0.055 | 0.862 ± 0.054 | 0.913 ± 0.044 | 0.912 ± 0.043 |
| **Mean IoU (ED)** | 0.781 ± 0.068 | 0.783 ± 0.065 | 0.868 ± 0.052 | 0.869 ± 0.050 |
| **Mean DSC (ED)** | 0.875 ± 0.046 | 0.877 ± 0.043 | 0.928 ± 0.032 | 0.929 ± 0.030 |
| **Mean IoU (ES)** | 0.740 ± 0.085 | 0.740 ± 0.084 | 0.816 ± 0.076 | 0.813 ± 0.074 |
| **Mean DSC (ES)** | 0.848 ± 0.060 | 0.848 ± 0.059 | 0.897 ± 0.049 | 0.895 ± 0.047 |

IoU: intersection over union, DSC: Dice similarity score, ED: end-diastolic, ES: end-systolic. Data expressed as mean± standard deviation.

**Supplemental Table 3. LVEF prediction task evaluated by R-square and MAE of zero-shot vs. fine-tuned SAM**

| | EchoNet-dynamic test | | CAMUS-A2C | | CAMUS-A4C | | Mayo-TTE | | Mayo-POCUS | |
|---|---|---|---|---|---|---|---|---|---|---|
| | Zero-shot | Fine-tuned | Zero-shot | Fine-tuned | Zero-shot | Fine-tuned | Zero-shot | Fine-tuned | Zero-shot | Fine-tuned |
| **R-squared** | 0.141 | 0.161 | 0.445 | 0.761 | 0.549 | 0.709 | 0.374 | 0.517 | 0.638 | 0.718 |
| **MAE (%)** | 11.7 | 7.52 | 9.75 | 6.11 | 9.67 | 7.31 | 6.28 | 5.47 | 6.38 | 6.70 |

The summation of disks method was used to calculate the LV ejection fraction (LVEF) from the end-diastolic and end-systolic segmentation masks. While the model was not trained for LVEF prediction, the model-derived LVEF measurements were compared to the ground truth-derived LVEF by R-square and mean absolute error (MAE).

Reference

1. Antoine, C. *et al.* Clinical Outcome of Degenerative Mitral Regurgitation. *Circulation* **138**, 1317–1326 (2018).

2. Matulevicius, S. A. et al. Appropriate Use and Clinical Impact of Transthoracic Echocardiography. *JAMA Intern Med* **173**, 1600–1607 (2013).

3. Ouyang, D. et al. Video-based AI for beat-to-beat assessment of cardiac function. *Nature* **580**, 252–256 (2020).

4. Lang, R. M. et al. Recommendations for Cardiac Chamber Quantification by Echocardiography in Adults: An Update from the American Society of Echocardiography and the European Association of Cardiovascular Imaging. *J Am Soc Echocardiogr* **28**, 1-39.e14 (2015).

5. Liu, J. et al. Contemporary Role of Echocardiography for Clinical Decision Making in Patients During and After Cancer Therapy. *JACC Cardiovasc Imaging* **11**, 1122–1131 (2018).

6. Fonseca, R., Jose, K. & Marwick, T. H. Understanding decision-making in cardiac imaging:

determinants of appropriate use. *European Hear J - Cardiovasc Imaging* **19**, 262–268 (2017).

7. Otto, C. M. et al. 2020 ACC/AHA Guideline for the Management of Patients With Valvular Heart Disease A Report of the American College of Cardiology/American Heart Association Joint Committee on Clinical Practice Guidelines. *J Am Coll Cardiol* (2020) doi:10.1016/j.jacc.2020.11.018.

8. Tam, J. W., Nichol, J., MacDiarmid, A. L., Lazarow, N. & Wolfe, K. What Is the Real Clinical Utility of Echocardiography? A Prospective Observational Study. *J Am Soc Echocardiogr* 12, 689–697 (1999).

9. Pellikka, P. A. et al. Variability in Ejection Fraction Measured By Echocardiography, Gated Single-Photon Emission Computed Tomography, and Cardiac Magnetic Resonance in Patients With Coronary Artery Disease and Left Ventricular Dysfunction. *JAMA Netw Open* **1**, e18**1**456–e181456 (2018).

10. Malm, S., Frigstad, S., Sagberg, E., Larsson, H. & Skjaerpe, T. Accurate and reproducible measurement of left ventricular volume and ejection fraction by contrast echocardiography A comparison with magnetic resonance imaging. J Am Coll Cardiol **44**, 1030–1035 (2004).

11. Cole, G. D. et al. Defining the real-world reproducibility of visual grading of left ventricular function and visual estimation of left ventricular ejection fraction: impact of image quality, experience and accreditation. *Int J Cardiovasc Imaging* **31**, 1303–1314 (2015).

12. Koh, A. S. et al. A comprehensive population‑based characterization of heart failure with mid‑range ejection fraction. *Eur J Heart Fail* **19**, 1624–1634 (2017).

13. Gilbert, A. et al. Generating Synthetic Labeled Data From Existing Anatomical Models: An Example With Echocardiography Segmentation. *IEEE T Med Imaging* **40**, 2783–2794 (2021).

14. Salte, I. M. et al. Artificial Intelligence for Automatic Measurement of Left Ventricular Strain in Echocardiography. JACC Cardiovasc Imaging 14, 1918–1928 (2021).

15. Huang, H. et al. Segmentation of Echocardiography Based on Deep Learning Model. *Electronics* **11**, 1714 (2022).

16. Amer, A., Ye, X. & Janan, F. ResDUnet: A Deep Learning-Based Left Ventricle Segmentation Method for Echocardiography. *IEEE Access* **9**, 159755–159763 (2021).

17. Jafari, M. H., Woudenberg, N. V., Luong, C., Abolmaesumi, P. & Tsang, T. Deep Bayesian Image Segmentation For A More Robust Ejection Fraction Estimation. *2021 IEEE 18th Int. Symp. Biomed. Imaging (ISBI)* **00**, **12**64–1268 (2021).

18. Mokhtari, M., Tsang, T., Abolmaesumi, P. & Liao, R. Medical Image Computing and Computer Assisted Intervention – MICCAI 2022. *Lect. Notes Comput. Sci.* 360–369 (2022) doi:10.1007/978-3-031-16440-8_35.

19. Yu, Y. et al. Techniques and Challenges of Image Segmentation: A Review. *Electronics* 12, 1199 (2023).

20. Dosovitskiy, A. et al. An Image is Worth 16x16 Words: Transformers for Image Recognition at Scale. *Arxiv* (2020) doi:10.48550/arxiv.2010.11929.

21. Vaswani, A. et al. Attention Is All You Need. Arxiv (2017) doi:10.48550/arxiv.1706.03762.

22. Liu, Y. et al. Summary of ChatGPT/GPT-4 Research and Perspective Towards the Future of Large Language Models. *Arxiv* (2023).

23. Bommasani, R. et al. On the Opportunities and Risks of Foundation Models. *Arxiv* (2021) doi:10.48550/arxiv.2108.07258.

24. Kirillov, A. et al. Segment Anything. *Arxiv* (2023).

25. Mazurowski, M. A. et al. Segment Anything Model for Medical Image Analysis: an Experimental Study. *Arxiv* (2023).

26. Ma, J. et al. Segment anything in medical images. *Nat. Commun.* **15**, 654 (2024).

27. Leclerc, S. et al. Deep Learning for Segmentation Using an Open Large-Scale Dataset in 2D Echocardiography. *IEEE Trans. Méd. Imaging* **38**, 2198–2210 (2019).

28. Kirillov, A. et al. Segment Anything. *Arxiv* (2023).

29. Ma, J. & Wang, B. Segment Anything in Medical Images. *arXiv* (2023) doi:10.48550/arxiv.2304.12306.

30. Isensee, F., Jaeger, P. F., Kohl, S. A. A., Petersen, J. & Maier-Hein, K. H. nnU-Net: a self-configuring method for deep learning-based biomedical image segmentation. *Nat. Methods* **18**, 203–211 (2021).

31. Kingma, D. P. & Ba, J. Adam: A Method for Stochastic Optimization. *Arxiv* (2014) doi:10.48550/arxiv.1412.6980.

32. Maier-Hein, L. et al. Metrics reloaded: Recommendations for image analysis validation. *Arxiv* (2022) doi:10.48550/arxiv.2206.01653.

33. Mintz, G. S. & Kotler, M. N. Clinical Value and Limitations of Echocardiography: Its Use in the Study of Patients With Infectious Endocarditis. *Arch. Intern. Med.* **140**, 1022–1027 (1980).

34. Mondillo, S., Maccherini, M. & Galderisi, M. Usefulness and limitations of transthoracic echocardiography in heart transplantation recipients. *Cardiovasc. Ultrasound* **6**, 2–2 (2008).

35. Abdulla, A. M., Frank, M. J., Canedo, M. I. & Stefadouros, M. A. Limitations of echocardiography in the assessment of left ventricular size and function in aortic regurgitation. *Circulation* **61**, 148–155 (2018).

36. Tromp, J. et al. Automated interpretation of systolic and diastolic function on the echocardiogram: a multicohort study. *Lancet Digital Heal* **4**, e46–e54 (2022).

37. Zhang, D. et al. Deep Learning for Medical Image Segmentation: Tricks, Challenges and Future Directions. *Arxiv* (2022) doi:10.48550/arxiv.2209.10307.

38. Bao, Y., Cheng, X., Vreede, T. D. & Vreede, G.-J. D. Investigating the relationship between AI and trust in human-AI collaboration. *Proc. 54th Hawaii Int. Conf. Syst. Sci.* (2021) doi:10.24251/hicss.2021.074.

39. Huang, G. S., Alviar, C. L., Wiley, B. M. & Kwon, Y. The Era of Point-of-Care Ultrasound Has Arrived: Are Cardiologists Ready? *Am J Cardiol* **132**, 173–175 (2020).

40. Kalagara, H. et al. Point-of-Care Ultrasound (POCUS) for the Cardiothoracic Anesthesiologist. *J Cardiothor Vasc An* **36**, 1132–1147 (2022).

41. Palmero, S. L. et al. Point-of-Care Ultrasound (POCUS) as an Extension of the Physical Examination in Patients with Bacteremia or Candidemia. *J Clin Medicine* **11**, 3636 (2022).

42. Kline, J., Golinski, M., Selai, B., Horsch, J. & Hornbaker, K. The effectiveness of a blended POCUS curriculum on achieving basic focused bedside transthoracic echocardiography (TTE) proficiency. A formalized pilot study. *Cardiovasc Ultrasoun* **19**, 39 (2021).

43. Wu, J. et al. Medical SAM Adapter: Adapting Segment Anything Model for Medical Image Segmentation. *Arxiv* (2023).