Microsatellite Instability Detection in Clinical Cancer Samples: A Multiplex qPCR

Approach without Matching Normal Samples

Wei Chen[1,2], Yan Helen Yan[1,2], Blake Young[1], Alessandro Pinto[1], Qi Jiang[1], Nanjia

Song[1], Weijie Yao[1], David Yu Zhang[1]*, Jinny Xuemeng Zhang[1]*

1NuProbe USA, Houston TX, USA   *2These authors contributed equally.*

*Corresponding authors

David Yu Zhang   Jinny Xuemeng Zhang

2575 West Bellfort Avenue, Suite 200   2575 West Bellfort Avenue, Suite 200

Houston, TX, 77054   Houston, TX, 77054

Genomic.dave@gmail.com   Jinnychang92@gmail.com

(713) 485-0148   (713) 485-0148

Keywords: Microsatellite Instability (MSI); qPCR; Colorectal Cancer; Lynch Syndrome

Abstract

Background: Microsatellite instability (MSI) indicates DNA mismatch repair deficiency in cancers like colorectal cancer. The current gold standard technique, PCR/capillary electrophoresis (CE), requires matching normal samples and specialized instrumentation. We developed VarTrace, a rapid and low-cost quantitative PCR (qPCR) assay, to evaluate MSI using solely the tumor sample DNA, obviating the requirement for matching normal samples.

Methods: 101 formalin-fixed paraffin-embedded (FFPE) tumor samples were tested using VarTrace and compared to the Promega OncoMate assay utilizing PCR-CE. Tumor percentage limit of detection was evaluated on contrived samples derived from clinical MSI-H samples. Analytical sensitivity, specificity, limit of detection and input requirements were assessed using synthetic commercial reference standards.

Results: VarTrace demonstrated 100% test success rate, 100% sensitivity and 98% specificity compared to OncoMate across 101 clinical FFPE samples. It detected MSI-H with 97% accuracy down to 10% tumor percentage. Analytical studies using synthetic samples showed a limit of detection of 5% variant allele frequency and a limit of input of 0.5 ng.

Conclusions: This study validates VarTrace as a swift, accurate and economical assay for MSI detection in samples with low tumor percentages without the need for matching normal DNA. VarTrace's capacity for highly sensitive MSI analysis holds potential for enhancing the efficiency of clinical workflows and broadening the availability of this crucial test.

Introduction

Microsatellite instability (MSI) is characterized by length changes in repetitive DNA sequences called microsatellites and is often caused by deficiencies in DNA mismatch repair (dMMR) (1, 2). MSI is an important biomarker in several cancers (3, 4), most notably colorectal cancer (1, 5) where approximately 15% of patients are classified as MSI-H (6). MSI can indicate dMMR tumors and is also an important screening factor for Lynch syndrome (6-8), an inherited disorder caused by germline mutations in MMR genes that confers high lifetime cancer risks. Patients with high MSI (MSI-H) tumors have favorable prognosis and derive great benefit from immunotherapy compared to aggressive chemotherapy(9-17).

Current standard methods for detecting MSI and dMMR in colorectal cancer include immunohistochemistry (IHC), PCR amplification of microsatellite markers followed by capillary electrophoresis (PCR-CE) (18, 19), or next-generation sequencing (NGS) (20-28) as recommended by the National Comprehensive Cancer Network. IHC can reflect MSI status through identifying loss of MMR protein expression but may miss some MSI tumors with intact protein expression (29, 30). NGS and PCR-CE directly classify tumors as MSI-H or microsatellite stable (MSS) but often require comparison to matched normal tissue (18-26). While NGS can accurately assess MSI and MMR gene status from minute inputs (25), it requires specialized instrumentation. PCR-CE is considered the gold standard for MSI detection due to its high efficiency and sensitivity in detecting repeat markers. The Bethesda Guidelines recommend a panel of five well-characterized homopolymer markers for MSI testing by PCR-CE: BAT-25, BAT-26, NR-21, NR-24, and MONO-27(31). However, neither IHC nor PCR-CE alone ensures

accurate diagnosis (18, 29, 30, 32, 33). Using both methods simultaneously increases

sample requirements and screening costs substantially. Moreover, the multi-step PCR-

CE process still demands advanced instrumentation, with only marginal cost savings

over NGS. Alternative methods such as droplet digital PCR (32) and targeted next-

generation sequencing panels (25-28) have also emerged for accurate MSI screening,

but require further validation.

To reduce costs and sample input while maintaining sensitivity for MSI detection, we

developed a rapid qPCR assay using specially designed primers and blockers to

preferentially amplify MSI alleles over wildtype sequences. Our assay requires only 0.5

ng DNA input and can directly detect MSI in DNA samples extracted from just 10%

tumor content without preamplification. Combined with automated data analysis, MSI

status can be determined within 8 hours of receiving a sample, much faster than

existing PCR-CE protocols. This qPCR approach could potentially be applied to both

tumor tissue and cell-free DNA pending clinical validation. By enabling rapid, robust MSI

detection from minute DNA quantities, this method could greatly expand test

accessibility in labs with qPCR instrument and facilitate Lynch syndrome screening and

precision medicine in colorectal cancer management.

Material and Methods

Samples

Formalin-fixed paraffin-embedded (FFPE) samples were obtained from several sources.

12 samples were purchased from BioChain Institute, of which 7 were confirmed to have

MSI status by IHC and 5 were confirmed by Promega OncoMate MSI Dx assay. 41

samples with a diagnosis of adenocarcinoma of the7colon/rectum were purchased from

OriGene, but MSI status was not determined. 48 samples with MSI status confirmed by

PCR were purchased from Discovery Life Sciences. All clinical samples information is

collected from vendors and summarized in Table 1 and Supplemental Table S1-3.

Seraseq® MSI Reference Panel Mix AF5% (Seracare, catalog number: 0710-1675) was

used as Commercial reference standard for sensitivity, specificity and limit of input study.

In addition, VarTrace Positive and Negative Controls was also included as a comparison

in for sensitivity, specificity and limit of input (LoI) study. VarTrace Negative Control is

background human genomic DNA GM25485, which is known to be microsatellite stable

(MSS). VarTrace Positive Control is a mixture of GM25485 and 5 synthetic MSI-H DNA

fragments, and variant for each marker was quantified as ~5% variant allele frequency

(VAF) using qPCR individually. DNA fragments used in Positive Control are plasmid

synthesized and purified in GenScript. Additional WT and variants templates used in

limit of blank (LoB) and limit of detection (LoD) study respectively are synthetic gBlock

fragments ordered from IDT, sequences are listed in Supplemental Table S4.

NuProbe VarTrace MSI assay

The VarTrace MSI Kit comprises 2 tubes of 4x oligo mixes that can detect totally 5

markers, a Positive Control and a Negative Control. The background DNA used in both

Positive and Negative Controls is cell line GM25485 (BioChain), which is known to be

MSS. TaqPath ProAmp Multiplex Master Mix (Applied Biosystems, Catalog number:

A30868) was used to conduct all VarTrace experiment. All VarTrace experiment was

prepared strictly following the user manual (Supplemental Methods) and perform qPCR

readout using Applied Biosystems 3500 Fast Dx (Table 2).

MSI Testing Using PCR-CE Method

Promega Oncomate MSI Dx is an assay cleared by FDA to determine MSI status, and well known as the "gold standard'. All FFPE samples underwent MSI testing using the Promega OncoMate MSI Dx Analysis assay per the manufacturer's protocol. Testing was performed by a third-party CLIA-certified lab, which carried out DNA extraction, quantitation, and Promega assay on the samples. Redundant DNA was returned to NuProbe and stored at 4°C for 2 years prior to additional test.

DNA Extraction and Quantification

For OriGene and BioChain samples, DNA was extracted from one slide per sample and no tumor preselection using the GeneRead FFPE Kit (Qiagen, Germany, Catalog number: 180134) per manufacture protocol with a slight modification (details in Supplemental Methods).

Considering only one slide per sample was obtained from Discovery, DNA extraction was performed in the third-party CLIA lab for Promega OncoMate assay test. Redundant DNA was returned and stored at 4°C. Prior to VarTrace Assay test, part of Discovery DNA samples was purified and concentrated for a second time due to magnetic bead residue.

All samples are quantified for DNA concentration prior to VarTrace tests using Invitrogen Qubit 4 Fluorometer (ThermoFisher Scientific, Catalog number: Q33238) and Qubit™ dsDNA Quantification Assay, High Sensitivity (ThermoFisher Scientific, Catalog number: Q32851).

Data Analysis

The QuantStudio 3/5 Real-Time PCR Design & Analysis 2 software is used to extract Cq values under default settings. In the analysis pipeline (Supplemental Method and Fig. S1), samples are labeled as failed if the Cq value of *GAPDH* is larger than 37 or no amplification occurred. For each sample and control, Cq Value of *GAPDH* is subtracted from Cq value of each marker obtain $\Delta Cq_{marker}^{Sample}$. A marker is considered positive if $\Delta Cq_{marker}^{Sample}$ is smaller than $\Delta Cq_{marker}^{PosCtrl}$ + 1. Samples with more than 2 positive markers were classified as MSI-H, those with 1 positive marker as MSI-L, and those with 0 positive markers as MSS. Both MSI-L and MSS are considered as Negative in final report, and only MSI-H is considered as Positive.

Results

Blocker displacement amplification enables qPCR detection of MSI at Low Variant Allele Frequencies

NuProbe developed VarTrace qPCR assay to detect MSI using a previous published blocker displacement amplification (BDA) technology without matching normal sample. BDA can manipulate reaction thermodynamics by introducing a rationally designed blocker oligonucleotide that perfectly binds to wildtype (WT) targets and shares a same binding region with the primer (Fig. 1A). Blocker has a stronger binding affinity towards WT templates comparing to primer, resulting in the inhibition of primer binding and extension. On the other hand, bulge formed between variant template and blocker can weaken the blocker binding stability and better amplification. Therefore, BDA can detect

low frequency variants by preferably amplify variants over WT templates and generating

fluorescence signals in qPCR setting.

The feasibility of single-plex qPCR with BDA in detecting microsatellite instability was

first demonstrated by amplifying NR-21 unstable variants spiked into microsatellite

stable (MSS) background DNA at variant allele frequencies (VAFs) down to 0.1% (Fig.

1B). There was significant separation of over 4 Cq cycles between the wildtype and 0.1%

VAF samples, while no template controls showed no amplification. All 5 markers were

tested individually with samples with a range of VAFs, and calibration curves for all 5

markers from 0.1% to 50% VAF showed high coefficients of determination (R^2) ranging

from 0.93 to 0.99 (Supplemental Fig. S2), indicating both accurate detection and

quantitation of MSI samples by qPCR using BDA.

Based on sensitive detection on each marker in single plex validation, VarTrace MSI

assay comprises a set of seven blockers to target each one of 5 MSI markers from

Bethesda guidelines: NR-21, NR-24, BAT-25, BAT-26 and MONO-27 (Fig. 1C). Primers

targeting *GAPDH* are included as a reference control that reflects sample input and

quality. *GAPDH* Cq value is utilized in the analysis algorithm. One set of blockers

correspond to sequences of WT 0, ±1, ±2, and ±3 bp in order to minimize stutter

products, short tandem repeats introduced by polymerase slippage. WTs (0) are

selected to maximize coverage of natural variation in wild-type microsatellite lengths,

enabling MSI detection without requiring a matched normal sample.

Detection of MSI in FFPE Tumor Samples using the VarTrace Assay

A total of 101 FFPE tumor samples obtained from commercial sources were tested

using the VarTrace assay, of which 60 had known MSI status from vendor. Successful

DNA extraction, quantitation, and qPCR testing was achieved in the first round for all 53

samples extracted in-house. The remaining 48 samples from Discovery was extracted

at an external provider, redundant DNA were sent back and stored at 4c for 2 years.

Rehydration, centrifugation and magnetic separation were introduced to samples that

have little liquid or with bead residues. All 48 samples were successfully tested after

necessary procedure and DNA quantitation. No failed or invalid results occurred during

screening of the 101 samples, and no retested needed.

The 101 samples encompassed a wide range of tumor percentages from 10% to 90%

(median 40%). The VarTrace assay detected MSI-H in 47 samples, with tumor

percentages ranging from 15% to 80% (median 45%) (Fig. 2A). Most MSI-H samples

(40/47) were positive for ≥4 markers, while the remaining 7 samples with <50% tumor

content were positive for only 2-3 markers (Fig. 2C). The MONO-27 marker showed no

amplification in MSI-negative samples, while BAT-25 showed positive signals in all MSI-

positive samples, indicating potential for optimization of marker specificity.

Comparison of VarTrace Assay to Promega OncoMate MSI Analysis

The MSI status of the 101 FFPE samples was also tested using the Promega

OncoMate assay at a third-party CLIA lab (Supplemental Table S5 and Supplemental

Fig. S3-10). However, 6 samples showed invalid results due to marker readout failure,

yielding a sampe test success rate of 94%. Of the remaining 95 samples, the VarTrace

and OncoMate assays demonstrated significant concordance, with only 1 sample

scored as MSI-H by VarTrace but MSI-L by OncoMate. The one discrepant sample was

confirmed as MSI-H by the vendor report, suggesting OncoMate likely missed low-

frequency MSI signals due to limitations in its pre-amplification efficiency and analysis

algorithm (Supplemental Fig. S11). Considering both MSI-L and MSS as negative and

MSI-H as Positive, the VarTrace assay demonstrated 99% concordance, 100% (95% CI

92%~100%) sensitivity and 98% (95% CI 89%~100%) specificity compared to

OncoMate (Fig. 2D). Comparison of individual marker results between the VarTrace and

Oncomate assays was also performed. The VarTrace assay demonstrated a specificity

above 95% for 3 markers (NR-21, BAT-26, MONO-27) and a sensitivity and

concordance above 95% for 3 markers (BAT-25, BAT-26, MONO-27).

Comparison of VarTrace and OncoMate Assays to Vendor Validation

Of the 101 samples, 60 had MSI status validated by the commercial vendor, with 44

determined to be MSI-H. Compared to the vendor results, VarTrace concordantly

classified 59/60 samples, with one sample scored as MSS by VarTrace but MSI-H by

the vendor (Supplemental Table S5 and S6). This yielded a sensitivity of 98% (95% CI

88%~100%) and specificity of 100% (95% CI 79%~100%) for VarTrace.

Similarly, the OncoMate assay has 2 samples scored negative when vendor claimed to

be MSI-H. After excluding 3 invalid samples from the validation dataset, the OncoMate

assay demonstrated 95% sensitivity (95% CI 84%~99%) and 100% specificity (95% CI

77%~100%) versus vendor validation (Supplemental Table S6). Therefore, the

VarTrace assay showed slightly better concordance with vendor results compared to

OncoMate.

Determination of Tumor Percentage Limit of Detection

Serial dilutions using wildtype DNA to make contrived samples at different tumor percentage were performed on 41 MSI-H samples (Fig. 3A and Supplemental Fig. S12-32), with each sample tested once at each dilution due to limited sample amount. The limit of detection of one sample is defined as the smallest tumor percentage that has positive result, and all tests with higher tumor percentage also report as positive. When evaluating limit of detection (LoD) of each individual sample, a tumor percentage LoD below 5% was achieved in 35/41 samples, with only 1 sample showing LoD above 10% (Fig. 3B and Supplemental Table S7). This sample originally carries a tumor percentage of 30% and tested positive in 3 markers, which may have sensitive marker performance and lose marker signal during serial dilution. All samples with 5 markers positive in undiluted DNA tests have tumor percentage LoD no bigger than 5%. Similarly, all samples with 4 markers positive at original testing showed tumor percentage LoD no bigger than 10%.

A total of 246 tests across varying tumor percentages were included in the analysis to evaluate the accuracy of the VarTrace assay at different tumor content levels, of which 41 tests using original undiluted sample are included. All results were considered equally, regardless of differences in starting material used to create the contrived diluted specimens. The positivity rate was calculated at different tumor percentage ranges to assess the performance of the VarTrace assay as tumor content decreased. VarTrace Assay demonstrated great concordance in detecting positive samples under low tumor percentage with a 100% positivity for samples with ≥15% tumor, 97% positivity from 10%

to15%, and 92% positivity from 5% to10% tumor percentage (Fig. 3C). Additionally, 5

tests were scored MSI-H at tumor percentages below 1% (Supplemental Table S8).

These results indicate reliable MSI detection by the VarTrace assay even at low tumor

content.

Analytical Performance of VarTrace Assay Using Synthetic Samples

The analytical sensitivity and specificity of the VarTrace assay was evaluated using two

types of synthetic samples: Seraseq® MSI Reference Panel Mix AF5% and internally

generated VarTrace positive controls (5%VAF) and negative controls. The VarTrace

assay demonstrated 100% sensitivity and 100% specificity in detecting all 5 markers

when tested with both 2 ng input (1 ng per reaction tube) of the Seraseq reference

samples and 10 ng of the internal controls (Table 3, Supplemental Table S9 and Fig.

S33-34). No false positive results occurred for any markers in the negative control

samples in either analytical study.

To assess the limit of blank, synthetic wild-type (WT) fragments for 5 markers were

tested in the following groups: WT (0), WT (+3) with 3 nt addition, and WT (-3) with 3 nt

deletions in the mononucleotide repeat region. Across 12 repeats of each WT group, no

WT genotypes demonstrated positive signals for any markers, indicating 100% correct

negative scoring within the wild-type allele length range (Table 3, Supplemental Table

S10 and Fig. S35-36).

The limit of detection was evaluated using variants with 7 nt deletions in the

mononucleotide repeat regions of all 5 markers (MSI-7 variant pool) at VAFs ranging

from 1% to 50%. These variants were designed to be near the WT allele length range

and difficult to distinguish from WT. The results showed the VarTrace assay could

detect and score all experimental repeats containing the -7 variant at 5% VAF as

positive (Table 3, Supplemental Table S11 and Fig. S37-40). This is consistent with

detection of the BAT-25, NR-24 and MONO-27 variants with 6 nt deletions present at 5%

VAF in the Seraseq reference sample from the analytical study.

The limit of input was assessed using the Seraseq Positive Reference (5% VAF) and

MSI-7 variant pool (5% VAF), as sensitive samples representing low tumor content with

short microsatellite deletions. Input amounts of 1 ng to 0.1 ng per tube of these 2

samples were tested in 6 repeats. The VarTrace assay yielded positive results for all 5

markers in all repeats with as little as 1 ng input (0.5 ng per tube) for both samples at 5%

VAF (Table 3, Supplemental Table S12 and Fig. S41-42). Evaluating MSI classification,

the VarTrace assay detected MSI-H status with occasional marker drop-out at all

repeats of both samples down to 0.5 ng input (0.25 ng per tube).

Discussion

VarTrace is a qPCR assay that can detect MSI status using 5 markers from Bethesda

Guidelines directly from tumor sample DNA without matching normal sample. Our study

demonstrated 100% sensitivity and 98% specificity for the VarTrace assay compared to

the OncoMate assay across 101 FFPE samples. VarTrace also showed better

performance than OncoMate when benchmarked to vendor pathology reports. Beyond

comparable accuracy and sample requirement benefit, VarTrace also showed

advantages in industrial adaptability over OncoMate. Unlike the two-step pre-

amplification and capillary electrophoresis process required for OncoMate, the VarTrace

assay platform enables direct qPCR readout from the sample. This simplification of the

workflow reduces instrumentation requirements, as shown through comparison of the

timelines in Fig. 2E. This enabled VarTrace to be completed in 8 hours, faster than the

10+ hour OncoMate timeframe and better suited for shift work. The higher throughput

scalability of qPCR vs capillary electrophoresis also enables VarTrace to process 3-fold

more samples per run (Table 2).

Studies of tumor percentage dilution and analytical performance using synthetic

samples showed the VarTrace MSI Assay could detect MSI with 97% positivity at tumor

content as low as 10% and DNA input as little as 0.5 ng.

However, when markers were considered individually, the sensitivity and specificity of

NR-24 was lower at 80% and 90% respectively compared to OncoMate. Discordant NR-

24 results showed deletions of 4-7 nucleotide mononucleotide repeats. In an additional

limit of detection study using an MSI-5 10% variant pool (Supplemental Table S11),

VarTrace also demonstrated 2 false negatives for NR-24 out of 6 repeats. This indicates

less sensitive detection of short deletions less than 7 nt by NR-24 in the VarTrace assay.

However, NR-24 did show positive amplification without reaching the calling threshold in

these discrepant samples. Re-evaluating the NR-24 calling threshold could potentially

improve accuracy of this marker.

Next-generation sequencing (NGS) panels can concurrently detect MSI status and

alterations in MMR genes from solid tumor samples. FoundationOne CDx (24), as a

representative NGS panel, became the first FDA-approved NGS test for pan-cancer

biomarker profiling and tumor mutational burden, including multiple microsatellite

markers to determine MSI status. It also enables discovery of novel microsatellite

markers across cancer types (3, 4, 20, 24, 26, 28), ovarian cancer (24) for instance.

While NGS provides additional benefits, its use for first-line MSI testing in colorectal

cancer may be prohibitively expensive and has limited clinical utility beyond

determination of MSI status.

The VarTrace assay platform has potential for broader applications through validation of

additional MSI markers identified by NGS, as well as alternative sample types like cell-

free DNA (cfDNA) (25, 34, 35). cfDNA enables minimally invasive MSI testing compared

to tissue biopsies, but detecting MSI in cfDNA requires even higher sensitivity due to

limited tumor DNA in plasma. VarTrace can better enrich and amplify minimal MSI

markers in cfDNA by simply suppressing *GAPDH* amplification to minimize consumption

of PCR reagents by this control gene. With further optimization and calibrated calling

thresholds, the same platform may be adapted for sensitive and specific MSI detection

from cfDNA pending clinical validation. Expanding the validated markers and sample

types would extend the utility of the rapid, robust VarTrace assay for MSI analysis

across clinical settings.

In conclusion, our findings demonstrate that the VarTrace Assay is a promising new tool

for fast, accurate and broadly applicable MSI testing across clinical settings. This assay

stands poised to redefine the standard for MSI detection, extending its relevance not

only to colorectal cancer but potentially to other applications as well. The

implementation of this assay holds the potential to facilitate Lynch syndrome screening

on a grand scale, significantly reduce turnaround times for surgical decision, and make

possible the assessment of MSI status even in constrained biopsy samples or cfDNA.

Further clinical validation and implementation studies will help translate these analytical

performance improvements into meaningful patient benefits.

Funding and Disclosures/Conflict of Interest

This work is fully funded by NuProbe USA. There is a patent pending on using BDA to detect MSI biomarker under U.S. Patent Application No. 17/499,536. Y.H.Y is the inventor and the patent applicant is NuProbe USA. D.Y.Z. declares a competing interest in the form of consulting for and significant equity ownership in NuProbe Global, Torus Biosystems, Biostate.AI, Pupil Bio and Pana Bio.

Acknowledgments

The authors thank Dr. Mingjie Dai for valuable suggestions to the manuscript. The authors wish to acknowledge the assistance of Jana Harvey and Jonathan Tran in experiment execution and material organization for this study.

References

1.   Markowitz SD, Bertagnolli MM. Molecular Basis of Colorectal Cancer. New England Journal of Medicine 2009;361:25:2449-60.

2.   Li G-M. Mechanisms and functions of DNA mismatch repair. Cell Research 2008;18:1:85-98.

3.   Hause RJ, Pritchard CC, Shendure J, Salipante SJ. Classification and characterization of microsatellite instability across 18 cancer types. Nature Medicine 2016;22:11:1342-50.

4.   Arulananda S, Thapa B, Walkiewicz M, Zapparoli GV, Williams DS, Dobrovic A, John T. Mismatch Repair Protein Defects and Microsatellite Instability in

Malignant Pleural Mesothelioma. Journal of Thoracic Oncology 2018;13:10:1588-

94.

5.   Muzny DM, Bainbridge MN, Chang K, Dinh HH, Drummond JA, Fowler G, et al.

Comprehensive molecular characterization of human colon and rectal cancer.

Nature 2012;487:7407:330-7.

6.   de la Chapelle A, Hampel H. Clinical Relevance of Microsatellite Instability in

Colorectal Cancer. Journal of Clinical Oncology 2010;28:20:3380-7.

7.   Thibodeau SN, Bren G, Schaid D. Microsatellite Instability in Cancer of the

Proximal Colon. Science 1993;260:5109:816-9.

8.   Ahadova A, Gallon R, Gebert J, Ballhausen A, Endris V, Kirchner M, et al. Three

molecular pathways model colorectal carcinogenesis in Lynch syndrome.

International Journal of Cancer 2018;143:1:139-50.

9.   Boland CR, Goel A. Microsatellite Instability in Colorectal Cancer.

Gastroenterology 2010;138:6:2073-87.e3.

10.   Ganesh K, Stadler ZK, Cercek A, Mendelsohn RB, Shia J, Segal NH, Diaz LA.

Immunotherapy in colorectal cancer: rationale, challenges and potential. Nature

Reviews Gastroenterology & Hepatology 2019;16:6:361-75.

11.   Xie Y-H, Chen Y-X, Fang J-Y. Comprehensive review of targeted therapy for

colorectal cancer. Signal Transduction and Targeted Therapy 2020;5:1:22.

12.   Sargent DJ, Marsoni S, Monges G, Thibodeau SN, Labianca R, Hamilton SR, et

al. Defective Mismatch Repair As a Predictive Marker for Lack of Efficacy of

Fluorouracil-Based Adjuvant Therapy in Colon Cancer. Journal of Clinical

Oncology 2010;28:20:3219-26.

13.   Goldstein J, Tran B, Ensor J, Gibbs P, Wong HL, Wong SF, et al. Multicenter
387   retrospective analysis of metastatic colorectal cancer (CRC) with high-level
388   microsatellite instability (MSI-H). Annals of Oncology 2014;25:5:1032-8.

389   14.   Goodman AM, Sokol ES, Frampton GM, Lippman SM, Kurzrock R. Microsatellite-
390   Stable Tumors with High Mutational Burden Benefit from Immunotherapy. Cancer
391   Immunology Research 2019;7:10:1570-3.

392   15.   Giovanni F, Romain C, Sara L, Kohei S, Maria Elena E, Marwan F, et al. Ascites
393   and resistance to immune checkpoint inhibition in dMMR/MSI-H metastatic
394   colorectal and gastric cancers. Journal for ImmunoTherapy of Cancer
395   2022;10:2:e004001.

396   16.   Tougeron D, Mouillet G, Trouilloud I, Lecomte T, Coriat R, Aparicio T, et al.
397   Efficacy of Adjuvant Chemotherapy in Colon Cancer With Microsatellite Instability:
398   A Large Multicenter AGEO Study. JNCI: Journal of the National Cancer Institute
399   2016;108:7:djv438.

400   17.   Lin A, Zhang J, Luo P. Crosstalk Between the MSI Status and Tumor
401   Microenvironment in Colorectal Cancer [Original Research]. Frontiers in
402   Immunology 2020;11.

403   18.   Bacher JW, Flanagan LA, Smalley RL, Nassif NA, Burgart LJ, Halberg RB, et al.
404   Development of a Fluorescent Multiplex Assay for Detection of MSI-High Tumors.
405   Disease Markers 2004;20:136734.

406   19.   Toh JW, Singh P, Tangirala VAASK, Limmer A, Spring KJ. A Simple, Accurate and
407   Cost-Effective Capillary Electrophoresis Test with Computational Methods to Aid
408   in Universal Microsatellite Instability Testing. Cells. 2021.

20. Long DR, Waalkes A, Panicker VP, Hause RJ, Jr., Salipante SJ. Identifying
Optimal Loci for the Molecular Diagnosis of Microsatellite Instability. Clinical
Chemistry 2020;66:10:1310-8.

21. Niu B, Ye K, Zhang Q, Lu C, Xie M, McLellan MD, et al. MSIsensor: microsatellite
instability detection using paired tumor-normal sequence data. Bioinformatics
2014;30:7:1015-6.

22. Vanderwalde A, Spetzler D, Xiao N, Gatalica Z, Marshall J. Microsatellite
instability status determined by next-generation sequencing and compared with
PD-L1 and tumor mutational burden in 11,348 patients. Cancer Medicine
2018;7:3:746-56.

23. Ratovomanana T, Cohen R, Svrcek M, Renaud F, Cervera P, Siret A, et al.
Performance of Next-Generation Sequencing for the Detection of Microsatellite
Instability in Colorectal Cancer With Deficient DNA Mismatch Repair.
Gastroenterology 2021;161:3:814-26.e7.

24. Trabucco SE, Gowen K, Maund SL, Sanford E, Fabrizio DA, Hall MJ, et al. A
Novel Next-Generation Sequencing Approach to Detecting Microsatellite
Instability and Pan-Tumor Characterization of 1000 Microsatellite
Instability–High Cases in 67,000 Patient Samples. The Journal of
Molecular Diagnostics 2019;21:6:1053-66.

25. Yu F, Leong KW, Makrigiorgos A, Adalsteinsson VA, Ladas I, Ng K, et al. NGS-
based identification and tracing of microsatellite instability from minute amounts
DNA using inter-Alu-PCR. Nucleic Acids Research 2021;49:4:e24-e.

26.   Middha S, Zhang L, Nafa K, Jayakumaran G, Wong D, Kim HR, et al. Reliable

432   Pan-Cancer Microsatellite Instability Assessment by Using Targeted Next-

433   Generation Sequencing Data. JCO Precision Oncology 2017:1:1-17.

434   27.   Salipante SJ, Scroggins SM, Hampel HL, Turner EH, Pritchard CC. Microsatellite

435   Instability Detection by Next Generation Sequencing. Clinical Chemistry

436   2014;60:9:1192-9.

437   28.   Waalkes A, Smith N, Penewit K, Hempelmann J, Konnick EQ, Hause RJ, et al.

438   Accurate Pan-Cancer Molecular Diagnosis of Microsatellite Instability by Single-

439   Molecule Molecular Inversion Probe Capture and High-Throughput Sequencing.

440   Clinical Chemistry 2018;64:6:950-8.

441   29.   Berrino E, Aquilano MC, Valtorta E, Amodio V, Germano G, Gusmini M, et al.

442   Unique Patterns of Heterogeneous Mismatch Repair Protein Expression in

443   Colorectal Cancer Unveil Different Degrees of Tumor Mutational Burden and

444   Distinct Tumor Microenvironment Features. Modern Pathology 2023;36:2.

445   30.   Palmieri G, Colombino M, Cossu A, Marchetti A, Botti G, Ascierto PA. Genetic

446   instability and increased mutational load: which diagnostic tool best direct

447   patients with cancer to immunotherapy? Journal of Translational Medicine

448   2017;15:1:17.

449   31.   Boland CR, Thibodeau SN, Hamilton SR, Sidransky D, Eshleman JR, Burt RW,

450   et al. A National Cancer Institute Workshop on Microsatellite Instability for Cancer

451   Detection and Familial Predisposition: Development of International Criteria for

452   the Determination of Microsatellite Instability in Colorectal Cancer. Cancer

453   Research 1998;58:22:5248-57.

454   32.   Funkhouser WK, Jr., Lubin IM, Monzon FA, Zehnbauer BA, Evans JP, Ogino S,

455   Nowak JA. Relevance, Pathogenesis, and Testing Algorithm for Mismatch

456   Repair–Defective Colorectal Carcinomas: A Report of the Association for

457   Molecular Pathology. The Journal of Molecular Diagnostics 2012;14:2:91-103.

458   33.   Cohen R, Hain E, Buhard O, Guilloux A, Bardier A, Kaci R, et al. Association of

459   Primary Resistance to Immune Checkpoint Inhibitors in Metastatic Colorectal

460   Cancer With Misdiagnosis of Microsatellite Instability or Mismatch Repair

461   Deficiency Status. JAMA Oncology 2019;5:4:551-5.

462   34.   Silveira AB, Bidard F-C, Kasperek A, Melaabi S, Tanguy M-L, Rodrigues M, et al.

463   High-Accuracy Determination of Microsatellite Instability Compatible with Liquid

464   Biopsies. Clinical Chemistry 2020;66:4:606-13.

465   35.   Duffy MJ, Crown J. Circulating Tumor DNA as a Biomarker for Monitoring

466   Patients with Solid Cancers: Comparison with Standard Protein Biomarkers.

467   Clinical Chemistry 2022;68:11:1381-90.

470 **Table1 Sample information**

		TOTAL	MSI-H[1]	MSI-L/ MSS[1]	NA
VARIABLES		101	45	50	6
SEX	Female	46	20	23	3
	Male	41	12	26	3
	Unknown	14	13	1	0
AGE	>=50	59	19	37	3
	<50	12	3	8	1
	Unknown	30	23	5	2
STAGE	I	12	9	3	0
	II	25	18	7	0
	III	31	13	14	4
	IV	29	2	25	2
	Unknown	4	3	1	0
TUMOR TISSUE TYPE	Primary	73	44	25	4
	Metastatic	28	1	25	2
TUMOR LOCATION	Colon	61	40	18	3
	Rectum	5	1	3	1
	Other (cecum, small intestine, stomach, liver)	32	1	29	2
	Unknown	3	3	0	0
TUMOR ORIGIN	Colon	87	41	42	4
	Rectum	7	1	4	2
	Other (cecum, small intestine, stomach)	4	0	4	0
	Unknown	3	3	0	0

1. *FFPE sample information is categorized by MSI status detected using Promega*

OncoMate MSI Analysis, as this is the current 'gold standard'.

FFPE samples are all commercially purchased from vendors, focusing on with

samples with known MSI status from BioChain and Discovery, and samples from

colorectal or other related cancer from OriGene.

Table2 Comparison of Operating Characteristics and Specification between

Promega Oncomate and NuProbe VarTrace

	Promega OncoMate	NuProbe VarTrace
Normal Sample Required	Yes	No
Instrument Setting[1]	ABI 7500 Fast (qPCR) **AND** ABI 3500 Dx (Sanger)	ABI 7500 Fast (qPCR)
Sample throughput[2]	12	46
Time to result	2-day result	Same-day result
Tumor Perc of FFPE Sample	>=30%	>=10%
Input requirement	1 ng	0.5 ~ 10 ng
Success Rate	94%	100%

1. ABI 7500 Fast (qPCR) is short for Applied Biosystems™ 7500 Fast Dx Real-

Time PCR; ABI 3500 Dx (Sanger) is short for Applied Biosystems™ 3500 Dx

Genetic Analyzer.

2. Sample throughput is calculated based on instrument specification.

ABI 3500 Dx (Sanger) throughput is 24 capillary per run, including matched

sample. So, results are available for 12 patient per run per instrument

ABI 7500 Fast (qPCR) throughput is 96 wells per run, including 2 tubes per

sample. Two Positive Control in plate will reduce the throughput. So, results are

available for up to 46 patient per run per instrument.

Table3 Analytical Performances of VarTrace using Synthetic Samples

Analytical Validation		
Samples	Seraseq MSI Reference	VarTrace Internal Ctrl
Input	2 ng	10 ng
Specificity	100% (CI: 83% to 100%)	100% (CI: 83% to 100%)
Sensitivity	100% (CI: 83% to 100%)	100% (CI: 83% to 100%)

Limit of Blank: 10 ng input			
Samples	WT (+3) 100%VAF	WT (0)	WT (-3) 100%VAF
Concordance	12/12(100%)	12/12(100%)	12/12(100%)

Limit of Detection: 10 ng input		
Concordance at Samples	Synthetic Ref Material (-5)	Synthetic Ref Material (-7)
50%VAF	6/6 (100%)	6/6 (100%)
10%VAF	6/6 (100%)	6/6 (100%)
5%VAF	4/6 (67%)	6/6 (100%)
1%VAF	0/6 (0%)	4/6 (67%)

Limit of Input		
Concordance at Input	Synthetic Ref Material (-7) 5% VAF	Seraseq MSI Reference 5%VAF
1ng	6/6 (100%)	6/6 (100%)
0.5ng	6/6 (100%)	6/6 (100%)
0.25ng	6/6 (100%)	6/6 (100%)
0.1ng	5/6(83%)	6/6 (100%)

Figures and Figure Legend

Fig. 1. Detecting microsatellite instable variants using qPCR with BDA. (A) Single plex

design diagram. Blocker sets (green) overlapping the forward primer binding site are

designed to perfectly match common wildtype (WT) sequences: WT (0, ±1, ±2, ±3). This

allows the blockers to outcompete forward primer binding to WT sequences. WT (0)

represents the reference sequence with the highest population frequency, and WT (+1)

represents sequences with a 1 nt addition in the mononucleotide repeat region relative

to WT (0). MSI variants (red) create a mismatch that enables forward primer binding and

PCR amplification. (B) qPCR detection of NR-21 variants in synthetic MSI-H samples

with different variant allele frequencies (VAFs). WT (0) and no-template controls are

shown in solid gray and dashed black, respectively. All variants and WT curves are

tested and showed in triplicate, NTC is a single test. (C) Overview of the VarTrace

assay composition and coverage. No blocker for *GAPDH*. Frequencies of genotypes

covered by WTs (-3 to +3) were determined from population sequence data reported in

previous publication (18).

Fig. 2. MSI Results of 101 FFPE tumor samples. Details on sample information are provided in the Methods and Table 1. (A) Tumor percentage distribution across all samples and MSI-H subset. Grey shading indicates samples below the 30% tumor percentage, which is below the Promega OncoMate specification for input samples. (B) Representative qPCR curves from an MSI-H sample (C504247) showing positive detection of all 5 markers, tube1 can detect 3 markers and the rest 2 markers are detected in tube2. *GAPDH* is present in both tubes and used as reference control. (C) Summary heatmap of MSI results from VarTrace for the 5 markers and overall MSI status. Samples are grouped by Positive (MSI-H) and Negative (MSI-L/MSS) status and sorted by decreasing tumor percentage within each group. White indicates positive detection or MSI-H status; black indicates no detection or MSI-L/MSS status. Sample C504247 in (B) is highlighted in magenta. (D) Concordance of MSI status between the

524   95 samples successfully tested by both VarTrace and OncoMate methods. 6 samples

525   failed because of loss of marker signal in OncoMate tests. Diagnostic sensitivity (95%

526   CI) is 100% (92%~100%), and diagnostic specificity (95% CI) is 98% (89%~100%). (E)

527   Timeline comparison of workflow steps between VarTrace and OncoMate assay. See

528   Table 2 for additional assay specification comparisons.

Fig. 3. Tumor Percentage (TP) Limit of Detection (LoD). (A) Contrived sample dilution demonstration. Wildtype genomic DNA from cell line GM25485 was used as the MSS background (yellow tube) for all contrived sample dilutions. DNA extracted from an FFPE slide (C504247, 40% tumor percentage) was quantified by *GAPDH* in the original VarTrace test (Fig. 2B) and used as the starting MSI-H material (purple tube). The WT gDNA and FFPE DNA were mixed to create contrived samples at desired tumor percentage (blue tube). Serial dilutions were made to generate LoD datasets for this sample, with each dilution tested once. LoD was defined as the lowest tumor percentage at which all samples with ≥ LoD tumor percentage scored as MSI-H. (B) TP LoD determined by diluting 41 MSI-H samples with WT sample. Dot colors indicate number of positive markers detected from the undiluted sample. (C) Positivity of VarTrace assay across samples grouped into tumor percentage brackets, with number of samples per bracket showed in bar graph.

Ethic Statement

This study was conducted in accordance with the Declaration of Helsinki. The research protocol was approved by the Institutional Review Board at NuProbe USA Inc (IRB number: NP257).

Preprint Statement

This work has also submitted as preprint at medRxiv.

Link: https://medrxiv.org/cgi/content/short/2023.11.07.23298217v1

DOI: https://doi.org/10.1101/2023.11.07.23298217

A.

Fwd Primer

Blocker Sets   TaqMan

WT
Template

Rev Primer

MSI
Template

Extension **–**
Signal **–**

Extension **+**
Signal **+**

NR-21 Blocker Sets

$A_{17}A$	-3
$A_{17}AA$	-2
$A_{17}AAA$	-1
A(21) $A_{17}AAAA$	0
$A_{17}AAAAA$	+1
$A_{17}AAAAAA$	+2
$A_{17}AAAAAAA$	+3

B. NR-21 Single-plex qPCR

Yield vs Cycles

— 50% VAF
— 10% VAF
— 2% VAF
— 0.5% VAF
— 0.1% VAF
— WT
-- NTC

C.

	Channel	WildType Target : 0	Tube	WT coverage in population		
				White	Asian	African
NR-21	FAM	A(21)	1	98%	94%	99%
NR-24	HEX(SUN)	A(23)	1	98%	98%	99%
BAT-26	Cy5	A(26)	1	98%	98%	90%
BAT-25	Cy5	A(25)	2	98%	98%	90%
MONO-27	FAM	A(27)	2	98%	98%	99%
GAPDH	ROX	Reference Control, both tube 1 and 2				

A. WT
1 ng/uL
TP = 0

MSI-H
1 ng/uL
original TP = 40%

15uL

5uL

Contrived Sample
1 ng/uL

TP = $\frac{5 \times 40\%}{5 + 15}$

= 10%

VarTrace
TP LoD Test

25%	MSI-H
10%	MSI-H
5%	MSI-H
2.5%	MSI-H

LoD of Sample