Capture of Group A Streptococcus by Open-Microfluidic
CandyCollect Device in Pediatric Patients

Wan-chen Tu, MS,[*a] Ingrid Jeacopello, BS,[*a] Andrea Blom,
BSN, RN,[b] Elena Alfaro, BA, MDiv,[b] Victoria A. M.
Shinkawa, BS,[a] Daniel B. Hatchett, BS,[a] Juan C. Sanchez,
BS,[a] Anika M. McManamen, MS,[a] Xiaojing Su, MS,[a] Erwin
Berthier, PhD,[a] Sanitta Thongpang, PhD,[a, c] Ellen R. Wald,
MD, F.A.A.P, [§b] Gregory P. DeMuri, MD, F.A.A.P, [§b]
Ashleigh B. Theberge, PhD[§a,d]

Affiliations: [a]Department of Chemistry, University of Washington, Seattle, WA,
USA; [b]Department of Pediatrics, University of Wisconsin School of Medicine and
Public Health, Madison, WI, USA; [c]Department of Biomedical Engineering, Faculty
of Engineering, Mahidol University, Nakorn Pathom, Thailand; [d]Department of
Urology, School of Medicine, University of Washington, Seattle, WA, USA

[*]Contributed equally as co-first authors
[§]Co-senior authors

Address correspondence to: Ashleigh B. Theberge, Department of Chemistry, University
of Washington, Box 351700, Seattle, WA 98195-1700, [abt1@uw.edu], 206-685-2330.

Manuscript word count: 2948

Key Points

Question: In pediatric patients with Group A Streptococcus pharyngitis, how do test results and user experience compare across three sampling methods—CandyCollect devices, mouth swabs, and pharyngeal swabs?

Findings: In this cohort study of 30 children, aged 5-14 years, saliva samples were collected with CandyCollect devices and mouth swabs and analyzed via qPCR. The results show CandyCollect, a pathogen collection tool preferred by children, had 100% concordance with the results from pharyngeal swabs positive with a rapid antigen detection test performed as part of their clinical care.

Meaning: With further development and testing, the CandyCollect device may potentially become an alternative sampling tool for the diagnosis of streptococcal pharyngitis.

Abstract

Importance: Obtaining high-quality samples to diagnose streptococcal pharyngitis in pediatric patients is challenging due to discomfort associated with traditional pharyngeal swabs. This may cause reluctance to go to the clinic, inaccurate diagnosis, or inappropriate treatment for children with sore throat.

Objective: Determine the efficacy of using CandyCollect, a lollipop-inspired open-microfluidic pathogen collection device, to capture Group A Streptococcus (GAS) and compare user preference for CandyCollect, conventional pharyngeal swabs, or mouth swabs among children with pharyngitis and their caregivers.

Design: Participants of this cohort study were recruited over a 7-month period in 2022 – 2023.

Setting: This study was conducted at an ambulatory care clinic that serves pediatric patients in the Madison, Wisconsin, metropolitan area.

Participants: Study participants were diagnosed with GAS pharyngitis using a traditional pharyngeal swab via rapid antigen detection test (RADT); those testing positive were approached or reached out to about participation in the study. A total of 74 caregiver/children dyads were contacted about the study: 23 declined to participate; 21 were not eligible; and 30 willing and eligible participants were admitted into the study. A caregiver provided verbal consent and parental permission, and all children provided verbal assent. Immediately after the standard of care visit in which the throat swab was obtained, a research nurse guided participants through collecting oral samples: CandyCollect device and mouth swab (ESwab[TM]). CandyCollect and mouth swab samples were analyzed for GAS by quantitative polymerase chain reaction (qPCR) at the University of Washington.

Exposure: Detection of salivary GAS using qPCR analysis of samples obtained from CandyCollect devices and mouth swabs.

Main Outcome(s) and Measure(s): The proportion of pediatric patients with GAS pharyngitis, as determined by a positive pharyngeal swab tested via a RADT, who were also positive using a CandyCollect and mouth swab analyzed by qPCR.

Results: All child participants (30/30) were positive for GAS by qPCR on both the mouth swab and CandyCollect. Caregivers ranked CandyCollect as a good sampling method overall (27/30), and all caregivers (30/30) would recommend the CandyCollect for children 5 years and older. Twenty-three of 30 children "really like" the taste and 24/30 would prefer to use the CandyCollect if a future test was needed. All caregivers (30/30) and most children (28/30) would be willing to use the CandyCollect device at home.

Conclusion and relevance: All participants tested positive for GAS on all three collection methods (pharyngeal swab, mouth swab, and CandyCollect). While both caregivers and children like the CandyCollect device, some caregivers would prefer a shorter collection

time. Future work includes additional studies with larger cohorts presenting with

pharyngitis of unknown etiology and shortening collection time, while maintaining the

attractive form of the device.

Trial Registration:

Registry name: ClinicalTrials.gov

ClinicalTrials.gov Identifier: NCT05175196

Weblink: https://classic.clinicaltrials.gov/ct2/show/NCT05175196

Conflict of Interest Disclosures (includes financial disclosures): Ashleigh
B. Theberge, Xiaojing Su, Erwin Berthier, and Sanitta Thongpang filed patent
63/152,103 (International Publication Number: WO 2022/178291 Al) through
the University of Washington on the CandyCollect oral sampling device.
Ashleigh B. Theberge reports filing multiple patents through the University of
Washington and receiving a gift to support research outside the submitted
work from Ionis Pharmaceuticals. Erwin Berthier is an inventor on multiple
patents filed by Tasso, Inc., the University of Washington, and the University
of Wisconsin. Sanitta Thongpang has ownership in Salus Discovery, LLC, and
Tasso, Inc. Erwin Berthier has ownership in Salus Discovery, LLC, and Tasso,
Inc. and is employed by Tasso, Inc. However, this research is not related to
these companies. Sanitta Thongpang, Erwin Berthier, and Ashleigh B.
Theberge have ownership in Seabright, LLC, which will advance new tools
for diagnostics and clinical research, potentially including the CandyCollect
device. The terms of this arrangement have been reviewed and approved by
the University of Washington in accordance with its policies governing
outside work and financial conflicts of interest in research. The other authors
have no conflicts of interest to disclose.

Funding/Support: All phases of this study were supported by National
Institutes of Health grants (R21AI166120, R35GM128648 (the latter
specifically supported some of the in-lab developments and procedure
developments)), the Washington Research Foundation, the Camille and Henry
Dreyfus Foundation, and an Alfred P. Sloan Research Fellowship. The content
is solely the responsibility of the authors and does not necessarily represent
the official views of the National Institutes of Health or other funding sources.

Contributors Statement Page

Wan-chen Tu designed the sample processing protocol, carried out the sample analysis,
performed data analysis, made figures for the manuscript, drafted the initial manuscript,
and critically reviewed and revised the manuscript.

Ingrid Jeacopello fabricated the CandyCollect devices and prepared devices for human
subject research, interpreted the data, drafted the initial manuscript, and critically reviewed
and revised the manuscript.

Andrea Blom advised on study design and execution, enrolled and interacted with the
participants, collected the samples and user surveys, accessioned and shipped the samples,
interpreted the data, and critically reviewed and revised the manuscript.

Elena Alfaro advised on study design and execution, advised on all regulatory aspects,
drafted the study protocol and survey instruments, obtained Institutional Review Board
(IRB) approval, interpreted the data, and critically reviewed and revised the manuscript.

Victoria A. M. Shinkawa, Daniel B. Hatchett, and Juan C. Sanchez fabricated the
CandyCollect devices and prepared devices for human subject research, including
designing experimental protocols/engineering designs, and critically reviewed and revised

148   the manuscript.
149
150   Anika M. McManamen and Xiaojing Su designed the sample processing protocol and
151   drafted and critically reviewed and revised the manuscript.
152
153   Erwin Berthier conceptualized the research, advised on study design and interpretation,
154   and critically reviewed and revised the manuscript.
155
156   Ashleigh B. Theberge, Gregory P. DeMuri, Ellen R. Wald, and Sanitta Thongpang
157   conceptualized the research, designed the study, oversaw study execution, interpreted the
158   data, and critically reviewed and revised the manuscript.
159
160   All authors approved the final manuscript as submitted and agree to be accountable for all
161   aspects of the work.
162
163

Introduction

Group A Streptococcus (GAS) accounts for over 600 million cases of bacterial

pharyngitis worldwide each year,[1] affecting 15% of school-aged children in developed

countries. In developing nations the burden is markedly greater, with rates 5-10 times

higher.[1] Though symptoms of GAS pharyngitis almost always resolve after 3-4 days,

antibiotics are recommended to shorten the course of the infection, lessen spread, and

prevent acute suppurative complications.[1] Accurate detection of GAS is imperative;

untreated and recurring GAS infections can lead to severe outcomes, such as acute

rheumatic fever, which causes carditis in 30-45% of cases and is the most common cause of

acquired heart disease in children worldwide.[1]

A standard method for diagnosing GAS pharyngitis is with a rapid antigen detection

test (RADT), which requires a posterior pharyngeal swab sample and provides a positive or

negative result within minutes.[2] If the RADT is negative, a throat culture is performed on a

posterior pharyngeal swab in order to detect false negatives, and laboratory analysis takes

more time.[2,3] Another option is a molecular test (e.g., quantitative polymerase chain

reaction (qPCR)), again requiring a posterior pharyngeal swab.[4] Pharyngeal swabs are

uncomfortable, often acting as a deterrent for children and adults; however, it is important

to confidently diagnose bacterial pharyngitis not only for the long-term health of the patient

but also to avoid over-prescribing antibiotics.[1] DeMuri, et al. established that mouth swabs

(sucked on like a lollipop) provide an adequate sample for the detection of GAS when

sensitive methods such as PCR are used.[5]

With developments in PCR and RADTs using saliva instead of posterior pharyngeal

swabs, more options are now available for salivary detection, such as spit tubes

(OMNIgene™), passive drooling, mouth swabs (ESwab™), and cotton rolls (Salivette®).[6-8]

The CandyCollect, a lollipop-inspired open-fluidic sampling device, was first introduced in

our 2021 in-lab study that showed its ability to capture and accumulate *Streptococcus*

pyogenes for analysis.[9] In 2022, we tested the device with a remote, at-home human

subjects study in which the CandyCollect devices effectively captured commensal bacteria

(*Streptococcus mutans* and *Staphylococcus aureus*) from healthy adults.[10] The

CandyCollect devices were shipped to the lab in dry tubes under ambient conditions to be

quantified using qPCR from the elution fluid.[10] Ours is not the only lollipop-inspired device;

there are at least three other collection or testing devices with a similar lollipop form

already on the market, although these alternative devices do not look like a traditional

candy lollipop and also do not provide flavor. The Self-LolliSponge™ by Copan has the

user keep the sponge stick in their mouth for a "few minutes" with a lemon scented cap to

improve the experience.[11] The V-Chek COVID-19 Saliva Antigen Test provides a

"'Lollipop' Saliva Swab" that the user places in their mouth for 90 seconds and then

attaches to a lateral flow assay cassette.[12] The Whistling COVID-19 Saliva Antigen Rapid

Test combines an absorbent material attached to a handle that the user places in their mouth

with an immunochromatographic assay.[12]

Obtaining high-quality samples to diagnose GAS pharyngitis in pediatric patients

is important for treatment; however, the current collection tools are not child friendly.

Here we aimed to demonstrate the feasibility of the CandyCollect device by collecting

salivary samples from 30 pediatric patients aged 5-17 years with known GAS pharyngitis.

Further we aimed to compare user preference for CandyCollect, conventional pharyngeal

swabs, or mouth swabs among children with pharyngitis and their caregivers.

Method

Mouth swabs and CandyCollect devices

Swabs and media: The pharyngeal swab used in the clinical visit was a standard of care double swab, BBL™ CultureSwab™ Transport Systems by Copan (FisherScientific, Cat# B4320109). The white-capped ESwab™ with nylon flocked swabs and Liquid Amies media collection kits by Copan (FisherScientific, Cat# R723480) were used as mouth swabs.

CandyCollect device stick fabrication: The CandyCollect device is milled from a sheet of 4 mm polystyrene (Goodfellow, Cat# 700-272-86) using the DATRON Neo computerized numerical control (CNC) mill as previously described.[9] The Fusion file and an engineering drawing of the CandyCollect device are included in the Supplemental Material (Figure S1). After sanding rough edges, devices were sonicated in isopropanol (IPA) (FisherScientific, A451-4) and 70% v/v ethanol (FisherScientific, Decon™ Labs, 07-678-004).

Plasma treatment of CandyCollect devices: A Zepto LC PC Plasma Treater (Diener Electronic GmbH, Ebhausen, Germany), using oxygen gas, increased hydrophilicity of the device surface. The process includes decreasing the chamber pressure to 0.20-0.25 mbar and adding oxygen gas while 70 W voltage is applied for 5 minutes.[9,10]

Preparation of CandyCollect devices for human subjects study: CandyCollect devices were prepared as previously described with the exception of storage containers and candy mass.[9,10] Half of the patients (1-15) were sampled with CandyCollects made with 1.2-1.34 g isomalt candy, while the second half (16-30) were sampled with 0.45-0.57 g; procedures are duplicated here for convenience. In accordance with the kitchen hygiene guidelines of the Washington State Cottage Food Operations Law (RCW 69.22.040(2b-f(ii-iv))), candy was applied to CandyCollect sticks by lab members who are trained in

food safety, have a Food Worker Card (WA State), and wore gloves and a mask during

food preparation. Candy was prepared by gradually adding isomalt to water while

heating to a target temperature of 171°C, then adding food coloring. Next, the isomalt

mixture was cooled in a room temperature water bath while strawberry candy flavoring

was added. Plasma treated CandyCollect sticks were cleaned using hot water and dish

soap. After the sticks were dry, small amounts of isomalt candy were re-melted and

added to silicone molds with the CandyCollect sticks. Once the candy was cooled, the

mass was recorded, and the device was heat-sealed in a polypropylene bag. Devices were

stored in Ziploc® storage bags with a food safe desiccant (Amazon, Cat# B00DYKTS9C)

until being sent to UW-Madison for the clinical study.

Clinical Study Procedures

Subjects and Specimens

Caregiver-child dyads were recruited from an ambulatory care clinic that serves children

in Madison, Wisconsin. Children were diagnosed with GAS pharyngitis using a

traditional pharyngeal swab via RADT. The study was approved by the UW-Madison

Health Sciences Institutional Review Board (#2021-1427). Study data were collected and

managed using REDCap (Research Electronic Data Capture) hosted at the University of

Wisconsin-Madison, School of Medicine and Public Health.[13,14]

A parent or legal guardian (aka "caregiver") provided verbal consent for their own

participation in the study and caregiver permission for their child, and all children

provided verbal assent. Eligible children needed to be able to suck on swabs and the

CandyCollect devices and could not have a sensitivity to sugar-free products;

additionally, a parent or legal guardian needed to be present and enroll in the study along

with the child. Immediately after the clinic visit, in which the traditional pharyngeal

swab was obtained, a research nurse guided participants through collecting four oral

samples: two commercially available mouth swabs (ESwab™) for 10 seconds each and

two CandyCollect devices until the candy was fully dissolved (Figure 1). The samples

were stored at -80°C and shipped on dry ice prior to qPCR analysis. Race and ethnicity

information was collected based on the requirements of NIH. Race and ethnicity

information was taken from the electronic health record, where it had previously been

entered based on self-report by the parent/legal guardian.

Testing Preference Survey

After sample collection was complete, the research nurse gave a paper-based survey to

both the child and their caregiver. All survey questions are included in supplementary

materials, Table S1-2. The research nurse provided instructions about the Wong-Baker

FACES® Pain Rating Scale to all child participants, explaining that each face represents

a person who has no pain (hurt), or some, or a lot of pain (Figure S2).[15] The nurse asked

the participants to choose the face that best describes the pain they experienced during

each type of testing. The research nurse assisted children with completing the form if

they needed help reading or understanding. The research nurse gave caregivers privacy

in completing their surveys, to encourage honest responses, only answering questions as

needed. Participants were also asked to provide suggestions and free responses about the

main reasons for their preferences. All free response questions are included in Table S3.

Specimen Processing and Laboratory Analysis

The detailed specimen processing and laboratory analysis procedures were previously

described.[10] In brief, the mouth swab and CandyCollect sample collection was directed

by a research nurse. Samples were stored in -20 °C and then transferred to storage at -80

°C within 12 days. Samples were shipped to the University of Washington on dry ice and

then stored at -80 °C before processing. All laboratory procedures were performed in

accordance with Biosafety Level-2 laboratory practices and the University of

Washington Site-Specific Bloodborne Pathogen Exposure Control Plan. CandyCollect

devices and mouth swabs were processed as previously described;[10] additional details are

in the Supplemental Material. Briefly, the material on the CandyCollect device was

eluted by phosphate buffered saline (PBS) (Gibco™, Cat# 10010023) with 5%

Proteinase K (Thermo Scientific™, Cat# EO0491); for mouth swab samples, *S. pyogenes*

DNA was isolated using MagMAX™ Total Nucleic Acid Isolation Kit (ThermoFisher

Scientific, Cat# AM1840) according to the protocol supplied by the manufacturer. After

elution and lysis, the samples were also enriched and purified using MagMAX™ Total

Nucleic Acid Isolation Kit following the protocol supplied by the manufacturer. [10]

qPCR was performed for the detection of *S. pyogenes* genomic DNA as previously

described.[9, 10] The results were expressed as Cycle threshold (Ct, number of cycles until

the sample was detected above the threshold). Any Ct value below or equal to Ct values

for the lowest concentration of DNA standards (5 fg) on the same plate was considered

positive; this cut-off was determined from standard curves performed with genomic

DNA isolated from *S. pyogenes*.[9]

Statistics: Statistical analysis was performed using GraphPad Prism 9 software. Paired t-

tests were used to evaluate the significance of pairwise comparisons.

Preparation of positive and negative controls

Preparation of positive controls (via in-lab capture of *S. pyogenes*) and negative controls

is described in our prior publications [9, 10] and the Supplemental Material. In brief, *S.*

pyogenes was prepared from *Streptococcus pyogenes* Rosenbach (American Type

Culture Collection, ATCC®, Cat# 700294™), and cultured in Todd-Hewitt Yeast

broth.[16] *S. pyogenes* culture was inoculated one day prior to experiments. For positive

controls, 50 µL of *S. pyogenes* culture media were added to the CandyCollect devices,

and mouth swabs were soaked in the *S. pyogenes* culture media for 10 seconds. PBS was

used as the negative control sample.

Results

Demographic Characteristics

In this study, 30 pediatric child-caregiver dyads were enrolled at a clinic in Madison,

Wisconsin. Enrolled child participants were aged 5 to 14 years with GAS pharyngitis,

which is consistent with the age range in which GAS pharyngitis is most common;[1] the

mean age of participants was 8.8 years (Figure S3 and Table S4). More than 90% of

child participants were white and not Hispanic. (Table S4).

qPCR analysis of GAS in CandyCollect devices and mouth swabs

Each sample was analyzed using qPCR and the Ct is shown in Figure 2. The samples

collected by CandyCollect devices have an average Ct of 31.28 (range=19.1-38.6) and

mouth swabs have an average Ct of 32.08 (range=19.5-38.1); the Ct from the two

sampling methods are not significantly different ($p>0.05$). Nine samples collected by

CandyCollect devices have a higher Ct than the corresponding mouth swabs and the Ct

of 13 samples collected by CandyCollect devices is lower than the corresponding mouth

swabs. Two different versions of the CandyCollect device (with 0.5 g and 1.3 g candy)

were used, with half of the participants using each size. The comparison of the Ct values

from two different sizes of candy, shown in Figure S4, are not significantly different

(p>0.05). There is also no notable correlation between the CandyCollect device

consumption time and the Ct values (Figure S5).

User Feedback of CandyCollect devices and two other commercial collection tools

All 30 children and their caregivers completed feedback surveys. Twenty-nine of 30

children responded that they did not feel any pain or discomfort while using the

CandyCollect devices; in comparison, more than half of participants felt pain/discomfort

with the traditional throat swabs (Figure 3). Twenty-seven of 30 caregivers ranked the

CandyCollect as a good sampling method overall; 29/30 caregivers indicated that using

the CandyCollect device was a pleasant experience for their child (Figure 3). The

majority of children reported that the CandyCollect device is easy to use (17/30), and

they "really like" the taste (n= 23/30) (Figure 4). The majority of caregivers reported that

the appearance of the CandyCollect was appealing (n= 26/30) (Figure 4). When asked to

consider using CandyCollect, mouth swab, or throat swab if they needed another GAS

pharyngitis diagnostic test next week, more than a half of the children (24/30) and

caregivers (17/30) prefer to use the CandyCollect device, with the remaining children

and caregivers preferring the mouth swab; no participants selected the throat swab

(Figure 4). For caregivers of children who consumed the 0.5 g CandyCollect device

(consumption time range 2.3-5.9 min; 3.3 min median), 12/15 preferred the

CandyCollect device and 3/15 preferred the mouth swab; for caregivers of children who

consumed the 1.3 g CandyCollect device (consumption time range 4.3-14.2 min; 6.5 min

median), 5/15 preferred the CandyCollect device and 10/15 preferred the mouth swab.

Almost all children (28 out of 30) and all caregivers (30 out of 30) stated that they would

be willing to use the CandyCollect device at home (Figure 4). All caregivers report they

would recommend using the CandyCollect device (Figure 4) for children 5 years and

older. All survey questions are included in table S1-2.

Discussion

The ultimate goal of the CandyCollect device is to capture bacterial pathogens for

laboratory detection. In our previous in-lab study, we developed the CandyCollect device

to capture *S. pyogenes* in saliva and established a qPCR assay to detect *S. pyogenes*

eluted from CandyCollect devices.[9] We also used CandyCollect devices to capture

commensal bacteria (*Streptococcus mutans* and *Staphylococcus aureus*) from healthy

adults in a home setting as proof-of-concept human subjects research.[10] The present

publication is the first human subjects research that used CandyCollect devices with

pediatric patients. The results show that all 30 children diagnosed with GAS pharyngitis

using a RADT were positive on CandyCollect devices and mouth swabs. As expected in

studies of human subjects, there was some variability in the amount of bacteria (Ct value)

across sampling methods and participants, which was also shown in a previous study

comparing throat swabs and mouth swabs.[5]

From the user feedback surveys, many caregivers noted the CandyCollect device is less

frightening for children than other sampling methods, especially in comparison to throat

swabs. Several caregivers noted the CandyCollect is easy, fun, and non-invasive. Some

caregivers indicated that the CandyCollect device could help children be less afraid of

going to the doctor or even be excited to do the test. All survey free responses and

suggestions are included in Table S3. When asked for suggestions to improve

CandyCollect, the most common suggestion was to shorten the sampling time. We

responded to this suggestion mid-study by using a CandyCollect that afforded a shorter

sampling time for the second half of participants. As noted in the Results, caregivers

indicated a preference for either the CandyCollect or the mouth swab, with no caregivers

(and no children) preferring the throat swab. Notably, more caregivers preferred their

children to use CandyCollect devices rather than mouth swabs when their child used

smaller CandyCollect devices with a faster dissolving time (12/15) in comparison to

caregivers of children who used the larger CandyCollect devices with a longer dissolving

time (5/15). Slightly more children also preferred using CandyCollect devices when they

used faster-dissolving CandyCollect devices (13/15 vs. 11/15). Schuster et al., also

reported the same reasoning for sampling preferences: 73% of participants preferred the

nasal swab compared to saliva collection by drooling due to the faster sampling time.[17]

We are currently working to modify the CandyCollect so as to have faster dissolving

candy.

There are three major limitations for this study. First, this pilot study was performed with

a relatively small sample size; we have ongoing studies with larger sample sizes in adults

and children with GAS and other respiratory illnesses. Second, participants were

recruited from a single site with relatively homogenous demographics, although we have

no reason to suspect that demographics influence capture of pathogens in saliva. Our

ongoing work includes remote nationwide studies with enrollment demographics

matching national distribution of race/ethnicities. Finally, to maximize the number of

positive cases tested with CandyCollect in this feasibility study, we only included

children who tested positive for GAS on pharyngeal swabs analyzed with RADT during

their clinic visits; future studies will include children with pharyngitis of unknown

etiology evaluated with standard clinical and research protocols.

Conclusion

The CandyCollect device, a user-friendly pathogen collection method, is specifically

designed to make both in-clinic and at-home sampling more accessible for children. In

our inaugural pediatric clinical study, we demonstrated the CandyCollect device

successfully captured GAS from 30 children in a clinical setting. Notably, a majority of

the children and their caregivers expressed a preference for the CandyCollect device as

their chosen method for collecting GAS (compared to traditional throat swabs and mouth

swabs). This study represents a significant step towards facilitating collection of samples

for the detection of GAS and other respiratory pathogens. Currently, we are modifying

the CandyCollect device based on user feedback to optimize its performance and user

experience, and we will enroll more participants to achieve larger sample sizes of

children with GAS pharyngitis. The present study paves the way for several exciting

research directions, including: (1) integrating existing lateral flow immunoassay

techniques with the CandyCollect device to enhance the diagnostic capabilities at home

and shorten the diagnostic time, and (2) expanding the range of pathogens that can be

captured using the CandyCollect device.

Acknowledgments
We would like to thank Jason Robertson for his review of the manuscript; Lochlan
Hickok and Paul Miller for helping facilitate shipping and logistics; Karen N.
Adams advising on work with human subjects study design and regulatory protocols;
Alexandra Lindstrom and Bridget L. Johnson advising on study design and
execution. Moreover, we thank the participants enrolled in this study.

References

1. Walker MJ, Barnett TC, McArthur JD, et al. Disease Manifestations and Pathogenic Mechanisms of Group A Streptococcus. Clinical Microbiology Reviews. 2014;27(2):264-301. doi:10.1128/cmr.00101-13

2. Centers for Disease Control and Prevention Updated June 27, 2022. Accessed July10, 2023. https://www.cdc.gov/groupastrep/diseases-hcp/strep-throat.html#:~:text=Clinicians%20need%20to%20use%20either,the%20gold%20standard%20diagnostic%20test

3. Shulman ST, Bisno AL, Clegg HW, et al. Clinical Practice Guideline for the Diagnosis and Management of Group A Streptococcal Pharyngitis: 2012 Update by the Infectious Diseases Society of America. Clinical Infectious Diseases. 2012;55(10):e86-e102. doi:10.1093/cid/cis629

4. Ralph AP, Holt DC, Islam S, et al. Potential for Molecular Testing for Group A Streptococcus to Improve Diagnosis and Management in a High-Risk Population: A Prospective Study. Open Forum Infect Dis. 2019;6(4):ofz097. doi:10.1093/ofid/ofz097

5. DeMuri G and Wald ER. Detection of Group A Streptococcus in the Saliva of Children Presenting With Pharyngitis Using the cobas Liat PCR System. Clin Pediatr (Phila). 2020;59(9-10):856-858. doi:10.1177/0009922820920936

6. Bellagambi FG, Lomonaco T, Salvo P, et al. Saliva sampling: Methods and devices. An overview. TrAC Trends in Analytical Chemistry. 2020;124:115781. doi:https://doi.org/10.1016/j.trac.2019.115781

7. All-in-one system for the collection and rapid stabilization of microbial DNA and RNA. Accessed August 15, 2023. https://www.dnagenotek.com/row/products/collection-microbiome/omnigene-oral/OME-505.html

8. Melo Costa M., Benoit N, Dormoi J, et al. Al-meras, L. Salivette, a relevant saliva sampling device for SARS-CoV-2 detection. J. Oral Microbiol. 2021; 13, 1920226. doi: 10.1080/20002297.2021.1920226

9. Lee UN, Su X, Hieber DL, et al. CandyCollect: at-home saliva sampling for capture of respiratory pathogens. Lab Chip. 2022;22(18):3555-3564. doi:10.1039/d1lc01132d

10. Tu W-c, McManamen AM, Su X, et al. At-Home Saliva Sampling in Healthy Adults Using CandyCollect, a Lollipop-Inspired Device. Analytical Chemistry. 2023;95(27):10211-10220. doi:10.1021/acs.analchem.3c00462

11. Ottaviano E, Parodi C, Borghi E, et al. Saliva detection of SARS-CoV-2 for mitigating company outbreaks: a surveillance experience, Milan, Italy, March 2021. Epidemiol Infect. 2021;149:e171. doi:10.1017/s0950268821001473

12. De Meyer J, Goris H, Mortelé O, et al. Evaluation of Saliva as a Matrix for RT-PCR Analysis and Two Rapid Antigen Tests for the Detection of SARS-CoV-2. Viruses. 2022;14(9)doi:10.3390/v14091931

13. Harris PA, Taylor R, Thielke R, Payne J, Gonzalez N, Conde JG. Research electronic data capture (REDCap)—A metadata-driven methodology and workflow process for providing translational research informatics support. Journal of Biomedical Informatics. 2009;42(2):377-381. doi:10.1016/j.jbi.2008.08.010

14. Harris PA, Taylor R, Minor BL, et al. The REDCap consortium: Building an

international community of software platform partners. Journal of Biomedical
Informatics. 2019;95:103208. doi:10.1016/j.jbi.2019.103208
15. Wong-Baker FACES® Pain Rating Scale https://wongbakerfaces.org/
16. Gera K, McIver KS. Laboratory growth and maintenance of *Streptococcus pyogenes*
(the Group A Streptococcus, GAS). Curr Protoc Microbiol. Oct 2 2013;30:9d.2.1-
9d.2.13. doi:10.1002/9780471729259.mc09d02s30
17. Schuster JE, Potts J, Selvarangan R, Mast DK, Goldman JL. A COVID-19 Testing
Preference Study in Schools. Pediatrics. 2023;152(Suppl 1) doi:10.1542/peds.2022-
060352H

487
488
489

490   **Figure 1. Group A Streptococcus sampling tools used in this study.**

491
492   Pharyngeal (throat) swab, ESwab™ (mouth swab), and the CandyCollect device (from left
493   to right). The components of the CandyCollect device include: 1) strawberry candy to
494   facilitate pleasant user experience of the CandyCollect device and 2) open channels which
495   capture bacterial pathogens.
496
497

498 **Figure 2. Comparison Ct values from CandyCollect and mouth swabs of 30 pediatric**
499 **patients.**

500
(A) Pooled cycle threshold (Ct) values of samples collected from 30 child participants with
GAS pharyngitis with the mouth swabs (left, light grey) and CandyCollect devices ((right,
dark grey). Ct from the two sampling methods are not significantly different (p>0.05). (B)
individual Ct values collected from the CandyCollect devices (purple, square) and mouth
swabs (green, circle). Data are plotted by the Ct of the CandyCollect devices; the
participant numbers were randomly re-numbered. Any Ct value below or equal to Ct
values for the lowest concentration of DNA standards (5 fg) on the same plate was
considered positive; this cut-off was determined from standard curves prepared with *S.*
pyogenes cultures; all points plotted are positive results.
510
511
512
513

514 **Figure 3. Survey responses to assess the comparison of three collection methods.**

(A) Children's ranking of pain/discomfort

(B) Caregivers' ranking of collection methods

(C) Caregivers' impression of collection methods

515
516   Three collection methods in this study included traditional pharyngeal swab, mouth swab,
517   and the CandyCollect device. All survey questions are included in Table S1 and S2: (A)
518   Table S1 questions 1-3 (B) Table S2 questions 2, 4, and 6 (C) Table S2 questions 1, 3, and
519   5.

520 Figure 4. Survey responses to assess the ranking of CandyCollect devices and user
521 preferences.

(A) Caregivers' ranking for CandyCollect

(C) Willingness to do CandyCollect at home

(D) Sampling method preference

(B) Children's ranking for CandyCollect

(E) Most suitable sampling method for children in general

(F) Recommend CandyCollect to children ages 5 and above

522
523 All survey questions are included in Table S1 and S2: (A) Table S2 question 10 (B) Table
524 S1 questions 6 and 7 (C) Table S1 question 5 and Table S2 question 9 (D) Table S1
525 question 4 and Table S2 question 7 (E) Table S2 question 8 (F) Table S2 question 11.