Quality measures of two-stage newborn hearing screening: Systematic review with a Bayesian meta-analysis.

Short title: Meta-analysis of two-stage newborn hearing screening

Kirsi Manz[1*], Uta Nennstiel[2], Carola Marzi[2], Ulrich Mansmann[1¶], Inken Brockow[2¶]

[1]Institute for Medical Information Processing, Biometry and Epidemiology (IBE), Faculty of Medicine, LMU Munich, Munich, Germany

[2]Bavarian Health and Food Safety Authority, Oberschleißheim, Germany

*Corresponding author

E-mail: manz@ibe.med.uni-muenchen.de (KM)

[¶]These authors contributed equally to this work.

Abstract

Background: Newborn Screening for hearing impairment (NHS) is a crucial public health issue worldwide. Often, a two-stage screening with two different testing approaches is used. We aimed to investigate the optimal screening algorithm, based on data from the literature published in the last 30 years. A particular focus of the study was to synthesize the existing evidence on two-stage newborn hearing screening regarding the refer rate (RFR), the percentage of children that did not pass the second test or were lost after the first test.

Methods: We searched MEDLINE for studies on two-stage NHS using transitory evoked otoacoustic emissions (TEOAE) or automated auditory brainstem response (AABR). All studies on newborns who received their first test as an inpatient and a second test up to one month later were eligible. Random effects meta-analysis and Bayesian modeling were performed to estimate RFR, effects of the second test phase on the RFR, and sensitivity and specificity of TEOAE and AABR, respectively. Risk of bias was assessed using QUADAS-II. The unfunded study was registered in PROSPERO (CRD42023403091).

Results: Fifty-seven study protocols, including over 704,000 newborns, met the inclusion criteria. Certainty in the evidence was rated as moderate. The RFR was higher when the test method was changed than without a change of method (AABR-AABR: RFR = 1.4% (95% confidence interval (CI): 0.9, 2.2%), TEOAE-TEOAE: RFR = 2.5% (CI: 2.0, 3.1%), TEOAE-AABR: RFR = 4.9% (CI: 3.1, 7.5%), AABR-TEOAE: RFR = 5.9% (CI: 5.0, 6.9%). Across all protocols, both methods demonstrated high sensitivity (AABR: 98.9% (95% credibility interval (CrI): 96.0, 100.0%), TEOAE: 96.9% (CrI: 92.5, 99.8%)) and high specificity (AABR: 92.7% (CrI: 92.5, 92.9%), TEOAE: 91.3% (CrI: 91.2, 91.4%)).

Conclusions: Strategies that did not involve changes to the screening method had lower RFR. Although both methods demonstrated high sensitivity, AABR appears to have slightly higher specificity compared to TEOAE.

Introduction

A properly functioning auditory system is essential for a child's acquisition of spoken language. Early intervention is critical for age-appropriate spoken language development in children with hearing impairment [1-5]. Therefore, the Joint Committee on Infant Hearing recommends that interventions begin no later than three months of age [6]. Since the late 1990s, two screening methods have been available which allow a very early diagnosis: the measurement of transitory evoked otoacoustic emissions (TEOAE) and automated auditory brainstem response (AABR) [7]. With these testing methods, a prevalence of about 1,3 per 1,000 newborns [8], good treatment options with hearing aids, cochlear implants and early intervention, and a positive cost-benefit balance [9-11] congenital hearing disorders are suitable as a target disease for newborn screening [12, 13]. Universal newborn hearing screening (UNHS) was included in standard care in Germany in 2009, as specified by the Federal Joint Committee (G-BA) in §§ 47 to 57 of the Pediatrics Directive [14]. The Pediatrics Directive outlines a two-step screening algorithm in well babies, beginning with initial TEOAE measurement in both ears, followed by bilateral AABR measurement if the initial test is not passed, and defines quality criteria. An important quality factor of a screening program is the refer rate (RFR). In the NHS, this refers to the percentage of screened infants who require referral to a pediatric audiologist for further diagnostic examinations after not passing the screening tests (positive screening). This includes newborns who are lost-to-follow-up after not passing the initial test, as well as those who receive a "fail" result on the second test. False positive findings are of particular concern as they can cause anxiety for affected families and increase the demand for scarce resources in pediatric audiology practices and outpatient clinics.

Evaluations of the German NHS for the years 2011/12 [8] and 2017/18 [15, 16] have shown that the recommended screening-algorithm is often not followed. In over 50% of cases, where infants do not pass the first test, a second TEOAE measurement is performed instead of the required AABR. Additionally, this second TEOAE test yielded a "fail" result in only about 10% of cases, compared to 20% for infants who underwent a second test using AABR. Upon analyzing these data, the failure rate of the second test was particularly high when the screening method was altered. The second test showed the lowest failure rate with the TEOAE / TEOAE algorithm at 9.62%, while the highest rate was observed with TEOAE / AABR at 26.59%.

Accordingly, international recommendations suggest performing a second TEOAE test after the initial TEOAE result was "fail" in newborns without risk factors for hearing impairment ("well-babies") [6, 17].

This study reviewed current literature to investigate the quality of available screening tests and investigate the optimal screening algorithm based on data published in the last 30 years. The study specifically focused on synthesizing existing evidence related to the RFR in two-stage NHS.

Materials and methods

Model

The first step involved formalizing the two-stage screening process on which our meta-analysis is based. Then, a meta-regression was conducted to assess the quality of the four possible testcombinations (TEOAE-TEOAE, TEOAE-AABR, AABR-TEOAE, AABR-AABR). Additionally, a meta-analysis of the positive rates in both stages was performed to infer the sensitivities and specificities of both methods.

The study population comprised of newborns with relevant hearing impairment ("diseased", D^+) and without it ("healthy", D^-). The prevalence of hearing impairment in the population is denoted by π. The first stage of screening is performed using a method with a sensitivity of SE1 and specificity of SP1. Therefore, the test positivity rate of the first stage (PR1), which includes true positives (TP1) and false positives (FP1), is given by: PR1 = TP1 + FP1 = π*SE1 + (1-π)*(1-SP1). These newborns are considered to have "*failed*" and will receive a second test. The rate of positive results from the first stage is also referred to as the "failure rate".

The proportion of newborns whose first test result was "pass" and who leave the screening procedure is the negative rate NR1 = π*(1-SE1) + (1-π)*SP1. However, not all newborns with a positive first test proceed to the second stage: The proportion of ρ of positively tested newborns is *lost*. It is assumed that this loss is independent of the hearing status (the newborns drop out for reasons unrelated to the first test result). Thus, a proportion of (1-ρ) newborns undergo the second stage screening test.

The proportion of SE2 is identified as positive among the true positives of the first test TP1 = π*SE1, Accounting for the loss this yields TP2 = π*SE1*(1-ρ)*SE2. Similarly, the proportion of (1-SP2) is falsely identified as positive among the false positives of the first test FP1 = (1-π)*(1-SP1). Accounting for the loss, we obtain FP2 = (1-π)*(1-SP1)*(1-ρ)*(1-SP2).

The proportion of newborns with "fail" results in both tests together with those who are lost after a first positive test form the refer rate (RFR):

$$\text{RFR} = \rho*[\pi*SE1 + (1-\pi)*(1-SP1)] + (1-\rho)*[\pi*SE1*SE2 + (1-\pi)*(1-SP1)*(1-SP2)]. \text{ (Eq. 1)}$$

Fig 1 shows the structure of the two-stage model together with the respective selection and loss to follow up processes.

Fig 1: Model for two-stage newborn hearing screening. π = prevalence of hearing impairment, D+ (D-) = newborns with (without) hearing impairment, T- (T+) =newborns with "pass" ("fail") test results, SE_1 (SE_2) = sensitivity of first (second) test, SP_1 (SP_2) = specificity of first (second) test, ρ = loss rate after the first test stage, L = number of newborns lost after the first test result yielded "fail", F.2=number of newborns whose second test result was "fail".

Theoretical consideration of factors influencing the refer rate
We consider the relationship between the first stage positive rate (PR1 = π*SE1 + (1-π)*(1-SP1)) and the refer rate (RFR).

$$RFR = \rho*[\pi*SE1 + (1-\pi)*(1-SP1)] + (1-\rho)*[\ \pi*SE1*SE2 + (1-\pi)*(1-SP1)*(1-SP2)] =$$
$$\rho*PR1 + (1-\rho)*[\pi*SE1*SE2 + (1-\pi)*(1-SP1)*(1-SP2)] =$$
$$PR1 * \{\rho + (1-\rho) * (\pi*SE1*SE2 + (1-\pi)*(1-SP1)*(1-SP2)) / (\pi*SE1 + (1-\pi)*(1-SP1))\} =$$
$$PR1* \{\rho + (1-\rho)*P(T2+|T1+) \}. \text{ (Eq. 2)}$$

In case of a small prevalence π (close to 0, inserting π=0 in Eq.2), it holds

$$RFR = PR1 * \{\rho + (1-\rho)*(1-SP2) \}. \quad \text{(Eq. 3)}$$

For small prevalence estimates, the RFR is linearly related to the failure (positive) rate of the first stage PR1. This linear relationship is determined by the loss rate ρ and the specificity of the second stage test SP2. Such a linear relationship between the RFR and failure rate can be exploited in a meta-regression. The quantity P(T2+|T1+) is called conditional stage 2 positive rate (cPR2).

Equations 2 and 3 split the RFR into a component which quantifies the influence of the first stage and a component which quantifies the influence of the procedures after stage 1 (selection to stage 2 and quality of the second test). The RFR is the product of both factors. It is of interest to quantify the post stage 1 influence. This will result from a meta-regression.

Simultaneous estimation of sensitivity and specificity in different two-stage screening settings

For the simultaneous estimation of the sensitivity and specificity we assume that sensitivities and specificities of each test method are equal across the studies under consideration. For a study-specific prevalence π_S and the two-stage procedure of the screening process (with equal sensitivities and specificities across all studies) the following holds:

Positive rate at stage 1: $PR1 = P(T1+) = \pi_S*SE1 + (1-\pi_S)*(1-SP1)$.

Conditional positive rate at stage 2:
$$cPR2 = P(T2+|T1+) =$$
$$(\pi_S*SE1*SE2 + (1-\pi_S)*(1-SP1)*(1-SP2)) / (\pi_S*SE1 + (1-\pi_S)*(1-SP1)).$$

The value of cPR2 is not influenced by the loss rate ρ.

Both positive rates can be estimated from the data. If multiple studies with different prevalences are given for each combination of tests used (TEOAE-TEOAE, TEOAE-AABR, AABR-TEOAE, AABR-AABR) it is possible to infer the corresponding sensitivities and specificities (see [18, 19]). It is also possible to estimate the specific loss rate of the study protocol (how many test positives of the first stage are examined in the second stage) and the RFR (see Eq. 2).

Study protocol
The meta-analysis was prospectively registered in PROSPERO (CRD42023403091). The review protocol can be found at and downloaded from https://www.crd.york.ac.uk/prospero/display_record.php?ID=CRD42023403091. No amendments to the protocol have been made.

Eligibility criteria
For this review, the following PICO criteria were applied:

(P) Population: (Well) babies undergoing a two-stage hearing screening: (1) Initial screening as an inpatient in the maternity clinic; (2) Second test up to a maximum of one month later; (3) No use of the test method distortion product otoacoustic emissions (DPOAE); (4) Not exclusively newborns from NICU (neonatal intensive care unit).

(I) Intervention: Two-stage hearing screening using TEOAE, AABR, or a combination of both

(C) Comparator: not applicable

(O) Outcome: RFR after two screening steps, sensitivity and specificity of TEOAE and AABR

The population should preferably consist of well babies. Studies that included newborns with risk factors for hearing impairment or babies from the NICU were considered if the study data did not clearly distinguish between well babies and these newborns. However, studies that only included newborns with risk factors or from the NICU were excluded in this review.

The review considered studies in which both the first and the second tests were completed within one month after birth.

Search method

We searched MEDLINE using PubMed for relevant articles without language or geographic restrictions from the time of its inception until January, 2023. The following search strategy was used and carried out on January 4th, 2023:

("newborn hearing screening") OR ("neonatal hearing screening") OR ("infant hearing screening") OR ((("newborn screening") OR ("neonatal screening") OR ("infant screening")) AND ("hearing")) OR (("hearing screening") AND ("newborn" OR "neonatal" OR "infant"))

A broader search strategy without specifying the test methods (TEOAE or AABR) was chosen to avoid missing relevant publications that may not specify the type of test in the title or abstract. This decision was based on a small pilot search conducted to test the search strategy, which showed that including the type of screening test in the search resulted in missing a few articles already known to be relevant to the review

All articles had an abstract in English. If the full texts of the selected articles were not written in a language that the authors speak, online translation tools (DeepL, Google Translator) were used to translate them into English, where possible. Only peer-reviewed publications were considered.

Study records

The search strategy was saved in Citavi version 6.11. The data from the selected publications were extracted into Excel spreadsheets.

One reviewer searched the information sources and screened the titles and abstracts of the identified studies for inclusion and classified each study as eligible or ineligible. The study was classified as potentially eligible if it could not be clearly excluded based on its title and abstract. The full texts of all (potentially) eligible studies were then retrieved and reviewed by two additional reviewers. Again, the studies were marked as eligible or ineligible for inclusion, and the selection was discussed with the first reviewer until all three reviewers agreed. The first reviewer extracted the required data and made the preliminary decision on study inclusion based on the availability of the data for extraction. The second reviewer double-checked the extracted data and made corrections, which were discussed with all three reviewers and led to the final inclusion of all studies with available data.

Data extraction and items

Data extracted for study description included the following: First Author and year of publication, screening test combination (i.e., TEOAE-TEOAE, TEOAE-AABR, AABR-TEOAE, AABR-AABR), years of screening, country, number of newborns screened, name of screening device(s), time of first test and second test, and whether the study was performed only in well babies or also in newborns with risk factors/ from the NICU.

Data extracted for quantitative analysis included the following: Number of newborns screened with the first test and type of first test (TEOAE/AABR), number of newborns who passed the first test, number of newborns who did not pass the first test, number of

newborns who did not pass the first test and did not return for the second test (lost-to-follow up), number of newborns screened for the second time and type of second test (TEOAE/AABR), number of newborns who passed the second test, number of newborns who did not pass the second test, and number of newborns who were referred after not passing both screening tests.

Studies that provided data on simultaneous TEOAE and AABR testing were added to both screening test combinations: TEOAE followed by AABR and AABR followed by TEOAE. Otherwise, such studies would have had to be excluded as the order of the tests was not sequential.

The following data were derived from the variables collected: The failure rate after the first test step was calculated as the quotient of the number of newborns who did not pass the first test and all newborns screened in the first step. Similarly, the failure rate after the second test step was calculated as the quotient of the number of newborns who did not pass the second test and all newborns screened in the second step. The RFR was calculated as the sum of the number of newborns who did not pass the second test and the number of newborns who did not pass the first test and did not return for the second test, divided by the total number of newborns screened.

Outcomes and prioritization
Data were sought to calculate the RFR after the two-step NHS, with the failure rate after the first test included in the analysis. For the analysis, it was essential to retrieve the number of infants who passed/did not pass each screening step. Therefore, studies that only reported the results after both screening steps could not be included in the review. This review focused mainly on well babies, but studies conducted in both well babies and newborns with risk factors or from the NICU were also included.

Risk of bias in individual studies and publication bias
The quality of the included studies was assessed using a modified version of the QUADAS-II tool, which was developed for diagnostic accuracy studies [20]. Of the original four domains, the "reference test" domain was not applicable to our study because we focused on the two-stage screening process without knowledge of the true hearing loss status. Therefore, we replaced the domain "index test" by "first test" and "reference test" by "second test". The risk of bias for each study included was assessed independently by two reviewers. Disagreements were resolved by discussion with a mediator. Risk of bias was assessed at the study level.

Publication bias was not expected to have a significant impact on the literature found. It was assumed that all studies on two-stage NHS would be worthy of publication, as they describe not only the quality of the NHS, but also its implementation and problems. Selective reporting within studies, e. g. favoring one of the two screening tests, is also unlikely to be a problem. Therefore, no methods were used to assess the risk of publication bias in this meta-analysis.

Confidence in cumulative evidence
The strength of the overall body of evidence was assessed using Grading of Recommendations, Assessment, Development and Evaluation [21].

Subgroup analyses

No subgroup analyses were specified in the study protocol. However, based on the characteristics of the included studies we decided to analyze the studies that exclusively comprised of well babies as a subgroup. An additional post hoc sensitivity analysis was carried out to assess the effect of an outlier in the TEOAE-TEOAE group.

Statistical analysis

Continuous variables were summarized by median (minimum - maximum) or shown as boxplots, and categorical variables were presented by frequency (%). Due to the explorative nature of our study, no adjustment for multiple testing was considered. Statistical significance was claimed at 5% level ($p < 0.05$) or for non-overlapping 95% credibility intervals. A random effect meta-analysis of the RFR was performed using the R-package *rmeta* and the DerSimonian-Laird approach. Heterogeneity indices Q and I^2 were calculated using the random effects estimate and random effects weights. For the visualization of the risk of bias we used the source code of the *rob_summary* function of the *robvis* package to produce a similar graph adapted to our needs. Calculations were performed using WinBUGS Version 1.4.3 [22] and R Version 3.6.3 [23]. All data and analysis scripts are available at the Open Science Framework repository at https://osf.io/nuk4p/.

Results

The PRISMA flow diagram for the search and study selection process is shown in Fig 2. Out of the 3202 records identified from PubMed, 12 duplicates were removed prior to screening. A total of 2951 records were excluded. A full-text search was conducted on the remaining 239 records. Four records could not be retrieved. Out of the 235 reports that were screened, 184 were excluded. In many cases this was due to a failure to provide the required data at each screening step (n=52), not performing the first or second test within the specified time frame of one month (n=41), or an unsuitable study design (n=38). Additional reasons for exclusion are listed in Fig 2. N=5 reports comprised more than one screening protocol (n=4 two protocols [24-27] and n=1 three protocols [28]). Of the reports with more than one protocol, three [26-28] provided data on simultaneous TEOAE and AABR testing and were included in both the TEOAE-AABR and AABR-TEOAE groups. A total of 51 reports with 57 study protocols were included in the meta-analysis.

Fig 2: PRISMA flow diagram of the search and study selection process. *Includes also studies where the time point of the first or second test was not specified. **Language barrier refers to reports that could not be automatically translated into English online. A detailed list of the 184 excluded reports is available from the authors upon request.

The analysis included 57 study protocols from 51 reports (5 of those with more than one study protocol) with 704381 newborns. Table 1 provides an overview of the study characteristics. Of the 57 study protocols, n = 35 (61.4%) studies examined the TEOAE-TEOAE test combination, n = 6 (10.5%) examined the TEOAE-AABR test combination n = 3 (5.3%) examined the AABR-TEOAE test combination, and n = 13 (22.8%) examined the

AABR-AABR test combination. The median study size across all study protocols was n = 3724 newborns (min - max: 64 - 245219). The median study size for TEOAE-TEOAE was n = 8671 newborns (min - max: 81 - 245219), for TEOAE-AABR it was n = 1578 newborns (min - max: 64 - 4718), for AABR-TEOAE it was n = 247 newborns (min - max: 64 - 2005), and for AABR-AABR it was n = 3142 newborns (min - max: 81 - 31984).

Table 1: Characteristics of the included study protocols

Autor	Year(s)	Country	No of screened newborns	Device 1st and 2nd test	Time of first test	Time of second test	High-risk/ NICU included*
TEOAE - TEOAE:							
Aidan et al. 1999 [29]	1995-1997	France	1421	ILO92	48 h	within the first month	no
Alanazi 2020 [30]	2020[J,D]	Saudi Arabia	20171	GSI AudioScreener	before hospital discharge	within two weeks	yes
Arslan et al. 2013 [31]	2007-2008	Turkey	2229	Accuscreen	within 7 d, before hospital discharge	within 15 days after 1st screening	yes
Benito-Orejas et al. 2008 [24]	2001-2003	Spain	2454	ILO92	within 48h, NICU: infants tested on discharge day	within the first month	yes
Bevilacqua et al. 2010 [32]	2010[J,I]	Brazil	11466	Audix	24 h	within 20 days after birth	yes
Calevo et al. 2007 (a) [33]	2001	Italy	3238	Echo Screen	48-72 h	within the third week of life	yes
Calevo et al. 2007 [34]	2002-2004	Italy	32502	Echo Screen	48-72 h	within the third week of life	no
Cavalcanti et al. 2012 [35]	2007-2009	Brazil	3724	AccuScreen	36-48 h	one week after first test	no[F]
Chapchap et al. 2001 [36]	1996-1999	Brazil	4196	ILO88	48-72 h (NICU: prior discharge)	within 30 days (no specification)	yes
Clarke et al. 2003 [25]	2001-2002	United Kingdom	81	Echocheck (ILO88)	21 h[A]	before hospital discharge	no
De Capua et al. 2007 [37]	1998-2006	Italy	19700	ILO 292	96 h[A], NICU: at post-menstrual age of 37-41 weeks	10-20 days after birth	yes
Escobar-Ipuz et al. 2019 [38]	2007-2017	Spain	9350	unknown	<48 h	before hospital discharge	yes
Farahani et al. 2017 [39]	2013	Iran	2784	AccuScreen	day 1 or day 2	1-2 weeks after first test	no
Ferlito et al. 2021 [40]	2018	Italy	37562	unknown	48-72 h	within the first month	yes
Gilbey et al. 2013 [41]	2010-2011	Israel	4958	unknown	before hospital discharge	day after 1st screening	no
Guastini et al. 2010 [42]	2006-2009	Italy	8671	ILO92	48-72 h, NICU: when stable general condition	two weeks after first test	yes
Habib et al. 2005 [43]	1996-2004	Saudi Arabia	11986	ILO88	<48 h	5th day	no
Hatzopoulus et al. 2007 [44]	2003-2004	Albania	450	EchoLab Plus	24-48 h	within 4 weeks of first test	no
Jakubikova et al. 2003 [45]	2003[J]	Sloval Republic	3048	ILO88	4th-12th day at hospital discharge	1 month after first test	yes
Korres et al. 2006 (1) [46]	2006[J,G]	Greece	22195	ILO88	48-72h	after 1 month	yes
Korres et al. 2006 (2) [47]	2006[J,H]	Greece	25032	ILO88	48-72h	after 1 month	yes
Kosmidou et al. 2021 [48]	2018 - 2020	Greece	1491	Accuscreen	first days of life	within the first month	no

Autor	Year(s)	Country	No of screened newborns	Device 1st and 2nd test	Time of first test	Time of second test	High-risk/ NICU included*
Lin et al. 2004 [49]	2000-2002	Taiwan	5938	EchoScreen	>24, before discharge	1 month later	no
Magnani et al. 2015 [50]	2010-2013	Italy	10359	AccuScreen	24-48 h	within 3weeks from birth	no
Molini et al. 2016 [51]	2010-2012	Italy	18796	unknown	24-36 h	1 month of age	no
Pastorino et al. 2005 [52]	1997-2001	Italy	19290	ILO88, Echoscreen	36-48 h (vaginal delivery), 3-5 d (C-section)	15 - 30 d after discharge	no
Prpic et al. 2007 [53]	2002-2006	Croatia	11746	AuDX	2. or 3. day (NICU: when stable)	3 weeks after first test	yes
Satish et al. 2019 [54]	2015-2017	India	26487	unknown	<48 h	one week after first test	yes
Sequi Canet et al. 2016 [55]	2002-2013	Spain	14015	Echocheck ILO88	as late as possible prior discharge	age of 1 month	yes
Sheng et al. 2021 [28]	2018-2019	China	1340	AccuScreen	<48 h	<48 h	no
Tasci et al. 2010 [56]	2007-2008	Turkey	15323	AccuScreen	24-48 h	within 10 days	no
Tatli et al. 2007 [57]	2002-2003	Turkey	711	Echocheck	last working day prior hospital discharge	one week after first test	yes
Unlu et al. 2015 [58]	2009-2013	Turkey	2933	Echoport ILO292	day 5	day 15	no
Vos et al. 2014 [59]	2007-2012	Belgium	245219	Accuscreen, Echo-Screen, Echocheck	48-72 h	day 3 or 4 (following day after first test)	no
Welzl-Müller et al. 1997 [60]	1997[J]	Austria	2338	Echoport	within the first days	within 1-2 days after first test	no
TEOAE - AABR:							
Dort et al. 2000 [26]	2000[J]	Canada	64	ILO88, Smart Screener	before hospital discharge	before hospital discharge	yes
Lin et al. 2007 [61]	2004-2005	Taiwan	3540	ILO292, ALGO III	>48 h	before hospital discharge	yes
Olusanya et al. 2008 [62]	2005-2006	Nigeria	1150	Echo-Screen, ALGO Portable	24-48 h	before hospital discharge	no
Olusanya et al. 2010 [63]	2005-2007	Nigeria	4718	Echo-Screen, ALGO Portable	24 h, SCBU[B]: shortly before discharge	before hospital discharge	yes
Ong et al. 2020 [27]	2018	Philippines	247	Ero-Scan, AccuScreen	before hospital discharge	before hospital discharge	yes
Sheng et al. 2021 [28]	2018-2019	China	2005	AccuScreen, AccuScreen	< 72 h	< 72 h	no
AABR-TEOAE:							
Dort et al. 2000 [26]	2000[J]	Canada	64	ILO 88, Smart Screener	before hospital discharge	before hospital discharge	yes
Ong et al. 2020 [27]	2018	Philippines	247	Ero-Scan, AccuScreen	before hospital discharge	before hospital discharge	yes
Sheng et al. 2021 [28]	2018-2019	China	2005	AccuScreen, AccuScreen	< 72 h	< 72 h	no
AABR - AABR:							
Ayas et al. 2021 [64]	2017-2020	United Arab Emirates	1821	Algo 3i	24-48 h or shortly before discharge	two weeks after first test	no
Benito-Orejas et al. 2008 [24]	2004-2006	Spain	3117	AccuScreen Pro	within 48h, NICU: tested on discharge day	within 1 month after birth	yes
Clarke et al. 2003 [25]	2001-2002	United Kingdom	81	ALGO-3	24 h[A]	before hospital discharge	no
Clemens et al. 2001 [65]	1999-2000	United States	3142	Algo 2/Algo 2e	before hospital discharge	within 12-24 h after first test,	no

Autor	Year(s)	Country	No of screened newborns	Device 1st and 2nd test	Time of first test	Time of second test	High-risk/ NICU included*
						prior discharge	
Fan et al. 2010 [66]	2005-2008	Taiwan	7139	ALGO 3i	before hospital discharge	at 1 month of age	no
Gupta et al. 2015 [67]	2011-2012	India	2265	Beraphone MB11	24-48 h, preterm babies not before 34 postmenstrual weeks	within 7 days after 1st test	yes
Huang et al. 2013 [68]	2009-2010	Taiwan	15790	unknown	24-36 h	36 - 60 h of age	yes
Iwasaki et al. 2003 [69]	2000-2001	Japan	4085	ALGO 2e	48-72 h	5 - 6 days after birth, prior discharge	yes
Kelly et al. 2021 [70]	2014-2016	United States	31984	ALGO	6h (vaginal delivery), 12 h (C-section)	before hospital discharge	no
Messner et al. 2001 [71]	1998-1999	United States	5771	ALGO-II	< 24 h	before hospital discharge	no
Shim et al. 2021 [72]	2005-2015	South Korea	3059	Abaer system	24 h	within 1 month	yes
Tanyeri Toker et al. 2023 [73]	2020-2021	Turkey	570	unknown	<72 h	7-15 days of age	no
Tsuchiya et al. 2006 [74]	1999-2004	Japan	8979	ALGO 2e	day 4	1 month of age	no

*Studies that include only healthy newborns or exclude newborns with risk factors were scored as "no", NICU: newborn intensive care unit, [A]median age at first test, [B]SCBU: Special care baby unit, [C]second half of the Wessex trial, [D]Data from three years were analyzed, [E]mean age at first test, [F]originally only well babies, newborns transferred to special care baby unit were counted as losses to follow-up, [G]Three years after program initiation, [H]Additional data from two additional years were presented, [I]during 3 years, [J]Unknown dates, the publication year of the article is given.

Fig 3A presents the loss rates of newborns who did not pass the first test and did not attend the second test for all test combinations. Depending on the study, loss rates of up to 72% were reported. With the exception of the TEOAE-TEOAE group, which had a median loss-rate of 11%, the median loss rates were zero for all groups. Fig 3B shows the RFR of all study protocols. Median RFRs of 2.3%, 6.0%, 6.5% and 1.2% were found for the TEOAE-TEOAE, TEOAE-AABR, AABR-TEOAE, and AABR-AABR combinations, respectively. All studies, except for one with a RFR of 33.3% ([25]), reported RFR below 10%. Out of the 57 study protocols, 37 (64.9%) showed a RFR below 4%.

Fig 3: Loss rates (ρ) (A) and refer rates (RFR) (B) for the four different screening test combinations. The dashed horizontal line indicates the threshold of 4% for RFR. AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission.

Random effect meta-analysis of RFR

Forest plots of the random effects meta-analysis of the RFR are presented in Fig 4 for the TEOAE-TEOAE test combination and in Fig 5 for the other test combinations. The summary of the results of the meta-analysis is displayed in Table 2. Forest plots for the meta-analysis of study protocols in well babies can be found in the Supporting Information, S1 Fig and S2 Fig.

TEOAE-TEOAE

Study	Failures	Total screened		RFR	95% CI
Aidan et al. 1999	124	1421		0.087	[0.073; 0.103]
Alanazi 2020	504	20171		0.025	[0.023; 0.027]
Arslan et al. 2013	55	2229		0.025	[0.019; 0.032]
Benito-Orejas et al. 2008	64	2454		0.026	[0.020; 0.033]
Bevilacqua et al. 2010	885	11466		0.077	[0.072; 0.082]
Calevo et al. 2007 (a)	58	3238		0.018	[0.014; 0.023]
Calevo et al. 2007	462	31796		0.015	[0.013; 0.016]
Calvacanti et al. 2012	290	3724		0.078	[0.069; 0.087]
Chapchap et al. 2001	38	4196		0.009	[0.006; 0.012]
Clarke et al. 2003	27	81		0.333	[0.232; 0.447]
De Capua et al. 2007	447	19700		0.023	[0.021; 0.025]
Escobar-Ipuz et al. 2019	204	9390		0.022	[0.019; 0.025]
Farahani et al. 2017	8	2784		0.003	[0.001; 0.006]
Ferlito et al. 2021	361	37562		0.010	[0.009; 0.011]
Gilbey et al. 2013	261	4958		0.053	[0.047; 0.059]
Guastini et al. 2010	122	8671		0.014	[0.012; 0.017]
Habib et al. 2005	300	11986		0.025	[0.022; 0.028]
Hazopoulus et al. 2007	22	450		0.049	[0.031; 0.073]
Jakubikova et al. 2003	59	3048		0.019	[0.015; 0.025]
Korres et al. 2006 (1)	535	22195		0.024	[0.022; 0.026]
Korres et al. 2006 (2)	534	25032		0.021	[0.020; 0.023]
Kosmidou et al. 2021	13	1491		0.009	[0.005; 0.015]
Lin et al. 2004	231	5938		0.039	[0.034; 0.044]
Magnani et al. 2015	91	10359		0.009	[0.007; 0.011]
Molini et al. 2016	1006	18796		0.054	[0.050; 0.057]
Pastorino et al. 2005	175	19290		0.009	[0.008; 0.011]
Prpic et al. 2007	179	11746		0.015	[0.013; 0.018]
Satish et al. 2019	408	26487		0.015	[0.014; 0.017]
Sequi Canet et al. 2016	182	14015		0.013	[0.011; 0.015]
Sheng et al. 2021	39	1340		0.029	[0.021; 0.040]
Tasci et al. 2010	322	15323		0.021	[0.019; 0.023]
Tatli et al. 2007	11	711		0.015	[0.008; 0.028]
Unlu et al. 2015	126	2933		0.043	[0.036; 0.051]
Vos et al. 2014	12333	245219		0.050	[0.049; 0.051]
Welzl-Müller et al. 1997	120	2338		0.051	[0.043; 0.061]
Random effects model	**20596**	**602538**		**0.025**	**[0.020; 0.031]**

Heterogeneity: $I^2 = 43.7\%$, $\tau^2 = 0.437$

Fig 4: Random effects meta-analysis of the refer rate (RFR) for the 35 study protocols on TEOAE-TEOAE. The table displays the number of newborns who did not pass the first and second test ("Failures"), the total number of screened newborns ("Total screened"), and the RFR with the 95% confidence interval (95% CI) for each study. The summary estimate, including the 95% CI is shown as a gray diamond. The vertical solid line indicates the threshold of 4% for the RFR. TEOAE = transitory evoked otoacoustic emission.

Fig 5: Random effects meta-analysis of the refer rate (RFR) for study protocols on TEOAE-AABR, AABR-TEOAE and AABR-AABR. The table displays the number of newborns who did not pass the first and second test ("Failures"), the total number of screened newborns ("Total screened"), and the RFR with the 95% confidence interval (95% CI) for each study. The summary estimate per test combination, including the 95% CI, is shown as a gray diamond. The vertical solid line indicates the threshold of 4% for the RFR. AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission.

Table 2: Results of the random effects meta-analysis of the refer rate.

Test combination	RFR (95% CI)	I^2	tau^2	Q	df	p
TEOAE-TEOAE	2.5% (2.0, 3.1%)	43.7%	0.437	60.37	34	**0.004**
TEOAE-AABR	4.9% (3.1, 7.5%)	16.2%	0.285	5.97	5	0.310
AABR-TEOAE	5.9% (5.0, 6.9%)	0%	0	1.76	2	0.416
AABR-AABR	1.4% (0.9, 2.2%)	30.7%	0.678	17.31	12	0.138

RFR = refer rate, CI = confidence interval, AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission. The p-value shown in the right column relates to the Cochrane's Q statistic (Q) for heterogeneity. Statistically significant p values ($p < 0.05$) are printed in bold.

Strategies that do not involve changing the screening test method exhibit the lowest RFR. The upper limits of their 95% confidence intervals are below the recommended quality threshold of 4%. If the screening test method is changed from stage 1 to stage 2, the corresponding 95% confidence intervals cover the 4% threshold, indicating that it has been exceeded.

Studies of both screening combinations TEOAE-TEOAE and AABR-AABR show a moderate degree of heterogeneity, as quantified by I^2 (43.7% and 30.7%, respectively). Removing the TEOAE-TEOAE study protocol with a remarkably high RFR of 33.3% [25], as shown in S3 Fig, reduces the summary estimate for the RFR from 2.5% (2.0, 3.1%) to 2.3% (1.9, 2.9%). Results on RFR for well babies (S1 Fig and S2 Fig) are comparable to those of all studies: AABR-AABR and TEOAE-TEOAE show the lowest RFR (below 4%). Changing the test method results in higher RFR.

The results of the meta-regression of the failure rate of the first test and the RFR are presented in S1 Table and S4 Fig. The meta-regression regresses the RFR on the positive rate of the first stage (Eq.2) and allows inference on the conditional positivity rate in stage 2.

Meta-analysis of the pooled studies on the two-stage screening process to infer test-specific sensitivities and specificities

Estimates for the sensitivity and specificity of the two screening methods were calculated applying our final model for the two-stage screening process and using the methodology of Joseph et al. [18], (Table 3). To this end, it is assumed that the sensitivities and specificities for each diagnostic test are the same in all studies and the prevalence is 1.0 per 1000 children according to prior knowledge.

Table 3: Results for sensitivity and specificity across all 57 study protocols (middle column) and for the 29 study protocols that include exclusively well babies (right column).

Parameter	All study protocols (n=57) Estimate (95% CrI)	Well-baby study protocols (n=29) Estimate (95% CrI)
Sensitivity AABR	98.9% (96.0, 100.0%)	99.2% (97.1, 100.0%)
Sensitivity TEOAE	96.9% (92.5, 99.8%)	99.7% (98.9, 100.0%)
Specificity AABR	92.7% (92.5, 92.9%)	90.9% (90.6, 91.2%)
Specificity TEOAE	91.3% (91.2, 91.4%)	90.0% (89.9, 90.1%)

CrI = credibility interval, AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission.

Sensitivities of 98.9% (CrI: 96.0%, 100.0%) and 96.9% (CrI: 92.5%, 99.8%) were found for AABR and TEOAE, respectively. The corresponding specificities were 92.7% (CrI: 92.5%, 92.9%) and 91.3% (CrI: 91.2%, 91.4%). As the credibility intervals overlap, both methods do not significantly differ in their sensitivity. In contrast to this, the specificity of AABR is statistically significantly higher than the specificity of TEOAE. Table 3 also shows the results of the subgroup analysis of 29 study protocols that exclusively involved well babies (right column). Slightly higher sensitivities and slightly lower specificities were found compared to all studies (middle column).

Bias assessment
Fig 6 shows the results of the bias assessment. While there was a low risk of bias in the use of the tests (first test: 96.5%, second test: 100%), almost half of the studies had a high risk of bias in patient selection, and in flow and timing (43.9% and 40.4%). In terms of applicability concerns, more than half of the studies had high concerns for patient selection (63.2%). Similar to the risk of bias, both tests showed only low applicability concerns (first test: 96.5%, second test: 100%). The study-level bias assessment for all domains can be found in S2 Table.

Fig 6: Bias assessment using QUADAS-II tool. The figure shows the percentage of the included 57 study protocols with low, high, or unclear risk of bias (A) and concerns regarding applicability (B).

Discussion

A meta-analysis was conducted to identify the optimal screening algorithm for a two-stage NHS using combinations of TEOAE or AABR tests. The study analyzed 57 study protocols with over 704000 newborns fulfilling the inclusion criteria. The results showed that the refer rate (RFR) was lower when there was no change in the screening method used. The aggregated RFR was 1.4% for the AABR-AABR test combination and 2.5% for TEOAE-TEOAE. The studies showed high sensitivity and specificity for both the AABR and TEOAE tests.

The RFR is an important quality parameter in the NHS, as screenings with a refer result must be followed up by a pediatric audiologist. These specialists are scarce in Germany (as well as in other countries), which can lead to long waiting times for families with children who have not passed the screening tests. In addition, false positive results can cause

unnecessary anxiety for parents [75]. A low RFR in the NHS can be achieved primarily through a multi-step screening algorithm. Therefore, when the NHS was introduced in 2009, the German Pediatrics Directive required an AABR control if the first hearing test was not passed and a RFR of less than 4% at discharge, in line with national and international quality targets at the time [14, 76]. The UK even requested lower RFRs (acceptable: 3%, achievable: 2%) [17].

The German NHS evaluation data for the years 2011/2012 and the follow-up evaluation data for 2017/2018 both indicate that a second hearing test following an initial screening test with a "fail" result considerably reduces the RFR, as this second test was passed in over 80% [8, 15]. However, it is worth noting, that this second measurement was performed with a TEOAE in more than half of the tests, which is contrary to the German Pediatrics Directive. In follow-up evaluation interviews this was explained by the longer measurement duration and the increased susceptibility to interference of the AABR. The follow-up evaluation analysis revealed a considerably higher rate of refer results in the second test when a different test method was used than in the first test [8, 15]. This is in line with the finding of this meta-analysis that the highest RFR were observed when the test method was changed (TEOAE-AABR or AABR-TEOAE).

In the German follow-up evaluation, the TEOAE-TEOAE algorithm had a lower RFR (9.62%) than AABR-AABR (13.98%) [15]. In contrast, in this meta-analysis the AABR-AABR test combination had the lowest RFR, even lower than that of the TEOAE-TEOAE algorithm, while also having the highest sensitivity and specificity. This was also observed in the subgroup analysis on "well babies" and may be attributed to the higher lost-to-follow-up rate in the studies reporting TEOAE-TEOAE results (see Fig 3). In addition, studies utilizing this test combination exhibited a significantly higher heterogeneity (see Table 2) and were often based on routine clinical data, whereas the data for the AABR-AABR test sequence mainly derived from clinical studies. These factors may have affected the estimation of sensitivity and specificity. Nevertheless, the TEOAE-TEOAE test combination also had a low RFR, while maintaining good specificity and sensitivity.

One limitation of our study is the assumption of homogeneity of sensitivities and specificities across all studies, which ignores the heterogeneity caused by differently qualified personnel and in different settings (i.e., quiet vs. noisy). As these factors influencing the screening result are only described in detail in a few reports, they could not be considered in the meta-analysis. Similarly, the reports often lack information on whether data from children with risk factors for hearing impairment were included. However, they usually provide information on whether children from the NICU were included. In the subgroup analysis including exclusively study protocols with newborns without risk factors ("well babies"), there were only minor differences in the sensitivity and specificity of the different methods. As in the meta-analysis of all studies, the AABR-AABR and TEOAE-TEOAE algorithms showed the lowest RFR (below 4%) also in the subgroup of well babies, as well as higher RFR after changing the test method. A further limitation of the study is the fact, that only MEDLINE was searched for literature. However, according to van Enst et al. 2014 [77], the search in MEDLINE is sufficient for systematic reviews on diagnostic test accuracy.

To our knowledge, this is the first systematic review on quality measures of the tests used in a two-stage NHS. The strengths of the study include the large number of newborns included and the relatively high number of study reports. While we analyzed the refer rate using a

standard frequentist random effects meta-analysis, our proposed method of estimating sensitivity and specificity using Bayesian methods may provide an interesting way for analysis in future studies.

In its most recent 2019 position paper, the Joint Committee on Infant Hearing recommends performing at least two screening attempts with the same method or an AABR after TEOAE before discharging a well baby. TEOAE testing after an initial AABR with a refer result is also acceptable for well babies, as the lost-to-follow-up rate for outpatient follow-up is very high [6, 78]. The UK screening program guidelines recommend performing a further TEOAE test with an interval of at least 5 hours if the first TEOAE test is not passed for well babies [17].

The results of the meta-analysis and the data analysis of the German follow-up evaluation should provide evidence for adjusting the German Pediatrics Directive regarding the method of the second test to improve the RFR and align with international recommendations. Staff compliance with performing a second test before discharge is expected to improve if TEOAE tests are allowed, as TEOAE tests are faster and easier to perform than an AABR measurement. In contrast, changing the method of the second hearing test after the initial test was not passed leads to higher RFR and does not appear to have any benefits. However, for children with risk factors for perinatal hearing impairment, both hearing tests should always be performed with an AABR test, as specified in the German Pediatrics Directive [14] and international guidelines.

Acknowledgments

We would like to thank Anja Friedrichs for the full-text retrieval of non-open access studies.

References

1.   Korver AM, Konings S, Dekker FW, Beers M, Wever CC, Frijns JH, et al. Newborn hearing screening vs later hearing screening and developmental outcomes in children with permanent childhood hearing impairment. JAMA. 2010;304(15):1701-8. Epub 2010/10/21. doi: 10.1001/jama.2010.1501. PubMed PMID: 20959580.

2.   McCann DC, Worsfold S, Law CM, Mullee M, Petrou S, Stevenson J, et al. Reading and communication skills after universal newborn screening for permanent childhood hearing impairment. Arch Dis Child. 2009;94(4):293-7. Epub 2008/11/19. doi: 10.1136/adc.2008.151217. PubMed PMID: 19015215.

3.   Pimperton H, Blythe H, Kreppner J, Mahon M, Peacock JL, Stevenson J, et al. The impact of universal newborn hearing screening on long-term literacy outcomes: a prospective cohort study. Arch Dis Child. 2016;101(1):9-15. Epub 2014/11/27. doi: 10.1136/archdischild-2014-307516. PubMed PMID: 25425604; PubMed Central PMCID: PMCPMC4717422.

4.   Wolff R, Hommerich J, Riemsma R, Antes G, Lange S, Kleijnen J. Hearing screening in newborns: systematic review of accuracy, effectiveness, and effects of interventions after screening. Arch Dis Child. 2010;95(2):130-5. Epub 2009/03/31. doi: 10.1136/adc.2008.151092. PubMed PMID: 19329444.

5.   Yoshinaga-Itano C, Manchaiah V, Hunnicutt C. Outcomes of Universal Newborn Screening Programs: Systematic Review. J Clin Med. 2021;10(13). Epub 2021/07/03. doi: 10.3390/jcm10132784. PubMed PMID: 34202909; PubMed Central PMCID: PMCPMC8268039.

6.   Joint committee on infant hearing. Year 2019 Position Statement: Principles and Guidelines for Early Hearing Detection and Intervention Programs. Journal of Early Hearing Detection and Intervention. 2019;4(2):1-44. doi: 10.15142/fptk-b748.

7.   Grandori F. European Consensus Statement on Neonatal Hearing Screening. J Laryngol Otol. 1998;112(12):1219. Epub 1999/04/21. doi: 10.1017/s002221510014294x. PubMed PMID: 10209630.

8.   Nennstiel U, Brockow I, Söhl K, Zirngibl A, Am Zehnhoff-Dinnesen A, Matulat P, et al. [Endbericht zur Evaluation des Neugeborenen-Hörscreenings 2011/2012.] 2017 [December 28, 2023]. Available from: https://www.g-ba.de/downloads/40-268-4395/2017-05-18_Kinder-RL_Annahme_Endbericht_NHS-Bericht.pdf. .

9.   Grosse SD, Mason CA, Gaffney M, Thomson V, White KR. What Contribution Did Economic Evidence Make to the Adoption of Universal Newborn Hearing Screening Policies in the United States? Int J Neonatal Screen. 2018;4(3):25. Epub 2018/08/21. doi: 10.3390/ijns4030025. PubMed PMID: 30123850; PubMed Central PMCID: PMCPMC6094389.

10.   Rohlfs AK, Wiesner T, Drews H, Muller F, Breitfuss A, Schiller R, et al. Interdisciplinary approach to design, performance, and quality management in a multicenter newborn hearing screening project. Discussion of the results of newborn hearing screening in Hamburg (part II). Eur J Pediatr. 2010;169(12):1453-63. Epub 2010/06/15. doi: 10.1007/s00431-010-1229-0. PubMed PMID: 20544359.

11.   Sharma R, Gu Y, Ching TYC, Marnane V, Parkinson B. Economic Evaluations of Childhood Hearing Loss Screening Programmes: A Systematic Review and Critique. Appl Health Econ Health Policy. 2019;17(3):331-57. Epub 2019/01/27. doi: 10.1007/s40258-018-00456-1. PubMed PMID: 30680698; PubMed Central PMCID: PMCPMC7279710.

12.   Wilson JMG, Jungner G, World Health O. Principles and practice of screening for disease. Geneva: World Health Organization; 1968.

13.   Andermann A, Blancquaert I, Beauchamp S, Dery V. Revisiting Wilson and Jungner in the genomic age: a review of screening criteria over the past 40 years. Bull World Health Organ. 2008;86(4):317-9. Epub 2008/04/29. doi: 10.2471/blt.07.050112. PubMed PMID: 18438522; PubMed Central PMCID: PMCPMC2647421.

14.   Gemeinsamer Bundesausschuss (G-BA). [Richtlinie des Gemeinsamen Bundesausschusses über die Früherkennung von Krankheiten bei Kindern (Kinder-Richtlinie).] 2020 [January 18, 2024]. Available from: https://www.g-ba.de/downloads/62-492-3190/Kinder-RL_2023-05-12_iK-2023-07-13.pdf.

15.   Nennstiel U, Brockow I, Hanuaer M, Heißenhuber A, Marzi C, Söhl K, et al. [Endbericht zur Folge-Evaluation des Endbericht der Evaluation des Neugeborenen-Hörscreenings 2017/2018.] 2022 [December 18, 2023]. Available from: https://www.g-ba.de/downloads/40-268-9045/2022-11-17_Kinder-RL_Abnahme-Endbericht-Folge-Evaluation-NHS_Bericht.pdf.

16.   Brockow I, Sohl K, Hanauer M, Heissenhuber A, Marzi C, Am Zehnhoff-Dinnesen A, et al. [Newborn hearing screening in Germany-results of the 2011/2012 and 2017/2018 evaluations]. Bundesgesundheitsblatt Gesundheitsforschung Gesundheitsschutz. 2023;66(11):1259-67. Epub 2023/10/16. doi: 10.1007/s00103-023-03779-0. PubMed PMID: 37843595; PubMed Central PMCID: PMCPMC10622351.

17.   Government Digital Service UK. Newborn hearing screening programme (NHPS) Patient journey from screen to referral. 2021 [January 18, 2024]. Available from: https://www.gov.uk/government/publications/newborn-hearing-screening-programme-nhsp-operational-guidance/6-patient-journey-from-screen-to-referral.

18.   Joseph L, Gyorkos TW, Coupal L. Bayesian estimation of disease prevalence and the parameters of diagnostic tests in the absence of a gold standard. Am J Epidemiol. 1995;141(3):263-72. Epub 1995/02/01. doi: 10.1093/oxfordjournals.aje.a117428. PubMed PMID: 7840100.

19.   Rutjes AW, Reitsma JB, Coomarasamy A, Khan KS, Bossuyt PM. Evaluation of diagnostic tests when there is no gold standard. A review of methods. Health Technol Assess. 2007;11(50):iii, ix-51. Epub 2007/11/21. doi: 10.3310/hta11500. PubMed PMID: 18021577.

20.   Whiting PF, Rutjes AW, Westwood ME, Mallett S, Deeks JJ, Reitsma JB, et al. QUADAS-2: a revised tool for the quality assessment of diagnostic accuracy studies. Ann Intern Med. 2011;155(8):529-36. Epub 2011/10/19. doi: 10.7326/0003-4819-155-8-201110180-00009. PubMed PMID: 22007046.

21.   Guyatt GH, Oxman AD, Vist GE, Kunz R, Falck-Ytter Y, Alonso-Coello P, et al. GRADE: an emerging consensus on rating quality of evidence and strength of recommendations. BMJ. 2008;336(7650):924-6. Epub 2008/04/26. doi: 10.1136/bmj.39489.470347.AD. PubMed PMID: 18436948; PubMed Central PMCID: PMCPMC2335261.

22.   Lunn DJ, Thomas A, Best N, Spiegelhalter D. WinBUGS - A Bayesian modelling framework: Concepts, structure, and extensibility. Statistics and Computing. 2000;10(4):325-37. doi: 10.1023/A:1008929526011.

23.   R Core Team. R: A Language and Environment for Statistical Computing. 2022.

24.   Benito-Orejas JI, Ramirez B, Morais D, Almaraz A, Fernandez-Calvo JL. Comparison of two-step transient evoked otoacoustic emissions (TEOAE) and automated auditory brainstem response (AABR) for universal newborn hearing screening programs. Int J Pediatr Otorhinolaryngol. 2008;72(8):1193-201. Epub 2008/06/14. doi: 10.1016/j.ijporl.2008.04.011. PubMed PMID: 18550180.

25.   Clarke P, Iqbal M, Mitchell S. A comparison of transient-evoked otoacoustic emissions and automated auditory brainstem responses for pre-discharge neonatal hearing screening. Int J Audiol. 2003;42(8):443-7. Epub 2003/12/09. doi: 10.3109/14992020309081514. PubMed PMID: 14658852.

26.   Dort JC, Tobolski C, Brown D. Screening strategies for neonatal hearing loss: which test is best? J Otolaryngol. 2000;29(4):206-10. Epub 2000/09/26. PubMed PMID: 11003070.

27.   Ong KMC, Rivera AS, Chan AL, Chiong CM. Determining concordance and cost impact of otoacoustic emission and automated auditory brainstem response in newborn

hearing screening in a tertiary hospital. Int J Pediatr Otorhinolaryngol. 2020;128:109704. Epub 2019/10/14. doi: 10.1016/j.ijporl.2019.109704. PubMed PMID: 31606683.

28.   Sheng H, Zhou Q, Wang Q, Yu Y, Liu L, Liang M, et al. Comparison of Two-Step Transient Evoked Otoacoustic Emissions and One-Step Automated Auditory Brainstem Response for Universal Newborn Hearing Screening Programs in Remote Areas of China. Front Pediatr. 2021;9:655625. Epub 2021/06/01. doi: 10.3389/fped.2021.655625. PubMed PMID: 34055691; PubMed Central PMCID: PMCPMC8160434.

29.   Aidan D, Avan P, Bonfils P. Auditory screening in neonates by means of transient evoked otoacoustic emissions: a report of 2,842 recordings. Ann Otol Rhinol Laryngol. 1999;108(6):525-31. Epub 1999/06/23. doi: 10.1177/000348949910800601. PubMed PMID: 10378518.

30.   Alanazi AA. Referral and Lost to System Rates of Two Newborn Hearing Screening Programs in Saudi Arabia. Int J Neonatal Screen. 2020;6(3):50. Epub 2020/10/31. doi: 10.3390/ijns6030050. PubMed PMID: 33123632; PubMed Central PMCID: PMCPMC7570083.

31.   Arslan S, Isik AU, Imamoglu M, Topbas M, Aslan Y, Ural A. Universal newborn hearing screening; automated transient evoked otoacoustic emissions. B-ENT. 2013;9(2):122-31. Epub 2013/08/06. PubMed PMID: 23909119.

32.   Bevilacqua MC, Alvarenga Kde F, Costa OA, Moret AL. The universal newborn hearing screening in Brazil: from identification to intervention. Int J Pediatr Otorhinolaryngol. 2010;74(5):510-5. Epub 2010/03/23. doi: 10.1016/j.ijporl.2010.02.009. PubMed PMID: 20303604.

33.   Calevo MG, Mezzano P, Zullino E, Padovani P, Scopesi F, Serra G, et al. Neonatal hearing screening model: an Italian regional experience. J Matern Fetal Neonatal Med. 2007;20(6):441-8. Epub 2007/08/04. doi: 10.1080/14767050701398090. PubMed PMID: 17674253.

34.   Calevo MG, Mezzano P, Zullino E, Padovani P, Serra G, Group S. Ligurian experience on neonatal hearing screening: clinical and epidemiological aspects. Acta Paediatr. 2007;96(11):1592-9. Epub 2007/10/17. doi: 10.1111/j.1651-2227.2007.00475.x. PubMed PMID: 17937684.

35.   Cavalcanti HG, Guerra RO. The role of maternal socioeconomic factors in the commitment to universal newborn hearing screening in the Northeastern region of Brazil. Int J Pediatr Otorhinolaryngol. 2012;76(11):1661-7. Epub 2012/08/28. doi: 10.1016/j.ijporl.2012.07.041. PubMed PMID: 22921603.

36.   Chapchap MJ, Segre CM. Universal newborn hearing screening and transient evoked otoacoustic emission: new concepts in Brazil. Scand Audiol Suppl. 2001;(53):33-6. Epub 2001/06/21. doi: 10.1080/010503901750166600. PubMed PMID: 11409775.

37.   De Capua B, Costantini D, Martufi C, Latini G, Gentile M, De Felice C. Universal neonatal hearing screening: the Siena (Italy) experience on 19,700 newborns. Early Hum Dev. 2007;83(9):601-6. Epub 2007/02/20. doi: 10.1016/j.earlhumdev.2007.01.001. PubMed PMID: 17307313.

38.   Escobar-Ipuz FA, Soria-Bretones C, Garcia-Jimenez MA, Cueto EM, Torres Aranda AM, Sotos JM. Early detection of neonatal hearing loss by otoacoustic emissions and auditory brainstem response over 10 years of experience. Int J Pediatr Otorhinolaryngol.

2019;127:109647. Epub 2019/08/31. doi: 10.1016/j.ijporl.2019.109647. PubMed PMID: 31470205.

39.   Farahani F, Hamidi Nahrani M, Seifrabiei MA, Emadi M. The Effect of Mode of Delivery and Hospital Type on Newborn Hearing Screening Results Using Otoacoustic Emissions: Based on Screening Age. Indian J Otolaryngol Head Neck Surg. 2017;69(1):1-5. Epub 2017/02/28. doi: 10.1007/s12070-016-0967-3. PubMed PMID: 28239569; PubMed Central PMCID: PMCPMC5305649.

40.   Ferlito S, Maniaci A, Cocuzza S, La Mantia I, Di Mauro P, Poli G, et al. Universal newborn hearing screening in the Italian Region of Sicily in 2018. Acta Otorhinolaryngol Ital. 2021;41(4):356-63. Epub 2021/09/18. doi: 10.14639/0392-100X-N1162. PubMed PMID: 34533539; PubMed Central PMCID: PMCPMC8448180.

41.   Gilbey P, Kraus C, Ghanayim R, Sharabi-Nov A, Bretler S. Universal newborn hearing screening in Zefat, Israel: the first two years. Int J Pediatr Otorhinolaryngol. 2013;77(1):97-100. Epub 2012/11/06. doi: 10.1016/j.ijporl.2012.10.004. PubMed PMID: 23122541.

42.   Guastini L, Mora R, Dellepiane M, Santomauro V, Mora M, Rocca A, et al. Evaluation of an automated auditory brainstem response in a multi-stage infant hearing screening. Eur Arch Otorhinolaryngol. 2010;267(8):1199-205. Epub 2010/02/12. doi: 10.1007/s00405-010-1209-z. PubMed PMID: 20148257.

43.   Habib HS, Abdelgaffar H. Neonatal hearing screening with transient evoked otoacoustic emissions in Western Saudi Arabia. Int J Pediatr Otorhinolaryngol. 2005;69(6):839-42. Epub 2005/05/12. doi: 10.1016/j.ijporl.2005.01.018. PubMed PMID: 15885338.

44.   Hatzopoulos S, Qirjazi B, Martini A. Neonatal hearing screening in Albania: results from an ongoing universal screening program. Int J Audiol. 2007;46(4):176-82. Epub 2007/04/25. doi: 10.1080/14992020601145310. PubMed PMID: 17454230.

45.   Jakubikova J, Kabatova Z, Zavodna M. Identification of hearing loss in newborns by transient otoacoustic emissions. Int J Pediatr Otorhinolaryngol. 2003;67(1):15-8. Epub 2003/02/01. doi: 10.1016/s0165-5876(02)00285-9. PubMed PMID: 12560144.

46.   Korres SG, Balatsouras DG, Nikolopoulos T, Korres GS, Ferekidis E. Making universal newborn hearing screening a success. Int J Pediatr Otorhinolaryngol. 2006;70(2):241-6. Epub 2005/07/21. doi: 10.1016/j.ijporl.2005.06.010. PubMed PMID: 16029898.

47.   Korres SG, Balatsouras DG, Gkoritsa E, Eliopoulos P, Rallis E, Ferekidis E. Success rate of newborn and follow-up screening of hearing using otoacoustic emissions. Int J Pediatr Otorhinolaryngol. 2006;70(6):1039-43. Epub 2005/12/02. doi: 10.1016/j.ijporl.2005.10.018. PubMed PMID: 16318876.

48.   Kosmidou P, Tzifas S, Lygeros S, Danielides G, Nikolopoulos T, Dimitriou G, et al. Newborn Hearing Screening: Analysing the Effectiveness of Early Detection of Neonatal Hearing Loss in a Hospital in Greece. Cureus. 2021;13(11):e19807. Epub 2021/12/28. doi: 10.7759/cureus.19807. PubMed PMID: 34956791; PubMed Central PMCID: PMCPMC8693701.

49.   Lin CY, Huang CY, Lin CY, Lin YH, Wu JL. Community-based newborn hearing screening program in Taiwan. Int J Pediatr Otorhinolaryngol. 2004;68(2):185-9. Epub 2004/01/17. doi: 10.1016/j.ijporl.2003.10.007. PubMed PMID: 14725985.

50.   Magnani C, Bacchi G, Borghini AM, Delmonte D, Fava G, Occasio AM, et al. Universal newborn hearing screening: the experience of the University Hospital of Parma. Acta Biomed. 2015;86(3):273-7. Epub 2015/12/24. PubMed PMID: 26694155.

51.   Molini E, Calzolaro L, Lapenna R, Ricci G. Universal newborn hearing screening in Umbria region, Italy. Int J Pediatr Otorhinolaryngol. 2016;82:92-7. Epub 2016/02/10. doi: 10.1016/j.ijporl.2016.01.007. PubMed PMID: 26857323.

52.   Pastorino G, Sergi P, Mastrangelo M, Ravazzani P, Tognola G, Parazzini M, et al. The Milan Project: a newborn hearing screening programme. Acta Paediatr. 2005;94(4):458-63. Epub 2005/08/12. doi: 10.1111/j.1651-2227.2005.tb01918.x. PubMed PMID: 16092461.

53.   Prpic I, Mahulja-Stamenkovic V, Bilic I, Haller H. Hearing loss assessed by universal newborn hearing screening--the new approach. Int J Pediatr Otorhinolaryngol. 2007;71(11):1757-61. Epub 2007/09/11. doi: 10.1016/j.ijporl.2007.07.015. PubMed PMID: 17825432.

54.   Satish HS, Anil Kumar R, Viswanatha B. Screening of Newborn Hearing at a Tertiary Care Hospital in South India. Indian J Otolaryngol Head Neck Surg. 2019;71(Suppl 2):1383-90. Epub 2019/11/22. doi: 10.1007/s12070-018-1454-9. PubMed PMID: 31750182; PubMed Central PMCID: PMCPMC6841892.

55.   Sequi Canet JM, Sala Langa MJ, Collar Del Castillo JI. [Results from ten years newborn hearing screening in a secondary hospital]. An Pediatr (Barc). 2016;85(4):189-96. Epub 2016/01/05. doi: 10.1016/j.anpedi.2015.11.003. PubMed PMID: 26725594.

56.   Tasci Y, Muderris, II, Erkaya S, Altinbas S, Yucel H, Haberal A. Newborn hearing screening programme outcomes in a research hospital from Turkey. Child Care Health Dev. 2010;36(3):317-22. Epub 2009/12/18. doi: 10.1111/j.1365-2214.2009.01029.x. PubMed PMID: 20015280.

57.   Tatli MM, Bulent Serbetcioglu M, Duman N, Kumral A, Kirkim G, Ogun B, et al. Feasibility of neonatal hearing screening program with two-stage transient otoacoustic emissions in Turkey. Pediatr Int. 2007;49(2):161-6. Epub 2007/04/21. doi: 10.1111/j.1442-200X.2007.02344.x. PubMed PMID: 17445032.

58.   Unlu I, Guclu E, Yaman H. When should automatic Auditory Brainstem Response test be used for newborn hearing screening? Auris Nasus Larynx. 2015;42(3):199-202. Epub 2014/12/03. doi: 10.1016/j.anl.2014.10.005. PubMed PMID: 25459495.

59.   Vos B, Lagasse R, Leveque A. Main outcomes of a newborn hearing screening program in Belgium over six years. Int J Pediatr Otorhinolaryngol. 2014;78(9):1496-502. Epub 2014/07/12. doi: 10.1016/j.ijporl.2014.06.019. PubMed PMID: 25012194.

60.   Welzl-Muller K, Boheim K, Stephan K, Schlogel H, Stadlmann A, Nekahm D. [Optimizing hearing screening by transient evoked otoacoustic emissions in newborn infants]. HNO. 1997;45(4):227-32. Epub 1997/04/01. doi: 10.1007/s001060050109. PubMed PMID: 9221263.

61.	Lin HC, Shu MT, Lee KS, Lin HY, Lin G. reducing false positives in newborn hearing screening program: how and why. Otol Neurotol. 2007;28(6):788-92. Epub 2007/10/20. doi: 10.1097/mao.0b013e3180cab754. PubMed PMID: 17948357.

62.	Olusanya BO, Wirz SL, Luxon LM. Hospital-based universal newborn hearing screening for early detection of permanent congenital hearing loss in Lagos, Nigeria. Int J Pediatr Otorhinolaryngol. 2008;72(7):991-1001. Epub 2008/04/25. doi: 10.1016/j.ijporl.2008.03.004. PubMed PMID: 18433883.

63.	Olusanya BO, Bamigboye BA. Is discordance in TEOAE and AABR outcomes predictable in newborns? Int J Pediatr Otorhinolaryngol. 2010;74(11):1303-9. Epub 2010/09/11. doi: 10.1016/j.ijporl.2010.08.010. PubMed PMID: 20828836.

64.	Ayas M, Yaseen H. Emerging Data from a Newborn Hearing Screening Program in Sharjah, United Arab Emirates. Int J Pediatr. 2021;2021:2616890. Epub 2021/07/15. doi: 10.1155/2021/2616890. PubMed PMID: 34257673; PubMed Central PMCID: PMCPMC8257347.

65.	Clemens CJ, Davis SA. Minimizing false-positives in universal newborn hearing screening: a simple solution. Pediatrics. 2001;107(3):E29. Epub 2001/03/07. doi: 10.1542/peds.107.3.e29. PubMed PMID: 11230610.

66.	Fan JY, Chen LS, Lai JC, Chen MK, Chen HC. A pre-paid newborn hearing screening programme: a community-based study. B-ENT. 2010;6(4):265-9. Epub 2011/02/10. PubMed PMID: 21302689.

67.	Gupta S, Sah S, Som T, Saksena M, Yadav CP, Sankar MJ, et al. Challenges of Implementing Universal Newborn Hearing Screening at a Tertiary Care Centre from India. Indian J Pediatr. 2015;82(8):688-93. Epub 2015/02/06. doi: 10.1007/s12098-015-1688-4. PubMed PMID: 25652547.

68.	Huang HM, Chiang SH, Shiau YS, Yeh WY, Ho HC, Wang L, et al. The universal newborn hearing screening program of Taipei City. Int J Pediatr Otorhinolaryngol. 2013;77(10):1734-7. Epub 2013/09/10. doi: 10.1016/j.ijporl.2013.08.004. PubMed PMID: 24012220.

69.	Iwasaki S, Hayashi Y, Seki A, Nagura M, Hashimoto Y, Oshima G, et al. A model of two-stage newborn hearing screening with automated auditory brainstem response. Int J Pediatr Otorhinolaryngol. 2003;67(10):1099-104. Epub 2003/10/11. doi: 10.1016/s0165-5876(03)00199-x. PubMed PMID: 14550964.

70.	Kelly AF, Kelly PK, Shah M. Auditory Brainstem Response Pass Rates Correlate with Newborn Hour of Life and Delivery Mode. J Pediatr. 2021;230:100-5. Epub 2020/10/26. doi: 10.1016/j.jpeds.2020.10.036. PubMed PMID: 33098840.

71.	Messner AH, Price M, Kwast K, Gallagher K, Forte J. Volunteer-based universal newborn hearing screening program. Int J Pediatr Otorhinolaryngol. 2001;60(2):123-30. Epub 2001/08/24. doi: 10.1016/s0165-5876(01)00507-9. PubMed PMID: 11518589.

72.	Shim J, Kim H, Kwon Y, Chang J, Park E, Im GJ. Results of a 10-year hearing screening using automated auditory brainstem response in newborns: The two-step AABR method. Int J Pediatr Otorhinolaryngol. 2021;151:110947. Epub 2021/10/27. doi: 10.1016/j.ijporl.2021.110947. PubMed PMID: 34700296.

73.   Tanyeri Toker G, Kumbul YC, Cetinkol AE, Aslan H, Baba P, Oncel MY. Is Gestational COVID-19 a Risk Factor for Congenital Hearing Loss? Otol Neurotol. 2023;44(2):115-20. Epub 2023/01/11. doi: 10.1097/MAO.0000000000003761. PubMed PMID: 36624586; PubMed Central PMCID: PMCPMC9835235.

74.   Tsuchiya H, Goto K, Yunohara N, Matsuoka M, Nishioka M, Nakamura Y, et al. Newborn hearing screening in a single private Japanese obstetric hospital. Pediatr Int. 2006;48(6):604-7. Epub 2006/12/16. doi: 10.1111/j.1442-200X.2006.02274.x. PubMed PMID: 17168982.

75.   Hewlett J, Waisbren SE. A review of the psychosocial effects of false-positive results on parents and current communication practices in newborn screening. J Inherit Metab Dis. 2006;29(5):677-82. Epub 2006/08/19. doi: 10.1007/s10545-006-0381-1. PubMed PMID: 16917730.

76.   Gemeinsamer Bundesausschuss (G-BA). [Tragende Gründe zum Beschluss des Gemeinsamen Bundesausschusses über eine Änderung der Kinder-Richtlinien: Einführung eines Neugeborenen-Hörscreenings] 2008 [January 23, 2024]. Available from: https://www.g-ba.de/downloads/40-268-641/2008-06-19-Kinder-Hörscreening_TrG.pdf.

77.   van Enst WA, Scholten RJ, Whiting P, Zwinderman AH, Hooft L. Meta-epidemiologic analysis indicates that MEDLINE searches are sufficient for diagnostic test accuracy systematic reviews. J Clin Epidemiol. 2014;67(11):1192-9. Epub 2014/07/06. doi: 10.1016/j.jclinepi.2014.05.008. PubMed PMID: 24996667.

78.   Hunter LL, Meinzen-Derr J, Wiley S, Horvath CL, Kothari R, Wexelblatt S. Influence of the WIC Program on Loss to Follow-up for Newborn Hearing Screening. Pediatrics. 2016;138(1). Epub 2016/06/17. doi: 10.1542/peds.2015-4301. PubMed PMID: 27307144; PubMed Central PMCID: PMCPMC4925076.

Supporting Information

S1 Checklist. PRISMA checklist.

Section and Topic	Item #	Checklist item	Location where item is reported
TITLE			
Title	1	Identify the report as a systematic review.	Title
ABSTRACT			
Abstract	2	See the PRISMA 2020 for Abstracts checklist.	Considered in abstract
INTRODUCTION			
Rationale	3	Describe the rationale for the review in the context of existing knowledge.	Introduction (paragraphs 1-4)
Objectives	4	Provide an explicit statement of the objective(s) or question(s) the review addresses.	Introduction (paragraph 4)
METHODS			
Eligibility criteria	5	Specify the inclusion and exclusion criteria for the review and how studies were grouped for the syntheses.	Materials and methods → Eligibility criteria
Information sources	6	Specify all databases, registers, websites, organisations, reference lists and other sources searched or consulted to identify studies. Specify the date when each source was last searched or consulted.	Materials and methods → Search method
Search strategy	7	Present the full search strategies for all databases, registers and websites, including any filters and limits used.	Materials and methods → Search method
Selection process	8	Specify the methods used to decide whether a study met the inclusion criteria of the review, including how many reviewers screened each record and each report retrieved, whether they worked independently, and if applicable, details of automation tools used in the process.	Materials and methods → Study records
Data collection process	9	Specify the methods used to collect data from reports, including how many reviewers collected data from each report, whether they worked independently, any processes for obtaining or confirming data from study investigators, and if applicable, details of automation tools used in the process.	Materials and methods → Data extraction and items
Data items	10a	List and define all outcomes for which data were sought. Specify whether all results that were compatible with each outcome domain in each study were sought (e.g. for all measures, time points, analyses), and if not, the methods used to decide which results to collect.	Materials and methods → Data extraction and items
	10b	List and define all other variables for which data were sought (e.g. participant and intervention characteristics, funding sources). Describe any assumptions made about any missing or unclear information.	Materials and methods → Data extraction and items, legend of Table 1
Study risk of bias assessment	11	Specify the methods used to assess risk of bias in the included studies, including details of the tool(s) used, how many reviewers assessed each study and whether they worked independently, and if applicable, details of automation tools used in the process.	Materials and methods → Risk of bias in individual studies and publication bias
Effect measures	12	Specify for each outcome the effect measure(s) (e.g. risk ratio, mean difference) used in the synthesis or presentation of results.	Materials and methods → Data extraction and items

Section and Topic	Item #	Checklist item	Location where item is reported
Synthesis methods	13a	Describe the processes used to decide which studies were eligible for each synthesis (e.g. tabulating the study intervention characteristics and comparing against the planned groups for each synthesis (item #5)).	Materials and methods → Statistical analysis
	13b	Describe any methods required to prepare the data for presentation or synthesis, such as handling of missing summary statistics, or data conversions.	Materials and methods → Data extraction and items
	13c	Describe any methods used to tabulate or visually display results of individual studies and syntheses.	Materials and methods → Statistical analysis
	13d	Describe any methods used to synthesize results and provide a rationale for the choice(s). If meta-analysis was performed, describe the model(s), method(s) to identify the presence and extent of statistical heterogeneity, and software package(s) used.	Materials and methods → Statistical analysis
	13e	Describe any methods used to explore possible causes of heterogeneity among study results (e.g. subgroup analysis, meta-regression).	Materials and methods → Subgroup analysis
	13f	Describe any sensitivity analyses conducted to assess robustness of the synthesized results.	Materials and methods → Subgroup analysis
Reporting bias assessment	14	Describe any methods used to assess risk of bias due to missing results in a synthesis (arising from reporting biases).	Materials and methods → Risk of bias in individual studies and publication bias
Certainty assessment	15	Describe any methods used to assess certainty (or confidence) in the body of evidence for an outcome.	Materials and methods → Confidence in cumulative evidence
RESULTS			
Study selection	16a	Describe the results of the search and selection process, from the number of records identified in the search to the number of studies included in the review, ideally using a flow diagram.	Fig 2
	16b	Cite studies that might appear to meet the inclusion criteria, but which were excluded, and explain why they were excluded.	Legend of Fig 2
Study characteristics	17	Cite each included study and present its characteristics.	Table 1
Risk of bias in studies	18	Present assessments of risk of bias for each included study.	S2 Table
Results of individual studies	19	For all outcomes, present, for each study: (a) summary statistics for each group (where appropriate) and (b) an effect estimate and its precision (e.g. confidence/credible interval), ideally using structured tables or plots.	Fig 3, S2 Table
Results of syntheses	20a	For each synthesis, briefly summarise the characteristics and risk of bias among contributing studies.	Fig 3
	20b	Present results of all statistical syntheses conducted. If meta-analysis was done, present for each the summary estimate and its precision (e.g. confidence/credible interval) and measures of statistical heterogeneity. If comparing groups, describe the direction of the effect.	See Figs 4 and 5 and Tables 2 and 3
	20c	Present results of all investigations of possible causes of	S3 Fig

Section and Topic	Item #	Checklist item	Location where item is reported
		heterogeneity among study results.	
	20d	Present results of all sensitivity analyses conducted to assess the robustness of the synthesized results.	S3 Fig
Reporting biases	21	Present assessments of risk of bias due to missing results (arising from reporting biases) for each synthesis assessed.	NA, since no reporting bias occurred.
Certainty of evidence	22	Present assessments of certainty (or confidence) in the body of evidence for each outcome assessed.	S1 File
DISCUSSION			
Discussion	23a	Provide a general interpretation of the results in the context of other evidence.	Discussion (paragraphs 2-4)
	23b	Discuss any limitations of the evidence included in the review.	Discussion (paragraph 5)
	23c	Discuss any limitations of the review processes used.	Discussion (paragraph 5)
	23d	Discuss implications of the results for practice, policy, and future research.	Discussion (last paragraph)
OTHER INFORMATION			
Registration and protocol	24a	Provide registration information for the review, including register name and registration number, or state that the review was not registered.	Materials and methods → Study protocol
	24b	Indicate where the review protocol can be accessed, or state that a protocol was not prepared.	Materials and methods → Study protocol
	24c	Describe and explain any amendments to information provided at registration or in the protocol.	Materials and methods → Study protocol
Support	25	Describe sources of financial or non-financial support for the review, and the role of the funders or sponsors in the review.	See Additional information
Competing interests	26	Declare any competing interests of review authors.	See Additional information
Availability of data, code and other materials	27	Report which of the following are publicly available and where they can be found: template data collection forms; data extracted from included studies; data used for all analyses; analytic code; any other materials used in the review.	Materials and Methods → Statistical analysis (last sentence)

NA = not applicable.

TEOAE-TEOAE

Study	Failures	Total screened		RFR	95% CI
Aidan et al. 1999	124	1421		0.087	[0.073; 0.103]
Calevo et al. 2007	462	31796		0.015	[0.013; 0.016]
Calvacanti et al. 2012	290	3724		0.078	[0.069; 0.087]
Clarke et al. 2003	27	81		0.333	[0.232; 0.447]
Farahani et al. 2017	8	2784		0.003	[0.001; 0.006]
Gilbey et al. 2013	261	4958		0.053	[0.047; 0.059]
Habib et al. 2005	300	11986		0.025	[0.022; 0.028]
Hazopoulus et al. 2007	22	450		0.049	[0.031; 0.073]
Kosmidou et al. 2021	13	1491		0.009	[0.005; 0.015]
Lin et al. 2004	231	5938		0.039	[0.034; 0.044]
Magnani et al. 2015	91	10359		0.009	[0.007; 0.011]
Molini et al. 2016	1006	18796		0.054	[0.050; 0.057]
Pastorino et al. 2005	175	19290		0.009	[0.008; 0.011]
Sheng et al. 2021	39	1340		0.029	[0.021; 0.040]
Tasci et al. 2010	322	15323		0.021	[0.019; 0.023]
Unlu et al. 2015	126	2933		0.043	[0.036; 0.051]
Vos et al. 2014	12333	245219		0.050	[0.049; 0.051]
Welzl-Müller et al. 1997	120	2338		0.051	[0.043; 0.061]
Random effects model	**15950**	**380227**		**0.032**	**[0.025; 0.043]**

Heterogeneity: $I^2 = 69.0\%$, $\tau^2 = 0.360$

S1 Fig: Random effects meta-analysis of the refer rate (RFR) for the 18 study protocols on TEOAE-TEOAE in well babies. Shown is the number of newborns who failed both first and second test ("Failures"), number of screened newborns ("Total screened") and the RFR with the 95% confidence interval (95% CI) for each study. The summary estimate including the 95% CI is shown as gray diamond. The vertical solid line shows the threshold of 4% for the RFR. TEOAE = transitory evoked otoacoustic emission.

Study	Failures	Total screened		RFR	95% CI
TEOAE-AABR					
Olusanya et al. 2008	55	1150		0.048	[0.036; 0.062]
Sheng et al. 2021	113	2005		0.056	[0.047; 0.067]
Random effects model	**168**	**3155**		**0.053**	**[0.046; 0.062]**
Heterogeneity: $I^2 = 0\%$, $\tau^2 = <0.001$					
AABR-TEOAE					
Sheng et al. 2021	113	2005		0.056	[0.047; 0.067]
AABR-AABR					
Ayas et al. 2021	24	1821		0.013	[0.008; 0.020]
Clarke et al. 2003	3	81		0.037	[0.008; 0.104]
Clemens et al. 2001	39	3142		0.012	[0.009; 0.017]
Fan et al. 2010	40	7139		0.006	[0.004; 0.008]
Kelly et al. 2021	1232	31984		0.039	[0.036; 0.041]
Messner et al. 2001	343	5771		0.059	[0.053; 0.066]
Tanyeri Toker et al. 2023	3	570		0.005	[0.001; 0.015]
Tsuchiya et al. 2006	31	8979		0.003	[0.002; 0.005]
Random effects model	**1715**	**59487**		**0.014**	**[0.008; 0.026]**
Heterogeneity: $I^2 = 36.2\%$, $\tau^2 = 0.669$					

S2 Fig: Random effects meta-analysis of the refer rate (RFR) for study protocols on TEOAE-AABR, AABR-TEOAE and AABR-AABR in well babies. Shown is the number of newborns who did not pass first and second test ("Failures"), number of screened newborns ("Total screened") and the RFR with the 95% confidence interval (95% CI) for each study. The summary estimate per test combination including the 95% CI is shown as gray diamond. The vertical solid line shows the threshold of 4% for the RFR. AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission.

TEOAE-TEOAE

Study	Failures	Total screened		RFR	95% CI
Aidan et al. 1999	124	1421		0.087	[0.073; 0.103]
Alanazi 2020	504	20171		0.025	[0.023; 0.027]
Arslan et al. 2013	55	2229		0.025	[0.019; 0.032]
Benito-Orejas et al. 2008	64	2454		0.026	[0.020; 0.033]
Bevilacqua et al. 2010	885	11466		0.077	[0.072; 0.082]
Calevo et al. 2007 (a)	58	3238		0.018	[0.014; 0.023]
Calevo et al. 2007	462	31796		0.015	[0.013; 0.016]
Calvacanti et al. 2012	290	3724		0.078	[0.069; 0.087]
Chapchap et al. 2001	38	4196		0.009	[0.006; 0.012]
De Capua et al. 2007	447	19700		0.023	[0.021; 0.025]
Escobar-Ipuz et al. 2019	204	9390		0.022	[0.019; 0.025]
Farahani et al. 2017	8	2784		0.003	[0.001; 0.006]
Ferlito et al. 2021	361	37562		0.010	[0.009; 0.011]
Gilbey et al. 2013	261	4958		0.053	[0.047; 0.059]
Guastini et al. 2010	122	8671		0.014	[0.012; 0.017]
Habib et al. 2005	300	11986		0.025	[0.022; 0.028]
Hazopoulus et al. 2007	22	450		0.049	[0.031; 0.073]
Jakubikova et al. 2003	59	3048		0.019	[0.015; 0.025]
Korres et al. 2006 (1)	535	22195		0.024	[0.022; 0.026]
Korres et al. 2006 (2)	534	25032		0.021	[0.020; 0.023]
Kosmidou et al. 2021	13	1491		0.009	[0.005; 0.015]
Lin et al. 2004	231	5938		0.039	[0.034; 0.044]
Magnani et al. 2015	91	10359		0.009	[0.007; 0.011]
Molini et al. 2016	1006	18796		0.054	[0.050; 0.057]
Pastorino et al. 2005	175	19290		0.009	[0.008; 0.011]
Prpic et al. 2007	179	11746		0.015	[0.013; 0.018]
Satish et al. 2019	408	26487		0.015	[0.014; 0.017]
Sequi Canet et al. 2016	182	14015		0.013	[0.011; 0.015]
Sheng et al. 2021	39	1340		0.029	[0.021; 0.040]
Tasci et al. 2010	322	15323		0.021	[0.019; 0.023]
Tatli et al. 2007	11	711		0.015	[0.008; 0.028]
Unlu et al. 2015	126	2933		0.043	[0.036; 0.051]
Vos et al. 2014	12333	245219		0.050	[0.049; 0.051]
Welzl-Müller et al. 1997	120	2338		0.051	[0.043; 0.061]
Random effects model	**20569**	**602457**		**0.023**	**[0.019; 0.029]**

Heterogeneity: $I^2 = 22.8\%$, $\tau^2 = 0.429$

RFR axis: 0   0.02 0.04 0.06 0.08 0.1

S3 Fig: Random effects meta-analysis of the refer rate (RFR) for 34 study protocols on TEOAE-TEOAE without study by Clarke et al. 2003. Shown is the number of newborns who did not pass first and second test ("Failures"), number of screened newborns ("Total screened") and the RFR with the 95% confidence interval (95% CI) for each study. The summary estimate including the 95% CI is shown as gray diamond. The vertical solid line shows the threshold of 4% for the RFR. TEOAE = transitory evoked otoacoustic emission.

S1 Table: Results of the meta-regression.

Test combination	Mean slope (95% CrI)
TEOAE-TEOAE	0.346 (0.341, 0.352)
TEOAE-AABR	0.310 (0.282, 0.339)
AABR-AABR	0.280 (0.267, 0.293)

SD = standard deviation, CrI = credibility interval, AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission.

S4 Fig: Refer rates in dependence of the failure rate of the first test for the test combinations AABR-AABR (n=13, blue triangles), AABR-TEOAE (n=3, green diamonds), TEOAE-AABR (n=6, red points) and TEOAE-TEOAE (n=35, gray squares). AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission. Meta-regression lines including credibility intervals are plotted with data from S1 Table. No meta-regression line is shown for AABR-TEOAE test combination with only three studies.

S2 Table: Study-level assessment of bias for all 57 study protocols.

Study	Risk of bias				Applicability concerns		
	Patient selection	First test	Second test	Flow and timing	Patient selection	First test	Second test
TEOAE-TEOAE:							
Aidan et al. 1999	low	low	low	high	high	low	low
Alanazi 2020	low	low	low	high	high	low	low
Arslan et al. 2013	high	low	low	low	high	low	low
Benito-Orejas et al. 2008	high	low	low	high	high	low	low
Bevilacqua et al. 2021	low	low	low	high	high	low	low
Calevo et al. 2007 (a)	unclear	low	low	low	low	low	low
Calevo et al. 2007	low	low	low	low	low	low	low
Calvacanti et al. 2012	high	low	low	high	high	low	low
Chapchap et al. 2001	high	low	low	high	high	low	low
Clarke et al. 2003	high	low	low	low	low	low	low
De Capua et al. 2007	high	low	low	high	high	low	low
Escobar-Ipuz et al. 2019	high	low	low	low	high	low	low

	Risk of bias				Applicability concerns		
Study	Patient selection	First test	Second test	Flow and timing	Patient selection	First test	Second test
Farahani et al. 2017	low	low	low	low	low	low	low
Ferlito et al. 2021	high	low	low	high	high	low	low
Gilbey et al. 2013	low	low	low	low	low	low	low
Guastini et al. 2010	high	low	low	low	high	low	low
Habib et al. 2005	low	low	low	low	low	low	low
Hatzopoulus et al. 2007	low	low	low	high	high	low	low
Jakubikova et al. 2003	high	low	low	high	high	low	low
Korres et al. 2006 (1)	high	high	low	high	high	high	low
Korres et al. 2006 (2)	low	low	low	low	low	low	low
Kosmidou et al. 2021	low	low	low	low	low	low	low
Lin et al. 2004	high	low	low	high	high	low	low
Magnani et al. 2015	low	low	low	low	low	low	low
Molini et al. 2016	low	low	low	high	high	low	low
Pastorino et al. 2005	low	low	low	high	high	high	low
Prpic et al. 2007	high	low	low	low	high	low	low
Satish et al. 2019	high	low	low	low	high	low	low
Sequi Canet et al. 2016	low	low	low	high	high	low	low
Sheng et al. 2021	low	low	low	high	high	low	low
Tasci et al. 2010	low	low	low	low	high	low	low
Tatli et al. 2007	high	low	low	high	high	low	low
Unlu et al. 2015	low	low	low	low	low	low	low
Vos et al. 2014	low	low	low	high	high	low	low
Welzl-Müller et al. 1997	low	low	low	high	high	low	low
TEOAE-AABR:							
Dort et al. 2000	high	low	low	low	high	low	low
Lin et al. 2007	low	low	low	low	low	low	low
Olusanya et al. 2008	low	low	low	low	low	low	low
Olusanya et al. 2010	high	low	low	high	high	low	low
Ong et al. 2020	high	low	low	low	high	low	low
Sheng et al. 2021	low	low	low	low	low	low	low
AABR-TEOAE:							
Dort et al. 2000	high	low	low	low	high	low	low
Ong et al. 2020	low	low	low	low	high	low	low
Sheng et al. 2021	low	low	low	low	low	low	low
AABR-AABR:							
Ayas et al. 2021	low	low	low	low	low	low	low
Benito-Orejas et al. 2008	low	low	low	high	high	low	low
Clarke et al. 2003	high	low	low	low	low	low	low
Clemens et al. 2001	low	low	low	low	low	low	low
Fan et al. 2010	high	low	low	low	high	low	low
Gupta et al. 2015	high	low	low	low	high	low	low
Huang et al. 2013	high	low	low	low	high	low	low
Iwasaki et al. 2003	low	low	low	low	low	low	low
Kelly et al. 2021	low	low	low	low	low	low	low
Messner et al. 2001	low	low	low	high	high	low	low
Shim et al. 2021	high	low	low	low	high	low	low
Tanyeri Toker et al. 2023	low	low	low	low	low	low	low
Tsuchiya et al. 2006	high	low	low	high	low	low	low

AABR = automated auditory brainstem response, TEOAE = transitory evoked otoacoustic emission.

Population: (Well) babies undergoing a two-stage hearing screening
Setting: Newborns from studies from all countries (high, middle and low-income countries) with the following criteria
(1) Initial screening as an inpatient in the maternity clinic
(2) Second test up to a maximum of one month later
(3) No use of distortion product otoacoustic emissions (DPOAE)
(4) Not exclusively NICU (neonatal intensive care unit).
Includes studies of well babies only and studies of well babies and NICU infants. Studies of exclusively NICU infants were not included in this review.
Intervention: Two-stage hearing screening using TEOAE, AABR, or combination of both
Comparison: not applicable

Test findings			Certainty of the evidence (GRADE)	Comments
AABR	Pooled sensitivity (95% CrI)	98.9% (96.0, 100.0%)	Moderate[1]	Sensitivity of a test for a condition with low prevalence needs to be high, assumption of equal sensitivities across all studies
	Pooled specificity (95% CrI)	92.7% (92.5, 92.9%)	Moderate[2]	Assumption of equal specificities across all studies
TEOAE	Pooled sensitivity (95% CrI)	96.9% (92.5, 99.8%)	Moderate[1]	Sensitivity of a test for a condition with low prevalence needs to be high, assumption of equal sensitivities across all studies
	Pooled specificity (95% CrI)	91.3% (91.2, 91.4%)	Moderate[2]	Assumption of equal specificities across all studies
Screening algorithm	**Pooled refer rate (95% CI)**	**Number of newborns (studies)**		
AABR-AABR	1.4% (0.9, 2.2%)	87803 (13)	Moderate[3]	Moderate heterogeneity, but larger number of studies
TEOAE-TEOAE	2.5% (2.0, 3.1%)	602538 (35)	Moderate[3]	Moderate heterogeneity, but large number of studies
TEOAE-AABR	4.9% (3.1, 7.5%)	11724 (6)	Moderate[4]	Low heterogeneity, but small number of studies
AABR-TEOAE	5.9% (5.0, 6.9%)	2316 (3)	Moderate[4]	No evidence of heterogeneity, but small number of studies

S1 File: Summary of findings table. CrI = Credibility interval, CI = confidence interval, GRADE = GRADE Working Group grades of evidence (high, moderate, low, very low).[1]Level of evidence was lowered by one level due to assumption of equal sensitivities in the Bayesian analysis. [2]Level of evidence was lowered by one level due to assumption of equal specificities in the Bayesian analysis. [3]Level of evidence was lowered by one level due to found moderate heterogeneity in the random effects meta- analysis. [4]Level of evidence was lowered by one level due to small number of pooled studies in the random effects meta-analysis.