Artificial Intelligence-Based Automated Interpretation of Images of Electrocardiograms: Development and Multinational Validation of ECG-GPT

Akshay Khunte BS[1,2*], Veer Sangha BS[1,3*], Evangelos K Oikonomou MD DPhil[1], Lovedeep S Dhingra MBBS[1], Arya Aminorroaya MD MPH[1], Andreas Coppi PhD[1,4], Sumukh Vasisht Shankar MS[1], Elijah Rockers MS[5], Bobak J Mortazavi PhD[4,6], Deepak L Bhatt MD MPH[7], Harlan M Krumholz MD SM[1,4,8], Sadeer Al-Kindi MD[5], Girish N Nadkarni MD MPH[9,10#], Akhil Vaid MD[9,10#], Rohan Khera MD MS[1,4,11#]

[1]Section of Cardiovascular Medicine, Department of Internal Medicine, Yale School of Medicine, New Haven, CT
[2]NYU Grossman School of Medicine, New York, NY
[3]Department of Engineering Science, Oxford University, Oxford, UK
[4]Center for Outcomes Research and Evaluation, Yale-New Haven Hospital, New Haven, CT
[5]Center for Cardiovascular Computational & Precision Health, Houston Methodist DeBakey Heart & Vascular Center, Houston, TX, USA
[6]Department of Computer Science & Engineering, Texas A&M University, College Station, TX
[7]Mount Sinai Fuster Heart Hospital, Icahn School of Medicine at Mount Sinai, New York, NY
[8]Department of Health Policy and Management, Yale School of Public Health, New Haven, CT
[9]The Charles Bronfman Institute for Personalized Medicine, Icahn School of Medicine at Mount Sinai, New York, NY, USA
[10]The Division of Data Driven and Digital Medicine, Department of Medicine, Icahn School of Medicine at Mount Sinai, New York, NY, USA
[11]Section of Health Informatics, Department of Biostatistics, Yale School of Public Health, New Haven, CT

*Contributed equally as co-first authors
#Contributed equally as co-senior authors

Correspondence to: Rohan Khera, MD, MS
195 Church St, 6[th] Floor, New Haven, CT 06510,
203-764-5885; rohan.khera@yale.edu; @rohan_khera

ABSTRACT

Background: Timely and accurate assessment of electrocardiograms (ECGs) is crucial for

diagnosing, triaging, and clinically managing patients. Current workflows rely on computerized

ECG interpretation tools built into ECG signal acquisition systems, which use rule-based

algorithms that are unreliable and frequently not available in low-resource settings. We

developed and validated a format-independent vision encoder-decoder model – ECG-GPT – that

can generate free-text, expert-level interpretations directly from 12-lead ECG images.

Methods: Using 12-lead ECGs and their corresponding diagnosis statements collected at the

Yale-New Haven Health System (YNHHS) between 2000 and 2022, we developed a vision-text

transformer model to generate interpretation statements from images of ECGs. Using structured

clinical assessment, semantic similarity, and conventional natural language generation metrics,

we validated ECG-GPT across 7 geographically distinct health settings. These include (1) 3 large

and diverse US health systems, (2) consecutive ECGs from a central reading system in Minas

Gerais, Brazil, (3) the prospective cohort study, UK Biobank, (4) a Germany-based, publicly

available repository, PTB-XL, and (5) a community hospital in Missouri.

Results: Overall, 2.9 million ECGs were used for model development. The model performed

well in clinical assessment across 26 extracted labels: for atrial fibrillation, sinus tachycardia,

sinus bradycardia, premature atrial contractions, and premature ventricular contractions,

AUROCs and AUPRCs ranged from 0.80-0.95 and 0.50-0.86, respectively. For left bundle

branch block, right bundle branch block, first degree atrioventricular block, left anterior

fascicular block, and left posterior fascicular block, AUROCs and AUPRCs ranged from 0.88-

0.96 and 0.23-0.86, respectively. Across all 26 conditions, diagnostic accuracy ranged between

0.93-0.99. ECG-GPT identified the full context of the diagnosis statements with allied

conditions. It had a median pairwise cosine similarity of 0.90 (IQR 0.83-0.97), significantly

greater than the median baseline similarity of 0.73 (IQR 0.67-0.78, p<0.001). This separation

between median pairwise and baseline similarity remained consistent across all 26 condition-

specific subsets. The results were comparable across external validation sites.

Conclusions: We developed and extensively validated a vision encoder-decoder model that

generates expert-level interpretations from ECG images. This represents a scalable and

accessible strategy for automated ECG analysis, especially in low-resource settings.

ABBREVIATIONS AND ACRONYMS

ECG     Electrocardiography

AI      Artificial Intelligence

YNHH    Yale New Haven Hospital

MSHS    Mount Sinai Health System

HMH     Houston Methodist Hospital

LRH     Lake Regional Hospital

AUROC    Area under receiving operation characteristics

AUPRC    Area under precision recall curve

CLINICAL PERSPECTIVE

What is New?

- ECG-GPT is a vision encoder-decoder model capable of generating full-text ECG

interpretations directly from ECG images, regardless of layout or format.

- The model was trained on over 2.7 million ECGs and externally validated across 3.8

million additional ECGs from demographically and geographically diverse populations.

78
What are the clinical implications?

- ECG-GPT enables automated, expert-level ECG interpretation directly from images,

eliminating the need for signal data or device integration.

- The model demonstrates consistent performance across diverse patient populations, ECG

formats, and care settings.

- This scalable, image-based approach may expand access to accurate ECG interpretation

in low-resource settings.

INTRODUCTION

Electrocardiography (ECG) is a widely available, first-line, noninvasive tool for diagnosing, triaging, and managing cardiovascular disease.[1] Traditional workflows often rely on computerized ECG interpretation algorithms to generate preliminary reads, which, despite limited accuracy,[2,3] provide diagnostic support and enable faster triage for high-risk conditions.[4] Such algorithms, however, are often proprietary, require raw signal data, and have substantial variability in accuracy,[4,5] with the potential to lead to patient harm.[6,7] This makes computerized pre-reads inaccessible to clinicians in rural and low-resource settings, a disparity further exacerbated by the lower availability of expert-level readers.[8,9] This lack of reliable and interoperable automated system-generated ECG reports in many low-resource settings globally highlights the need for an accurate, easily accessible ECG interpretation tool.

This is particularly pronounced in community health centers in rural areas of low- to middle-income countries, where junior health professionals and community health workers provide preventive and basic care, including the interpretation of ECGs. Further management frequently relies on telecommunications with a limited number of experts available to provide such consultation.[10] Moreover, the accuracy of ECG interpretations by physicians is variable even after educational interventions, with less than 75% accuracy reported for even cardiologists.[11] Therefore, there is an unmet need for accessible, reliable, and accurate tools to provide expert-level ECG interpretations.

Though recent advances in deep learning enable accurate classification of specific ECG abnormalities,[12–15] they are generally limited to a select number of commonly encountered abnormalities and do not address less common conduction and rhythm disorders or variations within common rhythms. However, even these less common abnormalities may have significant

therapeutic and prognostic implications despite their low prevalence. Additionally, while our

prior work has demonstrated that diagnostic models can directly identify information from ECG

images,[15,16] most traditional signal-based models, like computerized interpretation algorithms,

predominantly rely on raw signal data, limiting their scalability to the point-of-care and across

low-resource settings. The development of models exclusively for images is challenged by

variations in the layout of the leads, labeling of the leads, structure and design of the graphed

background, and the quality of the image acquisition.

In this study, we report the development of ECG-GPT (**Figure 1**), a novel vision-text

transformer model capable of generating diagnostic reports from ECG images regardless of the

layout, trained against the full breadth of expert-verified ECG interpretations across 2.7 million

12-lead ECGs collected over 21 years in a large US-based hospital system. ECG-GPT can be

accessed as a web-based application that can receive ECG images across formats and layouts as

the only input and generate diagnostic reports (demonstration hosted at https://www.cards-

lab.org/ecg-gpt). We pursued broad multinational validation in 3.8 million ECGs across

temporally and geographically distinct datasets, including two distinct US-based major referral

hospital systems, a Brazil-based large telehealth network, a UK-based prospective cohort study,

a publicly available ECG dataset from Germany, and a rural US-based community hospital.

METHODS

The study was reviewed by the Yale Institutional Review Board, which approved the study

protocol and waived the need for informed consent as the study represents secondary analysis of

existing data. The development dataset from Yale and the validation datasets from Mount Sinai,

Houston Methodist Hospital, and Lake Regional Hospital are not publicly available, given the

stipulations of the relevant Institutional Review Boards. The external datasets from Brazil, UK

Biobank, and PTB-XL are available directly from the respective groups and are outside the

purview of the authors. The model is publicly available online as a web-based application for

research use at https://www.cards-lab.org/ecg-gpt.

Data Source for Model Development

Raw voltage data were collected for all 12-lead ECGs with corresponding diagnosis statements

obtained at the Yale New Haven Health System (YNHHS) during 2000-2021. Each ECG was

recorded with a sampling frequency of 500 Hz using Philips PageWriter and GE MAC machines.

The subset of these ECGs with continuous recording across all 12 leads was then split at the

patient level into training, validation, and test sets (85%, 5%, 10%). For the test set, we restricted

to ECGs with no marked abnormalities and abnormal ECGs with confirmed reports certified by a

cardiologist. In the training and validation sets, we chose to include abnormal, unconfirmed

ECGs collected in the emergency setting, since these are interpreted at the point-of-care, and

may not be listed as a confirmed read in the system. Finally, in the training set, a random set of

ECGs with no marked abnormalities were removed to match the observed prevalence of such

ECGs in the original cohort (**Figure S1**).

Signal Preprocessing and ECG Image Generation

First, all ECGs that did not contain 10 seconds of continuous recording across all 12 leads were

excluded. To enable the generation of ECG images like those used in clinical settings, we further

preprocessed the signal before plotting. For this, we subtracted a one-second median filter from

the original raw voltage for each lead of 10-second ECGs to remove baseline wander, mirroring

the approach undertaken by ECG machines before printing clinical ECGs available to and

interpreted by physicians. All ECGs obtained in the YNHHS were plotted at their original

sampling rate of 500 Hz. ECGs used for external validation, which were recorded at sampling

frequencies between 300-600 Hz, were down sampled to 300 Hz before plotting, as described

previously.[15] While the ECGs in the development set were recorded using GE MAC and Philips

PageWriter machines, those in the external validation were frequently only GE MAC models,

even though the specific models used varied within individual validation datasets.

The preprocessed ECG signals were transformed into ECG images at 100 dots per inch (DPI)

using the python library ecg-plot.[17] We employed multiple strategies to ensure the robustness of

the model. First, we converted each signal waveform to multiple images using four different

layouts of the leads to account for different schemes of real-world ECGs. To ensure the model is

resilient to these other formats, it was trained using all these variations of ECG images, an

approach we previously developed and validated.[15] The four formats used in model development

(**Figure 1**) included (1) The standard printed ECG format in the US with lead I as the rhythm

strip. This format consists of four 2.5-second sequential columns, each containing a 2.5-second

strip from three leads. (2) The same as the standard US format that includes an additional rhythm

strip from lead II. (3) An alternate format with no rhythm strip comprising two 5-second

columns. The first column represents a simultaneous recording from limb leads, while the second

column represents a simultaneous recording from precordial leads. (4) A shuffled format in

which precordial leads are recorded in the first two columns and limb leads are represented in the

last two columns.

Second, the conversion of ECG signals to images was done independently of the model

development to ensure the model remains agnostic to preprocessing steps from ECG signals to

images. Third, all images were rotated by a random amount between -10 to 10 degrees prior to

training. Finally, we used Python Image Library (PIL v9.2.0) to convert all ECG images to

greyscale and down-sample them to the required size for input into the model regardless of their

initial resolution. Training the model using plotted ECG images instead of raw signal data

translates ECG image inference as a computer vision application rather than a signal extraction

or processing task. ECG-GPT leverages the entire ECG representation during inference,

incorporating background gridlines and lead labels. This approach, combined with the varied

formats used during training, enables generating interpretations directly from the ECG images

available to end-user clinicians.

Image Standardization for Model Inference

We have implemented a previously validated approach to standardize model inputs.[16] Inputs are

limited to 12-lead ECG tracings, which are vertically oriented, minimally rotated, have a uniform

background, and do not have peripheral annotations. Additionally, to mitigate the effects of

noise, a two-step preprocessing approach is applied to each image: first, the image is straightened

and cropped to correct for rotations and to remove any elements outside of the ECG tracing,

including any potential identifiers or interpretation summary. Second, the algorithm scales the

brightness and contrast of the ECGs to the mean values of the development population before

generating model predictions (**Figure S2**). ECGs with deviations in brightness and contrast 50%

greater or lower than those seen in the development set are flagged as requiring the image to be

recaptured in better quality before inference (**Figure S3**).

Diagnostic Statement Preprocessing

Cardiologist-confirmed diagnosis statements were preprocessed to remove all identifying

information using a rule-based approach. A string search was performed for all diagnostic

statements to identify and remove names, references to previous ECGs, all dates and times, and

punctuation. The Python PyEnchant package was then used to generate a list of all abbreviations

present in the processed diagnosis statements.[18] A pair of clinical experts manually generated a

dictionary containing each abbreviation and its expanded form for the 100 most common

abbreviations (**Table S1**). Each instance of these abbreviations in the diagnosis statement was

then replaced with its expansion. Additionally, the most common misspellings and all synonyms

for a condition were replaced with a single term. This processed diagnosis statement was then

used for model development and evaluation (**Figure S4**).

Model Development

We built a custom Vision Encoder-Decoder model using the HuggingFace framework.[19] For the

image encoder, we selected a BEiT transformer model, pretrained on the ImageNet dataset,[20] due

to its robust performance on the ImageNet-1K benchmark, relatively few trainable parameters

compared to other state-of-the-art models, and the 384x384 pixel input size.[21] We selected the

base-size version of the model, with ~84 million trainable parameters, which takes a linearly

embedded sequence of 16x16 pixel patches as the input.

We selected the Generative Pretrained Transfomer-2 (GPT-2) transformer for the text

decoder, initially developed for prompt-based text generation. This architecture, which contains

~124 million parameters, decoded the lower-dimensional representations generated from images

by the BEiT encoder into ECG interpretation statements with a maximum output token length of

48 tokens. To better tailor the model outputs to the text found in ECG diagnosis statements and

to reduce the number of tokens needed, thus lowering the computational costs of using the

model, we first fine-tuned the GPT-2 tokenizer on the ECG diagnosis statements used to train the

model. In addition to the lightweight nature of GPT-2 relative to other large language models

capable of text generation, the extensive pretraining of the GPT-2 decoder in combination with

the fine-tuned tokenizer enabled direct integration into the vision-text transformer architecture

without further fine-tuning.

Overall, the composed Vision Encoder-Decoder model, consisting of the BEiT encoder and

GPT-2 decoder, has over 239 million trainable parameters (**Figure 2**). The model was trained on

a generated image dataset at a learning rate of 5 x 10^{-5} for 20 epochs. The model randomly

selected a format each time an ECG was loaded during training to reduce overfitting. We used

the Adam optimizer, a minibatch size of 14, and a cross-entropy loss function to minimize the

error between the GPT-2 output and the tokenized reference diagnostic statements to train the

model.[22] Model development used the HuggingFace Transformers 4.28.1 framework with

PyTorch 2.0.0 and Python 3.11.3 on eight NVIDIA A100 80GB graphics processing units.

External Validation

We pursued validation on five generated ECG image datasets and three native ECG image

datasets acquired outside the YNHHS. These included A) one generated image dataset and one

distinct native image dataset from Mount Sinai Health System (MSHS) in New York, NY, B) a

native image dataset from Houston Methodist Hospital (HMH) in Houston, TX, C) a native

image dataset from Lake Regional Hospital (LRH) in Osage Beach, MO, and D) images

generated from four publicly available ECG signal datasets.

Both sets from MSHS and the set from HMH contained corresponding diagnosis statements.

The first set from MSHS included 1,434,455 ECGs collected from 2013 to 2023 that were then

plotted in the standard image format used in training. The second included 10,052 distinct, native

ECG images collected prospectively from May to July 2024. The set from HMH included 62,876

native ECG images collected from 2018 to 2024. A federated approach was implemented to

enable external validation within the MSHS and HMH. Both the model and all semantic and

clinical assessment code were containerized using Docker and securely deployed within the

infrastructure of each hospital. It accepted file paths as input, ensuring that patient data remained

within the hospital's system without the need for external data sharing. This approach facilitated

accurate prediction generation while safeguarding patient privacy. Notably, the lightweight

design of the model enables GPU-less inference on devices with at least 16GB of RAM. This,

combined with the containerized approach, allows for efficient local deployment on a wide array

of machines, enhancing accessibility in settings with limited internet connectivity.

The set from LRH included 64 native ECG images with diagnostic labels for six rhythm and

conduction disorders, including atrial fibrillation (AF), sinus tachycardia (ST), sinus bradycardia

(SB), left bundle branch block (LBBB), right bundle branch block (RBBB), and first degree

atrioventricular block (1dAVb). The set included 8-10 ECGs with each of these labels and ECGs

labeled as normal. Though the layout of these ECGs was similar to the standard layout used for

model development, there were multiple key distinctions. First, the V1 lead, instead of lead I,

was the rhythm lead. Second, the signal was black, as opposed to blue, and vertical lines were

separating the leads. Additionally, there were variations in background and grid color, as well as

in the position and font of the lead label.

Validation was also pursued in images generated from four geographically distinct, open-

source ECG signal datasets labeled with the same six rhythm and conduction disorders as the set

from LRH. First, we obtained 45,389 ECGs from the UK Biobank, under research application

#71033. The UK Biobank represents the largest population-based cohort of 502,468 people in

the United Kingdom with protocolized imaging, laboratory testing, and linked electronic health

records. We also used a set of 2,228,236 ECGs from the CODE15 study dataset, a set of ECG

recordings previously used for both signal- and image-based multilabel ECG classification

models.[12,15,23] As the primary CODE15 dataset consists of ECGs collected and annotated for six

rhythm and conduction disorders by individual clinicians during routine care, we also deployed

the model in a secondary cardiologist-validated dataset. This dataset contained 827 additional

ECGs collected in the Telehealth Network of Minas Gerais between April and September

2018.[12,15] For each of these ECGs, annotations for the six rhythm and conduction disorders were

made by two independent cardiologists following criteria from the American Heart

Association,[24] with disagreements resolved by a third cardiologist. Finally, the model

performance was assessed in PTB-XL, a previously described dataset of ECGs.[25] This dataset

contains 21,837 10-second, 12-lead recordings collected at 500 Hz from 18,885 patients in

Germany between 1989 and 1996. The records for each ECG, including diagnostic, form, and

rhythm statements, were used to extract labels for the same set of rhythm and conduction

disorders.

Model Evaluation

We evaluated ECG-GPT using three distinct strategies: (1) semantic similarity to assess the

model's ability to capture the full clinical context of reference interpretations, (2) syntactic

similarity to quantify the overlap between reference and generated statements, and (3) structured

label assessment to evaluate the model's accuracy for specific clinical labels extracted from the

statements. Clinical labels for 26 conditions, spanning key rhythm and conduction disorders

selected by two cardiologists, were extracted from each reference and model-generated statement

using a standardized string search approach. For each condition, basic string search was

performed to extract each label using a set of strings (**Table S2**). A condition was flagged as

positive if the statement contained a full match for any string in the set. ECGs were flagged as

negative if there was no match or if the match was preceded by a negation, including "no" or

"without".

First, for semantic similarity, we fine-tuned a lightweight DistilBERT model,[26,27] pretrained

on a large corpus of electronic health record notes,[28,29] in the same set of cardiologist-confirmed

diagnosis statements used to train the vision-text transformer model. The training mirrored the

standard approach for training masked language models, with a chunk size of 128 tokens and a

masking probability of 0.15. The model was trained at a 2×10^{-5} learning rate until validation

loss did not improve for three consecutive epochs. We used the Adam optimizer, a minibatch

size of 16, and a cross-entropy loss function to minimize the error between the masked and

generated tokens to train the model.[22] Model development used the HuggingFace Transformers

4.28.1 framework with Torch 2.0.0 and Python 3.11.3 on four RTX 3090 graphics processing

units.

After fine-tuning, we deployed the masked language model in the held-out test set and the

MSHS external validation set to generate 768-dimensional embeddings for each reference and

model-generated statement. We used an identical federated approach to the model-generated

statements to generate embeddings for the statements within the MSHS and HMH. Pairwise

similarity was computed as the median cosine similarity between the embeddings for each

reference statement and its paired model-generated statement. Baseline similarity was computed

as the median cosine similarity between the embeddings for 100,000 random pairings of

reference and model-generated statements. Of note, there are no benchmarks for semantic

similarity, as its interpretation varies based on the specific task and data context. Cosine

similarity values range from 0 to 1, with higher values indicating higher similarity between two

statements.

To assess the model's ability to diagnose specific conditions within their complete clinical

context, we created subsets for each of the 26 diagnostic labels identified from the reference

statements, consisting of all ECGs marked positive for that condition. Pairwise and baseline

similarity were computed identically to the approach used for the complete datasets. For subsets

too small to generate 100,000 random pairings, the baseline similarity was reported as the

median cosine similarity between embeddings for all possible reference and model-generated

statement pairings within the subset.

Next, we used four conventional NLG metrics to assess the syntactic similarity between the

original diagnosis reports and generated text. ROUGE and BLEU, which range from 0 to 1,

evaluate recall and precision, respectively, for the overlap of n-grams between generated and

reference statements, providing insight into content overlap and coherence.[30,31] METEOR, which

ranges from 0 to 1, incorporates both syntactic and semantic similarity by aligning word stems

and synonyms, enabling an evaluation of content relevance in addition to word overlap.[32]

Finally, CIDEr, which ranges from 0 to 5, measures consensus between generated text and

reference summaries through similarity to human consensus, enhancing evaluation robustness

across various linguistic styles.[33] Collectively, these metrics offer a comprehensive assessment of

the syntactic similarity between original diagnosis statements and model-generated statements.

Each of these metrics was deployed in both the internal held-out test set using the HuggingFace

Evaluate package for computing ROUGE, BLEU, and METEOR scores and the COCO Caption

Evaluation package for computing CIDEr scores, respectively.[34,35]

Finally, we also implemented a secondary analysis of model performance using the 26

extracted labels. For each reference and model-generated statement, we computed agreement

between the statements for each of the 26 rhythm and conduction disorders. For the open-source

datasets, we compared the corresponding reference labels for AF, ST, SB, RBBB, LBBB, and

1dAVb to the labels extracted from the model-generated statements for these ECGs. We used

multiple metrics to comprehensively evaluate ECG-GPT's performance on structured labels,

including the area under the receiver operating characteristic (AUROC) to measure model

discrimination between classes. 95% confidence intervals for AUROC were calculated using

DeLong's algorithm.[36,37] We also assessed the area under precision-recall curve (AUPRC),

which is especially informative for imbalanced classes as it emphasizes correct identification of

the positive class. We employed the bootstrap resampling method to estimate confidence

intervals for AUPRC.[38] Additionally, we evaluated accuracy, sensitivity, specificity, F1 score,

positive predictive value (PPV), and negative predictive value (NPV) across conditions to

provide a thorough assessment of the model's performance.

Statistical Analysis

Summary statistics are presented as counts and percentages for categorical elements and median

and interquartile range (IQR) for continuous elements. A paired t-test was used to compute the

probability of overlap between the pairwise and baseline cosine similarity of reference and

model-generated statements. All analyses were performed using Python 3.11.3, and the

significance level was set at an alpha of 0.05.

RESULTS

Study Population

We developed ECG-GPT in a set of 2,888,384 12-lead ECG recordings with accompanying

cardiologist-confirmed diagnosis statements performed on 601,616 unique patients at the Yale-

New Haven Health System (YNHHS) between 2000 and 2022. These ECGs reflected a wide

distribution of demographics, with a mean age of 63.2 (SD 18.0) years at the time of the ECG. In

the development population, 1,393,683 (48.2%) of the ECGs were from women, 1,904,662

(65.9%) ECGs were from non-Hispanic White, 437,244 (15.1%) non-Hispanic Black, 298,490

(10.3%) Hispanic, 34,855 (1.2%) non-Hispanic Asian, and 213,133 (7.4%) from patients from

other racial backgrounds (**Table S3**), reflecting broad and diverse representation. In a period

when direct linkage to hospital admission records data was possible (with the digitization of

health records, starting 2013), 37.5%, 13.5%, and 49.0% of ECGs were performed in the

inpatient, emergency, and outpatient settings, respectively.

After processing with a rule-based approach, the diagnosis statements corresponding to each

ECG had a median length of 81 characters (IQR 51, 120) and a median length of 12 tokens (IQR

8, 17) after tokenization. In the development cohort, most ECGs reported sinus rhythm

(1,935,574, 67.0%), followed by left atrial enlargement (359,747, 12.5%), left ventricular

hypertrophy (342,340, 11.9%), and atrial fibrillation (AF) (281,961, 9.8%). A total of 265,640

(9.2%) and 126,757 (4.4%) ECGs reported sinus tachycardia (ST) and sinus bradycardia (SB),

respectively. Right bundle branch block (RBBB) and left bundle branch block (LBBB) were

present in 195,474 (6.8%) and 94,928 (3.3%) ECGs, respectively, and 211,836 (7.33%) ECGs

reported first degree atrioventricular block (1dAVb). Though the presence of multiple rhythms

over a 10-second duration would be unusual, if the patient had a series of conduction and rhythm

disorders all present together (such as AF with premature ventricular contractions and bundle

branch blocks), all would be included in the original diagnosis statement and thus the extracted

labels. Accordingly, 104,451 (3.6%) ECGs featured multiple key rhythm disorders, including

AF, ST, SB, premature atrial contractions (PACs), and premature ventricular contractions

(PVCs). The proportion of ECGs with the 26 conditions, spanning rhythm and conduction

disorders extracted from the diagnosis statements is listed in **Table S3**.

Semantic Similarity

We used a fine-tuned DistilBERT language model to generate embeddings for each reference

and model-generated statements for each ECG in the 250,589 ECGs in the internal held-out test

set, obtained from 57,860 patients who had not contributed any data to the training or validation

set (**Table S3**). The median cosine similarity between the embeddings for reference statements

and their paired model-generated statements was 0.90 (IQR 0.83-0.97). This result was

significantly higher than the median cosine similarity between 100,000 randomly selected

combinations of reference and model-generated statements (0.73 (IQR 0.67-0.78, p<0.001).

In a secondary analysis to assess the model's ability to capture individual conditions within

their full clinical context, the median pairwise cosine similarity was significantly greater than the

respective median random cosine similarity across all 26 subsets (**Table 1**). Across all 26

conditions, pairwise and baseline similarities ranged from 0.84-0.97 and 0.72-0.84, respectively.

For key rhythm disorders – AF, ST, SB, premature atrial complexes (PACs), and premature

ventricular complexes (PVCs) – pairwise and random similarities ranged from 0.91-0.93 and

0.77-0.81, respectively. For conduction abnormalities – LBBB, RBBB, 1dAVb, left anterior

fascicular block (LAFB), and left posterior fascicular block (LPFB) – pairwise and random

similarities ranged from 0.91-0.95 and 0.79-0.84, respectively.

Syntactic Similarity

For NLG metrics, ECG-GPT matches or outperforms most state-of-the-art medical image

captioning models.[39–41] We report scores of 0.663 and 0.655 for ROUGE-1 and ROUGE-L,

respectively. For BLEU scores we report scores ranging from 0.595 for BLEU-1 to 0.419 for

BLEU-4. We also report a METEOR score of 0.693 and a CIDEr score of 3.32 (**Table S4**).

Structured Label Assessment

Model performance in the held-out test set for each of the 26 rhythm and conduction disorders,

including accuracy, positive and negative predictive values, specificity, sensitivity, AUROC,

AUPRC, and F1 scores, are recorded in **Table 2**. For AF, ST, SB, PACs, and PVCs, AUROCs

and AUPRCs ranged from 0.80-0.95 and 0.50-0.86, respectively. For LBBB, RBBB, 1dAVb,

LAFB, and LPFB, the AUROCs and AUPRCs ranged from 0.88-0.96 and 0.23-0.86,

respectively. Across all 26 conditions, diagnostic accuracy ranged between 0.93-0.99.

Furthermore, ECG-GPT performed comparably to two previously published multilabel CNN

models[14,42] across common labels (**Table S5**). There was no significant difference in the

performance for the labels of interest when comparing ECGs recorded in the inpatient,

emergency, and outpatient setting, with median AUROCs of 0.89 (IQR 0.85-0.92), 0.90 (IQR

0.87-0.94), and 0.89 (IQR 0.83-0.94), respectively (p=0.696).

External Validation – Mount Sinai Health System

We used a total of 1,434,455 ECG images generated from raw 12-lead ECG signal data and

10,052 prospectively acquired native ECG images from MSHS for validation. In assessing

semantic similarity for the set of generated ECG images, embeddings had a median pairwise

similarity of 0.86 (IQR 0.79-0.96), significantly greater than the median baseline similarity of

0.74 among 2 random statements (IQR 0.69-0.80, p<0.001). For the set of native ECG images,

embeddings had a median pairwise and baseline similarity of 0.80 (IQR 0.72-0.86) and 0.73

(IQR 0.68-0.78), respectively. This separation persisted across the 26 subsets corresponding to

each extracted rhythm and conduction disorder (**Table 3** and **Table 4**). For key rhythm disorders

– AF, ST, SB, PACs, and PVCs – pairwise and baseline similarity ranged from 0.86-0.89 and

0.78-0.81, respectively, for the generated ECG images and 0.78-0.87 and 0.75-0.80, respectively,

for the native ECG images. For key conduction abnormalities – LBBB, RBBB, 1dAVb, LAFB,

and LFPB – pairwise and baseline similarity ranged from 0.88-0.95 and 0.79-0.85, respectively,

for the generated ECG images and 0.84-0.85 and 0.77-0.81, respectively, for the native ECG

images.

The model performed well in clinical assessment across the 26 extracted labels. For AF, ST,

SB, PACs, and PVCs, AUROCs and AUPRCs ranged from 0.74-0.94 and 0.40-0.83,

respectively, for the generated ECG images and 0.67-0.91 and 0.30-0.76, respectively, for the

native ECG images. For LBBB, RBBB, 1dAVb, LAFB, and LFPB, AUROCs and AUPRCs

ranged from 0.87-0.97 and 0.14-0.83, respectively, for the generated ECG images and 0.71-0.97

and 0.05-0.63, respectively, for the native ECG images. Model performance in the native image

MSHS validation dataset across all 26 conditions is reported in **Table 5** and **Table 6**.

External Validation – Houston Methodist Hospital

We also deployed ECG-GPT in a set of 62,876 ECGs in their native image format from HMH.

Across conditions, embeddings had a median pairwise similarity of 0.81 (IQR 0.75-0.88),

significantly greater than the median baseline similarity of 0.77 among 2 random statements

(IQR 0.73-0.82, p<0.001). This separation persisted across the majority of the 26 subsets

corresponding to each extracted rhythm and conduction disorder (**Table 7**). In clinical

assessment, AUROCs and AUPRCs, respectively, for AF, ST, SB, PACs, and PVCs ranged from

0.65-0.96 and 0.13-0.68. For LBBB, RBBB, 1dAVb, LAFB, and LFPB, AUROCs and AUPRCs

ranged from 0.74-0.97 and 0.04-0.77, respectively. Model performance across all 26 conditions

is reported in **Table 8**.

External Validation – Lake Regional Hospital

Next, we used a real-world dataset of 64 ECG images collected at LRH in Osage Beach, MO. In these ECGs, the model had AUROCs of 0.95, 0.95, and 0.75 for AF, ST, and SB, respectively. For LBBB, RBBB, and 1dAVb, the model reported AUROCs of 0.82, 1.00, and 0.87, respectively.

External Validation – Open-source datasets

The model's diagnostic performance in each of the four open-source datasets used for validation is noted in **Table 9**. First, in 2,228,236 ECGs from the previously described CODE15 dataset collected by the Telehealth Network of Minas Gerais (TNMG), Brazil, between 2010 and 2017.[12,23] Here, AUROCs for rhythm disorders were 0.94, 0.92, and 0.92, and AUPRCs were 0.56, 0.54, and 0.11, for AF, ST, and SB, respectively. The model performed similarly for conduction abnormalities, with AUROCs of 0.90, 0.96, and 0.89, and AUPRCs of 0.62, 0.67, and 0.29 for LBBB, RBBB, and 1dAVb, respectively. When deployed to a smaller, cardiologist-validated dataset collected by TNMG in Brazil between April and September 2018, consisting of 827 ECGs manually annotated by two cardiologists with disputes resolved by a third,[15] AUROCs were higher across nearly all diagnostic labels. The model reported AUROCs of 1.00 (rounded from 0.998), 0.98, and 0.94 for AF, ST, and SB, respectively. For LBBB, RBBB, and 1dAVb, the model reported AUROCs of 0.94, 0.97, and 0.85, respectively.

The third of these four open-source datasets consisted of 45,389 ECGs obtained from patients enrolled in the UK Biobank. Here, the model had AUROCs ranging from 0.91 to 0.99 (**Table 9**), thus highlighting the reproducibility of our approach across clinical and non-clinical settings. Lastly, in 21,784 ECGs from the Germany-based PTB-XL dataset the model had AUROCs of

0.96, 0.95, and 0.91 for AF, ST, and SB, respectively, and 0.97, 0.98 and 0.84 for LBBB, RBBB,

and 1dAVb respectively.

DISCUSSION

We describe the development and external validation of ECG-GPT, a first-of-its-kind AI pipeline

that enables the direct generation of automated, complete ECG interpretation statements from

images of ECGs in any format. The model performs well against clinician-certified reports

across various natural language generation metrics and diagnostic labels spanning a wide array

of conduction, rhythm, and structural heart disorders. The model's scalability is further supported

by its robust performance across a range of demographically, temporally, and geographically

distinct cohorts, its online deployment, and the capacity to containerize and deploy the model

without sharing data in a federated approach.

Our work on format-independent, image-based ECG captioning represents a novel

development for vision-text machine learning models. Our model significantly outperforms prior

implementations for generating free-text reports from ECG signals, with CIDEr and METEOR

scores of 4.69 and 0.75, respectively, compared with scores of 2.55 and 0.27, respectively,

reported in a previous study suggesting high textual consistency between original and generated

reports.[41] Machine learning-based multilabel models have been previously developed to

simultaneously diagnose large sets of conditions, but these approaches are inherently limited to

those labels selected for training and do not capture the full diagnostic range of ECGs.[12,15,42]

Moreover, the utility of such signal models is limited to healthcare systems with the resources to

store signal data and incorporate models into the clinical workflow. We demonstrate consistent

performance across a range of external validation sets for classifying key rhythm and conduction

disorders. Of note, the performance for labels in external validation sets where specific labels

were explicitly available matches the performance on those labels in prior published reports. The

simplicity of this system, based on images, is that there is inherent interoperability and the

absence of a requirement to integrate with ECG machines to extract signals. This approach is

particularly advantageous in low-resource regions, where ECGs are currently not stored beyond

printing ECG images at the time of acquisition.[43] This approach also adds convenience and

provides access in any venue, including, for example, to emergency medical services providers

or in remote locations.

ECG-GPT can be used directly by clinicians at the point of care by uploading ECG images

from their phones or as scanned images to a web-based interface, with a demonstration

accompanying this study.[44] This applies anywhere that end-users may still lack access to

automated reads or require interpretation before a specialist's review. The image-based model

can also be more easily integrated into repositories of scanned ECGs, the most prevalent and

interoperable format for storing and sharing ECGs. Further, ECG-GPT has a unique combination

of diagnostic accuracy and range, demonstrating expert-level performance for key conditions

while also retaining the capability to generate statements for rare conditions frequently not

captured by standard multi-label models. This feature could make ECG-GPT a tool for

generating pre-reads and enabling more efficient triage globally in areas with insufficient access

to specialists and computerized ECG interpretation algorithms. The study here focuses on the

robustness of ECG-GPT in generating accurate interpretations from ECG images, but future

prospective studies are warranted to fully assess its integration into clinical workflows in care

settings where the tool is used.

Our study has several limitations. First, while the model accurately diagnosed the selected

conditions, it is impossible to determine and thus evaluate the performance for the full extent of

possible diagnoses the model could output for a given ECG image due to the size and variety of

the corpus of diagnosis statements used for model development. However, we report model

performance across various rhythm and conduction disorders of varying severity and prevalence,

suggesting that ECG-GPT's performance generalizes to other conditions. Moreover, using the

federated approach implemented for external validation, ECG-GPT could be continually fine-

tuned to improve performance in individual healthcare systems. This would ensure a reliable

pipeline with consistent performance for future ECGs within the specific patient populations in

which the model is deployed. Second, it currently only generates interpretation statements in

English, which may limit its utility in non-English-speaking settings.

Third, while four different formats were used during model development, we cannot

ascertain whether the model generalizes equally well to every other novel ECG image format.

However, the model's performance within the large, native image datasets from MSHS, HMH,

and LRH, each consisting of ECG images plotted in configurations distinct from those used in

model development, indicates the model can generate accurate interpretation statements for ECG

images with variations not seen during training. Finally, though strictly expert-validated

diagnosis statements were used to develop the model, these statements are not always completely

accurate, limiting the model's performance. As evidenced by the high diagnostic accuracy of the

model in the external set of ECGs manually annotated by two cardiologists, the model may

perform better if developed and evaluated in more rigorously validated diagnosis statements.

Furthermore, the current practice of clinicians over-reading the computer-generated reads and

correcting them without version control precludes the head-to-head assessment of ECG-GPT

against computer-generated reads. Future studies could compare the automated outputs of the

model to ECGs labeled by cardiologists without prior computerized interpretation, providing a

clearer assessment of the model's performance relative to human experts. Nevertheless, the

higher performance in labels assigned by more than one expert suggests that it likely performs at

or above the performance of the current computerized reads at the US health systems, especially

for the tested diagnoses.

CONCLUSIONS

We have developed and extensively validated a novel vision-text transformer capable of

generating complete diagnostic statements from ECG images in any lead layout and

configuration. Our approach represents a scalable and accessible strategy for generating accurate,

expert-level reports from photos of ECGs, enabling accurate ECG interpretation anywhere that

an ECG image, paper or digital, can be produced.

DATA AVAILABILITY

The development dataset from Yale and the validation datasets from Mount Sinai, Houston

Methodist, and Lake Regional Hospital are not publicly available, given the stipulations of the

relevant Institutional Review Boards. The external datasets from Brazil, UK Biobank, and PTB-

XL are available directly from the respective groups and are outside the purview of the authors.

CODE AVAILABILITY

The code for the study is available from the authors for review. Final code will be made available

for replication at the time of publication. However, due to intellectual property and IRB

stipulations at Yale, as well as patient privacy and data protection regulations, the model weights

will not be made publicly available. A completely functional version of the model is publicly

hosted at https://www.cards-lab.org/ecg-gpt.

ACKNOWLEDGEMENTS

Author contributions: RK conceived the study and accessed the data. AK, VS, and RK

developed the model. AK, VS, AV, and RK pursued the statistical analysis. AK, VS, and LSD

drafted the manuscript. All authors provided feedback regarding the study design and made

critical contributions to the manuscript writing. RK supervised the study, procured funding, and is

the guarantor.

Funding: This study was supported by research funding awarded to Dr. Khera by the Yale School

of Medicine and grant support from the National Institutes of Health (under awards

R01HL167858, R01AG089981, and K23HL153775) and the Doris Duke Charitable Foundation

(under award, 2022060). Dr. Oikonomou receives support from the National Heart, Lung, and

Blood Institute of the National Institutes of Health (under award 1F32HL170592). The funders

had no role in the design and conduct of the study; collection, management, analysis, and

interpretation of the data; preparation, review, or approval of the manuscript; and decision to

submit the manuscript for publication.

Competing Interests: Mr. Khunte, Mr. Sangha, and Dr. Khera are the coinventors of U.S.

Provisional Patent Application No. 63/428,569. Mr. Sangha and Dr. Khera are the coinventors of

U.S. Pending Patent Application No. 63/346,610, and are co-founders of Ensight-AI with Dr.

Krumholz. Dr. Khera is the coinventor of U.S. Provisional Patent Application No. 63/177,117

(unrelated to current work) and is a co-founder of Evidence2Health, a precision health platform

for evidence-based care. He is also an associate editor at JAMA, and received support from the

National Institutes of Health (under award R01HL167858, R01AG089981, and K23HL153775)

and the Doris Duke Charitable Foundation (under award, 2022060). He also receives research

support, through Yale, from Bristol-Myers Squibb, Novo Nordisk, and BridgeBio. Dr.

Oikonomou receives support from the National Heart, Lung, and Blood Institute of the National

Institutes of Health (under award 1F32HL170592). He is an academic co-founder of

Evidence2Health LLC, a co-inventor in patent applications (US17/720,068, 63/177,117,

63/580,137, 63/606,203, WO2018078395A1, WO2020058713A1) and has been an ad hoc

consultant for Caristo Diagnostics Ltd. Dr. Nadkarni is a founder of Renalytix, Pensieve, and

Verici and provides consultancy services to AstraZeneca, Reata, Renalytix, Siemens Healthineer,

and Variant Bio, and serves a scientific advisory board member for Renalytix and Pensieve. He

also has equity in Renalytix, Pensieve, and Verici. Dr. Krumholz works under contract with the

Centers for Medicare & Medicaid Services to support quality measurement programs, was a

recipient of a research grant from Johnson & Johnson, through Yale University, to support

clinical trial data sharing; was a recipient of a research agreement, through Yale University, from

the Shenzhen Center for Health Information for work to advance intelligent disease prevention

and health promotion; collaborates with the National Center for Cardiovascular Diseases in

Beijing; receives payment from the Arnold & Porter Law Firm for work related to the Sanofi

clopidogrel litigation, from the Martin Baughman Law Firm for work related to the Cook Celect

IVC filter litigation, and from the Siegfried and Jensen Law Firm for work related to Vioxx

litigation; chairs a Cardiac Scientific Advisory Board for UnitedHealth; was a member of the

IBM Watson Health Life Sciences Board; is a member of the Advisory Board for Element

Science, the Advisory Board for Facebook, and the Physician Advisory Board for Aetna; and is

the co-founder of Hugo Health, a personal health information platform, and co-founder of

Refactor Health, a healthcare AI-augmented data management company, and Ensight-AI, Inc.

Dr. Bhatt discloses the following relationships - Advisory Board: Angiowave, Bayer, Boehringer

Ingelheim, CellProthera, Cereno Scientific, Elsevier Practice Update Cardiology, High Enroll,

Janssen, Level Ex, McKinsey, Medscape Cardiology, Merck, MyoKardia, NirvaMed, Novo

Nordisk, PhaseBio, PLx Pharma, Stasys; Board of Directors: American Heart Association New

York City, Angiowave (stock options), Bristol Myers Squibb (stock), DRS.LINQ (stock

options), High Enroll (stock); Consultant: Broadview Ventures, GlaxoSmithKline, Hims, SFJ,

Youngene; Data Monitoring Committees: Acesion Pharma, Assistance Publique-Hôpitaux de

Paris, Baim Institute for Clinical Research (formerly Harvard Clinical Research Institute, for the

PORTICO trial, funded by St. Jude Medical, now Abbott), Boston Scientific (Chair, PEITHO

trial), Cleveland Clinic, Contego Medical (Chair, PERFORMANCE 2), Duke Clinical Research

Institute, Mayo Clinic, Mount Sinai School of Medicine (for the ENVISAGE trial, funded by

Daiichi Sankyo; for the ABILITY-DM trial, funded by Concept Medical; for ALLAY-HF,

funded by Alleviant Medical), Novartis, Population Health Research Institute; Rutgers

University (for the NIH-funded MINT Trial); Honoraria: American College of Cardiology

(Senior Associate Editor, Clinical Trials and News, ACC.org; Chair, ACC Accreditation

Oversight Committee), Arnold and Porter law firm (work related to Sanofi/Bristol-Myers Squibb

clopidogrel litigation), Baim Institute for Clinical Research (formerly Harvard Clinical Research

Institute; RE-DUAL PCI clinical trial steering committee funded by Boehringer Ingelheim;

AEGIS-II executive committee funded by CSL Behring), Belvoir Publications (Editor in Chief,

Harvard Heart Letter), Canadian Medical and Surgical Knowledge Translation Research Group

(clinical trial steering committees), CSL Behring (AHA lecture), Cowen and Company, Duke

Clinical Research Institute (clinical trial steering committees, including for the PRONOUNCE

trial, funded by Ferring Pharmaceuticals), HMP Global (Editor in Chief, Journal of Invasive

Cardiology), Journal of the American College of Cardiology (Guest Editor; Associate Editor),

K2P (Co-Chair, interdisciplinary curriculum), Level Ex, Medtelligence/ReachMD (CME

steering committees), MJH Life Sciences, Oakstone CME (Course Director, Comprehensive

Review of Interventional Cardiology), Piper Sandler, Population Health Research Institute (for

the COMPASS operations committee, publications committee, steering committee, and USA

national co-leader, funded by Bayer), WebMD (CME steering committees), Wiley (steering

committee); Other: Clinical Cardiology (Deputy Editor); Patent: Sotagliflozin (named on a

patent for sotagliflozin assigned to Brigham and Women's Hospital who assigned to Lexicon;

neither I nor Brigham and Women's Hospital receive any income from this patent); Research

Funding: Abbott, Acesion Pharma, Afimmune, Aker Biomarine, Alnylam, Amarin, Amgen,

AstraZeneca, Bayer, Beren, Boehringer Ingelheim, Boston Scientific, Bristol-Myers Squibb,

Cardax, CellProthera, Cereno Scientific, Chiesi, CinCor, Cleerly, CSL Behring, Eisai, Ethicon,

Faraday Pharmaceuticals, Ferring Pharmaceuticals, Forest Laboratories, Fractyl, Garmin, HLS

Therapeutics, Idorsia, Ironwood, Ischemix, Janssen, Javelin, Lexicon, Lilly, Medtronic, Merck,

Moderna, MyoKardia, NirvaMed, Novartis, Novo Nordisk, Otsuka, Owkin, Pfizer, PhaseBio,

PLx Pharma, Recardio, Regeneron, Reid Hoffman Foundation, Roche, Sanofi, Stasys, Synaptic,

The Medicines Company, Youngene, 89Bio; Royalties: Elsevier (Editor, Braunwald's Heart

Disease); Site Co-Investigator: Abbott, Biotronik, Boston Scientific, CSI, Endotronix, St. Jude

Medical (now Abbott), Philips, SpectraWAVE, Svelte, Vascular Solutions; Trustee: American

College of Cardiology; Unfunded Research: FlowCo. All other authors declare no relevant

competing interests.

REFERENCES

1. Schlant, R. C. *et al.* Guidelines for electrocardiography. A report of the American college of cardiology/American heart association task force on assessment of diagnostic and therapeutic cardiovascular procedures (committee on electrocardiography). *Circulation* **85**, 1221–1228 (1992).

2. Shah, A. P. & Rubin, S. A. Errors in the computerized electrocardiogram interpretation of cardiac rhythm. *J. Electrocardiol.* **40**, 385–390 (2007).

3. Guglin, M. E. & Thatai, D. Common errors in computer electrocardiogram interpretation. *Int. J. Cardiol.* **106**, 232–237 (2006).

4. Schläpfer, J. & Wellens, H. J. Computer-interpreted electrocardiograms: Benefits and limitations. *J. Am. Coll. Cardiol.* **70**, 1183–1192 (2017).

5. Garvey, J. L., Zegre-Hemsey, J., Gregg, R. & Studnek, J. R. Electrocardiographic diagnosis of ST segment elevation myocardial infarction: An evaluation of three automated interpretation algorithms. *J. Electrocardiol.* **49**, 728–732 (2016).

6. Estes, N. A. M., 3rd. Computerized interpretation of ECGs: supplement not a substitute. *Circulation. Arrhythmia and electrophysiology* vol. 6 2–4 (2013).

7. Bogun, F. *et al.* Misdiagnosis of atrial fibrillation and its clinical consequences. *Am. J. Med.* **117**, 636–642 (2004).

8. Jones, C., Park, T. & United Economic Research Service (Ers). *Health Status and Health Care Access of Farm and Rural Populations*. (Bibliogov, 2012).

9. Aneja, S. *et al.* US cardiologist workforce from 1995 to 2007: modest growth, lasting geographic maldistribution especially in rural areas. *Health Aff. (Millwood)* **30**, 2301–2309 (2011).

10. Soriano Marcolino, M. *et al.* The experience of a sustainable large scale Brazilian Telehealth Network. *Telemed. J. E. Health.* **22**, 899–908 (2016).

11. Cook, D. A., Oh, S.-Y. & Pusic, M. V. Accuracy of physicians' electrocardiogram interpretations. *JAMA Intern. Med.* **180**, 1461 (2020).

12. Ribeiro, A. H. *et al.* Automatic diagnosis of the 12-lead ECG using a deep neural network. *Nat. Commun.* **11**, 1760 (2020).

13. Hannun, A. Y. *et al.* Cardiologist-level arrhythmia detection and classification in ambulatory electrocardiograms using a deep neural network. *Nat. Med.* **25**, 65–69 (2019).

14. Hughes, J. W. *et al.* Performance of a convolutional neural network and explainability technique for 12-lead electrocardiogram interpretation. *JAMA Cardiol.* **6**, 1285–1295 (2021).

15. Sangha, V. *et al.* Automated multilabel diagnosis on electrocardiographic images and signals. *Nat. Commun.* **13**, 1583 (2022).

16. Sangha, V. *et al.* Detection of left ventricular systolic dysfunction from electrocardiographic images. *Circulation* (2023) doi:10.1161/CIRCULATIONAHA.122.062646.

17. Ecg-plot. *PyPI* https://pypi.org/project/ecg-plot/ (2021).

18. Pyenchant. *PyPI* https://pypi.org/project/pyenchant/ (2021).

19. Wolf, T. *et al.* Transformers: State-of-the-Art Natural Language Processing. in *Proceedings of the 2020 Conference on Empirical Methods in Natural Language Processing: System Demonstrations* 38–45 (Association for Computational Linguistics, Online, 2020).

20. Deng, J. *et al.* ImageNet: A large-scale hierarchical image database. in *2009 IEEE Conference on Computer Vision and Pattern Recognition* 248–255 (2009).

21. Bao, H., Dong, L., Piao, S. & Wei, F. BEiT: BERT Pre-Training of Image Transformers.
arXiv [cs.CV] (2021).
22. Loshchilov, I. & Hutter, F. Decoupled Weight Decay Regularization. *arXiv [cs.LG]* (2017).
23. Ribeiro, A. L. P. *et al.* Tele-electrocardiography and bigdata: The CODE (Clinical
Outcomes in Digital Electrocardiography) study. *J. Electrocardiol.* **57S**, S75–S78 (2019).
24. Kligfield, P. *et al.* Recommendations for the standardization and interpretation of the
electrocardiogram: part I: The electrocardiogram and its technology: a scientific statement
from the American Heart Association Electrocardiography and Arrhythmias Committee,
Council on Clinical Cardiology; the American College of Cardiology Foundation; and the
Heart Rhythm Society: endorsed by the International Society for Computerized
Electrocardiology. *Circulation* **115**, 1306–1324 (2007).
25. Wagner, P. *et al.* PTB-XL, a large publicly available electrocardiography dataset. *Scientific*
Data vol. 7 Preprint at https://doi.org/10.1038/s41597-020-0495-6 (2020).
26. Devlin, J., Chang, M.-W., Lee, K. & Toutanova, K. BERT: Pre-training of Deep
Bidirectional Transformers for Language Understanding. *arXiv [cs.CL]* (2018).
27. Sanh, V., Debut, L., Chaumond, J. & Wolf, T. DistilBERT, a distilled version of BERT:
smaller, faster, cheaper and lighter. *arXiv [cs.CL]* (2019).
28. Wang, G. *et al.* Optimized glycemic control of type 2 diabetes with reinforcement learning:
a proof-of-concept trial. *Nat. Med.* **29**, 2633–2642 (2023).
29. Alsentzer, E. *et al.* Publicly Available Clinical BERT Embeddings. *arXiv [cs.CL]* (2019).
30. Lin, C.-Y. ROUGE: A Package for Automatic Evaluation of Summaries. in *Text*
Summarization Branches Out 74–81 (Association for Computational Linguistics, Barcelona,
Spain, 2004).
31. Papineni, K., Roukos, S., Ward, T. & Zhu, W.-J. BLEU: A method for automatic evaluation
of machine translation. https://aclanthology.org/P02-1040.pdf.
32. Banerjee, S. & Lavie, A. METEOR: An Automatic Metric for MT Evaluation with
Improved Correlation with Human Judgments. in *Proceedings of the ACL Workshop on*
Intrinsic and Extrinsic Evaluation Measures for Machine Translation and/or
Summarization 65–72 (Association for Computational Linguistics, Ann Arbor, Michigan,
2005).
33. Vedantam, R., Zitnick, C. L. & Parikh, D. CIDEr: Consensus-based image description
evaluation. *arXiv [cs.CV]* 4566–4575 (2014).
34. von Werra, L. *et al.* Evaluate & evaluation on the Hub: Better best practices for data and
model measurements. *arXiv [cs.LG]* (2022).
35. Chen, X. *et al.* Microsoft COCO captions: Data collection and evaluation server. *arXiv*
[cs.CV] (2015).
36. Sun, X. & Xu, W. Fast Implementation of DeLong's Algorithm for Comparing the Areas
Under Correlated Receiver Operating Characteristic Curves. *IEEE Signal Process. Lett.* **21**,
1389–1393 (2014).
37. DeLong, E. R., DeLong, D. M. & Clarke-Pearson, D. L. Comparing the areas under two or
more correlated receiver operating characteristic curves: a nonparametric approach.
Biometrics **44**, 837–845 (1988).
38. Efron, B. Bootstrap methods: Another look at the jackknife. *Ann. Stat.* **7**, 1–26 (1979).
39. Selivanov, A. *et al.* Medical image captioning via generative pretrained transformers. *Sci.*
Rep. **13**, 4171 (2023).

40. Ayesha, H. *et al.* Automatic medical image interpretation: State of the art and future
directions. *Pattern Recognit.* **114**, 107856 (2021).
41. Bartels, M. G. G. *et al.* Learning to Automatically Generate Accurate ECG Captions. in
Proceedings of The 5th International Conference on Medical Imaging with Deep Learning
(eds. Konukoglu, E. et al.) vol. 172 86–102 (PMLR, 06--08 Jul 2022).
42. Kashou, A. H. *et al.* A comprehensive artificial intelligence-enabled electrocardiogram
interpretation program. *Cardiovasc Digit Health J* **1**, 62–70 (2020).
43. Siontis, K. C., Noseworthy, P. A., Attia, Z. I. & Friedman, P. A. Artificial intelligence-
enhanced electrocardiography in cardiovascular disease management. *Nat. Rev. Cardiol.* **18**,
465–478 (2021).
44. ECG-GPT. *CarDS Lab* cards-lab.org/ecg-gpt (2024).
755

756 **Figure 1. Overview of ECG-GPT's development and evaluation.**
757 Abbreviations: ECG, electrocardiogram; BEiT, Bidirectional Encoder representation from Image
758 Transformers; GPT-2, Generative Pretrained Transformer-2.

Figure 2. Vision Encoder-Decoder Model Architecture and Sample ECG-GPT Output Interpretation Statement. Abbreviations: ECG, electrocardiogram; BEiT, Bidirectional Encoder representation from Image Transformers; GPT-2, Generative Pretrained Transformer-2.

Table 1. Pairwise and baseline similarity between reference and model-generated ECG interpretation statements in the held-out test set. Abbreviations: AF, atrial fibrillation; ST, sinus tachycardia; SB, sinus bradycardia; LBBB, left bundle branch block; RBBB, right bundle branch block; LAFB, left anterior fascicular block; LPFB, left posterior fascicular block; SVT, supraventricular tachycardia; PAC, premature atrial complexes; PVC, premature ventricular complexes; LAE, left atrial enlargement; RAE, right atrial enlargement; LVH, left ventricular hypertrophy; RVH, right ventricular hypertrophy; MI, myocardial infarction; AVb, atrioventricular block; WPW, Wolff-Parkinson-White syndrome; Repol, repolarization. *Acute MI includes ST Elevation MI (STEMI).

Labels	Pairwise Similarity	Baseline Similarity	P-Value
Overall	0.896 (0.827-0.970)	0.727 (0.671-0.783)	<0.001
Sinus Rhythm	0.884 (0.815-0.968)	0.746 (0.691-0.800)	<0.001
AF	0.914 (0.859-0.968)	0.794 (0.748-0.839)	<0.001
Atrial Flutter	0.909 (0.844-0.961)	0.782 (0.740-0.839)	<0.001
ST	0.913 (0.849-0.975)	0.788 (0.739-0.837)	<0.001
SB	0.945 (0.871-1.000)	0.807 (0.755-0.860)	<0.001
Sinus Arrythmia	0.969 (0.911-1.000)	0.819 (0.743-0.884)	<0.001
LBBB	0.929 (0.846-1.000)	0.824 (0.768-0.880)	<0.001
RBBB	0.939 (0.885-0.994)	0.826 (0.778-0.874)	<0.001
LAFB	0.931 (0.875-0.975)	0.803 (0.756-0.850)	<0.001
LPFB	0.949 (0.897-0.989)	0.842 (0.787-0.891)	<0.001
SVT	0.899 (0.828-0.955)	0.778 (0.726-0.828)	<0.001
PAC	0.908 (0.848-0.956)	0.769 (0.715-0.821)	<0.001
PVC	0.927 (0.874-0.969)	0.784 (0.736-0.830)	<0.001
LAE	0.914 (0.841-0.968)	0.766 (0.713-0.822)	<0.001
RAE	0.920 (0.866-0.964)	0.770 (0.718-0.823)	<0.001
LVH	0.935 (0.879-0.979)	0.799 (0.745-0.853)	<0.001
RVH	0.954 (0.882-0.987)	0.794 (0.736-0.858)	<0.001
Low Voltage	0.914 (0.845-0.968)	0.756 (0.698-0.816)	<0.001
Left Axis Deviation	0.913 (0.851-0.969)	0.760 (0.712-0.810)	<0.001
Acute MI	0.902 (0.840-0.958)	0.764 (0.717-0.819)	<0.001
Lead Reversal	0.923 (0.844-0.969)	0.816 (0.732-0.889)	<0.001
1st degree AVb	0.913 (0.849-0.967)	0.787 (0.736-0.840)	<0.001
2nd degree AVb	0.898 (0.837-0.948)	0.815 (0.759-0.865)	<0.001
3rd degree AVb	0.870 (0.801-0.929)	0.753 (0.690-0.807)	<0.001
WPW	0.837 (0.719-0.935)	0.724 (0.670-0.838)	<0.001
Repol Abnormality	0.922 (0.858-0.971)	0.780 (0.726-0.840)	<0.001

Table 2. Clinical assessment of model-generated ECG interpretation statements in the held-
out test set. Abbreviations: PPV, Positive Predictive Value; NPV, Negative Predictive Value;
AUROC, area under the receiver operator characteristic; AUPRC, area under precision-recall
curve; AF, atrial fibrillation; ST, sinus tachycardia; SB, sinus bradycardia; LBBB, left bundle
branch block; RBBB, right bundle branch block; LAFB, left anterior fascicular block; LPFB, left
posterior fascicular block; SVT, supraventricular tachycardia; PAC, premature atrial complexes;
PVC, premature ventricular complexes; LAE, left atrial enlargement; RAE, right atrial
enlargement; LVH, left ventricular hypertrophy; RVH, right ventricular hypertrophy; MI,
myocardial infarction; AVb, atrioventricular block; WPW, Wolff-Parkinson-White syndrome;
Repol, repolarization. *Acute MI includes ST Elevation MI (STEMI).

Labels	Accuracy	PPV	NPV	Specificity	Sensitivity	AUROC	AUPRC	F1
Sinus Rhythm	0.950	0.971	0.913	0.944	0.954	0.949 (0.948-0.95)	0.956 (0.955-0.957)	0.962
AF	0.970	0.853	0.984	0.983	0.862	0.922 (0.92-0.925)	0.75 (0.745-0.755)	0.858
Atrial Flutter	0.968	0.381	0.997	0.970	0.868	0.919 (0.914-0.923)	0.333 (0.324-0.342)	0.529
ST	0.988	0.936	0.992	0.994	0.912	0.953 (0.951-0.955)	0.86 (0.856-0.865)	0.924
SB	0.988	0.902	0.992	0.995	0.867	0.931 (0.928-0.934)	0.79 (0.783-0.797)	0.885
Sinus Arrythmia	0.963	0.395	0.998	0.964	0.922	0.943 (0.94-0.947)	0.366 (0.358-0.373)	0.553
LBBB	0.990	0.853	0.995	0.995	0.855	0.925 (0.921-0.929)	0.735 (0.725-0.744)	0.854
RBBB	0.989	0.909	0.995	0.993	0.935	0.964 (0.962-0.965)	0.854 (0.849-0.859)	0.922
LAFB	0.971	0.701	0.989	0.980	0.808	0.894 (0.891-0.898)	0.577 (0.568-0.583)	0.751
LPFB	0.987	0.271	0.999	0.988	0.861	0.924 (0.915-0.933)	0.234 (0.222-0.248)	0.412
SVT	0.994	0.399	0.999	0.996	0.698	0.847 (0.833-0.861)	0.28 (0.259-0.303)	0.508
PAC	0.967	0.771	0.976	0.989	0.616	0.802 (0.798-0.806)	0.498 (0.49-0.506)	0.685
PVC	0.973	0.799	0.988	0.983	0.848	0.915 (0.913-0.918)	0.688 (0.681-0.694)	0.823
LAE	0.927	0.679	0.961	0.956	0.706	0.831 (0.828-0.834)	0.513 (0.508-0.519)	0.692
RAE	0.989	0.251	0.999	0.990	0.764	0.877 (0.864-0.889)	0.192 (0.178-0.207)	0.377
LVH	0.949	0.729	0.983	0.960	0.864	0.912 (0.910-0.914)	0.645 (0.641-0.651)	0.791
RVH	0.927	0.118	0.998	0.928	0.822	0.875 (0.868-0.882)	0.099 (0.095-0.104)	0.206
Low Voltage	0.945	0.660	0.977	0.963	0.762	0.862 (0.86-0.865)	0.524 (0.518-0.531)	0.708
Left Axis Deviation	0.940	0.649	0.975	0.959	0.753	0.856 (0.853-0.859)	0.511 (0.504-0.518)	0.697
Acute MI*	0.985	0.255	0.998	0.986	0.711	0.849 (0.838-0.86)	0.183 (0.171-0.194)	0.375
Lead Reversal	0.997	0.325	1.000	0.997	0.718	0.858 (0.836-0.879)	0.234 (0.206-0.261)	0.447
1st degree AVb	0.964	0.743	0.983	0.978	0.787	0.883 (0.88-0.886)	0.601 (0.594-0.608)	0.764
2nd degree AVb	0.998	0.379	0.999	0.999	0.643	0.821 (0.796-0.846)	0.244 (0.207-0.283)	0.477
3rd degree AVb	0.998	0.260	1.000	0.998	0.693	0.846 (0.816-0.875)	0.181 (0.149-0.21)	0.378
WPW	0.999	0.444	1.000	1.000	0.624	0.812 (0.771-0.853)	0.277 (0.218-0.353)	0.519
Repol Abnormality	0.930	0.515	0.973	0.951	0.661	0.806 (0.802-0.809)	0.365 (0.359-0.37)	0.579

Table 3. Pairwise and baseline similarity between reference and model-generated
statements for plotted ECG signals in the Mount Sinai Health System. Abbreviations: AF,
atrial fibrillation; ST, sinus tachycardia; SB, sinus bradycardia; LBBB, left bundle branch block;
RBBB, right bundle branch block; LAFB, left anterior fascicular block; LPFB, left posterior
fascicular block; SVT, supraventricular tachycardia; PAC, premature atrial complexes; PVC,
premature ventricular complexes; LAE, left atrial enlargement; RAE, right atrial enlargement;
LVH, left ventricular hypertrophy; RVH, right ventricular hypertrophy; MI, myocardial
infarction; AVb, atrioventricular block; WPW, Wolff-Parkinson-White syndrome; Repol,
repolarization. *Acute MI includes ST Elevation MI (STEMI).
796

Labels	Pairwise Similarity	Baseline Similarity	P-Value
Overall	0.861 (0.786-0.961)	0.740 (0.687-0.795)	<0.001
Sinus Rhythm	0.853 (0.777-0.964)	0.751 (0.696-0.806)	<0.001
AF	0.880 (0.811-0.946)	0.807 (0.764-0.852)	<0.001
Atrial Flutter	0.871 (0.807-0.933)	0.811 (0.762-0.854)	<0.001
ST	0.857 (0.790-0.938)	0.780 (0.733-0.830)	<0.001
SB	0.904 (0.814-1.000)	0.794 (0.738-0.845)	<0.001
Sinus Arrythmia	0.918 (0.794-1.000)	0.829 (0.744-0.890)	<0.001
LBBB	0.945 (0.854-1.000)	0.853 (0.798-0.897)	<0.001
RBBB	0.897 (0.839-0.952)	0.812 (0.770-0.853)	<0.001
LAFB	0.900 (0.838-0.948)	0.808 (0.766-0.848)	<0.001
LPFB	0.877 (0.811-0.932)	0.793 (0.745-0.839)	<0.001
SVT	0.831 (0.763-0.900)	0.783 (0.723-0.835)	<0.001
PAC	0.858 (0.799-0.922)	0.785 (0.740-0.831)	<0.001
PVC	0.893 (0.816-0.958)	0.807 (0.758-0.855)	<0.001
LAE	0.916 (0.816-0.984)	0.815 (0.752-0.866)	<0.001
RAE	0.884 (0.808-0.963)	0.795 (0.743-0.841)	<0.001
LVH	0.922 (0.819-0.978)	0.825 (0.770-0.877)	<0.001
RVH	0.832 (0.774-0.904)	0.778 (0.732-0.827)	<0.001
Low Voltage	0.840 (0.761-0.927)	0.764 (0.707-0.824)	<0.001
Left Axis Deviation	0.891 (0.810-0.957)	0.785 (0.737-0.829)	<0.001
Acute MI	0.796 (0.730-0.926)	0.760 (0.706-0.880)	<0.001
Lead Reversal	0.847 (0.736-0.902)	0.789 (0.686-0.866)	<0.001
1st degree AVb	0.907 (0.817-0.968)	0.819 (0.760-0.867)	<0.001
2nd degree AVb	0.871 (0.797-0.941)	0.812 (0.751-0.870)	<0.001
3rd degree AVb	0.817 (0.743-0.880)	0.767 (0.712-0.820)	<0.001
WPW	0.748 (0.671-0.942)	0.707 (0.660-0.808)	<0.001
Repol Abnormality	0.875 (0.810-0.954)	0.793 (0.719-0.856)	<0.001

Table 4. Pairwise and baseline similarity between reference and model-generated statements for prospectively acquired PDFs of ECGs in the Mount Sinai Health System. Abbreviations: AF, atrial fibrillation; ST, sinus tachycardia; SB, sinus bradycardia; LBBB, left bundle branch block; RBBB, right bundle branch block; LAFB, left anterior fascicular block; LPFB, left posterior fascicular block; SVT, supraventricular tachycardia; PAC, premature atrial complexes; PVC, premature ventricular complexes; LAE, left atrial enlargement; RAE, right atrial enlargement; LVH, left ventricular hypertrophy; RVH, right ventricular hypertrophy; MI, myocardial infarction; AVb, atrioventricular block; WPW, Wolff-Parkinson-White syndrome; Repol, repolarization. *Acute MI includes ST Elevation MI (STEMI).

Labels	Pairwise Similarity	Baseline Similarity	P-Value
Overall	0.796 (0.720-0.863)	0.728 (0.676-0.781)	<0.001
Sinus Rhythm	0.789 (0.702-0.857)	0.737 (0.686-0.791)	<0.001
AF	0.860 (0.799-0.908)	0.804 (0.766-0.846)	<0.001
Atrial Flutter	0.838 (0.780-0.900)	0.804 (0.755-0.849)	<0.001
ST	0.778 (0.734-0.833)	0.746 (0.707-0.792)	<0.001
SB	0.819 (0.751-0.882)	0.775 (0.727-0.821)	<0.001
Sinus Arrythmia	0.840 (0.794-0.908)	0.815 (0.766-0.856)	<0.001
LBBB	0.846 (0.804-0.928)	0.808 (0.754-0.855)	<0.001
RBBB	0.854 (0.807-0.900)	0.794 (0.754-0.834)	<0.001
LAFB	0.851 (0.807-0.896)	0.801 (0.760-0.839)	<0.001
LPFB	0.843 (0.781-0.903)	0.774 (0.717-0.829)	<0.001
SVT	0.835 (0.758-0.867)	0.798 (0.749-0.842)	0.008
PAC	0.838 (0.784-0.882)	0.792 (0.749-0.834)	<0.001
PVC	0.868 (0.811-0.919)	0.811 (0.762-0.853)	<0.001
LAE	0.825 (0.773-0.875)	0.788 (0.739-0.830)	<0.001
RAE	0.786 (0.737-0.854)	0.752 (0.706-0.799)	<0.001
LVH	0.847 (0.797-0.894)	0.801 (0.754-0.846)	<0.001
RVH	0.821 (0.774-0.872)	0.757 (0.719-0.805)	<0.001
Low Voltage	0.820 (0.746-0.876)	0.769 (0.716-0.818)	<0.001
Left Axis Deviation	0.842 (0.792-0.893)	0.778 (0.736-0.818)	<0.001
Acute MI	0.834 (0.769-0.954)	0.822 (0.764-0.940)	0.438
Lead Reversal	0.846 (0.751-0.904)	0.814 (0.732-0.875)	0.029
1st degree AVb	0.853 (0.794-0.902)	0.813 (0.764-0.854)	<0.001
2nd degree AVb	0.882 (0.822-0.909)	0.826 (0.787-0.869)	0.027
3rd degree AVb	0.844 (0.788-0.890)	0.792 (0.754-0.853)	0.098
WPW	0.658 (0.656-0.685)	0.656 (0.639-0.683)	0.144
Repol Abnormality	0.828 (0.773-0.878)	0.776 (0.719-0.829)	<0.001

Table 5. Clinical assessment of model-generated ECG interpretation statements on plotted
ECG signals in the Mount Sinai Health System. Abbreviations: PPV, Positive Predictive
Value; NPV, Negative Predictive Value; AUROC, area under the receiver operator
characteristic; AUPRC, area under precision-recall curve; AF, atrial fibrillation; ST, sinus
tachycardia; SB, sinus bradycardia; LBBB, left bundle branch block; RBBB, right bundle branch
block; LAFB, left anterior fascicular block; LPFB, left posterior fascicular block; SVT,
supraventricular tachycardia; PAC, premature atrial complexes; PVC, premature ventricular
complexes; LAE, left atrial enlargement; RAE, right atrial enlargement; LVH, left ventricular
hypertrophy; RVH, right ventricular hypertrophy; MI, myocardial infarction; AVb,
atrioventricular block; WPW, Wolff-Parkinson-White syndrome; Repol, repolarization. *Acute
MI includes ST Elevation MI (STEMI).

Labels	Accuracy	PPV	NPV	Specificity	Sensitivity	AUROC	AUPRC	F1
Sinus Rhythm	0.931	0.942	0.908	0.883	0.955	0.919 (0.919-0.92)	0.93 (0.929-0.93)	0.949
AF	0.978	0.89	0.984	0.992	0.793	0.893 (0.891-0.894)	0.721 (0.718-0.724)	0.839
Atrial Flutter	0.975	0.372	0.997	0.978	0.838	0.908 (0.905-0.91)	0.314 (0.31-0.318)	0.515
ST	0.979	0.91	0.988	0.989	0.898	0.944 (0.943-0.944)	0.828 (0.827-0.83)	0.904
SB	0.965	0.948	0.967	0.995	0.734	0.865 (0.863-0.866)	0.726 (0.724-0.728)	0.828
Sinus Arrythmia	0.941	0.412	0.987	0.951	0.728	0.84 (0.838-0.841)	0.312 (0.309-0.315)	0.526
LBBB	0.99	0.746	0.998	0.992	0.914	0.953 (0.952-0.954)	0.684 (0.679-0.688)	0.821
RBBB	0.987	0.876	0.996	0.99	0.94	0.965 (0.965-0.966)	0.827 (0.825-0.829)	0.907
LAFB	0.969	0.558	0.991	0.977	0.776	0.877 (0.875-0.878)	0.441 (0.437-0.445)	0.649
LPFB	0.985	0.179	0.999	0.986	0.753	0.869 (0.864-0.875)	0.136 (0.131-0.141)	0.289
SVT	0.993	0.279	0.999	0.995	0.644	0.819 (0.812-0.826)	0.181 (0.172-0.188)	0.389
PAC	0.961	0.763	0.968	0.991	0.479	0.735 (0.733-0.737)	0.396 (0.392-0.399)	0.589
PVC	0.972	0.749	0.988	0.981	0.826	0.904 (0.902-0.905)	0.629 (0.626-0.633)	0.785
LAE	0.923	0.564	0.986	0.927	0.877	0.902 (0.901-0.903)	0.506 (0.504-0.508)	0.686
RAE	0.985	0.284	0.999	0.986	0.839	0.913 (0.909-0.916)	0.239 (0.234-0.245)	0.424
LVH	0.951	0.749	0.988	0.955	0.922	0.938 (0.938-0.939)	0.701 (0.699-0.703)	0.827
RVH	0.937	0.084	0.996	0.94	0.607	0.774 (0.769-0.778)	0.054 (0.053-0.056)	0.147
Low Voltage	0.957	0.592	0.988	0.966	0.813	0.89 (0.888-0.891)	0.493 (0.49-0.495)	0.686
Left Axis Deviation	0.934	0.639	0.967	0.96	0.687	0.824 (0.822-0.825)	0.469 (0.466-0.471)	0.663
Acute MI*	0.974	0.22	0.997	0.977	0.672	0.825 (0.821-0.829)	0.151 (0.146-0.154)	0.331
Lead Reversal	0.993	0.175	0.999	0.994	0.512	0.753 (0.745-0.761)	0.091 (0.085-0.097)	0.261
1st degree AVb	0.97	0.699	0.99	0.977	0.845	0.911 (0.91-0.912)	0.599 (0.596-0.603)	0.765
2nd degree AVb	0.998	0.421	0.999	0.998	0.563	0.781 (0.772-0.79)	0.238 (0.22-0.253)	0.482
3rd degree AVb	0.998	0.277	1	0.999	0.655	0.827 (0.813-0.84)	0.182 (0.167-0.195)	0.389
WPW	0.999	0.533	1	1	0.608	0.804 (0.789-0.819)	0.325 (0.302-0.354)	0.568
Repol Abnormality	0.927	0.319	0.985	0.938	0.667	0.803 (0.801-0.805)	0.227 (0.225-0.229)	0.432

 Table 6. Clinical assessment of model-generated ECG interpretation statements on
prospectively acquired PDFs of ECGs in the Mount Sinai Health System. Abbreviations:
PPV, Positive Predictive Value; NPV, Negative Predictive Value; AUROC, area under the
receiver operator characteristic; AUPRC, area under precision-recall curve; AF, atrial
fibrillation; ST, sinus tachycardia; SB, sinus bradycardia; LBBB, left bundle branch block;
RBBB, right bundle branch block; LAFB, left anterior fascicular block; LPFB, left posterior
fascicular block; SVT, supraventricular tachycardia; PAC, premature atrial complexes; PVC,
premature ventricular complexes; LAE, left atrial enlargement; RAE, right atrial enlargement;
LVH, left ventricular hypertrophy; RVH, right ventricular hypertrophy; MI, myocardial
infarction; AVb, atrioventricular block; WPW, Wolff-Parkinson-White syndrome; Repol,
repolarization. *Acute MI includes ST Elevation MI (STEMI)

Labels	Accuracy	PPV	NPV	Specificity	Sensitivity	AUROC	AUPRC	F1
Sinus Rhythm	0.92	0.946	0.879	0.909	0.927	0.918 (0.912-0.924)	0.923 (0.917-0.929)	0.936
AF	0.96	0.693	0.986	0.971	0.821	0.896 (0.882-0.91)	0.582 (0.55-0.615)	0.752
Atrial Flutter	0.958	0.207	0.996	0.961	0.735	0.848 (0.811-0.886)	0.156 (0.126-0.189)	0.323
ST	0.959	0.797	0.978	0.977	0.804	0.89 (0.878-0.903)	0.661 (0.634-0.685)	0.8
SB	0.97	0.895	0.978	0.988	0.822	0.905 (0.894-0.916)	0.755 (0.732-0.781)	0.857
Sinus Arrythmia	0.909	0.325	0.993	0.911	0.865	0.888 (0.872-0.904)	0.288 (0.265-0.315)	0.473
LBBB	0.979	0.559	0.997	0.981	0.898	0.939 (0.921-0.958)	0.504 (0.45-0.555)	0.689
RBBB	0.961	0.645	0.998	0.96	0.973	0.966 (0.96-0.973)	0.629 (0.597-0.655)	0.776
LAFB	0.956	0.415	0.979	0.976	0.453	0.714 (0.689-0.74)	0.208 (0.176-0.246)	0.433
LPFB	0.977	0.084	0.999	0.978	0.625	0.802 (0.716-0.887)	0.054 (0.023-0.086)	0.148
SVT	0.985	0.206	0.999	0.986	0.841	0.913 (0.859-0.968)	0.174 (0.117-0.227)	0.33
PAC	0.95	0.738	0.957	0.992	0.347	0.669 (0.651-0.688)	0.298 (0.264-0.336)	0.472
PVC	0.963	0.694	0.986	0.975	0.801	0.888 (0.873-0.903)	0.57 (0.539-0.605)	0.744
LAE	0.9	0.436	0.967	0.922	0.654	0.788 (0.772-0.804)	0.314 (0.288-0.339)	0.523
RAE	0.988	0.235	0.996	0.992	0.369	0.681 (0.622-0.74)	0.091 (0.042-0.145)	0.287
LVH	0.813	0.377	0.927	0.851	0.572	0.712 (0.698-0.725)	0.274 (0.256-0.292)	0.455
RVH	0.925	0.023	0.997	0.928	0.378	0.653 (0.581-0.724)	0.011 (0.006-0.019)	0.043
Low Voltage	0.948	0.518	0.978	0.967	0.617	0.792 (0.771-0.812)	0.34 (0.304-0.379)	0.563
Left Axis Deviation	0.917	0.605	0.949	0.959	0.551	0.755 (0.739-0.77)	0.379 (0.351-0.408)	0.577
Acute MI*	0.89	0.04	0.998	0.891	0.75	0.82 (0.765-0.876)	0.031 (0.021-0.041)	0.075
Lead Reversal	0.96	0.052	0.998	0.962	0.525	0.743 (0.665-0.822)	0.029 (0.015-0.049)	0.095
1st degree AVb	0.942	0.55	0.988	0.949	0.843	0.896 (0.882-0.91)	0.475 (0.441-0.508)	0.666
2nd degree AVb	0.996	0.225	0.999	0.997	0.45	0.723 (0.612-0.835)	0.102 (0.032-0.196)	0.3
3rd degree AVb	0.996	0.216	0.999	0.997	0.471	0.734 (0.612-0.856)	0.103 (0.028-0.211)	0.296
WPW	0.999	0.333	0.999	1	0.143	0.571 (0.431-0.711)	0.048 (0-0.334)	0.2
Repol Abnormality	0.908	0.198	0.98	0.924	0.497	0.711 (0.685-0.736)	0.117 (0.101-0.14)	0.283

Table 7. Pairwise and baseline similarity between reference and model-generated statements in the Houston Methodist Hospital. Abbreviations: AF, atrial fibrillation; ST, sinus tachycardia; SB, sinus bradycardia; LBBB, left bundle branch block; RBBB, right bundle branch block; LAFB, left anterior fascicular block; LPFB, left posterior fascicular block; SVT, supraventricular tachycardia; PAC, premature atrial complexes; PVC, premature ventricular complexes; LAE, left atrial enlargement; RAE, right atrial enlargement; LVH, left ventricular hypertrophy; RVH, right ventricular hypertrophy; MI, myocardial infarction; AVb, atrioventricular block; WPW, Wolff-Parkinson-White syndrome; Repol, repolarization. *Acute MI includes ST Elevation MI (STEMI).

Labels	Pairwise Similarity	Baseline Similarity	P-Value
Overall	0.814 (0.753-0.882)	0.772 (0.733-0.817)	<0.001
Sinus Rhythm	0.797 (0.747-0.873)	0.772 (0.735-0.818)	<0.001
AF	0.884 (0.840-0.922)	0.855 (0.817-0.889)	<0.001
Atrial Flutter	0.878 (0.832-0.919)	0.860 (0.814-0.896)	<0.001
ST	0.856 (0.808-0.902)	0.826 (0.790-0.862)	<0.001
SB	0.828 (0.774-0.888)	0.800 (0.758-0.842)	<0.001
Sinus Arrythmia	0.892 (0.854-0.939)	0.866 (0.832-0.899)	<0.001
LBBB	0.902 (0.865-0.982)	0.865 (0.824-0.910)	<0.001
RBBB	0.903 (0.865-0.948)	0.853 (0.815-0.886)	<0.001
LAFB	0.895 (0.862-0.930)	0.851 (0.813-0.882)	<0.001
LPFB	0.867 (0.830-0.903)	0.823 (0.784-0.860)	<0.001
SVT	0.880 (0.828-0.921)	0.851 (0.813-0.886)	<0.001
PAC	0.871 (0.825-0.915)	0.836 (0.796-0.873)	<0.001
PVC	0.909 (0.857-0.952)	0.862 (0.820-0.898)	<0.001
LAE	0.879 (0.838-0.923)	0.859 (0.828-0.889)	<0.001
RAE	0.869 (0.831-0.921)	0.854 (0.823-0.885)	<0.001
LVH	0.913 (0.885-0.943)	0.880 (0.850-0.906)	<0.001
RVH	0.897 (0.854-0.934)	0.854 (0.817-0.890)	<0.001
Low Voltage	0.838 (0.797-0.885)	0.814 (0.782-0.850)	<0.001
Left Axis Deviation	0.890 (0.848-0.926)	0.841 (0.808-0.872)	<0.001
Acute MI	0.880 (0.840-0.909)	0.865 (0.832-0.895)	0.007
Lead Reversal	0.909 (0.863-0.935)	0.880 (0.827-0.910)	<0.001
1st degree AVb	0.895 (0.839-0.939)	0.864 (0.817-0.901)	<0.001
2nd degree AVb	0.902 (0.850-0.953)	0.890 (0.852-0.924)	0.041
3rd degree AVb	0.829 (0.779-0.883)	0.823 (0.780-0.872)	0.547
WPW	0.777 (0.755-0.871)	0.786 (0.755-0.866)	0.94
Repol Abnormality	0.878 (0.828-0.931)	0.833 (0.789-0.878)	<0.001

Table 8. Clinical assessment of model-generated ECG interpretation statements on PDFs of
ECGs in the Houston Methodist Hospital. Abbreviations: PPV, Positive Predictive Value;
NPV, Negative Predictive Value; AUROC, area under the receiver operator characteristic;
AUPRC, area under precision-recall curve; AF, atrial fibrillation; ST, sinus tachycardia; SB,
sinus bradycardia; LBBB, left bundle branch block; RBBB, right bundle branch block; LAFB,
left anterior fascicular block; LPFB, left posterior fascicular block; SVT, supraventricular
tachycardia; PAC, premature atrial complexes; PVC, premature ventricular complexes; LAE, left
atrial enlargement; RAE, right atrial enlargement; LVH, left ventricular hypertrophy; RVH, right
ventricular hypertrophy; MI, myocardial infarction; AVb, atrioventricular block; WPW, Wolff-
Parkinson-White syndrome; Repol, repolarization. *Acute MI includes ST Elevation MI
(STEMI).

Labels	Accuracy	PPV	NPV	Specificity	Sensitivity	AUROC	AUPRC	F1
Sinus Rhythm	0.888	0.89	0.881	0.701	0.962	0.832 (0.828-0.835)	0.883 (0.881-0.886)	0.925
AF	0.989	0.801	0.995	0.994	0.837	0.916 (0.907-0.924)	0.675 (0.652-0.698)	0.819
Atrial Flutter	0.989	0.312	0.999	0.990	0.839	0.914 (0.895-0.934)	0.263 (0.233-0.298)	0.455
ST	0.976	0.596	0.998	0.978	0.937	0.957 (0.952-0.962)	0.561 (0.542-0.579)	0.729
SB	0.923	0.977	0.917	0.997	0.589	0.793 (0.788-0.797)	0.65 (0.642-0.656)	0.735
Sinus Arrythmia	0.956	0.516	0.995	0.958	0.905	0.932 (0.926-0.937)	0.471 (0.457-0.486)	0.657
LBBB	0.996	0.767	0.999	0.997	0.905	0.951 (0.94-0.962)	0.696 (0.67-0.73)	0.831
RBBB	0.989	0.809	0.998	0.991	0.944	0.967 (0.963-0.972)	0.765 (0.752-0.78)	0.871
LAFB	0.975	0.451	0.993	0.981	0.698	0.84 (0.827-0.852)	0.321 (0.298-0.345)	0.548
LPFB	0.982	0.082	0.998	0.984	0.497	0.74 (0.704-0.777)	0.042 (0.03-0.056)	0.141
SVT	0.998	0.497	0.999	0.998	0.722	0.86 (0.822-0.898)	0.36 (0.3-0.431)	0.589
PAC	0.976	0.798	0.978	0.997	0.308	0.653 (0.642-0.663)	0.267 (0.249-0.289)	0.444
PVC	0.791	0.139	0.995	0.787	0.892	0.84 (0.833-0.846)	0.128 (0.123-0.133)	0.241
LAE	0.880	0.258	0.994	0.879	0.896	0.887 (0.882-0.893)	0.235 (0.227-0.244)	0.400
RAE	0.961	0.141	0.999	0.962	0.904	0.933 (0.919-0.947)	0.128 (0.113-0.143)	0.244
LVH	0.852	0.265	0.996	0.847	0.943	0.895 (0.891-0.899)	0.253 (0.246-0.261)	0.413
RVH	0.940	0.028	0.999	0.941	0.644	0.793 (0.756-0.83)	0.019 (0.014-0.024)	0.053
Low Voltage	0.953	0.664	0.967	0.983	0.498	0.74 (0.733-0.748)	0.362 (0.347-0.379)	0.569
Left Axis Deviation	0.950	0.543	0.978	0.969	0.623	0.796 (0.788-0.804)	0.359 (0.345-0.372)	0.580
Acute MI*	0.943	0.026	1.000	0.943	0.835	0.889 (0.855-0.923)	0.022 (0.016-0.027)	0.051
Lead Reversal	0.767	0.005	0.999	0.768	0.693	0.73 (0.685-0.776)	0.004 (0.003-0.005)	0.009
1st degree AVb	0.980	0.809	0.988	0.991	0.755	0.873 (0.866-0.881)	0.622 (0.606-0.642)	0.781
2nd degree AVb	0.999	0.484	1.000	0.999	0.608	0.804 (0.736-0.871)	0.295 (0.18-0.413)	0.539
3rd degree AVb	1.000	0.481	1.000	1.000	0.464	0.732 (0.638-0.826)	0.224 (0.089-0.388)	0.473
WPW	0.998	0.129	1.000	0.998	0.652	0.825 (0.726-0.925)	0.084 (0.035-0.147)	0.216
Repol Abnormality	0.931	0.115	0.994	0.935	0.6	0.768 (0.751-0.784)	0.074 (0.066-0.084)	0.193

	Labels	Accuracy	PPV	NPV	Specificity	Sensitivity	AUROC	AUPRC	F1
Cardiologist Validated CODE15	AF	0.996	0.765	1.000	0.995	1.000	0.998 (0.996-1.00)	0.765 (0.550-0.947)	0.867
	ST	0.993	0.884	0.999	0.994	0.974	0.984 (0.959-1)	0.862 (0.749-0.973)	0.927
	SB	0.890	0.139	1.000	0.888	1.000	0.944 (0.934-0.954)	0.139 (0.077-0.202)	0.244
	LBBB	0.996	1.000	0.995	1.000	0.871	0.935 (0.876-0.995)	0.875 (0.759-0.971)	0.931
	RBBB	0.990	0.829	0.998	0.992	0.944	0.968 (0.93-1)	0.785 (0.639-0.898)	0.883
	AVb	0.976	0.629	0.990	0.985	0.710	0.847 (0.766-0.929)	0.456 (0.293-0.603)	0.667
CODE15	AF	0.988	0.618	0.998	0.990	0.897	0.943 (0.942-0.945)	0.556 (0.551-0.559)	0.731
	ST	0.986	0.625	0.997	0.989	0.855	0.922 (0.92-0.923)	0.538 (0.534-0.543)	0.722
	SB	0.874	0.109	0.999	0.872	0.972	0.922 (0.921-0.923)	0.107 (0.106-0.108)	0.197
	LBBB	0.993	0.766	0.997	0.996	0.810	0.903 (0.901-0.905)	0.624 (0.619-0.629)	0.788
	RBBB	0.988	0.724	0.998	0.990	0.923	0.956 (0.955-0.957)	0.670 (0.667-0.673)	0.811
	AVb	0.974	0.353	0.997	0.977	0.811	0.894 (0.892-0.896)	0.289 (0.286-0.293)	0.492
UK Biobank	AF	0.997	0.939	0.998	0.999	0.876	0.937 (0.925-0.95)	0.824 (0.799-0.855)	0.906
	ST	0.999	0.862	1.000	1.000	0.815	0.907 (0.868-0.947)	0.703 (0.599-0.798)	0.838
	SB	0.924	0.973	0.894	0.981	0.853	0.917 (0.914-0.919)	0.894 (0.890-0.898)	0.909
	LBBB	0.999	0.871	1.000	0.999	0.982	0.991 (0.984-0.997)	0.855 (0.822-0.884)	0.923
	RBBB	0.996	0.832	1.000	0.996	0.979	0.987 (0.983-0.992)	0.814 (0.791-0.840)	0.899
	AVb	0.969	0.663	0.995	0.973	0.911	0.942 (0.936-0.947)	0.609 (0.592-0.627)	0.767
PTB-XL	AF	0.978	0.787	0.995	0.981	0.931	0.956 (0.949-0.962)	0.738 (0.716-0.758)	0.853
	ST	0.990	0.836	0.996	0.993	0.907	0.950 (0.940-0.960)	0.762 (0.732-0.787)	0.870
	SB	0.876	0.184	0.998	0.874	0.942	0.908 (0.899-0.917)	0.175 (0.161-0.187)	0.308
	LBBB	0.995	0.857	0.999	0.996	0.944	0.970 (0.960-0.980)	0.811 (0.783-0.846)	0.899
	RBBB	0.989	0.691	0.999	0.989	0.980	0.984 (0.978-0.99)	0.678 (0.645-0.713)	0.810
	AVb	0.954	0.419	0.989	0.963	0.711	0.837 (0.821-0.853)	0.309 (0.284-0.335)	0.528
LRH	AFIB	0.969	0.923	0.980	0.980	0.923	0.952 (0.874-1.00)	0.868 (0.665-1.00)	0.923
	ST	0.984	1.000	0.982	1.000	0.900	0.950 (0.852-1.00)	0.916 (0.714-1.00)	0.947
	SB	0.922	1.000	0.915	1.000	0.500	0.750 (0.587-0.913)	0.578 (0.245-0.873)	0.667
	LBBB	0.938	1.000	0.930	1.000	0.636	0.818 (0.669-0.967)	0.699 (0.427-0.873)	0.778
	RBBB	1.000	1.000	1.000	1.000	1.000	1.00 (1.00-1.00)	1.00 (1.00-1.00)	1.000
	AVb	0.906	0.692	0.961	0.925	0.818	0.871 (0.747-0.996)	0.598 (0.312-0.818)	0.750