1

The diagnostic accuracy of the point-of-care urine dipstick test in detecting urinary tract infection among symptomatic patients in Nairobi County, Kenya.

6 Hellen A. Onyango[1,2,3*], Derek J Sloan[1¶], Katherine Keenan[2¶], Mike Kesby[2¶], Caroline Ngugi[3&],

7 Humphrey Gitonga[4&], Robert Hammond[1¶]

9 [1]School of Medicine, University of St Andrews, Scotland, United Kingdom
10

11 [2]School of Geography and Sustainable Development, University of St Andrews, Scotland,
12 United Kingdom
13

14 [3]College of Health Sciences, Jomo Kenyatta University of Agriculture and Technology,
15 Nairobi, Kenya
16

17 [4]Centre for Microbiology Research, Kenya Medical Research Institute, Nairobi, Kenya.
18

20 *Corresponding author

21 E-mail: hao1@st-andrews.ac.uk (HAO)

23 ¶ These authors contributed equally to this work

24 & These authors also contributed equally to this work

Abstract

Background: Urinary tract infections (UTIs) are among the most frequently diagnosed bacterial infections and constitute a large proportion of workload in clinical microbiology laboratories. Urine culture is the confirmatory test for UTI. However, most primary care settings routinely use the more rapid, less labour-intensive dipstick. This study assesses the characteristics of a urine dipstick test in predicting a positive urine culture and how best it can be used in resource constrained settings despite its limitations.

Method: A cross-sectional study was conducted at two level-V health facilities in Nairobi County, Kenya. Adults and children presenting with clinical symptoms of UTI were enrolled after obtaining written informed consent. Midstream urine samples were collected. Urinary dipstick was used to identify Nitrites (NIT) and leucocyte esterase (LE) production. Urine was cultured on Cystine Lactose Electrolyte Deficient agar, blood agar and MacConkey agar; and incubated at 37^0C for 24 hours. Urine cultures with pure bacterial colony counts of $\geq 10^4$ cfu/ml were classified as "positive" for UTI.

Results: Of a total of 552 participants enrolled into the study, 124 (23%) were urine culture positive. Prior medication use was associated with culture negativity. With urine culture as the reference standard, urinary dipstick sensitivity was poor overall (using either LE+ or NIT+ to confer a positive dipstick result still only achieved sensitivity of 66.9%). Using combined NIT+ and LE+ to confer a positive dipstick result had the highest specificity (99.2%), alongside a positive predictive value of 91.1%, and a positive likelihood ratio of 35.6. A NIT+ test alone showed highest concordance with urine culture results (percentage agreement: 86%) but still had a Cohen's Kappa value of only 0.5, conferring weak agreement overall.

Conclusion: Dipstick test is a poor surrogate of urine culture. However, the test may be suitable as a 'rule-out' test to exclude UTI, and avoid antibiotic prescription, when both NIT

53 and LE are negative. Although dipstick continues to be in use in resource constrained settings,

54 poor concordance with urine culture results highlights a need for better near patient tests to

55 diagnose UTI and guide antibiotic decision-making.

56 **Keywords: Urinary tract infection, point of care, dipstick, urine culture, low-and-middle**

57 **income countries, accuracy.**

1. Introduction

UTI is an inflammatory response of the urothelium to bacterial invasion and is among the

most frequent community-acquired bacterial infection, affecting more than 150 million

individuals globally [1,2]. The disease is associated with morbidity and mortality, particularly

among high-risk subpopulations such as children, pregnant women, the elderly and

immunocompromised patients [3,4]. Allied to death and disability are prolonged hospital

stays, the need for second-line antimicrobial drugs, adverse impacts on the income and

livelihoods of individuals, and general economic harm at a national level [5,6].

Prompt diagnosis of urinary tract infection is critical for patient management. However, some

people with UTI present with atypical signs and symptoms, whilst others may have UTI

symptoms in the absence of infection. This makes the clinical diagnosis a challenge [7], and

illustrates the need for laboratory investigations. Quantitative urine culture is the gold

standard method for UTI diagnosis despite widely acknowledged limitations of cost, labour

intensity and prolonged time to result of between 24-72 hours [8]. These limitations have

made urine dipstick a more convenient and frequently used first line laboratory investigation

for UTI in primary care settings [9]. In many resource-limited settings it is the only test

available [10,11].

Dipstick test is a 10-parameter reagent strip designed to test markers of infection to a range

of medical conditions, based on colorimetric principles. Markers which are mainly helpful to

detect UTIs are leucocytes esterase (LE) and nitrite (NIT) which detects bacteriuria or pyuria

in urine [9]. Nitrite testing detects the presence of bacteria with enzymes that can convert

urinary nitrates to nitrites and is associated with members of the family Enterobacterales.

Other urinary pathogens such as *Staphylococcus* spp*., Pseudomonas* spp., and *Enterococcus*

spp. do not produce nitrate reductase enzymes [8]. This limits the microbiological range of

nitrite-mediated detection of UTI. Another limitation is the fact that a minimum bladder

incubation period of 4 hours is required for nitrates to be converted to nitrites at reliably

detectable levels [7,8].

The leucocyte esterase (LE) test detects whether leucocytes have produced proteins with

esterolytic activity that have then hydrolysed ester substrates. LE testing is likely to give a

false positive result when urine is highly contaminated with bacterial vaginal flora; when

specimens contain eosinophils or *trichomonas* species, both of which can produce esterase;

and when the strip is exposed to an oxidizing agent or formalin [7,12,13]. False negative

results may arise when urine has high levels of glycosuria and proteinuria; when urine is

preserved using boric acid; and when patient is on antibiotic treatment regimen [12]. In

addition to nitrites and leucocyte esterase, other urinary dipstick markers of UTI include

blood, increased pH, and proteins. However, these are less specific, as they also have many

other causes [8].

A previous meta-analysis conducted on the accuracy of urine dipstick test relative to

quantitative urine culture found moderate sensitivity (48%) and specificity (91%) in detecting

UTIs. The sensitivity was higher in inpatients (58% vs 45%), while the specificity was greater

in outpatients (96% vs 45%)[14]. In High Income Countries (HIC), there is widespread

recognition of the urine dipstick test's utility as a preliminary screening tool, guiding clinicians

on whether further culture is warranted. However, in Low-and-middle Income (LMICs), most

hospitals lack the laboratory capacity to perform urine culture and rely entirely on dipstick for

diagnosis. This may result in inappropriate use of antimicrobials. This study quantifies the

limitations of dipstick characteristics in settings where it is mostly used without culture back-

up and assesses how best it can be utilized despite its limitations.

105 2. Materials and methods

106 2.1. Study design

A cross-sectional study was conducted at Mama Lucy Kibaki Hospital (MLKH) and Mbagathi

District Hospital (MDH) between 25/05/2022 to 22/04/2023. The hospitals are located within

Nairobi County, Kenya, and mostly serve the urban populace. The Kenyan healthcare system

is structured in a hierarchical manner, from community-based primary healthcare services

through to specialized hospital care. The current structure consists of six levels (I-VI) in

ascending order. Both MLKH and MDH are government owned level (V) tertiary health

facilities.

114 2.2. Participant recruitment and sample collection

Adults (≥18 years) and children (5-17 years) were recruited at the outpatient department of

MLKH or MDH if they presented with one or more UTI symptoms (increased urinary frequency

or urgency, dysuria, burning sensation upon urination, lower abdominal pain, and/or

unexplained fever (≥38°C) in children). Participants were taken through the informed consent

document in their preferred language (English or Kiswahili). Written informed consent was

obtained from adult patients. Assent and consent were obtained for participants aged 13-17

years. Parents/guardians of participants aged ≤13 years consented on their behalf.

Participants/guardians who were unable to sign marked the consent form with a thumb print.

Potential participants who did not meet the criteria of a presumptive UTI case, or who

declined to consent were excluded from the study. A structured questionnaire was used to

collect self-reported socio-demographic details such as age, gender, level of education,

monthly income, and prior antibiotic intake. Parents/ guardians of children (5-17 years) filled

the questionnaire on their behalf. All data were collected electronically using Epicollect5

mobile application (https://five.epicollect.net) [15]. Consenting participants were instructed

on how to aseptically collect clean catch mid-stream urine into a sterile screw capped

universal bottle. All samples were assigned a unique study identification number, transported

to the microbiology laboratory, and processed within two hours of collection.

2.3. Dipstick test.

The dipstick test was conducted using combur-10 test M strips according to the

manufacturer's instructions (Combur-10, UK). The urine strip was dipped into approximately

10 ml of the urine specimen, removed immediately, and results read after waiting period of

2 minutes. The strip was held horizontally adjacent to the reagent colour blocks on the strip

container and colours carefully matched. LE was reported as negative, trace, 1+ small, 2+

moderate, 3+ large while NIT was recorded as either positive or negative. With reference to

the manufacturers guide for interpretation, dipstick testing that produced nitrites or

leucocyte esterase greater than trace was taken as positive for UTI. No nitrites or leucocyte

esterase were interpreted as negative.

2.4. Quantitative urine culture assay

A well-mixed 10 µl urine aliquot was plated directly on CLED agar, blood agar (BA) and

MacConkey agar (Oxoid, England), and incubated aerobically at 37^0C for 24 hours. A pure

bacterial growth yielding colony counts of 10^4 cfu/ml was deemed significant for a UTI

infection. Mixed urine cultures (with more than one colony type) or those with either low

bacterial colony counts of $<10^4$ cfu/ml, or without microbial growth were interpreted as

negative for UTI. Isolates were identified to the species level using colonial morphological

characteristics, Gram-staining technique and standard biochemical tests (catalase, coagulase,

urease, oxidase, sulfide indole motility, methyl red, citrate utilization)[16].

2.5. Statistical analysis

Data were downloaded from Epicollect to Microsoft excel for analysis (Microsoft Corp, Red-

Mond, Washington, USA). Characteristics of study participants were all presented as

categorical variables including age (which was separated into brackets of 5-10, 11-20, 21-30,

31-40, 41-50 and >50 years) and average monthly income (which was separated into brackets

of <3000, 3000-9900, 10000-24000, 25000-49990, 50000-74900, and 75000-100000 Ksh). Age

was also summarised as median and interquartile range (IQR).

For urinary dipstick, LE and NIT diagnostic yield were calculated in four ways: LE positive,

irrespective of NIT result (LE+), NIT positive irrespective of LE result (NIT+), both NIT and LE

positive (NIT+ and LE+), and either NIT or LE positive (NIT or LE+) [17]. Univariate logistic

regression was performed to assess the relationship between prior medication exposure and

urine culture positivity, with the result expressed as an Odds Ratio (OR) and 95% confidence

Interval (CI). The sensitivity, specificity, positive predictive value (PPV), negative predictive

value (NPV) and likelihood ratios (LR) were calculated with 95 % CI for dipstick parameters

using urine culture as the reference standard. Whilst there is no universal agreement on

target performance thresholds for UTI diagnostics, it is generally accepted that diagnostic

tests with very high positive +LR (>10) and very low -LR (<0.1) are most useful in ruling in or

ruling out a clinical diagnosis respectively[18]. Percentage agreement and Cohen's Kappa

coefficient were applied to assess the agreement between dipstick and culture results.

Cohen's Kappa results were interpreted as follows: 0-0.2, 'no' agreement; 0.21-0.39,

'minimal' agreement; 0.40-0.59, 'weak' agreement; 0.60-0.79, 'moderate' agreement; 0.80-

0.90, 'strong' agreement; >0.9, 'perfect' agreement [19].

173 2.6. Ethical statement

This study was approved by the University of St. Andrews Teaching and Research Ethics

Committee (UTREC), [Approval code. MD15749]; Jomo Kenyatta University of Agriculture and

Technology Institutional Ethics Review Board (JKUAT-IERB) [Approval no.

JKU/IERC/02316/0166]; National Commission for Science Technology and Innovation

(NACOSTI) [Approval no. P/21/12520]. Approvals to access study sites were also obtained

from the Nairobi metropolitan services, MLKH and MDH. Each participant in the study

provided either a written informed consent or marked the consent form with a thumb print.

Although the study involved children, their participation was limited to providing a urine

sample under the supervision of their parents/guardians. All questionnaires were completed

by parents/guardians on their behalf. The principal investigator was never alone with the

children; all interviews were conducted in the presence of an adult. Therefore, certification

for working with people in the vulnerable group or the local East African equivalent was not

required.

187 3. Results

188 3.1. Characteristics of study participants

A total of 622 participants were screened for enrolment into the study. However, 552 were

enrolled, 402 at Mama Lucy Kibaki Hospital, 150 at Mbagathi District hospital. Seventy were

excluded for the reasons outlined in **Fig 1.** The median age of the participants was 29 years

(IQR:24-36). There were proportionally more females 72% (398/552). Most of the participants

had at least a secondary level of education 270 (49%). Among the 552 individuals recruited,

236 (43%) had taken medication within two weeks preceding their enrolment. Of these

individuals, 168 (30%) had been on antibiotics, while the remaining 68 (12%) had taken other

types of medication. Details of participants characteristics are shown in **Table 1.**

Fig 1: **Participant recruitment flow diagram**. A total of 622 symptomatic adults and child

patients were screened for enrolment. Five hundred and fifty-two urine samples were

obtained from eligible participants tested using dipstick and culture.

Table 1: Social- demographic characteristics of study participants

Variable	Response	n=552
		n (%)
Gender	Male	154 (28)
	Female	398 (72)
Age	5-10	30 (5)
	11-20	49 (9)
	21-30	238 (43)
	31-40	128 (23)
	41-50	67 (12)
	>50	40 (7)
Marital status	Married	258 (52)
	Single	184(37)
	Widowed	37 (7)
	Divorced	15 (2)
Level of education	Pre-primary	17 (3)
	Primary	116 (21)
	Secondary	270 (49)
	Tertiary	149 (27)
Average monthly income (Ksh)	<3000	133 (24)
	3000-9900	110 (20)
	10000-24900	139 (25)
	25000-49900	94 (17)
	50000-74900	68 (12)
	75000-100000	8 (1)
Medication intake	No medication	316 (57)
	Yes-antibiotics	168 (30)
	Yes-others	68 (12)

3.2. Microbiological characteristics

Among 552 urine samples analysed, 124 (22.5%) were positive for UTI (a monoculture bacterial growth of $\geq 10^4$ cfu/ml). Two-hundred-and-thirty-six (43%) participants had taken medication in the two weeks prior to enrolment, and a negative culture result was more likely in those who had taken prior medication (OR: 1.3, 95% CI; 0.64-2.87).

The bacterial isolates from positive cultures were characterised and identified to the species

level. The proportion of infections attributable to Gram-negative organisms were 97 (78%).

The most prevalent uropathogen was *Escherichia coli*, 64 (51.6%). *E. coli* was followed in order

of prevalence by *Klebsiella pneumoniae*, 16 (12.9%); *Staphylococcus aureus*, 14 (11.3%);

coagulase negative Staphylococcus (CoNS), 7 (5.6%); *Enterococcus faecalis*, 6 (4.8%),

Klebsiella spp., 5 (4%); *Proteus mirabilis,* 5 (4%); and *Acinetobacter baumanni*, 1(0.8%).

Pseudomonas aeruginosa, *Proteus vulgaris*, and *Citrobacter koseri* each were 2 (1.6%).

The positivity rate from urinary dipstick was dependent on how LE and NIT results were

interpreted, as shown in **Fig 2**. If 'NIT+ or LE+' was considered, 187 (34%) of participants were

positive, translating to a higher positivity rate than culture. If only 'LE+' was used, 158 (29%)

were positive which also exceeded the culture positivity rate. 'NIT+' and 'NIT and LE+' analyses

yielded 63 (11%) and 34 (6%) positive results respectively which were lower than the culture

positivity rate. In general, dipstick positivity was higher when based on 'LE+' alone, and lower

when a 'NIT+' result was required.

Fig 2: UTI prevalence estimation based on dipstick test parameters. UTI prevalence based

on culture was 23%. The LE, NIT or LE estimated UTI prevalence was higher, while NIT, NIT

and LE were lower than prevalence based on culture.

3.3. Performance characteristics of dipstick using urine culture as a reference

Having identified positivity rates from different combinations of urinary dipstick results, it was

important to progress to assess performance in relation to the current gold standard urine

culture. Results are provided in **Table 2.** Having either NIT+ or LE+ yielded the highest

sensitivity of 67%, but overall sensitivity of urine dipstick was poor (ranging from 25-67%).

Having NIT+ and LE+ yielded the highest specificity of 99.3%, and specificity of urine dipstick

was generally higher when NIT+ was required as part of a positive result (97.9% for NIT+ and

99.3% for NIT+ and LE+). Similar to specificity, both PPV and +LR were highest when NIT+ was

required as part of a positive result (PPV of 85.7% for NIT+ and 91.2% for NIT+ and LE+; +LR

of 20.7 for NIT+ and 35.7 for NIT+ and LE+). Overall, NPV ranged from 82-9% and -LR ranged

from 0.43-0.66.

Table 2: Diagnostic performance of NIT and LE relative to quantitative urine culture.

Dipstick parameter		Urine culture		Dipstick performance (95% CI)					
		+	-	Sensitivity % (95% CI)	Specificity % (95% CI)	PPV % (95% CI)	NPV % (95% CI)	+ LR (95% CI)	- LR (95% CI)
LE	+	60	98	48.38 (39.38 – 57.49)	77.10 (72.76 – 80.94)	37.97 (30.48 – 46.06)	83.75 (79.65 – 87.18)	2.11 (1.64-2.17)	0.66 (0.56-0.79)
	-	64	330						
NIT	+	54	9	43.55 (34.76 – 52.73)	97.89 (95.90 – 98.97)	85.71 (74.10 – 92.86)	85.68 (82.19 -88.60)	20.7 (10.52-40.74)	0.57 (0.49-0.67)
	-	70	419						
NIT and LE	+	31	3	25.00 (17.85 – 33.72)	99.29 (97.78 – 99.81)	91.17 (75.18 – 97.69)	82.05 (78.40 – 85.20)	35.66 (11.09-14.69)	0.75 (0.68-0.84)
	-	93	425						
NIT or LE	+	83	104	66.93 (57.84 – 74.96)	75.70 (71.29 – 79.63)	44.38 (37.19 – 51.81)	88.76 (84.96 – 91.72)	2.75 (1.93-2.78)	0.43 (0.36-0.57)
	-	41	324						
Quantitative urine culture was used as the reference test; NIT - Nitrite; LE - Leucocyte Esterase; NIT and LE - Combined nitrites and Leucocyte Esterase; NIT or LE - Either nitrite or leucocyte Esterase; PPV-Positive Predictive Value; NPV- Negative Predictive Value; LR-Likelihood Ratio.									

251 None of age, gender or use of prior medication, including antibiotics, substantially altered any

252 performance characteristics of urinary dipstick compared to a reference of urine culture.

3.4. Concordance between dipstick and culture results.

Further to measuring dipstick performance with culture as a 'gold standard', direct

concordance of results was also assessed. Percentage agreement and Cohen's kappa results

were both calculated, as shown in **Table 3**, because Cohen's kappa considers the possibility

that some concordance may occur by chance. The highest percentage agreement between

dipstick and culture occurred when NIT+ was used to denote dipstick positivity (86%

agreement, 95% CI: 0.82-0.88) but even then, the Cohen's kappa of 0.5 described only 'weak'

agreement.

Table 3: Concordance between Dipstick parameters and quantitative urine culture.

Dipstick parameter	Concordance with urine culture		
	(%) Agreement (95% CI)	Cohen's Kappa (95% CI)	Kappa interpretation
LE	71 (0.66-0.74)	0.23 (0.13-0.33)	Minimal agreement
NIT	86 (0.82-0.88)	0.50 (0.40-0.60)	Weak agreement
NIT and LE	82 (0.78-0.85)	0.32 (0.20-0.44)	Minimal agreement
NIT or LE	73 (0.69-0.77)	0.36 (0.25-0.41)	Minimal agreement

4. Discussion

This study formally quantified the limitations of the urine dipstick test in accurately predicting a positive urine culture, particularly to inform it's use in resource-constrained settings where confirmatory culture may not be available. The overall results show that urine dipstick test is an inadequate tool to assess the probability of a positive urine culture in outpatients with UTI associated symptoms. However, given the test characteristics, the dipstick test may be considered as a rule-out tool for antibiotic prescriptions; a negative dipstick test result characterized by no reaction for both nitrites and leucocyte esterase has very high specificity to predict a negative urine culture, making the diagnosis of UTI unlikely and empirical antimicrobial therapy unnecessary. Implementation of this strategy in an outpatient setting where culture is rarely performed could lead to a reduction in the prescription of antibiotics, thereby contributing to antimicrobial stewardship.

Even though all 552 participants recruited to this study had typical symptoms of UTI, only 124 (23%) were confirmed by positive urine culture, illustrating the difficulty for healthcare providers in deciding on the need for antibiotic prescription based on clinical presentation alone. However, 30% of patients had already taken antibiotics before presentation to hospital out-patient department. This has been demonstrated as a clear predictor of a negative culture result in the present study and in prior literature [16], underlining the challenge of performing and interpreting microbiology test results in a heavily antibiotic pre-exposed population. The overall context highlights the need for development and expanded availability of new, reliable point of care tests to guide decision-making in UTI management.

The urine dipstick test is a complex read-out with different variables reported. In this study, consideration of urine dipstick positivity as 'either NIT+ or LE+' reported a positivity rate of

34%. This compares to prior dipstick positivity reports of 54% by Maina *et al* in a similar study

in Kenya[11], 73% by Katunzi *et al* in Tanzania[20] and 34% by Dadzie *et al* in Ghana[17].

Differences in the care setting and population-specific estimates have been cited as the main

sources of heterogeneity of dipstick results between studies[21].

In the context of clinical decision-making for infection management, diagnostic tests can be

used to 'rule-out' or to 'rule-in' the need for antibiotic therapy. From these data, the

requirement for both NIT+ and LE+ to confer dipstick positivity had potential as a good 'rule-

out' test due to its high specificity of 99%, albeit with a cautionary note to clinicians that of

518 patients who were dipstick negative on this basis, 93 (17.9%) had positive urine cultures.

Most uncomplicated UTIs are self-limiting in large patient groups[22], suggesting that an

initial missed diagnosis will often have minimal consequences. However, for certain patient

groups the balance of risk may be different; e.g., for pregnant women, a missed UTI diagnosis

could lead to poor perinatal and maternal outcomes [23]. Overall, the findings concur with a

meta-analysis including 14 studies that support the use of urine dipstick test in a rule-out

strategy[24]. A similar conclusion was drawn by Deville *et al* in a meta-analysis that included

70 publications, despite a considerable variation in settings of the different studies that

included both out- and in-patients, emergency department, ante-natal unit, and all levels of

care from the community to tertiary care [14]. These results are all contradictory to an older

meta-analysis which reported that in many clinical settings, the probability of a positive

culture with a negative dipstick test is too high to dismiss the probability of a UTI [25].

An alternative, and common, use of the urine dipstick test in resource limited settings is by

interpreting the positive results as a tool to support the diagnosis of UTI (rule-in strategy) and

as the key result necessary to justify to antibiotic initiation. Whilst 'either NIT+ or LE+' confers

the highest sensitivity for any dipstick positivity analysis in this study, the result of 67%

sensitivity is still unsatisfactory (1/3 of patients with culture confirmed UTI will be missed).

Additionally, the specificity of interpreting the dipstick result in this was low. These findings

mirror those of Maina *et al* and Dadzie *et al* who report an overestimation of UTI by 'either

NIT+ or LE+' test results compared to culture by 74% and 72% respectively. The consequence

may be that some patients who don't need antibiotics will receive them, whilst a concerning

proportion of those who need them may not have them prescribed.

The greatest overall agreement between urinary dipstick and culture was found when dipstick

positivity is reported based on a NIT+ test alone. Therefore, this approach to dipstick

interpretation might be the best surrogate for culture positivity if dipstick is the only test

available, but the kappa agreement of 0.5 remains weak. In prior studies, the presence of

nitrites in urine has been shown to be highly specific for bacteriuria (96.6%-97.5%), but has a

low sensitivity of 0-44% for bacteriuria between 10^3-10^5 cfu/ml [7,9,17,26,27]. The typically

low sensitivity value for nitrites can be explained by the process of nitrification that requires

approximately 4 hours for detectable nitrite levels to be produced in urine [28]. Alternative

explanations could be the lack of dietary nitrates, dilution of the nitrite in urine with diuretics,

and unreliability of the nitrite test in detecting Gram-positive organisms[8]. Therefore, the

presence of nitrites is highly predictive of UTI, but their absence does not exclude it. In a

marked contrast, a higher sensitivity of 90% for NIT+ alone, was reported by Sirasaporn *et al*

among 56 spinal code injury patients (SCI) in Thailand [29]. While UTI in the general

population is based on symptoms, urinalysis and culture, this approach is not applicable to

SCI patients as they already have an impaired neurological function and urinary

system[30,31]. Perhaps, the difference in target population, diagnostic criterion, and the

small sample size could explain the discrepancy between this study and others.

The LE+ test alone has a slightly higher sensitivity, and lower specificity when compared to

the individual performance of NIT+. The test has poor positive predictive value and minimal

agreement with culture (kappa, 0.23). The LE sensitivity of 62.2% reported by Maina *et al*

[11], 60% by Dadzie *et al* [17] and 48% by Anith *et al* [9] were all higher than the nitrite test,

which mirrors that of the present study. Despite being more sensitive than nitrites, the LE

test is not specific for UTI and may be associated with other inflammatory disorders affecting

the urinary tract such as vaginitis, chlamydial urethritis or other infections than can elicit an

immune response and production of white blood cells [32]. These factors may have

contributed to variability in the positive predictive value of LE test from 19% to 88% in prior

work, and may also undermine its usefulness in the present dataset [26,33].

The diagnostic gains of the urine dipstick assessed through the likelihood ratios (+LR and -LR),

ranged from 2.1-30.7 and 0.4-0.7 respectively. Likelihood ratios, which result from the

interplay   of sensitivity and specificity provide valuable insights into the probability of 'ruling

in' or 'ruling out' an infection. According to Guyatt *et al*, diagnostic tests with very high

positive +LR (>10) and very low -LR (<0.1) are very useful in ruling in or ruling out a diagnosis

respectively[18,34].   Based on this, NIT+, and combined 'NIT+ and LE+' are useful indicators

for ruling in a UTI. These findings are comparable to systematic review by John *et al* which

reported +LR ranging from 4.27-29.3 and -LR from 0.2-0.5 [24]. However, the ≤0.5 kappa

values recorded for all the dipstick parameters suggests a minimal or weak agreement, and

consequently, little confidence should be placed on results generated by dipstick in the

absence of culture results.

This study had some limitations. The study was conducted in two purposively chosen health

facilities in an outpatient setting. Generalisation to other settings or patient populations can

be challenging, particularly when symptoms associated with UTI may stem from other

underlying medical conditions. Optical (naked eye) reading of test strips, as employed in this

study may have introduced variations in results accuracy. Reports indicate that automated

reading methods using spectrophotometers deliver superior outcomes. By enabling

standardized readings, the methods enhance repeatability and reproducibility, indirectly

influencing the analytical performance of the tests. Despite the enrolment of over 500

participants, the number of positive reference test based on culture was limited by antibiotic

intake prior to recruitment into the study and a definition of positivity that required a

concentration of 10^4cfu/ml. As the majority (78%) of positive urine cultures were attributable

to gram negative bacterial infections, there was insufficient diversity of pathogens to study

whether different infecting organisms influenced the diagnostic performance of urinary

dipstick results.

4. Conclusion

This study corroborate the widely held view that dipsticks are a poor surrogate for urine

culture and are not suitable as definitive method for UTI diagnosis. However, particularly in

settings where urine culture is unavailable, the use of urine dipstick test (Negative for both

NIT and LE) as a rule out strategy has potential in lower risk patient groups to reduce

unnecessary antibiotic prescriptions. Future studies might combine specific UTI symptoms

and dipstick test results to explore possible approaches to increase the test's reliability.

Finally, the challenges of performing urine culture diagnostics in resource limited settings,

calls for research to develop and validate novel near or point of care technologies with better

performance characteristics than urinary dipstick.

Acknowledgements

The authors would like to thank and appreciate all participants for their participation in the

study. We would also like to thank MLKH and MDH for giving us an opportunity to conduct

this research, not forgetting the clinicians and laboratory personnel who supported us in data

collection.

REFERENCES

1. Öztürk R, Murt A. Epidemiology of urological infections: a global burden. World J Urol [Internet]. 2020;38(11):2669–79. Available from: https://doi.org/10.1007/s00345-019-03071-4

2. Li X, Fan H, Zi H, Hu H, Li B, Huang J, et al. Global and Regional Burden of Bacterial Antimicrobial Resistance in Urinary Tract Infections in 2019. J Clin Med. 2022;11(10):1–14.

3. Walsh C, Collyns T. The pathophysiology of urinary tract infections. Surg (United Kingdom) [Internet]. 2017;35(6):293–8. Available from: http://dx.doi.org/10.1016/j.mpsur.2017.03.007

4. Okojie RO, Omorokpe VO. A survey on urinary tract infection associated with two most common uropathogenic bacteria. African J Clin Exp Microbiol. 2018;19(3):111–5.

5. E Esteve-Palau, G Solande, F Sánchez, L Sorlí, M Montero, R Güerri, J Villar, S Grau JPH. Clinical and economic impact of urinary tract infections caused by ESBL-producing Escherichia coli requiring hospitalization: A matched cohort study. J Infect. 2015;71(6):667–74.

6. Foxman B. Epidemiology of urinary tract infections: Incidence, morbidity, and economic costs. Disease-a-Month. 2003 Feb 1;49(2):53–70.

7. Bartlett JG. Laboratory diagnosis of urinary tract infections in adult patients. Infect Dis Clin Pract. 2004;12(6):360–1.

8. Graham J C AG. Laboratory diagnosis of urinary tract infection. J Clin Pathol. 2001;54(1):911–9.

9. Mambatta A, Jayarajan J, Rashme V, Harini S, Menon S, Kuppusamy J. Reliability of dipstick assay in predicting urinary tract infection. J Fam Med Prim Care. 2015;4(2):265.

10. Masoud SS, Majigo M, Silago V, Kunambi P, Nyawale H, Moremi N, et al. Utility of dipstick urinalysis in the diagnosis of urinary tract infections among outpatients in Mwanza and Dar es Salaam regions in Tanzania. Bull Natl Res Cent [Internet]. 2024;48(1). Available from: https://doi.org/10.1186/s42269-023-01159-x

11. Maina J, Mwaniki J, Mwiti F, Kiiru S, Katana J, Wanja F, et al. Evaluation of the diagnostic

performance of the urine dipstick test for the detection of urinary tract infections in patients

treated in Kenyan hospitals. Access Microbiol. 2023;5(6):1–13.

12. Fuller CE, Threatte GA, Henry JB, Davey FR HC. Basic Examination of Urine. In: Clinical

diagnosis and management by laboratory methods. 2001. p. 367–402.

13. Nicolle LE, Bradley S, Colgan R, Rice JC, Schaeffer A, Hooton TM. 7 Infectious Diseases

Society of America Guidelines.pdf. 2005;

14. Devillé WLJM, Yzermans JC, Van Duijn NP, Bezemer PD, Van Der Windt DAWM, Bouter

LM. The urine dipstick test useful to rule out infections. A meta-analysis of the accuracy. BMC

Urol. 2004;4:1–14.

15. Aanensen DM, Huntley DM, Feil EJ, Al-Own F, Spratt BG. EpiCollect: Linking smartphones

to web applications for epidemiology, ecology and community data collection. PLoS One.

2009;4(9).

16. Chesbrough M. District Laboratory Practice in Tropical countries, Part 2, Second Edition.

Cambridge University Press, UK; 2006.

17. Dadzie I, Quansah E, Puopelle Dakorah M, Abiade V, Takyi-Amuah E, Adusei R. The

Effectiveness of Dipstick for the Detection of Urinary Tract Infection. Can J Infect Dis Med

Microbiol. 2019;2019.

18. Jaeschke R, Guyatt G LJ. Diagnostic tests. In: Guyatt G, Rennie D, eds. In: Chicago, editor.

Users' guides to the medical literature. AMA Press; 2002. p. 121–40.

19. McHugh ML. Lessons in biostatistics interrater reliability : the kappa statistic. Biochem

Medica [Internet]. 2012;22(3):276–82. Available from: https://hrcak.srce.hr/89395

20. George Msema B, Leonard K, James K, Lilian N, Akili M, Kennedy Daniel M. Diagnostic

accuracy of dipsticks test among clinically suspected urinary tract infected patients at National

Hospital, Tanzania. Open J Bacteriol. 2019;3(1):003–7.

21. Bafna P, Deepanjali S, Mandal J, Balamurugan N, Swaminathan RP, Kadhiravan T.
Reevaluating the true diagnostic accuracy of dipstick tests to diagnose urinary tract infection
using Bayesian latent class analysis. PLoS One [Internet]. 2020;15(12 December):1–12.
Available from: http://dx.doi.org/10.1371/journal.pone.0244870

22. Vik I, Bollestad M, Grude N, Bærheim A, Damsgaard E, Neumark T, et al. Ibuprofen versus
pivmecillinam for uncomplicated urinary tract infection in women — A double-blind ,
randomized non- inferiority trial. PLoS Med. 2018;15(5):1–21.

23. Schnarr J, Smaill F. Asymptomatic bacteriuria and symptomatic urinary tract infections in
pregnancy. Vol. 38, European Journal of Clinical Investigation. 2008. p. 50–7.

24. John AS, Boyd JC, Lowes AJ, Price CP. The Use of Urinary Dipstick Tests to Exclude Urinary
Tract Infection. Am J Clin Pathol. 2006;126(3):428–36.

25. Terry A. Hurlbut, III, M.D, Benjamin Littenberg M. The Diagnostic Accuracy of Rapid
Dipstick Tests to Predict Urinary Tract Infection. Am J Clin Pathol. 1991;96(5):582–8.

26. Semeniuk H, Church D. Evaluation of the leukocyte esterase and nitrite urine dipstick
screening tests for detection of bacteriuria in women with suspected uncomplicated urinary
tract infections. J Clin Microbiol. 1999;37(9):3051–2.

27. Prah JK, Amoah S, Ocansey DWK, Arthur R, Walker E, Obiri-Yeboah D. Evaluation of
urinalysis parameters and antimicrobial susceptibility of uropathogens among out-patients at
University of Cape Coast Hospital. Ghana Med J. 2019;53(1):44–51.

28. Mahyar A, Ayazi P, Froozesh M, Daneshi-Kohan MM, Barikani A. Can urinary nitrite results
be used to conduct antimicrobial option for urinary tract infection in children? Iran J Pediatr.
2012;22(2):237–40.

29. Duanngai K, Sirasaporn P, Ngaosinchai S. The reliability and validity of using the urine
dipstick test by patient self-assessment for urinary tract infection screening in spinal cord
injury patients. J Fam Med Prim Care. 2017;6(3):578.

30. Jeanne M Hoffman, Rohini Wadhwani, Eve Kelly, Bonnie Dixit DDC. Nitrite and leukocyte

dipstick testing for urinary tract infection in individuals with spinal cord injury. J Spinal Cord

Med. 2004;27(2):128–32.

31. García Leoni ME, De Ruz AE. Management of urinary tract infection in patients with spinal

cord injuries. Clin Microbiol Infect. 2003;9(8):780–5.

32. Mohanna AT, Alshamrani KM, SaemAldahar MA, Kidwai AO, Kaneetah AH, Khan MA, et al.

The Sensitivity and Specificity of White Blood Cells and Nitrite in Dipstick Urinalysis in

Association With Urine Culture in Detecting Infection in Adults From October 2016 to October

2019 at King Abdulaziz Medical City. Cureus. 2021;13(October 2016):1–7.

33. Young JL, Soper DE. Urinalysis and urinary tract infection: Update for clinicians. Infect Dis

Obstet Gynecol. 2001;9(4):249–55.

34. Deeks JJ, Altman DG. Diagnostic tests 4: likelihood ratios. Bmj. 2004;329(7458):168–9.

SUPPORTING INFORMATION

S1 Table. Performance of LE.

S2 Table. Performance of NIT.

S3 Table. Performance of NIT and LE

S4 Table. Performance of either NIT or LE

S1-S4 Tables. Performance characteristics of dipstick parameters across study variables. The dipstick performance was not influenced by age, gender or prior medication intake.

S5 File. Dipstick manuscript raw data

S6 File. Participants baseline demographic characteristics

Figure

Figure