Title:

Factors affecting anal sphincter recruitment during intraoperative pudendal nerve stimulation

Authors:

Amador C Lagunas[1,2], Po-Ju Chen[1,2], Luis Ruiz[1,2], Amolak S Jhand[1,2], Nystha Baishya[1,2], Scott F Lempka[1,2,3], Priyanka Gupta[4], Tim M Bruns[1,2]

Affiliations:

[1]Department of Biomedical Engineering, University of Michigan, Ann Arbor, MI, USA
[2]Biointerfaces institute, University of Michigan, Ann Arbor, MI, USA
[3]Department of Anesthesiology, University of Michigan, Ann Arbor, MI, USA
[4]Department of Urology, University of Michigan, Ann Arbor, MI, USA

Corresponding author:

Tim M. Bruns
Email: bruns@umich.edu
Phone: (734) 647-8727
Address: B10-A169 NCRC, 2800 Plymouth, Ann Arbor, MI 48109-2800

Financial Disclaimers/Conflict of Interest: This project was funded by an award from the National Institutes of Health SPARC program (OT2OD028191).

Prior Presentation: Society of Urodynamics, Female Pelvic Medicine & Urogenital Reconstruction 2022 Meeting, San Diego, February 23, 2022

Contributions:

AC Lagunas: Data collection, Data Analysis, Manuscript Writing
P Chen: Data Collection, Manuscript Editing
L Ruiz: Data Analysis, Manuscript Editing
AS Jhand: Data Analysis, Manuscript Editing
N Baishya: Data Analysis, Manuscript Editing
SF Lempka: Project Development, Manuscript Editing
P Gupta: Project Development, Manuscript Editing
TM Bruns: Project development, Data Collection, Manuscript Writing

Abstract

Introduction and Hypothesis

The relationship between pudendal neuromodulation and patient factors is not well understood. This study aimed to identify and quantify physiological, demographic, and stimulation factors that significantly affect external anal sphincter (EAS) recruitment and outcomes in participants receiving pudendal neuromodulation for treatment of lower urinary tract symptoms and pelvic pain.

Methods

Participants (N=16) provided demographic and diagnostic information upon entry to this observational study. EAS activation at different stimulation amplitudes and pulse widths was recorded during lead implantation. Magnetic resonance imaging and computed tomography were used to determine the distance of the electrodes on the implanted lead from the nerve. Linear mixed modeling was used to quantify the impact of each variable on EAS recruitment.

Results

Participant sex, age, and body-mass index did not significantly affect EAS recruitment with pudendal nerve stimulation. Participant diagnoses had significant relationships to EAS recruitment, likely due to unbalanced group sizes. A pulse width of 210 μs required less current than 60 μs (p = 0.005) and less charge than 450 μs (p = 0.02) to activate the EAS at threshold. Increased electrode-to-nerve distance decreased the magnitude of the EAS response (p = 0.0011), increased the EAS activation threshold (p < 0.001), and was related to reduced bladder symptom improvements.

Conclusions

Of the three tested pulse widths, 210 μs requires the least charge and is a good choice for stimulation. The distance between the electrode and pudendal nerve has an important role in nerve recruitment and minimizing this distance should be a priority during lead implantation.

Keywords

electromyography, external anal sphincter, linear mixed model, neuromodulation, pudendal nerve

Brief Summary:

This study identifies key factors in external anal sphincter recruitment during pudendal neuromodulation that should be considered during lead implantation.

Abbreviations:
EAS: External Anal Sphincter
LUT: Lower Urinary Tract
LUTD: Lower urinary tract disorders
CT: Computed tomography
MRI: Magnetic resonance imaging
AUASI: American Urological Association Symptom Index
 M-ISI: Michigan Incontinence Symptom Index
FSFI-6: Female Sexual Function Index-6
SHIM: Sexual Health Inventory for Men: SHIM
CRAD-8: Colorectal-Anal Distress Inventory-8
FGUPI: Female Genitourinary Pain Index
MGUPI: Male Genitourinary Pain Index
EMG: Electromyography
AUC: Area under the curve
IQR: Interquartile range
OAB: overactive bladder

Introduction

Chronic pelvic pain and lower urinary tract disorders (LUTD) affect millions of people in the United States [1,2]. These conditions result in diminished physical health, mental health, and productivity [3–5] placing a large and growing burden on the healthcare system [6]. Therapy options for chronic pelvic pain and LUTD are typically grouped on their level of invasiveness including non-invasive, pharmacologic, and minimally invasive therapies like botulinum toxin injections and sacral neuromodulation [7].

When patients fail to respond to traditional therapies they are left with few options. Pudendal nerve stimulation has emerged as a minimally invasive off-label therapy for pelvic pain and LUTD, adding an option for people who fail to respond to other therapies [8]. While pudendal neuromodulation most often improves symptoms, outcomes can vary and not all patients gain complete relief [9–12]. We need to know which factors are critical for nerve recruitment and treatment success, to further improve patient benefit from pudendal neuromodulation.

Typically, the neurostimulator implant procedure involves two surgical stages. In the stage-1 implant surgery, the stimulation lead is placed near the main trunk of the pudendal nerve and the pulse generator is placed in a later stage-2 implant surgery. Anal sphincter recruitment in response to pudendal stimulation during the stage-1 surgery is a key indicator that the lead is correctly positioned and recruiting fibers in the main trunk of the pudendal nerve [13]. However, to our knowledge, no previous study has investigated what factors influence recruitment of the anal sphincter. Our primary objective was to determine which stimulation parameters and physiological factors have a significant impact on the recruitment of the inferior rectal branch of the pudendal nerve during pudendal nerve stimulation.

Materials and Methods

Approval for this study was obtained from the Institutional Review Board (HUM00165005), and was registered on ClinicalTrials.gov (NCT04236596) before commencement of study procedures.

We recruited patients receiving pudendal neuromodulation for the treatment of bladder dysfunction and/or pelvic pain. This was a single center study with all surgeries performed by the same physician. After providing informed consent, participants completed clinically validated surveys to assess bladder health (American Urological Association Symptom Index: AUASI [14] and Michigan Incontinence Symptom Index: M-ISI [15]), sexual function (Female Sexual Function Index-6: FSFI-6 [16] or Sexual Health Inventory for Men: SHIM [17]), bowel function (Colorectal-Anal Distress Inventory-8: CRAD-8 [18]), and pelvic pain (Female Genitourinary Pain Index:FGUPI [19] or Male Genitourinary Pain Index: MGUPI [19]). Just before or after the final post-op clinic visit participants completed the same set of surveys, and we calculated survey score changes. Participants underwent a pre-surgery pelvic magnetic resonance imaging (MRI) scan, their scheduled stage-1 implant surgery, and a pelvic computed tomography (CT) scan at least three weeks after the stage-2 surgery. Individuals were excluded from the study if they were pregnant or planned to be pregnant during the study or had a neurogenic bladder. Participants were recruited from September 2020 to June 2023. We targeted a sample size of 20 participants based on the expected number of pudendal implants in our clinical practice within that timeframe and similar human subject neuromodulation studies [20].

Data was collected during the stage-1 surgery, in which an Interstim (Medtronic, MN, USA) electrode lead was inserted near the pudendal nerve, as described in Peters 2013 [21], using fluoroscopy and external anal sphincter (EAS) electromyography (EMG) to confirm lead position. Monopolar needle EMG electrodes (Ambu, MD, USA), were placed in the external anal sphincter at the 3 and 9 o'clock positions. The needle electrodes were connected to a Cascade Elite (Cadwell, WA, USA) intraoperative neurophysiological monitor on which the EAS EMG data was recorded with a 333 hertz (Hz) sampling rate. The EMG data was filtered by the Cascade system with 3 kHz low pass and 10 Hz high pass filters, and a 60 Hz notch filter when 60 Hz noise was observable.

After lead placement, 10-second stimulation trials were performed in which the stimulation amplitude, pulse width (among 60, 210, and 450 μs), and electrode combinations were varied. The implanted lead had four electrodes (E0 to E3) that could be activated individually (monopolar referenced to a surface electrode on the ankle) or used in any bipolar combination once the external pulse generator was connected. Stimulation was performed at either 3 Hz for monopolar stimulation or 3.1 Hz for multipolar stimulation. For each electrode and pulse width combination, threshold

amplitude was defined as the lowest stimulation amplitude that elicited an EAS contraction. Stimulation trials were performed at and above the threshold amplitude to a maximum of 5.5 mA. Charge per pulse was calculated by multiplying the applied current by the pulse width for a given trial.

We used volumetric imaging from the pre-surgery MRI (3T T2 weighted pelvis scan; Ingenia 3T system, Philips, Amsterdam, the Netherlands) and post-surgery helical CT (L4 vertebra to the proximal femur shaft with a slice thickness of 0.625 mm; Discover CT750 HD or Revolution scanner, GE Healthcare, IL, USA) scans to create patient-specific models of the pelvic anatomy. We used Mimics Research(v25.0; Materialise, Leuven, Belgium) and the Vascular Modeling Toolkit (v1.4.0; http://www.vmtk.org/) to identify and segment the pudendal nerve in the MRI scans and the implanted lead artifacts in the CT scan, and co-registered surface segmentations derived from the MRI and CT scans to align the lead and nerve. We recreated the lead geometry from the CT artifacts using the geometry functions of Comsol Multiphysics (v 5.6; Comsol, Stockholm, Sweden). We calculated the center-to-center distances between each electrode on the implanted lead and the pudendal nerve in Python (v3.7.0).

For each stimulation trial, we calculated the average area under the curve (AUC) of the rectified EAS EMG for five stimulation events using MATLAB (v2021a, MathWorks, Natick, Massachusetts, USA). This value was used to quantify the magnitude of the anal sphincter contraction, as done previously [22–24]. The AUC calculation interval started 10 milliseconds after the stimulation artifact onset and continued for 100 milliseconds (Supplemental Figure 1).

We performed all statistical analysis with R Statistical Software (v4.4.1; R Core Team 2024). We used a Shapiro-Wilk test to check for normality of each data set. A Kruskal-Wallis test was used to determine if groups had equal medians followed by a pairwise Mann-Whitney U test with a Bonferroni correction to determine statistically significant differences between groups. Where relevant, values are given for the median and interquartile range (IQR) or average and standard deviation. A significance level of $\alpha = 0.05$ was used. We used linear mixed models to estimate the EAS AUC and the threshold current based on participant age, sex, body mass index (BMI), lower urinary tract symptoms (overactive bladder (OAB), urinary retention, urinary incontinence, and pelvic pain), and stimulation parameters (stimulation amplitude, pulse width, electrode-to-nerve distance). We only included participants who had a lead migration distance of 5 mm or less in models incorporating electrode-to-nerve distance data, based on a comparison of each participant's stage-1 fluoroscopy images and CT images. We chose this distance because adjacent electrodes on the lead are roughly 5 mm apart. We used the R packages *lme4* and *lmerTest* to calculate the linear mixed model parameters. In the model, pulse width is a 3-level categorical variable with a 60 µs pulse width used as the baseline condition. Nerve-to-electrode distance, stimulation amplitude, age, and BMI were used as continuous variables with all other variables included as yes/no categoricals. ANOVA was used to compare models and determine which fixed effects to include. Participants were classified as having OAB, retention, incontinence, or pelvic pain based on the diagnoses given by their examining urologist. Diagnoses of urge urinary incontinence, stress urinary incontinence, or mixed urinary incontinence were all classified as incontinence. Diagnoses of pelvic pain, pudendal neuralgia, vulvodynia, and interstitial cystitis were all classified as pelvic pain in the statistical analysis.

Results

Of the 23 participants eligible for study inclusion, seven did not complete intraoperative data collection because of issues with insurance approval, unsuccessful device implantation, or study withdrawal and were not included in this analysis. Patient information for the included sixteen participants is summarized in Table 1. One female participant had their implanted lead removed and was reenrolled.

The monopolar EAS threshold current for 210 and 450 µs pulse widths were not normally distributed (p < 0.01) and had a median of 1.90 (IQR = 1.2), 0.75 (IQR = 0.45), and 0.55 (IQR = 0.45) mA for a pulse width of 60, 210, and 450 µs, respectively. The three pulse widths had significantly different (p = 0.001) median threshold levels, with the threshold current level at 60-µs significantly higher than 210 µs (p = 0.005) and 450 µs (p = 0.004) (Figure 1A). Similarly, the charge per pulse at EAS threshold was not normally distributed (p < 0.001). The median charge required for 60, 210, and 450 µs pulse widths was 114.0 (IQR = 72.0), 157.5 (IQR = 94.5), and 247.5 (IQR = 202.5) nC, respectively. These three groups did not have the same median (p = 0.001). There was a significant difference between the median charge required at anal sphincter threshold for 450 and 210 µs pulse widths (p = 0.02) and 450 and 60 µs pulse widths (p = 0.004) (Figure 1B). Supplemental Figure 2 provides individual participant monopolar threshold current and charge levels for each tested

pulse width. There was no significant difference between monopolar and bipolar thresholds (p = 0.98) (Supplemental Figure 3).

In a linear mixed model incorporating all monopolar trials, the estimated AUC was significantly impacted by stimulation current, stimulation pulse width, and diagnosis of OAB or urinary retention. The model estimate for each statistically significant fixed effect is given in Table 2. Patient demographic data (BMI, age, or sex) did not have a significant fixed effect. The relationship between EAS response and applied current across pulse widths and diagnoses can be seen in Figure 2. The EAS AUC for each diagnosed condition is shown in Supplemental Figure 4. Overall, we included 343 monopolar trials across sixteen participants in this model.

Eleven study participants (all female) had fluoroscopy and CT images that allowed for identifying lead migration distances. Five participants (all female) had a lead migration distance less than 5 mm, and their corresponding electrode-to-nerve distances are shown in Figure 3. We created two linear mixed models to investigate the impact of stimulation parameters on EAS AUC and EAS threshold for participants with valid electrode-to-nerve distances. The estimates for the fixed effect of applied current, pulse width, and electrode to nerve distance on EAS AUC were all statistically significant and are shown in Supplemental Table 1. Additionally, when controlling for other variables, each millimeter between the nerve and electrode decreased the expected EAS AUC by 0.27 µV*s (95% confidence interval [CI] = -0.41 to -0.13; p = 0.0011). A second mixed model showed that electrode-to-nerve distance and pulse width both significantly affect EAS threshold (Supplemental Table 2). Each millimeter between the nerve and electrode increased the expected threshold by 0.22 mA (95% CI = 0.15 to 0.29; p = 2.7e-7). We performed linear regressions between the average electrode-to-nerve distance and change in each survey score. There was a significant relationship (p = 0.049) between AUASI score change and average electrode-to-nerve distance but not with any other survey score change (Figure 4).

Discussion

In this study, we examined the influence of key patient data and stimulation parameters on pudendal nerve recruitment. To our knowledge, this is the first study which investigates these relationships in human subjects.

A pulse width of 210 µs is the current standard of care for pudendal neuromodulation but there is little documentation supporting this choice, particularly in human subjects. We show that when compared to 60 µs and 450 µs, a pulse width of 210 µs balances the minimum required current (Figure 1A) and charge (Figure 1B) and therefore requires less energy to recruit the pudendal nerve. These results agree with the findings of an earlier study [25] that showed a pulse width of 200 µs minimizes the energy required to activate afferent fibers of the pudendal nerve and support the use of common sacral neuromodulation stimulation parameters in pudendal neuromodulation.

We estimated the relationship between EAS activation and age, BMI, and LUTD diagnosis with a linear mixed model. Age, sex, BMI and diagnosis of incontinence or pelvic pain did not significantly affect anal sphincter recruitment. However, diagnosis of OAB or retention significantly reduced the EAS activation based on 95% confidence interval testing (Table 2). Participants with diagnoses of OAB or retention had lower EAS activation across all applied currents (Figure 2). If true, these results would suggest that people with these diagnoses might require more current than is typical to activate the EAS and recruit the pudendal nerve. However, it is important to note that there was an unbalanced distribution of diagnoses in our study population (Supplemental Figure 4) and diagnoses of neither OAB nor retention (n=2) or both OAB and retention (n=1) were rare, having a large effect on the observed AUC differences. Also, the model estimates for OAB and retention had large errors of 0.58 and 0.57, respectively, showing that the strength of the effect varied across participants. The model also showed that pulse width and current significantly impact EAS recruitment (Table 2) as does electrode-to-nerve distance (Supplemental Table 1). We saw that increasing the distance between the nerve and the electrode reduced the magnitude of the EAS response, agreeing with bioelectric principles. This relationship also held true when we examined the current required to activate the EAS at threshold (Supplemental Table 2).

We observed a significant relationship between electrode-to-nerve distance and AUASI scores (Figure 4), with improved bladder outcomes when the distance was minimized. We also observed that six of eleven participants had lead migration larger than 5 mm between the stage-1 implant surgery and CT imaging. Examination of lead migration was not

a primary goal of this study, however it is not commonly reported in the literature. Future studies should specifically investigate whether lead migration has a significant relationship to outcomes or changes in pelvic floor recruitment.

Generally, the goal when implanting a pudendal stimulation lead is to activate the anal sphincter with a current below 2 mA [26] with at least three electrodes [27]. In this study, 94% (15/16) of participants had EAS activation below 2 mA with a 210 μs pulse width (Supplemental Figure 2). While lead implantation at the pudendal nerve is a challenging procedure, these results show that pudendal nerve stimulation can consistently activate the EAS at low current when the lead is placed by an experienced surgeon. During the implant surgery, monopolar stimulation is used to locate the nerve and adjust the position of the foramen needle or lead. However, most pudendal neuromodulation patients use bipolar electrode combinations to manage their symptoms at home. In our data, we did not see a significant difference in the threshold current for monopolar trials and bipolar trials (Supplemental Figure 3) supporting the continued use of monopolar stimulation to confirm intraoperative lead location.

It's important to note some limitations in this study. Evoked EAS responses due to stimulation were monitored live and identified manually, leaving room for human judgement errors and misidentification of threshold current. However, as the amplitude step interval was only 0.1 mA, we expect any threshold errors were minimal. The relative depth of the EAS needles was not tracked across procedures. In general, the surgeon seeks to place the needles as deep as possible with a consistent position on opposite sides of the sphincter, however it is possible one or both needles shifted during a surgical procedure which could affect the EMG recording. There is an unbalanced grouping in the number of participants with OAB and the gender ratio of participants with OAB that limits the conclusions we can make in this study. Additional studies should be done with larger, balanced samples sizes and diagnoses to investigate the relationship between EAS recruitment and OAB diagnosis.

In conclusion, a 210 μs pulse width is commonly used during pudendal neuromodulation [25], but this value has not been clearly justified in humans with current stimulation technologies. This study highlights the benefits of a 210 μs stimulation pulse width, as compared to 60 and 450 μs, and provides support for its continued use in pudendal neuromodulation. Pudendal nerve stimulation is effective at activating the pudendal nerve and EAS in patients with a wide range of LUT symptoms and backgrounds. Minimizing the distance between the nerve and electrode is critical for recruitment of the pudendal nerve. Larger distances require more current to activate the nerve and may therefore reduce the battery lifespan of implanted devices and reduce the size of the therapeutic window. Future studies should examine how bipolar stimulation affects pudendal nerve recruitment at different distances from the electrode to the nerve and look deeper at how LUT dysfunction may affect recruitment of the EAS with pudendal nerve stimulation.

Acknowledgments

We thank Vanessa Pruitt and Mackenzie Moore for their contributions as study coordinators and Gaurang Shah for his help in reviewing MRI scans. We also thank Kelly Ridenour and Stephanie Jones for their EMG identification expertise in the operating room. This project was funded by an award from the National Institutes of Health SPARC program (OT2OD028191).

References

1. Gracely A, Gupta P. Pudendal neuromodulation for pelvic pain. Curr Bladder Dysfunct Rep. 2020 Sep;15(3):113–20.
2. Marinkovic SP. New technologies in the management of overactive bladder: current research and future prospects. Ther Adv Urol. 2019 Jan;11:1756287219844669.
3. Coyne KS, Wein AJ, Tubaro A, Sexton CC, Thompson CL, Kopp ZS, et al. The burden of lower urinary tract symptoms: evaluating the effect of LUTS on health-related quality of life, anxiety and depression: EpiLUTS. BJU Int. 2009 Apr;103:4–11.
4. Eapen R, Radomski S. Review of the epidemiology of overactive bladder. Res Rep Urol. 2016 Jun;71.
5. Huang G, Le AL, Goddard Y, James D, Thavorn K, Payne M, et al. A Systematic Review of the Cost of Chronic Pelvic Pain in Women. J Obstet Gynaecol Can. 2022 Mar;44(3):286-293.e3.
6. Puckrein G, Walker D, Xu L, Congdon P, Gooch K. The Prevalence and Forecast Prevalence of Overactive Bladder in the Medicare Population. Clin Med Insights Urol. 2019 Jan;12:117956111984746.
7. Cameron AP, Chung DE, Dielubanza EJ, Enemchukwu E, Ginsberg DA, Helfand BT, et al. The AUA/SUFU Guideline on the Diagnosis and Treatment of Idiopathic Overactive Bladder. J Urol. 2024 Apr 23;10.1097/JU.0000000000003985.
8. Peters KM, Killinger KA, Boguslawski BM, Boura JA. Chronic pudendal neuromodulation: Expanding available treatment options for refractory urologic symptoms: Chronic Pudendal Neuromodulation. Neurourol Urodyn. 2010 Sep;29(7):1267–71.
9. Jottard K, Bruyninx L, Bonnet P, Mathieu N, De Wachter S. Pilot study: pudendal neuromodulation combined with pudendal nerve release in case of chronic perineal pain syndrome. The ENTRAMI technique: early results. Int Urogynecology J. 2021 Oct;32(10):2765–70.
10. Peters KM, Killinger KA, Jaeger C, Chen C. Pilot Study Exploring Chronic Pudendal Neuromodulation as a Treatment Option for Pain Associated with Pudendal Neuralgia: Neuromodulation and Pudendal Neuralgia. LUTS Low Urin Tract Symptoms. 2015 Sep;7(3):138–42.
11. Heinze K, Hoermann R, Fritsch H, Dermietzel R, van Ophoven A. Comparative pilot study of implantation techniques for pudendal neuromodulation: technical and clinical outcome in first 20 patients with chronic pelvic pain. World J Urol. 2015 Feb;33(2):289–94.
12. Meier KM, Vecellio PM, Killinger KA, Boura JA, Peters KM. Pudendal Neuromodulation is Feasible and Effective After Pudendal Nerve Entrapment Surgery. J Sex Med. 2022 Jun 1;19(6):995–1001.
13. Peters KM, Feber KM, Bennett RC. Sacral versus pudendal nerve stimulation for voiding dysfunction: A prospective, single-blinded, randomized, crossover trial. Neurourol Urodyn. 2005;24(7):643–7.
14. Barry MJ, Fowler FJ, O'leary MP, Bruskewitz RC, Holtgrewe HL, Mebust WK, et al. The American Urological Association Symptom Index for Benign Prostatic Hyperplasia. J Urol [Internet]. 2017 Feb [cited 2022 Jun 29];197(2S). Available from: http://www.jurology.com/doi/10.1016/j.juro.2016.10.071
15. Suskind AM, Dunn RL, Morgan DM, DeLancey JOL, McGuire EJ, Wei JT. The Michigan incontinence symptom index (M-ISI): A clinical measure for type, severity, and bother related to urinary incontinence: Incontinence Symptom Index. Neurourol Urodyn. 2014 Sep;33(7):1128–34.
16. Isidori AM, Pozza C, Esposito K, Giugliano D, Morano S, Vignozzi L, et al. Original Research—Outcomes Assessment: Development and Validation of a 6-Item Version of the Female Sexual Function Index (FSFI) as a Diagnostic Tool for Female Sexual Dysfunction. J Sex Med. 2010 Mar 1;7(3):1139–46.
17. Rosen R, Cappelleri J, Smith M, Lipsky J, Peña B. Development and evaluation of an abridged, 5-item version of the International Index of Erectile Function (IIEF-5) as a diagnostic tool for erectile dysfunction. Int J Impot Res. 1999 Dec 1;11(6):319–26.
18. Barber MD, Walters MD, Bump RC. Short forms of two condition-specific quality-of-life questionnaires for women with pelvic floor disorders (PFDI-20 and PFIQ-7). Am J Obstet Gynecol. 2005 Jul;193(1):103–13.
19. Clemens JQ, Calhoun EA, Litwin MS, McNaughton-Collins M, Kusek JW, Crowley EM, et al. Validation of a Modified National Institutes of Health Chronic Prostatitis Symptom Index to Assess Genitourinary Pain in Both Men and Women. Urology. 2009 Nov;74(5):983-987.e3.
20. George AT, Dudding TC, Gurmany S, Kamm MA, Nicholls RJ, Vaizey CJ. Pudendal Nerve Stimulation for Bowel Dysfunction in Complete Cauda Equina Syndrome. Ann Surg. 2014 Mar;259(3):502–7.
21. Peters KM. Pudendal Neuromodulation for Sexual Dysfunction. J Sex Med. 2013 Apr;10(4):908–11.

22. Gregory WT, Lou JS, Stuyvesant A, Clark AL. Quantitative Electromyography of the Anal Sphincter After Uncomplicated Vaginal Delivery: Obstet Gynecol. 2004 Aug;104(2):327–35.

23. Stålberg E, van Dijk H, Falck B, Kimura J, Neuwirth C, Pitt M, et al. Standards for quantification of EMG and neurography. Clin Neurophysiol. 2019 Sep;130(9):1688–729.

24. Su X, Cutinella M, Koppes S, Agran JE, Dinsmoor DA. Electromyographic Responses Across Different Pulse-Widths of Sacral Neuromodulation in Sheep. Neuromodulation Technol Neural Interface. 2019 Aug;22(6):684–9.

25. Plevnik S, Vodušek DB, Vrtačnik P, Janež J. Optimization of pulse duration for electrical stimulation in treatment of urinary incontinence. World J Urol. 1986 Aug;4(1):22–3.

26. Vodušek DB, Plevnik S, Vrtačnik P, Janež J. Detrusor inhibition on selective pudendal nerve stimulation in the perineum. Neurourol Urodyn. 1987;6(5):389–93.

27. Gilleran JP, Alpert SA, editors. Adult and Pediatric Neuromodulation [Internet]. Cham: Springer International Publishing; 2018 [cited 2021 Jun 11]. Available from: http://link.springer.com/10.1007/978-3-319-73266-4

Tables

Table 1 Summary of participant characteristics
Average and (standard deviation) given when relevant.

Variable	Count
Sex	
Female	13
Male	3
Diagnosed Condition	
Overactive Bladder	12
Retention	5
Incontinence	7
Pelvic Pain	12
Age (years)	55.9 (10.8)
BMI (kg/m^2)	28.4 (3.8)

Table 2 Linear mixed model estimating anal sphincter activation:
Summary of the significant fixed effects for the linear mixed model estimating EAS AUC based on all monopolar trials. A representative equation for the corresponding linear model is shown in equation S1.

Fixed Effect	Estimate (µV*s)	Standard Error	95% Confidence Interval	p value
Intercept	2.73	0.69	(1.38, 4.08)	4.67e-4
Current	0.79	0.10	(0.60, 0.98)	1.5e-14
Pulse Width 210 µs	1.91	0.35	(1.23, 2.61)	8.27e-8
Pulse Width 450 µs	2.38	0.39	(1.60, 3.15)	4.22e-9
OAB yes	-2.79	0.58	(-3.91, -1.66)	2.66e-4
Retention yes	-2.28	0.56	(-3.39, -1.17)	8.94e-4

Figures

Figure 1: Characterization of external anal sphincter EMG activation by intraoperative monopolar pudendal nerve stimulation. A) The current required to activate the EAS at threshold depending on pulse width. B) Charge per pulse at the EAS activating current depending on pulse width. Box plots represent the IQR for the height of each box, with the bold central line indicating the median. Tips of whiskers extend to the value(s) within 1.5 x IQR of each box side. * indicates $p < 0.05$, ** indicates $p < 0.01$.

Figure 2: Anal sphincter response magnitude plotted against applied current for all monopolar trials. Data is separated by participant diagnosis.

Figure 3: Measured distances (in millimeters) between each electrode on the implanted quadripolar lead and the pudendal nerve for five participants. Electrodes are labeled from E0 to E3 with E0 representing the electrode most distal to the implantable pulse generator and E3 representing the electrode most proximal.

Figure 4: Linear regression fits between participant electrode-to-nerve distance and symptomatic survey score percent change. Surveys on asymptomatic functions (CRAD-8, FSFI) did not have relationships to electrode-to-nerve distance (p = 0.606, 0.706, respectively).