Accurate and cost-efficient whole genome sequencing of hepatitis B virus

using Nanopore

Joakim B. Stenbäck[1,2], Daniel Schmidt[3], Ulrika Noborg[1], Joel Gustafsson[1], Peter Norberg[1],

Maria E. Andersson[1,2], Michael X Fu[4], Heli Harvala[5] and Johan Ringlander[1,2]

[1]Department of Infectious Diseases, Institute of Biomedicine, University of Gothenburg,
Sweden
9

10 [2]Department of Clinical Microbiolgy, Sahlgrenska University Hospital, Gothenburg, Sweden
11

12 [3]Bioinformatics and Data Centre, University of Gothenburg, Sweden
13

14 [4]Nuffield Department of Medicine, University of Oxford, United Kingdom
15

16 [5]Radcliffe Department of Medicine, University of Oxford, United Kingdom
17

Abstract

Deep sequencing of the whole hepatitis B virus genome increases the analytical resolution

and has the potential to improve molecular epidemiology investigations. The aim of this work

was to develop and evaluate the performance of such deep sequencing using the Nanopore

technology. The method includes an initial PCR step to generate two overlapping amplicons

that cover the whole relaxed circular HBV genome found in circulating viral particles and

covalently closed circular DNA in infected hepatocytes, followed by sequencing using the

Nanopore rapid barcoding kit that allows parallel analysis of several samples in one reaction.

The libraries can be sequenced with the standard Nanopore flow cell on MiniIon or GridIon

devices, as well as the Flongle. The performance of the method was evaluated by comparing

Nanopore and Sanger sequences or qPCR results from 64 clinical samples. The Nanopore-

derived consensus sequences were, on average, 99.9% similar to those from Sanger

sequencing and the full HBV genome was determined in samples with HBV DNA levels of

approximately 3 $\log_{10}$ IU/mL with MagNA pure 96 extraction and < 2 $\log_{10}$ IU/mL using a

high-volume manual extraction protocol on a subset of samples from patients with very low

viral load (1.62-3.74 IU/mL). A perfect agreement with Sanger/qPCR-derived genotype was

seen. The cost of sequencing per genome using the Nanopore method is low, ranging 6-

37euros. We conclude that whole-genome sequencing of HBV with Nanopore is well suited

for genomic characterization, antiviral resistance mutation analysis and genotyping of HBV in

a routine laboratory setting.

Background

Chronic infection with hepatitis B virus (HBV) causes a heavy health burden worldwide (1).

Chronic hepatitis B poses a risk for long-term complications such as cirrhosis and cancer, a

risk that can be reduced by treatment with nucleot(s)ide analogs (2). HBV has nine genotypes

(A-I) with relatively distinct geographical distribution, differing from each other by at least

8% of the ≈ 3200 nucleotides that comprise the genome. In addition, there are a large number

of subgenotypes, some of which require whole genome comparison for accurate

classification. Whole genome sequencing can also support molecular epidemiology

investigations, for example, when evaluating potential transmission between subjects infected

with the same genotype (3, 4). Identification of genotypes can be achieved by simplified

methods such as real-time PCR (5), but new sequencing techniques have become almost as

easy and allow relatively fast genotyping at a relatively low cost. Also, the short circular HBV

genome enables full genome coverage by only two overlapping PCRs. Genotyping has

clinical applicability since infection with genotype C is a risk factor for liver complications,

and infection with genotype A is more susceptible to interferon treatment (6).

Although antiviral resistance against tenofovir and entecavir is relatively rare, resistance

analysis is quite often of interest in patients with poor therapeutic response. Entecavir

resistance typically occurs after previous treatment with lamivudine (7, 8). Resistance to

tenofovir is very rare, but mutations leading to substitution 236T in reverse transcriptase, in

combination with other mutations, have been suggested to lead to antiviral resistance (9, 10).

Nanopore sequencing has been widely used during the COVID-19 pandemic (11-14), and this

technique is also increasingly used in bacteriology for antimicrobial resistance genotyping

(15), 16s and whole-genome sequencing (16, 17). Many clinical microbiological laboratories

that previously did not perform sequencing are now in possession of one or multiple

Nanopore instruments (18). Apart from whole-genome sequencing of SARS-CoV-2 (19),

methods for many other microbes have been launched (20-22). The relatively low upfront cost

for the device, low cost of sequencing and simple protocols are benefits that have been

described. On the other hand, a higher error rate, the need for high computer calculation

capacity and specialized bioinformatics may discourage the introduction of the Nanopore

method in a routine laboratory. In this article, we describe the development of a user-friendly

method for Nanopore sequencing of HBV and a coherent bioinformatical pipeline adapted for

clinical use in routine laboratories.

Methods

Patients and samples

Patient samples from three different cohorts were part of this study. Cohort 1 consisted of 38

HBV positive sera from individual patients that had been analyzed with a Sanger sequencing-

based resistance testing method as part of routine diagnostics at the Department of Clinical

Microbiology, Sahlgrenska University Hospital, Gothenburg. One sample from this cohort

was quantified with qPCR and serially diluted to assess the analytical sensitivity. Cohort 2

was made out of 19 sera that had previously been genotyped using qPCR (5) in routine

diagnostics, also at Sahlgrenska University Hospital. Cohort 3 consisted of 7 plasma samples

from blood donor screening that all had low viral load and previously had been quantified

using ultrasensitive PCR and genotyped using a Sanger sequencing-based method in the

United Kingdom, as previously described (23) . Details on the number of samples, genotype

and viral load are found in Table 1.

Nucleic acid extraction and HBV DNA quantification

For cohorts 1 and 2, which had been previously analyzed in routine diagnostics at the

Sahlgrenska University Hospital, Gothenburg, Sweden, total NA extraction from 200 uL of

blood/plasma was performed using the MagNA pure 96 (Roche, Basel, Switzerland),

according to the manufacturer's instructions. For the low viral load panel (cohort 3), large-

volume extraction from 5 mL of plasma was made using a modified protocol for the Roche

High Volume NA extraction kit, as previously described (23). Four samples in cohort 3 were

analyzed using both extraction methods for comparison. For cohorts 1 and 2, HBV DNA

quantification was performed using the Cobas HBV test on the 6800 platform (Roche)

according to the manufacturer's instructions; for cohort 3, quantification was performed as

previously described (23).

Serial dilutions of one sample with an original concentration of 5.34 $\log_{10}$ IU/mL were

sequenced with Nanopore to further assess the sensitivity. The sample was diluted in H_2O to

1:10, 1:100, 1:1000 and 1:10000. Only the first sets of primers for the whole genome method

was used and the nested PCR covering the RT region was not applied. The dilutions were

sequenced on a flongle flow cell.

PCR and Nanopore sequencing

PCRs intended to generate two overlapping amplicons covering the whole genome were

designed. The primers used for amplicon 1 were 56F (CCTGCTGGTGGCTCCAGT) and

1827R (GAAAAAGTTGCATGGTGCTGGT), and for amplicon 2 primers 1689F

(ACCGACCTTGAGGCCTACTTCA) and 262R (CCACCACGAGTCTAGACTCT). Primers

were manually designed based on an alignment with whole genomes of genotypes A-G.

Mismatches in primers targets were allowed and the sequences were chosen to preferably

target A-E, the most common genotypes in Sweden and the UK. Identical settings and

reagents were used in the two amplifications. To the reaction, 5 µL of sample and 1 µL of

each primer in 10 µM dilutions were added to 10 µL of master mix containing Hot Start PCR

mix (2X) (ThermoFisher, Waltham, MA, USA) and 10 µL H_2O. A denaturation step at 95°C

was included prior to 40 cycles of PCR, with annealing at 60°C for 10 s and elongation at

72°C for 2 min.

Capillary gel electrophoresis was performed on a TapeStation 4200 (using the DT5000 kit,

Agilent, Santa Clara, CA, USA), and if no visible bands were seen, a nested PCR with inner

primers 251F (GTGGTGGACTTCTCTCAATTTTC) and 1801R

(CAGACCAATTTATGCCTACAGCCT) was performed with the same PCR settings as

described in the first PCR. In this study, only the amplicon covering the reverse transcriptase

(RT) region was nested.

Amplicons were barcoded, and libraries were prepared with the Rapid Barcoding Kit 96

SQK-RBK110.96, attaching barcodes and adapters randomly on the amplicons using

transposase (both kits from Oxford Nanopore Technologies, Oxford, UK). Samples were

pooled and sequenced with the MinIon sequencing device (Oxford Nanopore Technologies)

and the R10.4.1 flowcell (FLO-MIN114) or an R10.4.1 flongle (FLO-FLG114) (Oxford

Nanopore Technologies). Cohort 1 samples were sequenced five at a time on a flongle

flowcell to mimic the planned use in routine diagnostics. Cohort 2 and 3 samples were

sequenced on one standard flowcell each, and reruns for reproducibility testing of cohort 3

samples were performed using the flongle. One or two negative controls (H2O) were used in

PCR and subsequent Nanopore runs and bioinformatic analysis.

This protocol is available at protocols.io (*https://www.protocols.io/view/hbv-whole-genome-*

sequencing-using-oxford-nanopore-n92ldrow9g5b/v1).

Sanger sequencing

Parts of the reverse transcriptase (RT) region was sequenced with the Sanger method. For

cohort 1, Sanger sequencing was performed after PCR with primers 593F

(TGCACCTGTATTCCCATCCCATC) and 1010R (GRGCAGCAAADCCCAAAAGACC),

if no band was seen, semi-nested PCR was performed with 635F

(TTCCTATGGGAGTGGGCCTC) and the same reverse. For cohort 3, parts of the RT region was

covered by two amplicons and Sanger sequenced, as previously described (23).

Bioinformatics

Base calling of Nanopore fast5 files was made with the GPU-accelerated guppy program

(https://github.com/asadprodhan/GPU-accelerated-guppy-basecalling) with super accurate

mode settings. Subsequently, fastq files were run through a pipeline including following

steps: Reads were mapped to HBV reference genomes representing genotypes A-H using

Minimap2. Draft consensus sequences were created using samtools and polished with Medaka

(Oxford Nanopore Technologies) and the consensus sequence from the best matching

genotype was chosen. Minor variant analysis was performed with freebayes (24) and a an

inhouse script was used to call resistance mutations based on those listed at geno2pheno

(https://hbv.geno2pheno.org/). Samples were assigned as passed or failed, the latter if fewer

reads than the negative control. The pipeline is available at github

(https://github.com/ClinicalGenomicsGBG/hbv_nano). The Stanford HBV database

(https://hivdb.stanford.edu/HBV/HBVseq/development/hbvseq.pl?action=showSequenceFor

m) and geno2pheno tool (https://hbv.geno2pheno.org/) were used for genotyping and calling

antiviral resistance mutations from consensus sequences and to compare sequence quality in

the RT region, and showed consistent results with the pipeline. Splicing analysis was made

using FLAME (25).

Results

Antiviral resistance analysis and genotyping

In cohort 1, 18/38 samples had previously detected antiviral resistance mutations by Sanger

sequencing. The same mutations were found using Nanopore, with the addition of a mutation

causing the resistance-associated amino acid substitution 80I in one sample, that only was

found with the Nanopore method, as this position was not covered by the Sanger amplicon.

Samples with resistance mutations had a viral load of 5.0 $\log_{10}$ IU/mL in mean (range 3-7.92

$\log_{10}$ IU/mL) All minor variants found with the Sanger-based analysis were also detected by

and the antiviral resistance mutation with the lowest frequency was present in 25% of the

Nanopore reads. All samples but four could be analyzed regarding antiviral resistance without

performing nested PCR and those had low viral loads ranging 2.4-3.37 $\log_{10}$ IU/mL. Thus, the

clinical sensitivity for antiviral resistance mutations was 100%. There were no antiviral

resistance-associated mutations in cohorts 2 and 3, that previously not had been analyzed for

resistance mutations.

In cohort 1,2 and 3, genotyping using Sanger or qPCR (cohort 2) did not differ from the

Nanopore results. HBV genotypes A-F were represented among the samples (table 1). The

number of samples with mismatches in the Nanopore method primer targets was low and

typically only in one position. However, the four samples with an unsuccessful whole genome

sequencing might have had more mismatches in primers 56F and 1827R, that could not be

detected. Sequencing of the RT region of these samples after nested PCR was successful and

showed a lack of mutations in 1689F and 262R primer targets for all but the one sample with

genotype F that had multiple differences.

Sensitivity and reproducibility

In cohort 1, the entire HBV genome was obtained in all samples >3.27 log10 IU/mL, that was

33/38 samples. In the one sample with genotype F, the whole amplicon covering the RT

region, but not the whole genome, was obtained after the first PCR, thus, nested PCR was not

performed. The remaining four samples (all ≤3.27 log10 IU/mL with genotype D (n=3) and

genotype B (n=1)) were subject to nested PCR targeting the region spanning nt 251-

1801(including the RT), and the entire region was obtained in all samples.

A sample with an original concentration of 5.34 $\log_{10}$ IU/mL was diluted and sequenced.

Sequences were identical in dilutions down to 1:100 but the number of reads in 1:1000 and

1:10000 were insufficient for antiviral resistance analysis and assembly of a whole genome.

At dilution 1:100, the full genome was not captured but the whole RT region was, which was

identical to the RT region of the 1:1 and 1:10 dilutions. Thus, the analytical sensitivity of the

method, using the standard MagNA pure 96 extraction and when five samples were pooled

and run on one flongle flow cell, was approximately 3.34 $\log_{10}$ IU/mL. Table 2 shows results

from the sequencing of a serially diluted sample. In cohort 2 all samples were successfully

whole genome sequenced. For cohort 3, consisting of seven samples, of which four samples

were extracted with both the automated MagNA pure 96 protocol and the large-volume

manual extraction protocol, the full HBV genome was successfully Nanopore sequenced after

manual extraction in all samples. The full genome was captured in 3/4 of the samples after

MagNA pure 96 extraction, and in the fourth sample, only the amplicon covering the RT

region showed reads. The three successful samples had viral loads of 3.43, 3.27 and 1.74 $\log_{10}$

IU/mL, respectively, and the remaining sample had a viral load of 1.62 log IU/mL. Two of the

four high volume extracted samples in cohort 3 (with viral load 3.37 and 1.62 $\log_{10}$ IU/mL)

were run on both the flongle R10.4 flowcell (FLO-FLG114) and the regular R10.4 flowcell,

(FLO-MINI 114) to investigate reproducibility, and consensus sequences were identical from

both flowcells.

Sequence similarity

In cohort 1, the sequence similarity between Nanopore and Sanger was 99.9% (range 99.8-

100%). Cohort 2 had not been Sanger sequenced. In cohort 3, the Sanger sequences covering

the variable part of the RT were compared with the same regions in the Nanopore-derived

sequences and showed a 99.8% similarity in mean (99.13-100%). Differences from Sanger

were mainly caused by ambiguous bases in the Sanger-derived sequence.

HBV splice variants and core deletions

Splicing at conventional splice sites was not detected in frequencies ≥10% threshold in any of

the samples, however, reads suggesting splicing was found in 29 samples when the frequency

threshold was lowered to 1%. Some of the HBV splice sites may affect the PCR target, most

importantly the nt 2447-489 splicing junction, and could therefore potentially not be detected.

Detection of all known preS and preCore deletions (26, 27) is however possible. Notably, in

cohort 3, consisting of low viral load samples, one sample showed a preS1 deletion spaning nt

2847-3060, as the predominant genome variant, and one sample harboring an nt 2994-3054

deletion as a minor variant in approximately 25% of the reads (figure 2). No deletions were

seen in cohorts 1 and 2.

Nanopore sequencing cost

Nanopore sequencing with the method described in this article can be performed with up to

96 samples per flow cell run or approximately twelve samples using the flongle flow cell.

With current costs for conventional extraction and PCR (4€) and Nanopore reagents (R10.4.1

flow cell from 455€, flongle flow cell from 66€ and rapid barcoding kit cost per total

barcoded library of 99€) the reagent cost for each run is in the range of 169-558 €, thus for

each genome in the range of 6-37€, depending on flow cell used and the number of samples

pooled.

Discussion

Nanopore whole-genome sequencing of HBV was fast, sensitive and accurate, with 99.9%

sequence agreement with Sanger sequencing. By introducing barcodes with the Nanopore

rapid barcoding method, antiviral resistance, genotype and sequence analyses can be achieved

at low cost. The main advantages of Nanopore are its simplicity and its feature to perform

long-read sequencing. By amplifying and sequencing only two amplicons with lengths of

approximately 1771 nt (nt 56 to 1827) and 1773 nt (nt 1689 to 262), respectively, the whole

circular HBV genome was captured. The primer regions were chosen to obtain continuous

reads for the RT, precore-core region and for the regions potentially involved in vaccine

escape and antiviral resistance (28-30).

For this study, we used the Rapid Barcoding kit 96 V14 but alternative Nanopore kits are also

possible. The ligation sequencing-based kits were tested during the method development but

tend to be more laborious and time consuming, however, the amplicons remain intact and are

sequenced as one continuous read, which could be preferable in haplotype and structural

variation analysis. Despite this transposase effect, we believe that the rapid barcoding kit is

best suited for routine diagnostics, as it demands little hands-on time and may be used for

sequencing and pooling of several different microbes simultaneously, which may be either

PCR-enriched or cultured. The same chemistry has also been used for SARS-CoV-2

sequencing and has proven to be scalable to up to 96 samples in one run (31).

We tested two different nucleic acid extraction protocols in this study, and the modified

protocol for the manual high-volume kit from 5 mL of plasma by Fu *et al.* (23) proved to be

useful in low viral load samples, and full genome coverage was achieved from samples as low

as 1.62 log10 IU/mL, that could not be achieved with MagNA pure 96 extraction from 200

microliters of serum. Presumably, the high-volume kit yielded higher levels of intact DNA

and, therefore, was preferable for long-read sequencing. Serial dilution of one sample showed

that for concentrations ≥3.43 log10 IU/mL, Nanopore sequencing of MagNA pure 96 (Roche)

extracts showed reproducible results and allowed detection of resistance mutations. However,

sequencing of the whole RT gene was achieved in all samples of cohort 1 and in contrary to

the serial dilutions, full genomes were captured in eight samples with viral load <3.43 log10

IU/mL also after MagNA pure 96 extraction. The seemingly lower sensitivity in the serial

dilution run compared to the other Nanopore runs might be explained by the differences in

viral concentrations, causing amplicons from higher concentration dilutions to outcompete

those with lower in the pooled reaction. The sensitivity could possibly be increased if samples

with high concentrations are diluted, but this did not seem necessary in the other samples and

runs for which sufficient results were obtained.

We applied the GPU-based super accurate basecalling for fast yet high-quality generation of

reads. Nanopore basecalling has improved immensely in recent years and is the main factor

for lowered error rate and several alternative methods have been presented (32-35). In a

previous study, we used a pipeline for consensus sequence extraction from NGS data and

almost perfect sequences were achieved, but some false insertion/deletions (indels) remained

(3). To overcome this issue in more error prone long-reads data, the Oxford Nanopore

Technologies tool Medaka was used to create a consensus sequence from Minmap2-mapped.

The implementation of the Medaka tool removed false indels, as previously shown for SARS-

CoV-2 (19). Medaka creates consensus sequences based on the major virus variant and for

minor variant analysis, we applied freebayes (24). In our analysis of antiviral resistance all

variants that previously had been found using Sanger sequencing were found and no false

positives were detected applying a frequency threshold of 10%. None of the samples in the

study had resistance mutations constituting <25% of the viral population, supporting the

threshold used. Previous studies of Nanopore error-rate suggest that this could be lowered to

3% (36), something we aim to investigate in future studies.

Previous studies have presented sequencing of HBV combined with rolling-circle

amplification (37, 38). McNaughton *et al.* adapted the method for Nanopore sequencing and

minor variant analysis (38). Their approach with rolling-circle amplification resulted in

extremely long amplicons and very long reads with the whole HBV sequence repeated,

optimal for haplotype reconstruction and high accuracy. For routine diagnostics a standard

PCR approach might still be preferable, as it presumably yields higher sensitivity (37) and

less laborious protocols (38). The method presented in this article, based on conventional

PCR, had an almost 100% agreement between Nanopore and Sanger derived consensus

genomes, allowing for clinical use. If detection of ultra-rare minor variants, or *in haplo*

mutations, is needed for research purposes, we believe that the method presented by

McNaughton *et al*, Ion Torrent or Illumina-based methods, might be preferable (39).

This study showed that deletions in HBV genomes can be found using long-read sequencing.

Deletions have been shown in many studies but their clinical relevance is not fully known

(40). Also, a recent article suggests that splicing is associated with poorer interferon treatment

outcomes (41). HBV genome deletions, such as those we found in the S region in two samples

in this study, have been associated with an increased risk of cirrhosis (42). We present a

method that enables large-scale use and characterization of structural genomic features, which

allows for bigger studies. Other primers than those presented here could improve the detection

of splicing variants (42-44) and we plan to study this.

Other limitations of this study include the lack of Sanger sequencing covering the entire HBV

genome meaning only RT region could be compared with Nanopore. Also, the primers for the

second amplicon in the Nanopore method, covering the S region, are not optimal for

genotypes F-J as some strains have up to three mismatches in the primer target, discovered

after the completion of this study. However, the primers for the RT amplicon appear to be

conserved in all genotypes which enable pan-genotypic resistance and genotyping. The aim of

this article was to present a method suitable routine laboratories and thus only used a nested

PCR for the amplicon covering the RT region, however, the whole genome sensitivity may be

further increased for all genotypes if additional amplicons are sequenced (45) or nesting of the

second amplicon is applied.

Our study shows that Nanopore sequencing of two overlapping amplicons is a feasible

method for HBV whole-genome sequencing, genotyping and antiviral resistance mutation

analysis in a routine laboratory setting, showing adequate sensitivity and accuracy for clinical

purposes. Minor variant analysis of clinically relevant subpopulations was possible. We
believe that this method, providing simple and cost-efficient sequencing, may contribute to
increased whole-genome sequencing of HBV and new discoveries of clinically relevant
genomic features.

Author contributions

JR, HH and UN planned and designed the study. JR, MA and JBS designed the Nanopore
method. JR and PN financed the study. JBS, UN, MA and MF performed laboratory work. JG
designed hardware and set up GPU-based basecalling. DS, JR and JBS designed and
performed bioinformatical analyses. JR drafted the manuscript. All authors contributed to the
final version of the manuscript.

Funding

This study was funded by the Gothenburg Society of Medicine and the Sahlgrenska
University Hospital.

Data availability

The data that support the findings of this study are openly available at the European
Nucleotide Archive (ENA) https://www.ebi.ac.uk/ena/browser/home. Fastq files are uploaded
with the study accession number PRJEB87509.

References

1.   Cao G, Liu J, Liu M. Trends in mortality of liver disease due to hepatitis B in China from 1990 to 2019: findings from the Global Burden of Disease Study. Chin Med J (Engl). 2022;135(17):2049-55.

2.   Liaw YF. Reduction of cirrhosis and hepatocellular carcinoma with antiviral therapy in chronic hepatitis B. Hepatology. 2013;58(5):1856.

3.   Ringlander J, Andersson ME, Prakash K, Larsson SB, Lindh M. Deep sequencing of hepatitis B virus using Ion Torrent fusion primer method. J Virol Methods. 2022;299:114315.

4.   Eilard A, Andersson M, Ringlander J, Wejstal R, Norkrans G, Lindh M. Vertically acquired occult hepatitis B virus infection may become overt after several years. J Infect. 2019;78(3):226-31.

5.   Malmstrom S, Berglin-Enquist I, Lindh M. Novel method for genotyping hepatitis B virus on the basis of TaqMan real-time PCR. J Clin Microbiol. 2010;48(4):1105-11.

6.   Malmstrom S, Eilard A, Larsson SB, Hannoun C, Norkrans G, Lindh M. Genotype impact on long-term virological outcome of chronic hepatitis B virus infection. J Clin Virol. 2012;54(4):321-6.

7.   Cassino L, Benetti S, Fay F, Tanno H, Quarleri J. Unsuccessful therapy with adefovir and entecavir-tenofovir in a patient with chronic hepatitis B infection with previous resistance to lamivudine: a fourteen-year evolution of hepatitis B virus mutations. BMC Infect Dis. 2011;11:178.

8.   Singal AK, Fontana RJ. Meta-analysis: oral anti-viral agents in adults with decompensated hepatitis B virus cirrhosis. Aliment Pharmacol Ther. 2012;35(6):674-89.

9.   Mokaya J, McNaughton AL, Bester PA, Goedhals D, Barnes E, Marsden BD, et al. Hepatitis B virus resistance to tenofovir: fact or fiction? A systematic literature review and structural analysis of drug resistance mechanisms. Wellcome Open Res. 2020;5:151.

10.   Park ES, Lee AR, Kim DH, Lee JH, Yoo JJ, Ahn SH, et al. Identification of a quadruple mutation that confers tenofovir resistance in chronic hepatitis B patients. J Hepatol. 2019;70(6):1093-102.

11.   Davis JJ, Long SW, Christensen PA, Olsen RJ, Olson R, Shukla M, et al. Analysis of the ARTIC Version 3 and Version 4 SARS-CoV-2 Primers and Their Impact on the Detection of the G142D Amino Acid Substitution in the Spike Protein. Microbiol Spectr. 2021;9(3):e0180321.

12.   Luo J, Meng Z, Xu X, Wang L, Zhao K, Zhu X, et al. Systematic benchmarking of nanopore Q20+ kit in SARS-CoV-2 whole genome sequencing. Front Microbiol. 2022;13:973367.

13.   Mannsverk S, Bergholm J, Palanisamy N, Ellstrom P, Kaden R, Lindh J, et al. SARS-CoV-2 variants of concern and spike protein mutational dynamics in a Swedish cohort during 2021, studied by Nanopore sequencing. Virol J. 2022;19(1):164.

14.   Bull RA, Adikari TN, Ferguson JM, Hammond JM, Stevanovski I, Beukers AG, et al. Analytical validity of nanopore sequencing for rapid SARS-CoV-2 genome analysis. Nat Commun. 2020;11(1):6272.

15.   Taxt AM, Avershina E, Frye SA, Naseer U, Ahmad R. Rapid identification of pathogens, antibiotic resistance genes and plasmids in blood cultures by nanopore sequencing. Sci Rep. 2020;10(1):7622.

16.   Widen J, Moren J, Molling P, Fagerstrom A, Enblad P, Eriksson BM, et al.
Diagnosis of external ventricular drainage related infections with real-time 16S PCR and
third-generation 16S sequencing. Infect Dis (Lond). 2024;56(7):521-30.
17.   Salva-Serra F, Jaen-Luchoro D, Jakobsson HE, Gonzales-Siles L, Karlsson R,
Busquets A, et al. Complete genome sequences of Streptococcus pyogenes type strain reveal
100%-match between PacBio-solo and Illumina-Oxford Nanopore hybrid assemblies. Sci
Rep. 2020;10(1):11656.
18.   Schmidt J, Blessing F, Fimpler L, Wenzel F. Nanopore Sequencing in a Clinical
Routine Laboratory: Challenges and Opportunities. Clin Lab. 2020;66(6).
19.   Arevalo MT, Karavis MA, Katoski SE, Harris JV, Hill JM, Deshpande SV, et al.
A Rapid, Whole Genome Sequencing Assay for Detection and Characterization of Novel
Coronavirus (SARS-CoV-2) Clinical Specimens Using Nanopore Sequencing. Front
Microbiol. 2022;13:910955.
20.   Zhang Y, Lu X, Tang LV, Xia L, Hu Y. Nanopore-Targeted Sequencing
Improves the Diagnosis and Treatment of Patients with Serious Infections. mBio.
2023:e0305522.
21.   Chan WS, Au CH, Leung SM, Ho DN, Wong EYL, To MY, et al. Potential
utility of targeted Nanopore sequencing for improving etiologic diagnosis of bacterial and
fungal respiratory infection. Diagn Pathol. 2020;15(1):41.
22.   Chen Y, Mao L, Lai D, Xu W, Zhang Y, Wu S, et al. Improved targeting of the
16S rDNA nanopore sequencing method enables rapid pathogen identification in bacterial
pneumonia in children. Front Cell Infect Microbiol. 2022;12:1001607.
23.   Fu MX, Simmonds P, Andreani J, Baklan H, Webster M, Asadi R, et al.
Ultrasensitive PCR system for HBV DNA detection: Risk stratification for occult hepatitis B
virus infection in English blood donors. J Med Virol. 2023;95(10):e29144.
24.   Erik Garrison GM. Haplotype-based variant detection from short-read
sequencing. arXiv. 2012(1207.3907).
25.   Holmqvist I, Backerholm A, Tian Y, Xie G, Thorell K, Tang KW. FLAME:
long-read bioinformatics tool for comprehensive spliceome characterization. RNA.
2021;27(10):1127-39.
26.   Gerken G, Kremsdorf D, Capel F, Petit MA, Dauguet C, Manns MP, et al.
Hepatitis B defective virus with rearrangements in the preS gene during chronic HBV
infection. Virology. 1991;183(2):555-65.
27.   Marinos G, Torre F, Gunther S, Thomas MG, Will H, Williams R, et al.
Hepatitis B virus variants with core gene deletions in the evolution of chronic hepatitis B
infection. Gastroenterology. 1996;111(1):183-92.
28.   Wilson JN, Nokes DJ, Carman WF. Current status of HBV vaccine escape
variants--a mathematical model of their epidemiology. J Viral Hepat. 1998;5 Suppl 2:25-30.
29.   Huang X, Lu D, Ji G, Sun Y, Ma L, Chen Z, et al. Hepatitis B virus (HBV)
vaccine-induced escape mutants of HBV S gene among children from Qidong area, China.
Virus Res. 2004;99(1):63-8.
30.   Basuni AA, Carman WF. HBV vaccine-escape variants. Methods Mol Med.
2004;95:115-24.
31.   Kalendar R, Kairov U, Karabayev D, Aitkulova A, Tynyshtykbayeva N,
Daniyarov A, et al. Universal whole-genome Oxford nanopore sequencing of SARS-CoV-2
using tiled amplicons. Sci Rep. 2023;13(1):10334.
32.   Yeh YM, Lu YC. MSRCall: a multi-scale deep neural network to basecall
Oxford Nanopore sequences. Bioinformatics. 2022;38(16):3877-84.
33.   Wick RR, Judd LM, Holt KE. Performance of neural network basecalling tools
for Oxford Nanopore sequencing. Genome Biol. 2019;20(1):129.

34.   Neumann D, Reddy ASN, Ben-Hur A. RODAN: a fully convolutional
architecture for basecalling nanopore RNA sequencing data. BMC Bioinformatics.
2022;23(1):142.
35.   Napieralski A, Nowak R. Basecalling Using Joint Raw and Event Nanopore
Data Sequence-to-Sequence Processing. Sensors (Basel). 2022;22(6).
36.   Grubaugh ND, Gangavarapu K, Quick J, Matteson NL, De Jesus JG, Main BJ,
et al. An amplicon-based sequencing framework for accurately measuring intrahost virus
diversity using PrimalSeq and iVar. Genome Biol. 2019;20(1):8.
37.   Martel N, Gomes SA, Chemin I, Trepo C, Kay A. Improved rolling circle
amplification (RCA) of hepatitis B virus (HBV) relaxed-circular serum DNA (RC-DNA). J
Virol Methods. 2013;193(2):653-9.
38.   McNaughton AL, Roberts HE, Bonsall D, de Cesare M, Mokaya J, Lumley SF,
et al. Illumina and Nanopore methods for whole genome sequencing of hepatitis B virus
(HBV). Sci Rep. 2019;9(1):7081.
39.   Ringlander J, Fingal J, Kann H, Prakash K, Rydell G, Andersson M, et al.
Impact of ADAR-induced editing of minor viral RNA populations on replication and
transmission of SARS-CoV-2. Proc Natl Acad Sci U S A. 2022;119(6).
40.   Sommer G, Gunther S, Sterneck M, Otto S, Will H. A new class of defective
hepatitis B virus genomes with an internal poly(dA) sequence. Virology. 1997;239(2):402-12.
41.   Chen J, Wu M, Wang F, Zhang W, Wang W, Zhang X, et al. Hepatitis B virus
spliced variants are associated with an impaired response to interferon therapy. Sci Rep.
2015;5:16459.
42.   Gunther S, Sommer G, Iwanska A, Will H. Heterogeneity and common features
of defective hepatitis B virus genomes derived from spliced pregenomic RNA. Virology.
1997;238(2):363-71.
43.   Maslac O, Wagner J, Sozzi V, Mason H, Svarovskaia J, Tan S, et al. Secreted
hepatitis B virus splice variants differ by HBV genotype and across phases of chronic
hepatitis B infection. J Viral Hepat. 2022;29(8):604-15.
44.   Sommer G, van Bommel F, Will H. Genotype-specific synthesis and secretion
of spliced hepatitis B virus genomes in hepatoma cells. Virology. 2000;271(2):371-81.
45.   Lumley SF, Kent C, Jennings D, Chai H, Airey G, Waddilove E, et al. Whole
genome sequencing of hepatitis B virus (HBV) using tiled amplicon (HEP-TILE) and probe-
based enrichment on Illumina and Nanopore platforms. medRxiv.
2024:2024.09.11.24313306.

Figure legends

Figure 1. Method overview. A) Nanopore HBV whole-genome sequencing workflow. B) The

position of the primers on the partially double-stranded HBV relaxed circular DNA (rcDNA),

found in HBV particles. The method is designed to amplify both strands, including the shorter

plus strand.

Figure 2. HBV coverage profiles from Nanopore sequencing of samples with and without

structural aberrations. Colored lines indicate differences compared to the reference sequence

reads were mapped against and grey matching nucleotides, the height reflects the number of

reads covering the positions. The approximate primer locations are marked as green (forward)

and red (reverse) arrows. A) Depicts a typical coverage profile, B) shows a sample with a

preS1 deletion as a minor variant seen as a gap in a fraction of the reads at nt 2994-3054 and

C) a sample with a majority of the HBV genomes with a preS1 deletion nt 2847-3060.

A

B

Table 1. Number of patients, genotype and viral load among the study cohorts.

HBV genotype		Cohort 1	Cohort 2	Cohort 3	Total
	A	5	0	2	7
	B	2	6	1	9
	C	8	8	1	17
	D	19	4	1	24
	E	3	1	2	6
	F	1	0	0	1
	All	38	19	7	64
Viral load IU/mL $\log_{10}$					
	<3	2	0	4	5
	3-5	24	0	3	28
	>5	12	19	0	31
Same genotype as Sanger/qPCR		100%	100%	100%	100%
Antiviral resistance sensitivity		100%	NA	NA	100%
Viral load IU/mL $\log_{10}$, mean		4.5	7.7	2.7	4.7
Fraction of the HBV genome, mean		93.7%	100%	99.8%	99.9%
Samples with full genome coverage (N/Total N)		33/38	19/19	7/7	59/64

Table 2. Analytical sensitivity and reproducibility after MagNA pure total NA extraction when running five pooled dilutions of one sample on a flongle flow cell.

Dilution	$\log_{10}$ IU/mL	% of full genome	Similiarity	Antiviral resistance detected*
1:1	5,34	100	100	L180M, S202G, M204V
1:10	4,34	100	100	L180M, S202G, M204V
1:100	3,34	56	100	L180M, S202G, M204V
1:1000	2,34	0	NA	No
1:10000	1,34	0	NA	No

*Previous Sanger sequencing showed L180M, S202G and M204V.