Feasibility of Low-Cost Virtual Reality Motor Amplification for Stroke Rehabilitation

Aidan Fisk[1], Summer Fox[1], Jenna Floyd[1], Daniel Blustein[1,2]
[1]Department of Psychology, Acadia University, Wolfville, NS, Canada
[2]Valley Regional Hospital, Nova Scotia Health Authority, Kentville, NS, Canada

Corresponding Author:
    Daniel Blustein
    18 University Ave, Wolfville, NS B4P 2R6, Canada
    Phone: +1 902 585 1220
    blustein.neuro@gmail.com

Author Note
This work was supported by ResearchNS under the New Health Investigator Grant program. The authors report there are no competing interests to declare. This study was approved by the Acadia University Research Ethics Board (REB File #22-07). We thank Joy Ndirangu and Logan VanOostrum for Unity coding assistance and Colin Berkley for text editing. All data and analysis code are freely available at OSF (https://osf.io/45kad/?view_only=cc0f6a49e7dc49ae9af0392fb03262f0).

Abstract
Purpose: Stroke is a leading cause of disability worldwide with improved treatments leading to higher survival rates and a greater demand for post-stroke rehabilitation. Technological rehabilitation solutions show promise to meet this growing need, but for maximum impact they must be accessible and affordable. To inform the development of low-cost virtual reality (VR) rehabilitation systems, here we assess the feasibility of motor amplification, an effective stroke treatment solution, on a low-cost controller-free system.
Design: We developed the REVIVE system, a VR stroke rehabilitation system operating on the standalone Meta Quest line of Head Mounted Displays which are wireless, low cost, and require no external hardware. Hand tracking, voice recognition and an automated motor amplification algorithm enable accessible engagement for users with a range of physical and cognitive abilities. An animated coach in the virtual environment guides users through gamified exercises which simulate activities of daily living and functional movements. We tested the system with 60 healthy young adults in a simulated stroke rehabilitation session, with a primary goal of validating the feasibility of the controller-free amplification feature.
Findings: Users reported minimal visually induced motion sickness even when experiencing the visuomotor perturbation generated by amplification, and more positive attitudes toward VR technology after the experience. Additionally, we provide a normative dataset for several REVIVE tasks to serve as a healthy baseline for future clinical applications.
Value: Our findings suggest that this affordable consumer-grade system, suitable for home or clinical use, has considerable potential to improve access to post-stroke rehabilitation services.

    Keywords: assistive technology, virtual reality, user acceptance, stroke rehabilitation, motor amplification

Worldwide, stroke ranks among the top causes of disability (*Global Burden of Disease 2021*). Increasing survival rates over the past 30 years due to improved treatments have led to lower death rates but more years lived with disability ("GBD Results", 2024), indicating a growing demand for post-stroke rehabilitation care. But patients typically receive less care than they need (Stockley *et al.*, 2019).

Access to care is a barrier to maximal recovery, with barriers that include cost, distance to treatment centers, and limited transportation (Shanmugasegaram *et al.*, 2013). These accessibility barriers are particularly acute in rural settings and in developing nations. In Canada, rural patients were less likely to receive stroke unit care (36% versus 51%), to see a neurologist (26% versus 65%), or to see an occupational therapist (52% versus 72%) compared to urban patients (Koifman *et al.*, 2016). This mirrors treatment disparity globally with patients in Low- or Middle-Income Countries (LMICs) being less likely to receive standard care recommendations (Kayola *et al.*, 2023; Owolabi *et al.*, 2021).

Increasing access to intensive post-stroke rehabilitation therapies is an important goal that could have a substantial impact on patient outcomes worldwide. Telehealth and other technology-enabled solutions can help overcome these barriers to treatment and have been shown to be as effective and cost-efficient as in-person stroke care (Caughlin *et al.*, 2020), enabling an increase in therapeutic dosing and access ("Canadian Stroke Best Practice Recommendations", 2022). Public acceptance of telehealth solutions is high ("Canadian Stroke Best Practice Recommendations", 2022), including among adults 65 years and older who make up about 75% of stroke cases (Yousufuddin and Young, 2019). Limitations persist as technology-enabled stroke rehabilitation options can place additional time and training burdens on clinicians (Cormican *et al.*, 2023), be costly (Krakauer *et al.*, 2021) and/or require multi-component hardware setups (Tsekleves *et al.*, 2016; Uswatte *et al.*, 2021).

Efforts to use low-cost VR headsets, such as the Meta Quest 2, to deliver stroke rehabilitation have demonstrated beneficial impacts on stroke recovery (Amin *et al.*, 2024; Herrera *et al.*, 2023). Some evidence suggests that immersive VR confers greater benefit than conventional rehabilitation in measures of shoulder, wrist and hand function (Kiper *et al.*, 2023), and gross motor function (Soleimani *et al.*, 2024). Dynamic adjustments to task parameters , such as shifting interactable objects closer to a user with a limited range of motion, enable these systems to adapt to the different abilities of users (Bressler *et al.*, 2024; Herrera *et al.*, 2024). These approaches typically rely on changing the environment, whereas an alternative approach, motor amplification, amplifies the user's movements to reduce their motor deficit in the virtual world. Motor amplification has been shown to reduce paretic limb non-use in stroke patients by temporarily reducing post-stroke motor deficits, facilitating task completion and reducing frustration, thus leveraging the power of reinforcement learning (Ballester *et al.*, 2015, 2016). This approach confers advantages over environmental manipulation by enabling movements through visual space which are not limited by the user's range of movement, thus increasing access to users with a wider range of motor abilities.

Attempts to implement amplification on low-cost headsets have been limited, with one requiring the need for controllers and manual setting of parameters ("WalkinVR", n.d.), reducing its ease of use. While the benefits of amplification have been established in the lab using controller-based systems (Ballester *et al.*, 2015; Wang *et al.*, 2020), realizing the accessibility benefit of low-cost VR requires the dedicated validation of automated amplification on a controller-free, low-cost platform. Additionally, the user acceptance of amplification, which intentionally introduces a visual-motor perturbation, is less understood and requires validation.

The introduction of a visual-motor mismatch makes the assessment of visually induced motion sickness (VIMS) a critical component of this validation. Commonly explained by sensory conflict theory (LaViola, 2000; Oman, 1990), VIMS represents a mismatch between visual and vestibular or proprioceptive cues, just as amplification in VR introduces, and could be a barrier to adoption. Although low-cost VR rehabilitation systems have shown high levels of engagement and satisfaction (Mc Kittrick *et al.*, 2023), further insight into the user experience with movement amplification is needed to determine the approach's feasibility and usability. We therefore focused on measuring VIMS levels and user acceptance of the technology as primary outcomes of this validation.

To fill a need for more affordable and user-friendly stroke rehabilitation options, we developed the REVIVE virtual reality system (Fisk et al., 2025) (Fig. 1). The REVIVE system (Rehabilitation and Engagement through Virtual Immersive Environments) uses a head-mounted display and an animated virtual coach to guide patients through exercises, reducing the need for constant clinician presence. Features include gamified tasks for high-dose motor exercise, hand tracking to eliminate the need for handheld controllers, and wireless operation. The REVIVE software runs on a commercially available headset, the Meta Quest 2/3S/3, which costs about the same as four physiotherapy sessions, enabling widespread deployment.

This paper presents the methodological development and initial proof-of-concept validation of the REVIVE system. We assessed the feasibility of automated, controller-free motor amplification in a sample of 60 young adults without disability and generated a normative baseline dataset that can be used to make clinical comparisons. The primary focus of the feasibility testing was to determine participant receptiveness to the technology as measured by a VR attitudes survey before and after the simulated rehabilitation session. Demonstrating the technology as accessible is critical to increase its acceptance by patients, particularly in individuals who may have less experience with emerging technologies such as older adults (Ortiz-Fernandez *et al.*, 2021). We also elaborate on the prevalence of VIMS (Fisk et al., 2025), which could be a barrier to the adoption of VR-based stroke rehabilitation. Our findings will inform the ongoing development of the REVIVE system and contribute to the ongoing discourse related to low-cost rehabilitation options for stroke, particularly in LMICs and rural areas with reduced access to care (Kayola *et al.*, 2023; Owolabi *et al.*, 2021).

Figure 1. *REVIVE system overview*. A. User interacting with virtual environment through a Meta Quest 2 head mounted display. B. User view within the virtual environment during the Bubble Pop Game. C. User view during the Pong Game. D. User view during the Animal Feeding Task.

Methods

REVIVE System

We created a software program featuring 12 gamified tasks, with three tested in this study. These tasks were designed to either simulate daily activities, like pouring water into a glass, or to act as immersive games where the hand acts as a controller in such classic games as Pong or Snake. Each task targeted one of three movement categories: 1) Gross reaching; 2) Reach and grasp; or 3) Precision grasping.

In this study, the software was run on the Meta Quest 2 head mounted display, an immersive virtual reality system that can operate wirelessly (the software also runs similarly on the more recently released Meta Quest 3S and 3). Onboard cameras track the hands of the user, allowing the user to interact with objects in the virtual space without the need for handheld controllers or external camera setups. Real-world physics are simulated in the virtual environment using the Unity Game Engine. Some object interactions behave differently from

their real-world analogue, such as grasping an object, and these incongruencies are demonstrated and explained in an initial tutorial for first-time users. We did not run the interaction tutorial in this study since the tasks completed in this study involved only gross reaching movements. At the beginning of each session, the user completes a guided calibration protocol which measures the range of motion (ROM) of each arm and hand. Any bilateral difference in ROM can be reduced by amplifying the movements of the deficient limb in the virtual world. For example, if a stroke survivor has a reduced range of shoulder flexion in one limb in the real world, the tracked movements can be amplified and displayed to match the unaffected limb's range of motion. We apply a motor smoothing algorithm to reduce the jitteriness induced by amplified hand tracking noise.

Usability Study

Following the initial development of the REVIVE system (Fisk et al., 2025), we conducted preliminary testing with healthy young adult participants to simulate a stroke rehabilitation session. This study was approved by the Acadia University Research Ethics Board (REB File #22-07). Sixty students from Acadia University receiving course credit for their involvement were recruited for the study. As an initial proof-of-concept and feasibility study, this sample size (30 per group) was chosen to provide sufficient data to establish normative baselines and detect potential usability issues, consistent with similar early-stage technology assessments. Participants provided demographic data, including gender identity [M/F/other], handedness [L/R/other], and age.

Study Setup

Participants donned a Meta Quest 2 VR headset connected to a desktop computer that ran the VR exercises. The experimenter monitored the virtual environment displayed to the participants and offered verbal guidance as necessary. The VR session lasted 10 to 15 minutes and included an initial calibration phase to assess participants' range of motion. Surveys were administered to collect data on participants' attitudes toward VR (pre- and post-session), prior technology use (pre-session), and VIMS (post-session). Half of the participants experienced motor amplification of their hand movements, while the other half served as a control group with no amplification.

VR Task Overview

Participants began with a calibration task to establish their range of motion, after which motor amplification was applied for the experimental group. They then performed three tasks: a Bubble Pop game, a Pong game, and an Animal Feeding task. The order of the last two tasks was counterbalanced, with half of the participants in each group completing the Pong game first and the other half beginning with the Animal Feeding task. A description of each task is provided below:

Calibration Task. Participants reached for colored cubes positioned at different locations in their environment. The software recorded the range of motion for each hand and calculated the amplification factor for the experimental group based on the proportional difference between the two sides. This information was also used to set interaction boundaries for the subsequent tasks. Outcome measures included the volume of the ROM envelope and the amplification factor applied.

Bubble Pop Game. Participants popped virtual bubbles appearing in their field of view over a one-minute period. Hand positions were tracked at a frequency of 50 Hz to record movement trajectories. Outcome measures included the total number of bubbles popped, the overall distance traveled by the hands, and the 95th percentile hand velocity.

Pong Game. Participants controlled a virtual paddle by moving their hand forward or backward. The hand with the smaller range of motion was the controlled hand, with amplification applied in the experimental condition. A ball rolled across the table, and participants earned points by successfully deflecting the ball past the computer-controlled paddle. After a one-minute practice round, participants played a two-minute game. Outcome measures included the number of paddle-ball contacts, points scored, and opponent points scored.

Animal Feeding Task. Participants sat at a table in front of three cans of animal food arranged in a line from left to right. Each can was adorned with the image of an animal (dog, cat, bird) to indicate which animal's food it contained. A virtual animal appeared at the far end of the table, and participants were required to push the matching can across a designated line away from the participant to "feed" the animal. The line's distance was calibrated to the individual's range of motion. Once the task was completed, the animal changed appearance and emitted an identifying sound (e.g. a bark for the dog), prompting the participant to push a new can across the line. Participants practiced for one minute before completing a two-minute session. The outcome measure was the number of correctly pushed cans.

Surveys

After providing written informed consent, participants completed surveys that collected demographic information and pre-intervention data on VR experience and attitudes. One question asked, "How often have you used virtual reality systems?", with the options to select being *Very frequently*, *Frequently*, *Occasionally*, *Rarely*, *Never*. In the VR attitudes survey (Table I), participants responded to six statements using a Likert-scale [*Strongly Agree* (4), *Agree* (3), *Neutral* (2), *Disagree* (1), *Strongly Disagree* (0)]. The numerical equivalent scores reported for Statements 2 through 5 (Table I) were summed and their mean was used as an overall metric indicating attitudes towards VR technology. Scores for Statements 3 and 6 were reverse-coded.

Following the VR exercises, an abbreviated VIMS symptom assessment was administered and the VR attitudes survey was readministered. The difference between the pre-survey and post-survey mean scores was used to indicate the overall change in VR attitude attributed to the VR experience. The VIMS symptom assessment consisted of 5 of 9 items from the Virtual Reality Sickness Questionnaire (VRSQ) (Kim *et al.*, 2018): General Discomfort, Fatigue, Eyestrain, Difficulty Focusing, Headache. The five items were selected to minimize survey burden while capturing the most common and relevant symptoms of VIMS for a short VR session. We also sought to reduce redundancy across survey items and focus on aspects most likely to affect stationary VR users. The mean score across the five items was used to indicate VIMS immediately following the VR experience.

Table I. *Statements used to assess attitudes towards virtual reality with Likert-scale responses.*

Statement Number	Statement
1	I am very comfortable using technology (e.g., cellphones, computers, gaming systems).
2	I am very comfortable in virtual environments that are created using virtual reality systems.
3	I am concerned that I could damage the virtual reality system when I use it.
4	I think that virtual reality systems are easy to use.
5	I think that virtual reality could cause mental fatigue.
6	I am motivated to learn how to use virtual reality systems.

Analysis Plan

Range of motion and motor performance scores on each of the three VR tasks were compared between groups using a two-sample t-test. For data presenting unequal variances, Welch's version of the test was used. To determine if there was a change in VR attitude score across all participants following the VR experience, a Wilcoxon signed rank test was run (on the medians) with the alternative hypothesis that the true location is not equal to zero (median $\neq 0$) which would indicate a non-zero change in attitude score. A Wilcoxon signed rank test was also conducted on VIMS medians to determine if they exceeded minimal levels after the VR experience, with the alternative hypothesis that the true median is greater than 1, which would indicate VIMS levels reported as "slightly" or above (Kim *et al.*, 2018). Wilcoxon rank sum tests were run for comparisons between groups for both VR attitude change and reported VIMS. Medians (and thus the nonparametric Wilcoxon tests) were used with the non-normally distributed zero-inflated data that was expected. Analysis scripts in the R programming language along with all data are freely available at OSF: https://osf.io/45kad/?view_only=cc0f6a49e7dc49ae9af0392fb03262f0

Results

Sixty young adults without disability participated in the simulated VR rehabilitation exercises, either with or without motor amplification. We sought to generate a normative baseline of user performance, measure attitudes towards the technology and how they change with VR exposure, gather feedback on user experience including VIMS symptoms, and identify any effect of amplification on these metrics.

Participants ranged in age from 18 to 39 years old (mean age = 19.7 years, SD = 3.22 years), with 47 participants identifying as women, 10 as men and 3 as non-binary. Three self-reported as left-handed.

During the calibration task, a Welch two-sample t-test indicated that the range of motion (ROM) for the left hand ($\mu = 0.41m^2$, range: $0.13 – 1.17$ m^2) was not statistically different ($t(117.97) = -0.266$, $p = 0.79$) from the right-hand ROM ($\mu = 0.42m^2$, range: $0.12 – 1.07$ m^2) (Fig. 2, Supplementary Fig. 1). No difference between hands was observed for the total distance moved during the calibration task and bubble pop game (Supplementary Fig. 2).

Figure 2. *Range of motion (ROM) by hand during calibration.* See Supplementary Fig. 1 for a breakdown of ROM by hand and amplification status.

Based on an individual's difference in ROM between the hands, an amplification factor was calculated to make these ROMs equal. The amplification factor calculated for participants with smaller left-hand ROMs ($\mu = 1.056$, range: $1.003 – 1.201$, n = 25) was slightly larger than those with smaller right hand ROMs ($\mu = 1.035$, range: $1.002 – 1.164$, n = 35), but this was not a statistically significant difference ($t(38.19) = 1.77$, $p = 0.085$, Welch's t-test) (Fig. 3). After centering the amplification factors around zero (i.e. >0 indicates a right-hand amplification), a one-sample t-test indicated the amplification factors were not statistically different from zero ($\mu = -0.003$, $t(59) = -0.376$, $p = 0.71$), indicating no hand side bias in detected ROM in this able-bodied participant sample.

Figure 3. *Computed amplification factor for left and right hands across all participants.*

With these small amplification factors, we observed no statistically significant effect of amplification on motor performance indicators across the three games tested (Figs. 4-6). In the Bubble Pop task, the amplified participant group popped a mean of 116.4 bubbles in one minute [SD = 66.3, range: 10 - 299] and the control group popped 102.5 bubbles [SD = 44.9, range: 13 - 235] (Fig. 4). A Welch's t-test indicated this was not a statistically significant difference (t(50.99) = -0.95, p = 0.34).

Figure 4. *Bubble pop results for participants without and with amplification.*

In the Pong Game, the amplified participant group hit the ball a mean of 26.8 times in 2 minutes [SD = 3.00, range: 19 - 33] and the control group hit the ball 27.1 times [SD = 2.30, range: 23 - 32] (Fig. 5). A Welch's t-test indicated this was not a statistically significant difference (t(54.32) = 0.39, p = 0.70).

Figure 5. *Pong Game results for participants without and with amplification.*

In the Animal Feeding Task, the amplified participant group pushed a mean of 38.1 cans in two minutes [SD = 21.1, range: 6 - 68] and the control group pushed 39.8 cans [SD = 18.8, range: 2 - 73] (Fig. 6). A Welch's t-test indicated this was not a statistically significant difference (t(57.24) = 0.37, p = 0.74).

Across all participants in both conditions, there was an average increase in attitude towards VR of 0.23 (SD = 0.54) (Fig. 7A), representing a small effect size (Cohen's d = 0.44) that was statistically significant as indicated by a Wilcoxon signed rank test with continuity correction (V = 987, p < .001). The mean increases for both conditions were numerically equivalent at 0.23 per question change in VR attitude score (amplification group SD = 0.53, control group SD = 0.56) (Fig. 7B).

Figure 6. *Animal Feeding results for participants without and with amplification.*

Figure 7. *Attitude change towards VR following the VR experience.* A. Mean per question score change from the pre- to post-survey for all participants. A Wilcoxon signed rank test with continuity correction showed this change to be statistically significant (V = 987, p < .001). B. Same data as in A broken down by intervention group, with and without motor amplification.

Immediately following the VR experience there was a mean VIMS score of 0.57 (SD = 0.44, median = 0.4) across all participants (Fig. 8A). On this scale, where users report the prevalence of symptoms with a zero aligning with 'not at all' and 4 indicating "very" as in Kim *et al.*, 2018, this median score of 0.4 represents a minimal level of discomfort, well below the threshold of 1 ('slightly') that we tested against. Three participants did not answer all questions on the VIMS survey and their data were removed from the analysis. Of these three participants, one responded to only one question with a score of 1 ("slightly") and the other two each skipped one question and had average scores across the answered questions of 0.125 and 1.25. A Wilcoxon signed rank test with continuity correction was run with the alternative hypothesis of u>1 given that scores above 1 would indicate more than a minor effect ("slightly" or greater VIMS symptoms). This test did not show a statistically significant effect on VIMS (V = 95.5, p = 1). A Wilcoxon rank sum test with continuity correction indicated that the difference in VIMS scores between amplified (mean = 0.64, median = 0.6) and non-amplified (mean = 0.50, median = 0.4) participants was not statistically different (W = 326, p = .20) (Fig. 8B).

Figure 8. *Visually induced motion sickness scores immediately following the VR experience.* . A. Mean score per statement for all participants. A Wilcoxon signed rank test with continuity correction did not show a statistically significant effect (u > 1; V = 95.5, p = 1). B. Same data as in A broken down by intervention group, with and without motor amplification [adapted from (Fisk, A. *et al.*, 2025)].

Discussion

We tested a virtual reality stroke rehabilitation platform in healthy young adults to assess the feasibility of its use and to generate a pool of normative motor performance data. Importantly, we show that motor amplification on a controller-free low-cost system has limited adverse or detrimental impacts on motor performance and VIMS. Additionally, a short exposure to the system positively influences user attitudes towards the technology. These data, alongside forthcoming data from healthy older adults, will enable quantitative motor recovery comparisons in stroke patients.

Importantly, we observed no statistically significant differences in motor performance between the amplification and no amplification groups. Given that the amplification introduces a visuo-motor disruption, a negative impact on performance might have been expected. Our findings confirm that automated and smoothed visual amplification on a low-cost VR headset does not inadvertently degrade motor performance in healthy users. This is an important validation step suggesting that the feature may be feasible for use with populations such as stroke patients, who may show larger amplification factors (90% of this study's participants had amplification factors less than 10%).

Another important finding is that minimal VIMS was reported, though slightly higher VIMS levels were observed in the motor amplification group. This trend, possibly explained by sensory conflict theory, was not statistically significant but could foretell potential issues when greater amplification factors are used, such as in clinical populations. Greater amplification factors would result in a larger mismatch between visual and proprioceptive feedback, potentially exacerbating VIMS levels experienced by users. This highlights the need to monitor VIMS symptoms closely as this methodology is translated to clinical populations.

Participant attitudes towards VR technology improved following the VR experience, with similar positive shifts observed in both the amplification and no amplification groups. These results suggest the feasibility for more widespread clinical adoption of such VR exercises.

As an initial feasibility assessment of this new technology, our study has several limitations that should be addressed in future work. The focus here on a healthy young adult population, while necessary, restricts the generalizability of our findings to inform the system's future clinical deployment. A forthcoming study testing healthy older adults as well as ongoing work with stroke patients will address this limitation. Additionally, the VR exposure in this study was limited to three tasks in a single session; longer sessions or multiple sessions could induce more VIMS symptoms. However, VIMS symptoms have been shown to occur quite quickly, with one study reporting an average nausea onset time of 62 seconds (Mazloumi Gavgani *et al.*, 2018). Nevertheless, ongoing comprehensive testing is needed to assess the system's feasibility when implemented across multiple sessions in clinical contexts. Furthermore, we did not administer a pre-session sickness questionnaire, meaning any post-session sickness scores could be confounded by pre-existing symptoms, though the low post scores hold value on their own. Future studies should include a pre-test to isolate the symptoms induced by the VR experience itself.

Overall, the VR system presented here shows strong potential for clinical deployment as a stroke rehabilitation solution. Importantly, the use of a low-cost consumer headset could make this emerging treatment paradigm accessible to more patients than previous VR stroke rehabilitation systems. Before clinical deployment, next steps include a forthcoming feasibility study with healthy older adults and clinical studies with stroke survivors.

References

Amin, F., Waris, A., Syed, S., Amjad, I., Umar, M., Iqbal, J. and Omer Gilani, S. (2024), "Effectiveness of Immersive Virtual Reality-Based Hand Rehabilitation Games for Improving Hand Motor Functions in Subacute Stroke Patients", *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, Vol. 32, pp. 2060–2069, doi: 10.1109/TNSRE.2024.3405852.

Ballester, B.R., Maier, M., San Segundo Mozo, R.M., Castañeda, V., Duff, A. and M. J. Verschure, P.F. (2016), "Counteracting learned non-use in chronic stroke patients with reinforcement-induced movement therapy", *Journal of NeuroEngineering and Rehabilitation*, Vol. 13 No. 1, p. 74, doi: 10.1186/s12984-016-0178-x.

Ballester, B.R., Nirme, J., Duarte, E., Cuxart, A., Rodriguez, S., Verschure, P. and Duff, A. (2015), "The visual amplification of goal-oriented movements counteracts acquired non-use in hemiparetic stroke patients", *Journal of NeuroEngineering and Rehabilitation*, Vol. 12 No. 1, p. 50, doi: 10.1186/s12984-015-0039-z.

Bressler, M., Merk, J., Gohlke, T., Kayali, F., Daigeler, A., Kolbenschlag, J. and Prahm, C. (2024), "A Virtual Reality Serious Game for the Rehabilitation of Hand and Finger Function: Iterative Development and Suitability Study", *JMIR Serious Games*, Vol. 12, p. e54193, doi: 10.2196/54193.

Caughlin, S., Mehta, S., Corriveau, H., Eng, J.J., Eskes, G., Kairy, D., Meltzer, J., *et al.* (2020), "Implementing Telerehabilitation After Stroke: Lessons Learned from Canadian Trials", *Telemedicine and E-Health*, Vol. 26 No. 6, pp. 710–719, doi: 10.1089/tmj.2019.0097.

Cormican, A., Hirani, S.P. and McKeown, E. (2023), "Healthcare professionals' perceived barriers and facilitators of implementing clinical practice guidelines for stroke rehabilitation: A systematic review", *Clinical Rehabilitation*, Vol. 37 No. 5, pp. 701–712, doi: 10.1177/02692155221141036.

Fisk, A., Fox, S., Floyd, J. and Blustein, D. (2025), "A low-cost virtual reality stroke rehabilitation system: development and initial feasibility assessment", *International Conference on Rehabilitation Robotics*, presented at the International Conference on Rehabilitation Robotics, IEEE, Chicago, IL, p. In Press.

"GBD Results". (2024), , Seattle, WI.

Global Burden of Disease 2021: Findings from the GBD 2021 Study. (2024), , Institute for Health Metrics and Evaluation (IHME), Seattle, WA.

Herrera, V., Albusac, J., Schez-Sobrino, S., Reyes-Guzmán, A. and Glez-Morcillo, C. (2024), "Dynamic Adjustment of Interactive Objects in Virtual Environments for Upper Limb Rehabilitation: A Patient-Centred Solution", *Spanish Computer Graphics Conference (CEIG)*, The Eurographics Association, doi: 10.2312/CEIG.20241142.

Herrera, V., Vallejo, D., Castro-Schez, J.J., Monekosso, D.N., De Los Reyes, A., Glez-Morcillo, C. and Albusac, J. (2023), "Rehab-Immersive: A framework to support the development of virtual reality applications in upper limb rehabilitation", *SoftwareX*, Vol. 23, p. 101412, doi: 10.1016/j.softx.2023.101412.

Kayola, G., Mataa, M.M., Asukile, M., Chishimba, L., Chomba, M., Mortel, D., Nutakki, A., *et al.* (2023), "Stroke Rehabilitation in Low- and Middle-Income Countries: Challenges and Opportunities", *American Journal of Physical Medicine & Rehabilitation*, Vol. 102 No. 2S, pp. S24–S32, doi: 10.1097/PHM.0000000000002128.

Kim, H.K., Park, J., Choi, Y. and Choe, M. (2018), "Virtual reality sickness questionnaire (VRSQ): Motion sickness measurement index in a virtual reality environment", *Applied Ergonomics*, Vol. 69, pp. 66–73, doi: 10.1016/j.apergo.2017.12.016.

Kiper, P., Godart, N., Cavalier, M., Berard, C., Cieślik, B., Federico, S., Kiper, A., *et al.* (2023), "Effects of Immersive Virtual Reality on Upper-Extremity Stroke Rehabilitation: A Systematic Review with Meta-Analysis", *Journal of Clinical Medicine*, Vol. 13 No. 1, p. 146, doi: 10.3390/jcm13010146.

Koifman, J., Hall, R., Li, S., Stamplecoski, M., Fang, J., Saltman, A.P. and Kapral, M.K. (2016), "The association between rural residence and stroke care and outcomes", *Journal of the Neurological Sciences*, Vol. 363, pp. 16–20, doi: 10.1016/j.jns.2016.02.019.

Krakauer, J.W., Kitago, T., Goldsmith, J., Ahmad, O., Roy, P., Stein, J., Bishop, L., *et al.* (2021), "Comparing a Novel Neuroanimation Experience to Conventional Therapy for High-Dose Intensive Upper-Limb Training in Subacute Stroke: The SMARTS2 Randomized Trial", *Neurorehabilitation and Neural Repair*, Vol. 35 No. 5, pp. 393–405, doi: 10.1177/154596832110007.

LaViola, J.J. (2000), "A discussion of cybersickness in virtual environments", *SIGCHI Bull.*, Vol. 32 No. 1, pp. 47–56, doi: 10.1145/333329.333344.

Mazloumi Gavgani, A., Walker, F.R., Hodgson, D.M. and Nalivaiko, E. (2018), "A comparative study of cybersickness during exposure to virtual reality and 'classic' motion sickness: are they different?", *Journal of Applied Physiology*, Vol. 125 No. 6, pp. 1670–1680, doi: 10.1152/japplphysiol.00338.2018.

Mc Kittrick, A., Desselle, M.R., Padilha Lanari Bo, A., Zhang, B., Laracy, S. and Tornatore, G. (2023), "Patient Experience in Adjunct Controller-Free Hand Tracking Virtual Reality Tasks for Upper-Limb Occupational Therapy Rehabilitation", *Journal of Patient Experience*, Vol. 10, p. 23743735231211983, doi: 10.1177/23743735231211983.

Oman, C.M. (1990), "Motion sickness: a synthesis and evaluation of the sensory conflict theory", *Canadian Journal of Physiology and Pharmacology*, NRC Research Press, Vol. 68 No. 2, pp. 294–303, doi: 10.1139/y90-044.

Ortiz-Fernandez, L., Magnusson, C., Ruiz, A.G., Indart, L.M., Zabala, J.S., Menica, J.A.A. and Arri, E.A. (2021), "Exploring attitudes towards use of technology to support stroke survivors living at home: A quantitative and qualitative content analysis study in Spain", *Journal of Rehabilitation and Assistive Technologies Engineering*, SAGE Publications Ltd STM, Vol. 8, p. 20556683211019690, doi: 10.1177/20556683211019690.

Owolabi, M.O., Thrift, A.G., Martins, S., Johnson, W., Pandian, J., Abd-Allah, F., Varghese, C., *et al.* (2021), "The state of stroke services across the globe: Report of World Stroke Organization–World Health Organization surveys", *International Journal of Stroke*, Vol. 16 No. 8, pp. 889–901, doi: 10.1177/17474930211019568.

Shanmugasegaram, S., Oh, P., Reid, R.D., McCumber, T. and Grace, S.L. (2013), "Cardiac rehabilitation barriers by rurality and socioeconomic status: a cross-sectional study", *International Journal for Equity in Health*, Vol. 12 No. 1, p. 72, doi: 10.1186/1475-9276-12-72.

Soleimani, M., Ghazisaeedi, M. and Heydari, S. (2024), "The efficacy of virtual reality for upper limb rehabilitation in stroke patients: a systematic review and meta-analysis", *BMC Medical Informatics and Decision Making*, Vol. 24 No. 1, p. 135, doi: 10.1186/s12911-024-02534-y.

Stockley, R., Peel, R., Jarvis, K. and Connell, L. (2019), "Current therapy for the upper limb after stroke: a cross-sectional survey of UK therapists", *BMJ Open*, Vol. 9 No. 9, p. e030262, doi: 10.1136/bmjopen-2019-030262.

Tsekleves, E., Paraskevopoulos, I.T., Warland, A. and Kilbride, C. (2016), "Development and preliminary evaluation of a novel low cost VR-based upper limb stroke rehabilitation platform using Wii technology", *Disability and Rehabilitation: Assistive Technology*, Taylor & Francis, Vol. 11 No. 5, pp. 413–422, doi: 10.3109/17483107.2014.981874.

"Understanding the Needs of People Living With Heart Disease and Stroke During the Covid-19 Pandemic: A Public Opinion Survey. Unpublished. Toronto, Ontario, 2021". (2022), *Canadian Stroke Best Practice Recommendations: Virtual Stroke Rehabilitation Interim Consensus Statement*, Vol. 11, pp. 1076–1082.

Uswatte, G., Taub, E., Lum, P., Brennan, D., Barman, J., Bowman, M.H., Taylor, A., *et al.* (2021), "Tele-rehabilitation of upper-extremity hemiparesis after stroke: Proof-of-concept randomized controlled trial of in-home Constraint-Induced Movement therapy", *Restorative Neurology and Neuroscience*, Vol. 39 No. 4, pp. 303–318, doi: 10.3233/RNN-201100.

"WalkinVR". (n.d.). *WalkinVR*, available at: https://www.walkinvrdriver.com/ (accessed 21 February 2025).

Wang, J., Lum, P., Kuo, H., Phan, Thanh and Lee, S.W. (2020), "Visual Amplification of Hand Movements in VR to Promote Arm Use in Unsupported Reaching Tasks", *RESNA Annual Conference*.

Yousufuddin, M. and Young, N. (2019), "Aging and ischemic stroke", *Aging*, Vol. 11 No. 9, pp. 2542–2544, doi: 10.18632/aging.101931.