medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

**Title**

# A multicenter study to assess the performance of the point-of-care RT-PCR Cobas SARS-CoV-2 & Influenza A/B nucleic acid test for use on the Cobas Liat system in comparison with centralized assays across healthcare facilities in the United States

**Authors**

Elissa M. Robbins[a#]

Rasa Bertuzis[a]

Ho-Chen Chiu[b]

Lupe Miller[a]

Christopher Noutsios[a]

**Affiliations**

[a]Roche Molecular Systems, Inc., Pleasanton, California, USA

[b]Roche Sequencing Solutions, Inc., Indianapolis, Indiana, USA

**[#]Corresponding author**

Elissa M. Robbins

**Running title**

Cobas Liat SARS-CoV-2 & Influenza A/B test



NOTE: This preprint reports new research that has not been certified by peer review and should not be used to guide clinical practice.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

**Non-structured abstract**

Respiratory diseases can share many of the same symptoms, highlighting the need for timely and accurate differentiation to facilitate effective clinical management and reduce transmission. Compared with centralized testing, molecular point-of-care tests (POCT) can provide a faster time to result.

We evaluated the RT-PCR POCT Cobas® SARS-CoV-2 & Influenza A/B qualitative assay for use on the Cobas Liat® system (the Cobas Liat SARS-CoV-2 & Influenza A/B test) in nasal and nasopharyngeal swab samples from 10 diverse healthcare facilities in the US. A composite comparator design consisting of three centralized tests was used to analyze SARS-CoV-2, whilst performance versus a single centralized test was used for analysis of influenza A and B. Evaluations included performance stratified by sample type (prospective/retrospective and nasal/nasopharyngeal [paired by subject]), collection method (self/healthcare worker-collected [alternated and approximately balanced], symptom status (symptomatic/asymptomatic), and SARS-CoV-2 vaccination status, as well as assay inclusivity and system ease of use.

A total of 2,247 samples were tested. For SARS-CoV-2, the overall percent agreement (OPA) was 98.8% (95% CI: 97.9, 99.3) in nasal swab samples and 99.0% (95% CI: 98.2, 99.4) in nasopharyngeal swab samples. Regression analysis showed that Ct values from paired nasal and nasopharyngeal swab samples were highly correlated (correlation coefficient 0.83). The OPA was ≥99.5% (sample type dependent) and 100.0% for influenza A and B, respectively. The Cobas Liat SARS-CoV-2 & Influenza A/B test was easy to use.

These results support the use of the POCT in various sample types and by various operators in the intended-use setting.

**Summary statement/importance**

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

This study highlights the benefits of RT-PCR POCT, namely comparable performance to centralized testing in multiple sample types and ease of use. Utilizing assays such as the Cobas Liat SARS-CoV-2 & Influenza A/B test may improve timely differentiation of respiratory diseases that share similar symptoms.

## Introduction

More than 4 years since the start of the coronavirus disease 2019 (COVID-19) pandemic, much of the world has returned to normality, yet COVID-19 continues to cause severe illness and mortality in the United States (US) and elsewhere(1, 2). This applies particularly for the immunocompromised, who are at risk of severe outcomes (3) and comprise approximately 3% of the US adult population(4). Disease activity is year-round, but seasonal spikes occur from November to April(5), coinciding with circulation of other respiratory viruses, such as influenza(6). Respiratory diseases such as COVID-19 and influenza share similar symptoms(7), highlighting the need for timely and accurate differentiation of the causative agent to help facilitate effective clinical management and reduce onward transmission.

The current choice of testing for severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) is setting dependent(8), but standard laboratory-based (i.e., centralized) nucleic acid amplification tests are the mainstay of diagnosis(9). Numerous antigen or molecular point-of-care tests (POCT) are commercially available(10, 11), offering the benefit of a faster time to result(12). Both POCT formats have lower performance compared with centralized testing, particularly sensitivity (13) — a secondary analysis of antigen POCT found the performance to be commonly lower than stated within manufacturers' instructions for use (14) — but molecular POCT have relatively higher performance in comparison with antigen POCT, with pooled sensitivity values of 93% (99% specificity) versus 71% (99% specificity), respectively(13). Molecular POCT can provide actionable results in as little as 20 minutes depending on the assay and workflow(11).

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Influenza is a significant cause of severe illness and death(15-17), resulting in a substantial clinical burden to patients and the healthcare system alike(18). Molecular POCT for the detection of influenza A/B subtypes, and other respiratory viruses such as respiratory syncytial virus, have been available for several years(11, 12, 19-22). Clinical practice guidelines from the Infectious Diseases Society of America recommend that antivirals are initiated for certain patient groups as soon as possible regardless of illness duration, or within 48 hours of illness onset(23), supporting the use of rapid molecular POCT for influenza testing in preference to centralized laboratory testing. In the hospital setting, an interventional study of a molecular POCT for influenza found significant improvements in isolation practices and reductions in length of stay compared with centralized testing(24), whilst a modeling study found that the introduction of POCT could reduce time to diagnosis, hospital stay, and in-hospital costs(25). Molecular POCT for influenza can generate sensitivity estimates of ≥96–100% in real-world settings(22, 26).

The primary objective of this study was to evaluate the real-world clinical performance of the POC Cobas SARS-CoV-2 & Influenza A/B qualitative assay for use on the Cobas Liat system (herein referred to as the Cobas Liat SARS-CoV-2 & Influenza A/B test), a multiplexed RT-PCR test combining measurement of SARS-CoV-2 and influenza subtypes, in reference to comparator centralized PCR tests for SARS-CoV-2 and influenza A and B viruses. Sample types were nasal and nasopharyngeal swab specimens from individuals suspected of respiratory viral infection consistent with COVID-19, and asymptomatic individuals exposed to COVID-19, who presented at healthcare facilities across the US during the COVID-19 public health emergency. Secondary objectives were to evaluate the influence of collection method and the ease of use of the Cobas Liat system.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

## Materials and methods

### Sites

Testing was conducted at 11 sites: 10 geographically diverse healthcare facilities that were representative of intended-use sites (i.e., POC settings, such as emergency departments, urgent care, pediatric and primary care clinics, and drive-through COVID-19 testing sites), and one reference laboratory for centralized testing. After testing using the Cobas Liat SARS-CoV-2 & Influenza A/B test at the healthcare facilities, samples were shipped to the reference laboratory for analysis with the comparator assays (see below).

Healthcare facilities were selected based on sample availability, ability to provide adequate resources and operators, ability to adhere to good clinical practice, and with the aim of ensuring representation of the demographic diversity of the US population. Not all healthcare facilities perfomming testing were certified under Clinical Laboratory Improvement Amendments (CLIA) regulations to perform waived testing, but all facilities met the requirements for CLIA-waived settings (i.e., intended-use settings using untrained non-laboratory personnel). All healthcare facilities performed both sample collection and POC testing. POC operators at the 10 healthcare facilities (n=30 operators in total) had limited or no laboratory training and were representative of typical test operators in CLIA-waived settings (e.g., nurses, nursing assistants, and medical assistants). Operators at the reference laboratory were blinded to the results generated at the healthcare facilities.

### Study population

The symptomatic group comprised samples from individuals with suspected respiratory viral infection consistent with COVID-19 (see Signs/symptoms, Supplementary material). The asymptomatic group comprised individuals with self-reported 'recent' exposure (timeframe undefined) to SARS-CoV-2-positive individuals or any other reasons to suspect COVID-19. Patient clinical information was gathered, including, but not limited to, antiviral usage for up

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

to 7 days prior to and on the day of sample collection, gender, race/ethnic group, and age. Individuals presenting at the healthcare facilities were considered for study inclusion, subject to the eligibility criteria (see Eligibility criteria, Supplementary material).

**Study design**

The clinical performance of the POC Cobas Liat SARS-CoV-2 & Influenza A/B test (27) was evaluated by comparing results to a composite comparator method for the SARS-CoV-2 analyte using the following three highly sensitive centralized testing assays: Cobas SARS-CoV-2 qualitative assay for use on the Cobas 6800/8800 systems (herein referred to as the Cobas 6800/8800 SARS-CoV-2 test)(28); Cobas SARS-CoV-2 & Influenza A/B qualitative assay for use on the Cobas 6800/8800 systems (herein referred to as the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test)(29); and Hologic® Aptima® SARS-CoV-2 Assay(30).

The clinical performance for the influenza A/B analyte components of the Cobas Liat SARS-CoV-2 & Influenza A/B test was evaluated against a single centralized testing assay, the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test(29).

Samples that generated invalid results were repeated where sample volume permitted additional testing; samples generating a second invalid result upon retesting were reported as invalid. Interrogation of discrepant samples was not performed, though selected samples were further retested with the Cobas Liat SARS-CoV-2 & Influenza A/B test for exploratory purposes.

**Sample types and collection**

Sample types included fresh and frozen nasal and nasopharyngeal swab samples, collected in the same media formulations but under different brand names: Copan Universal Transport Medium (UTM) or BD Universal Viral Transport (UVT) medium.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Prospective fresh samples (UTM)

One nasal swab and one nasopharyngeal swab sample were collected from each study participant. A nasal swab of both nostrils was first collected either by the healthcare worker (HCW), herein referred to as 'HCW-collected', or by the study participant under instruction from the HCW, herein referred to as 'self-collected', as per manufacturer's instructions.(31) A nasopharyngeal swab was then collected by HCWs from the same study participant to generate paired samples. If a nasopharyngeal swab had already been collected (using one nostril) as part of the healthcare facility's standard of care, the study nasopharyngeal swab was taken from the opposite nostril.

All prospective samples were collected from patients presenting to healthcare facilities during February to June 2022, characteristic of the 2021/2022 winter respiratory season.

Retrospective frozen samples (archived or purchased from external vendors; UTM or UVT)

The low influenza prevalence in the prospective population was anticipated because of reduced international travel, new respiratory care paradigms, increased public health awareness, and use of interventions (such as masking policies or social distancing) that occurred during the COVID-19 pandemic and study enrollment period, in which influenza A virus had been the dominant virus type in circulation.(32) Therefore, frozen samples positive for influenza A virus and for influenza B virus were used to supplement the prospectively collected fresh samples. These retrospective samples were collected in the US during the 2013–2014, 2014–2015, and 2019–2020 influenza seasons. Baseline demographic data and patient characteristics related to these samples were unavailable.

Retrospective samples with known SARS-CoV-2 status, collected between March 29, 2021 and May 26, 2021, were also included in the SARS-CoV-2 analyses.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Whilst multiple freeze-thaw cycles have a limited effect on the viral titer(33), there is the potential for samples that were initially low-titer positives to appear negative following storage. Internal Roche data support the stability of viral titer for up to three freeze-thaw cycles, so retrospective samples were only eligible for inclusion if they had undergone no more than two freeze-thaw cycles. Retrospective positive and negative samples were blended in a standard POCT workflow and assessed alongside prospectively collected samples.

**Details of assays and instruments**

Cobas Liat SARS-CoV-2 & Influenza A/B test (POC RT-PCR)

The Cobas Liat analyzer is for *in vitro* diagnostic use(27). The analyzer automates all nucleic acid amplification test processes, including target enrichment, inhibitor removal, nucleic acid extraction, reverse transcription, DNA amplification, real-time detection, and result interpretation in approximately 20 minutes(27). The Cobas Liat system comprises the Cobas Liat analyzer in conjunction with the Cobas Liat assay tubes(27).

The Cobas Liat SARS-CoV-2 & Influenza A/B test utilizes a single-use disposable assay tube that contains all the reagents necessary for the detection of SARS-CoV-2 and hosts the sample preparation and PCR processes(34). This multiplex real-time RT-PCR test is intended for the simultaneous rapid *in vitro* qualitative detection and differentiation of SARS-CoV-2, influenza A, and influenza B virus RNA in healthcare provider-collected nasal and nasopharyngeal swab samples and self-collected nasal swab samples (collected in a healthcare setting with instruction by a healthcare provider) from individuals suspected of respiratory viral infection consistent with COVID-19(34).

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test (centralized testing method [CTM] 1)

The Cobas 6800/8800 systems consists of the sample supply module, the transfer module, the processing module, and the analytic module(29). Automated data management is performed by the Cobas 6800/8800 software, which assigns results for all tests. Results are available within less than 3.5 hours after loading the sample on the system.(29) The positive SARS-CoV-2 result is defined as positive on the SARS-CoV-2 channel (Target 2) and/or the pan-sarbeco channel (Target 3).

The Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test is an automated multiplex real-time RT-PCR assay intended for simultaneous qualitative detection and differentiation of SARS-CoV-2, influenza A virus, and/or influenza B virus RNA in healthcare provider-collected nasal and nasopharyngeal swab samples and self-collected nasal swab samples (collected in a healthcare setting with instruction by a healthcare provider) from individuals suspected of respiratory viral infection consistent with COVID-19(29).

Cobas 6800/8800 SARS-CoV-2 test (CTM 2)

The Cobas 6800/8800 SARS-CoV-2 test is an automated real-time RT-PCR assay intended for the qualitative detection of nucleic acids from SARS-CoV-2 in healthcare provider-instructed, self-collected anterior nasal (nasal) swab samples (collected onsite) and healthcare provider-collected nasal, nasopharyngeal, and oropharyngeal swab samples collected from any individuals, including those suspected of COVID-19 by their healthcare provider, and those without symptoms or with other reasons to suspect COVID-19(28). The positive SARS-CoV-2 result is defined as positive on the SARS-CoV-2 channel (Target 1) and/or the pan-sarbeco channel (Target 2).

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Hologic Aptima SARS-CoV-2 assay (tiebreaker; CTM 3)

As a commonly used molecular assay(30), CTM 3 was utilized as a tiebreaker. The test utilizes transcription-mediated amplification for the qualitative detection of RNA from SARS-CoV-2 isolated and purified from nasopharyngeal, nasal, mid-turbinate and oropharyngeal swab samples, nasopharyngeal wash/aspirate, or nasal aspirates obtained from individuals meeting COVID-19 clinical and/or epidemiological criteria(35).

**Analyses**

All data analyses were performed using SAS/STAT® software (v9.4 or higher of the SAS System for Linux). No formal sample size calculations were calculated for this study, but patient enrollment was adjusted to accommodate disease prevalence and to generate a minimum of 50 SARS-CoV-2 positives, 30 influenza A positives, and 10 influenza B positives for analysis.

The clinical performance of the Cobas Liat SARS-CoV-2 & Influenza A/B test was evaluated using estimates of positive percent agreement (PPA), negative percent agreement (NPA), and overall percent agreement (OPA), calculated with 2-sided 95% confidence intervals (CI) using the Wilson-score method(36). Comparison was against either the composite comparator for SARS-CoV-2 (CTM 1, CTM 2, and CTM 3 [tiebreaker if required]) or a single comparator for influenza A/B (CTM 1).

For SARS-CoV-2, concordant results from CTM 1 and CTM 2 established true positive or true negative status. In the event of discordance between CTM 1 and CTM 2, CTM 3 (the tiebreaker) was used to establish true status. Samples could be coded as uninterpretable in the event of invalid, failed, aborted or missing results that were not resolved upon retesting. If CTM 1 or CTM 2 or CTM 3 returned uninterpretable results, and the other two methods returned different results (i.e., CTM 1 positive, CTM 2 negative, and CTM 3 uninterpretable), then the status of the sample would be described as indeterminate.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

McNemar's mid-p test (P-value) was used to assess any differences between sample types. A P-value of <0.05 was considered statistically significant.

POCT operators were asked to complete a questionnaire to evaluate the ease of use of the Cobas Liat SARS-CoV-2 & Influenza A/B test.

*In silico* analysis of the Cobas Liat SARS-CoV-2 & Influenza A/B test was performed to assess assay design inclusivity on all available SARS-CoV-2 sequences (taxonomy ID 2697049) in the Global Initiative on Sharing All Influenza Data (GISAID) and National Center for Biotechnology Information (NCBI) databases up to July 2024 (>16,800,000 sequences in NCBI and >8,900,000 sequences in GISAID). The predicted impact of each variant on gene target 1 and gene target 2 primer and probe binding site sequence was evaluated using Roche proprietary software and Melting5 software. A SARS-CoV-2 sequence would be potentially not detected by the Cobas Liat SARS-CoV-2 & Influenza A/B test if a delay in cycle threshold greater than five cycles and/or a probe melting temperature of <65°C was predicted.

## Ethics statement

This study was a non-interventional evaluation of an *in vitro* diagnostic device (test results were not used to inform patient care decisions). This study was conducted in compliance with International Conference on Harmonisation Good Clinical Practice guidelines, regulations of the US Food and Drug Administration (FDA), and the Declaration of Helsinki. The study protocol was submitted to Institutional Review Boards (IRBs) in accordance with FDA and local regulatory requirements before the start of the study. The following IRBs gave permission for this study to be performed: NorthShore University Health System Institutional Review Board; University of Rochester, Research Subjects Review Board; and Western IRB.

## Results

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

## Participants and summary of testing results

Subjects from 10 geographically diverse healthcare facilities were enrolled in the study. The summary of testing results can be seen in **Supplemental Table A**. A total of 2,209 SARS-CoV-2 results from prospective samples (both nasal and nasopharyngeal swabs) were valid for inclusion in downstream analyses.

## Population and baseline characteristics

Baseline demographic data for study participants were available for the prospectively collected samples; details on the study participants are presented in **Table 1**. Of these subjects, 506 were male (47.8%) and 553 were female (52.2%). The median age was 35 years (range 0–86).

Patient characteristics are present in **Supplemental Table B**. In total, 60.4% of subjects in the prospective population had signs and symptoms of respiratory infection, with days from onset of first symptom ranging from 1 to 365 days. The remaining 39.6% of subjects were asymptomatic but were clinically suspected of SARS-CoV-2 infection by their healthcare provider due to recent exposure or other reason. The nasal swab samples were evenly distributed between HCW-collected and self-collected.

## Medical history

The medical history relating to vaccination status is presented in **Table 2**. Most participants received a COVID-19 vaccine (68.7%), with Pfizer as the most common vaccine for both first and second doses (69.5% and 72.2%, respectively).

## SARS-CoV-2

<u>Determination of the composite comparator status for SARS-CoV-2</u>

In symptomatic study participants, the composite comparator method was SARS-CoV-2 positive for 134 nasal and 131 nasopharyngeal swab samples (**Supplemental Table C**).

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

In asymptomatic study participants, the composite comparator method was SARS-CoV-2 positive for 39 nasal and 39 nasopharyngeal swab samples (**Supplemental Table C**).

Performance

The total number of samples that were positive, negative, or invalid, and how these align with the composite comparator status, can be seen in **Figure 1**. The diagnostic performance (total number of positive and negative results) for SARS-CoV-2 can be seen in **Table 3**. Agreement between the composite comparator and the Cobas Liat SARS-CoV-2 & Influenza A/B test for detection of SARS-CoV-2 in nasal swabs (total, prospective and retrospective, self-collected or HCW-collected), nasopharyngeal swabs (total, prospective and retrospective), in samples from vaccinated and unvaccinated participants, and in symptomatic and asymptomatic participants can be seen in **Supplemental Table D**. For nasal swab samples, the total OPA, PPA, and NPA was 98.8% (95% CI: 97.9, 99.3), 97.1% (95% CI: 93.3, 98.7), and 99.1% (95% CI: 98.2, 99.5), respectively. Of note, the OPA for HCW-collected nasal swab samples was 98.4% (95% CI: 97.0, 99.2) and 99.1% (95% CI: 98.0, 99.6) for self-collected nasal swab samples. The total OPA, PPA, and NPA for nasopharyngeal swab samples was 99.0% (95% CI: 98.2, 99.4), 96.4% (95% CI: 92.4, 98.4), and 99.4% (95% CI: 98.7, 99.8), respectively.

Using all paired results for nasal and nasopharyngeal swab samples by subject, the swab type did not make a difference in the reported result (McNemar's mid-p test, P=0.771).

Cycle threshold (Ct) value distributions

The range of Ct values in nasal and nasopharyngeal swab samples can be seen in **Figure 2**. Self-collected nasal swabs showed the highest average viral load followed by nasopharyngeal swabs, then HCW-collected nasal swabs.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

The range of Ct values in subject-paired nasal and nasopharyngeal swab samples can be seen in **Figure 3**. The Deming regression analysis showed that the two sample types were highly correlated (**Figure 3A**; r = 0.83). Exploration of the collection method showed that HCW-collected nasal swabs had greater concordance between the paired samples (**Figure 3B**; r = 0.91) than the self-collected nasal swabs (r = 0.77; **Figure 3C**).

Discordance

Fourteen nasal and 11 nasopharyngeal swab samples showed discordant results between the Cobas Liat SARS-CoV-2 & Influenza A/B test and the centralized testing methods (**Supplemental Table E**); 10 were positive only on the Cobas Liat SARS-CoV-2 & Influenza A/B test.

**Influenza A**

Performance

The agreement between the Cobas Liat SARS-CoV-2 & Influenza A/B test and the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test for detection of influenza A in nasopharyngeal swabs (total, prospective and retrospective) and nasal swabs (total, prospective and retrospective, HCW-collected or self-collected) can be seen in **Supplemental Table FError! Reference source not found.**. The OPA ranged from 98.9% to 100.0%.

Discordance

Two nasal and five nasopharyngeal swab samples showed discordant results between the Cobas Liat SARS-CoV-2 & Influenza A/B test and the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test for detection of influenza A (**Supplemental Table G**), four of which were positive only on the Cobas Liat SARS-CoV-2 & Influenza A/B test.

**Influenza B**

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Performance

The agreement between the Cobas Liat SARS-CoV-2 & Influenza A/B test and the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test for detection of influenza B in nasopharyngeal swabs (total, prospective and retrospective) and nasal swabs (total, prospective and retrospective, self-collected nasal or HCW-collected) can be seen in **Supplemental Table H**. The OPA for all comparisons was 100.0%.

Discordance

No discordant results between the Cobas Liat SARS-CoV-2 & Influenza A/B test and the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test for detection of influenza B were recorded.

**Evaluation of the ease of use for the Cobas Liat system**

Twenty-seven operators with experience of testing samples on the Cobas Liat completed the questionnaire.The operators' average scores indicating their agreement with the statements in the questionnaire are shown in **Supplemental Table I**. The overall score was 4.5 out of 5 for the operators' answers to all eight statements, indicating that the operators agreed that the Cobas Liat system was easy to use.

**Inclusivity analysis**

The Cobas Liat SARS-CoV-2 & Influenza A/B test targets the nucleocapsid (N) and ORF1a/b regions of the SARS-CoV-2 genome.(34) *In silico* analysis showed that 99.98% of NCBI and 99.99% of GISAID sequences for SARS-CoV-2 had no changes in the primer/probe binding sites of both target regions simultaneously. All sequences were predicted to be detected by at least one of the two sites.

**Discussion**

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

In this study, we evaluated the performance of the POC RT-PCR Cobas Liat SARS-CoV-2 & Influenza A/B test against up to three centralized assays, stratified by various parameters. The high performance of the Cobas Liat SARS-CoV-2 & Influenza A/B test in comparison with centralized testing methods utilizing nasopharyngeal or nasal samples has previously been reported(37-41), but this study is the first to report the performance of the Cobas Liat SARS-CoV-2 & Influenza A/B test across variables such as collection method (self/HCW, retrospective/prospective), vaccination status, symptom status, and ease of use. Using the comparator result to establish the status of infection, we found high agreement between the POC RT-PCR test and the centralized testing assays for all three analytes across all variables.

Previously, a meta-analysis of nucleic acid amplification testing reported that the sensitivity for detection of SARS-CoV-2 in nasal samples may be less than nasopharyngeal samples(42), whilst a study specifically assessing the Cobas Liat SARS-CoV-2 & Influenza A/B test reported similar differences in SARS-CoV-2 detection by sample type(37). However, we saw no evidence of a sample-type difference in sensitivity, with SARS-CoV-2 positivity identified in both nasal and nasopharyngeal swab samples to a similar extent (Cobas Liat SARS-CoV-2 & Influenza A/B test positivity rate approximately 15%; the difference in sensitivity [nasal minus nasopharyngeal] was +0.9%).

A small study from Denmark found almost equivalent sensitivity for detection of SARS-CoV-2 using self- and HCW-collected samples (84.2% vs 89.5%, respectively), and that patients preferred self-collection(43). In our study, whether nasal samples were collected by HCWs or by the patients themselves had little impact on the agreement values between the POC RT-PCR test and the comparators. Indeed, in all comparisons the agreement for self-collected samples was actually higher than for HCW-collected samples. For SARS-CoV-2, Ct values of paired nasal and nasopharyngeal swab samples as measured by the POC RT-PCR test were highly correlated, and Ct values were on average lower in nasal swab samples (i.e., higher viral RNA titer) compared with nasopharyngeal swab samples, with self-collected nasal

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

swabs exhibiting the lowest Ct values. Collectively, this indicates that the high performance level of the POC RT-PCR test is maintained across collection methods, and that patients are able to self-sample effectively, and are willing to do so.

We identified a small number of discrepant samples for the SARS-CoV-2 analyte in both nasal and nasopharyngeal swab samples, consistent with other studies(37, 38), most of which were false positives. The majority of the false positives were in the high Ct/low viral titer range (Cobas Liat SARS-CoV-2 & Influenza A/B test Ct values: lowest 28.2, median 35.2, highest 37.1) and thus close to or at the lower limit of detection of Ct 35.2 (USA-WA1/2020 strain) for the Cobas Liat SARS-CoV-2 & Influenza A/B test(34). Of the false negatives detected by the Cobas Liat SARS-CoV-2 & Influenza A/B test, all except one sample were in the high Ct range. Indeed, after exploratory re-testing of false-negative nasopharyngeal samples, two were found to be positive (re-tested Ct values were 34.8 and 35.4). Generally, low viral burden can be a cause of discrepant results(44), and a small external quality assessment (reproducibility) study reported that the performance of the Cobas Liat SARS-CoV-2 & Influenza A/B test is highest at low Ct values(45).

US legislation categorizes tests for complexity (moderate or high) using seven criteria, such as the need for training and experience, and assigns scores within each criterion(45). Test systems can be assigned as 'waived complexity' under certain conditions, and these tests require no formal operator training or competency(45). In addition to CLIA-certified laboratories, the Cobas Liat SARS-CoV-2 & Influenza A/B test is authorized for use in patient care settings that are CLIA waived(34). In conjunction with the fast turnaround offered by the Cobas Liat system and demonstrated in studies of the Cobas Liat SARS-CoV-2 & Influenza A/B test in real-world settings (40, 46) or studies evaluating workflow or processing time(41, 47), our study confirms that the Cobas Liat SARS-CoV-2 & Influenza A/B test is easy to use in CLIA-waived settings, and offers results in a timeframe conducive for rapid patient management.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

The strength of our multicenter study lies in the comprehensive nature of the variables assessed, encompassing a wide variety of patient characteristics likely to be encountered in healthcare settings. The study included a prospectively enrolled cohort of patients seeking care, which is representative of the intended-use population of the Cobas Liat SARS-CoV-2 & Influenza A/B test. Our study has some limitations. Firstly, we did not ascertain the SARS-CoV-2 variants present in the samples. Whilst the emergence of SARS-CoV-2 variants has the ability to affect the diagnostic performance of the test, previous studies have indicated that the Cobas Liat SARS-CoV-2 & Influenza A/B test detects both wild-type and variants of concern, such as Alpha (B.1.1.7) or Omicron (B.1.1.529)(39, 47). Bioinformatic analysis to assess inclusivity showed that the Cobas Liat SARS-CoV-2 & Influenza A/B test is predicted to bind all sequences available in the NCBI and GISAID databases as of July 2024. The dual target assay design helps to ensure that the assay is robust and safeguards against the emergence of variants that have the potential to affect assay performance and evade detection.

## Conclusion

A POC RT-PCR test combining both measurement of SARS-CoV-2 and influenza subtypes with performance equivalent to routine centralized testing would provide a critical tool to improve the diagnosis and management of COVID-19 and influenza. We found that the performance of the Cobas Liat SARS-CoV-2 & Influenza A/B test was comparable to centralized testing methods. With the ease of use and equivalent performance, it highlights the question as to whether centralized testing should be considered as the gold standard. A study of the Cobas Liat Influenza A/B test (a Liat RT-PCR test that detects only the influenza analyte) in the emergency department found that POC testing for influenza was useful in improving several metrics, including the indication for treatment with neuraminidase in positive cases(48). It is relevant to note that SARS-CoV-2 and influenza coinfection can increase the risk of severe outcomes compared with those infected with SARS-CoV-2 alone,

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

429 particularly for those coinfected with influenza A.(49) Our study highlights the benefits of
430 molecular multiplex POCT to help improve timely differentiation of respiratory diseases that
431 share similar symptoms and support efforts to improve patient management.

432 **Funding**

433 This study was funded by Roche Molecular Systems, Inc (Pleasanton, California, USA).

434 **Disclosures**

435 EMR, RB, HC, LM, and CN are employees of Roche Molecular Systems, Inc. EMR participates
436 in Roche Connect and is a shareholder.

437 Assay disclaimers

438 The Cobas SARS-CoV-2 & Influenza A/B qualitative assay for use on the Cobas Liat system
439 (herein referred to as the Cobas Liat SARS-CoV-2 & Influenza A/B test) was originally
440 approved under Emergency Use Authorization EUA201779 and has since been cleared and
441 Clinical Laboratory Improvement Amendments (CLIA) waived under 223591/CW220014,
442 respectively, in the US and is CE-IVD marked in the European Union. Sample collection in
443 the patient's home is not approved in the US or European Union.

444 The Cobas SARS-CoV-2 & Influenza A/B qualitative assay for use on the Cobas 6800/8800
445 systems (herein referred to as the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test) is
446 authorized only for use under Emergency Use Authorization (EUA) in the US, and is CE-IVD
447 marked in the European Union. The assay is not approved for use in asymptomatic patients
448 in the US or European Union. The assay is not approved for use as a point-of-care (POC)
449 test/near-patient test (NPT) in the US/European Union. Sample collection in the patient's
450 home is not approved in the US or European Union.

451 The Cobas SARS-CoV-2 qualitative assay for use on the Cobas 6800/8800 systems (herein
452 referred to as the Cobas 6800/8800 SARS-CoV-2 test) is authorized only for use under EUA

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

in the US, and is CE-IVD marked in the European Union. The assay is not approved for use as a POC test/NPT in the US/European Union. Sample collection in the patient's home is not approved in the US or European Union.

Hologic Aptima SARS-CoV-2 assay is an FDA EUA assay.(30)

**Data availability**

The datasets generated during and/or analyzed during the current study are not publicly available due to patient confidentiality. Any access requests from qualified researchers should be submitted directly to the Ethical Committee of each participating study site.

**Acknowledgments**

Medical writing support was provided by Corrinne Segal of Elements Communications Ltd (London, UK) and was funded by Roche Molecular Systems, Inc.

COBAS and LIAT are trademarks of Roche. All other product names and trademarks are the property of their respective owners.

We would like to thank Saima Shams and Vaishali Mody for their contributions to the development of the Cobas SARS-CoV-2 & Influenza A/B qualitative assay for use on the Cobas Liat system (herein referred to as the Cobas Liat SARS-CoV-2 & Influenza A/B test), and Jingtao Sun for his bioinformatics expertise.

## Figures

Figure 1: Detection of SARS-CoV-2 using the Cobas® Liat® SARS-CoV-2 & Influenza A/B test and the composite comparator (prospective and retrospective, symptomatic and asymptomatic subjects)[a]




Abbreviations: CC, composite comparator; POC RT-PCR, point-of-care RT-PCR; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2.

[a]The composite comparator comprises the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test, Cobas 6800/8800 SARS-CoV-2 test, and Hologic® Aptima® SARS-CoV-2 Assay; the POC RT-PCR test is the Cobas Liat SARS-CoV-2 & Influenza A/B test.

The Cobas SARS-CoV-2 & Influenza A/B nucleic acid test for use on the Cobas Liat system is herein referred to as the Cobas Liat SARS-CoV-2 & Influenza A/B test. The Cobas SARS-CoV-2 & Influenza A/B qualitative assay for use on the Cobas 6800/8800 systems is herein referred to as the Cobas 6800/8800 SARS-CoV-2 & Influenza A/B test. Cobas SARS-CoV-2 qualitative assay for use on the Cobas 6800/8800 systems is herein referred to as the Cobas 6800/8800 SARS-CoV-2 test.



medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

482 Figure 2: $C_t$ values for the SARS-CoV-2 analyte from the Cobas® Liat® SARS-CoV-2 & Influenza A/B test in nasal and nasopharyngeal swab

483 samples[a]




484

485 Abbreviations: Ct, cycle threshold; HCW, healthcare worker-collected; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2.

486 [a]False positives are represented by black circles. Cross represents mean Ct values. Nasal (total): n=179, mean Ct 23.1, median Ct 21.9,

487 minimum Ct 11.0, maximum Ct 36.8; HCW-collected nasal: n=87, mean Ct 23.9, median Ct 23.7, minimum Ct 12.3, maximum Ct 36.8; self-



medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

collected nasal: n=92, mean Ct 22.4, median Ct 21.1, minimum Ct 11.0, maximum Ct 36.1; nasopharyngeal: n=177, mean Ct 23.4, median Ct 22.8, minimum Ct 10.6, maximum Ct 37.1.

False positives are indicated by black circles. The Cobas SARS-CoV-2 & Influenza A/B nucleic acid test for use on the Cobas Liat system is herein referred to as the Cobas Liat SARS-CoV-2 & Influenza A/B test. Horizontal solid line represents median values. Upper and lower boundaries of box plot represent 75th and 25th percentile.



medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

Figure 3: Deming regression for the SARS-CoV-2 analyte from the Cobas® Liat® SARS-CoV-2 & Influenza A/B test in subject-paired nasal and nasopharyngeal swab samples[a]

A: Nasal (total)

B: HCW-collected nasal swab

C: Self-collected nasal swab




Abbreviations: Ct, cycle threshold; FP, false positive; HCW, healthcare worker-collected; NPS, nasopharyngeal swab; NS, nasal; NS FP, nasal false positive; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2.

[a]The Cobas SARS-CoV-2 & Influenza A/B nucleic acid test for use on the Cobas Liat system is herein referred to as the Cobas Liat SARS-CoV-2 & Influenza A/B test. Prospective symptomatic = open circle; prospective exposed = triangle. Dashed line represents perfect correlation. Solid line represents Deming regression. Red data points represent HCW-collected nasal swabs. Blue data points represent self-collected nasal swabs.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

512 **Tables**

513 Table 1: Demographics (prospective symptomatic and asymptomatic subjects)

| Characteristics | Prospective population |
|---|---|
| **Total** | |
| N | 1,059 |
| **Age (years)** | |
| Mean | 34.2 |
| Standard deviation | 19.88 |
| Median | 35.0 |
| Range (minimum–maximum)[a] | 0.0–86.0 |
| **Age group (years), n (%)** | |
| ≤18 | 287 (27.1) |
| 19–39 | 319 (30.1) |
| 40–64 | 391 (36.9) |
| ≥65 | 62 (5.9) |
| **Sex at birth, n (%)** | |
| Male | 506 (47.8) |
| Female | 553 (52.2) |
| **Ethnicity, n (%)** | |
| Hispanic/Latino | 212 (20.0) |
| Not Hispanic/not Latino | 589 (55.6) |
| Not reported[b] | 236 (22.3) |
| Unknown[c] | 22 (2.1) |
| **Race, n (%)** | |
| American Indian/Alaskan Native | 3 (0.3) |
| Asian | 37 (3.5) |
| Black/African American | 175 (16.5) |
| Native Hawaiian/Pacific Islander | 6 (0.6) |
| White | 520 (49.1) |
| Other | 56 (5.3) |
| Not reported[b] | 262 (24.7) |

514

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

[a]All subjects <1 year old are counted as age 0.

[b]A clinical site that was a mobile drive-through site was not able to collect race/ethnicity identification from subjects.

[c]Unknown category indicates subjects for whom the corresponding information is not available.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

521

522 Table 2: Subject medical history (prospective symptomatic and asymptomatic subjects)

523

| Characteristics | Population statistics |
|---|---|
| **Total** | |
| N | 1,059 |
| **Subject received the influenza vaccine within the last 6 weeks, n (%)** | |
| Yes | 14 (1.3) |
| No | 1,045 (98.7) |
| Unknown | 0 (0.0) |
| **Influenza vaccine route of administration, n (%)[a]** | |
| Intramuscular | 14 (100.0) |
| Intranasal | 0 (0.0) |
| Unknown | 0 (0.0) |
| **Subject received a COVID-19 vaccine, n (%)** | |
| Yes | 728 (68.7) |
| No | 303 (28.6) |
| Unknown | 28 (2.6) |
| **Type of COVID-19 vaccine received for first dose, n (%)[b]** | |
| Pfizer | 506 (69.5) |
| Moderna | 185 (25.4) |
| Johnson & Johnson (J & J) | 30 (4.1) |
| Other | 3 (0.4) |
| Unknown | 4 (0.5) |
| **Type of COVID-19 vaccine received for second dose, n (%)[b]** | |
| Pfizer | 488 (72.2) |
| Moderna | 182 (26.9) |
| Other | 3 (0.4) |
| Unknown | 3 (0.4) |
| **Type of COVID-19 vaccine received first booster, n (%)[b,c]** | |
| Pfizer | 217 (64.0) |
| Moderna | 115 (33.9) |

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

| Characteristics | Population statistics |
|---|---|
| Johnson & Johnson (J & J) | 5 (1.5) |
| Other | 1 (0.3) |
| Unknown | 1 (0.3) |

Abbreviations: COVID-19, coronavirus disease 2019

[a]Percentages calculated based on number of subjects who received the influenza vaccine within the last 6 weeks prior to enrollment and sample collection.

[b]Percentages calculated based on number of subjects who received COVID-19 vaccine or first booster.

[c]There were six subjects that reported receiving a second booster.

533 Table 3: Results for the detection of SARS-CoV-2 using the composite comparator compared with the Cobas® Liat® SARS-CoV-2 & Influenza

534 A/B test[a]

| | Composite comparator (+) | Composite comparator (-) | *Total* | PPA % (95% CI) | NPA % (95% CI) |
|---|---|---|---|---|---|
| **Nasal** | | | | | |
| Cobas Liat SARS-CoV-2 & Influenza A/B test (+) | 165 | 9 | 174 | ***97.1 (93.3, 98.7)*** | ***99.1 (98.2, 99.5)*** |
| Cobas Liat SARS-CoV-2 & Influenza A/B test (-) | 5 | 961 | 966 | | |
| ***Total*** | ***170*** | ***970*** | ***1,140*** | | |
| **Nasopharyngeal** | | | | | |
| Cobas Liat SARS-CoV-2 & Influenza A/B test (+) | 162 | 5 | 167 | ***96.4 (92.4, 98.4)*** | ***99.4 (98.7, 99.8)*** |
| Cobas Liat SARS-CoV-2 & Influenza A/B test (-) | 6 | 896 | 902 | | |
| ***Total*** | ***168*** | ***901*** | ***1,069*** | | |

535

536 Abbreviations: CI, Score confidence interval; NPA, negative percent agreement; PPA, positive percent agreement; SARS-CoV-2, severe acute

537 respiratory syndrome coronavirus 2.

538 [a]The Cobas SARS-CoV-2 & Influenza A/B nucleic acid test for use on the Cobas Liat system is herein referred to as the Cobas Liat SARS-CoV-2

539 & Influenza A/B test.



medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

## References


1. Mathieu E, Ritchie H, Rodés-Guirao L, Appel C, Giattino C, Hasell J, Macdonald B, Dattani S, Beltekian D, Ortiz-Ospina E, Roser M. 2020. Coronavirus pandemic (COVID-19). https://ourworldindata.org/coronavirus. Accessed Mar 06, 2024.
2. Ahmad FB, Cisewski JA, Xu J, Anderson RN. 2023. COVID-19 Mortality Update - United States, 2022. MMWR Morb Mortal Wkly Rep 72:493–496.
3. Meyerowitz EA, Scott J, Richterman A, Male V, Cevik M. 2023. Clinical course and management of COVID-19 in the era of widespread population immunity. Nat Rev Microbiol 22:75–88.
4. Antinori A, Bausch-Jurken M. 2023. The burden of COVID-19 in the immunocompromised patient: implications for vaccination and needs for the future. J Infect Dis 228:S4–S12.
5. Wiemken TL, Khan F, Puzniak L, Yang W, Simmering J, Polgreen P, Nguyen JL, Jodar L, McLaughlin JM. 2023. Seasonal trends in COVID-19 cases, hospitalizations, and mortality in the United States and Europe. Sci Rep 13:3886.
6. Moriyama M, Hugentobler WJ, Iwasaki A. 2020. Seasonality of Respiratory Viral Infections. Annu Rev Virol 7:83–101.
7. World Health Organization. 2024. Coronavirus disease (COVID-19): similarities and differences between COVID-19 and influenza. https://www.who.int/news-room/questions-and-answers/item/coronavirus-disease-covid-19-similarities-and-differences-with-influenza. Accessed Aug 05, 2024.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

8. Baldanti F, Ganguly NK, Wang G, Möckel M, O'Neill LA, Renz H, Dos Santos Ferreira CE, Tateda K, Van Der Pol B. 2022. Choice of SARS-CoV-2 diagnostic test: challenges and key considerations for the future. Crit Rev Clin Lab Sci 59:445–459.

9. Hayden MK, Hanson KE, Englund JA, Lee MJ, Loeb M, Lee F, Morgan DJ, Patel R, El Mikati IK, Iqneibi S, Alabed F, Amarin JZ, Mansour R, Patel P, Falck-Ytter Y, Morgan RL, Murad MH, Sultan S, Bhimraj A, Mustafa RA. 2023. The Infectious Diseases Society of America guidelines on the diagnosis of coronavirus disease 2019 (COVID-19): molecular diagnostic testing. Clin Infect Dis doi:10.1093/cid/ciad646.

10. Katzenschlager S, Brümmer LE, Schmitz S, Tolle H, Manten K, Gaeddert M, Erdmann C, Lindner A, Tobian F, Grilli M, Pollock NR, Macé A, Erkosar B, Carmona S, Ongarello S, Johnson CC, Sacks JA, Denkinger CM, Yerlikaya S. 2023. Comparing SARS-CoV-2 antigen-detection rapid diagnostic tests for COVID-19 self-testing/self-sampling with molecular and professional-use tests: a systematic review and meta-analysis. Sci Rep 13:21913.

11. Tolan NV, Horowitz GL. 2022. Clinical diagnostic point-of-care molecular assays for SARS-CoV-2. Clin Lab Med 42:223–236.

12. Shirley JD, Bennett SA, Binnicker MJ. 2023. Current regulatory landscape for viral point-of-care testing in the United States. J Clin Virol 164:105492.

13. Fragkou PC, Moschopoulos CD, Dimopoulou D, Ong DSY, Dimopoulou K, Nelson PP, Schweitzer VA, Janocha H, Karofylakis E, Papathanasiou KA, Tsiordras S, De Angelis G, Thölken C, Sanguinetti M, Chung HR, Skevaki C. 2023. Performance of point-of care molecular and antigen-based tests for

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

SARS-CoV-2: a living systematic review and meta-analysis. Clin Microbiol Infect 29:291–301.

14. Bigio J, MacLean EL, Das R, Sulis G, Kohli M, Berhane S, Dinnes J, Deeks JJ, Brümmer LE, Denkinger CM, Pai M. 2023. Accuracy of package inserts of SARS-CoV-2 rapid antigen tests: a secondary analysis of manufacturer versus systematic review data. Lancet Microbe 4:e875–e882.

15. Wang X, Li Y, O'Brien KL, Madhi SA, Widdowson MA, Byass P, Omer SB, Abbas Q, Ali A, Amu A, Azziz-Baumgartner E, Bassat Q, Abdullah Brooks W, Chaves SS, Chung A, Cohen C, Echavarria M, Fasce RA, Gentile A, Gordon A, Groome M, Heikkinen T, Hirve S, Jara JH, Katz MA, Khuri-Bulos N, Krishnan A, de Leon O, Lucero MG, McCracken JP, Mira-Iglesias A, Moïsi JC, Munywoki PK, Ourohiré M, Polack FP, Rahi M, Rasmussen ZA, Rath BA, Saha SK, Simões EA, Sotomayor V, Thamthitiwat S, Treurnicht FK, Wamukoya M, Yoshida LM, Zar HJ, Campbell H, Nair H. 2020. Global burden of respiratory infections associated with seasonal influenza in children under 5 years in 2018: a systematic review and modelling study. Lancet Glob Health 8:e497–e510.

16. Paget J, Spreeuwenberg P, Charu V, Taylor RJ, Iuliano AD, Bresee J, Simonsen L, Viboud C. 2019. Global mortality associated with seasonal influenza epidemics: new burden estimates and predictors from the GLaMOR Project. J Glob Health 9:020421.

17. GBD 2017 Influenza Collaborators. 2019. Mortality, morbidity, and hospitalisations due to influenza lower respiratory tract infections, 2017: an analysis for the Global Burden of Disease Study 2017. Lancet Respir Med 7:69–89.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

18. Maleki F, Welch V, Lopez SMC, Cane A, Langer J, Enstone A, Markus K, Wright O, Hewitt N, Whittle I. 2023. Understanding the global burden of influenza in adults aged 18-64 years: a systematic literature review from 2012 to 2022. Adv Ther 40:4166–4188.

19. Gibson J, Schechter-Perkins EM, Mitchell P, Mace S, Tian Y, Williams K, Luo R, Yen-Lieberman B. 2017. Multi-center evaluation of the Cobas® Liat® Influenza A/B & RSV assay for rapid point of care diagnosis. J Clin Virol 95:5–9.

20. Young S, Phillips J, Griego-Fullbright C, Wagner A, Jim P, Chaudhuri S, Tang S, Sickler J. 2020. Molecular point-of-care testing for influenza A/B and respiratory syncytial virus: comparison of workflow parameters for the ID Now and Cobas Liat systems. J Clin Pathol 73:328–334.

21. Azar MM, Landry ML. 2018. Detection of influenza A and B viruses and respiratory syncytial virus by use of Clinical Laboratory Improvement Amendments of 1988 (CLIA)-waived point-of-care assays: a paradigm shift to molecular tests. J Clin Microbiol 56:e00367-18.

22. Melchers WJG, Kuijpers J, Sickler JJ, Rahamat-Langendoen J. 2017. Lab-in-a-tube: real-time molecular point-of-care diagnostics for influenza A and B using the Cobas® Liat® system. J Med Virol 89:1382–1386.

23. Uyeki TM, Bernstein HH, Bradley JS, Englund JA, File TM, Fry AM, Gravenstein S, Hayden FG, Harper SA, Hirshon JM, Ison MG, Johnston BL, Knight SL, McGeer A, Riley LE, Wolfe CR, Alexander PE, Pavia AT. 2019. Clinical practice guidelines by the Infectious Diseases Society of America: 2018 update on

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

diagnosis, treatment, chemoprophylaxis, and institutional outbreak management of seasonal influenza. Clin Infect Dis 68:895–902.

24. Berry L, Lansbury L, Gale L, Carroll AM, Lim WS. 2020. Point of care testing of Influenza A/B and RSV in an adult respiratory assessment unit is associated with improvement in isolation practices and reduction in hospital length of stay. J Med Microbiol 69:697–704.

25. Rahamat-Langendoen J, Groenewoud H, Kuijpers J, Melchers WJG, van der Wilt GJ. 2019. Impact of molecular point-of-care testing on clinical management and in-hospital costs of patients suspected of influenza or RSV infection: a modeling study. J Med Virol 91:1408–1414.

26. Mikamo H, Koizumi Y, Yamagishi Y, Asai N, Miyazono Y, Shinbo T, Horie M, Togashi K, Robbins EM, Hirotsu N. 2022. Comparing the Cobas Influenza A/B Nucleic acid test for use on the Cobas Liat System (Liat) with rapid antigen tests for clinical management of Japanese patients at the point of care. PLoS One 17:e0276099.

27. Roche Molecular Systems Inc. 2021. Cobas® Liat® System User Guide.

28. Roche Molecular Systems Inc. 2021. Instructions for use. Cobas® SARS-CoV-2. Qualitative assay for use on the Cobas® 6800/8800 Systems. P/N: 09179917001-07EN.

29. Roche Molecular Systems Inc. 2021. Instructions for use. Cobas® SARS-CoV-2 & Influenza A/B. Qualitative assay for use on the Cobas® 6800/8800 Systems. P/N: 09233474190.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

30. Mostafa HH, Hardick J, Morehead E, Miller JA, Gaydos CA, Manabe YC. 2020. Comparison of the analytical sensitivity of seven commonly used commercial SARS-CoV-2 automated molecular assays. J Clin Virol 130:104578.

31. Copan Diagnostics Inc. 2020. Instructions for use. FLOQSwabs. HPC217 Rev.00 Date 2020.10. https://www.copanusa.com/wp-content/uploads/2021/03/HPC217-PI-FLOQSWAB-NO-CE-REV.00-2020.10.pdf. Accessed Aug 5, 2024.

32. Centers for Disease Control and Prevention. 2023. 2021-2022 flu season summary. https://www.cdc.gov/flu/season/faq-flu-season-2021-2022.htm#print. Accessed Aug 05, 2024.

33. Dzung A, Cheng PF, Stoffel C, Tastanova A, Turko P, Levesque MP, Bosshard PP. 2021. Prolonged unfrozen storage and repeated freeze-thawing of SARS-CoV-2 patient samples have minor effects on SARS-CoV-2 detectability by RT-PCR. J Mol Diagn 23:691–697.

34. Roche Molecular Systems Inc. 2020. Instructions for use. Cobas® SARS-CoV-2 & Influenza A/B. Nucleic acid test for use on the Cobas® Liat® System. P/N: 09211101190.

35. Hologic. 2020. Instructions for use. Aptima® SARS-CoV-2 Assay (Panther® System).

36. Wilson E. 1927. Probable inference, the law of succession, and statistical inference. J Am Stat Assoc 22:209–212.

37. Akashi Y, Horie M, Kiyotaki J, Takeuchi Y, Togashi K, Adachi Y, Ueda A, Notake S, Nakamura K, Terada N, Kurihara Y, Kiyasu Y, Suzuki H. 2022. Clinical

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

performance of the cobas Liat SARS-CoV-2 & Influenza A/B Assay in nasal samples. Mol Diagn Ther 26:323–331.

38. Hansen G, Marino J, Wang ZX, Beavis KG, Rodrigo J, Labog K, Westblade LF, Jin R, Love N, Ding K, Garg S, Huang A, Sickler J, Tran NK. 2021. Clinical performance of the point-of-care Cobas Liat for detection of SARS-CoV-2 in 20 minutes: a multicenter study. J Clin Microbiol 59:e02811-20.

39. Park K, Sung H, Kim MN. 2023. Evaluation of the cobas Liat detection test for SARS-CoV-2 and influenza viruses following the emergence of the SARS-CoV-2 Omicron variant. Diagn Microbiol Infect Dis 105:115891.

40. Er TK, Chou YC, Chen SY, Huang JW. 2021. Rapid cobas Liat SARS-CoV-2 assay in comparison with the laboratory-developed real-time RT-PCR test. Clin Lab 67.

41. Matic N, Lawson T, Ritchie G, Lowe CF, Romney MG. 2024. Testing the limits of multiplex respiratory virus assays for SARS-CoV-2 at high cycle threshold values: Comparative performance of Cobas 6800/8800 SARS-CoV-2 & Influenza A/B, Xpert Xpress SARS-CoV-2/Flu/RSV, and Cobas Liat SARS-CoV-2 & Influenza A/B. J Assoc Med Microbiol Infect Dis Can 8:328-335.

42. Lee RA, Herigon JC, Benedetti A, Pollock NR, Denkinger CM. 2021. Performance of saliva, oropharyngeal swabs, and nasal swabs for SARS-CoV-2 molecular detection: a systematic review and meta-analysis. J Clin Microbiol 59:e02881-20.

43. Therchilsen JH, von Buchwald C, Koch A, Dam Nielsen S, Rasmussen DB, Thudium RF, Kirkby NS, Raaschou-Pedersen DET, Bundgaard JS, Iversen K, Bundgaard H, Todsen T. 2020. Self-collected versus healthcare worker-

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

collected swabs in the diagnosis of severe acute respiratory syndrome coronavirus 2. Diagnostics 10:678.

44. Craney AR, Velu PD, Satlin MJ, Fauntleroy KA, Callan K, Robertson A, La Spina M, Lei B, Chen A, Alston T, Rozman A, Loda M, Rennert H, Cushing M, Westblade LF. 2020. Comparison of two high-throughput reverse transcription-PCR systems for the detection of severe acute respiratory syndrome coronavirus 2. J Clin Microbiol 58:e00890-20.

45. Buchta C, Zeichhardt H, Badrick T, Coucke W, Wojtalewicz N, Griesmacher A, Aberle SW, Schellenberg I, Jacobs E, Nordin G, Schweiger C, Schwenoha K, Luppa PB, Gassner UM, Wagner T, Kammel M. 2023. Classification of "near-patient" and "point-of-care" SARS-CoV-2 nucleic acid amplification test systems and a first approach to evaluate their analytical independence of operator activities. J Clin Virol 165:105521.

46. May L, Robbins EM, Canchola JA, Chugh K, Tran NK. 2023. A study to assess the impact of the Cobas point-of-care RT-PCR assay (SARS-CoV-2 and Influenza A/B) on patient clinical management in the emergency department of the University of California at Davis Medical Center. J Clin Virol 168:105597.

47. Jian MJ, Chung HY, Chang CK, Lin JC, Yeh KM, Chen CW, Li SY, Hsieh SS, Liu MT, Yang JR, Tang SH, Perng CL, Chang FY, Shang HS. 2021. Clinical comparison of three sample-to-answer systems for detecting SARS-CoV-2 in B.1.1.7 lineage emergence. Infect Drug Resist 14:3255–3261.

48. Perlitz B, Slagman A, Hitzek J, Riedlinger D, Möckel M. 2021. Point-of-care testing for influenza in a university emergency department: a prospective study. Influenza Other Respir Viruses 15:608–617.

medRxiv preprint doi: https://doi.org/10.1101/2024.09.06.24313202; this version posted September 6, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity. It is made available under a CC-BY 4.0 International license .

49. Yan X, Li K, Lei Z, Luo J, Wang Q, Wei S. 2023. Prevalence and associated outcomes of coinfection between SARS-CoV-2 and influenza: a systematic review and meta-analysis. Int J Infect Dis 136:29–36.