Deuterium metabolic imaging of the human abdomen at clinical field strength

Pascal Wodtke[1,2], Mary A McLean[1,3], Ines Horvat-Menih[1,2], Jonathan R Birchall[1], Maria J Zamora-Morales[1], Ashley Grimmer[1,2], Elizabeth Latimer[1], Marta Wylot[1], Rolf F Schulte[4], Ferdia A Gallagher[1,3]

Author Affiliations

[1]University of Cambridge, Department of Radiology, Cambridge CB2 0QQ, United Kingdom

[2]University of Cambridge, Cancer Research UK Cambridge Centre, Cambridge CB2 0RE, United Kingdom

[3]University of Cambridge, Cancer Research UK Cambridge Institute, Cambridge CB2 0RE, United Kingdom

[4]GE HealthCare, Munich, Germany

Abstract

Background

The Warburg effect is a hallmark of cancer and is characterized by increased glucose consumption and lactate formation. Deuterium metabolic imaging (DMI) is an emerging non-invasive MRI method for probing this metabolic reprogramming in the field of neuroimaging. Here we show the feasibility of the technique for abdominal imaging using a routine 3 T MRI system, which has previously presented significant technical challenges.

Purpose

This study aimed to translate abdominal DMI to clinical field strength by optimizing the radiofrequency coil setup, the administered dose of deuterium (^{2}H)-labelled glucose, and the data processing pipeline for quantitative characterization of DMI signals over time in the kidney and liver, establishing a basis for routine clinical studies in the future.

Materials and Methods

Five healthy volunteers were recruited and imaged on 2 or 3 occasions, with different ^{2}H-glucose doses (totalling 13 DMI scan sessions). DMI was performed at 3 T using a flexible 20 x 30 cm^2 ^{2}H-tuned transmit-receive surface coil. We have defined three novel quantitative parameters as metrics of metabolism and compared these between doses and organs.

Results

The careful positioning of a dedicated surface coil minimized unwanted gastric signals while maintaining excellent hepatic and renal measurements. The timecourses derived from the liver and kidney were reproducible and comparable across different doses, with a trend towards lower

36 quantitative measurements with decreasing dose. An increase in the ^{2}H-water signal over time

37 particularly in the liver, could be used as an indirect measure of metabolism.

Conclusion

39 DMI of the human abdomen is feasible using a routine MRI system and the metabolism measured in

40 the kidney and liver can serve as a reference for future clinical studies. The ^{2}H-glucose dose can be

41 reduced from 0.75 to 0.25 g/kg to minimize gastric signal without substantially affecting the

42 reliability of organ quantification.

Introduction

Aerobic glycolysis, or the Warburg effect, occurs in a wide range of cancers and describes the increase

in glucose breakdown which occurs even in the presence of abundant oxygen[1]. This effect is

characterized by upregulated glucose consumption and the formation of downstream lactate that

accumulates in the extracellular space, creating a hostile acidic extracellular tumor environment. A

number of novel metabolic MRI methods have recently emerged to exploit this metabolism which

have the potential for identifying more aggressive tumor subtypes and for assessing early metabolic

response to treatments[2–5].

In 2018, deuterium metabolic imaging (DMI) emerged as one of these non-invasive methods to assess

glycolytic metabolism in the brain and cerebral tumors[6]. Unlike conventional proton (^{1}H) MRI,

deuterium (^{2}H) is found at very low natural abundance (~0.01%), and therefore deuterium spectra

contain almost no background signal apart from the large peak arising from naturally-abundant, and

partially-deuterated water (^{2}H-water) signal, with some smaller deuterated peaks from adipose tissue.

DMI exploits this low background signal through the administration of exogenous ^{2}H-labelled

molecules such as ^{2}H-glucose ([6,6'-^{2}H$_2$]glucose)[7–11]. This labelled glucose is very safe[12] and can be

conveniently administered orally, thereby facilitating clinical use of the technique (Fig. 1a). Following

absorption in the gastrointestinal tract, ^{2}H-glucose metabolism can be followed spatially and

temporally using deuterium magnetic resonance spectroscopic imaging (^{2}H-MRSI)[6,9].

Despite having great potential for assessing glycolysis, DMI has faced several challenges in its

translation into clinical field strength[7,13]. Whilst DMI has recently been translated to 3 T in the

brain[9,14], DMI outside of the brain suffers from additional challenges including: increased tissue

inhomogeneity, motion, poor shimming, prominent gastric signal from the oral probe, and overlap

between the naturally-abundant lipid peak and the tumor ^{2}H-lactate signal, all of which reduce

spectral quality. For these reasons, abdominal DMI has only been conducted at 7 T to date[8,10].

In this study, we have addressed these challenges and conducted abdominal DMI in healthy volunteers

at 3 T. The oral ^{2}H-glucose dose has been varied to explore the optimal dose required for clinical

imaging, to reduce both the cost and the gastric signal arising from the oral ingestion of the probe,

without affecting the desired signal from the abdominal organs[15]. We demonstrate the feasibility of

quantitative abdominal DMI of the liver and kidney at clinical field strength and present a robust

pipeline for assessing ^{2}H-glucose spectra for future clinical use including novel metrics of

metabolism.

		²H-Glucose [g]	Dose [g/kg]	Dose label	Age range [a]	Sex	Height [cm]	Weight [kg]	BMI [kg/m²]
Vol 1	Visit 1	60.00	0.69	high				87.0	23.6
	Visit 2	40.00	0.46	medium	26-30	m	192	87.0	23.6
	Visit 3	20.00	0.23	low				87.0	23.6
Vol 2	Visit 1	20.00	0.19	low	26-30	m	189	108.0	30.2
	Visit 2	60.00	0.56	high				107.2	30.0
Vol 3	Visit 1	40.00	0.50	medium				80.0	22.9
	Visit 2	60.00	0.75	high	20-25	m	187	80.0	22.9
	Visit 3	19.55	0.25	low				78.2	22.4
Vol 4	Visit 1	40.88	0.75	high	26-30	f	164	54.5	20.3
	Visit 2	13.50	0.25	low				54.0	20.1
Vol 5	Visit 1	33.50	0.50	medium				67.0	19.8
	Visit 2	50.18	0.75	high	26-30	m	184	66.9	19.8
	Visit 3	17.00	0.25	low				68.0	20.1

Table 1 | Demographics. Details about the individual volunteers and the dose received at each visit.

Figure 1 | Study schema and experimental procedure. (a) DMI as a technique. A solution containing ²H-glucose is prepared and orally administered, followed by spectroscopic imaging. **(b)** n = 5 volunteers were imaged on two (n = 2) or three (n = 3) different days, totalling 13 imaging sessions. At each visit, the volunteers received one of three dose regimens of the ²H-glucose.

Materials and Methods

Study Design

The study was approved by a local research ethics committee (23/YH/0127) and all five recruited healthy volunteers (4 male, 1 female, age range = 20-30, Table 1) gave informed consent in writing. To assess the effect of dose reduction, the volunteers underwent up to three DMI studies on different days, receiving different doses of ^{2}H-glucose (Cambridge Isotope Laboratories Inc., Tewksbury, MA), dissolved in water for injection (B. Braun, Melsungen, Germany), as shown in Table 1. The three dose regimens chosen for this study include the current standard dose given to volunteers and patients in previously published studies[9,11] ("high"), as well as a dose reduction of one-third ("medium") and two-thirds ("low") of the standard dose. The "high" dose corresponds to 0.75 g/kg, with a maximum administration of 60 g. All five volunteers received the high and the low dose, while three volunteers also received the medium dose, totalling n = 13 DMI sessions. The volunteers were fasted for 6 h pre-scan. Prior to, and immediately after each exam, blood pressure, heart rate, blood oxygen saturation, and blood glucose level were recorded (Fig. S1).

^{2}H-MR Spectroscopy and Imaging

The volunteers were positioned supine in the MRI (3 T Premier, GE Healthcare, Chicago, IL) head-first. A flexible 20x30 cm^2 ^{2}H-transmit-receive surface coil (RAPID Biomedical, Rimpar, Germany) was wrapped around the right side of the abdomen, and fixed via Velcro straps, with the center of the coil positioned between the liver and kidney. An axial T_1-weighted volume (LAVA-flex) was acquired for co-registration using the built-in body coil. Subsequently, a series of Hamming-filtered density-weighted ^{2}H-MRSI (real matrix size = 10x10x10, 1678 interleaves, FOV = 40 cm, FA = 60°, TE = 0.7 ms, TR = 250 ms) were acquired for ≤90 min, interleaved between acquisitions, by the collection of unlocalized ^{2}H spectra (TE = 0.7 ms, TR = 600 ms, FA = 90°, 64 averages).

Data Processing and Analysis

Spectra were zero-filled by a factor of two in all three spatial dimensions and fitted automatically using a customised version of OXSA-AMARES[16,17] with an SNR threshold of five. For display in Fig. 5, spatial interpolation by a factor of 12.8 in plane and a factor of 2 in the z-dimension was conducted. Voxels were manually assigned to the respective organs using an overlay of the summed ^{2}H-signal and

the axial anatomical LAVA-flex [1]H localizer (FOV = 48 cm, FA = 12°, TEs = 1.1/2.2 ms, TR = 3.96

ms), as seen in Fig. 3. For liver, kidney, and duodenum, the prior knowledge of the fitting routine

included three peaks: ^{2}H-water, ^{2}H-glucose, and ^{2}H-lipid+lactate, as those commonly expected for DMI

studies in the abdomen[10]. For the stomach, the prior knowledge was set to detect only one peak, assumed

to be ^{2}H-glucose in the absence of metabolism. Peak amplitudes, as fitted by OXSA-AMARES, were

used to calculate the GGW and LLW (equation 1 and 2) ratio for every organ and timepoint using

MATLAB (MathWorks, Natick, MA). The GGW and LLW ratio represent the respective normalization

of ^{2}H-glucose and ^{2}H-lipid+lactate signals against the naturally abundant ^{2}H-water peak to compare

across scans as defined in equations 1 and 2:

$$GGW(t) = \frac{\text{gluc}(t)}{\text{gluc}(t)+\text{wat}(t)} \qquad\qquad \text{equation 1}$$

$$LLW(t) = \frac{\text{liplac}(t)}{\text{liplac}(t)+\text{wat}(t)} \qquad\qquad \text{equation 2}$$

with gluc(t), liplac(t) and wat(t) representing the ^{2}H-glucose, ^{2}H-lipid+lactate (whose peaks overlap)

and ^{2}H-water signals at the respective timepoints. The GGW provides a normalized metric of ^{2}H-glucose

uptake, which is comparable across participants as a potentially useful clinical measurement. LLW is a

similar comparative metric, which at baseline is dominated by the lipid signal and any increase over

time corresponds to lactate formation in the organ of interest. For quantitative analysis of the

timecourses, GGW_{max}, GGW_{AUC}, and $GGW_{mean\ plateau}$ were calculated. GGW_{max} was determined by

taking the maximum GGW value in each organ timecourse. For GGW_{AUC}, the GGW value of each

timepoint t_n was multiplied by the time difference t_n-t_{n-1} between the previous timepoint in minutes

(equation 3). The separate AUC(t) values were then summed until t = 74 min, which was a common

timepoint across volunteers and exams, to make the GGW_{AUC} comparable across volunteers and across

scans:

$$GGW_{AUC}(x) = \sum_{n=1}^{x} GGW(t_n) \cdot (t_n - t_{n-1})\,; \text{with } t_x = 74 \text{ min} \qquad \text{equation 3}$$

$GGW_{mean\ plateau}$ was determined by taking the mean GGW of all timepoints after signal plateauing. The

plateau time was determined to be t = 30 min for liver, and t = 40 min for kidney, broadly aligning with

previously published studies[10]. For unlocalized spectroscopy, the $GGW_{mean\ plateau}$ was replaced by

GGW_{mean} over the whole timecourse, since the detection of signal build-up, and subsequent plateauing,

was not possible for all scans, depending on the contribution of the stomach to the overall detected

signal. For statistical analysis, Analysis of Variance (ANOVA) was performed in GraphPad Prism

(GraphPad Software, Boston, MA), followed by Tukey's Honestly Significant Difference (HSD) test as

post-hoc analysis.

Results

Unlocalized Spectroscopy

Unlocalized spectra showed good SNR (> 5) for all three dose regimens, with the ^{2}H-water (4.7 ppm) and ^{2}H-glucose (3.7 ppm) peaks being clearly resolved at clinical field strength (Fig. 2a-c). Interestingly, the GGW ratio derived from these unlocalized spectra showed ^{2}H-glucose build-up steadily over time in some volunteers, while appearing early in others (Fig. 2d). This is because, depending on the coil positioning, the stomach and duodenum were detectable in some cases, resulting in an early and high ^{2}H-glucose signal in some cases. The low dose regimen exhibited a lower GGW across all metrics (Fig. 2e; $GGW_{max} = 0.44 \pm 0.03$; $GGW_{mean} = 0.35 \pm 0.05$; $GGW_{AUC} = 24.1 \pm 5.0$; n = 4) compared to the medium ($GGW_{max} = 0.55 \pm 0.07$, p = 0.0902; $GGW_{mean} = 0.48 \pm 0.05$, p = 0.0193; $GGW_{AUC} = 28.8 \pm 2.7$; p = 0.5097, n = 3) and the high doses ($GGW_{max} = 0.57 \pm 0.07$, p = 0.0247; $GGW_{mean} = 0.47 \pm 0.05$, p = 0.0148; $GGW_{AUC} = 33.8 \pm 6.4$, p = 0.0560; n = 5). Statistically significant results are indicated in Fig. 2e. Notably, there was no significant difference in quantitative values between the medium and the high dose across all the metrics. The dynamics of the fitted ^{2}H-glucose signal amplitude showed a signal build-up over 60 min in most cases (Fig. 2f). An increase in the ^{2}H-water signal over time could be observed in all cases, ranging from 10-56% when comparing the last timepoint with the first in each case (Fig. 2g-h, Fig. S2).

Figure 2 | Dynamic unlocalized spectroscopy acquired from the healthy abdomen after oral administration of ^{2}H-glucose. (a-c) Spectral timecourses demonstrating ^{2}H-water and ^{2}H-glucose peaks from a single volunteer (#3) after receiving the following ^{2}H-glucose doses: (a) high; (b) medium; (c) low. **(d)** Timecourse of GGW (see methods) for each volunteer at each visit. **(e)** The three different quantitative measures we have introduced to evaluate ^{2}H-glucose signal over time (see methods). **(f)** The total ^{2}H-glucose signal initially increases with respect to the first timepoint in all volunteers and starts to decline after ~50-60 minutes. **(g)** The ^{2}H-water grows moderately over time in all volunteers, hinting suggesting the production of water through metabolism. **(h)** First (start) vs. last (end) timepoint of the data shown in (g), to demonstrate that all volunteers exhibit a growth in the ^{2}H-water signal, thus suggesting a consistent effect.

Imaging

Anatomically constrained ^{2}H signal was identified in the liver and kidney on MRSI in all cases (n = 13; Fig. 3), and in the stomach and duodenum in most cases (n = 10 and 11 respectively). Only ^{2}H-glucose could be detected in the stomach, while the duodenum exhibited a small but detectable ^{2}H-

water peak. The kidney exhibited a small lipid peak which was probably secondary to partial volume

effects from the surrounding visceral fat.

The timecourses for the kidney (Fig. 4a) and the liver (Fig. 4b) showed a build-up in the GGW ratio,

with the liver plateauing earlier at ~30 min, compared to the kidney at ~40 min. GGW signal in the

duodenum remained high throughout the whole timecourse (Fig. 4c). The raw or non-normalized ^{2}H-

glucose signal was assessed in the stomach due to the low ^{2}H-water signal, which showed a decrease

over time in all cases as gastric emptying occurred (Fig. 4d).

Figure 3 | Representative images derived from spatially resolved spectroscopy. The summed ^{2}H
signal is shown as a colorscale image with single-voxel spectra from each organ (volunteer 5, t = 33
min after receiving the high (gold) and the medium (blue) doses. Stomach and duodenal ROIs are
dominated by the ^{2}H-glucose signal, whereas kidney and liver exhibit a prominent ^{2}H-water peak as
well as a small lipid peak in the kidney. The boxes in the images indicate non-interpolated voxel sizes.

Comparison between the liver and kidney, and between doses

Across all three dose regimens and all three quantitative metrics, the liver exhibited a significantly higher ^{2}H-glucose uptake than the kidney (Fig. 4e-h, p-values for GGW_{max}/$GGW_{mean\ plateau}$/GGW_{AUC}: 0.0097/0.0032/0.0044). Furthermore, all three values increased with dose in both the liver and kidney. Notably, this trend was statistically significant between the low dose and the other two doses, but not between the high and the medium dose. Relevant p-values are shown in Figure S3. Example imaging timecourses from volunteer 5 for all three doses are displayed in Figure 5.

^{2}H-lipid content and ^{2}H-lactate formation

The ^{2}H-lactate signal appears at a similar chemical shift as the ^{2}H-lipid signal (both at ~1.3 ppm)[18], and as lipid is particularly abundant in the abdomen[19,20] this presents a challenge for the accurate quantification of ^{2}H-lactate. Hence, it is more important to characterize the naturally abundant lipid signal on abdominal DMI, compared to the brain where the majority of previous DMI research has been undertaken.

A small ^{2}H-lipid peak was identified in voxels overlying the kidney, and to a lesser extent the liver, which was likely due to partial volume effects from the visceral fat surrounding these organs (Fig. 3). The LLW ratio was calculated and the respective timecourses for both organs are displayed in Fig. 4i/j. The lipid content was higher in the kidney (mean across all volunteers and timepoints: 0.27 ± 0.10) compared to the liver (0.21 ± 0.08).

Lactate plays an important role in many pathological processes and therefore it is key to quantify the background ^{2}H-lactate signal derived from normal organs after oral administration of ^{2}H-glucose as a reference. Fig. 6a shows the normalized ^{2}H-lipid+lactate signal for the three volunteers that were administered all three doses. A direct comparison of the last timepoint compared to the first for all volunteers can be seen in Fig. 6b-c. In the liver, the ^{2}H-lipid+lactate signal increased in 12/13 examinations, with an average increase of $38.5 \pm 24.2\%$ showing that the healthy liver exhibits detectable ^{2}H-lactate 60-80 min after administering oral ^{2}H-glucose. In contrast, the ^{2}H-lipid+lactate signal increased in only 8/13 examinations in the kidney, with an average increase of only $5.0 \pm 17\%$, suggesting very little lactate formation in this timescale.

kidney timecourse

liver timecourse

duodenum timecourse

stomach timecourse

a

b

c

d

legend: high dose (orange), medium dose (blue), low dose (green)

GGW$_{AUC}$

GGW$_{mean\ plateau}$

GGW$_{max}$

e

f

g

	GGW$_{AUC}$			GGW$_{mean\ plateau}$			GGW$_{max}$		
	high dose	medium dose	low dose	high dose	medium dose	low dose	high dose	medium dose	low dose
liver	31.3 ± 2.6	29.5 ± 1.8	26.6 ± 2.2	0.51 ± 0.01	0.48 ± 0.02	0.43 ± 0.02	0.54 ± 0.02	0.51 ± 0.02	0.48 ± 0.03
kidney	26.5 ± 2.6	25.3 ± 2.5	22.4 ± 2.6	0.46 ± 0.02	0.42 ± 0.02	0.34 ± 0.03	0.49 ± 0.04	0.46 ± 0.04	0.41 ± 0.05
duodenum	51.5 ± 5.0	53.9 ± 3.9	49.1 ± 3.6	0.76 ± 0.06	0.78 ± 0.07	0.70 ± 0.10	0.86 ± 0.07	0.85 ± 0.05	0.85 ± 0.04

h

kidney lipid timecourse

liver lipid timecourse

i

j

Figure 4 | Quantitative evaluation of imaging data. (a-c) GGW timecourses for ^{2}H-glucose signal in the kidney (a), liver (b) and duodenum (c). **(d)** ^{2}H-glucose signal in the stomach over time shows a decrease likely to correspond to gastric emptying. **(e-g)** Quantitative measures of GGW$_{AUC}$ (e), GGW$_{mean\ plateau}$ (f) and GGW$_{max}$ (g) exhibit dose dependence in the liver and kidney, whereas they remain elevated in the duodenum regardless of dose. **(h)** Summarized quantitative values displayed in (e-g). **(i,j)** ^{2}H-lipid signal, as evaluated by the LLW parameter (see methods) over time. The kidney (i) shows a generally higher LLW than the liver (j), indicating a higher lipid content inside the imaged kidney voxels, which is likely to arise from the surrounding visceral fat.

^{2}H-water signal increase as an indirect measure of metabolism

The unlocalized spectroscopy showed an increase in ^{2}H-water signal over time (Fig. 2g-h), which has been reported in previous preclinical and clinical studies[9,10,21]. Spectroscopic imaging was subsequently used to localize this signal increase; Fig. 6d shows the normalized ^{2}H-water signal for the three volunteers that were administered all three doses. A direct comparison of the last timepoint compared to the first for all volunteers can be seen in Figures 6e-f. In the kidney, the ^{2}H-water signal increased over time in 10/13 examinations, with an average increase of $10.3 \pm 12.8\%$. In the liver, the ^{2}H-water signal increased in all 13 examinations, with an average increase of $17.6 \pm 9.7\%$. Hence, the liver is likely driving the ^{2}H-water signal increase in the unlocalized spectroscopy data. Indeed, the three examinations with the largest ^{2}H-water signal increases in the unlocalized spectra were the same three examinations that exhibited the largest ^{2}H-water signal increase in the liver imaging data. This observation further confirms the conclusion, that the liver is causing the dominant part of the ^{2}H-water signal increase in the unlocalized spectra.

liver/stomach

a

kidney/duodenum

b

Figure 5 | Imaging timecourses for ^{2}H-glucose for all three dose regimens administered to volunteer 5. (a) A slice containing both the liver and stomach showing declining ^{2}H-glucose signal in the stomach and increasing signal in the liver. **(b)** A slice containing both the kidney and the duodenum: the kidney shows a build-up of ^{2}H-glucose, whereas the duodenum signal shows little variation in signal intensity over time. Timecourses of the ^{2}H-water signal, as well as the summed ^{2}H-signal can be viewed in the supplementary information (Fig. S4, S5).

Figure 6 | ^{2}H-water and ^{2}H-lipid+lactate signal derived from individual volunteers over time, extracted from the imaging data. (a) ^{2}H-lipid+lactate signal from the three volunteers that underwent all three imaging sessions. Unlike the kidney, the liver showed a trend towards an increasing signal over time, up to x2.5 the initial value. **(b,c)** Comparison of the first and last timepoints from all the volunteers reveals a trend towards increasing ^{2}H-lipid+lactate signal in the liver (c), whereas the kidney (b) shows no trend across volunteers. The liver data suggests the hepatic formation of ^{2}H-lactate from ^{2}H-glucose in the timecourse of the experiment. **(d)** Comparative ^{2}H-water signal of the same three volunteers. The liver shows a clear and consistent increase in ^{2}H-water signal over time. **(e,f)** Comparison of the first and last timepoints from all the volunteers reveals a

273 trend towards increasing ^{2}H-water signal in both kidney (e) and liver (f) over time. In the case of the
274 liver, the ^{2}H-water signal increased in all volunteers.

Discussion

Deuterium metabolic imaging is an emerging method for probing tissue metabolism with potential
applications in a wide range of diseases given the importance of glucose utilization in both health and
disease. Here we translate the technique to the abdomen at clinical field strength, showing its
feasibility for assessing hepatic and renal metabolism. We also optimize the dose of oral ^{2}H-glucose
for clinical use and present metrics to quantitatively evaluate its uptake and metabolism in tissues.

The signal from ^{2}H-glucose was evaluated up to 90 min after oral administration and could be used to
evaluate accumulation and absorption in the stomach and duodenum. Importantly, the use of a surface
coil in combination with its positioning around the right side of the abdomen was used to minimize
unwanted gastric signal when imaging the liver and kidney. The liver generally exhibited a higher ^{2}H-
glucose signal compared to the kidney, rapid plateauing of the signal, more reproducible time curves,
and a lower lipid signal, therefore making it a promising organ to be assessed using DMI.

The increase in ^{2}H-water signal over time is likely to arise from water derived from glucose
metabolism[22,23]. The liver showed a higher ^{2}H-glucose uptake than the kidney, as well as the formation
of a small ^{2}H-lipid+lactate signal, in keeping with increased hepatic metabolism. A confounder when
analyzing DMI data is that the long timecourse for acquisition allows for metabolites from other
organs to be washed into the organ under assessment so the measured signal may not be due to
metabolism in the organ under investigation.

The doses of ^{2}H-glucose used in DMI range from 60-75 g[8–11,24], with 75 g being the standard dose for
a clinical glucose tolerance test[25]. The blood glucose is transiently increased during these dose
regimens, and we show that this could be avoided by lowering the dose to ~20 g, allowing steady state
metabolism to be probed without a substantial loss in signal (Fig. S1). This reduces the cost of DMI as
well as the risk of unwanted gastric signal contaminating the quantification from other organs.

As an emerging technique, the best approaches for quantitative assessment and kinetic modelling of
DMI are yet to be determined. The challenges are similar to other time-resolved metabolic imaging
methods, such as those found in the analysis of positron emission tomography (PET)[26] and
hyperpolarized ^{13}C-MRI data[27,28]. However, DMI also presents some problems unique to the technique
and here we have proposed three novel quantitative biomarkers for assessment of this dynamic data:
GGW_{AUC}, $GGW_{mean\ plateau}$, and GGW_{max}. The results show that the hepatic and renal measurements are
influenced by the administered dose but do not scale with it, suggesting a rate limiting element in

delivery and metabolism. It is interesting to note that, despite the range of doses given across different
volunteers, the duodenal measurements are remarkably constant over time and between individuals.

The increase in the ^{2}H-lipid+lactate signal observed in the liver is likely to be from lactate formation
as lipogenesis occurs on a slower timescale[29]. The lipid signal demonstrates a low SNR and has a
broader linewidth compared to the ^{2}H-water and ^{2}H-glucose and may be overestimated while fitting
with the automated approach here, which calculates a peak area.

One limitation of this study is the small cohort size; extending this study into larger cohorts and
undertaking formal repeatability and reproducibility studies in a multisite setting will be important for
future research in this field. Normalized plots have been referenced to the first timepoint of the same
exam and this reference timepoint differed across examinations. It would be advantageous to
reference the ^{2}H-water and ^{2}H-lipid signals against the naturally abundant measurements at baseline,
but as the participants were repositioned after drinking in the upright position, the exact placement of
the coil could change, potentially resulting in misregistration. Future studies could overcome this by
maintaining the participant on the scanner with the coil in position and administering the ^{2}H-glucose
using a straw[8], allowing the natural abundance scan to be directly compared to the post-administration
images. Intravenous injection of ^{2}H-glucose could also be considered in the future, bypassing
gastrointestinal uptake, which would reduce the required dose further and lead to more rapid substrate
delivery. However, this may result in a transient supraphysiological dose of glucose which could alter
tissue metabolism.

Conclusion

In conclusion, these data demonstrate the feasibility of quantitative DMI in the kidney and liver at
clinical field strength, without signal interference by the stomach. Furthermore, we show that the
administered dose can be significantly reduced, reducing the costs of a DMI scan and the otherwise
dominant gastric signal, which are important considerations for future clinical use. Additionally, we
introduced new methodology to assess glucose uptake and metabolism, which could be used to
quantify tissue metabolism in disease settings.

Acknowledgements

PW acknowledges support from the Gates Cambridge Trust (#OPP1144). MM acknowledges support from the Cambridge Experimental Cancer Medicine Centre. This research was further supported by Cancer Research UK (CRUK; C19212/A27150; C19212/A29082), CRUK Cambridge Centre, Wellcome Trust, NIHR Cambridge Biomedical Research Centre (BRC-1215-20014). The views expressed are those of the authors and not necessarily those of the NIHR or the Department of Health and Social Care. The authors also acknowledge support from the National Cancer Imaging Translational Accelerator (NCITA) and The Mark Foundation Institute for Integrative Cancer Medicine (MFICM) at the University of Cambridge.

References

1. Warburg, O. ueber den Stoffwechsel der Carcinomzelle. *Naturwissenschaften* **12**, 1131–1137 (1924) doi: 10.1007/BF01504608.

2. Tekade, R. K. & Sun, X. The Warburg effect and glucose-derived cancer theranostics. *Drug Discov. Today* **22**, 1637–1653 (2017) doi: 10.1016/j.drudis.2017.08.003.

3. Hesketh, R. L. & Brindle, K. M. Magnetic resonance imaging of cancer metabolism with hyperpolarized ^{13}C-labeled cell metabolites. *Curr. Opin. Chem. Biol.* **45**, 187–194 (2018) doi: 10.1016/j.cbpa.2018.03.004.

4. Kreis, F. *et al.* Measuring Tumor Glycolytic Flux in Vivo by Using Fast Deuterium MRI. *Radiology* **294**, 289–296 (2020) doi: 10.1148/radiol.2019191242.

5. Low, J. C. M. *et al.* Deuterium metabolic imaging differentiates glioblastoma metabolic subtypes and detects early response to chemoradiotherapy. *Cancer Res.* (2024) doi:10.1158/0008-5472.CAN-23-2552 doi: 10.1158/0008-5472.CAN-23-2552.

6. De Feyter, H. M. *et al.* Deuterium metabolic imaging (DMI) for MRI-based 3D mapping of metabolism in vivo. *Sci. Adv.* **4**, eaat7314 (2018) doi: 10.1126/sciadv.aat7314.

7. De Feyter, H. M. & De Graaf, R. A. Deuterium metabolic imaging – Back to the future. *J. Magn. Reson.* **326**, 106932 (2021) doi: 10.1016/j.jmr.2021.106932.

8. Poli, S. *et al.* Interleaved trinuclear MRS for single-session investigation of carbohydrate and lipid metabolism in human liver at 7T. *NMR Biomed.* **37**, e5123 (2024) doi: 10.1002/nbm.5123.

9. Kaggie, J. D. *et al.* Deuterium metabolic imaging and hyperpolarized 13C-MRI of the normal human brain at clinical field strength reveals differential cerebral metabolism. *NeuroImage* **257**, 119284 (2022) doi: 10.1016/j.neuroimage.2022.119284.

10. Gursan, A. *et al.* Deuterium body array for the simultaneous measurement of hepatic and renal glucose metabolism and gastric emptying with dynamic 3D deuterium metabolic imaging at 7 T. *NMR Biomed.* **36**, e4926 (2023) doi: 10.1002/nbm.4926.

11. Khan, A. S. *et al.* Deuterium Metabolic Imaging of Alzheimer Disease at 3-T Magnetic Field Strength: A Pilot Case-Control Study. *Radiology* **312**, e232407 (2024) doi: 10.1148/radiol.232407.

12. Jones, P. J. H. & Leatherdale, S. T. Stable isotopes in clinical research: safety reaffirmed. *Clin. Sci.* **80**, 277–280 (1991) doi: 10.1042/cs0800277.

13. Graaf, R. A. *et al.* On the magnetic field dependence of deuterium metabolic imaging. *NMR Biomed.* **33**, e4235 (2020) doi: 10.1002/nbm.4235.

14. Adamson, P. M. *et al.* Deuterium metabolic imaging for 3D mapping of glucose metabolism in humans with central nervous system lesions at 3T. *Magn. Reson. Med.* **91**, 39–50 (2024) doi: 10.1002/mrm.29830.

15. Ahmadian, N. *et al.* Human Brain Deuterium Metabolic Imaging at 7 T : Impact of Different [6,6′-^{2}H$_2$]Glucose Doses. *J. Magn. Reson. Imaging* jmri.29532 (2024) doi:10.1002/jmri.29532 doi: 10.1002/jmri.29532.

16. Purvis, L. A. B. *et al.* OXSA: An open-source magnetic resonance spectroscopy analysis toolbox in MATLAB. *PLOS ONE* **12**, e0185356 (2017) doi: 10.1371/journal.pone.0185356.

17. Vanhamme, L., van den Boogaart, A. & Van Huffel, S. Improved Method for Accurate and Efficient Quantification of MRS Data with Use of Prior Knowledge. *J. Magn. Reson.* **129**, 35–43 (1997) doi: 10.1006/jmre.1997.1244.

18. Fan, T. W.-M. Metabolite profiling by one- and two-dimensional NMR analysis of complex mixtures. *Prog. Nucl. Magn. Reson. Spectrosc.* **28**, 161–219 (1996) doi: 10.1016/0079-6565(95)01017-3.

19. Htun, K. T. *et al.* Advanced Molecular Imaging (MRI/MRS/1H NMR) for Metabolic Information in Young Adults with Health Risk Obesity. *Life* **11**, 1035 (2021) doi: 10.3390/life11101035.

20. Tkáč, I. *et al.* In vivo [1] H NMR spectroscopy of the human brain at 7 T. *Magn. Reson. Med.* **46**, 451–456 (2001) doi: 10.1002/mrm.1213.

21. Montrazi, E. T., Sasson, K., Agemy, L., Scherz, A. & Frydman, L. Molecular imaging of tumor metabolism: Insight from pyruvate- and glucose-based deuterium MRI studies. *Sci. Adv.* **10**, eadm8600 (2024) doi: 10.1126/sciadv.adm8600.

22. Mellanby, K. Metabolic Water and Desiccation. *Nature* **150**, 21–21 (1942) doi: 10.1038/150021a0.

23. Sawka, M. N., Cheuvront, S. N. & Carter, R. Human Water Needs. *Nutr. Rev.* **63**, S30–S39 (2005) doi: 10.1111/j.1753-4887.2005.tb00152.x.

24. Bøgh, N., Vaeggemose, M., Schulte, R. F., Hansen, E. S. S. & Laustsen, C. Repeatability of deuterium metabolic imaging of healthy volunteers at 3 T. *Eur. Radiol. Exp.* **8**, 44 (2024) doi: 10.1186/s41747-024-00426-4.

25. Jagannathan, R. *et al.* The Oral Glucose Tolerance Test: 100 Years Later. *Diabetes Metab. Syndr. Obes. Targets Ther.* **Volume 13**, 3787–3805 (2020) doi: 10.2147/DMSO.S246062.

26. Graham, M. M., Peterson, L. M. & Hayward, R. M. Comparison of simplified quantitative analyses of FDG uptake. *Nucl. Med. Biol.* **27**, 647–655 (2000) doi: 10.1016/S0969-8051(00)00143-8.

27. Larson, P. E. Z. *et al.* Investigation of analysis methods for hyperpolarized 13C-pyruvate metabolic MRI in prostate cancer patients. *NMR Biomed.* **31**, (2018) doi: 10.1002/nbm.3997.

28. Daniels, C. J. *et al.* A comparison of quantitative methods for clinical imaging with hyperpolarized [13] C-pyruvate. *NMR Biomed.* **29**, 387–399 (2016) doi: 10.1002/nbm.3468.

417   29. Schwarz, J.-M., Clearfield, M. & Mulligan, K. Conversion of Sugar to Fat: Is Hepatic de Novo

418   Lipogenesis Leading to Metabolic Syndrome and Associated Chronic Diseases? *J. Osteopath.*

419   *Med.* **117**, 520–527 (2017) doi: 10.7556/jaoa.2017.102.

421