Diagnostic accuracy of Impedance Spectroscopy versus Digital Rectal Examination for

Obstetric Anal Sphincter Injuries: a postpartum post-hoc analysis.

Stefano Salvatore[1], Katarzyna Borycka[2,3], Alessandro Ruffolo[1,4], Marcel Młyńczak[2,5], Maciej

Rosoł[2,5], Kacper Korzeniewski[2,5], Piotr Iwanowski[2], Antonino Spinelli[6,7], Renaud DeTayrac[8],

Carlo Ratto[9], Stavros Athanasiou[10], Diaa Rizk[11], Andrea Stuart[12], Jan Baekelandt[13],

Małgorzata Uchman-Musielak[14], Hynek Heřman[15], Petr Janku[16,17], Peter Rosenblat[18],

Mariusz Grzesiak[19], Adam Dziki[20], Ruwan Fernando[21]

1 Obstetrics and Gynecology Unit, Vita-Salute San Raffaele University Milan, Italy IRCCS San Raffaele
Scientific Institute, Milan, Italy
2 Department of General Surgery, Lazarski University Faculty of Medicine, Warsaw, Poland
3 OASIS Diagnostics, R&D of Medical Technology, Warsaw, Poland
4 Department of Gynecology, Jeanne de Flandre University Hospital, Lille, France
5 Warsaw University of Technology, Faculty of Mechatronics, Institute of Metrology and Biomedical
Engineering, Warsaw, Poland
6 Department of Biomedical Sciences, Humanitas University, Via Rita Levi Montalcini 4, 20072 Pieve
Emanuele, Milan, Italy
7 IRCCS Humanitas Research Hospital, via Manzoni 56, 20089 Rozzano, Milan, Italy
8 Department of Obstetrics and Gynaecology, Nimes University Hospital, Nîmes, France
9 Proctology and Pelvic Floor Surgery Unit, Isola Tiberina Hospital, Gemelli Isola, Catholic University, Rome,
Italy
Institute for the Care of Mother and Child, Prague, Czech Republic
10 Urogynecology Unit, Alexandra Hospital, Kapodistrian University of Athens, Greece
11Department of Obstetrics and Gynecology and Clinical Research Centre, College of Medicine and Medical
Sciences, Arabian Gulf University, Bahrain
12 Department of Obstetrics and Gynecology, Lund University, Lund, Sweden
13 Department of Obstetrics and Gynaecology, Imelda Hospital, Bonheiden, Belgium
14 Department of Gynaecology and Obstetrics, Madalinski Hospital, Warsaw, Poland
15 3rd Medical Faculty, Charles University, Prague, Czech Republic
16 Department of Obstetrics and Gynaecology, University Hospital Brno and Medical Faculty, Masaryk
University, Brno, Czech Republic
17 Department of Health Sciences, Medical Faculty, Masaryk University, Brno, Czech Republic
18 Division of Urogynecology and Reconstructive Pelvic Surgery, Cambridge Department of Obstetrics and
Gynecology, Harvard Medical School, Boston, Massachusetts, USA
19 Institute of Polish Mother's Health Centre, Łódź, Poland
20 Department of General and Colorectal Surgery, Medical University of Lodz, Poland
21 Department of Urogynaecology, St Mary's Hospital, Imperial College Healthcare NHS Trust, London, UK
39

40   Corresponding author: Katarzyna Borycka, MD,PhD
41   Postal address: Department of General Surgery, Medicover Hospital, 5 Rzeczypospolitej
42   Street, Warsaw, Poland
43   Email address: kborycka@interia.pl
44

Abstract

Background: Accurate diagnosis of obstetric anal sphincter injuries (OASIs) is critical for timely repair and prevention of long-term morbidity, yet digital rectal examination (DRE) remains insufficiently sensitive.

Methods: In this post-hoc analysis of a prospective multicentre diagnostic study (NCT04903977), we compared the diagnostic performance of DRE and machine learning–assisted impedance spectroscopy with three-dimensional endoanal ultrasound (EAUS) as the reference. A total of 152 women who delivered vaginally were assessed up to 8 weeks postpartum.

Results: DRE demonstrated a sensitivity of 44.3%, specificity of 83.5%, and overall accuracy of 67.8%. Impedance spectroscopy achieved 90.6% sensitivity, 84.6% specificity, and 87.0% accuracy against EAUS.

Conclusions: DRE frequently fails to detect OASIs postpartum. A minimally invasive, operator-independent impedance-based method significantly improves diagnostic sensitivity and warrants further evaluation for use immediately after delivery to enable timely primary repair.

Keywords: obstetric anal sphincter injury, digital rectal examination, impedance spectroscopy, endoanal ultrasound, faecal incontinence, postpartum diagnostics

Introduction

Obstetric anal sphincter injuries (OASIs) represent a major clinical challenge in maternal care [1-10]. Despite being widely underdiagnosed, they are strongly associated with long-term complications, including faecal incontinence, pain, and pelvic floor dysfunction [1,8] The reported incidence of OASI ranges widely, from 0.1% to 6% in routine clinical records [4,5],

yet population-level imaging studies suggest a true prevalence exceeding 25% [11]. Early

detection and appropriate repair of OASI—ideally within 8–12 hours post-delivery—is critical

to achieving optimal outcomes and avoiding delayed surgical intervention[12-14].

Digital rectal examination (DRE), although widely accessible, and recommended by

guidelines [12-24], shows poor sensitivity and high operator dependency [2,25]. Endoanal

ultrasound (EAUS), considered the reference standard for OASI detection, is limited in

routine delivery settings due to its cost, availability, and technical challenges when used

immediately postpartum. In particular, studies have shown that edema, hematoma, and gas

artifacts compromise the reliability of EAUS in the hours following delivery [26]

Alternative imaging approaches, such as transperineal or introital ultrasound, are easier to

apply and use standard obstetric equipment [27]. However, they lack precision in grading

sphincter defects and are not currently endorsed by international guidelines [17, 28]. This

creates a critical diagnostic gap during the window when timely repair is most effective.

To address this gap, we evaluated a machine learning–based impedance spectroscopy

system designed to detect structural changes in the anal sphincter complex. While previous

pilot studies have demonstrated the feasibility of this technique [29-31], the current post-hoc

analysis from the pivotal study [32] assesses its diagnostic performance in a controlled early

postpartum setting, using 3D-EAUS as the reference standard.

Our objective was to compare the sensitivity, specificity, and diagnostic accuracy of

impedance spectroscopy and DRE, focusing on their ability to identify OASI in the

postpartum period when EAUS reliability is optimal. These results may inform future use of

impedance-based diagnostics as a complementary tool to clinical examination in postpartum

care.

Materials and Methods

Study Design and Population

This post-hoc analysis is based on data from a prospective, comparative, multicentre diagnostic study [32] conducted between May 2021 and December 2022 at five European centres. The study was designed to validate the diagnostic performance and safety of a machine learning–assisted impedance spectroscopy system for detecting OASI, using three-dimensional endoanal ultrasound (EAUS) as the reference standard (ClinicalTrials.gov identifier: NCT04903977).

Women aged 18–49 who delivered vaginally at ≥34 weeks gestation were eligible for inclusion. Both primiparous and multiparous women were enrolled. Exclusion criteria included: (1) pre-existing faecal incontinence unrelated to OASI; (2) prior perineal or anal surgery; (3) active inflammatory bowel disease; (4) cardiac devices or arrhythmias; and (5) any acute or chronic conditions that would interfere with study procedures.

The study adhered to ISO 14155:2020 for good clinical practice in medical device trials and followed the principles of the Declaration of Helsinki. Ethical approval was obtained at each centre (see Declarations)*, and written informed consent was obtained from all participants.

Study Procedures

Participants were recruited within 8 weeks postpartum, allowing assessments to be performed after the resolution of acute delivery-related oedema and bleeding. The study included three visits:

- − Visit 1: Clinical history, physical examination, 12-lead ECG, Wexner incontinence score, and 3D-EAUS.
- − Visit 2: Within 7 days of Visit 1, participants underwent impedance spectroscopy and digital rectal examination (DRE).
- − Visit 3: Follow-up safety assessment (not included in this analysis).

Reference and Index Tests

Three-dimensional endoanal ultrasound (EAUS) served as the reference standard. Scans

were performed using a 360° 3D endoprobe by experienced colorectal specialists. EAUS

results were evaluated using three validated scales: the OASIS classification (grades 3a–4),

the Norderval score (quantitative sphincter defect), and a modified Starck classification

(anatomical mapping of internal and external sphincter injury). Results were finalized and

electronically locked before impedance test was performed.

DRE was conducted bi-digitally following international guidelines [12], with the result

recorded as either "OASI suspected" or "OASI not suspected." All examiners had prior

experience in perineal trauma assessment.

Impedance spectroscopy was performed using a single-use endoanal probe connected to a

handheld spectrometer (ONIRY system). Measurements lasted under one minute in

lithotomy position. The embedded machine learning algorithm returned a binary output

("PASS" or "REFER") without operator interpretation or external processing.

Blinding

Clinicians performing EAUS were blinded to the results of both index tests. DRE

examinations were conducted by trained investigators who were not involved in EAUS

assessment. The impedance spectroscopy system provided fully automated classification

based on internal algorithms, eliminating the need for blinding. Operator-independent

processing ensures reproducibility and minimizes bias in impedance-based results.

Device Description and ML Model

The ONIRY system includes an impedance spectrometer and a single-use endoanal probe,

which records tissue impedance in a supine, gynaecological position over approximately one

minute. Impedance data are analysed using an embedded neural network trained to detect

tissue patterns associated with OASI. The device provides a binary output: "PASS" (no OASI)

or "REFER" (suspected OASI), along with probable injury localization. All ML processing

occurs within the device; no external software is required.

The model was validated using 10 iterations of 10-fold cross-validation on impedance data,

classified against the EAUS reference.

Independent Factors Assessed

To understand factors affecting impedance readings, potential confounders such as residual

fluids, blood, and inflammation (monitored via calprotectin levels) were independently

analysed. This approach ensures the robustness of the impedance data and supports future

refinement of ONIRY performance in early postpartum settings. Follow-up studies focused

exclusively on day 1 postpartum use are ongoing.

Post-Hoc Grouping

Participants were reclassified into two groups based on EAUS findings:

– Group I (No OASI): Including intact perineum, first- or second-degree perineal tears

– Group II (OASI present): Third- or fourth-degree sphincter injuries

Statistical Analysis

Diagnostic performance was calculated for each index test using EAUS as the reference

standard. Primary outcomes were sensitivity, specificity, and accuracy. They were

categorized as Success (True Positive, TP or True Negative, TN) or Failure (False Positive,

FP, or False Negative, FN). Secondary analysis included performance by OASI grade (3a–

4) and area under the receiver operating characteristic (ROC) curve.

Cross-validation metrics for the machine learning model were reported as mean ± standard

deviation across 10 iterations.

Results

A total of 152 women were included in the final analysis. Baseline demographic and clinical

characteristics are presented in Table 1.

Group reallocation based on EAUS revealed that 61 participants had confirmed OASI (Group

II), while 91 participants had no OASI (Group I).

The diagnostic accuracy, sensitivity, and specificity of the DRE and ONIRY systems

compared to EAUS as the reference standard were evaluated. The performance of DRE in

identifying OASIs indicated an accuracy of 67.8%, reflecting the proportion of total diagnoses

(both positive and negative) that were correctly identified. Sensitivity, which measured the

ability of DRE to correctly identify those with the OASI presence (true positive rate), was

relatively low at 44.3%. On the other hand, specificity, measuring the ability to correctly

identify those without the condition (true negative rate), was higher at 83.5%.

The machine learning-assisted impedance spectroscopy was evaluated by comparing the

trained ML models' output with EAUS results. The system achieved an accuracy of 87.0% ±

0.5%, which represents a significant enhancement in the overall ability to diagnose OASIs

correctly. The sensitivity of the ONIRY system was 90.6% ± 2.0%, indicating its strong

capability to identify true positive cases of OASIs effectively. Specificity was robust at 84.6%

± 1.9%, suggesting that the ONIRY system maintains a high level of accuracy in confirming

the absence of injuries, similar to DRE.

The ONIRY performance metrics for each iteration of 10-fold cross-validation and the overall

statistical analysis are given in Table 2.

The accuracy of the two compared methods in detecting OASI in each OASIS classification

grades 3a, 3b, 3c, and 4 is shown in Table 3.

The comparison of receiver operating characteristic (ROC) curves for ONIRY and DRE,

along with their respective area under the curve (AUC) values, is presented in Figure 1.

Discussion

This post-hoc analysis compared the diagnostic performance of DRE and machine learning–

assisted impedance spectroscopy in detecting OASIs, using three-dimensional endoanal

ultrasound (EAUS) as the reference standard. The results demonstrate a significant

difference in sensitivity between the two methods: DRE correctly identified only 44.3% of

confirmed OASIs, whereas impedance spectroscopy achieved a sensitivity of 90.6%, with

overall diagnostic accuracy exceeding 87%.

DRE remains a first-line tool in postpartum care due to its accessibility, yet its performance

is highly dependent on examiner training and experience. DRE was more effective in

confirming the absence of injury than detecting its presence and had the poorest accuracy

in assessing the extent of external sphincter injury (24% and 53.8% in OASIS grades 3a and

3b, respectively). Our results reinforce prior studies [25,33] showing that DRE frequently fails

to detect external sphincter injuries, especially grade 3a and 3b tears, which often present

subtly on palpation. This is particularly concerning given that accurate differentiation between

second- and third-degree tears is essential for determining the need for surgical repair. And

this is what is most challenging for clinicians in obstetric practice: how to differentiate

between perineal injury grade 2 which can be addressed by a midwife, and grade 3, when

the external sphincter is injured, and an experienced obstetrician (or colorectal surgeon) with

expertise in OASI is required for repair.

Impedance spectroscopy, by contrast, provides objective, operator-independent classification based on biophysical tissue properties. In this study, it was able to detect OASIs with high sensitivity across all injury grades, including 3a. The system's performance was validated using cross-validated machine learning on study-derived data, enhancing reproducibility. While EAUS remains the imaging reference standard, its practical limitations—especially immediately after delivery due to oedema, bleeding, and anatomical distortion—are well-documented [26], and limit its availability in real-world delivery settings.

Importantly, this study did not evaluate the performance of impedance spectroscopy in the immediate postpartum period. Instead, the analysis was deliberately conducted during the early postpartum window (up to 8 weeks post-delivery) to optimize EAUS reliability and enable robust training of the machine learning model. This design enhances internal validity but limits generalizability to the delivery room setting.

The clinical consequences of undiagnosed or misclassified OASIs are substantial. Delayed recognition often leads to incomplete or suboptimal repair, resulting in persistent sphincter defects and long-term morbidity. Meta-analyses by Okeahialam et al. [34] report rates of anal incontinence exceeding 18% following OASI, with higher rates among more severe tear grades (15.6% for 3a, 18.3% for 3b, 20.6% for 3c, and 28.4% for 4th-degree tears). Taithongchai et al. [35]and Roper et al.)[36] have shown that missed or underdiagnosed OASIs are associated with worse continence outcomes and higher need for secondary colorectal procedures. In contrast, structured repair immediately after delivery is associated with favourable continence (7% pof fecal incontinence, with an additive 24% rate of mild gas incontinence symptoms) and quality-of-life outcomes [37,38].

Cost-effectiveness analyses also favor early diagnosis demonstrated that primary repair leads to both improved quality-adjusted life years and lower long-term healthcare costs compared with delayed sphincteroplasty [39]. These findings support efforts to strengthen

postpartum OASI detection, particularly during the therapeutic window in which surgical

repair can be most effective.

This study's strengths include its multicentre design, use of a high-quality imaging reference

standard, and blinded EAUS interpretation. However, several limitations must be

acknowledged. First, the diagnostic accuracy of DRE may be artificially high in this study due

to examiner expertise, which may exceed that of typical maternity settings. Second, the

impedance spectroscopy model was trained and validated on the same study population;

while cross-validation mitigates overfitting, external validation is needed. Finally, the study

does not address the diagnostic performance of the impedance system in the immediate

postpartum window—a critical time for enabling primary repair.

Future research should focus on validating impedance spectroscopy in real-time delivery

settings, ideally in prospective cohorts where early intervention decisions depend on rapid

and reliable detection. Additionally, comparative studies with transperineal ultrasound may

help define the role of impedance-based methods within the broader spectrum of postpartum

diagnostic tools.

Conclusion

Digital rectal examination identified only 44% of obstetric anal sphincter injuries in the early

postpartum period, while impedance spectroscopy achieved over 90% sensitivity and 87%

accuracy compared with 3D endoanal ultrasound. As a minimally invasive, operator-

independent method, it may offer critical diagnostic support when injury cannot be confidently

excluded on clinical examination. These findings justify further investigation of impedance

spectroscopy in the immediate postpartum setting, where timely detection is essential to

enable primary repair and prevent long-term complications.

Acknowledgments

The authors would like to thank all the other investigators and co-investigators involved in

the clinical trial from centres in the Czech Republic (Brno and Prague), Slovakia (Kosice),

Poland (Warsaw), and Spain (Leon) for their invaluable contributions to patient recruitment,

clinical trials conduct, and assistance with results.

Disclosure of interests

A.S. is an independent Ethicon, Takeda, Pfizer, and Sofar consultant.

M.U.M, H.H. and P.J. received remuneration as study investigators.

S.S., R.DT., A.S., H.H., C.R., S.A., D.R., A.St., J.B., M.G., A.D., and R.F. are independent

consultants of OASIS Diagnostics' Scientific Advisory Board.

K.B. is an independent consultant, and a Takeda trainer; a founder and management board

member at OASIS Diagnostics, the author of the related patent.

Contribution to Authorship

Conceptualization: S.S., R.F. K.B., M.M., P.I.

Data curation: S.S., M.M., M.R., K.K., P.I.

Formal analysis: M.M., M.R., K.K.

Funding acquisition: K.B., M.M.

Investigation: M.U.M., H.H., P.J.

Methodology: S.S., R.DT., A.S., H.H., C.R., S.A., D.R., A.St., J.B., R.Fr., M.G., A.D., and

R.F.

Project administration: K.B., M.M.

Resources: H.H., P.J.

Software: M.R., K.K.

Supervision: R.DT., A.S., H.H., C.R., S.A., D.R., A.St., J.B., R.Fr., M.G., A.D., and R.F.

Validation: M.M., P.I.

Visualization: M.R., K.K.

Writing the original draft: S.S., K.B., A.R., A.S. M.M., M.R., K.K.

Editing the final draft: A.St., D.R., M.U.M., A.R.

*Details of Ethics Approval

The study received approval from the ethics committees of the respective study sites on the

following dates: 19 March 2021 by the Ethics Committee of the Institute for Maternal and

Child Care (approval no. 1/19.03.2021), 27 April 2021 by the Ethics Committee for Research

with Medicines of the Health Areas of León and Bierzo (approval no. 2186), 9 June 2021 by

the Ethics Committee of the University Hospital of Brno (approval no. 47/21Zdrav.), 14

October 2021 by the Ethics Committee at the Regional Medical Chamber in Warsaw

(approval no. KB/1362/21), and 25 July 2022 by the Ethics Committee at AGEL Hospital

Košice-Šaca (approval no. ONIRY 3/2/2020).

Funding

The study was financed by the European Union as part of the Fast Track program, conducted

in Poland by the Polish National Centre for Research and Development (POIR.01.01.01-00-

0726/18).

References

[1] Sultan AH, Kamm MA, Hudson CN, Thomas JM, Bartram CI. Anal sphincter disruption

during vaginal delivery. N Engl J Med 1993; 329:1905-11.

[2] Groom KM, Paterson-Brown S. Can we improve on the diagnosis of third degree

tears? Eur J Obstet Gynecol Reprod Biol 2002;101:19 21.

[3] Thiagamoorthy G, Johnson A, Thakar R, Sultan AH. A national survey of perineal

trauma and its subsequent management in the United Kingdom. Int Urogynecol J

2014;25:1621–7.

[4] Gurol-Urganci I, Cromwell DA, Edozien LC, et al. Third- and fourth-degree perineal

tears among primiparous women in England between 2000 and 2012: time trends

and risk factors. BJOG 2013;120:1516–25

[5] Baghestan E, Irgens LM, Bordahl PE, Rasmussen S. Trends in risk factors for

obstetric anal sphincter injuries in Norway. Obstet Gynecol 2010;116:25–34.

[6] Zetterström J, López A, Anzén B, Norman M, Holmström B, Mellgren A. Anal sphincter

tears at vaginal delivery: risk factors and clinical outcome of primary repair. Obstet

Gynecol. 1999 Jul;94(1):21-8. PMID: 10389712.

[7] Lewicky-Gaupp C, Leader-Cramer A, Johnson LL, Kenton K, Gossett DR. Wound

complications after obstetric anal sphincter injuries. Obstet Gynecol. 2015

May;125(5):1088-1093. doi: 10.1097/AOG.0000000000000833. PMID: 25932836.

[8] Andrews V, Sultan AH, Thakar R, Jones PW. Occult anal sphincter injuries--myth or

reality? BJOG. 2006;113:195–200.

[9] Serati M, Ruffolo AF, Scancarello C, Braga A, Salvatore S, Ghezzi F. When does

oasis cause de novo pelvic floor dysfunction? role of the surgeon's skills. Int

Urogynecol J. 2023 Feb;34(2):493-498. doi: 10.1007/s00192-022-05205-6. Epub

2022 Apr 25. PMID: 35467138; PMCID: PMC9036836.

[10]   Zimmo K, Laine K, Vikanes A, Foss E, Zimmo M, Ali H, et al. Diagnosis and

repair of perineal injuries: knowledge before and after expert training multicentre

observational study among Palestinian physicians and midwives. BMJ Open

2017;7(4):e014183, 2017.

[11]   Sideris, M., McCaughey T, Hanrahan JG, Arroyo-Manzano D, Zamora J, Jha

S, Knowles CH, Thakar R, Chaliha C, Thangaratinam S. Risk of obstetric anal

sphincter injuries (OASIS) and anal incontinence: a meta-analysis. (2020). European

Journal of Obstetrics & Gynecology and Reproductive Biology, 252, 303-312.

[12]   Royal College of Obstetrics and Gynaecology. Management of third and

fourth degree perineal tears. Greentop guideline number 29 2015. Available from:

https://www.rcog.org.uk/media/5jeb5hzu/gtg-29.pdf

[13]   Department for Health and Wellbeing. South Australian Perinatal Practice

Guideline - Third and fourth degree tear management 2018. Available from:

www.sahealth.sa.gov.au/wps/wcm/connect/1faf87004eedec4db635b76a7ac0d6e4/T

hird+and+Fourth+Degree+Tear+Management_PPG_v5_1.pdf?MOD=AJPERES&a

mp;CACHEID=ROOTWORKSPACE-1faf87004eedec4db635b76a7ac0d6e4-oc-

RC8a

[14]   Queensland Clinical Guidelines. Queensland clinical guidelines: Perineal

Care 2023. Available from:

https://www.health.qld.gov.au/__data/assets/pdf_file/0022/142384/g-pericare.pdf

[15]   Harvey MA, Pierce M, Alter JE, Chou Q, Diamond P, Epp A, et al. Obstetrical

Anal Sphincter Injuries (OASIS): Prevention, Recognition, and Repair. J Obstet

Gynaecol Can.2015;37(12):1131-48.

[16]   Danish Association of Obstetrics and Gynaecology. Guideline for obstetric anal

sphincter injury (OASIS) 2019. Available from: https://nfog.org/wp-

content/uploads/2019/03/190313-obstetric-anal-sphincter-injury.pdf

[17]   Kropshofer A, Aigmüller T, Beilecke K, et al. Management of Third and Fourth-

Degree Perineal Tears After Vaginal Birth. Guideline of the DGGG, OEGGG, and

359  SGGG (S2k-Level, AWMF Registry No. 015/079, December 2020). Geburtshilfe

360  Frauenheilkd. 2022 Dec 7;83(2):165-183. doi: 10.1055/a-1933-2647.

361 [18]  Royal College of Physicians of Ireland. Institute of obstetricians and

362  gynaecologists and, directorate of clinical strategy and programmes, health service

363  executive. Management of obstetric anal sphincter injuries (2012).

364 [19]  Ministry of Health Mexico. Prevención, diagnóstico y tratamiento de

365  episiotomía complicada. 2015. Available from: https://www.cenetec-

366  difusion.com/CMGPC/IMSS-608-13/ER.pdf

367 [20]  Netherlands Society of Obstetrics and Gynaecology. Total rupture during

368  childbirth 2013-2024.

369 [21]  The American College of Obstetricians and Gynecologists. Practice bulletin no.

370  198 Prevention and Management of Obstetric Lacerations at vaginal delivery. Obstet

371  Gynecol. 2018;132:e87–102 . https://pubmed.ncbi.nlm.nih.gov/30134424/

372 [22]  Ducarme G, Pizzoferrato AC, de Tayrac R, Schantz C, Thubert T, Le Ray C,

373  Riethmuller D, Verspyck E, Gachon B, Pierre F, Artzner F, Jacquetin B, Fritel X.

374  Perineal prevention and protection in obstetrics: CNGOF clinical practice guidelines.

375  J Gynecol Obstet Hum Reprod. 2019 Sep;48(7):455-460. doi:

376  10.1016/j.jogoh.2018.12.002. Epub 2018 Dec 12.

377 [23]  Marty N, Verspyck E. Perineal tears and episiotomy: Surgical procedure -

378  CNGOF perineal prevention and protection in obstetrics guidelines. Gynecol Obstet

379  Fertil Senol. 2018

380 [24]  Flemish Society of Obstetrics and Gynecology (VVOG). CLINICAL

381  GUIDANCE PAPER VVOG. Obstetrical Anal Sphincter Injuries – OASIS. 2024

382 [25]  Shek KL, Atan IK, Dietz HP. Can Anal Sphincter Defects Be Identified by

383  Palpation? Female Pelvic Med Reconstr Surg. 2016 Nov/Dec;22(6):472-475

384 [26]  Huber M, Larsson C, Harrysson M, Strigård K, Lehmann JP, Nordin P, Tunón

K. Use of endoanal ultrasound in detecting obstetric anal sphincter injury immediately after birth. Acta Obstet Gynecol Scand. 2023 Mar;102(3):389-395.

[27]   Dietz, H., Exo-anal imaging of the anal sphincters: a pictorial introduction. J Ultrasound Med, 2018. 2018: p. 263-280.

[28]   AIUM/IUGA practice parameter for the performance of Urogynecological ultrasound examinations. Developed in collaboration with the ACR, the AUGS, the AUA, and the SRU. International Urogynecology Journal 2019;30:1389–1400

[29]   Młyńczak M, Borycka-Kiciak K, Uchman-Musielak M, Dziki A. Impedance Spectroscopy Method to Detect Pelvic Floor Muscle Damage – A Feasibility Study," In World Congress on Medical Physics and Biomedical Engineering, Springer, Singapore, pp. 875–878, 2018.

[30]   Borycka-Kiciak K, Młyńczak M, Kiciak A, Pietrzak P, Dziki A. Noninvasive obstetric anal sphincter injury diagnostics using impedance spectroscopy. Scientific Reports 2019;9(7097):1–9.

[31]   Młyńczak M, Rosoł M, Spinelli A, Dziki A, Wlazlak E, Surkont G, et al. Obstetric Anal Sphincter Injury Detection Using Impedance Spectroscopy with the ONIRY Probe. Appl Sci 2021;11:637, 2021.

[32]   Borycka K, Młyńczak M, Rosoł, Korzeniewski K, Iwanowski P, Heřman H, Janku P, Uchman-Musielak M, Dosedla E, Diaz EG, Sudoł-Szopińska I, Mik M, Ratto C, Spinelli A. Detection of obstetric anal sphincter injuries using machine learning-assisted impedance spectroscopy: a prospective, comparative, multicentre clinical study. Sci Rep. 2025 Mar 4;15(1):7522.

[33]   O'Leary BD, Kelly L, Fitzpatrick M, Keane DP. Underdiagnosis of internal anal sphincter trauma following vaginal delivery. Ultrasound Obstet Gynecol. 2023;61(2):251-256.

[34]   Okeahialam NA, Taithongchai A, Thakar R, Sultan AH. The incidence of anal

incontinence following obstetric anal sphincter injury graded using the Sultan

classification: a network meta-analysis. Am J Obstet Gynecol. 2023. PMID:

36379266 Review.

[35]   Taithongchai A, Veiga SI, Sultan AH, Thakar R. The consequences of

undiagnosed obstetric anal sphincter injuries (OASIS) following vaginal delivery. Int

Urogynecol J 2020;31:635–41

[36]   Roper JC, Thakar R, Sultan AH. Underclassified obstetric anal sphincter

injuries. Int Urogynecol J 2022;33:1473–9.

[37]   Andrews V, Thakar R, Sultan AH. Outcome of obstetric anal sphincter injuries

(OASIS)-role of structured management", Int Urogynecol J Pelvic Floor Dysfunct

2009; 20(8):973-8.

[38]   Norderval S, Öian P, Revhaug A, Vonen B. Anal Incontinence After Obstetric

Sphincter Tears: Outcome of Anatomic Primary Repairs. Diseases of the Colon &

Rectum 2005; 48(5):p 1055-1061.

[39]   Tan EK, Jacovides M, Khullar V, Teoh TG, Fernando RJ, Tekkis PP. A cost-

effectiveness analysis of delayed sphincteroplasty for anal sphincter injury. Colorectal

Disease 2008; 10(7):653-662.

Table/Figure Caption List

Table 1. Characteristics of the study population.

Table 2. Performance metrics of the ONIRY device in the assessment relative to 3-D

Endoanal Ultrasound and OASIS classification (each row shows the statistics for a single 10-

fold cross-validation each performed with different random seed).

Table 3. Performance metrics of DRE and ONIRY device in the assessment in reference to

EAUS under OASIS classification.

Figure 1. ROC curves for the ONIRY system and DRE are presented as blue and red lines,

respectively, with the identity line depicted as a black dashed line.

Tables

	Group I	Group II	Total
	(N=91)	(N=61)	(N=152)
Age (years)			
Mean ± SD	31.7 ± 4.6	30.7 ± 4.6	31.3 ± 4.6
Range (Min/Max)	22 (18/40)	20 (22/42)	24 (18/42)
Age categorized (n, %)			
<26 years	9 (9.9)	7 (11.5)	16 (10.5)
26<35 years	53 (58.2)	41 (67.2)	94 (61.8)
≥35 years	29 (31.9)	13 (21.3)	42 (27.6)
Weight (kg)			
Mean ± SD	71.7 ± 11.8	68.1 ± 9.6	70.2 ± 11.0
Range (Min/Max)	56 (49/105)	60 (48/108)	60 (48/108)
BMI (kg/m^2)			
Mean ± SD	25.4 ± 3.6	24.3 ± 3.4	25.0 ± 3.6
Range (Min/Max)	18.9 (17.4/36.3)	15.6 (17.7/33.3)	18.9 (17.4/36.3)
Number of pregnancies (including the index one) (n)			
Median	2	1	1
Range (Min/Max)	5 (1/6)	3 (1/4)	5 (1/6)
Primipara/Multipara	49/42	46/15	95/57
Risk factors for Obstetric Anal Sphincter Injury from the index delivery (n, %)			
Prolonged second phase of delivery	6 (6.6)	10 (16.4)	16 (10.5)
Fetal shoulder dystocia	1 (1.1)	0 (0.0)	1 (0.7)
Birth weight of the neonate >4kg	9 (9.9)	6 (9.8)	15 (9.9)
Induction of delivery with oxytocin	15 (16.5)	9 (14.8)	24 (15.8)
Head circumference of the neonate ≥34 cm	56 (61.53)	49 (80.3)	105 (69.1)
Time between the index delivery and ONIRY examination (days)			

Median	4	28	14
Range (Min/Max)	57 (0/57)	55 (1/56)	57 (0/57)
Time between the index delivery and EAUS examination (days)			
Median	4	28	14
Range (Min/Max)	57 (0/57)	55 (1/56)	57 (0/57)

Tables

Seed	Accuracy	Sensitivity	Specificity
1	87.2%	90.9%	84.6%
2	86.6%	92.6%	82.3%
3	87.9%	86.7%	88.6%
4	86.9%	87.7%	86.4%
5	87.6%	92.7%	84.2%
6	87.6%	89.4%	86.2%
7	86.9%	91.9%	83.5%
8	86.2%	91.0%	82.9%
9	86.9%	91.9%	83.5%
10	86.6%	91.2%	83.6%
Mean ±SD	87.0% ± 0.5%	90.6% ± 2.0%	84.6% ± 1.9%

Tables

Grade	N	Accuracy of DRE	Accuracy of impedance spectroscopy
Grade 3a	25	24.0%	81.2%
Grade 3b	26	53.8%	90.8%
Grade 3c	9	77.8%	95.6%
Grade 4	1	0.0%	60%