Leveraging Illumina iSeq100 for Whole Genome Sequencing of *Salmonella Typhi*: a
practical approach for resource-limited setting

Nishan Katuwal[1,2,3], Shiva Ram Naga[2,3], Aastha Shrestha[1,2,3], Sabin Bikram Shahi[2,3], Dipesh
Tamrakar[2,3,4], Rajeev Shrestha[1,2,3,5*]

1 Molecular and Genome Sequencing Research Lab, Dhulikhel Hospital, Kathmandu
University Hospital, Dhulikhel, Nepal

2 Center for Infectious Disease Research and Surveillance, Dhulikhel Hospital, Kathmandu
University Hospital, Dhulikhel, Nepal

3 Research and Development Division, Dhulikhel Hospital, Kathmandu University Hospital,
Dhulikhel, Nepal

4 Department of Community Medicine, Kathmandu University School of Medical Sciences,
Dhulikhel, Nepal

5 Department of Pharmacology, Kathmandu University School of Medical Sciences,
Dhulikhel, Nepal

*Corresponding Author

Prof. Dr. Rajeev Shrestha

email: rajeev.shrestha@kusms.edu.np

Word Count (Manuscript): 2565

Word Count (Abstract): 284

Number of Figures: 2

Number of Tables: 2

ABSTRACT:

Bacterial whole genome sequencing helps to improve our understanding of epidemiology and pathogenesis of bacterial infections and allows comprehensive investigation on virulence, evolution and resistance mechanisms. Nepal, in recent times, has seen some increase in sequencing capabilities but faces numerous hurdles for optimum utilization. However, these hurdles can be alleviated with use of Illumina iSeq100. Therefore, this study aimed at performing whole genome sequencing of bacteria isolated utilizing the iSeq100 platform.

For this study, 6 banked *Salmonella enterica* serovar Typhi bacterial isolates were selected. These isolates were extracted for DNA, confirmed by qPCR and then, their libraries were prepared. The libaries were checked and loaded in Illumina iSeq100 at loading concentration of 200pM. The consensus was generated from the raw genomic data by reference-based assembly, mapping onto S. Typhi CT18. These consensus genomes and coverage parameters were compared to data from using HiSeq and NextSeq. The raw reads were also evaluated using pathogenwatch (v22.3.8) to observe for genotype, mutations and resistance genes.

The coverage parameters (coverage width and depth) of the genomes from this study were compared to same genomes sequenced using Illumina HiSeq and NextSeq. The average coverage width (96.81%) and depth (63.75x) of genomes sequenced in iSeq100 were comparable to that of HiSeq (width: 98.72% and depth: 69.87x). When the genomes sequenced were compared, the genotypes detected, number of SNPs and genetic determinants of AMR genes were identical.

The data from bacterial whole genome sequencing using the Illumina iSeq100 is equally informative when compared to other high-end sequencers. Therefore, the primary goal of this study is to advocate for optimum utlisation of iSeq100 while still ensuring a high standard of quality. This optimum utilization would create capacity to fill critical surveillance gaps.

Keywords: iSeq100, Nepal, salmonella, whole genome sequencing

BACKGROUND:

As part of most infection and outbreak management protocols, several phenotypic and molecular methods for pathogen characterization are conventionally utilized to monitor and control the spread of infections. [1] However, these conventional control approaches are slow and resource intensive while failing to distinguish between closely related strain, resistance and virulence factors, mostly due to limited genomic resolution and require multiple assays. [2-5] Therefore, for obtaining comprehensive information on phylogeny and improve outbreak investigations, full genomic information are needed. [2]

Bacterial whole genome sequencing (WGS) is one of the most promising approaches of this development, which helps to improve our understanding of epidemiology and pathogenesis of bacterial infections.[6] This approach due to its comprehensiveness on pathogen biology, mutations, drug resistance, transmission, evolution, community profiling, clinical metagenomics and pathogen discovery are already transforming the research landscape in microbiology.[7] This is evident by the large number of whole genomes stored in public repositories. [8,9] These approaches have known to be used in public health surveillance and control for several bacterial infections due to *Escherichia coli, Campylobacter jejuni, Legionella pneumophila and Mycobacterium tuberculosis*, outbreaks and monitor the source of healthcare associated infections. [10-14] Furthermore, continuous development in high throughput sequencing have pushed current clinical microbiology field due to its vast potential in identification of infectious agents, detection of pathogenicity, antimicrobial resistance, evoluation and epidemiological surveillance. [10,15-17] For instance, identification of single nucleotide polymorphisms (SNPs) can differentiate the evolution, with low frequency of SNPs indicating bacteria are genetically similar and recently originated from the same source. [2] Further, significance of WGS has been more evident by its use in COVID-19 pandemic. [18,19]

Although, WGS approach can comprehend all genomic information, it is only being utlilised in niches, because clinicians and researchers have shown reluctance, due higher costs, data interpretation and burdensome process of early sequencing technologies. [1,20,21] Further, some sequencing instruments have high operational costs, require large multiplexing for effective and cost continuous sequencing. [22] Nevertheless, the encouraging prospects are recent advancements in sequencing technologies (for instance, Sequencing by Synthesis from Illumina) and investigation tools have made the platform to have high throughput, increase output, decrease analysis time and reduced cost. [23,24]

Nepal, recently, has also seen some increase in sequencing potential, with some sites with next generation sequencing capabilities in Illumina, Nanopore and Thermo Fisher platforms. [25,26] However, not all the sequencers are at full potential due to technical sophistication, supply chain issue, limitation in capacity and funding and as a result, in most studies, samples are transported outside. [27] Nonetheless, these barriers of cost and technical complexity can be alleviated with optimum utilization Illumina iSeq 100, which is an inexpensive benchtop next-generation sequencer that minimizes the up-front instrument costs while maximizing simplicity of use and capability in obtaining bacterial whole genome, in country. [28,29]

Therefore, this study aimed at performing Whole Genome Sequencing (WGS) of bacteria isolated from Dhulikhel Hospital Kathmandu University Hospital, utilizing the Illumina iSeq100 platform.

METHODS:

Bacterial Isolates Selection

For this study, 6 banked *Salmonella enterica* serovar Typhi bacterial isolates were selected randomly from Surveillance for Enteric Fever in Asia Project (SEAP). These isolates were previously sequencing in Illumina HiSeq and NextSeq platforms. These glycerol stock isolates were subcultured on MacConkey Agar (Oxoid, Cat: CM0007) and reconfirmed biochemically and serologically. An average of 30 isolated colonies were picked for each isolate.

Processing, DNA Extraction and qPCR

The selected colonies were resuspended in 800ul of sterile normal saline, vortexed and centrifuged for 10 minutes at 8000rpm, to obtain cell pellets. The supernatant was discarded and then the pellets were resuspended in 200ul sterile 1XPBS. The pellets were subsequently revotexed and 100ul of the resuspension was taken for extraction.

The extraction was done as per manufacturer manual, using Qiagen DNeasy Blood and Tissue Kit. (Qiagen, Cat: 69504), with modification in elution. The final elution was done in 50ul, which was used for re-elution to increase the yield.

The S. Typhi isolated were re-confirmed by qPCR targeting Ty21a gene, using forward primer: 5'-CGCGAAGTCAGAGTCGACATAG-3', reverse primer 5'-AAGACCTCAACGCCGATCAC-3' and probe [6-FAM] CATTTGTTCTGGAGCAGGCTGACGG [BHQ1a-Q]. The assays were run in BioRad CFX 96 Dx qPCR machine with reaction mix that included: 10ul of 2X master mix (Quantabio Perfecta, Cat: 95113-012), 0.8ul of forward primer (400nM), 0.8 ul of reverse primer (400nM), 0.4ul of probe (200nM), 4ul of nuclease free water and 4ul of DNA template. The samples will Ct values <20.00 were selected for library preparation.

Library Preparation and Whole Genome Sequencing

The concentration of the samples was checked by Qubit Fluorometer 4 (Invitrogen, Cat: Q33226) using dsDNA HS Kit (Invitrogen, Cat: Q32854). All the samples were diluted to have input amount of 75ng. The libraries were subsequently prepared using NEBNext Ultra II FS DNA Library Preparation Kit (New England Biolabs, Cat: 7805L). The libraries were quality checked for library size and concentration using Agilent Tapestation 4150 using D5000 HS Assay Kit (Agilent Technologies, Cat: 5067-5593). The average library size was 222.5bp. The isolates were loaded in two batches of two and four isolates at loading concentration of 200pM. The sequencing was done using pair ended barcoding primers at 2x146bp.

Assembly, Consensus and Bioinformatical Investigation

The raw genomic data were trimmed for adapters (fastp), assembled (bowtie2, samtools), consensus generated, and coverage was calculated (bamCoverage, samtools depth, awk) and viewed in Integrated Genome Browser. The consensus was built using reference-based assembly and mapping on reference S. Typhi CT18. (Figure 1)

Figure 1: Bioinformatics workflow for generating and evaluating consensus genome

These consensus genomes and coverage parameters were compared to genomic data of same isolates previously sequenced using Illumina HiSeq (S3 to S6) and NextSeq (S1 and S2). These sequenced genomic data were downloaded from European Nucleotide Archive (ENA).

The raw reads were evaluated using Pathogenwatch (v22.3.8) to observe for genotype detected, number of Single Nucleotide Polymorphisms (SNPs), antimicrobial resistance determinants. Pathogenwatch is a platform which facilitates rapid identification of genomic markers of antimicrobial resistance (AMR) and includes latest analytics for typing along with epidemiological contextualization with public genomic data. [21]

Data Availability

The genomic data, from Illumina iSeq100, can be obtained from National Center for Biotechnology Information's BioProject PRJNA1110744 (Accession: SRX24530188 to SRX24530193).

RESULTS:

Extraction and qPCR

The S. Typhi genomic DNA when evaluated for Ty21a gene, had average Ct value of 15.14,
while the concentrations were variable, ranging from 1.58ng/ul to 20.6ng/ul.

Whole Genome Sequencing

After the reference-based assembly, the Salmonella genomes were mapped over the S. Typhi
CT18 strain in Integrated Genome Browser, with gaps observed in from ~1,033,000bp to
1,047,000bp and ~1,908,000bp to ~1,933,000bp and ~3,053,000bp to ~3,059000bp. (Figure 2)

Figure 2: Coverage Visualization of S. Typhi genomes (S1 to S6) against S. Typhi CT18 strain
reference genome, from Integrated Genome Browser

Coverage Parameters

The coverage parameters (coverage width and depth) of the genomes from this study were
compared to same genomes sequenced using Illumina HiSeq and NextSeq. The average
coverage width (96.81%) and depth (63.75x) of genomes sequenced in iSeq100 were
comparable to that of HiSeq and NextSeq (width: 98.72% and depth: 69.87x).

Table 1: Comparison of coverage depth and width between Illumina HiSeq and Illumina
iSeq100

	Illumina iSeq100		Illumina HiSeq/NextSeq	
Sample	Average Coverage Depth	Coverage Width (%) at 20x	Average Coverage Depth	Coverage Width (%) at 20x
S1	72.23x	98.91	50.43x	98.90
S2	101.91x	98.93	92.77x	99.97
S3	52.40x	98.41	65.47x	98.92

S4	34.23x	89.04	88.49x	98.94
S5	46.67x	97.00	46.81x	97.05
S6	75.02x	98.58	75.29x	98.58

Comparison of Genotype, AMR genes and SNPs

When the genomes sequenced from Illumina iSeq100 and HiSeq/NextSeq were compared, the genotypes detected, number of SNPs and genetic determinants of AMR genes were identical.

Table 2: Comparison of genetic determinants, genotype and SNPs among genomes, evaluated by Pathogenwatch scheme (v22.3.8) for *Salmonella enterica* serovar Typhi (*S*. Typhi)

Sample	Illumina iSeq100			Illumina HiSeq		
	Genetic Determinants of AMR Genes	**Genotype**	**SNPs**	**Genetic Determinants of AMR Genes**	**Genotype**	**SNPs**
S1	gyrA_S83F	4.3.1.2	87	gyrA_S83F	4.3.1.2	87
S2	gyrA_S83F	4.3.1.2	87	gyrA_S83F	4.3.1.2	87
S3	blaTEM-1D, catA1, gyrA_D87N, sul1; sul2, dfrA7	4.3.1.1	87	blaTEM-1D, catA1, gyrA_D87N, sul1; sul2, dfrA7	4.3.1.1	87
S4	blaTEM-1D, catA1, gyrA_S83F, sul1; sul2, dfrA7	4.3.1.1	87	blaTEM-1D, catA1, gyrA_S83F, sul1; sul2, dfrA7	4.3.1.1	87
S5	blaTEM-1D, catA1, gyrA_S83F, sul1; sul2, dfrA7	4.3.1.1	87	blaTEM-1D, catA1, gyrA_S83F, sul1; sul2, dfrA7	4.3.1.1	87
S6	blaTEM-1D, catA1, gyrA_S83F, sul1; sul2, dfrA7	4.3.1.1	87	blaTEM-1D, catA1, gyrA_S83F, sul1; sul2, dfrA7	4.3.1.1	87

Cost Evaluation

The sequencing was performed in two batches (two samples in first batch and four in second batch). When the collective cost of human resource, molecular and sequencing reagents and overhead was considered, the cost of sequencing bacterial genome in Illumina iSeq100 would be ~$158 per gb when 4 samples are pooled and ~$312 per gb when 2 samples are pooled.

DISCUSSION:

Despite the potential, widespread adoption of WGS has been hindered by various challenges, including high costs, time-consuming protocols, and data interpretation complexities. [30] The iSeq100 represents a significant advancement in sequencing technology, offering a cost-effective and user-friendly solution that overcomes many barriers associated with traditional NGS platforms.[29] This platform has been known to be more useful when used with amplicon-based sequencing wet-lab approach. [31-33] Therefore, this study addresses these hurdles by using a low-cost next generation sequencer, Illumina iSeq100 for WGS of *Salmonella Typhi.*

Whole Genome Sequencing and Coverage Parameters

The coverage parameters of our approach were high (63.75x) and comparable to results from Illumina HiSeq and NextSeq (69.87x), which we agree could depend on the pooling. Though the threshold on the coverage depth depends upon the target pathogen, studies have estimated that that 40X are required for high-quality detection of virulence genes, while 21 bp reads at 4X coverage are sufficient for target gene detection and coverage width should preferably be between 50x – 100x. [34-36] In addition, in another study, for antibiotic resistance gene, at the read level, is considered present with 20x-30x coverage depth. [37,38] A study that utilized iSeq100 for WGS of bacterial isolates observed coverage depth of 17x - 149x with average of 59x, which is slightly lower than our observation. [28]

Similarly, high coverage width of at least 90% is recommended, which is well below the observed average coverage width of 96.81% of this study. [39] This high coverage width is required for optimum quality data during bioinformatics analysis, as this threshold helps achieve correct gene identification. [38]

The genome gaps were observed in the regions of ~1,033,000bp to 1,047,000bp and ~1,908,000bp to ~1,933,000bp and ~3,053,000bp to ~3,059,000bp, which correspond to STY_RS04890 gene (DUF2213 domain-containing protein), WP_001197078.1 (baseplate J/gp47 family protein), STY_RS09550 (phage minor head protein), STY_RS09735 (tyrosine-type recombinase/integrase), STY_RS15100 (DUF1460 domain-containing protein) and STY_RS15130 (site-specific integrase). [40] These gaps could be a result of genomic rearrangements, horizontal gene transfer as well as sequencing and alignment errors and presence of low complexity regions. [41]

Pooling and cost calculation

The pooling of isolates with varying genome size results in unequal sample coverage even with equimolar pooling, while some protocols do not mandatorily require equimolar pooling. [28,42] Further, pooling also depends on the complexity of the genome. For instance, though Escheria coli and Salmonella enterica have similar genome size, E. coli is more complex and has variable accessory genome. Therefore, comparatively larger number of S. enterica genomes can be pooled. [42] However, the cost per run also depends on pooling. The number of samples pooled would be inversely proportional to cartridge/flow cell requirement, while the amount of library preparation reagents remains the same. The average cost of run per gb is $60.65 for HiSeq (using HiSeq SBS Kit V4 kit - 250 cycles PE 2x125bp, which has now been discontinued), while for iSeq would be ~$130 per gb when 4 samples are pooled and ~$87 when 6 samples are pooled. [29,43] The cost, at our site, is slightly higher (~$158 per gb when 4 samples are pooled), possibly because cost of sequencing is dependent on region and

respective supply chain logistics. [44] Unfortunately, we cannot confirm the number of samples
pooled for sequencing in HiSeq and NextSeq, to make more informed comparisons. During the
discussion of low cost NGS, Oxford Nanopore provides great argument. The flongle, which
gives output upto 2Gb, costs ~$90, while iSeq100 cartridge/flowcell costs ~$630 with 1.2Gb
output. [42] However, the shorter shelf life of flow cell and lower accuracy of individuals reads
are counter intuitive.

Comparing genotypes and genomic determinants of AMR genes

In this study, both genotypes (4.3.1.1 and 4.3.1.2) belonging to lineage I and II, respectively,
of H58 strain, were detected. Both strains were first detected in a pediatric study conducted in
Kathmandu.[45] Furthermore, the point mutations (S83F and D87N) were observed in
quinolone resistance determining region of gyrA gene, which is associated with
fluoroquinolone resistance. [46-48] Similarly, mutations were observed in genes conferring
resistance to ampicillin (blaTEM), chloramphenicol (catA1) and resistance to co-trimoxazole
(sul) and resistance to trimethoprim (dfrA7). [49-52] In this study, the genotypes, genomic
determinants and number of SNPs among the S. Typhi genomes were identical, as expected, in
genomes sequenced from iSeq100 and HiSeq. This adds to the fact that the genomes sequenced
have similar quality.

Our findings have demonstrated that iSeq platform is able to generate high-quality, accurate
data for bacterial WGS, and could cost-effective addition to the genomic investigation
capability. The genomes generated from iSeq have comparable coverage parameters to those
obtained from more established commercial sequencing platforms like the Illumina HiSeq or
NextSeq, which are not available, nor viable (due to its large number of sample requirements,
reagents cost, annual maintenance cost among others) in Nepal. The findings of this study
support premise that iSeq100 is well suited to small laboratories with low sample throughput,
which is mostly the case in Nepal due to small molecular and genomic research ecosystem.

CONCLUSION

The data from bacterial whole genome sequencing using the Illumina iSeq100 is equally
informative when compared to other high-end sequencers. We have shown that Illumina iSeq,
though primarily designed for microbial metagenomics, is able to perform whole genome
sequencing of bacteria when combined with a modified wet-lab protocol. This approach will
also provide high-quality genomes for other pathogens of human health importance, with some
limitations on throughput. Thus, this study also advocates for optimum utlisation of iSeq100
while still ensuring a high standard of quality.

IMPACT OF THE STUDY

Looking ahead, the findings of this study pave the way for broader applications of the Illumina
iSeq100 platform in clinical microbiology, outbreak surveillance, and antimicrobial resistance
monitoring in Nepal. Future research should focus on further optimizing sample preparation
protocols, expanding sequencing capacity, and integrating bioinformatics tools to further
enhance the utility and accessibility of WGS technologies in pathogen investigation. This
approach will uplift the current national sequencing capabilities to full potential.

DECLARATIONS

Ethics approval and consent to participate

This study investigated banked bacterial isolates and did not contact human subjects or obtain data associated with them. The ethical approval was obtained from Institutional Review Committee at Kathmandu University School of Medical Sciences (ref: 115/24).

Availability of Data and Materials

The raw reads are available as an NCBI BioProject PRJNA1110744 (Accession: SRX24530188 to SRX24530193).

Competing interests

The authors declare no competing interests.

Funding

This study was funded and supported by Center for Infectious Disease Research and Surveillance, Dhulikhel Hospital Kathmandu University Hospital.

Acknowledgements

We gratefully thank Jason R. Andrews and the SEAP team for providing accessions numbers (European Nucleotide Archive) for the *S. Typhi* isolates. We are also thankful to all the research staffs at Center for Infectious Disease Research and Surveillance, Dhulikhel Hospital Kathmandu University Hospital.

299 **REFERENCES:**

1. Vourli, S., Kontos, F., Pournaras, S. (2021). WGS for Bacterial Identification and Susceptibility Testing in the Clinical Lab. In: Moran-Gilad, J., Yagel, Y. (eds) Application and Integration of Omics-powered Diagnostics in Clinical and Public Health Microbiology. Springer, Cham. https://doi.org/10.1007/978-3-030-62155-1_3

2. Quainoo S, Coolen JPM, van Hijum SAFT, Huynen MA, Melchers WJG, van Schaik W, Wertheim HFL. Whole-Genome Sequencing of Bacterial Pathogens: the Future of Nosocomial Outbreak Analysis. Clin Microbiol Rev. 2017 Oct;30(4):1015-1063. doi: 10.1128/CMR.00016-17. Erratum in: Clin Microbiol Rev. 2017 Nov 1;31(1):e00082-17. doi: 10.1128/CMR.00082-17.

3. Leopold SR, Goering RV, Witten A, Harmsen D, Mellmann A. 2014. Bacterial whole-genome sequencing revisited: portable, scalable, and standardized analysis for typing and detection of virulence and antibiotic resistance genes. J Clin Microbiol **52**:2365–2370. doi: 10.1128/JCM.00262-14.

4. Allen, J.D.; Ross, T.M. H3N2 influenza viruses in humans: Viral mechanisms, evolution, and evaluation. Hum. Vaccines Immunother. **2018**, 14, 1840–1847

5. Vashisht V, Vashisht A, Mondal AK, Farmaha J, Alptekin A, Singh H, Ahluwalia P, Srinivas A, Kolhe R. Genomics for Emerging Pathogen Identification and Monitoring: Prospects and Obstacles. BioMedInformatics. 2023; 3(4):1145-1177. https://doi.org/10.3390/biomedinformatics3040069

6. Garaizar J, Laorden L. Bacterial Genomics and Epidemiology. Microorganisms. 2023 May 29;11(6):1428. doi: 10.3390/microorganisms11061428. PMID: 37374930; PMCID: PMC10303121.

7. Pallen MJ. Et al. High-throughput sequencing and clinical microbiology: progress, opportunities and challenges. Current Opinion in Microbiology. 2010. 13(5). Doi: https://doi.org/10.1016/j.mib.2010.08.003

8. Supratim Mukherjee, Dimitri Stamatis, Cindy Tianqing Li, Galina Ovchinnikova, Jon Bertsch, Jagadish Chandrabose Sundaramurthi, Mahathi Kandimalla, Paul A. Nicolopoulos, Alessandro Favognano, I-Min A. Chen , Nikos C. Kyrpides and T.B.K. Reddy. Twenty-five years of Genomes OnLine Database (GOLD): data updates and new features in v.9. Nucl. Acids Res. (2022) doi: doi.org/10.1093/nar/gkac974

9. Sayers EW, Bolton EE, Brister JR, Canese K, Chan J, Comeau DC, Connor R, Funk K, Kelly C, Kim S, Madej T, Marchler-Bauer A, Lanczycki C, Lathrop S, Lu Z, Thibaud-Nissen F, Murphy T, Phan L, Skripchenko Y, Tse T, Wang J, Williams R, Trawick BW, Pruitt KD, Sherry ST. Database resources of the national center for biotechnology information. Nucleic Acids Res. 2022 Jan 7;50(D1):D20-D26. doi: 10.1093/nar/gkab1112. PMID: 34850941; PMCID: PMC8728269.

10. Gautam SS, Kc R, Leong KW, Mac Aogáin M, O'Toole RF. A step-by-step beginner's protocol for whole genome sequencing of human bacterial pathogens. J Biol Methods. 2019 Mar 15;6(1):e110. doi: 10.14440/jbm.2019.276. PMID: 31453259; PMCID: PMC6706130.

11. Grad YH, Lipsitch M, Feldgarden M, Arachchi HM, Cerqueira GC, et al. (2012) Genomic epidemiology of the Escherichia coli O104:H4 outbreaks in Europe, 2011. Proc Natl Acad Sci USA 109: 3065-3070. doi: 10.1073/pnas.1121491109.

12. Llarena A, Taboada E, Rossi M. (2017) Whole-Genome Sequencing in Epidemiology of Campylobacter jejuni Infections. J Clin Microbiol 55: 1269-1275. doi: 10.1128/JCM.00017-17.

13. Reuter S, Harrison TG, Köser CU, Ellington MJ, Smith GP, et al. (2013) A pilot study of rapid whole-genome sequencing for the investigation of a Legionella outbreak. BMJ Open 3: e002175. doi: 10.1136/bmjopen-2012-002175.

14. Gautam SS, Mac Aogáin M, Cooley LA, Haug G, Fyfe JA, et al. (2018) Molecular epidemiology of tuberculosis in Tasmania and genomic characterisation of its first known multi-drug resistant case. PLoS One 13: e0192351. doi: 10.1371/journal.pone.0192351.

15. Walker TM, Kohl TA, Omar SV, Hedge J, Del Ojo Elias C, et al. (2015) Whole-genome sequencing for prediction of Mycobacterium tuberculosis drug susceptibility and resistance: a

retrospective cohort study. Lancet Infect Dis 15: 1193-1202. doi: 10.1016/S1473-
3099(15)00062-6.
16. Takiff HE, Feo O. (2015) Clinical value of whole-genome sequencing of Mycobacterium
tuberculosis. Lancet Infect Dis 15: 1077-1090. doi: 10.1016/S1473-3099(15)00071-7.
17. Harris SR, Cartwright EJP, Török ME, Holden MTG, Brown NM, et al. (2012) Whole-genome
sequencing for analysis of an outbreak of meticillin-resistant Staphylococcus aureus: a
descriptive study. Lancet Infect Dis 13: 130-136. doi: 10.1016/S1473-3099(12)70268-2.
18. Berggreen H, Løvestad AH, Helmersen K, Jørgensen SB, Aamot HV. Lessons learned: use of
WGS in real-time investigation of suspected intrahospital SARS-CoV-2 outbreaks. J Hosp
Infect. 2023 Jan;131:81-88. doi: 10.1016/j.jhin.2022.10.003. Epub 2022 Oct 29. PMID:
36404573; PMCID: PMC9617632.
19. Francis RV, Billam H, Clarke M, Yates C, Tsoleridis T, Berry L, Mahida N, Irving WL, Moore
C, Holmes N, Ball JK, Loose M, McClure CP; COVID-19 Genomics UK (COG-UK)
consortium. The Impact of Real-Time Whole-Genome Sequencing in Controlling Healthcare-
Associated SARS-CoV-2 Outbreaks. J Infect Dis. 2022 Jan 5;225(1):10-18. doi:
10.1093/infdis/jiab483. PMID: 34555152; PMCID: PMC8522425.
20. Margulies M, Egholm M, Altman WE, Attiya S, Bader JS, Bemben LA, Berka J, Braverman
MS, Chen Y-J, Chen Z, Dewell SB, Du L, Fierro JM, Gomes XV, Godwin BC, He W, Helgesen
S, Ho CH, Irzyk GP, Jando SC, Alenquer MLI, Jarvie TP, Jirage KB, Kim J-B, Knight JR, Lanza
JR, Leamon JH, Lefkowitz SM, Lei M, Li J, Lohman KL, Lu H, Makhijani VB, McDade KE,
McKenna MP, Myers EW, Nickerson E, Nobile JR, Plant R, Puc BP, Ronan MT, Roth GT,
Sarkis GJ, Simons JF, Simpson JW, Srinivasan M, Tartaro KR, Tomasz A, Vogt KA, Volkmer
GA, et al.. 2005. Genome sequencing in microfabricated high-density picolitre reactors. Nature
437:376–380.
21. Valouev A, Ichikawa J, Tonthat T, Stuart J, Ranade S, Peckham H, Zeng K, Malek JA, Costa
G, McKernan K, Sidow A, Fire A, Johnson SM. 2008. A high-resolution, nucleosome position
map of C. elegans reveals a lack of universal sequence-dictated positioning. Genome Res
18:1051–1063. doi: 10.1101/gr.076463.108.
22. Christensen KD, Dukhovny D, Siebert U, Green RC. Assessing the Costs and Cost-
Effectiveness of Genomic Sequencing. J Pers Med. 2015 Dec 10;5(4):470-86. doi:
10.3390/jpm5040470. PMID: 26690481; PMCID: PMC4695866.
23. Harris SR, Cartwright EJ, Török ME, Holden MT, Brown NM, Ogilvy-Stuart AL, Ellington
MJ, Quail MA, Bentley SD, Parkhill J, Peacock SJ. 2013. Whole-genome sequencing for
analysis of an outbreak of meticillin-resistant Staphylococcus aureus: a descriptive study.
Lancet Infect Dis 13:130–136. doi: 10.1016/S1473-3099(12)70268-2.
24. Köser CU, Ellington MJ, Cartwright EJP, Gillespie SH, Brown NM, Farrington M, Holden
MTG, Dougan G, Bentley SD, Parkhill J, Peacock SJ. 2012. Routine use of microbial whole
genome sequencing in diagnostic and public health microbiology. PLoS Pathog 8:e1002824.
doi: 10.1371/journal.ppat.1002824.
25. WHO International. Building In-country Capacity for Pathogen Genetic Sequencing in Nepal.
Available   at :   https://www.who.int/nepal/news/detail/13-12-2021-building-in-country-
capacity-for-pathogen-genetic-sequencing-in-nepal. Accessed on: 6 September 2024.
26. Duke NUS. Nepal Country Profile. https://www.duke-nus.edu.sg/docs/apgilibraries/default-
document-library/country-profile_nepal.pdf?sfvrsn=a59caecf_4. Accessed on: 28 September
2024
27. WHO   International.   Nepal.   Available   at:
https://www.who.int/about/accountability/results/who-results-report-2020-mtr/country-
story/2021/nepal. Accessed on: 28 September 2024
28. Bruzek S, Vestal G, Lasher A, Lima A, Silbert S. Bacterial Whole Genome Sequencing on the
Illumina iSeq 100 for Clinical and Public Health Laboratories. J Mol Diagn. 2020
Dec;22(12):1419-1429. doi: 10.1016/j.jmoldx.2020.09.003. Epub 2020 Sep 19. PMID:
32961318.
29. Illumina iSeq100 instrument manual and product description. Available at:
https://www.illumina.com/systems/sequencing-platforms/iseq.html. Accessed on: 7 September
2024

30. Lucassen A, Houlston RS. The challenges of genome analysis in the health care setting. Genes (Basel). 2014 Jul 22;5(3):576-85. doi: 10.3390/genes5030576. PMID: 25055201; PMCID: PMC4198918.

31. Hourdel V, Kwasiborski A, Balière C, Matheus S, Batéjat CF, Manuguerra JC, Vanhomwegen J, Caro V. Rapid Genomic Characterization of SARS-CoV-2 by Direct Amplicon-Based Sequencing Through Comparison of MinION and Illumina iSeq100^{TM} System. Front Microbiol. 2020 Sep 25;11:571328. doi: 10.3389/fmicb.2020.571328. PMID: 33101244; PMCID: PMC7546329

32. Tamrakar D, Katuwal N, Adhikari N, Madhup SK, Dhimal M, Shrestha R. Genomic analysis of SARS-CoV-2 Circulating during Second and Third Wave of COVID-19 in Nepal. J Nepal Health Res Counc. 2023 Sep 8;21(1):57-62. doi: 10.33314/jnhrc.v21i1.4489. PMID: 37742150.

33. Shrestha R , Katuwal N , Adhikari N , Vanaerschot M , Tamrakar D , Dhimal M , Gyanwali P , Bhattarai S , Madhup SK , Devkota B . Whole Genome Sequence Analysis to Identify SARS-CoV-2 Variant in Nepal. Kathmandu Univ Med J (KUMJ). 2021 Apr-Jun;19(74):137-142. PMID: 34819443.

34. Lindsey RL, Pouseele H, Chen JC, Strockbine NA, Carleton HA. Implementation of whole genome sequencing (WGS) for identification and characterization of Shiga toxin-producing Escherichia coli (STEC) in the United States. Front Microbiol. 2016;7:1–9. doi: 10.3389/fmicb.2016.00766.

35. Lambert D, Carrillo CD, Koziol AG, Manninger P, Blais BW. GeneSippr: A rapid whole-genome approach for the identification and characterization of foodborne pathogens such as priority Shiga toxigenic Escherichia coli . PLoS One. 2015;10:1–19

36. Sayeb Maroura. Guidance document for WGS-Benchmarking. 2021. EURL CPS. Available at : https://sitesv2.anses.fr/en/system/files/NGS%20EURL%20WG%20-%20Del5%20-%20Benchmarking.pdf. Accessed on: 7 September 2024

37. Sanabria, A.M., Janice, J., Hjerde, E. et al. Shotgun-metagenomics based prediction of antibiotic resistance and virulence determinants in Staphylococcus aureus from periprosthetic tissue on blood culture bottles. Sci Rep 11, 20848 (2021). https://doi.org/10.1038/s41598-021-00383-7

38. Vieira, A.A., Piccoli, B.C., y Castro, T.R. et al. Pipeline validation for the identification of antimicrobial-resistant genes in carbapenem-resistant Klebsiella pneumoniae. Sci Rep **13**, 15189 (2023). https://doi.org/10.1038/s41598-023-42154-6

39. Doyle, R. M. et al. Discordant bioinformatic predictions of antimicrobial resistance from whole-genome sequencing data of bacterial isolates: An inter-laboratory study. Microbial. Genom. **6**(2), 1–13. https://doi.org/10.1099/mgen.0.000335 (2020)

40. Parkhill J, Dougan G, James KD, Thomson NR, Pickard D, Wain J, Churcher C, Mungall KL, Bentley SD, Holden MT, Sebaihia M, Baker S, Basham D, Brooks K, Chillingworth T, Connerton P, Cronin A, Davis P, Davies RM, Dowd L, White N, Farrar J, Feltwell T, Hamlin N, Haque A, Hien TT, Holroyd S, Jagels K, Krogh A, Larsen TS, Leather S, Moule S, O'Gaora P, Parry C, Quail M, Rutherford K, Simmonds M, Skelton J, Stevens K, Whitehead S, Barrell BG. Complete genome sequence of a multiple drug resistant Salmonella enterica serovar Typhi CT18. Nature. 2001 Oct 25;413(6858):848-52. doi: 10.1038/35101607. PMID: 11677608.

41. Nagarajan, N., Pop, M. Sequence assembly demystified. Nat Rev Genet **14**, 157–167 (2013). https://doi.org/10.1038/nrg3367 AND Kaas RS, Leekitcharoenphon P, Aarestrup FM, Lund O. Solving the problem of comparing whole bacterial genomes across different sequencing platforms. PLoS One. 2014 Aug 11;9(8):e104984. doi: 10.1371/journal.pone.0104984. PMID: 25110940; PMCID: PMC4128722.

42. Mitchell PK, Wang L, Stanhope BJ, Cronk BD, Anderson R, Mohan S, Zhou L, Sanchez S, Bartlett P, Maddox C, DeShambo V, Mani R, Hengesbach LM, Gresch S, Wright K, Mor S, Zhang S, Shen Z, Yan L, Mackey R, Franklin-Guild R, Zhang Y, Prarat M, Shiplett K, Ramachandran A, Narayanan S, Sanders J, Hunkapiller AA, Lahmers K, Carbonello AA, Aulik N, Lim A, Cooper J, Jones A, Guag J, Nemser SM, Tyson GH, Timme R, Strain E, Reimschuessel R, Ceric O, Goodman LB. Multi-laboratory evaluation of the Illumina iSeq platform for whole genome sequencing of Salmonella, Escherichia coli and Listeria. Microb

Genom. 2022 Feb;8(2):000717. doi: 10.1099/mgen.0.000717. PMID: 35113783; PMCID: PMC8942033.

43. La Trobe University Genomics. Available at : https://www.latrobe.edu.au/__data/assets/pdf_file/0005/724748/LTU-NGS-prices-AH.pdf Accessed on: 7 September 2024

44. Getchell M, Wulandari S, de Alwis R, Agoramurthy S, Khoo YK, Mak TM, Moe L, Stona AC, Pang J, Momin MHFHA, Amir A, Andalucia LR, Azzam G, Chin S, Chookajorn T, Arunkumar G, Hung DT, Ikram A, Jha R, Karlsson EA, Le Thi MQ, Mahasirimongkol S, Malavige GN, Manning JE, Munira SL, Trung NV, Nisar I, Qadri F, Qamar FN, Robinson MT, Saloma CP, Setk S, Shirin T, Tan LV, Dizon TJR, Thayan R, Thu HM, Tissera H, Xangsayarath P, Zaini Z, Lim JCW, Maurer-Stroh S, Smith GJD, Wang LF, Pronyk P; Asia Pathogen Genomics Initiative (Asia PGI) consortium. Pathogen genomic surveillance status among lower resource settings in Asia. Nat Microbiol. 2024 Sep 24. doi: 10.1038/s41564-024-01809-4. Epub ahead of print. PMID: 39317773.

45. Holt, K. E. et al. High-throughput bacterial SNP typing identifies distinct clusters of Salmonella Typhi causing typhoid in Nepalese children. BMC Infect. Dis. 2010. 10, 144.

46. Jacoby GA. Mechanisms of resistance to quinolones. Clin Infect Dis. 2005;41:7.

47. Brown JC, Shanahan PM, Jesudason MV, Thomson CJ, Amyes SG. Mutations responsible for reduced susceptibility to 4 quinolones in clinical isolates of multi resistant Salmonella typhi in India. J Antimicrob Chemother. 1996;37:891–900.

48. Wain J, Hoa NT, Chinh NT, Vinh H, Everett MJ, Diep TS, et al. Quinolone resistant Salmonella typhi from Vietnam; molecular basis of resistance and clinical response to treatment. Clin Infect Dis. 1997;25:1404–10.

49. Tamang MD, Oh JY, Seol SY, Kang HY, Lee JC, Lee YC, Cho DT, Kim J. Emergence of multidrug-resistant Salmonella enterica serovar Typhi associated with a class 1 integron carrying the dfrA7 gene cassette in Nepal. Int J Antimicrob Agents. 2007 Oct;30(4):330-5. doi: 10.1016/j.ijantimicag.2007.05.009. Epub 2007 Jul 12. PMID: 17629465.

50. Kongsoi S, Changkwanyeun R, Yokoyama K, Nakajima C, Changkaew K, Suthienkul O, Suzuki Y. Amino acid substitutions in GyrA affect quinolone susceptibility in Salmonella typhimurium. Drug Test Anal. 2016 Oct;8(10):1065-1070. doi: 10.1002/dta.1910. Epub 2015 Oct 30. PMID: 26514939.

51. Yusof NY, Norazzman NII, Zaidi NFM, Azlan MM, Ghazali B, Najib MA, Malik AHA, Halim MAHA, Sanusi MNSM, Zainal AA, Aziah I. Prevalence of Antimicrobial Resistance Genes in Salmonella Typhi: A Systematic Review and Meta-Analysis. Trop Med Infect Dis. 2022 Sep 27;7(10):271. doi: 10.3390/tropicalmed7100271. PMID: 36288012; PMCID: PMC9611315.

52. da Silva KE, Tanmoy AM, Pragasam AK, Iqbal J, Sajib MSI, Mutreja A, Veeraraghavan B, Tamrakar D, Qamar FN, Dougan G, Bogoch I, Seidman JC, Shakya J, Vaidya K, Carey ME, Shrestha R, Irfan S, Baker S, Luby SP, Cao Y, Dyson ZA, Garrett DO, John J, Kang G, Hooda Y, Saha SK, Saha S, Andrews JR. The international and intercontinental spread and expansion of antimicrobial-resistant Salmonella Typhi: a genomic epidemiology study. Lancet Microbe. 2022 Aug;3(8):e567-e577. doi: 10.1016/S2666-5247(22)00093-3. Epub 2022 Jun 21. PMID: 35750070; PMCID: PMC9329132.