Multiplex PCR detection of enteric pathogens in a community-based birth cohort in Ecuador: comparison of xTAG-GPP and TaqMan array card assays

Stuart Torres Ayala[1] *, Lesly Simbaña Vivanco[1] *, Nikolina Walas[2] *, Kelsey Jesser[3] ,
Nicolette A. Zhou[3] , Christine S. Fagnant-Sperati[3] , Hadley R. Burroughs[4] , Gwenyth O. Lee[5] ,
Joseph N.S. Eisenberg[6] , Gabriel Trueba[1] , Karen Levy[3] , Benjamin F. Arnold[4,7,8] †

Affiliations:
[1] Instituto de Microbiología, Universidad San Francisco de Quito, Quito, Ecuador
[2] Department of Environmental Health Sciences, University of California, Berkeley, CA, USA
[3] Department of Environmental and Occupational Health Sciences, University of Washington, Seattle, WA, USA
[4] Francis I. Proctor Foundation, University of California, San Francisco, CA, USA
[5] Rutgers Global Health Institute, Rutgers, The State University of New Jersey, New Brunswick, NJ, USA
[6] Department of Epidemiology, University of Michigan School of Public Health, Ann Arbor, MI, USA
[7] Department of Ophthalmology, University of California, San Francisco, CA, USA
[8] Institute for Global Health Sciences, University of California, San Francisco, CA, USA

* Equal contribution (STA, LSV, NW)

† Corresponding Author:

Benjamin F. Arnold
Associate Professor
F.I. Proctor Foundation
University of California, San Francisco
ben.arnold@ucsf.edu
+1 415 - 514 - 3227

Abstract

We compared the performance of two multiplex platforms, Luminex xTAG Gastrointestinal Pathogen Panel® and TaqMan Array Card, against a panel of 14 enteric pathogen targets in a community-based birth cohort in Ecuador. We found high levels of agreement and similar prevalence estimates across most pathogens.

Introduction

Enteric pathogens account for a substantial burden of disease among children in low- and middle-income countries (1). Multiplex qPCR assays such as the Luminex xTAG Gastrointestinal Pathogen Panel® (GPP) and the TaqMan Array Card (TAC) enable efficient detection of pathogens in stool compared with single-pathogen PCR or qPCR testing and represent a major advance in enteric pathogen diagnostics (2,3). Multiplex assays were originally developed and have been used extensively in clinical settings for early infection diagnosis (4). Epidemiological studies are increasingly using these assays to characterize the burden of enteric infections, diarrheal etiology, and intervention impacts (5,6), and as the number of these population-based studies continues to grow, there is an increasing need to better understand the comparability of different multiplex assays. GPP is a commercial assay that screens for 15 pathogens while the TAC assay is customized by individual labs and, for this project, included 30 pathogens. We compared the performance of the two multiplex assays, GPP and TAC, against a panel of 14 overlapping viral, bacterial, and protozoan enteric pathogen targets in a community-based birth cohort in Ecuador, a high transmission setting.

Methods

Study Design

ECoMiD is an ongoing longitudinal birth cohort study based in Esmeraldas Province in northern-coastal Ecuador (7). The protocol was reviewed and approved by institutional review boards at University of Washington (#STUDY00014270), Emory University (#IRB00101202), Universidad San Francisco de Quito (#2018–022M), and University of California, San Francisco (#21-33932), and all participants provided informed consent, with re-consent for each stool sample collection.

The study has enrolled 521 children from communities across a rural-urban gradient and collected

periodic stool samples from study subjects from ages one week to 24 months. To potentially

increase the efficiency of the assay comparison, we considered samples that had been analyzed

by the TAC assay, run earlier in the study, and had a positive result for at least one pathogen in

the GPP assay (n=485 samples). We then selected a random sample stratified by age (6, 12, 18

months) and location (rural accessible by river, rural accessible by road, intermediate, and urban)

to be representative of the cohort (n=156, 13 per stratum). We estimated that 156 samples would

provide 80% power to determine a difference between assays per target using McNemar's test

assuming a 5% alpha and sensitivities of 95.8% (TAC) and 89.6% (GPP), with conditional sensitivity of

98.9% for TAC given a positive by GPP (8) based on TAC and GPP parameters for rotavirus in clinical

samples (2).

Laboratory Methods

Stool samples were collected by caregivers in a small, insulated container and field staff collected

samples within 1 hour of sample production if the sample was not refrigerated, or within 3 hours if

the sample was refrigerated and stored in $-196\ ^{o}$C portable liquid nitrogen tanks. Samples were

transported monthly to the Universidad San Francisco de Quito for long-term storage at $-80\ ^{o}$C.

Nucleic acids from stool samples (180-220 mg) were extracted using Qiagen QIAamp Fast

DNA Stool Mini Kit (QIAGEN, Germantown, MD) into a proprietary elution buffer, with an added

bead beating step during sample lysis (Jesser et al. *in review*). During the extraction process MS2

and PhHV were added as an external control assessment of extraction and amplification

efficiency. ZymoBIOMICS Spike-in Controls (Zymo Research) were used as positive controls.

Extracted DNA was aliquoted and stored at $-80\ ^{o}$C.

For GPP testing samples were amplified and hybridized according to the Luminex xTAG

GPP kit protocol. xTAG® RNAse-free water was used as a negative control and three stocks of

known pathogen DNA (ZeptoMetrix NATtrol™ GI Verification Panel 2) were used as positive

controls. GPP gene targets are proprietary but median fluorescence intensity (MFI) values used

to determine positivity are available (**Supplemental Table 1**).

For TAC testing, extracted nucleic acids from stool samples (20 µL) were combined with

AgPath-ID One-Step RT-PCR master mix (50 µL) (Applied Biosystems, Waltham, MA), AgPath-

ID One-Step RT-PCR enzyme (4 µL) (Applied Biosystems), and nuclease-free water (26 µL)

(Applied Biosystems) and analyzed for pathogen gene targets using TAC (ThermoFisher

Scientific, Waltham, MA) (**Supplemental Table 2**) with the following cycling conditions: 45°C for

20 minutes, 95°C for 10 minutes, then 40 cycles of 95°C for 15 seconds and 60°C for 1 minute

on a QuantStudio 7 Flex instrument (ThermoFisher Scientific). Positive controls included PhHV,

MS2 and the pan *E. coli* gene target *uidA* as well as customized plasmids expressing all known

assay targets (ThermoFisher Scientific and Azenta Life Sciences, South Plainfield, NJ). Nuclease-

free water was used as a no template control on each card. Samples with cycle threshold (Ct)

value ≤ 35 for any of the gene targets for a pathogen were classified as positive.

Statistical Methods

We estimated pathogen target prevalence and agreement with exact, binomial 95% confidence

intervals for the 14 targets. Agreement of pathogen target-level results between the two assays

was assessed using McNemar's test and Cohen's kappa (9). Tests did not adjust for multiple

comparisons. We examined MFI and Ct values for samples with discordant TAC and GPP results,

positive by one assay and negative by the other, to determine if discordance was more likely with

lower quantity of sample DNA detected. Analyses were conducted using R (v4.4.0; R Core Team

2024).

Results

Two selected samples failed on the GPP assay, so the analysis included 154 samples. Overall, infection prevalence was similar between assays and agreement was >85% for 13 of 14 pathogen targets (**Figure 1**). There were differences in detection between TAC and GPP assays for six targets (McNemar's P<0.05), with higher prevalence by TAC for rotavirus, *Campylobacter* spp., and ST-ETEC, and higher prevalence by GPP for STEC-stx2, *Shigella* spp., and *Salmonella* spp.. Accounting for agreement due to chance, six targets differed with a kappa coefficient below 0.6 (**Figure 1**), however kappa statistics are influenced by outcome prevalence so comparison between pathogens should be made with caution given the wide range of prevalence observed (10). There was very poor agreement between assays for *Salmonella*, where the GPP assay classified 81% of samples as positive while the TAC assay classified 8% positive (**Figure 1**, **Supplemental Table 3**). The rank order of prevalence was similar between assays with the exception of rotavirus, ST-ETEC, and *Salmonella*. Across targets, discordance between assays was more likely for pathogens with MFI values just over the positivity cutoff for GPP (for GPP+, TAC–, **Supplemental Figure 1**) or Ct value just below 35 for TAC (for TAC+, GPP–, **Supplemental Figure 2**).

Discussion

Prior diagnostic comparison studies of the GPP and TAC assays have focused on tests of diarrheal samples in clinical settings, and found that the assays were broadly comparable and had good test performance as clinical diagnostics (2,6). This study aimed to evaluate the assay performance using community-based samples from young children and found the two assays were broadly comparable.

Consistent negative and positive controls on all GPP plates ruled out lab contamination

as an explanation for the poor agreement between assays for *Salmonella*. Previous studies have

noted high rates of *Salmonella* false positives by GPP (11,12) and at least one large-scale study

excluded GPP *Salmonella* results on this basis (6). The discrepancy between GPP and TAC may

result from differences in the oligonucleotide primers for the pathogen targets used for

Salmonella.

This study had limitations. First, GPP uses proprietary target sequences — although we

assume that differences between assay target sequences was an important underlying cause for

larger discrepancies, such as for *Salmonella* and rotavirus, we could only infer this through

examination of MFI and Ct values (**Supplemental Figures 1, 2**). Because we had no gold

standard measure of infection across the 14 pathogens, we focused on agreement between the

TAC and GPP assays but were unable to estimate their diagnostic characteristics, such as

sensitivity and specificity. We focused on pathogen-specific comparisons between assays and

did not assess co-infections or number of pathogens detected, which could be of interest in high

transmission settings. We intentionally over-sampled stools that were positive by TAC to at least

one target on the GPP assay to increase power for the comparison, but our sampling approach

could inflate estimates of prevalence. Finally, we did not consider diarrhea symptoms in this

analysis, but results should be representative of pediatric samples (both symptomatic and

asymptomatic) in a high transmission setting.

Despite these caveats, this study had many strengths. We tested samples collected in a

community-based cohort, with children enrolled across an urban-rural gradient at the ages when

enteric pathogen burden is highest. The assays included pathogens thought to be major causes

of diarrheal disease burden in lower resource settings (5), and we observed a broad range of

pathogen prevalence in this study. The results thus should inform similar epidemiologic field

studies.

Conclusion: This comparative analysis provides important guidance on comparing data

from TAC and GPP assays in non-clinical, pediatric samples for both within and across cohort

analyses.

157
158

159 Acknowledgements

160   This work was funded by the National Institutes of Health (R01A137679 to KL and JNSE,

161   R01AI162867 to BFA). The authors declare no competing interests. Data and replication files

162   are available through the Open Science Framework: https://osf.io/jh64t/ .

References

1. Calvopiña M, Eisenberg JNS, Atherton R, Trueba G, Andrade T, Eguiguren M, et al. Identifying Etiological Agents Causing Diarrhea in Low Income Ecuadorian Communities. Am J Trop Med Hyg. 2014 Sep 3;91(3):563–9.

2. Chhabra P, Gregoricus N, Weinberg GA, Halasa N, Chappell J, Hassan F, et al. Comparison of three multiplex gastrointestinal platforms for the detection of gastroenteritis viruses. J Clin Virol. 2017 Oct;95:66–71.

3. Chang LJ, Hsiao CJ, Chen B, Liu TY, Ding J, Hsu WT, et al. Accuracy and comparison of two rapid multiplex PCR tests for gastroenteritis pathogens: a systematic review and meta-analysis. BMJ Open Gastroenterol. 2021 Feb;8(1):e000553.

4. Khare R, Espy MJ, Cebelinski E, Boxrud D, Sloan LM, Cunningham SA, et al. Comparative Evaluation of Two Commercial Multiplex Panels for Detection of Gastrointestinal Pathogens by Use of Clinical Stool Specimens. Gilligan PH, editor. J Clin Microbiol. 2014 Oct;52(10):3667–73.

5. Platts-Mills JA, Liu J, Rogawski ET, Kabir F, Lertsethtakarn P, Siguas M, et al. Use of quantitative molecular diagnostic methods to assess the aetiology, burden, and clinical characteristics of diarrhoea in children in low-resource settings: a reanalysis of the MAL-ED cohort study. Lancet Glob Health. 2018 Dec;6(12):e1309–18.

6. Knee J, Sumner T, Adriano Z, Anderson C, Bush F, Capone D, et al. Effects of an urban sanitation intervention on childhood enteric infection and diarrhea in Maputo, Mozambique: A controlled before-and-after trial. eLife. 2021 Apr 9;10:e62278.

7. Lee GO, Eisenberg JNS, Uruchima J, Vasco G, Smith SM, Van Engen A, et al. Gut microbiome, enteric infections and child growth across a rural–urban gradient: protocol for the ECoMiD prospective cohort study. BMJ Open. 2021 Oct;11(10):e046241.

8. Connor RJ. Sample size for testing differences in proportions for the paired-sample design. Biometrics. 1987 Mar;43(1):207–11.

9. McHugh ML. Interrater reliability: the kappa statistic. Biochem Medica. 2012;22(3):276–82.

10. Sim J, Wright CC. The Kappa Statistic in Reliability Studies: Use, Interpretation, and Sample Size Requirements. Phys Ther. 2005 Mar 1;85(3):257–68.

11. Duong VT, Phat VV, Tuyen HT, Dung TTN, Trung PD, Minh PV, et al. Evaluation of Luminex xTAG Gastrointestinal Pathogen Panel Assay for Detection of Multiple Diarrheal Pathogens in Fecal Samples in Vietnam. Richter SS, editor. J Clin Microbiol. 2016 Apr;54(4):1094–100.

12. Kellner T, Parsons B, Chui L, Berenger BM, Xie J, Burnham CAD, et al. Comparative Evaluation of Enteric Bacterial Culture and a Molecular Multiplex Syndromic Panel in Children with Acute Gastroenteritis. Ledeboer NA, editor. J Clin Microbiol. 2019 Jun;57(6):e00205-19.

 # Figures and Tables

Figure 1: Infection prevalence, percent agreement and kappa statistic between the Luminex xTAG Gastrointestinal Panel (GPP) and TaqMan Array Card (TAC) assays for 14 enteric pathogens. Analysis includes 154 samples from children in Ecuador ages 6 to 18 months. A * indicates a McNemar's P<0.05 for difference between the two assays. Supplemental Table 3 includes additional details, including numerical estimates. Created with script: https://osf.io/4dteq .

213 **Supplemental Table 1: Luminex Gastrointestinal Pathogen Panel (GPP) assay targets and**
214 **corresponding median fluorescence intensity (MFI) thresholds for positivity**
215

Analyte	Threshold for positivity (MFI)
Adenovirus *40/41*	≥ 150
Camplobacter	≥ 150
C. difficile Probe-1	≥ 150
C. difficile Probe-2	≥ 150
Cryptosporidium	≥ 250
E. coli O157	≥ 150
E. histolytica	≥ 250
ETEC probe-1	≥ 200
ETEC probe-2	≥ 200
Giardia	≥ 250
Norovirus Probe-1	≥ 200
Norovirus Probe-2	≥ 350
Rotavirus A	≥ 150
Salmonella Probe-1	≥ 100,000 (POS), < 300 (NEG)
Salmonella Probe-2	≥ 200
STEC Probe-1	≥ 150
STEC Probe-2	≥ 150
Shigella	≥ 150
V. cholerae	≥ 150

216
217

218 **Supplemental Table 2: TaqMan Array Card (TAC) assay gene targets and corresponding**
219 **forward (F) and reverse (R) primers and probe (P) sequences.**
220

Organism	Gene Target	Sequence
Adenovirus (40/41)	*Fiber gene*	F, AACTTTCTCTCTTAATAGACGCC R, AGGGGGCTAGAAAACAAAA P, FAM-CTGACACGGGCACTCT-MGB
Norovirus GI	*ORF1-ORF2*	F, CGYTGGATGCGNTTYCATGA R, CTTAGACGCCATCATCATTYAC P, FAM-TGGACAGGAGATCGC-MGB
Norovirus GII	*ORF1-ORF2*	F, CARGARBCNATGTTYAGR TGGATGAG R, TCGACGCCATCTTCATTCACA P, FAM-TGGGAGGGCGATCGCAATCT-MGB
Rotavirus	*NSP3*	F, ACCATCTWCACRTRACCCTCTATGAG R, GGTCACATAACGCCCCTATAGC P, FAM-AGTTAAAAGCTAACACTGTCAAA-MGB
Camplobacter jejuni	*cadF*	F, CWGCTAAACCATARAAATAAAATTTCTCAC R, YTTTGAAGGTAATTTAGATATGGATAATCG P, VIC-CATTTTGAYGATTTTTGGCTTGA-MGB
	hipO	F, CTTGCGGTCATGATGGACATAC R, AGCACCACCCAAACCCTCTTCA P, FAM-TGCTTGCTGCAAAGTATT-MGB
Camplobacter coli	*GlyA*	F, AAACCAAAGCTTATCGTGTGC R, AGTGCAGCAATGTGTGCAAT P, FAM-TAAGCTCCAACTTCATCCG-MGB
LT-ETEC	*LT*	F, TTCCCACCGGATCACCAA R, CAACCTTGTGGTGCATGATGA P, FAM-CTTGGAGAGAAGAACCCT-MGB
ST-ETEC	*STh*	F, GCTAAACCAGYAGRGTCTTCAAAA R, CCCGGTACARGCAGGATTACAACA P, FAM-TGGTCCTGAAAGCATGAA-MGB
	STp	F, TGAATCACTTGACTCTTCAAAA R, GGCAGGATTACAACAAAGTT P, FAM-TGAACAACACATTTTACTGCT-MGB
STEC	stx1	F, ACTTCTCGACTGCAAAGACGTATG R, ACAAATTATCCCCTGWGCCACTATC P, FAM-CTCTGCAATAGGTACTCCA-MGB
	stx2	F, CCACATCGGTGTCTGTTATTAACC R, GGTCAAAACGCGCCTGATAG P, FAM-TTGCTGTGGATATACGAGG-MGB
Shigella spp.	*ipaH*	F, CCTTTTCCGCGTTCCTTGA

		R, CGGAATCCGGAGGTATTGC P, VIC-CGCCTTTCCGATACCGTCTCTGCA-MGB
Salmonella enterica	*ttr*	F, CTCACCAGGAGATTACAACATGG R, AGCTCAGACCAAAAGTGACCATC P, FAM-CACCGACGGCGAGACCGACTTT-MGB
Salmonella enterica serovar Typhi	*tviB*	F, TGTGGTAAAGGAACTCGGTAAA R, GACTTCCGATACCGGGATAATG P, VIC-TGGATGCCGAAGAGGTAAGACGAGA-MGB
	sty0201	F, CGCGAAGTCAGAGTCGACATAG R, AAGACCTCAACGCCGATCAC P, FAM-CAGCCTGCTCCAGAACA-MGB
Cryptosporidium hominus	*LIB13*	F, TCCTTGAAATGAATATTTGTGACTCG R, AAATGTGGTAGTTGCGGTTGAAA P, FAM-CTTACTTCGTGGCGGCGT-MGB
Cryptosporidium parvum	*LIB13*	F, TCCTTGAAATGAATATTTGTGACTCG R, TTAATGTGGTAGTTGCGGTTGAAC P, FAM-TATCTCTTCGTAGCGGCGTA-MGB
Cryptosporidium spp.	*18S*	F, GGGTTGTATTTATTAGATAAAGAACCA R, AGGCCAATACCCTACCGTCT P, FAM-TGACATATCATTCAAGTTTCTGAC-MGB
Entamoeba histolytica	*18S*	F, ATTGTCGTGGCATCCTAACTCA R, GCGGACGGCTCATTATAACA P, FAM-TCATTGAATGAATTGGCCATTT-MGB
Giardia spp.	*18S*	F, GACGGCTCAGGACAACGGTT R, TTGCCAGCGGTGTCCG P, FAM-CCCGCGGCGGTCCCTGCTAG-MGB

221
222
223
224

Supplemental Table 3: Summary of multiplex PCR test results for Luminex xTAG Gastrointestinal Pathogen Panel (GPP) and TaqMan Array Card (TAC). Test results from 154 samples measured among children at ages 6, 12, and 18 months old in Esmeraldas Province, Ecuador, 2022-2023. Test results are summarized by whether they were positive (+) or negative (–) by GPP and TAC. Methods include details on estimation of agreement, Cohen's kKappa, and McNemar's test for differences between assays. Created with script: https://osf.io/4dteq .

Assay Target	GPP - TAC -	GPP + TAC -	GPP - TAC +	GPP + TAC +	GPP Prevalence % (95% Confidence Interval)	TAC Prevalence % (95% Confidence Interval)	% Agreement (95% Confidence Interval)	Kappa Agreement (95% Confidence Interval)	McNemar's p-value
Adenovirus (40/41)	137	4	3	10	9.1 (5.1, 14.8)	8.4 (4.6, 14)	95.5 (90.9, 98.2)	0.72 (0.56, 0.9)	0.71
Norovirus GI	147	0	3	4	2.6 (0.7, 6.5)	4.5 (1.8, 9.1)	98.1 (94.4, 99.6)	0.72 (0.57, 0.9)	0.08
Norovirus GII	138	1	6	9	6.5 (3.2, 11.6)	9.7 (5.6, 15.6)	95.5 (90.9, 98.2)	0.70 (0.54, 0.9)	0.06
Rotavirus	142	0	11	1	0.6 (0, 3.6)	7.8 (4.1, 13.2)	92.9 (87.6, 96.4)	0.14 (0.06, 0.2)	0.00
Campylobacter spp.	107	3	17	27	19.5 (13.5, 26.6)	28.6 (21.6, 36.4)	87.0 (80.7, 91.9)	0.65 (0.50, 0.8)	0.00
LT-ETEC	87	10	9	48	37.7 (30, 45.8)	37 (29.4, 45.2)	87.7 (81.4, 92.4)	0.74 (0.58, 0.9)	0.82
ST-ETEC	NA	1	15	8	5.8 (2.7, 10.8)	14.9 (9.7, 21.6)	89.6 (83.7, 93.9)	0.45 (0.32, 0.6)	0.00
STEC stx1	134	5	1	14	12.3 (7.6, 18.6)	9.7 (5.6, 15.6)	96.1 (91.7, 98.6)	0.80 (0.65, 1.0)	0.10
STEC stx2	142	6	1	5	7.1 (3.6, 12.4)	3.9 (1.4, 8.3)	95.5 (90.9, 98.2)	0.57 (0.42, 0.7)	0.06
Shigella spp.	128	7	1	18	16.2 (10.8, 23)	12.3 (7.6, 18.6)	94.8 (90.0, 97.7)	0.79 (0.63, 0.9)	0.03
Salmonella spp.	30	112	0	12	80.5 (73.4, 86.5)	7.8 (4.1, 13.2)	27.3 (20.4, 35.0)	0.04 (-0.00, 0.1)	0.00
Cryptosporidium spp.	130	11	7	6	11 (6.6, 17.1)	8.4 (4.6, 14)	88.3 (82.2, 92.9)	0.34 (0.18, 0.5)	0.35
Entamoeba histolytica	151	2	1	0	1.3 (0.2, 4.6)	0.6 (0, 3.6)	98.1 (94.4, 99.6)	-0.01 (-0.16, 0.1)	0.56
Giardia spp.	119	4	11	20	15.6 (10.2, 22.3)	20.1 (14.1, 27.3)	90.3 (84.4, 94.4)	0.67 (0.51, 0.8)	0.07

 Supplemental Figure 1: Median fluorescence intensity (MFI) values for pathogen targets
 detected by the Luminex GPP Assay. Results are categorized according to TaqMan Array Card
 (TAC) and Luminex xTAG Gastrointestinal Pathogen Panel (GPP) assay sample results, for
 positive (+) and negative (−) detection for each target. Created with script: https://osf.io/hfv5r .

Supplemental Figure 2: Cycle threshold (Ct) values for pathogen associated gene targets detected by the TAC Assay. Multiple gene targets were used for some enteric pathogens in the TaqMan Array Card (TAC) panel. In each comparison, the top row label identifies the Luminex xTAG Gastrointestinal Pathogen Panel (GPP) target, and the second row identifies the TAC target. Results are categorized according to TAC and GPP assay sample results, for positive (+) and negative (−) detection for each GPP target. Created with script: https://osf.io/hfv5r .