Multiplex PCR detection of enteric pathogens in a community-based birth cohort in Ecuador: comparison of xTAG-GPP and TaqMan array card assays

Stuart Torres Ayala[1] *, Lesly Simbaña Vivanco[1] *, Nikolina Walas[2] *, Kelsey Jesser[3] ,
Nicolette A. Zhou[3] , Christine S. Fagnant-Sperati[3] , Hadley R. Burroughs[4], Gwenyth O. Lee[5],
Joseph N.S. Eisenberg[6], Gabriel Trueba[1], Karen Levy[3] , Benjamin F. Arnold[4,7,8] #

Affiliations:
[1] Instituto de Microbiología, Universidad San Francisco de Quito, Quito, Ecuador
[2] Department of Environmental Health Sciences, University of California, Berkeley, CA, USA
[3] Department of Environmental and Occupational Health Sciences, University of Washington, Seattle, WA, USA
[4] Francis I. Proctor Foundation, University of California, San Francisco, CA, USA
[5] Rutgers Global Health Institute, Rutgers, The State University of New Jersey, New Brunswick, NJ, USA
[6] Department of Epidemiology, University of Michigan School of Public Health, Ann Arbor, MI, USA
[7] Department of Ophthalmology, University of California, San Francisco, CA, USA
[8] Institute for Global Health Sciences, University of California, San Francisco, CA, USA

* STA, LSV, and NW contributed equally

Corresponding Author:
Benjamin F. Arnold
Associate Professor
F.I. Proctor Foundation
University of California, San Francisco
ben.arnold@ucsf.edu
+1 415 - 514 - 3227

Running head: Comparison of TaqMan and xTAG-GPP assays

Keywords: multiplex PCR, bacteria, viruses, protozoans, enteric pathogens

Abstract

We compared the performance of two multiplex platforms, Luminex xTAG Gastrointestinal Pathogen Panel® and TaqMan Array Card, against a panel of 14 enteric pathogen targets in a community-based birth cohort in Ecuador. We found high levels of agreement and similar prevalence estimates across most pathogens.

Introduction

Enteric pathogens account for a substantial burden of disease among children in low- and middle-income countries (1). Multiplex qPCR assays such as the Luminex xTAG Gastrointestinal Pathogen Panel® (GPP) and the TaqMan Array Card (TAC) enable efficient detection of pathogens in stool compared with single-pathogen PCR or qPCR testing and represent a major advance in enteric pathogen diagnostics (2,3). Multiplex assays were originally developed and have been used extensively in clinical settings for early infection diagnosis (4). Epidemiological studies are increasingly using these assays to characterize the burden of enteric infections, diarrheal etiology, and intervention impacts (5,6), and as the number of these population-based studies continues to grow, there is an increasing need to better understand the comparability of different multiplex assays. GPP is a commercial assay that screens for 15 pathogens while the TAC assay is customized by individual labs and, for this project, included 30 pathogens. We compared the performance of the two multiplex assays, GPP and TAC, against a panel of 14 overlapping viral, bacterial, and protozoan enteric pathogen targets in a community-based birth cohort in Ecuador, a high transmission setting.

Methods

Study Design

ECoMiD is an ongoing longitudinal birth cohort study based in Esmeraldas Province in northern-coastal Ecuador (7). The protocol was reviewed and approved by institutional review boards at University of Washington (#STUDY00014270), Emory University (#IRB00101202), Universidad San Francisco de Quito (#2018–022M), and University of California, San Francisco (#21-33932), and all participants provided informed consent, with re-consent for each stool sample collection. The study has enrolled 521 children from communities across a rural-urban gradient and collected periodic stool samples from study subjects from ages one week to 24 months. To potentially increase the efficiency of the assay comparison, we considered samples that had been analyzed by the TAC assay, run earlier in the study, and had a positive result for at least one pathogen in the GPP assay (n=485 samples). We then selected a random sample stratified by age (6, 12, 18 months) and location (rural accessible by river, rural accessible by road, intermediate, and urban) to be representative of the cohort (n=156, 13 per stratum). We estimated that 156 samples would provide 80% power to determine a difference between assays per target using McNemar's test assuming a 5% alpha and sensitivities of 95.8% (TAC) and 89.6% (GPP), with conditional sensitivity

of 98.9% for TAC given a positive by GPP (8) based on TAC and GPP parameters for rotavirus in

clinical samples (2).

Laboratory Methods

Stool samples were collected by caregivers in a small, insulated container and field staff collected

samples within 1 hour of sample production if the sample was not refrigerated, or within 3 hours if the

sample was refrigerated and stored in –196 °C portable liquid nitrogen tanks. Samples were

transported monthly to the Universidad San Francisco de Quito for long-term storage at –80 °C.

Nucleic acids from stool samples (180-220 mg) were extracted using Qiagen QIAamp Fast DNA

Stool Mini Kit (QIAGEN, Germantown, MD) into a proprietary elution buffer, with an added bead

beating step during sample lysis (Jesser et al. *in review*). During the extraction process MS2 and

PhHV were added as an external control assessment of extraction and amplification efficiency.

ZymoBIOMICS Spike-in Controls (Zymo Research) were used as positive controls. Extracted

DNA was aliquoted and stored at –80 °C.

For GPP testing samples were amplified and hybridized according to the Luminex xTAG GPP kit

protocol. xTAG® RNAse-free water was used as a negative control and three stocks of known

pathogen DNA (ZeptoMetrix NATtrol™ GI Verification Panel 2) were used as positive controls. GPP

gene targets are proprietary but median fluorescence intensity (MFI) values used to determine

positivity are available (**Supplemental Table 1**).

For TAC testing, extracted nucleic acids from stool samples (20 µL) were combined with AgPath-

ID One-Step RT-PCR master mix (50 µL) (Applied Biosystems, Waltham, MA), AgPath-ID One-

Step RT-PCR enzyme (4 µL) (Applied Biosystems), and nuclease-free water (26 µL) (Applied

Biosystems) and analyzed for pathogen gene targets using TAC (ThermoFisher Scientific,

Waltham, MA) (**Supplemental Table 2**) with the following cycling conditions: 45°C for 20 minutes,

95°C for 10 minutes, then 40 cycles of 95°C for 15 seconds and 60°C for 1 minute on a

QuantStudio 7 Flex instrument (ThermoFisher Scientific). Positive controls included PhHV, MS2

and the pan *E. coli* gene target *uidA* as well as customized plasmids expressing all known assay

targets (ThermoFisher Scientific and Azenta Life Sciences, South Plainfield, NJ). Nuclease-free

water was used as a no template control on each card. Samples with cycle threshold (Ct) value

≤ 35 for any of the gene targets for a pathogen were classified as positive.

 Statistical Methods

We estimated pathogen target prevalence and agreement with exact, binomial 95% confidence

intervals for the 14 targets. Agreement of pathogen target-level results between the two assays

was assessed using McNemar's test and Cohen's kappa (9). Tests did not adjust for multiple

comparisons. We examined MFI and Ct values for samples with discordant TAC and GPP results,

positive by one assay and negative by the other, to determine if discordance was more likely with

lower quantity of sample DNA detected. Analyses were conducted using R (v4.4.0; R Core Team

2024).

Results

Two selected samples failed on the GPP assay, so the analysis included 154 samples. Overall,

infection prevalence was similar between assays (**Figure 1**) and agreement was >85% for 13 of

14 pathogen targets (**Table 1**). There were differences in detection between TAC and GPP

assays for five targets (McNemar's P<0.05), with higher prevalence by TAC for rotavirus,

Campylobacter spp., and ST-ETEC, and higher prevalence by GPP for *Shigella* spp., and

Salmonella spp.. Accounting for agreement due to chance, six targets differed with a kappa

coefficient below 0.6 (**Table 1**), however kappa statistics are influenced by outcome prevalence

so comparison between pathogens should be made with caution given the wide range of

prevalence observed (10). There was very poor agreement between assays for *Salmonella*,

where the GPP assay classified 81% of samples as positive while the TAC assay classified 8%

positive (**Figure 1**, **Table 1**). The rank order of prevalence was similar between assays with the

exception of rotavirus, ST-ETEC, and *Salmonella*. Across targets, discordance between assays

was more likely for pathogens with MFI values just over the positivity cutoff for GPP (for GPP+,

TAC–, **Supplemental Figure 1**) or Ct value just below 35 for TAC (for TAC+, GPP–,

Supplemental Figure 2).

Discussion

Prior diagnostic comparison studies of the GPP and TAC assays have focused on tests of

diarrheal samples in clinical settings, and found that the assays were broadly comparable and

had good test performance as clinical diagnostics (2,6). This study aimed to evaluate the assay

performance using community-based samples from young children and found the two assays

were broadly comparable.

Consistent negative and positive controls on all GPP plates ruled out lab contamination as an

explanation for the poor agreement between assays for *Salmonella*. Previous studies have noted

high rates of *Salmonella* false positives by GPP (11,12) and at least one large-scale study

excluded GPP *Salmonella* results on this basis (6). The discrepancy between GPP and TAC may

result from differences in the oligonucleotide primers for the pathogen targets used for

Salmonella.

This study had limitations. First, GPP uses proprietary target sequences — although we assume

that differences between assay target sequences was an important underlying cause for larger

discrepancies, such as for *Salmonella* and rotavirus, we could only infer this through examination

of MFI and Ct values (**Supplemental Figures 1, 2**). Because we had no gold standard measure

of infection across the 14 pathogens, we focused on agreement between the TAC and GPP

assays but were unable to estimate their diagnostic characteristics, such as sensitivity and

specificity. We focused on pathogen-specific comparisons between assays and did not assess

co-infections or number of pathogens detected, which could be of interest in high transmission

settings. We intentionally over-sampled stools that were positive by TAC to at least one target on

the GPP assay to increase power for the comparison, but our sampling approach could inflate

estimates of prevalence. Finally, we did not consider diarrhea symptoms in this analysis, but

results should be representative of pediatric samples (both symptomatic and asymptomatic) in a

high transmission setting.

Despite these caveats, this study had many strengths. We tested samples collected in a

community-based cohort, with children enrolled across an urban-rural gradient at the ages when

enteric pathogen burden is highest. The assays included pathogens thought to be major causes

of diarrheal disease burden in lower resource settings (5), and we observed a broad range of

pathogen prevalence in this study. The results thus should inform similar epidemiologic field

studies.

Conclusion: This comparative analysis provides important guidance on comparing data from

TAC and GPP assays in non-clinical, pediatric samples for both within and across cohort

analyses.

Footnote information

Funding: This work was funded by the National Institutes of Health (R01A137679 to KL and JNSE, R01AI162867 to BFA).

Competing interests: The authors declare no competing interests.

Data availability: Data and replication files are available through the Open Science Framework: https://osf.io/jh64t/ .

Conference presentation: The results will be presented at the 2024 meeting of the American Society of Tropical Medicine and Hygiene, Oral Presentation Abstract #8373, November 17, 2024, New Orleans, LA, USA.

References

1. Calvopiña M, Eisenberg JNS, Atherton R, Trueba G, Andrade T, Eguiguren M, et al. Identifying Etiological Agents Causing Diarrhea in Low Income Ecuadorian Communities. Am J Trop Med Hyg. 2014 Sep 3;91(3):563–9.

2. Chhabra P, Gregoricus N, Weinberg GA, Halasa N, Chappell J, Hassan F, et al. Comparison of three multiplex gastrointestinal platforms for the detection of gastroenteritis viruses. J Clin Virol. 2017 Oct;95:66–71.

3. Chang LJ, Hsiao CJ, Chen B, Liu TY, Ding J, Hsu WT, et al. Accuracy and comparison of two rapid multiplex PCR tests for gastroenteritis pathogens: a systematic review and meta-analysis. BMJ Open Gastroenterol. 2021 Feb;8(1):e000553.

4. Khare R, Espy MJ, Cebelinski E, Boxrud D, Sloan LM, Cunningham SA, et al. Comparative Evaluation of Two Commercial Multiplex Panels for Detection of Gastrointestinal Pathogens by Use of Clinical Stool Specimens. Gilligan PH, editor. J Clin Microbiol. 2014 Oct;52(10):3667–73.

5. Platts-Mills JA, Liu J, Rogawski ET, Kabir F, Lertsethtakarn P, Siguas M, et al. Use of quantitative molecular diagnostic methods to assess the aetiology, burden, and clinical characteristics of diarrhoea in children in low-resource settings: a reanalysis of the MAL-ED cohort study. Lancet Glob Health. 2018 Dec;6(12):e1309–18.

6. Knee J, Sumner T, Adriano Z, Anderson C, Bush F, Capone D, et al. Effects of an urban sanitation intervention on childhood enteric infection and diarrhea in Maputo, Mozambique: A controlled before-and-after trial. eLife. 2021 Apr 9;10:e62278.

7. Lee GO, Eisenberg JNS, Uruchima J, Vasco G, Smith SM, Van Engen A, et al. Gut microbiome, enteric infections and child growth across a rural–urban gradient: protocol for the ECoMiD prospective cohort study. BMJ Open. 2021 Oct;11(10):e046241.

8. Connor RJ. Sample size for testing differences in proportions for the paired-sample design. Biometrics. 1987 Mar;43(1):207–11.

9. McHugh ML. Interrater reliability: the kappa statistic. Biochem Medica. 2012;22(3):276–82.

10. Sim J, Wright CC. The Kappa Statistic in Reliability Studies: Use, Interpretation, and Sample Size Requirements. Phys Ther. 2005 Mar 1;85(3):257–68.

11. Duong VT, Phat VV, Tuyen HT, Dung TTN, Trung PD, Minh PV, et al. Evaluation of Luminex xTAG Gastrointestinal Pathogen Panel Assay for Detection of Multiple Diarrheal Pathogens in Fecal Samples in Vietnam. Richter SS, editor. J Clin Microbiol. 2016 Apr;54(4):1094–100.

12. Kellner T, Parsons B, Chui L, Berenger BM, Xie J, Burnham CAD, et al. Comparative Evaluation of Enteric Bacterial Culture and a Molecular Multiplex Syndromic Panel in Children with Acute Gastroenteritis. Ledeboer NA, editor. J Clin Microbiol. 2019 Jun;57(6):e00205-19.

Figures and Tables

Figure 1: Infection prevalence for 14 enteric pathogens measured by Luminex xTAG Gastrointestinal Panel (GPP) and TaqMan Array Card (TAC) assays. Analysis includes 154 samples from children ages 6 to 18 months in Esmeraldas Province, Ecuador 2022-2023. An asterisk indicates McNemar's P<0.05 for difference between the two assays. Supplemental Table 3 includes numerical estimates. Created with script: https://osf.io/4dteq .

226
227

Table 1: Summary of multiplex PCR test results for Luminex xTAG Gastrointestinal Pathogen Panel (GPP) and TaqMan Array Card (TAC). Test results from 154 samples measured among children at ages 6 to 18 months in Esmeraldas Province, Ecuador, 2022-2023. Test results are summarized by whether they were positive (+) or negative (–) by GPP and TAC. Methods include details on estimation of agreement, Cohen's Kappa, and McNemar's test for differences between assays. Created with script: https://osf.io/4dteq .

Assay Target	GPP - TAC -	GPP + TAC -	GPP - TAC +	GPP + TAC +	Agreement % (95% CI)	Kappa (95% CI)	McNemar's p-value
Viruses							
Adenovirus (40/41)	137	4	3	10	95.5 (90.9, 98.2)	0.72 (0.56, 0.9)	0.71
Norovirus GI	147	0	3	4	98.1 (94.4, 99.6)	0.72 (0.57, 0.9)	0.08
Norovirus GII	138	1	6	9	95.5 (90.9, 98.2)	0.70 (0.54, 0.9)	0.06
Rotavirus	142	0	11	1	92.9 (87.6, 96.4)	0.14 (0.06, 0.2)	<0.01
Bacteria							
Campylobacter spp.	107	3	17	27	87.0 (80.7, 91.9)	0.65 (0.50, 0.8)	<0.01
LT-ETEC	87	10	9	48	87.7 (81.4, 92.4)	0.74 (0.58, 0.9)	0.82
ST-ETEC	130	1	15	8	89.6 (83.7, 93.9)	0.45 (0.32, 0.6)	<0.01
STEC stx1	134	5	1	14	96.1 (91.7, 98.6)	0.80 (0.65, 1.0)	0.10
STEC stx2	142	6	1	5	95.5 (90.9, 98.2)	0.57 (0.42, 0.7)	0.06
Shigella spp.	128	7	1	18	94.8 (90.0, 97.7)	0.79 (0.63, 0.9)	0.03
Salmonella spp.	30	112	0	12	27.3 (20.4, 35.0)	0.04 (-0.00, 0.1)	<0.01
Protozoa							
Cryptosporidium spp.	130	11	7	6	88.3 (82.2, 92.9)	0.34 (0.18, 0.5)	0.35
Entamoeba histolytica	151	2	1	0	98.1 (94.4, 99.6)	-0.01 (-0.16, 0.1)	0.56
Giardia spp.	119	4	11	20	90.3 (84.4, 94.4)	0.67 (0.51, 0.8)	0.07

Supplementary Information

Supplemental Table 1: Luminex Gastrointestinal Pathogen Panel (GPP) assay targets and corresponding median fluorescence intensity (MFI) thresholds for positivity

Analyte	Threshold for positivity (MFI)
Adenovirus *40/41*	≥ 150
Camplobacter	≥ 150
C. difficile Probe-1	≥ 150
C. difficile Probe-2	≥ 150
Cryptosporidium	≥ 250
E. coli O157	≥ 150
E. histolytica	≥ 250
ETEC probe-1	≥ 200
ETEC probe-2	≥ 200
Giardia	≥ 250
Norovirus Probe-1	≥ 200
Norovirus Probe-2	≥ 350
Rotavirus A	≥ 150
Salmonella Probe-1	≥ 100,000 (POS), < 300 (NEG)
Salmonella Probe-2	≥ 200
STEC Probe-1	≥ 150
STEC Probe-2	≥ 150
Shigella	≥ 150
V. cholerae	≥ 150

245 Supplemental Table 2: TaqMan Array Card (TAC) assay gene targets and corresponding
246 forward (F) and reverse (R) primers and probe (P) sequences.
247

Organism	Gene Target	Sequence
Adenovirus (40/41)	*Fiber gene*	F, AACTTTCTCTCTTAATAGACGCC R, AGGGGGCTAGAAAACAAAA P, FAM-CTGACACGGGCACTCT-MGB
Norovirus GI	*ORF1-ORF2*	F, CGYTGGATGCGNTTYCATGA R, CTTAGACGCCATCATCATTYAC P, FAM-TGGACAGGAGATCGC-MGB
Norovirus GII	*ORF1-ORF2*	F, CARGARBCNATGTTYAGR TGGATGAG R, TCGACGCCATCTTCATTCACA P, FAM-TGGGAGGGCGATCGCAATCT-MGB
Rotavirus	*NSP3*	F, ACCATCTWCACRTRACCCTCTATGAG R, GGTCACATAACGCCCCTATAGC P, FAM-AGTTAAAAGCTAACACTGTCAAA-MGB
Camplobacter jejuni	*cadF*	F, CWGCTAAACCATARAAATAAAATTTCTCAC R, YTTTGAAGGTAATTTAGATATGGATAATCG P, VIC-CATTTTGAYGATTTTTGGCTTGA-MGB
	hipO	F, CTTGCGGTCATGATGGACATAC R, AGCACCACCCAAACCCTCTTCA P, FAM-TGCTTGCTGCAAAGTATT-MGB
Camplobacter coli	*GlyA*	F, AAACCAAAGCTTATCGTGTGC R, AGTGCAGCAATGTGTGCAAT P, FAM-TAAGCTCCAACTTCATCCG-MGB
LT-ETEC	*LT*	F, TTCCCACCGGATCACCAA R, CAACCTTGTGGTGCATGATGA P, FAM-CTTGGAGAGAAGAACCCT-MGB
ST-ETEC	*STh*	F, GCTAAACCAGYAGRGTCTTCAAAA R, CCCGGTACARGCAGGATTACAACA P, FAM-TGGTCCTGAAAGCATGAA-MGB
	STp	F, TGAATCACTTGACTCTTCAAAA R, GGCAGGATTACAACAAAGTT P, FAM-TGAACAACACATTTTACTGCT-MGB
STEC	stx1	F, ACTTCTCGACTGCAAAGACGTATG R, ACAAATTATCCCCTGWGCCACTATC P, FAM-CTCTGCAATAGGTACTCCA-MGB
	stx2	F, CCACATCGGTGTCTGTTATTAACC R, GGTCAAAACGCGCCTGATAG P, FAM-TTGCTGTGGATATACGAGG-MGB
Shigella spp.	*ipaH*	F, CCTTTTCCGCGTTCCTTGA

		R, CGGAATCCGGAGGTATTGC P, VIC-CGCCTTTCCGATACCGTCTCTGCA-MGB
Salmonella enterica	*ttr*	F, CTCACCAGGAGATTACAACATGG R, AGCTCAGACCAAAAGTGACCATC P, FAM-CACCGACGGCGAGACCGACTTT-MGB
Salmonella enterica serovar Typhi	*tviB*	F, TGTGGTAAAGGAACTCGGTAAA R, GACTTCCGATACCGGGATAATG P, VIC-TGGATGCCGAAGAGGTAAGACGAGA-MGB
	sty0201	F, CGCGAAGTCAGAGTCGACATAG R, AAGACCTCAACGCCGATCAC P, FAM-CAGCCTGCTCCAGAACA-MGB
Cryptosporidium hominus	*LIB13*	F, TCCTTGAAATGAATATTTGTGACTCG R, AAATGTGGTAGTTGCGGTTGAAA P, FAM-CTTACTTCGTGGCGGCGT-MGB
Cryptosporidium parvum	*LIB13*	F, TCCTTGAAATGAATATTTGTGACTCG R, TTAATGTGGTAGTTGCGGTTGAAC P, FAM-TATCTCTTCGTAGCGGCGTA-MGB
Cryptosporidium spp.	*18S*	F, GGGTTGTATTTATTAGATAAAGAACCA R, AGGCCAATACCCTACCGTCT P, FAM-TGACATATCATTCAAGTTTCTGAC-MGB
Entamoeba histolytica	*18S*	F, ATTGTCGTGGCATCCTAACTCA R, GCGGACGGCTCATTATAACA P, FAM-TCATTGAATGAATTGGCCATTT-MGB
Giardia spp.	*18S*	F, GACGGCTCAGGACAACGGTT R, TTGCCAGCGGTGTCCG P, FAM-CCCGCGGCGGTCCCTGCTAG-MGB

248
249
250

Supplemental Table 3: Infection prevalence for 14 enteric pathogens measured by Luminex xTAG Gastrointestinal Panel (GPP) and TaqMan Array Card (TAC) assays. Stool samples were tested from children at ages 6, 12, and 18 months old in Esmeraldas Province, Ecuador, 2022-2023. Created with script: https://osf.io/4dteq .

GPP Target	N	GPP pos	GPP Prev (95% CI)	TAC pos	TAC Prev (95% CI)
Viruses					
Adenovirus_40_41	154	14	9.1 (5.1, 14.8)	13	8.4 (4.6, 14.0)
Norovirus_GI	154	4	2.6 (0.7, 6.5)	7	4.5 (1.8, 9.1)
Norovirus_GII	154	10	6.5 (3.2, 11.6)	15	9.7 (5.6, 15.6)
Rotavirus_A	154	1	0.6 (0.0, 3.6)	12	7.8 (4.1, 13.2)
Bacteria					
Campylobacter	154	30	19.5 (13.5, 26.6)	44	28.6 (21.6, 36.4)
ETEC_LT	154	58	37.7 (30.0, 45.8)	57	37.0 (29.4, 45.2)
ETEC_ST	154	9	5.8 (2.7, 10.8)	23	14.9 (9.7, 21.6)
STEC_stx1	154	19	12.3 (7.6, 18.6)	15	9.7 (5.6, 15.6)
STEC_stx2	154	11	7.1 (3.6, 12.4)	6	3.9 (1.4, 8.3)
Shigella	154	25	16.2 (10.8, 23.0)	19	12.3 (7.6, 18.6)
Salmonella	154	124	80.5 (73.4, 86.5)	12	7.8 (4.1, 13.2)
Protozoa					
Cryptosporidium	154	17	11.0 (6.6, 17.1)	13	8.4 (4.6, 14.0)
Entamoeba_histolytica	154	2	1.3 (0.2, 4.6)	1	0.6 (0.0, 3.6)
Giardia	154	24	15.6 (10.2, 22.3)	31	20.1 (14.1, 27.3)

258 **Supplemental Figure 1: Median fluorescence intensity (MFI) values for pathogen targets**
259 **detected by the Luminex GPP Assay**. Results are categorized according to TaqMan Array Card
260 (TAC) and Luminex xTAG Gastrointestinal Pathogen Panel (GPP) assay sample results, for
261 positive (+) and negative (−) detection for each target. Created with script: https://osf.io/hfv5r .
262

263
264
265

Supplemental Figure 2: Cycle threshold (Ct) values for pathogen associated gene targets
detected by the TAC Assay. Multiple gene targets were used for some enteric pathogens in the
TaqMan Array Card (TAC) panel. In each comparison, the top row label identifies the Luminex
xTAG Gastrointestinal Pathogen Panel (GPP) target, and the second row identifies the TAC
target. Results are categorized according to TAC and GPP assay sample results, for positive (+)
and negative (−) detection for each GPP target. Created with script: https://osf.io/hfv5r .