From subthalamic local field potentials to the selection of chronic deep brain stimulation contacts in Parkinson's disease - A systematic review

Authors: Marjolein Muller [a,b], Mark F.C. van Leeuwen [a,c], Carel F. Hoffmann [d], Niels A. van der Gaag [d,e], Rodi Zutt [a], Saskia van der Gaag [a], Alfred C. Schouten[f], M. Fiorella Contarino[a,b*]

[a]   Department of Neurology, Haga Teaching Hospital, The Hague, the Netherlands

[b]   Department of Neurology, Leiden University Medical Centre, Leiden, the Netherlands

[c]   Educational program Technical Medicine; Leiden University Medical Centre, Delft University of

Technology & Erasmus University Medical Centre Rotterdam.

[d]   Department of Neurosurgery, Haga Teaching Hospital, The Hague, the Netherlands

[e]   Department of Neurosurgery, Leiden University Medical Centre, Leiden, the

Netherlands

[f]   Department of Biomechanical Engineering, Delft University of Technology, Delft, the

Netherlands

[*]   Corresponding author

Contact information corresponding author:

E-mail fiorellacont@yahoo.it;

Postal Adress dr. M.F.Contarino, Department of Neurology, Leiden University Medical

Centre, Albinusdreef 2, 2333 ZA Leiden, the Netherlands.

Abstract

Background: Programming deep brain stimulation (DBS) of the subthalamic nucleus for optimal symptom control in Parkinson's Disease (PD) requires time and trained personnel. Novel implantable neurostimulators allow local field potentials (LFP) recording, which could be used to identify the optimal (chronic) stimulation contact. However, literature is inconclusive on which LFP features and prediction techniques are most effective.

Objective: To evaluate the performance of different LFP-based physiomarkers for predicting the optimal (chronic) stimulation contacts.

Methods: A literature search was conducted across nine databases, resulting in 418 individual papers. Two independent reviewers screened the articles based on title, abstract, and full text. The quality of included studies was assessed using a modified Joanna Briggs Institute Critical Appraisal Checklist for Case Series. Results were categorised in four classes based on the predictive performance with respect to the *a priori* chance.

Results: Twenty-five studies were included. Single-factor beta-band predictions demonstrated positive performance scores in 94% of the outcomes. Predictions based on single non-beta-frequency factors yielded positive scores in only 25% of the outcomes, with positive results mainly for high frequency oscillations. Multi-factor predictions (e.g. machine learning) achieved accuracy scores within the two highest performance classes more often than single beta-based predictions (100% versus 39%).

Conclusion: Predicting the optimal stimulation contact based on LFP recordings is feasible and can improve DBS programming efficiency in PD. Single beta-band predictions show more promising results than non-beta-frequency factors alone, but are outperformed by multi-factor predictions. Future research should further explore multi-factor prediction for optimal contact identification.

Keywords

- Deep brain stimulation

- Parkinson's disease

- Local field potentials

- Clinical contact choice

- Subthalamic nucleus

Introduction

Deep brain stimulation (DBS) of the subthalamic nucleus (STN) is an effective treatment for patients with Parkinson's disease (PD)[1, 2].

An important determinant of DBS effect, next to accurate implantation, is appropriate DBS programming, which starts by selection of the stimulation contact(s). Currently, this contact selection is often achieved by performing a monopolar review (MPR) [3]. During MPR, the thresholds for stimulation effects, including symptom reduction and side effects, are individually assessed for each contact (or ring in case of segmented levels) by clinical evaluation, while gradually increasing the stimulation amplitude. Contacts are then ranked and chosen for chronic stimulation based on their clinical efficacy and therapeutic window. This contact selection process is time-consuming (20-40 minutes per hemisphere), requires highly trained personnel, and can be exhausting and uncomfortable for patients [4, 5]. This is especially true for the currently available directional leads which contain eight individual contacts, and for other existing and future lead designs which may include even more contacts [6-8].

The results of a monopolar review can be influenced by several confounding factors such as patient fatigue, patient feedback, and the fact that some therapeutic effects of DBS have a latency period. [9]. In addition, the 'stun' effect resulting from lead implantation can temporarily decrease or even resolve the patient's motor symptoms, complicating the MPR process [9]. Subsequently, the MPR is further complemented by an iterative process of stimulation and medication adjustments on an outpatient basis, which can take up to twelve months [10]. Hence, a faster and more objective solution is required.

One potentially promising technique is using the local field potential (LFP) recorded from the target structure [1, 11]. LFPs are electrical signals generated by the summed and synchronous electrical activity of individual neurons. The recorded signals are often interpreted in the frequency domain and can be recorded in a monopolar or bipolar manner

using one or two electrodes on each implanted lead. Novel implantable neurostimulators, introduced in 2020, even allow for chronic LFP recording, simultaneously to stimulation. (figure 1).

For LFP recordings performed within the STN, one frequency band in particular, the beta-frequency band (13-35Hz), has shown a strong correlation with PD motor symptoms [12, 13]. Several studies also indicated that delivering stimulation through contacts closest to the beta source is often clinically most effective [12, 14], and that suppression of beta activity is associated with improvement of motor symptoms in PD [12]. It is therefore hypothesised that beta LFPs could assist and optimise the DBS contact selection process [1, 11]. More recently, it has been suggested that other frequency bands may also be useful in the identification of the best contact for stimulation and that a combination of information from different frequency bands or non-frequency sources, such as anatomical landmarks or evoked resonant neural activity (ERNA) might increase the efficiency [15-17].

The objective of this systematic review was to evaluate the reported performance of different LFP-based neurophysiological biomarkers (physiomarkers) for predicting the optimal (chronic) stimulation contacts.

Figure 1. Schematic representation of possible techniques for local field potential (LFP) recording. LFPs can be recorded from the implanted traditional (4 cylindrical 'ring' electrodes) or directional lead (2 cylindrical 'ring' electrodes surrounding 2 levels split into 3 seperate 'segment' electrodes). Monopolar recordings can be obtained by means of an external system connected with externalised extensions; bipolar recordings can also be obtained by means of an implantable pulse generator (IPG).

Methods

Search strategy

The literature search was conducted in accordance with the Preferred Reporting Items for Systematic Reviews and Meta-Analysis (PRISMA) 2009 guideline [18]. The search was updated until November 1st , 2023. The following databases were included in our search: Pubmed, Embase, Web of Science, Cochrane Library, Emcare, PsycINFO, Academic Search Premier and Google Scholar. In Google Scholar only the first 50 most relevant articles were considered. All search terms covered three main subjects: PD; LFPs/neuronal oscillations and DBS programming. The complete search strings are reported in the supplementary methods (Supplementary file 1).

Study selection

Two independent reviewers (MM and MvL) screened the articles. Any discrepancies were resolved through discussion. The following in- and exclusion criteria were used for the selection:

- Inclusion criteria: studies involving patients with PD; reporting of clinical choice or effect of stimulation contact; and seperately reporting LFP recordings from multiple contacts/contact pairs from leads implanted in the STN.
- Exclusion criteria: non-English articles; non-human evaluations (animal or computer simulations); literature review articles; case-reports; and non-peer reviewed manuscripts.

Quality assessment

Two reviewers (MM and MvL) independently evaluated the methodological quality of each selected study using a self-adapted 17-points version of the Joanna Briggs Institute (JBI) Critical Appraisal Checklist for Case Series (Supplementary file 2) [19], in which questions regarding blinding, patient medication/activity state, and LFP acquisition, pre-processing and analysis techniques were added to better suit the purpose of this review.

Data analysis

The reported LFP-based performance in predicting the optimal (chronic) stimulation contact as identified through clinical evaluation (*clinical stimulaton contact*), was corrected for the *a priori* chance of choosing the correct optimal stimulation contact. The latter is based on the number of contact points considered for the prediction and the amount of predicted contacts used as outcome; for example, a different *a priori* chance is present when the prediction is calculated among 4 contact points versus 8 contact points, and when the outcome refers to the single best contact point only versus the two best contact points.

The following scoring system was developed and used to assign a score to each outcome comparing the LFP-based contact prediction to the *clinical stimulation contact*: [+][+][+]: >3 times *a priori* chance; [+][+]: ≤3 times but >2 times *a priori* chance OR significant correlation, p<0.005; [+]:≤2 times *a priori* chance but higher than *a priori* chance OR significant correlation, p<0.05; [-]: ≤ *a priori* chance OR non-significant correlation, p>0.05.

Primary outcomes were categorised into three groups: predictions based on a single frequency-oscillation factor; predictions based on information from multiple frequency factors (i.e. different combinations of theta-, alpha-, beta-, gamma- and high frequency oscillations (HFO)-based features); predictions based on a combination of frequency and non-frequency factors (i.e. ERNA-, and anatomy landmarks). Furthermore, a subdivision was made for outcome measures derived from recordings in monopolar, bipolar or pseudomonopolar configurations. Here, a pseudomonopolar configuration refers to signals recorded in a bipolar configuration and subsequently postprocessed to monopolar recordings by means of custom methodologies.

As secondary outcomes, potientially clinical relevant differences in methodology were further evaluated in subcategories. These included differences in recording device; recording electrodes; recording lead; recording mode; beta feature; beta (sub)band; timing of

recording; time difference between recording and clinical contact choice; timing of clinical contact choice with respect to surgery; and contact choice method. The distribution of primary outcome scores across each class of the secondary outcome subcategories was determined as a percentage per class (e.g. subcategory: recording device; subgroup 1: internal recording device; frequency of performance classes: 0% [-], 10% [+], 80% [+][+], 10% [+][+][+] versus subgroup 2: external recording device, frequency of performance classes: 0% [-], 5% [+], 70% [+][+], 25% [+][+][+]). Differences between subgroups for all classes combined per subcategory were evaluated by means of a Pearson Chi-Squared test, where a two-sided p-value < 0.05 was considered significant.

Finally, other relevant clinical information provided by the included articles was collected and summarised.

Results

Article selection

A total of 418 unique articles were initially identified. After application of the predetermined in- and exclusion criteria 25 articles were included and reviewed (figure 2).

Figure 2. Flow diagram of the study in- and exclusion steps.

Quality assessment

The median quality score was 13 (range 9-17), supplementary table 1 (Supplementary file 2). A summary of the quality assessment for each study is provided in tables 1, 2 and 3. The included studies have a large variety of recording procedures (e.g. OFF/ON medication,

clinician blinding, etc.) and processing methods (e.g. artefact extraction/noise reduction, normalisation, etc.).

Among the 25 included studies, 5 studies concerned retrospective research [16, 20-23], 11 studies analysed data from ten or less patients [1, 5, 13, 15, 24-30], 17 articles originated from the work of six independent research groups (group #1: [11, 13, 15, 21, 31, 32]; group #2: [17, 20, 33]; group #3: [1, 16]; group #4: [5, 24]; group #5: [22, 34]; group #6: [29, 30]). Some articles included data collected from partially overlapping patient populations [11, 13, 29, 30].

Study characteristics

This systematic review encompassed a total of 763 STN placed leads, with a median of 27 STN placed leads per study (IQR: 27). A comprehensive summary of the study characteristics, methods and outcomes of the included articles is provided in supplementary table 2 (Supplementary file 2). Recordings were collected by means of an externalised device (17 studies; 624 STN) or via an implantable pulse generator (IPG) (8 studies; 139 STN); using traditional leads (16 studies: 582 STN), directional leads (8 studies: 153 STN) or both (1 study: 1 STN traditional and 27 STN directional). The choice of the *clinical stimulaton contact* was based on either MPR (13 studies; 278 STN), the chronic stimulation settings (11 studies: 466 STN) or both (1 study: 17 STN MPR and 2 STN chronic stimulation settings). The median time of the *clinical stimulaton contact* choice ranged from several minutes to 19 months after primary lead placement. In general, the LFP recordings were performed several minutes up to 6 months after initial lead placement. In three studies (19 STN) recordings were performed during IPG replacements. Recordings and *clinical stimulaton contact* choices were performed at a median of 7 years (range 4-22 years) after initial lead placement [24, 27, 28].

Primary outcomes

When considering single contact predictions based on LFP recordings in the beta-frequency band, the large majority of predictions (94%) achieved either a predictive accuracy larger

than the *a priori* chance or a significant correlation with the *clinical stimulation contact* ([+], [+][+] or [+][+][+]; table 1, figure 3). Only two outcomes did not achieve significance ([-]): the first , (r=-0.067, p=0.645) used monopolar recordings and a low-beta-peak feature [31]; the second (p=0.062) used bipolar recordings and a z-score for high-beta-power [22]; notably in both studies other considered beta-related features were significantly correlated. For 39% a predictive accuracy of two times the *a priori* chance or more ([+][+] or [+][+][+]) was achieved. In this category a high amount (6 or 8) of contact points was more often used than a low (3 or 4) amount of contact points (67% vs. 33%, respectively).

Using the best two (and not single) *clinical stimulaton contacts* as clinical outcome always resulted in positive results ([+] and [+][+] - Supplementary table 3, Supplementary file 2). The results for predictions based on single frequency factors other than beta-band factors achieved positive scores in 25% of the outcomes (table 2, figure 3). Only studies evaluating HFO-based features consistently resulted in positive scores [17, 33]. Gamma-band factor predictions achieved a positive score in a single study across 4 STN, however, all other gamma-based predictions, including two larger studies, achieved a negative score [1, 31, 35, 36]. Predictions based on theta, alpha or both feature-bands combined always resulted in negative scores.

Interestingly, combining multiple factors, including frequency and/or non-frequency factors, for instance by means of machine learning, achieved a predictive accuracy of more than two times the *a priori* score or a significant correlation with p<0.005 in all cases (table 3, figure 4). Therefore, combinations of multiple (frequency-)factors obtained more often the top higher positive scores ([+][+]/[+][+][+]) compared to single beta-based features (100% vs. 39%, respectively).

Table 1. Prediction of the single optimal stimulation contact based on beta-frequency oscillations alone compared to the clinical stimulation contact.

Article	Prediction method	Outcome	Number of contacts	Score	Number of STN	Medication state	Blinding clinician	JBI score (Y/N/U)
Monopolar recordings								
[23]	Predictive accuracy β-peak	59.0%	8	[+][+][+]	28	U	U	12/4/2
[32]	Predictive accuracy β-peak	63% (CE)	6	[+][+][+]	19	OFF	Y	12/5/1
[31]‡	Correlation high-β-peak	r=-0.452, p<0.001	4	[+][+]	50	U	U	13/3/2
[37]	Agreement β-peak	73.3%	4#	[+][+]	45	OFF	Y	13/4/1
[33]	Agreement β-peak	54.2%	4	[+][+]	24	U	Y	13/3/2
[21]‡	Predictive accuracy low-β-peak	47.0% (TW)	6	[+][+]	27	OFF	Y	17/1/0
[21]‡	Predictive accuracy low-β-peak	30.0% (TW)	8	[+][+]	27	OFF	Y	17/1/0
[20]‡	Predictive accuracy average-β-power	50.0%	4	[+]	92	U	Y	12/5/1
[17]‡	Agreement average-β-power	50.0%	4	[+]	28	U	Y	12/4/2
[25]	Correlation β-peak	r²=0.36, p=0.02 (UPDRS)	4	[+]	12	OFF	U	14/3/1
[15]‡	Predictive accuracy low-β-peak	66.7% (TW)	3	[+]	4	OFF	U	11/5/2
[15]‡	Predictive accuracy β-peak	58.3% (TW)	3	[+]	4	OFF	U	11/5/2
[15]‡	Predictive accuracy high-β-peak	41.7% (TW)	3	[+]	4	OFF	U	11/5/2
[31]‡	Correlation low-β-peak	r=-0.067, p=0.645	4	[-]	50	U	U	13/3/2
Pseudomonopolar recordings *(bipolar LFP recordings transformed to monopolar configuration using custom technique)*								
[34]‡	Predictive accuracy β-suppression	50.0%	6	[+][+]	32	OFF	Y	15/2/1
[34]‡	Predictive accuracy β-peak	37.5%	6	[+][+]	32	OFF	Y	15/2/1
[30]‡	Predictive accuracy β-peak	71.0% (CE)	3	[+][+]	14	OFF	Y	16/1/1
[30]‡	Predictive accuracy β-peak	50.0% (CE)	4	[+]	14	OFF	Y	16/1/1
[27]	Agreement β-peak	33.0%	4	[+]	9	OFF	Y	15/2/1
Bipolar recordings								

[29]‡	Predictive accuracy low-β-suppression	40.0%	8	[+][+][+]	13	OFF	Y	16/1/1
[35]	Correlation low-(/high-)β-peak	r=-0.45, p=0.0001	6	[+][+]	28	OFF	Y	15/2/1
[22]‡	Difference z-score low-β-power vs. inactive	p=0.008	4	[+]	95	OFF	U	14/3/1
[22]‡	Difference z-score β-power vs. inactive	p=0.01	4	[+]	95	OFF	U	14/3/1
[11]	Correlation average-β-power change	r=0.35, p=0.01	4	[+]	54	OFF	Y	12/5/1
[24]	Agreement β-peak	47.0%	4*	[+]	19	OFF	U	10/7/1
[5]	Difference low-β-peak vs. empirical	p > 0.05 (CE)	3	[+]	16	OFF	Y	11/6/1
[29]‡	Predictive accuracy high-β-suppression	25.0%	8	[+]	13	OFF	Y	16/1/1
[26]	Agreement β-peak	90.0%	4*	[+]	10	OFF	Y	15/2/1
[28]	Agreement β-peak	87.5%	4*	[+]	8	ON	U	9/6/1†
[1]	Agreement β-peak	100.0%	4*#	[+]	4	OFF	Y	14/3/1
[22]‡	Difference z-score high-β-power vs. inactive	p=0.062	4	[-]	95	OFF	U	14/3/1

[+][+][+]: >3 times *a priori* chance; [+][+]: ≤3 times but >2 times *a priori* chance OR significant correlation, p<0.005; [+]: ≤2 times *a priori* chance but higher than *a priori* chance OR significant correlation, p<0.05; [-]: ≤ *a priori* chance OR non-significant correlation, p>0.05; *: both contacts from bipolar recording configuration considered as prediction; #: both contacts from bipolar stimulation configuration used as reference for predictions (supplementary results, Supplementary file 1); †: 2 JBI questions were not applicable; ‡: multiple outcomes presented in the table. *Abbreviations:* *Y: Yes; N: No; U: Unknown/Unclear; STN: subthalamic nucleus; TW: therapeutic window at predicted contact as reference; CE: clinical efficiency at predicted contact as reference; UPDRS: Unified Parkinson's Disease Rating Scale at predicted contact as reference; HFO: High frequency oscillation; IR: improvement rate.*

Table 2. Prediction of the single optimal stimulation contact based on non-beta-frequency oscillations compared to the clinical stimulation contact.

Article	Prediction methods	Outcome	Number of contacts	Score	Number of STN	Medication state	Clinician blinding	JBI score (Y/N/U)
Monopolar recordings								
[33]‡	Agreement HFO-peak	70.8%	4	[+][+]	24	U	Y	13/3/2
[17]	Agreement average-HFO-power	36.0%	4	[+]	28	U	Y	12/4/2
[33]‡	Agreement θ+α-peak	25.0%	4	[-]	40	U	Y	13/3/2
[31]‡	Correlation θ-peak	r=−0.039, p=0.786 (UPDRS)	4	[-]	50	U	Y	13/3/2
[31]‡	Correlation α-peak	r=-0.131, p=0.366 (UPDRS)	4	[-]	50	U	Y	13/3/2
[31]‡	Correlation low-γ-peak	r=-0.110, p=0.446 (UPDRS)	4	[-]	50	U	Y	13/3/2
Bipolar recordings								
[1]‡	Agreement γ-peak	100.0%	4*#	[+]	4	OFF	Y	14/3/2
[22]	Difference z-score α-power vs. inactive	p=0.112	4	[-]	95	OFF	U	14/3/1
[35]	Correlation γ-peak	r=0.17, p=0.43	6	[-]	28	OFF	Y	15/2/1
[36]	Z-score γ-power > empirical	p>0.05 (IR UPDRS)	4	[-]	20	OFF	Y	9/6/3
[1]‡	Agreement θ-peak	0.0%	4*#	[-]	4	OFF	Y	14/3/2
[1]‡	Agreement α-peak	0.0%	4*#	[-]	4	OFF	Y	14/3/2

[+][+][+]: >3 times *a priori* chance; [+][+]: ≤3 times but >2 times *a priori* chance OR significant correlation, p<0.005; [+]: ≤2 times *a priori* chance but higher than *a priori* chance OR significant correlation, p<0.05; [-]: ≤ *a priori* chance OR non-significant correlation, p>0.05; *: both contacts from bipolar recording configuration considered as prediction; #: both contacts from bipolar stimulation configuration used as reference for predictions (supplementary results, Supplementary file 1); ‡: multiple outcomes presented in the table. *Abbrevations:* *Y: Yes; N: No; U: Unknown/Unclear; STN: subthalamic nucleus; HFO: High frequency oscillation; (IR) UPDRS: (improvement rate of) Unified Parkinson's Disease Rating Scale at predicted contact as reference.*

Figure 3. Overview of outcomes according to the used LFP feature for studies with a single factor approach. The figure shows the predictive value of a specific LFP frequency factor (colour-coded) in predicting the single most optimal stimulation contact for chronic stimulation. The figure demonstrates that positive outcomes were solely achieved using beta-related features with the exception of two studies using HFO and one using gamma. Studies using other LFP features reported almost exclusively no correlation. [+][+][+]: >3 times *a priori* chance; [+][+]: ≤3 times but >2 times *a priori* chance OR significant correlation, $p<0.005$; [+]: ≤2 times *a priori* chance but higher than *a priori* chance OR significant correlation, $p<0.05$; [-]: ≤ *a priori* chance OR non-significant correlation, $p>0.05$. Corresponding borders of the boxes indicate results coming from the same study; The size of the box is proportional to the amount of STN nuclei reported; *studies including ≤ 4 subthalamic nuclei; β^s: segment recordings; β^c: combination of ring-level and segment recordings; $\beta^\#$: research group number. *Abbreviations: av.: average; supp.: suppression; L-/H-β: low/high-β; HFO: high frequency osscillations.*

Table 3. Prediction of the single optimal stimulation contact based on beta-frequency oscillations in combination with other (non-)frequency factors compared to the clinical stimulation contact.

Article	Prediction methods	Outcome	Number of contacts	Score	Number of STN	Medication state	Clinician blinding	JBI score (Y/N/U)
Multiple frequency factors only								
Monopolar recordings								
[21]‡	LASSO (α,low/high-β,(fast)-γ,HFO)	37.0% (TW)	6	[+][+]	27	OFF	Y	17/1/0
[21]‡	LASSO (α,low/high-β,(fast)-γ,HFO)	30.0% (TW)	8	[+][+]	27	OFF	Y	17/1/0
Bipolar recordings								
[16]	SVM (θ- and low-β-power)	91.0%	4	[+][+][+]	28	OFF	U	11/5/2
[36]	β/α-ratio z-score vs. empirical	p < 0.001 (IR UPDRS)	4	[+][+]	20	OFF	U	9/6/3
Multiple frequency and non-frequency factors								
Monopolar recordings								
[20]‡	RF (average-ERNA/-β+anatomy)	80.0%	4	[+][+][+]	92	U	Y	12/5/1
[17]‡	Ranking (average-ERNA,-β,-HFO)	r²=0.61, p<0.001 (UPDRS)	4	[+][+]	28	U	Y	12/4/2

[+][+][+]: >3 times *a priori* chance; [+][+]: ≤3 times but >2 times *a priori* chance OR significant correlation, p<0.005; [+]: ≤2 times *a priori* chance but higher than *a priori* chance OR significant correlation, p<0.05; [-]: ≤ *a priori* chance OR non-significant correlation, p>0.05; ‡: multiple outcomes presented in the table. *Abbreviations: Y: Yes; N: No; U: Unknown/Unclear; STN: subthalamic nucleus; LASSO: Least Absolute Shrinkage and Selection Operator regression; HFO: High frequency oscillation; SVM: support vector machine; RF: Random Forest; TW: Therapeutic window at predicted contact as reference; (IR) UPDRS: (improvement rate of) Unified Parkinson's Disease Rating Scale at predicted contact as reference.*

Figure 4. Overview of outcomes according to the used LFP feature for studies with a multi-factor approach. The figure shows the predictive value of a specific prediction technique and combination of (non-)frequency factors (colour-coded) in predicting the best contact point for chronic stimulation. Studies using multifactor predictions always reported (strongly) positive correlations. [+][+][+]: >3 times *a priori* chance; [+][+]: ≤3 times but >2 times *a priori* chance OR significant correlation, p<0.005; [+]: ≤2 times *a priori* chance but higher than *a priori* chance OR significant correlation, p<0.05; [-]: ≤ *a priori* chance OR non-significant correlation, p>0.05. The size of the box is proportional to the amount of STN nuclei reported; β^s: segment recordings; β^c: combination of ring-level and segment recordings; $\beta^\#$: research group number.
Abbreviations: L-/H-β: low-/high-β; ERNA: evoked resonant neural activity; HFO: high frequency osscillations; SVM: support vector machine; RF: random forest; LASSO: least absolute shrinkage and selection operator.

Secondary outcomes

Potiential clinical relevant differences in methodology were further evaluated using subcategory comparisons for all single beta-based predictions (table 4). Significant differences between subgroups according to the Pearson Chi-squared test were found in only one subgroup, i.e. the type of electrode used for recording (χ^2: 14.825, p = 0.022). The combination of ring-levels and segments (all 8 contacts on a directional lead) results in a higher occurrence of [+][+][+] scores, followed by (6) segment-only recordings with a higher frequency of [+][+] scores and finally (4) ring-only recordings with a higher occurrence of [+] scores. However, when comparing the type of lead by itself (directional or traditional) by considering the same number of used electrodes (3 segments only or 4 levels only) no significant difference between directional leads and traditional leads per se was found.

Several studies also performed a variety of clinical comparisons between LFP-based and empirical *clinical stimulaton contact* choices, see supplementary results for an overview (Supplementary file 1). Six studies showed that LFP-based programming can achieve a clinical efficiency that is similar to empirical programming, whilst requiring less programming iterations, programming time and stimulation voltage [5, 11, 27, 29, 30, 36]. Two studies looked at the effect of general anaesthesia on the predictive value of several neurophysiological factors, however, results remained inconclusive [23, 33]. Five studies evaluated the effect of active or passive movement on the predictive value of several neurophysiological factors: although various differences across the predictive value of different neurophysiological factors were presented, no ideal moving- or resting-state feature(s) could be identified [1, 15, 21, 26, 34]. Six studies evaluated the predictive value of two non-frequency factors, namely anatomy and ERNA and found that overall the predictive value of non-frequency and frequency factors was similar [5, 17, 20-22, 33]: two of these studies showed that anatomical information could potentially improve the yield of neurophysiology-based predictions [21, 23].

Table 4. Scores for methodological subcategories in studies using one contact prediction based on beta-oscillations alone

Subcategory (total scores)	[+][+][+]	[+][+]	[+]	[-]	Pearson Chi^2 (2-sided p-value)
Recording system					
Externalised (21)	2 (9.5%)	6 (28.6%)	11 (52.4%)	2 (9.5%)	χ^2: 1.216,
Internalised (11)	1 (9.1%)	3 (27.3%)	7 (63.6%)	0 (0.0%)	p = 0.749
Recording electrodes					
Rings (18)	0 (0.0%)	3 (16.7%)	13(72.2%)‡	2 (11.1%)	
Segments (10)	1 (10.0%)	5 (50.0%)‡	4 (40.0%)	0 (0.0%)	**χ^2: 14.825,**
Combination rings/segments (4)	2 (50.0%)‡	1 (25.0%)	1 (25.0%)	0 (0.0%)	**p = 0.022**
Recording lead					
Traditional, 4 contacts (17)	0 (0.0%)	3 (17.6%)	12 (70.6%)	2 (11.8%)	χ^2: 0.806,
Directional, 3/4 contacts (6)	0 (0.0%)	1 (16.7%)	5 (83.3%)	0 (0.0%)	p = 0.668
Recording mode					
Monopolar (14)	2 (14.3%)	5 (35.7%)	6 (42.9%)	1 (7.1%)	
Pseudomonopolar (5)[1]	0 (0.0%)	3 (60.0%)‡	2 (40.0%)	0 (0.0%)	χ^2: 6.933,
Bipolar (13)	1 (7.7%)	1 (7.7%)	10(76.9%)‡	1 (7.7%)	p = 0.327
Beta feature					
Beta-peak (23)	2 (8.7%)	8 (34.8%)	12 (52.2%)	1 (4.4%)	
Average-beta (6)	0 (0.0%)	0 (0.0%)	5 (83.3%)‡	1 (16.7%)	χ^2: 6.982,
Beta-suppression (3)	1 (33.3%)	1 (33.3%)	1 (33.3%)	0 (0.0%)	p = 0.322
Beta (sub)band					
Full-beta (20)	2 (10.0%)	5 (25.0%)	13(65.0%)†	0 (0.0%)	
Low-beta (8)	1 (12.5%)	3 (37.5%)	3 (37.5%)	1 (12.5%)	χ^2: 5.600,
High-beta (4)	0 (0.0%)	1 (25.0%)	2 (50.0%)	1 (25.0%)	p = 0.469
Timing of recording					
Intra-op. (15)	2 (13.3%)	4 (26.7%)	8 (53.3%)	1 (6.7%)	
< 4 m. post-op. (14)	1 (7.1%)	5 (35.7%)	7 (50.0%)	1 (7.1%)	χ^2: 3.098,
≥4 m. post-op. (3)	0 (0.0%)	0 (0.0%)	3(100.0%)‡	0 (0.0%)	p = 0.796
Time difference between recording and contact choice					
≤48h (15)	1 (6.7%)	4 (26.7%)	10 (66.7%)	0 (0.0%)	
>48h & < 4 m. (6)	0 (0.0%)	3 (50.0%)†	2 (33.3%)	1 (16.7%)	χ^2: 5.985,
≥4 m. (11)	1 (9.1%)	2 (18.2%)	6 (54.6%)	1 (9.1%)	p = 0.425
Timing of contact choice					
≤48h post-op. (3)	0 (0.0%)	0 (0.0%)	3(100.0%)‡	0 (0.0%)	
>48h & < 4 m. post-op. (15)	1 (6.7%)	7 (46.7%)‡	6 (40.0%)	1 (6.7%)	χ^2: 6.485,
≥4 m. post-op. (14)[2]	2 (14.3%)	2 (14.3%)	9 (64.3%)	1 (7.1%)	p = 0.371
Contact choice method					
Clinical use (22)	2 (9.1%)	6 (27.3%)	12 (54.5%)	2 (9.1%)	χ^2: 0.970,
Performance score (10)[3]	1 (10.0%)	3 (30.0%)	6 (60.0%)	0 (0.0%)	p = 0.809

[+][+][+]: >3 times *a priori* chance; [+][+]: ≤3 times but >2 times *a priori* chance OR significant correlation, p<0.005; [+]: ≤2 times *a priori* chance but higher than *a priori* chance OR significant correlation, p<0.05; [-]: ≤ *a priori* chance OR non-significant correlation, p>0.05; [1]: Bipolar recordings transformed to monopolar configuration with custom technique; [2]: All patients follow-up of at least 4 months post-op, if follow-up time unknown ≥4 months presumed; [3]: Performance scores include contact choices based on either the therapeutic window, the clinical efficiency or the Unified Parkinson's Disease Rating Scale score; †: ≥15&< 25 percental points higher than all other subgroups within same subcategory AND more than one study; ‡: ≥25 percental points higher than other subgroup within same subcategory AND more than one study.
Abbreviations: intra-op.: intra-operatively; post-op.: post-operatively; m.: months; h: hours; Pearson Chi^{2n}: Pearson Chi-Squared test.

Discussion

In this systematic review we provide an overview of different methods used in the current literature to identify LFP-based physiomarkers able to predict *clinical stimulaton contact* choices for DBS in PD.

The results indicate that 1) Beta-frequency information is most often used and achieves positive results ([+], [+][+], or [+][+][+]) in 94% of the outcomes, although methods and results vary largely between studies*;* 2) Frequency factors from other frequency bands provide positive scores in only 25% of the outcomes, all of which concerning either HFO or, less often, gamma frequency features; and 3) Combining various (non-) frequency factors shows highly positive prediction scores ([+][+] or [+][+][+]) more often than single beta-band features alone (100% versus 39%, respectively).

Beta-frequency

In general, this review confirms that predictions based on beta frequency, including low, high, or total beta, provide positive prediction scores ([+], [+][+] or [+][+][+]) in the majority of the outcomes (29 out of 31). Whether prediction scores of less than 3 times the *a priori* chance (i.e. less than [+][+][+]) are acceptable in clinical practice will depend on the intended use of the feature. Currently, clinical choice is still to be considered the gold standard, but the use of LFP can be a useful support to complement clinical observations in selected cases or in order to reduce the search window.

In some recent articles the predictive capacity of the low- and high-beta sub-bands are compared. Previous research has shown that these sub-bands may have different physiological meaning and provide different predictive capacity and patterns of association to PD motor symptoms [31, 38-40]. Consequently, we evaluated whether low- and high-beta had a difference in performance when predicting the best stimulation spot, however, current results are not conclusive in that sense.

Earlier articles mainly focussed on the location of the (beta-)frequency peaks, whereas more recent work also addressed the location of the highest average beta power or the location with the largest stimulation-induced beta suppression. The rationale for using these additional beta-related features is multifold: an average is less susceptible to noise and/or artefacts than single peaks, and might therefore produce more robust results, and the use of stimulation-induced beta suppression may serve as a more direct confirmation of clinical efficacy [5, 29]. From our data analysis however, we could not confirm the superiority of these beta features. It should be noticed that to date, only few articles have evaluated these non-peak beta features, probably making this comparison less reliable.

Other frequencies

In the existing literature factors related to the theta-, alpha-, gamma- or HFO-band provided poorer results than beta-band factors (negative scores for 9/12 outcomes), when considered as single predictors. Several reasons could explain this finding. For example, these frequencies may be related to specific symptoms which are usually not major determinants in the initial choice of the *clinical stimulaton contact*, such as dyskinesia, tremor, or othe motor- or non-symptoms. Nonetheless, these frequencies might still be relevant in a chronic setting or in specific subsets of patients, and therefore should not be dismissed based on the limited data available.

Combination of factors

All five studies reporting on the use of combined factors achieved high positive scores. Two of these studies also individually compared results to predictions based on beta-frequency factors alone and showed that combinations of various (non-)frequency factors provided superior predictions [16, 20]. The finding that a combination of frequency factors can outperform beta-based predictions alone is interesting and should be further explored. Due to the heterogeneity of clinical phenotypes and disease stages of PD patients undergoing DBS, it could be hypothesised that not all aspects of PD can be captured by beta-frequency alone and that taking different factors into account might produce better results in individual

patients [41]. Considering the wide variation in included factors and (machine learning) models used so far, and the potentially even wider variety of methods not yet explored in the available literature, it remains unclear which method and factors could provide the best results.

Methodological aspects

The validity of LFP-based predictions of the *clinical stimulaton contact* is possibly influenced by numerous methodological factors such as study size, recording modality, recording instruments, pre-processing method, timing of recording, used reference method and the applied outcome measure. We attempted at identifying factors that might influence the predictive value of LFPs, and found that only the type of recording contact (ring-level vs. segment vs. combination of ring-levels and segments) used for recording seemed to influence the results. Stimulation level predictions were found to be more accurate for ring-levels in segmented electrodes than for cylindrical ring electrodes alone. This could potentially be very relevant, as directional leads require longer clinical evaluations during MPR, and thus a semi-automated *clinical stimulaton contact* selection would be most welcome [5, 15, 21, 23, 32, 34]. However, we acknowledge that this difference may be inherent to our analysis method which takes the *a priori* chance into account and thus assigns a higer value to predictions based on a larger number of contacts. Another possible explanation may lie in the fact that some of the directional leads used are those of the last generation, expressely designed with features that facilitate optimal recordings with respect to the tradional leads. Nonetheless, in this research we did not find a significant difference in accuracy between predictions performed with directional or traditional leads.

Other methodological aspects relevant for future research

No other methodological aspect was significantly related to the prediction efficacy. It is important to notice that subanalysis was only possible for studies reporting results on beta-related factors because these were present in larger numbers. Most of these studies yield positive results and used a large variability of methods, which could make it difficult to

identify potential small effects of methodological factors. On the other end we cannot exclude that some of these methodological factors might be more relevant when investigating other LFP features.

One important aspect could be the use of monopolar recordings with respect to bipolar recordings, because the latter require a translation to a single stimulation contact which still remains a complex matter [30].

Growing evidence suggests that the features of LFP recordings can change over time, especially in the early post-operative period, probably due to the stun-effect [42] or other surgery-related factors (such as patient fatigue, possible sedation residues). While most of the earlier studies were bound to recordings during surgery or in the immediate postoperative period, due to de inavailability of implantable systems recording LFPs, a large number of the reviewed studies (27 out of 32 outcomes) analysed LFP recordings and clinical contact choices performed later on, which is probably a more robust design to consider for future studies.

Although recordings from externalised systems bear the possibility of more powerful analysis techniques (e.g. higher sampling frequency), in this review they did not result in higher accuracies.

Strengths and Limitations

In this review we chose to focus on the clinical contact choice based on LFPs, an aspect which is highly relevant for the current clinical practice. Since the introduction on the market of the new implantable DBS systems capable of LFP recordings, an increasing number of clinics are presently using or considering the integration of LFP recordings into their clinical stimulation adjustment protocols with the aim of reducing time and resources needed for this complex process. Research about this aspect is currently being intensified, also in the view of the imminent introduction of adaptive stimulation protocols in clinical practices worldwide.

Consequently, the findings of this review lay the groundwork for future research and, by extension, future clinical protocols in DBS programming.

Here, we systematically reviewed the use of various frequency (and non-frequency) features, models and feature combinations from a vast amount of studies from multiple databases. Furthermore, various methodological issues which are crucial for the design of future studies, such as the modality, pre-processing and timing of LFP recordings, were reviewed. Considering the variability of hardware and methods used in the literature, we adopted a novel more objective classification of outcomes in scoring classes, which takes into account the *a priori* chance of prediction.

Nonetheless, this review also has some limitations. Due to the variability of the available studies, each scoring class includes a range of different outcome scores, outcome measures, reference methods and study population sizes.

Another important aspect that may have biased the results of this review was the variation in LFP pre-processing and analysis methodologies within the included articles. Not all of the reviewed articles implemented normalisation, noise reduction, and/or artefact extraction. Although these factors were considered in the quality assessment, it remained challenging to evaluate if and how they influenced the results. A large part of the included studies were conducted across a small number of centres and for six research groups more than one article was included in this review, which may have led to bias**.** Nonetheless, as this is a new technique and these few centres have built up expertise regarding the subject, this may have also increased the reliability of the results. For the secondary outcomes, an important drawback is that several articles were included multiple times when presenting multiple single beta-based outcomes.

Finally, most of the included articles considered the chronic stimulation contact as a reference measure. However, it is important to notice that the chronic stimulation contact choice can be influenced by side-effects, which are not reflected in LFP recordings.

Furthermore, the chronic stimulation contact choice may also be biased by subjective improvement reported by the patient, the clinicians' preference, the lead position or the amount of time between contact choice and surgery.

Conclusions and future perspectives

Research focussing on the use of LFPs for determining the optimal *clinical stimulaton contact* increased exponentially since 2010 (figure 3). There is a wide variability of methods used concerning the pre-processing steps; methods for converting bipolar recordings to monopolar predictions; timing for LFP recordings and clinical contact choice; methods for clinical contact choice; applied prediction techniques/(machine learning) models; and applied outcome measures: all these factors should be accurately considered within future research.

While earlier articles focussed on the location of the (beta-)frequency peaks alone, more recent work also investigates other frequency factors and even dynamic factors such as stimulation-induced beta suppression. Single beta-frequency band factor predictions show more promising results than predictions by means of single factors from other frequency bands. Future research should focus on systematically comparing the predictive capacity of different beta (sub-)band factors.

Furthermore, in the most recent literature (spread throughout 2020-2022) there seems to be a trend toward multi-factor predictions and the use of machine learning methods (figure 4). Evidence suggests that beta-band based predictions may be improved by integrating factors from other frequencies or non-frequency factors. Future research should focus on identification of multi-factor combinations and models with the highest predictive capacity, and consider clustering results according to different clinical features of the included patients in order to obtain individualised predictions.

Results from this review could serve as the basis for future research and clinical protocols in DBS programming, adding objective LFP-based support to the clinical contact selection process.

Acknowledgements

We wish to acknowledge the contribution of J.W. (Jan) Schoones, librarian LUMC, for assisting with the search.

Funding

This work has been funded by the European Union's Horizon Europe research and innovation programme under grant agreement number 101070865 (MINIGRAPH).

CRediT Statement

- M. Muller: Conceptualization; Data curation; Formal analysis; Investigation; Methodology; Visualization; Writing – original draft

- M.F.C. van Leeuwen: Formal analysis; Investigation; Writing – original draft

- C.F. Hoffmann: Writing – review & editing

- N.A. van der Gaag: Writing – review & editing

- R. Zutt: Writing – review & editing

- S. van der Gaag: Writing – review & editing

- A.C. Schouten: Conceptualization; Supervision; Writing – review & editing

- M.F. Contarino: Conceptualization; Methodology; Visualization; Supervision; Writing – review & editing

References

[1]   Ince NF, Gupte A, Wichmann T, Ashe J, Henry T, Bebler M, et al. Selection of optimal programming contacts based on local field potential recordings from subthalamic nucleus in patients with Parkinson's disease. Neurosurgery. 2010;67(2):390-7.

[2]   Pintér D, Járdaházi E, Balás I, Harmat M, Makó T, Juhász A, et al. Antiparkinsonian Drug Reduction After Directional Versus Omnidirectional Bilateral Subthalamic Deep Brain Stimulation. Neuromodulation. 2023;26(2):374-81.

[3]   Volkmann J, Moro E, Pahwa R. Basic algorithms for the programming of deep brain stimulation in Parkinson's disease. Mov Disord. 2006;21 Suppl 14:S284-9.

[4]   Thompson JA, Lanctin D, Ince NF, Abosch A. Clinical implications of local field potentials for understanding and treating movement disorders. Stereotact Funct Neurosurg. 2014;92(4):251-63.

[5]   Binder T, Lange F, Pozzi N, Musacchio T, Daniels C, Odorfer T, et al. Feasibility of local field potential-guided programming for deep brain stimulation in Parkinson's disease: A comparison with clinical and neuro-imaging guided approaches in a randomized, controlled pilot trial. Brain Stimulation. 2023;16(5):1243-51.

[6]   Pollo C, Kaelin-Lang A, Oertel MF, Stieglitz L, Taub E, Fuhr P, et al. Directional deep brain stimulation: an intraoperative double-blind pilot study. Brain. 2014;137(Pt 7):2015-26.

[7]   Steigerwald F, Müller L, Johannes S, Matthies C, Volkmann J. Directional deep brain stimulation of the subthalamic nucleus: A pilot study using a novel neurostimulation device. Mov Disord. 2016;31(8):1240-3.

[8]   Contarino MF, Bour LJ, Verhagen R, Lourens MA, de Bie RM, van den Munckhof P, et al. Directional steering: A novel approach to deep brain stimulation. Neurology. 2014;83(13):1163-9.

[9]   Mestre TA, Lang AE, Okun MS. Factors influencing the outcome of deep brain stimulation: Placebo, nocebo, lessebo, and lesion effects. Movement Disorders. 2016;31(3):290-8.

[10]   Kühn AA, Volkmann J. Innovations in deep brain stimulation methodology. Movement Disorders. 2017;32(1):11-9.

[11]   Yoshida F, Martinez-Torres I, Pogosyan A, Holl E, Petersen E, Chen CC, et al. Value of subthalamic nucleus local field potentials recordings in predicting stimulation parameters for deep brain stimulation in Parkinson's disease. J Neurol Neurosurg Psychiatry. 2010;81(8):885-9.

[12]   Zaidel A, Spivak A, Grieb B, Bergman H, Israel Z. Subthalamic span of beta oscillations predicts deep brain stimulation efficacy for patients with Parkinson's disease. Brain. 2010;133(Pt 7):2007-21.

[13]   Chen CC, Pogosyan A, Zrinzo LU, Tisch S, Limousin P, Ashkan K, et al. Intra-operative recordings of local field potentials can help localize the subthalamic nucleus in Parkinson's disease surgery. Exp Neurol. 2006;198(1):214-21.

[14]   Holdefer RN, Cohen BA, Greene KA. Intraoperative local field recording for deep brain stimulation in Parkinson's disease and essential tremor. Mov Disord. 2010;25(13):2067-75.

[15]   Nguyen TAK, Schüpbach M, Mercanzini A, Dransart A, Pollo C. Directional Local Field Potentials in the Subthalamic Nucleus During Deep Brain Implantation of Parkinson's Disease Patients. Front Hum Neurosci. 2020;14:521282.

[16]   Connolly AT, Kaemmerer WF, Dani S, Stanslaski SR, Panken E, Johnson MD, et al., editors. Guiding deep brain stimulation contact selection using local field potentials sensed by a chronically implanted device in Parkinson's disease patients. 2015 7th International IEEE/EMBS Conference on Neural Engineering (NER); 2015: IEEE.

[17]   Xu SS, Sinclair NC, Bulluss KJ, Perera T, Lee WL, McDermott HJ, et al. Towards guided and automated programming of subthalamic area stimulation in Parkinson's disease. Brain Commun. 2022;4(1):fcac003.

[18]   Moher D, Liberati A, Tetzlaff J, Altman DG. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. PLoS Med. 2009;6(7):e1000097.

[19]   Munn Z, Barker TH, Moola S, Tufanaru C, Stern C, McArthur A, et al. Methodological quality of case series studies: an introduction to the JBI critical appraisal tool. JBI Evid Synth. 2020;18(10):2127-33.

[20]   Xu SS, Lee WL, Perera T, Sinclair NC, Bulluss KJ, McDermott HJ, et al. Can brain signals and anatomy refine contact choice for deep brain stimulation in Parkinson's disease? J Neurol Neurosurg Psychiatry. 2022.

[21]   Shah A, Nguyen TK, Peterman K, Khawaldeh S, Debove I, Shah SA, et al. Combining Multimodal Biomarkers to Guide Deep Brain Stimulation Programming in Parkinson Disease. Neuromodulation. 2023;26(2):320-32.

[22]   Horn A, Neumann WJ, Degen K, Schneider GH, Kühn AA. Toward an electrophysiological "sweet spot" for deep brain stimulation in the subthalamic nucleus. Hum Brain Mapp. 2017;38(7):3377-90.

[23]   di Biase L, Piano C, Bove F, Ricci L, Caminiti ML, Stefani A, et al. Intraoperative Local Field Potential Beta Power and Three-Dimensional Neuroimaging Mapping Predict Long-Term Clinical Response to Deep Brain Stimulation in Parkinson Disease: A Retrospective Study. Neuromodulation. 2023.

[24]   Thenaisie Y, Palmisano C, Canessa A, Keulen BJ, Capetian P, Castro Jimenez M, et al. Towards adaptive deep brain stimulation: clinical and technical notes on a novel commercial device for chronic brain sensing. Journal of neural engineering. 2021;13.

[25]   Tamir I, Wang D, Chen W, Ostrem JL, Starr PA, de Hemptinne C. Eight cylindrical contact lead recordings in the subthalamic region localize beta oscillations source to the dorsal STN. Neurobiol Dis. 2020;146:105090.

[26]   Kochanski RB, Shils J, Verhagen Metman L, Pal G, Sani S. Analysis of Movement-Related Beta Oscillations in the Off-Medication State During Subthalamic Nucleus Deep Brain Stimulation Surgery. J Clin Neurophysiol. 2019;36(1):67-73.

[27]   Lewis S, Radcliffe E, Ojemann S, Kramer DR, Hirt L, Case M, et al. Pilot Study to Investigate the Use of In-Clinic Sensing to Identify Optimal Stimulation Parameters for Deep Brain Stimulation Therapy in Parkinson's Disease. Neuromodulation. 2023.

[28]   Swinnen BEKS, Stam MJ, Buijink AWG, de Neeling MGJ, Schuurman PR, de Bie RMA, et al. Employing LFP recording to optimize stimulation location and amplitude in chronic DBS for Parkinson's disease: A proof-of-concept pilot study. Deep Brain Stimulation. 2023;2:1-5.

[29]   Strelow JN, Dembek TA, Baldermann JC, Andrade P, Fink GR, Visser-Vandewalle V, et al. Low beta-band suppression as a tool for DBS contact selection for akinetic-rigid symptoms in Parkinson's disease. Parkinsonism Relat Disord. 2023;112:105478.

[30]   Strelow JN, Dembek TA, Baldermann JC, Andrade P, Jergas H, Visser-Vandewalle V, et al. Local Field Potential-Guided Contact Selection Using Chronically Implanted Sensing Devices for Deep Brain Stimulation in Parkinson's Disease. Brain Sci. 2022;12(12).

[31]   Chen PL, Chen YC, Tu PH, Liu TC, Chen MC, Wu HT, et al. Subthalamic high-beta oscillation informs the outcome of deep brain stimulation in patients with Parkinson's disease. Front Hum Neurosci. 2022;16:958521.

[32]   Tinkhauser G, Pogosyan A, Debove I, Nowacki A, Shah SA, Seidel K, et al. Directional local field potentials: A tool to optimize deep brain stimulation. Mov Disord. 2018;33(1):159-64.

[33]   Sinclair NC, McDermott HJ, Lee WL, Xu SS, Acevedo N, Begg A, et al. Electrically evoked and spontaneous neural activity in the subthalamic nucleus under general anesthesia. J Neurosurg. 2021:1-10.

[34]   Busch JL, Kaplan J, Bahners BH, Roediger J, Faust K, Schneider GH, et al. Local Field Potentials Predict Motor Performance in Deep Brain Stimulation for Parkinson's Disease. Movement Disorders. 2023.

[35]   Fernandez-Garcia C, Monje MHG, Gomez-Mayordomo V, Foffani G, Herranz R, Catalan MJ, et al. Long-term directional deep brain stimulation: Monopolar review vs. local field potential guided programming. Brain Stimulation. 2022;15(3):727-36.

[36]   Dong W, Qiu C, Chang L, Sun J, Yan J, Luo B, et al. The guiding effect of local field potential during deep brain stimulation surgery for programming in Parkinson's disease patients. CNS Neuroscience and Therapeutics. 2023.

[37]   Milosevic L, Scherer M, Cebi I, Guggenberger R, Machetanz K, Naros G, et al. Online Mapping With the Deep Brain Stimulation Lead: A Novel Targeting Tool in Parkinson's Disease. Mov Disord. 2020;35(9):1574-86.

[38]   Morelli N, Summers RLS. Association of subthalamic beta frequency sub-bands to symptom severity in patients with Parkinson's disease: A systematic review. Parkinsonism & Related Disorders. 2023;110:105364.

[39]   Priori A, Foffani G, Pesenti A, Tamma F, Bianchi AM, Pellegrini M, et al. Rhythm-specific pharmacological modulation of subthalamic activity in Parkinson's disease. Exp Neurol. 2004;189(2):369-79.

[40]   Neumann WJ, Degen K, Schneider GH, Brücke C, Huebl J, Brown P, et al. Subthalamic synchronized oscillatory activity correlates with motor impairment in patients with Parkinson's disease. Mov Disord. 2016;31(11):1748-51.

[41]   Pozzi NG, Isaias IU. Adaptive deep brain stimulation: Retuning Parkinson's disease. Handb Clin Neurol. 2022;184:273-84.

[42]   Okun MS, Gallo BV, Mandybur G, Jagid J, Foote KD, Revilla FJ, et al. Subthalamic deep brain stimulation with a constant-current device in Parkinson's disease: an open-label randomised controlled trial. Lancet Neurol. 2012;11(2):140-9.