Title: Diagnostic-avoiding *Chlamydia trachomatis* variants detected in cervical and endometrial

specimens from women during 16S microbiome profiling.

Authors: Sangmi Jeong[1], Tammy Tollison[1], Hayden Brochu[1], Hsuan Chou[1], Tammy Yu[2],

Priyanka Baghaie[2], Kacy S. Yount[2], Toni Darville[2], Harold C Wiesenfeld[3], Sharon L Hillier[3], Xinxia

Peng[1] and Catherine M. O'Connell[2]

[1]North Carolina State University

[2]University of North Carolina at Chapel Hill

[3]University of Pittsburgh School of Medicine and the Magee-Womens Research Institute

Correspondence: Catherine O'Connell, PhD, University of North Carolina at Chapel Hill, 8340B

MBRB, CB# 7509, 111 Mason Farm Road, Chapel Hill, NC 27599-7509, phone (919) 962-3422,

fax:(919) 966-7277, catherine.oconnell@unc.edu

Alternate corresponding author: Xinxia Peng, PhD, North Carolina State University, Raleigh, NC

27695, phone (919) 515-4481, fax: (919)-513-6464, xpeng5@ncsu.edu

Keywords: NAAT, chlamydia, gonorrhea, microbiome, 16S rRNA, false negative, 23S

Abstract:

Background: Performance of a 16S rRNA analysis of the cervicovaginal microbiome of 220

participants recruited into the T Cell Response against Chlamydia (TRAC) cohort between

February 2011 and August 2014 in Allegheny County, Pennsylvania USA detected DNA encoding

chlamydial 16S rRNA in samples from seven participants whose tests were negative for

Chlamydia trachomatis (CT) and DNA encoding gonococcal 16S rRNA from five participants

whose tests were negative for *Neisseria gonorrhoeae* (NG) infection with the Aptima Combo2

assay (Hologic).

Methods: We used targeted PCR amplification followed by sequencing to characterize the

chlamydial 23S rRNA locus and qPCR to detect gonococcal DNA in residual diagnostic swab

eluates or DNA used to generate 16S rRNA libraries.

Results: Discrepant specimens that contained chlamydial DNA carried a diagnostic-avoidant,

G1526A variant in the 23S rRNA locus identical to variants previously detected in Finland,

Denmark, and the UK. PCR validation of gonococcal DNA was confirmed for all participants

whose tests were negative, with stochastic effects consistent with infection levels close to the limit

of detection by the diagnostic assay.

Conclusions: These data indicate that this probe-avoidant CT mutant, and possibly others, were

circulating in the northeastern US prior to their detection and characterization in 2019. Although

infrequent, documentation of false negative results for CT indicates a need for clinicians to

consider performance of a second test that uses alternate PCR targets if patients have persistent

symptoms or have known contact to an infected sex partner and their initial NAAT is negative.

Background:

Performing cross sectional or longitudinal research studies investigating natural immunity

or immunopathogenesis of sexually transmitted infections (STI) in humans can be a complex,

time consuming and expensive undertaking. Efforts to mine all data associated with infection or

disease cohorts are often extensive and protracted. Such is the case for the TRAC (T cell

Response Against Chlamydia) cohort, a group of 246 cisgender women at elevated risk for

chlamydial infection who enrolled between February 2011 and August 2014 while presenting to

the Allegheny County, PA Health Department's Sexually Transmitted Diseases Clinic for STI

testing and/or management. The TRAC study sought to profile cell mediated responses

associated with protection against ascending *C. trachomatis* (CT) infection (from the cervix to

endometrium) or against CT reinfection [1, 2]. Since completion of enrollment, data obtained from

participant samples and questionnaire responses have been used to profile antibody [1, 3-5], T

cell antigen-specific responses [2], and to determine risk factors for ascending infection. Using

modern molecular techniques, we quantified pathogen burden [1, 2], local cytokine secretion [6]

and used genetic profiling techniques to search for biomarkers predictive of disease or reinfection

risk [7-9]. Secondary investigations have also assessed the impact of *Mycoplasma genitalium*

infection in this population [10]. Fundamental to the success of these analyses has been accurate

classification of study participants according to infection status at enrollment and during

subsequent follow up, so enrollee infection status was assigned according to the outcome of an

FDA-cleared STI diagnostic NAAT test performed at the study visit, the results of which were used

to inform clinical care. Infrequently, discrepant results were observed between research PCR

assays detecting chlamydial loci and results obtained using the Aptima Combo 2® assay (AC2),

manufactured by Hologic, Inc., a commercially available nucleic acid amplification test (NAAT)

for CT and *Neisseria gonorrhoeae* (NG), targeting 23S ribosomal RNA (rRNA) and 16S rRNA in

these bacteria, respectively. However, we could not exclude the possibility of cross-contamination

from the environment or chlamydia-infected specimens, particularly with repurposed, reprocessed

convenience samples, e.g., leftover transport medium.

The availability of pristine "future use" vaginal swabs collected from the same cohort and

stored since enrollment provided an opportunity to reexamine these discrepant results while

determining cervicovaginal microbiome profiles that associated with ascending chlamydial

infection [11]. We performed a 16S rRNA gene genomic DNA based microbiome analysis on the

entire cohort and detected chlamydial DNA in samples from seven participants whose tests were

negative for CT and gonococcal DNA from five participants whose tests were negative for NG

with the AC2 assay. PCR amplification and subsequent sequencing using primers directed

against the 23S rRNA locus revealed that some of the specimens contained chlamydial DNA

carrying a diagnostic-avoidant single nucleotide polymorphism (SNP) that likely contributed to

misclassification. These data indicate that this mutant, and possibly others, were circulating in the

northeastern US prior to their detection and characterization in the United Kingdom and Denmark

in 2019 and subsequently [12-14]. Discrepant NG detection reflected very low gonococcal loads

that challenged the limit of resolution for the NAAT. Although infrequent, documentation of false

negative results for CT or NG may indicate a need for clinicians to consider performance of a

second test that uses alternate or additional PCR probes.

Methods:

T Cell Response against Chlamydia (TRAC) Cohort (246 participants). This cohort was composed

of young (median age, 21 years; range, 18-35 years) cis-gender women at high risk for the

acquisition of CT [1]. Criteria indicating high-risk status included > 3 sexual partners in the

previous 6 months, ≤ 14 years of age at sexual debut, history of pelvic inflammatory disease

(PID), the presence of mucopurulent cervicitis on exam, or sexual contact with a partner known

to be infected with CT or NG or non-gonococcal non-chlamydial urethritis. Women with a current

diagnosis of PID were excluded. Additional exclusion criteria were pregnancy, uterine procedure

or miscarriage in the preceding 60 days, menopause, hysterectomy, antibiotic therapy in the

preceding 14 days, and allergy to study medications. TRAC participants were enrolled into a

longitudinal study designed to investigate T cell responses important for protection from incident

chlamydial infection over 12 months of follow-up between February 2011 and August 2014 and

were recruited from the Allegheny County Health Department's Sexually Transmitted Diseases

Clinic, UPMC Magee-Womens Hospital (MWH) Ambulatory Care Clinic, and the Reproductive

Infectious Disease Research Unit at MWH in Pittsburgh PA. Participants provided informed

consent at the time of enrollment and completed questionnaires regarding obstetric/gynecologic

history, behavioral practices, sex exposure, contraceptive methods, and symptoms. Clinical,

histological, and microbiological testing was performed, and blood and endometrial samples were

obtained after enrollment, after which all participants received single-dose agents for gonorrhea

(ceftriaxone, 250 mg intramuscularly) and chlamydia (azithromycin, 1 g orally). Participants in this

cohort were assessed for cervical and endometrial infection using AC2 (Hologic, Marlborough,

MA) NAAT with overall infection rates of 67% and 8.5% for CT and NG, respectively. *M. genitalium*

(MG) infection was determined using the Aptima MG NAAT (Hologic).

16S library construction and characterization. Cervicovaginal swab samples available from 246

women enrolled in TRAC were processed for 16S rRNA gene microbiome profiling [11]. In brief,

cervicovaginal microbial communities present in participant specimens were profiled by

sequencing the V4 region of 16S rRNA gene using Illumina sequencing. The 16S rRNA amplicon

sequencing data was processed at the level of inferred amplicon sequence variants (ASVs) using

USEARCH v11 [15]. The raw sequencing data supporting the conclusions of this article is publicly

available at NCBI's Short Read Archive (SRA) under BioProject accession number

PRJNA1136868.

PCR template purification from residual diagnostic samples. DNA was extracted as previously

published using a Quick-DNA Universal Kit (Zymo Research, Irvine, California, USA) from 100 µL

to 300 µL of reserved cervical swab eluates. Chlamydial and/or mycoplasmal loads were

measured by quantitative PCR [1] targeting the 16SrRNA gene [16] or the highly conserved

MG190 locus [10], respectively. Primers directed against the *parA* pseudogene [17], 16S rRNA

and 23S rRNA loci [18] were similarly used to validate and/or quantify NG abundance.

PCR amplification and sequencing of 23S rRNA alleles. Amplification conditions for detection of

SNPs in *C. trachomatis* 23S were as described [19]. The melting temperature was raised to 98°C

for optimal activation of the high fidelity, proof-reading enzyme Phusion (ThermoFisher). Samples

were processed independently in a dedicated PCR hood to minimize risk of cross-contamination

during analysis. Two rounds of amplification were required to recover sufficient product which was

cleaned and concentrated using a DNA Clean and Concentrate kit (Zymo) before sequencing in

forward and reverse directions using V3 region primers [19]. Two patient samples that had tested

positively using the AC2 test were also analyzed as controls. Sequences generated via Sanger

sequencing were aligned against the published sequence using Clone Manager Professional

Version 11 (Sci Ed Software LLC, Westminster, CO). Residual cervical templates were used from

profiled participants except for 3-209, for whom only endometrium-extracted DNA remained

available for analysis.

Statistical analysis. The nonparametric Spearman rank correlation coefficient was calculated to

assess the overall correlation between CT 16S rRNA gene DNA copy numbers measured by

qPCR and the raw read counts of CT 16S rRNA gene from microbiome sequencing analysis using

R function 'cor' [20]. The LOESS (Locally Weighted Scatter-plot Smoother) curve was estimated

and plotted using a 'geom_smooth' function of the R 'ggplot2' package with default settings [21].

Abundance of NG-specific ASVs present in libraries from individuals testing positively with AC2

and those who tested negative were compared using a Mann-Whitney test (two tailed).

Results:

Detection of chlamydia-derived amplicon sequence variant (ASV) in 16S libraries derived from

vaginal swabs obtained from CT negative TRAC participants. Total DNA was extracted from

"future use" vaginal swabs from a total of 246 TRAC participants for inclusion in a 16S rRNA gene-

based sequencing analysis of the contribution of the cervicovaginal microbiome to ascending

chlamydial infection. Three samples were excluded based on DNA low integrity values after

extraction. We also excluded twenty-three samples due to the equivocal results of chlamydial

infection at the cervix and/or endometrium from clinical diagnostic testing. The remaining 220

samples were processed using the ASV-level analysis, resulting in 1,990 ASVs detected in total

from samples sequenced to the average depth of ~151,000 reads. Taxonomy annotations of

ASVs were based on sequence comparison against the 16S rRNA gene reference database RDP

version 19 [22] using the SINTAX algorithm [23]. We assessed the concordance of STI detection

by examining raw read counts of sexually transmitted pathogen-specific 16S rRNA genes vs

infections previously assessed by NAAT diagnostic tests (CT, NG, MG) (**Fig. 1a**) and quantified

via qPCR (CT [1], **Fig. 1b**). About 58% (83 of 143) samples from participants whose tests were

positive for CT at enrollment had non-zero 16S rRNA gene read counts (**Fig. 1a**). Similarly, 85%

(17 of 20) samples from participants whose tests were postive for NG, and 15.8% (6 of 38)

samples from MG positive enrollees had non-zero 16S rRNA gene read counts. Overall, there

was modest correlation (Spearman = 0.575) between the two CT quantification methods (**Fig.**

1b).

Based on clinical diagnostic results obtained at enrollment, the TRAC participants were

classified into three groups: 1) Uninfected; 2) Endo- participants with CT infection detected only

in their cervix; 3) Endo+ participants who tested positively for CT in cervix and endometrium.

There were no TRAC enrollees whose tests were positive only in the endometrial sample.

Comparing ASV abundances between CT-positive (Endo- and Endo+) and CT-negative (control)

women, we were surprised to detect 16S ASV within libraries derived from samples from women

who had been assigned "uninfected" (N=7) based on the results of NAAT testing (**Fig. 1a**). These

results correlated with estimated chlamydial abundance derived from a previously performed

quantitative 16S PCR assay (**Fig. 1b** and **Table 1**), suggesting that attributing the original results

to external or cross-contamination was incorrect. A review of these participants' clinical data

revealed that they were mostly asymptomatic (5/7), one had reported experiencing dysuria,

consistent with >10 leucocytes detected per high power field in her urine sample with irregular

vaginal bleeding and another reported dyspareunia and abnormal vaginal discharge. Clinical

findings upon pelvic examination were unremarkable for all. Three of the seven participants within

this group reported contact with a CT infected partner in the three months prior to their enrollment

but had yet to be treated.

Table 1. TRAC participants with discrepant clinical and molecular CT detection

Sample ID	Clinical testing (CT, MG, NG)	16S rRNA read counts in ASV analysis	16S qPCR*	
			Cx	Endo
3-101	CT-/MG-/NG-	1	78	919
3-144		16	420	87
3-145		146	128800	35
3-148		111	3839	457
3-201		55	1495	721
3-226		31	1640	105
3-209	CT-/MG+/NG-	4	13000	62680

*Chlamydial 16S rDNA copies/swab [1].

In the context of the TRAC cohort, reasons for false negative CT NAAT test results, other

than testing failures detected at or during performance e.g., compromised reagents or automated

run failure (batch specific), include inadequate sampling at the anatomical site or an unanticipated

flaw in the diagnostic test. All NAAT tests were run within temporal parameters appropriate for

clinical STI testing, so batch-specific effects did not serve as an explanation for these

discrepancies because the participants were not recruited contemporaneously. Testing of an

endometrial biopsy might result in a false negative because the entire endometrial lining is not

sampled, the amounts of tissue obtained may vary [7], and the region of specimen dedicated to

NAAT testing is small [10]. However, cervical testing, particularly when performed by highly

trained practitioners, is robust [24, 25]. Several publications from Northern Europe have reported

instances of false negative testing specifically associated with the AC2 test [12, 13, 26, 27] arising

from probe-avoidant mutations at the chlamydial 23S rRNA locus [19]. Speculating that this might

explain our discrepant results, we attempted to amplify the relevant V3 region from cervically or

endometrially derived templates that had been extracted from residual diagnostic swab eluates.

In all instances we were successful using the protocol and primers used by Hokynar et al. [19],

further confirming the presence of chlamydial DNA in the clinical samples. We detected a G-A

substitution at 1523 in the samples of all seven test-discrepant participants (**Fig. 2**) when these

products were sequenced. NAAT positively-testing samples (N=2) from the cohort were wild type

at this position (data not shown). These data indicated that at least one probe-avoidant variant of

CT was circulating in the region between May 2012 and February 2014, the time window between

enrollments of our earliest and latest discrepant TRAC participants.

16S rRNA gene microbiome analysis also revealed a total of five participants with

discrepant NG AC2 testing enrolled from April 2011 to July 2013 (**Fig. 1A, Table 2**), including one

participant, 3-209, who also tested discrepantly for CT (**Table 1**). One participant reported mild

urinary symptoms with abnormal vaginal discharge and irregular bleeding and one other

participant reported bleeding with intercourse. Clinical findings on pelvic examination were

noteworthy for bleeding with cervical contact for four of five participants but were otherwise

unremarkable. Only one of these participants reported contact with an infected partner within the

preceding three months but had not received treatment.

The genus *Neisseria* includes a group of commensal species, *N. subflava*, *N. mucosa*, *N.*

elongata, in addition to the important human pathogens, *N. gonorrhoeae* and *N. meningitidis* with

the potential for genetic exchange between these naturally competent microorganisms [28] so we

initially speculated that the ASV detected lacked discriminatory capacity. However, alignment of

the 16S V4 sequence from *N. gonorrhoeae* with the 35 *Neisseria* species included in the RDP

v18 database, revealed that ASV ZOtu95, detected in the discrepant samples, was highly specific,

aligning with 100% identity between bases 535-787 of the 1544 nucleotide V4 region. This

suggested that the 16S ASV detected in libraries generated from the vaginal swabs reflected true

gonococcal infection. Next, we compared the number of ZOtu95 reads from discordant individuals

with those detected in libraries generated from TRAC participants with concordant NG detection

(Table 2) as a surrogate for pathogen load. The number of reads counted in NG+ samples ranged

across four orders of magnitude but participant swab samples yielding libraries with ≤6 counts still

tested positively via AC2 and no difference was detected between the groups using the non-

parametric Mann-Whitney test (P=0.4116). Low levels of infection may have contributed to

differential sampling of bacterial populations present at the cervix (diagnostic) and in the vagina

(16S analysis). Quantifying NG by qPCR in residual template was difficult because of the minimal

amount of cervical DNA remaining and the potential for high genetic variability between strains.

Using primers that targeted multicopy (16S, 23S rRNA) and single copy (ParA pseudogene) loci,

with genomic DNA extracted from *N. gonorrhoeae* strain FA19 as a positive control [29], we

performed qPCR on diagnostic swab eluates or the 16S template that remained from discrepant

NG+ participants (Table 2). Evaluating the performance of the primer pairs using material

obtained from participants testing positively for NG infection, we observed that only 8 of the 17

samples consistently yielded PCR product for all 3 loci, while 6, were positive for 1-2. We were

unsuccessful in amplifying material from the remaining 3 samples. Overall, successful

amplification of the targets did not reflect abundance of 16S read counts in participants' 16S rRNA

libraries. For samples derived from participants who had tested negatively clinically but with

ZOtu95 reads detected via 16S rRNA microbiome profiling, all tested positively for at least one

multicopy locus.

With no source for cross-contamination available in the lab performing the 16S rRNA

analysis, we also looked retrospectively at participant enrollment dates to exclude the remote

possibility of cross- or environmental contamination [30] occurring during sample collection or

NAAT testing. In only one instance did we observe enrollment of a discrepant NG- (3-121) and a

confirmed clinically NG+ participant within the same 48-hour window (3-120). However, the

presence of additional, neisserial species ASVs in the library generated from 3-120 only (Table

2), is inconsistent with contamination of the discrepant NG- sample by material derived from this

diagnostic swab.

Discussion:

This study revealed that non-congruent CT infection outcomes that we detected between

clinical diagnostics and 16S analysis for cervicovaginal samples collected from TRAC enrollees

are best explained as resulting from false negative NAAT tests. The point mutation in 23S that we

detected in discordant TRAC participants has been associated with reduced probe binding with

the original AC2 assay. The G1523A variant was detected in the UK after clinical samples were

screened for probe-avoiding mutants that might be in circulation [12] and in Denmark [13]. This is

the first report of its detection in the United States although Katz et al., reported other 23S variants

that mapped to A1518G or G1526A escaping detection in specimens collected in a five-week

window, spanning late 2019 to early 2020 (2 of 401) [31]. Our retrospective analysis suggests

that diagnostic-avoidant CT was already circulating in western Pennsylvania, USA by 2012 or

possibly earlier. The first participants of the TRAC cohort enrolled in February 2011, but we did

not identify diagnostically discordant enrollment samples collected prior to May 2012. Provided

that this mutation does not impair chlamydial fitness, it is likely that it continues to circulate. The

first 23S SNP variant escaping detection by the AC2 assay, FI-nvCT (C1515T), is still present in

Finland at an estimated prevalence of 1.8% [32]. Fortunately, the deficit in the clinical test has

been recognized and remediated by the manufacturer via incorporation of an additional

chlamydial target [26] to improve reliability.

While discordancy between the clinical diagnostic assay for CT 23S RNA and detection

of CT 16S rRNA, whether by qPCR or in the 16S rRNA libraries generated for microbiome analysis

was ultimately elucidated by our identification of a genetic variant, the basis for the discrepancy

of detection of NG 16S ASVs in participant samples from study participants who were diagnosed

NG negative via the AC2 assay is not easily explained. The NG probe of the AC2 assay is directed

against 16S rRNA and is both sensitive and highly specific [33-35]. Multiplex diagnostic tests have

been reported to have high sensitivity and high concordance with standard nucleic acid

amplification testing [36, 37] with a 4-in-1 test for CT, NG, *Trichomonas vaginalis*, and *M.*

genitalium receiving recent clearance by the FDA [38]. False positive results for NG have been

obtained arising from recombination between 16S loci of *N. meningitidis*, now recognized as a

urogenital pathogen [39-42] and NG [43, 44]. The presence of multiple ribosomal RNA operons

in this genus [45] elevates the risk for even rare recombination events to become fixed in the

population. Should the inverse recombination event also occur, and NG acquire 16S copies from

other neisserial species, this could lead to reduced sensitivity of currently available diagnostic

tests, the consequences of which could negatively impact infection control and promote spread

of diagnostic-avoidant mutants as was observed with Swedish nvCT [46, 47]. Identification of

highly conserved, alternative targets for detection of NG has proven extremely challenging, so it

is not surprising that we were unable to positively confirm the presence of genomic DNA in all the

samples. However, we were also unable to amplify even multicopy targets from specimens

collected from participants that tested positively by clinical NAAT, raising the possibility of low

abundance bacteria that are asymmetrically distributed between the cervix and vaginal vault of

discrepant enrollees.

Fortunately, study participants of TRAC were not harmed by failure to detect their CT

and/or NG infection because all recipients were administered appropriate antibiotic therapy at

enrollment regardless of their NAAT result. Appropriately, health care providers are moving from

empiric syndromic treatment of patients at risk for STIs to pathogen-specific treatment guided by

results of molecular tests [48]. Untreated CT and NG infections can lead to severe complications

including PID, infertility, ectopic pregnancy and chronic pelvic pain in women. While NAATs are

highly accurate, our study demonstrates that clinicians and epidemiologists should remain vigilant

for potential CT or NG variants evading current approved NAATs such as detection of negative

test results in a recently exposed individual, or a decline in overall rates of infection in a particular

geographic location as occurred with Swedish nvCT [46, 47].

Declarations:

Ethics approval and consent to participate: The study protocol was approved by the

institutional review boards for human research of the University of Pittsburgh (#PRO10010159)

and the University of North Carolina (#13-3074). All participants provided written informed consent

at the time of enrollment.

Consent for publication: Not applicable.

Availability of data and materials: The raw sequencing data supporting the conclusions of this

article is publicly available at NCBI's Short Read Archive (SRA) under BioProject accession

number PRJNA1136868.

Competing interests: The authors declare that they have no competing interests.

Funding: This work was supported by the National Institute of Allergy and Infectious Diseases

via U19 AI084024 and R01 AI170959.

Authors' contributions: SJ, KSY, XP, and CMO designed research studies. HCW and SLH

directed collection of participant samples and performance of diagnostic testing. TT, HC, TY and

PB performed experiments. SJ, TT, HB, HC, and XP analyzed data. SJ, KSY, TD, XP, and CMO

wrote the manuscript. All authors reviewed the manuscript.

Acknowledgements: We thank the women who agreed to participate in this study; Ingrid Macio,

Melinda Petrina, Carol Priest, Abi Jett, and Lorna Rabe for their efforts in the clinic and the

microbiology laboratory; and the staff at the Allegheny County Health Department STD Clinic, for

their efforts. We would also like to acknowledge the contribution of Dr. Ian Huntress, who sadly

passed away during the performance of these studies.

References:

1. Russell AN, Zheng X, O'Connell CM, Taylor BD, Wiesenfeld HC, Hillier SL, Zhong W, Darville T: **Analysis of Factors Driving Incident and Ascending Infection and the Role of Serum Antibody in Chlamydia trachomatis Genital Tract Infection**. *J Infect Dis* 2016, **213**(4):523-531.

2. Russell AN, Zheng X, O'Connell CM, Wiesenfeld HC, Hillier SL, Taylor BD, Picard MD, Flechtner JB, Zhong W, Frazer LC *et al*: **Identification of Chlamydia trachomatis Antigens Recognized by T Cells From Highly Exposed Women Who Limit or Resist Genital Tract Infection**. *J Infect Dis* 2016, **214**(12):1884-1892.

3. Liu C, Mokashi Neha V, Darville T, Sun X, O'Connell Catherine M, Hufnagel K, Waterboer T, Zheng X: **A Machine Learning-Based Analytic Pipeline Applied to Clinical and Serum IgG Immunoproteome Data To Predict Chlamydia trachomatis Genital Tract Ascension and Incident Infection in Women**. *Microbiology Spectrum* 2023, **0**(0):e04689-04622.

4. Liu C, Hufnagel K, O'Connell CM, Goonetilleke N, Mokashi N, Waterboer T, Tollison TS, Peng X, Wiesenfeld HC, Hillier SL *et al*: **Reduced Endometrial Ascension and Enhanced Reinfection Associated With Immunoglobulin G Antibodies to Specific Chlamydia trachomatis Proteins in Women at Risk for Chlamydia**. *J Infect Dis* 2022, **225**(5):846-855.

5. Darville T, Albritton HL, Zhong W, Dong L, O'Connell CM, Poston TB, Quayle AJ, Goonetilleke N, Wiesenfeld HC, Hillier SL *et al*: **Anti-chlamydia IgG and IgA are insufficient to prevent endometrial chlamydia infection in women, and increased anti-chlamydia IgG is associated with enhanced risk for incident infection**. *Am J Reprod Immunol* 2019, **81**(5):e13103.

6. Poston TB, Lee DE, Darville T, Zhong W, Dong L, O'Connell CM, Wiesenfeld HC, Hillier

SL, Sempowski GD, Zheng X: **Cervical Cytokines Associated With Chlamydia**

trachomatis Susceptibility and Protection. *J Infect Dis* 2019, **220**(2):330-339.

7. Zheng X, O'Connell CM, Zhong W, Nagarajan UM, Tripathy M, Lee D, Russell AN,

Wiesenfeld H, Hillier S, Darville T: **Discovery of Blood Transcriptional Endotypes in**

Women with Pelvic Inflammatory Disease. *J Immunol* 2018, **200**(8):2941-2956.

8. Zheng X, Zhong W, O'Connell CM, Liu Y, Haggerty CL, Geisler WM, Anyalechi GE,

Kirkcaldy RD, Wiesenfeld HC, Hillier SL *et al*: **Host Genetic Risk Factors for**

Chlamydia trachomatis-Related Infertility in Women. *J Infect Dis* 2021,

224(Supplement_2):S64-s71.

9. Zheng X, O'Connell CM, Zhong W, Poston TB, Wiesenfeld HC, Hillier SL, Trent M,

Gaydos C, Tseng G, Taylor BD *et al*: **Gene Expression Signatures Can Aid**

Diagnosis of Sexually Transmitted Infection-Induced Endometritis in Women.

Front Cell Infect Microbiol 2018, **8**(307):307.

10. Taylor BD, Zheng X, O'Connell CM, Wiesenfeld HC, Hillier SL, Darville T: **Risk factors**

for Mycoplasma genitalium endometritis and incident infection: a secondary data

analysis of the T cell Response Against Chlamydia (TRAC) Study. *Sex Transm*

Infect 2018, **94**(6):414-420.

11. Jeong S, Tollison T, Brochu H, Chou H, Huntress I, Yount KS, Zheng X, Darville T,

O'Connell CM, Peng X: **Cervicovaginal microbial features predicts**

≪em>Chlamydia trachomatis≪/em> spread to the upper genital tract of

infected women. *bioRxiv* 2024:2024.2011.2026.625070.

12. Cole MJ, Davis GS, Fifer H, Saunders JM, Unemo M, Hadad R, Roberts DJ, Fazal M,

Day MJ, Minshull J *et al*: **No widespread dissemination of Chlamydia trachomatis**

diagnostic-escape variants and the impact of Neisseria gonorrhoeae positivity on

the Aptima Combo 2 assay. *Sex Transm Infect* 2022, **98**(5):366-370.

371 13. Hadad R, Jensen JS, Westh H, Gronbaek I, Schwartz LJ, Nielsen L, Vang TM, Nielsen

372   R, Weinreich LS, Skov MN *et al*: **A Chlamydia trachomatis 23S rRNA G1523A variant**

373   **escaping detection in the Aptima Combo 2 assay (Hologic) was widespread**

374   **across Denmark in July-September 2019**. *APMIS* 2020, **128**(6):440-444.

375 14. Roberts DJ, Davis GS, Cole MJ, Naik D, Maru H, Woodford N, Muir P, Horner P, Simms

376   I, Thickett G *et al*: **Prevalence of new variants of Chlamydia trachomatis escaping**

377   **detection by the Aptima Combo 2 assay, England, June to August 2019**. *Euro*

378   *Surveill* 2019, **24**(38).

379 15. Edgar RC: **Search and clustering orders of magnitude faster than BLAST**.

380   *Bioinformatics* 2010, **26**(19):2460-2461.

381 16. O'Connell CM, Brochu H, Girardi J, Harrell E, Jones A, Darville T, Sena AC, Peng X:

382   **Simultaneous profiling of sexually transmitted bacterial pathogens, microbiome,**

383   **and concordant host response in cervical samples using whole transcriptome**

384   **sequencing analysis**. *Microb Cell* 2019, **6**(3):177-183.

385 17. Hazel A, Ponnaluri-Wears S, Davis GS, Low BS, Foxman B: **High prevalence of**

386   **Neisseria gonorrhoeae in a remote, undertreated population of Namibian**

387   **pastoralists**. *Epidemiol Infect* 2014, **142**(11):2422-2432.

388 18. Mitchev N, Singh R, Ramsuran V, Ismail A, Allam M, Kwenda S, Mnyameni F, Garrett N,

389   Swe Swe-Han K, Niehaus AJ *et al*: **Assessment of Antibiotic Resistance and Efflux**

390   **Pump Gene Expression in Neisseria Gonorrhoeae Isolates from South Africa by**

391   **Quantitative Real-Time PCR and Regression Analysis**. *Int J Microbiol* 2022,

392   **2022**:7318325.

393 19. Hokynar K, Rantakokko-Jalava K, Hakanen A, Havana M, Mannonen L, Jokela P,

394   Kurkela S, Lappalainen M, Unemo M, Puolakkainen M: **The Finnish new variant of**

395   **Chlamydia trachomatis with a single nucleotide polymorphism in the 23S rRNA**

target escapes detection by the Aptima Combo 2 test. *Microorganisms* 2019,

7(8):227.

20. **R: A Language and Environment for Statistical Computing**. In. Edited by Team RC:

R Foundation for Statistical Computing; 2015.

21. Wickham H: **ggplot: Elegant graphics for data analysis**. New York: Springer-Verlag

2016.

22. Cole JR, Wang Q, Fish JA, Chai B, McGarrell DM, Sun Y, Brown CT, Porras-Alfaro A,

Kuske CR, Tiedje JM: **Ribosomal Database Project: data and tools for high**

throughput rRNA analysis. *Nucleic Acids Res* 2014, **42**(Database issue):D633-642.

23. Robert CE: **SINTAX: a simple non-Bayesian taxonomy classifier for 16S and ITS**

sequences. *bioRxiv* 2016:074161.

24. Cook RL, Hutchison SL, Ostergaard L, Braithwaite RS, Ness RB: **Systematic review:**

noninvasive testing for Chlamydia trachomatis and Neisseria gonorrhoeae. *Ann*

Intern Med 2005, **142**(11):914-925.

25. Centers for Disease C, Prevention: **Recommendations for the laboratory-based**

detection of Chlamydia trachomatis and Neisseria gonorrhoeae--2014. *MMWR*

Recomm Rep 2014, **63**(RR-02):1-19.

26. Unemo M, Hansen M, Hadad R, Puolakkainen M, Westh H, Rantakokko-Jalava K,

Thilesen C, Cole MJ, Boiko I, Lan PT *et al*: **Sensitivity, specificity, inclusivity and**

exclusivity of the updated Aptima Combo 2 assay, which provides detection

coverage of the new diagnostic-escape Chlamydia trachomatis variants. *BMC*

Infect Dis 2020, **20**(1):419.

27. Rantakokko-Jalava K, Hokynar K, Hieta N, Keskitalo A, Jokela P, Muotiala A, Jokiranta

TS, Kuusela R, Sarkkinen H, Aittoniemi J *et al*: **Chlamydia trachomatis samples**

testing falsely negative in the Aptima Combo 2 test in Finland, 2019. *Euro Surveill*

2019, **24**(22).

28.   Smith NH, Holmes EC, Donovan GM, Carpenter GA, Spratt BG: **Networks and groups**

423   **within the genus Neisseria: analysis of argF, recA, rho, and 16S rRNA sequences**

424   **from human Neisseria species**. *Mol Biol Evol* 1999, **16**(6):773-783.

425   29.   Abrams AJ, Trees DL, Nicholas RA: **Complete Genome Sequences of Three**

426   **Neisseria gonorrhoeae Laboratory Reference Strains, Determined Using PacBio**

427   **Single-Molecule Real-Time Technology**. *Genome Announc* 2015, **3**(5).

428   30.   Ramsden S, Ovens K, Griffiths J, Muir P, Steele-Nicholson A, Horner PJ:

429   **Environmental contamination by Chlamydia trachomatis and Neisseria**

430   **gonorrhoeae: is it time to change our infection control practices? Results of a**

431   **regional study**. *Sex Transm Infect* 2023, **99**(1):35-40.

432   31.   Katz SS, Danavall DC, Morris MR, Herrod BP, Dale SE, Nye MB, Kersh EN, Kirkcaldy

433   RD, Raphael BH: **Chlamydia trachomatis Variants Escaping Detection in the**

434   **Aptima Combo 2 Assay in the United States**. *Sex Transm Dis* 2022, **49**(6):448-452.

435   32.   Puolakkainen M, Mannonen L, Unemo M: **The diagnostic-escape Finnish new variant**

436   **of Chlamydia trachomatis first described in 2019 remains present in Finland three**

437   **years after the introduction of an updated Aptima Combo 2 assay**. *Clin Microbiol*

438   *Infect* 2023, **29**(5):658-659.

439   33.   Bristow CC, McGrath MR, Cohen AC, Anderson LJ, Gordon KK, Klausner JD:

440   **Comparative evaluation of two nucleic acid amplification tests for the detection of**

441   **Chlamydia trachomatis and Neisseria gonorrhoeae at extragenital sites**. *Sexually*

442   *transmitted diseases* 2017, **44**(7):398.

443   34.   Golparian D, Tabrizi SN, Unemo M: **Analytical specificity and sensitivity of the**

444   **APTIMA Combo 2 and APTIMA GC assays for detection of commensal Neisseria**

445   **species and Neisseria gonorrhoeae on the Gen-Probe Panther instrument**.

446   *Sexually Transmitted Diseases* 2013, **40**(2):175-178.

35. Tabrizi SN, Unemo M, Limnios AE, Hogan TR, Hjelmevoll S-O, Garland SM, Tapsall J:

Evaluation of six commercial nucleic acid amplification tests for detection of

Neisseria gonorrhoeae and other Neisseria species. *Journal of clinical microbiology*

2011, **49**(10):3610-3615.

36. Matysiak C, Cheng A, Kirby JE: **Evaluation of the Abbott Alinity m STI assay for**

diagnosis of the primary causes of sexually transmitted infections in the United

States. *Pract Lab Med* 2023, **36**:e00332.

37. Kriesel JD, Bhatia AS, Barrus C, Vaughn M, Gardner J, Crisp RJ: **Multiplex PCR**

testing for nine different sexually transmitted infections. *Int J STD AIDS* 2016,

27(14):1275-1282.

38. Administration USFD: **Alinity m CMV Assay - P210022**. In.; 2022.

39. Urra E, Alkorta M, Sota M, Alcalá B, Martínez I, Barrón J, Cisterna R: **Orogenital**

transmission of Neisseria meningitidis serogroup C confirmed by genotyping

techniques. *Eur J Clin Microbiol Infect Dis* 2005, **24**(1):51-53.

40. Katz AR, Chasnoff R, Komeya A, Lee MV: **Neisseria meningitidis urethritis: a case**

report highlighting clinical similarities to and epidemiological differences from

gonococcal urethritis. *Sex Transm Dis* 2011, **38**(5):439-441.

41. Kanemitsu N, Hayashi I, Satoh N, Hayakawa T, Mitsuya H, Hayase Y, Hiramoto K,

Kojima M: **Acute urethritis caused by Neisseria meningitidis**. *Int J Urol* 2003,

10(6):346-347.

42. Bazan JA, Peterson AS, Kirkcaldy RD, Briere EC, Maierhofer C, Turner AN, Licon DB,

Parker N, Dennison A, Ervin M *et al*: **Notes from the Field: Increase in Neisseria**

meningitidis-Associated Urethritis Among Men at Two Sentinel Clinics -

Columbus, Ohio, and Oakland County, Michigan, 2015. *MMWR Morb Mortal Wkly*

Rep 2016, **65**(21):550-552.

43.   Walcher M, Skvoretz R, Montgomery-Fullerton M, Jonas V, Brentano S: **Description of**

an unusual Neisseria meningitidis isolate containing and expressing Neisseria

gonorrhoeae-Specific 16S rRNA gene sequences. *J Clin Microbiol* 2013,

51(10):3199-3206.

44.   Sukhum KV, Jean S, Wallace M, Anderson N, Burnham CA, Dantas G: **Genomic**

Characterization of Emerging Bacterial Uropathogen Neisseria meningitidis,

Which Was Misidentified as Neisseria gonorrhoeae by Nucleic Acid Amplification

Testing. *J Clin Microbiol* 2021, **59**(2).

45.   Klappenbach JA, Saxman PR, Cole JR, Schmidt TM: **rrndb: the Ribosomal RNA**

Operon Copy Number Database. *Nucleic Acids Res* 2001, **29**(1):181-184.

46.   Ripa T, Nilsson P: **A variant of Chlamydia trachomatis with deletion in cryptic**

plasmid: implications for use of PCR diagnostic tests. *EuroSurveill* 2006,

11(11):E061109.

47.   Unemo M, Seth-Smith HM, Cutcliffe LT, Skilton RJ, Barlow D, Goulding D, Persson K,

Harris SR, Kelly A, Bjartling C *et al*: **The Swedish new variant of Chlamydia**

trachomatis: genome sequence, morphology, cell tropism and phenotypic

characterization. *Microbiology*, **156**(Pt 5):1394-1404.

48.   Karellis A, Naeem F, Nair S, Mallya SD, Routy JP, Gahagan J, Yansouni CP, Kim J, Pai

NP: **Multiplexed rapid technologies for sexually transmitted infections: a**

systematic review. *Lancet Microbe* 2022, **3**(4):e303-e315.

492
493

Table 2. *Neisseria sp.* transcripts detected by 16S ASV analysis and genes detected by qPCR in diagnostic-discrepant or NG+ TRAC participants.

Sample ID	Diagnostic test		16S rRNA read counts				qPCR†		
	Cervix	Endo	ZOtu95: N_gonorrhoeae	ZOtu141: N_subflava	ZOtu264: N_mucosa	ZOtu1566: N_elongata	16S	23S	ParA
3-121	-	-	130	0	0	0	+	+	-
3-140	-	-	123	0	0	0	+	+	+
3-146	-	-	106	0	0	0	+	+	ND
3-151	-	-	6	0	0	0	+	-	+
3-209	-	-	1	0	0	0	-	+	ND
3-012	+	-	471	0	0	0	+	+	+
3-026	+	-	10	0	0	0	-	-	-
3-052	+	-	10	0	0	0	-	+	-
3-089	+	-	24	0	0	0	-	+	+
3-091	+	-	6	0	0	0	+	+	+
3-112	+	-	1032	0	0	0	+	+	+
3-114	+	-	85	0	0	0	-	-	-
3-120	+	-	19	211	131	0	-	+	-
3-122	+	-	110	0	0	0	+	+	+
3-124	+	-	927	0	0	0	-	+	-
3-133	+	-	3	0	0	0	-	-	-
3-136	+	-	2	0	1	0	+	+	-
3-177	+	-	180	0	0	0	+	+	+
3-208	+	-	1151	0	0	0	-	+	-
3-212	+	-	9	0	0	0	+	+	+
3-223	+	-	206	0	0	0	+	+	+
3-235	+	-	349	0	0	0	+	+	+

†Assay performed with diagnostic swab eluate or vaginal swab template used for 16S rRNA microbiome analysis. ND not done because specimen was insufficient or unavailable.

Figure 1. STI-derived 16S transcripts detected in vaginal swabs. a) 16S rRNA gene read

counts of STIs stratified by infection status at enrollment. b) Concordance between CT 16S rRNA

gene read counts versus CT burden by qPCR, where samples are colored by infection status at

enrollment. A LOESS curve is shown in red and Spearman correlation in the bottom-right. CT,

Chlamydia trachomatis; GC, *N. gonorrhoeae*; MG, *M. genitalium*

Figure 2. Portion of sequence alignment of V3 amplified PCR products from discrepant

TRAC participant specimens revealing a g to a substitution at 1523 in the 23S rDNA locus.

PCR products from two representative CT+ samples, as determined by AC2 diagnostic test were

wild type at this position (data not shown).

513
514