Title: The relationships between cochlear nerve health and AzBio sentence scores in quiet and noise in postlingually deafened adult cochlear implant users

Authors: Zi Gao[1], PhD; Yi Yuan[2], PhD; Jacob J. Oleson[3], PhD; Christopher R. Mueller[1]; Ian C. Bruce[4], PhD; René H. Gifford[5], PhD; Shuman He[1], MD, PhD

Affiliations: [1]Department of Otolaryngology – Head and Neck Surgery, The Ohio State University, Columbus, OH 43212

[2]Department of Audiology, San José State University, San José, CA 95192

[3]Department of Biostatistics, The University of Iowa, Iowa City, IA 52242

[4]Department of Electrical and Computer Engineering, McMaster University, Hamilton, ON, L8S 4K1, Canada

[5]Department of Hearing and Speech Sciences, Vanderbilt School of Medicine, Nashville, TN 37232

Correspondence: Shuman He, MD, PhD

Eye and Ear Institute

Department of Otolaryngology – Head and Neck Surgery

The Ohio State University

915 Olentangy River Road, Suite 4000

Columbus, OH 43212

Phone: 614-293-5963

Fax: 614-293-7292

Email: Shuman.He@osumc.edu

24   **Conflict of Interest:** None.

25   **Source of Funding:** This work was supported by grants from the National Institutes of Health

26   awarded to SH [grant numbers 1R01 DC016038 and R21 DC019458].

Author Contributions: ZG participated in data analysis, drafted and approved the final version

of this paper. YY participated in data collection, provided critical comments, and approved the

final version of this paper. JJO conducted statistical analyses, provided critical comments, and

approved the final version of this paper. CRM participated in data collection and analysis,

provided critical comments, and approved the final version of this paper. ICB and RHG provided

critical comments and approved the final version of this paper. SH designed this study,

participated in data analysis, provided critical comments, and approved the final version of this

paper.

The relationships between cochlear nerve health and AzBio sentence scores in quiet and

noise in postlingually deafened adult cochlear implant users

39   **ABSTRACT**

Objectives: This study investigated the relationships between the cochlear nerve (CN) health

and sentence-level speech perception outcomes measured in quiet and noise in postlingually

deafened adult cochlear implant (CI) users.

Design: Study participants included 28 postlingually deafened adult CI users with a Cochlear®

Nucleus™ device. For each participant, only one ear was tested. Neural health of the CN was

assessed at three or four electrode locations across the electrode array using two parameters

derived from results of the electrically evoked compound action potential (eCAP). One

parameter was the phase locking value (PLV) which estimated neural synchrony in the CN. The

other parameter was the sensitivity of the eCAP amplitude growth function (AGF) slope to

changes in the interphase gap (IPG) of biphasic electrical pulses (i.e., the $IPGE_{slope}$). Speech

perception was tested using AzBio sentences in both quiet and a ten-talker babble background

noise with +5 dB and +10 dB signal-to-noise ratios (SNR). $IPGE_{slope}$ and PLV values were

averaged across electrodes for each subject, both with and without weighting by the frequency

importance function (FIF) of the AzBio sentences. Pearson and Spearman correlations were used

to assess the pairwise relationships between the $IPGE_{slope}$, the PLV, and age. Multiple linear

regression models with AzBio score as the outcome and the PLV and the $IPGE_{slope}$ as predictors

were used to evaluate the associations between the three variables while controlling for age.

Results: The IPGE_{slope} and the PLV demonstrated different patterns with regards to their

relationships with electrode location, age, and speech perception. The PLV, but not the IPGE_{slope},

differed significantly across electrodes, where the apical electrodes had larger PLVs (better

neural synchrony) than the basal electrodes. The IPGE_{slope}, but not the PLV, was significantly

correlated with participant's age, where smaller IPGE_{slope} values (poorer spiral ganglion neuron

density) were associated with more advanced age. The PLV, but not the IPGE_{slope}, was

significantly associated with AzBio scores in the +5 dB SNR condition, where larger PLVs

predicted better speech perception. Neither the PLV nor the IPGE_{slope} was significantly

associated with AzBio score in quiet or in the +10 dB SNR condition. The result patterns

remained the same regardless of whether the mean values of the IPGE_{slope} and the PLV were

weighted by the AzBio FIF. The result patterns generally did not change with fitting methods or

input/output scales of the AGF slopes.

Conclusions: The IPGE_{slope} and the PLV quantify different aspects of CN health. The positive

association between the PLV and AzBio scores in the +5 dB SNR condition suggests that neural

synchrony is important for speech perception in adult CI users in challenging listening conditions

with a relatively high noise level. The lack of association between age and the PLV indicates that

reduced neural synchrony in the CN is unlikely the primary factor accounting for the greater

deficits in understanding speech in noise observed in older CI users, as compared to middle-aged

CI users.

Key words: cochlear implants, cochlear nerve, neural synchrony, neural health, speech

perception

INTRODUCTION

The cochlear implant (CI), a prosthesis that partially restores hearing through stimulating the cochlear nerve (CN) via electrodes surgically implanted into the inner ear, is a standard treatment option for listeners with sensorineural hearing loss (for a review, see Zeng 2004). Since electrical stimulation takes place at the auditory periphery, subsequent transmission of the signals by the CN is a prerequisite for the central auditory system's ability to access and process the sound information. Therefore, it is believed that the neural health of the CN is crucial for the success of CI treatment (e.g., He et al. 2017; Zamaninezhad et al. 2023). The association between the CN health and hearing performance in CI users has been supported by post-mortem observations, where within-subject between-ear comparisons showed that the ear with a larger amount of spiral ganglion neurons (SGNs) consistently yielded a better word recognition performance (Seyyedi et al. 2014). However, due to the invasiveness of the histological procedures, direct examination of the CN is not feasible in living human CI users. Rather, non-invasive electrophysiological measures, such as the electrically evoked compound action potential (eCAP), have been developed to assess the CN health status of human CI users in research and clinical settings.

The eCAP is a near-field recorded, synchronized response of a population of CN fibers elicited by electrically stimulating a CI electrode, which has been used to evaluate neural encoding of electrical stimulation at the CN, such as spectral resolution (Won et al. 2014), neural adaptation (Hughes et al. 2012; He et al. 2023b), and amplitude modulation encoding (Tejani et al. 2017) in CI users. Morphologically, a typical eCAP waveform consists of a negative peak (N1) at around 0.2-0.4 ms after stimulus onset, followed by a positive peak (P2) at around 0.6-0.8 ms after stimulus onset (e.g., Brown et al. 1990). The amplitude of the eCAP waveform is

defined as the difference in voltages between P2 and N1, and it increases with the stimulation

level. The relationship between stimulation level and eCAP amplitude can be depicted using an

amplitude growth function (AGF), also known as the input/output (I/O) function.

The slope of eCAP AGF has been shown to be associated with the density of SGNs in

animal studies, where steeper slopes indicate higher SGN density in pharmaceutically deafened,

implanted animals (Pfingst et al. 2015). Aligned with the animal results, shallower eCAP slopes

were observed in pediatric CI users with cochlear nerve deficiency (CND) compared to those

with normal-sized CNs (He et al. 2018). However, since the raw eCAP responses are susceptible

to inter-patient and inter-electrode differences in non-neural factors (Brochier et al. 2021), such

as electrode impedance and electrode-modiolus distance (Imsiecke et al. 2021), researchers have

been seeking to overcome this drawback by using the differences between eCAP-derived

measurements under various stimulation conditions to assess CN health. One particularly

promising measurement is the sensitivity of the eCAP amplitude to changes in the interphase gap

(IPG) of biphasic, electrical pulses (i.e., the IPG effects). Theoretically, non-neural factors

should affect eCAP amplitude growth functions (AGFs) measured using different IPGs in a

similar way. As a result, the effects of non-neural factors should be reduced or even eliminated

when the differences between AGFs measured at different IPGs are calculated. Aligning with

this theoretical possibility, research studies in both animal models and human CI users have

demonstrated that the IPG effects are not affected by non-neural factors, such as electrode

impedance (Schvartz-Leyzac & Pfingst 2016, 2018; Imsiecke et al. 2021) and electrode-

modiolus distance (Schvartz-Leyzac et al. 2020b; Imsiecke et al. 2021).

In animal models, the IPG effects have been shown to be correlated with SGN survival.

Specifically, larger effects of IPG on eCAP amplitude (Prado-Guitierrez et al. 2006) and the

AGF slope (Ramekers et al. 2014; Schvartz-Leyzac et al. 2019; Schvartz-Leyzac et al. 2020a)

are associated with higher SGN density in guinea pigs. Consistent with these results measured in

animal models, the IPG effect on the AGF slope (IPGE$_{slope}$) has been shown to be positively

correlated with sentence and consonant recognition (Schvartz-Leyzac & Pfingst 2018) and

speech reception threshold (SRT; Zamaninezhad et al. 2023) in postlingually deafened adult CI

users. More importantly, IPGE$_{slope}$ was larger in children with normal-sized CNs than in children

with CND (He et al. 2020; Yuan et al. 2022), a patient population with substantially less SGNs

than other patient populations. Specifically, histological results of human temporal bone studies

have shown that the mean SGN count in children with CND is 5,739 (range: 0-15,714),

representing about 16.4% of the SGN count expected for the corresponding age, which is

roughly 35,500 (Wright et al. 1986; Glueckert et al. 2010; Nelson & Hinojosa 2001; Haginomori

et al. 2002; Chen et al. 2020; Da Costa Monsanto et al. 2022). In comparison, the mean SGN

count in hearing impaired children with normal-sized CNs ranges from 14,301 to 20,523 (Miura

et al. 2002). It should be noted that the axons of existing SGNs in children with CND are

myelinated, with no apparent sign of degeneration (Glueckert et al. 2010; Wright et al. 1986).

Therefore, the results reported in He et al. (2020) and Yuan et al. (2022) provide a strong piece

of evidence supporting the indicative value of IPGE$_{slope}$ for SGN count. Nevertheless, the

sensitivity of IPGE$_{slope}$ as an index of SGN count in CI users cannot be determined due to the

lack of direct and precise measurements of neural structure in living human listeners.

147   The IPG effect on stimulation level offset (IPGE$_{offset}$) has been proposed as another

candidate index of CN health. Based on computational modeling results, Brochier et al. (2021)

proposed that the IPGE$_{offset}$ outperformed the IPGE$_{slope}$ in controlling for non-neural factors.

Consistent with this proposed relationship between IPGE$_{offset}$ and CN health, Jahn and Arenberg

(2020b) observed greater $IPGE_{offset}$ in younger compared to older listeners. However, several

empirical studies failed to demonstrate associations between the $IPGE_{offset}$ and speech perception

measurements in CI users with either full-length electrode array (Zamaninezhad et al. 2023) or

short array to preserve residual low frequency acoustic hearing (Kim et al. 2010). Recent

computational modeling studies on the IPG effects suggested that $IPGE_{slope}$ and $IPGE_{offset}$ might

reflect different aspects of CN health. As noted in the discussion of Brochier et al. (2021),

$IPGE_{offset}$ likely reflects the properties and health status, rather than the number, of surviving

SGNs. Aligned with this distinction, a three-dimensional model of the cochlea by Kipping et al.

(2024) and Zhang et al. (2024) demonstrated correlations between $IPGE_{offset}$ and CN

degeneration, defined as reduced diameter in or entire loss of the peripheral axons. In contrast, a

two-dimension model by Takanen et al. (2024) demonstrated that the $IPGE_{slope}$ calculated as the

numerical difference between the AGF slopes on a linear I/O scale is dependent on neural

survival, and that non-neural factors such as electrode-neuron distance had little interference on

the $IPGE_{slope}$. Taken together, the existing evidence from animal, human and computational

research are converging in suggesting that $IPGE_{slope}$ is a promising indicator of CN density

among the currently available eCAP measurements. As a result, $IPGE_{slope}$ was used to assess

neural survival of the CN in this study.

While having sufficient CN fibers responding to auditory input is a prerequisite for

auditory perception, CN density alone does not guarantee good hearing functions in challenging

listening environments. In theory, effective and accurate representation of sound signals that

allows the listener to separate target signals from noise requires synchronous firing across

neurons, which in turn depends on the health status of the CN, as demonstrated in animal and

computational modeling studies (Kim et al. 2013; Heshmat et al. 2020). Neural

desynchronization leads to a smeared representation of temporal cues, so that even though the

ability to detect sound in quiet may be minimally affected, hearing performance in noise would

degrade drastically. This scenario is exemplified by some listeners with auditory neuropathy

spectrum disorder (ANSD), who have normal or relatively good behavioral audiometric

thresholds and speech perception in quiet, but disproportionally impaired signal detection and

speech perception in noise (Kraus et al. 2000; Zeng et al. 2005). The electrophysiological

measures of patients with ANSD are characterized by a relatively normal cochlear microphonic

and/or otoacoustic emission (OAE) response, and an abnormal or absent auditory brainstem

response (ABR), which has been interpreted as a lack of synchrony across CN fibers despite a

relatively normal hair cell function (e.g., Starr et al. 2008). The crucial role of neural synchrony

in acoustic hearing has been further supported by the compound action potential (CAP) recorded

in normal hearing (NH) listeners, where the level of neural synchrony, quantified with the phase

locking value (PLV) of trial-by-trial CAP measurements, was found to be a strong predictor of

recognition scores for speech in noise and time-compressed speech in quiet (Harris et al. 2021).

Due to the difference between acoustic and electrical hearing, the observations in

listeners with NH or ANSD may not be readily generalizable to CI users. For many CI users,

speech perception in noise is a challenging task despite excellent hearing performance in quiet

(Zaltz et al. 2020). Histological observations of SGN dystrophy and demyelination in listeners

with various hearing profiles (Nadol 1997; Wu et al. 2019) suggest that reduced neural

synchrony could be an underlying cause of poor speech perception in noise in CI users.

However, this proposed relationship has rarely been evaluated, largely due to the lack of

electrophysiological measures of neural synchrony in the CN. We recently developed a new

method to quantify peripheral neural synchrony in CI users, where the PLV of trial-by-trial

eCAP responses was used as an index to quantify the degree of neural synchrony in the

responses generated by CN fibers across multiple electrical stimulations (He et al. 2024). We

demonstrated that higher PLVs are associated with better temporal resolution and smaller effects

of noise on word recognition in post-lingually deafened adult CI users, consistent with the

hypothesized effect of neural synchrony on hearing performance in electrical hearing.

202   In summary, previous research has established both CN survival and neural synchrony as

crucial factors contributing to hearing performance in CI users. However, considering that nerve

damage can result in both lower neural density and poorer synchronization, as has been shown in

computational models (Heshmat et al. 2020), little is known about whether the contributions of

neural survival and synchrony to hearing performance are independent or overlapping.

Observations in NH listeners by Harris et al. (2021) suggest that neural engagement and

synchrony are two separate dimensions that vary differently with changes in stimulus level. In

human CI users, the number and synchrony of excited CN fibers have been modeled using

retrospective deconvolution performed on intraoperative eCAP recordings, both significantly

associated with postoperative speech recognition scores (Dong et al. 2023). However, the model

was built upon assumptions about the shape of unitary response from CN fibers, which have not

been directly validated in humans (Dong et al. 2020; Dong et al. 2023). While the $IPGE_{slope}$ and

the PLV in CI users have been measured postoperatively in separate experimental studies to

assess their association with speech perception (Schvartz-Leyzac & Pfingst 2018; Zamaninezhad

et al. 2023; He et al. 2024), it is unclear whether the biological underpinnings of these two

indices are orthogonal and impact speech perception differently depending on listening

conditions.

To address this critical knowledge gap, we measured the $IPGE_{slope}$, the PLV and speech

perception in the same group of postlingually deafened adult CI users and evaluated their

relationships. Speech perception was measured using AzBio sentences (Spahr et al. 2012), a

speech corpus consisting of multiple lists of everyday sentences with similar levels of difficulty.

Based on previous studies on the effects of neural survival and neural synchrony on speech

perception in listeners with various hearing profiles, we hypothesized that (1) the $IPGE_{slope}$ and

the PLV are two independent measures of CN health (Harris et al. 2021; Kraus et al. 2000); (2)

the $IPGE_{slope}$ is positively associated with speech perception in quiet (Zamaninezhad et al. 2023);

(3) the PLV is positively associated with speech perception in noise (Zeng et al. 2005; Harris et

al. 2021; He et al. 2024).

230   **MATERIALS AND METHODS**

Participants

Study participants included 28 postlingually deafened CI users (age range: 20.39-84.04

years, mean = 61.86 yrs, standard deviation SD = 14.78 yrs; 15 female, 13 male; nine bilateral CI

users). Twenty-one of them also participated in our previous study on the development and

validation of the eCAP PLV measurement (He et al. 2024), and their PLV data were reused in

the current study. All participants were native speakers of American English and used a

Cochlear® Nucleus™ device (Cochlear Ltd., New South Wales, Australia). All participants had

a full 22-electrode insertion with their devices, as confirmed by postoperative computerized

tomography scans. Only one ear was tested in each participant. For bilateral users, the tested ear

was decided through randomization. With the exception of two participants (S21 and S25), none

of the participants had any functional acoustic hearing in either ear. The lack of functional

acoustic hearing was self-reported by the patients and was verified by pure tone audiometric

thresholds of 70 dB HL or higher, averaged across octave frequencies between 250 Hz and 8,000

Hz. For the participants with functional hearing on the opposite side, an ear plug was used to

minimize its potential contribution during the AzBio speech testing. Demographic information

and hearing loss etiology of the participants are listed in Table 1. All participants provided

written informed consent at their initial visit to the lab prior to data collection and were

compensated for their time. The study was approved by the Biomedical Institutional Review

Board at the corresponding author's institution (No. 2017H0131).

Table 1. Demographic information of all participants. The participant IDs are not known to

anyone outside the research group, including the participants themselves. M, male; F, female; L,

left; R, right; SNHL, sensorineural hearing loss; CA, contour advance; DFNB8, deafness,

autosomal recessive 8; DFNB10, deafness, autosomal recessive 10.

Participant	Ear tested	Age range at testing (yrs)	Internal device and electrode array	Hearing loss etiology	Electrodes tested
S01	L	61-65	CI 512	Sudden SNHL	3, 9, 14, 20
S02	L	66-70	CI 512	Meniere's disease	3, 9, 15, 18
S03	R	66-70	CI 24RE(CA)	Hereditary	3, 9, 15, 21
S04	L	36-40	CI 24RE(CA)	Head Trauma	3, 9, 15, 21
S05	R	61-65	CI 24RE(CA)	Unknown	8, 12, 15, 18
S06	L	51-55	CI 532	Unknown	4, 9, 15, 21
S07	R	61-65	CI 522	Head Trauma	6, 9, 18, 20
S08	R	36-40	CI 24RE(CA)	Unknown	3, 9, 15, 21
S09	R	56-60	CI 24RE(CA)	Unknown	3, 9, 15, 21
S10	R	66-70	CI 532	Unknown	3, 9, 15, 21
S11	R	66-70	CI 532	Unknown	3, 9, 15, 21
S12	L	76-80	CI 422	Unknown	4, 9, 15, 21
S13	L	61-65	CI 632	Unknown	3, 9, 15, 21
S14	R	66-70	CI 24RE(CA)	Unknown	3, 15, 21

S15	L	66-70	CI 532	Vestibular Schwannoma	3, 9, 15, 21
S16	R	81-85	CI 532	Unknown	3, 7, 10, 17
S17	L	56-60	CI 532	Unknown	6, 9, 15, 21
S18	L	76-80	CI 622	Unknown	6, 9, 15, 21
S19	R	66-70	CI 632	Head Trauma	3, 9, 15, 21
S20	R	81-85	CI 632	Unknown	3, 9, 15, 21
S21	R	51-55	CI 632	Unknown	3, 9, 15, 21
S22	L	56-60	CI 632	Unknown	3, 15, 18
S23	L	56-60	CI 532	Usher	3, 9, 15, 21
S24	L	76-80	CI 632	Unknown	3, 9, 15, 21
S25	L	36-40	CI 612	Vestibular Schwannoma	6, 15, 21
S26	L	18-20	CI 522	DFNB8/10	3, 9, 15, 21
S27	L	51-55	CI 612	Unknown	5, 9, 15, 21
S28	R	76-80	CI 24RE(CA)	Unknown	9, 12, 16

Stimuli and Apparatus

For eCAP recordings, the stimulus was a charge-balanced, cathodic leading biphasic

pulse with a pulse-phase duration of 25 µs. The IPG between the cathodic and anodic phases, as

well as the presentation levels, varied across measurements, as detailed in the "Procedures"

section. All eCAP recordings were performed using the neural response telemetry function

implemented in the Custom Sound EP software v6.0 (Cochlear Ltd., New South Wales,

Australia).

For speech perception tests, the stimuli were meaningful sentences (e.g., "The vacation

was cancelled on account of weather.") from the AzBio sentence corpus (Spahr et al. 2012),

recorded by two female and two male native American English speakers. The sentences

presented to each participant and for each condition were evenly distributed across the four

speakers. The background noise was a ten-talker babble presented at two signal-to-noise ratios

(SNRs): +10 dB and +5 dB. All stimuli were delivered via a loudspeaker (RadioEar Corporation,

PA) placed 1 m in front of the participant at 0° azimuth in a sound-attenuated booth.

Procedures

Testing Electrodes

The default testing sites for eCAP measurements were electrodes 3, 9, 15, and 21 (i.e., e3,

e9, e15 and e21). These electrodes were selected to cover a wide range along the array with

relatively equal numerical separations in between, while keeping the testing time reasonable. In

the case of an open- or short-circuit or lacking a measurable eCAP at the maximum comfortable

level due to high impedance at a default electrode, a nearby alternative electrode was tested. In

rare cases when multiple neighboring electrodes are untestable, the sites were selected to cover a

majority of the testable regions while keeping relatively equal separations in between. Four

participants (S14, S22, S25, and S28) were tested at only three electrodes due to time constraints.

The electrodes tested for each participant can be found in Table 1.

Behavioral C Level Measures

The maximum comfortable level (C level) for eCAP stimuli at each IPG level (7 µs and

42 µs) was determined via subjective rating using an ascending procedure. Prior to the

measurement, participants were shown a visual scale of 1 ("barely audible") to 10 ("very

uncomfortable") and were instructed to give a loudness rating using verbal responses or hand

gestures following each stimulus presentation. Each presentation consisted of five pulses

delivered at a probe rate of 15 Hz. The stimuli were first presented at a relatively low level and

gradually increased in steps of 3-5 clinical levels (CLs) until a rating of "7" was reached, then in

steps of 1-2 CLs until a rating of "8" was reached. The lowest level that corresponds to a rating

of "8" ("maximal comfort") was recorded as the behavioral C level.

eCAP Measures

The eCAP was measured using a two-pulse, forward-masking paradigm (Brown et al.,

1990), where the masker pulse was always presented at 10 CLs higher than the probe pulse. The

masker pulses were delivered at the testing electrode, and the eCAP responses were recorded two

electrodes away from the testing electrode in the apical direction. There was an exception for

electrode 21, which was recorded two electrodes away in the basal direction (i.e., electrode 19).

The probe pulses were presented at a probe rate of 15 Hz with a masker-probe interval of 400 µs.

The total number of trials in each stimulation sequence differed across measurements, as detailed

below. Responses were recorded at a sampling rate of 20,492 Hz with a sampling delay of 122

µs, an amplifier gain of 50 dB, and a monopolar-coupled stimulation mode.

Measure of Neural Survival: The IPGE$_{slope}$

In this study, the IPGE$_{slope}$ was operationally defined as the numerical difference (in

µV/dB, where dB was calculated in reference to 1 nC) between the AGF slopes with IPGs of 42

µs and 7 µs, and therefore, its measurement involved acquiring an AGF and calculating its slope

at each IPG level for each participant. The IPG of 7 µs was chosen because it was the shortest

IPG value available for eCAP recording in the Custom Sound EP software, and the IPG of 42 µs

was selected to allow for comparisons with previous studies on IPG effects in CI users with a

Cochlear® device (He et al. 2020; Yuan et al. 2022). For both IPG levels, the maximum

presentation level of the stimuli was the behavioral C level measured with an IPG of 42 µs. To

acquire AGFs, the eCAP measurement started at the maximum presentation level and decreased

in steps of 1 CL for five steps, then in steps of 5 CUs until no peaks could be visually identified

in the eCAP waveform, i.e., when the threshold is reached. Additional presentation levels in

steps of 1 CL for five steps near and above the eCAP threshold were tested. At each presentation

level, the eCAP waveform was acquired by averaging the raw responses to 50 pulses. The visual

identification of eCAP peaks, or lack thereof, was performed by the experimenter at the time of

testing and rechecked by an expert researcher offline.

The AGF slope at each IPG level was calculated using the window method developed by

Skidmore et al. (2022), where linear regressions were performed on sliding windows along a

resampled AGF, and the largest slope among all windows was regarded as the AGF slope. The

window method was suitable for calculating the slope of AGFs with various shapes, both in

animals and human listeners. For the results in human listeners, the correlation between AGF

window slope and speech perception outperformed traditional fitting methods such as linear or

sigmoid fitting. In the present study, we focused on the $IPGE_{slope}$ results calculated using the

window method in the linear-logarithmic I/O scale of $\mu V/dB$, following the recommendations by

Skidmore et al. (2022). However, since linear fitting and the linear-linear I/O scale of $\mu V/nC$

were commonly used in previous research on IPG effects (Schvartz-Leyzac & Pfingst 2018;

Zamaninezhad et al. 2023), and that fitting method and I/O scale may affect how well eCAP

measurements correlate with neural health and speech perception (Brochier et al. 2021;

Skidmore et al. 2022), we also calculated the $IPGE_{slope}$ on a linear-linear scale ($\mu V/nC$) with both

linear and window fitting methods to evaluate the robustness of the results across fitting methods

and I/O scales, briefly reported at the end of the Results section. All calculations were performed

using MATLAB 2021b (MathWorks, MA).

Measure of Neural Synchrony: The PLV

For each participant and electrode, the PLV was derived from 400 eCAP trials measured

using biphasic pulses with an IPG of 7 µs presented at the behavioral C level, using the method

developed by He et al. (2024). The IPG of 7 µs was selected for PLV measurement because it is

the default and most used IPG in the programming of Cochlear® CI devices. The PLV is a

unitless value between 0 and 1, where 0 means that the phases were randomly distributed across

trials, and 1 means that the phases were perfectly correlated. To calculate the PLV, the eCAP

responses were re-referenced to the first sample of each trial and time-frequency decomposed at

six linearly spaced frequencies (788.2, 1576.3, 2364.4, 3152.6, 3941.0, and 4729.2 Hz) and

divided into six partially overlapped time frames with an onset-to-onset interval of 48.8 µs and a

length of 1561.6 µs. At each frequency and within each time frame, the unit vectors representing

the phases of the 400 individual trials were averaged, and the length of the averaged vector was

taken as the time-frequency-specific PLV. The formula for calculating the PLV at the time *t* and

the frequency *f* based on *N* individual trials is:

352
$$PLV(f,t) = \left| \frac{1}{N} \sum_{k=1}^{N} \frac{F_k(f,t)}{|F_k(f,t)|} \right|$$

The PLV of the electrode was then obtained by averaging all time-frequency-specific PLVs

calculated from the eCAP responses at that electrode. The time-frequency decomposition and the

calculation of the PLV were performed using MATLAB R2021b and the newtimef.m function

from EEGLAB v2022.1 (Delorme & Makeig 2004).

Measure of Speech Perception: AzBio Scores

Each participant was tested with AzBio sentences (Spahr et al. 2012) under three

conditions: in quiet and in a ten-talker babble background noise with SNRs of +10 and +5 dB,

respectively. The sentences were presented at 60 dB SPL in all conditions. The participants were

tested unilaterally on the side with CI. Specifically, bilateral users were required to take off the

CI on the opposite side. For the two participants with functional acoustic hearing in the opposite

ear, an ear plug was used to minimize the contribution of acoustic hearing. For each participant

and condition, a sentence list was randomly selected from Lists 1-8 of the AzBio corpus, each

consisting of 20 sentences. For each participant, different word lists were used for different

conditions. Participants were instructed to repeat back after each sentence and were encouraged

to guess if they were unsure about what they heard. An experimenter recorded the number of

words they correctly repeated in each sentence. The AzBio score was calculated as the number of

words in the list correctly repeated by the participant, divided by the total number of words in the

list. All words in the list, including prepositions, counted towards the score.

Averaging IPGE$_{slope}$ and PLV Values across Electrodes

As demonstrated in He et al. (2023a), using eCAP results measured at single CI electrode

locations to correlate with auditory perception outcomes in CI users can lead to inaccurate

conclusions. Therefore, the values of the IPGE$_{slope}$ and the PLV were averaged across all tested

electrodes for each participant as an overall representation of CN health across the cochlea (He et

al. 2023a). Both weighted and unweighted averages were calculated for both parameters. To

calculate the weighted average, the results were weighted based on the frequency importance

function (FIF) derived from AzBio scores under various spectral filtering conditions and SNRs

in NH listeners (Lee & Mendel 2017). The FIF was fitted to a four-parameter Weibull function

in SigmaPlot v15 (Grafiti LLC, CA). For each participant, the importance weight of each test

electrode was calculated using the fitted Weibull function based on the electrode's central

frequency derived from the frequency-to-electrode table of the participant's everyday

programming map. The individual values of the empirically measured AzBio FIF and the fitted

curve are shown in Figure 1. The unweighted average was calculated as the arithmetic means of

the IPGE$_{slope}$ and the PLV values across all tested electrodes for each participant.

Figure 1. Individual AzBio FIF values (black circles) measured by Lee and Mendel (2017) and

the fitted Weibull function (gray line)

Data Analysis

The IPGE$_{slope}$ analyses reported in the Results section are based on the AGF slopes

obtained from the window fitting method in the unit of µV/dB, unless otherwise stated. The

electrode-specific IPGE$_{slope}$ and PLV values were compared across electrode locations using

linear mixed-effect models. Pairwise comparisons between the electrodes were performed using

the Tukey method for *p*-value adjustment. The IPGE$_{slope}$ and PLV values were then averaged

across the tested electrodes, weighted by the AzBio FIF, as representations of the overall CN

neural health. The pairwise relationships between the $IPGE_{slope}$, the PLV, and age at testing were

assessed using either Pearson or Spearman correlation tests for variable pairs depending on the

results of the Shapiro-Wilk normality test. Multiple linear regressions with AzBio score as the

outcome and the $IPGE_{slope}$ and the PLV as predictors were performed to evaluate the associations

among the three variables under each testing condition (quiet, +10 dB, and +5 dB SNR).

Participant's age was added to the regression models as a covariate to control for the potential

effects of advanced age on speech perception and/or eCAP measurements, as have been

demonstrated in CI users (Roberts et al. 2013; Sladen & Zappler 2015; Xie et al. 2019; Jahn &

Arenberg 2020a). If the residuals were not approximately normally distributed, the outcome

variables were transformed with appropriate methods to ensure that the normal residual

assumption of linear regression is met. Correlation tests were performed in JASP v0.18.3 (JASP

Team 2024), and the regressions were performed in R v4.4.1 (R Core Team 2024), with the lme4

(Bates et al. 2015), emmeans (Lenth 2024), and lmerTest (Kuznetsova et al. 2017) packages used

for the linear mixed-effect models.

RESULTS

The individual $IPGE_{slope}$ and PLV values measured at each electrode are shown in Figure

2, with the range, mean, and standard deviation (SD) listed in Table 2. To assess the potential

differences in the two measurements across electrodes, two linear mixed-effect regressions were

performed with the $IPGE_{slope}$ and the PLV as the outcome variables, the electrode location as the

fixed effect, and participant as the random effect. The four categories of electrode locations

corresponded to and were named after the four default testing locations: e3, e9, e15, and e21. If

other electrodes were measured in lieu of the default electrodes, they were assigned to one of the

categories based on their locations relative to the other electrodes tested in the same participant.

For example, among the four electrodes tested in participant S05, e8 was the most basal

electrode and therefore categorized as "e3" in the regression model. The category "e3" was used

as the reference level in the regressions. The degrees of freedom were estimated using the

Satterthwaite's method. The results reveal an overall effect for PLV across electrodes ($F_{(3,\ 77.4)} =$

5.44, $p = .002$). Focusing on the pairwise comparisons, only two comparisons showed a

statistically significant difference in PLVs which were e3 compared with the PLVs measured at

e15 ($t_{(77.2)} = 3.47$, $p = .005$) and e21 ($t_{(77.2)} = 3.27$, $p = .009$). The other comparisons were not

significantly different which were e3 with e9 ($t_{(77.7)} = 1.28$, $p = .578$), e9 with e15 ($t_{(77.5)} = 2.09$, p

$=.167$), e9 with e21 ($t_{(77.5)} = 1.89$, $p =.240$), and e15 with e21 ($t_{(77.0)} = 0.20$, $p =.997$). For the

analysis of electrode location effect on the IPGE$_{slope}$, two data points (S08 e9 and S26 e21) were

excluded from the analysis because their IPGE$_{slope}$ values were more than three *SDs* away from

the mean of the corresponding electrode location. The IPGE$_{slope}$ did not significantly differ

across electrodes ($F_{(3,\ 75.2)} = 2.22$, $p = .092$). In the subsequent sections, the IPGE$_{slope}$ and the

PLV refer to the weighted averages of the corresponding values across electrodes for individual

participants unless otherwise stated. It is worth noting that no outliers were removed in the

subsequent analysis, as all PLV and IPGE$_{slope}$ values fell within three *SDs* from the mean when

averaged across all tested electrodes of each participant.

Figure 2. Individual values of (a) the IPGE$_{slope}$ and (b) the PLV by electrode location. The four

categories from basal to apical locations corresponded to, and were named after, the default

electrodes e3, e9, e15, and e21, respectively. Values measured at non-default electrodes were

categorized based on their locations relative to the other electrodes tested in the same participant.

Boxes show the range between the first and the third quartile of the data values. The horizontal

bars inside the boxes represent the median. The vertical whiskers show the range of values that

are within 1.5 interquartile range (IQR) from the boxes. Asterisks denote pairwise comparisons

with significant differences ($p < .05$). N.S., no significant differences.

Table 2. Descriptive statistics of the IPGE$_{slope}$ and the PLV values measured at different

electrode locations. The IPGE$_{slope}$ values were calculated from AGF slopes in µV/dB fitted using

the window method. Two electrodes were identified as outliers for IPGE$_{slope}$ (more than three

SDs away from the mean) and therefore excluded from the descriptive statistics. Values in each

cell are listed in the format of "min, max, mean (*SD*)".

Electrode location	e3	e9	e15	e21
PLV	Min = 0.117 Max = 0.590 0.318 (0.137)	Min = 0.092 Max = 0.734 0.360 (0.161)	Min = 0.109 Max = 0.782 0.409 (0.160)	Min = 0.160 Max = 0.808 0.404 (0.166)
$IPGE_{slope}$	Min = -20.003 Max = 52.438 7.394 (18.073)	Min = -4.982 Max = 41.041 12.136(16.650)	Min = -8.306 Max = 45.231 7.668 (14.652)	Min = -10.068 Max = 41.102 3.701 (2.227)

Correlations between the $IPGE_{slope}$, the PLV, and Age

The eCAP AGFs acquired at different IPG durations from one example participant (S06) are shown in Figure 3. Time-frequency specific PLV values of the same participant are illustrated in Figure 4. Figure 5 shows the standardized values of $IPGE_{slope}$ and PLV of individual participants (panel a), along with their age (panels b-c). All values were within three *SDs* from the mean. Spearman correlation tests showed that there is a non-significant trend that the $IPGE_{slope}$ and the PLV were positively correlated ($rho_{(26)} = 0.340$, $p = .077$). Although not a focus of this study, we assessed the relationships between age and the two CN health measures. Spearman correlation test results revealed a significant negative correlation between the $IPGE_{slope}$ and age ($rho_{(26)} = -0.583$, $p = .001$). Pearson correlation test results showed that PLV was not significantly correlated with age ($r_{(26)} = -0.068$, $p = .730$).

Figure 3. The AGFs measured with IPGs of 7 µs and 42 µs in one participant (S06), where the

maximum stimulation level was set to the C level measured with a 42-µs IPG at each electrode.

The amplitudes were normalized by dividing the amplitude of each trial by the maximum

amplitude among the trials with a 7-µs IPG. Please note that the ranges of the axes are different

across panels due to large variabilities in eCAP thresholds, C-levels, and amplitudes across

electrodes.

Figure 4. PLV values measured in one participant (S06). The time-frequency specific PLV

values are shown in the heatmaps in the lower panels. The overall PLV values of individual

electrodes are listed in the upper panels. Bolded black lines represent eCAP responses averaged

across 400 trials. Responses in individual trials are plotted with gray lines.

Figure 5. Correlations between (a) the IPGE$_{slope}$ and the PLV, (b) the IPGE$_{slope}$ and age, and (c)

the PLV and age. The axes were standardized by converting into z-scores to improve

comparability across panels. Each symbol indicates the result measured in one participant. The

results of correlation tests are shown in each panel.

The Associations among the IPGE$_{slope}$, the PLV and Speech Perception Scores

The relationships between AzBio results in quiet and noise and eCAP measurements are

illustrated in Figure 6. The AzBio scores in quiet were rank transformed (i.e., the lowest and

highest scores were transformed into 1 and 28, respectively) due to non-normally distributed

residuals when the raw scores were used in the linear regression model (not reported here). The

results of linear regressions "AzBio ~ PLV + IPGE$_{slope}$ + age" revealed significant associations

between the PLV and AzBio scores in the noise +5 dB SNR condition ($t_{(24)} = 2.36$, $p = .027$)

after adjusting for IPGE$_{slope}$ and age, where larger PLVs (better neural synchrony) are associated

with higher AzBio scores, but not in the noise +10 dB SNR ($t_{(24)} = 1.61$, $p = .121$) or the quiet

condition ($t_{(24)} = 0.71$, $p = .485$). No significant association between the IPGE$_{slope}$ and AzBio

scores was observed in any testing conditions ($p > .10$ in all cases) after adjusting for PLV and

age. Detailed results of the linear regression models are available in Table 3. It is worth noting

that rank-transformation of the AzBio scores in quiet did not change the pattern or statistical

significance of the results.

Figure 6. The residuals of AzBio results measured in quiet and in noise with +10 and +5 dB

SNRs calculated using reduced linear regression models, plotted against eCAP measurements. In

the top panels, vertical axes represent the residuals of "AzBio ~ PLV + age" to highlight the

relationships between IPGE$_{slope}$ and AzBio results. In the bottom panels, vertical axes represent

the residuals of "AzBio ~ IPGE$_{slope}$ + age" to highlight the relationships between PLV and AzBio

results. The unstandardized beta and *p*-values were obtained from the full model "AzBio ~

IPGE$_{slope}$ + PLV + age" (Table 3) for the corresponding eCAP measurements shown in the

horizontal axis of each panel. Each symbol represents the AzBio result measured in one

participant.

Table 3. Results of linear models examining the relationships between the $IPGE_{slope}$, the PLV,

and AzBio scores measured in quiet and in two noise conditions. The $IPGE_{slope}$ values were

calculated from AGF slopes in µV/dB fitted using the window method. The AzBio scores in the

quiet condition were rank transformed to meet the normal residual assumption of linear

regression.

Listening condition	Predictor	Unstandardized β (SE)	t	p	Multiple R^2
Quiet	$IPGE_{slope}$	0.1003 (0.1411)	0.711	.4839	
	PLV	8.4126 (11.8582)	0.709	.4849	0.2312
	age	-0.1561(0.1421)	-1.098	.2829	
Noise, +10 dB SNR	$IPGE_{slope}$	-0.0525(0.5075)	-0.104	.9184	
	PLV	68.6727 (42.6512)	1.610	.1205	0.1555
	age	-0.4376 (0.5111)	-0.856	.4004	
Noise, +5 dB SNR	$IPGE_{slope}$	-0.2777 (0.3490)	-0.796	.4340	
	PLV	69.1906 (29.3348)	2.359	.0268*	0.3161
	age	-0.7400 (0.3515)	-2.105	.0459*	

Electrode Weighting by AzBio FIF

To qualitatively evaluate whether weighting the $IPGE_{slope}$ and the PLV by the AzBio FIF

modifies their relationships with AzBio scores, in a separate set of linear regressions, we used

unweighted averages of the $IPGE_{slope}$ and the PLV in lieu of their weighted counterparts. Overall,

the result patterns and statistical significance remained the same in the unweighted version of the

linear regressions, where the relationships between the PLV and AzBio scores were statistically

significant in the noise +5 dB SNR condition ($t_{(24)} = 2.47$, $p = .021$) but not in the noise +10 dB

SNR ($t_{(24)} = 1.74$, $p = .094$) or the quiet condition ($t_{(24)} = 0.91$, $p = .374$), and no significant

relationship between the $IPGE_{slope}$ and AzBio scores was observed in any of the tested conditions

($p > .10$ in all cases).

Comparisons across AGF Fitting Methods and I/O Scales

To evaluate the potential effects of AGF fitting methods and I/O scales on the results, we

recalculated the $IPGE_{slope}$ based on AGF slope used linear fitting and the window method, both

in the linear-linear I/O scale of μV/nC. The statistical tests described above were then

reperformed using each version of the $IPGE_{slope}$ values. The result patterns and statistical

significance stayed largely the same across fitting methods and I/O scale, with a couple

exceptions. The non-significant trend of positive correlation between the $IPGE_{slope}$ and the PLV

was shared when the $IPGE_{slope}$ values were calculated using window fitting in μV/dB ($rho_{(26)} =$

0.340, $p = .077$) and linear fitting in μV/nC ($rho_{(26)} = 0.352$, $p = .067$), but became a significant

positive correlation when $IPGE_{slope}$ was calculated using window fitting in μV/nC ($rho_{(26)} =$

0.490, $p = .009$). The non-significant association between the PLV and the AzBio scores in noise

with +10 dB SNR was shared when the $IPGE_{slope}$ values were calculated using window fitting in

μV/dB ($t_{(24)} = 1.61$, $p = .120$) and in μV/nC ($t_{(24)} = 1.62$, $p = .119$), but became a significant

positive association when $IPGE_{slope}$ was calculated using linear fitting in μV/nC ($t_{(24)} = 2.35$, p

$= .028$).

546   **DISCUSSION**

This study assessed the relationships between two CN health measures, the $IPGE_{slope}$ and

the PLV, and evaluated their contributions to speech perception in quiet and noise in

postlingually deafened adult CI users. We hypothesized that the $IPGE_{slope}$ and the PLV are two

independent measures predictive of speech recognition scores in quiet and in noise, respectively.

The hypotheses were partially supported by the results showing that the $IPGE_{slope}$ and the PLV

had different relationships with electrode location and age, and that the speech perception scores

measured in noise with +5 dB SNR were positively associated with the PLV. However, contrary

to our hypothesis, we did not observe significant associations between the IPGE$_{slope}$ and speech

perception measured either in quiet or in noise. The result patterns stayed largely the same

regardless of fitting methods (window vs. linear) or I/O scales (linear-linear vs. linear

logarithmic) used in AGF slope calculation.

Peripheral Neural Survival and Synchrony

The IPGE$_{slope}$ and the PLV demonstrated different variabilities across electrode locations

and age, suggesting that they were measuring different aspects of CN health at least to some

extent. These results are aligned with the physiological process of neural degeneration of the CN

that occurs in patients with sensorineural hearing loss and are consistent with histological results

of human temporal bone studies. Detailed explanations are provided below.

The deterioration of the bipolar SGNs starts at the peripheral axon, which connects the

SGN soma to the organ of Corti in the cochlea (Xing et al. 2012). The degeneration of the

peripheral axon starts with deterioration of the myelin sheath (Xing et al. 2012) and is followed by

axonal dystrophy, segmental swelling (ballooning), and ultimately, loss of the peripheral axon (e.g.,

Chen et al. 2006; Heshmat et al. 2020; Kumar et al. 2022; Liu et al. 2015; Nadol 1990; Wu et al.

2019; Wu et al. 2023). Even after the complete loss of the peripheral axon, the soma and the central

axon of the SGN (i.e., unipolar SGN) can survive for decades (Rask-Andersen et al. 2010), and

the proportion of unipolar SGNs increases steadily with age (Wu et al. 2023). This process gives

rise to a key difference between acoustic and electric hearing. While the loss of peripheral SGN

axons likely contributes to impairment in acoustic hearing (Wu et al. 2019; Wu et al. 2020; Wu et

al. 2021), electric hearing can be achieved even without the peripheral axons, as the stimulation

can directly reach the somata and/or the central axons of SGNs (Javel & Shepherd 2000). Both

axonal dystrophy and demyelination alter many neural properties (e.g., Tasaki 1955; Waxman &

Ritchie 1993), which reduce the synchronized discharge of affected CN fibers (e.g., Gonzalez

2019; Heshmat et al. 2020; Kim et al. 2013; Tasaki 1955) as well as across the population of CN

fibers (Kandel 2002). Additionally, loss of the peripheral axon and altered membrane properties

can shift the action potential initiation site distally to the SGN soma or central axon in electrical

hearing (e.g., Hartmann et al. 1984; Javel & Shepherd 2000; Van Den Honert & Stypulkowski

1984), further reducing discharge synchronization across CN fibers (Javel & Shepherd 2000). As

explained in He et al. (2024), the PLV is influenced by temporal jitter in spike firing of individual

CN fibers as well as discharge synchronization among a population of activated CN fibers across

multiple stimulations, which is an ideal tool to assess neural synchrony in the CN in CI patients.

The results of human temporal bone studies demonstrated an age-related loss in both

SGN soma and the peripheral axon in adult listeners: Compared with the apical regions, the basal

area of the cochlea shows the greatest loss in both SGN soma and peripheral, with peripheral

axon showing a substantial greater loss than SGN soma axon (Liu et al., 2015; Makary et al.,

2011; Wu et al., 2019; Wu et al., 2023). In one CI case who was 70 years of age at the time of

death, a total loss of peripheral axons for all surviving SGNs at the basal half of the cochlea was

observed. In contrast, peripheral axon counts in the apical cochlea are similar to those of age-

matched, nonimplanted listeners (O'Malley et al. 2024).

Based on the literature reported above, smaller PLVs are expected at more basal electrode

locations where substantial degeneration in the peripheral axon occurs in middle-aged and older

CI patients, which make up a majority of the patient group tested in this study. Consistent with

598 this expectation, our data showed an electrode location effect on the PLV, with significantly

599 smaller PLVs measured at the basal electrode location than apical locations. Our results did not

600 show the effect of age on the PLV. However, this result does not necessarily conflict with those

601 reported in Wu et al. (2019) and Wu et al. (2023) due to the difference in cases/participants

602 included in these two studies. Specifically, the cases reported in Wu et al. (2019) and Wu et al.

603 (2023) were non-CI listeners with acoustic hearing, while the current study focused on CI users.

604 Further studies using animal models, histological studies in human temporal bones and

605 computational modeling techniques are warranted to determine the crucial factors (e.g., age,

606 surgical trauma, etiology) for neural health of the CN in implanted cases. Finally, the lack of

607 association between the PLV and age calls for further investigations on CI users with a wide age

608 range to determine the contribution of poor neural synchrony in the CN to speech perception

609 deficits in noise experienced by many older CI users.

610   Unlike the PLV, the $IPGE_{slope}$ is negatively correlated with participant age but not

611 affected by electrode location. The significant correlation between the $IPGE_{slope}$ and age is

612 consistent with the age-related decline in SGN counts in humans reported in different studies

613 (e.g., Liu et al. 2015; Kusunoki et al. 2004; Makary et al. 2011; Wu et al. 2019; Wu et al. 2023) .

614 The lack of difference in the $IPGE_{slope}$ across electrodes is consistent with histological results of

615 some human temporal bone studies reporting similar amounts of SGN loss across cochlear

616 locations (Makary et al., 2011; Wu et al., 2023). However, substantial variations in SGN loss at

617 different cochlear locations have also been reported in the literature (e.g., Hinojosa & Marion

618 1983; O'Malley et al. 2024). To better understand and interpret our results, we compared the

619 overall trend of $IPGE_{slope}$ across electrode locations with the SGN counts of 26 implanted cases

620 reported in Seyyedi et al. (2013). This temporal bone study was selected for comparison due to

the similarities in sample size (28 vs. 26) and etiology of hearing loss (unknown, Meniere's

disease, acoustic neuroma, and trauma), CI devices from the same manufacturer (Cochlear®

Ltd.), and the same number of locations/segments (i.e., 4) along the cochlea being investigated. It

should be noted that the cases reported in Seyyedi et al. (2013) range from 42 to 92 years of age

(mean: 73.46 yrs; *SD*: 12.22 yrs) at the time of death, which is significantly older than the

participants tested in our study (Independent sample *t*-test: $t_{(52)} = 3.11$, $p = .003$). Figure 7 shows

box plots of SGN counts for the implanted ears reported in Tables 5 and 6 in Seyyedi et al.

(2013) as a function of cochlear location, along with IPGE$_{slope}$ results measured in our study.

Despite the difference in the exact cochlear location where these two measures were taken, the

overall trends of these two datasets measured at all except for the most basal location are

remarkably similar. For the results measured at the most basal location, SGN count shows a

steeper decrease compared to IPGE$_{slope}$ results. Possible factors accounting for this difference

include the older age of the cases included in Seyyedi et al. (2013) and the greater degrees of

sensorineural hearing loss in these cases due to the stricter CI candidacy criteria used decades

ago compared to our study. Both factors result in greater SNG loss at the base of the cochlea than

more apical areas. These results, along with those reported in He et al. (2020) and Yuan et al

(2022), support the idea that IPGE$_{slope}$ is indicative of SGN survival. Unfortunately, the numeric

relationship between IPGE$_{slope}$ and SGN count cannot be determined based on data shown in

Figure 7 due to different participants included in our study vs those included in Seyyedi et al.

(2013).

It should be noted that both the PLV and the IPGE$_{slope}$ are eCAP-derived measurements.

The eCAP is a near-field recording of the aggregated activity of a group of electrically

stimulated CN fibers. As a result, it is affected by firing synchrony both within and across CN

fibers as well as the number of activated CN fibers. In addition, each eCAP response used in

AGF slope calculation in this study was an average of 50 single sweeps. Its amplitude can be

affected by the variability in the peak latencies of the single sweeps, resulting in covariation

between IPGE$_{slope}$, an amplitude-based eCAP measurement, and the PLV. Therefore, the trend of

positive correlation between the PLV and IPGE$_{slope}$ values observed in this study is not

surprising. However, the differences in their associations with electrode location and age indicate

that the PLV and the IPGE$_{slope}$ reflect different aspects of CN health. Further studies using

computational modeling techniques are warranted to determine the biological underpinnings of

these measures and how different types and degrees of neural degeneration of the CN affect

these measures.

Figure 7. Box plots showing the results of the IPGE$_{slope}$ measured in this study (wide gray boxes)

and spiral ganglion neuron counts (narrow red boxes) reported in Seyyedi et al. (2013) across

four cochlear regions. Boxes show the range between the first and the third quartile of the data

values. The horizontal bars inside the boxes represent the median. The vertical whiskers show

the range of values that are within 1.5 interquartile range (IQR) from the boxes. The dots show

the 5th and 95th percentile of the data.

Peripheral Neural Synchrony Contributes to Speech Perception in Noise

We observed significant positive associations between the PLV and AzBio scores measured in noise with +5 dB SNR, suggesting that neural synchrony is important for speech perception in high levels of background noise. These results are consistent with and complement the data from our previous study on CI users (He et al. 2024; Skidmore et al. 2023). This observation also agrees with and expands the findings by Harris et al. (2021), where the CAP-derived PLV was shown to be a strong predictor of the perception of time-compressed speech and speech in noise in NH listeners. These results highlight the importance of neural synchrony in the CN in speech perception in listeners with various hearing profiles. The difference between the PLV effects on speech perception across listening environments is likely due to the heightened importance of temporal cues for speech perception in noise (Nie et al. 2006), and the lack of neural synchrony is associated with poor performance in psychophysical tasks requiring fine temporal perception (Zeng et al. 2005). The current results can also provide validation for using the eCAP-derived PLV as a measure of neural synchrony in adult CI users (He et al. 2024). Future studies could measure neural synchrony in the CN in implanted deafened animals using both the eCAP-derived PLV and traditional single-neuron recording methods (e.g., Seki & Eggermont 2003) to further evaluate the validity of using the PLV as an index for neural synchrony in the CN.

Lack of Significant Associations between the $IPGE_{slope}$ and Speech Perception

The lack of significant associations between the $IPGE_{slope}$ and speech perception in quiet does not support our hypothesis on the contribution of neural survival to speech perception. This

result is consistent with those reported by Imsiecke et al. (2021) but not consistent with the

observations in a recent study by Zamaninezhad et al. (2023). This discrepancy could be due to

some critical differences between the testing materials and methods used in these studies.

Specifically, the German matrix sentences in Zamaninezhad et al. (2023) consisted of

syntactically correct but semantically unpredictable sentences, while the AzBio corpus used in

the current study and the Hochmair-Schulz-Moser (HSM) sentence test used in Imsiecke et al.

(2021) consisted of meaningful sentences on everyday topics; the Freiburg monosyllable test in

Zamaninezhad et al. (2023) required the listeners to repeat a single word at a time, while the

AzBio and HSM tests required them to repeat a full sentence in each trial. These differences

allow the AzBio and HSM sentence tests to better simulate real-life listening situations, but also

leave room for the effect of cognitive factors such as working memory (Ingvalson et al. 2015) to

modulate the speech perception performance on top of CN health condition. Therefore, it is

possible that the association between the $IPGE_{slope}$ and speech perception, if any, has been

eclipsed by the individual differences in cognitive factors in the present study. Future research

could test both cognitive abilities and the $IPGE_{slope}$ in the same group of CI users to evaluate

their relative contributions to speech perception.

700   In addition, the $IPGE_{slope}$ values in Zamaninezhad et al. (2023) were calculated as the

difference between the AGF slopes measured with IPGs of 30 µs and 2.1 µs, but the present

study used the AGF slopes between 42 µs and 7 µs for the same calculation. It is possible that

the sensitivity of the $IPGE_{slope}$ as an index of CN survival varies with the IPG levels used for

eCAP recordings, and further investigation is warranted for optimizing the parameters in the

$IPGE_{slope}$ measurement for the purpose of representing CN health condition.

Frequency Importance Function in Speech Perception Measures

708   While not a central focus of the present study, in the models evaluating the contributing

factors to AzBio scores, we calculated the values of the $IPGE_{slope}$ and the PLV as both

unweighted averages across the tested CI electrodes and weighted averages based on the AzBio

FIF. The results were similar regardless of whether the AzBio FIF weights were applied, which

seemingly contradicts the definition of the FIF (Lee & Mendel 2017). A possible explanation is

that the FIF used in the current study was measured in NH listeners (Lee & Mendel 2017), and

thus may not be fully generalizable to CI users due to the differences in the weighting of

frequency bands in speech perception between CI users and NH listeners (Sladen & Ricketts

2015) and potentially larger individual differences in CI users than NH listeners (Mehr et al.

2001; Bosen & Chatterjee 2016). Furthermore, the weights of adjacent frequencies could differ

considerably in some FIFs (Healy et al. 2013), so that estimating the weights of CI electrodes

based on a smooth curve fitted to discrete values of the empirically measured AzBio FIF may

have limited validity, even if the overall FIF shape of CI users is similar to that of NH listeners.

Future research could develop and validate methods for measuring the FIF in CI users and test

whether weighting eCAP-derived indices by the CI-based FIF can improve their capability to

predict speech perception.

Methodological Considerations

Fitting Method and I/O Scale

One methodological factor to consider when calculating the $IPGE_{slope}$ is the fitting

method and I/O scale of the AGF slopes. Skidmore et al. (2022) demonstrated that the window

method is broadly applicable to various AGF shapes, and that the window fitting results

calculated using linear-linear scale (µV/nC) and linear-logarithmic scale (µV/dB) were highly

correlated in human listeners. To further evaluate how the I/O scale affects the calculation of

AGF slope at the level of individual electrodes, we compiled the indices of the selected windows

(1 to 8, where 1 and 8 correspond to the lower and higher ends of the dynamic range,

respectively), and the range of stimulation levels (in CL) corresponding to the selected window

of each tested electrode, when calculated both in µV/nC and in µV/dB. When averaged across all

electrodes and participants, the selected ranges in CL were similar when using µV/nC (176.1-

186.8 CL) and µV/dB (178.1-188.0 CL), while the selected window index differed by 1 window

on average (µV/nC: 5.2, µV/dB: 6.2).

For most of the electrodes, similar window indices and CL ranges were selected when

using µV/dB and µV/nC. In particular, for those with a concaved AGF shape similar to the lower

half of the S-shaped sigmoid function, a larger window index (typically 7 or 8) was selected

when using both µV/dB and µV/nC. However, the inherent differences in the units of dB (re 1

nC) and nC would result in different slope values. When the same or similar windows were

selected, the direction of the numerical difference between AGF slopes in µV/dB and µV/nC

were dependent on the stimulation level of the selected window. A qualitative observation was

that when the stimulation level of the selected window was relatively high, with an upper limit

around 175 CL or higher, the slope value calculated in µV/dB was generally larger than µV/nC.

Alternatively, when the stimulation of the selected window was relatively low, with an upper

limit around 165 CL or lower, the slope value calculated in µV/dB was generally smaller than

µV/nC.

In rare cases, the selected window indices were very different between µV/dB and µV/nC

(e.g., for S03 e15 with an IPG of 42 us, windows 8 and 1 were selected when using µV/dB and

µV/nC, respectively). Visual inspection of their AGF curves revealed that their AGF was more

linearly shaped (rather than concaved) when the amplitude was plotted against CL (a logarithm

unit). The linear unit of nC changes at a smaller step when the stimulation changes by 1 CL at

lower levels compared to higher levels, while the logarithmic unit of dB (re 1 nC) changes at

roughly equal steps across stimulation levels. Such difference would result in larger slopes in

µV/nC at lower stimulation levels compared to higher stimulation levels for AGFs with a linear

shape when plotted against CL, so that a small window index (corresponding to low stimulation

levels) would be selected when using µV/nC.

In short, the numerical difference in AGF slopes between window fitting using µV/dB

and µV/nC is influenced by both the shape of the AGF curve and the stimulation range of the

electrode. Such subtle numerical differences at the level of individual electrodes may result in

different statistical significance in data analysis. However, given the overall similarity in the

AGF slopes calculated using the window method in µV/nC and µV/dB, it is unlikely that the

general result trends are drastically different across I/O scales when using the window fitting

method, as demonstrated by the IPGE$_{slope}$ results in the current study.

PLV Analysis Applied to eCAP Recordings

The eCAP PLV analysis methodology used in this study follows that of He et al. (2024),

which was adapted from a PLV analysis approach for click-evoked acoustic compound action

potentials (CAPs) developed by Harris et al. (2021). The parameters of the analysis need to be

adapted because of the difference between the morphologies of eCAPs and click-evoked acoustic

CAPs. Schvartz-Leyzac et al. (2025) have recently expressed concerns about the suitability of

this adapted PLV method regarding the sampling rate, time window duration, and lack of

measurement of a baseline PLV before the stimulation period. However, in human CI users, the

shortest possible N1-P2 interpeak latency of the eCAP response is around 0.2 ms (for a review,

see He et al. 2017). In a Fast Fourier Transform (FFT) analysis, this corresponds to the shortest

possible fundamental period of 0.4 ms or the highest possible fundamental frequency of 2.5 kHz.

The sampling rate of Cochlear™ Nucleus® device is 20,492 Hz, and therefore the fundamental

frequency and first few harmonics of these shortest possible eCAP waveforms fall below the

Nyquist frequency of 10,246 Hz. Thus, this sampling rate, despite being the lowest among all

three major CI devices, is sufficient for eCAP recording and is not a factor that could limit the

sensitivity of PLV measures. Additionally, in human CI users, the eCAP occurs within the first

1200 μs after stimulus onset (Botros et al. 2007). Therefore, a recording window of 1600 μs is

sufficient for capturing the entire eCAP response. It is worth noting that the duration of the time

frame used in the FFT analysis, rather than the recording window of the eCAP, affects the

frequency range used in the PLV calculation. As demonstrated in He et al. (2024), the number of

time frames used to calculate the PLV, when the recording window is fixed at 1600 μs, is not a

determining factor for the PLV method. More detailed explanations are in the "Methodological

Considerations" section of He et al. (2024).

It is true that the recording window for measuring the eCAP using the neural response

telemetry function implemented in the Custom Sound EP software v6.0 (Cochlear Ltd., New

South Wales, Australia) does not allow for measurement of a baseline period before the stimulus

presentation. However, the duration of the time window can be extended to include 64 samples

(around 3,200 μs), which provides an opportunity to compare PLVs calculated for the signal

(i.e., eCAP responses) with those calculated for the baseline noise recorded following the eCAP

waveform. To fully understand the limitations of the PLV method, we used this extended time

window to record the eCAP results in nine participants tested in this study. For the traces

measured at each electrode location in these participants, PLVs were calculated for two

nonoverlapping 32-sample time windows, consisting of the first and the second half of recorded

samples. As has been pointed out previously, the eCAP in human CI users occurs within the first

1200 μs after stimulus onset (Botros et al. 2007) , which is fully captured by the first time

window. Therefore, the PLV caculated for the first time window was defined as the "signal"

PLV, and the PLV calculated for the second time window represent the PLV of the noise floor.

The segments of the recorded traces corresponding to the second time window at four electrode

locations in the left ear of S06 and the heat maps of time-frequency-specific PLVs are shown in

Supplemental Digital Content 1. The segments of these traces corresponding to the first time

window and the PLV heat maps are shown in Figure 4. The PLVs measured for the eCAP signal

(i.e., the first time window) and the noise floor (i.e., the second time window) at each electrode

location measured in these nine participants are shown in Supplemental Digital Content 2. The

results of a LMM showed that PLVs calcaulted for the signal were significantly larger than those

measured for the noise floor ($F_{(1, 52.14)}$ = 106.60, $p < .001$), after controlling for electrode location

and individual differences.

Potential Study Limitations

This study has five potential limitations. First, identifying the best parameter and method

to quantify the IPG effects is an ongoing research topic. In this study, the IPGE$_{slope}$ was selected

to quantify the IPG effect based on the current literature. Future studies may suggest other

parameters or quantification methods as preferred. If this is the case, we will reanalyze these data

using the preferred parameter and/or method. Second, the PLV and the IPGE$_{slope}$ are eCAP-

derived measurements. As a result, the possibility that these measures are affected by non-neural

factors cannot be completely excluded. Third, only the association between the PLV and

IPGE$_{slope}$ was assessed and reported in this study. The potential associations between the PLV

and other parameters that have been used to quantify the IPG effects in published studies are not

reported because they are beyond the scope of this study. These associations will be reported in

another paper in the near future. Fourth, the present study is that only 28 post-lingually deafened

adult participants were included in the study. Most of them were within the age range of 50-85

yrs and had generally good speech perception outcomes. Therefore, the variance in speech

perception scores explained by the PLV or the IPGE$_{slope}$ may not represent the variance

explained in the entire CI patient population. Further studies in CI users with varied speech

perception outcomes are warranted to assess the generalizability of the current observations.

Finally, speech perception was measured only using AzBio sentences, which have high

ecological validity but prone to the effects of central auditory processing and cognitive factors.

In this study, only one ear was tested for each participant, including those who are bilateral CI

users. Therefore, in the current dataset, it is not possible to control for potential individual

differences in central auditory processing and cognitive abilities using within-participant

between-ear comparisons. Future research can test both ears of bilateral CI users and/or add

measurements for central auditory processing and cognitive abilities to pinpoint the crucial

factors contributing to speech perception in adult CI users. Finally, the 10-talker babble was used

as the competing background noise to assess speech perception performance. It does not fully

capture the challenge of understanding speech in more complex environments.

844   **CONCLUSIONS**

The $IPGE_{slope}$ and the PLV are two eCAP-derived indices for different aspects of CN

health. The significant positive associations between the PLV and AzBio scores measured in

noise with +5 dB SNR suggest that neural synchrony is important for speech perception in

challenging listening environments. The lack of association between age and the PLV calls for

further investigation to determine whether reduced neural synchrony in the CN is unlikely the

primary factor accounting for excessive speech perception deficits in older CI users. Future

studies can also investigate the contribution of cognitive factors to speech perception and how

they interact with the effects of CN health status, as well as use animal models or computational

modeling techniques to better understand the biological underpinnings of the $IPGE_{slope}$ and the

PLV.

REFERENCES

Bates, D., Mächler, M., Bolker, B., et al. (2015). Fitting linear mixed-effects models using lme4.

J. Stat. Softw., 67.

Bosen, A.K., Chatterjee, M. (2016). Band importance functions of listeners with cochlear

implants using clinical maps. *J. Acoust. Soc. Am.*, 140, 3718–3727.

Botros, A., Van Dijk, B., Killian, M. (2007). AutoNRT^{TM}: An automated system that measures

ECAP thresholds with the Nucleus® Freedom^{TM} cochlear implant via machine

intelligence. *Artif. Intell. Med.*, 40, 15–28.

Brochier, T., McKay, C.M., Carlyon, R.P. (2021). Interpreting the effect of stimulus parameters

on the electrically evoked compound action potential and on neural health estimates. *J.*

Assoc. Res. Otolaryngol., 22, 81–94.

Brown, C.J., Abbas, P.J., Gantz, B. (1990). Electrically evoked whole-nerve action potentials:

Data from human cochlear implant users. *J. Acoust. Soc. Am.*, 88, 1385–1391.

Chen, J.X., Nourmahnad, A., O'Malley, J., et al. (2020). Otopathology in CHARGE syndrome.

Laryngoscope Investig. Otolaryngol., 5, 157–162.

Chen, M.A., Webster, P., Yang, E., et al. (2006). Presbycusic neuritic degeneration within the

osseous spiral lamina. *Otol. Neurotol.*, 27, 316–322.

Da Costa Monsanto, R., Knoll, R.M., De Oliveira Penido, N., et al. (2022). Otopathologic

Abnormalities in CHARGE Syndrome. *Otolaryngol. Neck Surg.*, 166, 363–372.

Delorme, A., Makeig, S. (2004). EEGLAB: an open source toolbox for analysis of single-trial

EEG dynamics including independent component analysis. *J. Neurosci. Methods*, 134, 9–

21.

Dong, Y., Briaire, J.J., Biesheuvel, J.D., et al. (2020). Unravelling the temporal properties of

human eCAPs through an iterative deconvolution model. *Hear. Res.*, 395, 108037.

Dong, Y., Briaire, J.J., Stronks, H.C., et al. (2023). Speech perception performance in cochlear

implant recipients correlates to the number and synchrony of excited auditory nerve

fibers derived from electrically evoked compound action potentials. *Ear Hear.*, 44, 276–

286.

Glueckert, R., Rask-Andersen, H., Sergi, C., et al. (2010). Histology and synchrotron radiation-

based microtomography of the inner ear in a molecularly confirmed case of CHARGE

syndrome. *Am. J. Med. Genet. A.*, 152A, 665–673.

Gonzalez-Gonzalez, S.G. (2019). Myelination of the auditory nerve: Functions and pathology.

Sci. J. Res. Rev., 1.

Haginomori, S.-I., Miura, M., Sando, I., et al. (2002). Temporal bone histopathology in

CHARGE association. *Ann. Otol. Rhinol. Laryngol.*, 111, 397–401.

Harris, K.C., Ahlstrom, J.B., Dias, J.W., et al. (2021). Neural presbyacusis in humans inferred

from age-related differences in auditory nerve function and structure. *J. Neurosci.*, 41,

10293–10304.

Hartmann, R., Topp, G., Klinke, R. (1984). Discharge patterns of cat primary auditory fibers

with electrical stimulation of the cochlea. *Hear. Res.*, 13, 47–62.

He, S., Shahsavarani, B.S., McFayden, T.C., et al. (2018). Responsiveness of the electrically

stimulated cochlear nerve in children with cochlear nerve deficiency. *Ear Hear.*, 39, 238–

250.

He, S., Skidmore, J., Bruce, I.C., et al. (2024). Peripheral neural synchrony in postlingually

deafened adult cochlear implant users. *Ear Hear.*, 45, 1125–1137.

He, S., Skidmore, J., Koch, B., et al. (2023a). Relationships between the auditory nerve

sensitivity to amplitude modulation, perceptual amplitude modulation rate discrimination

sensitivity, and speech perception performance in postlingually deafened adult cochlear

implant users. *Ear Hear.*, 44, 371–384.

He, S., Teagle, H.F.B., Buchman, C.A. (2017). The electrically evoked compound action

potential: From laboratory to clinic. *Front. Neurosci.*, 11, 339.

He, S., Xu, L., Skidmore, J., et al. (2020). The effect of interphase gap on neural response of the

electrically stimulated cochlear nerve in children with cochlear nerve deficiency and

children with normal-sized cochlear nerves. *Ear Hear.*, 41, 918–934.

He, S., Yuan, Y., Skidmore, J. (2023b). Relationships between the auditory nerve's ability to

recover from neural adaptation, cortical encoding of and perceptual sensitivity to within-

channel temporal gaps in postlingually deafened adult cochlear implant users. *Ear Hear.*,

44, 1202–1211.

Healy, E.W., Yoho, S.E., Apoux, F. (2013). Band importance for sentences and words

reexamined. *J. Acoust. Soc. Am.*, 133, 463–473.

Heshmat, A., Sajedi, S., Johnson Chacko, L., et al. (2020). Dendritic degeneration of human

auditory nerve fibers and its impact on the spiking pattern under regular conditions and

during cochlear implant stimulation. *Front. Neurosci.*, 14, 599868.

Hinojosa, R., Marion, M. (1983). Histopathology of profound sensorineural deafness. *Ann. N. Y.*

Acad. Sci., 405, 459–484.

Hughes, M.L., Castioni, E.E., Goehring, J.L., et al. (2012). Temporal response properties of the

auditory nerve: Data from human cochlear-implant recipients. *Hear. Res.*, 285, 46–57.

Imsiecke, M., Büchner, A., Lenarz, T., et al. (2021). Amplitude growth functions of auditory

nerve responses to electric pulse stimulation with varied interphase gaps in cochlear

implant users with ipsilateral residual hearing. *Trends Hear.*, 25, 23312165211014137.

Ingvalson, E.M., Dhar, S., Wong, P.C.M., et al. (2015). Working memory training to improve

speech perception in noise across languages. *J. Acoust. Soc. Am.*, 137, 3477–3486.

Jahn, K.N., Arenberg, J.G. (2020a). Electrophysiological estimates of the electrode–neuron

interface differ between younger and older listeners with cochlear implants. *Ear Hear.*,

41, 948–960.

Jahn, K.N., Arenberg, J.G. (2020b). Identifying cochlear implant channels with relatively poor

electrode-neuron interfaces using the electrically evoked compound action potential. *Ear*

Hear., 41, 961–973.

JASP Team (2024). JASP. Available at: https://jasp-stats.org.

Javel, E., Shepherd, R.K. (2000). Electrical stimulation of the auditory nerve: III. Response

initiation sites and temporal fine structure. *Hear. Res.*, 140, 45–76.

Kandel, E.R. (2002). Disease of the motor unit. In *E. R. Kandel, J. H. Schwartz, T. M. Jessell*

(Eds.), Principles of Neural Science 4th ed. (*pp. 695–703*). McGraw-Hill, Health

Professions Division.

Kim, J.H., Renden, R., Von Gersdorff, H. (2013). Dysmyelination of auditory afferent axons

increases the jitter of action potential timing during high-frequency firing. *J. Neurosci.*,

33, 9402–9407.

Kim, J.-R., Abbas, P.J., Brown, C.J., et al. (2010). The relationship between electrically evoked

compound action potential and speech perception: A study in cochlear implant users with

short electrode array. *Otol. Neurotol.*, 31, 1041–1048.

Kipping, D., Zhang, Y., Nogueira, W. (2024). A computational model of the electrically or

acoustically evoked compound action potential in cochlear implant users with residual

hearing. *IEEE Trans. Biomed. Eng.*, 71, 3192–3203.

Kraus, N., Bradlow, A.R., Cheatham, M.A., et al. (2000). Consequences of neural asynchrony: A

case of auditory neuropathy. *J. Assoc. Res. Otolaryngol.*, 1, 33–45.

Kumar, P., Sharma, S., Kaur, C., et al. (2022). The ultrastructural study of human cochlear nerve

at different ages. *Hear. Res.*, 416, 108443.

Kusunoki, T., Cureoglu, S., Schachern, P.A., et al. (2004). Age-related histopathologic changes

in the human cochlea: A temporal bone study. *Otolaryngol. Neck Surg.*, 131, 897–903.

Kuznetsova, A., Brockhoff, P.B., Christensen, R.H.B. (2017). lmerTest package: Tests in linear

mixed effects models. *J. Stat. Softw.*, 82.

Lee, S., Mendel, L.L. (2017). Derivation of frequency importance functions for the AzBio

sentences. *J. Acoust. Soc. Am.*, 142, 3416–3427.

Lenth, R.V. (2024). emmeans: Estimated marginal means, aka least-squares means. Available at:

https://cran.r-project.org/web/packages/emmeans/index.html.

Liu, C., Glowatzki, E., Fuchs, P.A. (2015). Unmyelinated type II afferent neurons report

cochlear damage. *Proc. Natl. Acad. Sci.*, 112, 14723–14727.

Makary, C.A., Shin, J., Kujawa, S.G., et al. (2011). Age-related primary cochlear neuronal

degeneration in human temporal bones. *J. Assoc. Res. Otolaryngol.*, 12, 711–717.

Mehr, M.A., Turner, C.W., Parkinson, A. (2001). Channel weights for speech recognition in

cochlear implant users. *J. Acoust. Soc. Am.*, 109, 359–366.

Miura, M., Hirsch, B.E., Sando, I., et al. (2002). Analysis of spiral ganglion cell populations in

children with normal and pathological ears. *Ann. Otol. Rhinol. Laryngol.*, 111, 1059–

1065.

Nadol, J. (1997). Patterns of neural degeneration in the human cochlea and auditory nerve:

Implications for cochlear implantation. *Otolaryngol. Head Neck Surg.*, 117, 220–228.

Nadol, J.B. (1990). Degeneration of cochlear neurons as seen in the spiral ganglion of man.

Hear. Res., 49, 141–154.

Nelson, E.G., Hinojosa, R. (2001). Aplasia of the cochlear nerve: A temporal bone study. *Otol.*

Neurotol., 22, 790–795.

Nie, K., Barco, A., Zeng, F.-G. (2006). Spectral and temporal cues in cochlear implant speech

perception. *Ear Hear.*, 27, 208–217.

O'Malley, J.T., Wu, P., Kaur, C., et al. (2024). Delayed hearing loss after cochlear implantation:

Re-evaluating the role of hair cell degeneration. *Hear. Res.*, 447, 109024.

Pfingst, B.E., Hughes, A.P., Colesa, D.J., et al. (2015). Insertion trauma and recovery of function

after cochlear implantation: Evidence from objective functional measures. *Hear. Res.*,

330, 98–105.

Prado-Guitierrez, P., Fewster, L.M., Heasman, J.M., et al. (2006). Effect of interphase gap and

pulse duration on electrically evoked potentials is correlated with auditory nerve survival.

Hear. Res., 215, 47–55.

R Core Team (2024). R: A language and environment for statistical computing. Available at:

https://www.R-project.org/.

Ramekers, D., Versnel, H., Strahl, S.B., et al. (2014). Auditory-nerve responses to varied inter-

phase gap and phase duration of the electric pulse stimulus as predictors for neuronal

degeneration. *J. Assoc. Res. Otolaryngol.*, 15, 187–202.

Rask-Andersen, H., Liu, W., Linthicum, F. (2010). Ganglion cell and 'dendrite' populations in

electric acoustic stimulation ears. In P. Van De Heyning & A. Kleine Punte, eds.

Advances in Oto-Rhino-Laryngology. (*pp. 14–27*). S. Karger AG.

Roberts, D.S., Lin, H.W., Herrmann, B.S., et al. (2013). Differential cochlear implant outcomes

in older adults. *The Laryngoscope*, 123, 1952–1956.

Schvartz-Leyzac, K.C., Colesa, D.J., Buswinka, C.J., et al. (2019). Changes over time in the

electrically evoked compound action potential (ECAP) interphase gap (IPG) effect

following cochlear implantation in Guinea pigs. *Hear. Res.*, 383, 107809.

Schvartz-Leyzac, K.C., Colesa, D.J., Buswinka, C.J., et al. (2020a). How electrically evoked

compound action potentials in chronically implanted guinea pigs relate to auditory nerve

health and electrode impedance. *J. Acoust. Soc. Am.*, 148, 3900–3912.

Schvartz-Leyzac, K.C., Holden, T.A., Zwolan, T.A., et al. (2020b). Effects of electrode location

on estimates of neural health in humans with cochlear implants. *J. Assoc. Res.*

Otolaryngol., 21, 259–275.

Schvartz-Leyzac, K.C., McClaskey, C.M., Dias, J.W., et al. (2025). Contributions of auditory

nerve density and synchrony to speech understanding in older cochlear implant users. *J.*

Assoc. Res. Otolaryngol.

Schvartz-Leyzac, K.C., Pfingst, B.E. (2016). Across-site patterns of electrically evoked

compound action potential amplitude-growth functions in multichannel cochlear implant

recipients and the effects of the interphase gap. *Hear. Res.*, 341, 50–65.

Schvartz-Leyzac, K.C., Pfingst, B.E. (2018). Assessing the relationship between the electrically evoked compound action potential and speech recognition abilities in bilateral cochlear implant recipients. *Ear Hear.*, 39, 344–358.

Seki, S., Eggermont, J.J. (2003). Changes in spontaneous firing rate and neural synchrony in cat primary auditory cortex after localized tone-induced hearing loss. *Hear. Res.*, 180, 28–38.

Seyyedi, M., Eddington, D.K., Nadol, J.B. (2013). Effect of monopolar and bipolar electric stimulation on survival and size of human spiral ganglion cells as studied by postmortem histopathology. *Hear. Res.*, 302, 9–16.

Seyyedi, M., Viana, L.M., Nadol, J.B. (2014). Within-subject comparison of word recognition and spiral ganglion cell count in bilateral cochlear implant recipients. *Otol. Neurotol.*, 35, 1446–1450.

Skidmore, J., Oleson, J.J., Yuan, Y., et al. (2023). The relationship between cochlear implant speech perception outcomes and electrophysiological measures of the electrically evoked compound action potential. *Ear Hear.*, 44, 1485–1497.

Skidmore, J., Ramekers, D., Colesa, D.J., et al. (2022). A broadly applicable method for characterizing the slope of the electrically evoked compound action potential amplitude growth function. *Ear Hear.*, 43, 150–164.

Sladen, D.P., Ricketts, Todd.A. (2015). Frequency importance functions in quiet and noise for adults with cochlear implants. *Am. J. Audiol.*, 24, 477–486.

Sladen, D.P., Zappler, A. (2015). Older and younger adult cochlear implant users: speech recognition in quiet and noise, quality of life, and music perception. *Am. J. Audiol.*, 24, 31–39.

Spahr, A.J., Dorman, M.F., Litvak, L.M., et al. (2012). Development and validation of the AzBio

sentence lists. *Ear Hear.*, 33, 112–117.

Starr, A., Zeng, F.G., Michalewski, H.J., et al. (2008). Perspectives on auditory neuropathy:

disorders of inner hair cell, auditory nerve, and their synapse. In *The Senses: A*

Comprehensive Reference. (*pp. 397–412*). Elsevier.

Takanen, M., Strahl, S., Schwarz, K. (2024). Insights into electrophysiological metrics of

cochlear health in cochlear implant users using a computational model. *J. Assoc. Res.*

Otolaryngol., 25, 63–78.

Tasaki, I. (1955). New measurements of the capacity and the resistance of the myelin sheath and

the nodal membrane of the isolated frog nerve fiber. *Am. J. Physiol.*, 181, 639–650.

Tejani, V.D., Abbas, P.J., Brown, C.J. (2017). Relationship between peripheral and

psychophysical measures of amplitude modulation detection in cochlear implant users.

Ear Hear., 38, e268–e284.

Van Den Honert, C., Stypulkowski, P.H. (1984). Physiological properties of the electrically

stimulated auditory nerve. II. Single fiber recordings. *Hear. Res.*, 14, 225–243.

Waxman, S.G., Ritchie, J.M. (1993). Molecular dissection of the myelinated axon. *Ann. Neurol.*,

33, 121–136.

Won, J.H., Humphrey, E.L., Yeager, K.R., et al. (2014). Relationship among the physiologic

channel interactions, spectral-ripple discrimination, and vowel identification in cochlear

implant users. *J. Acoust. Soc. Am.*, 136, 2714–2725.

Wright, C.G., Meyerhoff, W.L., Brown, O.E., et al. (1986). Auditory and temporal bone

abnormalities in CHARGE association. *Ann. Otol. Rhinol. Laryngol.*, 95, 480–486.

Wu, P., O'Malley, J.T., De Gruttola, V., et al. (2020). Age-related hearing loss is dominated by

damage to inner ear sensory cells, not the cellular battery that powers them. *J. Neurosci.*,

40, 6357–6366.

Wu, P., O'Malley, J.T., Liberman, M.C. (2023). Neural degeneration in normal-aging human

cochleas: Machine-learning counts and 3D mapping in archival sections. *J. Assoc. Res.*

Otolaryngol., 24, 499–511.

Wu, P.Z., Liberman, L.D., Bennett, K., et al. (2019). Primary neural degeneration in the human

cochlea: Evidence for hidden hearing loss in the aging ear. *Neuroscience*, 407, 8–20.

Wu, P.-Z., O'Malley, J.T., De Gruttola, V., et al. (2021). Primary neural degeneration in noise-

exposed human cochleas: Correlations with outer hair cell loss and word-discrimination

scores. *J. Neurosci.*, 41, 4439–4447.

Xie, Z., Gaskins, C.R., Shader, M.J., et al. (2019). Age-related temporal processing deficits in

word segments in adult cochlear-implant users. *Trends Hear.*, 23, 2331216519886688.

Xing, Y., Samuvel, D.J., Stevens, S.M., et al. (2012). Age-related changes of myelin basic

protein in mouse and human auditory nerve. *PLoS ONE*, 7, e34500.

Yuan, Y., Skidmore, J., He, S. (2022). Interpreting the interphase gap effect on the electrically

evoked compound action potential. *JASA Express Lett.*, 2, 027201.

Zaltz, Y., Bugannim, Y., Zechoval, D., et al. (2020). Listening in noise remains a significant

challenge for cochlear implant users: evidence from early deafened and those with

progressive hearing loss compared to peers with normal hearing. *J. Clin. Med.*, 9, 1381.

Zamaninezhad, L., Mert, B., Benav, H., et al. (2023). Factors influencing the relationship

between cochlear health measures and speech recognition in cochlear implant users.

Front. Integr. Neurosci., 17, 1125712.

Zeng, F.-G. (2004). Trends in cochlear implants. *Trends Amplif.*, 8, 1–34.

Zeng, F.-G., Kong, Y.-Y., Michalewski, H.J., et al. (2005). Perceptual consequences of disrupted

auditory nerve activity. *J. Neurophysiol.*, 93, 3050–3063.

Zhang, Y., Kipping, D., Nogueira, W. (2024). Evaluating electrophysiological and behavioral

measures of neural health in cochlear implant users: a computational simulation study.

Available at: https://arxiv.org/abs/2409.14767 [Accessed March 3, 2025].

Supplemental Digital Content 1.

1084

1085

1086   Supplemental Digital Content 2.

1087