Original Paper
Jamie Nam, MS, MBA[1], Grace C Bellinger, PhD, MPH[1], Junyao Li, BS[2], Margaret A French, PT,
DPT, PhD[3], Ryan T Roemmich, PhD[1,4]*
[1]Department of Physical Medicine and Rehabilitation, Johns Hopkins University School of
Medicine, Baltimore, MD 21205
[2]Northwestern University Feinberg School of Medicine, Chicago, IL 60611
[3]Department of Physical Therapy and Athletic Training, University of Utah, Salt Lake City, UT
84108
[4]Center for Movement Studies, Kennedy Krieger Institute, Baltimore, MD 21205
*corresponding author
11
12   Twitter: @GraceCBellinger (GCB), @MaggieFrenchDPT (MAF), @RyanRoemmich (RTR)
13
Contact information:
Room 240
Kennedy Krieger Institute
716 North Broadway
Baltimore, MD 21205
rroemmi1@jhmi.edu
20
21   Abstract word count: 398   Text word count: 3642
22   Number of figures: 4   Number of tables: 4

Self-reported acceptance of a wearable activity monitor in persons

with stroke

Abstract
Background: Wearable activity monitors offer clinicians and researchers accessible,
scalable, and cost-effective tools for continuous remote monitoring of functional status.

These technologies can complement traditional clinical outcome measures by providing
detailed, minute-by-minute remotely collected data on a wide array of biometrics that
include, as examples, physical activity and heart rate. There is significant potential for the
use of these devices in rehabilitation after stroke if individuals will wear and use the
devices; however, the acceptance of these devices by persons with stroke is not well
understood.

Objective: In this study, we investigated the participant-reported acceptance of a
commercially available, wrist-worn wearable activity monitor (the Fitbit Inspire 2) for
remote monitoring of physical activity and heart rate in persons with stroke. We also
assessed relationships between reported acceptance and adherence to wearing the device.
Methods: Sixty-five participants with stroke wore a Fitbit Inspire 2 for three months, at
which point we assessed the acceptance of wearing the device using the Technology
Acceptance Questionnaire (TAQ) inclusive of its seven dimensions (Perceived Usefulness
(PU), Perceived Ease of Use (PEOU), Equipment Characteristics (EC), Privacy Concern (PC),
Perceived Risk (PR), Facilitating Conditions (FC), and Subjective Norm (SN). We then
performed Spearman's correlations to assess relationships between acceptance and
adherence to device wear, which we calculated as both the percentage of daily wear time
and percentage of valid days the device was worn during the three weeks preceding TAQ
administration.
Results: Most participants reported generally agreeable responses with high overall total
TAQ scores across all seven dimensions that indicated strong acceptance of the device;
"Agree" was the median response to 29 of the 31 TAQ statements. Participants
generally found the device beneficial for their health, efficient for monitoring, easy to use
and don/doff, and unintrusive to daily life. However, participant responses on the TAQ did
not show significant positive correlations with measures of actual device wear time (all
p>0.05).
Conclusions: This study demonstrates generally high self-reported acceptance of the Fitbit
Inspire 2 among persons with stroke. Participants reported general agreement across all
seven TAQ dimensions with minimal concerns interpreted as being directly relatable to
post-stroke motor impairment (e.g., donning and doffing devices, using the device
independently). However, the high self-reported acceptance scores did not correlate
positively with measures of real-world device wear. Accordingly, it should not be assumed
that persons with stroke will adhere to wearing these devices simply because they report
high acceptability.
76
77   **Key words:** stroke, wearables, mobility, technology acceptance, physical activity, remote
78   monitoring, rehabilitation

80 Introduction

81 Background

Wearable devices have the potential to advance how clinicians and researchers measure
functional status by offering accessible, scalable, and cost-effective remote monitoring
tools.[1,2] Traditional outcome measures provide only a discrete snapshot of an
individual's functional status;[3–5] emerging wearable technologies have the potential to
address this issue by providing minute-by-minute, longitudinal data on physical activity,
heart rate, and many other biometrics from daily life that extend beyond the clinical or
laboratory setting. The use of wearable devices for remote monitoring could offer
additional observational data on functional status measured directly in an individual's daily
environment.
Notably, wearable devices could provide valuable insight into recovery following
neurologic damage (e.g., after stroke) by enabling granular, longitudinal assessment of
activity, one of the components of the World Health Organization International
Classification of Functioning, Disability, and Health model.[6] Approximately 80% of
persons with stroke experience some form of motor impairment, and around 50% continue
to have significant functional limitations six months post-stroke.[7–9] These limitations
often translate into reduced daily activity, as persons with stroke generally walk
approximately half as many steps each day as individuals without stroke.[10] Wearable
devices provide an opportunity to monitor these functional changes remotely with insights
on daily physical activity. However, the ability to perform remote monitoring after stroke
using wearable devices is dependent on how well persons with stroke accept these devices.

102 Prior Work

Wearable activity monitors like Fitbit devices generally show high acceptance among
healthy individuals[11] and a range of patient groups, including older adults with cognitive
and motor impairment.[12,13] For older adults with cognitive impairment, Fitbit-based
interventions may improve motivation for physical activity and sleep, but success is
dependent on interfaces that are easy to use,[14] reduce cognitive load,[15] and are
beneficial to the user.[14] In studies of persons with minimal motor impairment, the lack of
sustained Fitbit usage and challenges to acceptance have largely been attributed to
technical issues – difficulties with empty batteries, broken devices, or lost devices–[16]
rather than usability concerns. For example, users with multiple sclerosis reported
frustration when syncing data between devices.[17]
113
To measure acceptance of wearable technologies across diverse clinical populations,
researchers have extensively used the Technology Acceptance Model (TAM) due to its
simplicity and strong empirical support[18,19], with the Technology Acceptance
Questionnaire (TAQ)[13] serving as an extension of this framework with additional factors
related to user acceptance. Numerous studies have validated this framework across many
contexts, demonstrating its predictive power in understanding technology adoption
behaviors.[18–22] While wearable activity monitors have demonstrated varying levels of
acceptance across the general population, their perceived usability and effectiveness in
individuals with stroke remain less explored, prompting our use of the TAQ to assess their
suitability in this specific patient group.[23]

Goal of This Study

Here, we aimed to assess the acceptance of wearable devices in persons with stroke. We also examined the relationship between acceptance and adherence to wearing the Fitbit device. We studied Fitbit devices specifically because these have been used extensively for remote monitoring of step count, heart rate, and energy expenditure – among other metrics – in many populations.[24] We hypothesized that 1) acceptance of the Fitbit devices would be variable across persons with stroke but would generally indicate that these technologies are acceptable, useful, and easy to use, and 2) acceptance would be significantly associated with real-world adherence to wearing the Fitbit device.

Methods

Recruitment

We recruited persons with stroke from the outpatient rehabilitation clinics at Johns Hopkins Hospital through clinician referrals through MyChart messages. The inclusion criteria for this study were: age 18 years and older; history of stroke (as confirmed by ICD-10 codes); ownership of a smartphone and in-home Wi-Fi to access the internet; walking as a primary form of mobility (with assistive devices allowed); and ability and willingness to install the Fitbit mobile application on their smartphone device. All participants provided oral consent, as approved by the Institutional Review Board at the Johns Hopkins University School of Medicine (IRB00247292).

After obtaining consent, the study team mailed a Fitbit Inspire 2 (Fitbit Inc., San Francisco, CA, USA) to participants and asked them to wear it for one year as part of a larger study.[25] This study focuses on a sub-analysis of the first three months of device wear, during which the Technology Acceptance Questionnaire (TAQ) was administered to assess participants' experiences and perceptions of the device at the three-month time point. As an incentive, we allowed participants to keep the Fitbit following completion of the study.

Study Instructions

We instructed participants to wear the Fitbit on their non-paretic (i.e., less impaired) wrist throughout the entire day, including during sleep; if participants had difficulty donning the device and lacked available assistance, we permitted them to wear it on the paretic wrist. We documented the paretic side and the wrist on which the Fitbit was worn. Then, we guided the participants over the phone to set up the device and install the Fitbit application on their smartphone. We instructed participants to remove the device only when showering or charging it. We also asked them to synchronize the device at least once per day using the Fitbit smartphone application. Data from the Fitbit were extracted by a custom-built application that is described elsewhere.[25,26] The study team incorporated notifications and reminders to assist with adherence to wearing the Fitbit, as detailed in our previous work.[25]

After a participant was enrolled for three months, the study team attempted to contact the participant up to three times to administer the TAQ, our metric of acceptance. Each contact attempt was documented in REDCap, which included the date and time of the call, the outcome of the attempt (e.g., reached, voicemail, no answer), the

participant's response to the assessment, and any notes or follow-up actions required. If
the participant was reached, the team member administered the questionnaire verbally.
However, if the participant was not reached after three attempts, no further attempts were
made to administer the TAQ.

Technology Acceptance Questionnaire (TAQ)
We used the TAQ – which includes dimensions for Perceived Usefulness (PU), Perceived
Ease of Use (PEOU), Equipment Characteristics (EC), Privacy Concern (PC), Perceived Risk
(PR), Facilitating Conditions (FC), and Subjective Norm (SN) – as established by Puri et
al.[13] The full questionnaire is included as Multimedia Appendix 1 in Puri et al.[13] The
TAQ consisted of 31 statements rated on a 5-point Likert scale where participants
indicated their levels of agreement or disagreement with each statement. These 31
statements covered dimensions of perceived usefulness (PU; n=5 statements), perceived
ease of use (PEOU, n=7 statements), facilitating conditions (FC; n=2 statements), subjective
norm (SN; n=3 statements), equipment characteristics (EC; n=8 statements), privacy
concern (PC; n=3 statements), and perceived risk (PR; n=3 statements). We summed the
scores within each dimension ("Strongly Disagree" = 1 point, "Disagree" = 2 points,
"Neutral" = 3 points, "Agree" = 4 points, "Strongly Agree" = 5 points). We note that, to
ensure the reliability and validity of the measure, certain statements in the TAQ were
negatively framed to mitigate acquiescence bias.[27] As is standard, we minimized possible
response biases where higher numerical values represent lower agreement by applying a
reverse coding procedure to these statements: each affected score was standardized by
subtracting it from 6 for statements 2, 10, 17, 25, and 27. This was done prior to summing
the scores within each dimension. We then calculated the total TAQ score by summing the
scores from all items across all dimensions. In addition to the 31 statements, there are also
six multiple choice questions addressing various aspects of device use that are not assigned
to any domain.

Fitbit Adherence
To assess adherence, we analyzed Fitbit data from the 21 days immediately preceding
completion of the TAQ. This timeframe was selected because the TAQ items specifically
reference user perceptions of their Fitbit over the prior three weeks. We identified Fitbit
wear minutes as any minute with a non-zero heart rate through the accelerometry
package[28] in R.[29] We calculated two primary adherence metrics. First, we calculated
the percentage of average daily wear time as the percentage of time the device was worn
each day by dividing the number of wear minutes by the total maximum amount of minutes
in a day (1440) and multiplied by 100. These daily percentages were then averaged across
the 21-day period. Second, we calculated percentage of valid wear days the number of days
within the three-week window with at least 75% of wear time (i.e., 1080 or more minutes).
The number of valid days was then divided by 21 (the total days in the assessment
window) and multiplied by 100 to determine the percentage of valid days.

Statistical Analysis
We calculated descriptive statistics for the 37 individual items of the TAQ (inclusive of the
31 Likert scale statements and 6 additional multiple-choice questions), the summed scores
on each of the seven dimensions from the TAQ, and the summed total score of the TAQ. We

also report descriptive statistics for the metrics of adherence. To assess relationships
between Fitbit adherence and TAQ responses, we used Spearman's correlations (due to the
ordinal nature of the TAQ responses). We performed all statistical analyses using R version
4.4.1[29] with α=0.05.
214

RESULTS

Participants

Of the 108 persons with stroke that we contacted, 98 participants enrolled in the study (the
remaining 10 individuals either did not meet the inclusion criteria or declined to
participate). Among the enrolled participants, four were lost to follow-up (i.e., they could
not be reached despite repeated attempts during the three-month study period) and eight
withdrew voluntarily (i.e., they chose to discontinue their involvement after enrolling).
Furthermore, 21 participants did not complete the TAQ three months post-enrollment. As a
result, we included 65 participants with stroke in the final analysis. We include a CONSORT
diagram in Figure 1 and summarized study and excluded participant characteristics in
Table 1.
226

Self-reported Acceptance of Wearable Devices in Participants with Stroke

We show percentages of responses (Strongly Disagree, Disagree, Neutral, Agree, Strongly
Agree) to each of the 31 TAQ statements rated by Likert scales organized by dimension
(Figure 2). The five statements with the highest proportions of "Agree" or "Strongly Agree"
responses were (in order; 2a, 2b, and so forth indicates that there were multiple
statements with the same proportions of responses):
1) Statement 1: "I think that monitoring my activity and health 24 hours a day, 7 days a
234   week, can be a good thing."
2a) Statement 7: "I was able to perform my daily tasks as usual while wearing the
device."
2b) Statement 11: "The battery life of the device meets my expectations."
4) Statement 18: "I was able to put the device on in a reasonable amount of time."
5) Statement 20: "I am comfortable with my health data being shared with equipment
manufacturers as long as it is shared anonymously."
241

The five statements with the lowest proportions of "Agree" or "Strongly Agree" responses
were (in order):
1a) Statement 25: "Wearing the device caused me to have joint pain."
1b) Statement 27: "I was embarrassed to wear the device in front of family members."
3a) Statement 2: "I was afraid that the device would discover a major health issue."
3b) Statement 17: "I was concerned that the device is not securely attached to me."
3c) Statement 26: "I was able to shower or bathe normally while wearing the device."
As statements 1a-3b are reverse-coded; thus, a lower proportion of "Agree" or "Strongly
Agree" responses to these statements indicates a more positive perception of the device.
With the exceptions of the five statements immediately above as well as statements 10 ("I
experience skin irritations while wearing the device") and 15 ("I find the display of the

device easy to read outdoors"), a majority of participants responded either "Agree" or
"Strongly Agree" to each of the remaining 24 statements. We also note that responses to
statement 26 were likely influenced by our instructions to the participants to remove the
device while bathing or showering.
257
We show group means, standard deviations, medians, and interquartile ranges of these
scores for each of the seven dimensions in Table 2 (alongside the maximum and minimum
possible scores for each dimension) and for each individual TAQ statement in Table 3.
Medians of the summed scores for each dimension ranged from 71% in PEOU (median=25,
maximum possible score=35) to 80% in all other dimensions (Table 2). Furthermore, all
dimensions – with the exception of PEOU – showed median scores of 4 (or "Agree") on each
individual statement (Table 3). Modestly lower scores in the PEOU dimension were largely
driven by generally less agreeable responses to statements 15 "I find the display of the
device easy to read outdoors" and 26 "I was able to shower or bathe normally while
wearing the device", again noting that we instructed the participants to remove the device
before bathing or showering. The median for the total TAQ score was 76% of the maximum
possible score (median=118, maximum possible score=155).
270
Next, we analyzed responses to the six multiple choice questions from the TAQ (questions
32-37) that did not have designated dimensions (Table 4). Most participants found the
device useful (question 32), with 95.4% rating the information provided as either "very
useful" or "somewhat useful". Nearly all participants (92.3%) expressed willingness to
continue using the device to monitor their health (question 33), and 96.9% reported
wearing the device for 15–21 days out of the 21-day period (question 34). In terms of
value, most participants indicated willingness to pay no more than $100 for the device
(question 35). Most (90.8%) participants reported looking at their health data provided by
their device (question 36). Finally, self-perception of activity levels varied among
participants with 64.6% considering themselves to be active (question 37).

Relationships Between Self-Reported Acceptance and Fitbit Adherence
Overall participants wore the Fitbit an average and standard deviation of 80.0%±24.7% of
the total minutes in a day, with a median wear time of 91% and an interquartile range of
22%. Wear time exceeded the threshold needed to be considered a valid wear day on
78.0%±25.8% of days with a median of 90% and an interquartile range of 33%. The
scatterplots in Figures 3 and 4 show relationships between the summed scores of the
different TAQ dimensions (as well as TAQ total scores) and the percentages of wear time
and valid wear days, respectively. Contrary to our hypothesis, there were no statistically
significant positive relationships between TAQ dimension summed scores or total scores
and Fitbit adherence metrics. We did, however, observe two small but statistically
significant negative associations between PU and percent wear time ($\rho=-0.27$, p=0.03) and
between PU and the percentage of valid wear days ($\rho=-0.26$, p=0.04).
293

Discussion

Principal Results

Our study examines self-reported perceptions of the Fitbit Inspire 2 wearable activity monitor among individuals with stroke, as measured by the Technology Acceptance Questionnaire (TAQ). A majority of the participants thought the device was beneficial for their health, efficient for monitoring their health, easy to use and don/doff, and unintrusive to daily life; one notable exception was the response to the statement "I find the display of the device easy to read outdoors". Generally, participants did not express significant concerns about privacy or data security, consistent with prior studies.[30,31] Contrary to our hypothesis, more agreeable responses to the TAQ statements were not associated with higher device percentage of average daily wear time and percentage of valid wear days at a statistically significant level.

Comparison with Prior Work

The findings of our study are consistent with previous literature demonstrating acceptance of wearable activity monitors in other populations.[13,32–36] Given the growing interest in using wearables for activity monitoring and telerehabilitation after stroke,[26,37–40] it is important to consider not only their potential benefits but also because of low levels of physical activity, cognitive and motor impairments that may affect this population's acceptance and engagement.[10,41,42] There were no commonly reported acceptance concerns that we deemed likely to be related to post-stroke motor impairment: as examples, participants widely agreed to the statements "I was able to wear the device easily without help from another person," "I find the device easy to use," "I was able to put the device on in a reasonable amount of time," and "I was able to remove the device easily without help from another person." These findings complement recent studies demonstrating the perceived value and user satisfaction of wearable technologies in stroke rehabilitation[43–45] and support a path toward scalable implementation of remote monitoring with wearable devices. This is likely due in part to the flexibility we allowed in permitting participants to wear the device on their paretic side if needed, accommodating individual motor abilities.
We also highlight that the study participants reported generally agreeable responses across all seven dimensions of the TAQ. Previous work highlighted that technical and usability issues (e.g., requiring a mobile application to sync the data from the device to the server) may affect the perceived usefulness of wearable devices;[16] however, we did not observe this in our sample. This is potentially because any technical difficulties were often addressed via interactions with the study team. Our findings across the different dimensions were largely similar to those reported in a previous sample of older adults.[13] It is important to emphasize that monitoring of device adherence may be necessary despite the high reported acceptance. Our findings did not support the hypothesis that higher user acceptance as measured by the TAQ would correlate with measures of adherence to wearing the device, as we did not observe statistically significant positive correlations between TAQ responses and our measures of adherence. This revealed that high reported acceptance of the device does not guarantee that a patient or research participant will necessarily adhere to wearing the device in everyday life. Technologies that help to

337 automate oversight of device wear and messaging to promote adherence will be important
338 for ensuring data quality.[46]
339
The correlational analyses indicated small but significant negative associations between
the perceived usefulness (PU) score and both adherence metrics. These results are
contrary to our hypotheses, which were grounded in the TAM and related literature, where
PU is typically positively associated with adoption behaviors.[47,48] Existing studies have
shown that higher PU is often associated with sustained use of technology across various
domains,[49–51] including healthcare settings.[52–55] Our findings may be attributed to
the specific context in which the wearable device was used. Unlike most prior technology
adoption studies where participants voluntarily adopted technology based on its
usefulness, our study cohort used the Fitbit as part of their participation in a research
study. This mandated context could have influenced the PU from typical motivational
factors driving technology adoption. It is possible that participants did not view Fitbit as
inherently useful for their health recovery goals but instead perceived it as a tool for
fulfilling study requirements. While survey responses suggest that most participants
agreed that the device could improve health and monitor well-being efficiently, these
endorsements may reflect a general perception of health technology utility rather than a
personalized after stroke recovery needs. Consequently, their assessment of PU may reflect
this externally driven motivation rather than genuine alignment with personal health
management goals.
358
As the push toward clinical use of wearables in stroke rehabilitation continues to move
forward, it is also important to consider the needs and perspectives of all key stakeholders
– patients, their family members and caregivers, and clinicians – in addition to the device
acceptance reported by patients in our study. For example, recent studies have provided
vital information regarding how persons with stroke prefer to receive data from wearables
and identified a set of metrics deemed most useful;[23] incorporated perspectives from
patients and clinicians on the value of using wearables and identified preferences for
incorporation into clinical care;[44] and demonstrated important design considerations for
adoption of the wearables and accompanying smartphone applications as outlined by
persons with stroke.[45] For example, to consider the difficulty of donning/doffing
providing an alternative strap such as a Velcro strap instead of the original buckle band.
Future work should consider these multifaceted aspects – patient acceptance, patient and
clinician data provision preferences, and device design – as we move closer to clinical
implementation of wearables in stroke rehabilitation.

Limitations
We acknowledge some limitations in our study. First, our sample was heterogeneous
regarding stroke chronicity. Accordingly, we did not design this study to assess how
wearable device acceptance may differ across stages of stroke recovery (e.g., acute,
subacute, chronic). However, this diversity in stroke recovery stages could be beneficial, as
it reflects the clinical reality where wearable devices in stroke rehabilitation should not
discriminate based on recovery stage but rather be accessible and valuable to patients at
various points in their recovery. Second, we only used the Fitbit Inspire 2 device. While we
anticipate that many of our findings may generalize across different models of

commercially available wearable devices due to the nature of the statements included in
the TAQ, we do not have data to support this directly. Third, we focused on the TAQ to
provide information about device acceptance in particular. We did not use other
instruments that could provide additional data on other aspects of patient perceptions
about wearable devices (e.g., the System Usability Scale for assessment of usability).

Conclusion

This study reported the perceived acceptance of a wrist-worn activity monitor in persons
with stroke. In response to statements on the TAQ, participants with stroke generally
reported the device to be beneficial for their health, useful for monitoring their health, easy
to use, and minimally intrusive. We observed generally agreeable responses to TAQ
statements across all seven dimensions of the instrument. Contrary to our hypothesis,
more agreeable responses to the TAQ statements were not positively correlated with
metrics of device wear, revealing that adherence to wearing the device should not be
assumed even when participants report high device acceptance. In summary, this study
detailed new information about the acceptance of wearable activity monitors in persons
with stroke and its association with real-world device wear.

Acknowledgments

399 We acknowledge funding from the American Heart Association (grants 23IPA1054140 and
400 935556 to RTR, grant 24POST1187285 to GCB) and the Sheikh Khalifa Stroke Institute at
401 Johns Hopkins Medicine.

Conflicts of Interest

403 The authors declare no relevant conflicts of interest.

Abbreviations

CONSORT: Consolidated standards of reporting trials
EC: Equipment characteristics
FC: Facilitating conditions
IQR: Interquartile range
PC: Privacy concern
PEOU: Perceived ease of use
PR: Perceived risk
PU: Perceived usefulness
SN: Subjective norm
TAM: Technology Acceptance Model
TAQ: Technology Acceptance Questionnaire

Data Availability

Data will be made available upon reasonable request.

Figure Captions
Figure 1. CONSORT diagram showing the enrollment and participation of participants at
various stages throughout the study.
Figure 2. Percentages of responses to each individual statement on the TAQ grouped by
the seven TAQ dimensions.
Figure 3. Scatterplots showing relationships between summed scores for each TAQ
dimension (as well as total TAQ scores) and Fitbit wear time. Asterisks indicate statistically
significant relationships ($p<0.05$).
Figure 4. Scatterplots showing relationships between summed scores for each TAQ
dimension (as well as total TAQ scores) and valid Fitbit wear days. Asterisks indicate
statistically significant relationships ($p<0.05$).

References

1. Patel MS, Asch DA, Volpp KG. Wearable Devices as Facilitators, Not Drivers, of Health Behavior Change. JAMA 2015 Feb 3;313(5):459. doi: 10.1001/jama.2014.14781

2. Piwek L, Ellis DA, Andrews S, Joinson A. The Rise of Consumer Health Wearables: Promises and Barriers. PLOS Med 2016 Feb 2;13(2):e1001953. doi: 10.1371/journal.pmed.1001953

3. Chang M, Büchel D, Reinecke K, Lehmann T, Baumeister J. Ecological validity in exercise neuroscience research: A systematic investigation. Eur J Neurosci 2022 Jan;55(2):487–509. doi: 10.1111/ejn.15595

4. Ingram JN, Wolpert DM. Naturalistic approaches to sensorimotor control. Prog Brain Res Elsevier; 2011. p. 3–29. doi: 10.1016/B978-0-444-53752-2.00016-3ISBN:978-0-444-53752-2

5. Peterson SM, Singh SH, Wang NXR, Rao RPN, Brunton BW. Behavioral and Neural Variability of Naturalistic Arm Movements. eneuro 2021 May;8(3):ENEURO.0007-21.2021. doi: 10.1523/ENEURO.0007-21.2021

6. Organization WH. International Classification of Functioning Disability and Health (ICF). Geneva: World Health Organization; 2001. ISBN:978-92-4-154542-6

7. Langhorne P, Bernhardt J, Kwakkel G. Stroke rehabilitation. The Lancet 2011 May;377(9778):1693–1702. doi: 10.1016/S0140-6736(11)60325-5

8. Lawrence ES, Coshall C, Dundas R, Stewart J, Rudd AG, Howard R, Wolfe CDA. Estimates of the Prevalence of Acute Stroke Impairments and Disability in a Multiethnic Population. Stroke 2001 Jun;32(6):1279–1284. doi: 10.1161/01.STR.32.6.1279

9. Li S. Stroke Recovery Is a Journey: Prediction and Potentials of Motor Recovery after a Stroke from a Practical Perspective. Life 2023 Oct 15;13(10):2061. doi: 10.3390/life13102061

10. Fini NA, Holland AE, Keating J, Simek J, Bernhardt J. How is physical activity monitored in people following stroke? Disabil Rehabil 2015 Sep 11;37(19):1717–1731. doi: 10.3109/09638288.2014.978508

11. Ringeval M, Wagner G, Denford J, Paré G, Kitsiou S. Fitbit-Based Interventions for Healthy Lifestyle Outcomes: Systematic Review and Meta-Analysis. J Med Internet Res 2020 Oct 12;22(10):e23954. doi: 10.2196/23954

12. McMahon SK, Lewis B, Oakes JM, Wyman JF, Guan W, Rothman AJ. Assessing the Effects of Interpersonal and Intrapersonal Behavior Change Strategies on Physical Activity in Older Adults: a Factorial Experiment. Ann Behav Med 2017 Jun;51(3):376–390. doi: 10.1007/s12160-016-9863-z

13. Puri A, Kim B, Nguyen O, Stolee P, Tung J, Lee J. User Acceptance of Wrist-Worn
Activity Trackers Among Community-Dwelling Older Adults: Mixed Method Study.
JMIR MHealth UHealth 2017 Nov 15;5(11):e8211. doi: 10.2196/mhealth.8211

14. Walker K, Griffiths C, Yates J, Birkett-Swan L, Hina F, Ang CS. The Experience and
Impact of a Fitbit-Based Physical Activity and Sleep Hygiene Intervention for Older
Adults Experiencing Cognitive Decline. Adv Aging Res 2024;13(03):41–59. doi:
10.4236/aar.2024.133004

15. Wilson S, Beswick E, Morrell R, Bhogal S, Tolley C, Whitfield T, Wing K, Ardle RM,
Walker Z, Slight S. Acceptance of wearable technology for the early detection of
dementia-causing diseases: perspectives from the CODEC II cohort. Alzheimers
Dement 2024 Dec;20(S2):e089597. doi: 10.1002/alz.089597

16. Hermsen S, Moons J, Kerkhof P, Wiekens C, De Groot M. Determinants for Sustained
Use of an Activity Tracker: Observational Study. JMIR MHealth UHealth 2017 Oct
30;5(10):e164. doi: 10.2196/mhealth.7311

17. Lavelle G, Norris M, Flemming J, Harper J, Bradley J, Johnston H, Fortune J, Stennett A,
Kilbride C, Ryan JM. Validity and Acceptability of Wearable Devices for Monitoring
Step-Count and Activity Minutes Among People With Multiple Sclerosis. Front Rehabil
Sci 2022 Jan 11;2:737384. doi: 10.3389/fresc.2021.737384

18. Lee Y, Kozar KA, Larsen KRT. The Technology Acceptance Model: Past, Present, and
Future. Commun Assoc Inf Syst 2003;12. doi: 10.17705/1CAIS.01250

19. Holden RJ, Karsh B-T. The Technology Acceptance Model: Its past and its future in
health care. J Biomed Inform 2010 Feb;43(1):159–172. doi: 10.1016/j.jbi.2009.07.002

20. Chen K, Chan AHS. Predictors of gerontechnology acceptance by older Hong Kong
Chinese. Technovation 2014 Feb;34(2):126–135. doi:
10.1016/j.technovation.2013.09.010

21. Kim J, Park H-A. Development of a Health Information Technology Acceptance Model
Using Consumers' Health Behavior Intention. J Med Internet Res 2012 Oct
1;14(5):e133. doi: 10.2196/jmir.2143

22. Scherer R, Siddiq F, Tondeur J. The technology acceptance model (TAM): A meta-
analytic structural equation modeling approach to explaining teachers' adoption of
digital technology in education. Comput Educ 2019 Jan;128:13–35. doi:
10.1016/j.compedu.2018.09.009

23. Demers M, Cain A, Bishop L, Gunby T, Rowe JB, Zondervan DK, Winstein CJ.
Understanding stroke survivors' preferences regarding wearable sensor feedback on
functional movement: a mixed-methods study. J NeuroEngineering Rehabil 2023 Nov
1;20:146. PMID:37915055

24. Fuller D, Colwell E, Low J, Orychock K, Tobin MA, Simango B, Buote R, Van Heerden D,
Luan H, Cullen K, Slade L, Taylor NGA. Reliability and Validity of Commercially
Available Wearable Devices for Measuring Steps, Energy Expenditure, and Heart Rate:
Systematic Review. JMIR MHealth UHealth 2020 Sep 8;8(9):e18694. doi:
10.2196/18694

25. French MA, Keatley E, Li J, Balasubramanian A, Hansel NN, Wise R, Searson P, Singh A,
Raghavan P, Wegener S, Roemmich RT, Celnik P. The feasibility of remotely
monitoring physical, cognitive, and psychosocial function in individuals with stroke or
chronic obstructive pulmonary disease. Digit Health 2023 Jan;9:205520762311761.
doi: 10.1177/20552076231176160

26. Koffman LJ, Crainiceanu CM, Roemmich RT, French MA. Identifying Unique Subgroups
of Individuals With Stroke Using Heart Rate and Steps to Characterize Physical
Activity. J Am Heart Assoc Wiley; 2023 Sep 19;12(18):e030577. doi:
10.1161/JAHA.123.030577

27. Schriesheim CA, Hill KD. Controlling Acquiescence Response Bias by Item Reversals:
The Effect on Questionnaire Validity. Educ Psychol Meas 1981 Dec;41(4):1101–1114.
doi: 10.1177/001316448104100420

28. Van Domelen DR. Functions for Processing Accelerometer Data. 2018.

29. Team R. R: A language and environment for statistical computing. Vienna, Austria;
2023.

30. Davis FD. Perceived Usefulness, Perceived Ease of Use, and User Acceptance of
Information Technology. MIS Q 1989 Sep;13(3):319. doi: 10.2307/249008

31. Venkatesh V, Davis FD. A Theoretical Extension of the Technology Acceptance Model:
Four Longitudinal Field Studies. Manag Sci 2000 Feb;46(2):186–204. doi:
10.1287/mnsc.46.2.186.11926

32. de Looff PC, Noordzij ML, Nijman HLI, Goedhard L, Bogaerts S, Didden R. Putting the
usability of wearable technology in forensic psychiatry to the test: a randomized
crossover trial. Front Psychiatry Frontiers; 2024 Jun 14;15. doi:
10.3389/fpsyt.2024.1330993

33. Drehlich M, Naraine M, Rowe K, Lai SK, Salmon J, Brown H, Koorts H, Macfarlane S,
Ridgers ND. Using the Technology Acceptance Model to Explore Adolescents'
Perspectives on Combining Technologies for Physical Activity Promotion Within an
Intervention: Usability Study. J Med Internet Res 2020 Mar 6;22(3):e15552. doi:
10.2196/15552

34. Jiang A, Al-Dajani N, King C, Hong V, Koo HJ, Czyz E. Acceptability and feasibility of
ecological momentary assessment with augmentation of passive sensor data in young

adults at high risk for suicide. Psychiatry Res 2023 Aug 1;326:115347. doi:
10.1016/j.psychres.2023.115347

35.   Michalsen H, Henriksen A, Pettersen G, Hartvigsen G, Wangberg S, Thrane G, Jahnsen
R, Anke A. Using mobile health to encourage physical activity in individuals with
intellectual disability: a pilot mixed methods feasibility study. Front Rehabil Sci
Frontiers; 2023 Aug 24;4. doi: 10.3389/fresc.2023.1225641

36.   Rosenberg D, Kadokura EA, Bouldin ED, Miyawaki CE, Higano CS, Hartzler AL.
Acceptability of Fitbit for physical activity tracking within clinical care among men
with prostate cancer. AMIA Annu Symp Proc 2017 Feb 10;2016:1050–1059.
PMID:28269902

37.   Lang CE, Waddell KJ, Barth J, Holleran CL, Strube MJ, Bland MD. Upper Limb
Performance in Daily Life Approaches Plateau Around Three to Six Weeks Post-stroke.
Neurorehabil Neural Repair 2021 Oct;35(10):903–914. PMID:34510934

38.   Waddell KJ, Strube MJ, Tabak RG, Haire-Joshu D, Lang CE. Upper Limb Performance in
Daily Life Improves Over the First 12 Weeks Poststroke. Neurorehabil Neural Repair
SAGE Publications Inc STM; 2019 Oct 1;33(10):836–847. doi:
10.1177/1545968319868716

39.   Waddell KJ, Patel MS, Clark K, Harrington TO, Greysen SR. Effect of Gamification With
Social Incentives on Daily Steps After Stroke: A Randomized Clinical Trial. JAMA
Neurol 2022 May 1;79(5):528–530. doi: 10.1001/jamaneurol.2022.0231

40.   Waddell KJ, Patel MS, Wilkinson JR, Burke RE, Bravata DM, Koganti S, Wood S, Morley
JF. Deploying Digital Health Technologies for Remote Physical Activity Monitoring of
Rural Populations With Chronic Neurologic Disease. Arch Rehabil Res Clin Transl 2023
Mar 1;5(1):100250. doi: 10.1016/j.arrct.2022.100250

41.   Derungs A, Schuster-Amft C, Amft O. Physical Activity Comparison Between Body
Sides in Hemiparetic Patients Using Wearable Motion Sensors in Free-Living and
Therapy: A Case Series. Front Bioeng Biotechnol 2018 Oct 17;6:136. doi:
10.3389/fbioe.2018.00136

42.   English C, Healy GN, Coates A, Lewis L, Olds T, Bernhardt J. Sitting and Activity Time in
People With Stroke. Phys Ther 2016 Feb 1;96(2):193–201. doi: 10.2522/ptj.20140522

43.   Demers M, Bishop L, Cain A, Saba J, Rowe J, Zondervan D, Winstein C. Wearable
technology to capture arm use of stroke survivors in home and community settings:
feasibility and early insights on motor performance. medRxiv; 2023. p.
2023.01.25.23284790. doi: 10.1101/2023.01.25.23284790

44.   Miller AE, Holleran CL, Bland MD, Fitzsimmons-Craft EE, Newman CA, Maddox TM,
Lang CE. Perspectives of Key Stakeholders on Integrating Wearable Sensor

Technology into Rehabilitation Care: A Mixed-Methods Analysis. medRxiv; 2024. p.
2024.11.25.24317911. doi: 10.1101/2024.11.25.24317911

45.   Stock R, Gaarden AP, Langørgen E. The potential of wearable technology to support
stroke survivors' motivation for home exercise – Focus group discussions with stroke
survivors and physiotherapists. Physiother Theory Pract Taylor & Francis; 2024 Aug
2;40(8):1779–1790. PMID:37246716

46.   French MA, Balasubramanian A, Hansel NN, Penttinen SK, Wise R, Raghavan P,
Wegener ST, Roemmich RT, Celnik PA. Impact of automated data flow and reminders
on adherence and resource utilization for remotely monitoring physical activity in
individuals with stroke or chronic obstructive pulmonary disease. medRxiv; 2024. p.
2024.04.15.24305852. doi: 10.1101/2024.04.15.24305852

47.   Adams DA, Nelson RR, Todd PA. Perceived Usefulness, Ease of Use, and Usage of
Information Technology: A Replication. MIS Q 1992 Jun;16(2):227. doi:
10.2307/249577

48.   Saini GS, Sharma S. Factors Affecting Consumers' Perception towards E-Payment
Systems in India. Indian J Comput Sci 2017 Jun 1;2(3):23. doi:
10.17010/ijcs/2017/v2/i3/115039

49.   Bhattacherjee A. Understanding Information Systems Continuance: An Expectation-
Confirmation Model. MIS Q 2001 Sep;25(3):351. doi: 10.2307/3250921

50.   Liu Y, Park Y, Wang H. The mediating effect of user satisfaction and the moderated
mediating effect of AI anxiety on the relationship between perceived usefulness and
subscription payment intention. J Retail Consum Serv 2025 May;84:104176. doi:
10.1016/j.jretconser.2024.104176

51.   Qi N, Koo JJ. A Composite Study of the Factors Influencing User Subscription Intention
to Artificial Intelligence Generated Content - Taking ChatGPT Users as an Example -.
Korean Soc Sci Art 2024 Mar 31;42(2):457–476. doi: 10.17548/ksaf.2024.03.30.457

52.   Hu PJ, Chau PYK, Sheng ORL, Tam KY. Examining the Technology Acceptance Model
Using Physician Acceptance of Telemedicine Technology. J Manag Inf Syst 1999
Sep;16(2):91–112. doi: 10.1080/07421222.1999.11518247

53.   Garavand A, Aslani N, Nadri H, Abedini S, Dehghan S. Acceptance of telemedicine
technology among physicians: A systematic review. Inform Med Unlocked
2022;30:100943. doi: 10.1016/j.imu.2022.100943

54.   Klingberg A, Sawe HR, Hammar U, Wallis LA, Hasselberg M. m-Health for Burn Injury
Consultations in a Low-Resource Setting: An Acceptability Study Among Health Care
Providers. Telemed E-Health 2020 Apr 1;26(4):395–405. doi: 10.1089/tmj.2019.0048

55.   Almojaibel AA, Munk N, Goodfellow LT, Fisher TF, Miller KK, Comer AR, Bakas T,
609   Justiss MD. Health Care Practitioners' Determinants of Telerehabilitation Acceptance.
610   Int J Telerehabilitation 2020 Jun 30;12(1):43–50. doi: 10.5195/ijt.2020.6308

Table 1. Study participant characteristics.

	Included (n=65) Mean±SD	Excluded (n=35) Mean±SD
Age (years)	62.4±12.4	60.5±12.5
Days post-stroke at date of enrollment	1053.9±1664.1	230.1±367.0
Days between date of enrollment and date of TAQ completion	127.8±42.9	N/A
	N (%)	**N (%)**
Sex		
Male	37 (56.9%)	20 (57.1%)
Female	28 (43.1%)	15 (42.9%
Race		
American Indian or Alaska Native	2 (3.1%)	0 (0.0%)
Asian	3 (4.6%)	1 (2.9%)
Black or African American	21 (32.3%)	22 (62.9%)
White/Caucasian	38 (58.5%)	10 (28.6%)
Multiple	1 (1.5%)	2 (5.7%)
Ethnicity		
Hispanic or Latino	4 (6.2%)	1 (2.9%)
Not Hispanic or Latino	61 (93.8%)	34 (97.1%)
Fitbit Wrist Worn*		
Paretic	15 (23.1%)	N/A
Non-paretic	49 (75.4%)	N/A
*Unknown	1 (1.5%)	N/A

*data not available for one participant

Table 2. Statistics of the responses to TAQ dimensions and the full TAQ.

	Mean±SD	Median	IQR	Max possible	Min possible
Perceived Usefulness	19.1±3.3	20	4	25	5
Perceived Ease of Use	25.7±3.9	25	5	35	7
Facilitating Conditions	7.7±1.5	8	1	10	2
Subjective Norm*	11.6±1.9	12	3	15	3
Equipment Characteristic*	32.5±3.5	32	5	40	8
Privacy Concern	11.8±2.0	12	3	15	3
Perceived Risk*	11.9±2.0	12	3	15	3
Technology Acceptance Questionnaire	120.4±12.8	118	14	155	31

*These dimensions have reverse items in the subscore. Note that PR consists of all reversed questions, therefore the entire dimension is reversed.

Table 3. Statistics of the responses to each individual statement of the TAQ.

Statement	Mean±SD	Median	IQR
1) 24/7 activity monitoring beneficial	4.4±0.6	4	1
2) Afraid of discovering health issue*	4.1±0.9	4	1
3) Comfortable with data storage	3.9±0.9	4	0
4) Assistance available if needed	3.8±1.1	4	0
5) Wore device easily unassisted	3.9±1.2	4	1
6) Removed device unassisted	4.0±1.1	4	1
7) Daily tasks performed normally	4.4±0.7	4	1
8) Device easily concealed	4.0±0.9	4	0
9) Forgot device was worn	4.0±1.0	4	1
10) Experienced skin irritation*	3.6±1.3	4	3
11) Battery life meets expectations	4.3±0.6	4	1
12) Smartphone app easy to use	3.8±0.9	4	0
13) Device easy to use	4.1±0.8	4	1
14) Display easy to read indoors	3.7±1.0	4	1
15) Display easy to read outdoors	3.1±1.2	3	2
16) Pleasant to wear during night	3.9±0.8	4	0
17) Concerned about secure attachment*	4.1±0.8	4	1
18) Device donning reasonable	4.0±0.9	4	0
19) No privacy concerns	3.9±1.1	4	1
20) Comfortable with data sharing	4.1±0.7	4	1

21) Necessary device knowledge	3.9±0.9	4	0
22) Efficient for monitoring health	3.5±1.1	4	1
23) Device motivated activity	3.6±1.0	4	1
24) Device can improve health	3.9±0.9	4	0
25) Device caused joint pain*	4.3±0.5	4	1
26) Bathed normally with device	2.8±0.8	3	0
27) Embarrassed wearing around family*	4.4±0.7	4	1
28) Friends would encourage device use	3.4±0.9	4	1
29) Family would encourage device use	3.8±1.0	4	1
30) Device extends home living	3.7±0.9	4	1
31) Usable in various locations	4.0±0.8	4	1

*These statements have been reverse coded to preserve directionality.

Table 4. Responses to TAQ multiple choice questions.

	% Responses

32. How useful did you find the information provided by the smart wearable device (such as step count, sleep data, heart rate) either on the wearable itself, or in the smartphone application?

Very useful	55.4%
Somewhat useful	40.0%
Not very useful	4.6%
Not at all useful	0.0%

33. Would you use the device you used during the last 21 days to continue to monitor or track your physical activity or health?

Yes	92.3%
No	7.7%

34. Over the last 21 days, how often do you think you wore the smart wearable device?

Never	0.0%
Between 0-7 days	0.0%
Between 8-14 days	3.1%
Between 15-21 days	96.9%

35. How much would you be willing to pay for the device you wore during the last 21 days?

$0	23.1%
$1-$50	26.2%
$51-$100	36.9%
$101-$200	13.9%
$201-$300	0.0%
$300-$400	0.0%

36. Did you find yourself looking at your health data in the smartphone application more/less often after the first few days?

No, I looked at the health data fairly consistently throughout the 21-day period	32.3%
Yes, I looked at the health data more often after the first few days of use	43.1%
Yes, I looked at the health data less often after the first few days of use	15.4%
I did not look at my health or am not interested in monitoring it	9.2%

37. Do you consider yourself to be an active person?

Yes	64.6%
No	35.4%