Clinical Development and Performance of the First To Know® Syphilis Self-Test for Over-the-Counter Usage: a De Novo Rapid Test for Treponemal Antibody

Kevin Clark[1], Sharon Rabine[1], Jianghong Qian[1], Diane Biddison[1] , Sydney Spradlin[1], Vicki Thompson[1], Qinwei Shi[1], Jing Xu[1], Raji Pillai[1], Lu Zhang[1], Alicia Brown[1], Lisa Nibauer[1], Antony George Joyee[1], Jody D. Berry[1]

Short Title: Performance of the First To Know® Syphilis Over-the-Counter Self-test

[1] Now Diagnostics, Inc. 1200 Stewart Pl, Springdale, AR 72764 USA

*Corresponding author: Jody D. Berry, Now Diagnostics, Inc. jody.berry@nowdx.com

ABSTRACT

Background: The resurgence of syphilis in the USA and around the globe has hastened the need for widely available syphilis testing. Early and rapid detection of syphilis is crucial for prompt treatment, avoiding serious clinical complications, and preventing the spread of infection. Self-administered rapid tests enhance access to testing and promote timely treatment for positive cases. The present report describes the evaluation of diagnostic performance of First to Know® Syphilis Test, the first over the counter (OTC) treponemal rapid test.

Methods: A prospective multi-site clinical study with 1270 subjects was conducted, where the subjects self-collected capillary blood from a finger stick and executed self-testing without the assistance of any site personnel. The test results were compared with a composite reference standard to evaluate the diagnostic performance of the First To Know® Syphilis Test. Additional analysis was performed in a panel of 125 pre-characterized clinically staged syphilis patient samples.

Results: The First to Know® test exhibited an overall sensitivity of 93.4% (95% CI, 87.0% - 96.8%) and specificity of 99.5% (95% CI, 98.9% - 99.8%) and accuracy of 99%. The overall agreement was 98.9%, with Cohen's Kappa (k) value 0.93 (95% CI, 0.90–0.97) showing excellent agreement between the reference standard and First to Know® test results. In addition, the test was validated in a panel of 125 clinically staged syphilis patient samples and the results showed 100% agreement in detecting anti-treponemal antibodies in all the samples tested.

Conclusions: The present study findings showed an excellent diagnostic performance of the First to Know® Syphilis Test demonstrating its utility in the screening and diagnosis of syphilis. The availability of this OTC test and its excellent performance at low prevalence, will have a profound impact on syphilis detection and prevention strategies and could reduce co-morbidities such as HIV and other STI transmission.

Keywords: Syphilis, *Treponema pallidum*, Diagnosis, Rapid syphilis test, Self-test, Over-the-Counter Syphilis Test

Introduction

Syphilis is a sexually transmitted infection (STI), caused by the spirochete bacterium, *Treponema pallidum.* It is transmitted by direct sexual contact or through vertical transmission from mother to child during pregnancy and can cause substantial morbidity and mortality. The number of syphilis cases in the U.S. is currently on the rise. According to the recent WHO report, infected cases have escalated globally by over 1 million in 2022, reaching a total of 8 million (1) that included 700,000 cases of congenital syphilis (2). The CDC reported that the cases increased by nearly 80% between 2018 and 2022 (from 115,000 to more than 207,000) (3). The rise in syphilis can be attributed to multiple factors including social and behavioral changes in this PrEP era*,* health disparities, particularly among sexual and gender minority populations, lack of timely testing and adequate treatment, intersections with the HIV and substance use epidemics, and increased maternal transmission leading to congenital syphilis infections (4). Syphilis is a chronic, multi-stage disease. The natural course of syphilis progresses through four stages including primary, secondary, latent and tertiary stages, with the primary and secondary stages being the most infectious. It must be treated in early stages to be cured. If left untreated, syphilis can lead to serious complications and permanent damage in the nervous and cardiovascular systems that can be life-threatening (5, 6, 7). Syphilis during pregnancy, if untreated or treated late, can result in serious adverse birth outcomes including still births, neonatal deaths, prematurity, low birth weight, and congenitally infected infants (8, 9). Most people are often asymptomatic or do not notice symptoms, which makes screening a crucial step in diagnosis. There is no natural immunity to syphilis and the antibodies generated from past infection are insufficient to confer protection against re-infection.

The laboratory diagnosis of syphilis is complex and necessitates a combination of clinical and laboratory criteria to differentiate current or past infection, and absence of infection (10, 11). Culture for *T. pallidum* is cumbersome and is available only in selected research laboratories and there is no FDA-cleared PCR test for syphilis at present. Serological testing remains the mainstay for diagnosing syphilis infection. Serology is divided into two different tests: nontreponemal tests (NTTs) such as rapid plasma regain [RPR] test and Venereal Disease Research Laboratory (VDRL) and treponemal tests (TTs) such as *T. pallidum* Hemagglutination Assay (TPHA). While the TT is often positive for life, NTT titers decrease after treatment, therefore the current TTs and NTTs have their limitations, hampering syphilis control and prevention efforts in real-world settings (9). A single serologic NTT or TT is not sufficient for diagnosis and therefore a combination of positive TT and reactive NTT is required for the diagnosis of syphilis infection. In the past, syphilis screening mainly followed a traditional testing algorithm, starting with NTT and the positive results require further confirmation through more specific TTs (9, 10). With increased availability of various treponemal-specific immunoassays including enzyme immunoassays (EIAs), chemiluminescence immunoassays (CIAs) etc., recently, many laboratories are increasingly adopting a 'reverse sequence screening' algorithm which begins the screening with a TT and the reactivity is confirmed with an NTT. When the NTT is non-reactive, a second but different treponemal test is performed to determine if the first treponemal test was a false positive (12). Regardless of which algorithm is used, it is important to consider the sensitivity and specificity of these assays in clinically characterized sera, stratified by stage of syphilis (13) and all screening results must be correlated with clinical diagnosis covering patients' symptoms, previous history and sexual risk factors to make an accurate diagnosis (14).

Frequent syphilis testing for timely and accurate diagnosis of active infections for appropriate clinical management is a key strategy to effectively prevent disease transmission (7). This could be achieved with rapid point-of-care tests which are easy to perform, and inexpensive so that repeated testing is practical, and test results are available immediately. Rapid over the counter (OTC) tests have become entrenched in society since the COVID-19 pandemic. There is a broad acceptance of being able to self-test at home. An in-home test for syphilis has the potential to greatly impact public health by improving access to timely detection and treatment.

Herein we describe the clinical performance of the first over-the-counter (OTC) treponemal rapid test. The First To Know® Syphilis Test (Now Diagnostics Inc.) is the first at-home, OTC test for syphilis, having received FDA marketing authorization under a de novo classification pathway in August 2024. It is a rapid buffer-less lateral flow immunoassay test device that detects *T. pallidum* antibodies in human whole blood (capillary) from individuals suspected of having a syphilis infection. Here, we report the results from a clinical study that evaluated the test performance of the First To Know® Syphilis Test.

Materials and methods

The First To Know® Syphilis Test is a simple buffer-less lateral flow visually-read test device that detects antibodies to syphilis antigens and requires a fingerstick and whole blood to run the test in as little as 15 minutes. There is no need for a swab nor a stand to run this test. Briefly, a drop of human capillary whole blood is added to the test device cassette, where the blood is drawn into the fill zone. As the sample flows into the porous test strip, syphilis-specific antibody in the sample

binds at the test band location. The appearance of a visible test band indicates the sample contains a detectable level of anti-treponemal antibody. The internal control and the appearance of the control band line assures that the sample is applied correctly, and the test components are working properly. The test is a chromatographic immunoassay that delivers results in 15 minutes.

Study subjects, sample collection, and Syphilis testing

To evaluate the performance of the First To Know® Syphilis Test, a prospective clinical study was conducted in six geographically diverse sites across the US (the sites were in Tempe, AAZ; Las Vegas, ANV; Los Angeles, MCR; Spokane, MWA; West Orange, PNJ and North Miami, SFL). Individuals who met the inclusion criteria were consecutively included in the study. The sample size for the test performance evaluation was determined based on the FDA guidance documents. For the clinical study, subjects from two cohorts were enrolled (September 1, 2021, through October 17, 2023); One cohort of sexually active persons (18-64 years old) and another cohort of expectant mothers (≥18 years old). The enrolled subjects self-collected capillary blood from a finger stick and executed self-testing according to the instructions for use with the First To Know® Syphilis Test. The clinical study protocol and the informed consent form were approved by an IRB. If the test results were incomplete or not performed or lab results were missing for any samples, they were excluded from the analysis. Of the 1345 subjects enrolled, samples from 75 subjects had invalid results and were excluded. A total of 1270 subjects from the two cohorts who obtained a valid result were included (1011 samples from the sexually active cohort and 259 samples from the expectant mothers cohort combined for the final evaluation) to evaluate the diagnostic performance of the First To Know® Syphilis Test. There were no adverse events in the performance study.

Composite reference result

Venous whole blood samples were also collected from the subjects, plasma and serum samples prepared from whole blood were sent and tested by a reference lab using FDA-cleared comparator tests. The comparator tests included two treponemal tests (BioPlex 2200 Syphilis Total and SERODIA- TP-PA Test) and one non-treponemal test (Wampole Impact RPR Test). The reference diagnosis of 'true positive' for syphilis was established if at least two of the three tests were positive, and this composite reference standard was used for comparison with the First To Know® Syphilis Test to evaluate test sensitivity and specificity. The participants did not have reference standard test results at the time of self-testing with the First To Know® Syphilis Test. The Standards for Reporting of Diagnostic Accuracy Studies (STARD) reporting guidelines were followed to ensure the completeness and transparency of reporting diagnostic accuracy studies (15).

Clinically staged syphilis patient samples and testing with First to Know® Syphilis Test

In addition to the clinical study evaluation, we validated the diagnostic accuracy of the First To Know® Syphilis Test in a panel of characterized serum samples of known positive clinically staged patients obtained from the CDC. The panel included serum samples from patients with primary syphilis (n=25), secondary syphilis (n=56), early latent syphilis (n=16), and late latent syphilis (n=28).

Statistical analysis

Statistical analysis was performed using GraphPad Prism 9 (GraphPad Software, Inc.) The performance of the First To Know® Syphilis Test was evaluated based on comparison with

composite reference standard using a dichotomous approach, two by two crosstab analysis for the calculation of sensitivity (positive percent agreement [PPA]), specificity (negative percent agreement [NPA]), positive predictive value (PPV), and negative predictive value (NPV) and accuracy. The concordance analysis for diagnostic performance was determined using the Cohen's kappa (κ) value according to the criteria proposed by Landis & Koch, e.g. values more than 0.8 were considered as "almost perfect agreement" (16).

Results

Samples from a total of 1270 study participants (that consisted of two cohorts including a sexually active cohort and a pregnant women cohort) were subjected to this evaluation study for the First To Know® Syphilis Test. The overall positivity for syphilis in this study was determined to be 8.3% (106/1270) based on the composite reference standard. The First to Know® test showed positivity in 99 out of the 106 positive samples. Among the 1164 negative samples, 1158 samples were negative by the First to Know® test. Based on the composite reference test results, the First to Know® test yielded an overall sensitivity of 93.4% (95% CI, 87.0% - 96.8%) and specificity of 99.5% (95% CI, 98.9% - 99.8%) (Table 1). The test demonstrated an accuracy of 99% in this evaluation. The overall agreement between the reference standard and the First to Know® test was 98.9%, with Cohen's Kappa (k) value 0.93 (95% CI, 0.90–0.97) showing within the scale of 'almost perfect agreement' (Kappa between 0.81 and 1.00).

When the cohorts in the study population were analyzed separately, the sensitivity and specificity of the First to Know® test for the sexually active cohort group samples (n=1011), were

94.9% (95% CI, 88.5% - 98.3%) and 99.6% (95% CI, 98.9% - 99.9%) respectively, while in the pregnant women cohort (n=259), the sensitivity and specificity of the test were 75.0% (95% CI, 40.9% - 92.9%) and 99.2% (95% CI, 97.1% - 99.8%) respectively. The First To Know® test results observed with different cohorts are shown in Table 2.

In addition to the above, we also evaluated the First To Know® Syphilis Test in a set of 125 characterized serum samples from known positive clinically staged patients. The First to Know® Syphilis Test showed 100% sensitivity in detecting anti treponemal antibodies in all the samples tested, having absolute agreement with reference standard. The breakdown of different clinically staged groups and test positivity is shown in Table 3.

Discussion

Syphilis is a preventable and curable bacterial sexually transmitted infection but if left undiagnosed and untreated, it can result in serious downstream health outcomes. Further, there is a higher risk of acquiring HIV and other STIs (17, 18). Therefore, rapid and accurate detection of syphilis is indispensable to ensure treatment and to control the transmission of the disease. In the present study, we evaluated the diagnostic performance of the First To Know Syphilis® Test, in a large real world lateral flow study. The findings of present evaluation based on self-testing by the study subjects showed excellent overall clinical performance of the First To Know® test (93.4.% sensitivity, 99.5% specificity and 99% accuracy) compared with the reference method. This also meets the WHO recommended criteria of minimum of 85% sensitivity and 95% specificity for syphilis rapid tests (19). The sensitivities of commercially available POC tests are in the range of

76% - 98%, with 90% to 99% specificities in varied settings. (20-25). In addition to the clinical performance study, we validated the diagnostic accuracy of the First To Know® test in a panel of characterized samples from clinically staged syphilis patients. Our results showed that the First To Know® test showed 100% sensitivity in detecting anti-treponemal antibodies in all the tested samples in the panel demonstrating diagnostic efficacy. These results suggest that the First To Know® test has the potential to aid accurate syphilis diagnosis in high-risk patient populations with specific clinical stages, however larger and more detailed studies are needed to confirm the present finding.

The current resurgence of syphilis in the US and the alarming spread of infections underscores the critical need for rapid diagnosis and containment (2, 3, 26). Currently, syphilis incidence is increasing more rapidly among men who have sex with men (MSM) than any other U.S. subpopulation (27, 28). Consistent with literature, we observed that syphilis test positivity in MSM was significantly higher than that of non-MSM (32.8% vs 6.9%, p<0.001) (data not shown). In addition, there is a dramatic increase in congenital syphilis over the 4-year period from 2018 to 2022 (3) which clearly necessitates early detection. Increasing the coverage and frequency of syphilis testing are crucial elements of any control effort to allow earlier detection and treatment, interrupting transmission which aids to control the spread of syphilis. Specifically, focusing testing on high-risk populations like men who have sex with men (MSM), female sex workers and routine syphilis testing during pregnancy can significantly impact syphilis control. The de novo FDA marketing authorization and availability of the First To Know® Syphilis Test represent a key advancement in the landscape of rapid diagnostics for syphilis that would enable increased access to testing and early diagnosis and will support the universal effort in preventing transmission of infections. Although POC tests were available prior to the First To Know® marketing authorization, currently, the First To

Know® Syphilis Test is the only at-home, OTC test for syphilis. It should be noted that positive test results from this test should be followed by additional laboratory testing through a health care provider to confirm the diagnosis. Nevertheless, it serves best as a screening tool to warn individuals that they may have an identifiable infection.

While syphilis rates have risen considerably in high-risk populations, its prevalence is still relatively low in the overall population. The setting of our clinical performance study reflects a relatively low-prevalence environment, resembling more of a community-based setting. In an analysis comparing First To Know® test with two other U.S. commercial POC tests, we observed that the PPVs of the First To Know® test remained the highest and did not show dramatic decrease with a decrease in prevalence rates compared to the other two POC rapid tests. The NPVs showed no statistical difference between the tests (supplemental, Figure S1). This suggests that First To Know® test could accurately identify and refer positives from testing to further care (ruling in those at risk) and also could identify those who are negative. Overall, these observations and above findings indicate the reliability of the First To Know® test in early detection and management of the disease that can lead to better population health outcomes. Real world studies in different prevalence settings/populations are further needed to solidify the diagnostic efficacy and clinical utility of the test.

The availability of an over-the-counter test with at-home results like the First To Know® Syphilis Test has important implications that extend beyond testing for convenience. By being readily available, it could increase the initial volume of screening, which is the first step towards diagnosing syphilis, which is crucial for early treatment and prevention of serious complications. Home based self-testing for STIs like syphilis provides several benefits over conventional clinic/laboratory-based testing. In addition to the simplicity in performing the test, it provides

privacy, without fear of judgement and therefore offers significant advantages over currently available POC tests. Self-testing at home can be a beneficial strategy to reach more at-risk individuals who are reluctant or otherwise may never seek testing. In fact, it has been proven effective in the case of HIV control programs and helped to counter barriers for testing those at high-risk who otherwise rarely get tested (29, 30). The recent CDC randomized clinical trial called Evaluation of Rapid HIV Self-testing Among MSM Project (eSTAMP) study showed that self-testing increased the uptake and frequency in testing and resulted in people seeking care or additional testing after obtaining a positive result (31). These programs and other studies in the U.S. exemplify the high level of acceptability and effectiveness of home testing for STIs (29). With the availability of an OTC test like the First To Know® test, it is feasible to implement similar programs for syphilis which could help control spread of infection. In this regard, the First To Know® Syphilis Test's ease of use with a simple fingerstick blood micro sampling and short time to obtain results with high sensitivity and specificity can be very advantageous for self-testing and to facilitate early diagnosis. Our data showed that the First To Know® test is robust for self-testing in different settings; the usability studies demonstrated that the error rate during self-testing with First To Know® was very minimal (data not shown) underscoring the ease of use and suitability for testing by lay users.

There are some limitations to our study; the study population represents a low prevalent setting, and the performance in high prevalence groups was not evaluated here. Therefore, the results from the current study might not reflect its performance in higher prevalence settings thus possibly underestimating the test performance. While our study included a pregnant cohort, the sample size was small. Considering the importance of syphilis testing in pregnant women, it is important to have

further detailed clinical studies in pregnant women. In addition, future research involving larger, clinically stratified patient populations will be important to strengthen our findings.

In conclusion, the present study observations showed an excellent overall diagnostic performance of the First To Know® Syphilis Test demonstrating its utility as a valuable tool in the screening and diagnosis of syphilis. As the incidence of syphilis continues to rise in the United States and worldwide, the availability of this OTC test can have a profound impact on syphilis detection and prevention strategies and could support reduction of co-morbidities such as HIV and other STI transmission.

Author Contributions

Conceptualization: KC, VT; Data curation: SR, DB, VT, LZ; Formal Analysis: DB, AGJ, LN, RP; Funding Acquisition: KC; Experimentation: KC, JQ, SS, VT, LZ, AB, QS; Methodology: JQ, KC,QS, SS, AB; Supervision: KC, VT; Validation: DB, SR; Resources: DB, AB; Project Administration: SR, RP; Writing Original Draft Preparation: AGJ, JDB; Writing Review & Editing: AGJ, JDB, KC, LN. All authors have read and agreed to the published version of the manuscript.

Funding

This work was supported by NOW Diagnostics

Competing Interest Statement

All authors are employees of Now Diagnostics. AGJ is a scientific consultant of Now Diagnostics. LN is marketing consultant and RP is project management consultant for Now Diagnostics. All other authors have disclosed that they have no conflicts of interest

Ethics Statement

This study was conducted in accordance with the ethical standards of the Helsinki Declaration. IRB approval for the multi-site study was obtained from Advarra IRB #Pro00032035.

The initial approval was on Dec 9, 2019, and latest approval was on July 25, 2023. Informed consent was obtained from all study subjects, and all the participant data were de-identified in the database.

Data Availability Statement

The study design and data used in this research are described in the text and additional details can be provided upon reasonable request.

Acknowledgments:

We thank Dave Kern, regulatory consultant, for his support for this work. We greatly thank Robert Weigle, Joshua Bafford, Thomas Grifa, Susan Ward, and Todd Grice for their continued support and technical inputs into the program and the manuscript. We thank Jing Shi, PhD for statistical guidance and inputs, Deonna Robertson for technical assistance and Todd Simpson for graphic arts support.

REFERENCES

1. Implementing the global health sector strategies on HIV, viral hepatitis and sexually transmitted infections, 2022–2030: report on progress and gaps 2024. WHO 22 May 2024 Situation report (https://www.who.int/publications/i/item/9789240094925)
2. WHO data on Syphilis 2024. https://www.who.int/data/gho/data/themes/topics/data-on-syphilis
3. Centers for Disease Control and Prevention. Sexually Transmitted Infections Surveillance 2022. Atlanta: US Department of Health and Human Services; 2024. https://www.cdc.gov/std/statistics/2022/default.htm
4. Papp JR, Park IU, Fakile Y, Pereira L, Pillay A, Bolan GA. CDC Laboratory Recommendations for Syphilis Testing, United States, 2024. MMWR Recomm Rep 2024;73(No. RR-1):1–32. DOI: http://dx.doi.org/10.15585/mmwr.rr7301a1.
5. World Health Organization. (2001). Global prevalence and incidence of selected curable sexually transmitted infections : overview and estimates. World Health Organization. https://iris.who.int/handle/10665/66818
6. Hook, E.W., 3rd. Syphilis. Lancet 2017, 389, 1550–1557
7. Pham MD, Ong JJ, Anderson DA, Drummer HE, Stoové M. Point-of-Care Diagnostics for Diagnosis of Active Syphilis Infection: Needs, Challenges and the Way Forward. Int J Environ Res Public Health. 2022 Jul 4;19(13):8172.

8. Gomez GB, Kamb ML, Newman LM, Mark J, Broutet N, Hawkes SJ. Untreated maternal syphilis and adverse outcomes of pregnancy: a systematic review and meta-analysis. Bull World Health Organ. 2013;91(3):217–26.

9. Tudor ME, Al Aboud AM, Leslie SW, et al. Syphilis. [Updated 2024 Aug 17]. In: StatPearls [Internet]. Treasure Island (FL): StatPearls Publishing; 2024 Jan-. Available from: https://www.ncbi.nlm.nih.gov/books/NBK534780/

10. Luo Y, Xie Y and Xiao Y (2021) Laboratory Diagnostic Tools for Syphilis: Current Status and Future Prospects. Front. Cell. Infect. Microbiol. 10:574806.

11. Spach DH, Ramchandani MS. 2024. Core Concepts - Syphilis - 2nd Edition - National STD Curriculum Syphilis. https://www.std.uw.edu/go/comprehensive-study/syphilis/core-concept/all

12. Morshed, M. G., and Singh, A. E. (2015). Recent trends in the serologic diagnosis of syphilis. Clin. Vaccine Immunol. CVI 22 (2), 137–147. doi: 10.1128/cvi.00681-14

13. Park IU, Tran A, Pereira L, Fakile Y. Sensitivity and Specificity of Treponemal-specific Tests for the Diagnosis of Syphilis. Clin Infect Dis. 2020 Jun 24;71(Suppl 1):S13-S20. doi: 10.1093/cid/ciaa349.

14. Ortiz DA, Shukla MR, Loeffelholz MJ. The Traditional or Reverse Algorithm for Diagnosis of Syphilis: Pros and Cons. Clin Infect Dis. 2020 Jun 24;71(Suppl 1):S43-S51. doi: 10.1093/cid/ciaa307.

15. Bossuyt PM, Reitsma JB, Bruns DE, Gatsonis CA, Glasziou PP, Irwig L, LijmerJG Moher D, Rennie D, de Vet HCW, Kressel HY, Rifai N, Golub RM, Altman DG, Hooft L, Korevaar DA, Cohen JF, For the STARD Group. STARD 2015: An Updated List of Essential Items for Reporting Diagnostic Accuracy Studies.

16. Landis, J.R.; Koch, G.G. The Measurement of Observer Agreement for Categorical Data. Biometrics 1977, 33, 159–174.

17. Boily MC, Baggaley RF, Wang L, et al. . Heterosexual risk of HIV-1 infection per sexual act: systematic review and meta-analysis of observational studies. Lancet Infect Dis 2009; 9:118–29.

18. Sellati TJ, Wilkinson DA, Sheffield JS, Koup RA, Radolf JD, Norgard MV. Virulent Treponema pallidum, lipoprotein, and synthetic lipopeptides induce CCR5 on human monocytes and enhance their susceptibility to infection by human immunodeficiency virus type 1. J Infect Dis 2000; 181:283–93.

19. Herring A, Ballard R, Mabey D, Peeling RW., WHO/TDR Sexually Transmitted Diseases Diagnostics Initiative. Evaluation of rapid diagnostic tests: syphilis. Nat Rev Microbiol. 2006 Dec;4(12 Suppl):S33-40.

20. Tudor ME, Al Aboud AM, Leslie SW, Gossman W. Syphilis. 2024 Aug 17. In: StatPearls [Internet]. Treasure Island (FL): StatPearls Publishing; 2024 Jan–. PMID: 30521201.

21. Jafari Y, Peeling RW, Shivkumar S, Claessens C, Joseph L, Pai NP. Are Treponema pallidum specific rapid and point-of-care tests for syphilis accurate enough for screening in resource limited settings? Evidence from a meta-analysis. PLoS One. 2013;8(2):e54695.

22. Bristow CC, Klausner JD, Tran A. Clinical Test Performance of a Rapid Point-of-Care Syphilis Treponemal Antibody Test: A Systematic Review and Meta-analysis. Clin Infect Dis. 2020 Jun 24;71(Suppl 1):S52-S57.

23. Fakile YF, Markowitz N, Zhu W, Mumby K, Dankerlui D, McCormick JK, Ham DC, Hopkins A, Manteuffel J, Sun Y, Huang YA, Peters PJ, Hoover KW. Evaluation of a Rapid Syphilis Test in an Emergency Department Setting in Detroit, Michigan. Sex Transm Dis. 2019 Jul;46(7):429-433.

24. Nakku-Joloba E, Kiragga A, Mbazira JK, Kambugu F, Jett-Goheen M, Ratanshi RP, Gaydos C, Manabe YC. Clinical Evaluation of 2 Point-of-Care Lateral Flow Tests for the Diagnosis of Syphilis. Sex Transm Dis. 2016 Oct;43(10):623-5

25. Causer LM, Kaldor JM, Fairley CK, Donovan B, Karapanagiotidis T, Leslie DE, Robertson PW, McNulty AM, Anderson D, Wand H, Conway DP, Denham I, Ryan C, Guy RJ. A laboratory-based evaluation of four rapid point-of-care tests for syphilis. PLoS One. 2014 Mar 11;9(3):e91504.

26. Amerson EH, Castillo Valladares HB, Leslie KS. Resurgence of Syphilis in the US—USPSTF Reaffirms Screening Guidelines. *JAMA Dermatol.* 2022;158(11):1241–1243.

27. Mayer KH. Old Pathogen, New Challenges: A Narrative Review of the Multilevel Drivers of Syphilis Increasing in American Men Who Have Sex With Men. Sex Transm Dis. 2018 Sep;45(9S Suppl 1):S38-S41.

28. Centers for Disease Control and Prevention. (2024). National Overview of STIs in 2023.

29. Melendez JH, Hamill MM, Armington GS, Gaydos CA, Manabe YC. Home-Based Testing for Sexually Transmitted Infections: Leveraging Online Resources During the COVID-19 Pandemic. Sex Transm Dis. 2021 Jan;48(1):e8-e10.

30. Steehler K, Siegler AJ. Bringing HIV Self-Testing to Scale in the United States: a Review of Challenges, Potential Solutions, and Future Opportunities. J Clin Microbiol 2019. 57(11):e00257–19.

31. MacGowan RJ, Chavez PR, Borkowf CB, Owen SM, Purcell DW, Mermin JH, Sullivan PS; eSTAMP Study Group. Effect of Internet-Distributed HIV Self-tests on HIV Diagnosis and Behavioral Outcomes in Men Who Have Sex With Men: A Randomized Clinical Trial. JAMA Intern Med. 2020 Jan 1;180(1):117-125.

Figure 1. First To Know® Syphilis test device showing visually read test results

Table 1. Evaluation of performance characteristics of First To Know® Syphilis test compared to the composite reference as authorized by the FDA for OTC

First To Know® assay	Reference (composite)		
	Positive	Negative	Subtotal
Positive	99	6	105
Negative	7	1158	1165
Subtotal	106	1164	1270

Sensitivity/PPA	93.4% (95% CI, 87.0% - 96.8%)
Specificity/NPA	99.5% (95% CI, 98.9% - 99.8%)
Positive Predictive Value (PPV)	94.3% (95% CI, 88.1% - 97.4%)
Negative Predictive Value (NPV)	99.4% (95% CI, 98.8% - 99.7 %)
Accuracy	99.0% (95% CI, 98.3% - 99.5%)

* For reference standard, positive and negative results were determined by a composite algorithm as described in the methods.

Table 2. Study cohorts and test performance

Study cohorts Total (n=1270)	First To Know® Syphilis Test	
	Sensitivity/PPA Test +/Ref + (%, 95% CI)	Specificity/NPA Test -/Ref - (%, 95% CI)
Sexually active cohort (n=1011)	93/98 (94.9%, 95% CI, 88.6% - 97.8%)	909/913 (99.6%, 95% CI, 98.9% - 99.8%)
Pregnant cohort (n=259)	6/8 (75.0%, 95% CI, 40.9%- 92.9%)	249/251 (99.2%, 95% CI, 97.1% - 99.8%)

Table 3. First To Know® Syphilis testing in Clinically staged syphilis patients (n=125).

Clinical stage of Syphilis Samples (n=125)	Number Tested	Test positivity Number (%)
Primary Syphilis	25	25 (100%)
Secondary Syphilis	56	56 (100%)
Early Latent Syphilis	16	16 (100%)
Late Latent Syphilis	28	28 (100%)

Supplemental

Figure S1. Comparative analysis of First To Know® Syphilis Test with other commercial tests on PPV and NPV with different prevalence settings based on sensitivities and specificities

*PPV and NPVs were based on sensitivities and specificities of different tests. FTK; First To Know® Syphilis Test