Title Page

Title:

Artificial Intelligence Generated Computed Tomography Segmentation of the Thoracoabdominal Aorta

Authors:

Devina Chatterjee[1,2], Nicholas T. Obey[1], Benjamin Shou[1,3,4], Sangmita Singh[4], Adrian R. Acuna Higaki[1], Adham Ahmed[1], Ely Erez[1], Michela Cupo[1,5], Nathaniel Price[6], Irbaz Hameed[1,5], Eric B. Schneider[1], Prashanth Vallabhajosyula[1,5]*, Chin Siang Ong[1,7]*

*These authors contributed equally to this work.

Affiliations:

[1] Department of Surgery, Yale School of Medicine, New Haven, Connecticut, USA

[2] University of Maryland School of Medicine, Baltimore, Maryland, USA

[3] Department of Cardiothoracic Surgery, Stanford University School of Medicine, Stanford, California, USA

[4] Johns Hopkins University School of Medicine, Baltimore, Maryland, USA

[5] Division of Cardiac Surgery, Department of Surgery, Yale School of Medicine, New Haven, Connecticut, USA

[6] Digital and Technology Solutions, Yale New Haven Hospital Health System, New Haven, Connecticut, USA

[7] Harvard T.H. Chan School of Public Health, Boston, MA, USA

Corresponding authors:

Chin Siang Ong, MBBS, PhD

Assistant Professor of Surgery

Division of Surgical Outcomes

Surgery Center for Health Services and Outcomes Research

Department of Surgery, Yale School of Medicine

100 Church Street South

New Haven, CT 06519

Phone: 203-432-4771

Email Address: chinsiang.ong@yale.edu

Prashanth Vallabhajosyula, MD, MS

Director, Yale Aortic Institute

Director, Yale Pulmonary Thromboendarterectomy Program

Division of Cardiac Surgery

Yale School of Medicine

330 Cedar Street, BB 204

New Haven, CT 06510

Phone: 203-785-5000

Email address: prashanth.vallabhajosyula@yale.edu

Meeting presentation:

Accepted for presentation at the 38th EACTS Annual Meeting, 9-12 October 2024,

Lisbon, Portugal.

Word Count:

1614

Graphical Abstract

Evaluation of Artificial Intelligence Generated Segmentation on Aortic Computed Tomography Data

Summary
P: 20 TEVAR patients 27 CTA scans I: Automated Aortic Segmentation, Multiplanar Reconstruction C: Manual Aortic Segmentation O: Median Dice Coefficient 0.96 (0.02)

Manual   Automated   Overlay

Multiplanar Reconstruction

Results from an example patient. TEVAR: Thoracic Endovascular Aortic Repair, CTA: Computed Tomography Angiography

Abstract

Objectives: The rising global burden of cardiovascular diseases (CV) highlights the critical need for efficiency in disease diagnosis and management. An important area for such improvement is utilization of artificial intelligence (AI) for streamlining time and resources in CV imaging workflow. We evaluate the performance of artificial intelligence (AI) segmentation for aortic segmentation on clinical computed tomography angiography (CTA) images and compare accuracy to manual methods. Such automation would markedly improve efficiency and accuracy of aortic surveillance.

Methods: This retrospective study included 27 scans from 20 patients who underwent thoracic endovascular aortic repair (TEVAR) between January 2020 and March 2022. An open-source AI model was applied to segment the aorta, and its performance was assessed by comparing AI-generated segmentations with manual segmentations using Dice similarity coefficients, volumetric analysis, and aortic dimensions. Centerline reconstructed images of thoracoabdominal aorta were processed to extract radiomic features, including maximum diameter and cross-sectional area, for analysis.

Results: The AI tool achieved a median Dice coefficient of 0.96 (0.02), indicating a high degree of concordance with manual segmentation. Multiplanar reconstruction was performed to visualize the aorta and extract measurements along its length using the automated centerline, and radiomic features, including maximum diameter and cross-sectional area, were subsequently extracted for analysis.

Conclusions: AI segmentation demonstrates strong potential for improving efficiency

and consistency in thoracoabdominal aortic segmentation, achieving high accuracy

compared to manual methods. These findings highlight the feasibility of AI integration into

clinical practice for diagnosis and surveillance of aortopathies, warranting further

validation on larger datasets to enable clinical translation.

Keywords: AI segmentation, aortic segmentation, computed tomography angiography,

artificial intelligence, radiomics, thoracic endovascular aortic repair.

Introduction

Imaging modalities play a critical role in diagnosis, surveillance, and clinical decision-making of cardiovascular (CV) pathologies, which affect a large portion of the population. Therefore, significant health system and health care provider (HCP) time and resources are dedicated to ever increasing CV imaging demands. Typically, clinicians manually outline regions of interest on images—is time-consuming, labor-intensive, and prone to variability [1-2]. At least 6 in 10 U.S. adults (61%), more than 184 million people, are expected to have some form of cardiovascular disease (CVD) within the next 30 years, contributing to a staggering $1.8 trillion in direct and indirect costs [3]. Given the prevalence of CVD and the necessity of regular imaging-based risk assessments, manual segmentation is becoming increasingly impractical for large-scale studies and clinical settings. It is time-consuming, prone to interpreter discordance and variation, increasing demands for 24 hour imaging review may introduce HCP fatigue and propensity for errors. These points underscore the critical need for reliable, automated systems to supplement and complement the current demands faced by CV radiologists and HCPs to achieve wide-scale application across our health system [4-6].

Automated segmentation offers a promising alternative, addressing the drawbacks of manual methods while maintaining, or even exceeding, their benefits, such as predictive value, risk stratification, and standardization [2]. Moreover, automated segmentation reduces observer bias, is more scalable, and requires less time [1]. For AI technologies to be implemented in clinical practice, they must be validated using real-world clinical data, as we aimed to do. One such area in CVD where clinical decision making is heavily

dependent on imaging modalities is the diagnosis and management of aortopathies.

Patients with aortic diseases also need life-long surveillance with serial imaging to monitor

for disease progression and integrity of aortic interventions. Therefore, in this study, we

assessed the potential utility of AI segmentation of thoracoabdominal aorta CT

angiography in patients undergoing complex hybrid aortic arch and descending

thoracoabdominal repairs to enable automated accurate dimensional analysis compared

to traditional, manual segmentation CT angiography analysis.

This study investigates the performance of AI segmentation tools applied to uncurated

clinical data. The primary objectives were: (1) to assess the fidelity of AI-generated aortic

segmentations against manual benchmarks, and (2) to identify the practical challenges

and opportunities for clinical adoption of AI segmentation tools.

Materials and Methods

Patient Cohort

Patients who underwent hybrid arch repair with zones 0 to 5 thoracic endovascular aortic

repair (TEVAR) at Yale School of Medicine from January 2020 to March 2022 were

included [8], contributing 27 scans.

AI Segmentation

CTA images were sourced in DICOM format and converted to NIfTI using the dicom2nifti

Python package and segmented using TotalSegmentator V2, an nnU-Net based model

designed for the segmentation of distinct anatomical structures in CT images, including

the aorta. This model had shown a high Dice similarity coefficient, measuring the similarity

between manual and automated measurements on a scale of 0 (no similarity) to 1

(identical sets), of 0.943 on diverse clinical data sets, indicating superior segmentation

accuracy. The model has been trained and validated on a dataset comprising 1,204 CT

examinations, which included images with various abnormalities and from different clinical

settings, thereby ensuring robustness and reliability [8].

Morphology Analysis

Segmentation measurements were performed using modules from the VMTK plug-in for

3DSlicer, an open-source platform for medical image informatics, image processing, and

three-dimensional visualization [9]. The Extract Centerline module was first used to

compute the centerline between endpoints placed at the aortic root and bifurcation.

Multiplanar reconstruction (MPR) was then performed to visualize the aorta in different

planes. The maximum inscribed sphere method was applied to measure aortic diameters

accurately across various sections, and the Cross-Section Analysis module was used to

generate the diameter and cross-section area of the segmentation at each point along

the centerline. The accuracy of AI-driven aortic segmentation was evaluated using the

Dice similarity coefficient, volumetric analysis, and the assessment of aortic dimensions.

Radiomic features of interest were compared against manual segmentation results for

two example patients to assess the capability of the automated method to achieve high

fidelity in anatomical replication. Dice scores were calculated using the Segment

Comparison module from the SlicerRT extension [10]. Statistics were generated using

the Segmentation Statistics module, presented as median and interquartile range (IQR).

Mann-Whitney U tests were used for comparisons, with p values less than 0.05

considered significant.

Results

Patient Cohort Characteristics

A total of 20 patients, predominantly males (70%), with a mean age of 58.5 ± 11.8 years,

who underwent hybrid arch repair with TEVAR were included in this study, with a total of

27 scans (Table 1). The majority of patients presented with type A aortic dissections

(70%), with common comorbidities such as hypertension (90%), dyslipidemia (50%), and

diabetes (30%). This patient population was chosen for the analysis because it represents

one of the most complex groups of patients to surveillance from an imaging standpoint,

and clinical decision making is heavily dependent on findings on CTA imaging.

AI segmentations achieved a median Dice coefficient of 0.96 (0.02) (Table 2). Radiomic

features obtained by AI segmentation were not significantly different from those obtained

by manual segmentation (Table 3). This close agreement indicates the robustness of AI

segmentation across varied patient anatomies.

Quantitative Assessment of Aortic Segmentation Accuracy

The Dice similarity coefficient of Patient 1 was 0.92, indicating a high degree of

concordance between the manual and AI-generated segmentations. The largest diameter

along the aorta was recorded at 49.5 mm, with a total aortic distance from the root of

145.5 mm. AI segmentation produced a total volume of 481.8 cm³, which was 78 cm³ less

than the volume determined manually. The voxel count for AI segmentation was

3,359,224, showing a reduction of 544,242 compared to the manual count.

The segmentation of Patient 2's aorta showed an improved Dice similarity score of 0.97.

The largest diameter was measured at 43.4 mm, and the total distance from the root was

68.5 mm. The AI model reported a volume of 450.2 cm³, 7 cm³ higher than the manual

segmentation. The voxel count, representing the number of 3D volume units used to

define the aorta, was 2,533,435—39,195 more than manual segmentation—indicating

greater precision. The overlays presented in Figure 1 visually substantiate the

segmentation accuracy, with AI-generated segmentations closely aligning with those

manually delineated.

Radiomic Feature Extraction from Segmented Models

Multiplanar reconstruction of the aorta was performed on Example Patient 2 using

automated segmentation techniques (Figure 2). The centerline is used as a guide for

extracting the aortic diameter at various points along its length, demonstrated in the right

panel. The automated centerline extraction, performed using the maximum inscribed

sphere method, is crucial for accurate measurement of the aortic diameter. This method

provides an advantage over the manual method, which relies on the three orthogonal

planes (sagittal, coronal, and axial) and cannot measure the aortic diameter based on the

maximum inscribed sphere method without reconstruction. A plot of the unprocessed

aortic diameter extracted from cross-sectional analysis along the centerline path.

Discussion

This study demonstrates the potential of AI to enhance the efficiency and accuracy of aortic segmentation. The high Dice coefficients observed indicate strong concordance between AI and manual segmentations. The differences in radiomic features obtained by AI and manual segmentation, e.g. volume and voxel counts, were not statistically significant. These results underscore AI's ability to streamline the segmentation process while maintaining expert-level precision, significantly reducing the time and variability associated with manual methods, and still achieving high spatial overlap, as indicated by the Dice coefficients.

Benefits of AI segmentation

AI segmentation offers several distinct advantages over manual methods. It eliminates inter-rater variability, providing consistent and reproducible results. Furthermore, it allows for high-throughput processing, enabling rapid analysis of large datasets, which is particularly useful in clinical settings with heavy workloads. AI's scalability also ensures applicability across diverse patient populations and clinical conditions. Our aforedescribed approach, validated in patients with both normal aortas and thoracoabdominal pathology, is uniquely positioned to streamline analyses of pathological aortas and patient diagnoses. Automated segmentation has been rigorously validated in pathological aortic conditions, not just healthy tissue, including cases of aneurysms, dissections, and atherosclerotic disease. This ensures its reliability and accuracy in capturing clinically relevant variations in diameter, including peaks and troughs indicating dilations and constrictions, which are critical for surgical planning and

risk assessment. These validations strengthen the clinical utility of the model, particularly

in high-risk populations (Figure 2).

Radiomics Advancements

Radiomics, supported by AI segmentation, offers detailed quantitative imaging features

that extend beyond traditional measurements. For example, the ability to compute the

maximum aortic diameter using a 3D centerline model enhances accuracy compared to

manual measurements taken on sagittal, coronal, or transverse planes, which may

misrepresent true cross-sectional diameter. Additionally, features such as shape

descriptors, intensity scores, and volumes can provide valuable insights into aortic

morphology and pathology. The high-throughput extraction of these features facilitates

their integration into machine learning models, paving the way for advanced predictive

analytics.

Limitations

This study is a proof of concept, limited by a small sample size (n=27), which restricts the

generalizability of the findings. Furthermore, integrating AI into clinical workflows poses

significant challenges, including approval from regulatory bodies, clinician training, and

compatibility with electronic health record systems. Additional validation and testing are

needed to ensure seamless operation without disrupting existing workflow. Addressing

these barriers is crucial for achieving widespread clinical adoption and ensuring effective

implementation of AI tools.

Conclusion

Pretrained, "out-of-the-box" AI tools exhibit promising accuracy in aortic segmentation from heterogeneous clinical data. These tools can potentially reduce the time and labor costs associated with manual segmentation, facilitating more detailed and rapid preoperative planning. Further studies are required to validate these findings across broader populations and to address the workflow and regulatory challenges identified.

Funding

No funding was received for this study.

Conflicts of Interest

All authors declare no financial or non-financial competing interests.

Author Contributions

Conceptualization: DC, BS, CSO, PV.

Data Curation: NTO, AAH, AA, IH, EE, MC, NP.

Formal Analysis: DC, NTO.

Investigation: DC, NTO.

Methodology: DC, BS, NTO, CSO.

Project Administration: CSO.

Resources: NP, EBS, CSO.

Software: DC, NTO.

Supervision: EBS, IH, PV, CSO.

Visualization: DC, NTO, SS.

Writing – Original Draft: DC, BS, SS.

Writing – Review & Editing: EBS, IH, PV, CSO.

Ethical Statement

This study was determined to be exempt from review by Yale University's Institutional

Review Board (IRB) under protocol number 2000032492, with Waiver of Consent/Waiver

of HIPAA Authorization included in the study protocol due to minimal risk to participants

and impracticality.

Data Availability Statement

The data underlying this article cannot be shared publicly due to the privacy of individuals

that participated in the study. Study data are available upon reasonable request to the

corresponding author, in accordance with institutional policies and any applicable data

sharing or data use agreements.

Figures

Example Patient 1
Dice: **0.92**
Largest diameter along aorta (mm): 49.5
Total Aortic Distance from root (mm): 145.5
Voxel Count: 3,359,224 [-544,242 from manual]
Volume (cm^3): 481.8 [-78 from manual]

Example Patient 2
Dice: **0.97**
Largest diameter along aorta (mm): 43.4
Total Aortic Distance from root (mm): 68.5
Voxel Count: 2,533,435 [+39,195 from manual]
Volume (cm^3): 450.2 [+7 from manual]

Manual   Automated   Overlay   Manual   Automated   Overlay

Figure 1: This figure illustrates the results of aortic segmentations from two example

patients using manual and automated methods. The left panel shows results for Example

Patient 1 and the right panel for Example Patient 2. Each panel is divided into three sub-

panels displaying the manually segmented model, the automated segmented model, and

an overlay of both, indicating comparative spatial alignment and segmentation

consistency.

Results 2: Multiplanar reconstruction (Example Patient 2)

Unprocessed aortic diameter extraction from cross-section analysis along the centerline path

Automated centerline extraction of the aorta and diameter measurement methods using maximum inscribed sphere

Figure 2: Multiplanar Reconstruction and Diameter Analysis for Example Patient 2. This

figure demonstrates the advanced capabilities of automated segmentation and

multiplanar reconstruction for detailed aortic analysis. Left Panel: Displays a 3D

reconstruction of the aorta with automated centerline extraction, using the maximum

inscribed sphere method to measure aortic diameters accurately across various sections.

Right Panel: Presents a graph of unprocessed aortic diameters, extracted from cross-

sectional analysis along the centerline. This graph delineates the changes in diameter

along the aorta's length, highlighting potential areas of pathological dilation or constriction

crucial for surgical assessment and planning.

Tables

Table 1. Patient Demographics

Characteristic	Count (%) *
Age, mean (SD), years	58.0 (11.8)
Male	14 (70%)
Black	4 (20%)
Hispanic	1 (5%)
White	15 (75%)
Height, mean (SD), cm	174.3 (10.5)
Weight, mean (SD), kg	96.0 (14.8)
Body mass index, mean (SD)	31.9 (6.0)
Diabetes	6 (30%)
Dyslipidemia	10 (50%)
Hypertension	18 (90%)
Smoking	11 (55%)
Chronic obstructive pulmonary disease	2 (12%)
Prior coronary artery bypass grafting surgery	2 (10%)
Prior aortic surgery	10 (50%)
Presenting etiology	

Aneurysm	1 (5%)
Type A aortic dissection	14 (70%)
Type A/type B aortic dissection	4 (20%)
Type B aortic dissection	1 (5%)
Acuity	
Acute	10 (50%)
Chronic	9 (45%)
Subacute	1 (5%)
Status	
Elective	8 (40%)
Emergent	4 (20%)
Urgent	8 (40%)

302   * Unless otherwise specified

303   Table 2: Comparison of AI and Manual Aortic Segmentation (Metrics)

Metric	Median (IQR)
Dice Similarity Coefficient	0.96 (0.02)
Percentage of True Positives	1.14 (7.41)
Percentage of True Negatives	98.77 (8.06)
Percentage of False Positives	0.04 (0.07)
Percentage of False Negatives	0.06 (0.32)

305   Table 3: Comparison of AI and Manual Aortic Segmentation (Radiomic Features, Median

306   (IQR))

Radiomic Feature	Manual Segmentation	AI Segmentation	p value
Voxel count	1,911,721 (1,205,663 - 3,397,613)	1,789,071 (1,141,627 - 3,113,231)	0.736
Volume (cm^3)	539.5 (419 - 619)	513.5 (420 - 602)	0.634
Surface Area (mm^2)	58,758.3 (52,136 - 69,342)	59,675.1 (51,299 - 72,394)	0.736

References

[1] C. Ozturk, D. H. Pak, L. Rosalia, D. Goswami, M. E. Robakowski, R. McKay, et al., "AI-powered multimodal modeling of personalized hemodynamics in aortic stenosis," *ArXiv Prepr. ArXiv240700535*, 2024.

[2] K. Liu, D. Zhao, L. Feng, Z. Zhang, P. Qiu, X. Wu, et al., "Unraveling phenotypic heterogeneity in stanford type B aortic dissection patients through machine learning clustering analysis of cardiovascular CT imaging," *Hellenic J. Cardiol.*, 2024.

[3] K. J. Maddox, M. S. Elkind, H. J. Aparicio, Y. Commodore-Mensah, S. D. de Ferranti, W. N. Dowd, *et al.*, "Forecasting the burden of cardiovascular disease and stroke in the United States through 2050—prevalence of risk factors and disease: a presidential advisory from the American Heart Association," *Circulation*, vol. 149, pp. e00–e00, 2024.

[4] C. Caradu, A.-L. Pouncey, E. Lakhlifi, C. Brunet, X. Bérard, and E. Ducasse, "Fully automatic volume segmentation using deep learning approaches to assess aneurysmal sac evolution after infrarenal endovascular aortic repair," *J. Vasc. Surg.*, vol. 76, no. 3, pp. 620-630.e3, 2022, doi: 10.1016/j.jvs.2022.03.891.

[5] C. Chen, C. Qin, H. Qiu, G. Tarroni, J. Duan, W. Bai, *et al.*, "Deep learning for cardiac image segmentation: a review," *Front. Cardiovasc. Med.*, vol. 7, p. 25, 2020.

[6] H. Kim, J. Jung, J. Kim, B. Cho, J. Kwak, J. Y. Jang, *et al.*, "Abdominal multi-organ auto-segmentation using 3D-patch-based deep convolutional neural network," *Sci. Rep.*, vol. 10, no. 1, p. 6204, 2020.

[7] I. Hameed, A. Ahmed, S. Pupovac, N. Nassiri, R. Assi, and P. Vallabhajosyula, "Aortic remodeling following hybrid arch repair with zone 0 to 5 thoracic endovascular aortic repairs for complex arch and descending thoracic aortic pathologies," *JTCVS Open*, vol. 17, pp. 23–36, Feb. 2024, doi: 10.1016/j.xjon.2023.12.004.

[8] J. Wasserthal, H.-C. Breit, M. T. Meyer, M. Pradella, D. Hinck, A. W. Sauter, *et al.*, "TotalSegmentator: Robust Segmentation of 104 Anatomic Structures in CT Images," *Radiol.*

Artif. Intell., vol. 5, no. 5, 2023, doi: 10.1148/ryai.230024.

[9] L. Antiga, M. Piccinelli, L. Botti, B. Ene-Iordache, A. Remuzzi, and D. A. Steinman, "An image-

based modeling framework for patient-specific computational hemodynamics," *Med. Biol.*

Eng. Comput., vol. 46, pp. 1097–1112, 2008.

[10]C. Pinter, A. Lasso, A. Wang, D. Jaffray, and G. Fichtinger, "SlicerRT: radiation therapy

research toolkit for 3D Slicer," *Med. Phys.*, vol. 39, no. 10, pp. 6332–6338, 2012.