Title

Innovative AI models for clinical decision-making: predicting blastocyst formation and quality from time-lapse embryo images up to embryonic day 3

Running title

AI-Based Blastocyst Formation Prediction

Authors

A Yanai, MD[1], A Horie, MD, PhD[1,2], A Sakurai, MD[1], S Imakita, MD[1], M Nakamura, MD[1], A Ikeda, MD[1], S Shitanaka, MD[1], T Ohara, MD[1], B Nakakita, MD[1], A Ueda, MD[1], Y Kitawaki, MD[1], Y Sagae, MD, PhD[1], A Okunomiya, MD, PhD[1], M Mandai, MD, PhD[1]

[1] Department of Gynecology and Obstetrics, Kyoto University Graduate School of Medicine, Kyoto, Japan

[2] Department of Obstetrics & Gynecology, Medical Research Institute KITANO HOSPITAL, PIIF Tazuke-kofukai, Osaka, Japan

Corresponding author:

Akihito Horie, MD, PhD,

ORCID iD: 0000-0002-0801-5835

Department of Gynecology and Obstetrics,

Kyoto University Graduate School of Medicine,

54 Shogoin Kawahara-cho, Sakyo, Kyoto 606-8507, Japan

Phone: 81-75-751-3269

Email: a_horie@kuhp.kyoto-u.ac.jp

Abstract

Study question: Can device-independent artificial intelligence (AI) models accurately predict
blastocyst formation and quality by embryonic day 3 using time-lapse images and age?

Summary answer: The AI-based prediction models achieved receiver operating
characteristic (ROC) area under the curve (AUC) ≥0.87 for predicting blastocyst formation,
good blastocysts, and poor blastocyst + arrested embryos, demonstrating high accuracy at
embryonic day 3.

What is known already: Previous blastocyst prediction models from embryonic day 3 time-
lapse images and parameters have achieved moderate accuracy (ROC AUCs: 0.73–0.82).
These models require labor-intensive manual parameter annotation or exhibit device
dependence, limiting their generalizability and clinical usefulness.

Study design, size, duration: This was a retrospective, multicenter study using 7,111 eligible
human embryos cultured in four types of time-lapse incubators across four facilities from
May 2018 to December 2022. Additionally, 23,852 time-lapse and stereomicroscopic images
of embryos during the same period were collected and annotated using 17-class supervised
classification labels for the AI auto-annotator.

Participants/materials, setting, methods: Embryo culture time-lapse images and clinical
data were retrospectively collected from Kyoto University Hospital and its cooperating
facilities. This study included 2PN embryos cultured individually per well up to embryonic
day 5 or 6. Degenerated embryos and those with poor Veeck grades (4 and 5) were excluded.
Time-lapse images were recorded every 15 minutes between 24- and 64-hours post-
insemination and preprocessed. Embryos with over 5% image loss or preprocessing errors
were excluded. The final dataset was used to construct AI models for blastocyst formation
and quality prediction. The dataset was split into 80% for training and cross-validation and
20% for independent testing. A pre-trained image classification AI (NASNet-A Large) was
fine-tuned as AI auto-annotator, using 17-class supervised classification dataset including cell
stages and Veeck grades. The AI auto-annotator converted time-lapse images into time-
series labels. These labels, the proportions of each stage and grade, and the age at egg
retrieval were analyzed using a gradient boosting algorithm (XGBoost) to predict blastocyst
formation and quality. The performance of models was evaluated using ROC and precision-
recall analyses with 5-fold validation and independent testing.

Main results and the role of chance: The AI auto-annotator achieved 85% accuracy in
classifying embryo images into 17 classes, with 95% cell stage classification accuracy. The
blastocyst formation prediction model had an ROC AUC of 0.86 for mean cross-validation
and 0.87 for test. The good blastocyst formation prediction model had a ROC AUC of 0.88 for
test. The poor blastocyst + arrested embryo prediction model had a ROC AUC of 0.87 for test
and achieved an AUC of 0.90 for precision recall, meeting the target accuracy for clinical
decision-making. The performance of models was evaluated through five-fold cross-

validation and independent testing. Subgroup analyses revealed minimal differences
between facilities and age groups, demonstrating the generalizability of the model.

Limitations, reasons for caution: The predictive models used Gardner criteria, widely
adopted but prone to inter-rater variability, potentially affecting consistency. Blastocyst
evaluations at each facility were unmodified, and single-rater evaluations were avoided to
prevent domain shift. The exclusion of embryos with defects or poor quality (Veeck grades 4
and 5) may limit the model's generalizability. Additionally, as this was a retrospective study,
prospective validation is required to confirm the findings.

Wider implications of the findings: This versatile AI model enables more accurate embryo-
by-embryo decisions by embryonic day 3, potentially guiding selection between continuing
blastocyst culture versus early transfer/freezing, and thus improving individualized patient
care in ART.

Study funding/competing interest(s): This work was supported by JSPS KAKENHI grant
number JP23K08821. The authors have no conflicts of interest to declare.

Trial registration number: N/A.

Keywords

Capsule: Artificial intelligence models accurately predicted blastocyst formation and quality

using time-lapse images and age on embryonic day 3, supporting clinical decision-making

regarding blastocyst culture or early embryo transfer.

Introduction

Assisted reproductive technology (ART) plays an important role in infertility treatment. In 2021, 9% of all births in Japan resulted from ART (Katagiri *et al.*, 2024). Within ART, there are two main timings for embryo transfer: cleavage-stage transfer (on the embryonic day 2 or 3) and blastocyst-stage transfer (on the embryonic day 5 or 6). In patients with a good prognosis, blastocyst-stage transfer is usually preferred owing to higher pregnancy and birth rates per transfer (Cameron *et al.*, 2020; Glujovsky *et al.*, 2022). According to the Japanese National Registry of ART ("the Japan Society of Obstetrics and Gynecology ART Databook 2021," n.d.; Katagiri *et al.*, 2024), 73% of transfer cycles in 2021 were blastocyst-stage transfers, and the pregnancy rate per embryo transfer was double that of cleavage-stage transfers (41% vs. 20%, respectively). Conversely, in patients with poor prognoses and few available embryos (e.g., over 40 years of age), prolonged culture sometimes fails to produce transferable blastocysts. Several reports (Vlaisavljević *et al.*, 2001; Kovačič *et al.*, 2002; De Croo *et al.*, 2022) showed that extending culture to the blastocyst stage did not improve implantation or cumulative pregnancy rates in certain scenarios. Additionally, one report (Xiao *et al.*, 2019) pointed out that the live birth rate of cleavage-stage embryo transfer was double that of blastocyst culture in patients with only one available embryo on embryonic day 3. Thus, cleavage-stage embryo transfer remains an important option for patients with a poor prognosis, such as older patients. In Japan, where donor eggs are difficult to obtain, 47% of patients aged >43 years undergo cleavage−stage transfer (Katagiri *et al.*, 2024).

In previous studies (Glujovsky *et al.*, 2022), the choice between early embryo freezing/transfer and blastocyst transfer was made on a patient-by-patient basis rather than an embryo-by-embryo basis. Embryo-by-embryo optimal selection requires accurate

prediction of morphological quality on embryonic day 5 or 6 by embryonic day 3. Although

one report (Jiang *et al.*, 2022) showed that static morphological assessment on embryonic

day 3 could predict blastocyst formation, the receiver operating characteristic area under

the curve (ROC AUC) of 0.78 was considered insufficient for reliable clinical decision-making.

Since 2008, time-lapse embryo culture (Leung *et al.*, 2016) has enabled the evaluation of

morphokinetic parameters, leading to multiple studies reporting predictions of blastocyst

formation (Wong *et al.*, 2010; Cruz *et al.*, 2012; Dal Canto *et al.*, 2012; Chamayou *et al.*,

2013; Conaghan *et al.*, 2013; Athayde Wirka *et al.*, 2014; Milewski *et al.*, 2015; Yang *et al.*,

2015; Motato *et al.*, 2016; Mizobe *et al.*, 2018; Zaninovic *et al.*, 2019) and good blastocyst

formation (Kirkegaard *et al.*, 2013; Cetinkaya *et al.*, 2015; Storr *et al.*, 2015; Yang *et al.*, 2015;

Mizobe *et al.*, 2018) using these parameters. While day 4 predictions of blastocyst formation

achieved validated AUCs as high as 0.85 (Motato *et al.*, 2016), clinically useful predictions at

day 3 remained lower, with AUCs of 0.73–0.81 (Milewski *et al.*, 2015; Zaninovic *et al.*, 2019).

Recent advances have included artificial intelligence (AI) analysis of time-lapse images (Tran

et al., 2019; Huang *et al.*, 2021; Lee *et al.*, 2021; Liao *et al.*, 2021; Barnes *et al.*, 2023), with

one reported AI-based model achieving an AUC of 0.82 for predicting blastocyst formation

on embryonic day 3 (Liao *et al.*, 2021). However, previously reported predictive models

(Milewski *et al.*, 2015; Zaninovic *et al.*, 2019; Liao *et al.*, 2021; Jiang *et al.*, 2022) face several

limitations. First, the prediction accuracy was insufficient for reliable clinical decisions

regarding whether extended blastocyst culture or early embryo transfer/freezing on

embryonic day 3. Next, the time-lapse parameters required laborious manual annotation

and exhibited interrater variability (Sundvall *et al.*, 2013). Last, AI prediction models often

lacked interpretability ("black box" issue) and were highly device-dependent, as they were

trained using images from only one type of time-lapse incubator (Tran *et al.*, 2019; Huang *et*

al., 2021; Lee *et al.*, 2021; Liao *et al.*, 2021; Barnes *et al.*, 2023).

This study aimed to develop an accurate AI blastocyst prediction model on embryonic day 3

with high versatility and interpretability for clinical use. To achieve this objective, we

analyzed time-lapse images acquired from multiple time-lapse incubators at multiple

facilities, constructed models from human-interpretable features, and visualized the AI

decision basis using SHapley Additive exPlanations (SHAP) (Lundberg and Lee, 2017;

Lundberg *et al.*, 2018).

Materials and Methods

Overview of the Blastocyst Prediction using AI

An overview is shown in Figure 1. We created a dataset for the blastocyst prediction model

(Figure 1A). We fine-tuned the ImageNet-1k pre-trained image analysis AI (NASNet-A Large

(Zoph *et al.*, 2018)), using a 17-class embryo image classification dataset to build an AI auto-

annotator (Figure 1B). AI extracted features from time-lapse images, and the age at oocyte

retrieval was analyzed using a gradient boosting method (XGBoost (Chen and Guestrin,

2016)) to build each prediction model for blastocyst formation, good blastocyst formation,

and poor blastocyst + arrested embryos (Figure 1E). Our AI blastocyst prediction model

predicted embryonic morphology on embryonic day 5 or 6 based on embryo culture time-

lapse images and age at oocyte retrieval.

Fine-Tuning the AI Auto-Annotator

A 17-class embryo image classification dataset (Figure 2A) was created from 23,852 embryo

images with 17-class supervised labels, including 1-cell stage with or without pronuclei (PN)

and 2-, 3-, 4-, 5–7-, and 8-cell stages with Veeck grades 1/2/3 (Veeck, 1988). The ImageNet-

1k pre-trained image AI, NASNet-A Large (Zoph *et al.*, 2018), was fine-tuned on this dataset

to create an AI auto-annotator (Supplementary Text 1). Time-lapse images were analyzed

using the AI auto-annotator and classified into 17-class morphological labels. The

proportions of each cell stage and Veeck grade were calculated in each embryo as features.

The Veeck grade was replaced by the mode frequency (or a better grade if the mode

frequency was the same) at each cell stage for each embryo.

Data Collection and Preparation for the Blastocyst Prediction Dataset

Embryo culture time-lapse images and clinical data recorded in daily practice were

retrospectively collected from Kyoto University Hospital and its cooperating facilities from

May 2018 to December 2022 (Supplementary Text 1 and Figure 1C).

This study included 2PN embryos with cleavage that were time-lapse cultured, with at least

embryonic day 5, single embryo per well. Degenerated and poor embryos with Veeck grades

4 and 5 were excluded. Time-lapse images were collected every 15 min at 24–64 hours post-

insemination (hpi). Time-lapse images were collected up to 64 hpi because most day 3

embryo transfer/freezing were performed after 64 hpi. Time-lapse videos were

preprocessed and converted to JPG images of 331 × 331 pixels. Embryos with more than 5%

image loss were excluded due to missing data, debris, misplacement or significant

preprocessing errors. (Supplementary Text 1).

The blastocyst prediction dataset was created by combining the time-lapse images at each

time point with the morphological evaluation on embryonic day 5 or 6, age at oocyte

retrieval, and facility.

Overall, 8,043 fertilized embryos cultured in four types of time-lapse incubators in 2,792

oocyte retrieval cycles at four facilities were included. After exclusions, 7,111 embryos

remained. The dataset was randomly split into 80% for training and cross-validation and 20%

(1,422 embryos) for testing (Figure 1A).

Blastocysts were evaluated for Gardner stage and grade; blastocyst formation was defined

as Gardner stage ≥3, and arrested embryos were defined as others. Good blastocysts were

defined as Gardner grades AA, AB, BA, and BB; poor blastocysts as CC; and others as

moderate blastocysts. The morphological evaluation is shown in Figure 1D.

Building the AI Blastocyst Prediction Models

The AI auto-annotator was used to classify the time-lapse images at each time point into

morphological labels, and the proportions of cell stages and Veeck grades were calculated.

Time-series labels, proportions, and ages at oocyte retrieval were analyzed using XGBoost

and models were built from the training and cross-validation dataset (Supplementary Text 1).

The targets for prediction were blastocyst formation, good blastocyst formation, and poor

blastocyst + arrested embryos. The models were five-fold cross-validated and tested on a

test dataset. The endpoint was ROC AUC and the AUC for precision-recall (PR). For clinical

decision-making, a target PR AUC of poor blastocyst + arrested embryos prediction was set

at 0.90. Subgroup analyses were performed by facility and age group (<35, 35−39, and ≥40

years).

Blastocyst Prediction Models from Single or Few Embryos Images

To examine the superiority of time-lapse image analysis over static image analysis,

prediction models were created from a single image at 24, 48, or 64 hpi, or three images at

24, 48, and 64 hpi, assuming a once-a-day observation, and ROC AUC values were compared.

In the first type of model, the AI auto-annotator classified static images into morphological

annotation labels, and XGBoost built prediction models from single or three labels and age.

For the next type of model, the fine-tuned NASNet-A Large was trained on single images

(grayscale, 331 × 331 pixels) at 24, 48, and 64 hpi in the training and cross-validation dataset

to create prediction models. Additionally, because NASNet-A Large is a color image analysis

AI that allows up to three channels of input for each image, we created merged images using

24-, 48-, and 64-hpi images for each channel and trained NASNet-A Large on the merged

images to create the prediction models.

SHAP-Based Model Interpretation

In our prediction models, the AI auto-annotator classified embryo images into human-

recognizable morphological labels, and all features inputted to XGBoost were interpretable.

However, because XGBoost has dozens to hundreds of decision trees, it is difficult to directly

interpret the basis for the prediction models. Therefore, we used SHAP (Lundberg and Lee,

2017; Lundberg *et al.*, 2018) to analyze the contribution and predictive direction of each

feature in the test dataset. Because SHAP values are positive and negative for positive and

negative predictions, respectively, we calculated the absolute mean of SHAP values for each

feature and extracted the highly contributing features. The correlation between the value

and SHAP value of each contributing feature was visualized and used to interpret the models.

Statistical Analyses

The model performance was evaluated using ROC and PR analyses and AUCs. The analyses

were performed using numpy 1.21.6, scipy 1.7.1, and sklearn 1.0.2 on Python 3.7.15. In the

comparison of blastocyst prediction models from time-lapse images and those from single or

few embryo images, ROC AUCs were compared, and p-values were calculated using the

DeLong test. The DeLong test was performed using the fast DeLong method in Python

(https://github.com/yandexdataschool/roc_comparison) (DeLong *et al.*, 1988; Sun and Xu,

2014). The significance level was set at $p<0.05$; however, because of nine-fold tests, the

significance level was set at $p<0.0055$ with Bonferroni correction.

Ethical Considerations

This study was conducted in accordance with the principles embodied in the Declaration of

Helsinki and was approved by the Ethics Committee of Kyoto University Graduate School and

Faculty of Medicine (approval number R3320). Informed consent was obtained in the form

of opt-out through the websites of facilities.

Results

AI Auto-Annotator Performance

We fine-tuned ImageNet-1k pre-trained image AI NASNet-A Large with the 17-class embryo

image classification dataset (Figures 1B and 2A). The training was early stopped at 30 epochs

(Figure 2B and C), and the test accuracy was 85%. The confusion matrix (Figure 2D) showed

that most misclassifications were due to errors in the Veeck grade, the cell stage accuracy

was excellent at 95% (Supplementary Figure 1).

The distribution of AI annotation labels at each time point in the blastocyst prediction

dataset showed the progression of cleavage from two to eight cells (Figure 2E). At 24 hpi, the

percentage of annotations misclassified as four or more cells owing to fragmentation or

other factors was small (2.0%).

Major time-lapse parameters were calculated from the AI annotation labels and compared

with those in previous reports (Cruz *et al.*, 2012; Dal Canto *et al.*, 2012; Milewski *et al.*, 2015;

Motato *et al.*, 2016; Canosa *et al.*, 2020) (Supplementary Table 1). AI-calculated t2 (time to

the 2-cell stage) was significantly smaller in the blastocyst formation group than in the

arrested embryo group (median: 24.8 vs. 27.0 hpi, $p<0.001$) and was within the mean or

median values in previous reports.

Blastocyst Formation Prediction Model

Using XGBoost and recursive feature elimination, 88 features were selected for the

blastocyst formation prediction model (Supplementary Table 2). The model achieved a test

AUC of 0.87 and a five-fold cross-validation mean AUC of 0.86 (Figure 3A). With a Youden's

index (Youden, 1950) threshold of 0.49, the test accuracy reached 0.79. These results

demonstrated the model's robustness in cross-validation and testing.

Good Blastocyst Formation Prediction Model

The good blastocyst formation prediction model achieved a test ROC AUC of 0.88 using 156

features. Owing to the low proportion of good blastocysts (28%), the test PR AUC was

relatively lower than other models (Figure 3B).

Poor Blastocyst + Arrested Embryo Prediction Model

The poor blastocyst + arrested embryo prediction model (Figure 3C), using 76 features,

achieved a test AUC of 0.87. Its PR AUC reached 0.90, indicating superior positive predictive

accuracy compared to other models and meeting the target performance criteria. Designed

as a clinical tool, this model assists in selecting early embryo transfer or freezing on

embryonic day 3 instead of blastocyst culture. Adjusting the decision threshold further

personalizes the clinical application, enabling a focus on achieving a 90% positive predictive

value (PPV) at 60% sensitivity or an 80% PPV at 80% sensitivity (Figure 3D).

Subgroup Analyses

Subgroup analyses by facility for all models indicated no large differences between facilities,

and the blastocyst formation prediction model was able to predict with an accuracy of AUC

≥0.83 for each facility (Figure 4A and Supplementary Figure 2A). Including age at oocyte

retrieval as a feature, subgroup analyses by age group showed small differences between

groups, and predictions were possible irrespective of the age group (Figure 4B and

Supplementary Figure 2B). These findings highlight the robustness of our models.

Comparison of Time-Lapse and Limited-Image Models

The AUC of the blastocyst formation prediction model built using XGBoost based on AI auto-

annotated labels from single or three images and age was 0.63–0.78; however, the AUC of

the model based on time-lapse image analysis was significantly higher, at 0.87 ($p<1.0\times10^{-5}$

for all comparisons; Figure 5A). The blastocyst formation prediction models based on image-

only analysis by NASNet-A Large had an AUC of 0.64–0.76, and the prediction model based

on time-lapse image analysis had a significantly higher AUC, at 0.87 ($p<1.0\times10^{-5}$ for all

comparisons) (Figure 5B). The models based on time-lapse image analysis were also

significantly better ($p<1.0\times10^{-5}$ for all comparisons) for good blastocyst formation prediction

and poor blastocyst + arrested embryo prediction (Supplementary Figure 3).

Feature Contribution and SHAP Analysis

The top 30 features for each model, ranked by absolute mean SHAP values (Figure 6A),

consistently included proportions (2-cell stage, 4-cell stage, Veeck grade 2, etc.) and age at

oocyte retrieval. While the time points of high contribution varied across models, a wide

range of features within the 24–64 hpi broadly contributed to the models.

The correlation between annotation labels and SHAP values (Figure 6B) at each time point in

the blastocyst formation prediction model indicated that whether each feature contributes

to a positive or negative prediction often varies at a single point because XGBoost deals with

the 17-class labels as discrete-value variables rather than categorical variables. The 5–7-cell

stage of grade 1/2 or 8-cell stage at 62.75 hpi, 2-cell stage or more at 29.75 hpi, and 4-cell

stage or less at 41.75 hpi were found to contribute to positive predictions. Focusing on 41.75

hpi, the 5-cell or more stage due to overestimation of poor-grade embryos contributed to

negative predictions, whereas the 3-cell or less stage, which was the most frequently

arrested stage, incorrectly contributed to positive predictions. In contrast, at 42.75 hpi, the

3-cell or less stage contributed appropriately to negative predictions, and the entire model

worked well.

The strong contribution of a very small proportion of 2-cell stage to negative predictions is

thought to represent abnormal direct cleavage.

Although our AI auto-annotator occasionally overestimated the cell stage in poor-grade

embryos, embryos with abnormally high proportion of 8-cell had negative SHAP values. Thus,

these blastocyst prediction models were able to accurately evaluate even poor-grade

embryos with morphological misclassification.

Discussion

Main Findings

In this study, 24–64-hpi time-lapse images recorded at various incubators and ages at oocyte

retrieval from multiple centers were analyzed using AI to build models for predicting

blastocyst morphology after the embryonic day 5. The AUCs for the blastocyst formation,

good blastocyst formation, and poor blastocyst + arrested embryo prediction models were

0.87, 0.88, and 0.87, respectively. The positive predictive accuracy of the poor blastocyst +

arrested embryo prediction model, which is envisioned as a clinical tool for deciding whether

blastocyst culture should be continued, or early embryo freezing/transfer should be

performed, achieved a PR AUC of 0.90, meeting the target accuracy.

Performance and Versatility of the AI Auto-Annotator

The newly developed AI auto-annotator offers several advantages over previously published

models. In particular, it is compatible with various types of incubators, attains the highest

level of cell stage classification accuracy without labor-intensive manual annotation, and

incorporates additional qualitative evaluations, including pronuclearity and Veeck grade. Our

cell stage accuracy of 95% compares favorably with previous studies (Khan *et al.*, 2016;

Dirvanauskas *et al.*, 2019; Raudonis *et al.*, 2019; Leahy *et al.*, 2020; Danardono *et al.*, 2022).

This performance was particularly remarkable considering that we included fragmented

embryos (Veeck grade 3) and more difficult-to-classify stages such as 3- and 5-7-cell. A

previous study reported a slightly higher accuracy (Dirvanauskas *et al.*, 2019), but their

model included fewer classification categories and a limited set of embryos, which may not

reflect real-world diversity. In contrast, our approach includes a broader range of embryo

states, including those with fragmentation, making it more generalizable to clinical practice.

Accuracy and Generalizability of the Prediction Models

We implemented versatile image pre-processing and classified time-lapse images at each

time point using the highly accurate AI auto-annotator, enabling detailed analysis of stage

and grade, calculation of parameters into robust prediction models.

Whereas prior day 3 blastocyst formation prediction models achieved validated ROC AUCs

ranging from 0.73 to 0.82 (Milewski *et al.*, 2015; Zaninovic *et al.*, 2019; Liao *et al.*, 2021; Jiang

et al., 2022), our model demonstrated superior performance with an AUC of 0.87. Similarly,

our good blastocyst formation model achieved an ROC AUC of 0.88, improving on reported

unvalidated AUCs of 0.73–0.82 (Kirkegaard *et al.*, 2013; Cetinkaya *et al.*, 2015; Storr *et al.*,

2015) . Moreover, we successfully created a poor blastocyst + arrested embryo model with a

PR AUC of 0.90, indicating its utility for identifying embryos unlikely to develop further. This

prediction can guide clinicians in deciding whether to proceed with day 3 transfer or freezing

the embryo rather than risk further culture.

Most previous models relied on time-lapse parameters or images captured within 24 hpi

(Wong *et al.*, 2010; Cruz *et al.*, 2012; Dal Canto *et al.*, 2012; Chamayou *et al.*, 2013;

Conaghan *et al.*, 2013; Kirkegaard *et al.*, 2013; Athayde Wirka *et al.*, 2014; Cetinkaya *et al.*,

2015; Storr *et al.*, 2015; Yang *et al.*, 2015; Milewski *et al.*, 2015; Mizobe *et al.*, 2018;

Zaninovic *et al.*, 2019; Liao *et al.*, 2021). However, our study included only images of the

embryos from 24 to 64 hpi, which are routinely recorded at most facilities, ensuring broader

applicability while maintaining superior accuracy.

In contrast to previous AI models, which were device-dependent (Tran *et al.*, 2019; Huang *et*

al., 2021; Lee *et al.*, 2021; Liao *et al.*, 2021), our study analyzed images from four types of

incubators across four facilities. By using a highly adaptable pre-processing method and a

versatile AI auto-annotator, our models effectively managed variations such as aspect ratios,

embryo area percentages, imaging contrasts, resolutions, and well geometries.

Data were collected irrespective of oocyte retrieval age, insemination method, or culture

medium, creating heterogeneity across cases. Nonetheless, subgroup analyses showed

minimal differences in AUC across facilities and age groups, confirming the model's

generalizability. These findings suggest that our models are versatile, robust, and suitable for

diverse clinical settings.

Interpretability of the prediction models

XGBoost outperformed Long Short-Term Memory models (data not shown) by analyzing

features in parallel, effectively capturing temporal aspects of morphokinetic development

from 24–64 hpi. SHAP analysis revealed how features, including human-recognizable

parameters and embryo age, contributed to predictions. Importantly, localized

misclassifications, such as overestimated cell stages in fragmented embryos, did not

compromise overall model accuracy due to integration with more reliable features.

By restricting features to interpretable parameters, our approach ensures transparency and

addresses the "black box" limitation often associated with AI. SHAP values allow clinicians to

visualize individual embryo predictions, providing a clear understanding of the rationale

behind the model's decisions. This interpretability fosters clinician confidence, accelerates

regulatory approval, and supports the clinical application of AI tools in embryo assessment.

Limitations

Some limitations should be considered. First, we relied on the Gardner criteria (Gardner,

1999; Gardner *et al.*, 2000) for blastocyst grading. While widely used, these criteria are

subjective and prone to inter-rater variability (Storr *et al.*, 2017). To minimize domain shift,

grading was conducted independently at each facility without standardization.

Second, embryos with missing records, misplacement, or poor-quality images were excluded,

limiting absolute generalizability. However, this approach ensured a focus on clinically

relevant embryos, enhancing practical applicability.

Fnally, clinical heterogeneity in insemination methods, culture media, and patient

demographics may have introduced variability. While our multicenter design mitigated these

effects to some extent, further validation in prospective cohorts is warranted.

Despite these limitations, the study demonstrates the potential of AI models to advance

embryo assessment in diverse clinical settings.

Conclusions

In this study, we developed an interpretable AI model capable of accurately predicting

blastocyst morphology based on time-lapse images taken between 24 and 64 hpi and

maternal age. The model's compatibility with various types of incubators ensures its broad

applicability across diverse clinical settings. Notably, the poor blastocyst + arrested embryo

prediction model offers significant clinical value, providing actionable insights for

determining whether to continue blastocyst culture or perform early embryo freezing or

transfer on day 3. By enabling embryo-specific decisions, this tool has the potential to

optimize in vitro fertilization workflows, enhance individualized care, and improve patient

outcomes.

Data availability: The AI code and required weights for this study are available at GitHub

(https://github.com/KU-ObGy-ART/blastocyst-prediction/). Owing to the nature of this

research, the participants of this study did not agree that their data, including time-lapse

images, should be shared publicly. Thus, images and clinical features of the fine-tuning and

blastocyst prediction datasets are not available.

Acknowledgments

The authors thank Dr. Aisaku Fukuda at the IVF Osaka Clinic, Dr. Kimihiko Nakamura at the

Amanogawa Ladies Clinic, and Dr. Masahide Shiotani and Noritoshi Enatsu at the Hanabusa

Women's Clinic for generously providing the time-lapse images and clinical data.

Author's roles

A Yanai: Conceptualization, Methodology, Software, Investigation, Resources, Writing -

Original Draft, Visualization, Formal analysis, Data Curation, Writing - Review & Editing. **A**

Horie: Conceptualization, Methodology, Writing - Review & Editing, Project administration.

A Sakurai: Investigation, Writing - Review & Editing. **S Imakita:** Investigation, Writing -

Review & Editing. **M Nakamura:** Investigation, Writing - Review & Editing. **A Ikeda:**

Investigation, Writing - Review & Editing. **S Shitanaka:** Validation, Writing - Review &

Editing. **T Ohara:** Validation, Writing - Review & Editing. **B Nakakita:** Conceptualization,

Methodology, Writing - Review & Editing. **A Ueda:** Methodology, Software, Resources,

Writing - Review & Editing. **Y Kitawaki:** Investigation, Writing - Review & Editing. **Y Sagae:**

Methodology, Funding acquisition, Writing - Review & Editing. **A Okunomiya:** Methodology,

Resources, Writing - Review & Editing, **M Mandai:** Supervision, Writing - Review & Editing.

Funding

This work was supported by JSPS KAKENHI grant number JP23K08821.

Conflict of interest

None declared.

References

Athayde Wirka K, Chen AA, Conaghan J, Ivani K, Gvakharia M, Behr B, Suraj V, Tan L, Shen S.

Atypical embryo phenotypes identified by time-lapse microscopy: High prevalence and

association with embryo development. *Fertil Steril* 2014;**101**:1637–1648.e5.

Barnes J, Brendel M, Gao VR, Rajendran S, Kim J, Li Q, Malmsten JE, Sierra JT, Zisimopoulos P,

Sigaras A, *et al.* A non-invasive artificial intelligence approach for the prediction of

human blastocyst ploidy: a retrospective model development and validation study.

Lancet Digit Heal 2023;**5**:e28–e40.

Cameron NJ, Bhattacharya S, McLernon DJ. Cumulative live birth rates following blastocyst-

versus cleavage-stage embryo transfer in the first complete cycle of IVF: A population-

based retrospective cohort study. *Hum Reprod* 2020;**35**:2365–2374.

Canosa S, Bergandi L, Macrì C, Charrier L, Paschero C, Carosso A, Segni N Di, Silvagno F,

Gennarelli G, Benedetto C, *et al.* Morphokinetic analysis of cleavage stage embryos and

assessment of specific gene expression in cumulus cells independently predict human

embryo development to expanded blastocyst: a preliminary study. *J Assist Reprod*

Genet 2020;**37**:1409–1420.

Cetinkaya M, Pirkevi C, Yelke H, Colakoglu YK, Atayurt Z, Kahraman S. Relative kinetic

expressions defining cleavage synchronicity are better predictors of blastocyst

formation and quality than absolute time points. *J Assist Reprod Genet* 2015;**32**:27–35.

Chamayou S, Patrizio P, Storaci G, Tomaselli V, Alecci C, Ragolia C, Crescenzo C, Guglielmino

A. The use of morphokinetic parameters to select all embryos with full capacity to

implant. *J Assist Reprod Genet* 2013;**30**:703–710.

Chen T, Guestrin C. XGBoost: A scalable tree boosting system. *Proc ACM SIGKDD Int Conf*

Knowl Discov Data Min 2016;**13**-**17**-**Augu**:785–794.

Conaghan J, Chen AA, Willman SP, Ivani K, Chenette PE, Boostanfar R, Baker VL, Adamson GD,

Abusief ME, Gvakharia M, *et al.* Improving embryo selection using a computer-

automated time-lapse image analysis test plus day 3 morphology: results from a

prospective multicenter trial. *Fertil Steril* 2013;**100**:412–419.e5.

Croo I De, Colman R, Sutter P De, Stoop D, Tilleman K. No difference in cumulative live birth

rates between cleavage versus blastocyst transfer in patients with four or fewer

zygotes: results from a retrospective study. *Hum Reprod Open* 2022;**2022**:1–9.

Cruz M, Garrido N, Herrero J, Pérez-Cano I, Muñoz M, Meseguer M. Timing of cell division in

human cleavage-stage embryos is linked with blastocyst formation and quality. *Reprod*

Biomed Online 2012;**25**:371–381.

Dal Canto M, Coticchio G, Mignini Renzini M, Ponti E De, Novara PV, Brambillasca F, Comi R,

Fadini R. Cleavage kinetics analysis of human embryos predicts development to

blastocyst and implantation. *Reprod Biomed Online* 2012;**25**:474–480

Danardono GB, Erwin A, Purnama J, Handayani N, Polim AA, Boediono A, Sini I. A

Homogeneous Ensemble of Robust Pre-defined Neural Network Enables Automated

Annotation of Human Embryo Morphokinetics. *J Reprod Infertil* 2022;**23**:250–256.

DeLong ER, DeLong DM, Clarke-Pearson DL. Comparing the areas under two or more

correlated receiver operating characteristic curves: a nonparametric approach.

Biometrics 1988;**44**:837–845.

Dirvanauskas D, Maskeliunas R, Raudonis V, Damasevicius R. Embryo development stage

prediction algorithm for automated time lapse incubators. *Comput Methods Programs*

Biomed 2019;**177**:161–174.

Gardner DK. In-vitro culture of human blastocyst. *Towar Reprod Certain Infertil Genet*

beyond 1999 1999;, p. 378–388. Parthenon Press.

Gardner DK, Lane M, Stevens J, Schlenker T, Schoolcraft WB. Blastocyst score affects

implantation and pregnancy outcome: Towards a single blastocyst transfer. *Fertil Steril*

2000;**73**:1155–1158.

Glujovsky D, Quinteiro Retamar AM, Alvarez Sedo CR, Ciapponi A, Cornelisse S, Blake D.

Cleavage-stage versus blastocyst-stage embryo transfer in assisted reproductive

technology. *Cochrane database Syst Rev* 2022;**5**:CD002118.

Huang B, Tan W, Li Z, Jin L. An artificial intelligence model (euploid prediction algorithm) can

predict embryo ploidy status based on time-lapse data. *Reprod Biol Endocrinol*

2021;**19**:1–10.

Jiang X, Cai J, Liu L, Liu Z, Wang W, Chen J, Yang C, Geng J, Ma C, Ren J. Does conventional

morphological evaluation still play a role in predicting blastocyst formation? *Reprod Biol*

Endocrinol 2022;**20**:1–12.

Katagiri Y, Jwa SC, Kuwahara A, Iwasa T, On M, Kato K, Kishi H, Kuwabara Y, Taniguchi F,

Harada M, *et al.* Assisted reproductive technology in Japan: A summary report for 2021

by the Ethics Committee of the Japan Society of Obstetrics and Gynecology. *Reprod*

Med Biol 2024;**23**:e12552.

Khan A, Gould S, Salzmann M. Deep convolutional neural networks for human embryonic cell

counting. *Lect Notes Comput Sci (including Subser Lect Notes Artif Intell Lect Notes*

Bioinformatics) 2016;**9913 LNCS**:339–348.

Kirkegaard K, Kesmodel US, Hindkjær JJ, Ingerslev HJ. Time-lapse parameters as predictors of

blastocyst development and pregnancy outcome in embryos from good prognosis

patients: A prospective cohort study. *Hum Reprod* 2013;**28**:2643–2651.

Kovačič B, Vlaisavljevi V, Reljič M, Gavri Lovrec V. Clinical outcome of day 2 versus day 5

transfer in cycles with one or two developed embryos. *Fertil Steril* 2002;**77**:529–536.

Leahy BD, Jang W-D, Yang HY, Struyven R, Wei D, Sun Z, Lee KR, Royston C, Cam L, Kalma Y,

et al. Automated Measurements of Key Morphological Features of Human Embryos for

IVF. *Med Image Comput Comput Assist Interv* 2020;**12265**:25–35.

Lee CI, Su YR, Chen CH, Chang TA, Kuo EES, Zheng WL, Hsieh WT, Huang CC, Lee MS, Liu M.

End-to-end deep learning for recognition of ploidy status using time-lapse videos. *J*

Assist Reprod Genet 2021;**38**:1655–1663.

Leung AS, Son WY, Dahan MH. Time-lapse imaging of embryos: current evidence supporting

its use. *Expert Rev Med Devices* 2016;**13**:881–883.

Liao Q, Zhang Q, Feng X, Huang H, Xu H, Tian B, Liu J, Yu Q, Guo N, Liu Q, *et al.* Development

of deep learning algorithms for predicting blastocyst formation and quality by time-

lapse monitoring. *Commun Biol* 2021;**4**:415.

Lundberg SM, Erion GG, Lee S-I. Consistent Individualized Feature Attribution for Tree

Ensembles. 2018;Available from: http://arxiv.org/abs/1802.03888.

Lundberg SM, Lee SI. A unified approach to interpreting model predictions. *Adv Neural Inf*

Process Syst 2017;4766–4775.

Milewski R, Kuć P, Kuczyńska A, Stankiewicz B, Łukaszuk K, Kuczyński W. A predictive model

for blastocyst formation based on morphokinetic parameters in time-lapse monitoring

of embryo development. *J Assist Reprod Genet* 2015;**32**:571–579.

Mizobe Y, Tokunaga M, Oya N, Iwakiri R, Yoshida N, Sato Y, Onoue N, Ezono Y. Synchrony of

the first division as an index of the blastocyst formation rate during embryonic

development. *Reprod Med Biol* 2018;**17**:64–70.

Motato Y, los Santos MJ de, Escriba MJ, Ruiz BA, Remohí J, Meseguer M. Morphokinetic

analysis and embryonic prediction for blastocyst formation through an integrated time-

lapse system. *Fertil Steril* 2016;**105**:376–384.e9.

Raudonis V, Paulauskaite-Taraseviciene A, Sutiene K, Jonaitis D. Towards the automation of

early-stage human embryo development detection. *Biomed Eng Online* 2019;**18**:1–20.

BioMed Central.

Storr A, Venetis CA, Cooke S, Kilani S, Ledger W. Inter-observer and intra-observer

agreement between embryologists during selection of a single Day 5 embryo for

transfer: A multicenter study. *Hum Reprod* 2017;**32**:307–314.

Storr A, Venetis CA, Cooke S, Susetio D, Kilani S, Ledger W. Morphokinetic parameters using

time-lapse technology and day 5 embryo quality: a prospective cohort study. *J Assist*

Reprod Genet 2015;**32**:1151–1160.

Sun X, Xu W. Fast Implementation of DeLong's Algorithm for Comparing the Areas Under

Correlated Receiver Operating Characteristic Curves. *IEEE Signal Process Lett* [Internet]

2014;**21**:1389–1393.

Sundvall L, Ingerslev HJ, Breth Knudsen U, Kirkegaard K. Inter- and intra-observer variability

of time-lapse annotations. *Hum Reprod* 2013;**28**:3215–3221.

the Japan Society of Obstetrics and Gynecology ART Databook 2021Available from:

https://www.jsog.or.jp/activity/art/2021_JSOG-ART.pdf.

Tran D, Cooke S, Illingworth PJ, Gardner DK. Deep learning as a predictive tool for fetal heart

pregnancy following time-lapse incubation and blastocyst transfer. *Hum Reprod*

2019;**34**:1011–1018.

Veeck LL. Oocyte Assessment and Biological Performance. *Ann N Y Acad Sci* 1988;**541**:259–

274.

Vlaisavljević V, Kovačič B, Reljič M, Lovrec VG, Sajko MČ. Is there any benefit from the

culture of a single oocyte to a blastocyst-stage embryo in unstimulated cycles? *Hum*

Reprod 2001;**16**:2379–2383.

Wong CC, Loewke KE, Bossert NL, Behr B, Jonge CJ De, Baer TM, Pera RAR. Non-invasive

imaging of human embryos before embryonic genome activation predicts development

to the blastocyst stage. *Nat Biotechnol* 2010;**28**:1115–1121. Nature Publishing Group.

Xiao JS, Healey M, Talmor A, Vollenhoven B. When only one embryo is available, is it better

to transfer on Day 3 or to grow on? *Reprod Biomed Online* 2019;**39**:916–923.

Yang ST, Shi JX, Gong F, Zhang SP, Lu CF, Tan K, Leng LZ, Hao M, He H, Gu YF, *et al.* Cleavage

pattern predicts developmental potential of day 3 human embryos produced by IVF.

Reprod Biomed Online 2015;**30**:625–634.

Youden WJ. Index for rating diagnostic tests. *Cancer* 1950;**3**:32–35.

Zaninovic N, Nohales M, Zhan Q, los Santos ZMJ de, Sierra J, Rosenwaks Z, Meseguer M. A

comparison of morphokinetic markers predicting blastocyst formation and implantation

potential from two large clinical data sets. *J Assist Reprod Genet* 2019;**36**:637–646.

Zoph B, Vasudevan V, Shlens J, Le Q V. Learning Transferable Architectures for Scalable

Image Recognition. *2018 IEEE/CVF Conf Comput Vis Pattern Recognit* [Internet] 2018;, p.

8697–8710. IEEEAvailable from: https://ieeexplore.ieee.org/document/8579005/.

Athayde Wirka K, Chen AA, Conaghan J, Ivani K, Gvakharia M, Behr B, Suraj V, Tan L, Shen S.

Atypical embryo phenotypes identified by time-lapse microscopy: High prevalence and

association with embryo development. *Fertil Steril* 2014;**101**:1637-1648.e5. Elsevier Inc.

Barnes J, Brendel M, Gao VR, Rajendran S, Kim J, Li Q, Malmsten JE, Sierra JT, Zisimopoulos P,

Sigaras A, *et al.* A non-invasive artificial intelligence approach for the prediction of

human blastocyst ploidy: a retrospective model development and validation study.

Lancet Digit Heal 2023;**5**:e28–e40.

Cameron NJ, Bhattacharya S, McLernon DJ. Cumulative live birth rates following blastocyst-

versus cleavage-stage embryo transfer in the first complete cycle of IVF: A population-

based retrospective cohort study. *Hum Reprod* 2020;**35**:2365–2374.

Canosa S, Bergandi L, Macrì C, Charrier L, Paschero C, Carosso A, Segni N Di, Silvagno F,

Gennarelli G, Benedetto C, *et al.* Morphokinetic analysis of cleavage stage embryos and

assessment of specific gene expression in cumulus cells independently predict human

embryo development to expanded blastocyst: a preliminary study. *J Assist Reprod*

Genet 2020;**37**:1409–1420.

Cetinkaya M, Pirkevi C, Yelke H, Colakoglu YK, Atayurt Z, Kahraman S. Relative kinetic

expressions defining cleavage synchronicity are better predictors of blastocyst

formation and quality than absolute time points. *J Assist Reprod Genet* 2015;**32**:27–35.

Chamayou S, Patrizio P, Storaci G, Tomaselli V, Alecci C, Ragolia C, Crescenzo C, Guglielmino

A. The use of morphokinetic parameters to select all embryos with full capacity to

implant. *J Assist Reprod Genet* 2013;**30**:703–710.

Chen T, Guestrin C. XGBoost: A scalable tree boosting system. *Proc ACM SIGKDD Int Conf*

Knowl Discov Data Min 2016;**13**-**17**-**Augu**:785–794.

Conaghan J, Chen AA, Willman SP, Ivani K, Chenette PE, Boostanfar R, Baker VL, Adamson GD,

Abusief ME, Gvakharia M, *et al.* Improving embryo selection using a computer-

automated time-lapse image analysis test plus day 3 morphology: results from a

prospective multicenter trial. *Fertil Steril* 2013;**100**:412-9.e5.

Croo I De, Colman R, Sutter P De, Stoop D, Tilleman K. No difference in cumulative live birth

rates between cleavage versus blastocyst transfer in patients with four or fewer

zygotes: results from a retrospective study. *Hum Reprod Open* 2022;**2022**:1–9.

Cruz M, Garrido N, Herrero J, Pérez-Cano I, Muñoz M, Meseguer M. Timing of cell division in

human cleavage-stage embryos is linked with blastocyst formation and quality. *Reprod*

Biomed Online 2012;**25**:371–381.

Dal Canto M, Coticchio G, Mignini Renzini M, Ponti E De, Novara PV, Brambillasca F, Comi R,

Fadini R. Cleavage kinetics analysis of human embryos predicts development to

blastocyst and implantation. *Reprod Biomed Online* [Internet] 2012;**25**:474–480.

Reproductive Healthcare Ltd.

Danardono GB, Erwin A, Purnama J, Handayani N, Polim AA, Boediono A, Sini I. A

Homogeneous Ensemble of Robust Pre-defined Neural Network Enables Automated

Annotation of Human Embryo Morphokinetics. *J Reprod Infertil* 2022;**23**:250–256.

DeLong ER, DeLong DM, Clarke-Pearson DL. Comparing the areas under two or more

correlated receiver operating characteristic curves: a nonparametric approach.

Biometrics 1988;**44**:837–845.

Dirvanauskas D, Maskeliunas R, Raudonis V, Damasevicius R. Embryo development stage

prediction algorithm for automated time lapse incubators. *Comput Methods Programs*

Biomed 2019;**177**:161–174. Elsevier B.V.

Gardner DK. In-vitro culture of human blastocyst. *Towar Reprod Certain Infertil Genet*

beyond 1999 1999;, p. 378–388. Parthenon Press.

Gardner DK, Lane M, Stevens J, Schlenker T, Schoolcraft WB. Blastocyst score affects

implantation and pregnancy outcome: Towards a single blastocyst transfer. *Fertil Steril*

2000;**73**:1155–1158.

Glujovsky D, Quinteiro Retamar AM, Alvarez Sedo CR, Ciapponi A, Cornelisse S, Blake D.

Cleavage-stage versus blastocyst-stage embryo transfer in assisted reproductive

technology. *Cochrane database Syst Rev* 2022;**5**:CD002118.

Huang B, Tan W, Li Z, Jin L. An artificial intelligence model (euploid prediction algorithm) can

predict embryo ploidy status based on time-lapse data. *Reprod Biol Endocrinol*

2021;**19**:1–10. BioMed Central.

Jiang X, Cai J, Liu L, Liu Z, Wang W, Chen J, Yang C, Geng J, Ma C, Ren J. Does conventional

morphological evaluation still play a role in predicting blastocyst formation? *Reprod Biol*

Endocrinol 2022;**20**:1–12. BioMed Central.

Katagiri Y, Jwa SC, Kuwahara A, Iwasa T, On M, Kato K, Kishi H, Kuwabara Y, Taniguchi F,

Harada M, *et al.* Assisted reproductive technology in Japan: A summary report for 2021

by the Ethics Committee of the Japan Society of Obstetrics and Gynecology. *Reprod*

Med Biol 2024;**23**:e12552.

Khan A, Gould S, Salzmann M. Deep convolutional neural networks for human embryonic cell

counting. *Lect Notes Comput Sci (including Subser Lect Notes Artif Intell Lect Notes*

Bioinformatics) 2016;**9913 LNCS**:339–348.

Kirkegaard K, Kesmodel US, Hindkjær JJ, Ingerslev HJ. Time-lapse parameters as predictors of

blastocyst development and pregnancy outcome in embryos from good prognosis

patients: A prospective cohort study. *Hum Reprod* 2013;**28**:2643–2651.

Kovačič B, Vlaisavljevi V, Reljič M, Gavri Lovrec V. Clinical outcome of day 2 versus day 5

transfer in cycles with one or two developed embryos. *Fertil Steril* 2002;**77**:529–536.

Leahy BD, Jang W-D, Yang HY, Struyven R, Wei D, Sun Z, Lee KR, Royston C, Cam L, Kalma Y,

et al. Automated Measurements of Key Morphological Features of Human Embryos for

IVF. *Med Image Comput Comput Assist Interv* 2020;**12265**:25–35.

Lee CI, Su YR, Chen CH, Chang TA, Kuo EES, Zheng WL, Hsieh WT, Huang CC, Lee MS, Liu M.

End-to-end deep learning for recognition of ploidy status using time-lapse videos. *J*

Assist Reprod Genet 2021;**38**:1655–1663. Journal of Assisted Reproduction and Genetics.

Leung AS, Son WY, Dahan MH. Time-lapse imaging of embryos: current evidence supporting

its use. *Expert Rev Med Devices* 2016;**13**:881–883. Taylor & Francis.

Liao Q, Zhang Q, Feng X, Huang H, Xu H, Tian B, Liu J, Yu Q, Guo N, Liu Q, *et al.* Development

of deep learning algorithms for predicting blastocyst formation and quality by time-

lapse monitoring. *Commun Biol* 2021;**4**:415.

Lundberg SM, Erion GG, Lee S-I. Consistent Individualized Feature Attribution for Tree

Ensembles. 2018;Available from: http://arxiv.org/abs/1802.03888.

Lundberg SM, Lee SI. A unified approach to interpreting model predictions. *Adv Neural Inf*

Process Syst 2017;4766–4775.

Milewski R, Kuć P, Kuczyńska A, Stankiewicz B, Łukaszuk K, Kuczyński W. A predictive model

for blastocyst formation based on morphokinetic parameters in time-lapse monitoring

of embryo development. *J Assist Reprod Genet* 2015;**32**:571–579.

Mizobe Y, Tokunaga M, Oya N, Iwakiri R, Yoshida N, Sato Y, Onoue N, Ezono Y. Synchrony of

the first division as an index of the blastocyst formation rate during embryonic

development. *Reprod Med Biol* 2018;**17**:64–70.

Motato Y, los Santos MJ de, Escriba MJ, Ruiz BA, Remohí J, Meseguer M. Morphokinetic

analysis and embryonic prediction for blastocyst formation through an integrated time-

lapse system. *Fertil Steril* 2016;**105**:376-84.e9.

Raudonis V, Paulauskaite-Taraseviciene A, Sutiene K, Jonaitis D. Towards the automation of

early-stage human embryo development detection. *Biomed Eng Online* 2019;**18**:1–20.

BioMed Central.

Storr A, Venetis CA, Cooke S, Kilani S, Ledger W. Inter-observer and intra-observer

agreement between embryologists during selection of a single Day 5 embryo for

transfer: A multicenter study. *Hum Reprod* 2017;**32**:307–314.

Storr A, Venetis CA, Cooke S, Susetio D, Kilani S, Ledger W. Morphokinetic parameters using

time-lapse technology and day 5 embryo quality: a prospective cohort study. *J Assist*

Reprod Genet 2015;**32**:1151–1160.

Sun X, Xu W. Fast Implementation of DeLong's Algorithm for Comparing the Areas Under

Correlated Receiver Operating Characteristic Curves. *IEEE Signal Process Lett* [Internet]

2014;**21**:1389–1393.

Sundvall L, Ingerslev HJ, Breth Knudsen U, Kirkegaard K. Inter- and intra-observer variability

of time-lapse annotations. *Hum Reprod* 2013;**28**:3215–3221.

the Japan Society of Obstetrics and Gynecology ART Databook 2021Available from:

https://www.jsog.or.jp/activity/art/2021_JSOG-ART.pdf.

Tran D, Cooke S, Illingworth PJ, Gardner DK. Deep learning as a predictive tool for fetal heart

pregnancy following time-lapse incubation and blastocyst transfer. *Hum Reprod*

2019;**34**:1011–1018.

Veeck LL. Oocyte Assessment and Biological Performance. *Ann N Y Acad Sci* 1988;**541**:259–

274.

Vlaisavljević V, Kovačič B, Reljič M, Lovrec VG, Sajko MČ. Is there any benefit from the

culture of a single oocyte to a blastocyst-stage embryo in unstimulated cycles? *Hum*

Reprod 2001;**16**:2379–2383.

Wong CC, Loewke KE, Bossert NL, Behr B, Jonge CJ De, Baer TM, Pera RAR. Non-invasive

imaging of human embryos before embryonic genome activation predicts development

to the blastocyst stage. *Nat Biotechnol* 2010;**28**:1115–1121. Nature Publishing Group.

Xiao JS, Healey M, Talmor A, Vollenhoven B. When only one embryo is available, is it better

to transfer on Day 3 or to grow on? *Reprod Biomed Online* 2019;**39**:916–923.

Yang ST, Shi JX, Gong F, Zhang SP, Lu CF, Tan K, Leng LZ, Hao M, He H, Gu YF, *et al.* Cleavage

pattern predicts developmental potential of day 3 human embryos produced by IVF.

Reprod Biomed Online 2015;**30**:625–634.

Youden WJ. Index for rating diagnostic tests. *Cancer* 1950;**3**:32–35.

Zaninovic N, Nohales M, Zhan Q, los Santos ZMJ de, Sierra J, Rosenwaks Z, Meseguer M. A

comparison of morphokinetic markers predicting blastocyst formation and implantation

potential from two large clinical data sets. *J Assist Reprod Genet* 2019;**36**:637–646.

Journal of Assisted Reproduction and Genetics.

Zoph B, Vasudevan V, Shlens J, Le Q V. Learning Transferable Architectures for Scalable

Image Recognition. *2018 IEEE/CVF Conf Comput Vis Pattern Recognit* [Internet] 2018;, p.

8697–8710. IEEEAvailable from: https://ieeexplore.ieee.org/document/8579005/.

Figure legends

Figure 1. Overview of our blastocyst prediction AI. (**A**) Flowchart of the blastocyst prediction dataset. (**B**) Flowchart of the 17-class embryo image classification dataset and fine-tuning of AI auto-annotator. (**C**) Proportion of each facility in the blastocyst prediction dataset. (**D**) Proportion of embryo morphology on embryonic day 5 or 6 in the blastocyst prediction dataset. (**E**) Schema of the blastocyst prediction AI structure. AI = artificial intelligence, CV = cross-validation, n = number of embryos, Fac. = facility, SHAP = SHapley Additive exPlanations

Figure 2. Fine-tuning of the AI auto-annotator. (**A**) Labels and number of images in the 17-class embryo image classification dataset. (**B, C**) Accuracy and loss of fine-tuning. (**D**) Confusion matrix of fine-tuning (17-class image classification). (**E**) Proportion of AI annotation labels at each time point in the blastocyst prediction dataset. AI = artificial intelligence.

Figure 3. Performance of the prediction models. (**A, B, C**) ROC and precision-recall analyses of three prediction models. The blue lines indicate the results of five-fold validations, whereas the red line denotes the test result. (**D**) Precisions and recalls at each threshold in the poor blastocyst + arrested embryo prediction model. AUC = area under the curve, CV = cross-validation, ROC = receiver operating characteristic, PPV = positive predictive value

Figure 4. ROC subgroup analyses stratified by facility (**A**) and age group (**B**). AUC = area

under the curve, ROC = receiver operating characteristic, y.o. = years old

Figure 5. Blastocyst prediction models from single or few time-lapse images. (**A**) Comparison

of time-lapse-based model with one or three label-based models. (**B**) Comparison of time-

lapse-based model with one of three image-based models. AUC = area under the curve, hpi

= hours post-insemination

Figure 6. Visualization of model decision basis. (**A**) Top 30 contributing features based on

SHAP. (**B**) SHAP values of highly contributing features of the blastocyst formation prediction

model. SHAP values of highly contributing features of the good blastocyst formation and

poor blastocyst + arrested embryo prediction models are shown in Supplementary Figure 4.

SHAP = SHapley Additive exPlanations

Figure 1. Overview of Our Blastocyst Prediction AI

A. Blastocyst Prediction Dataset

4 facilities
2,792 OPU cycles
n = 8,043 embryos

→ 24-64 hours post-fertilization images

Excluded   n = 931
· Low grade (< G3) n = 313
· Inadequate image quality or missing records n = 619

Blastocyst Prediction Dataset
n = 7,111

80% / 20%

Train / CV dataset
n = 5,689

Test dataset
n = 1,422

Build AI Model
Cross Validation

AI Model
Performance

B. Fine-tuning of Auto-annotator

17-class Embryo Image Classification Dataset
n = 23,852 images

· Cleavage stage
· Supervised label with stage and Veeck grade

90%   10%   10%

Train dataset   Validation dataset   Test dataset

Training with Data Augmentation   Early Stopping   AI Model Performance

Embryo Images
(331 x 331 pixel)

Image AI Auto-annotator
NASNet Large

Classification Labels
8cell G1
4cell G1
2cell G1
1cell

C. Proportion of Each Facility in the Blastocyst Prediction Dataset

Fac.1 14% | Fac.2 16% | Fac.3 39% | Fac.4 31%

D. Day5-6 Embryo Morphology of the Blastocyst Prediction Dataset

Moderate   Poor

Good 28% | 13% | 9% | Arrested 50%

Blastocyst

Classified by Gardner Grade

Good Blastocyst* : AA/AB/BA/BB
Moderate Blastocyst* : AC/CA/BC/CB
Poor Blastocyst*: CC   * Gardner Stage ≥ 3

E. Blastocyst Prediction AI

TIme-lapse Movie

preprocess →

Preprocessed Images
(331 x 331 pixel)

Image AI Auto-annotator
NASNet Large

Extracted Features Classification Labels
8cell G1
4cell G1
2cell G1
1cell

Age at oocyte retrieval

Classification labels on each time

Proportion of each cell stage / grade

Boosting Predictor

Gradient Boosting : XGBoost
dataset   model
train
test
error
Error
Iteration

Boosting Prediction

Prediction Value
Evaluation of Models

SHAP Value
Visualization of Model Decision Basis

Figure 2. Fine-tuning of Auto-annotator AI

A.

Label	n
1cell	1,188
1cell PN	1,516
2cell G1	1,289
2cell G2	1,275
2cell G3	1,393
3cell G1	1,119
3cell G2	1,059
3cell G3	1,112
4cell G1	1,315
4cell G2	1,376
4cell G3	1,357
5-7cell G1	1,530
5-7cell G2	1,909
5-7cell G3	1,720
8-cell G1	1,537
8-cell G2	1,570
8-cell G3	1,587

B. Accuracy

C. Loss

D. Confusion Matrix of Image Classification

E. The proportion of AI annotation labels on each time in the blastocyst prediction dataset

Hours Post-Insemination (hpi)

Figure 3. Performance of prediction models

A. Blastocyst formation prediction model

B. Good blastocyst formation prediction model

C. Poor blastocyst + arrested embryo prediction model

D. Poor blastocyst + arrested embryo prediction model

Threshold	Cross-validation mean		Test	
	Precision (PPV)	Recall (Sensitivity)	Precision (PPV)	Recall (Sensitivity)
0.9	0.95	0.45	0.95	0.44
0.8	0.92	0.60	0.90	0.61
0.7	0.89	0.69	0.87	0.70
0.6	0.85	0.76	0.85	0.76
0.5	0.83	0.81	0.82	0.81
0.4	0.79	0.87	0.79	0.87
0.3	0.74	0.92	0.75	0.92
0.2	0.68	0.96	0.69	0.97
Prior probability	0.59		0.59	

Figure 4. ROC subgroup analyses

A. Subanalysis by facility

Blastocyst formation

Test All (AUC : 0.87)
Facility 1 (AUC : 0.84)
Facility 2 (AUC : 0.90)
Facility 3 (AUC : 0.88)
Facility 4 (AUC : 0.83)

B. Subanalysis by age group

Blastocyst formation

Test All (AUC : 0.87)
~34 y.o. (AUC : 0.85)
35~39 y.o. (AUC : 0.88)
40~ y.o. (AUC : 0.86)

Figure 5.

A. Annotation-based models

B. Image-based models

Figure 6. Visualization of model decision basis

A. Absolute mean of SHAP values of the top 30 contributing features

B. SHAP values and highly contributive features of the blastocyst formation prediction model