AI-Driven Opportunistic CT Osteoporosis Screening and Establishment of Normative Values

Authors:

Malte Westerhoff [1], Soterios Gyftopoulos [2], Bari Dane [2], Emilio Vega [2], Daniel Murdock [2], Norbert Lindow [1], Felix Herter [1], Khaled Bousabarah [1], Michael P. Recht [2], Miriam A. Bredella[2]

[1] Visage Imaging, GmbH, Berlin, Germany

[2] Department of Radiology, NYU Langone Health and NYU Grossman School of Medicine, New York, NY, USA

Corresponding author:

Miriam A. Bredella, MD, MBA
NYU Langone Health and NYU Grossman School of Medicine
Translational Research Building 743
227 East 30th Street
New York, NY 10016
Tel: 646-501-2523

Funding information: None

Data Sharing: All data generated or analyzed during the study are included in the published paper

Abbreviations:

CT: Computed Tomography, AI: Artificial Intelligence, DXA: dual-energy X-ray absorptiometry, ROI: region of interest, HU: Hounsfield Units

Abstract

Background: Osteoporosis is underdiagnosed and undertreated. Traditional screening methods like dual-energy X-ray absorptiometry have limitations, prompting the exploration of opportunistic screening using CT and artificial intelligence (AI).

Purpose: To develop a reproducible and robust AI-based method to automatically assess spine trabecular attenuation on noncontrast CTs, irrespective of protocol or scanner model used during imaging acquisition, and to establish normative values in a large diverse population.

Methods: An AI-based method was developed to automatically quantify trabecular attenuation using a 3D region of interest (ROI) of the thoracic and lumbar spine on chest, abdomen, or spine CTs, adjusted for different tube voltages and scanner models. Normative values, thresholds for osteoporosis of trabecular attenuation of the thoracic and lumbar spine were established across a diverse population, stratified by age, sex, race, and ethnicity. Differences in trabecular attenuation were assessed using the Mann-Whitney U or the Chi-Square Test.

Results: 538,946 CT examinations from 283,499 patients (mean age 65 years ± 15, 51.2% women and 55.5% White), performed on 50 scanner models using six different tube voltages were analyzed. Hounsfield Units at 80 kVp versus 120 kVp differed by 23%, and different scanner models resulted in differences of values by < 10%. Automated ROI placement was validated by manual radiologist review, demonstrating 99% agreement. Mean trabecular attenuation was higher in young women (<50 years) than young men (p<.001) and decreased with age, with a steeper decline in postmenopausal women. In patients older than 50 years, trabecular attention was higher in male than female patients (p<.001). Trabecular attenuation was highest in Blacks, followed by Asians and lowest in Whites (p<.001). The threshold for L1 in diagnosing osteoporosis was 68 HU.

Conclusion: AI-based automated opportunistic osteoporosis screening can identify patients with low bone mineral density that undergo CT scans for clinical purposes on different scanners and protocols.

Introduction

Osteoporosis is characterized by low bone mineral density (BMD), which increases the risk of fractures (1). Even though osteoporosis is both preventable and treatable, it remains underdiagnosed and undertreated (2). Consequently, screening guidelines have been established to identify reduced BMD in at-risk individuals (3). Among patients eligible for osteoporosis screening, only one out of five patients undergo screening (4). In addition, significant disparities exist in osteoporosis screening and treatment across different racial and ethnic groups (5, 6).

Osteoporosis screening is typically performed by dual-energy x-ray absorptiometry (DXA), however, 50% of those who suffer a fracture have BMD values on DXA in the normal range (7, 8). BMD assessed by DXA depends on body size, which can lead to erroneous assessments of fracture risk for small patients (9) and those with obesity (10). DXA values in the spine can also be impacted by degenerative changes, such as osteophytes or vascular calcifications, which are common in the elderly. This can falsely elevate BMD and underestimate fracture risk (9). Moreover, DXA is a 2D planar technique that cannot separate cortical from trabecular bone (11).

Quantitative CT (QCT) using a calibration phantom is a more accurate method for measuring BMD compared to DXA. QCT is able to measure volumetric BMD, separate cortical from trabecular bone, and is less susceptible to overestimations of BMD due to degenerative diseases (11, 12). More recently, an "opportunistic screening" approach with AI-driven automatic assessment of BMD of the spine has been proposed to detect osteoporosis using non-contrast CT scans of the spine, chest or abdomen, which are performed for other clinical purposes (13, 14). Potential sources of inaccuracy with CT evaluation of BMD include variation in trabecular attenuation within a vertebra when using a 2D region of interest (ROI) on a single axial slice (13), and differences in scan parameters, such as difference in tube voltages, or scanner model (15, 16). There is also a lack of established normal values for BMD by opportunistic CT in racially and ethnically diverse populations. This can influence the performance of risk

prediction algorithms and possibly lead to unequal outcomes for individuals from diverse backgrounds (17, 18).

The purpose of this study was twofold. First, we aimed to develop a reproducible and robust AI-based method using a convolutional neural network that can automatically and accurately assess trabecular attenuation of the thoracic and lumbar spine on CT scans irrespective of the protocol or scanner model used during imaging acquisition. Second, we aimed to establish normative values for CT measurements of spine trabecular attenuation in a large diverse population over a wide age range.

Materials and Methods

Patient population

This study was approved by the institutional review board and was Health Insurance Portability and Accountability Act compliant. The requirement for informed consent was waived. A retrospective picture archiving and communication system (PACS; Visage Imaging, Inc., San Diego, CA, United States) search was performed for all patients between 20 and 94 years of age who underwent a non-dual energy, non-contrast CT of the chest, abdomen, thoracic or lumbar spine between 2003-2024 with a slice thickness of at least 5mm. The electronic health record was reviewed to obtain self-reported race and ethnicity information. As this study aimed to screen patients who would be diagnosed with osteoporosis or osteopenia by DXA, patients on medication that could affect bone metabolism were not excluded, similar to the National Health and Nutrition Examination Survey (NHANES), which serves as the reference database for DXA (19).

Computed Tomography Technique

Non-contrast chest, abdominal, and spine CT examinations were performed according to standard departmental protocols at 80-130 kVp. Tube current was 50 to 160 mAs, pitch 0.6-0.9, and rotation time 0.25-0.625s. Axial, coronal, and sagittal images at 0.5-5 mm were generated.

Automated Image segmentation

To automate ROI placement, a convolutional neural network was trained to segment the CT volumes into the major anatomical structures, specifically including and labeling the vertebral bodies (20). The standard method for manual measurement of bone attenuation on CT images typically involves placing a circular ROI within a single axial slice of the trabecular bone of a suitable vertebral body and measuring the mean attenuation in Hounsfield Units (HU). The choice of slice and ROI positioning can significantly impact the result because of trabecular inhomogeneities within a vertebral body (13). We

therefore used 3D spherical ROIs, which cover more voxels and therefore are less susceptible to trabecular inhomogeneities.

The volume of each lumbar and thoracic vertebra that was fully contained within the field of view was automatically calculated and a spherical ROI with a volume of 5% of the respective vertebral body was placed inside the vertebra. The ROI position was determined automatically to meet the following criteria: (i) the entire ROI was fully contained within the respective segmented vertebra, (ii) it had the lowest attenuation (13), and (iii) it was placed anterior to the basivertebral foramen (**Figure 1**). The median HU within the spherical ROI was then computed to provide a more robust measure against outliers than could be seen using a mean value. Finally, the HU values were averaged for each vertebral body that was fully imaged in more than one series.

Automated consistency checks for each vertebral body were performed to ensure that the segmentations were valid (**Supplement 1**). If they were not valid, the series was excluded

Three board certified musculoskeletal radiologists with ≥20 years of experience (SG, MPR, MAB) independently reviewed segmentations and ROI placements of 128 CT scans from 128 patients randomly selected that were not part of the training set.

To compare attenuation values obtained by our 3D method with the standard 2D method, we also placed a 2D elliptic ROI in the central slice of each vertebral body and calculated the mean attenuation as described by Jang et al. (13).

Adjustments for Tube Voltage and Scanner Model

Attenuation values in HU from different scanner models and different kVp values were calibrated to a common reference, as the included CT examinations were performed on multiple different CT scanner models with different imaging protocols (**Supplement 2**).

The most commonly used tube voltage for the included studies was 120 kVp (60% of studies) and thus served as the reference for our data. For scans with other tube voltages, the HU values were adjusted to the equivalent 120 kVp using a linear scaling $HU_{120} = A_x\ HU_x$ where HU_x is the HU measured at voltage x, HU_{120} is the equivalent value calibrated to a 120 kVp scan, and A_x is the mapping parameter determined by the statistical analysis from **Supplement 2**. After tube voltage calibration, scanner model calibration was performed using the same statistical method **(Supplement 2)**.

Statistical analysis

Statistical analysis was performed using software (SQL Server and Excel, Microsoft, Redmond, WA, and Python 3.12). Continuous variables were reported as mean, median, standard deviation, 10th and 90th percentiles. Categorical variables were reported using counts and percentages. Normative values for each lumbar and thoracic vertebral body were stratified by sex, race, and ethnicity, and age group in 5-year intervals. Differences in means of continuous variables were assessed using the Mann-Whitney U test and differences in categorical variables were assessed using the Chi-Square Test. A two☐sided p value <.05 was considered significant.

The standard criterion to diagnose osteoporosis by DXA is a BMD that is -2.5 standard deviations (SD) or lower than that of a healthy young population (T score -≤2.5), and the diagnosis of osteopenia is a BMD that is between -1 and -2.5 SD below the BMD of a healthy young population (21). Given that our proposed method is a different measurement technique than DXA, statistical measures such as standard deviation are different and therefore the same thresholds cannot be applied (22). To establish thresholds that correspond to the established DXA criteria of osteoporosis, osteopenia, or normal, the following approach was used: Prior studies have demonstrated that the prevalence of osteoporosis in 60-69 year-old female patients is 12% and the prevalence of osteopenia in that cohort is 53% (23). Consequently, the 12th and 53rd percentiles of CT-based attenuation (HU) were calculated for each vertebra in our female patient population. Those HU values are the thresholds that would classify 12% of patients as osteoporotic and 53% of patients as osteogenic, respectively. The International Society for Clinical Densitometry (ISCD)

recommended that a female reference database should be used to assess osteoporosis in men (21). We

therefore did not determine separate thresholds for men.

Results

Demographics

In our retrospective PACS search 592,878 CT examinations in 307,238 patients were identified of which 53,932 were excluded based on the automated consistency checks. Consistency checks identified technical reasons such as metal artifact, compression fractures, sclerotic or lytic vertebral lesions that precluded accurate assessment of trabecular attenuation, resulting in 538,946 evaluable CT examinations in 283,499 patients (mean age 65 years ± 15, 51.2 % women and 55.5 % White) **(Figure 2 and Table 1).**

Automated ROI placement, Scanner and kVp Model Adjustments

In 99% (127 out of 128 cases), the radiologists agreed with the automated ROI placement. The mean difference between 2D vs 3D ROI placement was 53HU and the median difference was 26 HU (14-26 IQR).

CTs were performed on 50 scanner models using six different tube voltages. Mapping parameters (Ax) for each tube voltage (kVp) to adjust HU to the equivalent of 120 kVp, the most commonly used tube voltage, are listed in **Table 2a**. HU at 80 kVp and 120 kVp differed by 23% (correction factor 0.77).

After tube voltage calibration, scanner model calibration was performed and the respective values for each scanner models are listed in **Table 2b.** The correction factor to calibrate individual scanner models to the selected reference scanner was less than 5% for most (41 of 50) scanner models and less than 10% for 47 or the 50 scanner models **(Table 2b).** This suggests that scanner model specific calibration is less critical than correction for tube kilovoltage difference.

Normative Values

Attenuation values of trabecular bone of thoracic and lumbar vertebra stratified by patient age and sex were calculated. The results for the L1 vertebra by age and sex are provided in **Tables 3a and b** and **Figure 3**.

The mean trabecular attenuation decreased with increasing age. In younger patients (premenopausal, < 50 years), the attenuation was higher in women than in men (mean attenuation of L1 in 30–35 year-old women versus 30-35 year-old men: 204±42 HU versus 190±41 HU, p<.001), but lower in older (postmenopausal, >50 years) women than in men (mean attenuation of L1 in 65–70 year-old women versus 65-70 year-old men: 105±44 HU versus 119±44 HU, p<.001) (**Figure 3**). The decrease of trabecular attenuation in men was almost linear with age while in women there was a steeper drop around menopause (**Figure 3**).

The attenuation measurements differed amongst the vertebrae. In general, the attenuation was lower in the more caudal vertebrae. It was highest in T1, and lowest in L3, though, in young men the difference between L1 to L4 was minimal. The age-related reductions in attenuation were greatest in the more caudal vertebrae, ranging from 2.3 HU/year in the T1 vertebra to 2.9 HU/year in the L5 vertebra in women, while the mean attenuation decreased from 1.2 HU/year (T1) and 2.1 HU/year (L5) in men (**Figure 4**).

The spread of trabecular attenuation was wide in both female and male patients, as shown as 10[th] and 90[th] percentile in **Tables 3a and b. Table 4** shows the 12[th] percentile and the 53[rd] percentile of the attenuation in each vertebra in females 60-69 years of age. As the incidence of osteoporosis and osteopenia in this age group is estimated to be 12% and 53%, respectively (23), the values in **Table 4** can be used in men and women as thresholds to classify osteoporosis and osteopenia that would result in a similar incidence (21, 23). For L1, that threshold was 68 HU for osteoporosis and 107 HU for osteopenia.

Normative Values by Race and Ethnicity

Analyses were performed by self-reported race and ethnicity. The three largest groups were White (55.5%), Black (8.5%) and Asian (4.9%). In 31% no information on race was available (**Table 1**). The characteristics of age-related bone loss were similar in all three groups in women and men. Trabecular attenuation was highest in Blacks followed by Asians and lowest in Whites (mean attenuation of L1 in 65–70 year-old Black, Asian, and White women: 131±44 HU, 108±42 HU, 98±32 HU, p<.001,

and in 65–70 year-old Black, Asian, and White men: 146±47 HU, 120±41 HU, 111±37 HU, p<.001) **(Figures 5a and b).** In our cohort, 7.8% were Hispanic and 53.8% were non-Hispanic, while in 38.4% no data on ethnicity were available **(Table 1).** The characteristics of age-related bone loss were similar in both groups in women and men. There was no statistically significant difference in trabecular attenuation in Hispanic and non-Hispanic women (mean attenuation of L1 in 65–70 year-old Hispanic and non-Hispanic women:103±34 HU versus 102±37 HU, p=.36.) **(Figure 6a)**, while Hispanic men had higher trabecular attenuation than non-Hispanic men (mean attenuation of L1 in 65–70 year-old Hispanic and non-Hispanic men: 123±50 HU versus 114±40 HU, p<.001) **(Figure 6b).**

Discussion

Osteoporosis is underdiagnosed and undertreated, and many patients at risk for osteoporosis are not referred for DXA screening, especially the underserved (2, 4-6). Our study showed that opportunistic osteoporosis screening, which leverages existing CT scans performed for other clinical purposes, can address these disparities and identify patients with low BMD who may be at risk for fractures.

There are technical challenges associated with opportunistic osteoporosis screening, such as variations in scan protocols or differences between scanner models (15, 16). This is mainly due to different compositions of the x-ray spectra with different tubes and different voltages, and the dependency of absorption rates of different materials on photon energy. These factors can lead to imprecise measurements of trabecular attenuation, potentially resulting in the overdiagnosis or underdiagnosis of osteoporosis (15). Phantom measurements, performed either synchronous with the clinical scan or asynchronous, can be used to determine the function that maps attenuation values obtained at one kVp setting on one scanner model to those measured at another (24, 25). However, phantom measurements are not practical for opportunistic screening applications and are generally not available for retrospective analysis. We therefore developed a novel statistical method of calibrating for different scan protocols and different scanner models. We found that tube voltage had a larger influence on attenuation values (23%) than scanner model (<10%), suggesting that scanner model specific calibration is less critical than correction for voltage difference. The correction factors for scanner model and tube voltage from this study can be used by others to analyze retrospective and prospective datasets that were obtained using different scan protocols and scanner models.

Another technical challenge of opportunistic osteoporosis screening is accurate placement of the ROI within trabecular bone. Differences in slice selection or ROI placement can lead to substantial variations in results. In a study by Jang et al (13), automated ROI placement in the L1 vertebral body was on average 21 HU higher than manual placement, highlighting the need for accurate and reproducible ROI placement. While 3D ROIs, such as spheres or cylinders, provide more robust statistical measures by

sampling multiple slices and more voxels, they are rarely used in clinical practice due to PACS limitations and the increased complexity of user interaction. We developed a novel method of automatically placing a 3D spherical ROI within trabecular bone which uses the lowest areas of trabecular attenuation, thereby avoiding bone islands or other lesions that can artifactually increase bone density. Our measurements were lower than previously reported thresholds. Compared to the 2D method described by Jang et al (13), the values using our method were a mean of 35 HU and median 26 HU lower, when averaged over all subjects. Thus, our threshold of 68 HU for diagnosing osteoporosis in a 3D ROI is comparable to a threshold of 90-100 HU using a 2D method reported in previous work (13).

Our study showed that the trabecular attenuation differed in different vertebrae, indicating that specific thresholds should be used for each vertebra as indicators for osteoporosis, and provide reference values for each vertebra. Of note, age-related bone loss was greatest in the lumbar spine, suggesting that using lumbar vertebrae may be more sensitive to detect bone loss. Most prior studies limited evaluation to a single vertebra, such as the L1 (13, 26). Our automated method of placing 3D-ROIs makes it feasible to measure trabecular attenuation in all vertebrae in the field of view. The automatic generation of our attenuation measurements will be an important contributing factor to its widespread adoption, as it will decrease potential resistance from radiologists who could find the task of manually performing these measurements on eligible CT examinations burdensome.

One major obstacle to using opportunistic osteoporosis screening is the lack of normative data in diverse populations. Our study examining 538,946 CTs in 283,499 patients from different races and ethnicities is the largest to date. Our values show higher mean trabecular attenuation in Blacks compared to Whites and Asians in both sexes, which is consistent with findings previously reported in literature (27). We also show higher trabecular attenuation in Asians compared to Whites. Hispanic men had higher trabecular attenuation compared to non-Hispanics, while there was no significant difference in women. Clinical implications of opportunistic osteoporosis screening and established CT based normative values include an increase in the number of patients who will get screened for osteoporosis, and referred for

DXA, enabling timely intervention, thereby preventing fractures, reducing morbidity, costs and improving overall patient outcomes (28-30).

Our study had several limitations. First, the generalizability of our study results may be limited as the data used to create the normative values were taken from facilities on the Northeastern United States. The inclusion of 283,499 patients should help negate the impact of this issue. We relied on self-reported data on race and ethnicity which were only available in a subset of patients. Second, we used a 3D-ROI that identified the lowest attenuation within a vertebral body, and our attenuation values were a mean of 35 and a median of 26 HU lower than reported values that were obtained using a 2D methods. Therefore, our thresholds should not be used with 2D-ROIs without adjustments. Third, we included patients without taking into account their past medical histories and medications which could potentially impact some of the measurements used to create our normative values. However, we modeled our methods after NHANES, which is used as a reference database for DXA and does not exclude patients on bone-active medications (19, 31). We estimate that this impact is small, due to the large number of included patients and the wide range of indications for the CT, most of which are not related to chronic bone disease.

In conclusion, we developed a reproducible and robust AI based method using a convolutional neural network that can automatically and accurately assess trabecular attenuation of the thoracic and lumbar spine on CT scans irrespective of the protocol or scanner model used during imaging acquisition. Age, sex, race and ethnicity stratified normative values were also established that can be used to diagnose osteoporosis. These are important steps towards the widespread adoption of opportunistic CT imaging for this patient population, which can lead to improved patient outcomes through opportunistic osteoporosis screening and decreased overall healthcare costs.

References

1. Nih Consensus Development Panel on Osteoporosis Prevention D, Therapy. Osteoporosis prevention, diagnosis, and therapy. JAMA. 2001;285(6):785-95.

2. Lems WF, Raterman HG. Critical issues and current challenges in osteoporosis and fracture prevention. An overview of unmet needs. Ther Adv Musculoskelet Dis. 2017;9(12):299-316.

3. Camacho PM, Petak SM, Binkley N, Diab DL, Eldeiry LS, Farooki A, et al. American Association of Clinical Endocrinologists/American College of Endocrinology Clinical Practice Guidelines for the Diagnosis and Treatment of Postmenopausal Osteoporosis- 2020 Update Executive Summary. Endocr Pract. 2020;26(5):564-70.

4. Gyftopoulos S, Pelzl CE, Da Silva Cardoso M, Xie J, Kwon SC, Chang CY. Bone Density Screening Rates Among Medicare Beneficiaries: An Analysis with a focus on Asian Americans. Skeletal Radiol. 2024;53(11):2347-55.

5. Miller RG, Ashar BH, Cohen J, Camp M, Coombs C, Johnson E, et al. Disparities in osteoporosis screening between at-risk African-American and white women. J Gen Intern Med. 2005;20(9):847-51.

6. Noel SE, Santos MP, Wright NC. Racial and Ethnic Disparities in Bone Health and Outcomes in the United States. J Bone Miner Res. 2021;36(10):1881-905.

7. Siris ES, Chen YT, Abbott TA, Barrett-Connor E, Miller PD, Wehren LE, et al. Bone mineral density thresholds for pharmacological intervention to prevent fractures. Arch Intern Med. 2004;164(10):1108-12.

8. Stone KL, Seeley DG, Lui LY, Cauley JA, Ensrud K, Browner WS, et al. BMD at multiple sites and risk of fracture of multiple types: long-term results from the Study of Osteoporotic Fractures. J Bone Miner Res. 2003;18(11):1947-54.

9. Kroger H, Vainio P, Nieminen J, Kotaniemi A. Comparison of different models for interpreting bone mineral density measurements using DXA and MRI technology. Bone. 1995;17(2):157-9.

10. Javed F, Yu W, Thornton J, Colt E. Effect of fat on measurement of bone mineral density. Int J Body Compos Res. 2009;7(1):37-40.

11.	Lin W, He C, Xie F, Chen T, Zheng G, Yin H, et al. Quantitative CT screening improved lumbar BMD evaluation in older patients compared to dual-energy X-ray absorptiometry. BMC Geriatr. 2023;23(1):231.

12.	Lin W, He C, Xie F, Chen T, Zheng G, Yin H, et al. Discordance in lumbar bone mineral density measurements by quantitative computed tomography and dual-energy X-ray absorptiometry in postmenopausal women: a prospective comparative study. Spine J. 2023;23(2):295-304.

13.	Jang S, Graffy PM, Ziemlewicz TJ, Lee SJ, Summers RM, Pickhardt PJ. Opportunistic Osteoporosis Screening at Routine Abdominal and Thoracic CT: Normative L1 Trabecular Attenuation Values in More than 20 000 Adults. Radiology. 2019;291(2):360-7.

14.	Pickhardt PJ, Nguyen T, Perez AA, Graffy PM, Jang S, Summers RM, et al. Improved CT-based Osteoporosis Assessment with a Fully Automated Deep Learning Tool. Radiol Artif Intell. 2022;4(5):e220042.

15.	Garner HW, Paturzo MM, Gaudier G, Pickhardt PJ, Wessell DE. Variation in Attenuation in L1 Trabecular Bone at Different Tube Voltages: Caution Is Warranted When Screening for Osteoporosis With the Use of Opportunistic CT. AJR Am J Roentgenol. 2017;208(1):165-70.

16.	Shepherd JA, Schousboe JT, Broy SB, Engelke K, Leslie WD. Executive Summary of the 2015 ISCD Position Development Conference on Advanced Measures From DXA and QCT: Fracture Prediction Beyond BMD. J Clin Densitom. 2015;18(3):274-86.

17.	Blue MNM, Tinsley GM, Ryan ED, Smith-Ryan AE. Validity of Body-Composition Methods across Racial and Ethnic Populations. Adv Nutr. 2021;12(5):1854-62.

18.	Orimoloye OA, Budoff MJ, Dardari ZA, Mirbolouk M, Uddin SMI, Berman DS, et al. Race/Ethnicity and the Prognostic Implications of Coronary Artery Calcium for All-Cause and Cardiovascular Disease Mortality: The Coronary Artery Calcium Consortium. J Am Heart Assoc. 2018;7(20):e010471.

19.   Prevention CfDCa. National Health and Nutrition Examination Survey Dual-energy X-ray

Absorptimetry (DXA) Procedure Manual Hyattsville [Available from:

https://wwwn.cdc.gov/nchs/data/nhanes/public/2007/manuals/manual_dexa.pdf.

20.   Wasserthal J, Breit HC, Meyer MT, Pradella M, Hinck D, Sauter AW, et al. TotalSegmentator:

Robust Segmentation of 104 Anatomic Structures in CT Images. Radiol Artif Intell. 2023;5(5):e230024.

21.   Watts NB, Leslie WD, Foldes AJ, Miller PD. 2013 International Society for Clinical

Densitometry Position Development Conference: Task Force on Normative Databases. J Clin Densitom.

2013;16(4):472-81.

22.   Mindways. QCT vs. DXA: What's the Score?2012. Available from:

https://www.qct.com/Downloads/Whitepaper-20130402_online.pdf.

23.   Wright NC, Looker AC, Saag KG, Curtis JR, Delzell ES, Randall S, et al. The recent prevalence

of osteoporosis and low bone mass in the United States based on bone mineral density at the femoral neck

or lumbar spine. J Bone Miner Res. 2014;29(11):2520-6.

24.   Bartenschlager S, Cavallaro A, Pogarell T, Chaudry O, Uder M, Khosla S, et al. Opportunistic

Screening With CT: Comparison of Phantomless BMD Calibration Methods. J Bone Miner Res.

2023;38(11):1689-99.

25.   Engelke K, Chaudry O, Bartenschlager S. Opportunistic Screening Techniques for Analysis of

CT Scans. Curr Osteoporos Rep. 2023;21(1):65-76.

26.   Langer FW, Torri GB, Schaffazick F, Maia G, Wiethan CP, Haygert CJ, et al. Opportunistic

Screening for Low Bone Mineral Density in Routine Computed Tomography Scans: A Brazilian

Validation Study. J Clin Densitom. 2024;28(1):101539.

27.   Hochberg MC. Racial differences in bone strength. Trans Am Clin Climatol Assoc.

2007;118:305-15.

28.   Gyftopoulos S, Pelzl CE, Chang CY. Quantifying the Opportunity and Economic Value of Bone

Density Screening Using Opportunistic CT: A Medicare Database Analysis. J Am Coll Radiol.

2025;22(3):349-57.

29.   Pickhardt PJ, Correale L, Hassan C. AI-based opportunistic CT screening of incidental cardiovascular disease, osteoporosis, and sarcopenia: cost-effectiveness analysis. Abdom Radiol (NY). 2023;48(3):1181-98.

30.   Pickhardt PJ, Lee SJ, Liu J, Yao J, Lay N, Graffy PM, et al. Population-based opportunistic osteoporosis screening: Validation of a fully automated CT tool for assessing longitudinal BMD changes. Br J Radiol. 2019;92(1094):20180726.

31.   Looker AC, Melton LJ, 3rd, Borrud LG, Shepherd JA. Changes in femur neck bone density in US adults between 1988-1994 and 2005-2008: demographic patterns and possible determinants. Osteoporos Int. 2012;23(2):771-80.

Tables

Table 1. Demographics of the study population

Variables	Mean (SD)
Age (years)	65 ± 15
Female (%)	145,021 (51.2%)
Male (%)	138,478 (48.8%)
Race (%)	
White	157,457 (55.5%)
Black	24,152 (8.5%)
Asian	13,938 (4.9%)
Other/unknown	87,952 (31.0%)
Ethnicity (%)	
Hispanic	22,114 (7.8%)
Non-Hispanic	152,596 (53.8%)
Other/unknown	108,789 (38.4%)

Table 2a. Adjustment of Hounsfield Units (HU) values to the equivalent 120 kVp using linear scaling

Tube Voltage (kVp)	Mapping parameter Ax
80	0.773
90	0.808
100	0.918
110	0.935
120	1.000
130	1.015

Table 2b. Scanner models and mapping parameter using linear scaling.

Scanner Model	Mapping parameter Ax
TOSHIBA Aquilion	0.824
TOSHIBA Aquilion/LB	0.875
GE LightSpeed QX/i	0.895
Siemens SOMATOM X.cite	0.900
TOSHIBA Aquilion PRIME	0.905
SIEMENS Emotion 16 (2007)	0.909
GE LightSpeed Plus	0.919
Siemens SOMATOM X.ceed	0.935
Siemens SOMATOM go.Top	0.937
SIEMENS Emotion Duo	0.944
SIEMENS SOMATOM go.Up	0.948
SIEMENS SOMATOM Edge Plus	0.959
SIEMENS SOMATOM Force	0.960
SIEMENS Perspective	0.964
GE Optima CT660	0.969
SIEMENS SOMATOM Perspective	0.969
SIEMENS SOMATOM go.All	0.974
SIEMENS Sensation 16 CT1	0.974
SIEMENS Sensation 4	0.975
SIEMENS SOMATOM Drive	0.979
SIEMENS Emotion 6	0.980
GE LightSpeed16	0.981
SIEMENS SOMATOM Definition AS	0.984
GE BrightSpeed	0.984
SIEMENS Emotion 16	0.985
SIEMENS Sensation 16	0.988
GE LightSpeed Pro 16	0.988
SIEMENS SOMATOM Definition Flash	0.996
SIEMENS SOMATOM Definition Edge	**1.000**
Philips Brilliance 64	1.000

SIEMENS Sensation 16 CT2	1.005
SIEMENS Biograph40	1.006
SIEMENS SOMATOM Definition AS+	1.012
SIEMENS Definition	1.013
SIEMENS Definition AS+	1.014
SIEMENS Sensation 64	1.020
SIEMENS SOMATOM Definition	1.021
SIEMENS Sensation 40	1.025
GE LightSpeed VCT	1.032
SIEMENS Biograph 40	1.083
TOSHIBA Aquilion	0.824
TOSHIBA Aquilion/LB	0.875
GE LightSpeed QX/i	0.895
Siemens SOMATOM X.cite	0.900
TOSHIBA Aquilion PRIME	0.905
SIEMENS Emotion 16 (2007)	0.909
GE LightSpeed Plus	0.919
Siemens SOMATOM X.ceed	0.935
Siemens SOMATOM go.Top	0.937
SIEMENS Emotion Duo	0.944

Table 3a. Normative attenuation values of L1 trabecular bone in Hounsfield Units (HU) in females

Age (years)	Mean	10th percentile	Median	90th percentile
20 - 24	213	166	213	262
25 - 29	207	157	206	259
30 -34	204	154	202	257
35 -39	200	149	198	254
40 - 44	191	139	189	246
45 - 49	177	123	176	233
50 - 52	152	100	148	208
55 -59	130	84	127	180
60 -64	116	71	112	165
65 - 69	105	61	102	153
70 -74	97	53	93	144
75 - 79	88	44	84	135
80 - 84	81	36	76	127
85 - 89	73	28	68	120
90 -95	66	20	62	114

Table 3b. Normative attenuation values of L1 trabecular bone in Hounsfield Units (HU) in men

Age (years)	Mean	10th percentile	Median	90th percentile
20 - 24	206	156	204	261
25 - 29	199	149	197	251
30 -34	190	142	189	241
35 -39	183	134	180	234
40 - 44	173	124	171	224
45 - 49	164	115	161	215
50 - 52	151	103	148	204
55 -59	139	91	136	190
60 -64	129	81	126	182
65 - 69	119	70	116	171
70 -74	110	62	107	161
75 - 79	101	52	97	152
80 - 84	93	44	88	144
85 - 89	82	35	78	135
90 -95	76	25	71	131

Table 4 12[th] and 53[rd] percentile of trabecular attenuation in HU in female patients 60-69 years

Vertebra	12[th] percentile Attenuation (HU)	53[rd] percentile Attenuation (HU)
T1	111	157
T2	104	149
T3	97	143
T4	93	138
T5	90	135
T6	86	130
T7	83	125
T8	80	123
T9	82	125
T10	86	128
T11	85	125
T12	76	115
L1	68	107
L2	61	100
L3	57	95
L4	58	97
L5	66	107

Figures and Figure legends

Figure 1. Placement of 3-dimensional region of interests within thoracic and lumbar vertebral bodies. a) straightened spine (curved view) with vertebrae segmentation, b) close-up of vertebra segmentation with axial ROI center slice position indicator, c)   axial ROI placement, and d) sagittal ROI placement.

Figure 2. Flow diagram

Figure 3. Mean trabecular attenuation in Hounsfield Units (HU) of L1 by sex and age. The decrease of trabecular attenuation in men was almost linear with age while in women there was a steeper drop around menopause. In all figures showing age, data points are aggregated in 5 year intervals.

Figure 4. Age-related mean trabecular attenuation of selected thoracic and lumbar vertebra in Hounsfield Units (HU) in females. Attenuation differed amongst the vertebrae with lower attenuation in the more caudal vertebrae which also showed greater age-related reductions in attenuation.

Figure 5a. Mean trabecular attenuation in Hounsfield Units (HU) of the L1 vertebra in White, Black, and Asian females. Trabecular attenuation was highest in Blacks followed by Asians and lowest in Whites.

Figure 5b. Mean trabecular attenuation in Hounsfield Units (HU) of the L1 vertebra in White, Black, and Asian males. Trabecular attenuation was highest in Blacks followed by Asians and lowest in Whites.

Figure 6a. Mean trabecular attenuation in Hounsfield Units (HU) of the L1 vertebra in Hispanic and Non-Hispanic women. There was no statistically significant difference in trabecular attenuation in Hispanic and non-Hispanic women.

Figure 6b. Mean trabecular attenuation in Hounsfield Units (HU) of the L1 vertebra in Hispanic and Non-Hispanic men. Hispanic men had higher trabecular attenuation than non-Hispanic men.