Title: A sensitive multiplex RT-qPCR assay to detect SARS-CoV-2 in respiratory

samples

Authors: Alice Halliday[1,2,#,$], Begonia Morales-Aza[1,2,#], Kaltun Duale[1,2, #], Philippa Lait[3,4], Bristol

Vaccine Centre Microbiology & Molecular Laboratories*, Rachel Milligan[2], Stephanie Hutchings[5],

Hannah Pymont[5], Barry Vipond[5], Peter Muir[5], Christy Waterfall[6], Jane Coghill[6], Jeremy Carr[7], Emma

Plested[7], Jennifer Oliver[1,8], Katy Fidler[9,10], Caroline Relton[11,12], Matthew D Snape[7,13] on behalf of the

'Be on the Team' Investigators*, Andrew D. Davidson[2], Christopher Helps[3,4], & Adam Finn[1,2,8,14].

Affiliations:

1 – Bristol Vaccine Centre, University of Bristol

2 – School of Cellular and Molecular Medicine, University of Bristol.

3 - Langford Vets, University of Bristol

4 – School of Veterinary Sciences, University of Bristol

5 - UK Health Security Agency (UKSHA)

6 - Genomics Facility, Faculty of Health & Life Sciences, University of Bristol

7 - Oxford Vaccine Group, University of Oxford Department of Paediatrics.

8 - Population Health Sciences, Bristol Medical School, University of Bristol, UK

9 - Brighton and Sussex Medical School, Falmer, UK

10 - Royal Alexandra Children's Hospital, University Hospital Sussex NHS Foundation Trust

11 - MRC Integrative Epidemiology Unit, Population Health Sciences, Bristol Medical School,

University of Bristol, UK

12- London School of Hygiene & Tropical Medicine, London, UK

13 - Oxford NIHR Biomedical Research Centre.

14 - Paediatric Immunology and Infectious Diseases, Bristol Royal Hospital for Children

*See Appendix 1 for a full list of investigators from both the BVC Microbiology and Molecular

Laboratories and from the 'Be on the Team' Investigators.

# These authors contributed equally

$ Corresponding author: alice.halliday@bristol.ac.uk

Abstract

We developed an in-house RT-qPCR assay for the detection of SARS-CoV-2 from different sample

types. Novel primer and probe sets were designed to amplify 2 regions of the nucleocapsid (N) gene.

Two dual-target multiplex assays were then evaluated for performance in detection of SARS-CoV-2

in known samples. An assay combining one novel primer pair for N (N6), a published region for the

envelope protein (E) and an extraction control target (MS2), was found to have good efficiency and a

low limit of detection, and to be sensitive and specific for SARS-CoV-2 in a sample set of known

positive and negative clinical samples. Assay performance was maintained for the main variants of

concern (Alpha, Delta, Gamma and Omicron BA.1, BA.2, BA4, JN.1). The assay was used to test for

SARS-CoV-2 in throat swab samples collected in schools from teenagers in the UK as part of an

observational study of two Meningococcal B Vaccines in Jan-March 2020. The low prevalence of

positives identified (0.44-0.47% in two weeks of March 2020) suggests that there was minimal

circulation of SARS-CoV-2 in UK schools prior to the first lockdown. The N6/E/MS2 assay has since

been used to detect SARS-CoV-2 in samples from multiple studies. Continual monitoring will ensure

that performance of this novel multiplex assay is maintained as novel sub-variants emerge.

Introduction

The emergence of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) which causes coronavirus infectious disease 2019 (COVID-19) has led to a pandemic with severe health, social and economic consequences globally. Some early attempts to establish diagnostic PCR assays to detect SARS-CoV-2 using a range of gene targets including those encoding the RNA-dependent RNA polymerase (RdRp) and nucleocapsid (N), envelope (E) and Spike (S) proteins (1, 2) were only moderately successful with relatively low sensitivity and significant numbers of false negatives. Subsequently a range of high performing assays have been developed and deployed by government and academic labs and numerous commercial assays have also become available.

Lateral flow antigen tests, which are less sensitive but reasonably predictive of infectiousness have the important advantages of low cost, rapid results with no requirement for laboratory processing and have been widely deployed to support public health efforts to control spread of the infection. Nevertheless, RT-qPCR remains the gold standard testing methodology for individual diagnosis of SARS-CoV-2 infection with high sensitivity and specificity.

Here we describe development, evaluation and deployment of a novel SARS-CoV-2 RT-qPCR assay which uses Envelope (E) and Nucleocapsid (N) gene targets and delivers high technical and diagnostic performance. We also present data from the testing of throat swabs samples collected from teenagers in England in early 2020, at the very start of the UK SARS-CoV-2 epidemic.

Methods

Samples, Ethics and Consent

Frozen SARS-CoV-2 positive and negative nasopharyngeal/oropharyngeal swabs (NPS/OPS) in viral transport medium (VTM) collected in early 2020 were taken from individuals undergoing testing for SARS-CoV-2 across the Bristol area. These were tested at Severn Pathology Laboratories, Bristol, UK

Health Security Agency (UKHSA), and anonymised remnant samples and laboratory test data were

subsequently transferred to the University of Bristol to assist in the validation of two candidate RT-

qPCR assays.

For field testing the assay on unknown samples, oropharyngeal swabs (OPS) from healthy 16–17-

year-olds enrolled in the "Be on the Team" (BOTT) study were used. This was a large vaccine study in

progress at the time evaluating the impact of two protein antigen meningococcal vaccines (Bexsero

(GSK) and Trumenba (Pfizer)) on respiratory carriage (3, 4). Signed informed consent was obtained in

all cases for surplus material from swab samples to be used for additional research and permission

to use them for detection of SARS-CoV-2 was obtained from the ethics committee who first

approved the parent study before this work was done (18/SC/0055). Oropharyngeal swabs were

collected into 1.5 mL skim milk, tryptone, glucose, and glycerin (STGG) broth (E&O Laboratories)

from BOTT at sites in 16 regions across the UK between January and March 2020. These swabs were

shipped to the Bristol laboratories for total nucleic acid extraction and RT-qPCR.

Positive control material for SARS-CoV-2 Delta and Omicron strains were obtained with appropriate

informed consent via Bristol Biobank (20/WA/0273), from infected adult volunteer donors working

in the Biomedical Sciences Building, University of Bristol. Collected specimens were held at 4°C and

transported to the laboratories before freezing at -70 °C within 4 hours.

Extraction

Samples were kept frozen at −70 °C and were defrosted on ice prior to nucleic acid extraction. All

samples were first vortexed for 30 seconds, then chemically inactivated by adding 500 µL of L6

Buffer (Severn Biotech, 20-8600-15) to 90 µL sample (on ice), then vortexed for a further 3 seconds

and incubated at room temperature for 15 minutes (as per UKHSA Guidelines). Positive and negative

samples provided by UKHSA and BOTT samples underwent nucleic acid extraction using the

QIAsymphony DSP Virus/Pathogen Mini Kit (QIAGEN), in combination with the QIAsymphony SP

instrument, according to the manufacturer's instructions (Extraction Protocol: Cell_Free 200 with

elute volume of 60 µl). Additional extraction protocols (i.e. Macherey Nagel Nucleospin RNA Kit;

Kingfisher) were also used and compared for extraction of RNA from control materials and found to

provide comparable results to the QIASymphony (results not shown). For all extractions, samples

were spiked with the RNA bacteriophage MS2 (genome accession V00642.1, provided by Langford

Vets) at a final dilution of 10^{-8} of a lab grown stock to give a CT value of ~24. This was used as an

internal control for the extraction process, RT and qPCR.

RT-qPCR Assay Design and Initial Optimisation

With the exception of the E gene, where primer and probe sequences were unchanged from those

described previously (5), RT-qPCR assays and novel primers were designed for different genes of

SARS-CoV-2 as previously described (6), with details provided in Table 1. Novel SARS-CoV-2 primers

and probes from different regions of the nucleocapsid (N gene; denoted N2 and N6 (in-house

numbering)) were designed using Primer3 (7) using a consensus multiple sequence alignment of 658

SARS-CoV-2 N gene sequences downloaded from GenBank (sequences downloaded April 2020 and

Alignments performed in May 2020). Amplicon secondary structure was predicted using Mfold (8).

One fluorophore/quencher combination (FAM-BHQ-1) was included for the probe of each of the

SARS-CoV-2 target genes. SARS-CoV-2 primer/probes were multiplexed in pair combinations with

MS2 bacteriophage primers and probe (with a Cy5-BHQ3 fluorophore/quencher) to allow for

detection of an internal extraction control. All primers and probes were HPLC purified and ordered

from Metabion Ltd. (Germany).

Table 1.   **Primer and Probe Sequences. Full sequences and modifications for primers and**
probes in the SARS-CoV-2 N6/E/MS2 Multiplex Assay as well as the final concentration used in
RT-qPCR. * Position indicated is relative to SARS-CoV-2 isolate Wuhan-Hu-1 genome (accession
MN908947.3) or MS2 phage genome (accession V00642.1). Probes were labelled with 6-
Carboxyfluorescein (6-FAM) and Black Hole Quencher 1 (BHQ1) or Cyanine 5 (CY5) and Black
Hole Quencher 3 (BHQ3).

Name	Position *	Sequence (5'-3')	Concentration used in RT-qPCR
CoV2_N2_F	28499-28520	ACCAATAGCAGTCCAGATGACC	200nM
CoV2_N2_R	28560-28583	GATCTTTCATTTTACCGTCACCAC	200nM
CoV2_N2_P	28534-28559	6-FAM-CCGAAGAGCTACCAGACGAATTCGTG -BHQ-1	100nM
CoV2_N6_F Forward Primer	28919-28938	GATGCTGCTCTTGCTTTGCT	200nM
CoV2_N6_R Reverse Primer	28980-29001	CCTTGTTGTTGTTGGCCTTTAC	200nM
CoV2_N6_P Probe (FAM)	28940-28968	6-FAM-CTGCTTGACAGATTGAACCAGCTTGAGAG-BHQ-1	100nM
E_Sarbeco_F1 Forward Primer	26269-26294	ACAGGTACGTTAATAGTTAATAGCGT	200nM
E_Sarbeco_R2 Reverse Primer	26360-26381	ATATTGCAGCAGTACGCACACA	200nM
E_Sarbeco_P1 Probe (FAM)	26332-26357	6-FAM- ACACTAGCCATCCTTACTGCGCTTCG-BHQ-1	100nM
MS2 F Forward primer	2163-2180	ATGGGGCACAAGTTGCAG	50nM
MS2 R Reverse primer	2226-2245	GGGTAACGGTTGCTTGTTCA	50nM
MS2 P Probe (CY5)	2183-2206	Cy5-TGCAGCGCCTTACAAGAAGTTCGC-BHQ-3	50nM

A one-step RT-qPCR assay was used for all candidate target combinations, where each reaction well

contained 6.25 µL TaqPath 1- Step RT-qPCR Master Mix, CG (ThermoFisher Scientific), 1 µL of 25X

primer and probe mix, 7.75 µL molecular grade water and 10 µL total nucleic acid extract. Each RT-

qPCR run contained the following controls: nucleic acid extracted SARS-CoV-2 positive control of the

ancestral Wuhan strain (1:10,000 dilution; see Table 2 below); extracted L6 Buffer (negative control);

molecular grade water (W4502-1L, Sigma-Aldrich; non-template control).

Total nucleic acid and master mixes were pipetted using the automated QIAgility HEPA / UV

Instrument (QIAGEN) before being run on the QuantStudio 7 Real-Time PCR System (Applied

Biosystems). Thermal cycling consisted of 25°C for 2 mins, 50°C for 15 mins and 40 cycles of 95°C for

10 sec, 60°C for 30 sec (as per guidelines for TaqPath, ThermoFisher Scientific).

Reaction Efficiencies and Limit of Detection

Reaction efficiencies were determined using dilutions of SARS-CoV-2 ancestral strain (Wuhan, see

Table 1) RNA purified from tissue culture supernatant of SARS-CoV-2 infected cells as described

previously (9). Limit of detection (LoD) was determined using RNA extracted from a commercial

source of synthetic SARS-CoV-2 of known copy number (AccuPlex SARS-CoV-2 reference material,

4162 copies/ml, SeraCare Life Sciences Inc, USA). Purified RNA was diluted two-fold and between 8

and 16 replicates run at each dilution.

Sensitivity and Specificity

Diagnostic sensitivity and specificity were determined using a panel of 60 SARS-CoV-2 positive and

59 SARS-CoV-2 negative samples kindly supplied by UKSHA Bristol and previously tested in their

laboratories using a commercial kit (RealStar®SARS-CoV-2 RT-PCR Kit 1.0, Altona Diagnostics,

Hamburg), which acted as the reference standard. To calculate confidence intervals for sensitivity

and specificity values, the Clopper-Pearson method was used.

SARS-CoV-2 Variants and Sequencing

To test the performance of the N6/E/MS2 and N2/N6/MS2 assays on different SARS-CoV-2 variants,

viruses were obtained from multiple sources; details of these are provided in Table 2. In all cases,

the specific strain ID was determined using Illumina mRNA sequencing (performed by staff at the

Genomics Facility, University of Bristol) using the following kits/methods. A total volume of 8ul of

total RNA sample was taken directly into the NEBNext® Artic SARS-CoV-2 Library Prep Kit Part

Number E7650S/L (New England Biolands, Inc. Ipswich, MA) or the NEBNext® Artic FS SARS-CoV-2

Library Prep Kit Part Number E7658S/L followingthe manufacturer's instructions without deviation.

Briefly, the protocol involved cDNA synthesis, followed by targeted cDNA amplification using two

primer pair pools, A and B, to span the SARS-CoV-2 virus genome (primer schemes based on the

original   work   of   the   ARTIC   Network   (https://github.com/joshquick/artic-

ncov2019/blob/master/primer_schemes/nCoV-2019/V3/nCoV-2019.tsv). Primer schemes evolved

throughout the running of this experiment and the latest version was used to represent current

variants. Resulting amplicons from the two PCRs (A and B) per sample were pooled and cleaned up

using a bead-based method (beads supplied in library preparation kit) before checking resulting

amplicon size (1 in 10 dilution) on the Agilent TapeStation High Sensitivity DNA 1000 Screentape

Assay (Agilent Technologies, Santa Clara, CA) on the Agilent 2200 TapeStation instrument. The

amplicons were end-repaired prior to NEBNext 'Adaptor for Illumina' ligation, a further bead-based

clean-up, PCR enrichment of the adaptor-ligated DNA to incorporate a unique dual-index barcode

into each sample, and a final bead-based clean-up of the final libraries.

Final libraries were quantified using the Thermofisher High Sensitivity dsDNA Qubit assay

(Thermofisher Scientific, Inc., Waltham, MA) and validated using the TapeStation (Agilent) High

Sensitivity DNA1000 screentape assay on the Agilent 2200 TapeStation instrument. The libraries

were normalised to 4nM, pooled equimolarly, and diluted to 10pM for sequencing on the Illumina

MiSeq instrument using an Illumina Version 3, 600-cycle sequencing kit (Illumina Inc, San Diego, CA)

and MiSeq Control Software Version 2.5.0.5. A spike-in of a PhiX control library Version 3 (Illumina

Inc., San Diego, CA) at 10 percent was added to the sequencing run to assist with accurate base

calling for a biased library. Primary data analysis was completed by on-board MiSeq Software (RTA

Version 1.18.54.0) and FASTQ Creator (MiSeq Reporter Version 2.5.1.3). Data visualisation and QC

metrics were assessed using Illumina BaseSpace Sequence Hub (basespace.illumina.com).

Table 2.   Details of sequences and sources of the various SARS-CoV-2 strains used to
validate RT-qPCR assays and as positive control material.

Strain name	Details of Strain ID/sequence/annotation	Source
Ancestral strain (Wuhan)	REMRQ0001 GenBank accession number MW041156.1	Isolated in April, 2020 as previously described (9)
Alpha	hCoV-19/England/204690005/2020; GISAID ID: EPI_ISL_693401;	Kindly provided by Professor Wendy Barclay, Imperial College, London and Professor Maria Zambon, UK Health Security Agency
Delta	hCoV-19/England/BRS-UoB-233/2021; GISAID ID: EPI_ISL_15250227,	Isolated from a donor sample as previously described (10)
Gamma	hCoV-19/England/520336_B1_P0/2021, GISAID ID: EPI_ISL_2080492	
Omicron BA.1	hCoV-19/England/BRS-UoB-942/2022; GISAID ID: EPI_ISL_18150332	Isolated from donor sample obtained from the Bristol Biobank
Omicron BA.2	hCoV-19/England/BRS-UoB-317/2022; GISAID ID: EPI_ISL_18303155	Isolated from donor sample obtained from the Bristol Biobank
Omicron BA.4	hCoV-19/England/BRS-UoB-64/2022; GISAID ID: EPI_ISL_18151340	Isolated from donor sample obtained from the Bristol Biobank
Omicron JN.1	Only Spike gene sequence and confirmed to have JN.1 lineage defining mutations.	Isolated from donor sample obtained from the Bristol Biobank

Positive samples identified from the field testing of BOTT OPS collected in 2020, were sequenced

and strain-typed in Exeter by COG-UK using MinION sequencing as described in the published

protocol (11).

Data Analysis

Data analysis and presentation were performed with GraphPad Prism (v10). The limit of detection

was defined as the copy number at which >95% of replicates were detected, as determined by fitting

a 4 parameter logistic regression (least squares fit) on a plot of fraction samples positive against

known copies of virus per RT-qPCR, and interpolating the value (and 95% confidence intervals where

fraction positive = 0.95). RT-qPCR data from test samples were exported using either the

QuantStudio Real-Time PCR Software v1.3 or the Qtower3. Any CT in the FAM channel for candidate

assays indicated SARS-CoV-2 RNA had been detected. Cut-off values were determined using

multiple repeats of serial dilutions of positive control material, and for all assays CT values <=35

were considered positive. Thresholds were set to 0.01 for SARS-CoV-2 N6 and E, which are

indistinguishable since they both use the FAM channel and 0.01 for the MS2 bacteriophage used the

Cy5 channel.

Results

Initial characterisation of N6/E/MS2 and N2/N6/MS2 RT-qPCR assays

We initially developed single plex assays to detect different genes from the SARS-CoV-2 genome,

including two areas of the gene for nucleocapsid (N2, N6) and one each of envelope (E) and the RNA-

dependent-RNA-polymerase (RdRp). Of these potential targets, primers binding to the E and N genes

were found to provide the best efficiencies (data not shown). Two candidate dual target assays were

developed, to improve on the sensitivity over single plex assays), including incorporation of control

target (MS2) to assess for extraction, RT and qPCR performance. The candidate assays selected for

further characterisation were: N6/E/MS2 and N2/N6/MS2.

We sought to determine limits of detection (LoD) and reaction efficiencies of the two candidate

assays, using the ancestral (Wuhan) SARS-CoV-2. Both assays showed suitable efficiencies as

multiplex assays and were compared to use of the targets in a single plex assay (Supplementary

Figure 1). However, the N2/N6/MS2 assay had a greater reaction efficiency at 98% with R^2= 0.995,

compared to N6/E/MS2 at 95% (R^2=0.998) (Supplementary Figure 1). The LoD was determined using

multiple aliquots of samples with known copy numbers (Figure 1) and was defined as the copy

number at which 95% of replicates were detected.

215
216
Figure 1. Limit of detection (LoD) was determined using RNA extracted from a commercial
source of synthetic SARS-CoV-2 of known copy number (Accuplex SARS-CoV-2 reference
material, 4162 copies/ml, Seracare Life Sciences Inc, USA).   Purified RNA was diluted
two-fold and between 8 and 16 replicates were run at each dilution. A) N6/E/MS2 and
B) N2/N6/MS2 plots show the RNA copies per reaction on the axis and the % positive
readings on each assay on the y axis with a curve plotted using 4 parameter logistic
regression (least squares) with 90% confidence intervals also plotted. The LoD is defined
as the copy number of virus at which 95% of replicates are detected as determined by
interpolating the values when the fraction positive =0.95; these are stated in a panel
within each plot, along with the R squared (R^2) for each assay.

The N6/E/MS2 assay displayed a lower LoD of 1.39 RNA copies per RT-qPCR (1.25-1.55, 90% CI)

compared to N2/N6/MS2 at 2.79 (1.87 –6.35, 90% CI). At the highest dilution of virus at 0.65 copies

per 10 µl reaction, the N6/E/MS2 assay detected almost double the number of positive results with

very low viral copies compared to N2/N6/MS2 assay (75% compared to 37.5%).

Sensitivity and Specificity

To determine assay performance of the two candidate assays, we performed a blind comparison

with a gold standard dual target RT-qPCR kit which was commonly used at the time of assay

development (2020), the Altona kit for monoplex detection of the S and E genes. Using a randomised

set of 60 positive and 59 negative dual NPS and OPS samples provided by UKHSA Bristol (Severn

Pathology, Southmead Hospital), we extracted RNA and tested the two candidate RT-qPCR assays.

Both assays had high sensitivity for the positive samples (98.33% (59/60) for N6/E/MS2 and 96.7%

(58/60) for N2/N6/MS2), and 100% (59/59) specificity, see Table 3).

Table 3. Diagnostic performance of candidate RT-qPCR assays. The sensitivity and specificity of the N6/E/MS2 and N2/N6/MS2 assays were compared to the results from the Altona assay (reference standard) used by the UKHSA.

Candidate Assay	Positive Samples Detected as Positive out of Total	Negative Samples Detected as Positive out of Total	Sensitivity (95% CI)	Specificity (95% CI)
N6/E/MS2	59/60	0/59	98.3% (91.06-99.96)	100% (93.94-100)
N2/N6/MS2	58/60	0/59	96.7% (88.47-99.59)	100% (93.94-100)

Thus, both assays provided comparable diagnostic performance when compared to a widely used commercial assay. When comparing the two candidate assays, the lower limit of detection and slightly higher observed sensitivity led us to select the N6/E/MS2 assay as the better of the two in-house assays.

Use of N6/E/MS2 assay for detection of SARS-CoV-2 from samples collected in early 2020

The N6/E/MS2 PCR assay was used to detect SARS-CoV-2 in 1803 OPS samples with unknown SARS-CoV-2 infection status collected from healthy (i.e. asymptomatic) 16-18 year-old secondary school students in several sites across the UK between January and March 2020. These samples were collected as part of a large study of the impact of two meningococcal vaccines on throat colonisation in healthy teenagers (3). Of these samples, 4/1803 tested positive on the assay with each positive sample coming from a different site across the UK (Bristol, Cardiff, Glasgow and Maidstone) collected in March. The prevalence across the full study sampling period was 0.2% (95% CI 0.06 – 0.57) but in March the incidence was 0.45%; 2 of the 427 (0.47%, 95% CI 0.06% - 1.7%) swabs taken during the week beginning 9th March were positive and 2 of the 459 (0.44%, 95% CI 0.05% - 1.57%) swabs taken during the week beginning 16th March were positive (Supplementary Figure 2). To confirm detection of SARS-CoV-2 and identify strains, sequencing was performed on these samples by COG-

UK in Exeter, which identified the samples as being of 3 lineages/clades of the ancestral Wuhan virus

(12), results are shown in Table 4. These results suggest that the virus was in circulation at very low

levels within secondary schools across England before the first lockdown period was implemented in

March 2020.

Table 4.   **Details of the 4 OPS samples identified as positive from samples taken from**
children in March 2020 (total of n=1803 tested) as part of the Be on The Team (BOTT) study.

BOTT Site	Sample Date	N6/E PCR C_T	Virus lineage as identified by Sequencing (COG-UK, 2020)
Bristol	12/03/2020	28.448	B
Cardiff	16/03/2020	26.981	B.1.1
Glasgow	16/03/2020	28.574	B
Maidstone	10/03/2020	24.697	B.40

Detection of SARS-CoV-2 Variants

To confirm comparable detection of the key SARS-CoV-2 variants of interest (relating to different

stages of the pandemic) in different respiratory samples, we initially performed 10-fold serial

dilutions in culture media of four SARS-CoV-2 variants (Wuhan, Delta, Alpha and Omicron BA.1)

which were important at the time. Samples were extracted using QIAsymphony and run on the

N6/E/MS2 RT-qPCR assay, and were found to have good detection across the Ct range and suitable

efficiency (Supplementary Figure 2). This was repeated more recently using Gamma and Omicron

subvariants, including a virus of the JN.1 lineage; similar assay performance was observed for the

variants (Figure 2B).

Figure 2. Detection of different SARS-CoV-2 Omicron and Gamma sub-variants with the
N6/E/MS2 RT-qPCR assay. Detection of 10-fold serial dilutions of stock of SARS-CoV-2
Gamma strain (light blue dots), Omicron BA.1 (black squares), Omicron BA.2 (red
triangles), Omicron BA.4 (green upside-down triangles), and Omicron JN.1 (dark blue
diamonds) viruses. The X-axis represents the 10-fold dilution and the y axis the Ct value.
The average Ct value for each dilution (3-6 replicates per dilution) is plotted along with
the line of best fit after performing a simple linear regression.

The results demonstrate comparable performance for detection for each of the variants tested (also

see supplementary Table 1), indicating applicability of the assay for a range of SARS-COV-2 variants

representing several time periods since the start of the COVID-19 pandemic.

Discussion

Quantitative RT-qPCR assays remain the gold standard approach for detection of SARS-CoV-2 in

clinical samples. Here we present the development, performance characteristics and deployment of

a novel dual-target SARS-CoV-2 RT-qPCR assay, N6/E/MS2. The assay demonstrates good accuracy

for detection of SARS-CoV-2 in control and clinical samples as well as preserved ability to detect the

major SARS-CoV-2 variants including the recently circulating and highly transmissible Omicron JN.1

subvariant (13).

Since the start of the COVID-19 pandemic, multiple commercial and in-house RT-qPCR assays have

been developed for the detection of SARS-CoV-2. As others have reported, we found that sensitivity

for detection can be improved by targeting multiple genes (14).. It is thus common for SARS-COV-2

RT-qPCR assays to be multiplex, with most currently-used commercial or recommended assays

targeting a combination of the E, S, ORF1a and RdRp genes (14, 15). Different regions of the N gene

are also common targets, such as in the assays published by the CDC (16). However, the N gene

primers presented here are unique and offer high technical performance. With the use of M-Fold

and primer3, we demonstrate how use of prediction software packages for the design of novel RT-

qPCR assays can achieve high efficiencies and lead to good assay performance. We have also

demonstrated the ability of the N6/E/MS2 assay to detect key SARS-CoV-2 variants of concern,

including more recently circulating Omicron variant lineages. Conserved performance in these

distinct variants of concern may relate to the choice of two distinct target gene regions (N6, E) which

have remained relatively conserved in comparison with the S gene (17).

To field-test the assay in the early stages of the first UK wave of infections, ~1800 samples from

school age children collected in early 2020 (BOTT) were tested, and only four were found to be

positive for SARS-CoV-2 (0.2%). Interestingly, these four positive samples were each identified from

different sites, suggestive of minimal but widespread transmission of SARS-CoV-2 in UK in this age

group at that time, despite the reported number of cases in the UK being only approximately 250 on

March 10[th] rising to ~1,500 cases on March 15th i.e. when the samples were collected (18). This

dispersion of isolated cases is supported by global transmission modelling studies that predict that

community transmission was probably common in multiple parts of Europe as early as early January

2020 (19), and suggest that large numbers of asymptomatic infections went undetected at this time.

Furthermore, sequencing data and modelling showed that there were multiple importation events

of index cases into the UK in late February/Early March of 2020 from other parts of Europe (20). This

is also compatible with our detection of isolated positive cases in 4 distinct locations in the UK more

or less simultaneously. The sampling dates of the positive samples were close to the start of the

introduction of a national lock down in the UK, on 23rd March 2023 (18), which included school

closures. However, given the low rate of infections detected in this cohort, it is unlikely that schools

were driving a lot of the transmission at that time.

This study has some limitations. Whilst we have tested for sensitivity and specificity compared to a

gold standard assay, these data were collected in the early stages of the pandemic, with the

ancestral Wuhan virus, and we do not have assay comparison data for later variants. We included an

internal RNA control (MS2) to monitor the efficiency of the extraction process. However, unlike

other assays we did not include a human control probe to control for effective sample taking. Thus,

some negative results observed may have been due to inadequate sampling.

In summary, the N6/E/MS2 assay is sensitive and specific for SARS-CoV-2 and has since been used to

detect SARS-CoV-2 in respiratory samples collected as part of multiple studies, including a

surveillance project in Bristol schools ("CoMMinS") (21), and hospital and community-based

surveillance studies of adult lower respiratory tract disease in Bristol, UK ("AvonCAP") (22). As more

recent SARS-CoV-2 variants have shown genetic drift, it is important that those using the assay

perform regular checks on assay performance with current/recent variants.

References

1.   Regen F, Eren N, Heuser I, Hellmann-Regen J. A simple approach to optimum pool size for
pooled SARS-CoV-2 testing. International Journal of Infectious Diseases. 2020;100:324-6.
2.   Vogels CBF, Brito AF, Wyllie AL, Fauver JR, Ott IM, Kalinich CC, et al. Analytical sensitivity and
efficiency comparisons of SARS-CoV-2 RT-qPCR primer-probe sets. Nat Microbiol. 2020;5(10):1299-
305.
3.   Carr JP, MacLennan JM, Plested E, Bratcher HB, Harrison OB, Aley PK, et al. Impact of
meningococcal ACWY conjugate vaccines on pharyngeal carriage in adolescents: evidence for herd
protection from the UK MenACWY programme. Clinical Microbiology and Infection.
2022;28(12):1649.e1-.e8.

4.   Carr J, Plested E, Aley P, Camara S, Davis K, MacLennan JM, et al. 'Be on the TEAM' Study
(Teenagers Against Meningitis): protocol for a controlled clinical trial evaluating the impact of
4CMenB or MenB-fHbp vaccination on the pharyngeal carriage of meningococci in adolescents. BMJ
Open. 2020;10(10):e037358.

5.   Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DK, et al. Detection of 2019
novel coronavirus (2019-nCoV) by real-time RT-PCR. Eurosurveillance. 2020;25(3):2000045.

6.   Peters IR, Helps CR, Hall EJ, Day MJ. Real-time RT-PCR: considerations for efficient and
sensitive assay design. Journal of Immunological Methods. 2004;286(1):203-17.

7.   Untergasser A, Cutcutache I, Koressaar T, Ye J, Faircloth BC, Remm M, et al. Primer3—new
capabilities and interfaces. Nucleic Acids Research. 2012;40(15):e115-e.

8.   Zuker M. Mfold web server for nucleic acid folding and hybridization prediction. Nucleic
Acids Research. 2003;31(13):3406-15.

9.   Daly JL, Simonetti B, Klein K, Chen K-E, Williamson MK, Antón-Plágaro C, et al. Neuropilin-1 is
a host factor for SARS-CoV-2 infection. Science. 2020;370(6518):861-5.

10.   Erdmann M, Williamson MK, Jearanaiwitayakul T, Bazire J, Matthews DA, Davidson AD.
Development of SARS-CoV-2 replicons for the ancestral virus and variant of concern Delta for
antiviral screening. bioRxiv. 2022:2022.10.11.511804.

11.   Quick J. nCoV-2019 sequencing protocol v3 (LoCost) V.3 2020 [Available from:
https://www.protocols.io/view/ncov-2019-sequencing-protocol-v3-locost-bp2l6n26rgqe/v3.

12.   Harper H, Burridge A, Winfield M, Finn A, Davidson A, Matthews D, et al. Detecting SARS-
CoV-2 variants with SNP genotyping. PLOS ONE. 2021;16(2):e0243185.

13.   Yang S, Yu Y, Xu Y, Jian F, Song W, Yisimayi A, et al. Fast evolution of SARS-CoV-2 BA.2.86 to
JN.1 under heavy immune pressure. The Lancet Infectious Diseases. 2024;24(2):e70-e2.

14.   Peñarrubia L, Ruiz M, Porco R, Rao SN, Juanola-Falgarona M, Manissero D, et al. Multiple
assays in a real-time RT-PCR SARS-CoV-2 panel can mitigate the risk of loss of sensitivity by new
genomic variants during the COVID-19 outbreak. Int J Infect Dis. 2020;97:225-9.

15.   Mostafa HH, Hardick J, Morehead E, Miller J-A, Gaydos CA, Manabe YC. Comparison of the
analytical sensitivity of seven commonly used commercial SARS-CoV-2 automated molecular assays.
Journal of Clinical Virology. 2020;130:104578.

16.   Lu X, Wang L, Sakthivel SK, Whitaker B, Murray J, Kamili S, et al. US CDC Real-Time Reverse
Transcription PCR Panel for Detection of Severe Acute Respiratory Syndrome Coronavirus 2. Emerg
Infect Dis. 2020;26(8):1654-65.

17.   Basak S, Kayet P, Ghosh M, Chatterjee J, Dutta S. Emergence of Genomic Diversity in the
Spike Protein of the "Omicron" Variant. Viruses. 2023;15(10).

18.   Wikipedia. Timeline of the COVID-19 pandemic in the United Kingdom (January–June 2020):
Wikipedia; 2024 [Available from: https://en.wikipedia.org/wiki/Timeline_of_the_COVID-
19_pandemic_in_the_United_Kingdom_(January%E2%80%93June_2020)#cite_note-56.

19.   Davis JT, Chinazzi M, Perra N, Mu K, Pastore YPA, Ajelli M, et al. Cryptic transmission of SARS-
CoV-2 and the first COVID-19 wave. Nature. 2021;600(7887):127-32.

20.   du Plessis L, McCrone JT, Zarebski AE, Hill V, Ruis C, Gutierrez B, et al. Establishment and
lineage dynamics of the SARS-CoV-2 epidemic in the UK. Science. 2021;371(6530):708-12.

21.   Looker KJ, McClenaghan E, Judd A, Pierotti L, Downing H, Phelan J, et al. Acute and long-term
outcomes of SARS-CoV-2 infection in school-aged children in England: Study protocol for the joint
analysis of the COVID-19 schools infection survey (SIS) and the COVID-19 mapping and mitigation in
schools (CoMMinS) study. PLoS One. 2024;19(5):e0303892.

22.   Hyams C, Begier E, Garcia Gonzalez M, Southern J, Campling J, Gray S, et al. Incidence of
acute lower respiratory tract disease hospitalisations, including pneumonia, among adults in Bristol,
UK, 2019, estimated using both a prospective and retrospective methodology. BMJ Open.
2022;12(6):e057464.

Acknowledgements

405   We acknowledge both the Bristol Genomics Facility and COGG UK (Exeter) as the sequencing

406   providers used for this research. We also acknowledge the Bristol Biobank for support in acquiring

407   control samples for assay development.

Funding

This work was funded in part by University of Bristol Emergency funding awarded to AF and Langford

Vets Ltd research funds awarded to PL and CH. AF, ADD, BMA & AH acknowledge funding towards

this work from Elizabeth Blackwell Institute, funded in part by the Wellcome Trust [Grant number

204813/Z/16/Z] with additional support from Bristol Alumni and Friends (DARO). We also

acknowledge funding via a 2020/21 COVID 19 Research Award Funded by the NIHR Oxford

Biomedical Research Centre (BRC) for the testing of the Be On The Team (BOTT) OPS STGG samples.

'Be on the TEAM', 'Evaluating the effect of 'MenB' vaccination on meningococcal carriage', PR-R18-

0117- 21001 (BOTT) was funded by the National Institute for Health Research (NIHR) Policy Research

Programme. We also acknowledge Meningitis Now funding which enabled the storage and curation

of throat swab samples from the Be on the Team study, and The NIHR Oxford Biomedical Research

Centre and the Rockinghorse charity, Brighton (RH20-13) who co-funded the SARS-CoV-2 qPCR

analysis of the samples.

Competing Interests

AF was lead or co-investigator of several trials of COVID-19 vaccines including the following:

Oxford-AstraZeneca, Valneva, Sanofi-GSK. He also was lead or co-investigator on

epidemiological studies of SARS CoV2 infection funding by the UK Government and Pfizer.

All related research funding was paid to his employer. AF also undertakes paid consultancy

for several vaccine manufacturers and developers, although only work for Sanofi relates to

COVID-19.