Title: A multitude of dyes but unsatisfactory staining: Physico-chemical profile and factors associated with staining quality of Romanowsky-type stains used for malaria microscopy in Plateau State, Nigeria.

Authors: Stephen Eghelakpo Akar[1,2,3,5], David Magbagbeola Dairo[2], Remi Kehinde[2], Tosan Erhabor[5], Mary Ogheneruemu Akar[4], Godwin Ntadom[8], Treasure Obianuju Okoye Wariboko[10], Patrick Nguku, Shakir Balogun, Oladosu Oladipo[7,12], Wellington Oyibo[7], Olufemi. Ajumobi, [1,9], Ikeoluwapo Ajayi[2]

Author's affiliation:

1. Nigeria Field Epidemiology and Laboratory Training Program, Abuja, Nigeria
2. Department of Epidemiology and Medical Statistics, Faculty of Public Health, College of Medicine, University of Ibadan, Ibadan, Nigeria
3. Department of Eco-epidemiology, Institute of Tropical Medicine, Nagasaki University, Nagasaki, Japan
4. Department of Laboratory Services, Asokoro General Hospital, Abuja, Nigeria
5. Medical Laboratory Science Council of Nigeria
6. Nigeria Centre for Disease Control
7. ANDI Centre of Excellence for Malaria Diagnosis, National Malaria Quality Assurance Laboratory, College of Medicine, University of Lagos
8. School of Community Health Sciences, University of Nevada Reno, USA
9. Henry Jackson Foundation for Medical Research International, Abuja, Nigeria
10. Department of Medical Microbiology and Parasitology, College of Medicine, University of Ibadan, Ibadan, Nigeria
11. Federal Ministry of Health, Abuja, Nigeria
12. Pure and Applied Biology Programme, Bowen University, Iwo, Osun State, Nigeria

Correspondence author email: akarstephen@gmail.com

Abstract

Introduction: Blood smear microscopy using Romanowsky-type remains the gold standard for malaria diagnosis in resource-poor settings. Good quality Romanowsky-type stains are a major determinant of the quality of malaria microscopy. The quality of malaria microscopy done in Nigeria has been queried due to reported high positivity rate compared to RDT results. There is paucity of studies done to assess the composition and staining quality of Romanosky-type stains used for malaria microscopy in Nigeria. We therefore determined the physico-chemical characteristics and factors associated with staining outcome in blood smears.

Methods: We conducted a cross-sectional study among 92 laboratory facilities selected by multi-stage sampling across three senatorial districts in Plateau State. We analyzed 94 stain samples, including reference stains for comparison. We subjected the stains to ultraviolet-spectroscopy, high performance liquid chromatography, gas chromatography, ashing and pH measurement. We conducted univariate, bivariate and multivariate binary logistic regression analysis at $p < 0.05$.

Result: Of the 94 stains samples, 38(40.4%) were Field stains, 36(38.3%) were Giemsa stains, and 20(21.3%) were Leishman stains. Median pH of stock stain solution was 6.6 (IQR = 6.5-6.7), 7.4 (IQR = 7.1-7.5) and 7.0 (IQR = 6.7-7.1, $p < 0.001$) for Field, Giemsa and Leishman stains respectively. In 20.0% (4/20) of Leishman stains, absolute methanol was used as solvent and in 63.9% (23/36) of Giemsa stains, ratio of glycerol to absolute methanol was 1:1. Overall, 23.7% (9/38), 33.3% (12/36) and 25.0% (5/20) of Field, Giemsa and Leishman stains yielded satisfactory staining. Independent predictors of staining quality were . Proportion of Azure B>20% (aOR = 15.1, 95% CI = 2.62-115, p=0.004); Azure B to Eosin Y ratio >1.5 (aOR = 4.55, 95% CI = 1.03-25.10, p=0.057); and using alcoholic solvent for stain preparation (aOR = 27.70, 95% CI = 2.43-462.00, p=0.012).

Conclusion: Romanowsky-type stains used for malaria microscopy in Plateau State were of variable composition and mostly lacked the RGE-relevant dyes required for good nuclear and cytoplasmic differentiation of malaria parasites in blood smears. A stain prepared from component dye species yielded superior staining quality in blood smear.

Introduction

Malaria, a mosquito-borne parasitic disease caused by *Plasmodium* species, has remained a major global public health concern (Poespoprodjo et al., 2023). Worldwide, the number of malaria cases was estimated to be 263 million with an incidence of 60.4 cases per 1000 population, accounting for an estimated 597 000 deaths, with a mortality rate of 13.7 per 100000 in 2022. In 2023 alone, the 11 high burden to high impact (HBHI) countries- Burkina Faso, Cameroon, the Democratic Republic of the Congo, Ghana, India, Mali, Mozambique, Niger, Nigeria, Uganda and the United Republic of Tanzania - accounted for 66% of all malaria cases and 68% of deaths globally. Between 2017 and 2023, the incidence of malaria cases in these 11 countries increased by 13.8%, from 152 million to 173 million, and mortality increased by 2.3%, from 399 000 to 408 000. (WHO, 2024). In 2022, the African Region bore the heaviest malaria burden, with 94% of cases and 95% of deaths globally, representing 233 million malaria cases and 580 000 of deaths, a small reduction compared to 2021("Malaria | WHO | Regional Office for Africa," n.d.). Nigeria accounts for the highest incidence of global malaria burden compared to any other country, with 27% of the global estimated malaria cases and 31% of the estimated deaths, as well as an estimated 55% of malaria cases in West Africa in 2022. Nigeria's large population, combined with other factors such as sanitation and ecological characterisitcs that favor mosquito breeding, account for the endemicity of malaria transmission ("Malaria in Nigeria: Statistics & Facts | Severe Malaria Observatory," n.d.). WHO has recommended prompt malaria diagnosis either by microscopy or rapid diagnostic tests (RDTs) for all patients with suspected malaria before administering treatment. Early and accurate diagnosis is essential both for effective case management and for malaria surveillance. Parasite-based diagnostic testing significantly reduces illness and death by enabling health providers to swiftly distinguish between malarial and non-malarial fevers and select the most appropriate treatment. It improves the overall management of patients with febrile illnesses and may also help reduce the emergence and spread of drug resistance ("Global Malaria Programme," n.d.). Although the introduction of rapid diagnostic tests (RDTs) has demonstrated the potential to ameliorate some diagnostic limitations associated with light microscopy, RDT kits are still faced with challenges such as inability to quantify parasite load for patient treatment monitoring and follow up (NMIS,

2015). Hence, Giemsa stain microscopy has remained the gold standard for routine parasite-based malaria diagnosis in resource-poor countries (Idowu Olufunmilayo et al., 2015; Smith et al., 2001). However, Mokuolu et al. (2016) have questioned the higher positivity rate of malaria diagnosis by microscopy in Nigeria compared to RDT as they opined that higher prevalence due to microscopy does not reflect the progress and gains that have been made by extant control efforts (Mokuolu et al., 2016). Microscopy involves the staining of either thick or thin blood films with the Romanowsky-type stains such as Giemsa, Leishman or Field stain, however, the WHO recommends Giemsa stain (WHO, 2016a). It has been shown that there is a strong correlation between the physico-chemical properties of these stains and their performance when applied to blood smears (Friedrich K, Seiffert W, 1990; Krafts K, 1993; Marshall, 1978a; Wittekind, 2002). The staining process follows an ordered chemical interaction between stain components and cellular constituents in blood (R. W. Horobin, 2011; Horobin and Walter, 1987a; Wittekind DH and Gehring T, 1985). Unlike malaria RDT kits, there is no known program for evaluating the quality of microscopy reagents like the WHO-FIND program that certifies RDT kits (Ashraf et al., 2012; Steven A. Harvey et al., 2017; WHO-FIND, 2014). In North America, the Biological Stains Commission certifies biological stains for diagnostic use, but such is non-existent in WHO Africa Region. Evidence from the field shows that many facilities source their microscopy reagents, including stains from the open market. The supply of in vitro diagnostic reagents and equipment for microscopy and other laboratory tests is poorly regulated in developing countries, where evidence of stocking, sales and supply of sub-standard diagnostic products has been demonstrated (Mori et al., 2011; Nora Engel, Keri Wachter, Madhukar Pai, Jim Gallarda and Isabelle Celentano, 2016; Rugera et al., 2014; Ruth McNerney and W.Peeling, 2015). The accuracy of malaria microscopy depends on maintaining a high level of staff competence and performance, ensuring good-quality reagents and equipment at all levels and regular external quality assessment (Ilah Garba et al., 2016; Munene et al., 2017; Ria R. Ghai et al., 2016; WHO, 2016b). Microscopy stain solutions are either prepared by laboratory personnel locally using component dye powders (in-house stains) or purchased as ready-to-use aqueous stock solution (commercially prepared). The quality of the Romanowski-type stains used for malaria microscopy cannot be guaranteed if their physico-chemical properties are compromised (R. W. Horobin, 2011; Marshall et al., 1978, 1975a; Penney et al., 2002; Wittekind, 2002). Poor quality stains have the potential to compromise the quality of malaria microscopy, even if personnel are competent and have access to efficient

equipment (Marshall et al., 1978, 1975a). Poor quality microscopy could result in either over

diagnosis or under-diagnosis of cases with an attendant unnecessarily increase disease

prevalence, over prescription of ACTs, and increase in drug resistance. Under-diagnosis of cases,

on the other hand, may lead to under-reporting of cases and increased morbidity and mortality

(Altaras et al., 2016; Kahama-maro et al., 2011; Miller and Sikes, 2015). In most secondary and tertiary

facilities, microscopy continues to be the preferred method for parasite-based diagnosis of

malaria (Mokuolu et al., 2016). However, internal quality assurance procedures tend to vary

between facilities with an attendant variation in test results when compared from laboratory to

laboratory. Reagents for diagnostics are sourced from various sources in the open market, a large

proportion of which do not fulfill regulatory requirements for the importation, stocking, storage

and supply of in vitro diagnostics (MLSCN, 2018). The quality of staining quality produced by the

Romanowski-Giemsa stains depends on a number of factors, including stain composition, both in

the solid and liquid states, which are related to the physico-chemical properties of the stain,

manner of constitution for use, storage conditions and duration of time of applying it to the

biological material (Penney et al., 2002; WHO, 2016c). The Romanowski-Giemsa staining

process follows an ordered chemical reaction in which the dye components interact with each

other and cellular components of biological materials. There is, therefore, a correlation between

the physico-chemical properties of the stains and their staining quality (Friedrich et al., 1990a;

Friedrich K, Seiffert W, 1990; R. Horobin, 2011; Horobin and Walter, 1987b; Marshall et al.,

1975b; P. N. Marshall, 1975a, 1975b; Wittekind and Gehring, 1985a). A major problem

encountered in the routine use of commercial Romanowsky-type stains is batch-to-batch

variations in staining properties, which is usually caused by variations in dye composition and

the presence of contaminating metal salts (Marshall et al., 1975b; P. N. Marshall, 1974) .Stains

standardization is therefore necessary to ensure consistently good stain performance and to

facilitate stained slides comparison between laboratories (R. W. Horobin, 2011; Marshall et al.,

1975a; Schulte, 1991; Scott, R.E. & French, 1924). Uniformly stained blood smears reduce diagnostic

errors and allow morphological features to be correctly identified either by reference to atlases or

slide comparison (Khan et al., 2011; Maguire et al., 2006; Marshall et al., 1978, 1975c). Liquid

chromatographic technique such as thin layer chromatography (TLC) and high-performance

liquid chromatography (HPLC), uv-spectrophotometry, sulphated ash percent determination,

and staining test have been demonstrated to be the most effective techniques for quality control

of stains (Horobin, R.W. and Murgatroyd, 1967; Marshall et al., 1975b; P N Marshall, 1976a;

Marshall and Lewis, 1974; P. N. Marshall, 1974; Proctor and Horobin, 1985; W. Galbraith,

1980). Unfortunately, these techniques appear not to be employed by most commercial

manufacturers of biological stains to ascertain the quality of their products (Gordon B. Proctor

and Richard W. Horobin, 1985; Horobin and Walter, 1987b; Krafts et al., 2010; Wittekind,

2002). By using these techniques, the specifications of successful Romanowski-type stains such

as Giemsa and Leishman have been determined. The simplest commercial stains were found to

comprise Methylene Blue, Azure B and Eosin Y (R. Horobin, 2011; P. N. Marshall, 1976;

Penney et al., 2002) . The presence of certain other dyes, such as Azure A, sym-

Dimethylthionine, Azure C, thionine, 2',4',5'-tribromofluorescein, (Fluorescein), although not

necessary for the Romanowsky-Giemsa effect, have been shown not to have a detrimental effect

on stain performance and characteristics; whereas the presence of methylene violet Bernthsen,

methyl thionoline and high concentrations of metallic salts have been proved to be harmful to

stain performance as regards the Romanowsky-Giemsa effect (Marshall, 1978b; Marshall et al.,

1978; Roe, M.A., Wilcox, A. and Lillie, 1941). The objective of the study, therefore, was to

describe the physico-chemical characteristics of Romanowski-type stains used for malaria

microscopy in Plateau State and determine factors that affect the quality of colour imparted on

malaria parasites in stained blood smears. To our knowledge, this is the first attempt to conduct

this kind of study in Nigeria.

Results

General Stain Characteristics

Of the 94 stains samples, 40.4% (38/94) were Field stains, 38.3% (36/94) were Giemsa stains,

21.3% (20/94) were Leishman stains, and 1.1% (1/94) was a reference stain (Marshall and Lewis

stain). Overall, 21.3% (20/94) of stains were manufactured by BEMA Scientific (Field stain:

28.9%; Giemsa stain: 16.7%; and Leishman stain: 15.0%). Batch number was indicated on the

bottles of 43.6% (41/94) of all stains (Field stain: 21.1%; Giemsa stain: 63.9%; and Leishman

stain: 50.0%). Expiry date was indicated on 42.6% (40/94) of all stains (Field stain: 21.1%;

Giemsa stain: 63.9%; and Leishman stain: 45.0%. Commercially prepared stains accounted for

68.1% (64/94) of all stain solution (Field stain: 86.8%; Giemsa stain: 47.2%; and Leishman stain:

70%). The location (company address) of stain manufacturer was known for 68.1% (64/94) of all

stains (Field stain: 63.2%; Giemsa stain: 72.2% and Leishman stains: 70.0%) (**Table 1**).

Ultraviolet Spectroscopic Characteristics of Stain Samples

The four most prominent groups of dyes present in the stain samples were detected in their order

of relative abundance across stain samples and identified based on the maximum absorbance

wavelengths (λmax) of the dye species present within each group. The first most conspicuous

dyes had a median λmax of 624.0nm (IQR: 607.2, 658.0nm) for all stains, which corresponds to

the λmax of Brilliant Cresyl Blue. When disaggregated by stain type, the median λmax was

658.5nm (IQR: 653.2, 660.0nm), corresponding to that of Methylene Blue for Field stain;

612.0nm (IQR: 517.0, 624.0nm), corresponding to Indigo carmine for the Giemsa stains; and

608.5nm (IQR: 516.8, 619.8nm), corresponding to Aniline Blue for the Leishman stains

(p<0.001). Thus, Methylene Blue emerged as the most abundant dye accounting for 38.3% of

dyes found in the first most prominent dyes in all stains but differ by stain type (Field stain:

81.6%; Giemsa stain: 13.9%; and Leishman stain: 0.0%, p<0.001). The highest prevalence of

dye species irrelevant for the Romanowski-Giemsa effect (RGE) was found in the Leishman

stains (Alcian Blue: 5.0%; Alizarin Red: 5.0%; Aniline Blue: 15.0%; Brilliant cresyl Blue: 5.0%;

Sym-Dimethyl Thionine: 10.0%) and the Giemsa stains (Aniline Blue: 11.1%; Brilliant Cresyl

Blue: 19.4%; Indigo Carmine: 11.1%; and Sym-Dimethyl Thionine: 2.8%). The second most

prominent group of dyes also had a median λ_{max} of 624.0nm (IQR: 609.2, 624.0nm) for all

stains, which corresponds to the λ_{max} of Brilliant Cresyl Blue. When disaggregated by stain type,

the median λ_{max} was 624.0nm (IQR: 624.0, 624.0nm), corresponding to that of Brilliant Cresyl

Blue for Field stain; 624.0nm (IQR: 590.8, 636.0nm), corresponding to Brilliant Cresyl Blue for

the Giemsa stains; and 560.5nm (IQR: 516.8, 624.0nm), corresponding to Pyronine derivatives

for the Leishman stains (p=0.004). Among the third most prominent group of dyes, the median

λ_{max} was 539.0nm (IQR: 516.2, 600.5nm) for all stains, which corresponds to the λ_{max} of Neutral

Red. The median λmax also differed by stain type. For the Field stains, median λ_{max} was

516.0nm (IQR: 515.0, 517.0nm), corresponding to the λ_{max} of Eosin B or Y; the median λ_{max} was

626.0nm (IQR: 544.2, 629.0nm) for the Giemsa stains, corresponding to the λ_{max} of Azure A;

and among the Leishman stains, the median λmax was 549.0nm (IQR: 541.5, 599.5nm,

p<0.001), which corresponds to the λmax of Phloxine B. It was also found that the presence of

RGE-irrelevant dyes was more among the Leishman stains (Neutral red: 30.0%; Basic Fuchsin:

20.0%; Pyronine derivatives: 15.0%; and methylene Violet Bernthsen: 5.0%). Among the fourth

most abundant dyes, the median λ_{max} was 542.0nm (IQR: 419.0, 624.0nm), corresponding to the

λ_{max} of Neutral Red, which tend to differ across stain types (Field stain: 624.0nm [624.0,

624.0]nm; Giemsa stain: 517.5nm [410.0, 626.0nm]; and Leishman stain: 414.0nm [385.0,

543.5nm]; p<0.001). These groups of dyes were found to have a preponderance of unknown dyes

(all stains: 36.2%; Field stain: 21.1%; Giemsa stain: 38.9%; Leishman stain: 60.0%). The most

prevalent single dye detected in the stains were Methylene Blue in the Field stains (81.6%) and

Eosin in the Giemsa stains (30.6%) and Leishman stains (35.0%). Azure b was detected as the

most prevalent single dye in 7.4% of all stains (Field stain: 13.3%; Giemsa stain: 5.6%; and

Leishman stain: 0.0%). The presence of RGE-antagonizing dyes was found in 8.5% of all stains

(Field stain: 0.0%; Giemsa stain: 11.1%; and Leishman stain: 20.0%). (**Table 1, Fig 1-4**). The

identities of the four most prominent dyes detected in each of the stain samples based on their

λ_{max} is shown Supplementary Tables 1-3 for Field, Giemsa, and Leishman stains respectively.

Table 1. Physico-chemical profile of Rowmanowsky-type stains used for malaria microscopy in

Plateau State, Nigeria stratified by stain type

Variable	Level	Overall	Field	Giemsa	Leishman	p-value
No.		94	38	36	20	
Physical characteristics						
Manufacturer	Afis Biochemicals	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	**0.040**
	AFIS Biochemicals	4 (4.3)	4 (10.5)	0 (0.0)	0 (0.0)	
	ANDi Lab	2 (2.1)	0 (0.0)	2 (5.6)	0 (0.0)	
	Antek Diagnostics Ltd, UK	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Apex Lab Reagents	2 (2.1)	0 (0.0)	2 (5.6)	0 (0.0)	
	BEMA Scientific	20 (21.3)	11 (28.9)	6 (16.7)	3 (15.0)	
	Chem Light	1 (1.1)	1 (2.6)	0 (0.0)	0 (0.0)	
	Ferntec ASN	3 (3.2)	1 (2.6)	1 (2.8)	1 (5.0)	
	God's Defence Med Lab	1 (1.1)	1 (2.6)	0 (0.0)	0 (0.0)	
	Hembarge Chemical	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Hembarge Chemicals	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Javic Chemicals	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	JUTH Lab	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Kaycee Chemicals	5 (5.3)	1 (2.6)	1 (2.8)	3 (15.0)	
	Kem Light Laboratories	4 (4.3)	0 (0.0)	3 (8.3)	1 (5.0)	
	Kenac Reagents	1 (1.1)	1 (2.6)	0 (0.0)	0 (0.0)	
	Kenol Pharmacy & Laboratories	1 (1.1)	1 (2.6)	0 (0.0)	0 (0.0)	
	Kermel	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Lab Tech Chemicals	3 (3.2)	1 (2.6)	1 (2.8)	1 (5.0)	
	Medicom Laboratories	3 (3.2)	0 (0.0)	1 (2.8)	2 (10.0)	
	Merck	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	NINO Scientific Company	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	

Variable	Level	Overall	Field	Giemsa	Leishman	p-value
	P. Park Scientific	1 (1.1)	1 (2.6)	0 (0.0)	0 (0.0)	
	PSSH Lab	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Ranjo Medix	12 (12.8)	9 (23.7)	2 (5.6)	1 (5.0)	
	Raymond A Lamb, London	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Marshsl and Lewis	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	S. Merck Chemicals	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Searle, UK	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Sigma Aldriech, Germany	2 (2.1)	0 (0.0)	0 (0.0)	2 (10.0)	
	Suzinal Chemicals	2 (2.1)	2 (5.3)	0 (0.0)	0 (0.0)	
	Trust Care Laboratories	2 (2.1)	1 (2.6)	0 (0.0)	1 (5.0)	
	Trust Chemicals Laboratories Ltd	8 (8.5)	3 (7.9)	5 (13.9)	0 (0.0)	
	Unknown	3 (3.2)	0 (0.0)	2 (5.6)	1 (5.0)	
Manufacturer location known	No	30 (31.9)	14 (36.8)	10 (27.8)	6 (30.0)	0.690
	Yes	64 (68.1)	24 (63.2)	26 (72.2)	14 (70.0)	
Mode of preparation	Commercial	64 (68.1)	33 (86.8)	17 (47.2)	14 (70.0)	**0.001**
	In-house	30 (31.9)	5 (13.2)	19 (52.8)	6 (30.0)	
Batch number known	No	53 (56.4)	30 (78.9)	13 (36.1)	10 (50.0)	**0.001**
	Yes	41 (43.6)	8 (21.1)	23 (63.9)	10 (50.0)	
Expiry data indicated	No	54 (57.4)	30 (78.9)	13 (36.1)	11 (55.0)	**0.001**
	Yes	40 (42.6)	8 (21.1)	23 (63.9)	9 (45.0)	
UV spectroscopic characteristics						
λmax of 1st most prominent dyes (median [IQR])		624.0 [607.2, 658.0]	658.5 [653.2, 660.0]	612.0 [517.0, 624.0]	608.5 [516.8, 619.8]	**<0.001**
Identity of 1st most prominent dye	Alcian blue	2 (2.1)	1 (2.6)	0 (0.0)	1 (5.0)	**<0.001**
	Alizarin red	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Aniline blue	7 (7.4)	0 (0.0)	4 (11.1)	3 (15.0)	
	Azure B	7 (7.4)	5 (13.2)	2 (5.6)	0 (0.0)	
	Azure C	3 (3.2)	0 (0.0)	1 (2.8)	2 (10.0)	
	Brilliant cresyl blue	10 (10.6)	0 (0.0)	7 (19.4)	3 (15.0)	
	Eosin B or Y	18 (19.1)	0 (0.0)	11 (30.6)	7 (35.0)	
	Indigo carmine	5 (5.3)	1 (2.6)	4 (11.1)	0 (0.0)	
	Methylene blue	36 (38.3)	31 (81.6)	5 (13.9)	0 (0.0)	
	Sym-dimethyl thionine	3 (3.2)	0 (0.0)	1 (2.8)	2 (10.0)	
	Unknown	2 (2.1)	0 (0.0)	1 (2.8)	1 (5.0)	
λmax of 2nd most prominent dyes (median [IQR])		624.0 [609.2, 624.0]	624.0 [624.0, 624.0]	624.0 [590.8, 636.0]	560.5 [516.8, 624.0]	0.004
Identity of 2nd most prominent dye	Alcian blue	5 (5.3)	1 (2.6)	3 (8.3)	1 (5.0)	**<0.001**
	Aniline blue	3 (3.2)	0 (0.0)	1 (2.8)	2 (10.0)	
	Azocarmine	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Azure A	1 (1.1)	1 (2.6)	0 (0.0)	0 (0.0)	
	Azure B	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Azure C	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Brilliant cresyl blue	43 (45.7)	29 (76.3)	10 (27.8)	4 (20.0)	
	Eosin B or Y	15 (16.0)	0 (0.0)	8 (22.2)	7 (35.0)	
	Erythrocin B derivative	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Indigo carmine	2 (2.1)	2 (5.3)	0 (0.0)	0 (0.0)	
	Methyl thionine	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Methylene blue	8 (8.5)	0 (0.0)	8 (22.2)	0 (0.0)	
	Safranine	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Sym-dimethyl thionine	3 (3.2)	1 (2.6)	2 (5.6)	0 (0.0)	
	Unknown	8 (8.5)	4 (10.5)	2 (5.6)	2 (10.0)	
λmax of 3rd most prominent dyes (median [IQR])		539.0 [516.2, 600.5]	516.0 [515.0, 517.0]	626.0 [544.2, 629.0]	549.0 [541.5, 599.5]	**<0.001**
Identity of 3rd most prominent dye	Azocarmine	1 (1.1)	1 (2.6)	0 (0.0)	0 (0.0)	<0.001
	Azure A	11 (11.7)	0 (0.0)	9 (25.0)	2 (10.0)	
	Azure B	3 (3.2)	0 (0.0)	3 (8.3)	0 (0.0)	
	Basic fuchsin	6 (6.4)	0 (0.0)	2 (5.6)	4 (20.0)	
	Congo Red	2 (2.1)	2 (5.3)	0 (0.0)	0 (0.0)	
	Darrow red	3 (3.2)	3 (7.9)	0 (0.0)	0 (0.0)	
	Eosin B or Y	34 (36.2)	31 (81.6)	3 (8.3)	0 (0.0)	

Variable	Level	Overall	Field	Giemsa	Leishman	p-value
	Erythrosin B	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Methyl thionine	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Methylene blue	3 (3.2)	0 (0.0)	2 (5.6)	1 (5.0)	
	Methylene Violet Bernthsen	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Neutral red	11 (11.7)	0 (0.0)	5 (13.9)	6 (30.0)	
	Pyronine derivative	7 (7.4)	0 (0.0)	4 (11.1)	3 (15.0)	
	Thionine derivative	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Toluidine blue	2 (2.1)	0 (0.0)	2 (5.6)	0 (0.0)	
	Unknown	7 (7.4)	1 (2.6)	4 (11.1)	2 (10.0)	
λmax of 4th most prominent dyes (median [IQR])		542.0 [419.0, 624.0]	624.0 [624.0, 624.0]	517.5 [410.0, 626.0]	414.0 [385.0, 543.5]	<0.001
Identity of 4th most prominent dyes	Aniline blue	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	<0.001
	Azure A	10 (10.6)	0 (0.0)	8 (22.2)	2 (10.0)	
	Azure B	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Brilliant cresyl blue	30 (31.9)	30 (78.9)	0 (0.0)	0 (0.0)	
	Eosin B or Y	9 (9.6)	0 (0.0)	8 (22.2)	1 (5.0)	
	Ethyleosin	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Methylene blue	1 (1.1)	0 (0.0)	1 (2.8)	0 (0.0)	
	Methylene Violet Bernthsen	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Neutral red	4 (4.3)	0 (0.0)	3 (8.3)	1 (5.0)	
	Pyronine derivative	2 (2.1)	0 (0.0)	1 (2.8)	1 (5.0)	
	Safranine	1 (1.1)	0 (0.0)	0 (0.0)	1 (5.0)	
	Unknown	33 (35.1)	8 (21.1)	13 (36.1)	12 (60.0)	
Absorbance of 1st major dyes (median [IQR])		1.1 [0.7, 1.5]	1.0 [0.8, 1.5]	1.3 [0.9, 1.9]	0.8 [0.5, 1.0]	0.006
Absorbance of 2nd major dyes (median [IQR])		0.9 [0.6, 1.3]	0.9 [0.7, 1.4]	1.2 [0.7, 1.3]	0.6 [0.4, 0.8]	0.002
Absorbance of.3rd major dye (median [IQR])		1.1 [0.5, 2.0]	2.8 [1.4, 4.0]	0.8 [0.4, 1.1]	0.4 [0.2, 0.7]	<0.001
Absorbance of 4th major dyes (median [IQR])		0.8 [0.1, 1.8]	2.8 [1.4, 4.0]	0.4 [0.1, 0.8]	0.1 [0.1, 0.2]	<0.001
Methylene Blue as most prominent dye	No	58 (61.7)	7 (18.4)	31 (86.1)	20 (100.0)	<0.001
	Yes	36 (38.3)	31 (81.6)	5 (13.9)	0 (0.0)	
Antagonistic dyes present	No	86 (91.5)	38 (100.0)	32 (88.9)	16 (80.0)	0.027
	Yes	8 (8.5)	0 (0.0)	4 (11.1)	4 (20.0)	
Aniline Blue most prominent dye	No	87 (92.6)	38 (100.0)	32 (88.9)	17 (85.0)	0.067
	Yes	7 (7.4)	0 (0.0)	4 (11.1)	3 (15.0)	
Eosin B or Y as most prominent dye	No	76 (80.9)	38 (100.0)	25 (69.4)	13 (65.0)	<0.001
	Yes	18 (19.1)	0 (0.0)	11 (30.6)	7 (35.0)	
Brilliant Cresyl Blue as most prominent dye	No	84 (89.4)	38 (100.0)	29 (80.6)	17 (85.0)	0.020
	Yes	10 (10.6)	0 (0.0)	7 (19.4)	3 (15.0)	
Azure B most prominent dye	No	87 (92.6)	33 (86.8)	34 (94.4)	20 (100.0)	0.166
	Yes	7 (7.4)	5 (13.2)	2 (5.6)	0 (0.0)	
Number of dye components (median [IQR]		9.0 [8.0, 11.0]	10.0 [8.2, 12.0]	9.0 [8.0, 10.2]	7.0 [6.0, 8.0]	<0.001
Number of irrelevant dyes (median [IQR]		6.0 [5.0, 6.0]	6.0 [5.0, 8.0]	6.0 [5.0, 6.0]	5.0 [4.0, 6.0]	0.015
Number of unknown dyes		4.8 (2.0)	6.1 (2.4)	4.1 (1.1)	3.8 (1.1)	<0.001
High Performance Liquid Chromatographic characteristics						
Peak area of Methylene blue (median [IQR])		22.5 [0.0, 95.4]	92.7 [51.1, 148.9]	17.7 [0.0, 34.6]	0.0 [0.0, 0.0]	<0.001
Peak area of Azure A (median [IQR])		66.7 [31.1, 146.6]	42.1 [20.5, 70.3]	145.9 [81.9, 191.6]	58.0 [27.9, 116.9]	<0.001
Peak Area of Azure B (median [IQR.])		36.9 [0.0, 101.7]	34.1 [13.0, 51.1]	115.6 [49.3, 242.8]	0.0 [0.0, 0.0]	<0.001
Peak area of Eosin B (median [IQR])		0.0 [0.0, 31.4]	0.0 [0.0, 0.0]	27.4 [0.0, 72.0]	27.7 [5.0, 35.5]	<0.001
Peak area of Eosin Y (median [IQR])		63.1 [9.6, 123.6]	8.8 [7.3, 11.2]	104.7 [75.1, 131.4]	126.8 [93.1, 166.5]	<0.001
Azure A proportion (median [IQR])		14.6 [10.5, 23.9]	12.6 [10.6, 14.6]	23.2 [14.0, 29.9]	16.3 [5.4, 26.0]	0.001
Azure B proportion (median [IQR])		9.2 [0.0, 15.2]	12.9 [6.6, 14.8]	12.5 [4.5, 22.2]	0.0 [0.0, 0.0]	<0.001
Eosin B proportion (median [IQR])		0.8 [0.0, 5.0]	0.0 [0.0, 0.6]	3.2 [0.0, 6.6]	6.1 [0.7, 7.9]	<0.001
Eosin Y proportion (median [IQR])		7.6 [2.6, 21.4]	2.2 [1.2, 4.0]	10.7 [6.9, 15.7]	27.5 [23.9, 28.8]	<0.001
Methylene blue proportion (median [IQR])		2.9 [0.0, 30.5]	32.5 [23.2, 38.1]	1.6 [0.0, 2.6]	0.0 [0.0, 0.0]	<0.001
Unknown dyes proportion (median [IQR])		39.5 [33.7, 50.1]	37.8 [33.5, 46.3]	41.0 [33.8, 46.3]	45.5 [34.3, 66.5]	0.343

Variable	Level	Overall	Field	Giemsa	Leishman	p-value
Proportion of RGE relevant dyes (median [IQR])		35.4 [25.4, 48.6]	50.5 [37.3, 55.9]	29.0 [20.7, 39.3]	27.9 [24.7, 30.1]	**<0.001**
Proportion of RGE irrelevant dyes (median [IQR])		64.6 [51.3, 74.6]	49.5 [44.1, 62.7]	71.1 [60.7, 79.3]	72.1 [69.9, 75.2]	**<0.001**
Ratio of Methylene blue to Azure B (median [IQR]		0.1 [0.0, 2.6]	2.7 [0.4, 3.2]	0.1 [0.0, 0.2]	0.0 [0.0, 0.0]	**<0.001**
Ratio of Azure B to Eosin Y (median [IQR]		1.0 [0.0, 4.2]	4.3 [0.0, 6.0]	1.2 [0.5, 2.2]	0.0 [0.0, 0.0]	**<0.001**
Other chemical characteristics						
pH of stock stain (median [IQR])		6.9 [6.6, 7.3]	6.6 [6.5, 6.7]	7.4 [7.1, 7.5]	7.0 [6.7, 7.1]	**<0.001**
Sulphated ash percent (median [IQR])		34.3 [15.9, 47.8]	66.2 [45.4, 73.7]	35.3 [29.8, 46.5]	11.9 [6.6, 29.1]	0.064
Right pH	No	44 (46.8)	19 (50.0)	10 (27.8)	15 (75.0)	**0.003**
	Yes	50 (53.2)	19 (50.0)	26 (72.2)	5 (25.0)	
Right solvent	No	29 (30.9)	0 (0.0)	13 (36.1)	16 (80.0)	**<0.001**
	Yes	65 (69.1)	38 (100.0)	23 (63.9)	4 (20.0)	

Fig 1. Spectroscopic characteristics of the first prominent group of dyes present in Romanowski-type stains used for malaria microscopy in Plateau state, Nigeria disaggregated by mode of preparation",

Fig 2. Spectroscopic characteristics of the second most prominent dyes present in Romanowski-type stains used for malaria microscopy in Plateau state, Nigeria disaggregated by stain type",

Fig 3. Spectroscopic characteristics of the third most prominent group of dyes present in Romanowsky-type stains used for malaria microscopy in Plateau state, Nigeria disaggregated by stain type.

Data Source: Cross-sectional survey conducted by Akar Stephen in Plateau State, Nigeria, 2018

Fig 4. Spectroscopic characteristics of fourth most prominent group of dye present in
Romanowsky-type stains used for malaria microscopy in Plateau state, Nigeria disaggregated by
stain type.

Determination of Dye Proportions by High Performance Liquid Chromatography (HPLC)

HPLC was done to determine the proportion of the dyes essential for the Romanowsky-Giemsa
effect (RGE) in each of the stain samples. The dyes analyzed were pure methylene blue, Azure
A, Azure B, eosin B and eosin Y. The presence and concentrations of these dyes were
determined using the retention times of the standard dyes. Individual solutions of methylene
blue, Azure A, Azure B, eosin B and eosin Y prepared from pure powders were
chromatographed in triplicates by HPLC to determine the individual mean retention time of the
eluents. The retention times were used to identify the presence of the dyes in the samples of
stains collected. The retention time obtained for each dye species were as follows: Eosin Y, 1.4
minutes, Eosin B, 1.404 minutes, methylene blue, 2.920 minutes and Azure B, 2.923 minutes.
The dye peak area under the HPLC graph output, which was calculated as the area of a triangle,
was used to estimate the proportion of each dye species present in the stain samples.

The median proportion of Azure A for all stains was found to be 14.6% (IQR: 10.5-23.9%),

which in turn differ significantly by stain type (Field stain: median = 12.6%, IQR = 10.6-14.6%;

Giemsa stain: median = 23.2%, IQR = 14.0-29.9%; and Leishman stain: median = 16.3%; IQR =

5.4-26.0%, p=0.001). The median proportion of Azure B for all stains was 9.2% (IQR: 0.0-

15.2%) and vary significantly by stain type (Field stain: median = 12.9%, IQR = 6.6-14.8%);

Giemsa stain: median = 12.5%, IQR= 4.5-22.2%; and Leishman stain: median = 0.0, IQR= 0.0-

0.0%, p<0.001). For Eosin B, median proportion for all stains was 0.8% (IQR: 0.0-6.6%), which

was found to differ across stain types (Field stain: median = 0.0%, IQR = 0.0-0.6%); Giemsa

stain: median = 3.2%, IQR= 0.0-6.6%; and Leishman stain: median = 6.1, IQR= 0.7-7.9%,

p<0.001). The median proportion of Eosin Y was found to be 7.6% (IQR: 2.6-21.4%) which also

differ significantly across stain types (Field stain: median = 2.2%, IQR = 1.2-4.0%); Giemsa

stain: median = 10.7%, IQR= 6.9-15.7%; and Leishman stain: median = 27.5, IQR= 23.9-28.8%,

p<0.001). For Methylene Blue, the median proportion was found to be 2.9% (IQR: 0.0-30.5%)

for all stains, which was found to differ by stain type (Field stain: median = 32.5%, IQR = 23.2-

38.1%); Giemsa stain: median = 1.6%, IQR= 0.0-2.6%; and Leishman stain: median = 0.0, IQR=

0.0-0.0%, p<0.001). The proportion of unidentified dyes (unknown dyes) was found not to vary

significantly across tains types (All stains: medina =39.5% [IQR: 33.7, 50.1%]; Field stain:

median = 37.8% [IQR: 33.5, 46.3%]; Giemsa stain: median = 41.0% [IQR: 33.8, 46.3%]; and

Leishman stain: median = 45.5% [IQR: 34.3, 66.5%], p=0.343. However, the proportion of

RGE-relevant dyes differ significantly across stain types (All stains: median = 35.4% [IQR: 25.4,

48.6%]; Field stain: median = 50.5% [IQR: 37.3, 55.9%]; Giemsa stain: median = 29.0% [20.7,

39.3%]; and Leishman stain: median = 27.9% [24.7, 30.1%], p<0.001). The ratio of Azure b to

Eosin Y, which are the crucial ingredients of the RGE reaction were also found to differ

significantly across the stain types (All stain: median = 0.1 [IQR: 0.0, 2.6]; Field stain: 2.7 [0.4,

3.2]; Giemsa stain: 0.1 [0.0, 0.2]; and Leishman stain: 0.0 [0.0, 0.0], p<0.001) (**Table 1**). On

comparing the dye proportions by mode of stain preparation, it was found that commercially

prepared Giemsa stain solution had higher proportion of Azure A compared to in-house solutions

(**Fig 5a**). The proportion of Azure B was found to be lowest among commercially prepared

Leishman stain solutions and highest in in-house Giemsa stain solutions (**Fig 5b**). Both

commercial and in-house Field stain solutions contained higher proportion of Methylene Blue

compared to Giemsa and Leishman stains (**Fig 5c**). The Field stains contained lesser proportion

of Eosin B compared to Giemsa and Leishman stains irrespective of mode of stain preparation

(**Fig 5d**). Both commercial and in-house solutions of Leishman stains were found to contain

more Eosin Y compared to Field and Giemsa stains (**Fig 5e**), but the Leishman stains exhibited

the least Azure B to Eosin Y ratio while the Field stains were highest (**Fig 5f**). The proportion of

Azure A, Azure B, Eosin B, Eosin Y, and Methylene Blue present in the individual stain samples

are shown in **Supplementary Tables 3-4** for Field, Giemsa, and Leishman stains respectively.

Fig 5. Comparison of the proportion Azure A, Azure B, Methylene Blue, Eosin B, Eosin Y and

the ratio of Azure B to Eosin Y present in Field, Giemsa, and Leishman stains used for malaria

microscopy in Plateau State, Nigeria.

Stain pH, Sulphated Ash Percent and Solvent Used for Preparation Stain solution

Median pH of stock stain solution was 6.9 (IQR = 6.6 – 7.3) for all stains (Field stains: median =

6.6 [IQR: 6.5 – 6.7]; Giemsa stains: median = 7.4 [IQR = 7.1–7.5]; and Leishman stains: median

= 7.0[6.7-7.1], $p<0.001$). The median sulphated ash percent of dried powders from which in-

house stain solutions were made was 34.3% (15.9– 47.8%) for all stains (Field stain: median =

66.2% [IQR = 45.4– 73.7%]; Giemsa stain: median = 35.3% [29.8-46.5%]; and Leishman stain:

median = 11.9% [IQR = 6.6 – 29.1%], p=0.064) (**Table 1**). All Field stain solutions were found
to be made of water as solvent. Only 20% (4/20) of Leishman stains were found to be made of
absolute methanol. Up to 63.9% (23/36) of Giemsa stains were found to be composed of a
glycerol to absolute methanol ratio of 1:1 with a median of 1.3 (0.9 – 1.9). Glycerol to methanol
ratio was found to be higher in in commercially prepared stain solution (median = 1.8; IQR: 0.9-
2.7) compared to in-house stain solution (median = 1.2; IQR: 1.1-1.5) of Giemsa stains (**Fig 6a**).
320

322 **Fig 6.** Comparison of Glycerol to methanol ratio of Giemsa stains used for malaria microscopy
323 in Plateau State, Nigeria.
324

Assessment of Staining Quality of Stain Samples
The staining quality of the stain samples was assessed using thin blood film microscopy. Known
malaria positive blood smears were stained and examined by WHO certified malaria
microscopists. Staining quality of each stain sample was measured by the quality of colour
imparted by the stain samples on the cytoplasm and nucleus of malaria parasites. Staining quality
was graded excellent if the cytoplasm was stained blue and nucleus pink; good if cytoplasm was
stained blue and nucleus red or pale pink; fair if cytoplasm was stained purple-blue and nucleus
deep red or pale pink; and poor if cytoplasm was pale blue, purple or unstained and nucleus.
Of the two stain samples that yielded excellent staining, one was a the reference Giemsa stain
sample obtained from ANDi laboratory, Lagos and the other a Giemsa-type stain described by
Marshall et al. (1975). Among stains that were graded as "good", 75.0% were Giemsa and 25.0%
were Field stains. None of the Leishman stains yield "good" staining. Stains that yielded "fair"
(81.2%) or "poor" (70.6%, p=0.027) staining outcome were largely commercially prepared

solutions, which were also had expiry dates missing from the labels on the containers (75.0% and 57.4% respectively). "Poor" performing stains were found to contain RGE-irrelevant dyes such as Alcian Blue, Alizarin Red, Aniline Blue, Azure C, Brilliant Cresyl Blue, Indigo Carmine, Methylene Blue, Sym-Dimethyl Thionine, Methyl Thionine, Safranine, Basic Fuchsin, Congo Red, Darrow Red, Erythrosin B, Methylene Violet Bernthsen, Neutral Red, Pyronine derivatives, Toluidine Blue, and unidentified (unknown) dyes (**Table 2**).

Table 2. Evaluation of the performance of Romanowsky-type stains used for malaria microscopy in Plateau State, Nigeria.

Variable	level	Overall	Excellent	Fair	Good	Poor	p-value
No.		94	2	16	8	68	
Manufacturer location known	No	30 (31.9)	0 (0.0)	5 (31.2)	3 (37.5)	22 (32.4)	0.786
	Yes	64 (68.1)	2 (100.0)	11 (68.8)	5 (62.5)	46 (67.6)	
Stain type	Field	38 (40.4)	0 (0.0)	7 (43.8)	2 (25.0)	29 (42.6)	0.125
	Giemsa	36 (38.3)	2 (100.0)	4 (25.0)	6 (75.0)	24 (35.3)	
	Leishman	20 (21.3)	0 (0.0)	5 (31.2)	0 (0.0)	15 (22.1)	
Mode of preparation	Commercial	64 (68.1)	0 (0.0)	13 (81.2)	3 (37.5)	48 (70.6)	**0.027**
	In-house	30 (31.9)	2 (100.0)	3 (18.8)	5 (62.5)	20 (29.4)	
Batch number known	No	53 (56.4)	0 (0.0)	12 (75.0)	3 (37.5)	38 (55.9)	0.111
	Yes	41 (43.6)	2 (100.0)	4 (25.0)	5 (62.5)	30 (44.1)	
Expiry data indicated	No	54 (57.4)	0 (0.0)	12 (75.0)	3 (37.5)	39 (57.4)	0.111
	Yes	40 (42.6)	2 (100.0)	4 (25.0)	5 (62.5)	29 (42.6)	
λmax of 1^{st} most prominent dyes (median [IQR])		624.0 [607.2, 658.0]	649.0 [648.5, 649.5]	624.0 [614.2, 654.2]	618.0 [588.0, 654.0]	624.0 [573.8, 659.2]	0.828
Identity of 1st most prominent dye	Alcian blue	2 (2.1)	0 (0.0)	1 (6.2)	0 (0.0)	1 (1.5)	0.073
	Alizarin red	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Aniline blue	7 (7.4)	0 (0.0)	1 (6.2)	0 (0.0)	6 (8.8)	
	Azure B	7 (7.4)	2 (100.0)	1 (6.2)	0 (0.0)	4 (5.9)	
	Azure C	3 (3.2)	0 (0.0)	1 (6.2)	0 (0.0)	2 (2.9)	
	Brilliant cresyl blue	10 (10.6)	0 (0.0)	2 (12.5)	1 (12.5)	7 (10.3)	
	Eosin B or Y	18 (19.1)	0 (0.0)	1 (6.2)	2 (25.0)	15 (22.1)	
	Indigo carmine	5 (5.3)	0 (0.0)	2 (12.5)	2 (25.0)	1 (1.5)	
	Methylene blue	36 (38.3)	0 (0.0)	6 (37.5)	3 (37.5)	27 (39.7)	
	Sym-dimethyl thionine	3 (3.2)	0 (0.0)	1 (6.2)	0 (0.0)	2 (2.9)	
	Unknown	2 (2.1)	0 (0.0)	0 (0.0)	0 (0.0)	2 (2.9)	
λmax 2^{nd} most prominent dyes (median [IQR])		624.0 [609.2, 624.0]	576.5 [547.2, 605.8]	624.0 [518.0, 624.0]	624.0 [624.0, 655.2]	624.0 [608.0, 624.0]	0.141
Identity of 2nd most prominent dye	Alcian blue	5 (5.3)	0 (0.0)	0 (0.0)	1 (12.5)	4 (5.9)	0.960
	Aniline blue	3 (3.2)	0 (0.0)	0 (0.0)	0 (0.0)	3 (4.4)	
	Azocarmine	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Azure A	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Azure B	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Azure C	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Brilliant cresyl blue	43 (45.7)	0 (0.0)	7 (43.8)	4 (50.0)	32 (47.1)	
	Eosin B or Y	15 (16.0)	1 (50.0)	5 (31.2)	0 (0.0)	9 (13.2)	
	Erythrocin B derivative	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Indigo carmine	2 (2.1)	0 (0.0)	0 (0.0)	0 (0.0)	2 (2.9)	
	Methyl thionine	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Methylene blue	8 (8.5)	0 (0.0)	1 (6.2)	3 (37.5)	4 (5.9)	
	Safranine	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Sym-dimethyl thionine	3 (3.2)	0 (0.0)	1 (6.2)	0 (0.0)	2 (2.9)	

Variable	level	Overall	Excellent	Fair	Good	Poor	p-value
	Unknown	8 (8.5)	1 (50.0)	2 (12.5)	0 (0.0)	5 (7.4)	
λmax of 3rd most prominent dyes (median [IQR])		539.0 [516.2, 600.5]	537.5 [528.2, 546.8]	539.0 [517.0, 547.5]	627.0 [517.0, 632.5]	535.5 [516.0, 573.5]	0.371
Identity of 3rd most prominent dye	Azocarmine	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	0.769
	Azure A	11 (11.7)	0 (0.0)	2 (12.5)	2 (25.0)	7 (10.3)	
	Azure B	3 (3.2)	0 (0.0)	0 (0.0)	2 (25.0)	1 (1.5)	
	Basic fuchsin	6 (6.4)	0 (0.0)	2 (12.5)	0 (0.0)	4 (5.9)	
	Congo Red	2 (2.1)	0 (0.0)	0 (0.0)	0 (0.0)	2 (2.9)	
	Darrow red	3 (3.2)	0 (0.0)	0 (0.0)	0 (0.0)	3 (4.4)	
	Eosin B or Y	34 (36.2)	1 (50.0)	7 (43.8)	3 (37.5)	23 (33.8)	
	Erythrosin B	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Methyl thionine	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Methylene blue	3 (3.2)	0 (0.0)	0 (0.0)	0 (0.0)	3 (4.4)	
	Methylene Violet Bernthsen	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Neutral red	11 (11.7)	0 (0.0)	4 (25.0)	0 (0.0)	7 (10.3)	
	Pyronine derivative	7 (7.4)	1 (50.0)	0 (0.0)	0 (0.0)	6 (8.8)	
	Thionine derivative	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Toluidine blue	2 (2.1)	0 (0.0)	0 (0.0)	1 (12.5)	1 (1.5)	
	Unknown	7 (7.4)	0 (0.0)	1 (6.2)	0 (0.0)	6 (8.8)	
λmax 4th most prominent dyes (median [IQR])		542.0 [419.0, 624.0]	473.0 [444.0, 502.0]	476.0 [405.8, 624.0]	624.0 [533.8, 626.0]	577.5 [421.0, 624.0]	0.070
Identity of 4th most prominent dye	Aniline blue	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	**<0.001**
	Azure A	10 (10.6)	0 (0.0)	0 (0.0)	3 (37.5)	7 (10.3)	
	Azure B	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Brilliant cresyl blue	30 (31.9)	0 (0.0)	5 (31.2)	2 (25.0)	23 (33.8)	
	Eosin B or Y	9 (9.6)	0 (0.0)	3 (18.8)	2 (25.0)	4 (5.9)	
	Ethyleosin	1 (1.1)	1 (50.0)	0 (0.0)	0 (0.0)	0 (0.0)	
	Methylene blue	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Methylene Violet Bernthsen	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Neutral red	4 (4.3)	0 (0.0)	0 (0.0)	1 (12.5)	3 (4.4)	
	Pyronine derivative	2 (2.1)	0 (0.0)	0 (0.0)	0 (0.0)	2 (2.9)	
	Safranine	1 (1.1)	0 (0.0)	0 (0.0)	0 (0.0)	1 (1.5)	
	Unknown	33 (35.1)	1 (50.0)	8 (50.0)	0 (0.0)	24 (35.3)	
Absorbance of 1st major dye (median [IQR])		1.1 [0.7, 1.5]	2.5 [2.2, 2.9]	1.1 [0.9, 1.5]	1.4 [1.3, 1.5]	1.0 [0.7, 1.5]	0.245
Absorbance of 2nd major dye (median [IQR])		0.9 [0.6, 1.3]	1.6 [1.4, 1.7]	0.9 [0.7, 1.1]	1.3 [1.2, 1.3]	0.8 [0.6, 1.2]	0.116
Absorbance of 3rd major dye (median [IQR])		1.1 [0.5, 2.0]	0.6 [0.5, 0.6]	1.0 [0.4, 3.0]	1.2 [1.1, 1.3]	1.1 [0.5, 2.0]	0.566
Absorbance of 4th major dyes (median [IQR])		0.8 [0.1, 1.8]	0.3 [0.2, 0.5]	0.9 [0.1, 2.4]	1.1 [0.9, 1.2]	0.7 [0.1, 1.9]	0.471
Methylene Blue as most prominent dye	No	58 (61.7)	2 (100.0)	10 (62.5)	5 (62.5)	41 (60.3)	0.728
	Yes	36 (38.3)	0 (0.0)	6 (37.5)	3 (37.5)	27 (39.7)	
Antagonistic dyes present	No	86 (91.5)	2 (100.0)	15 (93.8)	8 (100.0)	61 (89.7)	0.726
	Yes	8 (8.5)	0 (0.0)	1 (6.2)	0 (0.0)	7 (10.3)	
Aniline Blue as most prominent dye	No	87 (92.6)	2 (100.0)	15 (93.8)	8 (100.0)	62 (91.2)	0.795
	Yes	7 (7.4)	0 (0.0)	1 (6.2)	0 (0.0)	6 (8.8)	
Eosin B or Y as most prominent dye	No	76 (80.9)	2 (100.0)	15 (93.8)	6 (75.0)	53 (77.9)	0.433
	Yes	18 (19.1)	0 (0.0)	1 (6.2)	2 (25.0)	15 (22.1)	
Brilliant Cresyl Blue as most prominent dye	No	84 (89.4)	2 (100.0)	14 (87.5)	7 (87.5)	61 (89.7)	0.953
	Yes	10 (10.6)	0 (0.0)	2 (12.5)	1 (12.5)	7 (10.3)	
Peak area of Methylene blue (media [IQR])		22.5 [0.0, 95.4]	0.0 [0.0, 0.0]	22.8 [0.0, 108.8]	17.5 [12.1, 32.5]	23.3 [0.0, 98.4]	0.386
Peak area of Azure A (media [IQR])		66.7 [31.1, 146.6]	47.0 [46.6, 47.5]	76.3 [40.6, 164.5]	86.9 [29.9, 138.7]	63.9 [28.3, 146.5]	0.913
Peak Area of Azure B (median [IQR])		36.9 [0.0, 101.7]	139.0 [136.7, 141.4]	36.0 [0.0, 108.9]	179.6 [84.2, 297.6]	33.4 [0.0, 61.4]	**0.022**

Variable	level	Overall	Excellent	Fair	Good	Poor	p-value
Peak area of Eosin B (median [IQR])		0.0 [0.0, 31.4]	0.0 [0.0, 0.0]	3.4 [0.0, 30.3]	0.0 [0.0, 53.6]	1.5 [0.0, 30.2]	0.646
Peak area of Eosin Y (median [IQR])		63.1 [9.6, 123.6]	150.1 [137.2, 162.9]	61.4 [10.0, 126.0]	72.1 [17.2, 131.4]	57.6 [8.6, 116.2]	0.259
Azure A proportion (median [IQR])		14.6 [10.5, 23.9]	7.8 [7.4, 8.1]	15.5 [11.9, 19.0]	11.9 [5.4, 19.2]	14.6 [10.6, 24.7]	0.266
Azure B proportion (median [IQR])		9.2 [0.0, 15.2]	23.0 [22.0, 24.0]	9.4 [0.0, 14.2]	22.4 [18.4, 24.2]	6.6 [0.0, 13.9]	**0.001**
Eosin B proportion (median [IQR])		0.8 [0.0, 5.0]	0.0 [0.0, 0.0]	0.0 [0.0, 6.2]	0.0 [0.0, 4.4]	1.6 [0.0, 5.4]	0.386
Eosin Y proportion (median [IQR])		7.6 [2.6, 21.4]	25.7 [22.2, 29.2]	7.2 [1.5, 15.8]	10.2 [4.5, 15.8]	7.7 [2.8, 21.5]	0.358
Methylene blue proportion (median [IQR.]])		2.9 [0.0, 30.5]	0.0 [0.0, 0.0]	18.0 [0.0, 31.7]	2.7 [1.4, 9.4]	3.2 [0.0, 31.7]	0.360
Unknown dyes proportion (median [IQR])		39.5 [33.7, 50.1]	43.2 [38.5, 47.9]	38.0 [34.3, 48.2]	44.1 [40.6, 46.1]	39.2 [32.1, 51.2]	0.959
Proportion of RGE-relevant dyes (median [IQR])		35.4 [25.4, 48.6]	48.7 [44.2, 53.2]	38.9 [31.2, 44.0]	38.4 [36.1, 42.7]	30.5 [23.1, 49.3]	0.256
Proportion of RGE-irrelevant dyes (median [IQR])		64.6 [51.3, 74.6]	51.3 [46.8, 55.8]	61.1 [56.0, 68.8]	61.6 [57.3, 63.9]	69.6 [50.7, 76.8]	0.258
Ratio of Azure B to Eosin Y (median [IQR])		1.1 [0.0, 4.8]	0.9 [0.9, 1.0]	3.2 [0.0, 5.8]	2.5 [1.3, 4.3]	0.6 [0.0, 4.7]	0.371
Ratio of Methylene blue to Azure B (median [IQR]		1.9 [0.1, 3.1]	0.0 [0.0, 0.0]	2.5 [0.3, 3.1]	0.1 [0.1, 0.6]	2.4 [0.1, 3.4]	**0.048**
Number of dye components (median [IQR])		9.0 [8.0, 11.0]	9.0 [8.5, 9.5]	9.0 [7.0, 11.0]	10.0 [9.0, 11.0]	9.0 [8.0, 11.0]	0.844
Number of irrelevant dyes (median [IQR])		6.0 [5.0, 6.0]	6.0 [5.5, 6.5]	5.5 [4.8, 6.0]	5.5 [5.0, 6.0]	6.0 [5.0, 6.2]	0.868
Number unknown dyes		4.8 (2.0)	6.0 (1.4)	4.7 (2.1)	4.4 (1.2)	4.8 (2.1)	0.774
pH of stock stain (median [IQR])		6.9 [6.6, 7.3]	7.4 [7.3, 7.4]	6.8 [6.5, 7.1]	7.2 [7.0, 7.5]	6.8 [6.6, 7.4]	0.092
Sulphated ash percent (median [IQR])		34.1 (20.5)	19.8 (14.1)	25.0 (16.3)	35.0 (10.1)	36.9 (23.0)	0.609
Sulphated ash percent (median [IQR])		35.3 [29.8, 46.5]	19.8 [14.8, 24.7]	34.3 [33.3, 35.2]	34.2 [28.8, 40.5]	40.8 [34.3, 48.5]	0.252
Glycerol.to methanol ratio (median [IQR])		1.3 [0.9, 1.9]	1.2 [1.2, 1.3]	2.2 [1.6, 3.2]	1.4 [0.9, 1.4]	1.2 [0.9, 2.2]	0.538

The distribution of the maximum absorbance wavelengths of the four most prominent group of dyes by the staining performance is shown in supplementary figures 1-4. They also had a higher presence of known RGE antagonistic dyes (10.3%). "Poor" stains were found to contain higher proportion of Azure A (median = 14.6%; IQR: 10.6-24.7%); significantly lower proportion of Azure B (median = 6.6%; IQR: 0.0-13.9%, p=0.001); higher proportion of Eosin B (median = 1.6%; IQR: 0.0-5.4%); lower Eosin Y (median = 7.7%; IQR: 2.8-21.5%); lower proportion of RGE-relevant dyes (median = 30.5%; IQR: 23.1-49.3%); higher Methylene Blue to Azure B ratio (median = 2.4; IQR: 0.1-3.4, p=0.048); and higher sulphated ash percent (median = 40.8%; IQR: 34.3-48.5%) (Table 2). A comparison of the proportion of the Azure, Azure B, Eosin B,

Eosin Y, and Methylene Blue present in the stains by staining outcome and stain type is shown in

Fig 7 while **Fig 8** shows that for Azure B to Eosin Y ratio, Methylene Blue to Azure B ratio,

proportion of RGE-relevant dyes, and proportion of RGE-irrelevant dyes. "Good" performing

Field stains had a median stock pH of 6.8 while "poor" stains exhibited a median stock pH of

6.6. "Excellent" and "good" performing Giemsa stains had a median stock pH of about 7.2 with

"good" performing stains exhibiting a much wider inter-quartile range. For the Leishman stains,

the "poor" performing stain samples exhibited a higher pH compared to those that performed

"fairly" (**Fig 9a and b**). It was found that the higher the sulphated ash percent of the stain

powder the poorer the performance of the stain solution, with the Field stains exhibiting the

highest sulphated ash percent (**Fig 9c and d**). For the Giemsa stains, "excellent" and "good"

performing stains had a much narrower interquartile range. Variation in glycerol to methanol

ratio was found be wider among commercially prepared Giemsa stain solutions (**Fig 6b and c**).

Fig 7. Comparison of proportion of Azure A, Azure B, Methylene Blue, Eosin B, and Eosin Y

present in Romanowsky-type stains used for malaria microscopy in Plateau State by staining

quality.

Fig 8. Comparison of proportion of Azure B to Eosin Y ratio, Methylene Blue to Azure B ratio, Proportion of RGE-relevant dyes, and Proportion of RGE-irrelevant dyes present in Romanowsky-type stains used for malaria microscopy in Plateau State by staining quality.

Fig 9. Comparison of pH of stock solution of Romanowsky-type stains used for malaria microscopy in Plateau State, Nigeria by stain type and staining quality.

Factors Associated with Quality of Staining

In determining the factors associated with satisfactory staining outcome (quality of staining), we combined stains with "fair" and "good" staining outcome as "satisfactory" (coded as 1) and those with "poor" staining outcome as "unsatisfactory" (coded as 0). At bivariate analysis, it was found that having a proportion of Azure B greater than 20% was the only factor associated with satisfactory staining outcome (crude OR = 7.88, 95% CI: 2.45-28.40, p<0.001). However, stain type being Giemsa (crude OR = 1.61, 95% CI: 0.58-4.57, p=0.40); in-house prepared stains (crude OR = 1.5, 95% CI: 0.57-3.58, =0.40); alcohol-based stains (crude OR= 1.40, 95% CI: 0.56-3.71, 0.50); most prominent dye being Azure B (crude OR= 2.09, 95% CI: 0.39-10.20, p=0.40); first four most prominent dyes inclusive of RGE-relevant dyes (crude OR= 1.64, 95% CI: 0.32-7.25, p= 0.50); proportion of Azure A less than 20% (crude OR= 2.20, 95% CI: 0.82-6.64, p=0.14); proportion of Eosin B less than 5% (crude OR= 1.63, 95% CI: 0.57-5.42, p=0.40); Azure B to Eosin Y ratio greater than 1.5 (crude OR= 2.23, 95% CI: 0.89-5.79, p=0.09); Methylene Blue to Azure B ratio less than 1 (crude OR= 1.54, 95% CI: 0.57-4.20, p= 0.40); pH of stock stain solution (crude OR= 1.23; 0.60-2.67, p=0.60), and sulphated ash percent of less than 30% (crude OR= 1.92, 95% CI: 0.45-7.57, p-value= 0.4) were associated with satisfactory staining outcome but not statistically significant. Independent predictors of satisfactory staining outcome were proportion of Azure B greater than 20% (adjusted OR= 15.10, 95% CI: 2.62-115.00, p= 0.007); Azure B to Eosin Y ratio greater than or equal to 1.5 (aOR= 4.55; 95% CI: 1.03-25.1, p=0.004); and stains solution formulated with an alcohol-based solvent (adjusted OR= 27.70, 95% CI: 2.43-462.00, p= 0.012) (**Table 3**). Similar predictors were obtained when the analysis was stratified by mode of stain preparation (**Table 4**).

Table 3. Factors associated with satisfactory staining outcome among Romanowsky-type stains used for malaria microscopy in Plateau State, Nigeria.

Characteristic	CrudeOR[1]	95% CI[1]	p-value	aOR[1]	95% CI[1]	LRT p-value
Stain type						
Field	1.00	Ref				
Giemsa	1.61	0.58, 4.57	0.4			
Leishman	1.07	0.29, 3.71	>0.9			
Methylene blue to Azure B ratio <1						
No	1.00	Ref	0.4			
Yes	1.54	0.57, 4.20				
Proportion of Eosin Y >15%						

Characteristic	CrudeOR[1]	95% CI[1]	p-value	aOR[1]	95% CI[1]	LRT p-value
No	—	—	0.6			
Yes	0.76	0.28, 1.97				
Azure B Eosin Y ratio >=1.5						
No	1.00	Ref	0.090	1.00	Ref	**0.057**
Yes	2.23	0.89, 5.79		4.55	1.03, 25.1	
Proportion of Eosin B <5%						
No	1.00	Ref	0.4			
Yes	1.63	0.57, 5.42				
Proportion of methylene blue >25%						
No	1.00	Ref	0.8	1.00	Ref	0.12
Yes	0.87	0.32, 2.25		4.12	0.73, 29.1	
Proportion of Azure A <20%						
No	1.00	Ref	0.14	1.00	Ref	0.197
Yes	2.20	0.82, 6.64		3.06	0.65, 16.5	
Proportion of Azure B >20%						
No	1.00	Ref	**<0.001**	1.00	Ref	**0.004**
Yes	7.88	2.45, 28.4		15.1	2.62, 115	
1ˢᵗ four dyes include both RGE dyes						
No	1.00	Ref	0.5			
Yes	1.64	0.32, 7.25				
Methylene blue as most prominent dye						
No	1.00	Ref	0.7			
Yes	0.80	0.30, 2.03				
Antagonistic dyes present						
No	1.00	Yes	0.3	1.00	Ref	0.252
Yes	0.35	0.02, 2.10		0.25	0.01, 2.49	
Aniline blue as most prominent dye						
No	1.00	Ref	0.4			
Yes	0.41	0.02, 2.59				
Eosin B or Y as most prominent dye						
No	1.00	Ref	0.3	1.00	Ref	0.100
Yes	0.46	0.10, 1.57		0.22	0.02, 1.41	
BCB as most prominent dye						
No	1.00	Ref	0.9			
Yes	1.14	0.23, 4.48				
Azure A as most prominent dye						
No	1.00	Ref	0.7			
Yes	0.80	0.30, 2.03				
Azure B as most prominent dye						
No	1.00	Ref	0.4			
Yes	2.09	0.39, 10.2				
Number of unknown dyes <7						
No	1.00	Ref	>0.9	1.00	Ref	0.072
Yes	1.01	0.29, 3.04		13.3	0.70, 571	
Number of RGE-irrelevant dyes >7						
No	1.00	Ref	0.7	1.00	Ref	0.081
Yes	0.77	0.23, 2.27		0.09	0.00, 1.07	
pH	1.23	0.60, 2.67	0.6	1.26	0.46, 4.18	0.7
Mode of preparation						
Commercial	1.00	Ref				
In-house	1.50	0.57, 3.85	0.40			
Solvent type						

Characteristic	CrudeOR[1]	95% CI[1]	p-value	aOR[1]	95% CI[1]	LRT p-value
Water-based	1.00	Ref	0.5	1.00	Ref	**0.012**
Alcohol-based	1.40	0.56, 3.71		27.7	2.43, 462	
Sulphated ash percent category						
Unknown	1.00	Ref		1.00	Ref	0.569
< 30%	1.92	0.45, 7.57	0.4	0.80	0.08, 5.92	
>30%	1.11	0.32, 3.44	0.9	0.23	0.03, 1.23	

412 **Table 4.** Predictors of satisfactory staining outcome among Romanowsky-type stains used for

413 malaria microscopy in Plateau State, Nigeria stratified by mode of stain preparation

	Commercially Prepared Stains			In-house Prepared Stains		
Characteristic	OR[1]	95% CI[1]	p-value	OR[1]	95% CI[1]	p-value
Azure B to Eosin Y ratio >=1.5						
No	1.00	Ref	0.2	1.00	Ref	0.2
Yes	3.40	0.84, 15.8	0.095	4.55	1.03, 25.1	0.057
Proportion of Azure A <20						
No	1.00	Ref	0.12	1.00	Ref	0.2
Yes	3.30	0.76, 16.6		3.06	0.65, 16.5	
Proportion of Methylene blue >2						
No	1.00	Ref	0.2	1.00	Ref	0.12
Yes	3.07	0.59, 19.1		4.12	0.73, 29.1	
Presence of antagonistic dyes						
No	1.00	Ref	0.3	1.00	Ref	0.3
Yes	0.27	0.01, 2.62		0.25	0.01, 2.49	
Eosin B or Y as most prominent dye						
No	1.00	Ref	0.3	1.00	Ref	0.15
Yes	0.35	0.04, 1.95		0.22	0.02, 1.41	
Proportion of Azure B >20%						
No	1.00	Ref	**0.006**	1.00	Ref	**0.004**
Yes	11.4	2.24, 78.5		15.1	2.62, 115	
Number of unknown dyes >7%						
No	1.00	Ref	0.076	1.00	Ref	0.11
Yes	16.3	1.01, 625		13.3	0.70, 571	
Number of irrelevant dyes >7%						
No	1.00	Ref	**0.049**	1.00	Ref	0.081
Yes	0.07	0.00, 0.75		0.09	0.00, 1.07	
pH	1.19	0.47, 3.50	0.7	1.26	0.46, 4.18	0.7
Solvent type						
Water-based	1.00	Ref	**0.035**	1.00	Ref	**0.018**
Alcohol-based	13.3	1.51, 158	**0.028**	27.7	2.43, 462	**0.012**
Sulphated ash percent category						
Uknown				1.00	Ref	0.11
<30%				0.80	0.08, 5.92	
>30%				0.23	0.03, 1.23	
[1]OR = Odds Ratio, CI = Confidence Interval						

Discussion

For the first time, to our knowledge, an attempt has been made to profile the physico-chemical characteristics of the Romannowsky-type (Rowmanowsky-Giemsa stains) used for malaria microscopy in a malaria endemic setting in sub-Saharan Africa. We found that laboratories utilize Romanowsky-type stains other than the WHO recommended Giemsa stain for malaria microscopy. This may probably indicate lack of access to or non-adherence to the WHO malaria microscopy. However, this finding is similar to that reported in a study in India in which preference is given to the use of Leishman rather than Giemsa stain (Sathpathi et al., 2014). In about half of stain samples, it was found that batch number and expiration dates were missing from the product label and some manufacturers failed to indicate their company address. None of the manufacturers listed the dye components of their stains on the container label of the stains. Issues of proper labelling of commercially available Romanowsky-type stains have been reported previously (P. N. Marshall, 1974; Schulte, 1991). To achieve certification, the Biological Stains Commission (BSC) requires stain manufacturers to these critical details on their product labels (Penney et al., 2002; Schulte, 1991). Most manufacturers were found to have demonstrated evidence of non-compliance with these requirements in this study. These findings highlight the need for adequate regulation of biological stains used for clinical laboratory diagnostics in Nigeria. Ultraviolet spectroscopic and HPLC assays have revealed a high degree of variability of dye composition and the presence of RGE-irrelevant and RGE-antagonizing dyes in the stain samples analyzed in this study. Whereas the two most prominent dyes in our reference stain (Marshall et al., 1975c) were Azure B and Eosin Y, which are actually the only two dye species needed to produce the Romanowsky-Giemsa Effect (Friedrich et al., 1990a; Friedrich K, Seiffert W, 1990; R. W. Horobin, 2011; WITTEKIND, 2008; Wittekind DH and Gehring T, 1985), majority of the stain samples from various laboratories exhibited a wide range of variability with regards to the presence and proportion of these two essential dyes. This finding is similar to the results of previous commercial stains evaluation (Gordon B. Proctor and Richard W. Horobin, 1985; Horobin and Walter, 1987a; Marshall et al., 1975a; Schulte, 1991). Such variability in dye composition and concentrations has been incriminated in capricious and poor performance of Romanowsky-type stains (Marshall, 1978a; S. A. Bentley, 1980; Schulte, 1991). Variation in the proportion of the RGE-relevant dyes was conspicuous even amongst stains produced by the same manufacturer. This is not surprising as previous studies have demonstrated evidence of batch-to-batch variation in the dye composition of commercial Romanowsky-type

stains (Lubrano et al., 1977; Marshall et al., 1978, 1975c; P N Marshall, 1976b). It has been reported that

large-scale manufacture of Romanowsky-Giemsa stains rely heavily on the polychroming of

Methylene blue (Krafts et al., 2010; Lillie, 1942), which has been shown to be more difficult to control

at industrial scale, resulting in the formation various RGE-irrelevant dyes in addition to Azure B

(R. Horobin, 2011; Lillie, 1942; Löhr et al., 1974; P N Marshall, 1976b). This may have been responsible

for the presence of various spurious dyes detected in the stain samples, such as Aniline Blue and

Brilliant Cresyl Blue. We identified several RGE-irrelevant dye species present in the stain

solutions based on their uv-spectroscopic characteristics. For example, the Field stains were mostly

composed of Methylene blue, Azure A, and Brilliant cresyl blue. The Giemsa stains contained

Azure C, Neutral red, Basic fuchsine, Pyronine derivatives, Brilliant cresyl blue, Indigo carmine,

Alcian blue, Safranine O, Aniline blue and a wide range of unknown dyes. The Leishman stains

appeared to be most spurious in composition as RGE-relevant dyes were found to have been

replaced by irrelevant dyes such as Aniline blue, Neutral red, Thionine, Cresyl violet, Safranine,

Phloxine, Azure C, Alizarin red, Brilliant cresyl blue and a range of unknown dyes. These findings

are in keeping with the results obtained by Marshall et al. in evaluation of some commercial stains

in the United Kingdom (Marshall et al., 1975a). Although levels of Azure B and Eosin Y were

detected in some of the stains during HPLC assay despite the presence of spurious dyes, it became

apparent that their concentrations were inadequate to provide an adequate ratio of Azure B to Eosin

Y to produce the Romanowsky-Giemsa effect needed for satisfactory cytoplasmic and nuclear

differentiation of malaria parasites. In this current study, excellent staining of parasites was

obtained with the reference stain (Marshall et al., 1978), which yielded 25.0% and 32.8% of Azure

B and Eosin Y respectively with an attendant Azure B to Eosin Y ratio of 0.8:1.0 A similarly

performing stain was a Giemsa sample from the ANDi Laboratory, Lagos, which gave a ratio of

1.1. These findings are at variance with previous studies that demonstrated satisfactory staining

with Azure B to Eosin Y ratio ranging from 5:1 – 10:1 (Wittekind, 1979). We, however, noted in

this current study, a wide range of Azure B to Eosin Y ratio among the stain samples obtained

from various laboratories and suppliers. It has been documented that commercial manufacturers

of Romanowsky-Giemsa stains prefer an Azure B to Eosin Y ratio of 2:1 when formulating their

stains (Marshall et al., 1978; Wittekind, 1979). In this current study, a number of stain samples

exhibited Azure B to Eosin Y ratio around this commercially preferred ratio and higher.

Furthermore, this study revealed that higher ratios of Azure B to Eosin Y does not necessarily

confer excellent staining quality on the stains. This ratio was found to be quite high amongst the
Field stains (yet none exhibited excellent staining), lower in the Giemsa stains and almost
completely missing in the Leishman stains. These findings are a pointer to the fact that stain
performance does not depend on Azure Bto Eosin Y ratio alone. The Leishman stains performed
most poorly. Previous workers have demonstrated that the Leishman stain is highly capricious
both in composition and performance due to its method of preparation (R. W. Horobin, 2011;
Marshall, 1978a; S. A. Bentley, 1980). In this current study, the Leishman stains were found to contain
dyes other than Azure B and Eosin Y in order of concentration. Previous studies have demonstrated
that combinations of Methylene blue and Eosin Y and Azure A and Eosin Y are unable to produce
the RGE, resulting in the inability of the microscopist to differentiate the cytoplasm and the
chromatin dots of malaria parasites (Horobin and Walter, 1987b; Wittekind, 1983; Wittekind and
Gehring, 1985b). The finding of the presence of Eosin B in some of the stains in this current study
may not be considered problematic as previous studies have demonstrated that Eosin B could be
substituted for Eosin Y in the formulation of the Romanowsky-Giemsa stains (Friedrich et al., 1990b;
Marshall et al., 1978, 1975c; Wilson, 1907; Wittekind, 1983).

Generally, the Rowmanowsky-Giemsa stains are procured by laboratories either as commercial
ready-to-use solutions or as powders for reconstitution in appropriate solvents. In this study, we
attempted to compare the dye compositions of commercial solutions versus those constituted from
powder in-house against the reference stain of Marshall et al.(Marshall et al., 1975c). We have
demonstrated that irrespective of mode of preparation, stains that have high amounts of Azure A,
low Azure B, low Eosin Y and high Methylene blue combinations fail to give excellent staining
results. This finding agrees with the results of previous Romanowsky-Giemsa stains researchers
(Gordon B. Proctor and Richard W. Horobin, 1985; Horobin and Walter, 1987a; Lubrano et al., 1977;
Marshall, 1978a; Marshall et al., 1975c; Marshall and Galbraith, 1984). In addition, the higher the
proportion of RGE-relevant dyes in the stains, the better the performance; stains that had pH lower
than 7.0 and those with high sulphated ash percent tend to perform poorly. Specifically, among
the Giemsa stains, we noted variation in the ratio of glycerol to methanol, the solvent used to
constitute the stain. The higher the glycerol to methanol ratio, the poorer the performance of the
stains. These findings tend to be in line with factors that have been previously demonstrated to
influence performance of the Romanowsky-Giemsa stains (Marshall et al., 1975a, 1975c; P. N.
Marshall, 1975a; WITTEKIND, 2008; Wittekind DH and Gehring T, 1985). We have shown that minimum

amounts of the RGE-relevant dyes are required to achieve the Romanowsky-Giemsa effect, which enables adequate cytoplasmic and nuclear colour differentiation in a stained malaria parasite. A minimum of 25% RGE-relevant dyes (Azure B and Eosin Y) was required for satisfactory staining. However, performance was pH-dependent and also affected by mode of stain preparation. For optimum performance, pH of 7.1-7.4 was required for the working solution, prepared from powder in-house. Although the proportion of Eosin Y present in the stains seems to be unimportant as found in this current study, the ratio of Azure B to Eosin Y is a crucial factor. We noted a ratio of at least 1.4:1.0 for commercially available stains to achieve satisfactory staining of malaria parasites. It appears that this ratio is also pH-dependent and could be easily achieved with in-house prepared stains. Previous studies have canvassed the need for laboratories to produce their stains locally ("in-house") from scratch using commercial pure dyes (Marshall et al., 1975c, 1975a; S. A. Bentley, 1980). It must be noted that in this current study, the reference stain used was a Romanowsky-Giemsa stain formulated by Marshall et al (Marshall et al., 1975c) using commercial pure dyes. The stain was found to yield adequate amounts of Azure B and Eosin Y and the presence RGE-irrelevant dyes was minimal compared to the commercial stains sampled from laboratories. As demonstrated in this study, the dissolution of the RGE-relevant dyes in glycerol-methanolic solvent yields better stains, hence the use of Field and Leishman stains are not the best options for malaria microscopy. We recommend training laboratory personnel on the preparation of malaria microscopy stains from scratch using the pure RGE-relevant dyes and the evaluation of the stains for their composition. Perhaps the stain of Marshall et al. which was used as the reference Romanowsky-type stain in this study could be a possible choice as it can be readily prepared from individual dye components (Azure A, Azure B, Methylene Blue and Eosin Y). The malaria microscopist only needs to procure the individual powder dyes and dissolve in appropriate mixture of glycerol and methanol solvent and allow to brew. This would guarantee a stain of constant composition as previously demonstrated by the authors. There is urgent need to strengthen regulation of stains used for malaria microscopy and other IVDs both in Plateau State and in Nigeria at large. Further studies should be done to establish the role of impurities in the quality of colour imparted by the Romanowsky-type stains on malaria parasite and blood cells as we were unable to obtain sufficient number of powder stain samples from the facilities that participated in this study. The main independent predictors of staining quality seen in this study were the proportion of Azure B present in the stain; Azure B to Eosin Y ratio; and the solvent used to

constitute the stain. Stains containing an Azure B proportion of 20% or more were over ten times

more likely to produce satisfactory staining outcome; those with Azure B to Eosin Y ratio of 1.5

or more were about five times more likely to yield satisfactory staining; and stains constituted with

alcoholic solvents were over twenty times more likely to produce satisfactory staining outcome.

We noted that factors such as presence of RGE-antagonistic dyes, Eosin Y or B being the most

abundant dye, presence of seven or more RGE-irrelevant dyes, and higher sulphated ash content

were inimical to satisfactory staining outcome.

In conclusion, we have, for the first time, profiled the physico-chemical characteristics of

Romanowsky-type stains in Nigeria and elucidated on the factors that influence their performance

in malaria parasite staining. The continuous stocking, supply and utilization of Romanowsky-type

stains of variable composition, loaded with RGE-irrelevant dyes for malaria parasite cytoplasmic

and nuclear differentiation (the so-called Romanowsky-Giemsa effect) in stained blood films has

grievous consequences for malaria testing in Nigeria. As the NMEP presses towards malaria

elimination in Nigeria, there is urgent need to develop and implement policies targeted at

improving the quality of malaria microscopy reagents in both public and private laboratories to

curtail either over- or under-estimation of the burden of malaria. This could possibly be achieved

in collaboration with relevant regulatory agencies. The training of laboratory personnel on the

preparation of a simple and stable Romanowsky-type stain from component dyes is pivotal for as

long as microscopy remains the gold standard for malaria testing in this setting.

Methods

Study area

Plateau State is in North-central, Nigeria. The state has an estimated population of 4.5million

people (Ria R. Ghai et al., 2016). Malaria is endemic in the state with a current prevalence

among children aged 6 – 59 months estimated to be 13 – 22% by microscopy (Ilah Garba et al.,

2016). This study was carried out in the 17 Local Government Areas in Plateau State. The LGAs

are grouped into three senatorial districts- Plateau North (six LGAs), Plateau Central (five

LGAs) and Plateau South (six LGAs). Plateau north consists mostly of metropolitan LGAs while

the central and southern senatorial districts consist of majorly rural LGAs. Both public and

private health facilities (including stand-alone laboratories) exist in all three senatorial districts of

the State. There are 1492 health facilities in Plateau State. Among these are 1, 250, 220 and 22

primary, secondary and tertiary level facilities respectively. Overall, 237 (15.9%) facilities

render malaria microscopy services in both public and private facilities across the three

senatorial districts in Plateau State. The majority of primary health facilities use malaria RDT

kits for testing.

Study design and population

This was   a facility-based onsite cross-sectional survey conducted among laboratories in public

and private health facilities in Plateau State. It involved the collection of stain samples used for

malaria microscopy (Field stain, Giemsa stain and Leishman stain). Laboratory services is

provided by both public and private facilities in Plateau State, including malaria-testing services.

Malaria microscopy is done mostly by secondary and tertiary level facilities. The majority of

primary health facilities use malaria RDT kits for testing. There is no centralized stain distribution

system, so each facility source for its own stain from suppliers in the open market. While most

public facility laboratories use the WHO-recommended Giemsa stain, those in private facilities

use other variants of Romanoski-type stains such as Field stain and Leishman stain for malaria

microscopy. Moreover, while most public laboratories prefer to prepare their stains from

commercial dye powder, majority of private laboratories prefer commercially prepared, ready-to-

use stain solutions.

Sample Size and Sampling Procedure

The sample size for laboratory facilities was calculated using the formula by Yamane (1967) (Yamane, 1967), which gave a minimum sample size of 144 laboratory facilities. A multi-stage sampling technique was used for sampling laboratory facilities offering malaria microscopy services across the three senatorial districts in Plateau State. A sampling frame was generated from data of laboratory facilities obtained from the Medical Laboratory Science Council of Nigeria (MLSCN), Plateau State Zonal office and Plateau State Ministry of Health. Only facilities that had a functional microscope and Romanowsky-type stains in current use were included in the study. Facilities that employed both malaria RDT kits and microscopy were excluded. Sampling was done by first selecting facilities on the basis of senatorial district (Plateau North = 117, Plateau Central = 70, Plateau South = 50). Facilities were then further selected based on ownership either as "public" (n= 81) or "private" (n= 156). In every senatorial district, all public teaching hospitals and general hospitals were selected (census) while primary health centers and private facilities (including stand-alone laboratories) were selected by simple random sampling (balloting) until the minimum sample size was completed. In all, 92 facilities had adequate quantity of stains to offer for the study.

Data Collection

An interviewer-administered stain information questionnaire was used to capture relevant data, including date, stain type, batch number, expiry date, mode of stain preparation, form of stain and name of stain manufacturer. Twenty milliliters (20mL) of the stain currently used for malaria microscopy was collected in a labeled sterile universal bottle alongside 20mL of the water used for dilution (for those using Giemsa stain). The stain solution was immediately kept in a locked mobile cupboard to prevent stain exposure to light. For facilities, that prepared their own stains solution from powder, 20grams of the powder form was also collected. All stain samples and diluents were carefully labelled with a facility number.

Laboratory Techniques

The dye components of the stains were determined according to the methods of the Biological Stains Commission (Penney et al., 2002), using high performance liquid chromatography (HPLC) and ultra violet spectrophotometry. . The Agilent Series 1050 HPLC system with Agilent

Chemstation software was used for HPLC assay, Agilent 8453 UV-VIS spectrophotometer was

used for spectrophotometric analysis and Agilent 7697A gas chromatography analyzer was used

for stains solvent analysis. The method of stain performance evaluation described by (Marshall et

al., 1975a)) was used: Venous blood was obtained in 5ml EDTA container from three subjects-

hematological normal, megaloblastic anemia and acute granulocytic leukemia. Blood sample

confirmed to be positive for *Plasmodium falciparum* in 5mL EDTA container was obtained and

used to make thick and thin blood films on a clean, grease-free glass slide and air-dried

according to the standard method (WHO, 2016b). Methanol fixing was done for films meant for

Giemsa and Field stains staining. The number of slides made for each sample corresponded to

the number of stain samples collected and the reference stain. The dried films were stained with

the working solutions of the sampled stains according to standard hematology procedure, dried

and examined microscopically by WHO-certified expert microscopists using a pre-calibrated

microscope.

Data processing and analysis

The main outcome variable was quality of colour imparted on malaria parasites in the blood smear,

which was defined as "satisfactory" if the colour was graded as "excellent", "fair" or "good" and

"unsatisfactory" if the colour was graded as "poor" by the WHO-certified microscopists. The main

independent variables were stain type, the mode of stain preparation, proportion of methylene

blue, proportion of Azure B, proportion of Eosin Y, Azure B:Eosin Y ratio, stain pH, proportion

of RGE-relevant dye, and type of solvent used. Data was entered using Microsoft Word Excel

version 2016 and analyzed using R statistical software. We summarized categorical variables using

frequencies and percentages, while numerical variables were summarized using the median and

interquartile range. Comparison of percentages between outcome and explanatory variables was

done using Pearson's Chi-square statistic and Fisher's exact test for categorical variables and

Kruskal-Wallis rank sum test for continuous variables with p-values <0.05 considered to be

statistically significant. To estimate the strength of the association between explanatory variables

and the outcome variable, we calculated prevalence odds ratios with their corresponding 95%

confidence intervals and p-values. We used a simple logistic regression model to demonstrate the

effect of individual determinants on the likelihood of good quality staining. By estimating the

crude prevalence odds ratios and their associated 95% confidence intervals, we showed the

strength and direction of the association between each predictor and the likelihood of "satisfactory" staining outcome. We used a directed acyclic graph (DAG) (Textor et al., 2016) to illustrate the relationship between the outcome variable and the independent variables (**Supplementary Fig 1**) and a stepwise multivariable binary logistic regression approach (Akar et al., 2025; Tuboi et al., 2007) to assess the effects of multiple predictors of satisfactory staining quality simultaneously. Here, we used an Akaike Information Criterion-based stepwise model to systematically eliminate the predictors. We corrected the output of the final model for highly colinear predictors by removing variables with high variance inflation factor (VIF > 5). We determined the independence of each predictor using the likelihood ratio test. We fitted a final binary logistic regression model (estimated using maximum likelihood) to predict "satisfactory" staining outcome with proportion of Azure B >20%, proportion of Methylene Blue >25%, Eosin B or Y as most prominent dye, Azure B to Eosin Y ratio <2, presence of antagonistic dyes, number of unknown dyes <7, number of RGE-irrelevant dyes >7, pH, solvent type, and sulphated ash percent category. The model's explanatory power was moderate (Tjur's $R2 = 0.298$).` Standardized parameters were obtained by fitting the model on standardized dataset version. 95% Confidence Intervals (CIs) and p-values were computed using a Wald z-distribution approximation. This simultaneous modeling allows the calculation of adjusted prevalence odds ratios (aOR) as a measure of the contribution of individual predictors after adjusting for other factors included in the model.

Acknowledgement

We wish to appreciate Dr John Kernan and Prof Richard Horobin, both of the Biological Stains Commission for providing useful literature for this work. We appreciate the team of laboratory technologists at the Department of Chemistry, Faculty of Science, Bowen University, Iwo, Osun State, Nigeria for their insightful guidance in developing the laboratory analysis protocols used in this study.

Contributors

S.E.A., M.D., A.K., I.O., O.O., and W.O. contributed with literature review, study design and development of study tool. M.D. and A.K. supervised the field activities. S.E, M.D., A.K., O.A., O.O. and W.O. did data analysis and result interpretation. O.A, G.N. and W.T. provided relevant operational and field information and guidelines to enable implementation of the study. All authors were involved in manuscript review for publication

Funding

The study was funded out-of-pocket.

Competing interests

None declared.

Ethical approval

Ethical approval for the study was obtained from the University of Ibadan/University College Hospital Ethics Committee (Ref:UI/EC/17/0352), Plateau State Ethical Review Committee (Ref: MOH/MIS/202/VOL.T/X), Plateau State Hospitals Management Board (ADM/554/vol.1/141) and Jos University Teaching Hospital Ethics Committee (Ref: JUTH/DCS/ADM/127/XXVII/704).

Provenance and peer review

Not commissioned; externally peer reviewed.

Data availability statement

Data are available in public, open access repository.

Open access

This is an open access article distributed in accordance with the Creative Commons Attribution 4.0 Unported (CC BY 4.0) license, which permits others to copy, redistribute, remix, transform and build upon this work for any purpose, provided the original work is properly cited, a link to the licence is given, and indication of whether changes were made. See: https://creativecommons.org/licenses/by/4.0/.

References

Akar SE, Nwachukwu W, Adewuyi OS, Ahumibe AA, Akanimo I, Okunromade O, Babatunde O, Ihekweazu C, Hitachi M, Kato K, Takamatsu Y, Hirayama K, Kaneko S. 2025. Epidemiology of Yellow Fever in Nigeria: Analysis of Climatic, Ecological, Socio-Demographic, and Clinical Factors Associated with Viral Positivity Among Suspected Cases Using National Surveillance Data, 2017-2023. *J Epidemiol Glob Health* **15**:2. doi:10.1007/S44197-025-00341-W/TABLES/3

Altaras R, Nuwa A, Agaba B, Street E, Tibenderana JK, Strachan CE. 2016. Why do health workers give anti-malarials to patients with negative rapid test results? A qualitative study at rural health facilities in western Uganda. *Malar J* **15**:23. doi:10.1186/s12936-015-1020-9

Ashraf S, Kao A, Hugo C, Christophel EM, Fatunmbi B, Luchavez J, Lilley K, Bell D. 2012. Developing standards for malaria microscopy : external competency assessment for malaria microscopists in the Asia-Pacific. *Malar J* **11**:1. doi:10.1186/1475-2875-11-352

Friedrich K, Seiffert W ZH. 1990. Romanowsky dyes and Romanowsky-Giemsa effect. 5. Structural
investigations of the purple DNA-AB-EY dye complexes of Romanowsky-Giemsa staining.
Histochemistry **93**:247–256.

Friedrich K, Seiffert W, Zimmermann HW. 1990a. Histochemistry Romanowsky dyes and Romanowsky-
Giemsa effect. *Histochemistry* **93**:247–256.

Friedrich K, Seiffert W, Zimmermann HW. 1990b. Romanowsky dyes and Romanowsky-Giemsa effect. 5.
Structural investigations of the purple DNA-AB-EY dye complexes of Romanowsky-Giemsa staining.
Histochemistry **93**:247–56.

Global Malaria Programme. n.d. https://www.who.int/teams/global-malaria-programme/case-
management/diagnosis

Gordon B. Proctor and Richard W. Horobin. 1985. A Widely Applicable Analytical System for Biological
Stains: Reverse-Phase Thin Layer Chromatography. *Stain Technol* **60**:1–6.
doi:10.3109/10520292709115648

Horobin R. 2011. How Romanowsky stains work and why they remain valuable — including a proposed
universal Romanowsky staining mechanism and a rational troubleshooting scheme. *Biotechnic &*
Histochemistry **86**:36–51. doi:10.3109/10520295.2010.515491

Horobin RW. 2011. How Romanowsky stains work and why they remain valuable — including a
proposed universal Romanowsky staining mechanism and a rational. *Biotechnic & Histochemistry*
86:36–51. doi:10.3109/10520295.2010.515491

Horobin, R.W. and Murgatroyd LB. 1967. A comparison of rapid electro- phoretic and chromatographic
methods for the investigation of common histo- logical dyes. *Histochemie* **11**:141–151.

Horobin RW, Walter KJ. 1987a. Understanding Romanowsky staining. *Histochemistry* **86**:331–336.
doi:10.1007/BF00490267

Horobin RW, Walter KJ. 1987b. Understanding Romanowsky staining - I: The Romanowsky-Giemsa effect
in blood smears. *Histochemistry* **86**:331–336. doi:10.1007/BF00490267

Idowu Olufunmilayo A, Odeyemi Oluwatosin M, Babalola Ayodele S. 2015. Assessment of Accuracy and
Effectiveness of Rapid Diagnostic Test for Malaria Diagnosis at Primary Health Centres in Abeokuta
Nigeria. *International Journal of TROPICAL DISEASE & Health Nigeria* **8**:90–97.
doi:10.9734/IJTDH/2015/17725

Ilah Garba B, Sakajiki Muhammad A, Musa A, Edem B, Yusuf I, Kaura Bello N, Oladiran Adeniji A,
Kolawole T. 2016. Diagnosis of Malaria: A Comparison between Microscopy and Rapid Diagnostic
Test among Under-five Children at Gusau, Nigeria. *SubSaharan African Journal of Medicine* **3**:96–
101. doi:10.4103/2384-5147.184371

Kahama-maro J, Acremont VD, Mtasiwa D, Genton B, Lengeler C. 2011. Low quality of routine
microscopy for malaria at different levels of the health system in Dar es Salaam. *Malar J* **10**:332.
doi:10.1186/1475-2875-10-332

Khan Muhammad Amir, Walley JD, Munir MA, Khan Muhammad Aslam, Khokar NG, Tahir Z, Nazir A,
Shams N. 2011. District level external quality assurance (EQA) of malaria microscopy in Pakistan:
pilot implementation and feasibility. *Malar J* **10**:45. doi:10.1186/1475-2875-10-45

Krafts K. 1993. The Ehrlich-Chenzinsky-Plehn-Mala- chowski-Romanowsky-Nocht-Jenner-May- Gr ü
nwald- Leishman-Reuter-Wright-Giemsa-Lillie-Roe-Wilcox stain: the mystery unfolds. *Clin Lab Med*
13:759–771.

Krafts KP, Hempelmann E, Oleksyn BJ. 2010. The color purple : from royalty to laboratory , with
apologies to Malachowski. *Journal of Biotechnic and Histochemistry* **86**:7–35.
doi:10.3109/10520295.2010.515490

Lillie RD. 1942. Studies on polychrome Methylene Blue. 1I. Acid oxidation methods of polychroming.
Stain TechnoL **17**:97–101.

Löhr W, Sohmer I, Wittekind D. 1974. The azure dyes: Their purification and physicochemical properties.
I. Purification of azure a. *Biotechnic and Histochemistry* **49**:359–366.
doi:10.3109/10520297409117012

Lubrano GJ, Dean WW, Heinsohn HG, Stastny M. 1977. The analysis of some commercial dyes and
Romamowsky stains by high-performance liquid chromatography. *Stain Technol* **52**:13–23.

Maguire JD, Lederman ER, Barcus MJ, Wendy A, Meara PO, Jordon RG, Duong S, Muth S, Sismadi P,
Bangs MJ, Prescott WR, Baird JK, Wongsrichanalai C. 2006. Production and validation of durable ,
high quality standardized malaria microscopy slides for teaching , testing and quality assurance
during an era of declining diagnostic proficiency. *Malar J* **5**:92. doi:10.1186/1475-2875-5-92

Malaria | WHO | Regional Office for Africa. n.d. https://www.afro.who.int/health-topics/malaria

Malaria in Nigeria: Statistics & Facts | Severe Malaria Observatory. n.d.
https://www.severemalaria.org/countries/nigeria

Marshall PN. 1978a. Romanowsky-type stains in haemotology. *Histochem J* **10**:1–29.
doi:10.1007/BF01003411

Marshall PN. 1978b. Romanowsky-type stains in haematology. *Histochem J* **10**:1–29.

Marshall P N. 1976a. The composition of stains produced by the oxidation of Methylene Blue. *Histochem*
J **8**:431–42.

Marshall P. N. 1976. The composition of stains produced by the oxidation of methylene blue. *Histochem*
J **8**:431–42.

Marshall P N. 1976b. The composition of stains produced by the oxidation of Methylene Blue. *Histochem*
J **8**:431–42.

Marshall PN, Bentley SA, Lewis SM. 1978. Staining properties and stability of a standardised
Romanowsky stain. *J Clin Pathol* **31**:280–282.

Marshall PN, Bentley SA, Lewis SM. 1975a. An evaluation of some commerical Romanowsky stains. *J Clin*
Pathol **28**:680–685. doi:10.1136/jcp.28.8.680

Marshall PN, Bentley SA, Lewis SM. 1975b. An evaluation of some commerical Romanowsky stains. *J Clin*
Pathol **28**:680–685. doi:10.1136/jcp.28.8.680

Marshall PN, Bentley SA, Lewis SM. 1975c. A standardized Romanowsky stain prepared from purified
dyes. *J Clin Pathol* **28**:920–923.

Marshall PN, Galbraith W. 1984. Microspectrophotometric studies of Romanowsky stained blood cells.
III. The action of methylene blue and azure B. *Stain Technol* **59**:17–36.

Marshall PN, Lewis SM. 1974. A rapid thin-layer chromatographic system for Romanowsky blood stains.
Stain Technol **49**:235–240. doi:10.3109/10520297409116984

Miller E and, Sikes HD. 2015. Addressing Barriers to the Development and Adoption of Rapid Diagnostic
Tests in Global Health. *Nanobiomedicine (Rij)*. doi:10.5772/61114

MLSCN. 2018. Medical Laboratory Science Council of Nigeria Guideline on Sample Management.

Mokuolu OA, Ntadom GN, Ajumobi OO, Alero RA, Wammanda RD, Adedoyin OT, Okafor HU, Alabi AD,
Odey FA, Agomo CO, Edozieh KU, Fagbemi TO, Njidda AM, Babatunde S, Agbo EC, Nwaneri NB,
Shekarau ED, Obasa TO, Ezeigwe NM. 2016. Status of the use and compliance with malaria rapid
diagnostic tests in formal private health facilities in Nigeria. *Malar J* **15**:4. doi:10.1186/s12936-015-
1064-x

Mori M, Ravinetto R, Jacobs J. 2011. Quality of medical devices and in vitro diagnostics in resource-
limited settings. *Tropical Medicine & International Health* **16**:1439–1449. doi:10.1111/j.1365-
3156.2011.02852.x

Munene S, Songok J, Munene D, Carter J. 2017. Implementing a regional integrated laboratory
proficiency testing scheme for peripheral health facilities in East Africa Special issue: External
Quality Assessment in Laboratory Medicine Original papers. *Biochem Med (Zagreb)* **27**:110–3.
doi:10.11613/BM.2017.014

NMIS. 2015. Nigeria Malaria Indicator Survey (MIS). Abuja.

Nora Engel, Keri Wachter, Madhukar Pai, Jim Gallarda CB, Isabelle Celentano RW. 2016. Addressing the
challenges of diagnostics demand and supply: insights from an online global health discussion
platform. *BMC* **1**:132.

P. N. Marshall SABSML. 1975a. Staining properties and stability of a standardized Romanowsky stain. *J*
Clin Path **31**:280–2.

P. N. Marshall SABSML. 1975b. A standardized Romanowsky stain prepared from purified dyes. *J Clin*
Path **28**:920–3.

P. N. Marshall SML. 1974. Batch variations in commercial dyes employed for Romanowsky-type staining:
a thin-layer chromatographic study. *Stain Technol* **49**:351–8.

Penney DP, Powers JM, Frank M, Willis C, Churukian C. 2002. Analysis and testing of biological stains —
The Biological Stain Commission Procedures. *Biotechnic & Histochemistry* **77**:237–275.

Poespoprodjo JR, Douglas NM, Ansong D, Kho S, Anstey NM. 2023. Malaria. *The Lancet* **402**:2328–2345.
doi:10.1016/S0140-6736(23)01249-7

Proctor GB, Horobin RW. 1985. A widely applicable analytical system for biological stains: reverse-phase
thin layer chromatography. *Stain Technol* **60**:1–6.

Ria R. Ghai MIT, Azza El Bakry, Colin A. Chapman and TLG. 2016. Multi-method assessment of patients
with febrile illness reveals over-diagnosis of malaria in rural Uganda. *Malar J* **15**:460.

Roe, M.A., Wilcox, A. and Lillie RD. 1941. Eosinates of the Azures and Methylene Blue in preparation of a
satisfactory Giemsa stain from dyes of American manufacture. *Publ Hlth Rep, Wash* **56**:1906–1909.

Rugera SP, McNerney R, Poon AK, Akimana G, Mariki RF, Kajumbula H, Kamau E, Mpawenimana S, Said
SY, Toroitich A, Ronoh W, Sollis KA, Sonoiya S, Peeling RW. 2014. Regulation of medical diagnostics
and medical devices in the East African community partner states. *BMC Health Serv Res* **14**:524.
doi:10.1186/s12913-014-0524-2

Ruth McNerney, W.Peeling and R. 2015. Regulatory In Vitro Diagnostics Landscape in Africa: Update on
Regional Activities. *Clin Infect Dis* **61**:135–40. doi:10.1093/cid/civ553

S. A. Bentley PNMFET. 1980. Standardization of the Romanowsky staining procedure. *Analyt Quant Cytol*
2:15–8.

Sathpathi S, Mohanty AK, Satpathi P, Mishra SK, Behera PK. 2014. Comparing Leishman and Giemsa
staining for the assessment of peripheral blood smear preparations in a malaria-endemic region in
India. *Malar J* **13**:512.

Schulte EKW. 1991. Standardization of biological dyes and stains: pitfalls and possibilities. *Histochemistry*
95:319–328. doi:10.1007/BF00266958

Scott, R.E. & French RW. 1924. Standardization of biological stains. II. Methylen Blue. *Milit surg* **55**:337–
352.

Smith TA, Leuenberger Regula, Lengeler Christian, Getie Sisay, Getnet Gebeyaw, Guelig Dylan, Burton R,
LaBarre Paul, Pillai DR, Marcus B, Babcock H, Smith T, Leuenberger R, Lengeler C, Richards S, Milne

L, Kyi M, Chiodini P, Warhurst D, Haditsch M, Wongsrichanalai C, Barcus M, Muth S, Sutamihardja
A, Wernsdorfer W, Golassa L, Enweji N, Erko B, Assefa A, Swedberg G, Alemu A, Fuehrer HP, Getnet
G, Kassu A, Getie S, Noedl H, Cordray M, Richards-Kortum R, Li P, Zhao Z, Wang Y, Xing H, Parker D,
Yang Z, Baum E, Li W, Sattabongkot J, Sirichaisinthop J, Li S, Yan G, Cui L, Fan Q, Hänscheid T,
Grobusch M, Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K, Amino N, Hase T,
Polley S, Mori Y, Watson J, Perkins M, González I, Notomi T, Chiodini PL, Sutherland C, Yager P,
Edwards T, Fu E, Helton K, Nelson K, Tam M, Weigl BH, Weigl B, Domingo G, LaBarre P, Gerlach J,
Weigl B, Domingo G, Gerlach J, Tang D, Harvey D, Talwar N, Fichtenholz A, Lew B, LaBarre P,
LaBarre P, Hawkins KR, Gerlach J, Wilmoth J, Beddoe A, Singleton J, Boyle D, Weigl B, LaBarre P,
Gerlach J, Wilmoth J, Beddoe A, Singleton J, Weigl B, Singleton J, Osborn J, Lillis L, Hawkins K,
Guelig D, Price W, Johns R, Ebels K, Boyle D, Weigl B, LaBarre P, Curtis K, Rudolph D, Nejad I,
Singleton J, Beddoe A, Weigl B, LaBarre P, Owen S, Han E, Alemu A, Muluye D, Mihret M, Adugna
M, Gebeyaw M, Alemu A, Fuehrer H, Getnet G, Tessema B, Noedl H, Snounou G, Viriyakosol S, Zhu
X, Jarra W, Pinheiro L, Rosario V do, Thaithong S, Brown K, Singh B, Bobogare A, Cox-Singh J,
Snounou G, Abdullah M, Rahman H, Abdul-Ghani R, Al-Mekhlafi A, Karanis P, Polley S, González I,
Mohamed D, Daly R, Bowers K, Watson J, Mewse E, Armstrong M, Gray C, Perkins M, Bell D, Kanda
H, Tomita N, Kubota Y, Mori Y, Chiodini PL, Sutherland C, Hopkins H, González I, Polley S, Angutoko
P, Ategeka J, Asiimwe C, Agaba B, Kyabayinze D, Sutherland C, Perkins M, Bell D, Lee P, Ji D, Liu C,
Rampao H, Rosario V do, Lin I, Shaio M, Pöschl B, Waneesorn J, Thekisoe O, Chutipongvivate S,
Panagiotis K, Paris D, Imwong M, Faiz A, Hasan M, Yunus E, Silamut K, Lee S, Day N, Dondorp A.
2001. Child mortality and malaria transmission intensity in Africa. *Trends Parasitol* **17**:145–149.
doi:10.1016/S1471-4922(00)01814-6

Steven A. Harvey SI, Nina Martin, Cristina Lussiana, Elizabeth Streat, Stephanie Dolan, Nora
Champouillon, Daniel J. Kyabayinze, Robert Mugerwa, Grace Nakanwagi, Nancy Njoki,
Ratsimandisa Rova, and Jane Cunningham. 2017. Quality issues with malaria rapid diagnostic test
accessories and bufer packaging: fndings from a 5-country private sector project in Africa. *Malar J*
16:160.

Textor J, van der Zander B, Gilthorpe MS, Liśkiewicz M, Ellison GT. 2016. Robust causal inference using
directed acyclic graphs: the R package 'dagitty.' *Int J Epidemiol* **45**:1887–1894.
doi:10.1093/IJE/DYW341

Tuboi SH, Costa ZGA, da Costa Vasconcelos PF, Hatch D. 2007. Clinical and epidemiological
characteristics of yellow fever in Brazil: analysis of reported cases 1998-2002. *Trans R Soc Trop*
Med Hyg **101**:169–175. doi:10.1016/j.trstmh.2006.04.001

W. Galbraith PNMJWB. 1980. Microspectrophotometric studies of Romanowsky stained blood cells I.
subtraction analysis of a standardized procedure. *J Microsc* **119**:313–30.

WHO. 2024. World Malaria Report. Geneva: World Health Organization.

WHO. 2016a. Malaria Microscopy Quality Assurance Manual – Version 2, Version 2. Geneva.

WHO. 2016b. Malaria Microscopy Standard Operating Procedure - MM-SOP-07A (No. MM-SOP-07A), 1.
Geneva.

WHO. 2016c. Quality Control of Giemsa Stock Solution and Buffered Water (No. Version 1), 1. Geneva.

WHO-FIND. 2014. Malaria rapid diagnostic test performance: results of WHO product testing of malaria
RDTs: round 6 (2014-2015). Geneve.

Wilson TM. 1907. On the Chemistry and Staining Properties of Certain Derivatives of the Methylene Blue
Group when Combined with Eosin. *J Exp Med* **1**:645–670.

WITTEKIND D. 2008. On the nature of Romanowsky dyes and the Romanowsky-Giemsa effect. *Clin Lab
Haematol* **1**:247–262. doi:10.1111/j.1365-2257.1979.tb01090.x

Wittekind D. 1979. On the nature of Romanowsky dyes and the upplemental Content Full text links. *Clin
Lab Haematol 1979;1(4):247-62* **1**:247–262.

Wittekind DH. 2002. Romanowsky-Giemsa Stains In: Horobin, editor. Conn's Biological Stains. Oxford:
BIOS Scientific Publishers. pp. 303–312.

Wittekind DH. 1983. On the nature of Romanowsky-Giemasa staining and its significance for
cytochemistry and histochemistry: an overall view. *Histochem J* **15**:1029–1047.
doi:10.1007/BF01002498

Wittekind DH and Gehring T. 1985. On the nature of Romanowsky-Giemsa staining and the
Romanowsky-Giemsa effect. 1. Model experiments on the specificity of azure B-eosin Y stain as
compared with other thiazine dye-eosin Y combinations. *Histochemical Journal* **17**:263–296.

Wittekind DH, Gehring T. 1985a. On the nature of Romanowsky-Giemsa staining and the Romanowsky-
Giemsa effect. I. Model experiments on the specificity of Azures B-Eosin Y stain as compared with
other thiazine dye-Eosin Y combinations. *Histochem J* **17**:263–289. doi:10.1007/BF01004591

Wittekind DH, Gehring T. 1985b. On the nature of Romanowsky-Giemsa staining and the Romanowsky-
Giemsa effect. I. Model experiments on the specificity of Azures B-Eosin Y stain as compared with
other thiazine dye-Eosin Y combinations. *Histochem J* **17**:263–289. doi:10.1007/BF01004591

Yamane T. 1967. Statistics: An Introductory Analysis, 2nd Editio. ed. New York: Harper and Row.

Data Source: Cross-sectional survey conducted by Akar Stephen in Plateau State, Nigeria, 2018

Data Source: Cross-sectional survey conducted by Akar Stephen in Plateau State, Nigeria, 2018

Data Source: Cross-sectional survey conducted by Akar Stephen in Plateau State, Nigeria, 2018

Data Source: Cross-sectional survey conducted by Akar Stephen in Plateau State, Nigeria, 2018

(a) Glycerol methanol ratio versus mode of stain preparation

(b) Glycerol methanol ratio versus staining quality

(c) Glycerol methanol ratio versus staining quality and stain type

(a) Proportion of Azure A

(b) Proportion of Azure B

(c) Proportion of Methylene Blue

(d) Proportion of Eosin B

(e) Proportion of Eosin Y