From Spectral Resolution to Speech Perception: A Review of Findings in Postlingually Deafened Adult Cochlear Implant Listeners

Samin Ashjaei[1], Reed Farrar[2], Madison Paxton[1], Kathryn Morgan[2], Meisam Arjmandi[1,3]*

[1] Department of Communication Sciences and Disorders, University of South Carolina, 1705 College Street, Columbia, SC, USA

[2] Department of Psychology, University of South Carolina, 1512 Pendleton Street, Columbia, SC, USA

[3] Institute for Mind and Brain, University of South Carolina, Barnwell Street, Columbia, SC, USA

[*]Correspondence should be addressed to:

Meisam K. Arjmandi, Ph.D.

University of South Carolina

Close-Hipp Building, Room 268

1705 College Street, Columbia, SC 29208

Email: meisam@mailbox.sc.edu

Phone Number: 803-777-6574

Competing Interests Statement: Authors declare that there are no competing interests

associated with this manuscript.

Abstract

Reduced spectral resolution ability limits speech recognition in cochlear implant (CI) listeners. While several studies have examined the association between spectral resolution and speech perception, uncertainties persist regarding the strength of this link and related methodological and clinical factors. This review synthesizes prior findings on this relationship in postlingually deafened adult CI listeners using psychophysical measures of spectral resolution, evaluated based on four criteria: (1) whether they consider the categorization nature of speech recognition task, (2) whether they account for modulation frequency within and across spectral channels while capturing essential spectral modulations in speech, (3) whether they assess spectral resolution globally or at a channel-specific level, and (4) their relative time efficiency for clinical application. Many studies report a significant association, with some measures—such as test of spectral ripple discrimination threshold and its modified versions—demonstrating superior predictive capacity, yet no single measure meets all criteria. Our review highlights the critical role of methodological factors and calls for more refined, effective, and channel-specific assessment techniques that better capture the link between spectral resolution and speech perception in CI listeners. These approaches may improve clinical CI programming by addressing poorly functioning electrodes and enhancing perceptual outcomes in postlingually deafened adult CI listeners.

Keywords: Spectral resolution ability; speech recognition; postlingually deafened adult cochlear implant listeners

1. Introduction

Cochlear implants (CIs) are the most effective neural prostheses for individuals with moderate to profound sensorineural hearing loss who gain no-to-limited benefit from hearing aids (Svirsky et al., 2000; Holden et al., 2013; Caswell-Midwinter et al., 2022). CI listeners experience significant improvements in hearing, speech, and language outcomes (Svirsky et al., 2000; Holden et al., 2013; Caswell-Midwinter et al., 2022; Arjmandi et al., 2022a). However, challenges persist, including substantial individual variability and significant difficulties with speech understanding in noise (e.g., Svirsky et al., 2000; Holden et al., 2013; Caswell-Midwinter et al., 2022; Arjmandi et al., 2022a, Ashjaei et al., 2022; Ashjaei et al., 2024). Spectral resolution ability is a key focus in CI research and one of the most critical abilities for effective speech perception, particularly in noisy environments (O'Neill et al., 2019; Fu & Nogaki, 2005; Won et al., 2007; Friesen et al., 2001). It refers to listeners' abilities to resolve spectral components of complex acoustic signals such as speech (e.g., Won et al., 2007; Litvak et al., 2007). Speech signals contain numerous co-varying acoustic cues, including spectral features such as formant frequencies (primarily the first and second formants; e.g., Sagi & Svirsky, 2017; Arjmandi et al., 2018), formant transitions (which also have a temporal component; e.g., Hedrick & Carney, 1997), spectral tilt (e.g., Alexander & Kluender, 2008), spectral cues for place of articulation (e.g., Dorman & Loizou, 1996), and the spectral composition of the release burst (e.g., Repp, 1984) in vowel and consonant sounds. The ability to perceptually distinguish between speech phonemes that differ in frequency information is essential for speech perception (e.g., Winn, 2020; DiNino et al., 2016; Nie, et al., 2005; Arjmandi et al., 2022a). Specifically, understanding speech in noise heavily relies on resolving spectral information, as important spectral cues for

speech perception can be masked in the presence of competing noise (e.g., McArdle & Wilson, 2009).

Prior research on speech perception in CI listeners indicates that errors in word and phoneme recognition are heavily driven by their suboptimal access to spectral contrasts (e.g., DiNino et al., 2016; Cucis et al., 2021; Winn et al., 2016; Arjmandi et al., *in press*). The present work reviews previous findings on the relationship between spectral resolution ability and speech perception in postlingually deafened adult CI listeners, highlighting current trends and key factors contributing to variations in results. This review focuses on studies using psychophysical methods to assess spectral resolution to provide a cohesive synthesis of prior findings while minimizing variability from differing methodologies. A future review can separately summarize findings from physiological and neurophysiological approaches. We further discuss the impact of test type in assessing both spectral resolution and speech perception. Additionally, we discuss key factors that contribute to a robust relationship between spectral resolution and speech recognition in postlingually deafened adult CI listeners. Finally, we discuss the clinical implications of current approaches regarding the relationship between spectral resolution ability and speech perception.

1.1. Framework Linking Spectral Resolution and Speech Perception in CI Listeners

Although the focus of the present review is on prior findings on the relationship between spectral resolution abilities and speech perception outcomes in postlingually deafened CI listeners, we propose a theoretical framework summarizing key constructs, factors, and processes influencing this relationship (Figure 1). Future studies are warranted to examine how these factors may mediate or moderate the relationship between spectral resolution ability and speech perception outcomes. One factor is the number of active CI electrodes. CIs replace thousands of hair cells

on the typical human cochlea with only 12 to 22 electrodes (e.g., Zeng et al., 2008; Loizou, 1998; O'Donoghue, 2013; Dorman & Wilson, 2004; Svirsky, 2017; Goupell, 2015), leading to a coarse representation of the sound spectrum and limitations in distinguishing between closely spaced frequency components (harmonics) (e.g., Shannon, 1995; Goupell, 2015; Henry, & Turner, 2003; Friesen et al., 2001; Holmes et al., 1987; Cychosz et al., 2024). The electrode number is inherently limited by physical space within the scala media of the cochlea (e.g., Goycoolea et al., 1990). With this limited number of electrodes, there is a likelihood of current spread or channel interaction, where electrical stimulation from one electrode overlaps with adjacent electrodes, blurring the distinct spectral cues in speech and making it harder to perceive fine frequency differences for speech perception (e.g., Goehring et al., 2021; Jones et al., 2013; Cucis et al., 2021; Bierer & Litvak, 2016; Abbas et al., 2004).

>>>>>>>>>>>>>>>> *Figure 1 about here* <<<<<<<<<<<<<<<<<

Figure 1. Our proposed framework for the relationship between spectral resolution ability and speech perception outcomes in postlingually deafened CI listeners. The potential factors impacting this relationship are summarized. Overall, there are six categories of factors affecting this ability in postlingually deafened CI listeners: (a) the number of active electrodes/spectral channels, (b) the parameters that influence CI mapping, (c) the parameters that influence the quality by which each electrode stimulates target auditory neurons (electrode-neuron interface; ENI quality), (d) the electrode array type, (e) the audiological factors, and (f) individual differences in central auditory processing and cognitive functions. The solid arrows represent factors that can influence spectral resolution ability and speech perception, and the dashed arrows indicate intermediate connections between these factors. Note that empirical evidence for

some proposed relationships is yet limited or even nonexistent, which highlights the need for further research to examine these connections.

The representation of spectral information through CI and its impact on speech perception can also be influenced by *CI mapping parameters*. For example, monopolar stimulation strategy creates broader electrical fields compared to bipolar stimulation strategy, which could increase channel interaction (e.g., Berenstein et al., 2008; Zhu et al., 2012; Kwon et al., 2006), reduce spectral resolution (e.g., Jones et al., 2013) and speech perception outcomes (e.g., Cucis et al., 2021; Throckmorton et al., 2002). *CI speech coding strategies* could also impact spectral resolution and speech perception depending on how they encode spectral information in speech (e.g., Wouters et al., 2015; Liepins et al., 2020). Envelope-based strategies, such as continuous interleaved sampling (CIS) and advanced combination encoder (ACE), use a spectral encoding approach based on the function of auditory filter bank and envelope detection, which are limited in representing spectral fine structure cues in speech. In contrast, strategies like fine structure processing (FSP) and HiRes120 aim to transmit fine structure cues (e.g., Wouters et al., 2015). These differences can affect the accessibility of frequency information for CI listeners, particularly influencing speech perception in noise (e.g., Wouters et al., 2015). Other factors of CI mapping, including pulse width (e.g., Skinner et al., 2002; Henry & Turner, 2003), stimulation rate (e.g., Azadpour et al., 2018), and dynamic range (e.g., Loizou & Poroy, 2001), can also impact spectral resolution and consequently, speech perception, although findings have been inconsistent.

The quality by which individual CI electrodes stimulate their target auditory nerves in the cochlea is defined as the *electrode-neuron interface (ENI) quality* (e.g., Arenberg Bierer, 2010; Arjmandi et al., 2022a). It is identified as a key determinant of spectral resolution as it can cause

excessive channel interaction (e.g., Arjmandi et al., 2022a; DiNino et al., 2019; Arenberg Bierer, 2010; DeVries & Arenberg, 2018a). *Optimal electrode placement* with deeper insertion depth and closer proximity to the modiolus can improve spectral fidelity and speech perception (e.g., Wanna et al, 2014; DeVries & Arenberg, 2018b). Optimal *electrode-neuron distance* provides more focused stimulation to the auditory nerves and decreases the spread of the neural excitation, which in turn increases the frequency specificity and spectral resolution (e.g., DeVries & Arenberg, 2018b; Long et al., 2014; Goldwyn et al., 2010; Jahn & Arenberg, 2020). In addition, CI's frequency allocation map may not perfectly match the natural tonotopic organization of the cochlea, which leads to *place-frequency mismatch*. This misalignment can distort spectral cues, particularly in cases of shallow electrode insertion or anatomical differences in cochlear length (Xu et al., 2020; Canfarotta et al., 2020).

Another dimension of ENI quality is *neural health*, which involves the survival and health of the auditory nerve fibers (e.g., Goldwyn et al., 2010; Jahn & Arenberg, 2020). Even when electrodes are optimally positioned in the target region, a low number of surviving or healthy neurons may prevent the effective transmission of spectral information, limiting spectral resolution and speech perception. The electrical current could spread beyond the targeted neural population due to the conductive properties of cochlear fluids and tissues (*tissue growth and ossification*), which impacts the impedance of each channel between electrodes and the auditory nerve. This current spread reduces the ability to stimulate discrete neural populations, contributing to poorer spectral resolution and thus, reduced speech perception. *Electrode array design* may help mitigate challenges related to channel interaction and residual hearing preservation by improving electrode placement relative to the auditory nerves. For example, *perimodiolar arrays* position electrodes near the modiolus, *lateral wall arrays* align them along

the outer wall of the cochlear tympani, and *mid-scala arrays* situate them centrally within the scala tympani. Deeper insertion depth, although within the safe limits to avoid traumas, covers a wider range of frequencies along the basilar membrane, which can potentially reduce the risk of frequency-place mismatch (e.g., Zeng et al., 2008). Despite these arguments, the current evidence of their effects on spectral resolution and speech perception in CI listeners is not yet compelling.

The relationship between spectral resolution ability and speech perception outcomes can be influenced by *individual differences in central auditory processing* of spectral information and speech perception (e.g., Abbas et al., 2004; Scheperle & Abbas 2015a; Scheperle & Abbas 2015b; see also Farrar et al., 2024 for a review). This involves how the central auditory system interprets incoming auditory information from CI electrodes, including its ability to process, differentiate, and integrate frequency information, which may either facilitate (compensatory role) or hinder (maladaptive role) speech understanding. Even when peripheral factors like the CI mapping and ENI quality are optimized to some extent, individual variability in how the central auditory system processes incoming electrical signals can lead to differing perceptual outcomes (e.g., Scheperle & Abbas 2015a; Scheperle & Abbas 2015b; Farrar et al., 2024). The relationship between spectral resolution ability and speech perception can be influenced by audiological factors including etiology of hearing loss, age at implantation, pre-implantation duration of auditory deprivation, duration of CI use, residual hearing, and receiving auditory rehabilitation services (e.g., Green et al., 2007). These factors may directly or indirectly affect neural health (e.g., Otte et al., 1978; Leake & Hradek, 1988), tissue growth or ossification (e.g., Kamakura & Nadol, 2016; Wilk et al., 2016), and central auditory processing of spectral information and its connection to speech perception. For instance, higher age at implantation can negatively impact

spectral resolution (e.g., Landsberger et al., 2018; Gifford et al., 2018), which could be due to reduced neural plasticity and increased neural degeneration. Prolonged under-stimulation of the auditory nerve fibers (long duration of haring loss) can exacerbate the degeneration of auditory nerves (e.g., Otte et al., 1978; Leake & Hradek, 1988) and central auditory nervous system, which could limit the brains' ability to distinguish spectral components important for speech perception.

1.2. Importance of Spectral Resolution in Speech Recognition of CI Listeners

Studies also used speech perception scores from typical hearing (TH) listeners in response to stimuli with manipulated spectral information to simulate how spectral information is transmitted through CI signal processing algorithms. Spectral resolution can be manipulated using noise vocoders (e.g., Shannon et al., 1995; Friesen et al., 2001; Xu et al., 2005), sine vocoders (e.g., Dorman and Loizou, 1997), gaussian envelope tone vocoders (e.g., Goupell et al., 2013), and other methods to simulate spectral smearing (e.g., ter Keurs et al., 1993). After generating vocoded signals, changing the number of spectral channels, level of channel interaction, and electrode array insertion depth can simulate the spectral resolution ability in CI listeners (see Cychosz et al., 2024 for a review). The effect of the number of spectral channels in CI-simulated studies indicates that higher numbers of channels typically provide CI listeners with more distinct spectral information and are accompanied by more accurate phoneme (e.g., Shannon, 1995), words (e.g., Loizou & Dorman, 1999) and sentence recognition (e.g., Gonzalez & Oliver, 2005), specifically in background noise. However, this improvement plateaus after a certain number of channels, with scores remaining unchanged beyond a certain threshold (Shannon, 1995; Loizou & Dorman, 1999; Gonzalez & Oliver, 2005; Cychosz et al., 2024). The level of channel interaction is typically simulated using varying spectral bandwidths or filter slopes when

processing speech signal into continuous spectral bands (e.g., Başkent & Shannon, 2006; Shannon, 1995). Overall, the findings are consistent in showing less overlap between spectral channels in their electrical stimulation provides listeners with more distinct frequency information (e.g., Mehta et al., 2020; Cychosz et al., 2024). However, the filter slope values used to simulate channel interaction in CI-simulated studies may not necessarily replicate the effects of channel interaction on speech perception in actual CI listeners (e.g., Goupell, 2015). Therefore, speech recognition scores from CI-simulated studies may not be directly comparable to outcomes observed in actual CI listners.

Results on the effects of simulated insertion depth are mixed and appear to be influenced by factors such as patients' ability to recalibrate their spectral perception after a frequency shift (e.g., Zhou et al., 2010; Geurts & Wouters, 2005). Findings from these simulation studies suggest that frequency-place mismatch is a more detrimental issue than insertion depth alone (Canfarotta et al., 2020) and that modifying this feature does not fully mitigate its effects. Simulation studies provide valuable insights, as they allow researchers to manipulate parameters in controlled environments and isolate specific effects (e.g., Cucis et al., 2021; Xu et al., 2020; DiNino et al., 2016), which would be challenging in studies including CI listeners due to sample size limitations and individual variability. However, the main limitation of simulation studies is that they cannot fully capture all aspects of real-world CI listening. Individual factors, such as age at implantation and the duration of auditory deprivation, are not accounted for in these simulations. Furthermore, evidence suggests that CI listeners differ from normal-hearing individuals in their cortical processing of speech, with reduced activity in the superior and middle temporal gyri (see Farrar et al., 2024, for a review), an aspect not well represented in simulation studies. Given these strengths and limitations, the current study prioritizes findings from actual CI listeners to

provide clinically relevant conclusions while recognizing the complementary role of simulations in advancing our understanding of CI speech processing.

To enhance speech recognition outcomes in CI listeners by improving spectral resolution, it is crucial to understand the relationship between these two factors. Gaining this insight will help identify the most effective methods for assessing spectral resolution in relation to speech perception and guide the development of optimal strategies for its improvement. The primary objective of this review is to synthesize prior findings on the relationship between spectral resolution and speech perception in postlingually deafened adult CI listeners. We have examined findings in both quiet and noisy conditions, highlighted inconsistencies in the results, addressed methodological challenges in establishing this relationship, and outlined practical recommendations for future research. Limitations of existing methods have also been explored, providing insights for developing more effective measures to predict speech outcomes, emphasizing their clinical implications. By reporting the correlation coefficients for the spectral resolution–speech perception relationship in each study, we have assessed the strength of these methods in establishing the link between spectral resolution and speech perception. Additionally, we reviewed different speech perception measures employed across studies to evaluate their effectiveness in establishing this relationship. In discussing the methods for measuring spectral resolution, we considered the following criteria: 1) whether they recognize that speech recognition is a categorization task, not a discrimination task (Holt & Lotto, 2010), 2) whether they account for modulation frequency within and across spectral channels (Zeng et al., 2008; Loizou, 1998; Dorman & Wolson, 2004) and represent the most important spectral modulations in speech (Singh & Theunissen, 2003), 3) whether they provide a global measure of spectral resolution or assess it on a channel-specific basis, 4) to what extent the approach is time efficient

for clinical application. We have used these main criteria to understand the effectiveness and clinical applicability of different spectral resolution measures. For clinical applications, our goal is to evaluate whether they are sufficiently fast, single-channel-based (capable of assessing spectral resolution at the level of individual CI channels), reliable, sensitive, and accurate enough to capture the association between spectral resolution and speech perception in postlingually deafened adult CI listeners. These criteria are important for developing an effective measure of spectral resolution that evaluates individual CI channels and allows for selective manipulation (deactivation, current focusing, current steering) of channels with relatively poor spectral resolution to improve speech recognition through individualized CI programming.

2. Methods

2.1. Eligibility Criteria

We included all studies published by January 2025. The search process is explained under the study selection section. A PIO framework was adopted for this review, where P (patient) referred to postlingually deafened adult CI listeners, I (intervention) was the direct assessment measure(s) for spectral resolution ability based on psychophysical approaches, and O (outcome) was the speech perception score(s). For this review, spectral resolution assessment is considered an intervention within our PIO framework, which is used to structure our research questions, literature review, and conduct an evidence-based analysis. Only studies that assessed spectral resolution based on psychophysical approaches were included while studies using physiological and neurophysiological approaches were excluded. Psychophysical methods are perceptual approaches in which participants directly respond to auditory stimuli with varying frequency content in behavioral tasks. CI simulation studies were excluded as discussed in the introduction.

2.2. Information Sources and Search

Six electronic databases were searched to identify relevant publications, namely Academic Search Complete, Google Scholar, MEDLINE (ProQuest), PubMed, Scopus, and Taylor & Francis. These databases were accessed through the University of South Carolina's library website. Key search terms were identified based on the main goal of this study, which was to understand previous findings on the relationship between spectral resolution ability and speech perception in postlingually deafened adult CI listeners. Synonyms and abbreviations of the identified key search terms were used to ensure that all relevant publications were included. These five databases were consistently searched using the following combination of key search terms: "*speech perception*" OR "*speech*" OR "*speech recognition*" OR "*word recognition*" OR "*vowel identification/recognition*" OR "*consonant identification/recognition*" OR "*sentence recognition*" OR the same key search terms combined with "*in quiet*" or "*in noise*", AND "*adult cochlear implant listeners/users/recipients*" OR "*postlingually deafened adult cochlear implant listeners/users/recipients*" AND "*spectral resolution*" OR "*frequency selectivity*".

2.3. Study Selection

Using keywords, our search initially identified 111 papers. The study selection process began with a title and abstract screening, during which duplicate studies were removed. Studies were included if they met the following criteria: (1) participants were postlingually deafened adults with CIs, (2) spectral resolution ability was assessed using a psychophysical method, (3) a speech recognition evaluation was conducted, (4) the relationship between spectral resolution and speech perception was statistically tested, and (5) the study was published in English. Exclusion criteria included studies where (1) participants were not actual CI listeners (e.g., studies using vocoder simulations in typical-hearing listeners), (2) the focus was on temporal

resolution not spectral resolution, and (3) the study was a review paper rather than an original research article. After applying these criteria, twenty-three papers remained for full-text review. Busby (1993) was excluded because it did not statistically analyze the relationship between spectral resolution and speech perception. Among twenty-three studies, seven studies also included a control group of participants with typical hearing; however, this review focuses exclusively on their findings from CI listeners. Table 1 describes the key characteristics and findings of each study.

>>>>>>>>>>>>>>>> *Table 1 about here* <<<<<<<<<<<<<<<<<<

Table 1. Summary of the findings from prior studies investigating the relationship between measures of spectral resolution and speech recognition in postlingually deafened adult CI listeners. CI: cochlear implant, TH: typical hearing. yo: years old, R2: coefficient of determination, SNR: signal-to-noise ratio.

3. Results

3.1. Participants Characteristics

The average sample size across studies was 42 participants per study (SD = ± 93.86). However, after excluding one study with a relatively large sample size of 477 (Gifford et al., 2018), the average is 21 participants per study (SD= ±13.4), with sample sizes ranging from 8 to 31 participants. This shows that the number of participants varied largely across studies, which may influence the statistical power of studies and the robustness of their conclusions. The mean age at implantation across these studies ranged from 47 to 62.5 years old with an average duration of CI experience ranging from 9.78 months to 9.69 years. In terms of the participants' hearing configuration (unilateral versus bilateral versus bimodal), four studies included only unilateral CI listeners (Arjmandi et al., *in press*; Gifford et al., 2018; DeVries & Arenberg 2018a; Won et al.,

2015) and six studies involved bilateral or bimodal listeners as well (Won et al., 2007; Hughes & Stille 2008; Gifford et al., 2014; Winn et al., 2016; Lawler et al., 2017; DiNino et al., 2020). One study included only bilateral CI listeners (Winn & Litovsky 2015), and the remainder of the reviewed papers did not clearly specify participants' listening configuration.

3.2. Key Findings

Eighteen studies reported a statistically significant association between their measures of spectral resolution and speech recognition test scores (Arjmandi et al., *in press*; DiNino et al., 2020; Gifford et al., 2018; Lawler et al., 2017; Winn et al., 2016; Winn & Litovsky, 2015; Won et al., 2015; Jeon et al., 2015; Drennan et al., 2014; Gifford et al., 2014; Anderson et al., 2012; Anderson et al., 2011; Saoji et al., 2009; Litvak et al., 2007; Won et al., 2007; Henry et al., 2005; Henry & Turner, 2003; Donaldson & Nelson, 2000; Nelson et al., 1995). Various methods used to assess spectral resolution have demonstrated significant association with speech perception outcomes. These measures include *spectral ripple discrimination threshold* (SRD; Winn et al., 2016; Jeon et al., 2015; Anderson et al., 2011; Won et al., 2007; Henry & Turner, 2003) and its specific modifications for spectral or temporal modulations (DiNino et al., 2020; Lawler et al., 2017; Won et al., 2015; Litvak et al., 2007; Henry et al., 2005). These modified versions of SRD include spectral peak resolution threshold (Henry et al., 2005), clinical SRD (Drennan et al., 2014), *spectral-temporally modulated ripple test* (SMRT; DiNino et al., 2020; Lawler et al., 2017), two subcategories of SRD called *spectrotemporal modulation detection threshold* (STM; Won et al., 2015), *spectral modulation detection threshold* (SMD or SMT; Litvak et al., 2007; Saoji et al., 2009), as well as a quick version of SMD (Gifford et al., 2014; Gifford et al., 2018). Among these approaches, SRD was the most used approach for evaluating spectral resolution (Winn et al., 2016; Jeon et al., 2015; Anderson et al., 2011; Won et al., 2007; Henry & Turner,

2003). Electrode pitch ranking also showed a significant correlation with speech perception in two earlier studies (Zwolan et al., 1997; Nelson et al., 1995).

Speech perception in the reviewed studies was assessed using recognition of sentences (Gifford et al., 2018), key words in sentences (Lawler et al., 2017; Won et al., 2015; Anderson et al., 2011; Anderson et al., 2012; Nelson et al., 2011; Zwolan et al., 1997), monosyllabic (Gifford et al., 2018; DeVries & Arenberg, 2018a; Winn et al., 2016; Winn & Litovsky, 2015; Drennan et al., 2014; Gifford et al., 2014; Anderson et al., 2011; Anderson et al., 2012; Hughes & Stille, 2009; Won et al., 2007; Zwolan et al., 1997) or disyllabic words (Jeon et al., 2015; Won et al., 2007), and phonemes (Arjmandi et al., *in press*; DiNino et al., 2020; Jeon et al., 2015 ; Anderson et al., 2012; Anderson et al., 2011; Nelson et al., 2011; Hughes & Stille, 2008; Litvak et al., 2007; Henry et al., 2005 ;Henry & Turner, 2003; Donaldson & Nelson, 2000; Zwolan et al., 1997; Nelson et al., 1995). These assessments were conducted either in quiet conditions (Gifford et al., 2018; DeVries & Arenberg, 2018a; Winn et al., 2016; Won et al., 2015; Jeon et al., 2015; Anderson et al., 2011; Anderson et al., 2012; Nelson et al., 2011; Hughes & Stille, 2009; Won et al., 2007; Henry et al., 2005; Henry & Turner, 2003) or in the presence of various types of background noise, including multi-talker babble (Gifford et al., 2018; Lawyer et al., 2017; Won et al., 2015; Jeon et al., 2015; Anderson et al., 2011; Anderson et al., 2012), steady-state noise (Won et al., 2015; Won et al., 2007), and speech-shaped noise (Nelson et al., 2011).

Several studies using SRD found that this measure accounted for between 25% to 59% of the variation in sentences (Anderson et al., 2011), words (Won et al., 2007; Winn et al., 2016; Jeon et al., 2015; Drennan et al., 2014), or phoneme recognition scores (Jeon et al., 2015; Anderson et al., 2011; Henry & Turner, 2003), both in quiet and noisy listening conditions, as indicted by the reported coefficient of determination (R^2) values (Jeon et al., 2015; Anderson et

al., 2011; Won et al., 2007). The coefficient of determination (R^2) is a statistical measure that explains the proportion of variance in the dependent variable (speech perception variable) that is predictable from the independent variable (spectral resolution variable). However, some studies did not evaluate speech perception in both listening conditions (Winn et al., 2016; Henry & Turner, 2003), or the significant relationship was observed only in quiet or noise (Anderson et al., 2011). Four studies focused on vowel identification in /hVd/ context (Henry & Turner 2003; Dinino et al., 2020) and CNC word recognition only in quiet (Drennan et al., 2014; Winn et al., 2016), which found a significant relationship between SRD and speech recognition score. One study using SRD found no significant correlation with vowel identification in either quiet or noise, nor with any other speech materials in noise, except for word recognition in sentences presented in quiet (Anderson et al., 2011). However, in their later study, Anderson et al., (2012) reported significant correlations in both listening conditions when using spectral ripple detection thresholds, referred to as spectral modulation transfer function (SMTF). These findings suggest that SMTF may be a more effective measure compared to SRD for establishing the relationship between spectral resolution and speech perception. This finding also emphasizes the difference between two tasks of spectral ripple discrimination versus detection threshold; however, their differences and correlations were not fully discussed in other papers (Saoji et al., 2009; Litvak et al., 2007) and the discrimination task is more widely used. Overall, the findings from using SRD indicate that this measure is a relatively reliable measure of spectral resolution when examining the association between the ability to resolve spectral information and speech recognition scores in postlingually deafened adult CI listeners in both quiet and noisy conditions.

The extent to which this measure explains speech recognition ability has been highly variable (*min* = 0.25; Won et al., 2007, *max* = 0.59; Anderson et al., 2011, *SD* = 0.145).

Furthermore, although performance on this test can be affected by the contributions of individual electrodes to transmitting spectral information, it is a global measure of spectral resolution rather than a channel-specific measure. As such, it does not provide detailed information about the performance of individual CI channels, limiting the ability to identify specific channels with poor spectral resolution. This is important for any targeted modifications (e.g., channel deactivation or current focusing) in CI programming to improve speech recognition. It is worth noting that SRD captures the amplitude modulation aspects of speech perception in CI listeners; however, it does not provide a pure assessment of spectral resolution ability in this population (Aronoff & Landsberger, 2013). Instead, the *spectral-temporally modulated ripple test* (SMRT) introduces a modified version of SRD and a more precise method of assessing spectral resolution. Lawler et al. (2017) utilized SMRT and the Arizona Biomedical Institute (AzBio) sentence test in noise at an 8 dB signal-to-noise ratio (SNR). In addition, one study examined this relationship with vowel identification in quiet (DiNino et al., 2020). Both studies found a significant association between SMRT and their test of speech perception.

Studies that used SMD test to evaluate spectral resolution ability reported significant correlations with phoneme recognition in quiet ($R^2 = 0.7$ for vowel identification and $R^2 = 0.67$ for consonant identification; Litvak et al., 2007; Saoji et al., 2009). A modified, clinically-validated version of the test, the quick spectral modulation detection (QSMD), has also shown significant relationships with speech recognition scores. Gifford et al. (2014) reported an R^2 of 0.66 between QSMD scores and CNC word recognition in quiet, whereas Gifford et al. (2018) observed a substantially lower R^2 of 0.27 for the same measure. Additionally, AzBio sentence recognition in quiet and in noise at a 5 dB SNR was correlated with QSMD scores ($R^2 = 0.26$; Gifford et al., 2018). Although the QSMD addresses the limitation of the original SMD test in

being time-consuming, the marked variability in R^2 values is noteworthy, especially given the similarities in participant demographics and test materials across studies (see Table 1). In particular, the significant drop in R^2 in the larger sample study (n = 477) compared to the smaller sample (n = 76) suggests that increased variability across participants may be associated with attenuation in the observed relationships.

Electrode pitch ranking is another test of spectral resolution in CI listeners that was used to assess spectral resolution ability. It explained a significant portion of the variation in consonant recognition scores in quiet (Donaldson & Nelson 2000; Nelson et al., 1995), particularly when the association between spectral resolution and consonant recognition scores were evaluated based on the relative transmitted information on place of articulation (Donaldson & Nelson 2000). It should be noted that the perceptual separation of electrodes, rather than their fixed physical distance along the array, played a role in CI listeners' ability to rank pitch. A distinct pitch sensitivity and significant improvement in consonant recognition was achieved only at certain spatial distance between electrodes. In Donaldson & Nelson (2000), the average distance was above 4.5 mm, while Nelson et al. (1995) reported a wide range, from 0.47 mm to 8.71 mm. Furthermore, in Donaldson & Nelson (2000), the significant correlation between pitch sensitivity and consonant recognition was observed only among participants who were experienced users of the spectral peak (SPEAK) speech processing strategy with at least one year of CI use. This strategy was designed to provide more detailed spectral characterization of the speech waveform (Skinner et al., 1994). This significant association was not, however, observed in the Multipeak Speech Processor (MPEAK) users who had recently switched to SPEAK, which emphasizes the importance of both programming strategy and duration of hearing through CI.

On the other hand, a group of studies found no statistically significant relationship between their measures of spectral resolution and speech perception. Forward masking psychophysical tuning curve (PTC) or forward masking spatial tuning curve (fmSTC), and the electrode discrimination test did not show a significant correlation with any of the phoneme, word, and sentence recognition tests used with adult CI listeners (DeVries & Arenberg 2018a; Anderson et. al., 2011; Nelson et al., 2011; Hughes & Stille 2009; Zwolan et al., 1997) except for preliminary findings from a recently published study, which indicated a significant correlation between PTC and vowel identification in noise ($R^2 = 0.21$; Arjmandi et al., *in press*). Notably, the index quantified from PTCs in this study was the normalized slope of tuning curve rather than peak, which may explain the finding. Although these approaches provide channel-specific measures of spectral resolution, the inconsistent findings suggest that further research is needed to determine whether they can establish a strong relationship between spectral resolution and speech perception.

Overall, the findings on the relationship between spectral resolution and speech perception appear to be strongly influenced by the specific methods used for measurement. While some approaches, such as SRD, have successfully demonstrated a robust correlation (Winn et al., 2016; Jeon et al., 2015; Anderson et al., 2011 & 2012; Won et al., 2007; Henry & Turner, 2003), others, like PTC (DeVries & Arenberg 2018a; Anderson et. al., 2011; Nelson et al., 2011; Hughes & Stille 2009), have failed to establish a meaningful link, although a recent study showed a significant association between PTCs and vowel identification in noise (Arjmandi et al., *in press*). Even among the successful approaches, variability in R^2 values suggests a range of performance in their predictive power and practical utility, highlighting inconsistencies in their overall effectiveness. Beyond predictive accuracy, it is critical to evaluate

these measures based on several key criteria. Time efficiency is particularly important, as tests that require excessive time or resources may be impractical for clinical use. Another important consideration is whether a measure provides global spectral resolution information or channel-specific evaluation. Global measures, while valuable for assessing overall performance, may overlook variations across individual electrodes that are, in fact, critical for targeted adjustments of the CI programming such as deactivation of poor electrodes. In fact, several studies showed that ENI quality, channel interaction, and frequency selectivity vary across CI channels (Arjmandi et al., 2022a; Jahn et al., 2020; Abbas et al., 2004). Additionally, a test of spectral resolution should be evaluated based on its ability to capture the categorization nature of speech perception beyond pure sound discrimination (Holt & Lotto, 2010; Winn & Litovsky, 2015). Speech perception involves interpreting continuous acoustic signals and mapping them onto distinct linguistic categories, such as phonemes or words. This categorization process relies on access to both coarse- and fine-grained acoustic information and the ability to group similar sounds into meaningful units (Holt & Lotto, 2006; Holt & Lotto, 2010). Traditional measures of spectral resolution, however, often focus on the ability to distinguish between isolated frequencies or tones and may not fully capture how listeners categorize complex, speech-like signals. As a result, these measures might overlook important aspects of how the spectral resolution ability of CI listeners relates to their ability to process, perceive, and understand speech. Finally, an effective measure should account for the unique nature of signal transmission in CIs, which conveys modulation frequency within and across spectral bands (e.g., Zeng et al., 2008; Goupell, 2015). This distinction is important, as measures that fail to align with this characteristic may not accurately capture how variation in spectral resolution ability is associated with variation in speech perception outcomes.

Taken together, these considerations emphasize the need for selecting or developing measures that balance predictive accuracy with clinical practicality. Such measures should not only establish a robust relationship between spectral resolution and speech perception, but also assess spectral resolution at the level of individual CI channels within a clinically feasible timeframe. In addition, the observed variability across results may indicate the influence of several factors, including the choice of speech stimuli (sentence or words or phonemes), noise type (static or multi-talker babbling or speech-shaped noises), specific features of spectral resolution tests such as ripple density in SRD, as well as individual differences in cortical processing of spectral information in speech (Bálint et al., 2024; Farrar et al., 2024; Finke et al., 2016). The latter remains understudied, and there is limited information available on cortical processes in CI listeners in response to variations in spectral resolution and speech-evoked cortical activities (see Farrar et al., 2024 for a review).

3.3. Synthesis of Results

Several factors may contribute to the inconsistent and variable findings observed across these studies. These factors can be broadly categorized into the assessment methods used to measure spectral resolution and those used to evaluate speech perception. In the following sections, we will examine these factors in detail and compare findings to better understand their potential influences on the relationship between spectral resolution and speech perception in adult CI listeners.

3.3.1. *Potential Impacts of Spectral Resolution Measures*

There are fundamental differences in the methodologies used to evaluate spectral resolution in the reviewed studies. The R^2 values suggest that some of these approaches are more reliable in exploring the relationship between spectral resolution ability and speech perception. The SRD

test has been extensively used for this purpose (Winn et al., 2016; Jeon et al., 2015; Anderson et al., 2011 & 2012; Won et al., 2007; Henry & Turner, 2003). In this test, the listener is presented with broadband noises with amplitude-modulated (rippled) power spectra (e.g., Supin et al., 1994; Supin et al., 1997; Supin et al., 1999). The task is to differentiate between a reference and a target sound, where the target has its spectral peaks and valleys inverted in a phase compared to the reference sound. As the definition implies, this test includes three manipulable components: *modulation depth*, *modulation rate* (frequency), and *ripple density* (frequency of ripples per octave, RPO). Modulation depth and rate are used to assess individuals' access to temporal information in the signal envelope, while ripple density involves a spectral manipulation applied to the envelope of the stimuli. Discrimination of a smaller modulation depth, higher temporal modulation rate, and a larger number of ripples per octave is typically interpreted as a sign of better spectral resolution ability (Winn et al., 2022; Winn et al., 2016). Many of these studies used a typical spectral ripple discrimination task which is conducted using a modulation depth of 30 dB and temporal modulation rate of 5 or 10 Hz with adaptive ripple frequency or density starting from 0.5 ripple per octave (Lawler et al., 2017; Won et al., 2015; Jeon et al., 2015; Drennan et al., 2014; Anderson et al., 2011; Henry & Turner 2003).

To examine the effect of temporal or spectral modulation of signal on spectral resolution results, each of these features can be manipulated while keeping the other features constant. To investigate an effect of spectral ripple frequency, some studies looked at a limited number of densities including 0.5, 1 and 2 RPO, which are considered low densities (Winn et al., 2016; Won et al., 2015; Litvak et al., 2007). Another study evaluated these low densities with more detail at 0.25, 0.75 and 1.5 RPO in addition to the 0.5, 1 and 2 RPO densities (Anderson et al. 2012). To investigate the effect of higher ripple densities, some studies explored a wide range of ripple

frequencies per octave with 13 (Henry & Turner, 2003) or 14 different RPOs, including 0.125, 0.176, 0.25, 0.354, 0.5, 0.707, 1.0, 1.414, 2.0, 2.828, 4.0, 5.657, 8.0, and 11.314 RPO (Jeon et al., 2015; Drennan et al., 2014; Won et al., 2007; Henry et al., 2005). While a general finding from these studies show a significant correlation exists between ripple discrimination threshold task (ripple detection in Anderson et al., 2012) with speech perception, it is also evident that high ripple densities result in some level of aliasing (Anderson et al., 2011; Winn & O'Brien, 2021) and temporal smearing of the amplitude modulation (Lawler et al., 2017), particularly in CI listeners. Also, some of the observed relationships between our target variables, spectral resolution and speech perception, were only evident when SRD was measured by low ripple densities (Won et al., 2015; Anderson et al., 2012). Moreover, this measurement is indicated to be time-intensive with low feasibility for clinical applications. To address this limitation, some studies used the quick spectral modulation detection (QSMD) measure as a quick version of spectral ripple detection measure that requires the listener to discriminate a flat spectrum noise from one with spectral peaks and valleys. In this design, target signal is presented with five modulation depths (10, 11, 13, 14, and 16 dB) and two modulation frequencies (0.5 and 1.0 cycle/oct) (Gifford et al., 2014 & 2018). They found significant correlations with CNC word recognition and AzBio sentence recognition in quiet and with multi-talker babble noise in a +5 dB SNR. Another study modified the conventional SRD test with randomized spectral ripple envelope phase, 12 ripple densities (0.125, 0.176, 0.25, 0.354, 0.5, 0.707, 1.0, 1.414, 2.0, 2.828, 4.0, 5.657 RPO) and denser noise, introduced as clinical SRD. After being validated and showing high correlation with the conventional approach (Drennan et al., 2014), clinical SRD successfully explained an average of 58% variation in CNC word recognition scores in 28 postlingually deafened adult CI listeners (Drennan et al. 2014).

As highlighted by Azadpour and McKay (2012), several factors could complicate the interpretation of SRD test results. One such factor is the emergence of local loudness cues. For example, if a listener concentrates on a narrow frequency band that does not cover a complete ripple cycle, any inherent loudness difference between the target and reference stimuli in that band may serve as an unintended cue. Additionally, cues might arise from the extreme frequency edges of the stimuli. When listeners focus on either the high or low frequency limits, differences in the maximum or minimum audible frequencies between the two stimuli could inadvertently signal the distinction. Moreover, the overall balance of frequencies, quantified as the spectral centroid or weighted mean frequency, might shift between stimuli. If listeners are sensitive to this shift, they may perceive an audible difference even when the primary manipulation is intended to affect only the ripple structure. Collectively, such confounding factors can lead to overestimation of spectral resolution measurement via SRD in CI listeners (Aronoff & Landsberger, 2013; Winn & O'Brien, 2022). Arguably, SMRT eliminates the role of these extraneous cues in spectral resolution by drifting the modulation phase over time in both target and reference stimuli (Aronoff & Landsberger, 2013). SMRT showed significant association with vowel identification in quiet (Dinino et al., 2020) and with AzBio sentence recognition in noise (Lawler et al., 2017). Won et al. (2015) found that their STM measure remained significantly associated with sentence recognition scores even after controlling for the effect of temporal modulation detection. Notably, the R^2 values for STM and sentence recognition slightly increased after accounting for this effect (from 0.38 to 0.42 for K-CID and from 0.22 to 0.28 for K-HINT), though this increase may not be statistically meaningful. This suggests that a substantial portion of the variability in speech recognition can be attributed to participants' spectral modulation detection ability.

Another psychophysical approach used for the purpose of spectral resolution assessment is forward-masking PTCs, which involves presenting a brief probe stimulus at a constant, low current level (10%–30% of the dynamic range), which is expected to activate a small region of neural elements associated with the test electrode (i.e., probe electrode; Nelson, & Freyman, 1984; Shannon, 1990). Forward masking is used to determine the masker level at which the masker signal just barely masks the probe signal. The masker signal can vary in location along the electrode array, ranging from the probe electrode (on-frequency masker) to electrodes positioned on either apical or basal sides relative to the probe electrode. The forward-masking spatial tuning curve (fmSTC) or psychophysical tuning curve (PTC) is a graph that illustrates the masker level on different electrodes that just mask a probe stimulus fixed at a specific level (Moore, 1978; Arenberg et al., 2022). In other words, it shows how the level of a masking stimulus, presented shortly before the probe, must vary across electrodes to effectively mask the perception of the probe stimulus as a separate tone. Typically, masker levels are lowest near the probe electrode and increase with distance from the probe electrode, both apically and basally, resulting in a characteristic tuning curve shape (i.e., PTC). The sharpness of PTC is typically quantified as an indicator of spectral resolution on the probe electrode, with sharper PTCs indicating greater spectral resolution for the location of probe electrode along the electrode array (e.g., Nelson et al., 2008; Nelson et al., 2011). Despite the precise information that this approach provides about channel specific spectral resolution ability, so far, no significant correlation has been reported between PTC/STC sharpness and speech perception measures, including CNC word recognition in quiet (DeVries & Arenberg 2018a; Hughes & Stille 2009), phoneme recognition in quiet, and sentence recognition in both quiet and noise (Nelson et al., 2011; Hughes & Stille 2009). The only exception is a recent study where mean PTC slope obtained

across apical and basal sides was associated with vowel identification in noise, which is a task that heavily relies on resolving spectral information (Arjmandi et al., *in press*). Although almost all of the studies reviewed here did not find a significant relationship between PTCs and speech perception, this approach still offers valuable insights into channel-specific spectral resolution. However, for predicting speech perception, factors such as the type of speech stimuli used and the methods for quantification of PTCs, such as slope rather than peak, may provide greater predictive value.

Some of the earlier methods for assessing spectral resolution focused on individuals' sensitivity to place-pitch, that is, the psychophysical perception of the place of stimulation within the cochlea, which serves as a key cue for pitch perception (Donaldson & Nelson, 2000; Nelson et al., 1995). This approach, which is called electrode pitch ranking, involves asking participants to rank and judge the tones' frequency presented in a pair of electrodes within certain spatial distances of 0.75, 1.5, 3.0, and 4.5 mm. In some subjects who required more electrode separation, they were tested with extended distances of 6.0 and 7.5 mm (Donaldson & Nelson, 2000). Accurately ranking two closely spaced electrodes indicates good neural survival and spectral resolution ability (Townshend et al., 1987; Nelson et al., 1995). This approach was specifically tested for its association with consonant identification, with a focus on analyzing how well place of articulation information is transmitted (Donaldson & Nelson, 2000; Zwolan et al.1997; Nelson et al. 1995). Use of place of articulation cues in consonant identification requires precise access to spectral information (Sydral, 1983). The analysis was centered on groups of consonants where differentiation relies heavily on their place of articulation, such as /m/ versus /n/ sounds or /f/ versus /v/ sounds. These studies showed considerable variability among CI listeners in place-pitch sensitivity (Donalson & Nelson, 2000). However, a general trend

observed was that CI listeners with better place-pitch sensitivity, measured as accurate pitch ranking in closely spaced electrodes, tended to have better consonant recognition due to improved utilization of spectral cues like place of articulation (Donaldson & Nelson 2000; Nelson et al., 1995). It was found that participants needed varying amounts of spatial separation between stimulation points on the cochlea, ranging from 0.47 mm to 8.71 mm, to achieve a certain level of sensitivity to place-pitch differences. This degree of spatial separation explained 37% of the variability in how accurately participants perceived the place-of-articulation features of consonants (e.g., distinguishing between sounds like /b/, /d/, and /g/; Donaldson & Nelson 2000). This significant correlation was achieved only with spatial separation above 4.5 mm in one group (Nelson et al., 1995).

Another method for evaluating place-pitch sensitivity involved including perceptually discriminable electrodes in individual's everyday CI map (Zwolan et al., 1997). This method uses an adaptive electrode discrimination task, where participants are asked to determine whether two intervals with different frequencies are perceived as the same or different. Discriminable electrodes were included in an experimental map and a wide range of speech perception tasks were assessed (phoneme, word, and sentence recognition). The results varied widely among the participants, but a subgroup of participants showed a significant increase in their speech perception scores when using only the discriminable electrodes in their CI programming. This finding demonstrates the value of considering electrode-specific tasks when measuring the relationship between spectral resolution and speech perception. However, even when using electrode-specific tasks to assess spectral resolution, the ability to predict speech perception may vary depending on the type of speech stimuli and individual differences among participants. These spectral resolution assessment approaches have limitations that constrain their

effectiveness as comprehensive tools for exploring the link between spectral resolution and speech recognition ability in postlingually deafened adult CI listeners.

Despite the significant correlation found between SRD and its modified versions with speech perception scores (Winn et al., 2016; Jeon et al., 2015; Anderson et al., 2011; Anderson et al., 2012; Won et al., 2007; Henry & Turner, 2003; Drennan et al., 2014; Lawler et al., 2017; Dinino et al., 2020), they provide a global measure of spectral resolution. That is, they are not able to provide channel-specific assessment of spectral resolution ability. Also, SRD relies on a non-speech stimulus (i.e., broadband noise) and, therefore, cannot capture all aspects of auditory processing that link spectral resolution to speech perception in CI listeners. Moreover, except the QSMD, of which there is very little evidence available with variable predictive capacity (Gifford et al., 2014; Gifford et al., 2018), other SRD methods are time-consuming with limited feasibility in clinical settings. Similar to SRD, PTCs and electrode pitch ranking and discrimination are non-speech tasks that are based on pure tone stimuli. However, they can measure spectral resolution at the level of individual CI channels, which makes channel-specific assessments of spectral resolution possible. However, to this date, these methods have not shown promising results to effectively establish the link between spectral resolution ability and speech recognition in adult CI listeners. These challenges, combined with the relatively long testing duration, have limited the clinical use of PTCs and electrode pitch-ranking methods for assessing spectral resolution and identifying poorly performing electrodes for selective manipulations, such as channel deactivation (Bierer & Litvak, 2016; Noble et al., 2014; Gifford et al., 2016; Noble et al., 2013), current focusing (Srinivasan et al., 2013; Berenstein et al., 2008; de Jong et al., 2019; Devries & Arenberg, 2018b), or current steering (Berenstein et al., 2008). Despite these

limitations, methods such as PTCs remain the gold standard for assessing spectral resolution and evaluating how effectively each CI electrode transmits its designated frequency information.

In an attempt to address these challenges and capture the nature of speech perception as a categorization task using speech cues, studies have proposed using speech-based tasks that involve categorization of speech sounds that heavily rely on resolving spectral information (Winn et al., 2016; Winn & Litovsky, 2015; Dinino et al., 2020). These tasks involve access to formant transitions (a fine spectral resolution cue) and spectral tilt (a coarse spectral resolution cue) for vowel categorization, and place-of-articulation cues for consonant categorization. Studies have shown that the ability of CI listeners to use fine spectral information, such as formant transitions, is significantly correlated with their ability in performing various speech perception tasks. For example, Winn & Litovsky (2015) and Winn et al. (2016) found correlations between formant-based spectral resolution tasks and word recognition scores. Similarly, DiNino et al. (2020) linked formant cue sensitivity to vowel identification. These findings indicate that speech-based assessments of spectral resolution may provide a promising approach for investigating the relationship between spectral resolution ability and speech perception in a more reliable and clinically relevant manner. Such assessments leverage speech stimuli, offering the potential to directly relate spectral resolution to real-world communication tasks. However, current evidence supporting the effectiveness and generalizability of this approach remains limited. To date, studies have primarily utilized a narrow set of speech stimuli, such as /ba/ versus /da/ pairs (Winn et al., 2016; Winn and Litovsky, 2015; DiNino et al., 2020), which do not encompass the broader range of frequency information transmitted through CI electrodes for different vowel and consonant sounds. Additionally, consonant categorization using spectral tilt cue failed to show a relationship with either word or vowel recognition and the

strength of this approach in capturing the relationship with speech perception dropped noticeably from one study (R^2 = 0.5, Winn et al., 2016) to another (R^2 = 0.27, Winn & Litovsky, 2015). Investigating a more comprehensive approach with a larger set of speech stimuli may help address these limitations, allowing for a more complete characterization of variations in spectral resolution across CI channels, as well as their impact on different levels of speech perception.

3.3.2. Potential Impacts of Speech Perceptions Measures

The specific type of speech test being used can affect individuals' ability to perceive and understand speech. Speech recognition scores may vary depending on whether individuals are asked to identify vowels, consonants, monosyllabic words, key words in sentences, or entire sentences. Different tasks engage different aspects of speech perception, so results may not always be directly comparable across tests. That is, speech perception is a task-dependent ability, and the context essentially affects speech perception scores. Therefore, observed differences in the relationship between measures of spectral resolution and speech perception across studies may be partly due to the variations in the speech materials used. We have not conducted a statistical analysis to determine the extent to which variations in speech perception measures contribute to the discrepancies observed in establishing a link between spectral resolution and speech perception. In fact, such an analysis would be inherently limited considering the effect of differences in participant characteristics across studies, including variability in demographic, clinical, or cognitive factors. These differences make it challenging, if not impossible, to disentangle the effects of speech perception measures from the influence of participant-related variables. Therefore, we provide an overview of different types of speech stimuli used in the reviewed studies.

Studies used variable speech materials including sentences (Gifford et al., 2016; Lawler et al., 2017; Won et al., 2015; Anderson et al., 2011 & 2012; Nelson et al., 2011; Hughes & Stille, 2009; Zwolan et al., 1997), words (Zwolan et al., 1997; Won et al., 2007; Hughes & Stille, 2009; Gifford et al., 2014; Drennan et al., 2014; Jeon et al., 2015; Winn & Litovsky, 2015; Winn et al., 2016; DeVries & Arenberg, 2018a; Gifford et al., 2018), and phoneme sounds (vowels or consonants; Arjmandi et al., *in press*; DiNino et al., 2020; Anderson et al., 2011 & 12; Nelson et al., 2011; Hughes & Stille, 2009; Litvak et al., 2007; Henry et al., 2005; Henry & Turner, 2003; Zwolan et al., 1997). Tests of sentence recognition included AzBio, Central Institute for the Deaf (CID), Hearing-In-Noise Test (HINT), City University of New York (CUNY), Institute of Electrical and Electronics Engineering (IEEE), and Bamford-Kowal-Bench (BKB) standard sentences. In nearly all sentence recognition tasks, the percentage of correct responses was determined based on the number of accurately identified key words. For word recognition tasks, monosyllabic CNC word lists from Peterson & Lehiste, 1962, Northwestern University Auditory Test Number Six (NU-6), and Maryland University were the most used. Additionally, two studies assessed word recognition ability using spondee word lists (Won et al., 2007; Jeon et al., 2015). Phoneme recognition was also used for this purpose in which participants were tested for vowel recognition in h-vowel-d (/hVd/) contexts (DiNino et al., 2020; Anderson et al., 2011; Anderson et al., 2012; Nelson et al., 2011; Henry et al., 2005; Henry & Turner, 2003; Zwolan et al., 1997) or b-vowel-t (/bVt/) contexts in one study (Litvak et al., 2007) and consonant identification within an a-consonant-a (/aCa/) context (Jeon et al., 2015; Nelson et al., 2011; Litvak et al., 2007; Henry et al., 2005; Donaldson & Nelson, 2000; Zwolan et al., 1997; Nelson et al., 1995). In addition, the percentage of correct phonemes were evaluated during word recognition in some of these studies (Hughes & Stille, 2009; Zwolan et al., 1997). Finally, in order to functionally

capture the categorization nature of speech perception, three studies manipulated spectral cues in six consonant vowel syllables, which included liquids */ra/* versus */la/*, fricatives */sa/* versus /□*a/*, and plosives */ba/-/da/*, but their group analysis was restricted to responses to syllables in the */ba/-/da/* continuum (Winn et al., 2016; Winn & Litovsky 2015; DiNino et al., 2020).

Arguably, conventional speech perception measures are not sensitive enough to capture spectral resolution ability. Speech is inherently a multidimensional stimulus, with multiple cues working together to shape perception abilities (e.g., Toscano & McMurray, 2010; Holt & Lotto, 2006; Arjmandi et al., 2018; Rosenblum, 2008; Christensen & Humes, 1997). Considering both temporal and spectral cues involved in speech perception, even when a target sound is accurately identified, it does not guarantee that spectral cues were fully resolved. Listeners with CI may use compensatory strategies in the temporal domain to achieve good speech identification performance despite limited access to spectral cues (e.g., Moberly et al., 2014; Fleming & Winn, 2022; Peng et al., 2012; Winn et al., 2012). Cues such as consonant burst spectrum and formant transitions in consonants or formant frequencies in vowels are reliable representatives of spectral cues (Hillenbrand et al., 1995; Lehiste, 1965). Therefore, vowel identification and consonant categorization tasks that manipulate spectral cues (e.g., the first and second formant frequencies, cues for place of articulation) can reveal how effectively listeners use these spectral cues for speech perception. This approach focuses on CI listeners' functional use of spectral information in speech perception, rather than solely measuring their basic psychophysical ability to detect or resolve spectral details. As an example from CI-simulation studies, DiNino et al. (2016) showed that altering specific aspects of CI channel function, such as deactivating certain channels (creating holes) or making more overlap between channels (shallower filter slope), directly influence participants' abilities to correctly identify vowels, and the errors they made were

systematic and predictable based on the type of manipulation. Therefore, speech-based measures of spectral resolution may provide a more clinically relevant and precise way to establish the relationship between spectral resolution ability and speech perception outcomes.

Further bolstering clinical relevance, many studies have incorporated speech stimuli presented in background noise to evaluate the role of spectral resolution in understanding speech, making their findings more applicable to real-world settings (Won et al., 2007; Nelson et al., 2011; Anderson et al., 2011 & 12; Jeon et al., 2015; Won et al., 2015; Lawler et al., 2017; Gifford et al., 2018; Arjmandi et al., *in press*). The presence of noise highlights the importance of spectral resolution, as noise can either distort the transmitted information or mask it with additional energy (e.g., Harmon, 2021). In addition, compared to speech perception in quiet, understanding speech in noisy environments requires enhanced access to spectral information (e.g., Baer & Moore, 1993). Evaluating speech perception in noise also improves the ecological validity of the assessment, making it more representative of real-life listening situations. In the reviewed papers, types of noise included were steady state (Won et al., 2007), multi-talker babble (Won et al., 2007; Arjmandi et al., *in press*; Gifford et al., 2018; Jeon et al., 2015), and speech shaped noise (Nelson et al., 2011) with variable SNRs ranging from -6 dB to +21 dB and more focus on a 0 dB to 10 dB SNR (Arjmandi et al., *in press*; Gifford et al., 2018; Lawler et al., 2017; Nelson et al., 2011). Notably, studies examining the relationship between spectral resolution and speech perception in both quiet and noisy conditions have found a significant correlation, but the strength of this correlation differed between the two listening conditions. Many of them found a stronger relationship in quiet compared to noise (Anderson et al., 2011; Anderson et al., 2012; Jeon et al., 2015; Won et al., 2015), although, the R^2 did not differ substantially in Gifford et al. (2018). By contrast, Won et al., (2007) used the steady-state noise for the noisy condition and

found stronger correlation between SRD and spondee word recognition compared to CNC word recognition in quiet.

Our review of prior findings on the relationship between spectral resolution and speech perception suggests that the type of speech materials used may influence study outcomes (Henry et al., 2005; Anderson et al., 2011; Winn et al., 2016; Winn & Litovsky 2015; Won et al., 2007). In addition to examples from the effect of listening condition, checking the relationship between the measure of spectral resolution with both vowel and consonant recognition revealed a stronger relationship with consonants (Henry et al., 2005). This stronger correlation with consonant recognition was also observed compared to word recognition in noise (Jeon et al., 2015). Additionally, in another study, while a strong correlation was observed between keyword in sentence recognition and SRD, no significant correlation was observed for vowels (Anderson et al., 2011). This could be due to factors such as differing spectral fluctuations among stimuli or varying levels of information redundancy, such as those found in phonemes compared to sentences (Anderson et al., 2011), which can impact speech perception. However, based on the current findings from the reviewed studies, we cannot attribute non-significant relationships or variability in correlations solely to differences in speech materials. This conclusion is supported by studies using the same speech tests that have produced varying, or sometimes significantly different, R^2 values across different research (Henry et al., 2005; Anderson et al., 2011; Winn et al., 2016; Winn & Litovsky 2015; Won et al., 2007). For example, testing vowel identification using the same /hVd/ contexts with SRD demonstrated a relatively strong correlation in Henry et al., (2005; $R^2 = 0.42$), while no significant correlation was observed in another study (Anderson et al., 2011). The strength of observed relationship across studies also varied noticeably. For instance, the correlation between the categorization of the spectrally manipulated consonants

with word recognition varied from R^2 of 0.5 in one study (Winn et al., 2016) to 0.27 in another one (Winn & Litovsky, 2015). Some studies employed a comprehensive battery of speech tests to assess spectral resolution and found no correlation between their spectral resolution measures and any of the speech tests, regardless of the specific test characteristics (Nelson et al., 2011; Hughes & Stille 2009; Zwolan et al., 1997). Overall, the findings suggest that differences in spectral resolution assessment methods have a greater impact on capturing the relationship between spectral resolution ability and speech perception than differences in speech perception assessment methods.

4. Discussion

We reviewed and synthesized prior findings on the relationship between spectral resolution ability and speech recognition in postlingually deafened adult CI listeners. The primary objective was to evaluate studies that employed psychophysical methods to assess spectral resolution and its association with various speech perception tasks. Among the reviewed studies, seventeen (80%) reported a statistically significant relationship between measures of spectral resolution and speech perception, with R^2 values ranging from 0.21 to 0.69 (Arjmandi et al., *in press*; DiNino et al., 2020; Gifford et al., 2018; Lawler et al., 2017; Winn et al., 2016; Winn & Litovsky, 2015; Won et al., 2015; Jeon et al., 2015; Drennan et al., 2014; Gifford et al., 2014; Anderson et al., 2012; Anderson et al., 2011; Litvak et al., 2007; Won et al., 2007; Henry et al., 2005; Henry & Turner, 2003; Donaldson & Nelson, 2000; Nelson et al., 1995). This variation indicates that the strength of the relationship is inconsistent across studies. Some studies found no significant association (Zwolan et al., 1997; Hughes & Stille, 2009; Nelson et al., 2011; DeVries & Arenberg, 2018a). We critically evaluated the spectral resolution and speech perception measures used in these studies. Specifically, we evaluated whether these measures (a) account for the

nature of speech perception as a categorization task (Holt & Lotto, 2010); (b) represent the most important spectral modulation cues in speech—spectral information available in the speech envelope within and across different spectral channels (Singh & Theunissen, 2003); (c) provide a global or channel-specific assessment of spectral resolution; and (d) assess the efficiency of test time. These criteria collectively determine their overall suitability for clinical application. This comprehensive evaluation highlights the strengths and limitations of current approaches and their potential for improving our understanding of the link between spectral resolution and speech perception in CI listeners.

4.1. Relationship Between Spectral Resolution and Speech Perception

Spectral resolution reflects a listener's ability to distinguish fine differences in frequency information (Litvak et al., 2007; Won et al., 2007), which is critical for decoding speech sounds, such as distinguishing vowels and consonants. Speech perception, particularly in noisy environments, relies heavily on spectral resolution ability to access important acoustic cues like formant frequencies, formant transitions, spectral tilt, and place of articulation (e.g., Xu et al., 2005). Accurate understanding of the relationship between spectral resolution and speech perception helps pinpoint the mechanisms that contribute to better speech recognition, particularly in the presence of background noise where many CI listeners struggle significantly (e.g., Arjmandi et al., 2022a; Oxenham, & Kreft, 2014; Zeng, 2004; Ashjaei et al., 2022). In addition, spectral resolution varies among individual CI listeners (e.g., Arjmandi et al., 2022a; DeVries & Arenberg, 2018a; Caswell-Midwinter et al., 2022) and even across individual CI channels (Arjmandi et al., 2022a; Arenberg et al., 2018) due to the loss and/or smearing of spectral information caused by channel interaction (e.g., Jones et al., 2013; White et al., 1984; Cucis et al., 2021; Throckmorton & Collins, 2002). Better understanding of this relationship is

important for researchers and clinicians to develop reliable tools to identify poorly performing electrodes that contribute to spectral degradation and poor speech perception. This allows clinicians to implement targeted interventions by adjusting CI programming strategies (i.e., deactivation) targeted at channels with poor transmission of spectral information.

The observed association between spectral resolution ability and speech perception varied across studies, which could be attributed to differences in the assessment methods, as well as other audiological factors and the demographics of the study population (Figure 1). Among spectral resolution methods, SRD emerges as the most prevalent method for evaluating spectral resolution and accounts for a considerable percentage of the variance in speech recognition scores (25% to 59%). Although SRD shows reliable performance in assessment of spectral resolution and its association with speech recognition, its effectiveness varies depending on the technical details such as ripple density, test condition (quiet versus noise), and participants' experience with CI (Winn et al., 2016). Modified versions of SRD, particularly SMRT and SMD, also provide insight into this relationship. For instance, studies like Lawler et al. (2017) showed that SMRT is significantly correlated with sentence recognition performance in noise, highlighting its utility in more complex auditory environments. Moreover, while SMD demonstrated significant correlations with phoneme recognition (Litvak et al., 2007; Saoji et al., 2009), its time-consuming nature prompted the development of a faster version, QSMD, which still maintains a relative association with word recognition (Gifford et al., 2018; Gifford et al., 2014).

The measure of electrode pitch ranking presents an intriguing divergence from channel-specific assessments of spectral resolution, as it has shown significant correlations with consonant recognition, particularly when considering the spatial arrangement of electrodes and

their influence on pitch perception (Nelson et al., 1995; Donaldson & Nelson, 2000). However, its effectiveness appears highly dependent on background factors such as experience with specific speech processing strategies, as demonstrated by differences in outcomes between SPEAK and MPEAK users (Donaldson & Nelson, 2000). At the same time, several studies have identified methods that lack significant predictive power for speech perception (Zwolan et al., 1997; Hughes & Stille, 2009; Nelson et al., 2011; DeVries & Arenberg, 2018a). For example, PTCs and electrode discrimination tests did not yield consistent correlations with speech perception, highlighting limitations within these assessment frameworks. The observed discrepancies across findings underscore the impact of both methodological variations and individual differences among participants. These inconsistencies highlight the need for further refined experimental approaches that better account for the heterogeneous auditory profiles of CI listeners, as well as more controlled studies with carefully selected participant groups to minimize variability and enhance the reliability of findings.

4.2. Methodological Insights and Challenges

Observed variabilities in predictive capacity of different measures of spectral resolution can be partly explained by speech materials used in these studies. Sentences, words, and phonemes differ markedly in the level of redundancy and spectral information they provide to listeners (Lorenzi et al., 2006), influencing their perceptual demands and the cues needed for accurate recognition (Xu et al., 2005; Repp, 1982). Sentences offer the highest level of redundancy due to their syntactic and semantic context, prosodic features, and intonational patterns (e.g., Gardner et al., 1975), which collectively provide multiple cues that can compensate for gaps in spectral resolution (e.g., Shannon et al., 2004). Recognition of isolated words, while still benefiting from lexical and contextual cues, require more precise spectral resolution for accurate recognition, as

the listener relies heavily on clear representation of formant transitions and phonemic contrasts within a limited context. Recognition of phonemes, on the other hand, involves the least contextual cues, as they lack contextual or prosodic support and depend almost entirely on distinct spectral and temporal features, such as formant frequencies and their transition over time. This gains particular importance when discussing access to spectral information in CI listeners, which is mainly through access to modulation frequency within and across CI channels. The use of speech stimulus such as sentences rich in temporal cues can lead to overestimation of CI listeners' spectral resolution ability. Consequently, the choice of stimulus can substantially impact test outcomes and its association with spectral resolution ability. Higher redundancy in sentences may underestimate spectral resolution deficits, while phoneme-based tasks can reveal more granular details about individuals ability to resolve spectral information. Although testing with sentences better reflects real-life scenarios, more sensitive stimuli such as phonemes are necessary for fine-tuning CI programming, as they provide a more precise assessment of spectral resolution, minimize cognitive compensation, and help optimize electrode stimulation parameters and strategies for a subset of electrodes with poor spectral resolution.

Furthermore, the type of spectral resolution test and the psychophysical and functional aspects it targets can influence its relationship with speech perception. Speech perception is fundamentally a process of categorization rather than discrimination (Holt & Lotto, 2010); however, some of the current approaches for spectral resolution assessment fail to capture this characteristic (DeVries & Arenberg 2018a; Anderson et. al., 2011; Nelson et al., 2011; Hughes & Stille 2009; Zwolan et al., 1997; Arjmandi et al., *in press*; Gifford et al., 2018; Lawler et al., 2017; Won et al., 2015; Jeon et al., 2015; Drennan et al., 2014; Gifford et al., 2014; Anderson et al., 2012; Anderson et al., 2011; Litvak et al., 2007; Won et al., 2007; Henry et al., 2005; Henry

& Turner, 2003; Donaldson & Nelson, 2000; Nelson et al., 1995). In contrast, speech-based tasks of spectral resolution assessment are proposed to address this limitation (e.g., Winn et al., 2016; Winn & Litovsky, 2015; DiNino et al., 2016). Findings suggest that categorization of syllable pairs that are spectrally manipulated (e.g., formant cues, spectral tilt, and place of articulation cues) has the potential to provide an effective approach for spectral resolution assessment (e.g., Winn et al., 2015 & 2016). While these tasks remove the potential assistance of semantic, spectral, and temporal redundancies, they evaluate the perceptual spectral cue weighting ability in CI listeners. Unlike traditional measures that focus on absolute frequency discrimination (e.g., DeVries & Arenberg 2018a; Nelson et al., 2011), cue weighting reflects the functional ability to perceive and prioritize critical acoustic information within speech contrasts. This approach is particularly advantageous because it accounts for the real-world application of spectral cues in auditory-linguistic processing, capturing subtle differences between individuals with varying hearing abilities who may otherwise perform similarly on broader speech recognition tasks. As such, acoustic-phonetic cue weighting emerges as a promising tool for exploring functional spectral resolution and its impact on speech perception. However, there is limited research on this approach, and its capacity to capture the relationship between spectral resolution and speech perception has been limited thus far. Another important methodological consideration, particularly regarding SRD and SMD, is that they may not exclusively measure spectral resolution. Instead, listeners might rely on other cues, such as local loudness differences between the target and reference stimuli (Azadpour and McKay, 2012; Aronoff & Landsberger, 2013; Winn, M. B., & O'Brien, G. 2022). This underscores the need to select assessment methods that accurately isolate spectral resolution as a key limitation of CI devices.

A critical feature of spectral resolution measures is their capacity to assess spectral resolution at the individual channel level. This capability is particularly important when the objective is to identify channels with relatively poor spectral resolution that significantly contribute to errors in speech recognition scores. Some of the existing methods are global measures of spectral resolution such as SRD (Winn et al., 2016; Jeon et al., 2015; Anderson et al., 2011; Anderson et al., 2012; Won et al., 2007; Henry & Turner, 2003), SMRT (DiNino et al., 2020; Lawler et al., 2017) or SMT/SMD (Litvak et al., 2007; Gifford et al., 2014). While they have been shown to have consistent predictive ability for speech recognition scores, they do not have the capacity to determine the extent to which each electrode may differentially contribute to their global spectral resolution measure. On the other hand, methods such as PTC (Arjmandi et al., *in press*; DeVries & Arenberg 2018a; Anderson et. al., 2011; Nelson et al., 2011; Hughes & Stille 2008) and electrode discrimination (Donaldson & Nelson 2000; Nelson et al., 1995) provide detailed channel-specific insight into the role that each electrode plays in transmission of spectral information to the auditory nerve. However, as discussed, these measures do not consistently show significant correlations with speech perception and may be influenced by individual differences in attention and other cognitive abilities.

Performance on many of the spectral resolution measurements and speech perception tasks as well as the results for their link may be affected by individuals' cortical auditory processing and cognitive abilities (Figure 1). More specifically, many of the measures discussed involve attention during the task. This is also true in speech in noise perception tasks as results can be significantly impacted by the level of both cognitive processing and efferent inhibitory neural pathways (e.g., Price & Moncrieff, 2021; Beckers et al., 2023). Although limited, evidence from studies comparing the speech evoked cortical activities between CI listeners and

TH individuals indicate weaker brain activity in bilateral superior temporal gyrus (STG) and middle temporal gyrus (MTG) in temporal lobe (Farrar et al., 2024; Levin et al. 2022; Steinmetzger et al., 2022), which may be attributed to suboptimal spectral resolution in CI listeners (e.g., Arjmandi et al. 2022a; DeVries and Arenberg 2018a; DiNino et al. 2016; Levin et al. 2022; Loizou & Poroy 2001). However, there is very limited evidence on the cortical processing of spectral information and its role on spectral resolution processing on speech perception outcomes (Levin et al., 2022; Steinmetzger et al., 2022; Farrar et al., 2024).

5. Conclusions

In conclusion, many of the spectral resolution measurements appear promising for predicting speech perception capabilities in postlingually deafened adult CI listeners. Some methods outperform the rest such as SRD, which could be considered as a reliable tool to assess the global spectral resolution ability and its association with speech perception performance. Furthermore, the development of several time-efficient approaches such as QSMD has helped overcome the limited clinical feasibility of some of the discussed techniques (Gifford et al., 2018). The emergence of speech-based tasks as more functional and clinically feasible approaches is beneficial for both researchers and clinicians in understanding the connection between spectral resolution and speech perception. However, some weak correlations observed in these studies necessitate further investigation to clarify the underlying mechanisms. (e.g., Winn & Litovsky, 2015).

For spectral resolution measures to be effective and clinically applicable, they need to satisfy multiple criteria. First, they need to be designed to capture the nature of CI signal processing in coding spectral information in speech and account for modulation frequency within and across spectral bands (Zeng et al., 2008; Loizou, 1998; Dorman & Wolson, 2004) and

represent the most important spectral modulations in speech. Second, they are expected to provide a channel-specific measure of spectral resolution, which is essential for selectively manipulating CI channels, whether through deactivation, current focusing, or current steering, to reduce channel interaction, enhance spectral resolution, and ultimately improve speech recognition. Finally, these measures need to be fast enough for clinical application. Some, like the QSMD test (Gifford et al., 2018), are quick, taking only a few minutes, and are readily available as downloadable software (https://www.ear-lab.org/software-downloads.html). However, they are not channel-specific. Others, such as PTCs, are channel-specific but are laborious and have shown limited ability to capture the relationship between spectral resolution and speech perception (DeVries & Arenberg, 2018a; Anderson et al., 2011; Nelson et al., 2011).

5.1. Future Research Directions

Findings from this review provide valuable insights into future research on assessing spectral resolution and its relationship with speech perception. Speech-based tasks that focus on spectral cues, such as formant frequencies and place of articulation, could complement existing approaches by specifically evaluating speech-related spectral resolution abilities in CI listeners. Additionally, channel-specific measures capable of identifying poorly performing electrodes may provide targeted insights for optimizing CI programming. An important but less frequently addressed topic is about the distinct contributions of peripheral (processing at the level of the cochlea and auditory nerves) and central spectral resolution abilities (processing at higher auditory centers in the brain) to speech perception outcomes. Although distinguishing between the contributions of peripheral and central auditory pathways in the relationship between spectral resolution and speech perception is inherently difficult and methodologically challenging, doing so is crucial for identifying the origins of spectral resolution limitations and understanding their

impact on speech understanding. If peripheral spectral resolution is the primary limitation, interventions can focus on improving peripheral limitations such as channel interaction (e.g., DeVries & Arenberg, 2018b; Bierer & Litvak, 2016; Caswell-Midwinter & Arenberg, 2023) and frequency-place mismatch (e.g., Sagi, & Svirsky, 2024). If central spectral resolution plays a larger role, training or rehabilitative therapies targeting neural plasticity could be more effective. Expanding research into the cortical processing of spectral information and speech sounds can provide deeper insights into the neural mechanisms underlying spectral resolution ability and speech perception, individual variability on these abilities, and potential avenues to maximize CI benefits. Lastly, future studies are warranted to examine how individual differences in cognitive abilities may facilitate or hinder the link between spectral resolution and speech perception in CI listeners, providing a better understanding of the factors that explain the connection between spectral resolution ability and speech perception outcomes. In addition, while this review focuses on postlingually deafened adult CI listeners, a future review on children with CIs is warranted. Research on the relationship between spectral resolution and speech recognition in children remains limited (e.g., DeFreese et al., 2024; Arjmandi et al., *in press*; Jahn et al., 2022). This is particularly important considering that children with CIs develop largely variable speech recognition abilities (e.g., Arjmandi et al., 2022b; Davidson et al., 2011; Dunn et al., 2014; Tomblin et al., 2005) and language outcomes (Niparko et al., 2010; Svirsky et al., 2004; Arjmandi et al., 2022c; Dilley et al., 2020). Future studies in children should also address challenges such as developmental and cognitive factors that may influence the link between spectral resolution ability and speech perception.

Acknowledgements

We gratefully acknowledge the support provided by startup funds from the University of South Carolina to M.A.; a Presidential Fellowship awarded to S.A.; the Magellan Scholar Grant, awarded to R.F. and K.M., from the Office of the Vice President for Research at USC; and the Honors College Research Grant, awarded to R.F. and K.M. at USC, as well as the assistance of the USC Institute for Mind and Brain in supporting this work.

References

Abbas, P. J., Hughes, M. L., Brown, C. J., Miller, C. A., & South, H. (2004). Channel interaction in cochlear implant users evaluated using the electrically evoked compound action potential. *Audiology and Neurootology*, 9(4), 203–213.

Alexander, J. M., & Kluender, K. R. (2008). Spectral tilt change in stop consonant perception. *The Journal of the Acoustical Society of America*, *123*(1), 386-396.

Apoux, F., & Healy, E. W. (2012). Use of a compound approach to derive auditory-filter-wide frequency-importance functions for vowels and consonants. *The Journal of the Acoustical Society of America*, 132(2), 1078-1087.

Arenberg Bierer, J. (2010). Probing the electrode-neuron interface with focused cochlear implant stimulation. *Trends in amplification*, 14(2), 84-95.

Arenberg, J. G., Parkinson, W. S., Litvak, L., Chen, C., Kreft, H. A., & Oxenham, A. J. (2018). A dynamically focusing cochlear implant strategy can improve vowel identification in noise. *Ear and Hearing*, 39(6), 1136-1145.

Arenberg J.G., Jahn K.N., Charles H., & Arjmandi M. K. (2022). Psychophysical Tuning Curves as a Measure of Spectral Resolution in Children and Adults with Cochlear Implants. *19th*

International Symposium on Hearing: Psychoacoustics, Physiology of Hearing, and Auditory Modelling, from the Ear to the Brain (ISH2022), Lyon, France.

Arjmandi, M. K., Jahn, K. N., Hem, C. B., & Arenberg, J. G. (in press). Relationship between psychophysical tuning curves and vowel identification in noise in children and adults with cochlear implants. *Journal of Speech, Language, and Hearing Research*.

Arjmandi, M. K., Jahn, K. N., & Arenberg, J. G. (2022a). Single-channel focused thresholds relate to vowel identification in pediatric and adult cochlear implant listeners. *Trends in Hearing*, 26, 23312165221095364.

Arjmandi, M., Dilley, L. C., & Wagner, S. E. (2018). Investigation of acoustic dimension use in dialect production: machine learning of sonorant sounds for modeling acoustic cues of African American dialect. In *11th International Conference on Voice Physiology and Biomechanics* (Vol. 12, p. 13). USA: East Lansing.

Arjmandi, M. K., Herrmann, B. S., Caswell-Midwinter, B., Doney, E. M., & Arenberg, J. G. (2022b). A modified pediatric ranked order speech perception score to assess speech recognition development in children with cochlear implants. *American Journal of Audiology*, *31*(3), 613-632.

Arjmandi, M. K., Houston, D., & Dilley, L. C. (2022c). Variability in quantity and quality of early linguistic experience in children with cochlear implants: Evidence from analysis of natural auditory environments. *Ear and hearing*, *43*(2), 685-698.

Anderson, E. S., Oxenham, A. J., Nelson, P. B., & Nelson, D. A. (2012). Assessing the role of spectral and intensity cues in spectral ripple detection and discrimination in cochlear-implant users. *The Journal of the Acoustical Society of America*, 132(6), 3925-3934.

Anderson, E. S., Nelson, D. A., Kreft, H., Nelson, P. B., & Oxenham, A. J. (2011). Comparing spatial tuning curves, spectral ripple resolution, and speech perception in cochlear implant users. *The Journal of the Acoustical Society of America*, 130(1), 364-375.

Aronoff, J. M., & Landsberger, D. M. (2013). The development of a modified spectral ripple test. *The Journal of the Acoustical Society of America*, 134(2), EL217-EL222.

Ashjaei, S., Jalilvand, H., Hadipour, H., & Sadeghi, R. (2022). Acceptable Noise Level in Children with Unilateral Cochlear Implant and Bimodal Hearing. *Auditory and Vestibular Research*.

Ashjaei, S., Behroozmand, R., Fozdar, S., Farrar, R., & Arjmandi, M. (2024). Vocal Control and Speech Production in Cochlear Implant Listeners: A Review within Auditory-Motor Processing Framework. *Hearing Research*, 109132.

Azadpour, M., & McKay, C. M. (2012). A psychophysical method for measuring spatial resolution in cochlear implants. *Journal of the Association for Research in Otolaryngology*, 13, 145-157.

Azadpour, M., McKay, C. M., & Svirsky, M. A. (2018). Effect of pulse rate on loudness discrimination in cochlear implant users. *Journal of the Association for Research in Otolaryngology*, 19, 287-299.

Baer, T., & Moore, B. C. (1993). Effects of spectral smearing on the intelligibility of sentences in noise. *The Journal of the Acoustical Society of America*, 94(3), 1229-1241.

Bálint, A., Wimmer, W., Rummel, C., Caversaccio, M., & Weder, S. (2024, July). Neural Correlates of Speech Comprehension in Normal Hearing Individuals and Cochlear Implant Users-An fNIRS Study in Quiet and Noisy Environments. In *2024 46th Annual*

International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC) (pp. 1-5). IEEE.

Bálint, A., Wimmer, W., Rummel, C., Caversaccio, M., & Weder, S. (2024). Neural Correlates of Speech Comprehension in Normal Hearing Individuals and Cochlear Implant Users-An fNIRS Study in Quiet and Noisy Environments. In *2024 46th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)* (pp. 1-5). IEEE.

Berenstein, C. K., Mens, L. H., Mulder, J. J., & Vanpoucke, F. J. (2008). Current steering and current focusing in cochlear implants: Comparison of monopolar, tripolar, and virtual channel electrode configurations. *Ear and Hearing*, 29(2), 250–260.

Başkent, D., and Shannon, R. V. (2006). "Frequency transposition around dead regions simulated with a noise band vocoder," *Journal of Acoustic Society of America*. 119(2), 1156–1163.

Beckers, L., Tromp, N., Philips, B., Mylanus, E., & Huinck, W. (2023). Exploring neurocognitive factors and brain activation in adult cochlear implant recipients associated with speech perception outcomes—A scoping review. *Frontiers in Neuroscience*, 17, 1046669.

Bierer, J. A., & Litvak, L. (2016). Reducing channel interaction through cochlear implant programming may improve speech perception: Current focusing and channel deactivation. Trends in Hearing, 20, 233121651665338. Boëx, C., de Balthasar, C., Kós, M. I., & Pelizzone, M. (2003). Electrical field interactions in different cochlear implant systems. *The Journal of the Acoustical Society of America*, 114(4), 2049-2057.

Busby, P. A., Tong, Y. C., & Clark, G. M. (1993). Electrode position, repetition rate, and speech perception by early- and late-deafened cochlear implant patients. *Journal of the Acoustical Society of America*, 93(2), 1058–1067.

Canfarotta, M. W., Dillon, M. T., Buss, E., Pillsbury, H. C., Brown, K. D., & O'Connell, B. P. (2020). Frequency-to-place mismatch: characterizing variability and the influence on speech perception outcomes in cochlear implant recipients. *Ear and hearing*, 41(5), 1349-1361.

Caswell-Midwinter, B., Doney, E. M., Arjmandi, M. K., Jahn, K. N., Herrmann, B. S., & Arenberg, J. G. (2022). The relationship between impedance, programming and word recognition in a large clinical dataset of cochlear implant recipients. *Trends in Hearing*, 26, 23312165211060983.

Caswell-Midwinter, B., & Arenberg, J. G. (2023). Comparing Fixed and Individualized Channel Interaction Coefficients for Speech Perception With Dynamic Focusing Cochlear Implant Strategies. *Trends in Hearing*, *27*, 23312165231176157.

Christensen, L. A., & Humes, L. E. (1997). Identification of multidimensional stimuli containing speech cues and the effects of training. *The Journal of the Acoustical Society of America*, 102(4), 2297-2310.

Cucis, P. A., Berger-Vachon, C., Thaï-Van, H., Hermann, R., Gallego, S., & Truy, E. (2021). Word recognition and frequency selectivity in cochlear implant simulation: Effect of channel interaction. *Journal of Clinical Medicine*, 10(4), 679.

Cychosz, M., Winn, M. B., & Goupell, M. J. (2024). How to vocode: Using channel vocoders for cochlear-implant research. *The Journal of the Acoustical Society of America*, 155(4), 2407-2437.

Davidson, L. S., Geers, A. E., Blamey, P. J., Tobey, E., & Brenner, C. (2011). Factors contributing to speech perception scores in long-term pediatric cochlear implant users. *Ear and Hearing*, 32(1), 19S–26S.

DeVries, L., & Arenberg, J. G. (2018a). Psychophysical tuning curves as a correlate of electrode position in cochlear implant listeners. *Journal of the Association for Research in Otolaryngology*, *19*, 571-587.

DeVries, L., & Arenberg, J. G. (2018b). Current focusing to reduce channel interaction for distant electrodes in cochlear implant programs. *Trends in hearing*, 22, 2331216518813811.

DeFreese, A., Camarata, S., Sunderhaus, L., Holder, J., Berg, K., Lighterink, M., & Gifford, R. (2024). The impact of spectral and temporal processing on speech recognition in children with cochlear implants. *Scientific Reports*, 14(1), 14094.

de Jong, M. A., Briaire, J. J., & Frijns, J. H. (2019). Dynamic current focusing: a novel approach to loudness coding in cochlear implants. *Ear and Hearing*, 40(1), 34-44.

Dilley, L., Lehet, M., Wieland, E. A., Arjmandi, M. K., Kondaurova, M., Wang, Y., ... & Bergeson, T. (2020). Individual differences in mothers' spontaneous infant-directed speech predict language attainment in children with cochlear implants. *Journal of Speech, Language, and Hearing Research*, *63*(7), 2453-2467.

DiNino, M., O'Brien, G., Bierer, S. M., Jahn, K. N., & Arenberg, J. G. (2019). The estimated electrode-neuron interface in cochlear implant listeners is different for early-implanted children and late-implanted adults. *Journal of the Association for Research in Otolaryngology*, 20, 291-303.

DiNino, M., Wright, R. A., Winn, M. B., & Bierer, J. A. (2016). Vowel and consonant confusions from spectrally manipulated stimuli designed to simulate poor cochlear implant electrode-neuron interfaces. *The Journal of the Acoustical Society of America*, 140(6), 4404-4418.

DiNino, M., Arenberg, J. G., Duchen, A. L., & Winn, M. B. (2020). Effects of age and cochlear implantation on spectrally cued speech categorization. Journal of Speech, Language, and *Hearing Research*, 63(7), 2425-2440.

Donaldson, G. S., Rogers, C. L., Cardenas, E. S., Russell, B. A., & Hanna, N. H. (2013). Vowel identification by cochlear implant users: Contributions of static and dynamic spectral cues. *The Journal of the Acoustical Society of America*, 134(4), 3021-3028.

Donaldson, G. S., Rogers, C. L., Johnson, L. B., & Oh, S. H. (2015). Vowel identification by cochlear implant users: Contributions of duration cues and dynamic spectral cues. *The Journal of the Acoustical Society of America*, 138(1), 65-73.

Donaldson, G. S., & Nelson, D. A. (2000). Place-pitch sensitivity and its relation to consonant recognition by cochlear implant listeners using the MPEAK and SPEAK speech processing strategies. *The Journal of the Acoustical Society of America*, 107(3), 1645-1658.

Dorman, M. F., & Wilson, B. S. (2004). The design and function of cochlear implants: fusing medicine, neural science and engineering, these devices transform human speech into an electrical code that deafened ears can understand. *American Scientist*, 92(5), 436-445.

Dorman, M. F., & Loizou, P. C. (1996). Relative spectral change and formant transitions as cues to labial and alveolar place of articulation. *The Journal of the Acoustical Society of America*, *100*(6), 3825-3830.

Drennan, W. R., Anderson, E. S., Won, J. H., & Rubinstein, J. T. (2014). Validation of a clinical assessment of spectral-ripple resolution for cochlear implant users. *Ear and hearing*, 35(3), e92-e98.

Dunn, C. C., Walker, E. A., Oleson, J., Kenworthy, M., Van Voorst, T., Tomblin, J. B., Ji, H., Kirk, K., McMurray, B., Hanson, M., & Gantz, B. J. (2014). Longitudinal speech perception

and language performance in pediatric cochlear implant users: The effect of age at implantation. *Ear and Hearing*, 35(2), 148–160

Farrar, R., Ashjaei, S., & Arjmandi, M. K. (2024). Speech-evoked cortical activities and speech recognition in adult cochlear implant listeners: a review of functional near-infrared spectroscopy studies. *Experimental Brain Research*, 1-22.

Feng, L., & Oxenham, A. J. (2018). Effects of spectral resolution on spectral contrast effects in cochlear-implant users. *The Journal of the Acoustical Society of America*, 143(6), EL468-EL473.

Finke, M., Büchner, A., Ruigendijk, E., Meyer, M., & Sandmann, P. (2016). On the relationship between auditory cognition and speech intelligibility in cochlear implant users: An ERP study. *Neuropsychologia*, 87, 169-181.

Fleming, J. T., & Winn, M. B. (2022). Strategic perceptual weighting of acoustic cues for word stress in listeners with cochlear implants, acoustic hearing, or simulated bimodal hearing. *The Journal of the Acoustical Society of America*, 152(3), 1300-1316.

Friesen, L. M., Shannon, R. V., Baskent, D., & Wang, X. (2001). Speech recognition in noise as a function of the number of spectral channels: Comparison of acoustic hearing and cochlear implants. *The Journal of the Acoustical Society of America*, 110(2), 1150–1163.

Fu, Q. J., & Nogaki, G. (2005). Noise susceptibility of cochlear implant users: the role of spectral resolution and smearing. *Journal of the Association for Research in Otolaryngology*, *6*, 19-27.

Gardner, H., Albert, M. L., & Weintraub, S. (1975). Comprehending a word: The influence of speed and redundancy on auditory comprehension in aphasia. *Cortex*, 11(2), 155-162.

Geurts, L., & Wouters, J. (2004). Better place-coding of the fundamental frequency in cochlear implants. *The Journal of the Acoustical Society of America*, 115(2), 844-852.

Gifford, R. H., Noble, J. H., Camarata, S. M., Sunderhaus, L. W., Dwyer, R. T., Dawant, B. M., ... & Labadie, R. F. (2018). The relationship between spectral modulation detection and speech recognition: Adult versus pediatric cochlear implant recipients. *Trends in hearing*, 22, 2331216518771176.

Gifford, R. H., Hedley-Williams, A., & Spahr, A. J. (2014). Clinical assessment of spectral modulation detection for adult cochlear implant recipients: A non-language based measure of performance outcomes. *International journal of audiology*, 53(3), 159-164.

Goehring, T., Archer-Boyd, A. W., Arenberg, J. G., & Carlyon, R. P. (2021). The effect of increased channel interaction on speech perception with cochlear implants. *Scientific Reports*, 11(1), 10383.

Goehring, T., Arenberg, J. G., & Carlyon, R. P. (2020). Using spectral blurring to assess effects of channel interaction on speech-in-noise perception with cochlear implants. *Journal of the Association for Research in Otolaryngology*, 21(4), 353-371.

Goldwyn, J. H., Bierer, S. M., & Bierer, J. A. (2010). Modeling the electrode–neuron interface of cochlear implants: Effects of neural survival, electrode placement, and the partial tripolar configuration. *Hearing research*, 268(1-2), 93-104.

Gonzalez, J., & Oliver, J. C. (2005). Gender and speaker identification as a function of the number of channels in spectrally reduced speech. *The Journal of the Acoustical Society of America*, 118(1), 461-470.

Goupell, M. J. (2015). Pushing the envelope of auditory research with cochlear implants. *Acoustics Today*, 11(2), 26-33.

Goycoolea, M. V., Muchow, D. C., Schirber, C. M., Goycoolea, H. G., & Schellhas, K. (1990). Anatomical perspective, approach, and experience with multichannel intracochlear implantation. *The Laryngoscope*, 100(S50), 1-18.

Green, K. M., Bhatt, Y. M., Mawman, D. J., O'Driscoll, M. P., Saeed, S. R., Ramsden, R. T., & Green, M. W. (2007). Predictors of audiological outcome following cochlear implantation in adults. *Cochlear implants international*, 8(1), 1-11.

Harmon, T. G., Dromey, C., Nelson, B., & Chapman, K. (2021). Effects of Background Noise on Speech and Language in Young Adults. *Journal of speech, language, and hearing research*, 64(4), 1104–1116.

Hedrick, M. S., & Carney, A. E. (1997). Effect of relative amplitude and formant transitions on perception of place of articulation by adult listeners with cochlear implants. *Journal of Speech, Language, and Hearing Research*, 40(6), 1445-1457.

Henry, B. A., & Turner, C. W. (2003). The resolution of complex spectral patterns by cochlear implant and normal-hearing listeners. *The Journal of the Acoustical Society of America*, 113(5), 2861-2873.

Henry, B. A., Turner, C. W., & Behrens, A. (2005). Spectral peak resolution and speech recognition in quiet: Normal hearing, hearing impaired, and cochlear implant listeners. *The Journal of the Acoustical Society of America*, 118(2), 1111-1121.

Hillenbrand, J., Getty, L. A., Clark, M. J., & Wheeler, K. (1995). Acoustic characteristics of American English vowels. *The Journal of the Acoustical Society of America*, 97(5), 3099–3111.

Holden, L. K., Finley, C. C., Firszt, J. B., Holden, T. A., Brenner, C., Potts, L. G., ... & Skinner, M. W. (2013). Factors affecting open-set word recognition in adults with cochlear implants. *Ear and hearing*, 34(3), 342-360.

Holmes, A. E., Kemker, E. J., & Merwin, G. E. (1987). The effects of varying the number of cochlear implant electrodes on speech perception. *Otology & Neurotology*, 8(3), 240-246.

Holt, L. L., & Lotto, A. J. (2006). Cue weighting in auditory categorization: Implications for first and second language acquisition. *The Journal of the Acoustical Society of America*, 119(5), 3059-3071.

Holt, L. L., & Lotto, A. J. (2010). Speech perception as categorization. *Attention, Perception, & Psychophysics*, 72(5), 1218-1227.

Horn, D. L., Dudley, D. J., Dedhia, K., Nie, K., Drennan, W. R., Won, J. H., ... & Werner, L. A. (2017). Effects of age and hearing mechanism on spectral resolution in normal hearing and cochlear-implanted listeners. *The Journal of the Acoustical Society of America*, 141(1), 613-623.

Hughes, M. L., & Stille, L. J. (2008). Psychophysical versus physiological spatial forward masking and the relation to speech perception in cochlear implants. *Ear and hearing*, 29(3), 435-452.

Jahn, K. N., & Arenberg, J. G. (2020). Electrophysiological estimates of the electrode–neuron interface differ between younger and older listeners with cochlear implants. *Ear and Hearing*, 41(4), 948-960.

Jahn, K. N., Arenberg, J. G., & Horn, D. L. (2022). Spectral resolution development in children with normal hearing and with cochlear implants: A review of behavioral studies. *Journal of Speech, Language, and Hearing Research*, 65(4), 1646-1658.

Jahn, K. N., & Arenberg, J. G. (2020). Identifying cochlear implant channels with relatively poor electrode-neuron interfaces using the electrically evoked compound action potential. *Ear and Hearing*, 41(4), 961-973.

Jeon, E. K., Turner, C. W., Karsten, S. A., Henry, B. A., & Gantz, B. J. (2015). Cochlear implant users' spectral ripple resolution. *The Journal of the Acoustical Society of America*, 138(4), 2350-2358.

Jones, G. L., Ho Won, J., Drennan, W. R., & Rubinstein, J. T. (2013). Relationship between channel interaction and spectral-ripple discrimination in cochlear implant users. *The Journal of the Acoustical Society of America*, 133(1), 425-433.

Kamakura, T., & Nadol Jr, J. B. (2016). Correlation between word recognition score and intracochlear new bone and fibrous tissue after cochlear implantation in the human. *Hearing research*, 339, 132-141.

Kwon, B. J., & van den Honert, C. (2006). Effect of electrode configuration on psychophysical forward masking in cochlear implant listeners. *The Journal of the Acoustical Society of America*, 119(5), 2994-3002.

Landsberger, D. M., Padilla, M., Martinez, A. S., & Eisenberg, L. S. (2018). Spectral-temporal modulated ripple discrimination by children with cochlear implants. *Ear and Hearing*, 39(1), 60-68.

Lawler, M., Yu, J., & Aronoff, J. M. (2017). Comparison of the spectral-temporally modulated ripple test with the Arizona Biomedical Institute Sentence Test in cochlear implant users. *Ear and hearing*, 38(6), 760-766.

Lehiste, I. (1965). Acoustical Characteristics of Selected English Consonants. *Language*, 41, 332.

Leake, P. A., & Hradek, G. T. (1988). Cochlear pathology of long term neomycin induced deafness in cats. *Hearing research*, 33(1), 11-33.

Levin, M., Balberg, M., & Zaltz, Y. (2022). Cortical activation in response to speech differs between prelingually deafened cochlear implant users with good or poor speech-in-noise understanding: an fNIRS study. *Applied Sciences*, 12(23), 12063.

Lin, Y. S., Lu, H. P., Hung, S. C., & Chang, C. P. (2009). Lexical tone identification and consonant recognition in acoustic simulations of cochlear implants. *Acta oto-laryngologica*, 129(6), 630-637.

Liepins, R., Kaider, A., Honeder, C., Auinger, A. B., Dahm, V., Riss, D., & Arnoldner, C. (2020). Formant frequency discrimination with a fine structure sound coding strategy for cochlear implants. *Hearing Research*, 392, 107970.

Litvak, L. M., Spahr, A. J., Saoji, A. A., & Fridman, G. Y. (2007). Relationship between perception of spectral ripple and speech recognition in cochlear implant and vocoder listeners. *The Journal of the Acoustical Society of America*, 122(2), 982-991.

Loizou, P. C. (1998). Mimicking the human ear. *IEEE signal processing magazine*, 15(5), 101-130.

Loizou, P. C., Dorman, M., & Tu, Z. (1999). On the number of channels needed to understand speech. *The Journal of the Acoustical Society of America*, 106(4), 2097-2103.

Loizou, P. C., & Poroy, O. (2001). Minimum spectral contrast needed for vowel identification by normal hearing and cochlear implant listeners. *The Journal of the Acoustical Society of America*, 110(3), 1619-1627.

Long, C. J., Holden, T. A., McClelland, G. H., Parkinson, W. S., Shelton, C., Kelsall, D. C., & Smith, Z. M. (2014). Examining the electro-neural interface of cochlear implant users using

psychophysics, CT scans, and speech understanding. *Journal of the Association for Research in Otolaryngology*, 15, 293-304.

Lorenzi, C., Gilbert, G., Carn, H., Garnier, S., & Moore, B. C. (2006). Speech perception problems of the hearing impaired reflect inability to use temporal fine structure. *Proceedings of the National Academy of Sciences*, 103(49), 18866-18869.

Mehta, A. H., Lu, H., & Oxenham, A. J. (2020). The Perception of Multiple Simultaneous Pitches as a Function of Number of Spectral Channels and Spectral Spread in a Noise-Excited Envelope Vocoder. *Journal of the Association for Research in Otolaryngology*, 21(1), 61–72.

Mehta, A. H., & Oxenham, A. J. (2017). Vocoder Simulations Explain Complex Pitch Perception Limitations Experienced by Cochlear Implant Users. *Journal of the Association for Research in Otolaryngology*, 18(6), 789–802.

McArdle, R., & Wilson, R. H. (2009). Speech perception in noise: The basics. Perspectives on hearing and hearing disorders: *research and diagnostics*, 13(1), 4-13.

Moberly, A. C., Lowenstein, J. H., Tarr, E., Caldwell-Tarr, A., Welling, D. B., Shahin, A. J., & Nittrouer, S. (2014). Do adults with cochlear implants rely on different acoustic cues for phoneme perception than adults with normal hearing?. *Journal of Speech, Language, and Hearing Research*, 57(2), 566-582.

Moore, B. C. (1978). Psychophysical tuning curves measured in simultaneous and forward masking. *The Journal of the Acoustical Society of America*, 63(2), 524-532.

Nelson, D. A., & Freyman, R. L. (1984). Broadened forward☐masked tuning curves from intense masking tones: Delay☐time and probe☐level manipulations. *The Journal of the Acoustical Society of America*, 75(5), 1570-1577.

Nelson, D. A., Van Tasell, D. J., Schroder, A. C., Soli, S., & Levine, S. (1995). Electrode ranking of "place pitch" and speech recognition in electrical hearing. *The Journal of the Acoustical Society of America*, 98(4), 1987-1999.

Nelson, D. A., Donaldson, G. S., & Kreft, H. (2008). Forward-masked spatial tuning curves in cochlear implant users. *The Journal of the Acoustical Society of America*, 123(3), 1522-1543.

Nelson, D. A., Kreft, H. A., Anderson, E. S., & Donaldson, G. S. (2011). Spatial tuning curves from apical, middle, and basal electrodes in cochlear implant users. *The Journal of the Acoustical Society of America*, 129(6), 3916-3933.

Nie, K., Barco, A., & Zeng, F. G. (2006). Spectral and temporal cues in cochlear implant speech perception. *Ear and hearing*, 27(2), 208-217.

Nie, K., Stickney, G., & Zeng, F. G. (2005). Encoding frequency modulation to improve cochlear implant performance in noise. *IEEE Transactions on Biomedical Engineering*, 52(1), 64–73.

Niparko, J. K., Tobey, E. A., Thal, D. J., Eisenberg, L. S., Wang, N. Y., Quittner, A. L., ... & CDaCI Investigative Team. (2010). Spoken language development in children following cochlear implantation. *Jama*, *303*(15), 1498-1506.

Noble, J. H., Gifford, R. H., Hedley-Williams, A. J., Dawant, B. M., & Labadie, R. F. (2014). Clinical evaluation of an image-guided cochlear implant programming strategy. *Audiology and Neurotology*, 19(6), 400–411.

Noble, J. H., Hedley-Williams, A. J., Sunderhaus, L., Dawant, B. M., Labadie, R. F., Camarata, S. M., &Gifford, R. H. (2016). Initial results with image-guided cochlear implant programming in children. *Otology & Neurotology*, 37(2), e63–e69.

Noble, J. H., Labadie, R. F., Gifford, R. H., & Dawant, B. M. (2013). Image-guidance enables new methods for customizing cochlear implant stimulation strategies. *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, 21(5), 820–829

O'Donoghue, G. (2013). Cochlear implants—science, serendipity, and success. *N Engl J Med*, 369(13), 1190-1193.

O'Neill, E. R., Kreft, H. A., & Oxenham, A. J. (2019). Speech perception with spectrally non-overlapping maskers as measure of spectral resolution in cochlear implant users. *Journal of the Association for Research in Otolaryngology*, 20(2), 151-167.

Otte J., Schuknecht H. F., & Kerr A. G. (1978). Ganglion cell populations in normal and pathological human, implications for cochlear implantation. *The Laryngoscope*, 88(8), 1231–1246.

Oxenham, A. J., & Kreft, H. A. (2014). Speech perception in tones and noise via cochlear implants reveals influence of spectral resolution on temporal processing. *Trends in Hearing*, 18, 2331216514553783.

Peterson, G. E., & Lehiste, I. (1962). Revised CNC lists for auditory tests. *The Journal of speech and hearing disorders*, 27, 62–70.

Peng, S. C., Chatterjee, M., & Lu, N. (2012). Acoustic cue integration in speech intonation recognition with cochlear implants. *Trends in Amplification*, 16(2), 67-82.

Price, C. N., & Moncrieff, D. (2021). Defining the Role of Attention in Hierarchical Auditory Processing. *Audiology research*, 11(1), 112–128.

Repp, B. H. (1982). Phonetic trading relations and context effects: New experimental evidence for a speech mode of perception. *Psychological bulletin*, *92*(1), 81.

Repp, B. H. (1984). The role of release bursts in the perception of [s]-stop clusters. *The Journal of the Acoustical Society of America*, *75*(4), 1219-1230.

Rosenblum, L. D. (2008). Speech perception as a multimodal phenomenon. *Current directions in psychological science*, 17(6), 405-409.

Sagi, E., & Svirsky, M. A. (2017). Contribution of formant frequency information to vowel perception in steady-state noise by cochlear implant users. *The Journal of the Acoustical Society of America*, *141*(2), 1027-1038.

Sagi, E., & Svirsky, M. A. (2024). A Level-Adjusted Cochlear Frequency-to-Place Map for Estimating Tonotopic Frequency Mismatch With a Cochlear Implant. *Ear and Hearing*, 10-1097.

Saoji, A. A., Litvak, L., Spahr, A. J., & Eddins, D. A. (2009). Spectral modulation detection and vowel and consonant identifications in cochlear implant listeners. *The Journal of the Acoustical Society of America*, 126(3), 955-958.

Scheperle, R. A., & Abbas, P. J. (2015). Peripheral and central contributions to cortical responses in cochlear implant users. *Ear and hearing*, *36*(4), 430-440.

Scheperle, R. A., & Abbas, P. J. (2015). Relationships among peripheral and central electrophysiological measures of spatial and spectral selectivity and speech perception in cochlear implant users. *Ear and Hearing*, 36(4), 441-453.

Shannon, R. V. (1990). Forward masking in patients with cochlear implants. *The Journal of the Acoustical Society of America*, 88(2), 741-744.

Shannon, R. V., Zeng, F. G., Kamath, V., Wygonski, J., & Ekelid, M. (1995). Speech recognition with primarily temporal cues. *Science*, 270(5234), 303-304.

Shannon, R. V., Galvin III, J. J., & Baskent, D. (2001). Holes in hearing. *Journal of the Association for Research in Otolaryngology*, 3(2), 185.

Shannon, R. V., Fu, Q. J., Galvin, J., & Friesen, L. (2004). Speech perception with cochlear implants. In *Cochlear implants: Auditory prostheses and electric hearing* (pp. 334-376). New York, NY: Springer New York.

Skinner, M. W., Clark, G. M., Whitford, L. A., Seligman, P. M., Staller, S. J., Shipp, D. B., ... & Beiter, A. L. (1994). Evaluation of a new spectral peak coding strategy for the Nucleus 22 channel cochlear implant system. *Otology & Neurotology*, 15, 15-27.

Skinner, M. W., Holden, L. K., Whitford, L. A., Plant, K. L., Psarros, C., & Holden, T. A. (2002). Speech recognition with the nucleus 24 SPEAK, ACE, and CIS speech coding strategies in newly implanted adults. *Ear and hearing*, 23(3), 207-223.

Singh, N. C., & Theunissen, F. E. (2003). Modulation spectra of natural sounds and ethological theories of auditory processing. *The Journal of the Acoustical Society of America*, 114(6), 3394-3411.

Srinivasan, A. G., Padilla, M., Shannon, R. V., & Landsberger, D. M. (2013). Improving speech perception in noise with current focusing in cochlear implant users. *Hearing Research*, 299, 29-36.

Steinmetzger, K., Meinhardt, B., Praetorius, M., Andermann, M., & Rupp, A. (2022). A direct comparison of voice pitch processing in acoustic and electric hearing. *NeuroImage: Clinical*, *36*, 103188.

Supin, A. Y., Popov, V. V., Milekhina, O. N., & Tarakanov, M. B. (1994). Frequency resolving power measured by rippled noise. *Hearing Research*, 78(1), 31-40.

Supin, A. Y., Popov, V. V., Milekhina, O. N., & Tarakanov, M. B. (1997). Frequency-temporal resolution of hearing measured by rippled noise. *Hearing Research*, 108(1-2), 17-27.

Supin, A. Y., Popov, V. V., Milekhina, O. N., & Tarakanov, M. B. (1999). Ripple depth and density resolution of rippled noise. *The Journal of the Acoustical Society of America*, 106(5), 2800-2804.

Svirsky, M. A., Robbins, A. M., Kirk, K. I., Pisoni, D. B., & Miyamoto, R. T. (2000). Language development in profoundly deaf children with cochlear implants. *Psychological science*, 11(2), 153-158.

Svirsky, M. (2017). Cochlear implants and electronic hearing. *Physics Today*, 70(8), 52-58.

Svirsky, M. A., Sagi, E., Meyer, T. A., Kaiser, A. R., & Teoh, S. W. (2011). A mathematical model of medial consonant identification by cochlear implant users. *The Journal of the Acoustical Society of America*, 129(4), 2191-2200.

Syrdal, A. K. (1983). Perception of consonant place of articulation. *Speech and Language*, 9, 313-349.

Tomblin, J. B., Barker, B. A., Spencer, L. J., Zhang, X., & Gantz, B. J. (2005). The effect of age at cochlear implant initial stimulation on expressive language growth in infants and toddlers. *Journal of Speech, Language, and Hearing Research*, 48(4), 853–867.

Teoh, S. W., Neuburger, H. S., & Svirsky, M. A. (2003). Acoustic and electrical pattern analysis of consonant perceptual cues used by cochlear implant users. *Audiology and Neurotology*, 8(5), 269-285.

Ter Keurs, M., Festen, J. M., & Plomp, R. (1993). Effect of spectral envelope smearing on speech reception. II. *The Journal of the Acoustical Society of America*, 93(3), 1547-1552.

Throckmorton, C. S., & Collins, L. M. (2002). The effect of channel interactions on speech recognition in cochlear implant subjects: Predictions from an acoustic model. *The Journal of the Acoustical Society of America*, 112(1), 285–296.

Townshend, B., Cotter, N., Van Compernolle, D., & White, R. L. (1987). Pitch perception by cochlear implant subjects. *The Journal of the Acoustical Society of America*, 82(1), 106-115.

Toscano, J. C., & McMurray, B. (2010). Cue integration with categories: Weighting acoustic cues in speech using unsupervised learning and distributional statistics. *Cognitive science,* 34(3), 434-464.

Wanna, G. B., Noble, J. H., Carlson, M. L., Gifford, R. H., Dietrich, M. S., Haynes, D. S., ... & Labadie, R. F. (2014). Impact of electrode design and surgical approach on scalar location and cochlear implant outcomes. *The Laryngoscope*, 124(S6), S1-S7.

Wilk, M., Hessler, R., Mugridge, K., Jolly, C., Fehr, M., Lenarz, T., & Scheper, V. (2016). Impedance changes and fibrous tissue growth after cochlear implantation are correlated and can be reduced using a dexamethasone eluting electrode. *PloS one*, 11(2), e0147552.

Winn, M. B., & Litovsky, R. Y. (2015). Using speech sounds to test functional spectral resolution in listeners with cochlear implants. *The Journal of the Acoustical Society of America*, 137(3), 1430-1442.

Winn, M. B., Won, J. H., & Moon, I. J. (2016). Assessment of spectral and temporal resolution in cochlear implant users using psychoacoustic discrimination and speech cue categorization. *Ear and hearing*, 37(6), e377-e390.

Winn, M.B. (2020). Accommodation of gender-related phonetic differences by listeners with cochlear implants and in a variety of vocoder simulations. *The Journal of the Acoustical Society of America,* 147 1, 174 .

Winn, M. B., & O'Brien, G. (2022). Distortion of spectral ripples through cochlear implants has major implications for interpreting performance scores. *Ear and hearing*, 43(3), 764-772.

Winn, M. B., Chatterjee, M., & Idsardi, W. J. (2012). The use of acoustic cues for phonetic identification: effects of spectral degradation and electric hearing. *The Journal of the Acoustical Society of America*, 131(2), 1465–1479.

White, M. W., Merzenich, M. M., & Gardi, J. N. (1984). Multichannel cochlear implants: Channel interactions and processor design. *Archives of Otolaryngology*, 110(8), 493-501.

Won, J. H., Drennan, W. R., & Rubinstein, J. T. (2007). Spectral-ripple resolution correlates with speech reception in noise in cochlear implant users. *Journal of the Association for Research in Otolaryngology*, 8(3), 384–392.

Won, J. H., Moon, I. J., Jin, S., Park, H., Woo, J., Cho, Y. S., ... & Hong, S. H. (2015). Spectrotemporal modulation detection and speech perception by cochlear implant users. *PloS one*, 10(10), e0140920.

Wouters, J., McDermott, H. J., & Francart, T. (2015). Sound coding in cochlear implants: From electric pulses to hearing. *IEEE Signal Processing Magazine*, 32(2), 67-80.

Xu, K., Willis, S., Gopen, Q., & Fu, Q. J. (2020). Effects of spectral resolution and frequency mismatch on speech understanding and spatial release from masking in simulated bilateral cochlear implants. *Ear and hearing*, 41(5), 1362-1371.

Xu, L., Thompson, C. S., & Pfingst, B. E. (2005). Relative contributions of spectral and temporal cues for phoneme recognition. *The Journal of the Acoustical Society of America*, 117(5), 3255-3267.

Zeng, F. G. (2004). Trends in cochlear implants. Trends in amplification, 8(1), 1-34.

Zeng, F. G., Rebscher, S., Harrison, W., Sun, X., & Feng, H. (2008). Cochlear Implants: System Design, Integration, and Evaluation. *IEEE Reviews in Biomedical Engineering*, 1, 115–142.

Zhou, N., Xu, L., & Lee, C. Y. (2010). The effects of frequency-place shift on consonant confusion in cochlear implant simulations. *The Journal of the Acoustical Society of America*, 128(1), 401-409.

Zhu, Z., Tang, Q., Zeng, F. G., Guan, T., & Ye, D. (2012). Cochlear-implant spatial selectivity with monopolar, bipolar and tripolar stimulation. *Hearing research*, 283(1-2), 45-58.

Zwolan, T. A., Collins, L. M., & Wakefield, G. H. (1997). Electrode discrimination and speech recognition in postlingually deafened adult cochlear implant subjects. *The Journal of the Acoustical Society of America*, 102(6), 3673-3685.