Title: Evaluating the N1-P2 interpeak latency of the eCAP and its inter-

trial variability as potential indicators of neural synchrony in the

cochlear nerve of cochlear implant users

Authors: Shuman He[1], MD, PhD; Ian C. Bruce[2], PhD; Zi Gao[1], PhD; Ross A.

Aiello[3], BS; Christopher R. Mueller[1], BS

Affiliations: [1]Department of Otolaryngology – Head and Neck Surgery, The

Ohio State University, 915 Olentangy River Road, Columbus, OH

43212

[2]Department of Electrical and Computer Engineering, McMaster

University, Hamilton, ON, L8S 4K1, Canada

[3]Department Mechanic Engineering & Material Science,

Washington University in St. Louis, St. Louis, MO 63105

Correspondence: Shuman He, MD, PhD

Eye and Ear Institute

Department of Otolaryngology – Head and Neck Surgery

The Ohio State University

915 Olentangy River Road, Suite 4000

Phone: 614-293-5963

Fax: 614-293-7292

Email: Shuman.He@osumc.edu

Source of Funding: This work was partially supported by grants from the National

Institutes of Health awarded to SH [grant numbers: 1

R01DC016038, 1R01 DC017846, R21 DC019458 and

R01DC022051-01A1]. Ian C. Bruce was supported by NSERC

Discovery Grant #RGPIN-2024-05888. The funding organizations

played no role in designing or conducting the study, collecting or

analyzing the data, interpreting the results, preparing the

manuscript, reviewing or approving its content, or deciding to

submit it for publication.

ABSTRACT

Objective: This study evaluated interpeak latency (IPL) and its inter-trial variability (VIL)

of the electrically evoked compound action potential (eCAP) as potential alternatives to

the phase-locking value (PLV) for quantifying cochlear nerve (CN) synchrony in

cochlear implant (CI) users.

Design: The IPL was assessed in postlingually deafened adults and three pediatric

populations: children with auditory neuropathy spectrum disorder, cochlear nerve

deficiency, and typical sensorineural hearing loss. VIL was evaluated only in adults.

Their associations with temporal resolution and speech perception outcomes were

evaluated. Frequency analysis was conducted to understand the impacts of eCAP

recording noise on IPL, VIL, and PLV. Simulations of inter-trial jitter in the eCAP were

performed to quantify how the IPL, VIL, and PLV metrics varied with increased temporal

jitter.

Results: eCAP traces recorded in all patient groups showed a multi-peak issue

affecting the accuracy of IPL and VIL assessments. Temporal resolution and speech

perception outcomes were significantly correlated with VIL but not with IPL metrics. The

PLV was impacted less by recording noise than either the IPL or the VIL. Simulation

results revealed that the IPL was less sensitive to the amount of inter-trial jitter in the

eCAP than were the VIL and the PLV.

Conclusions: The IPL is not a reliable indicator of CN synchrony. The VIL is indicative

of neural synchrony in the CN but is affected more by the eCAP recording noise than

the PLV. The PLV is therefore the preferred measure for quantifying neural synchrony in

the CN in CI users.

Key Words: electrical stimulation, cochlear nerve, neural synchrony, interpeak latency

Statements and Declarations

Conflict of Interest: None.

IRB information: The data reported in this study were collected for the projects that

were approved by the biomedical Institutional Review Board (IRB)

of The Ohio State University (IRB study #: 2017H0131, 2018H0344

and 2018N0005; PI: Shuman He), and the IRB of the University of

North Carolina at Chapel Hill (IRB study #: 12–1737; PI: Shuman

He).

Author Contributions: SH designed this study, participated in data analysis, drafted

and approved the final version of this paper. ICB participated in

study design and data analysis, conducted computational modeling

work, drafted and approved the final version of this paper. ZG

participated in data analysis, provided critical comments, and

approved the final version of this paper. RAA participated in data

analysis and approved the final version of this paper. CRM

participated in data collection and approved the final version of this

paper.

Data Availability Statement: The data that support the findings of this study are

available from the authors upon reasonable request with

permissions from The Ohio State University and the University of

North Carolina at Chapel Hill.

INTRODUCTION

The ability of human listeners to detect, discriminate and recognize complex

auditory signals depends on temporally organized neural activity in the auditory system

that is time-locked to the auditory input (i.e., neural synchrony) [e.g., [1-9]]. For cochlear

implant (CI) users, the cochlear nerve (CN) is the first neural structure to receive

electrical stimulation from the device. Consequently, the fidelity with which the CN

encodes and processes this stimulation critically influences both the quality and quantity

of auditory information conveyed to higher-order auditory centers. In animal models, the

discharge patterns of CN fibers have been shown to play a key role in encoding

temporal cues [10-12]. Consistent with these findings from single-fiber recordings,

simulation studies using computational models have demonstrated that poor neural

synchrony in the CN results in a degraded neural representation of temporal envelope

cues, leading to deficits in the perception of these cues [1,2]. Given the limited

availability of temporal fine structure cues and the reduced spectral resolution provided

by current CIs [13,14], CI users primarily depend on temporal envelope cues for speech

understanding [e.g., [15,16]]. Consequently, neural synchrony in the CN is expected to

play a critical role in determining speech perception outcomes in this population.

However, the lack of non-invasive methods for assessing CN neural synchrony has

historically hindered its investigation in CI users.

Recently, He et al. [17] reported the use of the phase-locking value (PLV) to

assess neural synchrony in the CN in CI patients. In this method, the PLV is calculated

based on trial-to-trial phase coherence in the summated activity of CN fibers evoked by

electrical stimulation (i.e., the electrically evoked compound action potential [eCAP]).

Their findings demonstrated the feasibility of applying this method in both adult and

pediatric CI populations [17-19], including children with auditory neuropathy spectrum

disorder (ANSD). Moreover, consistent with the literature on acoustic hearing [1,2], PLV

was found to correlate with temporal resolution (He et al. 2024b) and speech perception

in noise [17,18] in postlingually deafened adult CI users. These results suggest that PLV

holds strong promise as an objective indicator of perceptual outcomes in CI users.

However, as noted by He et al. [17], this method is affected by the sampling rate of CI

devices, which varies substantially across CI manufacturers. The extent to which this

variability affects PLV measurements remains unclear, posing a challenge for studies

involving participants with different CI systems.

As part of the effort to develop a noninvasive method for assessing neural

synchrony in the CN in human CI users , we evaluated, in addition to the PLV, the inter-

trial variability of the eCAP interpeak latency (VIL) as a potential alternative indicator of

neural synchrony [20]. More recently, Schvartz-Leyzac et al. [21] proposed that the

interpeak latency (IPL) of the average eCAP waveform could potentially serve as a

marker of neural synchrony. Both indicators are based on the morphological

characteristics of the eCAP. As a result, their computation does not depend on

manufacturer-specific parameters. Therefore, if the validity of the IPL and the VIL as

indicators of CN neural synchrony is established, they can effectively address a key

limitation of the PLV method.

In the present study, the IPL was defined as the difference between the N1 and P2

latencies of the eCAP waveform for a given electrode in each participant. The inter-trial

variability of the IPL (VIL) was defined as the standard deviation of the IPL across 400

eCAP sweeps (i.e., trials) measured at the same electrode in the same participant. The

selection of these two metrics was guided by findings in the acoustic hearing literature,

which demonstrate that decreased neural synchrony is associated with delayed

response peaks and increased inter-trial variability [22-26]. For healthy spiral ganglion

neurons (SGNs), spikes elicited by direct electrical stimulation at the single-unit level

exhibit considerably less temporal jitter compared to those evoked by synaptic inputs

from hair cells (Miller et al. 1999; Miller et al. 2003; Miller et al. 2001). However, axonal

dystrophy and demyelination of SGNs – a common issue observed in CI users [e.g.,

[27-33]] – alter many neural properties [e.g., [34,35]]. These alternations lead to

complex changes in spike timing (Resnick et al. 2018), action potential initiation sites

[e.g., [36-39]], and slowed action potential conduction. As a result, spike

synchronizations within individual CN fibers across repeated stimulations as well as

among different CN fibers are reduced, potentially resulting in delayed response peaks

and increased inter-trial variability of neural responses during electrical hearing. This

scientific basis supports the use of the VIL to quantify temporal jitter of individual CN

fibers, as measured using single-fiber recordings in animal models, where smaller VIL

values indicate less temporal jitter and thus better neural synchrony [40]. However, as

shown by simulation results reported in Resnick et al.[39], the effects of demyelination

on fiber sensitivity and spike timing at the single-unit level are complex and highly

heterogeneous. Furthermore, these effects strongly depend on the degree of

demyelination, which has been shown to vary substantially along the cochlea, even

within individual CI patients [33]. As a near-field recorded compound response in CI

patients, the eCAP is not only affected by neural factors (e.g., neural synchrony,

discharge rate of individual CN fibers, neural survival of the CN) but also affected by

non-neural factors (e.g., impedance, electrode-to-modiolus distance). This distinction is

important compared to the far-field recorded compound action potential in listeners with

acoustic hearing. Therefore, rather than directly extrapolating findings from acoustic

hearing to CI patients, it is essential to empirically evaluate whether reduced spike

synchronization observed at the single-unit level translates into measurable changes in

IPL and VIL in human CI users. This evaluation is a critical, scientifically rigorous

prerequisite before applying these metrics in clinical and research contexts involving CI

populations.

Evaluations were conducted by 1) comparing results across four patient

populations with different pathophysiological conditions affecting the CN, and 2)

assessing the associations between the IPL, the VIL, temporal resolution, and speech

perception outcomes in CI users. The rationale for this approach is that a robust

indicator needs to be sensitive to the pathophysiological condition it probes and predicts

the resulting auditory perception outcomes. The first patient population consisted of

children with ANSD, a condition characterized by poor neural synchrony that cannot be

fully compensated for by the enhanced neural synchronization provided by electrical

stimulation [1,41]. The second patient population included children with cochlear nerve

deficiency (CND), a group with substantially fewer but otherwise healthy SGNs

compared to children with idiopathic sensorineural hearing loss (SNHL) and normal-

sized CNs [42,43]. The third patient population was children with typical SNHL. The

fourth patient population was postlingually deafened adult CI users, who exhibit greater

loss of SGN somas and peripheral processes at the cochlear base relative to the apex

[32,44].

Data from all four patient populations were analyzed to evaluate the IPL. Given the

robust correlation between the VIL and the PLV revealed by our pilot results [20], VIL

analyses were restricted to the adult CI user dataset. Additionally, relationships between

IPL and age, as well as between IPL, VIL, and PLV, were examined in the adult cohort.

A frequency analysis of eCAP recording noise was performed to determine how the

noise spectrum impacts the frequency-domain PLV metric compared to the time-domain

IP and VIL measures. Finally, computational modeling was employed to understand the

effect of the amount of temporal jitter on IPL, VIL, and PLV metrics. Based on the

literature reviewed previously, shorter IPLs, smaller VILs, and larger PLVs were

expected to be indicative of stronger neural synchrony in the CN.

We hypothesized that reduced spike synchronization within and across CN fibers

would lead to greater inter-trial variability of the eCAP, which can be quantified using the

VIL. This study also tested the hypothesis proposed by Schvartz-Leyzac et al. [21]

regarding the effect of reduced neural synchrony on the IPL. Based on these

hypotheses, children with typical SNHL were expected to show shorter IPLs than

children with ANSD and adult CI users. Among adult CI users, longer IPLs and larger

VILs were predicted at basal electrode locations compared to more apical sites.

Furthermore, longer IPLs and larger VILs were anticipated to correlate with poorer

temporal resolution and lower speech perception scores in noise. Finally, strong

correlations were expected among IPL, VIL, and PLV metrics measured within the same

patient cohort.

MATERIALS AND METHODS

Datasets

These evaluations utilized three existing datasets collected across five studies

[17,18,41,45,46], along with one new dataset obtained from a subgroup of 14 adult CI

users (A003L, A008R, A009L, A020R, A024R, A035L, A052L, A061R, A070R, A075R,

A077L, A080L, A091L, and A093R) who also participated in the study reported in Gao

et al. [18]. For the first dataset, the eCAP was recorded as the averaged response

across 50 sweeps using a system gain of 50 dB, which are the default parameters used

in Custom Sound EP software (Cochlear Ltd, Macquarie, NSW, Australia). In the

second dataset, the eCAP was recorded as the averaged response across 100 sweeps

with a system gain of 40 dB. It has been shown that the number of averaged sweeps

needs to be doubled (i.e., 100 sweeps) when the system gain is set at 40 dB to

preserve the required signal quality for eCAP recordings [47]. For the third and fourth

datasets, the recorded eCAP response was 400 single sweeps with a default system

gain of 50 dB.

The first dataset included eCAPs evoked by single-pulse stimulation measured at

three electrode locations across the electrode array (default electrodes, listed from

basal to apical locations: electrodes 3, 12, and 19) measured in each of the 30 ears

tested in 24 children with ANSD and in each of the 29 ears in 26 children with typical

SNHL. These patients participated in the research study comparing temporal response

properties of the CN that has been reported in He et al. [45]. The results measured in

the right ear of one child with SNHL (S1) were not included due to data file corruption. A

subgroup of 10 children with ANSD also participated in a research study assessing the

association between objective within-channel gap detection threshold (GDT) measured

at electrode 12 and the Phonetically Balanced Kindergarten (PBK) word score (He et al.

2013).

229   The second dataset included eCAPs measured at the maximum comfortable

230   level using biphasic pulses with an interphase gap (IPG) of 7 µs at one basal, one

231   middle and one apical electrode locations in 30 children with CND. For each participant,

232   only one ear was tested. These children participated in a study evaluating the effect of

233   increasing the IPG of biphasic pulses on eCAPs [46].

234   The third dataset included eCAP results and AzBio sentence scores measured in

quiet and in a 10-talker babble noise with signal-to-noise ratios (SNRs) of +10 and +5

dB in 28 postlingually deafened adult CI users. For each participant, 400 eCAP sweeps

were recorded at the maximum comfortable level using biphasic pulses with an IPG of 7

µs at four electrode locations across the electrode array (default electrodes, listed from

basal to apical locations: electrodes 3, 9, 15, and 21). The IPL was calculated based on

the average responses of the 400 sweeps, and the VIL was calculated based on the

latencies of individual sweeps, as detailed in the next section. Psychophysical within-

channel GDTs were measured at two electrode locations in each of a subgroup of 15

adult participants. These adult patients participated in two of our previous studies

[17,18].

245   The fourth (i.e., the new) dataset included eCAPs measured at 0 clinical units

with an extended recording window to include 64 samples (~3,200 µs) in a subgroup of

14 adult CI users. The default recording window implemented in Custom Sound EP for

measuring the eCAP includes 32 samples (~ 1600 µs). All other eCAP stimulating and

recording parameters, as well as the electrode locations tested were the same as those

used in the third dataset for the same group of patients. These extended-window data

were utilized exclusively in a frequency analysis to assess the likely impact of eCAP

recording noise on IPL, VIL and PLV.

All study participants used a Cochlear™ Nucleus® device (Cochlear Ltd,

Macquarie, NSW, Australia) in the test ear. Demographic information for all participants,

the electrode locations tested in these experiments, as well as detailed procedures for

measuring the eCAP, the PLV, GDT, and speech perception outcomes have been

reported in our studies cited above. Therefore, these pieces of information were not

included here to avoid redundancy.

Calculation of the IPL and the VIL

As illustrated in panels (a) □ (d) of Figure 1 and panel (a) of Figure 6, both

single-sweeps and averaged eCAP waveforms frequently exhibited multiple sampled

points within the designated latency windows for the N1 and P2 peaks that had nearly

identical minimum or maximum voltage values. The time windows containing these

points are highlighted in the bolded black rectangles. Closer inspection revealed that the

voltage differences among these points were less than 0.5 μV, substantially smaller

than the estimated recording noise for Cochlear™ Nucleus® devices, which is

approximately 2–5 μV [48-50]. However, the latency difference between these sampled

points was 97.6 μs (i.e., two samples) or greater. Consequently, this "multi-peak" issue

introduced variability in the calculation of IPL and its VIL, depending on which sampled

point was selected to represent the peak latency.

<u>Insert Figure 1 about here</u>

Traditionally, N1 and P2 latencies have been determined by identifying the

minimum and maximum voltage values within designated latency windows or sample

points [51]. Using this approach, single N1 and P2 peaks are identified even when the

voltage differences between neighboring samples are as small as 0.0001 μV (i.e., the

smallest measurable voltage difference using the setup described in the current study),

despite the fact that these small voltage differences are well below the noise floor of CI

devices and therefore not clinically meaningful. As a result, the presence of "multi-peak"

issues in eCAP waveforms and their impact on latency estimation have not been well

documented or systematically addressed. To mitigate this issue, we applied a peak-

picking method that took this issue into consideration to a certain extent. Specifically, N1

and P2 latencies were identified within defined latency windows (150–600 μs for N1;

400–1000 μs for P2) by first rounding all voltage values to the nearest integer (in μV)

prior to peak selection. This rounding introduced an average voltage resolution of ~0.5

μV and ensured that the extreme values within a peak latency window differing by less

than 1 μV would be considered equivalent. This threshold is more conservative than the

lowest noise floor reported for Cochlear™ Nucleus® eCAP recordings (~ 2 μV) [48] and

therefore served as a conservative estimation for the impact of the multi-peak issue on

study results. The IPL was defined as the difference between these two peak latencies

(i.e., P2 minus N1). For traces with the multi-peak issue, the shortest IPL (IPL_{short}) and

longest IPL (IPL_{long}) were measured. The IPL_{short} was calculated as the difference

between the shortest P2 latency and the longest N1 latency, and the IPL_{long} was

calculated as the difference between the longest P2 latency and the shortest N1

latency. For traces without the multi-peak issue, the same IPL value was used as the

IPL_{short} and IPL_{long} in the subsequent analysis. The variabilities of the IPL_{short} (VIL_{short}),

the IPL_{long} (VIL_{long}), as well as the IPL averaged across all possible N1-P2 pairs (VIL_{all}),

were defined as the standard deviation of the corresponding IPL values of the 400

eCAP single sweeps recorded at each electrode location in each adult participant.

To compare IPLs across patient groups, IPLs were calculated for eCAPs

recorded in three pediatric patient groups and the averaged eCAP across the first 50

sweeps measured in adult CI users. The IPL of the averaged eCAP response across all

400 single-sweep eCAPs was used to evaluate associations with psychophysical GDT,

AzBio sentence scores measured in different listening conditions, the PLV, and age in

adult CI users. Separate analyses were performed for both the IPL_{short} and IPL_{long}, as

defined in the previous section. To calculate the VIL, the IPL for each of the 400 single-

sweep eCAPs recorded in adult CI users was calculated, allowing for evaluation of

variability across trials.

Statistical Analyses

All statistical analyses were conducted separately for the IPL_{short} and IPL_{long}.

Linear mixed-effects models (LMMs) with patient group and electrode location as fixed

effects were used to compare IPLs at three electrode locations between children with

ANSD and children with typical SNHL, as well as to compare IPLs at one basal and one

apical electrode location across all four patient groups included in this study. LMMs with

electrode location as fixed effect were used to compare IPLs and VIL_{all} values across

electrode locations for adult CI users. All LMMs used a correlated regression model with

an unstructured correlation matrix to account for repeated observations per participant.

The two ears of bilateral users, if both tested, were treated as separate participants.

Estimations were obtained using restricted maximum likelihood with Satterthwaite

degrees of freedom. Bonferroni correction was used to adjust for multiple comparisons.

Spearman rank correlation tests were used to assess the associations between

the IPL and objective GDT and between the IPL and PBK word scores measured in 10

children with ANSD, as well as between PLVs and IPLs in adult CI users. A Pearson

correlation test was used to assess the associations between the VIL$_{all}$ and the PLV. To

better compare our results with those reported in Schvartz-Leyzac et al. [5], a Pearson

correlation test was also used to evaluate the association between participant age and

the IPL at different electrode locations in adult participants. For these tests, Bonferroni

correction was used to adjust for multiple comparisons as needed.

Multiple linear regressions (MLRs) with AzBio score as the outcome and either

the IPL or the VIL$_{all}$ as the predictor were performed to evaluate the association

between speech perception outcomes and each of these two parameters under each

testing condition (quiet, +10 dB, and +5 dB SNR). The regression model included

participant's age as a covariate to control its potential effects on speech perception

and/or eCAP measurements.

The MLRs were performed in R v4.4.1 (R Core Team 2024). All other statistical

analyses for this study were performed using SPSS Statistics (v. 28.0.0) (IBM, Armonk,

NY). Statistical significance was determined at the 95% confidence level (i.e., $p < .05$).

Recording Noise Analysis

For each electrode tested in each participant in the fourth dataset (i.e., the 0

clinical unit recordings), the single-sweep eCAP recording noise frequency spectrum

averaged across 400 trials was calculated using the fast Fourier transform. The

percentages of recording noise power within and below the frequency range of the PLV

analysis (0.789 to 4.73 kHz) were computed.

Simulation of the Effects of Inter-trial eCAP Jitter on the IPL, VIL, and PLV.

Figure 2 of Schvartz-Leyzac et al. [21] illustrates how inter-trial jitter in the single-

sweep eCAP waveform can lead to smearing out of the average eCAP waveform and a

lengthening of the IPL of the average waveform. However, they only considered two

extreme cases in their schematic illustration to justify the use of the IPL as a measure of

neural synchrony. Specifically, the case shown in their Figure 2A has a very small

amount of jitter such that the average waveform is very similar to the template single-

sweep waveform. In comparison, the second case, shown in their Figure 2B has a much

larger amount of jitter such that the N1 peaks for some sweeps overlap with the P2

peaks on other sweeps, which leads to the smearing of the average eCAP waveform. It

is unclear how much of this smearing of the average eCAP waveform occurs in cases

with intermediate amounts of jitter and thus how sensitive the IPL of the average

waveform is to various degrees of neural synchrony. To systematically investigate this

issue, we have extended and modified the illustration of Schvartz-Leyzac et al. [21] in

the following ways.

First, Schvartz-Leyzac and colleagues implemented a horizontal shift in their jitter

simulation, but this leads to an identical single-sweep IPL for each trial. In our

simulations we have implemented the jitter as a horizontal linear scaling of the template

eCAP waveform, such that the single-sweep IPL varies directly as a function of the

scaling factor. This scaling-based approach to applying temporal jitter also leads to the

P2 latency having a larger standard deviation than then N1 latency, consistent with the

results of our latency analysis for the third dataset.

Second, we analyzed the IPL distributions for this dataset and found that they

were generally well fit by a gamma distribution. Panel (a) of Figure 2 shows the

histogram of single-sweep IPLs for an example electrode, with the fitted gamma

distribution probability density function (pdf) shown by the red curve. We used the

function `fitdist` from the Matlab Statistics and Machine Learning Toolbox to fit a

gamma distribution pdf to the single-sweep IPL data for each electrode from each

participant in the third dataset. Outliers were excluded using the Matlab function

`isoutlier`. Panels (b) and (c) of Figure 2 show the resulting histograms of the gamma

distribution parameters *a* and *b*, respectively. The median values of the gamma

distribution parameters were found to be $a = 18.6$ and $b = 20.4$.

Third, we designated the gamma distribution pdf with these median parameter

values to be our baseline single-sweep IPL distribution for our simulations. Alternative

gamma distributions with different amounts of jitter were then created by scaling the

baseline pdf around its mean by a *jitter factor*, such that a jitter factor of 0.5 produces a

single-sweep IPL standard deviation half that of the baseline pdf, while a jitter factor of 2

produces a standard deviation twice that of the baseline distribution. Jitter factors were

tested from 0 to 2 in steps of 0.1, and ten simulations were run at each value of the jitter

factor. Panel (d) of Figure 2 shows pdfs of a subset of the gamma distributions that

were used in the inter-trial jitter simulations. For each run, 400 stimulus repetitions

(trials) were implemented by generating a random number drawn from the appropriate

gamma distribution using the function `gamrnd` from the Matlab Statistics and Machine

Learning Toolbox, and applying the jittered linear scaling of the template eCAP

waveform via the Matlab function `interp1`.

Fourth, we created a template eCAP waveform for the simulations based on the

average eCAP waveform from electrode 21 of participant A008R, as this electrode had

the highest PLV value from the third dataset and a high SNR — see Fig. 2 of He et al

[17]. A smoothed version of this waveform was generated by upsampling from 20.492 to

500 kHz using piecewise cubic spline interpolation via the Matlab function `interp1`. A

linear horizontal scaling was then applied via `interp1` so that the waveform's IPL

matched the mean IPL from the gamma distribution pdfs shown in panel (d) of Figure 2.

The resulting template eCAP waveform is plotted in panel (e) of Figure 2. In the

simulations, jittered single-sweep waveforms were excluded as outliers if the N1 peak

did not occur within the analysis window, consistent with the methodology used for the

data analysis.

399   <u>Insert Figure 2 about here</u>

RESULTS

Evaluations of the IPL

Using a voltage rounding threshold of 1 µV, the multi-peak issue was observed in

22 out of 86 (25.58%), 15 out of 82 (18.29%), 15 out of 93 (16.13%), and 39 out of 107

(36.44%) eCAPs recorded in children with ANSD, children with SNHL, children with

CND and adult CI users, respectively. The difference between the IPL_{short} and IPL_{long}

ranged between 48.80 µs (one sample) and 146.40 µs (three samples) (mean: 86.50

µs, SD: 33.44 µs), between 48.80 µs (one sample) and 341.60 µs (seven samples)

(mean: 91.50 µs, SD: 79.44 µs), between 48.8 µs (one sample) and 195.2 µs (four

samples) (mean: 103.70 µs, SD: 58.76 µs), and between 48.8 µs (one sample) and

146.4 µs (three samples) (man: 58.56 µs, SD: 25.20 µs) for children with ANSD,

children with SNHL, children with CND, and adult CI users, respectively. These findings

demonstrate that the multi-peak issue introduces variability in IPL measurements across

patient groups and can span a range of up to seven samples (341.60 µs). Detailed

summary statistics (range, mean, and standard deviation) the IPL_{short} and IPL_{long}

measured at different electrode locations in different patient groups are included in

Table 1.

Table 1. Ranges (means, *SD*s) of the shortest and the longest interpeak latencies measured at different electrode locations in four patient groups. ANSD, auditory neuropathy spectrum disorder; SNHL, sensorineural hearing loss; CND, cochlear nerve deficiency; IPL, interpeak latency; IPL_{short}, the shortest IPL; IPL_{long}, the longest IPL.

Dataset	Patient Group	Electrode Location	IPL_{short} (µs)	IPL_{long} (µs)
He et al. (2016)	Children with ANSD	3	244.00 – 585.60 (355.54, 83.85)	244.40 – 634.40 (372.97, 98.69)
		12	195.20 – 585.60 (341.60, 98.47)	195.20 – 585.60 (370.21, 98.05)
		19	195.20 – 585.60 (393.77, 120.20)	195.20 – 634.40 (413.96, 118.55)
	Children with SNHL	3	244.00 – 585.60 (368.71, 98.84)	244.00 – 585.60 (377.75, 97.22)
		12	195.20 – 488.00 (345.21, 83.35)	244.00 – 585.60 (374.13, 91.80)
		19	244.40 – 780.80 (416.54, 129.78)	244.00 – 780.80 (432.23, 199.95)
He et al. (2020)	Children with CND	Basal	195.20 – 488.00 (305.39, 73.44)	244.00 – 488.00 (325.86, 78.07)
		Middle	195.20 – 488.00 (313.25, 74.30)	195.20 – 536.80 (321.14, 85.25)
		Apical	97.60 – 536.80 (312.32, 106.92)	195.20 – 634.40 (338.35, 103.27)
Gao et al. (2024)	Adult	3	195.20 – 439.20 (304.06, 57.42)	244.00 – 488.00 (332.22, 51.68)
		9	146.40 – 683.20 (339.65, 99.09)	195.20 – 683.20 (359.17, 94.37)
		15	195.20 – 439.20 (343.34, 64.36)	195.20 – 439.20 (367.74, 73.33)
		21	244.00 – 585.60 (371.23, 92.34)	244.00 – 585.60 (386.91, 93.85)

Effects of participant group and electrode location

The results of the IPL_{short} and IPL_{long} measured at three electrode locations in children with ANSD and children with typical SNHL are shown in panels (a) and (b) of Figure 3, respectively. For both sets of results, LMMs revealed no significant effect of patient group (IPL_{short}: $F_{(1, 54.73)} = 0.34$, $p = .561$; IPL_{long}: $F_{(1, 54.94)} = 0.12$, $p = .733$), or the

interaction between participant group and electrode location (IPL_{short}: $F_{(2, 55.17)} = 0.18$, p

$= .839$; IPL_{long}: $F_{(2, 54.52)} = 0.17$, $p = .848$) on the IPL. However, there was a significant

effect of electrode location on the IPL (IPL_{short}: $F_{(2, 55.17)} = 6.88$, $p = .002$; IPL_{long}: $F_{(2, 54.50)}$

$= 6.06$, $p = .004$). Post-hoc pairwise comparisons showed that IPLs measured at

electrode 19 were significantly shorter than those measured at both electrode 3 (IPL_{short}:

$p = .032$; IPL_{long}: $p = .015$) and electrode 12 (IPL_{short}: $p = .002$; IPL_{long}: $p = .006$) across

both patient groups. There was no significant difference between IPLs measured at

electrodes 3 and 12 (IPL_{short}: $p = .628$; IPL_{long}: $p = 1.000$).

435   <u>Insert Figure 3 about here</u>

Panels (c) and (d) of Figure 3 show the results of the IPL_{short} and IPL_{long} measured

at four electrode locations in adult participants. Overall, the data reveal an increasing

trend in IPL as electrode location moves toward more apical sites for both measures.

LMMs confirmed this visual observation, showing a significant effect of electrode

location on both the IPL_{short} ($F_{(3, 25.34)} = 5.98$, $p = .003$) and IPL_{long} ($F_{(3, 24.80)} = 3.65$, $p =$

$.026$). Post-hoc paired comparisons indicated that IPL_{short} values measured at electrode

3 were significantly shorter than those measured at electrodes 15 ($p = .023$) and 21 ($p =$

$.002$). IPL_{long} values measured at electrode 3 were also significantly shorter than those

measured at electrode 21 ($p = .020$). No other significant differences between electrode

pairs were observed for either measure.

To better understand the significant electrode location effect on the IPL observed

in adult participants, we compared IPL_{short} and IPL_{long} values measured at two electrode

locations across four patient groups. These results are shown in panels (e) and (f) of

Figure 3. For both the IPL_{short} and IPL_{long}, LMMs revelated significant effects of patient

group (IPL$_{short}$: $F_{(3, 111.52)}$ = 7.37, p < .001; IPL$_{long}$: $F_{(3, 111.27)}$ = 4.43, p = .006) and

electrode location (IPL$_{short}$: $F_{(1,110.27)}$ = 11.99, p < .001; IPL$_{long}$: $F_{(1,109.57)}$ = 9.55, p = .003).

The interaction of these two factors on the IPL was not statistically significant (IPL$_{short}$:

$F_{(3, 110.32)}$ = 1.29, p =.280; IPL$_{long}$: $F_{(3, 109.64)}$ = 0.80, p = .497). Regardless of how the IPL

was quantified, IPLs measured at the basal electrode location were significantly shorter

than those measured at the apical electrode location for all patient groups (IPL$_{short}$: p <

001; IPL$_{long}$: p = .002). For the group effects on the IPL$_{short}$ and IPL$_{long}$, post-hoc pairwise

comparisons showed slightly different results. Specifically, children with CND had

significantly smaller IPL$_{short}$ values than both children with ANSD (p = .009) children with

typical SNHL (p < .001). IPL$_{short}$ values measured in adult CI users were significantly

smaller than those measured in children with typical SNHL (p = .010). There was no

significant difference in the IPL$_{short}$ between any other two patient groups (p > .05). For

the results of the IPL$_{long}$, children with CND showed significantly shorter IPLs than

children with ANSD (p = .033) and children with typical SNHL (p = .009). There was no

significant difference in the IPL$_{long}$ between any other two patient groups (p > .05). The

lack of significant difference in the IPL$_{short}$ or IPL$_{long}$ between adult CI users and children

with CND should be noted and highlighted here due to its importance for data

interpretation.

<u>Correlations with the PLV</u>

Panels (g) and (h) of Figure 3 show the IPL$_{short}$ and IPL$_{long}$ measured in adult

participants plotted as a function of the PLV, respectively. The results of Spearman rank

correlation tests showed a significant correlation between the PLV and the IPL$_{short}$ ($\rho_{(105)}$

= 0.24, p = .014) with larger PLV values associated with longer IPL$_{short}$ latencies.

However, this significant correlation was not observed for the results of the IPL_{long} ($\rho_{(105)}$

= 0.07, p = .478).

<u>Associations with temporal resolution, speech perception and advanced age</u>

The associations between GDTs and IPLs quantified in both ways measured at

electrode 12 in a subgroup of 10 children with ANSD are shown in panels (a) and (b) of

Figure 4. Panels (c) and (d) show PBK word scores measured in the same group of

patients plotted against the IPL_{short} and the IPL_{long}, respectively. These results showed a

lack of association between the IPL, regardless of how it was quantified, and either of

the two auditory perception outcomes, which is supported by the results of Spearman

rank correlation tests listed in each panel.

483   <u>Insert Figure 4 about here</u>

Panels (e) and (f) of Figure 4 show psychophysical GDTs measured at two

electrode locations per test ear in a subgroup of 15 adult participants as a function of

the IPL_{short} and the IPL_{long}, respectively. These results clearly demonstrated a lack of

consistent trend in the association between the IPL quantified either way and GDT

across participants. For the results of the IPL_{short}, the two CI electrode locations tested

in the right ear of A024 showed equal IPLs but different GDTs. In the left ear of A040,

the same GDTs were measured for the two electrode locations with an IPL difference of

97.6 μs (two samples). For the rest of participants, while smaller GDTs were observed

at the electrode locations with shorter IPLs in six participants, results measured in the

other seven participants showed an opposite pattern. For the results of the IPL_{long}, the

two CI electrode locations tested in the right ear of A024 and in the left ear of A079

showed equal IPLs but different GDTs. In the left ear of A040, the same GDTs were

measured for the two electrode locations with an IPL difference of 146.4 µs (three

samples).   For the rest of participants, while smaller GDTs were observed at the

electrode locations with shorter IPLs in five participants, results measured in the other

seven participants showed an opposite pattern. Interestingly, in the left ear of participant

A052, the association between GDT and IPL differed depending on how the IPL was

quantified—showing opposite patterns for IPL_{short} and IPL_{long}. For all other participants,

the method used to quantify IPL (i.e., IPL_{short} vs. IPL_{long}) did not alter the direction of its

relationship with GDT. In contrast, the same set of psychophysical GDT data

demonstrated consistent trend of association with PLV values (lower GDTs ~ higher

PLV) across a majority of the participants (He et al. 2024b), with the exception of the

data measured in the right ear of A029 and in the left ears of A040 and A080.

Panels (g) ☐ (j) of Figure 4 show the associations between the numerically

averaged IPL across four electrode locations and AzBio sentence scores measured in

quiet and in noise in 28 adult participants, respectively. These data demonstrate a lack

of association between the IPL and AzBio sentence scores measured in any listening

conditions. This observation was supported by the results of MLRs after controlling for

the effect of participant age. Table 2 includes the results of MLRs calculated for IPLs

quantified in both ways and AzBio sentence scores measured in different listening

conditions.

Table 2. Results of Multiple linear regressions for assessing the associations between

AzBio sentence scores measured in different listening conditions and the interpeak

latency (IPL) results quantified in different ways after controlling the age effect. SNR:

signal to noise ratio; SE: standard error.

IPL Calculation Method	Listening Condition	Predictor	Unstandardized β (SE)	t	p
Min	Quiet	IPL	-0.0381 (0.0836)	-0.456	.653
		Age	0.1413 (0.3222)	0.439	.665
	Noise, SNR +10 dB	IPL	0.0496 (0.1109)	0.447	.659
		Age	-0.5375 (0.4277)	-1.257	.221
	Noise, SNR +5 dB	IPL	0.0688 (0.0791)	0.869	.393
		Age	-0.7113 (0.3051)	-2.331	.028*
Max	Quiet	IPL	-0.0982 (0.0780)	-1.259	.220
		Age	0.2425 (0.3048)	0.795	.434
	Noise, SNR +10 dB	IPL	-0.0061 (0.1068)	-0.057	.955
		Age	-0.4307 (0.4172)	-1.032	.312
	Noise, SNR +5 dB	IPL	0.0141 (0.0770)	0.183	.857
		Age	-0.6033 (0.3007)	-2.006	.056
Schvartz-Leyzac et al. (2025) 400 Sweeps	Quiet	IPL	0.1828 (0.2578)	0.709	.485
		Age	0.0229 (0.2841)	0.081	.936
	Noise, SNR +10 dB	IPL	0.0951 (0.3450)	0.276	.785
		Age	-0.4647 (0.3803)	-1.222	.233
	Noise, SNR +5 dB	IPL	0.0546 (0.2490)	0.219	.828
		Age	-0.5916 (0.2744)	-2.156	.041*
Schvartz-Leyzac et al. (2025) 50 Sweeps	Quiet	IPL	0.3906 (0.2190)	1.784	.087
		Age	-0.1144 (0.2826)	-0.405	.689
	Noise, SNR +10 dB	IPL	0.5123 (0.2911)	1.760	.091
		Age	-0.6802 (0.3756)	-1.811	.082
	Noise, SNR +5 dB	IPL	0.2869 (0.2151)	1.334	.194
		Age	-0.7120 (0.2775)	-2.566	.017*

Figure 5 shows IPLs measured at different electrode locations in adult participants

plotted as a function of age. The upper and lower panels show the results of the IPL_{short}

and IPL_{long}, respectively. Each panel represents the results measured at one electrode

location. Overall, IPLs measured at all four electrode locations showed a trend toward

longer latencies with increasing age. However, this trend only reached statistical

significance for the IPL$_{short}$ measured at electrode 21, as determined by Pearson

correlation tests with Bonferroni correction.

529   Insert Figure 5 about here

Evaluations of the VIL

Panel (a) of Figure 6 shows example single eCAP sweeps with the multi-peak

issue that affects the precision of the VIL. This issue occurred in all participants, even in

cases where eCAP amplitude was 1000 µV or larger (i.e., the grey trace with circle

marks). The percentage of traces with this multi-peak issue ranges from 0 to 57.65%

(mean: 29.13%, SD: 13.93%) and from 5.51 to 63.96% (mean: 46.59%, SD: 7.08%) for

the N1 and the P2 peaks, respectively. Panel (b) of Figure 6 shows the percentages of

eCAP traces with multiple points identified for the N1 (black circles) and the P2 peaks

(red triangles) plotted as a function of eCAP amplitude. These data demonstrate that

the percentage decreases with larger eCAP amplitudes for both peaks, which indicates

the impact of recording noise.

541   Insert Figure 6 about here

Panel (c) of Figure 6 shows the results of the VIL$_{short}$, VIL$_{long}$ and VIL$_{all}$. LMM

revealed significant main effect of VIL type, after controlling electrode location ($F_{(2, 284.95)}$

= 42. 37, $p < .001$). Post-hoc analysis revealed that the VIL$_{all}$ was significantly smaller

than both the VIL$_{short}$ and the VIL$_{long}$ ($p < .001$). There was no significant difference

between the VIL$_{short}$ and the VIL$_{long}$ ($p = .117$). Due to the strong correlations among

these VIL results, as demonstrated by panel (d) Figure 6, the VIL$_{all}$ results were used in

the subsequent assessments.

Panel (e) of Figure 6 shows VIL$_{all}$ values measured at four electrode locations in

adult participants. These results display an overall decreasing trend in the VIL_{all} as the

testing electrode moved to more apical electrode locations. Consistent with this

observation, the results of a LMM revealed a significant effect of electrode location on

the VIL_{all} ($F_{(3, 23.68)}$ = 3.56, p = .029). Post-hoc analysis revealed that VIL_{all} results

measured at electrode 3 were significantly larger than those measured at electrode 15

(p = .023). There was no significant difference in the VIL_{all} between any other electrode

location pairs (p > .05).

The association between psychophysical GDTs and VIL_{all} values is shown in panel

(g) of Figure 6. All except for four participants (A029, A040, A065 and A080) showed

lower GDTs at the electrode location with smaller VIL_{all} values. The results measured in

the right ear of A029 and in the left ears of A065 and A080 showed an opposite pattern.

For A040, the same GDTs were measured at the two electrode locations with different

VIL_{all} values. Panels (h) and (i) of Figure 6 show the association between the

numerically averaged VIL_{all} across four electrode locations and AzBio sentences scores

measured in quiet and in noise, respectively. These results demonstrate the lack of

association between AzBio sentence scores measured in quiet and the VIL_{all}. AzBio

sentence scores measured in the two noise conditions were lower for larger VIL_{all}

values. These observations are supported by the MLRs after controlling for the effect of

age (Quiet: $t_{(25)}$ = -2.043, p = .052; +10 dB SNR:　$t_{(25)}$ = -2.102, p = .046; +5 dB SNR:

$t_{(25)}$ = 2.342, p = .027).

Finally, panel (i) of Figure 6 shows the VIL_{all} plotted against the PLV. The result of

a Pearson correlation test showed a significant correlation between these two

parameters ($r_{(105)}$ = -0.923, p < .001), with smaller VILs correlated with larger PLV

values.

Analysis of Recording Noise

Panel (a) of Figure 7 shows the individual (thin lines) and population average (thick

black line) recording-noise spectrum for each of 69 electrodes across the 14 adult

participants in dataset four. For the population average noise spectrum, only 22.5% of

the noise power is within the frequency range of the PLV analysis (0.789 to 4.73 kHz),

indicated by the red vertical dashed lines. In contrast, 58% of the population average

recording noise power is below this frequency range. Thus, the effective SNR of the

PLV measure is substantially better than that of the time-domain IPL and VIL measures

which are impacted by the full noise bandwidth. Panels (b) and (c) show histograms for

the percentage of the recording noise power that is within the frequency range of the

PLV analysis (panel b) and the percentage of power below that frequency range (panel

c). Across all the electrodes in dataset four, the percentage of power within the PLV

analysis frequency range extends from 10% to 40%, whereas the percentage of power

below the PLV analysis range is between 30% and 90%. This indicates that electrodes

with a greater amount of total recording noise power than the population average only

have slightly more than average noise power within the PLV frequency range but have

substantially more noise than the average at lower frequencies. Therefore, the PLV is

more robust than the IPL and VIL over a greater range of recording noise powers.

592   Insert Figure 7 about here

Simulation of the Effects of Inter-trial eCAP Jitter

The results of the eCAP jitter simulations are shown in Figure 8. Panels (a) □ (d)

provide example results of the eCAP inter-trial jitter simulations, for jitter factors of 0.1,

0.3, 1, and 2, respectively. It can be observed that increasing the jitter factor from 0.1 to

0.3 produces only minimal temporal smearing in the mean eCAP waveform (thick black

lines) when compared to the template waveform (thick red dash lines), so the IPL is

practically unchanged. When the jitter factor is increased to a value of 1 (i.e., to the

baseline gamma distribution pdf that has a median amount of jitter) somewhat more

temporal smearing occurs in the mean eCAP waveform and its IPL is elongated

compared to the template waveform's IPL. Increasing the jitter factor to 2 produces

substantially more temporal smearing and a greatly lengthened IPL in the mean

waveform.

Panels (e) ☐ (g) of Figure 8 show the mean values of the estimated IPL of the

mean eCAP waveform, the VIL over the 400 trials, and the PLV over 400 trials,

respectively, as a function of the jitter factor. Error bars show the standard deviation in

the value over ten simulation repetitions. These results clearly show that the IPL is very

insensitive to the jitter factor over the range of 0 to around 0.5, and then from 0.5 to 1

the IPL starts increasing and has a roughly linear relationship with the jitter factor for

values between 1 and 2. However, the standard deviation of the estimates (indicated by

the error bars) also grows in this region, indicating that it lacks somewhat in reliability

even within this range of jitter factor values. In contrast, the VIL and the PLV show a

nearly linear relationship with the jitter factor over the entire range from 0 to 2, with only

a moderate increase in the standard deviation of the estimates at larger jitter factor

values. These simulation results strongly suggest that the VIL and PLV are much more

reliable measures of CN synchrony than the IPL of the mean waveform.

618   Insert Figure 8 about here

DISCUSSION

This study evaluated the IPL and the VIL as potential indicators of neural

synchrony in the CN in CI users. Both metrics require accurately identifying latencies of

the N1 and P2 peaks of the eCAP. In this section, two major factors affecting the

accuracy of both measures were discussed first, followed by discussions of the results

of each assessment.

Recording noise and sampling rate

In current CIs, the level of recording noise for eCAP measurements varies across

manufacturers and is dependent on the number of sweeps averaged to obtain the

eCAP. The noise floor is reported to be approximately 20 μV in older for Cochlear™

Nucleus® implants and 2 ▢ 5 μV in new for Cochlear™ Nucleus® implants [48] and 20

▢ 50 μV (Glassman & Hughes 2012) in Advanced Bionics devices. For MED-EL

devices, the noise floor is frequency dependent, with an estimated overall noise level of

2.6 ▢ 10 μV [52,53]. In our analysis of measured recording noise of the fourth dataset,

we found that the single-sweep noise amplitudes for these adult Cochlear™ Nucleus®

users is also frequency dependent — see panel (a) of Figure 7. As demonstrated by our

eCAP results, this recording noise introduces variability in IPL and VIL assessments.

The temporal resolution of peak latency measurements depends on the sampling

rate used for eCAP recordings. Higher sampling rates enable more precise identification

of response peaks, leading to more accurate IPL and VIL calculations. In Cochlear™

Nucleus® devices, the sampling rate is 20,492 Hz, resulting in a step size (i.e., the time

interval between consecutive samples) of 48.8 μs. This relatively large step size can

increase both local and global errors in peak latency estimation. Advanced Bionics

devices offer a sampling rate of 56 kHz, corresponding to a finer temporal resolution of

approximately 17.85 µs. However, the benefit of this higher resolution is offset by

device's relatively high noise floor. Theoretically, the high sampling rate (i.e.,1.2 MHz)

and the relatively low noise floors of MED-EL devices could improve the accuracy of IPL

and VIL assessments. More studies are warranted to test this theoretical possibility.

Overall, these two hardware-related factors—the noise floor and the sampling rate—

limit the accuracy of both IPL and VIL metrics.

Interpeak latency of the eCAP

One aim of this study was to test the hypothesized effect of reduced neural

synchrony of the CN on the IPL of the averaged eCAP across 50-100 sweeps, as

proposed by Schvartz-Leyzac et al. [21]. Overall, our results do not provide compelling

evidence to support the use of IPL as an indicator of neural synchrony in the CN. The

key findings supporting this conclusion are summarized and discussed below.

Finding 1*:* There was no significant group difference in the IPL between children

with ANSD, a patient population characterized by poor neural synchrony, and children

with typical SNHL. This lack of group difference between these two patient populations

questions the sensitivity of IPL as a reliable indicator of neural synchrony in the CN in

human CI users.

Finding 2*:* In adult CI users, IPLs increased as the testing electrode moved toward

more apical locations. If the IPL exclusively reflects neural synchrony in the CN, these

results would suggest that adult CI users have better neural synchrony at the base than

at more apical regions of the cochlea. However, this interpretation conflicts with

histological findings from human temporal bone studies in CI users showing greater

neural degeneration at the base than at the apex of the cochlea [32,44], as well as

neural degenerated patterns observed in age-related hearing loss [e.g., [54]].

Moreover, children with CND, a patient population with substantially fewer but otherwise

healthy SGNs, showed significantly shorter IPLs than both children with ANSD and

children with typical SNHL. This suggests that SGN count may affect the IPL, with

reduced SGN leading to shorter IPLs. The lack of significant difference in IPLs between

adult CI users and children with CND, combined with the well-documented SGN loss in

patients with SNHL [27-33,54], further support the potential impact of SGN count on the

IPL. Collectively, these findings raise questions about the biological mechanisms that

IPL measurements actually reflect.

Finding 3. In adult CI users, IPLs measured at all four electrode locations showed

an increasing trend with advancing age, with a stronger trend at the apical electrode

location. These age-related increases are inconsistent with prior evidence of greater CN

degeneration at the cochlear base than apex in older adults [54,55].

Finding 4. No significant associations were found between IPL and either temporal

resolution or speech perception outcomes in children with ANSD or in adult CI users.

These results contradict the established relationship between poor neural synchrony in

the CN and deficits in temporal resolution and speech perception in noise [e.g., [1]].

Finding 5. Consistent with the experimental data, our simulation results

demonstrated IPL assessment's relative lack of sensitivity to the amount of neural

synchrony and vulnerability to high measurement error, which further supported this

conclusion.

In summary, our data show that the IPL, as a proposed indicator of neural

synchrony in the CN, lacks both sensitivity to the relevant underlying neural pathology

and predictive value for the resulting auditory perceptual outcomes. Therefore, the

current findings do not support the use of IPL as a valid marker of neural synchrony in

the CN in CI users.

Study results comparisons

In a recent study, Schvartz-Leyzac et al. [21] proposed that decreased neural

synchrony should lead to large IPL values. In their study, a significant moderate

correlation between the IPL and participant age in data pooled across all electrodes in

all adult participants was found, with longer IPLs correlated with more advanced age

(see their Figure 5). Our data, as shown in Figure 5, showed a similar overall pattern,

with this positive correlation being stronger at more apical electrode locations. As

discussed above, this pattern is not consistent with neural degenerated patterns

observed in age-related hearing loss [e.g., [54]]. At this point, factors accounting for this

observed association between age and the IPL remain unknown. It should be noted that

our simulation results do not exclude the possibility that the IPL of the mean eCAP

waveform can be affected by pathologies that cause a deterministic spreading out of the

spike times in a population of CN fibers, leading to a common, broadened eCAP

waveform across trials. In these cases, longer IPLs will be observed but they are not

due to reduced neural synchrony in the CN. We speculate that the IPL is likely affected

by both deterministic changes in the eCAP waveform (common across trials) and inter-

trial jitter in the eCAP, suggesting that reduced neural synchrony is unlikely to be the

only factor underlying the age-related increase in the IPL.

Furthermore, Schvartz-Leyzac et al. [21] reported a negative correlation between

AzBio sentence scores at + 10 dB SNR and the IPL (see their Fig. 8), suggesting that longer IPLs were associated with poorer speech perception in noise. In contrast, our study found no such association between the IPL and AzBio sentence scores under any listening condition, regardless of how the IPL was quantified. We identified three key methodological differences in IPL quantification between Schvartz-Leyzac et al. [21] and our study. First, the number of sweeps included in the averaged eCAP responses differed. Schvartz-Leyzac et al. [21] used 50–100 sweeps, whereas we used 400 sweeps, potentially improving the SNR in our data. Second, there were differences in the designated latency windows for identifying the N1 and P2 peaks. Schvartz-Leyzac et al. [21] reported using latency windows of 150–300 µs for N1 and 400–600 µs for P2. However, we noted a discrepancy between these reported time windows and the IPL values shown in their Figures 5 and 6. Based on their defined latency windows, all IPLs should fall within 450 µs or less, yet several IPLs reported in their study exceeded 450 µs, with one even surpassing 500 µs. The reason for this inconsistency is unclear. In our study, broader latency windows were used (i.e., 150–600 µs for N1 and 400–1000 µs for P2) to better accommodate variabilities in eCAP morphology. Finally, Schvartz-Leyzac et al. [21] did not use any rounding up process to control for the multi-peak issue.

To determine whether the discrepancies between our findings and those of Schvartz-Leyzac et al. [21] were attributable to methodological differences, we reanalyzed our data using the time windows reported in their Methods section to identify N1 and P2 peak latencies. Specifically, IPLs were recalculated for averaged eCAPs across both the first 50 sweeps and the full set of 400 sweeps, without applying any

rounding-up procedure. We then reexamined the associations between IPL and

psychophysical GDTs, participant age, and AzBio sentence scores obtained under

various listening conditions in adult CI users. Supplemental Digital Content 1 shows the

association between the IPL, psychophysical GDT and AzBio sentence scores

measured in quiet and in noise. For the IPL calculated for the averaged eCAP across 50

sweeps, five participants showed equal IPLs at the two electrode locations with different

psychophysical GDTs. Smaller GDTs measured at the electrodes associated with

shorter IPLs in five participants. An opposite pattern was observed in four participants.

When using IPLs derived from 400-sweep averages, nine participants had equal IPLs at

two electrode locations with different GDTs, effectively ruling out the IPL as a reliable

indicator of GDT. In addition, no associations were found between the IPL and AzBio

sentence scores under any listening condition, regardless of the number of sweeps

used in eCAP averaging. The MLR results supporting this observation are presented in

Table 2. Supplemental Digital Content 2 shows IPLs measured at different electrode

locations as a function of participant age. These data revealed no correlation between

IPL and age at any electrode location, regardless of the sweep count used in the

averaged eCAP. This lack of association is further supported by Pearson correlation

results, as indicated in each panel. In summary, our reanalysis showed that the

discrepancies between our findings and those of Schvartz-Leyzac et al. (2025) cannot

be explained by the three methodological differences identified earlier. At present, the

factors accounting for this inconsistency remain unknown.

The inter-trial variability of the interpeak latency of the eCAP

The robust correlation between the VIL and the PLV indicates that both measures

likely assess common underlying biological mechanisms that produce temporal jitter in

the single-sweep eCAP waveform, as was simulated in Figure 8. This interpretation is

further supported by the similarity in their results, including the effects of electrode

location and their associations with temporal resolution and speech perception

outcomes, as reported in our previous studies [17,18]. Consequently, these findings are

not repeated or discussed in detail in the current report. Instead, we summarize the

advantages and limitations of each measure below.

The VIL

Pros: This measure can be applied across CI devices regardless of the

manufacturer. The computational method is straightforward and easy to understand. No

parameter adjustment is needed when testing patients with CIs from different

manufacturers. The VIL is only sensitive to the variability in peak latencies across trials,

in contrast to the IPL's additional dependence on the mean peak latencies (i.e., latency

differences that are common across trials).

Cons: As with IPL, VIL can be affected by the presence of multiple peaks in eCAP

waveforms, which can compromise precision. In the results presented in this paper, the

prevalence of multiple peaks is greater in the VIL method than the IPL approach,

because the VIL method is based on single sweeps and therefore does not average out

the recording noise. Additionally, our analysis of the eCAP recording-noise spectrum

indicates that the recording noise has a greater impact on the VIL than on the PLV. The

varying recording noise levels across CI devices may hinder the consistent application

of VIL as a cross-platform metric.

The PLV

Pros: The PLV is not affected by the ambiguity introduced by multiple peaks in

eCAP waveforms and has a better effective SNR than the IPL and VIL. The PLV is a

well-established concept for assessing neural synchrony in literature and is only

sensitive to variability in the eCAP waveform across trials, not the basic shape of the

single-sweep waveform arising from factors that are common across trials, in contrast to

the IPL.

Cons: Parameters used in mathematical computations need to be adjusted based

on the sampling rate of CI devices from each manufacturer. Current parameter settings

are optimized for Cochlear™ Nucleus® devices only. Specific parameters for other CI

systems remain undetermined. This limitation complicates direct comparison of PLV

results between users of different CI devices.

CONCLUSIONS

There is no compelling evidence from our data or simulations supporting the IPL

as a reliable indicator of neural synchrony in the CN. However, its inter-trial variability

could be used as an alternative to the PLV to evaluate neural synchrony of the CN in CI

users. The precision of this measure is affected by the sampling rate and recording

noise of CI devices.

REFERENCES

1.   Zeng FG, Kong YY, Michalewski HJ, Starr A. (2005). Perceptual consequences of disrupted auditory nerve activity. *Journal of Neurophysiology.* 93(6):3050-3063. https://doi.org/10.1152/jn.00985.2004

2.   Zeng FG, Oba S, Garde S, Sininger Y, Starr A. (1999). Temporal and speech processing deficits in auditory neuropathy. *Neuroreport.* 10(16):3429-35. https://doi.org/10.1097/00001756-199911080-00031

3.   Kraus N, Nicol T. (2003). Aggregate neural responses to speech sounds in the central auditory system. *Speech Commun.* 41(1):35-47. https://doi.org/10.1016/S0167-6393(02)00091-2

4.   Harris KC, Ahlstrom JB, Dias JW, et al. (2021). Neural presbyacusis in humans inferred from age-related differences in auditory nerve function and structure. *J Neurosci.* 41(50):10293-10304. https://doi.org/10.1523/JNEUROSCI.1747-21.2021

5.   Harris KC, Dubno JR. (2017). Age-related deficits in auditory temporal processing: unique contributions of neural dyssynchrony and slowed neuronal processing. *Neurobiol Aging.* 53:150-158. https://doi.org/10.1016/j.neurobiolaging.2017.01.008

6.   Rance G, Barker EJ. (2008). Speech perception in children with auditory neuropathy/dyssynchrony managed with either hearing aids or cochlear implants. *Otol Neurotol.* 29(2):179-182. https://doi.org/DOI 10.1097/mao.0b013e31815e92fd

7.   Anderson S, Skoe E, Chandrasekaran B, Kraus N. (2010). Neural timing is linked to speech perception in noise. *J Neurosci.* 30(14):4922-4926. https://doi.org/10.1523/Jneurosci.0107-10.2010

8.   Steinschneider M, Volkov IO, Fishman YI, Oya H, Arezzo JC, Howard MA.

(2005). Intracortical responses in human and monkey primary auditory cortex support a

temporal processing mechanism for encoding of the voice onset time phonetic

parameter. *Cereb Cortex*. 15(2):170-186. https://doi.org/10.1093/cercor/bhh120

9.   Steinschneider M, Volkov IO, Noh MD, Garell PC, Howard MA, 3rd. (1999).

Temporal encoding of the voice onset time phonetic parameter by field potentials

recorded directly from human auditory cortex. *J Neurophysiol*. 82(5):2346-57.

https://doi.org/10.1152/jn.1999.82.5.2346

10.   Delgutte B. (1980). Representation of speech-like sounds in the discharge

patterns of auditory-nerve fibers. *J Acoust Soc Am*. 68(3):843-857. https://doi.org/Doi

10.1121/1.384824

11.   Delgutte B, Kiang NYS. (1984). Speech coding in the auditory-nerve .4. Sounds

with consonant-like dynamic characteristics. *J Acoust Soc Am*. 75(3):897-907.

https://doi.org/Doi 10.1121/1.390599

12.   Litvak L, Delgutte B, Eddington D. (2001). Auditory nerve fiber responses to

electric stimulation: Modulated and unmodulated pulse trains. *J Acoust Soc Am*.

110(1):368-379. https://doi.org/Doi 10.1121/1.1375140

13.   Wilson BS, Finley CC, Lawson DT, Wolford RD, Zerbi M. (1993). Design and

evaluation of a continuous interleaved sampling (CIS) processing strategy for

multichannel cochlear implants. *J Rehabil Res Dev*. 30(1):110-116.

14.   Carroll J, Tiaden S, Zeng FG. (2011). Fundamental frequency is critical to

speech perception in noise in combined acoustic and electric hearing. *J Acoust Soc Am*.

130(4):2054-2062. https://doi.org/10.1121/1.3631563

15.   Fu QJ. (2002). Temporal processing and speech recognition in cochlear implant

users. *Neuroreport*. 13(13):1635-9. https://doi.org/10.1097/00001756-200209160-00013

16.   Shannon RV, Zeng FG, Kamath V, Wygonski J, Ekelid M. (1995). Speech

recognition   with   primarily   temporal   cues.   *Science*.   270(5234):303-304.

https://doi.org/DOI 10.1126/science.270.5234.303

17.   He S, Skidmore J, Bruce IC, Oleson JJ, Yuan Y. (2024). Peripheral neural

synchrony in postlingually deafened adult cochlear implant users. *Ear Hear*. 45(5):1125-

1137. https://doi.org/10.1097/AUD.0000000000001502

18.   Gao Z, Yuan Y, Oleson JJ, et al. (2025). The relationships between cochlear

nerve health and AzBio sentence scores in quiet and noise in postlingually deafened

adult cochlear implant users. *medRxiv*. https://doi.org/10.1101/2024.11.16.24317332

19.   He S, Chao X, Yuan Y, Skidmore J, Uhler KM. (2024). Assessing neural

synchrony in the cochlear nerve to electrical stimulation in children with auditory

neuropathy   spectrum   disorder.   *Ear   Hear*.

https://doi.org/10.1097/AUD.0000000000001567

20.   Skidmore J, Bruce, I.C., Yuan, Y., and He, S. Quantifying neural synchrony at the

level of the auditory nerve in cochlear implant users with recordings of the electrically

evoked compound action potential. presented at: The 46th Annual MidWinter Meeting of

the Association for Research in Otolaryngology; 2023; Orlando, FL.

21.   Schvartz-Leyzac KC, McClaskey CM, Dias JW, Pfingst BE, Harris KC. (2025).

Contributions of auditory nerve density and synchrony to speech understanding in older

cochlear   implant   users.   *Jaro-J   Assoc   Res   Oto*.   26(3):317-329.

https://doi.org/10.1007/s10162-025-00984-3

22.   Don M, Allen AR, Starr A. (1977). Effect of click rate on the latency of auditory

brain stem responses in humans. *Ann Otol Rhinol Laryngol.* 86(2 pt. 1):186-95.

https://doi.org/10.1177/000348947708600209

23.   Schaette R, Gollisch T, Herz AVM. (2005). Spike-train variability of auditory

neurons in vivo: Dynamic responses follow predictions from constant stimuli. *Journal of*

Neurophysiology. 93(6):3270-3281. https://doi.org/10.1152/jn.00758.2004

24.   Lichtenhan JT, Chertoff ME. (2008). Temporary hearing loss influences post-

stimulus time histogram and single neuron action potential estimates from human

compound   action   potentials.   *J   Acoust   Soc   Am.*   123(4):2200-2212.

https://doi.org/10.1121/1.2885748

25.   Santarelli R, Scimemi P, La Morgia C, Cama E, del Castillo I, Carelli V. (2021).

Electrocochleography in auditory neuropathy related to mutations in the OTOF or OPA1

gene. *Audiol Res.* 11(4):639-652. https://doi.org/10.3390/audiolres11040059

26.   Santarelli R, Starr A, Michalewski HJ, Arslan E. (2008). Neural and receptor

cochlear potentials obtained by transtympanic electrocochleography in auditory

neuropathy.   *Clin   Neurophysiol.*   119(5):1028-1041.

https://doi.org/10.1016/j.clinph.2008.01.018

27.   Fayad J, Linthicum FH, Jr., Otto SR, Galey FR, House WF. (1991). Cochlear

implants: histopathologic findings related to performance in 16 human temporal bones.

Ann   Otol   Rhinol   Laryngol.   100(10):807-11.

https://doi.org/10.1177/000348949110001004

28.   Fayad JN, Linthicum FH, Jr. (2006). Multichannel cochlear implants: relation of

histopathology   to   performance.   *Laryngoscope.*   116(8):1310-20.

https://doi.org/10.1097/01.mlg.0000227176.09500.28

29.   Nadol JB. (1997). Patterns of neural degeneration in the human cochlea and

auditory nerve: Implications for cochlear implantation. *Otolaryng Head Neck.*

117(3):220-228. https://doi.org/Doi 10.1016/S0194-5998(97)70178-5

30.   Merchant SN, Adams JC, Nadol JB. (2005). Pathology and pathophysiology of

idiopathic sudden sensorineural hearing loss. *Otol Neurotol.* 26(2):151-160.

https://doi.org/Doi 10.1097/00129492-200503000-00004

31.   Linthicum FH, Fayad JN. (2009). Spiral ganglion cell loss is unrelated to

segmental cochlear sensory system degeneration in humans. *Otol Neurotol.* 30(3):418-

422. https://doi.org/10.1097/MAO.0b013e31819a8827

32.   'Malley JTO, Wu PZ, Kaur C, et al. (2024). Delayed hearing loss after cochlear

implantation: Re-evaluating the role of hair cell degeneration. *Hearing Res.*

447https://doi.org/ARTN 109024

10.1016/j.heares.2024.109024

33.   Rask-Andersen H, Liu W, Linthicum F. (2010). Ganglion cell and 'dendrite'

populations in electric acoustic stimulation ears. *Adv Oto-Rhino-Laryng.* 67:14-27.

34.   Tasaki I. (1955). New measurements of the capacity and the resistance of the

myelin sheath and the nodal membrane of the isolated frog nerve fiber. *Am J Physiol.*

181(3):639-650. https://doi.org/DOI 10.1152/ajplegacy.1955.181.3.639

35.   Waxman SG, Ritchie JM. (1993). Molecular dissection of the myelinated axon.

Ann Neurol. 33(2):121-136. https://doi.org/DOI 10.1002/ana.410330202

36.   Hartmann R, Topp G, Klinke R. (1984). Discharge patterns of cat primary

auditory fibers with electrical stimulation of the cochlea. *Hear Res.* 13(1):47-62.

https://doi.org/10.1016/0378-5955(84)90094-7

37.   Javel E, Shepherd RK. (2000). Electrical stimulation of the auditory nerve. III.

Response initiation sites and temporal fine structure. *Hear Res.* 140(1-2):45-76.

https://doi.org/10.1016/s0378-5955(99)00186-0

38.   Vandenhonert C, Stypulkowski PH. (1984). Physiological-properties of the

electrically stimulated auditory-nerve .2. Single fiber recordings. *Hearing Res.*

14(3):225-243. https://doi.org/Doi 10.1016/0378-5955(84)90052-2

39.   Resnick JM, O'Brien GE, Rubinstein JT. (2018). Simulated auditory nerve axon

demyelination alters sensitivity and response timing to extracellular stimulation. *Hearing*

Res. 361:121-137. https://doi.org/10.1016/j.heares.2018.01.014

40.   Miller CA, Abbas PJ, Robinson BK, Rubinstein JT, Matsuoka AJ. (1999).

Electrically evoked single-fiber action potentials from cat: responses to monopolar,

monophasic   stimulation.   *Hearing   Res.*   130(1-2):197-218.   https://doi.org/Doi

10.1016/S0378-5955(99)00012-X

41.   He S, Grose JH, Teagle HF, et al. (2013). Gap detection measured with

electrically evoked auditory event-related potentials and speech-perception abilities in

children with auditory neuropathy spectrum disorder. *Ear Hear.* 34(6):733-44.

https://doi.org/10.1097/AUD.0b013e3182944bb5

42.   Glueckert R, Rask-Andersen H, Sergi C, et al. (2010). Histology and synchrotron

radiation-based microtomography of the inner ear in a molecularly confirmed case of

CHARGE syndrome. *Am J Med Genet A.* 152A(3):665-73.

https://doi.org/10.1002/ajmg.a.33321

43. Wright CG, Brown OE, Meyerhoff WL, Rutledge JC. (1986). Auditory and

temporal bone abnormalities in CHARGE association. *Ann Otol Rhinol Laryngol.* 95(5 Pt

1):480-6. https://doi.org/10.1177/000348948609500509

44. Seyyedi M, Eddington DK, Nadol JB. (2013). Effect of monopolar and bipolar

electric stimulation on survival and size of human spiral ganglion cells as studied by

postmortem histopathology. *Hearing Res.* 302:9-16.

https://doi.org/10.1016/j.heares.2013.04.007

45. He S, Abbas PJ, Doyle DV, McFayden TC, Mulherin S. (2016). Temporal

response properties of the auditory nerve in implanted children with auditory neuropathy

spectrum disorder and implanted children with sensorineural hearing loss. *Ear Hear.*

37(4):397-411. https://doi.org/10.1097/AUD.0000000000000254

46. He S, Xu L, Skidmore J, et al. (2020). The effect of interphase gap on neural

response of the electrically stimulated cochlear nerve in children with cochlear nerve

deficiency and children with normal-sized cochlear nerves. *Ear Hear.* 41(4):918-934.

https://doi.org/10.1097/AUD.0000000000000815

47. He S, Chao X, Wang R, et al. (2020). Recommendations for measuring the

electrically evoked compound action potential in children with cochlear nerve deficiency.

Ear Hear. 41(3):465-475. https://doi.org/10.1097/AUD.0000000000000782

48. Patrick JF, Busby PA, Gibson PJ. (2006). The development of the Nucleus

Freedom Cochlear implant system. *Trends Amplif.* 10(4):175-200.

https://doi.org/10.1177/1084713806296386

49.   van Dijk B, Botros AM, Battmer RD, et al. (2007). Clinical results of AutoNRT™,

a completely automatic ECAP recording system for cochlear implants. *Ear Hearing*.

28(4):558-570. https://doi.org/DOI 10.1097/AUD.0b013e31806dc1d1

50.   Holstad BA, Sonneveldt VG, Fears BT, et al. (2009). Relation of electrically

evoked compound action potential thresholds to behavioral T- and C-levels in children

with   cochlear   implants.   *Ear   Hear*.   30(1):115-27.

https://doi.org/10.1097/AUD.0b013e3181906c0f

51.   Botros A, van Dijk B, Killian M. (2007). AutoNRT™: An automated system that

measures ECAP thresholds with the Nucleus® Freedom™ cochlear implant via

machine   intelligence.   *Artif   Intell   Med*.   40(1):15-28.

https://doi.org/10.1016/j.artmed.2006.06.003

52.   Neustetter C, Zangerl M, Spitzer P, Zierhofer C. (2012). In-vitro characterization

of a cochlear implant system for recording of evoked compound action potentials.

Biomedical Engineering Online. 11https://doi.org/Artn 22

10.1186/1475-925x-11-22

53.   Gartner L, Lenarz T, Buchner A. (2018). Fine-grain recordings of the electrically

evoked compound action potential amplitude growth function in cochlear implant

recipients. *Biomed Eng Online*. 17(1):140. https://doi.org/10.1186/s12938-018-0588-z

54.   Wu PZ, O'Malley JT, Liberman MC. (2023). Neural degeneration in normal-aging

human cochleas: Machine-learning counts and 3D mapping in archival sections. *Jaro-J

Assoc Res Oto*. 24(5):499-511. https://doi.org/10.1007/s10162-023-00909-y

55.   Wu PZ, Liberman LD, Bennett K, de Gruttola V, O'Malley JT, Liberman MC.

(2019). Primary neural degeneration in the human cochlea: Evidence for hidden hearing

loss   in   the   aging   ear.   *Neuroscience.*   407:8-20.

https://doi.org/10.1016/j.neuroscience.2018.07.053

FIGURE LEGENDS

Figure 1. Examples of averaged eCAP response recorded in one patient from each

patient group with the multi-peak issue. The time windows where multiple

sampled points showed the same values are indicated using rectangles.

Figure 2. Methodology for simulation of eCAP inter-trial jitter. Panel (a) shows the

histogram of single-sweep IPLs for an example electrode. This distribution is well

fit by a gamma distribution with the probability density function (pdf) shown by the

red curve. Panels (b) and (c) show histograms of the gamma distribution

parameters *a* and *b*, respectively, from fits to each electrode in each adult

participant from dataset three. Panel (d) shows pdfs of example gamma

distributions that are used in the inter-trial jitter simulations. The thick black curve

shows the baseline distribution (jitter factor = 1.0) obtained from the median

values of *a* and *b* from the histograms in panels (b) and (c). The other gamma

distribution pdfs are obtained by scaling the baseline pdf around its mean by the

value of *jitter factor*. Panel (e) shows the template eCAP waveform that is used

for the inter-trial jitter simulations.

Figure 3. N1-P2 interpeak latencies (IPLs) of the eCAP measured in different patient

groups, as well as their associations with phase locking values (PLV). Panels (a)

and (b) show box plots of IPL_{short} and IPL_{long} values measured at three electrode

locations in children with auditory neuropathy spectrum disorder (ANSD) and

children with idiopathic sensorineural hearing loss (SNHL), respectively. Panels

(c) and (d) show IPL_{short} and IPL_{long} results measured at four electrode locations

in adult participants, respectively. For these four panels, lines and asterisks

indicate groups or electrode locations with statistically significant differences.

Panels (e) and (f) show box plots of IPL_{short} and IPL_{long} values measured at two

electrode locations in all four patient groups, respectively. For all box plots

included in this figure, boxes show the range between the first and the third

quartiles of the data values. The horizontal bars inside the boxes represent the

median. The vertical whiskers show the range of values that are within 1.5

interquartile range (IQR) from the boxes. The dots show the 5th and 95th

percentile of the data. Panels (g) and (h) show the association between IPLs

quantified in two ways and the PLV measured in adult participants. Each dot

indicates the result measured at one electrode location in one participant. The

results of Spearman rank correlation tests are indicated in these two panels.

Figure 4. The associations between IPLs quantified in two ways and auditory

perception outcomes. Panels listed in the left and right columns represent the

results of the IPL_{short} and the IPL_{long}, respectively. Panels (a) and (b) show scatter

plots of the association between the IPL and objective gap detection threshold

(GDT) measured in a subgroup of 10 children with ANSD. Panels (c) and (d)

show the Phonetically Balanced Kindergarten (PBK) word score plotted against

the IPL for the same group of children with ANSD. The results of Spearman

correlation tests are shown in these four panels. Panels (e) and (f) display IPLs

and psychophysical GDTs measured at two electrode locations in each of 15

adult participants. Lines connect the data measured at the two electrode

1025   locations tested in the same ear. Panels (g) and (h) show scatter plots of the

1026   association between the IPL and AzBio sentence scores measured in quiet in

1027   adult participants tested in this study. Panels (i) and (l) show scatter plots of the

1028   association between the IPL and AzBio sentence scores measured in the two

1029   noise conditions in the same group of adult participants. Each symbol indicates

1030   the result measured in one participant in these four panels.

1031 **Figure 5.** The association between age and the interpeak latency for adult participants.

1032   Upper and lower panels show the results of the IPL_{short} and the IPL_{long},

1033   respectively. Each panel represents the results measured at one electrode

1034   location. Each symbol indicates the result measured in one participant. Solid

1035   lines represent the results of linear regression. The results of Pearson correlation

1036   tests are indicated in each panel.

1037 **Figure 6.** Group results and individual examples of the inter-trial variability of the

1038   interpeak latency (VIL) measurement. Panel (a) shows example traces measured

1039   at electrode 21 in three participants. The time windows where multiple sampled

1040   points showed the same values are indicated using rectangles. Panel (b) shows

1041   the percentage of traces showing multiple equal-value points for the N1 (red

1042   triangles) and P2 (black dots) plotted against the eCAP amplitude for the results

1043   measured at all electrode locations in all participants tested in this study. Panel

1044   (c) shows the box plots for VILs calculated based on the longest (VIL_{long}) and

1045   shortest (VIL_{short}) interpeak latencies, as well as for the averaged interpeak

1046   latency across all possible interpeak latency values measured for trials with

1047   multiple equal-value points (VIL_{all}). Panel (d) shows the VIL_{long} and VIL_{short} plotted

against VIL_{all}. Panel (e) shows VIL_{all} results measured at four electrode locations

in adult participants. Panel (f) displays VIL_{all} values and psychophysical GDTs

measured at two electrode locations in each of 15 adult participants. Lines

connect the data measured at the two electrode locations tested in the same ear.

Panels (g) and (h) show scatter plots of the association between the VIL_{all} and

AzBio sentence scores measured in quiet and in noise in adult participants tested

in this study, respectively. Panel (i) shows the association between the VIL_{all} and

the phase locking value measured in adult participants. Boxes displayed in

panels (c) and (e) show the range between the first and the third quartiles of the

data values. The horizontal bars inside the boxes represent the median. The

vertical whiskers show the range of values that are within 1.5 interquartile range

(IQR) from the boxes. The dots show the 5^{th} and 95^{th} percentile of the data.

Asterisks indicate statistically significant differences. In panels (b), (d) and (h),

each symbol represents the result measured at one electrode location in one

participant. In panels (g) and (h), each symbol indicates the result measured in

one participant. In panels (h) and (i), solid lines represent the results of linear

regression. The result of a Pearson correlation test is indicated in panel (i).

Figure 7. Recording noise analysis. Panel (a) shows the individual (thin lines) and

population average (thick black line) recording-noise spectrum for each of 44

electrodes across the 14 adult participants in dataset four. The vertical dashed

red lines indicate the frequency range of the PLV analysis (0.789 to 4.73 kHz).

Panels (b) and (c) show histograms for the percentage of the total recording

noise power that is within the frequency range of the PLV analysis (panel b) and

the percentage of power below that frequency range (panel c).

Figure 8. Jitter simulation results. Panels (a)□(d) show example results of the eCAP

inter-trial jitter simulations, for jitter factors of 0.1, 0.3, 1, and 2, respectively. Thin

lines show the eCAP waveforms for single trials, thick black lines show the mean

eCAP over all the trials, and thick dashed red lines show the template eCAP

waveform. Panels (e)□(g) show the estimated IPL of the mean eCAP waveform,

the VIL, and the PLV, respectively, as a function of the jitter factor. Each data

point is the mean value from ten simulations run at that jitter factor, with error

bars showing the standard deviation in the value over the ten repetitions.

1080

1081

1082

1083

1084

1085

1086

1087

1088

1089