Cholera rapid diagnostic tests at the host-microbe interface:
Key Considerations for Global Deployments

Authors
S. M. Ahmed[1*†], Md. Abu Sayeed[2†], I. Sriguha[2], E. T. Cato[2], A. Creasy-Marrazzo[2], K. Islam[3], Md. I.UL. Khabir[3,4], Md. T. R. Bhuiyan[3], Y. Begum[3], Md. Taufiqul Islam[3], Zahid Hasan Khan[3], E. Freeman[2], A. Vustepalli[2], L. Brinkley[2], M. Kamat[5], L. S. Bailey[5], N. Madi[6,7], F. Qadri[3], B. J. Shapiro[6,7], D. T. Leung[9], K. B. Basso[5], D. A. Sack[8], J. R. Andrews[10], A. I. Khan[3*], E. J. Nelson[2*]

Affiliations
[1] Department of Epidemiology, Emory University, Atlanta, GA, USA
[2] Departments of Pediatrics and Environmental and Global Health, University of Florida, Gainesville, FL, USA
[3] Infectious Diseases Division, International Centre for Diarrhoeal Disease Research, Bangladesh (icddr,b), Dhaka, Bangladesh
[4] Alabama State University, Montgomery, AL, USA
[5] Department of Chemistry, University of Florida, Gainesville, FL, USA
[6] Department of Microbiology & Immunology, McGill University, Montréal, Canada
[7] McGill Centre for Microbiome Research, McGill University, Montréal, Canada
[8] Department of International Health, Johns Hopkins University Bloomberg School of Public Health, Baltimore, Maryland, USA
[9] Department of Internal Medicine, Spencer Fox Eccles School of Medicine at the University of Utah, Salt Lake City, Utah, USA
[10] Division of Infectious Diseases and Geographic Medicine, School of Medicine, Stanford University, Stanford, California

[†] Co-first authorship
[*] Corresponding authors

Email: sharia.m.ahmed@emory.edu, ashrafk@icddrb.org; eric.nelson@ufl.edu

Keywords
Cholera; diarrhoea; diarrhea, *Vibrio cholerae*; bacteriophage; phage; antibiotics; antimicrobial resistance; AMR; rapid diagnostic test; RDT; Bangladesh; outbreak

ABSTRACT

Background. Effective cholera outbreak response requires accurate bedside rapid diagnostic tests (RDTs) because access to laboratories is often limited. Formative studies suggest cholera diagnostics have multiple vulnerabilities, including antibiotics and predation by bacteriophage ('phage') specific to *Vibrio cholerae* (*Vc*).

Methods. We conducted a prospective nationwide study in Bangladesh among over 2000 patients with diarrhoeal disease to characterize how these vulnerabilities impact RDT performance. Assays included culture, qPCR and mass spectrometry.

Findings. With the current gold standard of culture or qPCR *Vc* positivity, we found no effect of phage on RDT performance. When the diagnostic criteria were expanded to include phage, there was a small decrease in RDT sensitivity. In contrast, large increases in sensitivity and specificity were observed among patients with moderate and severe dehydration. Using the expanded definition, the odds of RDT positivity decreased among cholera patients with phage exposure. The effect was most robust among patients with severe dehydration. Antibiotic were detected in over 80% of samples by LC-MS/MS which limited testing for effects on RDTs. Applying these findings, we estimated that restricting RDT use to severe patients with no reported antibiotic exposure increases sensitivity by 50% compared to unrestricted use. If phage were a diagnostic proxy for *Vc*, we estimate RDT would miss an additional 17% of cholera cases.

Interpretation. Cholera RDTs have critical limitations that require consideration in global deployments. Inclusion of phage detection in diagnostic criteria may improve case detection which requires further study. The impact of these findings likely extends to other diseases where diagnostics share similar vulnerabilities.

RESEARCH IN CONTEXT

Evidence before this study. We conducted two Pubmed searches for reports published after Jan 1, 2000, in all languages. The first search terms were [bacteriophage OR phage] AND [pathogen] AND [diagnostic test]. The primary search criteria identified 57 publications. A secondary criterion excluded reviews and papers on phage display technology. Among twenty remaining articles, bacteriophage were used as proxies for pathogens within diverse genera (Yersinia, Mycobacterium, Burkholderia, Staphylococcus, Salmonella, Bacillus, Escherichia, Acinetobacter, Vibrio); phage profiles were also used as biomarkers for infection. The second search terms were [rapid diagnostic test OR RDT] AND [Vibrio]. After excluding reviews, 101 papers were identified that covered immunologic (antibody-based lateral flow assays), molecular (PCR, qPCR, nl-qPCR) and mass spectrometry-based assays; only 5 articles related to vibriophage detection of which three were permutations of the same initiative.

Added value of this study. To our knowledge, this is the first study to investigate the vulnerabilities of RDT performance (e.g., sensitivity and specificity) to virulent bacteriophage and antibiotics in a large multi-site study. The hypothesis that antimicrobials and disease severity might impact RDT performance by decreasing or increasing the target number, respectively, is not novel. However, there is a lack of literature from large clinical studies that rigorously tests this hypothesis. Therefore, the added value of this study is a pragmatic evaluation of this hypothesis and a proof of concept for how to test these complex questions for other less tractable diseases.

Implications of all the available evidence. Inside the cholera field, we applied the findings by estimating that changing to a mode of restricted RDT use in which RDTs are reserved for severely dehydrated patients who report no antibiotic exposure,

49   significantly improves the sensitivity. Furthermore, we estimate that an additional 17%

50   of cases would be missed by RDT if the case definition were expanded to include phage

51   detection as a proxy for the pathogen. In this context, the RDT remains with limitations

52   that require consideration when optimizing global RDT deployments. Implications likely

53   extend to other diseases where diagnostics share similar vulnerabilities.

INTRODUCTION

Cholera causes an estimated 1.3 – 4 million symptomatic cases and 21,000 – 143,000 deaths annually[1]. Since the COVID-19 pandemic, cholera outbreaks expanded to over 44 countries[2] and the World Health Organization (WHO) declared a global emergency at its highest level (grade 3)[3]. In both endemic and non-endemic regions facing this crisis, access to 'gold-standard' culture and molecular based diagnostics at conventional laboratories is limited which can impede effective response. Accurate bedside rapid diagnostic tests (RDTs) offer a promising solution to this problem. In response, international agencies have subsequently operationalized large RDT deployments[4] for epidemiologic applications that include surveillance and optimization of vaccination campaigns[5,6].

Cholera is an acute watery diarrhoeal disease caused by the Gram-negative facultative anaerobe *Vibrio cholerae (Vc)*. Case fatality rates historically rose above 20% before intravenous fluids, but in the modern era can fall to less than 1% with use of oral rehydration solution (ORS) alone[7]. The current paradigm is that susceptible individuals become infected after exposure to contaminated food, fomites, or water. The exact infectious dose in humans is unknown and likely varies. In an infant mouse model, the infectious dose may be as low as 20 cells when the inoculum is human-shed *Vc*[8]. The infection may manifest in asymptomatic or symptomatic disease. The ratio of symptomatic to asymptomatic cases is thought to vary widely. In the endemic setting of Bangladesh, it is estimated there are 820 asymptomatic infections (or exposures) per symptomatic case[5].

Factors that influence cholera severity across the asymptomatic to symptomatic

spectrum includes the degree of prior exposure to Vc[7], antimicrobial exposures[9], and

effective predation by virulent bacteriophage (phage)[10]. The WHO Global Task Force on

Cholera Control (GTFCC) recommends antibiotic treatment (doxycycline (DOX),

azithromycin (AZI), or ciprofloxacin (CIP)) for more severe cases as a supportive

maneuver [11-14]. In practice, mass spectrometry revealed that nearly all cholera patients

in a setting like Bangladesh shed diverse antibiotics despite only half reporting

exposure[10,15,16]. Antibiotic resistance is common and spreads globally as cholera

spreads[17,18]. Virulent phages specific to Vc are relatively common and limited in number

(ICP1, 2, 3)[19-23]. Vc can harbor arrays of phage defense mechanisms, including BREX,

on a mobile integrative and conjugative element (ICE) which also harbors antibiotic

resistance factors. Phage in turn can harbor cognate antidefenses[24-27]. Phage can also

act synergistically or antagonistically with antibiotics to kill bacterial prey[28,29]. While a

phenotype called hyper-infectivity is thought to accelerate outbreaks, phage predation is

proposed to quench outbreaks although data are limited[30-35].

Culture and molecular (e.g., PCR, qPCR) assays are considered 'gold-standards' for

cholera diagnostics[36,37]. Adjunct commercialized RDTs typically are lateral flow devices

('dipsticks') that rely on O-specific polysaccharide (OSP) specific antibodies (Abs)[37-39].

Guidelines restrict RDTs to epidemiologic and not clinical applications because of

variable performance with sensitivity ranging from 58-100% and specificity from 71-

100%[37,40-42]. Attempts to increase RDT performance by altering protocols and assembly

have had limited gains to extend their use to clinical decision making[40,41,43].

Explanations for the varied sensitivity include variations in disease severity,

antimicrobial exposure, and phage predation[36,37,9,44]. We previously found that the odds

of RDT positivity among cholera patients decreased by more than 80% when phage or

antibiotics were detected[9]. This vulnerability calls into question the value of RDTs in

their current form and method of use. Determining antimicrobial effects on performance

may offer opportunities to optimize deployment and extend use to clinical decision-

making.

Our objective herein was to evaluate two primary hypotheses: (i) Virulent phage and

antibiotic exposure negatively impact RDT performance. (ii) RDT performance positively

correlates with disease severity, defined using the WHO dehydration classifications of

mild, moderate and severe[45]. These two hypotheses are connected by target

concentration. The mechanisms might be a decrease in Vc targets below the RDT limit

of detection (LOD) for the first hypothesis versus an increase in Vc targets above the

LOD for the second hypothesis [10,46]. To test these hypotheses, we analyzed data from a

prospective clinical study conducted across Bangladesh whereby stool samples were

collected from patients admitted with diarrhoeal disease at district hospitals[47]. Clinical

metadata were paired with microbiologic and molecular data to detect and quantify $V.$

$cholerae$ and associated phage; mass spectrometry was conducted on select samples

to objectively assess antibiotic effects. The results support our hypotheses exposing

critical RDT limitations while also revealing potential approaches that could optimize

global RDT deployments.

MATERIALS AND METHODS

Ethics Statement. Samples, and associated clinical data, were collected within two

previously published IRB approved clinical studies in Bangladesh: (i) The National

Cholera Surveillance (NCS) study (icddr,b ERC/RRC PR-15127)[48] and (ii) The mHealth

Diarrhea Management (mHDM) cluster randomized controlled trial (IEDCR

IRB/2017/10; icddr,b ERC/RRC PR-17036; University of Florida IRB 201601762) [47].

Study Design. The parent study for this secondary analysis was a prospective

longitudinal study of patients admitted with diarrhea at five Bangladesh Ministry of

Health and Family Welfare district hospitals (both NCS and mHDM sites) and two

centralized NCS hospitals (BITID; icddr,b) (**Fig S1**). Sites were distributed nation-wide[49].

Samples were collected from March 2018 to December 2018. Samples were collected

until April 2019 at one site (icddr,b) where all assays were performed except RDTs.

Laboratory procedures. Stool samples were collected at admission. Aliquots were

stabbed into Cary-Blair transport media for subsequent culture at the icddr,b central

laboratory via standard methods[50]. Aliquots for molecular analysis were preserved in

75% RNAlater (Invitrogen); total nucleic acid (tNA) was extracted using standard

methods. Criteria and methodology for liquid chromatography-tandem mass

spectrometry (LC-MS/MS) to objectively assess antibiotics among all culture positive

and select culture negative samples were described previously[10]. In brief, criteria were

(i) culture positive for *Vc*, (ii) phage (ICP1,2,3) positive by PCR among culture-negative

samples, or (iii) *Vc* positive by PCR (*ompW* target) among a random 10% of those

samples negative by culture for *Vc* and by PCR for phage (**Fig S2**). Common antibiotics

targeted for LC-MS/MS were based on prior research[10,15]. Ciprofloxacin,

doxycycline/tetracycline, and azithromycin, metronidazole and nalidixic acid were

analyzed qualitatively. For quantitative analysis of ciprofloxacin,

doxycycline/tetracycline, and azithromycin, standard curves were prepared by dilution

series of the native and isotopic forms of targets. Each stool sample was spiked with the

isotopic mixture as an internal reference control. Quantitative results were binned by

effective concentration and ineffective concentration; the definition of effective

concentration was based on the lowest minimal inhibitory concentration (MIC) in a prior

strain collection[51]. To provide additional insight after the PCR analysis, quantitative PCR

(qPCR) was subsequently performed on all available samples to enumerate absolute

abundances of *Vc* (*tcpA*), ICP1 (GP58.2), ICP2 (GP24), and ICP3 (GP19) using

methods previously described[10]. Estimates for CFU/ml were derived using the qPCR Ct

values from standard curves of serial dilutions of cultured *V. cholerae.* Stool samples

were tested with the Cholkit (Incepta Pharmaceuticals, Bangladesh), an RDT that

targets the detection of *V. cholerae* O1 OPS using a standardized method[39,52]. In brief,

five drops of watery stool were added to the diluent supplied in a microcentrifuge tube.

The RDT strip was then dipped into the stool-diluent mixture for 15 minutes. *V. cholerae*

O1 detection was confirmed 'by eye' for the appearance of both test and control lines[39].

Solid stools were not tested; cholera RDTs in general, including the Cholkit, are not

designed to test solid stools.

Statistical analyses. Statistics and visualizations were done in GraphPad Prism

version 10.2.3 and R version 4.4.2[53]. The "Eulerr" package was used to create **Figure**

1A for samples with all relevant test results available[54]. Test performance metrics

(sensitivity, specificity, positive predictive value, negative predictive value) and

corresponding exact binomial confidence limits were calculated in the "epiR" package[55].

Standardized predictive values were calculated by using the observed diagnostic

sensitivity and specificity in the given subgroup and calculating the predictive value that

would have been observed if disease prevalence was 50% in that subgroup. We used

unadjusted generalized linear models with a binomial error distribution and logit link to

estimate odds ratios and corresponding 95% confidence intervals. Fisher's exact test

was used to calculate p-values where indicated.

RESULTS

Study overview. A total of 2574 participants were enrolled at seven hospitals across

Bangladesh from March to December 2018, with one stool sample collected from each

participant (**Figs S1 and S2**). Enrollment continued at one site (icddr,b) until April 2019.

Although RDTs were not performed at the icddr,b, samples were included to better

assess phage and *Vc* relationships by qPCR. RDTs were performed on each of the

remaining 2180 samples of which all had culture and 2078 had *Vc* qPCR (**Table 1**).

Specific to testing our two hypotheses: (i) ICP1, ICP2, and ICP3 were detected among

137, 10, and 86 participants, respectively, and (ii) the number of participants with mild,

moderate and severe dehydration was 392, 1193 and 493, respectively. LC-MS/MS for

antibiotic detection was attempted and successful on all culture positive samples and a

subset of culture negative samples (446) [10,15,51] (**Fig S2).**

Characterization and comparison of assays for *Vc* detection. Among samples with

all assays performed, there was discordance in *Vc* detection by assay type (**Fig 1A**).

Out of 2078 samples with both *Vc* culture and qPCR, only 5.8% (120) of samples were

positive for *Vc* by both culture and qPCR. An additional 13.2% (274) of samples were

negative by culture and positive by qPCR, and 2.6% (55) of samples were positive by

culture and negative by qPCR. The concentration of *Vc* at which 50% of the RDTs

tested positive (diagnostic dose 50 or 'DD$_{50}$') was 1.93 x 10^8 CFU/ml; The percentage

of positive RDTs increased with *Vc* concentration binned by order of magnitude (**Fig**

1B).

Distribution of phage among samples with and without *Vc*. Among samples with all

assays performed, ICP1, ICP2 and ICP3 were detected by qPCR in 137, 10 and 76

samples, respectively (**Table 1**). These analyses include samples that were positive or

negative for *Vc*. In the absence of *Vc* (culture or qPCR), ICP1, ICP2, and ICP3 were

detected in 2.1% (43), 0.2% (4), and 3.1% (64) of samples respectively (**Fig 1A**); unless

specified, analyses focused on ICP1 and ICP3 because of the larger sample sizes. We

next analyzed quantitative associations between phage and *Vc*. Data are visualized in

quadrants by *Vc* and/or phage detection using a Ct threshold of 28 (**Fig 1C, D**).

Samples that did not have RDT results (gray 'x') were included to better elicit patterns in

the data. For both ICP1 and ICP3, increasing *Vc* detection was associated with

increasing phage detection. Among those samples with RDT results, the percentages of

RDT positive samples are shown in each quadrant. The quadrants with the least

percentage of RDT positive samples were those with phage detection alone.

Effect of phage predation on RDT performance. Among cholera samples defined

conventionally by *Vc* qPCR or culture, RDT sensitivity (46%) and specificity (85%) did

not differ when phage positive samples were excluded (**Table 2**). PPV and NPV were

respectively higher and lower among samples with ICP1 compared to other groups

(e.g., samples without phage). However, this effect was driven by differences in cholera

prevalence between groups, and resolved once predictive values were assessed at a

standardized prevalence. Among samples with ICP3, sensitivity decreased compared to

samples without phage yet confidence intervals overlapped. While culture and qPCR

are the current reference standards, the detection of cholera-specific phage among

samples negative for *Vc* suggested the conventional definition may inappropriately

exclude cases. Therefore, we explored an expanded definition that included phage

detection as a proxy for *Vc* for situations when the primary *Vc* target might be highly

degraded. The inclusion of phage in the diagnostic criteria resulted in a modest effect

on the RDT (**Table S1**); the sensitivity decreased from 0.46 (95%CI 0.41–0.51) to 0.38

(95%CI 0.34-0.42). Despite this seemingly small effect on sensitivity, we estimated

RDTs would miss an additional 17% of 'true' cholera cases if the diagnostic reference

standard were expanded to include "phage only" as cholera cases.

Effect of dehydration severity on RDT performance. We evaluated the impact of

dehydration severity on RDT performance. The median concentrations of *Vc* (CFU/ml)

derived from those samples positive by qPCR at mild, moderate or severe levels of

dehydration were 3.25×10^7 (IQR 8.60×10^6 to 1.14×10^9), 1.05×10^8 (IQR 1.06×10^7

to 2.16×10^9) and 1.00×10^8 (IQR 8.34×10^6 to 4.41×10^9), respectively. In this context,

RDT performance increased sequentially from patients with mild to moderate to severe

dehydration (**Table 3**). RDT sensitivity was 0.25 (95% CI: 0.16, 0.37) in those mildly

dehydrated, compared to 0.63 (95% CI: 0.54, 0.71) in samples from severely

dehydrated patients. Specificity also increased between mild cases and cases with

moderate or severe dehydration. Standardized predictive values, both positive and

negative, also increased significantly in severely dehydrated patients compared to

mildly dehydrated patients.

Effect of phage on the RDT performance at different dehydration levels. We

explored if the effect of phage on RDT performance varied by dehydration status

(possible effect modification) **(Table 4)**. Among *Vc* positive patients (culture or qPCR),

phage exposure was not associated with RDT positivity as an aggregate or by severity

level. When the case definition was expanded to include phage detection, the odds of

RDT positivity were significantly lower in those with phage compared to those without

phage (OR 0.377; 95%CI 0.217-0.642). This effect was observed among samples from

moderately and severely dehydrated cases, but not from mildly dehydrated cases. The

effect also differed by phage type used in the case definition. The effect of phage on

RDT positivity was most prominent among those samples positive by culture, qPCR, or

ICP3 (OR 0.099; 95%CI 0.023-0.292). The impact of dehydration status on this

relationship between phage presence and RDT performance (effect modification) was

also strongest with this expanded definition.

Effect of antibiotic exposure on RDT performance. Large discordance between self-

reported and LC-MS/MS-detected antibiotic exposures was observed. Among those

with complete diagnostics, 46% of patients (956/2078) self-reported antibiotic exposure

and 87% (220/252) of select samples assayed by LC-MS/MS (**Fig S2**) contained

antibiotics. Among the WHO recommended antibiotics, ciprofloxacin was the most

commonly self-reported antibiotic (27%; 568) followed by azithromycin (14%; 289) with

no participant self-reporting doxycycline use (**Table S2**). RDT performance was

consistent across all types of self-reported antibiotic exposure, including no self-

reported antibiotic exposure. To take a more objective approach, LC-MS/MS was used

to detect exposure to common antibiotics (CIP, AZI, DOX, metronidazole (MET), or

nalidixic acid (NAL) which were detected in 81% (178/220) of samples tested (**Table**

S2). The small proportion of samples that did not have antibiotics detected by LC-

MS/MS limited analyses on the effects of antibiotics on RDT performance. Despite this

challenge, RDT sensitivity among samples with AZI at or above the 'effective

concentration' was 0.33 (95%CI 0.13-0.59) compared to 0.67 (95%CI 0.47-0.83) among

samples without antibiotic (**Table S2**). Among *Vc* positive samples, antibiotic resistance

had no effect on the odds of RDT positivity with or without AZI exposure, however

samples sizes were small (**Table S3**).

Estimates of RDT performance in different modes of deployment. We sought to

estimate and compare RDT performance in different modes of deployment (**Table 5**).

We considered an unrestricted mode in which all patients are tested. We compared this

to a more restricted mode that would rely on data accessible at the bedside through the

clinical history (self-reported antibiotic exposure) or the physical exam (dehydration

status). Despite phage detection not being accessible at the bedside, it is reasonable to

consider it as a potential component to include in molecular diagnostic panels in the

future. LC-MS/MS is too expensive and complex for consideration. In this context, we

found that RDT sensitivity increases from 0.46 (95%CI: 0.41-0.51) among all patients to

0.69 (95%CI: 0.57-0.80) when use is restricted to patients that report no antibiotic

exposure and have severe dehydration. When the case definition is expanded to

include phage, RDT sensitivity also increases between restricted and unrestricted

modes. Modest effects were observed on PPV and NPV depending on case definition

and mode of use.

DISCUSSION

Effective cholera control is aided by accurate bedside rapid diagnostic tests (RDTs).

Here we conducted a secondary analysis of a prospective cohort study of patients with

diarrhoeal disease to test if virulent bacteriophage (phage) impact RDT performance

and determine if there are tractable approaches to optimize RDT use. With the current

gold standard of culture or qPCR *Vc* positivity, we found no significant effect of phage

on RDT performance. When the reference standard was expanded to include phage,

there was a small decrease in RDT sensitivity. In contrast, significantly higher sensitivity

and specificity were found among patients with moderate or severe dehydration. Using

the expanded definition of including phage detection, the odds of RDT positivity among

cholera patients decreased by 57% with phage exposure and the effect was most

robust among patients with severe dehydration. These findings align with our guiding

hypotheses as well as pose intriguing questions on how one might apply these results

to optimize RDT deployments.

Our first hypothesis was that virulent phage and antibiotic exposure negatively

impact RDT diagnostic performance. RDT sensitivity and specificity were not

appreciatively impacted when samples with phage were removed from the analysis.

Among ICP1 positive samples, PPV and NPV increased and decreased respectively

using the current gold standard case definition. This likely reflects an increased pretest

probability for *Vc*; standardizing rates of *Vc* removed the effect. Among ICP3 positive

samples, sensitivity and PPV were decreased however confidence intervals overlapped;

standardizing rates of *Vc* removed the effect on PPV. We next expanded the case

definition to include phage detection. Inclusion of phage in the diagnostic criteria had

modest effects that aligned with our guiding hypotheses.

Our second hypothesis was that RDT performance positively correlates with

disease severity. The rationale is based on our prior study of the predator (phage) --

prey (*Vc*) dynamics along the disease severity spectrum; the prior study found high

abundance of *Vc* was correlated with severe disease[10]. In contrast to the modest effects

of phage, significant increases in sensitivity and specificity were found among patients

with moderate or severe dehydration. This finding likely reflects that samples from more

severe patients had higher concentrations of *Vc* at levels that we showed above have a

higher percentage of RDT positivity.

The lack of congruence between the assays performed (culture, qPCR, RDTs)

may reflect complex underlying biologic events and not simply variation in assay

assembly and manipulation. Mechanistically, we propose that false positive RDTs may

be more common among mild cases because the concentration of *Vc* is low and non-*Vc*

microbiota is high which may increase cross-reactivity of the immunologic epitopes.

Although cross-reactivity may occur using molecular assays, cross-reactivity for the

qPCR primer set used was likely low based on our prior study[10]; that said degradation of

the nucleic acid impacts molecular assays which we attempted to mitigate by stabilizing

the samples. Culture is a definitive assay for *Vc* but sensitivity can decrease because of

processing delays, antibiotic exposure and possibly phage predation. In this context, we

evaluated the interactions between phage predation and disease severity. Using the

expanded case definition that includes phage detection, we found that the odds of RDT

positivity among cholera patients decreased by 63% with phage exposure. The effect

was most robust among patients with severe dehydration and varied by phage type;

ICP3 dramatically decreased the odds of RDT positivity among cholera patients. One

explanation for this effect of among severely dehydration patients with ICP3 is that there

is simply more prey for the predator among severely dehydration patients and the effect

is more easily detected. Among patients with mild disease, there may simply be

insufficient *Vc* for the RDT to detect independent of phage predation.

Both clinicians and epidemiologists face two key questions based on these

findings. The first question is whether to include phage detection in cholera diagnostic

panels as a proxy for detection of the pathogen when phage degrade the pathogen

below the limit of detection for a given assay. The benefits (e.g. overall increased case

detection) versus risks (e.g. increased cost) need consideration by both clinicians and

epidemiologists. The second question is how to best validate and incorporate these

findings as RDTs are being deployed globally to combat cholera. For validation, studies

in diverse clinical and epidemiologic scenarios are needed. For deployment, we

estimated the effects of restricting RDT use to patients who report no antibiotic use and

have severe dehydration. In this context, sensitivity had a relative increase of

approximately 50%, independent of consideration of phage predation. This restriction

may represent a strategy to conserve RDTs in resource limited situations. However, this

restriction may also miss non-severe disease that play an important role in transmission

or mortality in the convalescent period.

These findings should be considered within the limitations of the study. First, LC-

MS/MS was not performed on all samples because of logistical limitations. Among *Vc* or

phage positive samples, LMCS found antibiotics were common and the distributions of

samples were insufficient to disaggregate the effects of phage and antibiotics on RDT

performance; proposing to restrict RDT use to those without antibiotic exposure is in

some aspects empiric. Second, the recruitment and enrollment of patients within the

parent clinical trial was based on the census of patients that sought care. This may a

potential recruitment bias. Those patients that sought care are not considered

representative of all infected individuals in the community. Third, clinical exams were

conducted in district hospitals using WHO and Bangladesh Ministry of Health and

Family Welfare guidelines. These guidelines have varied accuracy; this study would

likely be improved by incorporating more accurate measures of dehydration[56]. Fourth,

the approach had experimental limitations: The analyses excluded host factors such as

innate and acquired immunity. Control assays were not performed on

asymptomatic/healthy participants in the community. Solid stools were not assayed.

Despite these limitations, the findings provide important actionable insights. Future time-

series studies of patients over the disease course and recovery period in endemic and

non-endemic settings would assist testing our guiding hypotheses and address study

limitations.

CONCLUSION

Herein, we found RDT performance was minimally impacted by phage overall and

significantly increased among severely dehydrated patients. Expanding the case

definition to include phage, the odds of RDT test positivity decreased significantly. We

favor the expanded case definition that includes phage while recognizing the need for

future investigation. Lastly, we provide estimates on the applied effects of restricted vs

non-restricted RDT deployment. The significance of these findings may extend broadly

to other disease where diagnostics share similar vulnerabilities.

Data availability. Available upon request.

Code availability. Available upon request.

Acknowledgements: We thank the patients for their participation. We thank the diligent

clinical and laboratory team members who were essential to the success of the study.

We are grateful to the Institute of Epidemiology, Disease Control and Research

(IEDCR), Ministry of Health and Family Welfare, and Government of Bangladesh who

collaborated on the original parent clinical studies. We are also grateful to the

Governments of Bangladesh and Canada for providing unrestricted support to the

icddr,b. We are grateful to B. Johnson, T. Linn, N. Rushing and K. Berquist for their

administrative guidance at the University of Florida.

Funding: The laboratory procedures were funded by a Wellcome Trust grant to EJN

[DFID grant 215676/Z] and National Institutes of Health grants to KBB [S10 OD021758-

01A1] and DS [R01 AI123422-06A1]. Statistical analysis was funded in part by a grant

from the Thrasher Research Fund to SMA. The parent clinical study was supported by

National Institutes of Health grants to EJN [R21TW010182]. The funders had no role in

the study design, data collection, statistical analyses, decision to publish, or preparation

of the manuscript.

Author contributions:

Conceptualization: SMA, MAS, FQ, JA, AIK, EJN

Implementation of clinical study: MAS, FQ, AIK, EJN

Lab procedures: MAS, ETC, ACM, KI, MIUK, MTRB, YB, EF, AV, LB, MK, LSB, EJN

Statistical analyses and visualizations: SMA, NM, BJS, JA, EJN

Funding acquisition: SMA, DS, KBS, EJN

Supervision: KBS, BJS, FQ, AIK, EJN

Writing – original draft: SMA, EJN

Writing – review & editing: SMA, MAS, DS, AKD, JA, FQ, AIK, EJN

Competing interests: Authors declare that they have no competing interests.

FIGURE 1

Fig. 1. Comparison of target detection by different diagnostics. A Euler plot displaying
overlaps between positive diagnostics. Given limitations with Euler plots, the following positive
values are not shown in the figure: only Vc culture (n= 35/175 total culture positives; 25%), only
ICP2 qPCR (n=4/10 total ICP2 positives; 40%), only Vc culture and Vc qPCR tcpA (n=18), only
RDT and ICP1 qPCR (n=4), and only Vc qPCR tcpA and ICP3 qPCR (n=9). **B**. Linear
regression analysis of relationship between $\log_{10}$ *Vc* abundance (CFU/ml) derived from *tcpA*
qPCR and RDT positivity. Samples were binned into five groups based on *Vc* concentration:
less than 1×10^7, $1\text{-}9 \times 10^7$, $1\text{-}9 \times 10^8$, $1\text{-}9 \times 10^9$, and above 1×10^{10} CFU/ml. In a given group,
each symbol is plotted by the median concentration (X coordinate) by the proportion of samples
RDT positive (Y coordinate). The *Vc* dose at which 50% of RDTs ('DD$_{50}$') were positive (gray
dotted line at 1.93×10^8 CFU/ml). Line of best fit is shown and fit using least squares regression
(95% CIs shown). **C**. Comparison of qPCR Ct values for *Vc* (*tcpA*) with ICP1. **D**. Comparison of
qPCR Ct values for *Vc* (*tcpA*) with ICP3 (GP19). For C and D: Quadrants are demarcated by
phage (Ph) and *Vc* positivity (+) with the percentage RDT positive per quadrant. Gray dotted
lines = thresholds for positive (Ct = 28). "X" = no RDT performed; hollow colored circle = RDT+,
hollow gray circle = RDT-. Tick shows dose at which 50% of all RDTs are positive (DD$_{50}$).

419 **Table 1. Overview of participants characteristics by RDT result**

Characteristic	RDT Negative n= 1632[a] (total)	RDT Positive n= 446[a] (total)
Site		
Habiganj[b] (#2)	328 (20.1%)	177 (39.7%)
Naogaon[b] (#4)	227 (13.9%)	48 (10.8%)
Patuakhali[b] (#5)	188 (11.5%)	40 (9.0%)
Comilla[b] (#15)	169 (10.4%)	18 (4.0%)
Kusthia[b] (#17)	458 (28.1%)	99 (22.2%)
Chittagong (BITID; #10)	262 (16.1%)	64 (14.3%)
Sex, n (%)		
Male (#1)	810 (49.6%)	226 (50.7%)
Female (#2)	822 (50.4%)	220 (49.3%)
Age (months), median (IQR)	300 (15.0-480)	240 (14.0-420)
Age, n (%)		
0-11 months	270 (16.5%)	78 (17.5%)
1-4 years	319 (19.5%)	82 (18.4%)
5-64 years	968 (59.3%)	277 (62.1%)
65+ years	75 (4.6%)	9 (2.0%)
Dehydration status		
Mild	302 (18.5%)	90 (20.2%)
Some	965 (59.1%)	228 (51.1%)
Severe	365 (22.4%)	128 (28.7%)
Culture, *n* (%)		
Negative	1574 (96.4%)	329 (73.8%)
Positive	58 (3.6%)	117 (26.2%)
qPCR, *n* (%)		
Vc (*tcpA*) negative	1426 (87.4%)	258 (57.8%)
Vc (*tcpA*) positive	206 (12.6%)	188 (42.2%)
ICP1 (GP58.2) negative	1544 (94.6%)	397 (89.0%)
ICP1 (GP58.2) positive	88 (5.4%)	49 (11.0%)
ICP2 (GP24) negative	1627 (99.7%)	441 (98.9%)
ICP2 (GP24) positive	5 (0.3%)	5 (1.1%)
ICP3 (GP19) negative	1557 (95.4%)	435 (97.5%)
ICP3 (GP19) positive	75 (4.6%)	11 (2.5%)
Self-reported antibiotic use		
No	877 (53.7%)	245 (54.9%)
Yes	755 (46.3%)	201 (45.1%)

[a] Values for those samples in which cultures, qPCR (*tcpA*, ICP1, ICP2, ICP3)
and RDT were performed out of the given total. Excludes samples from Dhaka
samples (n=394) because RDTs were not performed at this site
[b] Ministry of Health and Family Welfare, Government of Bangladesh District
Hospitals. Icddr,b = International Centre for Diarrhoeal Disease Research,
Bangladesh. BITID = Bangladesh Institute of Tropical Infectious Diseases.

Table 2. Effect of phage predation on RDT performance

Sample type	All	No phage [a]	ICP1 phage positive	ICP3 phage positive
Vc+/total (%)[b]	449 / 2078 (21.6)	345 / 1852 (18.6)	89 / 137 (65.0)	15 / 86 (17.4)
Sensitivity	0.46 (0.41, 0.51)	0.46 (0.40, 0.51)	0.49 (0.39, 0.60)	0.27 (0.08, 0.55)
Specificity	0.85 (0.84, 0.87)	0.85 (0.83, 0.87)	0.90 (0.77, 0.97)	0.90 (0.81, 0.96)
PPV_{obs}[c]	0.46 (0.42, 0.51)	0.41 (0.36, 0.46)	**0.90 (0.78, 0.97)**	0.36 (0.11, 0.69)
NPV_{obs}[c]	0.85 (0.83, 0.87)	0.87 (0.85, 0.89)	**0.49 (0.38, 0.60)**	0.85 (0.75, 0.92)
PPV_{50}[c]	0.76 (0.72, 0.79)	0.75 (0.71, 0.79)	0.83 (0.68, 0.93)	0.73 (0.45, 0.92)
NPV_{50}[c]	0.61 (0.59, 0.64)	0.61 (0.58, 0.64)	0.64 (0.53, 0.73)	0.55 (0.43, 0.67)

[a] ICP1, ICP2, and ICP3 not detected
[b] Number positive for Vc by culture or qPCR ($tcpA$) as the gold reference standard over total number samples in grouping
[c] Subscript indicates the proportion of samples that were gold-standard positive in the given group for the calculation of each predictive value, with "obs"=observed and "50"=50%. For example, PPV_{50} or standardized positive predictive value indicates the positive predictive value of RDT in each group if 50% of that group was gold-standard positive for cholera.

Table 3. Effect of dehydration severity on RDT performance

	Mild (comparator)[a]	Moderate[a]	Severe[a]
Vc+/total (%)[b]	75 / 392 (19.1)[c]	236 / 1193 (19.8)[c]	138 / 493 (28.0)[c]
Sensitivity	0.25 (0.16, 0.37)	0.43 (0.36, 0.49)	**0.63 (0.54, 0.71)**
Specificity	0.78 (0.73, 0.82)	**0.87 (0.84, 0.89)**	**0.88 (0.85, 0.92)**
PPV_{obs}[d]	0.21 (0.13, 0.31)	**0.44 (0.38, 0.51)**	**0.68 (0.59, 0.76)**
NPV_{obs}[d]	0.81 (0.77, 0.86)	0.86 (0.84, 0.88)	0.86 (0.82, 0.89)
PPV_{50}[d]	0.53 (0.42, 0.63)	**0.76 (0.71, 0.81)**	**0.85 (0.78, 0.89)**
NPV_{50}[d]	0.51 (0.45, 0.57)	**0.60 (0.57, 0.64)**	**0.71 (0.65, 0.76)**

[a] Mild, moderate and severe dehydration approximate 0-4%, 5-9% and 10%+ loss to normal body due to diarrhoeal fluid loss per WHO guidelines.
[b] Number positive for Vc by culture or qPCR ($tcpA$) as the gold reference standard over total number samples in grouping
[c] The concentration of Vc derived from qPCR data among patients with mild, moderate and severe dehydration is 3.25×10^7 (IQR 8.60×10^6 to 1.14×10^9), 1.05×10^8 (1.06×10^7 to 2.16×10^9) and 1.00×10^8 (8.34×10^6 to 4.41×10^9), respectively.
[d] Defined in Table 2.

448 **Table 4. Effect of phage exposure on RDT positivity among cholera patients**

Case definition	n	% RDT positive (n/ total)		OR[a]	CI[a]	P value[a]
		Phage Exposed	Unexposed			
Vc positive[a]						
All	197	46 (22/48)	46 (69/149)	0.981	0.508-1.88	0.954
Mild	43	6 (1/16)	15 (4/27)	0.383	0.019-2.91	0.411
Moderate	86	57 (8/14)	42 (30/72)	1.867	0.589-6.21	0.291
Severe	68	72 (13/18)	70 (35/50)	1.114	0.349-3.97	0.859
Vc positive or phage positive (ICP1,2,3)[a]						
All	259	25 (27/110)	46 (69/149)	**0.377**	**0.217-0.642**	**<0.001**
Mild	62	14 (5/35)	15 (4/27)	0.958	0.229-4.25	0.953
Moderate	113	20 (8/41)	42 (30/72)	**0.339**	**0.130-0.809**	**0.019**
Severe	84	41 (14/34)	70 (35/50)	**0.300**	**0.118-0.737**	**0.010**
Vc positive or ICP1 positive[a]						
All	218	32 (22/69)	46 (69/149)	**0.543**	**0.294-0.980**	**0.046**
Mild	57	13 (4/30)	15 (4/27)	0.885	0.189-4.13	0.872
Moderate	93	33 (7/21)	42 (30/72)	0.700	0.240-1.90	0.494
Severe	68	61 (11/18)	70 (35/50)	0.673	0.220-2.14	0.491
Vc positive or ICP3 positive[a]						
All	187	8 (3/38)	46 (69/149)	**0.099**	**0.023-0.292**	**<0.001**
Mild	32	20 (1/5)	15 (4/27)	1.44	0.065-13.6	0.770
Moderate	91	5 (1/19)	42 (30/72)	**0.078**	**0.004-0.409**	**0.015**
Severe	64	7 (1/14)	70 (35/50)	**0.033**	**0.002-0.188**	**0.002**

449 [a]Reference standard to define positive cholera cases for analysis within a given group. "*Vc* positive" is defined as
450 positive by culture or qPCR. *P* values were computed using logistic regression (see methods).

451

452

453

455 **Table 5. Estimates of RDT performance in different modes of deployment**

Approach	Unrestricted (comparator)	Use clinical exam	Use patient history and exam	Use clinical exam	Use patient history and exam
Restriction	None	Moderate and severe cases	Moderate and severe cases AND no reported antibiotic use	Severe cases	Severe cases AND no reported antibiotics use
Case definition (current gold standard): culture or qPCR for *Vc*					
Cases/total (%)	449 / 2078 (21.6)	374 / 1686 (22.2)	168 / 863 (19.5)	138 / 493 (28.0)	72 / 266 (27.1)
Sensitivity	0.46 (0.41, 0.51)	0.50 (0.45, 0.55)	0.54 (0.46, 0.61)	**0.63 (0.54, 0.71)**	**0.69 (0.57, 0.80)**
Specificity	0.85 (0.84, 0.87)	0.87 (0.85, 0.89)	0.86 (0.83, 0.89)	0.88 (0.85, 0.92)	0.86 (0.80, 0.91)
PPV_{obs}[b]	0.46 (0.42, 0.51)	0.53 (0.47, 0.58)	0.48 (0.41, 0.56)	**0.68 (0.59, 0.76)**	**0.65 (0.53, 0.75)**
NPV_{obs}[b]	0.85 (0.83, 0.87)	0.86 (0.84, 0.88)	0.88 (0.86, 0.91)	0.86 (0.82, 0.89)	0.88 (0.83, 0.93)
PPV_{50}[b]	0.76 (0.72, 0.79)	0.80 (0.76, 0.83)	0.79 (0.74, 0.84)	0.85 (0.78, 0.89)	0.83 (0.75, 0.90)
NPV_{50}[b]	0.61 (0.59, 0.64)	0.64 (0.61, 0.66)	0.65 (0.61, 0.69)	**0.71 (0.65, 0.76)**	**0.74 (0.66, 0.81)**
Case definition (expanded standard): culture or qPCR for *Vc* OR qPCR for ICP1/ICP3 phage					
Cases/ total (%)	567 / 2078 (27.3)	469 / 1686 (27.8)	216 / 863 (25.0)	173 / 493 (35.1)	89 / 266 (33.5)
Sensitivity	0.38 (0.34, 0.43)	0.41 (0.37, 0.46)	0.43 (0.36, 0.50)	**0.51 (0.43, 0.59)**	**0.57 (0.46, 0.68)**
Specificity	0.85 (0.83, 0.87)	0.87 (0.85, 0.89)	0.85 (0.82, 0.88)	0.88 (0.83, 0.91)	0.85 (0.79, 0.90)
PPV_{obs}[b]	0.49 (0.44, 0.54)	0.54 (0.49, 0.60)	0.49 (0.42, 0.57)	**0.69 (0.60, 0.77)**	**0.66 (0.55, 0.77)**
NPV_{obs}[d\b]	0.79 (0.77, 0.81)	0.79 (0.77, 0.81)	0.82 (0.79, 0.85)	0.77 (0.72, 0.81)	0.80 (0.73, 0.85)
PPV_{50}[b]	0.72 (0.68, 0.76)	0.76 (0.71, 0.80)	0.74 (0.68, 0.80)	0.80 (0.73, 0.86)	0.80 (0.70, 0.87)
NPV_{50}[b]	0.58 (0.55, 0.60)	0.60 (0.57, 0.62)	0.60 (0.56, 0.64)	0.64 (0.59, 0.69)	0.67 (0.59, 0.74)

456 [a] Shaded text represents the current unrestricted standard of RDT use using qPCR and culture for *Vc* as
457 the reference standard. Bold text compares performance horizontally between unrestricted and restricted
458 approaches; underlined text vertically compares sensitivity when the reference standard is expanded to
459 include the detection of phages by qPCR (ICP1/3)
460 [b] Defined in Table 2.
461
462

463 **SUPPLEMENTARY MATERIALS**
464
465 **Supplementary Figures**
466
467 **Figure S1. Study Map**

468
469
470 **Fig S1.** Study site locations across Bangladesh. Red

471 circles = RDT performed; gray circle = RDT not

472 performed. Numbers in brackets correlate with site

473 numbers in Table 1.

474   **Figure S2. Sample processing diagram**

475
476
477

478
479
480
481   **Figure S2. Sample processing diagram.** Schematic of the number of samples collected from patients admitted with
482   diarrhoeal disease (one per patient). * Those samples with RDTs (2180) are represented in the flow diagram; RDTs were
483   not performed at the icddr,b. ** Seven samples were excluded from further analysis (AMR testing and LC-MS/MS)
484   because they were misclassified as culture negative.

Supplementary Tables

Table S1. Comparison of RDT performance using different diagnostic criteria for cholera

	Vc (culture or qPCR)	*Vc* or ICP1 (qPCR)	*Vc* or ICP3 (qPCR)	*Vc* or ICP1 or ICP3[a]
Cases/total (%)[b]	449 / 2078 (21.6)	497 / 2078 (23.9)	520 / 2078 (25.0)	567 / 2078 (27.3)
Sensitivity	0.46 (0.41, 0.51)	0.43 (0.38, 0.47)	0.41 (0.37, 0.45)	0.38 (0.34, 0.43)
Specificity	0.85 (0.84, 0.87)	0.85 (0.83, 0.87)	0.85 (0.83, 0.87)	0.85 (0.83, 0.87)
PPV_{obs}[c]	0.47 (0.42, 0.51)	0.48 (0.43, 0.52)	0.48 (0.43, 0.53)	0.49 (0.44, 0.54)
NPV_{obs}[c]	0.85 (0.83, 0.87)	0.83 (0.81, 0.84)	0.81 (0.79, 0.83)	0.79 (0.77, 0.81)
PPV_{50}[c]	0.76 (0.72, 0.79)	0.74 (0.71, 0.78)	0.73 (0.70, 0.77)	0.72 (0.68, 0.76)
NPV_{50}[c]	0.61 (0.59, 0.64)	0.60 (0.57, 0.62)	0.59 (0.57, 0.62)	0.58 (0.55, 0.60)

[a] Independent of ICP2 detection

[b] Number positive for *Vc by* culture or qPCR (*tcpA*) as the gold reference standard over total number samples in grouping

[c] Subscript indicates the proportion of samples that were gold-standard positive in the given group for the calculation of each predictive value, with "obs"=observed and "50"=50%. For example, PPV_{50} or standardized positive predictive value indicates the positive predictive value of RDT in each group if 50% of that group was gold-standard positive for cholera.

495Table S2. RDT performance as a function of antibiotic exposure

Exposure:	Reference No antibiotic	Antibiotic (any) Exposure	Ciprofloxacin Reported vs Detected	At effective Concentration[b]	Azithromycin Reported vs Detected	At effective Concentration[b]	Doxycycline/Tetracycline[a] Reported vs Detected	At effective Concent.[b]
Self-reported (all patients)								
Vc+/total(%)[c]	210 / 1122 (18.7)	239 / 956 (25.0)	159 / 568 (28.0)	---	58 / 289 (20.1)	---	---	---
Sensitivity	0.46 (0.39, 0.53)	0.46 (0.40, 0.53)	0.48 (0.40, 0.56)	---	0.43 (0.30, 0.57)	---	---	---
Specificity	0.84 (0.81, 0.86)	0.87 (0.85, 0.90)	0.89 (0.86, 0.92)	---	0.87 (0.82, 0.91)	---	---	---
PPV_{obs}[d]	0.39 (0.33, 0.46)	0.55 (0.48, 0.62)	0.63 (0.54, 0.71)	---	0.45 (0.32, 0.59)	---	---	---
NPV_{obs}[d]	0.87 (0.85, 0.89)	0.83 (0.80, 0.86)	0.81 (0.78, 0.85)	---	0.86 (0.81, 0.90)	---	---	---
PPV_{50}[d]	0.74 (0.69, 0.78)	0.79 (0.73, 0.83)	0.81 (0.75, 0.87)		0.77 (0.66, 0.85)			
NPV_{50}[d]	0.61 (0.57, 0.64)	0.62 (0.58, 0.66)	0.63 (0.58, 0.68)		0.60 (0.53, 0.67)			
Detected by Mass Spectrometry (cholera patients only[e])								
Vc+/total(%)[a]	30 / 32 (93.8)	178 / 220 (80.9)	118 / 148 (79.7)	86 / 116 (74.1)	47 / 68 (69.1)	18 / 32 (56.3)	41 / 45 (91.1)	6 / 7 (85.7)
Sensitivity	0.67 (0.47, 0.83)	0.60 (0.53, 0.67)	0.55 (0.46, 0.64)	0.58 (0.47, 0.69)	0.49 (0.34, 0.64)	0.33 (0.13, 0.59)	0.61 (0.45, 0.7(	0.83 (0.36, 1.00)
Specificity	1.00 (0.16, 1.00)	0.83 (0.69, 0.93)	0.83 (0.65, 0.94)	0.83 (0.65, 0.94)	0.76 (0.53, 0.92)	0.71 (0.42, 0.92)	0.75 (0.19, 0.9!	0.00 (0.00, 0.97)
PPV_{obs}[d]	1.00 (0.83, 1.00)	0.94 (0.88, 0.97)	0.93 (0.84, 0.98)	0.91 (0.80, 0.97)	0.82 (0.63, 0.94)	0.60 (0.26, 0.88)	0.96 (0.80, 1.0(	0.83 (0.36, 1.00)
NPV_{obs}[d]	0.17 (0.02, 0.48)	0.33 (0.24, 0.43)	0.32 (0.22, 0.44)	0.41 (0.29, 0.54)	0.40 (0.25, 0.57)	0.45 (0.24, 0.68)	0.16 (0.03, 0.4(	0.00 (0.00, 0.97)
PPV_{50}[d]	1.00 (0.71, 1.00)	0.78 (0.68, 0.87)	0.77 (0.63, 0.87)	0.78 (0.63, 0.89)	0.67 (0.46, 0.85)	0.54 (0.21, 0.84)	0.71(0.46, 0.8!	0.45 (0.10, 0.85)
NPV_{50}[d]	0.75 (0.52, 0.91)	0.68 (0.59, 0.75)	0.65 (0.55, 0.74)	0.67 (0.55, 0.77)	0.60 (0.44, 0.74)	0.52 (0.30, 0.73)	0.66(0.45, 0.8:	0.00 (0.00, 1.00)

496[a] Mass spectrometry approach taken does not discriminate between tetracycline and doxycycline.
497[b] Concentration at or above the lowest minimal inhibitory concentration (see methods) and prior descriptions[51]
498[c] Reference standard to define positive cholera cases is culture or qPCR ($tcpA$). Approach independent of phage exposure.
499[d] Subscript indicates the proportion of samples that were gold-standard positive in the given group for the calculation of each predictive value, with "obs"=observed and
500"50"=50%. For example, PPV_{50} or standardized positive predictive value indicates the positive predictive value of RDT in each group if 50% of that group was gold-
501standard positive for cholera.
502[e] Patients who met the following criteria had subsequent mass spectrometry: (i) culture positive for Vc, (ii) phage (ICP1,2,3) positive by PCR among culture-negative
503samples, or (iii) Vc positive by PCR among a random 10% of samples negative by culture for Vc and by PCR for phage (see methods).
504
505

Table S3. Among *Vc* positive samples, impact of antibiotic exposure and resistance on RDT positivity (phage removed)[a]

Condition	RDT+	RDT- (false negatives)	OR (CI)	*P* value[b]
Ciprofloxacin (LC-MS/MS qual detection)[c]				
Exposed	30% (47)	22% (34)	0.705 (0.365-1.352)	0.3234
Not exposed	32% (49)	16% (25)		
Azithromycin (LC-MS/MS qual detection)				
Exposed	10% (15)	10% (16)	0.498 (0.223-1.105)	0.099
Not exposed	52% (81)	28% (43)		
Doxycycline (LC-MS/MS qual detection)[c]				
Exposed	12% (19)	7% (11)	1.077 (0.478-2.523)	1.000
Not exposed	50% (77)	31% (48)		
No azithromycin detected by qualitative LC-MS/MS[d]				
Isolate AZI resistant	25% (28)	14% (16)	0.750 (0.337-1.679)	0.54
Isolate AZI sensitive	43% (49)	18% (21)		
Azithromycin detected by qualitative LC-MS/MS[d]				
Isolate AZI resistant	29% (8)	21% (6)	1.333 (0.300 6.100)	1.000
Isolate AZI sensitive	25% (7)	25% (7)		

[a] Criteria to define cholera patients for subsequent mass spectrometry were (i) culture positive for *Vc*, (ii) phage (ICP1,2,3) positive by PCR among culture-negative samples, or (iii) *Vc* positive by PCR among a random 10% of samples negative by culture for *Vc* and by PCR for phage (see methods). Analyses removed samples with ICP1, 2, 3 detected. Those samples without RDT performed were excluded.
[b] Fisher's exact test
[c] Insufficient rates of resistance/sensitive strains for stratified analyses
[d] Among cholera patients in which the culture was positive and an isolate was obtained for AMR testing.

514 **Table S4. RDT performance by severity and age group with culture or qPCR (*tcpA*) as the reference standard**

	Mild (comparator)	Moderate	Severe	Moderate/Severe
0-11months				
Cases/total(%)[a]	25 / 167 (15.0)	13 / 142 (9.2)	4 / 39 (10.3)	17 / 181 (9.4)
Sensitivity	0.16 (0.05, 0.36)	0.38 (0.14, 0.68)	0.25 (0.01, 0.81)	0.35 (0.14, 0.62)
Specificity	0.73 (0.64, 0.80)	0.81 (0.73, 0.87)	0.89 (0.73, 0.97)	0.82 (0.76, 0.88)
PPV_{obs}[b]	0.09 (0.03, 0.22)	0.17 (0.06, 0.35)	0.20 (0.01, 0.72)	0.17 (0.07, 0.34)
NPV_{obs}[b]	0.83 (0.75, 0.89)	0.93 (0.86, 0.97)	0.91 (0.76, 0.98)	0.92 (0.87, 0.96)
PPV_{50}[b]	0.37 (0.21, 0.54)	0.66 (0.50, 0.80)	0.69 (0.27, 0.95)	0.67 (0.52, 0.80)
NPV_{50}[b]	0.46 (0.38, 0.55)	0.57 (0.46, 0.67)	0.54 (0.36, 0.72)	0.56 (0.47, 0.65)
1-4 years				
Cases/total(%)[a]	17 / 125 (13.6)	29 / 202 (14.4)	11 / 74 (14.9)	40 / 276 (14.5)
Sensitivity	0.06 (0.00, 0.29)	0.38 (0.21, 0.58)	0.45 (0.17, 0.77)	0.40 (0.25, 0.57)
Specificity	0.78 (0.69, 0.85)	0.82 (0.76, 0.87)	0.84 (0.73, 0.92)	0.83 (0.77, 0.87)
PPV_{obs}[b]	0.04 (0.00, 0.20)	0.26 (0.14, 0.42)	0.33 (0.12, 0.62)	0.28 (0.17, 0.42)
NPV_{obs}[b]	0.84 (0.75, 0.91)	0.89 (0.83, 0.93)	0.90 (0.79, 0.96)	0.89 (0.84, 0.93)
PPV_{50}[b]	0.21 (0.06, 0.47)	0.68 (0.54, 0.80)	0.74 (0.52, 0.90)	0.70 (0.58, 0.80)
NPV_{50}[b]	0.45 (0.36, 0.55)	0.57 (0.48, 0.65)	0.61 (0.46, 0.74)	0.58 (0.51, 0.65)
5-64 years				
Cases/total(%)[a]	32 / 97 (33.0)	185 / 797 (23.2)	118 / 351 (33.6)	303 / 1148 (26.4)
Sensitivity	0.41 (0.24, 0.59)	0.45 (0.38, 0.52)	0.67 (0.58, 0.75)	0.53 (0.48, 0.59)
Specificity	0.88 (0.77, 0.95)	0.89 (0.86, 0.91)	0.89 (0.85, 0.93)	0.89 (0.87, 0.91)
PPV_{obs}[b]	0.62 (0.38, 0.82)	0.55 (0.46, 0.63)	0.76 (0.67, 0.84)	0.63 (0.57, 0.69)
NPV_{obs}[b]	0.75 (0.64, 0.84)	0.84 (0.81, 0.87)	0.84 (0.79, 0.89)	0.84 (0.82, 0.87)
PPV_{50}[b]	0.77 (0.56, 0.91)	0.80 (0.74, 0.85)	0.86 (0.79, 0.91)	0.83 (0.79, 0.86)
NPV_{50}[b]	0.60 (0.47, 0.71)	0.62 (0.58, 0.66)	0.73 (0.67, 0.79)	0.66 (0.62, 0.69)
65+ years				
Cases/total(%)[a]	1 / 3 (33.3)	9 / 52 (17.3)	5 / 29 (17.2)	14 / 81 (17.3)
Sensitivity	1.00 (0.03, 1.00)	0.22 (0.03, 0.60)	0.40 (0.05, 0.85)	0.29 (0.08, 0.58)
Specificity	1.00 (0.16, 1.00)	0.95 (0.84, 0.99)	0.92 (0.73, 0.99)	0.94 (0.85, 0.98)
PPV_{obs}[b]	1.00 (0.03, 1.00)	0.50 (0.07, 0.93)	0.50 (0.07, 0.93)	0.50 (0.16, 0.84)
NPV_{obs}[b]	1.00 (0.16, 1.00)	0.85 (0.72, 0.94)	0.88 (0.69, 0.97)	0.86 (0.76, 0.93)
PPV_{50}[b]	1.00 (0.09, 1.00)	0.83 (0.39, 0.99)	0.83 (0.39, 0.99)	0.83 (0.54, 0.97)
NPV_{50}[b]	1.00 (0.09, 1.00)	0.55 (0.40, 0.70)	0.60 (0.38, 0.80)	0.57 (0.44, 0.69)

515 [a] Gold standard to define positive cholera cases is culture or qPCR (*tcpA*). Approach does not account for phage or antibiotics
516 exposures.
517 [b.] Subscript indicates the proportion of samples that were gold-standard positive in the given group for the calculation of each predictive value, with
518 "obs"=observed and "50"=50%. For example, PPV_{50} or standardized positive predictive value indicates the positive predictive value of RDT in each
519 group if 50% of that group was gold-standard positive for cholera.

REFERENCES

1.   Ali M, Nelson AR, Lopez AL, Sack DA. Updated global burden of cholera in
endemic countries. *PLoS Negl Trop Dis* 2015; **9**(6): e0003832.

2.   Cholera Upsurge. World Health Organization (WHO). 2024.
https://www.who.int/emergencies/situations/cholera-upsurge (accessed Feb. 20, 2025).

3.   Disease Outbreak News; Cholera - Global situation. World Health Organization
(WHO); 2023. https://www.who.int/emergencies/disease-outbreak-news/item/2023-
DON437 (accessed May 13, 2025).

4.   Global deployment of rapid diagnostic tests to boost fight against cholera. World
Health Organization (WHO). 2024. https://www.who.int/news/item/05-04-2024-global-
deployment-of-rapid-diagnostic-tests-to-boost-fight-against-cholera (accessed Feb 20,
2025).

5.   Hegde ST, Khan AI, Perez-Saez J, et al. Clinical surveillance systems obscure
the true cholera infection burden in an endemic region. *Nat Med* 2024; **30**(3): 888-95.

6.   Xu H, Zou K, Dent J, et al. Enhanced cholera surveillance to improve vaccination
campaign efficiency. *Nat Med* 2024; **30**(4): 1104-10.

7.   Nelson EJ, Harris JB, Morris JG, Jr., Calderwood SB, Camilli A. Cholera
transmission: the host, pathogen and bacteriophage dynamic. *Nat Rev Microbiol* 2009;
7(10): 693-702.

8.   Nelson EJ, Chowdhury A, Flynn J, et al. Transmission of *Vibrio cholerae* is
antagonized by lytic phage and entry into the aquatic environment. *PLoS Pathog* 2008;
4(10): e1000187.

9. Nelson EJ, Grembi JA, Chao DL, et al. Gold Standard Cholera Diagnostics Are

Tarnished by Lytic Bacteriophage and Antibiotics. *J Clin Microbiol* 2020; **58**(9).

10. Madi N, Cato ET, Abu Sayeed M, et al. Phage predation, disease severity, and

pathogen genetic diversity in cholera patients. *Science* 2024; **384**(6693): eadj3166.

11. Nelson EJ, Nelson DS, Salam MA, Sack DA. Antibiotics for both moderate and

severe cholera. *N Engl J Med* 2011; **364**(1): 5-7.

12. Islam MR. Single dose tetracycline in cholera. *Gut* 1987; **28**(8): 1029-32.

13. Gotuzzo E, Seas C, Echevarria J, Carrillo C, Mostorino R, Ruiz R. Ciprofloxacin

for the treatment of cholera: a randomized, double-blind, controlled clinical trial of a

single daily dose in Peruvian adults. *Clin Infect Dis* 1995; **20**(6): 1485-90.

14. Use of antibiotics for the treatment and control of cholera. World Health

Organization (WHO) Global Task Force on Cholera Control (GTFCC). 2018.

https://www.gtfcc.org/wp-content/uploads/2019/10/gtfcc-technical-note-on-use-of-

antibiotics-for-the-treatment-of-cholera.pdf (accessed Feb 18, 2025).

15. Alexandrova L, Haque F, Rodriguez P, et al. Identification of Widespread

Antibiotic Exposure in Patients With Cholera Correlates With Clinically Relevant

Microbiota Changes. *J Infect Dis* 2019; **220**(10): 1655-66.

16. Saber MM, Shapiro BJ. Benchmarking bacterial genome-wide association study

methods using simulated genomes and phenotypes. *Microb Genom* 2020; **6**(3).

17. Rouard C, Collet L, Njamkepo E, et al. Long-Distance Spread of a Highly Drug-

Resistant Epidemic Cholera Strain. *N Engl J Med* 2024; **391**(23): 2271-3.

18.   Lassalle F, Al-Shalali S, Al-Hakimi M, et al. Genomic epidemiology reveals

multidrug resistant plasmid spread between Vibrio cholerae lineages in Yemen. *Nat*

Microbiol 2023; **8**(10): 1787-98.

19.   Seed KD, Yen M, Shapiro BJ, et al. Evolutionary consequences of intra-patient

phage predation on microbial populations. *Elife* 2014; **3**: e03497.

20.   Seed KD, Bodi KL, Kropinski AM, et al. Evidence of a dominant lineage of *Vibrio*

cholerae-specific lytic bacteriophages shed by cholera patients over a 10-year period in

Dhaka, Bangladesh. *MBio* 2011; **2**(1): e00334-10.

21.   Seed KD, Faruque SM, Mekalanos JJ, Calderwood SB, Qadri F, Camilli A.

Phase variable O antigen biosynthetic genes control expression of the major protective

antigen and bacteriophage receptor in *Vibrio cholerae* O1. *PLoS Pathog* 2012; **8**(9):

e1002917.

22.   Yen M, Cairns LS, Camilli A. A cocktail of three virulent bacteriophages prevents

Vibrio cholerae infection in animal models. *Nat Commun* 2017; **8**: 14187.

23.   Beckman DA, Waters CM. *Vibrio cholerae* phage ICP3 requires O1 antigen for

infection. *Infect Immun* 2023; **91**(9): e0002623.

24.   Patel KM, Seed KD. Sporadic phage defense in epidemic *Vibrio cholerae*

mediated by the toxin-antitoxin system DarTG is countered by a phage-encoded

antitoxin mimic. *mBio* 2024; **15**(10): e0011124.

25.   Oshiro RT, Dunham DT, Seed KD. The vibriophage-encoded inhibitor OrbA

abrogates BREX-mediated defense through the ATPase BrxC. *J Bacteriol* 2024;

206(11): e0020624.

26.   LeGault K, Hays SG, Angermeyer A, et al. Temporal shifts in antibiotic resistance

elements govern phage-pathogen conflicts. *Science* 2021; **373**(6554):

2020.12.16.423150.

27.   Seed KD, Lazinski DW, Calderwood SB, Camilli A. A bacteriophage encodes its

own CRISPR/Cas adaptive response to evade host innate immunity. *Nature* 2013;

494(7438): 489-91.

28.   Brenzinger S, Airoldi M, Ogunleye AJ, Jugovic K, Amstalden MK, Brochado AR.

The Vibrio cholerae CBASS phage defence system modulates resistance and killing by

antifolate antibiotics. *Nat Microbiol* 2024; **9**(1): 251-62.

29.   McKitterick AC, Seed KD. Anti-phage islands force their target phage to directly

mediate island excision and spread. *Nat Commun* 2018; **9**(1): 2348.

30.   Jensen MA, Faruque SM, Mekalanos JJ, Levin BR. Modeling the role of

bacteriophage in the control of cholera outbreaks. *Proc Natl Acad Sci U S A* 2006;

103(12): 4652-7.

31.   Faruque SM, Naser IB, Islam MJ, et al. Seasonal epidemics of cholera inversely

correlate with the prevalence of environmental cholera phages. *Proc Natl Acad Sci U S*

A 2005; **102**(5): 1702-7.

32.   Faruque SM, Islam MJ, Ahmad QS, et al. Self-limiting nature of seasonal cholera

epidemics: Role of host-mediated amplification of phage. *Proc Natl Acad Sci U S A*

2005; **102**(17): 6119-24.

33.   King AA, Ionides EL, Pascual M, Bouma MJ. Inapparent infections and cholera

dynamics. *Nature* 2008; **454**(7206): 877-80.

34. Pasricha CL, Monte AJ, Gupta SK. Season variations of cholera bacteriophage in

natural waters and in man in Calcutta during the year 1930. *Indian Med Gaz* 1931; **66**:

543-50.

35. Silva-Valenzuela CA, Camilli A. Niche adaptation limits bacteriophage predation

of Vibrio cholerae in a nutrient-poor aquatic environment. *Proc Natl Acad Sci U S A*

2019; **116**(5): 1627-32.

36. Page AL, Alberti KP, Mondonge V, Rauzier J, Quilici ML, Guerin PJ. Evaluation

of a rapid test for the diagnosis of cholera in the absence of a gold standard. *PLoS One*

2012; **7**(5): e37360.

37. Dick MH, Guillerm M, Moussy F, Chaignat CL. Review of two decades of cholera

diagnostics--how far have we really come? *PLoS Negl Trop Dis* 2012; **6**(10): e1845.

38. Sayeed MA, Paisie T, Alam MT, et al. Development of a Monoclonal Antibody to

a Vibriophage as a Proxy for Vibrio cholerae Detection. *Infect Immun* 2022; **90**(8):

e0016122.

39. Sayeed MA, Islam K, Hossain M, et al. Development of a new dipstick (Cholkit)

for rapid detection of Vibrio cholerae O1 in acute watery diarrhoeal stools. *PLoS Negl*

Trop Dis 2018; **12**(3): e0006286.

40. George CM, Rashid MU, Sack DA, et al. Evaluation of enrichment method for the

detection of *Vibrio cholerae* O1 using a rapid dipstick test in Bangladesh. *Trop Med Int*

Health 2014; **19**(3): 301-7.

41. Debes AK, Ateudjieu J, Guenou E, et al. Clinical and Environmental Surveillance

for Vibrio cholerae in Resource Constrained Areas: Application During a 1-Year

Surveillance in the Far North Region of Cameroon. *Am J Trop Med Hyg* 2016; **94**(3):

537-43.

42. Ley B, Khatib AM, Thriemer K, et al. Evaluation of a rapid dipstick (Crystal VC)

for the diagnosis of cholera in Zanzibar and a comparison with previous studies. *PLoS*

One 2012; **7**(5): e36930.

43. Gupta P, Mankere B, Chekkoora Keloth S, Tuteja U, Pandey P, Chelvam KT.

Increased antibiotic resistance exhibited by the biofilm of *Vibrio cholerae* O139. *J*

Antimicrob Chemother 2018; **73**(7): 1841-7.

44. Perez-Saez J, Bhuiyan TR, Hegde ST, et al. Rethinking cholera diagnostic test

performance, interpretation and evaluation: a field-based latent-class analysis in

Bangladesh. *medRxiv* 2024: 2024.11.19.24317512.

45. The treatment of diarrhoea -- A manual for physicians and other senior health

workers. 4th Rev. ed. Geneva, Switzerland: World Heath Organization. 2005.

46. Sayeed MA, Nabil IK, Bhattacharjee P, et al. Assembly and performance of a

cholera RDT prototype that detects both Vibrio cholerae and associated bacteriophage

as a proxy for pathogen detection. *J Clin Microbiol* 2025; **63**(2): e0144324.

47. Khan AI, Mack JA, Salimuzzaman M, et al. Electronic decision support and

diarrhoeal disease guideline adherence (mHDM): a cluster randomised controlled trial.

Lancet Digit Health 2020; **2**(5): e250-e8.

48. Khan AI, Qadri F. Epidemiology of cholera in Bangladesh: Findings from

Nationwide Hospital-based Surveillance, 2014-2018. *CID* 2019.

49.   Li D, Liu CM, Luo R, Sadakane K, Lam TW. MEGAHIT: an ultra-fast single-node

solution for large and complex metagenomics assembly via succinct de Bruijn graph.

Bioinformatics 2015; **31**(10): 1674-6.

50.   Manual of Clinical Microbiology. 8th ed. Washington, D.C.: American Society for

Microbiology Press; 2003.

51.   Creasy-Marrazzo A, Saber MM, Kamat M, et al. Genome-wide association

studies reveal distinct genetic correlates and increased heritability of antimicrobial

resistance in *Vibrio cholerae* under anaerobic conditions. *Microb Genom* 2022; **8**(12).

52.   Islam MT, Khan AI, Sayeed MA, et al. Field evaluation of a locally produced rapid

diagnostic test for early detection of cholera in Bangladesh. *PLoS Negl Trop Dis* 2019;

13(1): e0007124.

53.   Team RC. R: A language and environment for statistical computing. R

Foundation for Statistical Computing. 2021. http://www.R-project.org/ (accessed May

26, 2019.

54.   J L. Eulerr: Area-Proportional Euler and Venn Diagrams with Ellipses. R

package. 2024. https://CRAN.R-project.org/package=eulerr>.

55.   M S, E S. EpiR: Tools for the Analysis of Epidemiological Data. R package

version 2.0.78. <https://CRAN.R-project.org/package=epiR> (accessed March 16,

2025).

56.   Levine AC, Gainey M, Qu K, et al. A comparison of the NIRUDAK models and

WHO algorithm for dehydration assessment in older children and adults with acute

diarrhoea: a prospective, observational study. *Lancet Glob Health* 2023; **11**(11): e1725-

e33.