Efficacy and feasibility of synergy-based multichannel functional electrical stimulation for

chronic stroke gait rehabilitation: a pilot study

Levine JT[1,2], Yu XS[1,3], Muñoz R[2], Fiorenza A[2], Smith T[5], Djuraskovic I[2,4], Peiffer JD[1,2],

Ambrosini E[7], Ferrante S[7], Webster R[5], Sakai J[5], Robison J[5], Roth EJ[2,3], Laczko J[2,6], Cotton

RJ[1,2,3], Pedrocchi A[7], Pons JL[1,2,3]

Affiliations:

1. Northwestern University, McCormick School of Engineering, Evanston, IL USA

8   2. Shirley Ryan AbilityLab, Chicago, IL USA

9   3. Northwestern University, Department of Physical Medicine and Rehabilitation, Chicago,

10   IL USA

11   4. Northwestern University, Department of Neuroscience, Evanston, IL USA

12   5. CIONIC Inc, San Francisco, CA USA

13   6. Pazmany Peter Catholic University, Faculty of Information Technology and Bionics,

14   Budapest, Hungary

15   7. Politecnico Di Milano, Department of Electronics, Milan, Italy

Corresponding Author: Jackson T. Levine levinej@u.northwestern.edu

Disclosures: Robison J, Webster R, Smith T, and Sakai J are employed by CIONIC Inc. and may

hold equity in the company. These individuals contributed to device design and provided technical

support but were not involved in the study design, participant selection, therapist or participant

training, outcome assessments or data analysis, in order to minimize potential bias.

Author Contributions: L.J.T., M.R. and P.J.L conceived and designed research; R.J., W.R.,

S.T., and S.J. modified device design and provided technical support; L.J.T., X.S.Y., M.R. and

F.A. performed experiments; L.J.T., X.S.Y., M.R., F.A., D.I. and P. J.D. analyzed data; L.J.T.,

X.S.Y., M.R., and F.A. interpreted results of experiments; L.J.T. prepared figures; L.J.T. drafted

manuscript; X.S.Y., M.R., F.A., A.E., F.S., L.J..,C.R.J., P.A., and P.J.L. edited and revised

manuscript; All authors read and approved the final manuscript

Abstract

Introduction: Chronic stroke gait disorders are characterized by impaired motor coordination

(i.e., muscle synergies), which are related to functional impairment. While high-intensity gait

training (HIGT) is the standard of care, and functional electrical stimulation (FES) to the tibialis

anterior can address foot drop, evidence has shown that extending FES to multiple muscles

improves functional and biomechanical outcomes. The Cionic Neural Sleeve provides a wearable

platform that allows customizable stimulation to muscles of the leg with clinically feasible time to

prepare the device and a user-friendly interface. We investigated the efficacy and clinical

implementation of a personalized intervention based on the individual's specific motor

coordination impairment–muscle synergy-based multichannel FES (MFES)–paired with HIGT to

improve gait speed, endurance, biomechanics, and similarity of synergies to normative data.

Methods: Fourteen individuals (six females) with chronic stroke gait impairments were recruited

and randomly divided into two groups, both receiving six weeks of either HIGT or MFES+HIGT.

Assessments of gait speed, endurance, balance, gait biomechanics, and muscle synergies were

conducted at baseline, midpoint, post, and one month following completion of training. Feasibility

was assessed with time to prepare the system and participant and therapist questionnaires. **Results:**

Despite a small sample size, only the MFES+HIGT group had sustained improvements in fast

walking speed and endurance. Additionally, while both groups improved in impaired limb step

length, MFES+HIGT led to reduced step width while ankle dorsiflexion actually got worse after

HIGT. Finally, average similarity of muscle synergy activation profiles to normative data,

measured through circular cross correlation, improved only for MFES+HIGT. Time to prepare the

device plateaued at nine sessions at 4.53 minutes, and both participants and therapists rated the

intervention highly feasible, acceptable, and usable. **Discussion:** Muscle synergy-based

MFES+HIGT demonstrated improved outcomes in gait speed, endurance, biomechanics, and

motor coordination with positive feedback on a highly usable and feasible stimulation interface.

This research supports further pursuit of this personalized intervention through a randomized

controlled clinical trial with larger sample sizes that can better characterize how individuals with

varying baseline function respond to synergy-based MFES+HIGT.

Keywords: muscle synergies, stroke, gait, rehabilitation, functional electrical stimulation,

personalized rehabilitation

Introduction

Among stroke survivors, gait disorders are one of the most common and debilitating outcomes,

affecting up to 80% of survivors and severely impairing mobility and independence.(1) Post-stroke

gait is well characterized by impaired motor coordination, which can be quantified through muscle

synergy analysis–a mathematical dimensionality reduction technique (e.g., nonnegative matrix

factorization, principal component analysis, etc.) to determine from multi-channel

electromyography (EMG) the sources of common input to multiple muscles during a complex,

multi-joint task. Muscle synergies during gait of neurologically intact individuals are highly

consistent across individuals, terrains (e.g., treadmill, overground, etc.), speeds, and inclines, and

typically consist of four synergies with clear biomechanical functions: weight acceptance, knee

flexion, propulsion, foot clearance.(2-4) After a stroke, a reduction in the complexity of the

synergies (i.e., number of synergies required to explain muscle activity) and similarity of muscle

synergies (i.e., muscle weights and activation profiles) with respect to normative data from

neurologically intact individuals are highly related to functional impairment metrics, including

walking speed, step length, and step width.(5, 6) Furthermore, improvement in gait spatiotemporal

metrics after gait rehabilitation is accompanied by changing muscle synergy patterns in individuals

with chronic stroke,(6, 7) suggesting that not only is it possible to alter muscle coordination after

the neuroplastic window is closed but neuroplastic changes may contribute to changes in gait

outcomes and biomechanics.

When paired with task-oriented movement intent, functional electrical stimulation (FES) can

induce peripheral benefits(8) and enhance central nervous system neuroplasticity(9-13) by

providing timed muscle activation and sensory feedback and helping to reshape neural control

through repetitive training.(14, 15) FES is commonly used to stimulate the tibialis anterior

(TA)/peroneal nerve during swing phase to address foot drop, one of the major gait impairments

post-stroke.(16, 17) However, FES to a single muscle has demonstrated limited improvement over

high intensity gait training (HIGT)–the standard of care–for improving chronic stroke gait.(18)

Conversely, extending stimulation to multiple muscles of the lower limb can target a greater

breadth of motor impairments that collectively impair gait. Several studies have investigated

additional stimulation to the gastrocnemius (GAS),(19, 20) gluteus medius (GM),(21, 22) and

quadriceps/hamstrings,(23, 24) showing improvements in motor outcomes. We hypothesize that

an intervention stimulating all of the key muscles for gait may further improve outcomes, but the

challenge is to identify an evidence-based approach to objectively determine when to stimulate

each muscle and to develop an easy-to-learn and quick implementation of multi-channel,

customizable FES.

Recently, our team pioneered a novel, personalized intervention: muscle synergy-based

multichannel functional electrical stimulation (MFES) gait rehabilitation for chronic stroke.(25)

This technique quantifies motor impairment of each individual post-stroke by identifying which

synergies are impaired and designing a personalized stimulation strategy of each muscle based

only on the impaired synergies. Additionally, the muscle synergy-based stimulation technique has

been shown to reduce FES-induced fatigue in the upper limb compared to other MFES stimulation

techniques.(26) Our team(25) and other colleagues(27) have developed pilot implementations of

the novel intervention on two individuals post-stroke each in a four-week training protocol and

found improved coordination and gait outcomes. These preliminary findings underscore the need

for further exploration of its clinical implementation with a comprehensive evaluation of

effectiveness through clinical, biomechanical, and neuromuscular assessments. CIONIC, Inc. has

developed the Cionic Neural Sleeve, a FDA-approved MFES wireless sleeve that can record and

stimulate from the major muscle groups of the leg and is currently approved for home use in

individuals with upper motor neuron injuries.(28, 29) Setup and use are streamlined through a

user-friendly application interface that allows users to select muscle targets and fine-tune

stimulation parameters—including amplitude, pulse width, and frequency—for each muscle

independently. The Cionic Neural Sleeve has been shown to improve ankle dorsiflexion in

individuals with foot drop, based on a study involving 32 participants with various neurological

impairments.(30) The study also reported improvements in quality of life, along with high

feasibility and safety.(31)

In this study, we aim to investigate the efficacy of a six-week synergy-based MFES combined with

HIGT protocol to improve functional gait outcomes, biomechanics, and muscle coordination

compared to HIGT, the standard of care, in chronic stroke survivors. Furthermore, we intend to

implement this novel, personalized intervention in the clinical setting via the Cionic Neural Sleeve that was adapted for this study. We assessed clinical feasibility by tracking the time required to don the device and configure stimulation parameters, and by collecting feedback from both participants and therapists experienced in inpatient stroke gait rehabilitation. We hypothesized that improvements in gait outcomes would correspond with biomechanical and muscle coordination changes resembling those of neurologically intact individuals. Additionally, we expected that, following therapist familiarization, preparation times would support clinical feasibility and that the intervention would be rated as usable, acceptable, and appropriate by both participants and clinicians.

Methods

Device

This study used the Cionic Neural Sleeve (Figure 1C), an FDA-approved device (510(k) number K21362) intended to provide ankle dorsiflexion and/or plantarflexion in adult individuals with foot drop and/or to assist knee flexion or extension in adult individuals with muscle weakness related to upper motor neuron disease/injury (e.g. stroke, damage to pathways to the spinal cord).(28, 29) Available in multiple sizes to accommodate different body types, the sleeve integrates recording and biphasic stimulation electrodes targeting the tibialis anterior (TA), gastrocnemius (GAS), rectus femoris (RF), vastus lateralis (VL), semitendinosus (ST), and biceps femoris (BF). It also includes two embedded inertial measurement units (IMUs) on the shank and thigh. To enable stimulation of the gluteus medius (GM)—a key muscle for gait not covered by the standard sleeve—Cionic custom-developed an external hip unit for this study, equipped with an IMU and EMG/stimulation electrodes (depicted in right panel of Figure 1C). EMG signals were sampled at 2000 Hz and IMU data at 100 Hz. The study protocol was approved by the Northwestern

University Institutional Review Board (STU00218244) and conducted in accordance with the

Declaration of Helsinki.

Participant Training

Population

Fourteen individuals with a history of a single, unilateral, supratentorial stroke of either ischemic

or hemorrhagic etiology occurring at least six months prior were recruited for this study.

Individuals were eligible if they experienced reduced walking ability, but could still walk

independently for at least ten meters with or without an assistive device. Exclusion criteria

included other neurological conditions or peripheral nerve injuries, comorbidities impacting gait

(e.g., severe arthritis or recent fracture), significant spasticity (greater than or equal to 3 on the

Modified Ashworth Scale), contraindications to FES, and current enrollment in other lower limb

therapy or research studies. Of the eighteen individuals recruited for the study, three were

unenrolled due to not meeting criteria upon in-person assessment and one dropped out. Prior to the

initial visit, participants were randomly assigned to one of two groups: 1) HIGT or 2) HIGT in

combination with synergy-based MFES (MFES+HIGT). All participants provided informed

consent at the initial visit and then completed a baseline assessment. See Table 1 for general

characteristics of participants. Three levels of ambulation are reported based on the self-selected

walking speed during the 10MWT at baseline: household (<0.4 m/s), limited community (0.4-0.8

m/s), and community (>0.8 m/s).(32)

To assess clinical feasibility, four physical therapists from Shirley Ryan AbilityLab were recruited

to deliver the MFES+HIGT training protocol. All therapists provided informed consent and

completed two hours of training on both the rehabilitation protocol and use of the Cionic Neural

Sleeve prior to being paired with a post-stroke participant for the six-week intervention. On

average, therapists had 4.5 years of experience in post-stroke gait rehabilitation and were familiar

with single-channel FES for foot drop (primarily using the Bioness device to stimulate the TA).

Intervention

Participants in both groups completed three sessions per week for six weeks, for a total of eighteen

training sessions. During training, participants ambulated on a treadmill with a harness donned for

safety only (no external support provided), per recommendations by the clinical practice

guidelines.(33) They were allowed to use upper extremity support on the handrails of the treadmill

if needed. Training consisted of a five-minute warmup period, 20 minutes of HIGT with or without

synergy-based MFES, and a five-minute cool down. During warmup and cool down, participants

walked at their self-selected velocity (SSV), which was determined during the 10 Meter Walk Test

(10MWT) at their baseline assessment. In both groups, therapists were instructed to challenge

participants during the training session to reach 70-85% of their maximum heart rate (HR),(33)

which was measured using a Polar OH1 sensor over the wrist extensors or biceps for real time

feedback.(34, 35) Maximum HR was calculated using the following equation(36):

$$Max\ HR = 208 - 0.7 * (Age)$$

If the participant was on beta-blockers, the HR range was reduced by 15 bpm to account for the

effects of the medication. Therapists could modulate speed and incline to increasingly challenge

the participants.(37-39) For those unable to achieve the required intensity in their HR range (N=2),

the Borg Rating of Perceived Exertion (RPE) scale,(40) which ranges from 6-20, was used to

monitor intensity with a goal RPE rating of 15 to 17 ("hard" to "very hard").

For participants in the MFES+HIGT group, therapists first donned the Cionic Neural Sleeve and

the additional external hip unit on participants' impaired limb. For each muscle, the maximum

comfortable stimulation amplitude that produced a strong visible contraction was determined

individually (range: 15–100 mA), while stimulation frequency and pulse width were standardized

at 30 Hz and 300 μs, respectively. Stimulation amplitude could be modified throughout training

by the therapist as necessary. Time taken to don the device and set stimulation parameters was

recorded during each session to assess clinical feasibility. Parameters set at the initial training visit

were saved, and thus only tested and modified if needed at subsequent visits. Participants received

personalized stimulation (see next section) only during the 20-minute high-intensity block of

training.

Personalized Stimulation Control

The personalized stimulation control has been designed as described in our previous work,(25)

validating the concept (depicted in Figure 1A with examples in 1B).

Participant EMG Processing and Muscle Synergies

Participant EMG was recorded at baseline during treadmill walking at SSV, which was determined

during the baseline 10MWT. Signals were bandpass filtered (3rd order, 40-400 Hz cutoff,

Butterworth filter), rectified, and low pass filtered (3rd order, 5 Hz cutoff, Butterworth). They were

then segmented into gait cycles using peak sagittal shank IMU angle as a proxy for heel strike.

Segmented EMG signals for each muscle were interpolated from 0 to 100% of the gait cycle and

normalized to the peak to their peak-to-peak amplitude. An average EMG profile was computed

for each muscle across all acceptable strides (Figure 1A Step 1).

Normative muscle synergy weights (W) and activation profiles (H) were computed as the average

of fourteen neurologically intact individuals (age: 46.9±17.0) walking at ten speeds ranging from

0.45 m/s to 1.8 m/s for one minute each using nonnegative matrix factorization (NNMF) and a

variance accounted for threshold of 0.9 to define the number of synergies (four synergies were

extracted on average).(41) To ensure accurate comparison between participant and normative

synergies—and to prevent incorrect matching between non-corresponding components—a

modified nonnegative matrix reconstruction algorithm was used.(25) The average normative

weights (W_{Norm}) and activation profiles (H_{Norm}) were used in combination with the average

participant EMG (M_{Pt}, 7 by 101 matrix) to perform two non-negative matrix reconstructions to

obtain the participant weights (W_{Pt}) and activation profiles (H_{Pt}). Participant weights (W_{Pt}) were

computed by fixing the normative activation profiles (H_{Norm}) and only allowing the participant

weights to be updated in each iteration i of the algorithm below:

213
$$W_{Pt}^{i+1} = W_{Pt}^{i} \frac{(M_{Pt'} \cdot H_{Norm}^{T})}{(W_{Pt}^{i} \cdot H_{Norm} \cdot H_{Norm}^{T})}$$

214   Similarly, participant activation profiles (H_{Pt}) were computed by fixing the normative synergy

215   weights (W_{Norm}) and only allowing the participant activation profiles to be updated in each

216   iteration of the algorithm below:

217
$$H_{Pt}^{i+1} = H_{Pt}^{i} \frac{(W_{Norm}^{T} \cdot M_{Pt'})}{(W_{Norm}^{T} \cdot W_{Norm} \cdot H_{Pt}^{i})}$$

Both iterative algorithms were run together for a minimum of 30 and maximum of 300 iterations.

If the reduction in error from the reconstructed EMG using W_{Pt} and H_{Pt} and the average participant

EMG did not exceed 10^{-5} and the minimum number of iterations was reached, the iterative

algorithm was concluded. Once the iterative algorithm was complete, the unitary norm of

participant synergies was taken for comparison with normative synergies (Figure 1A Step 2).

Identification of Impaired Participant Synergies

Participant synergies were compared to average normative synergies, and the degree of deviation

was quantified using three metrics: cosine similarity of weights **W**, circular cross correlation

coefficient of activation profiles **H**, and circular cross correlation time lag of activation profiles

H.(25) The time lag was computed as $Tlag = 1 - \left|\frac{lag}{100}\right|$, where *lag* represents a percentage of the

gait cycle (value between -50 and 50%). The mean and standard deviation for all three metrics of

muscle synergies of neurologically intact individuals with the average normative synergies were

first computed. For each participant synergy, comparisons to normative averages were computed,

and a synergy was classified as impaired if any of the three metrics deviated by more than two

standard deviations from the normative average (Figure 1A Step 3).

Personalized Participant Stimulation Profiles

Only the impaired synergies were used to drive the personalized stimulation profile development.

The normative weights and activation profiles of synergies that were deemed impaired for each

participant were used for nonnegative matrix reconstruction to compute a stimulation profile for

each muscle (*m*) according to the following equation:

$$Stimulation_m = \sum_{i=1}^{n} W_{Norm,i} * H_{Norm,i}$$

where **n** is the set of impaired synergies and **i** is the specific impaired synergy. The stimulation

profile was then normalized such that the minimum value was set to zero and maximum value was

set to one. To prevent continuous stimulation throughout the gait cycle and minimize fatigue, only

stimulation values exceeding 0.2 were included in the stimulation profile. These values were then

scaled to the maximum stimulation amplitude set by the therapist through the interface (Figure 1A

Step 4).

Gait Event Detection

The duration of each gait cycle was continuously updated using an exponential moving average

based on the previous ten cycles. Because participants may experience prolonged phases within

the gait cycle, relying solely on heel strike to trigger stimulation could lead to inappropriately

timed stimulation. To address this, key gait events were identified in the fourteen neurologically

intact individuals using sagittal plane IMU angles from the shank and thigh. These events and their

average timing within the gait cycle were: peak shank (0/100%), negative zero crossing (positive

to negative) shank (28%), minimum thigh (57%), positive zero crossing (negative to positive) thigh

(68%), minimum shank (73%), and positive zero crossing (negative to positive) shank (91%). Gait

events that occurred nearest in time (% gait cycle) to the desired onset of the stimulation ramp-up

and ramp-down phases were identified from each participant's gait cycle. These events were used

to trigger stimulation onset of ramp-up and ramp-down, respectively, with brief delays applied to

each to ensure alignment with the personalized stimulation timing profile (Figure 1A Step 5).

A. Steps for Personalized Stimulation

1. At baseline, record EMG and compute *participant synergies* with nonnegative matrix factorization iterative algorithm

2. Compare *participant synergies* to *average normative synergies* from neurologically intact individuals

3. Identify *impaired* participant synergies with *cosine similarity* and *circular cross correlation*

4. Nonnegative matrix reconstruction of each muscle with *only impaired synergies*

5. Provide *precisely timed* stimulation based on reconstructed activation profiles during *HIGT* block

B. Example Personalized Stimulation

C. Cionic Neural Sleeve

Figure 1: Personalized Synergy-Based Stimulation. A) Steps for identifying impaired synergies,

computing personalized stimulation based on the identified impaired synergies, and precisely

timing stimulation. B) Example of personalized stimulation profiles for two participants with

varying baseline synergy impairments. C) Cionic Neural Sleeve with electrode placement, sleeve

fit, and the application interface with the ability to modify frequency, pulse width, and intensity

(current). The placement of the external hip unit is also displayed, with labels for the stimulation

electrodes and sensor for recording electromyography (EMG) and inertial measurement unit

(IMU) signals. GM = Gluteus Medius, GAS = Gastrocnemius, TA = Tibialis Anterior, VL =

Vastus Lateralis, RF = Rectus Femoris, BF = Biceps Femoris, ST = Semitendinosus.

Assessment Visits and Outcome Measures

Assessment Visit

Assessments were performed at baseline, midpoint (after completing nine training sessions), post

(after completing all training sessions), and one month after completing the training. The baseline

assessment was completed one week prior to the beginning of training. During this week, the

participant wore a step-counting device (ActiGraph, USA)(42-44) to track average daily number

of steps taken as a measure of participation. This was repeated immediately following completion

of all training sessions.

At each assessment point, participants completed three clinical assessments: the 10-Meter Walk

Test (10MWT) at both SSV and fast velocity (FV), the 6-Minute Walk Test (6MWT), and the

Berg Balance Scale (BBS). The 10MWT was repeated twice at both speeds while participants were

recorded using a markerless motion capture system developed by our team,(45) enabling extraction

of spatiotemporal gait parameters and joint kinematics. For both the 10MWT and 6MWT,

participants were permitted to use the same assistive device for all assessment points (e.g., cane or

walker), but ankle-foot orthoses were not allowed to assess intrinsic ankle motor control

impairments.

Following completion of the clinical assessments, at each assessment visit, all participants donned

the Neural Sleeve on the affected limb with the addition of the external hip sensors. Following a

two-minute warm-up, participants walked on a treadmill for two minutes at their SSV while the

Neural Sleeve recorded IMU angular data and EMG from the same muscles used for stimulation.

No stimulation was administered during assessment visits.

Gait Biomechanics Outcome Measures

Markerless motion capture data taken during 10MWT SSV was processed according to Cotton et

al.(45, 46) Videos were collected from multiple calibrated and synchronized cameras. Participant

tracking using EasyMocap(47) and 2D joint location estimation using MeTRAbs-ACAE(48) was

facilitated using PosePipe.(49) We used an end-to-end reconstruction approach to jointly learn the

body scale parameters and joint angles of a differentiable biomechanical model most consistent

with video observations.(46) From these biomechanical reconstructions, we extracted gait event

timings(50, 51) and calculated the following spatiotemporal measures: step length, step width, and

stance percent. Kinematic analysis was focused on joint movements and phases specific to stroke

impairment on the impaired limb: peak ankle dorsiflexion during swing (foot clearance), peak hip

abduction during swing (hip circumduction compensatory mechanism for impaired foot

clearance), peak hip extension during stance (propulsion), peak knee extension during stance

(stance stability and hyperextension), and knee flexion during swing (limb advancement and

clearance).

Muscle Synergy Outcome Measures

Cosine similarity and circular cross correlation coefficient and lag were assessed for the impaired

limb compared to normative synergies. These measures were computed for each synergy, and the

average among synergies was reported. Additionally, the number of impaired synergies was

reported.

Feasibility Outcome Measures

To obtain participant feedback on the experimental intervention, participants in both groups

completed the Patient Perceived Usefulness Questionnaire*,* derived from the Global Rating of

Change Scale,(52) with the prompt "*With respect to your walking ability, how would you describe*

yourself now compared to before starting your rehabilitation program?" at the follow-up

assessment via a scale from -5 (Much Worse) to 5 (Completely Recovered). Participants also had

the opportunity to provide written feedback and any comments on how the training affected their

outside life at the follow-up assessment.

To assess clinical feasibility, the time required to don the Cionic Neural Sleeve and set stimulation

parameters was recorded for each session. To assess the ability of therapists to challenge their

participants to achieve the requirements of HIGT, mean HR, peak HR, percentage of time spent in

the target HR zone, and RPE were evaluated for each session. Therapists also completed a

questionnaire, the Therapist Perceived Usefulness Questionnaire, derived from the Global Rating

of Change Scale,(52) with the prompt "*With respect to the rehabilitation program used for this*

patient, how likely would you be to integrate this into your clinical practice?", ranging from -5

(Very Unlikely) to 5 (Very Likely). This form also provided space to explain their choices and

provide any additional feedback. Additionally, therapists completed two more surveys: Brooke

System Usability Scale(53) and Weiner Acceptability, Appropriateness and Feasibility of

Intervention Measure.(54) All therapist forms were collected after completing at least six training

sessions.

Statistics

SPSS 26.0.0.1 was used for statistical analysis. To compare participant demographics between

groups, a linear mixed model was used with group (HIGT, MFES+HIGT) as a fixed effect. To

ensure that the training sessions were equally challenging between groups, measures of heart rate

and RPE were compared using a linear mixed model with Session Number and Group as effects

and Participant, Therapist, Ambulation Level, Age, and Years Since Stroke as random effects.

Because of the small and variable sample collected in this study, statistical analysis of training

effects included a within group (HIGT, MFES+HIGT) linear mixed effects model to account for

changes over time (pre, mid, post, follow-up) as a result of each type of training and a between

group linear mixed effects model to compare the change from baseline between groups at each

assessment point (mid, post, follow-up). Random effects included Participant, Ambulation Level

(household, limited community, community), Age and Years Since Stroke. A significance value

of 0.05 was used for all of our models. Simple linear regression was used to assess whether muscle

synergy-related metrics (cosine similarity, cross correlation coefficient, cross correlation lag)

predicted clinical outcomes at baseline (10MWT SSV, 10MWT FV, 6MWT). A linear plateau

model was applied in RStudio 2025.05.0 to determine the inflection point at which setup time

plateaued. Setup durations from 57 total sessions conducted by four physical therapists (PT1: 13

sessions, PT2: 15 sessions, PT3: 15 sessions, PT4: 14 sessions) were analyzed. Assumptions of

linearity, independence, homoscedasticity, normality of residuals, and absence of multicollinearity

were verified prior to analysis and data were transformed if residuals were not normal according

to Shapiro-Wilk's test. If after transformation, residuals were not normal, then a generalized linear

mixed effects model was used.

Results

Participant Demographics

Seven participants per group (four males, three females) completed the training, with

demographics displayed in Table 1. There was no significant difference between groups in age

(HIGT: 60.6±10.7, MFES+HIGT: 57.7±10.2, p=0.331), years since stroke (p=0.703), baseline

10MWT SSV (p=0.625), baseline 6MWT (p=0.462), nor baseline BBS (p=0.909). Only one

individual (who had been assigned to the HIGT group) dropped out of the study due to pre-existing

medical issues before mid-assessment and therefore was not included in the analysis.

358   Table 1: Participant Demographics

ID	Group	Gender	Hemi Side	Years Since Stroke	Stroke Type	Baseline 10MWT SSV (m/s)	Baseline 6MWT (m)	Baseline BBS (/56)	Baseline Ambulatory Group
1	HIGT	Female	Right	21	Hemorrhagic	0.49	208	42	Limited
2	HIGT	Female	Right	5	Ischemic	0.69	250	48	Limited
3	HIGT	Male	Left	6	Ischemic	0.91	371	50	Community
4	HIGT	Male	Right	4	Ischemic	0.58	243	44	Limited
5	HIGT	Male	Right	15	Unknown	0.73	201	42	Limited
6	HIGT	Female	Left	2	Ischemic	1.01	366	54	Community
7	HIGT	Male	Left	6	Hemorrhagic	0.23	79	41	Household
7	Mean ± Std	4M/3F	4R/3L	8.4 ± 6.9		0.66 ± 0.26	245.4 ± 101.2	45.9 ± 4.9	1 House, 4 Limited, 2 Community
8	MFES+HIGT	Male	Left	3	Hemorrhagic	1.35	502	48	Community
9	MFES+HIGT	Female	Left	8	Hemorrhagic	0.4	133	49	Limited
10	MFES+HIGT	Male	Right	12	Ischemic	0.81	394	54	Community
11	MFES+HIGT	Male	Left	10	Hemorrhagic	0.15	47	33	Household
12	MFES+HIGT	Male	Left	2	Hemorrhagic	0.41	164	50	Limited
13	MFES+HIGT	Female	Left	15	Ischemic	0.11	43	34	Household
14	MFES+HIGT	Female	Left	3	Ischemic	0.75	58	50	Limited
7	Mean ± Std	4M/3F	1R/6L	7.6 ± 5.1		0.57 ± 0.44	191.6 ± 183.6	45.4 ± 8.4	2 House, 3 Limited, 2 Community

HIGT = High-Intensity Gait Training, MFES = Multichannel Functional Electrical Stimulation, Hemi = Hemiplegic, 10MWT

SSV = 10 Meter Walk Test Self-Selected Velocity, 6MWT = 6 Minute Walk Test, BBS = Berg Balance Scale

Training Measures

Training and cardiovascular intensity results demonstrate that, on average, both groups were able

to maintain HR intensities of 70-85% of maximum HR for 50% of each session, which is consistent

with other HIGT studies.(55) In addition, average RPE was 14.1±3.2 for participants in HIGT and

14.9±2.9 for participants in MFES+HIGT, indicating that all participants perceived the intensity

of the intervention close to 'Hard'. Only two participants (both in the MFES+HIGT group) were

unable to achieve their target HR zone for more than 30% of the training time; however, their

average RPE ratings were 16.6±1.5 and 13.2±1.0, respectively. There was no main effect of group,

session, nor group*session interaction for mean HR (as a percentage of maximum predicted HR),

peak HR (as a percentage of maximum predicted HR), percentage of time in HR zone, nor mean

RPE. P-values are reported along with mean and standard deviations for all metrics (see Table 2).

There were no adverse reactions in either group to training.

Table 2: Training and Cardiovascular Intensity Details

	HIGT	MFES+HIGT	P-value
Mean HR (% Predicted Max)	73.9 ± 6.0 %	70.0 ± 8.4 %	0.31
Peak HR (% Predicted Max)	82.1 ± 7.6 %	77.5 ± 8.9 %	0.28
Percent of Time in HR Zone	72.6 ± 30.2 %	49.2 ± 37.6 %	0.17
Mean RPE	14.1 ± 3.2	14.9 ± 2.9	0.65

HIGT = High-Intensity Gait Training, MFES = Multichannel Functional Electrical Stimulation, HR = Heart Rate,
RPE = Borg Rating of Perceived Exertion (Range = 6-20).

Baseline Regression of Synergy Metrics with Clinical Assessments

Regression of synergy metrics and clinical assessments was run with baseline data for all

participants, irrespective of training group, with the goal of validating the rationale behind driving

changes in muscle synergies to improve clinical outcomes. As predicted, average cosine similarity

and circular cross correlation coefficient at baseline were positively correlated with 10MWT SSV,

10MWT FV, and 6MWT. BBS was correlated with average cosine similarity and circular cross

correlation lag. Correlation coefficients and p values are displayed in Table 3, and regression

analysis is depicted in Figure 2 for 10MWT FV with cosine similarity and 6MWT with circular

cross correlation coefficient. On average, 2.1±1.6 and 2.3±0.8 synergies were impaired at baseline

for the HIGT group and MFES+HIGT group, respectively (see Supplementary Table 1). In both

groups, the synergy related to propulsion (primary activators: gastrocnemius) was the most

frequently impaired (HIGT: 5/7, MFES+HIGT: 6/7) while the synergy related to weight

acceptance (primary activators: gluteus medius, rectus femoris, vastus lateralis) was the least

frequently impaired (2/7 in both groups).

Figure 2: Regression of Baseline Clinical Assessments with Synergy Metrics. Linear regression

analysis compared clinical assessments and muscle synergy metrics at baseline to confirm their

relationship prior to training. Exemplary correlations are demonstrated with 10 Meter Walk Test

Fast Velocity (10MWT FV) vs average cosine similarity and 6 Minute Walk Test (6MWT) vs

circular cross correlation coefficient. All data are displayed in Table 3.

Table 3: Baseline Regression of Synergy Metrics with Clinical Assessments

	Cosine Similarity	Cross Correlation Coefficient	Cross Correlation Lag
10MWT SSV	R = .633, P = 0.015*	R = 0.585, P = 0.028*	R = 0.487, P = 0.077
10MWT FV	R = .628, P = 0.016*	R = 0.557, P = 0.038*	R = 0.483, P = 0.080
6MWT	R = 0.572, P = 0.032*	R = 0.582, P = 0.029*	R = 0.351, P = 0.219
BBS	R = 0.678, P = 0.008*	R = 0.513, P = 0.061	R = 0.686, P = 0.007*

Clinical Assessments

Clinical assessments are displayed in Figure 3. No differences were identified between groups at baseline (all p>0.05). For the 10MWT SSV, within the HIGT group, there was a main effect of time (p=0.043), with increases observed at post-intervention (0.75±0.31 m/s, p=0.005) and follow-up (0.74±0.32 m/s, p=0.011) compared to baseline (pre: 0.66±0.26 m/s). In the MFES+HIGT group, there was a main effect of time (p=0.004), with SSV higher at mid (0.64±0.44 m/s, p=0.022), post (0.67±0.45 m/s, p=0.003), and follow-up (0.69±0.50 m/s, p=0.001) compared to baseline (0.57±0.43 m/s). No differences in changes from baseline were identified between groups. For the 10MWT FV, within the HIGT group, no main effect of time was detected (avg=1.05±0.52 m/s, p=0.318). In the MFES+HIGT group, there was a main effect of time (p=0.022), with greater FV at post (0.84±0.56 m/s, p=0.015) and follow-up (0.87±0.61 m/s, p=0.004) compared to baseline (0.71±0.53 m/s). The improvement in fast velocity at follow-up was greater for MFES+HIGT than HIGT (p=0.040). For the 6MWT distance, within the HIGT group, there was a main effect of time (p=0.043), with greater distance at mid (277±122 m) compared to baseline (pre: 245±102 m, p=0.006). In the MFES+HIGT group, there was a main effect of time (p=0.014), with greater distance at mid (237±173 m, p=0.049), post (264±187 m, p=0.004) and follow-up (259±188 m, p=0.006) compared to baseline. No differences in changes from baseline were identified between groups. For the BBS, within the HIGT group, there was a main effect of time (p=0.039), with a higher score at mid (48.9±3.9) compared to baseline (pre: 45.9±4.91, p=0.005), while the MFES+HIGT group exhibited no effect of time (avg=45.1±7.8, p=0.521). This improvement from baseline for the HIGT group at midpoint was greater than the improvement for the MFES+HIGT group (p=0.017), however improvements were not maintained at post and follow-up. Average

number of steps per day had no main effect of time in either group (HIGT: 1908±1295, p=0.203;

MFES+HIGT: 3035±2285, p=0.375) and the improvement did not differ (p=0.150).

Figure 3: Change in Clinical Assessments from Baseline. The 10 Meter Walk Test (10MWT) at

self-selected velocity (SSV) and fast velocity (FV) and the 6 Minute Walk Test (6MWT) are

presented. Box plots display group data and individual points display participant data on the

difference from the baseline visit. Group is depicted by color and baseline ambulatory category

is depicted in the shape (see legend). * represents a within group significant difference across

time with the color of the * matching the group. ǂ represents a difference between groups at a

424   specific time point. A significance threshold of p < 0.05 was used.

Gait Spatiotemporal Metrics and Biomechanics

Spatiotemporal metrics and joint kinematics during SSV are reported in Figure 4 and

Supplementary Table 2. No differences were identified between groups at baseline for

spatiotemporal metrics and joint angles (all p>0.05). Both groups demonstrated a significant effect

of time on step length of the impaired limb (HIGT: p=0.027, MFES+HIGT: p=0.033), with both

exhibiting improvements at post (HIGT: p=0.033, MFES+HIGT: p=0.013) and follow-up (HIGT:

p=0.007, MFES+HIGT: p=0.010) compared to baseline. No differences in changes from baseline

were identified between groups (all p>0.05). For step width, within the HIGT group, no main effect

of time was found (p=0.213), but a main effect of time was identified in the MFES+HIGT group

(p=0.028), with a significant reduction from baseline to mid (p=0.048), post (p=0.042), and follow-

up (p=0.004). The change at follow-up was significantly different between groups (p=0.019). For

stance percent of the impaired limb, within the HIGT group, no main effect of time was found

(p=0.602), but a main effect of time was identified in the MFES+HIGT group (p=0.018), with

stance percent getting worse (less) at post compared to baseline (p=0.004). No between group

differences in changes from baseline were identified (all p>0.05).

In the HIGT group, the maximum dorsiflexion angle during swing had a main effect of time

(p=0.050), with participants exhibiting worse dorsiflexion (less) at post compared to baseline

(p=0.008), while no main effect of time was found in the MFES+HIGT group (p=0.682). For hip

extension during stance (a proxy measure of propulsion), while neither group exhibited a change

over time (HIGT: p=0.590, MFES+HIGT: p=0.083), participants in the MFES+HIGT group

exhibited a moderate increase in extension at post (14.3±.10.3°) compared to baseline (9.7±.13.2°).

No other kinematic metrics (hip abduction during swing, knee extension during stance, knee

flexion during swing) changed with time in either group (all p>0.05). No differences in changes

from baseline were identified between groups for all kinematic metrics (all p>0.05).

Figure 4: Change in Biomechanics. Top panel: change in impaired limb step length, step width, and impaired limb stance percent. Bottom panel: change in peak ankle dorsiflexion angle during swing, peak knee extension angle during stance, and peak hip extension angle during stance. Positive orientation for each joint are as follows: ankle = dorsiflexion, knee = extension, hip = flexion. Box plots display group data and individual points display participant data on the difference from the baseline visit. Group is depicted by color and baseline ambulatory category is depicted in the shape (see legend). * represents a within group significant difference across

time with the color of the * matching the group. ╪ represents a difference between groups at a

specific time point. A significance threshold of p < 0.05 was used. DF = Dorsiflexion, PF =

Plantarflexion, Flex = Flexion, Ext = Extension.

Muscle Synergies

On average, four muscle synergies were extracted from neurologically intact individuals, and

therefore, four synergies were used for the average normative data. Muscle synergies of

participants were compared with average normative data and results are displayed in Figure 5. No

differences were identified between groups at baseline (all p>0.05). The average cosine similarity

across all impaired-side synergies showed no significant effect of time in either group (HIGT:

avg=0.93±0.07, p=0.527; MFES+HIGT: avg=0.91±0.06, p=0.670). Similarly, the average circular

cross-correlation coefficient did not change over time in the HIGT group (avg=0.83±0.10,

p=0.695). However, in the MFES+HIGT group, a significant main effect of time on circular cross

correlation coefficient was observed (p=0.038), with increases at post (0.82±0.04, p=0.036) and

follow-up (0.81±0.04, p=0.027) compared to baseline (0.79±0.05). The absolute time lag of the

circular cross-correlation showed no effect of time in either group (HIGT: avg=0.88±0.11,

p=0.863; MFES+HIGT: avg=0.86±0.06, p=0.993). Number of impaired synergies showed no

effect of time in either group (HIGT: avg=2.39±1.26, p=0.129; MFES+HIGT: avg=3.39±0.74,

p=0.564). No differences between the changes from baseline were identified between groups for

all synergy metrics (all p>0.05).

Figure 5: Change in Average Muscle Synergy Metrics of Impaired Limb. Cosine similarity and

circular cross correlation coefficient were computed between participant synergies and average

normative data. Line plots display individual data for each participant on the left, with averages

depicted by thicker lines. Box plots display group data and individual points display participant

data on the difference from the baseline visit on the right. Group is depicted by color and

baseline ambulatory category is depicted in the shape (see legend). * represents a within group

significant difference across time where $p < 0.05$ with the color of the * matching the group.

Feasibility

Participants completed the Patient Perceived Usefulness Questionnaire, rating between -5 ("Much

Worse") and 5 ("Completely Recovered"). On average, the MFES+HIGT group rating was

+3.43±1.51, while the HIGT group rating was +3.14±0.38, which were not significantly different

(p = 0.636). Below are responses from three participants in the MFES+HIGT group, emphasizing

their greater independence, perceived improvement in endurance, and improvement in quality of

life:

"I am walking much better than before…I was relying on a brace…walking and now I don't use

any device at all. My balance is much improved and I am more independent in living my life. My

493   *wife says that I am walking much better."*

494   *"I am walking much (more) without my brace and moving much faster."*

495   *"Confidently walking at home without AFO"*

Participants in the HIGT group also reported improvements in endurance and balance, but did not

report the reduction in dependence on assistive devices as much as the MFES+HIGT group.

The time to don the device and set FES parameters was recorded and assessed for clinical

feasibility, as displayed in Figure 6. The inflection point, where the plateau began, was identified

at 8.48 sessions and the slope of the learning phase was -0.758 minutes per session. The curve

plateaued at 4.53 minutes. Therapists completed the Therapist Perceived Usefulness Questionnaire

on whether they would implement synergy-based MFES+HIGT into their own clinical practice,

rating between -5 ("Very Unlikely") and 5 ("Very Likely"). Therapists reported a mean score of

+3.6±0.48. Therapist feedback included ease of use to don and doff the device and update

stimulation parameters throughout training. For the Implementation Outcome Measures,(54)

which ranges from 1 to 5 and measures feasibility, appropriateness, and acceptability of

intervention, therapists reported an average score of 3.9±0.13, 4.1±0.13, and 4±0.00, respectively.

Lastly, on the 10-item System Usability Scale, measured on a 5-point Likert scale with a total

score out of 100, average therapist score for synergy-based MFES was 72.15±7.5 indicating an

"OK" rating bordering "Good".

Figure 6: Preparation Time vs Number of Sessions. The time to don the sleeve and set

parameters was recorded for each session for each of the four therapists, who are each depicted

with a different color as specified in the legend. An inflection point analysis was conducted to

identify the learning region and the plateau region. The slope of the learning region was -0.7578

minutes per session as depicted by the solid line. The preparation time plateaued at 8.48 sessions

517   at a value of 4.53 minutes.

Discussion

In this study, we investigated the use of muscle synergy-based MFES paired with HIGT in a six-

week chronic stroke gait rehabilitation compared to standard of care, HIGT, and the feasibility of

implementing this intervention into the clinical setting. At baseline, clinical assessments were

positively correlated with synergy metrics, suggesting a strong relationship between muscle

synergy organization and timing with gait function. The addition of synergy-based MFES did not

hinder the ability of the therapists to challenge participants to reach the training metrics set for

standard HIGT, and training metrics did not differ between groups. Only the MFES+HIGT group

demonstrated sustained improvements in fast walking speed and endurance, with fast walking

speed greater at follow-up in MFES+HIGT compared to HIGT. Furthermore, while both groups

improved impaired limb step length, MFES+HIGT also led to reduced step width, suggesting more

stable gait, while HIGT led to worse dorsiflexion during swing. Muscle synergy average circular

cross correlation coefficient to normative data in only the MFES+HIGT group improved while

other synergy metrics did not change in either group. Participant rating of the perceived effects of

MFES+HIGT were positive, with a perception of improved confidence and reduced reliance on

assistive devices after training. Therapists administering MFES+HIGT (n=4) rated the intervention

highly with respect to its feasibility, appropriateness, acceptability, and usability. Therapist

comfort with the device improved rapidly, requiring only nine sessions to plateau in preparation

time to an average of 4.53 minutes, which is feasible for clinical implementation. We recognize

that statistical analysis is limited due to the small and highly variable population recruited in this

study. Despite this limitation, we believe that these positive preliminary results support the need

to extend this intervention into a larger randomized controlled clinical trial, stratified for

individuals with diverse baseline functions and time since stroke to effectively identify responders

and non-responders, and potentially investigate a greater number of sessions to produce greater

clinically meaningful differences.

Training

On average, each group was able to achieve the target HR zone (70-85%) for 50% of each session. This was consistent with prior studies,(55) where participants maintained the 70-85% HR zone for at least 30% of a training session. Two participants (both in MFES+HIGT) were unable to achieve the target HR zone for at least 30% despite adjustments to speed and incline of the treadmill, leading to a non-significant but large difference between groups in time spent in target HR zone (HIGT: 72.6±30.2%; MFES+HIGT: 49.2±37.6%). These participants were both household ambulators with much lower SSV and greater levels of gait impairment who were unable to achieve faster speeds and inclines on the treadmill, limiting their intensity. However, RPE rating indicated that they were significantly challenged, with all ratings above 13. The two household ambulators in the MFES+HIGT group compared to the one in the HIGT group likely brought down the average time spent in target HR zone.

Clinical Outcomes

The present study found that while both groups improved SSV, only the MFES+HIGT group demonstrated a significant and sustained improvement in FV and endurance, with the improvement in FV at follow-up significantly greater for the MFES+HIGT group than HIGT group. The HIGT group did significantly improve in BBS and to a greater extent than the MFES+HIGT group at the midpoint, but this was not maintained after training or at follow-up, and no changes in average daily number of steps were observed. FV relies heavily on the strength of the hip flexors and extensors.(56, 57) Therefore, the greater improvement in fast walking speed observed in the MFES+HIGT group may be attributed to enhanced power generation at the paretic hip, likely facilitated by stimulation of gluteus medius, quadriceps, and hamstrings. While HIGT has been shown to elicit improvements in walking speed, endurance, balance, and steps per day through

task repetition for chronic stroke rehabilitation,(58) it is not as effective as in the subacute phase

post-stroke,(55) and long-term recovery remains uncertain for chronic stroke.(59) Furthermore,

single-channel FES to the TA does not consistently improve gait speed and endurance compared

to HIGT,(60) while greater improvements can be elicited by stimulating an additional muscle (e.g.,

gluteus medius(21, 22), gastrocnemius(19, 20)) or by stimulating all major muscle groups.(25, 27)

Importantly, our results highlight the maintenance of improvements in FV and endurance at one

month following the completion of only MFES+HIGT. Therefore, addressing impairments to

coordination of the entire leg may be the key to enhancing gait rehabilitation, but further

investigation of the long-term effects is necessary.

We hypothesize that MFES+HIGT may elicit more efficient neuromuscular control by optimizing

muscle coordination, minimizing compensatory strategies, and reducing energy expenditure, thus

facilitating the improved endurance found in this study. This interpretation is supported by

previous work combining FES with task-specific training, which has been shown to preserve

muscle mass in patients with moderate to severe acute stroke.(61) Furthermore, administering FES

to the TA and gastrocnemius during FV has been shown to reduce energy expenditure post-

stroke,(19, 20) and a study comparing synergy-based multichannel FES to other approaches in the

upper limb found that synergy-based stimulation elicits reduced fatigue during training.(26)

Endurance remains a critical challenge in stroke rehabilitation,(62) and greater walking endurance

has been linked to better community reintegration outcomes.(63) Together, these findings suggest

that while both groups gained from repetitive, high-intensity walking, the added sensorimotor

stimulation in the MFES+HIGT group may have yielded more efficient, fatigue-resistant gait,

thereby improving FV and endurance.

Biomechanics

Biomechanical gait analysis identified an increase in impaired limb step length for the both groups while only the MFES+HIGT group exhibited a reduction in step width, suggesting improved dynamic balance.(64, 65) However, MFES+HIGT also led to reduced impaired limb stance percent, which suggests that participants are spending less time in stance on their impaired limb. This could be a result of decreased hip extension, compromising the propulsive phase of stance and leading to reduced propulsive force, a common impairment post-stroke.(17, 66) The reduced torques produced to propel an individual post-stroke during gait are suggested to be a result of abnormal muscle coordination.(67) Reduced stance percent may also be due to greater knee hyperextension, which can reduce walking speed, decrease gait efficiency, and increase energy expenditure.(68) However, participants in the MFES+HIGT group exhibited non-significant, yet moderate improvement in hip extension during stance and no change in knee extension during stance (Figure 4) in addition to the improvements in walking speed and endurance, suggesting that joint kinematic compensations were not driving the reduction in stance phase.

HIGT also led to a negative effect, with participants exhibiting reduced ankle dorsiflexion during swing after training compared to baseline while no change was found after MFES+HIGT. Lack of dorsiflexion during swing phase leads to foot drop, which can often result in compensatory movements, such as hip circumduction and hip hiking.(16, 17) Therefore, we hypothesize that HIGT may have reinforced compensatory mechanisms at the cost of improving gait speed, leading to worse foot drop. The goal of HIGT is to challenge patients to reach a target HR zone by increasing speed and incline, forcing the individual to devise a strategy to achieve the task. In an unguided kinematic space, this could include compensatory mechanisms that may not be optimal due to stress on other joints or less efficient movement patterns. Pairing HIGT with FES to the TA

has been shown to improve dorsiflexion during swing, the target of the stimulation.(16) However,

its limited ability to improve broader coordination across joints combined with improved

spatiotemporal metrics after stimulation to both gluteus medius and TA(22) provides the rationale

for extending stimulation to multiple muscles and personalizing stimulation to the individual's

impairment. In this study, while the MFES+HIGT group did not significantly improve any joint

kinematic metrics, they at least maintained baseline function and, in the case of hip extension

during stance, exhibited moderate improvements. Therefore, individuals undergoing targeted

training to address abnormal muscle synergies may better facilitate changes that not only improve

propulsion, which would enable faster walking speeds with decreased energy expenditure, but also

avoid compensatory joint kinematics. The stimulation strategy may have guided movement within

a kinematic space that promotes muscle coordination patterns more similar to normative

recruitment (as discussed in the next section), thereby limiting the emergence of abnormal patterns.

Ultimately, joint kinematic improvements were limited in both interventions. Additional

investigation of comparison to normative data, inter-joint coordination, and inter-limb

coordination may provide a deeper understanding of the altered kinematic strategies used to

achieve the improvements in fast walking speed and endurance.

Muscle Synergies

At baseline, we observed a significant relationship between clinical assessments and muscle

synergy metrics, particularly, for cosine similarity and circular cross correlation coefficient,

providing support for an intervention that drives changes in muscle synergies to elicit changes in

gait function. While, there were no changes in the spatial organization of muscle synergies (i.e.,

cosine similarity) or the number of impaired synergies for either group, MFES+HIGT improved

average similarity of the activation profile to normative data (i.e., circular cross correlation

coefficient) and maintained this improvement at follow-up. The improvement in physiological

timing of muscle activation may have been induced by the enhanced afferent feedback provided

by FES during MFES+HIGT.(9-13) These findings align with previous literature demonstrating

that muscle synergy analysis can quantify neuromuscular impairment of individuals with chronic

stroke, and that these impairments relate to functional capacity. In the lower limb, post-stroke gait

is characterized by a reduction in number of synergies in the impaired limb, which correlates with

impairments in gait speed, step length, and propulsion.(5) Prior work from our group has shown

that combining muscle synergy and biomechanical features improves the prediction of walking

function post-stroke.(69)

Moreover, stroke gait rehabilitation has been shown to increase the number of synergies and

enhance the timing of impaired synergies, resulting in improved biomechanics and gait

outcomes.(6, 7) Two pilot studies (one by our group) on two individuals each receiving four weeks

of synergy-based MFES gait training resulted in moderate improvements in number of synergies

and similarity of synergies to normative data, which were accompanied by moderate improvements

in walking outcomes and gait metrics.(25, 27) We have also demonstrated that cycling training

combined with FES for individuals post-stroke can improve clinical outcomes, gait spatiotemporal

metrics, and similarity of gait synergies to normative patterns.(70) Although no changes emerged

with respect to the spatial organization of synergies (i.e., cosine similarity), the improvements in

similarity of the synergy activation timing to normative data highlight the ability of this targeted

intervention to modify neural control. Furthermore, the lack of changes in both joint kinematics

and cosine similarity may reflect the relatively low intervention dosage (18 sessions of 20 minutes)

and high variability in baseline impairment and synergy composition. A larger sample and a higher

training dose (e.g., more or longer sessions) may be required to elicit group-level effects on

synergy structure and kinematics.

Feasibility

Training with MFES+HIGT resulted in no dropouts, while only one dropout occurred in the HIGT

group (due to preexisting medical conditions), with no adverse reactions in either group.

Participants rated both interventions equally well with a score of 3.14 (HIGT) and 3.43

(MFES+HIGT) on a scale of -5 to 5. Participants in the MFES+HIGT group reported that after

training, they had improved endurance, increased confidence, and reduced reliance on assistive

devices, with some stating that they no longer use their ankle foot orthosis. Qualitatively, this is a

very promising finding, which should be systematically investigated in future studies.

This study also aimed to assess the usability and clinical feasibility of the novel MFES+HIGT

protocol amongst practicing physical therapists. With two hours of training and after

approximately nine sessions to administer the protocol, therapists were able to don the Cionic

Neural Sleeve and set FES parameters in 4.53 minutes on average. These therapists were

experienced with post-stroke gait training for at least three years and had some level of experience

with single-channel FES for foot drop. Although there is not current data to compare these results

to other FES devices, this appears to be a very feasible intervention for clinical implementation.

In practice, an individual receiving outpatient physical therapy will typically be scheduled for 45

minutes, two to three times per week. Therapy sessions include time for subjective report, objective

measures, monitoring vitals, and set up for intervention. With limited time for direct administration

of HIGT interventions, it is imperative that technology-assisted interventions do not take away

from the time administering an intervention in order for therapists to be able to achieve at least 20

minutes of HIGT per session.(33) Upon further evaluation of the System Usability Scores,

therapists addressed a concern in the following question: "I think that I would need the support of

a technical person to be able to use this system," with two agreeing with this statement and two

neutral on the topic. Modifications to the sleeve to include the external channels that allow

stimulation and recording of gluteus medius led to greater setup times, lack of confidence when

manually placing electrodes, and a few disconnects between the system and the external unit,

which is when a researcher would provide technical support.

Feedback from therapists administering the MFES+HIGT protocol overall indicated that the

intervention was usable. The System Usability Scale score on average was 71.5, passing the

threshold for acceptable usability (68/100).(71, 72) In addition, the average response from the

therapists on the Perceived Usefulness questionnaire was 3.6 indicating that they were likely to

implement the intervention into their clinical practice. Notably, all therapists were practicing in an

inpatient acute rehabilitation setting and primarily treated individuals in the subacute phase of

stroke recovery. Their responses may have been cautious, as they observed positive outcomes

using the intervention with chronic stroke participants but were uncertain about its feasibility in

their current clinical environment. Further research is needed to determine whether synergy-based

MFES could enhance HIGT during the acute inpatient phase and whether it can be feasibly

implemented in that setting.

Variability of Baseline Ambulatory Function

Gait speed is a valuable predictor of ambulation ability and has been shown to be a reliable method

for classification of patients post-stroke.(32) In the HIGT group, there was one household

ambulator (<0.4 m/s), four limited community ambulators (0.4-0.8 m/s), and two community

ambulators (>0.8 m/s). In the MFES+HIGT group, there were two household ambulators, three

limited community ambulators, and two community ambulators. Baseline ambulatory category

was included as a random effect in statistical testing, as the groups had different limitations in

training and outcomes, but the following comparisons between groups are purely qualitative,

especially for the HIGT household ambulation category which only had one participant. The two

household-level ambulators in the MFES+HIGT group did not on average reach their target HR

zone for at least 30% of their training session, as therapists were limited in how much they could

challenge these participants with speed and incline. As noted, these participants still provided the

rating of perceived exertion as challenging. Clinical assessments revealed that household

ambulators, which had lower scores on speed and endurance at baseline, also tended to have

smaller improvements in these metrics, which were not retained at follow-up. The changes in

biomechanics did not appear to follow any trends in ambulation level. For the synergy metrics, the

HIGT household ambulator exhibited marked reductions in scores whereas the MFES+HIGT

household ambulators exhibited improvements similar to the group average. With only one

participant in HIGT group, no conclusions can be made, but this poses the question of which

ambulation category may respond best to which treatment. Furthermore, household ambulators

tended to have lower cosine similarity and correlation coefficients than community ambulators,

while limited community ambulators exhibited a wide range of cosine similarity. We highlight

these differences among groups to emphasize the need to better characterize how differences in

baseline ambulatory function impact recovery potential and response to MFES+HIGT.

Limitations and Future Directions

A major limitation of this study is the small sample size (N=7 per group) with individuals post-

stroke of highly variable baseline function, ranging from baseline walking speed of 0.11 m/s to

1.35 m/s, despite all exhibiting some level of gait impairment. Furthermore, the years since stroke

onset ranged from two to 21, which can impact the ability of participants to improve. Due to this

small sample size, a comprehensive model could not be used to quantify impairments between

groups and across time. A larger sample size is required to run this statistical analysis with

sufficient power and to identify which ambulation category may benefit most from the

intervention. This study also did not include blinded assessors, and only individuals in the

MFES+HIGT group wore the sleeve during training, potentially leading to confounding effects of

improvements due to wearing the sleeve. However, these preliminary results still support the

pursuit of a larger study comparing this novel personalized intervention–synergy-based

MFES+HIGT—in a single-blinded randomized controlled trial, accounting for individuals with

varying baseline function and years since stroke. This follow-up study may also explore changing

the participant stimulation profiles as the participant recovers and may investigate a longer

rehabilitation program and greater dosage, as no plateau effect was observed. It may also be more

impactful to target individuals sooner after the stroke, in the acute/subacute phase (< 6 months),

which is often described as the neuroplastic window. Furthermore, this study is limited to analysis

of the final output of the nervous system at the periphery, but we suspect that MFES+HIGT is

driving changes to neural coordination and activation at the central level through afferent feedback

of MFES.(9-13) Therefore, investigation of the central mechanisms governing these changes in

muscle coordination at the spinal, subcortical, and cortical levels will be essential to a deeper

understanding of the mechanisms of recovery and could potentially provide rationale behind

responders and non-responders to treatment. Additionally, combining strategies like synergy-

based MFES with soft robotics in a hybrid solution may further enhance muscle coordination and

joint kinematics with training.(73-75) Finally, FES is not the only approach to modify muscle

synergies. Recent advancements in real time extraction of synergies have demonstrated the ability

to modulate synergies with operant conditioning and visual feedback through control of a cursor

and game play, providing another avenue for personalized, precision interventions for driving

neuromuscular changes to impaired muscle synergies.(76-81) For clinical translation, ongoing

assessment of product usability should be considered to keep setup time acceptable within the

clinical workflow.

Conclusion

In a small, variable sample of individuals living with lasting gait impairments due to chronic

stroke, a personalized stimulation approach–muscle synergy-based multichannel functional

electrical stimulation–paired with high-intensity gait training resulted in improved fast walking

speed, endurance, gait biomechanics, and similarity of the timing of muscle synergies to normative

data. Furthermore, this intervention was highly rated by both participants and clinicians, and the

learning curve for the setup time plateaued after nine visits to 4.53 minutes, a feasible amount of

time for clinical implementation. The results of this study support the pursuit of a larger clinical

trial of synergy-based multichannel functional electrical stimulation with high-intensity gait

training to identify responders and non-responders based on time since stroke and baseline

ambulatory function, which we observed to have differences in our small sample size.

References

1.   Strilciuc S, Grad DA, Radu C, Chira D, Stan A, Ungureanu M, et al. The economic

burden of stroke: a systematic review of cost of illness studies. J Med Life. 2021;14(5):606-19.

2.   Neptune RR, Clark DJ, Kautz SA. Modular control of human walking: A simulation

study. Journal of Biomechanics. 2009;42(9):1282-7.

3.   Ting H, Lena, Chiel J, Hillel, Trumbower D, Randy, Allen L, Jessica, Mckay L, J.,

Hackney E, Madeleine, et al. Neuromechanical Principles Underlying Movement Modularity and

Their Implications for Rehabilitation. Neuron. 2015;86(1):38-54.

4.   Gonzalez-Vargas J, Sartori M, Dosen S, Torricelli D, Pons JL, Farina D. A predictive

model of muscle excitations based on muscle modularity for a large repertoire of human

locomotion conditions. Frontiers in Computational Neuroscience. 2015;9.

5.   Clark DJ, Ting LH, Zajac FE, Neptune RR, Kautz SA. Merging of healthy motor

modules predicts reduced locomotor performance and muscle coordination complexity post-

stroke. J Neurophysiol. 2010;103(2):844-57.

6.   Routson RL, Clark DJ, Bowden MG, Kautz SA, Neptune RR. The influence of locomotor

rehabilitation on module quality and post-stroke hemiparetic walking performance. Gait &

Posture. 2013;38(3):511-7.

7.   Tan CK, Kadone H, Watanabe H, Marushima A, Yamazaki M, Sankai Y, et al.

Corrigendum: Lateral Symmetry of Synergies in Lower Limb Muscles of Acute Post-stroke

Patients After Robotic Intervention. Front Neurosci. 2020;14:113.

8.   Houghton PE, Campbell KE, Fraser CH, Harris C, Keast DH, Potter PJ, et al. Electrical

stimulation therapy increases rate of healing of pressure ulcers in community-dwelling people

with spinal cord injury. Arch Phys Med Rehabil. 2010;91(5):669-78.

9.   Gandolla M, Ward NS, Molteni F, Guanziroli E, Ferrigno G, Pedrocchi A. The Neural

Correlates of Long-Term Carryover following Functional Electrical Stimulation for Stroke.

Neural Plast. 2016;2016:4192718.

10.   Gandolla M, Niero L, Molteni F, Guanziroli E, Ward NS, Pedrocchi A. Brain Plasticity

Mechanisms Underlying Motor Control Reorganization: Pilot Longitudinal Study on Post-Stroke

Subjects. Brain Sci. 2021;11(3).

11.   Barsi GI, Popovic DB, Tarkka IM, Sinkjaer T, Grey MJ. Cortical excitability changes

following grasping exercise augmented with electrical stimulation. Exp Brain Res.

2008;191(1):57-66.

12.   Iftime-Nielsen SD, Christensen MS, Vingborg RJ, Sinkjær T, Roepstorff A, Grey MJ.

Interaction of electrical stimulation and voluntary hand movement in SII and the cerebellum

during simulated therapeutic functional electrical stimulation in healthy adults. Human Brain

Mapping. 2012;33(1):40-9.

13.   Milosevic M, Masugi Y, Obata H, Sasaki A, Popovic MR, Nakazawa K. Short-term

inhibition of spinal reflexes in multiple lower limb muscles after neuromuscular electrical

stimulation of ankle plantar flexors. Exp Brain Res. 2019;237(2):467-76.

14.   Marquez-Chin C, Popovic MR. Functional electrical stimulation therapy for restoration

of motor function after spinal cord injury and stroke: a review. BioMedical Engineering OnLine.

2020;19(1).

15.   Shin HE, Kim M, Lee D, Jang JY, Soh Y, Yun DH, et al. Therapeutic Effects of

Functional Electrical Stimulation on Physical Performance and Muscle Strength in Post-stroke

Older Adults: A Review. Ann Geriatr Med Res. 2022;26(1):16-24.

16.   Allen JL, Ting LH, Kesar TM. Gait Rehabilitation Using Functional Electrical

Stimulation Induces Changes in Ankle Muscle Coordination in Stroke Survivors: A Preliminary

Study. Front Neurol. 2018;9:1127.

17.   Chen CL, Chen HC, Tang SF, Wu CY, Cheng PT, Hong WH. Gait performance with

compensatory adaptations in stroke patients with different degrees of motor recovery. Am J Phys

Med Rehabil. 2003;82(12):925-35.

18.   Pereira S, Mehta S, McIntyre A, Lobo L, Teasell RW. Functional electrical stimulation

for improving gait in persons with chronic stroke. Top Stroke Rehabil. 2012;19(6):491-8.

19.   Awad LN, Reisman DS, Pohlig RT, Binder-Macleod SA. Reducing The Cost of

Transport and Increasing Walking Distance After Stroke: A Randomized Controlled Trial on

Fast Locomotor Training Combined With Functional Electrical Stimulation. Neurorehabil Neural

Repair. 2016;30(7):661-70.

20.   Hakansson NA, Kesar T, Reisman D, Binder-Macleod S, Higginson JS. Effects of fast

functional electrical stimulation gait training on mechanical recovery in poststroke gait. Artif

Organs. 2011;35(3):217-20.

21.   Cho M-K, Kim J-H, Chung Y, Hwang S. Treadmill gait training combined with

functional electrical stimulation on hip abductor and ankle dorsiflexor muscles for chronic

hemiparesis. Gait & posture. 2015;42(1):73-8.

22.   Chung Y, Kim J-H, Cha Y, Hwang S. Therapeutic effect of functional electrical

stimulation-triggered gait training corresponding gait cycle for stroke. Gait & posture.

2014;40(3):471-5.

23.   Van Bloemendaal M, Bus SA, Nollet F, Geurts ACH, Beelen A. Feasibility and

Preliminary Efficacy of Gait Training Assisted by Multichannel Functional Electrical

Stimulation in Early Stroke Rehabilitation: A Pilot Randomized Controlled Trial.

Neurorehabilitation and Neural Repair. 2021;35(2):131-44.

24.   Sijobert B, Azevedo C, Pontier J, Graf S, Fattal C. A Sensor-Based Multichannel FES

System to Control Knee Joint and Reduce Stance Phase Asymmetry in Post-Stroke Gait.

Sensors. 2021;21(6):2134.

25.   Ferrante S, Chia Bejarano N, Ambrosini E, Nardone A, Turcato AM, Monticone M, et al.

A Personalized Multi-Channel FES Controller Based on Muscle Synergies to Support Gait

Rehabilitation after Stroke. Frontiers in neuroscience. 2016;10.

26.   Bala S, Vishnu VY, Joshi D. Muscle Synergy-Based Functional Electrical Stimulation

Reduces Muscular Fatigue in Post-Stroke Patients: A Systematic Comparison. IEEE Trans

Neural Syst Rehabil Eng. 2023;31:2858-71.

27.   Lim J, Lim T, Lee J, Sim J, Chang H, Yoon B, et al. Patient-specific functional electrical

stimulation strategy based on muscle synergy and walking posture analysis for gait rehabilitation

of stroke patients. Journal of International Medical Research. 2021;49(5):030006052110167.

28.   Cionic. Cionic Neural Sleeve 2023 [Available from: https://cionic.com.

29.   Administration USFaD. 510(k) Premarket Notification: Cionic Neural Sleeve (K211651)

2022 [Available from:

https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm?ID=K211651.

30.   Robison J, Gibbons R, Achelis D, Bent B, Wajda D, Webster R. Augmenting gait in a

population exhibiting foot drop with adaptive functional electrical stimulation. 2022.

31.   Wang B, Wajda D, Sakai J, Yee J, Webster R. PCR145 Evaluating the Impact of the

Cionic Neural Sleeve™ on Quality of Life in Multiple Sclerosis Patients: A Pilot Study. Value in

Health. 2024;27(12):S534.

32.   Bowden MG, Balasubramanian CK, Behrman AL, Kautz SA. Validation of a speed-

based classification system using quantitative measures of walking performance poststroke.

Neurorehabil Neural Repair. 2008;22(6):672-5.

33.   Hornby TG, Reisman DS, Ward IG, Scheets PL, Miller A, Haddad D, et al. Clinical

Practice Guideline to Improve Locomotor Function Following Chronic Stroke, Incomplete

Spinal Cord Injury, and Brain Injury. Journal of Neurologic Physical Therapy. 2020;44(1):49-

100.

34.   Navalta JW, Davis DW, Malek EM, Carrier B, Bodell NG, Manning JW, et al. Heart rate

processing algorithms and exercise duration on reliability and validity decisions in biceps-worn

Polar Verity Sense and OH1 wearables. Sci Rep. 2023;13(1):11736.

35.   Oy PE. Polar OH1 Optical Heart Rate Sensor Kempele, Finland: Polar Electro Oy;

[Available from: https://www.polar.com/en/products/accessories/oh1-optical-heart-rate-sensor.

36.   Tanaka H, Monahan KD, Seals DR. Age-predicted maximal heart rate revisited. Journal

of the American College of Cardiology. 2001;37(1):153-6.

37.   Holleran CL, Straube DD, Kinnaird CR, Leddy AL, Hornby TG. Feasibility and Potential

Efficacy of High-Intensity Stepping Training in Variable Contexts in Subacute and Chronic

Stroke. Neurorehabilitation and Neural Repair. 2014;28(7):643-51.

38.   Hornby TG, Holleran CL, Hennessy PW, Leddy AL, Connolly M, Camardo J, et al.

Variable Intensive Early Walking Poststroke (VIEWS). Neurorehabilitation and Neural Repair.

2016;30(5):440-50.

39.   Straube DD, Holleran CL, Kinnaird CR, Leddy AL, Hennessy PW, Hornby TG. Effects

of dynamic stepping training on nonlocomotor tasks in individuals poststroke. Phys Ther.

2014;94(7):921-33.

40.   Borg G. Borg's Perceived Exertion and Pain Scales. Champaign, IL: Human Kinetics;

1998.

41.   Rabbi MF, Pizzolato C, Lloyd DG, Carty CP, Devaprakash D, Diamond LE. Non-

negative matrix factorisation is the most appropriate method for extraction of muscle synergies

in walking and running. Scientific Reports. 2020;10(1).

42.   Corp A. ActiGraph GT9X Link Activity Monitor Pensacola, FL, USA: ActiGraph Corp;

[Available from: https://www.actigraphcorp.com/.

43.   Ngueleu AM, Barthod C, Best KL, Routhier F, Otis M, Batcho CS. Criterion validity of

ActiGraph monitoring devices for step counting and distance measurement in adults and older

adults: a systematic review. J Neuroeng Rehabil. 2022;19(1):112.

44.   Suau Q, Bianchini E, Bellier A, Chardon M, Milane T, Hansen C, et al. Current

Knowledge about ActiGraph GT9X Link Activity Monitor Accuracy and Validity in Measuring

Steps and Energy Expenditure: A Systematic Review. Sensors (Basel). 2024;24(3).

45.   Cotton JR, DeLillo A, Cimorelli A, Shah K, Peiffer JD, Anarwala S, et al. Markerless

Motion Capture and Biomechanical Analysis Pipeline. arxiv. 2023.

46.   Cotton RJ. Differentiable Biomechanics Unlocks Opportunities for Markerless Motion

Capture arXiv preprint. 2024.

47.   EasyMocap: Make human motion capture easier: Github; 2021 [Available from:

https://github.com/zju3dv/EasyMocap.

48.   Sárándi I, Hermans A, Leibe B, editors. Learning 3D Human Pose Estimation from

Dozens of Datasets using a Geometry-Aware Autoencoder to Bridge Between Skeleton Formats.

IEEE/CVF Winter Conference on Applications of Computer Vision (WACV); 2023: IEEE.

49.   Cotton RJ. PosePipe: Open-Source Human Pose Estimation Pipeline for Rehabilitation

Research. Archives of physical medicine and rehabilitation. 2022;103(12):e161-e2.

50.   Cotton RJ, Mcclerklin E, Cimorelli A, Patel A, Karakostas T, editors. Transforming Gait:

Video-Based Spatiotemporal Gait Analysis. 2022 44th Annual International Conference of the

IEEE Engineering in Medicine & Biology Society (EMBC); 2022 2022-07-11: IEEE.

51.   Cimorelli A, Patel A, Karakostas T, Cotton RJ. Validation of portable in-clinic video-

based gait analysis for prosthesis users. Scientific Reports. 2024;14(1).

52.   Kamper SJ, Maher CG, Mackay G. Global Rating of Change Scales: A Review of

Strengths and Weaknesses and Considerations for Design. Journal of Manual & Manipulative

Therapy. 2009;17(3):163-70.

53.   Brooke J. SUS—A Quick and Dirty Usability Scale. Usability Evaluation in Industry.

1996;189:4-7.

54.   Weiner BJ, Lewis CC, Stanick C, Powell BJ, Dorsey CN, Clary AS, et al. Psychometric

assessment of three newly developed implementation outcome measures. Implementation

Science. 2017;12(1).

55.   Moore JL, Nordvik JE, Erichsen A, Rosseland I, Bø E, Hornby TG, et al. Implementation

of High-Intensity Stepping Training During Inpatient Stroke Rehabilitation Improves Functional

Outcomes. Stroke. 2020;51(2):563-70.

56.   Milot MH, Nadeau S, Gravel D. Muscular utilization of the plantarflexors, hip flexors

and extensors in persons with hemiparesis walking at self-selected and maximal speeds. J

Electromyogr Kinesiol. 2007;17(2):184-93.

57.   Jonkers I, Delp S, Patten C. Capacity to increase walking speed is limited by impaired hip

and ankle power generation in lower functioning persons post-stroke. Gait Posture.

2009;29(1):129-37.

58.   Hornby TG, Plawecki A, Lotter JK, Scofield ME, Lucas E, Henderson CE. Gains in

Daily Stepping Activity in People With Chronic Stroke After High-Intensity Gait Training in

Variable Contexts. Phys Ther. 2022;102(8).

59.   Kireyev IV ZN, Andriukhin DA, Sydorchuk MY, Tsemenko KV, Liashok AL. Efficiency

of High-Intensity Gait Training in the Rehabilitation of Post-Stroke Patients. Acta Balneologica.

2023;65(5):271-5.

60.   Jaqueline da Cunha M, Rech KD, Salazar AP, Pagnussat AS. Functional electrical

stimulation of the peroneal nerve improves post-stroke gait speed when combined with

physiotherapy. A systematic review and meta-analysis. Ann Phys Rehabil Med.

2021;64(1):101388.

61.   Nozoe M, Kanai M, Kubo H, Takeuchi Y, Kobayashi M, Yamamoto M, et al. Efficacy of

neuromuscular electrical stimulation for preventing quadriceps muscle wasting in patients with

moderate or severe acute stroke: A pilot study. NeuroRehabilitation. 2017;41(1):143-9.

62.   Mayo NE, Wood-Dauphinee S, Ahmed S, Gordon C, Higgins J, McEwen S, et al.

Disablement following stroke. Disabil Rehabil. 1999;21(5-6):258-68.

63.   Pang MY, Eng JJ, Miller WC. Determinants of satisfaction with community reintegration

in older adults with chronic stroke: role of balance self-efficacy. Phys Ther. 2007;87(3):282-91.

64.   Donlin MC, Ray NT, Higginson JS. User-driven treadmill walking promotes healthy step

width after stroke. Gait Posture. 2021;86:256-9.

65.　Chen G, Patten C, Kothari DH, Zajac FE. Gait differences between individuals with post-

stroke hemiparesis and non-disabled controls at matched speeds. Gait Posture. 2005;22(1):51-6.

66.　Arnold AS, Anderson FC, Pandy MG, Delp SL. Muscular contributions to hip and knee

extension during the single limb stance phase of normal gait: a framework for investigating the

causes of crouch gait. Journal of biomechanics. 2005;38(11):2181-9.

67.　Cruz TH, Dhaher YY. Evidence of abnormal lower-limb torque coupling after stroke: an

isometric study. Stroke. 2008;39(1):139-47.

68.　Geerars M, Minnaar-van der Feen N, Huisstede BMA. Treatment of knee hyperextension

in post-stroke gait. A systematic review. Gait & posture. 2022;91:137-48.

69.　Barroso FO, Torricelli D, Molina-Rueda F, Alguacil-Diego IM, Cano-de-la-Cuerda R,

Santos C, et al. Combining muscle synergies and biomechanical analysis to assess gait in stroke

patients. J Biomech. 2017;63:98-103.

70.　Ambrosini E, Parati M, Peri E, De Marchis C, Nava C, Pedrocchi A, et al. Changes in leg

cycling muscle synergies after training augmented by functional electrical stimulation in

subacute stroke survivors: a pilot study. Journal of NeuroEngineering and Rehabilitation.

2020;17(1).

71.　Bangor A, Kortum PT, Miller JT. An Empirical Evaluation of the System Usability

Scale. International Journal of Human-Computer Interaction. 2008;24(6):574-94.

72.　Bangor A, Kortum P, Miller J. Determining what individual SUS scores mean: Adding an

adjective rating scale. Journal of Usability Studies. 2009;4(3):114-23.

73.　Dell'Eva F, Dalla Gasperina S, Gandolla M, Pedrocchi A, Ambrosini E, editors. A hybrid

FES-motor cooperative control over a knee joint movement: A feasibility study International

Functional Electrical Stimulation Society 2022; 2022; Rotterdam, Netherlands.

74.   Tricomi E, Missiroli F, Xiloyannis M, Lotti N, Zhang X, Stefanakis M, et al. Soft robotic

shorts improve outdoor walking efficiency in older adults. Nature Machine Intelligence.

2024;6(10):1145-55.

75.   Ambrosini E, Dell'Eva F, Guanziroli E, Camerini V, Gandolla M, Brignole L, et al. A

Multifaceted Hybrid ES-Robotic Device for Gait Training in Individuals with Neurological

Disorders. 2024.

76.   Ghassemi M, Triandafilou K, Barry A, Stoykov ME, Roth E, Mussa-Ivaldi FA, et al.

Development of an EMG-Controlled Serious Game for Rehabilitation. IEEE Trans Neural Syst

Rehabil Eng. 2019;27(2):283-92.

77.   Berger DJ, d'Avella A. Effective force control by muscle synergies. Front Comput

Neurosci. 2014;8:46.

78.   Berger DJ, D'Avella A. Towards a Myoelectrically Controlled Virtual Reality Interface

for Synergy-Based Stroke Rehabilitation. Biosystems & Biorobotics: Springer International

Publishing; 2017. p. 965-9.

79.   Seo G, Kishta A, Mugler E, Slutzky MW, Roh J. Myoelectric interface training enables

targeted reduction in abnormal muscle co-activation. J Neuroeng Rehabil. 2022;19(1):67.

80.   Hong YNG, Ballekere AN, Fregly BJ, Roh J. Are muscle synergies useful for stroke

rehabilitation? Current opinion in biomedical engineering. 2021;19:100315.

81.   Torricelli D, De Marchis C, D'Avella A, Tobaruela DN, Barroso FO, Pons JL.

Reorganization of Muscle Coordination Underlying Motor Learning in Cycling Tasks. Frontiers

in Bioengineering and Biotechnology. 2020;8.