Facilitating Multi-Center Consistency in PET/CT Assessment: Transferability and Harmonization of 3D-Printed Phantom Inserts

Sai Kiran Kumar Nalla[1], Quentin Maronnier[3], Tala Palchan-Hazan[4], Olivier Caselles[2, 3*], John A. Kennedy[4, 5]

1. IUTC - Institut universitaire du cancer de Toulouse, Université de Toulouse, Toulouse, France

2. Tachyons SAS, Toulouse, France

3. Department of Medical Physics, Institut Claudius Regaud, Toulouse, France

4. Department of Nuclear Medicine, Rambam Health Care Campus, P.O.B. 9602, 3109601 Haifa, Israel

5. Ruth and Bruce Rappaport Faculty of Medicine, Technion Israel Institute of Technology, Haifa, Israel

***Correspondance :** Olivier Caselles (**caselles.olivier@iuct-oncopole.fr**)

ORCID IDS :

Sai Kiran Kumar Nalla : 0000-0002-8044-6330

Quentin Marronnier : 0000-0002-2266-9122

Tala Palchan-Hazan: 0009-0004-2632-1168

Olivier Caselles: 0000-0002-5404-5066

John A Kennedy**:** 0000-0002-2043-4397

Abstract:

Objective: Three-dimensional (3D) printing is widely explored to build cost effective, customizable and realistic phantoms in medical imaging but their adaptability across facilities is under explored. Evidence on applicability of standardization protocols and quantitative metrics of positron emission tomography (PET) to printed phantoms is minimal. This study aims to evaluate the harmonization and transferability of one such 3D printed phantom design across two facilities.

Approach: Phantom inserts (40mm) featured with spherical targets (0, 10, 15, and 20 mm) following a prior common design, were fabricated at two different facilities using fused deposition modeling (FDM) printers and positioned within a Jaszczak cylinder phantom, filled with radiotracer and soap/water solution. Acquisitions were performed using two PET/CT systems on different days with common protocols. Quantitative metrics, including Hounsfield Units (HU), Recovery Coefficient (RC) and TBR were extracted and statistically analysed to assess the influence of facility, scanner, and acquisition day.

Main Results: CT images revealed minor differences across facilities and scanners, with notable HU improvement upon prolonged immersion. HU_{mean} showed significant differences for facility ($p = 0.002$) and acquisition day ($p < 0.001$), but not scanner ($p = 0.87$). PET images successfully captured contrasts and targets. Average TBR_{Max} (mean±std. dev.) were 1.86 ± 0.03 (expected 2.0), 2.35 ± 0.05 (expected 2.5), and 3.10 ± 0.05 (expected 3.33). RC_{Max} and RC_{Mean} were similarly evaluated. No statistical differences were found for TBR_{Max}, RC_{Mean}, or RC_{Max} ($p \geq 0.07$ for Mann Whitney; $p \geq 0.57$ for Friedman) and bland-altman mean bias $< 10\%$, though wilcoxon post-hoc tests identified few outliers.

Significance: Our study demonstrates that a simple shared design file allows for easily reproducible 3D-printed phantoms, with consistency in PET quantitative metrics. This establishes crucial transferability and standardization, offering a new, easy way to share robust phantom designs globally.

Keywords: Quantitative PET/CT, Rapid prototyping, Radionuclide imaging, Standardization, Quality Control, Additive manufacturing

1. Background:

Quantitative Measurements of molecular level biological processes are key tools for diagnosis of diverse diseases in clinical oncology [1,2]. Positron emission tomography (PET) in conjunction with computed tomography (CT) is a widely used modality due its ability to provide both structural and functional information at high quantitative accuracy [3]. Consistency and reliability of PET/CT systems is ensured by rigorous quality control (QC) standards provided by the National Electrical Manufacturers Association (NEMA) [4], the European Federation of Organisations for Medical Physics (EFOMP) [5], and the European Association of Nuclear Medicine Research 4 Life (EANM EARL) [6] by means of test phantoms. Such standards are used to facilitate harmonization and transferability across facilities. Harmonization is generally focused on standardising acquisition protocols, reconstruction algorithms and quantitative analysis methods ensuring minimal inter-scanner variability [7–9]. Whereas, transferability ensures development of robust phantoms and data reproducibility across facilities [7,8,10].

Recent advancements in PET/CT systems have enabled more precise imaging especially in terms of spatial resolution [11,12]. Additionally, the rate of adoption of artificial intelligence (AI) into clinical workflows is rapidly increasing resulting in faster and accurate diagnostic interpretation [13,14]. Some other significant advances include the growing contribution of novel radiotracers and greater application of quantitative PET/CT for personalized medicine and theranostics [15,16]. While existing standards provide a foundation, this fast paced innovation calls for new phantom designs and capabilities to push the limits of available testing techniques. Furthermore, phantoms designed to simulate complex anatomical structures and varying levels of tracer uptake are the need of the hour.

Three dimensional (3D) printing is emerging as a powerful tool to create customizable and anthropomorphic phantoms [17]. Furthermore, it allows precise control over geometry and material composition which are essential for transferability. Various 3D printing based studies have confirmed that both lesions and tissue uptakes can be modelled for scanner assessment. One such methodology is the usage of permeable grids which can be filled with radioactive tracer [18,19]. The internal structure of the grid based on the porosity level can modulate the activity concentration (AC) thus allowing for the creation of varying radiotracer contrast within a single phantom with single tracer injection [20]. Multiple studies in both PET and SPECT imaging confirmed the compatibility of such inserts with existing simulation methods [18]. Also, 3D printing has been used to create simulated lesions with heterogeneous uptake, aiding in the development of radiomic metrics to bring them one step closer to clinical implementation [21]. Additionally, anthropomorphic phantoms mimicking patient geometries that can aid in testing AI algorithms were also produced using this technology [22]. Apart from that, 3D printing has enabled the physical creation of complex digitally-designed phantoms with many

features, allowing for a significant reduction in the number of experiments that might need to be performed, aiding the clinical application of new protocols.

A key challenge for adoption of 3D printing is the lack of multi-facility studies to illustrate the ability to rapidly and cost-effectively distribute 3D printed phantoms. More comprehensive studies especially from molecular imaging perspective are needed to establish harmonization protocols and data transferability. Material standardization is paramount, requiring reproducible results across different printers especially the characterization of attenuation properties and radiotracer activity distribution. Geometric accuracy is equally critical, necessitating strict quality control during the printing process to avoid dimensional variations that can impact quantitative measurements. A bigger challenge though for widespread adoption of 3D-printing is lack of universally accepted metrics. The choice and variability in metrics can have significant impact on the interpretation of results especially in multi-center studies [23]. Standardized uptake value (SUV) is a widely used metric in which the radiotracer concentration per millilitre is normalized by the total activity per weight (as a volume surrogate) in grams [7]. Consequently, the average SUV of a patient, or of the background of a phanmtom, is 1 g/ml. The adaptation of existing SUV thresholds to 3D printed grids can be impractical due to contrast bias provided by the grids to the expected value [20], thus other metrics are to be explored.

For harmonization between sites, one can rely on each site to accurately prepare the test phantoms to the prescribed contrasts, or one can ship an active phantom from site to site. The former method introduces variability by requiring several activity and/or volume measurements, whereas the latter requires long-lived radioisotopes and special shipping. Printed phantoms lend themselves to having the "same" phantom shipped digitally, with pre-determined contrast features. In this study we perform a dual-facility investigation with phantom design specifications provided by [20]. The main goal was to evaluate both harmonization and transferability concurrently. Under the same design, phantom preparation and acquisition protocols we attempt to quantify the variability existing in PET/CT measurements between different days, scanners, and 3D printers. Moreover, we evaluate the reproducibility of measured quantitative parameters derived from standard image analysis such as recovery coefficients and target-to-background ratios to establish their utility for 3D printing based studies.

2. Methods:

2.1. 3D Printed Inserts Fabrication, Phantom Preparation and Image Acquisition Protocols:

Phantom design specifications were followed based on a recent study [20]. Gyroid cubic inserts (40mm) designed in Autodesk Fusion 360 θ with variations in solidity S (50, 60, 70%) and pore distance D_P (i.e. unit cell distance: 5, 7, 10 mm) were used. Spherical target voids T (10, 15, 20 mm) were incorporated to assess quantitation and detectability. The total set contained 36 cubes of which 9 were homogeneous (also referred as reference cubes) and 27 with targets.

Additionally, cube holders of a 3×3 configuration were also printed. Detailed design parameters, including rationales

behind S, D_P, and T configurations, as well as the insert support structure design, are discussed in **[20]**. The primary

objective of this study was to evaluate the transferability of 3D-printed inserts across different facilities using a single

standardized design file. The inserts designed were exported in .STL file format which acted as the common shared design

file. Both facilities used their own FDM printers: a Prusa MK3S+ Printer (Prusa Research, Prague, Czech Republic) for

Facility –1 and a Stratasys F170 (Stratasys, Rehovot, Israel) for Facility – 2, with acrylonitrile butadiene styrene (ABS) as

the material. For the phantom preparation, specifications from same reference were used. The cubes were placed in a

cylinder from a Jaszczak phantom and filled with soap and water solution 4-6 hours prior to injection. Two studies

corresponding to Day 0 and Day 8 were performed at each facility. Injection on each acquisition day was carried out 45-

55 mins prior to scan. Tracer was administered by removing 10-50ml of background solution close to the phantom cap

prior and was added back prior to scan. The effective activity (**Table 1**) at scan time was maintained as the base criteria

for standardisation between facilities. At both facilities, image acquisitions were performed on 5-ring Discovery MI and a

32 cm OMNI Legend (*GE Healthcare, Chicago, IL, USA*). Acquisition parameters were maintained across facilities for

each scanner. A helical CT scan using three different slice thickness reconstructions were acquired for each scanner –

2.5mm CT (clinical), 1.25mm CT (high resolution) and a 3.5mm CT for the PET attenuation correction. The clinical CT

was the main reconstruction presented in this study. All clinical CT's were acquired at 120 keV with a 512×512 matrix

and a 50 cm transverse field of view (pixel spacing 0.98mm) for both scanners. PET acquisition parameters are presented

in **Table 1**. The scans were performed first on DMI followed by OMNI with a 25 minute break in between. All PET

acquisitions were carried out across 5 consecutive frames. To equalize for different count rates between scanners, the scan

duration for each OMNI acquisitions was scaled based on the effective decay and scan duration used for DMI. The exact

scan durations are laid out in **Table 1**.

138   **Table 1.** PET Acquisition and reconstruction parameters used by both facilities

Parameter	DMI	OMNI
Matrix Size	256×256	
Slice Thickness	2.8mm	2.07mm
No. of Slices	89	153
Scan Time	5×300 s	5×420 s
Reconstruction Diameter	600 cm	
Pixel Dimensions	[2.34 2.34] mm	
Voxel Volume	15.43 mm^3	11.33 mm^3

Reconstruction Algorithm	QCFX (time-of-flight BPL*)	QCHD (BPL*)
Beta	500	350
Injected Activity at Scan in MBq	Facility – 1: 51.0 (*Day 0*), 68.3 (*Day 8*) Facility – 2: 51.1 (*Day 0*), 49.9 (*Day 8*)	
Effective Activity at Scan in MBq	Facility – 1: 37.1 (*Day 0*), 37.1 (*Day 8*) Facility – 2: 37.7 (*Day 0*), 37.5 (*Day 8*)	

*Bayesian penalized-likelihood reconstruction with regularization parameter beta.

(a)   **(b)**

Figure 1. Jaszczak cylinder phantom with 36 cube inserts assembled in 4 layers prior to filling with a soap and distilled water solution – (a) Facility 1 , (b) Facility 2 – exact cube layout for each layer can be found in **[20]**

2.2 Data Extraction and Analysis:

Visual comparison of CT and PET Images was performed for qualitative differences. From Volume Viewer - PETVCAR® (*GE Healthcare, Chicago, IL, USA*), DICOM Images were exported for quantitative data extraction. Hounsfield Units (HU) for CT were extracted manually in 3D Slicer based on workflow used in **[24]**. Quantitative PET data extraction was carried out in MATLAB (*MathWorks INC, MA, USA*) using a semi-automatic program developed in house. The best transverse slice for each layer was identified automatically for meticulous region of interest (ROI) placement. Exact ROI placement is detailed out in **Figure 2**. Target ROI size is based on converted nearest pixel size from corresponding theoretical dimensions. To mitigate spill-out artefacts and exclude the gradient region where partial volume effects are prominent, a second ROI dilated from target ROI was positioned, and a no measurement zone was established based on Boolean subtraction of the 2 ROI's. For non-target cubes, 20mm was the reference size for ROI. The extracted activity concentration (kBq/ml) were then converted to standardized uptake value (SUV) with phantom solution weight of 4.5 kg based on Equation 1, with the activities decay corrected to the acquisition start time.

$$\text{SUV}_{(mean / max)} = \frac{\text{Activity Concentration Inside ROI in } \frac{\text{Bq}}{\text{ml}} \text{ (Mean/Max)}}{(\text{Injected Activity in Bq})/(\text{Weight of Phantom in g})} \quad (1)$$

SUV_{Mean} and SUV_{Max} were calculated for the entire data. Background SUV of the solution was averaged over 4 extracted

values for each layer. Subsequently to quantify the contrast, target to background ratio (TBR_{Max} and TBR_{Mean}) were

calculated using equation 2. Expected TBR values based on design parameters (S: 50, 60, 70) are 2, 2.5 and 3.33.

$$TBR_{(mean/max)} = \frac{\text{SUV Inside Target ROI (Mean/Max)}}{\text{SUV in the Adjacent Background Mesh (Mean)}} \quad (2)$$

For target quantitation, as per standard practice, recovery coefficient (RC_{Max} and RC_{Mean}) were also calculated using

Equation 3. Expected value of RC for targets is 1 and for homogeneous cubes is 0.5, 0.4, 0.3 for S: 50, 60, 70.

$$RC_{(mean/max)} = \frac{\text{SUV Inside the Target ROI (Mean/Max)}}{\text{SUV in the Adjacent Background Water ROI (Mean/Max)}} \quad (3)$$

167
168 **Figure 2.** Phantom ROIs. CT: (a) 25 mm Spherical VOI Placement centred on axial plane for HU extraction; (b) ROI

169 placement for each layer on transverse PET slice – background water (yellow), target (orange), homogeneous core (blue),

170 background mesh non target (red) and background mesh target (violet); (c) Exploded view for non-target inserts – central

171 core (region A), corner mesh (region B); (d) Exploded view for target cubes – target (region – 1), no measurement zone

172 (region 2), background mesh (region 3).

173
Data extraction was carried out on each frame individually and then averaged over 5 frames for each acquisition set. This

study investigated the influence of three key parameters on quantitative results: facility (printer), scanner, and acquisition

day. We utilized a hierarchical grouping method to determine the effect of each parameter. For each set, first normality

test was performed based on a Shapiro Wilk test followed by non-parametric Mann Whitney test for statistical differences

in data distribution [24]. Additionally, Friedman tests [20] were also carried out to assess pairwise discrepancies within the

datasets. Cross facility validation also followed all the aforementioned tests with the addition of a post hoc Wilcoxon sign

rank test with Bonferroni adjustment to identify outliers [25]. All the statistical tests were performed in JASP (*JASP Team,*

Netherlands) and R Studio (*PBC, Boston, MA*) with significance level of 0.05 for the *p values* and ± 10% for the bias and

variability.

3. Results:

3.1 CT HU Interpretation:

Visually minimal differences in CT across facilities and scanners can be seen in **Figure 3a**. The improvement in CT image

quality based on prolonged immersion is also noticeable for all images. Day 8 results of Facility-1 for D_P: 5mm are better

than Facility -2 with minimal target visibility. Notably the HU_{Min} value within the ROI, revealed the presence of small air

bubbles (approximately 2 voxels in size) with very low HU values (-800 to -925) for cubes: 70-5-10 and 70-5-20 on Day

0 (Facility -2), which dissolved over immersion period by Day 8 (**Figure A1**). Statistical tests yielded *p* values < 0.001 for

Shapiro Wilk establishing non-normality also as seen in **Figure 3b** and **3c**. Mann-Whitney test resulted in statistically

significant HU_{Mean} based on facility (*p value – 0.002*), acquisition day (*p value < 0.001*) and no differences based on

scanner (*p value – 0.87*). For DMI at Facility 1, the mean HU was -11.65 ± 14 on Day 0 and changed to -10.8 ± 14 on Day

8. At Facility 2, the initial value was -11.89 ± 14, changing to -6.61 ± 14. For OMNI, the mean values at Facility 1 were -

11.89 ± 14 (*Day 0*) and -10.68 ± 14 (*Day 8*) while at Facility 2 they were -11.61 ± 14 (*Day 0*) and -6.31 ± 14 (*Day 8*).

3.2 PET Quantitation:

For the PET, all slices were able to capture the contrasts and targets visually as illustrated in **Figure 4**. The target to

background ratios are presented in **Table 2**. For Facility 1, we observed the average noise in the image to be 4.19% for

both scanners on Day 0 and 4.54 and 4.3% for DMI and OMNI respectively on Day 8. The average noise values for each

scanner respectively at Facility 2 were 3.64 and 3.57% on Day 0, 3.45 and 3.56% on Day 8. SUV variability along the

phantom axis based on pairwise variability between layers was well within threshold of 10% with maximum variability of

4.1% (Facility - 1) and 8.7% (Facility - 2). Based on the values in **Table 2,** Cross facility TBR_{Max} (Mean ± SD) for 50%

contrast cubes ($TBR_{Theoretical} \sim 2$) were 1.82 ± 0.24 and 1.85 ± 0.23 for DMI. The values of same contrast for OMNI were

observed to be 1.85 ± 0.21 and 1.84 ± 0.23. Inserts with a $TBR_{Theoretical}$ of 2.5 resulted in values (Facility 1, Facility 2) of

2.13 ± 0.31 and 2.23 ± 0.34 for DMI, 2.22 ± 0.32 and 2.23 ± 0.39 for OMNI. The 70% solidity cubes with $TBR_{Theoretcial}$ of

3.33 yielded DMI values of 2.8 ± 0.5 and 2.77 ± 0.48 while OMNI values were 2.89 ± 0.53 and 2.81 ± 0.5. RC_{Max} and

RC_{Mean} were averaged for various insert sets with respect to target sizes for each facility and are presented in **Table 2**.

Additionally, the RC values were grouped together, averaged for each scanner and plotted in **Figure 4**.

Figure 3. CT analysis. (a) Black areas are air. Transverse slice of CT for both scanners and facilities viewed on abdominal windowing (W: -40, L: 400; abdominal windowing); raincloud HU_{Mean} data distribution plots for both scanners separated as function of (b) Facility; (c) Acquisition day

3.3 Parametric statistical tests:

Statistical tests on parameters (day, facility and scanner) resulted in non – normality and no significance was observed for the 3 PET metrics (TBR_{max}, RC_{mean} and RC_{max}) based on any grouping parameter. Mann Whitney p-values were ≥ 0.07 for most cubes except for RC_{mean} of reference cubes across scanners (~ 0.04). Additionally, parametric friedman pairwise analysis resulted p values of ≥ 0.72 (TBR_{Max}), ≥ 0.76 (RC_{Mean}) and ≥ 0.63 (RC_{Max}) for target cubes and ≥ 0.57 (TBR_{Max}), ≥ 0.07 (RC_{Mean}) and ≥ 0.16 (RC_{Max}) for reference cubes. Data distribution of the grouped metrics are represented in **Figure 6 and Figure 7**. Both centres observed no statistical differences with all intra facility *p – values > 0.05* and intraclass

correlation (ICC) values > 90%. Cross facility statistical results are summarised in **Table 3**. All the metrics have $p > 0.05$

thus no statistical differences for both median distribution and pairwise Friedman. Bland Altman mean bias fall within the

10% threshold but the SD band width varies for each pair. Post hoc Wilcoxon revealed some outliers with *p < 0.05* (**Table**

3) despite alpha rate Bonferroni adjustment.

Figure 4. Central Axial Slice of PET for both scanners: Day 0 (a) Facility -1; (b) Facility -2 and Day 8 (c) Facility -1; (d)

Facility -2 *(Note: For Layer -3, the 20mm (3 × 1) and 15mm (3 X 3) cubes of D_P:10 are misplaced between facilities);*

RC_{mean} averaged on over all data as function of target size (Bottom Left); RC_{max} averaged on over all data as function of

target size (Bottom right)

Table 2. PET Quantitative Metrics Calculated from Extracted Data

Insert Set (Solidity- Target Size)	TBR$_{Max}$ Day 0 (Mean ± SD); Day 8 (Mean ± SD)			
	Facility - 1		Facility - 2	
	DMI	OMNI	DMI	OMNI
50-10	1.53 ± 0.09	1.59 ± 0.07	1.53 ± 0.10	1.54 ± 0.06
	1.51 ± 0.05	1.59 ± 0.09	1.60 ± 0.06	1.57 ± 0.04
50-15	1.99 ± 0.08	1.91 ± 0.03	1.96 ± 0.06	1.91 ± 0.08
	1.89 ± 0.09	1.91 ± 0.04	1.94 ± 0.13	1.92 ± 0.05
50-20	2.03 ± 0.12	2.08 ± 0.09	2.00 ± 0.05	2.04 ± 0.04
	1.98 ± 0.04	2.02 ± 0.11	2.07 ± 0.04	2.06 ± 0.03
60-10	1.74 ± 0.12	1.81 ± 0.06	1.80 ± 0.09	1.75 ± 0.10
	1.72 ± 0.05	1.83 ± 0.11	1.79 ± 0.13	1.74 ± 0.10
60-15	2.33 ± 0.17	2.34 ± 0.18	2.37 ± 0.08	2.36 ± 0.08
	2.23 ± 0.27	2.40 ± 0.17	2.40 ± 0.09	2.38 ± 0.11
60-20	2.41 ± 0.07	2.50 ± 0.08	2.52 ± 0.03	2.54 ± 0.07
	2.34 ± 0.08	2.43 ± 0.08	2.49 ± 0.04	2.61 ± 0.09
70-10	2.16 ± 0.07	2.27 ± 0.28	2.09 ± 0.07	2.14 ± 0.19
	2.19 ± 0.17	2.19 ± 0.27	2.21 ± 0.17	2.17 ± 0.12
70-15	3.02 ± 0.10	3.03 ± 0.06	3.07 ± 0.08	3.16 ± 0.11
	3.07 ± 0.25	3.17 ± 0.19	3.12 ± 0.05	3.14 ± 0.12
70-20	3.18 ± 0.08	3.35 ± 0.10	3.05 ± 0.09	3.08 ± 0.08
	3.22 ± 0.13	3.34 ± 0.04	3.08 ± 0.08	3.16 ± 0.12
Insert Set (Solidity- Target Size)	Averaged RC$_{Max}$ (Mean ± SD); RC$_{Mean}$ (Mean ± SD)			
	Facility - 1		Facility - 2	
	DMI	OMNI	DMI	OMNI
50-10	0.82 ± 0.02	0.81 ± 0.03	0.83 ± 0.04	0.81 ± 0.03
	0.73 ± 0.01	0.72 ± 0.02	0.74 ± 0.03	0.72 ± 0.01
50-15	1.04 ± 0.04	0.99 ± 0.03	1.02 ± 0.05	0.98 ± 0.07
	0.86 ± 0.03	0.82 ± 0.03	0.84 ± 0.03	0.81 ± 0.05
50-20	1.09 ± 0.02	1.09 ± 0.03	1.08 ± 0.02	1.07 ± 0.02
	0.89 ± 0.01	0.88 ± 0.02	0.89 ± 0.02	0.87 ± 0.03
60-10	0.77 ± 0.03	0.76 ± 0.02	0.77 ± 0.05	0.70 ± 0.05
	0.67 ± 0.01	0.66 ± 0.01	0.68 ± 0.04	0.61 ± 0.03
60-15	1.02 ± 0.11	1.02 ± 0.07	1.03 ± 0.03	1.00 ± 0.03
	0.84 ± 0.14	0.83 ± 0.11	0.81 ± 0.02	0.77 ± 0.02
60-20	1.06 ± 0.05	1.06 ± 0.04	1.11 ± 0.01	1.09 ± 0.04
	0.80 ± 0.14	0.79 ± 0.12	0.87 ± 0.01	0.86 ± 0.02
70-10	0.73 ± 0.05	0.71 ± 0.09	0.74 ± 0.06	0.69 ± 0.05
	0.62 ± 0.04	0.59 ± 0.07	0.62 ± 0.04	0.56 ± 0.04
70-15	1.04 ± 0.02	1.02 ± 0.04	1.05 ± 0.01	1.01 ± 0.05
	0.78 ± 0.02	0.76 ± 0.02	0.78 ± 0.01	0.74 ± 0.03
70-20	1.08 ± 0.05	1.07 ± 0.07	1.10 ± 0.02	1.08 ± 0.02
	0.84 ± 0.03	0.83 ± 0.01	0.85 ± 0.02	0.83 ± 0.02

238
239
240 **Figure 6.** Raincloud Plots for 3 quantitative PET metrics of Target Cubes – TBR_{Max}, RC_{Mean} and RC_{Max} (Rows) Data

241 assorted together based on Facility, Scanner, and Acquisition Day (Columns); (Note: *Density Data has been overlaid on*

242 *each other for each plot with respect to its parameter to demonstrate the degree of overlap*

243
244
245
246
247 **Figure 7.** Raincloud Data Distribution Plots for 2 quantitative PET metrics for Non Target Cubes– Rows: RC_{Mean} and

248 RC_{Max}; Data assorted together based on – Columns: Facility, Scanner, and Acquisition Day. (*Expected Values are 0.3,*

249 *0.4, 0.5*)

Table 3. Cross Facility Statistical Tests Summary for the PET Metrics of Targets

Comparison Pair	TBR_{Max}			RC_{Max}			RC_{Mean}		
	Mann Whitney	Wilcoxon Bonferroni (True, Adjusted)	Bland Altman	Mann Whitney	Wilcoxon Bonferroni (True, Adjusted)	Bland Altman	Mann Whitney	Wilcoxon Bonferroni (True, Adjusted)	Bland Altman
DMI -1 vs DMI - 2	1.00	0.16, 0.64	2.4 (21, -16)	0.86	0.81, 1	0.6 (11, -10)	0.55	0.05, 0.2	-1.5 (16, -13)
OMNI -1 vs OMNI - 2	0.79	0.09, 0.36	-1.4 (11, -14)	0.47	0.03, 0.12	-1.9 (10, -13)	0.72	0.02, 0.08	-1.4 (13, -16)
DMI -1 vs OMNI - 2	0.59	0.15, 0.60	2.5 (20, -15)	0.21	<0.001	-3.1 (8, -14)	0.24	<0.001	-3.3 (13, -19)
DMI -2 vs OMNI - 1	0.33	0.10, 0.40	1.5 (14, -11)	0.33	0.06, 0.24	-1.7 (9, -13)	0.15	<0.001	-3.4 (11, -18)

DMI 1, OMNI 1 correspond to combined data of Facility -1 and DMI 2, OMNI 2 correspond to combined data of Facility – 2

Mann Whitney and Wilcoxon Sign Rank with Bonferroni Correction represents the *p values*

Bland Altman represents % deviation (Mean Bias, ± 1.96 SD)

4. Discussion:

We evaluated the standardization and transferability of 3D-printed inserts based on the design specifications outlined by [20]. This study was carried out across two different facilities, where the shared elements prior to image acquisition were solely the design files (.STL) and the phantom preparation protocol. The quantitative metrics presented a comprehensive assessment of cross facility consistency and potential variability in both CT and PET imaging employing two different scanners (DMI and OMNI). To our knowledge, this is the first time that a PET phantom with pre-determined contrast ratios was transmitted electronically between sites and successfully scanned.

From harmonization perspective, the explicit non visibility of targets is obligatory, especially if visual detectability is the goal, as mentioned by [20]. Although transaxial CT slices (**Figure 3**) showed no significant visual changes, with this regard Day 8 for Facility-1 performed better over Facility-2. The change in HU_{Mean} values by Day 8 from an identical starting point on Day 0 underscores significance of the facility, most likely due to the printer used. This significance is further quantified by *p-values* of 0.002 and 0.0001 observed between the different facilities and acquisition days. The statistical difference across acquisition days is expected and is inherent to the study design. While [20] suggested the phantom's

immersion history as a key criterion for CT attenuation, our observations highlight its amplified importance in relation to

the printer used. Even while using the same printer and settings, the quality of the 3D-prints, and consequently their

attenuation properties can vary. Additionally, with wear of the printed objects over time, the immersion time needed also

changes. Although minimal, the presence of small air bubbles within 2 targets on Day 0 for Facility-2 is undesirable. The

importance of the chosen immersion strategy is emphasized again, as these bubbles effectively dissipate by Day 8. The

lack of statistical significance in HU_{Mean} based on scanner is an encouraging finding. Within the framework of this study,

both scanners provide comparable results thus reducing a potential source of variability when combining data acquired on

different models. Finally, for cross facility harmonization if equivalent HU_{Mean} is desired, the immersion period needs to

be tailored based on the printer or filament manufacturer. Furthermore, the combination of HU_{Mean} within a range of ± 25

HU_{Water} and qualitative visibility, is a better harmonization criterion with respect to CT performance. Conversely, if the

objective is primarily localization, the effective attenuation correction of PET for 3D-printed grids from CT is sufficiently

accurate even for high density materials [18].

PET visual contrasts and target visibility are in agreement with the base phantom design [20] and indicated the

fundamental PET functionality at both facilities. The noise levels within the image are also below the 5% threshold used

in standard practice [5,6]. SUV variability along the phantom axis based on pairwise variability suggested uniform mixing

for both facilities. This approach was adopted because the acceptable threshold (10%) [5] needed to be consistently

achieved regardless of layer location as this is crucial for accurate target assessment without bias. We adapted the PET

preparation protocol to suggestions made by [20] by removing and adding back certain volume of the solution during

injection and mixing. This change was implemented as we fixed the post injection decay time (45-55 minutes), whereas

the original protocol didn't have such criteria. This reduced SUV variations in the background and increased reproducibility

for the same injected activity. Additionally, adjusting the scan duration for the activity at scan time also played a crucial

role in this regard and can help mitigating injection errors.

We observe that the of mean values of practically achieved TBR_{Max} by 3D printed grids (**Table 3**) are slightly

lower than the true values, although they remain within the range defined by the standard deviation. In our study this is

also attributed to significant partial volume effects in the 10 mm target cubes. Notably, standard deviation increases with

increase in contrast because the percentage of activity inside the mesh of the inserts reduces. Such bias was also previously

reported in grid studies [18–20]. However, a positive finding is that this bias is consistently captured across facilities and

scanners (**Table 2 and Figure 7**). Hence, the reproducibility of the bias suggests the potential for a global

correction/calibration factor(s) determined based on grid design parameters. Furthermore, we did not report the TBR_{Mean}

for the same reason as it would be unreliable metric for cross facility comparisons. As mentioned previously, we employed

a hierarchical statistical assessment for the PET metrics (TBR_{Max}, RC_{Mean}, RC_{Max}) with initial tests focused on the individual impact of facility, acquisition day, and scanner. This was followed by median tests on intra-facility and inter-facility measurements. Finally, pairwise tests provided insights into performance relative to differences individualized for each insert. The rationale behind such choice is to isolate sources of variability and identify bottlenecks to harmonization at each level to ensure consistency.

Agreement in TBR_{Max} values suggested by statistical *p-values* and mean percentage of variability ($< 10\%$) for different contrasts indicate the independence on sensitivity of these values with respect to acquisition day, facility and scanner. This is a critical to achieve quantitative PET harmonization and assessment of recovery coefficients. Building on this, statistical consistency observed in RC_{Mean} and RC_{Max} suggest that not only the physical setup but also the quantitation can be readily reproduced across facilities by simply sharing the design file and adhering to a common preparation protocol. One outlier observed was comparison of RC_{Mean} for reference cubes across scanners which showed significance in Mann Whitney ($p \sim 0.04$) but no significance in Friedman ($p \sim 0.07$). Additionally, the pairwise tests for some intra facility data warrant some attention to few outliers. More specifically, pairwise results of Facility-2 showed significance especially pertaining to a few Day 0 metrics which can be understood in conjunction with presence of air bubbles within 2 targets.

Moving to cross facility tests, global tests on the metrics both Mann Whitney and Friedman showed no statistical significance. Also the Bland-Altman mean bias fell within 10% threshold for all inter facility tests suggesting strong evidence for overall harmonization. TBR_{Max} was consistent even for pairwise Wilcoxon tests establishing the transferability of the phantom setup. In contrast, for recovery coefficients only 2 Wilcoxon tests resulted in acceptable *p values*. Furthermore, DMI 1 vs OMNI 2 resulted in bad correlations even with after alpha rate Bonferroni adjustment. This indicates that there are inherent statistical differences for certain cubes when 2 individual pairs are compared especially for different scanners. We filtered the metrics with a 10% absolute variability between pairs to quantify the outliers. In overall percentage, this was equivalent to 12% (26 outliers out of the 216 comparisons) of the RC_{Mean} data and 10% (21 outliers out of the 216) of the RC_{Max} data. The pair-wise *p-values* excluding outliers were all greater than 0.05.

Though we address the data extraction limitation of [20], the fundamental size of the ROI is still an issue. The strategy of using physical dimensions to create the ROI is not an ideal because of printing inaccuracies. This becomes even more challenging if heterogeneity is added to the targets. Hence an automatic segmentation method could be better suited for data extraction when more complexity is added. While we acknowledge the influence of target size on metrics as shown in **Figure 5**, a detailed statistical analysis in this regard was not performed. As mentioned previously, our main objective was to evaluate the harmonization and transferability of the 3D-printed inserts and protocols across different facilities and scanners, rather than to comprehensively investigate the impact of specific design parameters on PET metrics which was

already addressed in **[20]**. A key limitation for comparison across scanners were the count rates of DMI and OMNI were not the same since they were performed sequentially on same setup. This was due to the availability of scanners for experiments and also a different day for each device would make the CT results incomparable. Another limitation of this study is that consistency in the protocol with other radiotracers was not addressed.

The findings of this study emphasize the easy transferability of 3D printed phantoms across facilities with just a single standard design file. Utilisation of generic low-cost desk FDM printers further enhances the cost-effectiveness of this phantom design with average production cost less than 100 euros. With the capacity of 36 cubes per acquisition, this opens up the possibility of a wide spectrum of experiments with a single tracer injection. The major limitation that still exists in this study is the slight underperformance of grids in accurately generating intended contrast. Also, the effect of count rates (activity concentration) on this bias is to be explored. Moving forward, with further studies focusing on wide range of grid design parameters and printers this bias can be globally calibrated. Furthermore, exploring post-print processing techniques can reduce and standardize the immersion time needed for CT attenuation stability. Alternative printing materials, such as PLA, could also lead to improved tissue-equivalent performance on CT **[18]**; however, the immersion vs HU trade off needs an evaluation. Additionally, a dedicated target detectability study, encompassing a broader range of target sizes (e.g., 5mm to 35mm), is needed to benchmark the performance of these 3D-printed phantoms against established standards. Finally, relevance of the metrics used in this study should also be investigated when heterogeneity is introduced within the target shape or activity distribution.

Conclusion:

Based on a dual facility study, transferability of 3D printed targets for PET/CT assessment of varying radiotracer contrasts solely based on the transfer of a standardised design file has been comprehensively established. CT results showed some potential variability that needs to be addressed in the choice of the printer or manufacturer but the quantitative PET results showed a high degree of harmonization. With a shared preparation protocol, phantom results can be easily reproduced in terms of quantitative performance. We further demonstrated the compatibility of standard CT and PET metrics with 3D printed phantoms and also laid out some modifications to statistical guidelines by a hierarchical approach for such phantoms

Abbreviations:

ABS	Acrylonitrile Butadiene Styrene
AC	Activity Concentration in kBq/ml
DMI	Discovery MI
D_P	Distance between pores
FDM	Fused Deposition Modelling
FDG	Fluorodeoxyglucose
FOV	Field of View
HU	Hounsfield Units
NEMA	National Electrical Manufacturing Association
NM	Nuclear Medicine
PET / CT	Positron Emission Tomography / Computed Tomography
QC	Quality Control
RC	Recovery Coefficient
ROI	Region of Interest
S	Solidity
SLA	Stereolithography
SPECT/CT	Single Photon Emission Computed Tomography / Computed Tomography
SUV	Standardized Uptake Value
TOF	Time of Flight
TBR	Target to Background Ratio
VOI	Volume of Interest

References:

1. Sokolenko AP, Imyanitov EN. Molecular Diagnostics in Clinical Oncology. Front Mol Biosci [Internet]. 2018 [cited 2025 Mar 31];5. Available from: https://www.frontiersin.org/journals/molecular-biosciences/articles/10.3389/fmolb.2018.00076/full

2. Das S, Dey MK, Devireddy R, Gartia MR. Biomarkers in Cancer Detection, Diagnosis, and Prognosis. Sensors. 2023;24:37.

3. Trotter J, Pantel AR, Teo B-KK, Escorcia FE, Li T, Pryma DA, et al. Positron Emission Tomography (PET)/Computed Tomography (CT) Imaging in Radiation Therapy Treatment Planning: A Review of PET Imaging Tracers and Methods to Incorporate PET/CT. Adv Radiat Oncol [Internet]. 2023 [cited 2025 Mar 31];8. Available from: https://www.advancesradonc.org/article/S2452-1094(23)00041-6/fulltext

4. Houda Baghous NE, Kafhali ME, Tahmasbi M, Chakir EM, Kessioui AE, Elkhatib A, et al. Evaluating long-term
performance and quality control of the uMI 550 positron emission tomography- computed tomography (PET-CT) system:
A comprehensive scientific analysis. Radiography. 2025;31:102920.

5. Matheoud R, Boellaard R, Pike L, Ptacek J, Reynés-Llompart G, Soret M, et al. EFOMP's protocol quality controls in
PET/CT and PET/MR. Phys Med. 2023;105:102506.

6. Aide N, Lasnon C, Veit-Haibach P, Sera T, Sattler B, Boellaard R. EANM/EARL harmonization strategies in PET
quantification: from daily practice to multicentre oncological studies. Eur J Nucl Med Mol Imaging. 2017;44:17–31.

7. Akamatsu G, Tsutsui Y, Daisaki H, Mitsumoto K, Baba S, Sasaki M. A review of harmonization strategies for
quantitative PET. Ann Nucl Med. 2023;37:71–88.

8. Fahey F, Christian P, Zukotynski K, Sexton-Stallone B, Kiss C, Clarke B, et al. Use of a Qualification Phantom for PET
Brain Imaging in a Multicenter Consortium: A Collaboration Between the Pediatric Brain Tumor Consortium and the
SNMMI Clinical Trials Network. J Nucl Med. 2019;60:677–82.

9. Monsef A, Ay MR, Sheikhzadeh P, Geramifar P, Rahmim A, Ghafarian P. Harmonization based on quantitative analysis
of standardized uptake value variations across PET/CT scanners: a multicenter phantom study. Nucl Med Commun.
2022;43:1004.

10. Multicenter CT phantoms public dataset for radiomics reproducibility tests - Kalendralis - 2019 - Medical Physics -
Wiley   Online   Library   [Internet].   [cited   2025   Mar   31].   Available   from:
https://aapm.onlinelibrary.wiley.com/doi/10.1002/mp.13385

11. PET/CT imaging in cancer: Current applications and future directions - Farwell - 2014 - Cancer - Wiley Online Library
[Internet]. [cited 2025 Mar 31]. Available from: https://acsjournals.onlinelibrary.wiley.com/doi/10.1002/cncr.28860

12. Hussain D, Abbas N, Khan J. Recent Breakthroughs in PET-CT Multimodality Imaging: Innovations and Clinical
Impact. Bioengineering. 2024;11:1213.

13. Wang Y-R (Joyce), Baratto L, Elizabeth Hawk K, Theruvath AJ, Pribnow A, Thakor AS, et al. Artificial Intelligence
enables whole body Positron Emission Tomography Scans with minimal radiation exposure. Eur J Nucl Med Mol Imaging.
2021;48:2771–81.

14. Qi C, Wang S, Yu H, Zhang Y, Hu P, Tan H, et al. An artificial intelligence-driven image quality assessment system
for whole-body [18F]FDG PET/CT. Eur J Nucl Med Mol Imaging. 2023;50:1318–28.

15. Dhoundiyal S, Srivastava S, Kumar S, Singh G, Ashique S, Pal R, et al. Radiopharmaceuticals: navigating the frontier
of precision medicine and therapeutic innovation. Eur J Med Res. 2024;29:26.

16. Królicki L, Kunikowska J. Theranostics – present and future. Bio-Algorithms Med-Syst. 2021;17:213–20.

17. Filippou V, Tsoumpas C. Recent advances on the development of phantoms using 3D printing for imaging with CT,
MRI, PET, SPECT, and ultrasound. Med Phys. 2018;45:e740–60.

18. Jessen L, Gustafsson J, Ljungberg M, Curkic-Kapidzic S, Imsirovic M, Sjögreen-Gleisner K. 3D printed non-uniform
anthropomorphic phantoms for quantitative SPECT. EJNMMI Phys. 2024;11:8.

19. Theisen A-L, Lassmann M, Tran-Gia J. Toward a Patient-Specific Traceable Quantification of SPECT/CT-Based
Radiopharmaceutical Distributions. J Nucl Med. 2022;63:1108.

20. Nalla SKK, Maronnier Q, Hazan TP, Kennedy J, Caselles O. Enhancing PET/CT Assessment with Porous 3D Printed Grids: A Pilot Study [Internet]. medRxiv; 2025 [cited 2025 Apr 2]. p. 2025.04.01.25325022. Available from: https://www.medrxiv.org/content/10.1101/2025.04.01.25325022v1

21. Gallivanone F, Interlenghi M, D'Ambrosio D, Trifirò G, Castiglioni I. Parameters Influencing PET Imaging Features: A Phantom Study with Irregular and Heterogeneous Synthetic Lesions. Contrast Media Mol Imaging. 2018;2018:5324517.

22. Gear JI, Cummings C, Craig AJ, Divoli A, Long CDC, Tapner M, et al. Abdo-Man: a 3D-printed anthropomorphic phantom for validating quantitative SIRT. EJNMMI Phys. 2016;3:17.

23. Fahey FH, Kinahan PE, Doot RK, Kocak M, Thurston H, Poussaint TY. Variability in PET quantitation within a multicenter consortium. Med Phys. 2010;37:3660–6.

24. Maronnier Q, Courbon F, Caselles O. Evaluation of a method based on synthetic data inserted into raw data prior to reconstruction for the assessment of PET scanners. EJNMMI Phys. 2022;9:68.

25. Graef J, Senger C, Wetz C, Baur ADJ, Kluge AK, Lukas M, et al. Shortened Tracer Uptake Time in GA-68-DOTATOC-PET of Meningiomas Does Not Impair Diagnostic Accuracy and PET Volume Definition. Diagnostics. 2020;10:1084.

Appendix / Additional Files:

Figure A1. Snapshots of ROI with air bubbles on Day 0 and Day 8 for Facility 2