Assessing Cerebral Visual Impairment Through Non-Verbal Responses: Protocol Development and Feasibility of EchoCue™ [Registered Report, stage 1]

Author

Edwige Smague[1]

Affiliations:

[1]N.E.T.A: Neuronal Epileptotranscriptomic, Sydney, Australia

Contact:

edwige@netadiscovery.org

ORCID ID: 0009-0002-8182-7888

Abstract: This Stage 1 Registered Report outlines the protocol for a feasibility study of EchoCue™, a novel paradigm designed to assess cerebral visual impairment in children aged 2 to 6, through non-verbal response patterns. The paradigm minimises the reliance on verbal instruction or active cooperation, making it suitable for populations traditionally considered difficult to assess. The method pairs auditory and visual stimuli with the recording of reflexive responses, such as gaze shifts, blinks, and startle reactions, using an integrated infrared eye-tracking system. The primary objective of this study is to assess the feasibility of the data collection procedures and to generate robust estimates of effect sizes with 95% confidence intervals for the protocol's core measures, in order to inform the design of a future large-scale confirmatory trial.

Keywords: Cerebral visual impairment, screening tool, reflex response, autism differential diagnosis, non-verbal responses

Introduction

Cerebral visual impairment (CVI) is a major cause of childhood disability in Australia and high-income countries. Most existing screening tools rely on task-based protocols requiring active participation, rendering them inaccessible to the most severely affected children (1). The difficulty is compounded by the neurological origin of CVI involving damage to the visual pathways or the cortex, therefore necessitating evaluation methods that differ from traditional ophthalmic assessments (2).

This clinical challenge is further amplified by the symptomatic overlap with other neurodevelopmental disorders, such as autism spectrum disorder (ASD), including peculiarities in eye contact, stereotypical behaviours, and atypical responses to sensory stimuli, which can delay or even prevent the correct identification of visual impairments, thus depriving these children of appropriate and early care (3). There is, consequently, an urgent clinical and scientific need for an evaluation methodology capable of objectively testing the foundational sensory processes that underpin a child's perceptual abilities (4). Such an approach must minimise reliance on the child's active participation to provide objective markers of neurosensory integrity. EchoCue^TM has been designed as an

objective screening and assessment tool intended to provide critical

evidence within a comprehensive clinical evaluation.

Developed in this context, EchoCue^{TM} represents a new

methodological approach based on the presentation of calibrated

sensory stimuli (auditory and visual) and the synchronised capture

of reflex responses (eye movements, blinks, startles) using an

integrated eye-tracking system and a motion sensor. While the

primary aim of this study is to assess the method's initial

feasibility, future studies with larger cohorts will be required to

establish normative data.

Method

EchoCue™ is an original screening framework designed by

Edwige Smague in 2025 as part of a doctoral research initiative.

The device is not yet commercialised and is currently under

academic development for feasibility testing in Australia. The

methodology presented in this article is based on the EchoCue™

conceptual framework, which enables the objective screening of

reflexive and bottom-up neurovisual functions using strategically

calibrated   multimodal   sensory   stimuli.

Table 1. Research Questions, Variables, Planned Analyses, and Estimation Goals.

Research Question (RQ)	Exploratory Focus/Expectation	Key Variables	Planned Analyses	Estimation Goal/Feasibility Indicator
Do mean saccadic reaction times (SRT) differ across CVI, ASD, and TD groups?	Expect longer SRTs on average in the CVI group relative to ASD and TD groups.	IV: Group (CVI, ASD, TD) DV: SRT.	LMM estimation group differences in SRT with 95% CIs.	Report estimated effect sizes and 95% CIs for group differences; evaluate data completion and model convergence as feasibility metrics.
How does audiovisual enhancement of the startle reflex differ across groups (CVI, ASD, TD)?	Expect a smaller audiovisual enhancement effect in CVI vs other groups.	IV: Group Condition: Auditory-only vs Audiovisual DV: Startle Amplitude	LMM estimating Group × condition interaction with 95% CIs; follow-up estimated contrasts.	Report interaction effect and contrast estimates with 95% CIs, assess data quality and participant tolerance as feasibility indicators.
Can objective metrics (SRT, Startle	Expect measurable separation between CVI and non-CVI	Predictors: SRT, Startle Amplitude	Feature selection (LASSO) → ROC analysis with bootstrapped	Report estimated AUC 95% CI as preliminary indicator of screening

amplitude) differentiate CVI from non-CVI participants?	groups on combined metrics.	Outcome: Group Status (CVI / non-CVI)	95% CIs.	potential; evaluate feasibility via data completeness and classification stability.

Note. AUC: Area Under the Curve; DV: Dependent Variable; IV: Independent Variable; LMM: Linear Mixed-Effects Model; CI:

Confidence Interval; ROC: Receiver Operating Characteristic; SRT: Saccadic Reaction Time; TD: Typically Developing.

Sample Size Justification: The planned sample size (n=15 per group) is determined by recruitment feasibility for this initial proof-of-

concept study. A brief internal feasibility check (n=5 per group; CVI, ASD, TD) will be conducted at the outset to confirm task

tolerability, optimise calibration, and verify data-collection logistics before full recruitment. Although this study is exploratory, it is

designed to detect a large effect size (Cohen's d ≥ 0.8), as a clinically useful screening tool must be sensitive to clear, robust

differences in neurosensory function. Data from this study will provide effect size estimates to inform power calculations for future

validation trials designed to meet stricter statistical thresholds (e.g., 90% power). The primary goal of this study is to establish the

feasibility of the protocol and generate robust estimates of effect size and confidence intervals (CIs) to guide the design of a future

confirmatory trial. If the preliminary feasibility check indicates poor tolerability, calibration instability, or insufficient data capture

quality, protocol parameters (e.g., stimulus intensity, trial duration, eye-tracking calibration procedures) will be modified accordingly,

and feasibility will be re-verified before progressing to full recruitment. A preliminary power simulation (SIMR) indicates an

estimated power of approximately 0.70 for the planned comparison, consistent with the exploratory aim of this Stage I design.

Power Simulation

A power simulation using the SIMR package (Green & MacLeod, 2016) indicated an estimated statistical power of approximately 0.70 for the planned LMM analyses. As this value falls below conventional thresholds for confirmatory inference, the present study is therefore considered exploratory and estimation-focused.

Participants

A total of 45 participants, within the preschool age range of 2 to 6 years, will be recruited to ensure developmental homogeneity for this initial feasibility study. Participants will be divided into three age and sex-matched groups (n=15 per group): (1) children diagnosed with CVI; (2) a clinical control group with a diagnosis of ASD and no known visual impairment; and (3) a typically developing control group. All participants' adaptive functioning will be assessed and statistically controlled for (see the Measures and Statistical Model specification sections).

Feasibility Metrics

In line with the primary goal, feasibility will be evaluated using three primary indicators:

1. **Completion Rate**: The proportion of participants who complete all core testing blocks (1-3).

2. **Tolerance**: Observational coding of behavioural signs of distress (e.g., crying, agitation) during testing.
3. **Parents' Acceptability**: A brief post-session questionnaire rating the perceived comfort and ease of participation on a 5-point Likert scale (1=very low, 5=very high).

Items assess:

- Perceived comfort of the child during the task
- Clarity of instructions provided to caregivers
- Ease of participation
- Acceptability of session duration
- Willingness to participate again or recommend the procedure.

The mean score will be calculated per participant, with high acceptability defined as ≥4/5 at the group level.

Feasibility will be supported when ≥80% of participants complete all blocks, <10% discontinue due to distress, and the mean parent acceptability score ≥4/5.

Recruitment and Inclusion/Exclusion Criteria

Participants will be recruited from key paediatric institutions and support networks in Brisbane, Australia, including the Queensland Children's Hospital (QCH), Mater Children's Private Brisbane, National Disability Insurance Scheme (NDIS), Vision Australia, Autism Cooperative Research Centre (Autism CRC), Next Sense, Statewide Vision Impairment Services (via the Department of

Education Queensland), Queensland University of Technology

(QUT), and The University of Queensland (UQ).

This broad recruitment network, tapping into major clinical and

community support services, provides a realistic basis for

achieving the recruitment target of 15 per group for this specific

population. This target appears highly feasible; for context, a

recent study at just one of these Australian tertiary referral centres

successfully recruited 102 children with CVI or related

neurodevelopmental delays for a cross-sectional study (5).

Inclusion criteria for the CVI group will include: (1) a formal

clinical diagnosis of CVI from a qualified professional (e.g.,

paediatric neurologist or ophthalmologist), confirming the visual

impairment is of neurological origin; and (2) an age between 2 and

6 years.

Inclusion criteria for the ASD group will include: (1) a formal

clinical diagnosis of autism spectrum disorder from a qualified

professional (e.g., paediatrician, psychologist), based on

established criteria such as the DSM-5; (2) no known or suspected

visual impairment of either ocular or cerebral origin; and (3) an

age range between 2 and 6 years.

Inclusion criteria for the TD group will include: (1) no known

developmental, neurological, or psychiatric diagnoses; (2) no

parental or clinical concerns regarding vision, hearing, or general

development; (3) born full-term with no significant perinatal

complications; and (4) an age range between 2 and 6 years.

Notably, there will be no exclusion criterion based on cognitive or

intellectual level (e.g., IQ score <70). A key objective of this study

is to develop and assess the feasibility of a tool for the full range of

children with CVI, including those with co-occurring intellectual

disability.

Exclusion criteria for all groups will include: (1) the presence of a

severe, uncorrected ocular pathology that could primarily account

for visual deficits; (2) a diagnosed severe or profound hearing

impairment that would prevent perception of the auditory stimuli;

or (3) an unstable medical condition precluding participation in a

15-minute session.

Written informed consent will be obtained from the parents or

legal guardians of all participants.

Measures

Developmental and Adaptive Functioning

To assess the developmental and adaptive functioning level of each

participant, parents or legal guardians will complete the Vineland

Adaptive Behavior Scales, Third Edition, Vineland-3 (6).

The Vineland-3 is a standardised assessment completed by parents

or primary caregivers that evaluates communication, daily living

skills, socialisation, and motor skills. The Adaptive Behavior

Composite score will be used as a covariate to control for

developmental differences across groups.

This measure was deliberately selected because, as a parent-report

tool, it minimises the performance-based demands that are often

prohibitive for young or clinically affected children, thereby

aligning with the protocol's feasibility objective. Although the

Vineland provides an indirect rather than performance-based

index, it represents a robust and pragmatic method for statistically

controlling overall developmental level in this population.

Theoretical Rationale for Using Auditory Probes to Assess

Neurovisual Function

The EchoCue^{TM} protocol is based on an objective assessment

method that tests the child's neurosensory functions without the

need for the child to understand instructions or actively participate

(7). The approach involves exposing the child to a series of finely

calibrated auditory and visual stimuli while synchronously

recording their involuntary reflex responses (8), such as blinks,

micro-saccades, or startles, and visual orientation responses.

A key element of this method is using a wide range of sound

frequencies, from 20 to 8000 Hz. Each frequency range allows

distinct neurological circuits, both subcortical and cortical, to be

explored in order to better understand how the brain perceives,

integrates, and responds to multisensory stimuli (9). Low-

frequency sounds (20—250 Hz) act as a subcortical marker. This

range is particularly relevant for assessing reflexive response in

non-instructional paradigms (10).

These frequencies are perceived by the human body as both

vibrotactile stimulation and sound. They are transmitted partly by

bone conduction, directly stimulating the cochlea and vestibular

system, thus bypassing any disorders of the outer or middle ear

(11). More important still, they are processed primarily by the

archaic neural circuits of the brainstem due to their simple acoustic

configuration and their strong ability to arouse sensory alert

systems (12, 13). They are particularly effective at triggering

involuntary reflexes managed by the brainstem, such as the

stapedius reflex (an involuntary auditory reflex that protects the

inner ear from loud sounds) and the startle reflex (14, 15). The use

of these frequencies thus makes it possible to measure a basic

neurological response, a sign of the integrity of the subcortical

pathways, even before the information reaches the higher cortical

areas. This approach is consistent with findings from auditory

neuroscience showing that low-frequency auditory stimuli can

promote autonomic regulation, including parasympathetic

activation, through mechanisms such as vagal nerve modulation

(16). These effects have been associated with reduced anxiety and

increased physiological markers of calm in both clinical and

developmental populations (17, 18).

Beyond their function as a subcortical probe, the use of these low-

frequency sounds is also informed by emerging research on their

potential therapeutic effects (18). Studies have suggested that

certain frequencies can have a regulating or calming effect on the

nervous system (19), which may help in managing anxiety in

paediatric populations, including children with ASD during

stressful procedures (20). This dual function, as both an objective

probe and a potentially regulating stimulus, is a key feature of the

EchoCue™ protocol's design to maximise participant comfort.

Medium to high frequencies (250—8000 Hz) are used for cortical

analysis. This range is essential for assessing the classical auditory

pathway and higher cognitive functions (21, 22). It is necessary for

multisensory integration tasks, where the brain has to compare and

merge auditory information with visual information (23).

Difficulty processing these sounds or integrating them correctly

with a visual stimulus may reveal dysfunction in associative

cortical areas (e.g., the temporoparietal junction), which are often

involved in CVI (24).

To establish a reliable baseline of involuntary orienting behaviour,

the EchoCue™ protocol begins with an auditory-only condition

(Block 1) using lateralised, high-intensity sounds (25). This

condition is not intended to elicit voluntary saccadic eye movements alone, but rather to evoke a multimodal orienting reflex that includes head turning (26), postural adjustment, and, in some cases, eye movement (27). This reflex is mediated by the superior colliculus (28), a subcortical structure that integrates spatial auditory and visual inputs and directs attention and movement toward salient stimuli (29), even in complete darkness (30). The orienting response to sound occurs early in development and does not rely on cortical-level processing or explicit behavioural intention (31). The measurement system is specifically designed to capture this multimodal response. The TrueDepth facial tracking system (iPad Pro) captures subtle gaze and head movements, while a millimetre-wave (mmWave) radar device tracks full-body posture and micro-adjustments in real time. Together, these technologies provide a robust, non-invasive index of the brain's responsiveness to lateralised sound.

The extent to which a simultaneous visual stimulus (Block 3: Congruent Audiovisual condition) modulates this auditory-driven orienting response serves as an indirect index of functional visual processing. This approach captures the visual modulation of a non-visual reflex, offering an objective, instruction-free marker of visual function, particularly valuable in CVI, where conventional behavioural measures are often unavailable.

The Standardised Screening Protocol

The test session is brief, lasting between 12 and 15 minutes. It is structured in sequential blocks (see Table 2). While the broader EchoCue[TM] framework is designed to assess responses across a 20—8000 Hz range, this initial study uses a key subset of these frequencies (500 Hz and 2000 Hz) to address the pre-registered research questions.

Safety and Distress Protocol

Participant safety and comfort are paramount. All auditory stimuli will be calibrated and will not exceed a maximum intensity of 70 dB SPL, a level considered safe for brief exposures in paediatric populations (32).

To ensure auditory safety in the event of positional variation, participant distance from the device will be visually monitored throughout testing. Although stimuli are calibrated to 70 dB SPL at the head position, even substantial forward movement is expected to increase intensity only moderately (approximately +6dB per halving of distance), remaining within accepted short-duration paediatric safety limits. Sessions may be paused and seating position adjusted if a child approaches the speaker closely enough to materially alter sound exposure.

Throughout the session, the researcher will continuously monitor the child for any signs of distress (e.g., crying, agitation, prolonged

283   gaze aversion). The protocol includes flexible breaks between

284   blocks. The session will be terminated immediately if the child

285   exhibits clear signs of distress that cannot be readily soothed, or if

286   the parent/guardian requests discontinuation.

Table 2. Structured Sequence of the EchoCue^{TM} Screening Protocol (Blocks for Stage 1 Analysis)

Block	Stimulation Method	Frequencies / Modalities	Parameters	Clinical Goal	Key Markers to Measure
1.	Auditory-only condition	500 Hz and 2000 Hz, pure tones	Directional sounds (L/R), 100 ms duration, 70 dB SPL	Explore cortical auditory pathways, detect startle reflexes and lateralised orientation	Startle reflex, micro eye movements, orientation response towards sound
2.	Visual-only condition	LED flash (40 ms)	Unilateral hemifield L/R Without sound	Assess the primary visual pathway and visual attentional orientation	Pupillary reflex, head or eye movement towards the flash
3.	Congruent audiovisual stimuli	Synchronised 500 Hz and 2000 Hz tones with LED flash (40ms)	In the same hemifield (e.g., L+L), sounds at 100 ms duration, 70 dB SPL	Measure harmonious multisensory integration (aligned and congruent)	Enhanced reflex response, rapid and smooth orientation

288   Note.   L/R: left/Right   Hemifield;   Hz: Hertz;   SPL: Sound   Pressure   Level.

System Architecture and Measurement Parameters

The implementation of this protocol requires a simple, accessible, and non-invasive technological platform capable of fulfilling three essential functions: stimulus generation, eye-tracking, and motion detection.

Stimulus Generation and Primary Device

A central unit, an iPad Pro (model 2021 or later), generates the protocol's audio-visual stimuli. Stimulus presentation and data acquisition are managed by a custom-built application developed in Swift using Apple's ARKit API (version 5 or later)(33).

Eye-Tracking and Saccadic Reaction Time (SRT) Measurement

To ensure reliable data capture, the protocol uses the iPad's integrated infrared TrueDepth camera system. This system operates at a 60 Hz frame rate, providing a temporal resolution of 16.7 ms, which is considered sufficient for reliably measuring saccadic reaction times. Furthermore, it is coupled with the ARKit API, which offers the significant advantage of simultaneous head and face tracking, allowing for the automatic correction of gaze for head movements.

Gaze data is used to calculate Saccadic Reaction Time (SRT), defined as the time elapsed between the onset of a visual target and the onset of the saccade towards it (34).

- **Saccade Definition**: A saccade is detected using a velocity-based algorithm. An eye movement is classified as a saccade when its angular velocity exceeds a threshold of 30° per second and is considered complete when the velocity drops below this threshold. This distinguishes saccades from slower movements like ocular drift.

- **Onset Definition**: The onset of the saccade is the first video frame at which the eye's angular velocity exceeds the 30°/s threshold.

- **Valid Timeframe**: For an SRT calculation, a saccade's onset must occur between 150 ms and 700 ms after the visual target appears.

Motion Detection and Startle Amplitude Measurement

A single, external mmWave radar sensor (e.g., Acconeer XM122) is used to quantify fine motor responses from a distance of at least 60 cm (35). Given the novelty of this technique in a paediatric clinical context, this measure is considered exploratory.

The radar signal is used to measure the Startle Amplitude, which is operationally defined as a rapid, involuntary flinch in response to an auditory stimulus.

- **Measurement**: The radar provides a continuous measure of motion energy. The Startle Amplitude for a given trial is the

peak motion energy recorded within a pre-specified time window.

- **Timeframe**: The analysis window for detecting a startle is from 40 ms to 200 ms following stimulus onset.

- **Units and Normalisation**: The value is in arbitrary units (a.u.) of motion energy. It is normalised by dividing it by the standard deviation (SD) of the motion energy recorded during a 100 ms silent baseline period immediately preceding the stimulus.

- **Threshold**: A response is classified as a valid startle reflex only if its peak amplitude exceeds three SD above the mean of the pre-stimulus baseline energy. This startle-detection criterion is conceptually distinct from the trial-level exclusion rule, which uses the same 3 SD threshold solely to identify excessive pre-stimulus head movement.

Stimulus Presentation Protocol

The pre-registered research questions are addressed using data from Blocks 1, 2, and 3 (see Table 2). The details for these blocks are as follows:

- **General Parameters**: Auditory stimuli will be presented via the iPad's internal stereo speakers, calibrated to deliver a peak intensity of 70 dB SPL at the participant's head position. Visual stimuli are a white LED flash (40 ms duration) presented on

the left or right side of the screen. Each trial begins with a 500

ms central fixation point, and the inter-trial interval will vary

randomly between 1500 ms and 2500 ms.

▪ **Block 1 (Auditory-only)**: Consists of 20 trials (10 left, 10

right). For the trial relevant to the core research questions, a

subset of the available frequencies (500 Hz and 2000 Hz) will

be presented in a randomised order.

▪ **Block 2 (Visual-only)**: Consists of 20 trials (10 left, 10 right),

presented in a randomised order.

▪ **Block 3 (Congruent Audiovisual)**: Consists of 20 trials (10

left, 10 right). The auditory frequency (500 Hz or 2000 Hz) and

location will be randomised. The visual flash will always be

presented at the same location as the sound.

Data Quantification and Analysis

In line with the system's philosophy, the EchoCue^{TM} software does

not provide an interpretation or diagnosis. Its role is to act as a

quantification engine that applies signal processing algorithms to

transform raw data from the various sensors into the objective

metrics required for analysis, as detailed in the System

Architecture and Measurement Parameters section. These

quantified data are then compiled into a clinical report, allowing

the practitioner to base their evaluation on objective and

reproducible evidence (36).

The pre-registered research questions outlined in Table 1 are directly linked with specific blocks of the EchoCue^{TM} protocol. Data from Blocks 2 and 3 will be used to estimate group differences in Saccadic Reaction Time (SRT), while data from Blocks 1 and 3 will be used to estimate group × condition effects on Startle Amplitude. Data from Blocks 4 and 5 are collected for documentation and potential future exploratory validation but are not part of the current Stage 1 analyses.

Data-Based Exclusion Criteria

Pre-specified criteria will be applied to ensure data quality.

- **Trial-Level Exclusion:** Individual trials will be excluded from analysis if: (a) eye-tracking data quality is poor, operationally defined as loss of pupil detection for more than 50% of the trial duration, or(b) excessive head motion is detected, operationally defined as motion energy (from the mmWave radar) exceeding 3 SD from that trial's silent baseline period. This exclusion criterion applies specifically to pre-stimulus movement. Post-stimulus motion energy peaks exceeding the same 3 SD threshold within the reflex window (40-200 ms after auditory onset) will be treated as valid startle responses rather than artefactual movement. For SRT estimation, trials will also be excluded if gaze data are missing within the critical window for detecting the first saccade (150-700 ms after target onset),

rendering saccade onset unreliable. In practice, this means that trials in which gaze is lost between target onset and completion of the first saccade will be considered invalid for SRT estimation, even if total gaze loss remains below 50%. These thresholds are chosen to ensure that any analysed trial contains sufficient and valid data to reliably compute the outcome metrics.

- **Participant-Level Exclusion:** A participant's entire dataset will be excluded if fewer than 60% of their total trials are deemed valid after trial-level cleaning. This threshold is set to ensure that each participant provides enough data for their individual performance to be a stable and meaningful estimate, while balancing the challenges of data collection in this population. A participant's data will also be excluded in the event of a significant equipment malfunction (e.g., radar failure) that prevents the recording of a primary outcome for a majority of the protocol.

These thresholds are consistent with common standards in developmental eye-tracking research, where a minimum of 50% usable data per trial is often required to ensure the interpretability of gaze-based metrics (37). The 60% threshold at the participant level reflects a pragmatic trade-off between data quality and participant retention and is in line with prior

studies involving young or clinical populations, where strict

thresholds would otherwise lead to substantial data loss (38).

Model Estimation and Reporting

Models will be estimated using restricted maximum likelihood

(REML), with Satterthwaite approximations for degrees of

freedom where required for interval estimation. Regression

coefficients (ß), standard errors, and 95% CIs will be reported for

all fixed effects. For interpretability, effects will be expressed both

in raw units (e.g., milliseconds for SRT) and as estimated marginal

means (EMMs), with 95% CIs.

Residuals will be checked for normality, homoscedasticity, and

influence. Outliers beyond ± 3 SD will be examined but not

automatically excluded; any exclusion of outlier values will rely on

a predefined, transparent rule and will be reported separately from

the trial-level and participant-level exclusion criteria defined

above.

Trial-level data loss due to tracking errors will be handled at the

trial level, while participant-level missingness will be documented.

Completion rate and usable-trial proportions will be summarised to

assess data quality.

Analyses focus on estimation rather than null-hypothesis testing,

with interpretation based on the direction and magnitude of effects

and their 95% CIs rather than binary significance thresholds.

Screening Accuracy Analysis

To address the third research question, whether objective metrics (SRT, Startle Amplitude) can differentiate CVI from non-CVI participants, an estimation-based approach will be used:

1. Feature selection: A principled method (e.g., LASSO regression) may be used to identify a parsimonious set of predictors (e.g., SRT, Startle Amplitude) and handle potential collinearity.

2. ROC Analysis: The performance of the final model will be assessed using Receiver Operating Characteristic (ROC) analysis. The primary outcome will be the estimated Area Under the Curve (AUC) and its 95% CIs, derived via bootstrapping with 1,000 resamples.

Limitations and Future Directions

This study has several limitations inherent to its exploratory proof-of-concept design. First, the sample size is small and focused on a narrow age range, meaning that the results will mainly serve to generate effect size estimates for future larger-scale trials. Effect size estimates derived from small samples can be unreliable and are expected to have wide CIs; therefore, results will be interpreted with appropriate caution.

Second, the protocol relies on consumer-grade hardware, the built-in infrared camera of an iPad, for eye tracking. While this approach

greatly improves accessibility, its precision for capturing subtle

saccades and pupillary metrics at a typical clinical distance (e.g.,

60-100 cm) in a challenging paediatric population is not yet well

established in the scientific literature. Therefore, a key secondary

objective of this study is to assess the feasibility of this technology.

The results will provide important preliminary data on the viability

of using this accessible hardware for clinical research. Future work

should aim to validate these measures by comparing them to

research-grade eye tracking systems.

Similarly, the use of mmWave radar to quantify startle reflex

amplitude is an innovative but emerging technique for which

specific validation in paediatric neurology is lacking. Therefore, a

key secondary aim of this study is to provide crucial preliminary

data on the feasibility and sensitivity of this method for capturing

non-contact neurological reflexes in children.

Participants will be primarily recruited from specialist clinical

services, which may introduce selection bias toward more severe

or complex CVI cases. As a result, the sample may not fully

represent the broader heterogeneity of children with CVI

encountered in community or educational settings. Future studies

should aim to expand recruitment to more diverse populations to

improve generalisability.

The current study relies on prior CVI diagnoses provided by referring clinicians. While this reflects real-world diagnostic pathways, it introduces the risk of misclassification due to variability in diagnostic standards. This limitation is acknowledged, and future studies will include standardised diagnostic assessments and independent verification procedures to improve diagnostic reliability.

Finally, while the protocol controls for developmental level (Vineland-3), other key factors such as cognitive level (IQ) and visual acuity differences remain uncontrolled. These variables could potentially influence the results, and their absence represents a limitation of this feasibility study. Future research should aim to measure and characterise these variables more thoroughly.

Supplementary Materials A. Optional EEG and Blocks 4-5

Note: Details regarding the optional EEG configuration and Blocks 4-5, which are not part of the present Stage 1 analyses, are provided here for completeness.

Neural Activity Monitoring (Optional)

A flexible, front-mounted EEG headband may be integrated to record electrophysiological signals. For this proof-of-concept study, EEG use is optional to ensure that a child's refusal to wear the headband does not become a barrier to participation. Any

collected EEG data will be analysed descriptively to inform future

studies.

Experimental Blocks 4-5

These blocks include higher-level audiovisual integration tasks

intended for later validation phases. They are not part of the

current feasibility or estimation analyses but are documented here

for transparency.

Block 4 (Incongruent Audiovisual Stimuli):

This block introduces spatial incongruence between auditory and

visual inputs (e.g., sound on the left, flash on the right) to probe the

child's ability to resolve cross-modal conflict, a process often

impaired in neurodevelopmental disorders.

Block 5 (Naturalistic Stimuli)

This block incorporates ecologically valid, semantically rich

stimuli (e.g., familiar voices and animals) to assess social attention

and semantic processing, thereby maintaining engagement while

enabling the detection of higher-order cognitive markers.

Block	Stimulation method	Frequencies / Modalities	Parameters	Clinical goal	Key markers to measure
4.	Incongruent audiovisual stimuli	Sound and flash (1000 Hz to 3000 Hz)	Presented in opposite hemifields (e.g., L + R)	Evaluate sensory processing capacity under	Increased reaction time, hesitation or absence of

				cross-modal conflict	response
5.	Naturalistic stimuli	Familiar voices, animal sounds (250-4000 Hz) synchronised with animated images	Familiar voices, animal sounds, synchronised with an animated image	Probe social attention and semantic recognition	Prolonged fixation, smiling, vocalisation, motor imitation

Pre-Registration

This study protocol has been pre-registered on the Open Science

Framework (OSF). The final, time-stamped registration will be

made public upon acceptance in principle. A private, anonymised

view-only link to the project repository, containing the protocol,

analysis scripts, and study materials, has been provided to the

recommender.

Declarations

Ethics Statement: This research protocol has been developed in

preparation for a future doctoral project. Formal ethics approval

will be sought from the responsible Human Research Ethics

Committee (HREC) at the author's host institution prior to the

commencement of any research activities involving human

participants.

Data and Materials Availability

In accordance with TOP guidelines, all processed anonymised data,

analysis scripts, and digital study materials required to reproduce

the study's findings will be made publicly available on a repository

such as the OSF. The EchoCue™ software script developed for

stimulus delivery in this non-commercial, academic research

project will be included in this public deposition.

Data Availability Statement

The preregistration, materials, and data for this study are available

on the Open Science Framework.

References

1.   Smague E. Literature Review of the Influence of Neurovisual Disorders on the Diagnosis of Autism in Australian Children. Current Developmental Disorders Reports. 2025;12(1):3.

2.   Chokron S, Dutton GN. From vision to cognition: potential contributions of cerebral visual impairment to neurodevelopmental disorders. Journal of Neural Transmission. 2023;130(3):409–24.

3.   Gordon S, Kerr A, Wiggs C, Chiang MF. What Is Cerebral/Cortical Visual Impairment and Why Do We Need a New Definition? Ophthalmology. 2024;131(12):1357–8.

4.   Chang MY, Borchert MS. Methods of visual assessment in children with cortical visual impairment. Current Opinion in Neurology. 2021;34(1):89–96.

5.   Health. Q. Profile of children with cerebral visual impairment and cerebral palsy - An Australian cohort. 2022.

6.   Schoenmakers DH, van Beelen I, Voermans MMC, Perik D, Stellingwerff MD, Wolf NI, et al. Adaptive behavior assessed by Vineland-3 as comprehensive outcome measure in vanishing white matter. Ann Clin Transl Neurol. 2024;11(3):650–61.

7.   Esmail H, Arblaster G, Haslam L. The Assessment of Vision in Children with Severe Learning Difficulties: A Systematic Review. Br Ir Orthopt J. 2024;20(1):94–104.

8.   Gruters KG, Groh JM. Sounds and beyond: multisensory and other non-auditory signals in the inferior colliculus. Front Neural Circuits. 2012;6:96.

9.   Xiong Y, Zhang Y, Yan J. The neurobiology of sound-specific auditory plasticity: A core neural circuit. Neuroscience & Biobehavioral Reviews. 2009;33(8):1178–84.

10.   Moerel M, De Martino F, Uğurbil K, Yacoub E, Formisano E. Processing of frequency and location in human subcortical auditory structures. Scientific Reports. 2015;5(1):17048.

11.   Brandt JP, Winters R. Bone Conduction Evaluation. StatPearls. Treasure Island (FL): StatPearls Publishing
Copyright © 2025, StatPearls Publishing LLC.; 2025.

12.   Felix RA, 2nd, Gourévitch B, Portfors CV. Subcortical pathways: Towards a better understanding of auditory disorders. Hear Res. 2018;362:48–60.

13.   Kraus N, Nicol T. Brainstem origins for cortical 'what' and 'where' pathways in the auditory system. Trends Neurosci. 2005;28(4):176–81.

14.   Lauer AM, Behrens D, Klump G. Acoustic startle modification as a tool for evaluating auditory function of the mouse: Progress, pitfalls, and potential. Neurosci Biobehav Rev. 2017;77:194–208.

15.   Borg E. On the neuronal organization of the acoustic middle ear reflex. A physiological and anatomical study. Brain Res. 1973;49(1):101–23.

16.   Porges SW. Polyvagal Theory: Current Status, Clinical Applications, and Future Directions. Clin Neuropsychiatry. 2025;22(3):169–84.

17.   Yeomans JS, Frankland PW. The acoustic startle reflex: neurons and connections. Brain Res Brain Res Rev. 1995;21(3):301–14.

18.   Chaieb L, Wilpert EC, Reber TP, Fell J. Auditory beat stimulation and its effects on cognition and mood States. Front Psychiatry. 2015;6:70.

19.   Nakajima Y, Tanaka N, Mima T, Izumi SI. Stress Recovery Effects of High- and Low-Frequency Amplified Music on Heart Rate Variability. Behav Neurol. 2016;2016:5965894.

20.   Park Y, Guzick AG, Schneider SC, Fuselier M, Wood JJ, Kerns CM, et al. Dental Anxiety in Children With Autism Spectrum Disorder: Understanding Frequency and Associated Variables. Front Psychiatry. 2022;13:838557.

21.   Zatorre RJ, Belin P, Penhune VB. Structure and function of auditory cortex: music and speech. Trends in Cognitive Sciences. 2002;6(1):37–46.

22.   Näätänen R, Kujala T, Escera C, Baldeweg T, Kreegipuu K, Carlson S, et al. The mismatch negativity (MMN) – A unique window to disturbed central auditory processing in ageing and different clinical conditions. Clinical Neurophysiology. 2012;123(3):424–58.

23.   Stein BE, Stanford TR. Multisensory integration: current issues from the perspective of the single neuron. Nature Reviews Neuroscience. 2008;9(4):255–66.

24.   Silva PR, Farias T, Cascio F, Dos Santos L, Peixoto V, Crespo E, et al. Neuroplasticity in visual impairments. Neurol Int. 2018;10(4):7326.

25.   Rule SJ, Nickolaychuk BR. Lateralization of high frequency sounds as a function of interaural amplitude disparity. Can J Exp Psychol. 1995;49(3):368–75.

26.   Gerdfeldter B, Greičiūtė M, Wiens S. Examining experienced lateralization of sounds over headphones with electroencephalography. Neuropsychologia. 2025;207:109064.

27.   Reznik D, Henkin Y, Schadel N, Mukamel R. Lateralized enhancement of auditory cortex activity and increased sensitivity to self-generated sounds. Nat Commun. 2014;5:4059.

28.   Middlebrooks JC, Knudsen EI. Changes in external ear position modify the spatial tuning of auditory units in the cat's superior colliculus. J Neurophysiol. 1987;57(3):672–87.

29.   Novembre G, Pawar VM, Bufacchi RJ, Kilintari M, Srinivasan M, Rothwell JC, et al. Saliency Detection as a Reactive Process: Unexpected Sensory Events Evoke Corticomuscular Coupling. J Neurosci. 2018;38(9):2385–97.

30.   Wallace MT, Stevenson RA. The construct of the multisensory temporal binding window and its dysregulation in developmental disabilities. Neuropsychologia. 2014;64:105–23.

31.   Litovsky R. Development of the auditory system. Handb Clin Neurol. 2015;129:55–72.

32.   Davidson LS. Effects of stimulus level on the speech perception abilities of children using cochlear implants or digital hearing aids. Ear Hear. 2006;27(5):493–507.

33.   Marin-Campos R, Dalmau J, Compte A, Linares D. StimuliApp: Psychophysical tests on mobile devices. Behavior Research Methods. 2021;53(3):1301–7.

34.   Nissen L, Hübner J, Klinker J, Kapsecker M, Leube A, Schneckenburger M, et al. Towards Preventing Gaps in Health Care Systems through Smartphone Use: Analysis of ARKit for Accurate Measurement of Facial Distances in Different Angles. Sensors (Basel). 2023;23(9).

35.   Gao Z, Ali L, Wang C, Liu R, Wang C, Qian C, et al. Real-Time Non-Contact Millimeter Wave Radar-Based Vital Sign Detection. Sensors (Basel). 2022;22(19).

36.   Bossuyt PM, Reitsma JB, Bruns DE, Gatsonis CA, Glasziou PP, Irwig LM, et al. Towards complete and accurate reporting of studies of diagnostic accuracy: the STARD initiative. Standards for Reporting of Diagnostic Accuracy. Clin Chem. 2003;49(1):1–6.

37.   Hessels RS, Hooge ITC. Eye tracking in developmental cognitive neuroscience – The good, the bad and the ugly. Developmental Cognitive Neuroscience. 2019;40:100710.

38.   Franchak JM, Kretch KS, Soska KC, Adolph KE. Head-mounted eye tracking: a new method to describe infant looking. Child Dev. 2011;82(6):1738–50.