The Development and Validation of EchoCue™, a novel tool for assessing Cortical Visual Impairment in uncooperative children [Registered Report, stage I]

Author

Edwige Smague, edwige@netadiscovery.org +14252293045,

ORCID ID: 0009-0002-8182-7888

Department

Neuroscience

Institution

N.E.T.A: Neuronal Epileptotranscriptomic, Sydney, Australia

Abstract: This Stage 1 Registered Report presents EchoCue™, an innovative tool designed to detect cortical visual impairments in uncooperative children aged 0 to 18, without requiring verbal instruction or active participation. The protocol delivers extreme-frequency auditory stimuli (very low and high), paired with directional visual flashes. Reflexive responses, such as eye blinks, gaze shifts, pupil dilation, and startle reactions, are recorded using an eye-tracking system and a flexible frontal EEG device. The assessment is entirely passive, standardised, and suitable for children who are otherwise difficult to evaluate using conventional methods.

Keywords: Cortical visual impairment, screening tool, reflex

response, autism differential diagnosis, uncooperative children.

Introduction

Cortical Visual Impairment (CVI) is a major cause of childhood

disability in Australia and high-income countries. There are many

tools available for screening CVI, but none of them are suitable for

uncooperative children (1). The difficulty is compounded by the

fact that the origin of these disorders is not ocular but neurological,

involving damage to the visual pathways or the cortex, which

requires specific evaluation methods (2). Current tools for CVI rely

on the child's ability to follow instructions or perform tasks,

excluding many of the most affected children from standardised

assessment (3).

This methodological impasse leads to a high risk of underdiagnosis

and misdiagnosis. Indeed, there is significant symptomatic overlap

between CVI and autism spectrum disorders (ASD), including

peculiarities in eye contact, stereotypical behaviours, and atypical

responses to sensory stimuli. This clinical overlap delays or even

prevents the correct identification of visual impairments of cortical

origin, thus depriving these children of appropriate and early care

(4). There is therefore an urgent clinical and scientific need for an

evaluation methodology capable of passively testing children's

perceptual abilities (5). Such an approach must be able to dispense

with the active participation of the child in order to provide

objective markers of the integrity of their neurosensory system.

It is in this context that EchoCue[TM] was conceived, a new

methodological approach based on the integration of multimodal

data such as visual (light, images), auditory (sounds, hertz

frequencies, voices), vestibular (balance and movement), and

proprioceptive (body position) for the screening of CVI. The

protocol is based on the presentation of calibrated sensory stimuli

(auditory and visual) and the synchronised capture of reflex

responses (eye movements, blinks, startles) using an integrated eye

tracking system and a motion sensor. While the primary goal of

this protocol is to establish the method's initial validity, a

recognised limitation is that future work on larger cohorts will be

necessary to establish normative data.

Method

EchoCue™ is an original screening framework designed by

Edwige Smague in 2025 as part of a doctoral research initiative.

The device is not yet commercialised and is currently under

academic development for validation in Australia. The

methodology presented in this article is based on the EchoCue™

conceptual framework, which enables the passive screening of

neurovisual functions using strategically calibrated multimodal

sensory stimuli. This section outlines the theoretical foundations of

the approach, the standardised experimental protocol, the

technological setup required, and the data quantification process.

Table 1. Hypotheses, Variables, Planned Analyses, and Success Criteria for Stage 1 Registered

Report

Key hypothesis	Hypothesis	Key variables	Planned statistical analysis	Success criterion
H1a Confirmatory	The CVI group will exhibit significantly longer saccadic reaction times (SRT) compared to the ASD and Typically Developing (TD) groups.	**Independent Variable (IV):** Group (CVI, ASD, TD) **Dependent Variable (DV):** SRT (in ms)	A linear mixed-effects model (LMM) will be fitted to analyse SRT data, with Group as a fixed effect and Participant ID as a random effect.	A statistically significant main effect of Group on SRT ($p < .025$).
H1b Exploratory	CVI and ASD groups will exhibit different patterns of startle reflex modulation by audiovisual stimuli, compared to the TD group.	**IV**: Group, Stimulus Condition **DV**: Startle Amplitude	A linear mixed-effects model (LMM) will be used to analyse startle amplitude, testing for a Group × Condition interaction.	A statistically significant Group × Condition interaction effect ($p < .025$).
H2a Confirmatory screening accuracy test	The objective metrics measured will be able to reliably classify participants as CVI vs. non-CVI.	**Predictors**: SRT, Startle Amplitude. **Outcome**: Group Status (CVI / non-CVI)	Receiver Operating Characteristic (ROC) curve analysis will be conducted.	An Area Under the Curve (AUC) value greater than 0.75.

Note. AUC: Area Under the Curve; DV: Dependent Variable; H: Hypothesis; IV: Independent Variable;

LMM: Linear Mixed-Effects Model; p: p-value; ROC: Receiver Operating Characteristic; SRT:

Saccadic Reaction Time; TD: Typically Developing.

Sample Size Justification: The planned sample size (n=15/group) is based on recruitment feasibility for

this initial proof-of-concept study. Smallest Effect Size of Interest (SESOI): The primary confirmatory

test (H1a) is conceptualised to detect a large effect size (Cohen's $d \geq 0.8$), as a clinically useful screening

tool must be sensitive to clear, robust differences in neurosensory function. Data from this study will

provide effect size estimates to power future, larger-scale validation trials designed to meet stricter

statistical thresholds (e.g., 90% power).

Participants

A total of 45 participants will be recruited, divided into three age- and sex-matched groups (n=15 per group): (1) children diagnosed with CVI; (2) a clinical control group with a diagnosis of ASD and no known visual impairment; and (3) a typically developing control group.

Power analysis and sample size justification

Given the novelty of the EchoCue™ method, no prior data is available to conduct a precise a priori power analysis. Therefore, the sample size for this initial proof-of-concept study is justified based on two principles: feasibility and the detection of a clinically meaningful effect size as informed by the existing literature.

To align with the high standards of top-tier journals such as Cortex, our ultimate goal is to power studies to detect a large effect size (Cohen's $d \geq 0.8$). This threshold was chosen because a clinically useful screening tool, such as the one proposed, must be sensitive to clear and robust differences in neurosensory function rather than subtle statistical variations. An effect of this magnitude would represent a readily observable impairment, providing a strong and unambiguous signal for clinicians.

102 A formal power calculation based on these parameters indicates

103 that a sample of 33 participants per group would be required to

104 achieve 90% power at a strict alpha level of .025. The currently

105 planned target of 15 participants per group is based on recruitment

106 feasibility for this initial study. I acknowledge this sample provides

107 approximately 80% power at a standard alpha of .05. The primary

108 goal of this first stage is therefore to establish the viability of the

109 protocol, refine the methodology, and generate robust effect size

110 estimates from the data. These estimates will be used to power

111 future, larger-scale validation trials designed to meet the stricter

112 90% power threshold required for confirmatory research.

113 **Recruitment and inclusion/exclusion criteria**

114 Participants will be recruited from key paediatric institutions and

115 support networks in Brisbane, Australia, including the Queensland

116 Children's Hospital (QCH), Mater Children's Private Brisbane,

117 National Disability Insurance Scheme (NDIS), Vision Australia,

118 Autism Cooperative Research Centre (Autism CRC), CVI

119 Australia, Statewide Vision Impairment Services (via the

120 Department of Education Queensland), Queensland University of

121 Technology (QUT), and The University of Queensland (UQ).

122 Inclusion criteria for the CVI group will be: (1) a formal clinical

123 diagnosis of CVI from a qualified professional (e.g., paediatric

124	neurologist or ophthalmologist); (2) evidence that the visual
125	impairment is of neurological origin and not primarily explained
126	by a significant ocular pathology; and (3) an age between 0 and 18
127	years.
128	Exclusion criteria for all groups will be: (1) the presence of a
129	severe, uncorrected ocular pathology that could primarily account
130	for visual deficits; (2) a diagnosed severe or profound hearing
131	impairment that would prevent perception of the auditory stimuli;
132	or (3) an unstable medical condition precluding participation in a
133	15-minute session. Written informed consent will be obtained from
134	the parents or legal guardians of all participants.
135	
136	**Theoretical basis: the link between audition and vision in CVI**
137	**screening**
138	The EchoCue^{TM} protocol is based on a passive assessment method
139	that tests the child's neurosensory functions without the need for
140	the child to understand instructions or actively participate(7). The
141	approach involves exposing the child to a series of finely calibrated
142	auditory and visual stimuli while synchronously recording their
143	involuntary reflex responses (8)(such as blinks, micro-saccades, or
144	startles) and visual orientation responses. A key element of this
145	method is using a wide range of sound frequencies, from 20 to
146	8000 Hz. Each frequency range allows distinct neurological

neurologist or ophthalmologist); (2) evidence that the visual
impairment is of neurological origin and not primarily explained
by a significant ocular pathology; and (3) an age between 0 and 18
years.

Exclusion criteria for all groups will be: (1) the presence of a
severe, uncorrected ocular pathology that could primarily account
for visual deficits; (2) a diagnosed severe or profound hearing
impairment that would prevent perception of the auditory stimuli;
or (3) an unstable medical condition precluding participation in a
15-minute session. Written informed consent will be obtained from
the parents or legal guardians of all participants.

Theoretical basis: the link between audition and vision in CVI
screening

The EchoCue[TM] protocol is based on a passive assessment method
that tests the child's neurosensory functions without the need for
the child to understand instructions or actively participate(7). The
approach involves exposing the child to a series of finely calibrated
auditory and visual stimuli while synchronously recording their
involuntary reflex responses (8)(such as blinks, micro-saccades, or
startles) and visual orientation responses. A key element of this
method is using a wide range of sound frequencies, from 20 to
8000 Hz. Each frequency range allows distinct neurological

147   circuits, both subcortical and cortical, to be explored in order to

148   better understand how the brain perceives, integrates, and responds

149   to multisensory stimuli,(9).

150   Low-frequency sounds (20—250 Hz) act as a subcortical marker.

151   This range is particularly important for uncooperative children

152   (10). These frequencies are perceived by the human body as both

153   vibrotactile stimulation and sound (11). They are transmitted partly

154   by bone conduction, directly stimulating the cochlea and vestibular

155   system, thus bypassing any disorders of the outer or middle ear

156   (12). More important still, they are processed primarily by the

157   archaic neural circuits of the brainstem due to their simple acoustic

158   configuration and their strong ability to arouse sensory alert

159   systems (13) (14). They are particularly effective at triggering

160   involuntary reflexes managed by the brainstem, such as the

161   stapedius reflex (an involuntary auditory reflex that protects the

162   inner ear from loud sounds) and the startle reflex (15, 16). The use

163   of these frequencies thus makes it possible to measure a basic

164   neurological response, a sign of the integrity of the subcortical

165   pathways, even before the information reaches the higher cortical

166   areas (17).

167   Medium to high frequencies (250—8000 Hz) are used for cortical

168   analysis. This range is essential for assessing the classical auditory

169   pathway and higher cognitive functions (18) (19). It is necessary

for multisensory integration tasks, where the brain has to compare and merge auditory information with visual information (20). Difficulty processing these sounds or integrating them correctly with a visual stimulus may reveal dysfunction in associative cortical areas (e.g., the temporoparietal junction), which are often involved in CVI (21).

The use of an audio method for screening CVI is based on the brain's multisensory integration function (8) (22). Since CVI is a brain disorder that affects visual processing and not an eye condition, we can use a powerful auditory stimulus to establish a reliable baseline response. The way a simultaneous visual stimulus alters, or fails to alter, this baseline response can be used to objectively quantify the brain's capacity to process visual information (23). This is an indirect method of testing brain visual function without ever asking the child, "Did you see that?"

The standardised screening protocol

The test session is brief, lasting between 12 and 15 minutes. It is structured in five sequential blocks, each targeting a distinct aspect of sensory or neurological processing (24). The specific pure tones used in the protocol (e.g., 125 Hz, 500 Hz, 2000 Hz) are aligned with standard audiometric test frequencies, ensuring compatibility with established clinical practices (25) (26).

Table 2. Structured sequence of the EchoCue^{TM} multisensory screening protocol

Block	Stimulation method	Frequencies / Modalities	Parameters	Clinical goal	Key markers to measure
1.	Auditory-only condition	125 Hz (infrabasse), 500 Hz (grave), 2000 Hz (médium), 8000 Hz	Directional sounds (L/R)	Explore subcortical (low-frequency) and cortical (high-frequency) auditory pathways, detect startle reflexes and lateralised orientation	Startle reflex, micro eye movements, orientation response towards sound
2.	Visual-only condition	LED flash (40 ms)	Unilateral hemifield L/R Without sound	Assess the primary visual pathway and visual attentional orientation	Pupillary reflex, head or eye movement towards the flash
3.	Congruent audiovisual stimuli (audio and visual cues aligned in spatial direction)	Synchronized sound and flash 500 Hz and 2000 Hz	In the same hemifield (e.g., L + L)	Measure harmonious multisensory integration (aligned and congruent)	Enhanced reflex response, rapid and smooth orientation
4.	Incongruent audiovisual stimuli (audio and visual cues presented in conflicting directions)	Sound and flash 1000 Hz to 3000 Hz	Presented in opposite hemifields (e.g., L + R), Sound (L) vs Flash (R) and vice versa	Evaluate sensory processing capacity under cross-modal conflict	Increased reaction time, hesitation or absence of response
5.	naturalistic stimulu	Familiar voices, animal sounds 250-4000 Hz synchronized with animated images	Familiar voices, animal sounds, synchronised with an animated image	Probe social attention and semantic recognition. Simulate ecological situations to capture spontaneous attention	Prolonged fixation, smiling, vocalisation, motor imitation

193   **Note.** L/R = left/right hemifield; Hz = Hertz. Each block presents a specific sensory configuration to

194   evaluate reflexive and orienting responses across auditory, visual, and multisensory conditions. Block 5

195   introduces ecologically valid stimuli to assess spontaneous social and semantic attention.

The EchoCue™ protocol is structured as a progressive sequence of assessments, beginning with unimodal conditions to establish baseline responses and progressing towards increasingly complex multisensory integrations. Blocks 1 and 2 independently assess the integrity of subcortical and cortical auditory and visual pathways, establishing reference markers for reflexive and orienting responses (17). The use of frequencies such as 125 Hz, 500 Hz, and 2000 Hz aligns with standard clinical audiometric testing (27) (28). Block 3 evaluates the capacity for harmonious multisensory integration through spatially congruent audiovisual stimuli, a hallmark of intact sensory convergence (20). In contrast, Block 4 introduces spatial incongruence between auditory and visual inputs to probe the child's ability to resolve cross-modal conflict, a process often impaired in neurodevelopmental disorders (29). Block 1, in particular, leverages low-frequency tones that can stimulate vestibular and cochlear structures via bone conduction and trigger reflexes such as the auditory startle response (30) (31). Finally, Block 5 incorporates ecologically valid, semantically rich stimuli (e.g., familiar voices and animals) to assess social attention and semantic processing, thereby maintaining engagement while enabling the detection of higher-order cognitive markers. This design ensures the protocol's clinical relevance, even in children without verbal or behavioural responsiveness.

System architecture and data acquisition

The implementation of this protocol requires a simple, accessible, and non-invasive technological platform capable of fulfilling three essential functions:

- **Stimulus Generation:** A central unit, such as an iPad Pro-type digital tablet, generates the protocol's audio-visual stimuli with precise temporal synchronisation (32).

- **Eye-Tracking:** The tablet's integrated infrared camera captures gaze dynamics (saccades, blinks, pupillary reflex) through software APIs (e.g., ARKit). This method removes the necessity for costly external hardware (33).

- **Motion detection:** A single, external wireless sensor is necessary to quantify the fine motor responses that the tablet cannot measure. A millimetre-wave (mmWave) radar is used to accurately detect micro-movements, startle reflexes, and variations in respiratory rate from a distance of at least 60 cm (34).

- **Neural activity monitoring (optional):** A flexible, front-mounted EEG headband may be integrated to record low-density electrophysiological signals. This setup enables the detection of auditory and visual evoked potentials (e.g., MMN, P300), providing an additional objective layer of cortical response data (35, 36) (37).

The most critical function of the system is to ensure the perfect

synchronisation of all data streams (stimuli, ocular data, motion

data) on a single timeline (38).

Data quantification and analysis

In line with the system's philosophy, the EchoCue^{TM} software does

not provide an interpretation or diagnosis. Its role is to act as a

quantification engine that applies signal processing algorithms to

transform raw data into objective metrics. These metrics are then

presented factually to the clinician (39).

The main calculated metrics are:

▪ The average saccadic reaction time (in ms) (40).

▪ The success rate of visual target tracking (41) (42).

▪ The amplitude of the startle reflex (as measured by radar) (43).

▪ The average respiratory rate and its variations during the blocks

(44).

▪ The percentage of sensory dominance (visual vs. auditory) during

conflicting trials (45).

To ensure technical rigour, metrics are defined operationally. For

instance, the startle reflex is identified as a sharp increase in

motion energy exceeding a predefined threshold within 200 ms of

the stimulus (43), distinguishing it from slower, ambient

movements. Sensory dominance is calculated as the percentage of

trials in the conflicting block where the initial saccade was directed

towards the auditory versus the visual stimulus (45). These

quantified data are compiled into a clinical report, allowing the

practitioner to base their evaluation on objective and reproducible

evidence (40).

Data-based exclusion criteria

Pre-specified criteria will be applied to ensure data quality.

▪ Trial-level exclusion: Individual trials will be excluded from

analysis if eye-tracking data quality is poor (e.g., the tracker

loses the pupil for more than 50% of the trial duration).

▪ Participant-level participation: A participant's entire dataset

will be excluded if fewer than 60% of their trials are deemed

valid after trial-level cleaning, or in the event of a significant

equipment malfunction.

Statistical plan

To test for group differences in performance, a mixed-effects

model will be used with Group (CVI, ASD, Control) as a between-

subjects factor and Stimulus Location (Left, Right) as a within-

subjects factor. The dependent variables will be Saccadic Reaction

Time (SRT), pupil response magnitude, and startle reflex

amplitude.

Diagnostic accuracy

Receiver Operating Characteristic (ROC) analysis will be used to

determine the ability of our measured variables to classify

participants correctly into the CVI vs. non-CVI groups. The Area

Under the Curve (AUC) will be the primary metric.

Criteria for success

This validation study will be considered successful if the following

criteria are met: (1) The group difference in the primary dependent

variable (SRT) is statistically significant ($p < .025$), and (2) the

AUC for the ROC analysis is greater than 0.75, indicating good to

excellent diagnostic accuracy.

Declarations

Ethics Statement: This research protocol has been developed in

preparation for a future doctoral project. Formal ethics approval

will be sought from the responsible Human Research Ethics

Committee (HREC) at the author's host institution prior to the

commencement of any research activities involving human

participants.

Data and materials availability

In accordance with TOP guidelines, all processed anonymised data,

analysis scripts, and digital study materials required to reproduce

the study's findings will be made publicly available on a repository

such as the Open Science Framework (OSF). The EchoCue™

software script developed for stimulus delivery in this non-

commercial, academic research project will be included in this

public   deposition.

References

1.   Smague E. Literature Review of the Influence of

Neurovisual Disorders on the Diagnosis of Autism in Australian

Children. Current Developmental Disorders Reports. 2025;12(1):3.

2.   Chokron S, Dutton GN. From vision to cognition: potential

contributions of cerebral visual impairment to neurodevelopmental

disorders. Journal of Neural Transmission. 2023;130(3):409–24.

3.   McConnell EL, Saunders KJ, Little J-A. What assessments

are currently used to investigate and diagnose cerebral visual

impairment (CVI) in children? A systematic review. Ophthalmic &

physiological optics : the journal of the British College of

Ophthalmic Opticians (Optometrists). 2021;41(2):224–44.

4.   Gordon S, Kerr A, Wiggs C, Chiang MF. What Is

Cerebral/Cortical Visual Impairment and Why Do We Need a New

Definition? Ophthalmology. 2024;131(12):1357–8.

5.   Chang MY, Borchert MS. Methods of visual assessment in

children with cortical visual impairment. Current Opinion in

Neurology. 2021;34(1):89–96.

6.   Piasta SB, Justice LM. Encyclopedia of Research Design.

7.   Esmail H, Arblaster G, Haslam L. The Assessment of

Vision in Children with Severe Learning Difficulties: A Systematic

Review. Br Ir Orthopt J. 2024;20(1):94–104.

8.   Gruters KG, Groh JM. Sounds and beyond: multisensory

and other non-auditory signals in the inferior colliculus. Front

Neural Circuits. 2012;6:96.

9.   Xiong Y, Zhang Y, Yan J. The neurobiology of sound-

specific auditory plasticity: A core neural circuit. Neuroscience &

Biobehavioral Reviews. 2009;33(8):1178–84.

10.   Werner LA. The development of auditory behavior (or

what the anatomists and physiologists have to explain). Ear Hear.

1996;17(5):438–46.

11.   Moerel M, De Martino F, Uğurbil K, Yacoub E, Formisano

E. Processing of frequency and location in human subcortical

auditory structures. Scientific Reports. 2015;5(1):17048.

12.   Brandt JP, Winters R. Bone Conduction Evaluation.

StatPearls. Treasure Island (FL): StatPearls Publishing

Copyright © 2025, StatPearls Publishing LLC.; 2025.

13.   Felix RA, 2nd, Gourévitch B, Portfors CV. Subcortical

pathways: Towards a better understanding of auditory disorders.

Hear Res. 2018;362:48–60.

14.   Kraus N, Nicol T. Brainstem origins for cortical 'what' and 'where' pathways in the auditory system. Trends Neurosci. 2005;28(4):176–81.

15.   Lauer AM, Behrens D, Klump G. Acoustic startle modification as a tool for evaluating auditory function of the mouse: Progress, pitfalls, and potential. Neurosci Biobehav Rev. 2017;77:194–208.

16.   Borg E. On the neuronal organization of the acoustic middle ear reflex. A physiological and anatomical study. Brain Res. 1973;49(1):101–23.

17.   Yeomans JS, Frankland PW. The acoustic startle reflex: neurons and connections. Brain Res Brain Res Rev. 1995;21(3):301–14.

18.   Zatorre RJ, Belin P, Penhune VB. Structure and function of auditory cortex: music and speech. Trends in Cognitive Sciences. 2002;6(1):37–46.

19.   Näätänen R, Kujala T, Escera C, Baldeweg T, Kreegipuu K, Carlson S, et al. The mismatch negativity (MMN) – A unique window to disturbed central auditory processing in ageing and different clinical conditions. Clinical Neurophysiology. 2012;123(3):424–58.

20.   Stein BE, Stanford TR. Multisensory integration: current

issues from the perspective of the single neuron. Nature Reviews

Neuroscience. 2008;9(4):255–66.

21.   Silva PR, Farias T, Cascio F, Dos Santos L, Peixoto V,

Crespo E, et al. Neuroplasticity in visual impairments. Neurol Int.

2018;10(4):7326.

22.   Wallace MT, Stevenson RA. The construct of the

multisensory temporal binding window and its dysregulation in

developmental disabilities. Neuropsychologia. 2014;64:105–23.

23.   van Atteveldt N, Formisano E, Goebel R, Blomert L.

Integration of Letters and Speech Sounds in the Human Brain.

Neuron. 2004;43(2):271–82.

24.   Moore JK, and Linthicum Jr FH. The human auditory

system: A timeline of development. International Journal of

Audiology. 2007;46(9):460–78.

25.   Luo Q, Shen Y, Chen T, Zheng Z, Shi H, Feng Y, et al.

Effects of SoundBite Bone Conduction Hearing Aids on Speech

Recognition and Quality of Life in Patients with Single-Sided

Deafness. Neural Plast. 2020;2020:4106949.

26.   Musiek FE, Baran JA. The auditory system : anatomy,

physiology, and clinical correlates. Second edition. ed. San Diego,

CA: Plural Publishing; 2020.

27.   Guidelines for manual pure-tone threshold audiometry.

Asha. 1978;20(4):297–301.

28.   Lee SY, Seo HW, Jung SM, Lee SH, Chung JH. Assessing

the accuracy and reliability of application-based audiometry for

hearing evaluation. Scientific Reports. 2024;14(1):7359.

29.   Kawakami S, Uono S, Otsuka S, Yoshimura S, Zhao S,

Toichi M. Atypical Multisensory Integration and the Temporal

Binding Window in Autism Spectrum Disorder. Journal of Autism

and Developmental Disorders. 2020;50(11):3944–56.

30.   Jones GP, Lukashkina VA, Russell IJ, Lukashkin AN. The

vestibular system mediates sensation of low-frequency sounds in

mice. J Assoc Res Otolaryngol. 2010;11(4):725–32.

31.   Anne Small S, and Richard Stapells D. Maturation of bone

conduction multiple auditory steady-state responses. International

Journal of Audiology. 2008;47(8):476–88.

32.   Marin-Campos R, Dalmau J, Compte A, Linares D.

StimuliApp: Psychophysical tests on mobile devices. Behavior

Research Methods. 2021;53(3):1301–7.

33.   Nissen L, Hübner J, Klinker J, Kapsecker M, Leube A,

Schneckenburger M, et al. Towards Preventing Gaps in Health

Care Systems through Smartphone Use: Analysis of ARKit for

Accurate Measurement of Facial Distances in Different Angles.

Sensors (Basel). 2023;23(9).

34.   Gao Z, Ali L, Wang C, Liu R, Wang C, Qian C, et al. Real-

Time Non-Contact Millimeter Wave Radar-Based Vital Sign

Detection. Sensors (Basel). 2022;22(19).

35.   Polich J. Updating P300: An integrative theory of P3a and

P3b. Clinical Neurophysiology. 2007;118(10):2128–48.

36.   Näätänen R, Paavilainen P, Rinne T, Alho K. The

mismatch negativity (MMN) in basic research of central auditory

processing: a review. Clin Neurophysiol. 2007;118(12):2544–90.

37.   Lau-Zhu A, Lau MPH, McLoughlin G. Mobile EEG in

research on neurodevelopmental disorders: Opportunities and

challenges.   Developmental   Cognitive   Neuroscience.

2019;36:100635.

38.   Delorme A, Makeig S. EEGLAB: an open source toolbox

for analysis of single-trial EEG dynamics including independent

component analysis. J Neurosci Methods. 2004;134(1):9–21.

39.   Bossuyt PM, Reitsma JB, Bruns DE, Gatsonis CA,

Glasziou PP, Irwig LM, et al. Towards complete and accurate

reporting of studies of diagnostic accuracy: the STARD initiative.

Standards for Reporting of Diagnostic Accuracy. Clin Chem.

2003;49(1):1–6.

40.   Shimizu N. [Neurology of eye movements]. Rinsho

Shinkeigaku. 2000;40(12):1220–3.

41.   Ramat S, Leigh RJ, Zee DS, Optican LM. Ocular

oscillations generated by coupling of brainstem excitatory and

inhibitory saccadic burst neurons. Exp Brain Res. 2005;160(1):89–

106.

42.   Bonnen K, Burge J, Yates J, Pillow J, Cormack LK.

Continuous psychophysics: Target-tracking to measure visual

sensitivity. J Vis. 2015;15(3).

43.   Yeomans JS, Li L, Scott BW, Frankland PW. Tactile,

acoustic and vestibular systems sum to elicit the startle reflex.

Neurosci Biobehav Rev. 2002;26(1):1–11.

44.   Mosleh S, Coder JB, Scully CG, Forsyth K, Al Kalaa MO.

Monitoring Respiratory Motion With Wi-Fi CSI: Characterizing

Performance and the BreatheSmart Algorithm. IEEE Access.

2022;10:131932–51.

45.   Zimmer U, Roberts KC, Harshbarger TB, Woldorff MG.

Multisensory conflict modulates the spread of visual attention

across a multisensory object. NeuroImage. 2010;52(2):606–16.