Assessing Cerebral Visual Impairment through passive responses: Development and

validation of EchoCue™ [Registered Report, stage I]

Author

Edwige Smague[1, 2]

Affiliations:

[1]N.E.T.A: Neuronal Epileptotranscriptomic, Sydney, Australia

[2]Swedish Simulation Centre. Swedish Cherry Hill Campus, Seattle,

USA

Contact:

edwige@netadiscovery.org

ORCID ID: 0009-0002-8182-7888

Abstract: This Stage 1 Registered Report outlines the protocol for

a feasibility and preliminary validation study of EchoCue™, a

novel passive assessment method for cerebral visual impairments

(CVI) in children aged 2 to 6. The protocol is designed to minimise

reliance on verbal instruction or active cooperation, making it

suitable for populations traditionally considered difficult to assess.

The method pairs auditory and visual stimuli with the recording of

reflexive responses, such as gaze shifts, blinks, and startle

reactions, using an integrated infrared eye-tracking system and

optional EEG. The primary aim of this study is to evaluate the

feasibility of data collection procedures and assess the preliminary

reliability of the protocol's core measures, in preparation for future

large-scale validation.

Keywords: Cerebral visual impairment, screening tool, reflex

response, autism differential diagnosis, passive responses

Introduction

Cerebral Visual Impairment (CVI) is a major cause of childhood

disability in Australia and high-income countries. There are

numerous tools available for screening for CVI, but these are

predominantly task-based protocols that require active

participation from the child and are inaccessible to the most

severely affected children (1). The difficulty is compounded by the

fact that the origin of these disorders is not ocular but neurological,

involving damage to the visual pathways or the cortex, which

requires specific evaluation methods (2). Current tools for CVI rely

on the child's ability to follow instructions or perform tasks,

excluding many of the most affected children from standardised

assessment (3).

The assessment of Cerebral Visual Impairment (CVI) is a major

clinical challenge, leading to a high risk of underdiagnosis and

misdiagnosis. This difficulty is compounded by two key factors, a

lack of suitable assessment tools for all children and a significant

symptomatic overlap with other neurodevelopmental disorders

such as autism spectrum disorders (ASD), including peculiarities

in eye contact, stereotypical behaviours, and atypical responses to

sensory stimuli. This clinical overlap delays or even prevents the

correct identification of visual impairments of cerebral origin, thus

depriving these children of appropriate and early care (4). There is,

therefore, an urgent clinical and scientific need for an evaluation

methodology capable of passively testing the foundational sensory

processes that underpin a child's perceptual abilities (5).

Such an approach must be able to dispense with the active

participation of the child in order to provide objective markers of

the integrity of their neurosensory system. It is important to note

that EchoCue[TM] is designed as an objective screening and

assessment tool to provide key evidence within a comprehensive

clinical evaluation. It is in this context that EchoCue[TM] was

conceived, a new methodological approach based on the

integration of multimodal data such as visual (light, images),

auditory (sounds, hertz frequencies, voices), vestibular (balance

and movement), and proprioceptive (body position) for the

screening of CVI. The protocol is based on the presentation of

calibrated sensory stimuli (auditory and visual) and the

synchronised capture of reflex responses (eye movements, blinks,

startles) using an integrated eye tracking system and a motion

sensor. While the primary goal of this protocol is to establish the

method's initial validity, a recognised limitation is that future work

on larger cohorts will be necessary to establish normative data.

Method

EchoCue™ is an original screening framework designed by

Edwige Smague in 2025 as part of a doctoral research initiative.

The device is not yet commercialised and is currently under

academic development for validation in Australia. The

methodology presented in this article is based on the EchoCue™

conceptual framework, which enables the passive screening of

reflexive and bottom-up neurovisual functions using strategically

calibrated multimodal sensory stimuli. This section outlines the

theoretical foundations of the approach, the standardised

experimental protocol, the technological setup required, and the

data   quantification   process.

Table 1. Hypotheses, Variables, Planned Analyses, and Success Criteria for Stage 1 Registered Report

Key hypothesis	Hypothesis	Key variables	Planned statistical analysis	Success criterion
H1a Exploratory	The CVI group will exhibit significantly longer saccadic reaction times (SRT) compared to the ASD and Typically Developing Groups (TD).	IV: Group (CVI, ASD, TD) DV: Saccadic Reaction Time (SRT).	A linear mixed-effects model (LMM) will be fitted to test for a main effect of Group on SRT.	A statistically significant main effect of Group on SRT ($p < .05$).
H1b Exploratory	The audiovisual enhancement of the startle reflex (Congruent Audiovisual vs. Auditory-only conditions) will be significantly smaller in the CVI group compared to the ASD and TD groups.	IV: Group (CVI, ASD, TD), Stimulus Condition (Auditory-only, Congruent Audiovisual) DV: Startle Amplitude.	A linear mixed-effects model (LMM) will test for a Group x Condition interaction. Follow-up planned contrasts will be used to test the specific directional hypothesis.	A statistically significant planned contrast showing that the audiovisual enhancement effect is smaller in the CVI group compared to the other groups ($p < .05$).
H2a Exploratory screening accuracy	The objective metrics measured will be able to	Predictors: SRT, Startle Amplitude.	A multi-step approach will be used: (1) feature selection (e.g.,	The primary outcome will be the estimated Area Under the Curve

	reliably differentiate between participants as CVI vs. non-CVI.	Outcome: Group Status (CVI / non-CVI)	LASSO regression) followed by (2) Receiver Operating Characteristic (ROC) analysis with bootstrapping.	(AUC) and its 95% confidence interval to provide a preliminary estimate of screening accuracy.

Note. AUC: Area Under the Curve; DV: Dependent Variable; H: Hypothesis; IV: Independent Variable; LMM: Linear Mixed-Effects

Model; p: p-value; ROC: Receiver Operating Characteristic; SRT: Saccadic Reaction Time; TD: Typically Developing.

Sample Size Justification: The planned sample size (n=15/group) is based on recruitment feasibility for this initial proof-of-concept

study. Although this study is exploratory, it is conceptualised to detect a large effect size (Cohen's d $\geq$ 0.8), as a clinically useful

screening tool must be sensitive to clear, robust differences in neurosensory function. Data from this study will provide effect size

estimates to power future, larger-scale validation trials designed to meet stricter statistical thresholds (e.g., 90% power). The primary

goal of this initial study is to establish the feasibility of the protocol and generate robust estimates of effect size and confidence

intervals to inform a future larger-scale confirmatory trial.

Participants

A total of 45 participants focusing on a preschool age range, 2 to 6 years, will be recruited to ensure developmental homogeneity for this initial validation study. Participants will be divided into three age and sex-matched groups (n=15 per group): (1) children diagnosed with CVI; (2) a clinical control group with a diagnosis of ASD and no known visual impairment; and (3) a typically developing control group. All participants' adaptive functioning will be assessed and statistically controlled for (see Measures and Statistical Plan sections).

Power analysis and sample size justification

Given the novelty of the EchoCue™ method, no prior data are available to conduct a precise a priori power analysis for the proposed linear mixed models. The planned sample of 15 participants per group is justified primarily by recruitment feasibility for this initial proof-of-concept study.

To estimate the statistical power of this design, G*Power was used to conduct a sensitivity analysis. This analysis was based on a simplified between-group comparison (independent-samples t-test, two-tailed). With the planned sample size (n=15 per group) and a standard alpha of .05, the analysis indicates that the study will have approximately 56% power to detect a large effect size (Cohen's d = 0.8).

While this power level is below the conventional 80% target, it is deemed sufficient for the exploratory goals of this study, which are to establish feasibility and generate robust effect size estimates to power future, larger-scale validation.

Recruitment and inclusion/exclusion criteria

Participants will be recruited from key paediatric institutions and support networks in Brisbane, Australia, including the Queensland Children's Hospital (QCH), Mater Children's Private Brisbane, National Disability Insurance Scheme (NDIS), Vision Australia, Autism Cooperative Research Centre (Autism CRC), CVI Australia, Statewide Vision Impairment Services (via the Department of Education Queensland), Queensland University of Technology (QUT), and The University of Queensland (UQ).

Inclusion criteria for the CVI group will be: (1) a formal clinical diagnosis of CVI from a qualified professional (e.g., paediatric neurologist or ophthalmologist), confirming the visual impairment is of neurological origin; and (2) an age between 2 and 6 years.

Inclusion criteria for the ASD group will be: (1) a formal clinical diagnosis of autism spectrum disorder from a qualified professional (e.g., paediatrician, psychologist), based on established criteria such as the DSM-5; (2) no known or suspected visual impairment of either ocular or cerebral origin; and (3) an age between 2 and 6 years.

Inclusion criteria for the TD group will be: (1) no known developmental, neurological, or psychiatric diagnoses; (2) no parental or clinical concerns regarding vision, hearing, or general development; (3) born full-term with no significant perinatal complications; and (4) an age between 2 and 6 years.

Notably, there will be no exclusion criterion based on cognitive or intellectual level (e.g., IQ score <70). A key objective of this study is to validate a tool for the entire spectrum of children with CVI, including those with co-occurring intellectual disability.

Exclusion criteria for all groups will be: (1) the presence of a severe, uncorrected ocular pathology that could primarily account for visual deficits; (2) a diagnosed severe or profound hearing impairment that would prevent perception of the auditory stimuli; or (3) an unstable medical condition precluding participation in a 15-minute session.

Written informed consent will be obtained from the parents or legal guardians of all participants.

Measures

Developmental and adaptive functioning

To assess the developmental and adaptive functioning level of each participant, parents or legal guardians will complete the Vineland Adaptive Behavior Scales, Third Edition, Vineland-3.

The Vineland-3 is a standardised assessment completed via parent

or carer report that measures communication, daily living skills,

socialisation, and motor skills. The Adaptive Behavior Composite

score will be used as a measure of overall adaptive functioning and

included as a covariate in statistical analyses to control for

developmental differences.

Theoretical rationale using auditory probes to assess

neurovisual function

The EchoCue^{TM} protocol is based on a passive assessment method

that tests the child's neurosensory functions without the need for

the child to understand instructions or actively participate (6). The

approach involves exposing the child to a series of finely calibrated

auditory and visual stimuli while synchronously recording their

involuntary reflex responses (7), such as blinks, micro-saccades, or

startles, and visual orientation responses. A key element of this

method is using a wide range of sound frequencies, from 20 to

8000 Hz. Each frequency range allows distinct neurological

circuits, both subcortical and cortical, to be explored in order to

better understand how the brain perceives, integrates, and responds

to multisensory stimuli (8).

Low-frequency sounds (20—250 Hz) act as a subcortical marker.

This range is particularly important for difficult-to-test children

(9).

183	These frequencies are perceived by the human body as both
184	vibrotactile stimulation and sound (10). They are transmitted partly
185	by bone conduction, directly stimulating the cochlea and vestibular
186	system, thus bypassing any disorders of the outer or middle ear
187	(11). More important still, they are processed primarily by the
188	archaic neural circuits of the brainstem due to their simple acoustic
189	configuration and their strong ability to arouse sensory alert
190	systems (12, 13). They are particularly effective at triggering
191	involuntary reflexes managed by the brainstem, such as the
192	stapedius reflex (an involuntary auditory reflex that protects the
193	inner ear from loud sounds) and the startle reflex (14, 15). The use
194	of these frequencies thus makes it possible to measure a basic
195	neurological response, a sign of the integrity of the subcortical
196	pathways, even before the information reaches the higher cortical
197	areas. This approach is consistent with findings from auditory
198	neuroscience showing that low-frequency auditory stimuli can
199	promote autonomic regulation, including parasympathetic
200	activation, through mechanisms such as vagal nerve modulation
201	(Ahn et al., 2021; Porges, 2011). These effects have been
202	associated with reduced anxiety and increased physiological
203	markers of calm in both clinical and developmental populations
204	(16, 17).

Beyond their function as a subcortical probe, the use of these low-

frequency sounds is also informed by emerging research on their

potential therapeutic effects (17). Studies have suggested that

certain frequencies can have a regulating or calming effect on the

nervous system (18), which may help in managing anxiety in

paediatric populations, including children with ASD during

stressful procedures (19). This dual function, as both an objective

probe and a potentially regulating stimulus, is a key feature of the

EchoCue™ protocol's design to maximise participant comfort.

Medium to high frequencies (250—8000 Hz) are used for cortical

analysis. This range is essential for assessing the classical auditory

pathway and higher cognitive functions (20, 21). It is necessary for

multisensory integration tasks, where the brain has to compare and

merge auditory information with visual information (22).

Difficulty processing these sounds or integrating them correctly

with a visual stimulus may reveal dysfunction in associative

cortical areas (e.g., the temporoparietal junction), which are often

involved in CVI (23).

To establish a reliable baseline of involuntary orienting behaviour,

the EchoCue™ protocol begins with an auditory-only condition

(Block 1) using lateralised, high-intensity sounds (24). This

condition is not intended to elicit voluntary saccadic eye

movements alone, but rather to evoke a multimodal orienting

reflex that includes head turning (25), postural adjustment, and, in

some cases, eye movement (26). This reflex is mediated by the

superior colliculus (27), a subcortical structure that integrates

spatial auditory and visual inputs and directs attention and

movement toward salient stimuli (28), even in complete darkness

(King, 2004; Stein & Meredith, 1993; Wallace & Stein, 1997). The

orienting response to sound occurs early in development and does

not rely on cortical-level processing or explicit behavioural

intention (29). The measurement system is specifically designed to

capture this multimodal response. The TrueDepth facial tracking

system (iPad Pro) captures subtle gaze and head movements, while

a millimetre-wave (mmWave) radar device tracks full-body

posture and micro-adjustments in real time. Together, these

technologies provide a robust, non-invasive index of the brain's

responsiveness to lateralised sound. The degree to which a

simultaneous visual stimulus (Block 3: Congruent Audiovisual

condition) enhances or fails to enhance this auditory-driven

orienting response provides an indirect measure of functional

visual processing, which is particularly valuable in populations

such as CVI where behavioural reports are unreliable or

unavailable.

The standardised screening protocol

The test session is brief, lasting between 12 and 15 minutes. It is

structured in five sequential blocks (see table 2), each targeting a

distinct aspect of sensory or neurological processing (30). The

specific pure tones used in the protocol (e.g., 125 Hz, 500 Hz,

2000 Hz) are aligned with standard audiometric test frequencies,

ensuring compatibility with established clinical practices (31, 32).

While the broader EchoCue^{TM} framework is designed to assess

responses across a 20—8000Hz range, this initial study uses this

key subset of frequencies to test the pre-registered hypothesis.

Safety and distress protocol

Participant safety and comfort are paramount. All auditory stimuli

will be calibrated and will not exceed a maximum intensity of 70

dB SPL, a level considered safe for brief exposures in paediatric

populations (33).

Throughout the session, the researcher will continuously monitor

the child for any signs of distress (e.g., crying, agitation, prolonged

gaze aversion). The protocol includes flexible breaks between

blocks. The session will be terminated immediately if the child

shows clear signs of distress that cannot be easily soothed, or if the

parent/guardian requests to stop.

Table 2. Structured sequence of the EchoCue^{TM} multisensory screening protocol

Block	Stimulation method	Frequencies / Modalities	Parameters	Clinical goal	Key markers to measure
1.	Auditory-only condition	125 Hz, 500 Hz, 2000 Hz, 8000 Hz pure tones	Directional sounds (L/R), 100 ms duration, 70 dB SPL	Explore subcortical (low frequency) and cortical auditory pathways (high frequency), detect startle reflexes and lateralised orientation	Startle reflex, micro eye movements, orientation response towards sound
2.	Visual-only condition	LED flash (40 ms)	Unilateral hemifield L/R Without sound	Assess the primary visual pathway and visual attentional orientation	Pupillary reflex, head or eye movement towards the flash
3.	Congruent audiovisual stimuli	Synchronised 500 Hz & 2000 Hz tones with LED flash (40ms)	In the same hemifield (e.g., L+L), sounds at 100 ms duration, 70 dB SPL	Measure harmonious multisensory integration (aligned and congruent)	Enhanced reflex response, rapid and smooth orientation
4.	Incongruent audiovisual	Sound and flash (1000 Hz to 3000	Presented in opposite hemifields	Evaluate sensory processing capacity under	Increased reaction time, hesitation or absence of

	stimuli	Hz)	(e.g., L + R)	cross-modal conflict	response
5.	Naturalistic stimuli	Familiar voices, animal sounds (250-4000 Hz) synchronised with animated images	Familiar voices, animal sounds, synchronised with an animated image	Probe social attention and semantic recognition.	Prolonged fixation, smiling, vocalisation, motor imitation

Note. L/R = left/right hemifield; Hz = Hertz. Each block presents a specific sensory configuration to

evaluate reflexive and orienting responses across auditory, visual, and multisensory conditions. Block 5

introduces ecologically valid stimuli to assess spontaneous social and semantic attention.

The EchoCue^{TM} protocol is structured as a progressive sequence of

assessments, beginning with unimodal conditions to establish

baseline responses and progressing towards increasingly complex

multisensory integrations. Blocks 1 and 2 independently assess the

integrity of subcortical and cortical auditory and visual pathways,

establishing reference markers for reflexive and orienting

responses (16). The use of frequencies such as 125 Hz, 500 Hz,

and 2000 Hz aligns with standard clinical audiometric testing (34,

35). Block 3 evaluates the capacity for harmonious multisensory

integration through spatially congruent audiovisual stimuli, a

hallmark of intact sensory convergence (22). In contrast, Block 4

introduces spatial incongruence between auditory and visual inputs

to probe the child's ability to resolve cross-modal conflict, a

process often impaired in neurodevelopmental disorders (35).

Block 1, in particular, leverages low-frequency tones that can

stimulate vestibular and cochlear structures via bone conduction

and trigger reflexes such as the auditory startle response (36, 37).

Finally, Block 5 incorporates ecologically valid, semantically rich

stimuli (e.g., familiar voices and animals) to assess social attention

and semantic processing, thereby maintaining engagement while

enabling the detection of higher-order cognitive markers. This

design ensures the protocol's clinical relevance, even in children

without verbal or behavioural responsiveness.

System architecture and measurement parameters

The implementation of this protocol requires a simple, accessible, and non-invasive technological platform capable of fulfilling three essential functions: stimulus generation, eye-tracking, and motion detection.

Stimulus Generation and Primary Device

A central unit, an iPad Pro (model 2021 or later), generates the protocol's audio-visual stimuli. Stimulus presentation and data acquisition are managed by a custom-built application developed in Swift using Apple's ARKit API (version 5 or later)(38).

Eye-Tracking and Saccadic Reaction Time (SRT) Measurement

To ensure reliable data capture, the protocol uses the iPad's integrated infrared TrueDepth camera system. This system operates at a 60 Hz frame rate, providing a temporal resolution of 16.7 ms, which is considered sufficient for reliably measuring saccadic reaction times. Furthermore, it is coupled with the ARKit API, which offers the significant advantage of simultaneous head and face tracking, allowing for the automatic correction of gaze for head movements.

Gaze data is used to calculate Saccadic Reaction Time (SRT), defined as the time elapsed between the onset of a visual target and the onset of the saccade towards it (39).

- Saccade Definition: A saccade is detected using a velocity-based algorithm. An eye movement is classified as a saccade when its angular velocity exceeds a threshold of 30° per second and is considered complete when the velocity drops below this threshold. This distinguishes saccades from slower movements like ocular drift.

- Onset Definition: The onset of the saccade is the first video frame at which the eye's angular velocity exceeds the 30°/s threshold.

- Valid Timeframe: For an SRT calculation, a saccade's onset must occur between 150 ms and 700 ms after the visual target appears.

Motion Detection and Startle Amplitude Measurement

A single, external mmWave radar sensor (e.g., Acconeer XM122) is used to quantify fine motor responses from a distance of at least 60 cm (40). Given the novelty of this technique in a paediatric clinical context, this measure is considered exploratory.

The radar signal is used to measure the Startle Amplitude, which is operationally defined as a rapid, involuntary flinch in response to an auditory stimulus.

- **Measurement:** The radar provides a continuous measure of motion energy. The Startle Amplitude for a given trial is the peak motion energy recorded within a pre-specified time window.
- **Timeframe:** The analysis window for detecting a startle is from 40 ms to 200 ms following stimulus onset.
- **Units and Normalisation:** The value is in arbitrary units (a.u.) of motion energy. It is normalised by dividing it by the standard deviation of the motion energy recorded during a 100 ms silent baseline period immediately preceding the stimulus.
- **Threshold:** A response is classified as a valid startle reflex only if its peak amplitude exceeds three standard deviations above the mean of the pre-stimulus baseline energy.

Neural Activity Monitoring (Optional)

A flexible, front-mounted EEG headband may be integrated to record electrophysiological signals. For this proof-of-concept study, EEG is optional to ensure that a child's refusal to wear the headband does not become a barrier to participation. Any collected EEG data will be analysed descriptively to inform future studies.

The most critical function of the system is to ensure the perfect synchronisation of all data streams (stimuli, ocular data, motion data) on a single timeline.

Stimulus Presentation Protocol

The pre-registered hypotheses are tested using data from Blocks 1,

2, and 3 (see Table 2). The details for these blocks are as follows:

▪ General Parameters: Auditory stimuli will be presented via the

iPad's internal stereo speakers, calibrated to deliver a peak

intensity of 70 dB SPL at the participant's head position. Visual

stimuli are a white LED flash (40 ms duration) presented on

the left or right side of the screen. Each trial begins with a 500

ms central fixation point, and the inter-trial interval will vary

randomly between 1500 ms and 2500 ms.

▪ Block 1 (Auditory-only): Consists of 20 trials (10 left, 10

right). For the hypothesis testing trial, a subset of the available

frequencies (500 Hz and 2000 Hz) will be presented in a

randomised order.

▪ Block 2 (Visual-only): Consists of 20 trials (10 left, 10 right),

presented in a randomised order.

▪ Block 3 (Congruent Audiovisual): Consists of 20 trials (10 left,

10 right). The auditory frequency (500 Hz or 2000 Hz) and

location will be randomised. The visual flash will always be

presented at the same location as the sound.

Data quantification and analysis

In line with the system's philosophy, the EchoCue^{TM} software does

not provide an interpretation or diagnosis. Its role is to act as a

quantification engine that applies signal processing algorithms to

transform raw data from the various sensors into the objective

metrics required for hypothesis testing, as detailed in the System

architecture and measurement parameters section. These quantified

data are then compiled into a clinical report, allowing the

practitioner to base their evaluation on objective and reproducible

evidence (41).

The pre-registered hypotheses detailed in Table 1 are directly

linked to specific blocks of the EchoCue™ protocol.

Hypothesis H1a will be tested using data from Block 2 and Block

3. Hypothesis H1b will be tested by comparing data from Block 1

and Block 3. Data from Block 4 and Block 5 are collected for

future exploratory analyses that are beyond the scope of the current

study.

Data-based exclusion criteria

Pre-specified criteria will be applied to ensure data quality.

- **Trial-level exclusion:** Individual trials will be excluded from
  analysis if: (a) eye-tracking data quality is poor (e.g., the
  tracker loses the pupil for more than 50% of the trial duration),
  or (b) excessive head motion is detected by the mmWave radar
  during stimulus presentation. These thresholds are chosen to
  ensure that any analysed trial contains sufficient and valid data
  to reliably compute the outcome metrics.

- **Participant-level exclusion:** A participant's entire dataset will be excluded if fewer than 60% of their total trials are deemed valid after trial-level cleaning. This threshold is set to ensure that each participant provides enough data for their individual performance to be a stable and meaningful estimate, while balancing the challenges of data collection in this population. A participant's data will also be excluded in the event of a significant equipment malfunction (e.g., radar failure) that prevents the recording of a primary outcome for a majority of the protocol.

  These thresholds are consistent with common standards in developmental eye-tracking research, where a minimum of 50% usable data per trial is often required to ensure the interpretability of gaze-based metrics (42). The 60% threshold at the participant level reflects a pragmatic trade-off between data quality and participant retention and is in line with prior studies involving young or clinical populations, where strict thresholds would otherwise lead to substantial data loss (43).

Statistical plan

All analyses will be conducted in R using the lme4 package for mixed models. The Vineland-3 Adaptive Behaviour Composite score will be included as a covariate in all models to control for

432 developmental differences. The significance level for all
433 hypothesis tests is set at α = .05.

Hypothesis H1a: Group Differences in Saccadic Reaction Time

- Design: Between-subjects factor: Group (CVI, ASD, TD).
- Dependent Variable: Saccadic Reaction Time (SRT).
- Analysis: A Linear Mixed-Effects Model (LMM) will be
fitted to test for a main effect of Group. The model will
include a random intercept for each participant to account
for individual differences.

Hypothesis H1b: Group Differences in Startle Reflex Modulation

- Design: Within-subjects factor: Stimulus Condition
(Auditory-only vs. Congruent Audiovisual), Between-
subjects factor: Group (CVI, ASD, TD).
- Dependent Variable: Startle Amplitude.
- Analysis: An LMM will be fitted to test for a Group ×
Condition interaction, with a random intercept for each
participant.
- Follow-up: If the interaction is significant, planned
contrasts will be conducted to test if the audiovisual
enhancement of the startle reflex is significantly smaller in
the CVI group compared to the other two groups combined.

Hypothesis H2a: Screening Accuracy

- Goal: To provide a preliminary estimate of the protocol's accuracy in differentiating the CVI group from non-CVI participants (ASD and TD combined).
- Predictor Variables: Measures such as SRT, Startle Amplitude, etc.
- Outcome Variable: Group status (CVI vs. non-CVI).
- Analysis: A multi-step approach will be used:

1. Feature Selection: A principled method (e.g., LASSO regression) will be used to build a parsimonious predictive model and handle potential collinearity.
2. ROC Analysis: The performance of the final model will be assessed using Receiver Operating Characteristic (ROC) analysis. The primary outcome will be the Area Under the Curve (AUC) and its 95% confidence interval, derived via bootstrapping (1,000 resamples).

Limitations and future directions

This study has several limitations inherent to its exploratory proof-of-concept nature. First, the sample size is small and focused on a narrow age range, meaning that the results will mainly serve to generate effect size estimates for future larger-scale trials. Second, the protocol relies on consumer-grade hardware, the built-in infrared camera of an iPad, for eye tracking. While this approach greatly improves accessibility, its precision for capturing subtle

saccades and pupillary metrics at a typical clinical distance (e.g.,

60-100 cm) in a challenging paediatric population is not yet well

established in the scientific literature. Therefore, a key secondary

objective of this study is to assess the feasibility of this technology.

The results will provide important preliminary data on the viability

of using this accessible hardware for clinical research. Future work

should aim to validate these measures by comparing them to

research-grade eye tracking systems.

Similarly, the use of mmWave radar to quantify startle reflex

amplitude is an innovative but emerging technique for which

specific validation in paediatric neurology is lacking. Therefore, a

key secondary aim of this study is to provide crucial preliminary

data on the feasibility, reliability, and sensitivity of this method for

capturing non-contact neurological reflexes in children.

Participants will be primarily recruited from specialist clinical

services, which may introduce selection bias toward more severe

or complex CVI cases. As a result, the sample may not fully

represent the broader heterogeneity of children with CVI

encountered in community or educational settings. Future studies

will aim to expand recruitment to more diverse populations to

improve generalisability.

The current study relies on prior CVI diagnoses provided by

referring clinicians. While this reflects real-world diagnostic

pathways, it introduces the risk of misclassification due to

variability in diagnostic standards. This limitation is

acknowledged, and future studies will include standardised

diagnostic assessments and independent verification procedures to

improve diagnostic reliability.

Finally, while the protocol now controls for developmental level,

other factors such as socio-economic status and differences in non-

visual sensory processing (e.g., auditory hypersensitivity in the

ASD group) are not being matched or measured. These variables

could potentially influence the results, and their absence represents

a limitation of this study. Future research should aim to

characterise these variables more thoroughly.

Pre-registration

This study protocol has been pre-registered on the Open Science

Framework (OSF). The final, time-stamped registration will be

made public upon acceptance in principle. A private, anonymised

view-only link to the project repository, containing the protocol,

analysis scripts, and study materials, has been provided to the

recommender.

Declarations

Ethics Statement: This research protocol has been developed in

preparation for a future doctoral project. Formal ethics approval

will be sought from the responsible Human Research Ethics

Committee (HREC) at the author's host institution prior to the

commencement of any research activities involving human

participants.

Data and materials availability

In accordance with TOP guidelines, all processed anonymised data,

analysis scripts, and digital study materials required to reproduce

the study's findings will be made publicly available on a repository

such as the Open Science Framework (OSF). The EchoCue™

software script developed for stimulus delivery in this non-

commercial, academic research project will be included in this

public deposition.

Data availability statement

The preregistration, materials, and data for this study are available

on the Open Science Framework at

https://osf.io/6tj3f/?view_only=5c49d666e2c048aebbc110a53ff5bf

80.

References

1. Smague E. Literature Review of the Influence of Neurovisual Disorders on the Diagnosis of Autism in Australian Children. Current Developmental Disorders Reports. 2025;12(1):3.
2. Chokron S, Dutton GN. From vision to cognition: potential contributions of cerebral visual impairment to neurodevelopmental disorders. Journal of Neural Transmission. 2023;130(3):409–24.
3. McConnell EL, Saunders KJ, Little J-A. What assessments are currently used to investigate and diagnose cerebral visual impairment (CVI) in children? A systematic review. Ophthalmic & physiological optics : the journal of the British College of Ophthalmic Opticians (Optometrists). 2021;41(2):224–44.
4. Gordon S, Kerr A, Wiggs C, Chiang MF. What Is Cerebral/Cortical Visual Impairment and Why Do We Need a New Definition? Ophthalmology. 2024;131(12):1357–8.
5. Chang MY, Borchert MS. Methods of visual assessment in children with cortical visual impairment. Current Opinion in Neurology. 2021;34(1):89–96.
6. Esmail H, Arblaster G, Haslam L. The Assessment of Vision in Children with Severe Learning Difficulties: A Systematic Review. Br Ir Orthopt J. 2024;20(1):94–104.
7. Gruters KG, Groh JM. Sounds and beyond: multisensory and other non-auditory signals in the inferior colliculus. Front Neural Circuits. 2012;6:96.
8. Xiong Y, Zhang Y, Yan J. The neurobiology of sound-specific auditory plasticity: A core neural circuit. Neuroscience & Biobehavioral Reviews. 2009;33(8):1178–84.
9. Werner LA. The development of auditory behavior (or what the anatomists and physiologists have to explain). Ear Hear. 1996;17(5):438–46.
10. Moerel M, De Martino F, Uğurbil K, Yacoub E, Formisano E. Processing of frequency and location in human subcortical auditory structures. Scientific Reports. 2015;5(1):17048.
11. Brandt JP, Winters R. Bone Conduction Evaluation. StatPearls. Treasure Island (FL): StatPearls Publishing
Copyright © 2025, StatPearls Publishing LLC.; 2025.
12. Felix RA, 2nd, Gourévitch B, Portfors CV. Subcortical pathways: Towards a better understanding of auditory disorders. Hear Res. 2018;362:48–60.
13. Kraus N, Nicol T. Brainstem origins for cortical 'what' and 'where' pathways in the auditory system. Trends Neurosci. 2005;28(4):176–81.
14. Lauer AM, Behrens D, Klump G. Acoustic startle modification as a tool for evaluating auditory function of the

mouse: Progress, pitfalls, and potential. Neurosci Biobehav Rev. 2017;77:194–208.

15. Borg E. On the neuronal organization of the acoustic middle ear reflex. A physiological and anatomical study. Brain Res. 1973;49(1):101–23.

16. Yeomans JS, Frankland PW. The acoustic startle reflex: neurons and connections. Brain Res Brain Res Rev. 1995;21(3):301–14.

17. Chaieb L, Wilpert EC, Reber TP, Fell J. Auditory beat stimulation and its effects on cognition and mood States. Front Psychiatry. 2015;6:70.

18. Nakajima Y, Tanaka N, Mima T, Izumi SI. Stress Recovery Effects of High- and Low-Frequency Amplified Music on Heart Rate Variability. Behav Neurol. 2016;2016:5965894.

19. Park Y, Guzick AG, Schneider SC, Fuselier M, Wood JJ, Kerns CM, et al. Dental Anxiety in Children With Autism Spectrum Disorder: Understanding Frequency and Associated Variables. Front Psychiatry. 2022;13:838557.

20. Zatorre RJ, Belin P, Penhune VB. Structure and function of auditory cortex: music and speech. Trends in Cognitive Sciences. 2002;6(1):37–46.

21. Näätänen R, Kujala T, Escera C, Baldeweg T, Kreegipuu K, Carlson S, et al. The mismatch negativity (MMN) – A unique window to disturbed central auditory processing in ageing and different clinical conditions. Clinical Neurophysiology. 2012;123(3):424–58.

22. Stein BE, Stanford TR. Multisensory integration: current issues from the perspective of the single neuron. Nature Reviews Neuroscience. 2008;9(4):255–66.

23. Silva PR, Farias T, Cascio F, Dos Santos L, Peixoto V, Crespo E, et al. Neuroplasticity in visual impairments. Neurol Int. 2018;10(4):7326.

24. Rule SJ, Nickolaychuk BR. Lateralization of high frequency sounds as a function of interaural amplitude disparity. Can J Exp Psychol. 1995;49(3):368–75.

25. Gerdfelder B, Greičiūtė M, Wiens S. Examining experienced lateralization of sounds over headphones with electroencephalography. Neuropsychologia. 2025;207:109064.

26. Reznik D, Henkin Y, Schadel N, Mukamel R. Lateralized enhancement of auditory cortex activity and increased sensitivity to self-generated sounds. Nat Commun. 2014;5:4059.

27. Middlebrooks JC, Knudsen EI. Changes in external ear position modify the spatial tuning of auditory units in the cat's superior colliculus. J Neurophysiol. 1987;57(3):672–87.

28. Novembre G, Pawar VM, Bufacchi RJ, Kilintari M, Srinivasan M, Rothwell JC, et al. Saliency Detection as a Reactive

Process: Unexpected Sensory Events Evoke Corticomuscular Coupling. J Neurosci. 2018;38(9):2385–97.

29.	Litovsky R. Development of the auditory system. Handb Clin Neurol. 2015;129:55–72.

30.	Moore JK, and Linthicum Jr FH. The human auditory system: A timeline of development. International Journal of Audiology. 2007;46(9):460–78.

31.	Luo Q, Shen Y, Chen T, Zheng Z, Shi H, Feng Y, et al. Effects of SoundBite Bone Conduction Hearing Aids on Speech Recognition and Quality of Life in Patients with Single-Sided Deafness. Neural Plast. 2020;2020:4106949.

32.	Musiek FE, Baran JA. The auditory system : anatomy, physiology, and clinical correlates. Second edition. ed. San Diego, CA: Plural Publishing; 2020.

33.	Davidson LS. Effects of stimulus level on the speech perception abilities of children using cochlear implants or digital hearing aids. Ear Hear. 2006;27(5):493–507.

34.	Guidelines for manual pure-tone threshold audiometry. Asha. 1978;20(4):297–301.

35.	Kawakami S, Uono S, Otsuka S, Yoshimura S, Zhao S, Toichi M. Atypical Multisensory Integration and the Temporal Binding Window in Autism Spectrum Disorder. Journal of Autism and Developmental Disorders. 2020;50(11):3944–56.

36.	Jones GP, Lukashkina VA, Russell IJ, Lukashkin AN. The vestibular system mediates sensation of low-frequency sounds in mice. J Assoc Res Otolaryngol. 2010;11(4):725–32.

37.	Anne Small S, and Richard Stapells D. Maturation of bone conduction multiple auditory steady-state responses. International Journal of Audiology. 2008;47(8):476–88.

38.	Marin-Campos R, Dalmau J, Compte A, Linares D. StimuliApp: Psychophysical tests on mobile devices. Behavior Research Methods. 2021;53(3):1301–7.

39.	Nissen L, Hübner J, Klinker J, Kapsecker M, Leube A, Schneckenburger M, et al. Towards Preventing Gaps in Health Care Systems through Smartphone Use: Analysis of ARKit for Accurate Measurement of Facial Distances in Different Angles. Sensors (Basel). 2023;23(9).

40.	Gao Z, Ali L, Wang C, Liu R, Wang C, Qian C, et al. Real-Time Non-Contact Millimeter Wave Radar-Based Vital Sign Detection. Sensors (Basel). 2022;22(19).

41.	Bossuyt PM, Reitsma JB, Bruns DE, Gatsonis CA, Glasziou PP, Irwig LM, et al. Towards complete and accurate reporting of studies of diagnostic accuracy: the STARD initiative. Standards for Reporting of Diagnostic Accuracy. Clin Chem. 2003;49(1):1–6.

42.   Hessels RS, Hooge ITC. Eye tracking in developmental
cognitive neuroscience – The good, the bad and the ugly.
Developmental Cognitive Neuroscience. 2019;40:100710.
43.   Franchak JM, Kretch KS, Soska KC, Adolph KE. Head-
mounted eye tracking: a new method to describe infant looking.
Child Dev. 2011;82(6):1738–50.