Full title:

LAMP-coupled CRISPR-Cas12a assays for upgrading molecular

detection of *Leishmania* infections

Short title:

LAMP-CRISPR mediated *Leishmania* detection

Eva Dueñas[1,#a¶], Ingrid Tirado[1,2¶], Percy Huaihua[3], Ariana Parra del Riego[1], Luis

Cabrera-Sosa[1,4,#b], Jose A. Nakamoto[1,#c], María Cruz[5], Carlos M. Restrepo[6], Jorge

Arévalo[3,4], Vanessa Adaui[1,4*]

[1] Laboratory of Biomolecules, Faculty of Health Sciences, Universidad Peruana de

Ciencias Aplicadas (UPC), Lima, Peru; [2] Facultad de Ciencias Biológicas, Universidad

Nacional Mayor de San Marcos, Lima, Peru; [3] Laboratorio de Patho-antígenos,

Laboratorios de Investigación y Desarrollo, Facultad de Ciencias e Ingeniería,

Universidad Peruana Cayetano Heredia, Lima, Peru; [4] Instituto de Medicina Tropical

Alexander von Humboldt, Universidad Peruana Cayetano Heredia, Lima, Peru; [5]

Facultad de Ciencias de la Salud, Universidad Andina del Cusco, Cusco, Peru; [6] Centro

de Biología Molecular y Celular de Enfermedades, Instituto de Investigaciones

Científicas y Servicios de Alta Tecnología (INDICASAT AIP), Panama City, Panama

[#a]Current Address: Life Science Research Centre, Faculty of Science, University of

Ostrava, Ostrava, Czech Republic

[#b]Current Address: Department of Biomedical Sciences, Institute of Tropical Medicine

Antwerp, Antwerp, Belgium

#cCurrent Address: Laboratory of Protein Evolution, Department of Experimental

Medical Science, Lund University, Lund, Sweden

*Corresponding author: vanessa.adaui@upc.pe (VA)

¶These authors contributed equally to this work.

Keywords: leishmaniasis, *Leishmania*, CRISPR-Cas, nucleic acid detection, isothermal

amplification, LAMP, kDNA, 18S rDNA

Abstract

Background: Tegumentary leishmaniasis is a parasitic disease endemic in the Americas. Its clinical management and control rely on early and accurate diagnosis and adequate treatment. PCR-based molecular diagnostics offer high sensitivity and specificity over microscopy or culture but are less accessible in low-resource settings. New molecular tools for detecting *Leishmania* infections are needed in rural endemic regions. A promising tool harnessing CRISPR-Cas technology enables highly specific and sensitive detection of nucleic acid targets, offering an exciting potential for portable molecular diagnostics. Previously, we developed CRISPR-Cas12a-based assays coupled to PCR preamplification for *Leishmania* detection. Here, we adapted our assays, which target the 18S rDNA and kinetoplast DNA (kDNA) minicircles, by replacing PCR with loop-mediated isothermal amplification (LAMP).

Methodology/Principal Findings: LAMP-CRISPR assays were optimized for fluorescence-based and lateral flow readouts. The assays could detect as low as 0.2 genome equivalents per reaction using *L. braziliensis* M2904 strain genomic DNA. The kDNA assay reliably detected all tested species of the *Leishmania* (*Viannia*) subgenus, while the 18S assay showed pan-*Leishmania* detection capability. There was no cross-reactivity with other protozoan (*Trypanosoma cruzi* and *Plasmodium falciparum*) and bacterial (*Mycobacterium tuberculosis*) pathogen DNA, or with human DNA. When applied to 90 clinical samples (skin lesions) from the Cusco region of Peru and compared to kDNA real-time PCR, LAMP-CRISPR assays with a fluorescence readout achieved a sensitivity of 90.9% for kDNA and 72.7% for 18S rDNA, both with 100% specificity. Overall, lateral flow strip results agreed with fluorescence-based detection in 18 tested samples, with one discrepancy observed in the 18S assay associated with low parasite load.

60   **Conclusions/Significance**: These new assays, being amenable to further simplification

61   and optimization for their adoption in low-resource settings, hold promise as a new

62   generation of accurate molecular tools for leishmaniasis diagnosis and surveillance,

63   supporting One Health strategies for disease control.

Author Summary

Tegumentary leishmaniasis affects poverty-related populations in the Americas and

encompasses skin and mucosal lesions that can cause disfigurement and social stigma.

The disease is caused by several species of the protozoan parasite *Leishmania*. PCR-

based molecular diagnostics are currently the most sensitive and specific diagnostic

tools. Yet, these require specialized infrastructure and trained personnel that are not

readily available in low-resource settings. New tools are required to meet the diagnostic

needs in rural endemic areas. A promising tool leveraging CRISPR-Cas technology

enables cost-effective, *in vitro* nucleic acid detection, paving the way for diagnostic

solutions that could be made available to patients at, or near, the point of care. Here, we

harnessed the CRISPR-Cas12a system combined with loop-mediated isothermal

amplification (LAMP) to develop assays capable of detecting multiple *Leishmania*

species of medical importance. Our assays employ multi-copy targets widely used in

molecular diagnostics: the 18S rDNA for pan-*Leishmania* detection and a kDNA

minicircle region conserved among *L.* (*Viannia*) species. Results can be read with either

fluorescence detection or lateral flow strips. Both assays showed satisfying performance

in both analytical validation and clinical sample testing under laboratory conditions.

These new tools show promise to improve diagnosis and surveillance of leishmaniasis.

Introduction

Leishmaniasis is one of the most serious neglected infectious diseases, with 0.7 to 1 million new cases each year and between 20,000 and 30,000 deaths [1, 2]. As a vector-borne disease, leishmaniasis poses a substantial threat to millions of individuals in regions where the disease is endemic, which include Asia, South and Central America, Africa and the Mediterranean basin [3]. Leishmaniasis results from an infection with protozoan parasites of the genus *Leishmania* that are transmitted to humans (and other mammals) through the bite of infected sandflies and manifests in two main clinical forms: cutaneous (or tegumentary) and visceral leishmaniasis, each presenting unique challenges in terms of diagnosis, treatment, and control [4, 5]. In the Americas, tegumentary leishmaniasis is widely distributed in 18 endemic countries and comprises a spectrum of clinical conditions ranging from localized skin ulcers to severe disfiguring mucosal lesions [5]. This clinical pleomorphism is largely dependent on the immune responses of the human host as well as on factors intrinsic to the different infecting *Leishmania* species [6]. Diverse *Leishmania* species from the *Viannia* and *Leishmania* subgenera circulate in several Central and South American countries, with the former being responsible for the majority of leishmaniasis cases [4]. While efforts directed toward the development of effective prophylactic and therapeutic vaccines against human leishmaniasis are ongoing, early and accurate diagnosis paired with timely chemotherapy remain the most vital tools in the fight against the leishmaniases [7].

The diagnostic tools currently available for the detection of *Leishmania* spp. infections include direct smear examination (light microscopy), *in vitro* culture, immunological and molecular techniques [8]. Among the nucleic acid amplification techniques (NAATs), the polymerase chain reaction (PCR) and its variants, such as

real-time PCR, currently provide the most sensitive and specific diagnostic tools available, and several assays have been developed for the detection and identification of *Leishmania* species [9, 10]. While accessible in research and reference laboratories, PCR-based methods may not be readily available in rural endemic areas since they require infrastructure, resources, and technical expertise [11]. Because of this, (near) point-of-care (PoC) diagnostic tests have become a noteworthy concept, as they would allow easier, faster and less expensive diagnostic alternatives that need only minimal laboratory setup [12]. Advancements towards simplifying molecular diagnostic tools gave rise to multiple isothermal NAATs, such as loop-mediated isothermal amplification (LAMP) and recombinase polymerase amplification (RPA), which hold potential to decentralize molecular diagnostic testing [13]. Of these, LAMP has gained prominence due to its simplicity and effectiveness in detecting nucleic acid sequences with high specificity. LAMP is different from the traditional PCR as it operates at a constant temperature (60-65°C), eliminating the need for complex thermal cycling equipment [14, 15]. The LAMP reaction uses 4 to 6 primers that target specific DNA sequences, along with the Bst DNA polymerase, an enzyme that displays a strong strand displacement activity [14]. LAMP amplifies the target DNA producing a concatenated DNA product with stem-loop DNA structures, yielding up to 10^9 copies in less than an hour [14]. The results can be easily visualized, e.g. on the basis of the turbidity caused by reaction by-products [16], or with the use of sequence-specific probes [17, 18]. Due to its user-friendly nature and rapidity, LAMP holds great promise as a tool for the diagnosis of infectious diseases in resource-limited settings [19]. Several studies have reported LAMP assays for genus-, complex- or species-specific detection of *Leishmania*; these achieved high sensitivity and specificity and showed good diagnostic performance for human cutaneous leishmaniasis (CL) and visceral leishmaniasis (VL)

(reviewed in [20, 21]). A commercially available LAMP kit (Loopamp™ *Leishmania*

Detection Kit, developed by Eiken Chemical, Japan) to detect *Leishmania* DNA has

also been evaluated for the diagnosis of human CL and VL and showed comparable

performance to PCR-based methods [22].

Clustered regularly interspaced short palindromic repeats (CRISPR)/CRISPR-

associated proteins (Cas) technology has recently been harnessed as a tool for molecular

detection and the development of CRISPR-based diagnostics [23], based on the

discovery that once Cas12a and Cas13a enzymes bound to CRISPR RNA (crRNA)

recognize specific nucleic acid targets, they become activated, thereby catalyzing

cleavage of both the targets (*cis*-cleavage) as well as non-target nucleic acids (*trans*-

cleavage) [24, 25]. This collateral *trans*-cleavage activity can be applied for nucleic acid

detection by coupling it with a reporter molecule. For fluorescence-based assays, the

reporter molecule consists of a DNA or RNA oligomer carrying a fluorophore and a

quencher. So, upon target recognition, reporter cleavage by the activated Cas enzyme

generates a detectable fluorescence signal that indicates that the target nucleic acid

sequence is present in a sample. Alternatively, visualization of the Cas detection

reaction can be achieved by a lateral flow strip readout using a reporter molecule

labeled with FAM or digoxin and biotin [26, 27]. CRISPR-based assay platforms

reported thus far enable a highly sensitive detection of nucleic acids in the attomolar

range when combined with a target preamplification step using either isothermal

NAATs (e.g. DETECTR [24] and SHERLOCK [25]) or PCR-based NAATs (e.g.

HOLMES [28]). Notwithstanding that this two-step CRISPR-based assay format

increases assay complexity and poses the risk of cross-contamination during sample

transfer to the CRISPR reaction, it is a convenient way to pursue in the first

development phase of CRISPR-based assays to demonstrate their potential for further

assay development [29]. More recent efforts aim for amplification-free detection of nucleic acid targets with CRISPR, and the proof-of-concept was demonstrated for direct detection of pathogen RNA and DNA with quantitative ability [30-33].

The combination of CRISPR-Cas and LAMP technologies has advanced the progress of nucleic acid-based PoC diagnostic test development. Coupling LAMP reactions to downstream CRISPR-based detection improves specificity [19, 23], since LAMP is prone to non-specific amplification [34]. Integrated LAMP-CRISPR assays have been developed for the detection of several pathogens, including viruses (e.g. SARS-CoV-2 [26, 35-37], human papillomavirus [38]), bacteria (e.g. *Mycobacterium tuberculosis* [39], *Klebsiella pneumoniae* [40]), and fungi (e.g. *Verticillium dahliae* [41]). CRISPR-based assays have also been developed for the detection of parasites of medical and veterinary importance, such as the protozoa *Plasmodium* spp. [29, 42], *Toxoplasma gondii* [43], *Leishmania* spp. [44-47], *Trypanosoma brucei* [48, 49], and *Trypanosoma cruzi* [50]; and the trematode *Schistosoma* species [51, 52]. These assays, aiming at sensing parasite nucleic acid sequences in biological specimens, mostly combined CRISPR with RPA preamplification. Applications of LAMP-CRISPR for parasite detection are increasing, as illustrated by the detection of *Schistosoma* infection in host samples [53], the *in silico* design of LAMP primers and Cas12a crRNA sequences for identifying genetic polymorphisms associated with drug resistance in *P. falciparum* infections [54], the detection of *T. gondii* in environmental samples [55], and the detection of *T. cruzi* in infants for diagnosing acute congenital infection [56].

At present, leishmaniasis diagnosis in endemic areas is impaired mainly by the lack of infrastructure, equipment, trained human resources, and low access (if any) to effective diagnostic tools. New molecular diagnostic tools that are accurate, easy to use, accessible and affordable for use in resource-limited settings are required. In previous

work, we developed CRISPR-Cas12a-based assays coupled to PCR preamplification for

detection of *Leishmania* spp. in clinical samples [45]. Here, we have adapted our assays,

which target the *Leishmania* 18S ribosomal RNA gene (18S rDNA) and minicircle

kinetoplast DNA (kDNA), by replacing PCR with LAMP in the preamplification step,

to facilitate the road towards the development of a nucleic acid-based near-PoC

diagnostic tool for leishmaniasis. We optimized LAMP-CRISPR assay conditions with

a fluorescence signal readout in a plate reader as well as a visual readout on lateral flow

strips. We report the evaluation of LAMP-CRISPR assay diagnostic performance

against a kDNA real-time PCR assay on stored clinical samples from the Cusco region,

a highly endemic area for tegumentary leishmaniasis in south-central Peru where

Leishmania species of the *Viannia* subgenus circulate [57, 58].

Methods

DNA samples from reference strains of *Leishmania* and other microbial pathogens

Genomic DNA (gDNA) samples extracted from cultured promastigotes of reference

strains of *Leishmania* spp. analyzed here were retrieved from the DNA biobank of the

Leishmania research group at the Molecular Epidemiology Unit of the Instituto de

Medicina Tropical Alexander von Humboldt (IMTAvH), Universidad Peruana

Cayetano Heredia (UPCH) in Lima, Peru, and from the collection hosted by the Institute

for Scientific Research and High Technology Services (INDICASAT-AIP) in Panama.

gDNA samples isolated from cultured laboratory reference strains of other microbial

pathogens (*Trypanosoma cruzi*, *Plasmodium falciparum*, *Mycobacterium tuberculosis*)

were provided by the Infectious Diseases Research Laboratory at the UPCH and

INDICASAT-AIP. The designation and source of the strains whose gDNA were

analyzed in this study are listed in Table 1.

At the IMTAvH and UPCH, DNA was isolated using the High Pure PCR

Template Preparation Kit (Roche, Mannheim, Germany). At the INDICASAT-AIP, the

Wizard® Genomic DNA Purification Kit (Promega, Madison, WI, USA) was used.

DNA was isolated according to the manufacturer's protocol, suspended in elution buffer

or TE buffer and stored at -20°C until use.

Table 1. Reference strains used in this work.

Strain name	WHO code[¥]	Source
Leishmania (Viannia) braziliensis	MHOM/BR/75/M2904	IMTAvH-UPCH
Leishmania (Viannia) braziliensis	MHOM/BR/75/M2903	INDICASAT-AIP
Leishmania (Viannia) braziliensis	MHOM/PE/91/LC2043	IMTAvH-UPCH
Leishmania (Viannia) braziliensis	MHOM/PE/91/LC2177	IMTAvH-UPCH
Leishmania (Viannia) peruviana	MHOM/PE/90/HB86	IMTAvH-UPCH
Leishmania (Viannia) peruviana	MHOM/PE/90/LCA08	IMTAvH-UPCH
Leishmania (Viannia) peruviana	MHOM/PE/2005/WR-2771	INDICASAT-AIP
Leishmania (Viannia) guyanensis	MHOM/BR/1975/M4147	INDICASAT-AIP
Leishmania (Viannia) lainsoni	MHOM/BR/1981/M6426	INDICASAT-AIP
Leishmania (Viannia) panamensis	MHOM/PA/2018/BD-02	INDICASAT-AIP
Leishmania (Leishmania) amazonensis	MHOM/BR/1973/M2269	INDICASAT-AIP
Leishmania (Leishmania) mexicana	MHOM/BZ/1982/BEL21	INDICASAT-AIP
Leishmania (Leishmania) aristidesi	MORY/PA/1969/GML	INDICASAT-AIP
Leishmania (Leishmania) major	MHOM/SA/1991/WR-1088	INDICASAT-AIP
Leishmania (Leishmania) infantum	MCAN/BR/1997/P142	INDICASAT-AIP
Leishmania (Leishmania) donovani	MHOM/IN/2006/WR-2801	INDICASAT-AIP
Trypanosoma cruzi	Sylvio X10	UPCH
Trypanosoma cruzi	VT-1	UPCH
Trypanosoma cruzi	Tulahuen LacZ clone C4	UPCH
Plasmodium falciparum	HB3	INDICASAT-AIP
Mycobacterium tuberculosis	PA155-18C	INDICASAT-AIP

219   [¥]The World Health Organization (WHO) designation code for *Leishmania* strains is as follows: host (M

220   for Mammalia; HOM = *Homo sapiens*; CAN = *Canis familiaris*; ORY = *Oryzomys capito*)/country of

221   origin (BR = Brazil; BZ = Belize; IN = India; PA = Panama; PE = Peru; SA = Saudi Arabia)/year of

222   isolation/name of strain.

Patient DNA samples

Ethics statement

This study used anonymized stored DNA samples that were assigned a code and had

been isolated from different human skin ulcerative lesion specimen types, including

biopsy, lancet scraping, cytological brush, swab, and filter paper. The respective

patients had provided written informed consent for future research use of their

specimens and clinical data. Patients with clinically suspected CL were recruited at the

Hospital Nacional Adolfo Guevara Velasco (HNAGV) in Cusco, a region in Peru with

endemic circulation of *Leishmania* (*Viannia*) parasites, most frequently *L. braziliensis*,

L. guyanensis and *L. lainsoni* [57, 58], as part of another study between the HNAGV

and UPCH (contract 095-2018-FONDECYT-BM-IADT-AV to JA, financed by

CONCYTEC and The World Bank). The study protocol and informed consent

(registration number 103155) were reviewed and approved by the ethics committee of

the UPCH (IRB approval letter 063-05-19 dated 01/30/2019, latest renewed on

05/02/2023 with letter R-149-17-23).

DNA extraction from clinical samples prior to this study

DNA from clinical samples was isolated with the High Pure PCR Template Preparation

Kit (Roche), according to the manufacturer's protocol, suspended in elution buffer and

stored at -20°C until use.

LAMP primer design

The 18S rDNA gene (TriTrypDB ID: LbrM.27.2.208540; at

https://tritrypdb.org/tritrypdb/) and kDNA minicircles (GenBank accession no.

KY698803.1) from *L. braziliensis* MHOM/BR/75/M2904 were selected for LAMP

primer design. FASTA sequences up to 200 bp downstream and upstream to previously

reported crRNA target recognition sites for each gene [45] were uploaded to

PrimerExplorer V5 software (Eiken Chemical Co., Ltd). Three sets of external (F3 and

B3) and internal (FIP and BIP) primers were obtained for 18S rDNA, but only one for

kDNA. To select the final 18S primer set, the following features were considered: the

less likelihood of primer dimer formation (highest ΔG value), the 3' end stability for F2

and B2 ($\Delta G \leq$ -4 kcal/mol), the 5' end stability for F1c and B1c ($\Delta G \leq$ -4 kcal/mol) and

the less overlap between primers and the recognition site. After that, loop primers (LF

and LB) for both target genomic regions were also designed using the PrimerExplorer

V5 software. Final primer sets were ordered from Macrogen Inc. (Seoul, South Korea)

and are listed in the S1 Table.

To investigate the *in silico* specificity of the designed LAMP primers, we

downloaded all publicly available DNA sequences for each target from NCBI in May

2025. For *Leishmania* kDNA minicircle sequence analysis, a total of 1,060 sequences

from the *L.* (*Viannia*) subgenus and 965 sequences from the *L.* (*Leishmania*) subgenus

were retained and analyzed following a quality filtering process. Sequences that were

excluded contained ambiguous base pairs (e.g. 'N') exceeding 10% of the total length or

did not cover the primer binding sites. The eight LAMP primer binding regions and the

crRNA target site were manually annotated using Jalview v2.11.4.1 [59] following

multiple sequence alignment with Clustal Omega [60]. Sequence alignments are

available in the S1 File. A representative subset of 59 sequences (24 *Viannia*, 35

Leishmania) is illustrated in S1 Fig to indicate sequence conservation or variations

across the two analyzed *Leishmania* subgenera. As for the 18S rDNA, 37 sequences

from *Leishmania* spp. and 272 from *Trypanosoma* spp. were aligned and analyzed using

the same bioinformatics pipeline (S2 File). A subset of 31 representative sequences (16

Leishmania spp., 15 *Trypanosoma* spp.) was selected to denote sequence conservation

or variations across the two analyzed trypanosomatid genera (S2 Fig).

crRNA guide sequences and preparation

We used previously reported LbCas12a crRNA guide sequences targeting 18S rDNA

and kDNA minicircle sequences [45]. crRNA preparation from double-stranded DNA

templates (see S1 Table for nucleotide sequences) was performed as described

previously [45].

Loop-mediated isothermal amplification (LAMP) assay

We first optimized the LAMP assay reaction conditions using a commercial Bst 2.0$^{®}$

DNA polymerase (New England Biolabs Inc (NEB), Massachusetts, USA; Cat no.

M0537), following the LAMP protocol as suggested by the manufacturer (NEB). The

only modification was the final total concentration of $MgSO_4$ in the reaction (6 mM

Mg^{2+} for 18S rDNA and 5 mM Mg^{2+} for kDNA). These reaction conditions were used

to evaluate the incubation time needed for optimal amplification of the target DNA,

which was tested at 2×10^2, 2×10^{-1}, and 2×10^{-3} parasite genome equivalents (GE) per

reaction to cover a wide range of target abundance. The LAMP reaction products were

analyzed by 2% agarose gel electrophoresis using SYBR Gold staining and tested

downstream using CRISPR-Cas12a assays with fluorescence readout (see below).

To facilitate master mix preparation for sample processing, LAMP reactions for

the preamplification of *Leishmania* DNA targets were performed using the WarmStart®

LAMP Kit (DNA & RNA) (NEB; Cat no. E1700). Each reaction contained 12.5 µL of

WarmStart LAMP 2X Master Mix, 8 µL of molecular biology grade water, and 2.5 µL

of 10X LAMP primer mix for a single target (kDNA or 18S rDNA). Like for the LAMP

protocol using the separate Bst enzyme, the six target-specific primers (3 sets) were

added at a final concentration of 1.6 µM for forward inner and backward inner primers

(FIP/BIP), 0.4 µM for loop forward and loop backward primers (LF/LB), and 0.2 µM

for forward external and backward external primers (F3/B3). After that, we carefully

added 20 µL of immersion oil over the LAMP mix to create two layers to prevent cross

contamination. Then, 2 µL of DNA template (0.26 – 40 ng, for analytical testing) was

added to the first (bottom) layer, corresponding to the LAMP mixture, bringing the final

reaction volume to 25 µL. The LAMP reaction was incubated on a heating block at

65°C for 30 min for kDNA or for 50 min for 18S rDNA, and then stopped by incubating

at 82°C for 3 min.

A positive control (*L. braziliensis* M2904 gDNA, $10^2 – 10^4$ GE as indicated in

the figure legends) and a negative amplification control (LAMP reaction without DNA

template, i.e., No-Template Control (NTC)) were included in all experiments. In the

analytical specificity testing, a negative control consisting of human gDNA from

peripheral blood mononuclear cells (PBMC) of a healthy donor (40 ng of input DNA)

was also included.

CRISPR-Cas12a assay with fluorescence readout

LbCas12a-based detection reactions were performed as described previously [45, 61,

62], with the only modification being that a different reporter probe was used, which

allows fluorescence signal visualization on a LED transilluminator. We prepared the

ribonucleoprotein (RNP) complex 10X by mixing recombinant LbCas12a (produced in-

house [63]) with the target-specific folded crRNA (incubated at 65°C for 10 min

followed by room temperature (RT) incubation for 10 min) in a 1:1.5 molar ratio and 2

µM single-stranded DNA (ssDNA) reporter probe (/56-FAM/TTATT/3IABkFQ/; [24])

in reaction buffer 1X (NEBuffer 2.1 recipe, prepared in-house, without adding $MgCl_2$).

The labeled fluorophore/quencher probe was purchased from Macrogen Inc. (Seoul,

South Korea). A volume of 10 µL of the 10X RNP complex (pre-incubated for 10 min

at RT) was loaded into each well of a flat-bottom, black 96-well microplate (Thermo

Fisher Scientific, Waltham, MA, United States; Cat. no. 237107). Subsequently, 90 µL

of the diluted LAMP product –prepared by diluting 2 µL of the LAMP product in 106

µL of 1X reaction buffer containing 16.7 mM $MgCl_2$– was added. The 100 µL

LbCas12a *trans*-cleavage reaction, with a final $MgCl_2$ concentration of 15 mM, was

incubated and analyzed in a fluorescence plate reader (Synergy™ H1 hybrid multi-

mode reader; BioTek Instruments, Winooski, VT, United States) for 121 min at 25°C.

Fluorescence measurements were recorded every 1 min (λex: 490 nm; λem: 525 nm)

with a fluorescence gain setting of 120. For the experiments shown in S3 Fig,

measurements used the Varioskan™ LUX multimode microplate reader (Thermo Fisher

Scientific).

Optimization of the CRISPR-Cas12a-based detection assay coupled with lateral

flow readout

The lateral flow assay (LFA) readout was optimized using the HybriDetect - Universal

lateral flow assay kit (Milenia Biotec GmbH, Giessen, Germany; Cat. no. MGHD 1)

according to the manufacturer's instructions. First, to optimize the reporter probe

concentration (for reducing the background signal on the T-line of the LFA strip in

negative controls), we tested the reporter at a final concentration (1X) of 500 nM, 250

nM, and 100 nM per LFA. Second, we tested the reaction buffer 1X without $MgCl_2$

(prepared in-house following the NEBuffer 2.1 recipe) as compared to the HybriDetect

universal assay buffer. Third, we tested two incubation temperatures of the Cas12a

reaction: either 25°C or 37°C, for a reaction time of 30 min or 60 min. Fourth, we

evaluated the addition of polyethylene glycol (PEG)-8000 at 2% and 4% final

concentration to the completed CRISPR-Cas12a reaction for increasing the viscosity of

the buffer, which in turn slows down lateral flow speed.

Optimized CRISPR-Cas12a-based lateral flow strip assay for clinical samples and reference strain genomic DNA

After completion of the preamplification step, 2 µL of the LAMP product was mixed

with 106 µL of 1X reaction buffer with 16.7 mM $MgCl_2$ in a 1.5 mL microcentrifuge

tube. Then, we took 90 µL of the diluted LAMP product in a new microcentrifuge tube

and added 10 µL of the 10X pre-incubated RNP complex (100 nM Cas12a: 150 nM

crRNA: 1000 nM reporter probe). The lateral flow cleavage reporter probe

(5'/Bio/TTATTATT/6-FAM/3'; with terminal biotin and FAM labels in the opposite

order relative to [26]) was purchased as an HPLC-purified oligo from biomers.net

GmbH (Ulm, Germany). The 100 µL LbCas12a *trans*-cleavage reaction, with a final

$MgCl_2$ concentration of 15 mM, was incubated at 37°C for 60 min in the dark.

Afterward, 5 µL of PEG-8000 (2% final concentration) was added to the reaction tube

and homogenized by pipetting up and down. A lateral flow strip (Milenia HybriDetect)

was placed into the reaction tube and incubated for 5 min at RT. Lateral flow strips

were interpreted visually. LFA strip images shown here were taken with an iPhone 13

Pro camera.

Analytical sensitivity testing

Serial dilutions of *L. braziliensis* M2904 gDNA (extracted from a promastigote culture)

to a final range of 2×10^4 to 2×10^{-3} parasite GE per reaction (encompassing the same

range of parasite GE per reaction of the standard curve used in the kDNA qPCR assay

[64]) were tested as input DNA in LAMP reactions to amplify the target gene, followed

by LbCas12a-based detection assays as described above, using a plate-based

fluorescence readout and a colorimetric lateral flow-based readout. The analytical

sensitivity for each target gene was determined based on 3 independent experiments.

Analytical specificity testing

The analytical specificity of the LAMP-CRISPR assays targeting *Leishmania* kDNA

(*Viannia* subgenus) or 18S rDNA was tested using gDNA from representative

laboratory reference strains of New World and Old World *Leishmania* species, as well

as other microbial pathogens (Table 1). Target genes were amplified by LAMP (using

20–40 ng of input DNA) and detected by LbCas12a-based detection assays as

described above, using fluorescence-based and lateral flow-based readouts. Two

independent LAMP-CRISPR experiments were performed.

Evaluation of LAMP-CRISPR assay performance on clinical samples

Patient DNA samples extracted from skin lesion specimens were tested blindly by the

LAMP-CRISPR assays in groups of ten, plus positive and negative controls. The DNA

input for the LAMP preamplification step was 5 µL of diluted or undiluted DNA

samples (6–285 ng of input DNA; see S1 Dataset) to capture data from samples with

low *Leishmania* parasite load levels. The LAMP products were then detected by

LbCas12a-based detection assays as described above.

Quantitative real-time PCR (qPCR)

A qPCR assay targeting *Leishmania* kDNA minicircles was performed to detect and quantify *Leishmania* (*Viannia*) parasites in clinical samples, using primers and cycling conditions as described [64]. The human *ERV-3* gene was amplified in parallel using the primers reported by Yuan *et al.* [65]; it served as an indicator of specimen adequacy and to estimate the parasite load normalized to the number of human cell equivalents per sample [64].

Data processing and analysis

The data collected from LAMP-CRISPR assays were analyzed as described previously [45]. Briefly, the raw fluorescence data from each well of the LbCas12a-based assay plate were exported to Microsoft Excel. Based on the inspection of fluorescence time-course data from samples and controls that were run in parallel, the time point of fluorescence accumulation for data analysis was defined (at 25-min time point for both kDNA and 18S rDNA). The raw fluorescence data were normalized by dividing target reaction fluorescence accumulated at the defined time point of the test sample to that of the NTC reaction included in parallel in the same Cas12a assay plate (further called 'fluorescence ratio'). In the analytical validation, a result was considered 'detected' if a target reaction produced a fluorescence ratio ≥ 2 for the test sample above background (NTC). To set a threshold cutoff value for positive signal (i.e., detection of *Leishmania*) in clinical samples, we required that the fluorescence ratio be larger than the mean and three standard deviations of negative clinical samples (see S1 Dataset) selected as a separate panel for determining the cutoff. Graphs, numerical data analyses, and statistical analyses were performed using GraphPad Prism version 10.2 (GraphPad Software, San Diego, CA, United States).

422   For the diagnostic performance evaluation of the LAMP-CRISPR assays with

423 fluorescence readout, the sensitivity and specificity and their 95% confidence intervals

424 were calculated with MedCalc software [66] using kDNA real-time PCR as the

425 reference test. To evaluate the accuracy of the LAMP-CRISPR assays to discriminate a

426 positive (detected) from a negative (not detected) *Leishmania* infection status, a receiver

427 operating characteristic (ROC) curve was drawn and the area under the curve (AUC)

428 with its 95% confidence interval was calculated using GraphPad Prism version 10.2. In

429 all statistical tests, a *P* value of < 0.05 was considered statistically significant.

430   To assess the applicability of the LAMP-CRISPR assays with an LFA readout,

431 18 clinical samples selected for pilot testing were arbitrarily categorized into three

432 groups based on parasite load: low (<100 parasites/10^6 human cells), intermediate (100

433 – 10,000 parasites/10^6 human cells), and high (>10,000 parasites/10^6 human cells). LFA

434 strips were interpreted blinded to the fluorescence results.

436 **Results**

437 **Design and optimization of LAMP-CRISPR assays**

438 We developed two assays for the detection of *Leishmania* spp., which are based on the

439 most frequently used multicopy genetic markers in molecular diagnostics for

440 leishmaniasis: kDNA minicircles and 18S rDNA. Our assays combine LAMP with

441 CRISPR-LbCas12a-mediated DNA target recognition and ssDNA reporter cleavage,

442 with results assessed either with a fluorescence-based or lateral flow strip readout (Fig

443 1). These two readouts are made possible through the use of different dual-labeled

444 ssDNA reporter molecules: one labeled with a fluorophore-quencher pair for real-time

445 monitoring of fluorescence curves over the Cas12a reaction time course on a plate

reader, and the other labeled with biotin and FAM for results to be visually read with

paper-based lateral flow strips [26].

Fig 1. Schematic representation of the workflow of LAMP-CRISPR assays for

detection of *Leishmania* spp. in clinical samples. Sample preparation is performed

from skin lesion specimens taken from patients with suspected CL. The DNA samples

are subjected to LAMP amplification for just 30 min (kDNA) or 50 min (18S rDNA),

followed by Cas12a-based reactions with either a fluorescence-based or lateral flow

strip readout. Figure created with BioRender.com.

We designed LAMP primers flanking previously reported crRNA target

recognition sites [45] within *Leishmania* kDNA minicircles (S1 Fig) and 18S rDNA (S2

Fig). According to the *in silico* analyses, the kDNA primer set targets a region most

conserved within the *L.* (*Viannia*) subgenus and less conserved within the *L.*

(*Leishmania*) subgenus (S1 Fig and S1 File). *In silico* analysis of the selected 18S

primer set showed the expected conserved nature of the targeted genomic region at the

Leishmania genus level, whereas sequence variations were more frequent within the

Trypanosoma genus (S2 Fig and S2 File).

A schematic depicting the mechanism of LAMP amplification, with the resulting

LAMP amplicon containing the Cas12a crRNA target sequence is shown in Fig 2. We

optimized the LAMP assay reaction conditions using a commercial Bst 2.0 DNA

polymerase and tested the incubation time required for optimal amplification of the

target DNA. The CRISPR-Cas12a assay was used as the readout. For kDNA, the LAMP

reaction was evaluated at 4 different incubation times (10, 20, 30, and 40 min) (S3A-D

Fig). While 10 min were sufficient to amplify the target DNA sequence present at a

high quantity $(2 \times 10^2$ GE) (S3A Fig), at least 20 min were required for the

amplification of low abundant target DNA $(2 \times 10^{-1}$ GE) (S3B Fig). On the other hand,

an incubation time of 40 min enabled the amplification of a very low quantity of the

target DNA $(2 \times 10^{-3}$ GE) but also resulted in a positive signal of the NTC reaction

(S3D Fig), the latter indicative of contamination. These results were corroborated

through visualization of the LAMP reaction products by agarose gel electrophoresis

(S3F Fig). We thus chose an incubation time of 30 min as the most optimal for the

kDNA LAMP reaction (S3C and S3E Fig) for further experiments. For the 18S rDNA

gene, we tested the LAMP reaction at 5 different incubation times (10, 30, 40, 50, and

60 min) (S3G-K Fig). At least 40 min were required to amplify the target DNA present

at varying levels of abundance $(2 \times 10^2$ and 2×10^{-1} GE) (S3I Fig). Extending the

incubation time of the LAMP reaction to 60 min allowed the amplification of very low-

abundant target DNA $(2 \times 10^{-3}$ GE) but also resulted in NTC contamination (1 out of 2

replicates of the NTC reaction showed a positive signal) (S3K Fig). CRISPR-Cas12a-

based detection of LAMP amplicons could effectively discriminate between specific

amplification products and non-specific amplification. The latter was observed, for

instance, at the incubation times of 10 min and 40 min in the NTC reaction by agarose

gel electrophoresis (S3M Fig) but did not result in false positives in the Cas12a assay

(S3G and S3I Fig). On the basis of these results, we selected an incubation time of 50

min for the 18S LAMP reaction (S3J and S3L Fig) for further experiments.

Fig 2. Schematic representation of the LAMP reaction, highlighting the location of

the crRNA target sequence in the target DNA. (A-E) Steps in the LAMP reaction to

amplify the target DNA. The eight annealing locations of the six LAMP primers are

shown. The primers were designed to target unique sequences within a conserved region

of *Leishmania* kDNA minicircles (S1 Fig) or nuclear 18S rDNA (S2 Fig). The crRNA

target sequence and protospacer adjacent motif (PAM) site in the target DNA are

shown. For simplicity, the locations of both crRNA target sites are depicted in the same

graph (actually, the LAMP reaction is performed separately for each target). Figure

adapted from [14, 67] and created with BioRender.com.

The CRISPR-Cas12a assay is based on purified recombinant LbCas12a [63]. For

the fluorescence-based assay, we used optimized conditions as described previously

[45, 61, 62] and a quenched fluorescent ssDNA reporter [24]. We optimized here the

performance of the Cas12a-based LFA readout by testing varying concentrations of the

biotin-FAM reporter probe as well as the influence of the assay buffer, incubation

temperature and reaction time of the Cas12a assay, and PEG-8000 addition. The

optimal conditions identified involved: using the reporter probe at a final concentration

(1X) of 100 nM in our in-house Cas12a reaction buffer 1X (which performed

comparable to the HybriDetect universal assay buffer); incubating the Cas12a reaction

at 37°C for 60 min; and adding 2% PEG-8000 to the completed Cas12a reaction (S4

Fig).

Analytical validation of LAMP-CRISPR assays

We then assessed the analytical sensitivity of each primer set-crRNA combination using

serial dilutions of *L. braziliensis* M2904 gDNA extracted from a promastigote culture.

The kDNA LAMP-CRISPR assay displayed positive *L. braziliensis* detection down to 2

$\times$ 10^{-1} GE per reaction, both with a fluorescence readout (Fig 3A) and a LFA readout

(Fig 3B). Visualization of the kDNA LAMP products on an agarose gel confirmed the

presence of ladder-like bands indicative of a successful LAMP reaction (Fig 3C),

consistent with the Cas12a-based detected amplicon samples (Fig 3A and 3B). We

found a similar analytical sensitivity for the 18S LAMP-CRISPR assay, as detection of

at least 2×10^{-1} GE/reaction was achieved with a fluorescence-based readout (Fig 3D)

and a LFA readout (Fig 3E). The analysis of the 18S LAMP products by agarose gel

electrophoresis (Fig 3F) was consistent with the Cas12a assay results (Fig 3D and 3E).

The no-template control (NTC) tested in parallel in each LAMP-CRISPR assay showed

a background signal in the Cas12a fluorescence-based assay (Fig 3A and 3D, see the

two lowest GE input amounts whose fluorescence ratio is 1 given that the denominator

is the NTC signal). Likewise, the NTC did not exhibit a band (or the band observed was

faint) at the test line in the lateral flow strips (Fig 3B and 3E). This ruled out the

possibility of cross-contamination and non-specific amplification.

Fig 3. Analytical sensitivity testing of kDNA and 18S LAMP-CRISPR assays. Serial

dilutions of *L. braziliensis* M2904 gDNA encompassing 2×10^{4} to 2×10^{-3} genome

equivalents (GE) per reaction were subjected to LAMP amplification, followed by

Cas12a-based reactions for the kDNA (A-C) or 18S rDNA (D-F) target. (A, D)

Normalized fluorescence signals (fluorescence ratio) from Cas12a reactions (i.e.,

fluorescence signals taken at $t = 25$ min in the test sample relative to the NTC).

Fluorescence measurements were taken on the Synergy H1 plate reader. Data are

represented as mean ± SD (n = 3 independent amplification and detection runs). The X-

axis has a logarithmic scale. A fluorescence ratio ≥ 2 is considered detected. (B, E)

Cas12a reactions were performed using a biotin-FAM reporter molecule and visualized

on lateral flow strips. Uncleaved reporter molecules are captured at the first detection

line (control line), whereas collateral Cas12a cleavage activity generates a signal at the

second detection line (test line). The arrow indicates the direction of flow. Two

independent amplification and detection runs (r1 and r2) were analyzed. (C, F) LAMP

reaction products (5 µL) visualized by 2% agarose gel electrophoresis ran at 100 V for 1

h using SYBR Gold staining. L, GeneRuler 100 bp (C) or 100 bp Plus (F) DNA ladder.

NC, negative control (NTC, no-template control). Since space was tight on the agarose

gel images, we had to abbreviate the labels for each lane. For instance, "10^4" actually

represents "2×10^4" as indicated in the fluorescence (A, D) and LFA (B, E) panels.

Assay specificity testing (S5 Fig) was performed using gDNA samples isolated

from cultured laboratory reference strains of *Leishmania* spp., phylogenetically close

pathogens, and pathogens that cause similar lesions to CL or can coinfect the same host

(Table 1). The kDNA LAMP-CRISPR assay consistently detected representative strains

of the *Leishmania* (*Viannia*) species present in our sample set (10 of 10; Fig 4A). This

was expected given that our designed kDNA-crRNA targets a region of kDNA

minicircles conserved within the *L.* (*Viannia*) subgenus ([45] and S1 Fig). Additionally,

the assay detected 2 out of 6 tested strains belonging to the *L.* (*Leishmania*) subgenus

(Fig 4A). The 18S LAMP-CRISPR assay showed pan-*Leishmania* detection ability (Fig

4A). In both kDNA and 18S LAMP-CRISPR assays, no cross-reactivity was observed

for related trypanosomatid (*T. cruzi*) strains and other protozoan (*P. falciparum*) and

bacterial (*M. tuberculosis*) pathogens tested (Fig 4A). Also, the negative control with

human gDNA as template was not detected (Fig 4A). The analytical specificity was also

assessed with an LFA readout, where strips were interpreted by visual inspection on the

basis of the presence or absence of test bands. Three *L.* (*Viannia*) strains used as

positive controls were detected (i.e., showing a strong T-line and a very weak or

missing C-line) by both LAMP-CRISPR assays (Fig 4B, strips # 3-5). The negative

controls (human gDNA and NTC) had a negative test result (Fig 4B, strips # 1 and 2).

There was no cross-reactivity to other microbial pathogens tested (Fig 4B, strips # 6-10,

which show a strong C-line and a faint (or absent) T-line, similar to the negative

controls).

Fig 4. Analytical specificity testing of kDNA and 18S LAMP-CRISPR assays. The

specificity of the kDNA and 18S LAMP-CRISPR assays was evaluated using gDNA

from reference strains of *Leishmania* spp. and other microbial pathogens (see Table 1

and S5 Fig). (A) Heat maps depicting LAMP-CRISPR fluorescence-based detection

data from tested reference strain samples. The color scale represents normalized

fluorescence signals from Cas12a reactions (i.e., fluorescence signals taken at $t = 25$

min in the test sample relative to the NTC) for *Leishmania* kDNA and 18S rDNA

targets. The fluorescence ratio values from 2 independent assay repeats were entered

into GraphPad; the mean of these replicates is displayed. A fluorescence ratio ≥ 2 is

considered detected. (B) Cas12a reactions were performed using a biotin-FAM reporter

molecule and visualized on lateral flow strips. Uncleaved reporter molecules are

captured at the first detection line (control line), whereas collateral Cas12a cleavage

activity generates a signal at the second detection line (test line). The arrow indicates

the direction of flow. Top panel: kDNA LAMP-CRISPR assay; bottom panel: 18S

LAMP-CRISPR assay. Negative controls: HC, human control (gDNA from PBMCs of a

healthy donor) (strip # 1); NC, no-template control (NTC) (strip # 2). Positive controls:

Lbra, *L. braziliensis* M2904 (strip # 3); Lguy, *L. guyanensis* M4147 (strip # 4); Lper, *L.*

peruviana LCA08 (strip # 5). Non-related microbes: Mtb, *M. tuberculosis* PA155-18C

(strip # 6); Pf, *P. falciparum* HB3 (strip # 7). *T. cruzi* (Tc) strains: Tulahuen, Sylvio

X10, and VT-1 (strips # 8-10).

Performance evaluation of LAMP-CRISPR assays with clinical samples

Positive patient samples exhibited robust fluorescence curves in the Cas12a assay

indicating the presence of *Leishmania* kDNA (S6A Fig) or 18S rDNA (S6B Fig)

molecules. These results were consistent with the respective positive LAMP reactions

as visualized on an agarose gel (S6C and S6D Fig). We noted that the analysis of

LAMP products by agarose gel electrophoresis can sometimes give ambiguous (or

difficult to interpret) results. For instance, ladder-like bands were observed for samples

S72 and S09 amplified by the 18S LAMP assay (S6D Fig), although the band pattern

was different from the one observed for true positive reactions. This was clearly

resolved by the Cas12a-based assay on those amplicons, which confirmed that S72 and

S09 were negative for *Leishmania* 18S rDNA molecules (S6B Fig).

We next evaluated the performance of our LAMP-CRISPR fluorescence-based

detection assays using gDNA from 90 clinical samples of skin lesions. An in-house

qPCR assay [64] targeting a kDNA minicircle region conserved among *Leishmania*

(*Viannia*) species [68] served as the reference test. Of the 90 samples, 55 were

Leishmania-positive and 35 were negative by qPCR.

To determine the cutoff value for discriminating between *Leishmania*-positive

and –negative samples by the LAMP-CRISPR assays, we used the mean of the

fluorescence ratio of negative samples + 3SD. To this end, an independent panel of

negative samples was used: n = 19 for the kDNA assay and n = 10 for the 18S assay (S1

Dataset). A higher number of negative samples was needed for the former to capture the

inter-assay variation in raw fluorescence signal from the Cas12a reaction that was

evident over different periods of sample testing (S1 Dataset). The determined cutoff

value was 1.46 for the kDNA assay and 1.30 for the 18S assay (S1 Dataset).

The kDNA LAMP-CRISPR assay detected 50 (90.9%) of 55 qPCR-positive

samples (Fig 5 and 6A, Table 2, and S1 Dataset). Of these, the 5 samples not detected

by kDNA LAMP-CRISPR (coded 3, 5, 6, 28 and 71) harbored low parasite loads ($\leq$20

parasites/10^6 human cells) (Fig 6B and S1 Dataset), corresponding to 10^{-2} to 10^{-3}

parasite GE/reaction (S1 Dataset, see "Parasite equivalents" data in the worksheet

"qPCR_CS"). The 18S LAMP-CRISPR assay detected 40 (72.7%) of 55 qPCR-positive

samples (Fig 5 and 6C, Table 2, and S1 Dataset). Of these, the 15 samples not detected

by 18S LAMP-CRISPR contained a parasite load of $\leq$150 parasites/10^6 human cells

(Fig 6D and S1 Dataset), which corresponded to 10^{-1} to 10^{-3} parasite GE/reaction (S1

Dataset, see "Parasite equivalents" data in the worksheet "qPCR_CS"). Both LAMP-

CRISPR assays achieved 100% specificity (Table 2). The complete dataset of this study

is presented in the S1 Dataset.

A ROC curve analysis was used to assess the performance of each assay in

distinguishing *Leishmania*-positive and -negative clinical samples, with AUC values of

0.973 for the kDNA LAMP-CRISPR assay (Fig 7A) and 0.876 for the 18S LAMP-CRISPR assay (Fig 7B).

Fig 5. Evaluation of kDNA and 18S LAMP-CRISPR assay performance on clinical samples. Heat maps showing LAMP-CRISPR fluorescence-based detection data from tested clinical samples (n = 90). The color scale represents normalized fluorescence signals from Cas12a reactions (i.e., fluorescence signals taken at t = 25 min in the test sample relative to the NTC) for *Leishmania* kDNA and 18S rDNA targets. Samples are coded as 1-90 ('Code_UPC' as indicated in the S1 Dataset). PC, positive control (*L. braziliensis* M2904 gDNA, $10^3 – 10^4$ genome equivalents); NC, no-template control (NTC) of the LAMP reaction. The cutoff value was 1.46 for the kDNA assay and 1.30 for the 18S assay (mean of the fluorescence ratio of negative clinical samples used to calculate cutoff values + 3SD; see the S1 Dataset). The Cq values corresponding to the kDNA qPCR assay (reference test) are shown. For Cq values, the letter "U" denotes undetermined; X, not applicable.

Fig 6. Evaluation of kDNA and 18S LAMP-CRISPR assay performance on clinical samples: analysis across the range of parasite load levels. (A, C) Scatter plots showing normalized fluorescence signals from Cas12a reactions (i.e., fluorescence signals taken at t = 25 min in the test sample relative to the NTC) for *Leishmania* kDNA (A) and 18S rDNA (C) targets in clinical samples (n = 90; data shown in Fig 5 and S1 Dataset) versus Cq values determined by kDNA qPCR. (B, D) Scatter plots showing normalized fluorescence signals from Cas12a reactions for *Leishmania* kDNA (B) and

667 18S rDNA (D) targets in clinical samples versus the estimated parasite load in

668 quantifiable samples (n = 55). The parasite load is expressed as the number of

669 *Leishmania* parasites/10^6 human cells. The horizontal dotted lines represent the LAMP-

670 CRISPR assay cutoff value (mean of the fluorescence ratio of negative clinical samples

671 + 3SD; see the S1 Dataset): 1.46 for the kDNA assay and 1.30 for the 18S assay.

677 **Table 2. Diagnostic performance of LAMP-CRISPR assays with fluorescence**

678 **readout for *Leishmania* DNA detection compared with the kDNA qPCR assay**

Test	kDNA qPCR		% (95% CI)	
	Positive (n = 55)	Negative (n = 35)	Sensitivity	Specificity
kDNA LAMP-CRISPR			90.91 (80.05-96.98)	100.00 (90.00-100.00)
Positive	50	0		
Negative	5	35		
18S LAMP-CRISPR			72.73 (59.04-83.86)	100.00 (90.00-100.00)
Positive	40	0		
Negative	15	35		

Results of the statistical analysis of diagnostic test performance conducted with the MedCalc software

version 23.0.1. (https://www.medcalc.org/calc/diagnostic_test.php)

Fig 7. Evaluation of kDNA and 18S LAMP-CRISPR assay performance on clinical

samples: ROC curve analysis. ROC curve analysis performed using the fluorescence

ratio values of the kDNA LAMP-CRISPR assay (A) or the 18S LAMP-CRISPR assay

(B) from tested clinical samples (n = 90; data shown in Fig 5 and S1 Dataset), with the

Leishmania infection status (positive/negative) as determined by kDNA qPCR

(reference test). The area under the curve (AUC) and corresponding 95% confidence

interval (calculated using the Wilson/Brown method) are shown. The diagonal dotted

line corresponds to the null hypothesis with AUC of 0.5, in which the accuracy of a

diagnostic test is no different from random chance.

We also assessed the applicability of the LAMP-CRISPR assays with an LFA

readout on clinical samples. To this end, we performed pilot testing on gDNA extracted

from a representative group of clinical samples (n = 18) encompassing a wide range of

parasite load levels as determined by the kDNA qPCR assay. Lateral flow testing results

were consistent across two assay repeats per sample (n = 8; S7 Fig), indicating robust

assay reproducibility. Subsequently, we expanded the evaluation to include 10

additional clinical samples, thus analyzing a total of 18 specimens (Fig 8). Samples that

tested positive by the kDNA LAMP-CRISPR assay (16/18 = 88.9%) included the 3

parasite load categories: low (4 of 6 tested), intermediate (6 of 6), and high (6 of 6) (Fig

8). These LFA results were 100% concordant with the fluorescence-based readout (Fig

8 and S1 Dataset). The 18S LAMP-CRISPR assay with LFA readout detected

Leishmania DNA in 72.2% (13/18) of the examined samples, which had parasite load

levels defined as: low (2 of 6 tested), intermediate (5 of 6), and high (6 of 6) (Fig 8).

One sample (coded 12) out of the 18 tested samples (5.6%) by the 18S assay showed

discordant results between LFA and fluorescence-based readouts (Fig 8), with this

associated with a low parasite load (56.6 parasites/10^6 human cells) (S1 Dataset). The

two examined DNA samples from non-leishmaniasis patients (i.e., kDNA qPCR-negative; samples 09 and C7) tested negative by both LAMP-CRISPR assays (Fig 8 and S7 Fig).

Fig 8. Pilot testing of kDNA and 18S LAMP-CRISPR assays on clinical samples using a lateral flow readout. A representative subset of clinical samples with varying parasite load levels was tested with kDNA and 18S LAMP-CRISPR assays using a LFA readout. Visualization on the lateral flow strips (Milenia GenLine HybriDetect strips) was achieved through the cleavage of the biotin-FAM reporter molecule. C, control line; T, test line. The results of the fluorescence-based readout of the LAMP-CRISPR assays (+, detected; -, not detected, on the basis of the estimated cutoff; see S1 Dataset) as well as the Cq values of the kDNA qPCR are indicated below each strip. Controls included DNA samples from two non-leishmaniasis patients (i.e., kDNA qPCR-negative; samples 09 and C7), a positive control (PC, 2×10^2 genome equivalents from *L. braziliensis* M2904 gDNA), a human negative control (HC, 20 ng of human PBMC gDNA), and a no-template control (NTC) of the LAMP reaction.

Discussion

The pressing need for new diagnostic tools has been identified as a priority in the World Health Organization (WHO) 2030 road map to support the control and elimination of neglected tropical diseases [69]. Among them, leishmaniasis belongs to the category of the most neglected ones as it entails a group of diseases with a large negative impact on people living in endemic rural areas, who are often economically disadvantaged and do not have access to an early, correct disease diagnosis that would enable timely adequate

treatment. PCR-based molecular diagnostics provide the most sensitive and specific tools for *Leishmania* detection and are pivotal for early disease diagnosis, discrimination of *Leishmania* species, detection of asymptomatic infections, and epidemiological surveillance [70]. Besides their use in centralized laboratories, the progress in the automation and simplification of the workflow from sample preparation to target amplification and detection (result readout) steps has made molecular diagnostics more suitable and accessible for resource-limited settings in the form of near-PoC tests [71]. Examples of promising formats include the *Leishmania* OligoC-TesT (a PCR-oligochromatographic test) [72, 73] and a PCR-based method applied directly on EDTA blood for the diagnosis of VL, where the PCR is run on a portable miniPCR device and the labeled PCR products are detected with a lateral flow immunoassay [74]. Other recent advances have been facilitated by the incorporation of isothermal NAATs, such as LAMP and RPA, combined with lateral flow detection or other portable devices [75-79]. Among the most promising approaches is the integration of isothermal NAATs with CRISPR technology, which has fostered the development of a new generation of molecular diagnostics with the potential to advance toward near-PoC applications in low-resource settings [80].

In the present study, we aimed to contribute with the development of LAMP-coupled CRISPR-Cas12a assays for molecular detection of *Leishmania* spp. We designed new LAMP primer sets that flank previously reported crRNA target sites [45] within two commonly used multicopy regions of the *Leishmania* genome: the 18S rDNA and kDNA minicircles. We harnessed the CRISPR-Cas12a system as the detection method to avoid possible false positive results caused by non-specific amplification in the LAMP reaction and thereby improve the assay's specificity. The assay process can be completed within 1.5 h, with results readable via either of the two

optimized readouts: a fluorescence-based readout (suitable for laboratories with

minimal molecular biology equipment such as a blue light transilluminator or a

handheld fluorimeter) and a visual readout on lateral flow strips (to simplify the use of

these assays). We report on the analytical validation and initial performance evaluation

of our assays in clinical samples under controlled conditions in a research laboratory

setting. With these achievements, compared to our previous proof-of-concept PCR-

CRISPR assays [45], we aim to pave the way towards making CRISPR-based

diagnostics accessible in a testing kit format for the diagnosis of leishmaniasis in rural

endemic areas.

Here, in keeping with our *in silico* primer and crRNA design within conserved

target genomic regions, the analytical validation demonstrated the ability of the 18S

LAMP-CRISPR assay to achieve pan-*Leishmania* detection, whereas the kDNA

LAMP-CRISPR assay reliably detected species belonging to the *Leishmania* (*Viannia*)

subgenus. The kDNA assay also detected certain strains from the analyzed species of

the *L.* (*Leishmania*) subgenus, specifically *L. mexicana* from Central America and *L.*

donovani from the Old World. These results reflect both the similarity and variations in

mitochondrial kDNA minicircle sequences between the two subgenera (see S1 Fig).

Furthermore, our assays showed no cross-reactivity against other pathogens tested,

which co-circulate in the same geographic regions and can thus coinfect the same

human host, such as *T. cruzi* (a closely related trypanosomatid), *P. falciparum*, and *M.*

tuberculosis, or cause cutaneous lesions similar to CL (*M. tuberculosis*) [81]. Our

LAMP-CRISPR assays achieved high analytical sensitivity on serially diluted DNA

from the *L. braziliensis* M2904 strain, with a detection level equivalent to 0.2 parasites

per reaction. This detection sensitivity (below one parasite genome equivalent)

compares well with that of the reference kDNA qPCR assay (5×10^{-3} genome

equivalents per reaction; [64]), which is clinically relevant.

To explore the performance of our LAMP-CRISPR assays to detect *Leishmania*

DNA in clinical samples, we evaluated a blinded panel of human skin lesion samples

(n= 90) from patients enrolled in the Cusco region of Peru, where CL is caused by *L.*

(*Viannia*) species. Both LAMP-CRISPR assays with a fluorescence readout attained

good diagnostic performance, with sensitivity rates of 90.9% for kDNA and 72.7% for

18S rDNA, and a specificity of 100%, compared to the reference kDNA qPCR assay.

Discrepancies between the two methods corresponded to samples with low parasite

loads (S1 Dataset). Overall, LAMP-CRISPR assay results for the subgroup of samples

tested with both fluorescence-based and lateral flow strip readouts were concordant. The

18S LAMP-CRISPR assay showed a somewhat reduced sensitivity when applied to

clinical samples. This is likely related to the differences in target copy number:

Leishmania parasites present an estimated 10 to 170 copies of the nuclear 18S rDNA

region per genome [82, 83], while they harbor about 10,000 copies of

extrachromosomal kDNA minicircles per cell [84].

Several studies have developed prototype LAMP assays for detecting

Leishmania infections in diagnosing human VL, CL, and post-kala-azar dermal

leishmaniasis (PKDL), as well as in domestic animal reservoirs (dogs) and sand flies.

They target multicopy genomic regions, such as kDNA minicircles [85-87], the 18S

rRNA gene [88-90], the ITS1 region in the rDNA array [91], and cysteine proteinase B

(*cpb*) genes [92]. These assays proved to be highly sensitive for the detection of VL-

and CL-causing *Leishmania* species (at the genus, species complex, or species level,

depending on the target used), underscoring their potential value as diagnostic and

surveillance tools in endemic settings. With the genus-specific targets, because of their

high sequence conservation among the Trypanosomatidae, some nonspecific results

were observed. For instance, the study of Adams *et al.* [88] reported on a LAMP assay

based on 18S rDNA that was unable to discriminate between infections with

Leishmania, *T. brucei*, and *T. cruzi*. Since the diseases caused by these related

pathogens differ in clinical presentation, their differential diagnosis can be made on

clinical symptoms and confirmed using a species-specific LAMP assay designed for this

purpose [88]. In this study, when we tested our 18S LAMP primer set - crRNA

combination, no cross-reactivity was observed with DNA from 3 *T. cruzi* strains

belonging to different genetic lineages known as discrete typing units (DTUs), namely

DTU TcI (Sylvio X10 strain), DTU TcIV (VT-1 strain), and DTU TcVI (Tulahuen

LacZ clone C4 strain). Our results demonstrate the high analytical specificity of the

LAMP-CRISPR assay towards *Leishmania* spp. We also noted that integrating LAMP

amplification with downstream CRISPR-based detection improved discrimination

between positive and negative amplification results, as shown in previous studies [37,

56]. This outcome is due to the LAMP-coupled CRISPR assays offering two layers of

specificity: the first one enabled by using 4 to 6 LAMP primers that selectively target 6

to 8 distinct regions in the target DNA, and the second one due to the crRNA-guided

Cas12a-mediated target sequence recognition [23]. An optimization to improve the

LAMP assay specificity involves incorporating a probe-specific detection readout [18],

as illustrated in the study by Ruang-Areerate *et al.* [93] aimed at detecting

asymptomatic *Leishmania* infections in HIV-infected patients.

The LAMP technology is commercially available and delivered, such as the

Leishmania detection kit (Loopamp™ *Leishmania* Detection Kit; Eiken Chemical Co.,

Ltd., Japan), which does not require cold-chain storage. This kit targets kDNA

minicircles and 18S rDNA in the same reaction to enable pan-*Leishmania* detection

based on the LAMP assay developed by Adams *et al*. [94]. The diagnostic performance

evaluation of this kit in different endemic areas around the world has demonstrated its

high sensitivity and specificity for the diagnosis of VL [22, 76, 94-96], CL [22, 94, 97,

98], and PKDL [99], underscoring its potential for implementation in primary

healthcare facilities. Undoubtedly, the DNA extraction method can impact LAMP assay

performance when testing clinical samples. As shown in the study by Ghosh *et al*. [99],

a lower efficiency of LAMP to diagnose PKDL was observed when the LAMP reaction

was performed using crude DNA obtained by a simple in-house boil and spin protocol

compared to DNA extracted by a commercial silica column-based (Qiagen) kit (67.2%

vs. 89.7% sensitivity, respectively). Further, a recent study found a compromised

sensitivity (48.4%) of the *Leishmania* LAMP kit for the diagnosis of CL in Ethiopia,

possibly due to a primer mismatch with the *L. aethiopica* 18S rDNA target [100]. This

finding points out the need to extensively validate existing LAMP assays in diverse

clinical contexts and epidemiological scenarios in different geographical areas to

confirm their effectiveness in detecting local parasite variants that are circulating. The

same applies to our new LAMP-CRISPR assays.

Combining highly sensitive isothermal amplification with the notable highly

specific CRISPR-Cas12a system is an approach that has been applied for detecting

pathogenic trypanosomatids. For example, Wiggins *et al*. [46] developed proof-of-

concept RPA-coupled CRISPR-Cas12a assays that target kDNA maxicircles for

Leishmania spp. detection. They have recently adapted their assays into a

multichambered device to obtain a rapid nucleic acid diagnostic tool that does not

require electricity or instrumentation, a promising development for diagnostic

applications in low-resource settings [101]. In another study, Yang *et al*. [47] developed

a one-tube RPA-CRISPR/Cas12a assay that targets the *Leishmania Kmp11* gene. The

assay process can be completed within 50 min, with the fluorescent signal of reaction

tubes observed visually under blue light illumination. Their assay has been evaluated on

simulated clinical samples, yielding a sensitivity of 97.9% and a specificity of 100%.

Ortiz-Rodríguez *et al*. [50] developed a platform for *T. cruzi* detection, which was

validated on samples from reservoirs and insect vectors. It employs PCR or RPA

followed by CRISPR-Cas12a-based detection. Of 3 nuclear target genes evaluated,

namely cytochrome B (*Cytb*), the 18S rRNA gene, and histone H2A, only the former

(*Cytb*) resulted in a workable CRISPR-based assay since no mutations were found in the

chosen target site during testing of *T. cruzi* DNA extracted from the intestine of

triatomine vectors. Furthermore, the authors designed a portable prototype device

(called TropD-Detector) for fluorescence results discrimination, suitable for low-

resource settings. Guarnizo *et al*. [56] designed a LAMP-CRISPR/Cas12a assay that

targets a highly conserved region in the nuclear *HSP70* gene to specifically detect *T.*

cruzi in the context of acute Chagas disease diagnosis in endemic countries. They

evaluated the assay in clinical samples from 100 infants born to Chagas-positive

mothers in Bolivia, achieving a sensitivity of 77.27% and a specificity of 100%

compared to a TaqMan-based qPCR assay that targets satellite DNA.

Our study shows that LAMP-coupled CRISPR-Cas12a assays are an effective

alternative method for detecting *Leishmania* infections in different non-invasive and

invasive specimens (S1 Dataset) from human skin lesions. While kDNA LAMP primers

previously reported were for a *L. donovani* complex-specific assay [94] or for a *L.*

amazonensis-specific assay [87], our kDNA LAMP-CRISPR assay (specifically the

crRNA design) exploited a kDNA minicircle region conserved in the New World

species of the *Leishmania* (*Viannia*) subgenus ([45] and S1 Fig). Our pan-*Leishmania*

18S LAMP-CRISPR assay is expected to be capable of detecting various parasite

species from both the New and Old World (S2 Fig). Our findings encourage further

evaluation and validation of both assays in different geographical areas. Besides, our

assays have the potential to be adapted in future work to address the

ASSURED/REASSURED criteria for a diagnostic test suitable for adoption across

different healthcare levels, which require assays to be affordable, sensitive, specific,

user-friendly, robust and rapid, as well as equipment-free (or operating with minimal

equipment), deliverable to those who need them, with real-time connectivity, and

workable on easily collected specimens [102, 103]. Sample preparation is one of the

challenges in molecular assay implementation in resource-limited settings [104].

Further research is needed to optimize rapid DNA extraction protocols, such as boil-

spin [99] and SpeedXtract (a magnetic bead-based method) [105], combined with a

streamlined CRISPR assay workflow to move forward to a near-PoC nucleic acid-based

diagnostic assay for CL and other forms of leishmaniasis. Both LAMP and CRISPR

reactions are compatible with freeze-drying (lyophilization), which improves reagent

stability for easy transport and storage at room temperature and may reduce

contamination risks [106]. This improvement can also facilitate the preparation of the

assays in a pre-designed kit format, enabling a portable assay setup for resource-limited

settings. Therefore, the impact of freeze-drying reagents on the performance of our

assays needs to be evaluated in future work. Incorporating these refinements into our

assay procedure, along with user-friendly devices, such as the LFA strips or a handheld

fluorimeter, would allow its adoption in minimally-equipped laboratories in endemic

regions, and may be incorporated in a mobile suitcase laboratory. The latter approach

has been explored with a RPA assay in field studies conducted in Asia, showing

promising performance for VL, PKDL, and CL diagnosis in endemic settings [105, 107,

108]. After further optimization and validation, we envisage that our LAMP-CRISPR

assays may become valuable tools for providing rapid testing results directly in the affected regions, thereby mitigating delays associated with testing in centralized laboratories. CRISPR-based assays could also have a considerable role in epidemiological studies, in providing real-time surveillance data. Further, an approach for future work that offers the potential to be impactful in resource-limited endemic settings entails the integration of our assays with recent developments in lab-on-chip systems, such as a paper-based LAMP-CRISPR integrated diagnostic platform (PLACID) [37]. This paper-based device uses freeze-dried LAMP and CRISPR reagents, where nucleic acid amplification and CRISPR reactions are performed successively without fluid transfer using paper microfluidics. A smartphone-based protocol allows the fluorescent detection of nucleic acid targets. This platform showed robust performance for bacterial and viral pathogen target detection, with no false positives observed [37]. These are well-known refinements to consider in future assay development efforts. Indeed, we noticed that our current two-step LAMP-CRISPR open assay format intrinsically carries the risk of cross-contamination during sample transfer from tube-based (LAMP) to plate-based (CRISPR) reactions, as reported in other studies [106].

In conclusion, we developed novel LAMP-CRISPR assays for the molecular detection of *Leishmania*, targeting two multicopy genomic regions: 18S rDNA for pan-*Leishmania* detection and a kDNA minicircle region conserved among New World *L.* (*Viannia*) species. By integrating LAMP with CRISPR-Cas12a technology, the assays enable accurate detection, with results readout via fluorescence or lateral flow strips, the latter aimed at facilitating a simpler workflow. Analytical validation and initial performance testing on clinical samples demonstrated high sensitivity and specificity, underscoring the assays' potential as effective alternatives for detecting *Leishmania*

infections. With further optimization and validation, these prototype assays hold promise as next-generation molecular tools with great potential to improve the diagnosis and surveillance of leishmaniasis in endemic regions, thereby supporting One Health strategies for disease control.

References

1. World Health Organization. Leishmaniasis. World Health Organization: WHO. 2023. Available from: https://www.who.int/news-room/fact-sheets/detail/leishmaniasis

2. Pan American Health Organization. Leishmaniasis. Pan American Health Organization: PAHO. 2017. Available from: https://www3.paho.org/hq/dmdocuments/2017/2017-cha-leishmaniasis-factsheet-work.pdf

3. Alvar J, Vélez ID, Bern C, Herrero M, Desjeux P, Cano J, et al. Leishmaniasis worldwide and global estimates of its incidence. PLoS One. 2012;7(5):e35671.

4. Burza S, Croft SL, Boelaert M. Leishmaniasis. Lancet. 2018 Sep 15;392(10151):951-970.

5. Pan American Health Organization. Leishmaniasis: Epidemiological Report for the Americas. No.12. 2023 Dec 21. Available from: https://iris.paho.org/handle/10665.2/59155

6. Scott P, Novais FO. Cutaneous leishmaniasis: immune responses in protection and pathogenesis. Nat Rev Immunol. 2016 Sep;16(9):581-92.

7. Kaye PM, Cruz I, Picado A, Van Bocxlaer K, Croft SL. Leishmaniasis immunopathology-impact on design and use of vaccines, diagnostics and drugs. Semin Immunopathol. 2020 Jun;42(3):247-264.

8. Reimão JQ, Coser EM, Lee MR, Coelho AC. Laboratory Diagnosis of
Cutaneous and Visceral Leishmaniasis: Current and Future Methods.
Microorganisms. 2020 Oct 22;8(11):1632.

9. Van der Auwera G, Dujardin JC. Species typing in dermal leishmaniasis. Clin
Microbiol Rev. 2015 Apr;28(2):265-94.

10. Akhoundi M, Downing T, Votýpka J, Kuhls K, Lukeš J, Cannet A, et al.
Leishmania infections: Molecular targets and diagnosis. Mol Aspects Med. 2017
Oct;57:1-29.

11. Mesa LE, Manrique R, Muskus C, Robledo SM. Test accuracy of polymerase
chain reaction methods against conventional diagnostic techniques for
Cutaneous Leishmaniasis (CL) in patients with clinical or epidemiological
suspicion of CL: Systematic review and meta-analysis. PLoS Negl Trop Dis.
2020 Jan 21;14(1):e0007981.

12. Bharadwaj M, Bengtson M, Golverdingen M, Waling L, Dekker C. Diagnosing
point-of-care diagnostics for neglected tropical diseases. PLoS Negl Trop Dis.
2021 Jun 17;15(6):e0009405.

13. Oliveira BB, Veigas B, Baptista PV. Isothermal Amplification of Nucleic Acids:
The Race for the Next "Gold Standard." Front. Sens. 2021 Sep 28;2:752600.

14. Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K, Amino N, et
al. Loop-mediated isothermal amplification of DNA. Nucleic Acids Res. 2000
Jun 15;28(12):E63.

15. Nagamine K, Watanabe K, Ohtsuka K, Hase T, Notomi T. Loop-mediated
isothermal amplification reaction using a nondenatured template. Clin Chem.
2001 Sep;47(9):1742-3.

16. Mori Y, Nagamine K, Tomita N, Notomi T. Detection of loop-mediated
isothermal amplification reaction by turbidity derived from magnesium
pyrophosphate formation. Biochem Biophys Res Commun. 2001 Nov
23;289(1):150-4.

17. Mori Y, Hirano T, Notomi T. Sequence specific visual detection of LAMP
reactions by addition of cationic polymers. BMC Biotechnol. 2006 Jan 10;6:3.

18. Phillips EA, Moehling TJ, Bhadra S, Ellington AD, Linnes JC. Strand
Displacement Probes Combined with Isothermal Nucleic Acid Amplification for
Instrument-Free Detection from Complex Samples. Anal Chem. 2018 Jun
5;90(11):6580-6586.

19. Atçeken N, Yigci D, Ozdalgic B, Tasoglu S. CRISPR-Cas-Integrated LAMP.
Biosensors (Basel). 2022 Nov 17;12(11):1035. doi: 10.3390/bios12111035.

20. Nzelu CO, Kato H, Peters NC. Loop-mediated isothermal amplification
(LAMP): An advanced molecular point-of-care technique for the detection of
Leishmania infection. PLoS Negl Trop Dis. 2019 Nov 7;13(11):e0007698.

21. Erber AC, Sandler PJ, de Avelar DM, Swoboda I, Cota G, Walochnik J.
Diagnosis of visceral and cutaneous leishmaniasis using loop-mediated
isothermal amplification (LAMP) protocols: a systematic review and meta-
analysis. Parasit Vectors. 2022 Jan 24;15(1):34.

22. Ibarra-Meneses AV, Chicharro C, Sánchez C, García E, Ortega S, Ndung'u JM,
et al. Loop-Mediated Isothermal Amplification Allows Rapid, Simple and
Accurate Molecular Diagnosis of Human Cutaneous and Visceral Leishmaniasis
Caused by *Leishmania infantum* When Compared to PCR. Microorganisms.
2021 Mar 16;9(3):610.

23. Kaminski MM, Abudayyeh OO, Gootenberg JS, Zhang F, Collins JJ. CRISPR-

based diagnostics. Nat Biomed Eng. 2021 Jul;5(7):643-656.

24. Chen JS, Ma E, Harrington LB, Da Costa M, Tian X, Palefsky JM, et al.

CRISPR-Cas12a target binding unleashes indiscriminate single-stranded DNase

activity. Science. 2018 Apr 27;360(6387):436-439.

25. Gootenberg JS, Abudayyeh OO, Lee JW, Essletzbichler P, Dy AJ, Joung J, et al.

Nucleic acid detection with CRISPR-Cas13a/C2c2. Science. 2017 Apr

28;356(6336):438-442.

26. Broughton JP, Deng X, Yu G, Fasching CL, Servellita V, Singh J, et al.

CRISPR-Cas12-based detection of SARS-CoV-2. Nat Biotechnol. 2020

Jul;38(7):870-874.

27. Wang X, Ji P, Fan H, Dang L, Wan W, Liu S, et al. CRISPR/Cas12a technology

combined with immunochromatographic strips for portable detection of African

swine fever virus. Commun Biol. 2020 Feb 11;3(1):62.

28. Li SY, Cheng QX, Wang JM, Li XY, Zhang ZL, Gao S, et al. CRISPR-Cas12a-

assisted nucleic acid detection. Cell Discov. 2018 Apr 24;4:20.

29. Cunningham CH, Hennelly CM, Lin JT, Ubalee R, Boyce RM, Mulogo EM, et

al. A novel CRISPR-based malaria diagnostic capable of *Plasmodium* detection,

species differentiation, and drug-resistance genotyping. EBioMedicine. 2021

Jun;68:103415.

30. Nguyen LT, Smith BM, Jain PK. Enhancement of trans-cleavage activity of

Cas12a with engineered crRNA enables amplified nucleic acid detection. Nat

Commun. 2020 Sep 30;11(1):4906.

31. Fozouni P, Son S, Díaz de León Derby M, Knott GJ, Gray CN, D'Ambrosio
MV, et al. Amplification-free detection of SARS-CoV-2 with CRISPR-Cas13a
and mobile phone microscopy. Cell. 2021 Jan 21;184(2):323-333.e9.

32. Liu TY, Knott GJ, Smock DCJ, Desmarais JJ, Son S, Bhuiya A, et al.
Accelerated RNA detection using tandem CRISPR nucleases. Nat Chem Biol.
2021 Sep;17(9):982-988.

33. Nalefski EA, Patel N, Leung PJY, Islam Z, Kooistra RM, Parikh I, et al. Kinetic
analysis of Cas12a and Cas13a RNA-Guided nucleases for development of
improved CRISPR-Based diagnostics. iScience. 2021 Aug 18;24(9):102996.

34. Wang DG, Brewster JD, Paul M, Tomasula PM. Two methods for increased
specificity and sensitivity in loop-mediated isothermal amplification. Molecules.
2015 Apr 7;20(4):6048-59.

35. Joung J, Ladha A, Saito M, Kim NG, Woolley AE, Segel M, et al. Detection of
SARS-CoV-2 with SHERLOCK One-Pot Testing. N Engl J Med. 2020 Oct
8;383(15):1492-1494.

36. Wang R, Qian C, Pang Y, Li M, Yang Y, Ma H, et al. opvCRISPR: One-pot
visual RT-LAMP-CRISPR platform for SARS-cov-2 detection. Biosens
Bioelectron. 2021 Jan 15;172:112766.

37. Sen A, Masetty M, Weerakoon S, Morris C, Yadav JS, Apewokin S, et al.
Paper-based loop-mediated isothermal amplification and CRISPR integrated
platform for on-site nucleic acid testing of pathogens. Biosens Bioelectron. 2024
Aug 1;257:116292.

38. Mukama O, Yuan T, He Z, Li Z, Habimana J de D, Hussain M, et al. A high
fidelity CRISPR/Cas12a based lateral flow biosensor for the detection of HPV16

and HPV18. Sensors and Actuators B: Chemical. 2020 Aug;316:128119. pii: S092540052030469X. doi: 10.1016/j.snb.2020.128119.

39. Wang Y, Li J, Li S, Zhu X, Wang X, Huang J, et al. LAMP-CRISPR-Cas12-based diagnostic platform for detection of *Mycobacterium tuberculosis* complex using real-time fluorescence or lateral flow test. Mikrochim Acta. 2021 Sep 20;188(10):347.

40. Qiu X, Liu X, Ma X, Wang R, Chen S, Li F, et al. One-Pot Isothermal LAMP-CRISPR-Based Assay for *Klebsiella pneumoniae* Detection. Microbiol Spectr. 2022 Aug 31;10(4):e0154522.

41. Fang Y, Liu L, Zhao W, Dong L, He L, Liu Y, et al. Rapid and Sensitive Detection of *Verticillium dahliae* from Soil Using LAMP-CRISPR/Cas12a Technology. Int J Mol Sci. 2024 May 10;25(10):5185.

42. Lee RA, Puig H, Nguyen PQ, Angenent-Mari NM, Donghia NM, McGee JP, et al. Ultrasensitive CRISPR-based diagnostic for field-applicable detection of *Plasmodium* species in symptomatic and asymptomatic malaria. Proc Natl Acad Sci U S A. 2020 Oct 13;117(41):25722-25731.

43. Wang X, Cheng M, Yang S, Xing C, Li Q, Zhu Y, et al. CRISPR/Cas12a combined with RPA for detection of *T. gondii* in mouse whole blood. Parasit Vectors. 2023 Jul 30;16(1):256.

44. Bengtson M, Bharadwaj M, Franch O, van der Torre J, Meerdink V, Schallig H, et al. CRISPR-dCas9 based DNA detection scheme for diagnostics in resource-limited settings. Nanoscale. 2022 Feb 3;14(5):1885-1895.

45. Dueñas E, Nakamoto JA, Cabrera-Sosa L, Huaihua P, Cruz M, Arévalo J, et al. Novel CRISPR-based detection of *Leishmania* species. Front Microbiol. 2022 Sep 15;13:958693.

46. Wiggins TJ, Peng R, Bushnell RV, Tobin JK, MacLeod DA, Du K, et al. Instrument-Free Point-of-Care Diagnostic for *Leishmania* Parasites. Diagnostics (Basel). 2024 Dec 5;14(23):2744.

47. Yang K, Bi M, Mo X. CRISPR/Cas12a, combined with recombinase polymerase amplification (RPA) reaction for visual detection of *Leishmania* species. Microchem J. 2024 Dec;207:112283. doi: 10.1016/j.microc.2024.112283

48. Sima N, Dujeancourt-Henry A, Perlaza BL, Ungeheuer MN, Rotureau B, Glover L. SHERLOCK4HAT: A CRISPR-based tool kit for diagnosis of Human African Trypanosomiasis. EBioMedicine. 2022 Nov;85:104308.

49. Álvarez-Rodríguez A, Li Z, Jin BK, Stijlemans B, Geldhof P, Magez S. A CRISPR-Cas-based recombinase polymerase amplification assay for ultra-sensitive detection of active *Trypanosoma brucei evansi* infections. Front Mol Biosci. 2025 Feb 14;12:1512970.

50. Ortiz-Rodríguez LA, Cabanzo R, Jaimes-Dueñez J, Mendez-Sanchez SC, Duque JE. TropD-detector a CRISPR/LbCas12a-based system for rapid screening of *Trypanosoma cruzi* in Chagas vectors and reservoirs. Sci Rep. 2025 May 31;15(1):19107.

51. Cherkaoui D, Mesquita SG, Huang D, Lugli EB, Webster BL, McKendry RA. CRISPR-assisted test for *Schistosoma haematobium*. Sci Rep. 2023 Mar 27;13(1):4990.

52. MacGregor SR, McManus DP, Sivakumaran H, Egwang TG, Adriko M, Cai P, et al. Development of CRISPR/Cas13a-based assays for the diagnosis of Schistosomiasis. EBioMedicine. 2023 Aug;94:104730.

53. Liu J, Wang X, Sheng F, Giri BR, Li S, Xia T, et al. Metagenomic sequencing for identifying pathogen-specific circulating DNAs and development of diagnostic methods for schistosomiasis. iScience. 2023 Jul 27;26(9):107495.

54. Parikesit AA, Hermantara R, Gregorius K, Siddharta E. Designing hybrid CRISPR-Cas12 and LAMP detection systems for treatment-resistant *Plasmodium falciparum* with in silico method. Narra J. 2023 Dec;3(3):e301.

55. Liang Y, Xie SC, Lv YH, He YH, Zheng XN, Cong W, et al. A novel single-tube LAMP-CRISPR/Cas12b method for rapid and visual detection of zoonotic *Toxoplasma gondii* in the environment. Infect Dis Poverty. 2024 Dec 10;13(1):94.

56. Guarnizo SAG, Basma L, Equilia S, Condori BJ, Malaga E, Defazio S, et al. A specific, stable, and accessible LAMP assay targeting the HSP70 gene of *Trypanosoma cruzi*. medRxiv [Preprint]. 2025 Jan 13:2025.01.12.25320185. doi: 10.1101/2025.01.12.25320185.

57. Kato H, Cáceres AG, Seki C, Silupu García CR, Holguín Mauricci C, Castro Martínez SC, et al. Further insight into the geographic distribution of *Leishmania* species in Peru by cytochrome b and mannose phosphate isomerase gene analyses. PLoS Negl Trop Dis. 2019 Jun 20;13(6):e0007496.

58. Sandoval-Juárez A, Minaya-Gómez G, Rojas-Palomino N, Cáceres O. Identification of *Leishmania* species in patients derived to the National Institute of Health, Peru. Rev Peru Med Exp Salud Publica. 2020 Jan-Mar;37(1):87-92.

59. Waterhouse AM, Procter JB, Martin DM, Clamp M, Barton GJ. Jalview Version 2--a multiple sequence alignment editor and analysis workbench. Bioinformatics. 2009 May 1;25(9):1189-91.

60. Sievers F, Wilm A, Dineen D, Gibson TJ, Karplus K, Li W, et al. Fast, scalable

generation of high-quality protein multiple sequence alignments using Clustal

Omega. Mol Syst Biol. 2011 Oct 11;7:539.

61. Alcántara R, Peñaranda K, Mendoza-Rojas G, Nakamoto JA, Martins-Luna J,

Del Valle-Mendoza J, et al. Unlocking SARS-CoV-2 detection in low- and

middle-income countries. Cell Rep Methods. 2021 Oct 21;1(7):100093.

62. Alcántara R, Peñaranda K, Mendoza-Rojas G, Nakamoto JA, Dueñas E, Alvarez

D, et al. UnCovid: A versatile, low-cost, and open-source protocol for SARS-

CoV-2 RNA detection. STAR Protoc. 2021 Dec 17;2(4):100878.

63. Mendoza-Rojas G, Sarabia-Vega V, Sanchez-Castro A, Tello L, Cabrera-Sosa

L, Nakamoto JA, et al. A low-cost and open-source protocol to produce key

enzymes for molecular detection assays. STAR Protoc. 2021 Nov

1;2(4):100899.

64. Jara M, Adaui V, Valencia BM, Martinez D, Alba M, Castrillon C, et al. Real-

time PCR assay for detection and quantification of *Leishmania* (*Viannia*)

organisms in skin and mucosal lesions: exploratory study of parasite load and

clinical parameters. J Clin Microbiol. 2013 Jun;51(6):1826-33.

65. Yuan CC, Miley W, Waters D. A quantification of human cells using an ERV-3

real time PCR assay. J Virol Methods. 2001 Feb;91(2):109-17.

66. MedCalc Software Ltd. Diagnostic test evaluation calculator. Available from:

https://www.medcalc.org/calc/diagnostic_test.php (Version 23.0.1; accessed

August 21, 2024)

67. Parida M, Sannarangaiah S, Dash PK, Rao PV, Morita K. Loop mediated

isothermal amplification (LAMP): a new generation of innovative gene

amplification technique; perspectives in clinical diagnosis of infectious diseases. Rev Med Virol. 2008 Nov-Dec;18(6):407-21.

68. Lopez M, Inga R, Cangalaya M, Echevarria J, Llanos-Cuentas A, Orrego C, et al. Diagnosis of *Leishmania* using the polymerase chain reaction: a simplified procedure for field work. Am J Trop Med Hyg. 1993 Sep;49(3):348-56.

69. World Health Organization. Ending the neglect to attain the sustainable development goals: a road map for neglected tropical diseases 2021–2030. World Health Organization. 2020. Available from: https://iris.who.int/handle/10665/338565.

70. Gow I, Smith NC, Stark D, Ellis J. Laboratory diagnostics for human *Leishmania* infections: a polymerase chain reaction-focussed review of detection and identification methods. Parasit Vectors. 2022 Nov 5;15(1):412.

71. Gow I, Smith N, Stark D, Ellis J. Molecular Detection of Neglected Tropical Diseases: The Case for Automated Near-Point-of-Care Diagnosis of Leishmaniasis. Am J Trop Med Hyg. 2022 Nov 30;108(1):2-6.

72. Deborggraeve S, Laurent T, Espinosa D, Van der Auwera G, Mbuchi M, Wasunna M, et al. A simplified and standardized polymerase chain reaction format for the diagnosis of leishmaniasis. J Infect Dis. 2008 Nov 15;198(10):1565-72.

73. Carson C, Quinnell RJ, Holden J, Garcez LM, Deborggraeve S, Courtenay O. Comparison of *Leishmania* OligoC-TesT PCR with conventional and real-time PCR for Diagnosis of canine *Leishmania* infection. J Clin Microbiol. 2010 Sep;48(9):3325-30.

74. van Dijk NJ, Hagos DG, Huggins DM, Carrillo E, Ajala S, Chicharro C, et al. Simplified molecular diagnosis of visceral leishmaniasis: Laboratory evaluation

of miniature direct-on-blood PCR nucleic acid lateral flow immunoassay. PLoS Negl Trop Dis. 2024 May 7;18(5):e0011637.

75. Dixit KK, Verma S, Singh OP, Singh D, Singh AP, Gupta R, et al. Validation of SYBR green I based closed tube loop mediated isothermal amplification (LAMP) assay and simplified direct-blood-lysis (DBL)-LAMP assay for diagnosis of visceral leishmaniasis (VL). PLoS Negl Trop Dis. 2018 Nov 15;12(11):e0006922.

76. Ibarra-Meneses AV, Cruz I, Chicharro C, Sánchez C, Biéler S, Broger T, et al. Evaluation of fluorimetry and direct visualization to interpret results of a loop-mediated isothermal amplification kit to detect *Leishmania* DNA. Parasit Vectors. 2018 Apr 17;11(1):250.

77. Puri M, Brar HK, Mittal N, Madan E, Srinivasan R, Rawat K, et al. Rapid diagnosis of *Leishmania* infection with a portable loopmediated isothermal amplification device. J Biosci. 2021;46(4):92.

78. Travi BL, De los Santos MB, Shelite TR, Santos RP, Rosales LA, Castellanos-Gonzalez A, et al. Diagnostic Efficacy of Recombinase-Polymerase-Amplification Coupled with Lateral Flow Strip Reading in Patients with Cutaneous Leishmaniasis from the Amazonas Rainforest of Perú. Vector Borne Zoonotic Dis. 2021 Dec;21(12):941-947.

79. Kobialka RM, Ceruti A, Roy M, Roy S, Chowdhury R, Ghosh P, et al. Portable smartphone-based molecular test for rapid detection of *Leishmania* spp. Infection. 2024 Aug;52(4):1315-1324.

80. Qian X, Xu Q, Lyon CJ, Hu TY. CRISPR for companion diagnostics in low-resource settings. Lab Chip. 2024 Oct 9;24(20):4717-4740.

81. Martínez DY, Verdonck K, Kaye PM, Adaui V, Polman K, Llanos-Cuentas A, et al. Tegumentary leishmaniasis and coinfections other than HIV. PLoS Negl Trop Dis. 2018 Mar 1;12(3):e0006125.

82. Leon W, Fouts DL, Manning J. Sequence arrangement of the 16S and 26S rRNA genes in the pathogenic haemoflagellate *Leishmania donovani*. Nucleic Acids Res. 1978 Feb;5(2):491-504.

83. Inga R, De Doncker S, Gomez J, Lopez M, Garcia R, Le Ray D, et al. Relation between variation in copy number of ribosomal RNA encoding genes and size of harbouring chromosomes in *Leishmania* of subgenus *Viannia*. Mol Biochem Parasitol. 1998 May 1;92(2):219-28.

84. Simpson L. The mitochondrial genome of kinetoplastid protozoa: genomic organization, transcription, replication, and evolution. Annu Rev Microbiol. 1987;41:363-82.

85. Takagi H, Itoh M, Islam MZ, Razzaque A, Ekram AR, Hashighuchi Y, et al. Sensitive, specific, and rapid detection of *Leishmania donovani* DNA by loop-mediated isothermal amplification. Am J Trop Med Hyg. 2009 Oct;81(4):578-82.

86. Verma S, Avishek K, Sharma V, Negi NS, Ramesh V, Salotra P. Application of loop-mediated isothermal amplification assay for the sensitive and rapid diagnosis of visceral leishmaniasis and post-kala-azar dermal leishmaniasis. Diagn Microbiol Infect Dis. 2013 Apr;75(4):390-5.

87. Celeste JLL, Caldeira RL, Pires SDF, Silveira KD, Soares RP, de Andrade HM. Development and evaluation of a loop-mediated isothermal amplification assay for rapid detection of *Leishmania amazonensis* in skin samples. Exp Parasitol. 2019 Aug;203:23-29.

88. Adams ER, Schoone GJ, Ageed AF, Safi SE, Schallig HD. Development of a

reverse transcriptase loop-mediated isothermal amplification (LAMP) assay for

the sensitive detection of *Leishmania* parasites in clinical samples. Am J Trop

Med Hyg. 2010 Apr;82(4):591-6.

89. Nzelu CO, Gomez EA, Cáceres AG, Sakurai T, Martini-Robles L, Uezato H, et

al. Development of a loop-mediated isothermal amplification method for rapid

mass-screening of sand flies for *Leishmania* infection. Acta Trop. 2014

Apr;132:1-6.

90. Nzelu CO, Cáceres AG, Guerrero-Quincho S, Tineo-Villafuerte E, Rodriquez-

Delfin L, Mimori T, et al. A rapid molecular diagnosis of cutaneous

leishmaniasis by colorimetric malachite green-loop-mediated isothermal

amplification (LAMP) combined with an FTA card as a direct sampling tool.

Acta Trop. 2016 Jan;153:116-9.

91. Abbasi I, Kirstein OD, Hailu A, Warburg A. Optimization of loop-mediated

isothermal amplification (LAMP) assays for the detection of *Leishmania* DNA

in human blood samples. Acta Trop. 2016 Oct;162:20-26.

92. Chaouch M, Mhadhbi M, Adams ER, Schoone GJ, Limam S, Gharbi Z, et al.

Development and evaluation of a loop-mediated isothermal amplification assay

for rapid detection of *Leishmania infantum* in canine leishmaniasis based on

cysteine protease B genes. Vet Parasitol. 2013 Nov 15;198(1-2):78-84.

93. Ruang-Areerate T, Sukphattanaudomchoke C, Thita T, Leelayoova S, Piyaraj P,

Mungthin M, et al. Development of loop-mediated isothermal amplification

(LAMP) assay using SYBR safe and gold-nanoparticle probe for detection of

Leishmania in HIV patients. Sci Rep. 2021 Jun 9;11(1):12152.

94. Adams ER, Schoone G, Versteeg I, Gomez MA, Diro E, Mori Y, et al. Development and Evaluation of a Novel Loop-Mediated Isothermal Amplification Assay for Diagnosis of Cutaneous and Visceral Leishmaniasis. J Clin Microbiol. 2018 Jun 25;56(7):e00386-18.

95. Mukhtar M, Ali SS, Boshara SA, Albertini A, Monnerat S, Bessell P, et al. Sensitive and less invasive confirmatory diagnosis of visceral leishmaniasis in Sudan using loop-mediated isothermal amplification (LAMP). PLoS Negl Trop Dis. 2018 Feb 14;12(2):e0006264.

96. Hossain F, Picado A, Owen SI, Ghosh P, Chowdhury R, Maruf S, et al. Evaluation of Loopamp™ *Leishmania* Detection Kit and *Leishmania* Antigen ELISA for Post-Elimination Detection and Management of Visceral Leishmaniasis in Bangladesh. Front Cell Infect Microbiol. 2021 Apr 26;11:670759.

97. Vink MMT, Nahzat SM, Rahimi H, Buhler C, Ahmadi BA, Nader M, et al. Evaluation of point-of-care tests for cutaneous leishmaniasis diagnosis in Kabul, Afghanistan. EBioMedicine. 2018 Nov;37:453-460.

98. Schallig HDFH, Hu RVP, Kent AD, van Loenen M, Menting S, Picado A, et al. Evaluation of point of care tests for the diagnosis of cutaneous leishmaniasis in Suriname. BMC Infect Dis. 2019 Jan 7;19(1):25.

99. Ghosh P, Chowdhury R, Maruf S, Picado A, Hossain F, Owen SI, et al. Gauging the skin resident *Leishmania* parasites through a loop mediated isothermal amplification (LAMP) assay in post-kala-azar dermal leishmaniasis. Sci Rep. 2022 Oct 27;12(1):18069.

1272   100.   Taye B, Melkamu R, Tajebe F, Ibarra-Meneses AV, Adane D, Atnafu S,

1273   et al. Evaluation of Loopamp *Leishmania* detection kit for the diagnosis of

1274   cutaneous leishmaniasis in Ethiopia. Parasit Vectors. 2024 Oct 15;17(1):431.

1275   101.   Peng R, Yuqing F, Wiggins TJ, Bahadori N, Rogers JS, Waitkus J, et al.

1276   Multiple-Channel Funnel Adapted Sensing Tube (MFAST) for the Simple and

1277   Duplex Detection of Parasites. ACS Sens. 2025 Apr 25;10(4):2678-2688.

1278   102.   Mabey D, Peeling RW, Ustianowski A, Perkins MD. Diagnostics for the

1279   developing world. Nat Rev Microbiol. 2004 Mar;2(3):231-40.

1280   103.   Land KJ, Boeras DI, Chen XS, Ramsay AR, Peeling RW. REASSURED

1281   diagnostics to inform disease control strategies, strengthen health systems and

1282   improve patient outcomes. Nat Microbiol. 2019 Jan;4(1):46-54.

1283   104.   Dineva MA, MahiLum-Tapay L, Lee H. Sample preparation: a challenge

1284   in the development of point-of-care nucleic acid-based assays for resource-

1285   limited settings. Analyst. 2007 Dec;132(12):1193-9.

1286   105.   Gunaratna G, Manamperi A, Böhlken-Fascher S, Wickremasinge R,

1287   Gunawardena K, Yapa B, et al. Evaluation of rapid extraction and isothermal

1288   amplification techniques for the detection of *Leishmania donovani* DNA from

1289   skin lesions of suspected cases at the point of need in Sri Lanka. Parasit Vectors.

1290   2018 Dec 22;11(1):665.

1291   106.   Nguyen LT, Rananaware SR, Pizzano BLM, Stone BT, Jain PK. Clinical

1292   validation of engineered CRISPR/Cas12a for rapid SARS-CoV-2 detection.

1293   Commun Med (Lond). 2022 Jan 12;2:7.

1294   107.   Mondal D, Ghosh P, Khan MA, Hossain F, Böhlken-Fascher S,

1295   Matlashewski G, et al. Mobile suitcase laboratory for rapid detection of

Leishmania donovani using recombinase polymerase amplification assay.

Parasit Vectors. 2016 May 13;9(1):281.

108.   Ghosh P, Sharma A, Bhattarai NR, Abhishek K, Nisansala T, Kumar A,

et al. A Multi-Country, Single-Blinded, Phase 2 Study to Evaluate a Point-of-

Need System for Rapid Detection of Leishmaniasis and Its Implementation in

Endemic Settings. Microorganisms. 2021 Mar 12;9(3):588.

109.   Milenia Biotec GmbH. A short guide for improved readout performance.

Available from: https://www.milenia-biotec.com/uploads/2019/07/improved-

CRISPR-readout_final-1.pdf

1305

Supporting information captions

S1 Fig. Multiple sequence alignment of representative kDNA minicircle sequences

from members of the *L*. (*Viannia*) and *L*. (*Leishmania*) subgenera, with LAMP

primer binding regions and crRNA target site highlighted. A total of 24 *L*. (*Viannia*)

and 35 *L*. (*Leishmania*) sequences were aligned using Clustal Omega (alignments were

performed separately for each subgenus) and visualized in Jalview v2.11.4.1. The

locations of primer binding regions as well as of the crRNA target site and PAM

sequence in the target DNA are highlighted with colors to indicate sequence

conservation within the *L*. (*Viannia*) subgenus and sequence divergence within the *L*.

(*Leishmania*) subgenus. Nucleotide variants are represented by spaces in the alignment.

The kDNA minicircle sequence from the *L*. (*V*.) *braziliensis* M2904 strain (GenBank

accession no. KY698803.1) exhibited 100% identity across all primer and crRNA target

regions. The sequences shown are from selected representative strains/isolates of a

given *Leishmania* species; additional species and sequences can be examined in the full

alignments included in the S1 File. Figure created with BioRender.com.

S2 Fig. Multiple sequence alignment of representative 18S rDNA sequences from

members of the *Leishmania* and *Trypanosoma* genera, with LAMP primer binding

regions and crRNA target site highlighted. A total of 31 sequences (16 from

Leishmania spp. and 15 from *Trypanosoma* spp.) were aligned using Clustal Omega and

visualized in Jalview v2.11.4.1. Primer binding regions as well as the crRNA target site

and PAM sequence in the target DNA are highlighted with colors to indicate sequence

conservation within the *Leishmania* genus and sequence variations within the

Trypanosoma genus. Nucleotide variants are represented by spaces in the alignment.

The sequences shown are from selected representative strains/isolates of a given

species; additional species and sequences can be examined in the full alignments

included in the S2 File. Figure created with BioRender.com.

S3 Fig. Optimization of the incubation time of the LAMP reaction. To determine the

minimum necessary incubation time of the LAMP reaction for consistent DNA

amplification, tests were performed at different reaction times for *Leishmania* kDNA

(10, 20, 30, and 40 min) and 18S rDNA (10, 30, 40, 50, and 60 min) targets. The

number and text enclosed in a rectangle indicates the LAMP reaction time. The

amplification efficiency was assessed using varying amounts of template DNA (*L.*

braziliensis M2904 gDNA), corresponding to approximately 2×10^2, 2×10^{-1}, and $2 \times$

10^{-3} genome equivalents. A negative control (NTC) reaction was tested in parallel. Raw

fluorescence signals from Cas12a reactions are shown for kDNA (A-D) and 18S rDNA

(G-K) for one representative technical replicate. Normalized data (fluorescence ratio at *t*

= 25 min) from two technical replicates are shown in panels E (kDNA) and L (18S

rDNA). A fluorescence ratio ≥ 2 is considered detected. Fluorescence measurements in

this figure were taken on the Varioskan LUX plate reader. LAMP reaction products (5

µL each) of kDNA (F) and 18S rDNA (M) were analyzed by 2% agarose gel

electrophoresis ran at 100 V for 1 h using SYBR Gold staining. L, 100 bp DNA ladder (ABclonal). Since space was tight on the agarose gel images, we had to abbreviate the labels for each lane. For instance, "10^2" actually represents "2×10^2" as indicated in other panels of this figure.

S4 Fig. Optimization of CRISPR-based detection assays using a lateral flow readout. (A) Design of the Milenia GenLine HybriDetect test strips and readout interpretation. (Left) In the absence of the specific genetic target, the intact biotin-FAM labeled reporter molecule flows to the control line (C-line). This is interpreted as a negative test result. (Right) Upon recognition of the genetic target, the CRISPR RNP complex cleaves the reporter molecule, which flows to the test line (T-line). The signal intensity of the C-line is weakened. If cleavage of the reporter molecule is partial, both the C-line and T-line may appear with comparable intensity. Both scenarios are interpreted as a positive test result. Figure adapted from [109] and created with BioRender.com. (B-F) Different parameters influencing the LFA readout performance were evaluated to achieve consistent results: reporter concentration (B), assay buffer (C), incubation temperature (D) and reaction time (E) of the Cas12a assay, and the concentration of PEG-8000 (F). Results shown here correspond to the kDNA LAMP-CRISPR assay.

S5 Fig. Workflow from sample preparation to LAMP-CRISPR assays for analytical specificity testing. Figure created with BioRender.com.

S6 Fig. Results of LAMP-CRISPR detection with fluorescence readout in a group of clinical samples. Raw fluorescence curves generated by Cas12a detection of LAMP amplicons from *Leishmania* kDNA (A) and 18S rDNA (B) targets in a representative group of clinical samples harboring varying parasite load levels (see S1 Dataset) and ran in the same assay plate per genomic target. Out of the 8 tested samples shown, 6

showed robust fluorescence curves indicating the presence of *Leishmania* kDNA

molecules (panel A; samples coded S16, S23, S31, S52, S72, and S75) whereas the 18S

rDNA target was detected in 4 of them (panel B; samples coded S16, S23, S52, and

S75). Fluorescence measurements in this figure were taken on the Synergy H1 plate

reader. The respective LAMP amplicons (5 µL) targeting kDNA (C) or 18S rDNA (D)

were analyzed by 2% agarose gel electrophoresis (100 V, 60 min) using SYBR Gold

staining. Controls included a positive control (PC, 5×10^4 genome equivalents from *L.*

braziliensis M2904 gDNA), a human negative control (HC, human PBMC gDNA), and

a no-template control (NTC) of the LAMP reaction.

S7 Fig. Results of LAMP-CRISPR detection with LFA readout in a group of

clinical samples. Clinical samples chosen for this pilot testing (n = 8) were categorized

according to the parasite load levels and tested with kDNA and 18S LAMP-CRISPR

assays. Lateral flow testing was performed on two assay repeats, R1 and R2, per

sample. Visualization on the lateral flow strips (Milenia GenLine HybriDetect strips)

was achieved through the cleavage of the biotin-FAM reporter molecule. Following

LFA analysis, the strips were placed into the cassette housing for image acquisition. C,

control line; T, test line. Controls included DNA samples from two non-leishmaniasis

patients (i.e., kDNA qPCR-negative; samples 09 and C7), a positive control (PC, $2 \times$

10^2 genome equivalents from *L. braziliensis* M2904 gDNA), a human negative control

(HC, 20 ng of human PBMC gDNA), and a no-template control (NTC) of the LAMP

reaction.

S1 Table. Primer and crRNA template sequences used in this study.

S1 File. *In silico* alignment of *Leishmania* kDNA minicircle sequences from

***Viannia* and *Leishmania* subgenera, and location of primer and crRNA binding**

sites. This file contains sequence alignments and visualizations used to assess the *in*

silico specificity of the designed LAMP primers and Cas12a crRNA within conserved

regions of *Leishmania* kDNA minicircles. A total of 1,060 aligned sequences from

strains/isolates belonging to the *L.* (*Viannia*) subgenus and 965 aligned sequences from

strains/isolates of the *L.* (*Leishmania*) subgenus were included. Primer and crRNA

binding regions were manually annotated using Jalview v2.11.4.1 to enhance

visualization and facilitate identification of conserved target sites. Additionally, a subset

of representative sequences, 24 *L.* (*Viannia*) and 35 *L.* (*Leishmania*), is provided in

separate alignment project files to illustrate kDNA minicircle sequence conservation

and divergence across the two subgenera (see also S1 Fig).

S2 File. *In silico* alignment of 18S rDNA sequences from *Leishmania* and

***Trypanosoma* genera, and location of primer and crRNA binding sites.** This file

contains the alignments and visualizations used to evaluate the *in silico* specificity of

the designed LAMP primers and Cas12a crRNA targeting the 18S rDNA. It includes 37

aligned sequences of *Leishmania* spp. and 272 of *Trypanosoma* spp. Primer and crRNA

binding regions were manually annotated using Jalview v2.11.4.1 to enhance

visualization and identification. Additionally, an alignment of a subset of 31

representative sequences (16 *Leishmania*, 15 *Trypanosoma*) is provided for

visualization of 18S rDNA sequence conservation and divergence across the two genera

(see also S2 Fig).

S1 Dataset. Excel file containing the data on tested clinical samples generated

during the current study. This file contains the LAMP-CRISPR and qPCR data on

tested clinical samples, as well as the determination of threshold cutoff values for

fluorescence-based CRISPR-Cas12a assay result interpretation.

1419

Author Contributions

Conceptualization: Vanessa Adaui.

Formal analysis: Eva Dueñas, Percy Huaihua, Vanessa Adaui.

Funding acquisition: Vanessa Adaui.

Investigation: Eva Dueñas, Ingrid Tirado, Percy Huaihua, Ariana Parra del Riego, Luis

Cabrera-Sosa, Jose A. Nakamoto, Carlos M. Restrepo.

Project administration: Vanessa Adaui.

Resources: Luis Cabrera-Sosa, Jose A. Nakamoto, María Cruz, Carlos M. Restrepo,

Jorge Arévalo, Vanessa Adaui.

Supervision: Vanessa Adaui, Jorge Arévalo.

Visualization: Eva Dueñas, Ingrid Tirado.

Writing – original draft: Eva Dueñas, Ariana Parra del Riego, Vanessa Adaui.

Writing – review & editing: Eva Dueñas, Ingrid Tirado, Percy Huaihua, Ariana Parra

del Riego, Luis Cabrera-Sosa, Jose A. Nakamoto, María Cruz, Carlos M. Restrepo,

Jorge Arévalo, Vanessa Adaui.

1435

Acknowledgments

We thank Pohl Milón (Universidad Peruana de Ciencias Aplicadas) for inspiring

discussions; Yomara Romero (Universidad Peruana de Ciencias Aplicadas) for

assistance with sequence alignments and advice on their presentation; Edith Málaga and

Manuela Verástegui (Universidad Peruana Cayetano Heredia) for providing genomic

DNA from the *Trypanosoma cruzi* strains tested here; and the Office of Research at the

Universidad Peruana de Ciencias Aplicadas for the support provided to carry out this

research work.

1444

Funding Statement

This work was supported by ProCiencia, the Peruvian National Council for Science,

Technology, and Technological Innovation (CONCYTEC) - The World Bank (contract

036-2019-FONDECYT-BM-INC.INV) and by the Universidad Peruana de Ciencias

Aplicadas (internal fund C-005-2024). I.T. was supported by a postgraduate scholarship

of ProCiencia (contract PE501084865-2023). The funders had no role in study design,

data collection and analysis, decision to publish, or preparation of the manuscript.

1452

1453

DNA extraction

Skin ulcer

18S rDNA

kDNA

Biopsy Swab Filter paper

LAMP amplification

Immersion oil

Bst DNA pol

LAMP primers

LAMP products

65°C for 30-50 min

CRISPR/Cas12a detection

crRNA

Cas12a

Reporter

25°C for 25 min

Readout

i. Lateral flow strips (LFA)

ii. Fluorescence signal

Fluorescence (RFU)

Detected

Non detected

Time (min)

DNA target

A

B

C

D

E

Forward and backward external primers: F3 B3   Forward and backward internal primers: FIP: F1c F2 BIP: B1c B2   Forward and backward loop primers: LF LB

crRNA target sequence: Kinetoplast (kDNA) minicircles   18S rDNA   PAM sequence: 5' VAAA 3'

18S crRNA target sequence: 5' TGACTCAATTCATTCCGTGCGAAA 3'

kDNA crRNA target sequence: 5' TGGCATACAGAAACCCCGTTCAAA 3'

A

	kDNA	18S rDNA
L.(V.)braziliensis M2904		
L.(V.)braziliensis M2903		
L.(V.)braziliensis LC2177		
L.(V.)braziliensis LC2043		
L.(V.)peruviana WR-2771		
L.(V.)peruviana HB86		
L.(V.)peruviana LCA08		
L.(V.)guyanensis M4147		
L.(V.)lainsoni M6426		
L.(V.)panamensis BD-02		
L.(L.)amazonensis M2269		
L.(L.)mexicana BEL21		
L.(L.)aristidesi GML		
L.(L.)major WR-1088		
L.(L.)donovani WR-2801		
L.(L.)infantum P142		
T.cruzi Tulahuen		
T.cruzi VT-1		
T.cruzi Sylvio X10		
M.tuberculosis PA155-18C		
P.falciparum HB3		
Human control		

2   3   2   3

B

Human control / NTC — Negative controls (lanes 1, 2); Lbra M2904 / Lgvy M4147 / Lper LCA08 — Positive controls (lanes 3, 4, 5); Mt PA155-18C / Pf HB3 — Non-related microbes (lanes 6, 7); Tc Tulahuen / Tc Sylvio X10 / Tc VT-1 — *T. cruzi* strains (lanes 8, 9, 10)

T

C

Parasite load levels

High   Intermediate   Low

kDNA

25   75   21   16   51   76   19   20   07   23   53   47   12   31   35   52   03   05

T
C

LFA readout: +   +   +   +   +   +   +   +   +   +   +   +   +   +   +   +   −   −

RFU readout: +   +   +   +   +   +   +   +   +   +   +   +   +   +   +   +   −   −

Cq: 10.74   12.13   16.82   17.90   17.96   18.36   20.76   21.26   24.38   25.91   25.97   27.30   27.56   27.57   28.76   28.89   29.74   33.88

18S

25   75   21   16   51   76   19   20   07   23   53   47   12   31   35   52   03   05

T
C

LFA readout: +   +   +   +   +   +   +   +   +   +   +   +   −   −   −   −   −   −

RFU readout: +   +   +   +   +   +   +   +   +   +   +   +   +   −   +   +   −   −

Controls

kDNA

09   C7   PC   HC   NTC

T
C

18S

09   C7   PC   HC   NTC

T
C