1

Muscle fatigue during spinal stimulation and resistive ankle exoskeleton use in children with cerebral palsy

Charlotte R. DeVol[1], Victoria M. Landrum[1], Siddhi R. Shrivastav[2], Heather A. Feldner[3,4], Kristie F. Bjornson[5,6], Chet T. Moritz[2,3,7,8], Katherine M. Steele[1,4]*

[1]Department of Mechanical Engineering, University of Washington, Seattle, WA, USA

[2]Department of Electrical and Computer Engineering, University of Washington, Seattle, WA, USA

[3]Department of Rehabilitation Medicine, University of Washington, Seattle, WA, USA

[4]Center for Research and Education on Accessible Technology and Experiences, University of Washington, Seattle, WA, USA

[5]Department of Pediatrics, University of Washington, Seattle, WA, USA

[6]Seattle Children's Research Institute, Seattle, WA, USA

[7]Department of Neurobiology and Biophysics, University of Washington, Seattle, WA, USA

[8]Center for Neurotechnology, Seattle, WA, USA

Corresponding Author: Dr. Katherine M. Steele, PhD; email: kmsteele@uw.edu

Abstract

Evaluating fatigue during rehabilitation can help prevent overexertion to improve motor learning.
The purpose of this study was to quantify how walking with transcutaneous spinal cord stimulation
(tSCS) impacts muscle fatigue during treadmill training with and without a resistive ankle exoskeleton
(Exo) in children with cerebral palsy (CP). Nine children with CP (4-14 years old) participated in four
walking conditions: (1) No Device, (2) tSCS only, (3) Exo only, and (4) tSCS+Exo. Plantarflexor
maximum voluntary contraction (MVC) was performed before and after each condition. Soleus
amplitude root mean square (RMS) and median frequency (MDF) from electromyography recordings
were used as biomarkers for fatigue. Participants only showed signs of soleus fatigue in the first five
minutes of the Exo only condition with increased RMS ($p < 0.001$) and decreased MDF ($p < 0.001$).
After ten minutes of walking, there was a slight decrease in MDF in the No Device and tSCS only
conditions ($p = 0.002$), but no change in RMS. There were no significant differences in changes in
MVC force, RMS, or MDF between conditions. Walking with tSCS may reduce the impact of the Exo
on soleus fatigue for children with CP.

Keywords: cerebral palsy, gait training, spinal stimulation, biomechanics, muscle fatigue,
exoskeletons

Introduction

One of the main challenges among ambulatory people with cerebral palsy (CP) is fatigue in daily life. Individuals with CP use two to three times the energy of nondisabled peers to walk[1–3]. Fatigue limits mobility for both children and adults with CP, affecting their ability to perform tasks of daily living and interact with peers[4–6]. One component of fatigue is muscle fatigue, defined as a reduction in force output[7,8]. Prior to an observable decrease in force output, there are myoelectric manifestations of muscle fatigue. This includes a decrease in the conduction velocity of muscle fibers, which can be quantified as a simultaneous decrease in the frequency and increase in amplitude of electromyographic (EMG) recordings[9]. When performing repetitive tasks, such as walking, the muscles of children with CP fatigue faster than their peers[10–12]. Changes in muscle properties and reductions in the ability to activate all available muscle fibers may make children with CP more likely to rely on weaker, high endurance muscle fibers compared to peers and, thus, make them more prone to muscle fatigue[13].

Considering muscle fatigue is important when developing rehabilitation interventions. When physical training, such as walking, is performed with rest breaks to prevent over-exertion, the nervous system is more adaptable to learning new information, a process known as neuroplasticity[14]. Neuroplasticity is key for children with CP who are learning new tasks during physical therapy. High muscular effort can cause sustained muscle fatigue during a training session and may reduce neuroplasticity and performance[14,15]. Despite the importance of muscle fatigue on neuroplasticity, fatigue is often not quantified during the development or evaluation of rehabilitation interventions. Understanding fatigue during training can provide useful information for determining optimal session length and level of exertion during an intervention, thus helping develop more effective and individualized training.

Several new rehabilitation interventions have recently been introduced for children with CP that may influence muscle fatigue, neuroplasticity, and function. Exoskeletons that provide ankle resistance during push-off have improved walking speed and planarflexor control[16–18]. However, resistance applied during gait training may lead to more muscle fatigue compared to standard gait training. Understanding the effects of a resistive ankle exoskeleton on muscle fatigue may inform treatment session length or resistance levels to optimize conditions for neuroplasticity. Similarly, neuromodulation may also be a viable tool for reducing muscle fatigue and optimizing performance. Specifically, transcutaneous spinal cord stimulation (tSCS) combined with physical training, such as gait training, has recently shown promising improvements in muscle co-contraction, gross motor function, spasticity, and gait kinematics for children with CP[19–24]. Applying tSCS during walking can reduce muscle co-contraction and improve coordination between hip and knee motion during gait[21]. Through improved coordination, tSCS may reduce muscle demand during training to mitigate fatigue and improve conditions for neuroplasticity. Understanding the impact of tSCS on muscle fatigue may assist in its translation into clinical care.

The purpose of this study was to quantify muscle fatigue before, during, and after gait training with and without tSCS and a resistive ankle exoskeleton (Exo) targeting plantarflexor muscles for children with CP. We evaluated muscle activity before, during, and after walking on a treadmill under four conditions: No Device, tSCS only, Exo only, and tSCS + Exo. Compared to walking with No Device, we hypothesized that the Exo would increase muscle fatigue. We further hypothesized that the combination of tSCS with the Exo would reduce the rate of fatigue compared to the Exo alone.

Methods

Participants

Nine children with CP were recruited to evaluate gait training with tSCS and an Exo (Table 1).
Prior to enrollment, informed age-appropriate assent and consent were obtained from participants and
their caregivers, respectively. All visits were completed at the University of Washington. This study
was approved by the University of Washington Human Subjects Division (IRB identifier:
STUDY00014877) and registered at ClinicalTrials.gov (NCT05520359; registered on 25 August
2022). The study methods were performed in accordance with the approved protocol and University
of Washington Institutional Review Board guidelines and regulations. Individuals were eligible to
participate if they were ambulatory, Gross Motor Function Classification System levels I-II, had spastic
CP, had no prior selective dorsal rhizotomy, had not undergone a lower extremity surgery or botulinum
toxin injections in the past year, did not have uncontrolled seizures, were able to follow 2-3 step
commands, and had no implants that might interfere with electrical stimulation. Informed consent was
obtained from all participants' legal guardians for the use of any photos within the manuscript that
show a participant.

Table 1. Participant Characteristics:

ID	Sex	Age Range (years)	Height (cm)	Weight (kg)	Spastic CP Diagnosis	Side	GMFCS Level	Speed (m/s)	tSCS Amplitude (mA)	
									T11	L1
P01	M	10-14	169	64.3	Unilateral	R	I	1.2	30	20
P02	M	0-4	109	17.2	Bilateral	L	I	0.8	35	25
P03	M	10-14	155	38.6	Bilateral	R	II	1.0	20	30
P04	M	15-19	160	48.2	Bilateral	L	II	0.9	35	25
P05	F	5-9	117	20.5	Bilateral	R	II	0.6	10	20
P06	F	5-9	113	19.5	Bilateral	R	II	0.7	10	15-20
P07	F	10-14	157	51.3	Bilateral	L	I	1.0	20	30
P08	M	15-19	163	58.3	Bilateral	L	II	1.1	30	20
P09	M	10-14	135	39.0	Unilateral	R	II	0.8	10	20

ID = participant identifier; Side = more affected side based on the side with more spasticity at baseline and parent reports; GMFCS = Gross Motor Function Classification System; The tSCS amplitude applied to T11 = thoracic spinous process 11, L1 = lumbar spinous process 1.

Experimental Protocol

Participants completed four visits where they walked with or without an Exo and tSCS (Figure 1).
These walking conditions included: No Device, tSCS only (tSCS), resistive exoskeleton (Exo), or the
combination of tSCS and Exo (tSCS+Exo). P06 and P07 did not complete the tSCS+Exo visit due to
a change in eligibility. Participants walked on the treadmill at a preferred speed that was determined at
the first visit and held constant at subsequent visits. Walking conditions were performed on an inground
treadmill (Bertec, Columbus, OH) on separate days at least five days apart in a pseudo-randomized
order. Rest breaks were provided as needed with a maximum of two rest breaks during each session.

For the tSCS and tSCS+Exo visits, tSCS was administered via the Spinal Cord Neuromodulator
(SCONE[TM], SpineX, Inc.) following previously reported protocols[21]. Stimulation was applied using
adhesive gel electrodes placed just below the T11 and L1 spinous processes using 3.2 cm round
electrodes. The ground electrodes were 5.1 x 8.6 cm rectangular electrodes placed over the anterior
superior iliac spine (Figure 1C). The amplitude for the subthreshold stimulation was determined for

Figure 1. Overview of study design, data processing, and devices evaluated. A) Participants completed four study visits walking for a maximum of 20 minutes with: No Device, spinal stimulation (tSCS), a resistive ankle exoskeleton (Exo), and both tSCS and Exo (tSCS+Exo). Participants performed a maximum voluntary contraction (MVC) of the plantarflexor muscles while lying in a supine position before and after walking. Soleus muscle activity was recorded during each MVC and walking condition. B) Processing pipeline for electromyography (EMG) data to calculate median frequency (MDF) and amplitude root mean square (RMS) as biomarkers of muscle fatigue. C) Placement of electrodes for the Spinal Cord Neuromodulator (SCONE™) from SpineX. D) The SPARK exoskeleton by Biomotum provided a torque at the ankle to resist during push-off. FSR = force sensitive resistor.

each individual at their first visit using tSCS and held constant at subsequent visits, with the exception
of P06 who needed a slight adjustment of tSCS due to comfort during their first use of tSCS (Table 1).
The amplitude of tSCS was determined based on children's self-reported ease of walking, sensation
beneath the cathodes, and a clinical interventionist's observation of gait quality and participant's
behavior within the first ten minutes of use. The tSCS was then turned off and a break was provided
until walking began. The amplitude of tSCS was also lowered or turned off during seated breaks and
turned back on during walking, as sitting can change the sensation at the site of stimulation compared
to standing[25].

For the Exo and tSCS+Exo walking conditions, we used the SPARK exoskeleton by Biomotum
(Portland, OR) to provide resistance at the ankle during the push-off phase of gait[17]. The SPARK is an
untethered exoskeleton with a waist strap that carries the battery pack, Bowden cables that connect the
battery and motor to an ankle actuator, and foot plates inside the users' shoe. The SPARK records
timing of gait events from a force sensitive resistor in the foot plate and uses this information to time
when it applies a torque at the ankle to resist push-off (Figure 1D). The ankle resistance torque of the
Exo was set at 12% of each participant's bodyweight, similar to previous studies[16,17,26]. The SPARK
pairs with a mobile application to control the level of resistance, which can also provide audiovisual

feedback on the torque applied at the ankle. The audiovisual feedback feature was not used during the
first five minutes of walking with the Exo. Data after this period, when the audiovisual feedback was
used, was excluded from this analysis because it may impact soleus muscle activity[26].

During all walking conditions, electromyography (EMG) data of the soleus muscle were recorded
at 2000 Hz (Delsys, Inc., Natick MA). We focused on the soleus muscle because the Exo is designed
to target the plantarflexor muscles, and other plantarflexors are not accessible to record with surface
EMG with the Exo's calf cuff in place[17]. We analyzed the more-affected lower limb for all participants,
defined as the side of the body with more spasticity. Before (pre) and after (post) each walking
condition, maximum voluntary contraction (MVC) of the more-affected plantarflexors were obtained
using an isometric manual muscle test with a hand-held dynamometer (Activbody Inc., San Diego,
CA). Participants laid supine and were instructed to keep their knees fully extended while
plantarflexing their foot as strongly as possible for 5 seconds against a researcher providing resistance
with the hand-held dynamometer[18]. Participants were given visual feedback of their force value output
and instructed to raise the number as high as possible and continue pushing until time was up. The
MVC was repeated three times before and after each walking condition and EMG data of the soleus
muscle were recorded throughout and averaged across the three trials. Trials were dropped if there was
an outlier in force value, as judged visually by the research team during a session, and an extra MVC
was taken to include three total MVC's before and after each walking condition. P06 did not complete
the MVC for any walking conditions.

Data Analysis

Soleus muscle median frequency (MDF) and root mean square (RMS) were used as biomarkers
for myoelectric manifestations of muscle fatigue[9,27]. EMG data for walking conditions were analyzed
using a custom MATLAB script (MathWorks, Natick, MA). Soleus EMG signals were bandpass
filtered from 20 – 400 Hz using a 4th order Butterworth filter. Soleus firing rate was quantified by
segmenting bandpass filtered soleus EMG using a Gabor transform moving window centered at 15-
second intervals and calculating the MDF within each window[10,28]. The Gabor transform filters EMG
data in the time domain using a Gaussian function,

156
$$g(t) = e^{-a(t-b)^2}, \tag{1}$$

where $a = -0.01$ is the window width of the filter, t is the time, and $b = 15$ seconds, the distance
between centers of the moving window. The MDF was calculated using the built-in MATLAB
function, medfreq(). EMG data were then zero-centered, rectified, and low-pass filtered (4th order
Butterworth; 10 Hz; Figure 2) before calculating RMS amplitude in 15-second intervals[27]. Soleus firing
rate and RMS were normalized to the calculated firing rate and peak activation of the first 15 seconds
of walking, respectively, so reported values are a percent of the start value.

During the five second MVC task, EMG data were also recorded. The same EMG analyses
described above were repeated, except that MDF and RMS calculations were made over the entire five
second window.

Statistical Analysis

We used linear mixed effects models (LMM) to quantify the rate of change of EMG RMS and
MDF across five minutes of treadmill walking for all four walking conditions: No Device, tSCS only,
Exo only, and tSCS+Exo. Only the first five minutes were used for all four conditions because after
this point, participants began taking breaks or using audiovisual feedback in the conditions that

included Exo, potentially altering muscle fatigue in a nonlinear manner[26]. Use of an LMM was chosen
since all walking conditions contained the same participants. The equations for the EMG RMS and
MDF LMMs were:

180
$$EMG\ Outcome_{ij} = \beta_0 + \beta_1 \cdot Condition_i + \beta_2 \cdot Time_i$$
$$+ \beta_3 \cdot (Condition_i \cdot Time_i) + u_j + \epsilon_{ij}, \tag{2}$$

where $EMG\ Outcome_{ij}$ is the observed value of EMG RMS or MDF for the i-th observation of the
j-th subject. These equations had the best model fit under the assumption that both walking condition
and time influence fatigue rate. We also extended the LMM to included 10 minutes of treadmill
walking for tSCS and No Device conditions, during which no participants took rest breaks. Time was
measured in seconds, resulting in the fixed effects units as percent per second (%/s) and reported with
the 95-percent confidence intervals in brackets.

We compared percent change in force output, EMG RMS, and EMG MDF during the post-post
MVC task recorded before and after each walking condition. Data sets were not normally distributed,
as tested by the Kolmogorov-Smirnov test and a Kruskal-Wallis test was used to measure significant
differences between walking conditions. All statistical tests were performed in MATLAB.

Results

Fatigue During Five Minutes of
Treadmill Walking: All Conditions

Muscle fatigue, quantified as an
increase in RMS and a decrease in MDF,
occurred only during the first five minutes
of the Exo condition (Figure 3). Walking
with No Device for five minutes did not
cause a significant change in either RMS
or MDF (p = 0.67 and 1.0, respectively).
Walking with the Exo, there was a
significant increase in RMS (β_3 = 0.066
%/s [0.029, 0.10]) and decrease in MDF
(β_3 = -0.081 %/s [-0.11, -0.049])
compared to walking with No Device
(p<0.001). Walking with tSCS only or
Exo + tSCS had no effect on RMS (tSCS
p = 0.40; Exo + tSCS p = 0.43) or MDF
(tSCS p = 0.46; Exo + tSCS p = 0.89)
compared to walking with No Device.

Fatigue During Ten Minutes of Treadmill
Walking: No Device and tSCS Only

We also quantified changes in RMS
and MDF during ten minutes of
consecutive walking with No Device and
tSCS only to see if longer trials would
induce muscle fatigue (Figure 4).

Figure 2. Soleus EMG over the gait cycle for each walking
condition: No Device, tSCS only, Exo Only, and tSCS+Exo.
Each plot shows the gait cycles averaged over a 15-second
duration at minute zero, five, and ten. Thin grey lines are for
each participant and the thick black lines are the average
across participants. The vertical axis is the EMG amplitude
normalized to the RMS over the first 15-seconds.

Walking for 10 minutes with No Device led to no changes in RMS (p-value = 0.81) and a small but
significant reduction in MDF (β_3 = -0.0088 %/s [-0.014, -0.0031]; p-value < 0.002). Walking with
tSCS for ten minutes resulted in no significant differences in RMS or MDF compared to walking with
No Device (RMS p-value = 0.88; MDF p-value = 0.68).

Fatigue during Pre-Post MVC

Despite the observed muscle fatigue within the first 5 minutes of the Exo condition, there were no
significant changes in MVC force after twenty minutes of walking for any condition (Figure 5; p =
0.20). The force produced during the MVC task had a median increase after the Exo only condition of
6% (-15 to 34 Newtons, N), and decreased after training with No Device, tSCS only, and tSCS+Exo
with reductions of 18% (-51 to 23 N), 3% (-15 to 41 N) and 9% (-43 to 7 N), respectively. Both RMS
and MDF of the EMG signal increased slightly during the MVC after each walking trial, but the
changes were similar between conditions (RMS p = 0.29; MDF p = 0.95). The EMG RMS during
MVC increased by only 6%, 2%, 3%, and 7% after training with No Device, tSCS only, Exo only, and
tSCS+Exo, respectively. The EMG MDF during MVC increased 6%, 6%, 7%, and 5% after training

Figure 3. Myoelectric manifestations of muscle fatigue during five minutes treadmill walking under the four different conditions: No Device, tSCS only, Exo only, and tSCS+Exo. A) The linear mixed effects model (LMM) for RMS of the EMG signal normalized to the peak RMS during the first 15 seconds of each condition. B) The RMS fixed effects (β_3 in Eqn. 2) of the interaction between time and condition with respect to the No Device condition show a significant increase with Exo only. C) The LMM for MDF of the EMG signal, normalized to the MDF in the first 15 seconds of each condition. D) The MDF fixed effects (β_3 in Eqn. 2) of the interaction between time and condition with respect to the No Device condition show a significant decrease with Exo only. Areas indicating an increase in fatigue relative to No Device are indicated with grey arrows and shaded boxes. * significant difference relative to No Device condition (p < 0.05).

with No Device, tSCS only, Exo only, and tSCS+Exo, respectively.

Discussion

As hypothesized, training with the Exo induced the greatest rate of muscle fatigue within the first
five minutes of walking, as indicated with a simultaneous increase in soleus RMS and decrease in
MDF. Soleus fatigue was not observed during the tSCS+Exo condition, suggesting that tSCS may
reduce the impact of the Exo on muscle fatigue during training. Walking with No Device or with tSCS
only were very similar in their rates of muscle fatigue, suggesting that five to ten minutes of treadmill
training at a comfortable speed induced minimal myoelectric manifestations of muscle fatigue.
Changes in the soleus RMS and MDF during MVC also did not indicate muscle fatigue.

Children with CP have greater muscle fatigue during five minutes of walking compared to typically
developing (TD) peers. Eken et al., (2019) reported a significantly greater rate of fatigue via increased
EMG RMS and decreased MDF in the gastrocnemius and soleus muscles on the more-affected side of
children with CP compared to TD peers[29]. Their study evaluated overground walking, while here we
utilized walking on a treadmill, which may offload the plantarflexors and put more demand on the
quadriceps[30]. Parent et al., (2019) reported significant reductions in rectus femoris MDF across six
minutes of treadmill walking in children with CP, but no change in other muscles[10]. Similarly, we did
not observe simultaneous changes in RMS and MDF that would indicate muscle fatigue in the soleus
during five minutes of treadmill walking with No Device. The exoskeleton provides a dorsiflexion
torque that counters the desired action at push-off, to encourage the user to increase plantarflexor
engagement[18]. The increased rate of fatigue during training with the Exo suggests that the Exo

successfully engaged the soleus and
significantly increased the rate of fatigue. These
findings are specifically for a resistive ankle
exoskeleton. Assistive ankle exoskeletons have
been shown to reduce muscle demand for
children with CP and may result in reduced
muscle fatigue[31].

In ten consecutive minutes of walking with
No Device and tSCS only, we observed a small
but significant decline in MDF with no change
in RMS. Changes in force output can also
induce changes in RMS and MDF[27]. It is
possible that in addition to some muscle fatigue,
participants also had a simultaneous decline in
force output that masked myoelectric
manifestations of muscle fatigue in the RMS.
We aimed to control for differences in force
produced at the ankle by keeping walking speed

Figure 4. Myoelectric manifestations of muscle fatigue during ten minutes treadmill walking under two different conditions: No Device and tSCS only. A) The LMM for RMS of the EMG signal normalized to the peak RMS during the first 15 seconds of each condition. B) The LMM for MDF of the EMG signal, normalized to the MDF in the first 15 seconds of each condition. There were no significant differences between the tSCS and No Device condition for either RMS or MDF ($p > 0.05$).

consistent between walking conditions. But participants were allowed to use handrails for safety, which
could have offloaded some bodyweight from the legs. Future work measuring force output either with
an exoskeleton or instrumented treadmill could examine changes in force alongside fatigue.

We observed minimal changes in the EMG RMS and MDF during the MVC task performed before
and after each walking condition. Prior work evaluating fifteen minutes of treadmill training has shown
to reduce MVC force of the quadricep muscles by 11% in children with CP, with no change among
TD peers[32]. In our study, we focused only on the soleus muscles, which did not indicate any fatigue
with assessment of the MVC tasks alone. The increased sensory feedback when using either Exo or
tSCS may improve coordination of muscles and increase the ability of the user to maximally engage
in the MVC task, despite muscle fatigue[16,33]. Thus, future work should consider muscle fatigue and
coordination of activity across multiple muscles to identify the possible mechanisms altering fatigue
during walking, but not during the pre-post MVC tasks. Another possibility for no change during the
MVC testing is that this task is not as physically demanding as walking for children with CP and occurs
after the children have had a break from walking. Prior research has suggested that children with CP
show similar and even reduced rates of fatigue compared to nondisabled peers during tasks requiring
maximal effort, but this maximal effort (i.e. force output) is also reduced[29,34]. One contributing factor
to reduced maximal effort in children with CP is the potential activation of fewer muscle fibers during
volitional maximal effort tasks compared to TD peers[35]. As a result, we may observe muscle fatigue
more accurately during walking, where the demand on muscles could be higher than during the MVC
task.

There are several limitations of this study. First, we evaluated the first five minutes of walking in
our initial linear mixed effects model of all walking conditions, similar to prior studies on fatigue
during walking in children with CP[10,28]. We were unable to include the conditions with the Exo for
more than five minutes due to participants taking rest breaks and using the audiovisual feedback after
the first five minutes to focus on one leg at a time[26]. Future research should consider alternative analysis
methods to capture these additional cofounders and evaluate fatigue over longer time periods[36].
Second, we quantified fatigue in the soleus muscle, as the exoskeleton primarily targets the
plantarflexors and covers the gastrocnemius. The soleus is generally considered the more fatigue
resistant plantarflexor muscle[37], so greater changes in fatigue may be observed in the gastrocnemius.
The exoskeleton may also increase the engagement of other muscles, such as the rectus femoris.
Additional EMG processing will be needed to remove signal artifact from tSCS for more proximal
muscles[38]. Finally, MVC was quantified with a manual dynamometer with resistance to plantarflexion
provided by a researcher. Measuring MVC with a manual dynamometer is a common practice[18,29,39,40],
but use of an isokinetic dynamometer, such as a Biodex system, would produce more accurate force
measurements. Further, a direct recording of forces produced during walking could clarify changes in
RMS that may also be due to force output.

In conclusion, walking with an Exo increased the rate of muscle fatigue during the first five minutes
of treadmill walking compared to walking with No Device. The addition of tSCS reduced fatigue in
the soleus muscle and likely increased force output and, thus, engagement with the device. These
results suggest that tSCS may be a viable tool for reducing muscle fatigue during training to maximize
rehabilitation benefits of physical therapy for children with CP.

Data Availability

The datasets generated during and/or analyzed during the current study are available from the
corresponding author on reasonable request.

References

1. Bolster, E. A. M., Balemans, A. C. J., Brehm, M. A., Buizer, A. & Dallmeijer, A. J. Energy cost during walking in association with age and body height in children and young adults with cerebral palsy. *Gait Posture* **54**, 119–126 (2017).

2. Waters, R. L. & Mulroy, S. The energy expenditure of normal and pathologic gait. *Gait Posture* **9**, 207–231 (1999).

3. Rose, J., Gamble, J. G., Burgos, A., Medeiros, J. & Haskell, W. L. Energy expenditure index of walking for normal children and for children with cerebral palsy. *Dev Med Child Neurol* **32**, 333–340 (1990).

4. Jahnsen, R., Stanghelle, J. K. & Holm, I. Fatigue in adults with cerebral palsy in Norway compared with the general population. *Dev Med Child Neurol* **45**, 296–303 (2003).

5. Hilberink, S. R. *et al.* Health issues in young adults with cerebral palsy: Towards a life-span perspective. *J Rehabil Med* **39**, 605–611 (2007).

6. Dominici, N. *et al.* Systematic Review of Fatigue in Individuals With Cerebral Palsy. *Front Hum Neurosci* **15**, 1–16 (2021).

7. Gandevia, S. C. Spinal and Supraspinal Factors in Human Muscle Fatigue. *Physiol Rev* **81**, 1725–1789 (2001).

8. Bigland Ritchie, B., Johansson, R., Lippold, O. C. J. & Woods, J. J. Contractile speed and EMG changes during fatigue of sustained maximal voluntary contractions. *J Neurophysiol* **50**, 313–324 (1983).

9. De Luca, C. J. Myoelectrical manifestations of localized muscular fatigue in humans. *Crit Rev Biomed Eng* **11**, 151–279 (1984).

10. Parent, A. *et al.* Muscle fatigue during a short walking exercise in children with cerebral palsy who walk in a crouch gait. *Gait Posture* **72**, 22–27 (2019).

11. Eken, M. M. *et al.* Squat test performance and execution in children with and without cerebral palsy. *Clinical Biomechanics* **41**, 98–105 (2017).

12. Moreau, N. G., Li, L., Geaghan, J. P. & Damiano, D. L. Fatigue Resistance During a Voluntary Performance Task Is Associated With Lower Levels of Mobility in Cerebral Palsy. *Arch Phys Med Rehabil* **89**, 2011–2016 (2008).

13. Stackhouse, S. K., Binder-Macleod, S. A. & Lee, S. C. K. Voluntary Muscle Activation, Contractile Properties, and Fatigability in Children With and Without Cerebral Palsy. *Muscle Nerve* **31**, 594–601 (2005).

14. Xu, Y. *et al.* Rehabilitation Effects of Fatigue-Controlled Treadmill Training After Stroke: A Rat Model Study. *Front Bioeng Biotechnol* **8**, 1–17 (2020).

15. Ma, J., Chen, H., Liu, X., Zhang, L. & Qiao, D. Exercise-Induced Fatigue Impairs Bidirectional Corticostriatal Synaptic Plasticity. *Front Cell Neurosci* **12**, 1–14 (2018).

16. Conner, B. C., Schwartz, M. H. & Lerner, Z. F. Pilot evaluation of changes in motor control

after wearable robotic resistance training in children with cerebral palsy. *J Biomech* **126**, 110601
(2021).

17.   Conner, B. C., Luque, J. & Lerner, Z. F. Adaptive ankle resistance from a wearable robotic
device to improve muscle recruitment in cerebral palsy. *Ann Biomed Eng* **48**, 1309–1321
(2020).

18.   Conner, B. C., Remec, N. M., Orum, E. K., Frank, E. M. & Lerner, Z. F. Wearable adaptive
resistance training improves ankle strength, walking efficiency and mobility in cerebral palsy:
A pilot clinical trial. *IEEE Open J Eng Med Biol* **1**, 282–289 (2020).

19.   DeVol, C. R. *et al.* Effects of spinal stimulation and short-burst treadmill training on gait
biomechanics in children with cerebral palsy. *Gait Posture* **118**, 25–32 (2025).

20.   Hastings, S. *et al.* A pilot study combining noninvasive spinal neuromodulation and activity-
based neurorehabilitation therapy in children with cerebral palsy. *Nat Commun* **13**, 1–7 (2022).

21.   Gad, P. *et al.* Transcutaneous spinal neuromodulation reorganizes neural networks in patients
with cerebral palsy. *Neurotherapeutics* **18**, 1953–1962 (2021).

22.   Solopova, I. A. *et al.* Effects of spinal cord stimulation on motor functions in children with
cerebral palsy. *Neurosci Lett* **639**, 192–198 (2017).

23.   Shrivastav, S. R. *et al.* Transcutaneous spinal stimulation and short-burst interval treadmill
training in children with cerebral palsy: a pilot study. *IEEE Trans Biomed Eng* **72**, 1–12 (2025).

24.   Sachdeva, R., Girshin, K., Shirkhani, Y., Gad, P. & Edgerton, V. R. Combining spinal
neuromodulation and activity based neurorehabilitation therapy improves sensorimotor
function in cerebral palsy. *Frontiers in Rehabilitation Sciences* **4**, 1–5 (2023).

25.   Keller, A. *et al.* Noninvasive spinal stimulation safely enables upright posture in children with
spinal cord injury. *Nat Commun* **12**, 1–11 (2021).

26.   Spomer, A. M., Conner, B. C., Schwartz, M. H., Lerner, Z. F. & Steele, K. M. Audiovisual
biofeedback amplifies plantarflexor adaptation during walking among children with cerebral
palsy. *J Neuroeng Rehabil* **20**, 1–17 (2023).

27.   Cifrek, M., Medved, V., Tonković, S. & Ostojić, S. Surface EMG based muscle fatigue
evaluation in biomechanics. *Clinical Biomechanics* **24**, 327–340 (2009).

28.   Eken, M. M. *et al.* Lower limb muscle fatigue during walking in children with cerebral palsy.
Dev Med Child Neurol **61**, 212–218 (2019).

29.   Eken, M. M., Dallmeijer, A. J., Houdijk, H. & Doorenbosch, C. A. M. Muscle fatigue during
repetitive voluntary contractions: A comparison between children with cerebral palsy, typically
developing children and young healthy adults. *Gait Posture* **38**, 962–967 (2013).

30.   Lee, S. J. & Hidler, J. Biomechanics of overground vs. treadmill walking in healthy individuals.
J Appl Physiol **104**, 747–755 (2008).

31.   Lerner, Z. F., Damiano, D. L. & Bulea, T. C. The effects of exoskeleton assisted knee extension
on lower-extremity gait kinematics, kinetics, and muscle activity in children with cerebral palsy.
Sci Rep **7**, 1–12 (2017).

32. Vitiello, D. *et al.* Walking-induced muscle fatigue impairs postural control in adolescents with
unilateral spastic cerebral palsy. *Res Dev Disabil* **53–54**, 11–18 (2016).

33. Edgerton, V. R., Hastings, S. & Gad, P. N. Engaging Spinal Networks to Mitigate Supraspinal
Dysfunction After CP. *Front Neurosci* **15**, 1–5 (2021).

34. Leunkeu, A. N., Keefer, D. J., Imed, M. & Ahmaidi, S. Electromyographic (EMG) Analysis of
Quadriceps Muscle Fatigue in Children With Cerebral Palsy During a Sustained Isometric
Contraction. *J Child Neurol* **25**, 287–293 (2010).

35. Rose, J. & McGill, K. C. Neuromuscular activation and motor-unit firing characteristics in
cerebral palsy. *Dev Med Child Neurol* **47**, 329–336 (2005).

36. DeVol, C. R. *et al.* Effects of interval treadmill training on spatiotemporal parameters in
children with cerebral palsy: A machine learning approach. *J Biomech* **177**, 1–8 (2024).

37. Cronin, N. J., Avela, J., Finni, T. & Peltonen, J. Differences in contractile behaviour between
the soleus and medial gastrocnemius muscles during human walking. *Journal of Experimental
Biology* **216**, 909–914 (2013).

38. Kim, M. *et al.* A novel technique to reject artifact components for surface EMG signals recorded
during walking with transcutaneous spinal cord stimulation: A pilot study. *Front Hum Neurosci*
15, 1–14 (2021).

39. O'Brien, S. M., Carroll, T. J., Barber, L. A. & Lichtwark, G. A. Plantar flexor voluntary
activation capacity, strength and function in cerebral palsy. *Eur J Appl Physiol* **121**, 1733–1741
(2021).

40. Tedroff, K., Knutson, L. M. & Soderberg, G. L. Co-activity during maximum voluntary
contraction: a study of four lower-extremity muscles in children with and without cerebral
palsy. *Dev Med Child Neurol* **50**, 377–381 (2008).

Acknowledgements

The authors thank the children and their families for the time they dedicated to the research. We also thank the fellow researchers who supported our data collection visits and preliminary data analysis, especially Katherine Green and Yusuke Maruo. Additional researchers and clinicians who supported data collection included Roozbeh A. Katiraie, Michael Hall, Kathleen Landwehr-Prakel, and Lindsey Jouett.

Author Contributions

All authors conceived of and designed the research. C.R.D., V.M.L., and S.R.S. performed data collection. C.R.D., V.M.L., and K.M.S. carried out all data analysis and interpreted experimental results. C.R.D. and V.M.L. prepared figures and drafted the manuscript. All authors edited and revised the manuscript and approved the final version.

Funding

This work was supported by Seattle Children's Hospital CP Research Pilot Study Fund 2022 Award, NSF Graduate Research Fellowship Program Award DGE-1762114, and the National Institute

of Neurological Disorders and Stroke of the National Institutes of Health under award number
R01NS091056.

Declarations

Competing interests

C.T.M. serves as a clinical advisor to the company SpineX, who provided the stimulator for the
study. SpineX also licensed IP generated by the team at the University of Washington including
C.R.D., S.R.S., C.T.M., and K.M.S..