Postoperative Determination of Directional Deep Brain Stimulation (DBS) Lead Orientation Using Rotational Fluoroscopy: Interobserver Agreement and Comparison with CT-Based Software

Authors:

Paresh K Doshi*1, M.Ch.; Chandresh Karnavat2, D.N.B. ; Raj V Agarbattiwala1 M.Ch.

Affiliations:

1. Stereotactic and Functional Neurosurgery Department, Jaslok Hospital and Research Center, Mumbai, Maharashtra, India

2. Radiology Department, Jaslok Hospital and Research Center, Mumbai, Maharashtra, India

*Corresponding Author: Prof. Paresh K Doshi

Address:

Jaslok Hospital and Research Center,

15, G Deshmukh Marg, Pedder Road, Mumbai-40026.

Phone Number: +912266573185

Email: pareshkd@gmail.com

Abstract Word Count: 206

Text Word Count: 3090

Number of References: 30

Number of Figures: 4

Number of Tables: 1

Number of Videos: 0

Keywords: Directional Leads, Deep Brain Stimulation, Circular Statistics, Lead

Orientation, Iron Sight

Abstract

Background:

Directional deep-brain stimulation (DBS) requires knowledge of lead orientation to maximize benefit and minimize side effects. While CT-based software is widely used, its accuracy may decrease with the use of oblique leads. Rotational fluoroscopy using the Iron Sight method offers an alternative; however, its reliability in clinical practice has not been fully established.

Objective:

To assess the reliability of the Iron Sight method for postoperative lead orientation and compare it with CT-based software (Brainlab Elements™).

Methods:

We prospectively analyzed 70 directional leads implanted at our center (2022–2024). Orientation was measured on postoperative rotational fluoroscopy by two independent observers and compared with postoperative CT analyzed in Brainlab Elements™. The agreement was evaluated using established reliability statistics.

Results:

Observers using the Iron Sight method produced highly consistent measurements, differing on average by <1° and rarely more than ±7°. Compared with CT-based analysis, the Iron Sight method showed good agreement, with differences typically

within ±15°. These margins correspond to <1.5 mm displacement around the lead —

well within clinically acceptable limits for programming.

Conclusion:

Rotational fluoroscopy, combined with the Iron Sight method, provides clinicians with

a reliable and reproducible tool to confirm lead orientation. It may be especially

valuable when CT-based software is unavailable, inconclusive, or requires

corroboration.

Introduction

Deep brain stimulation (DBS) is a recognized treatment for movement disorders like

Parkinson's disease, tremors, and dystonia, as well as drug-resistant epilepsy and

obsessive-compulsive disorder, the latter still being experimental (1–4). The clinical

scope of DBS continues to expand, with ongoing research exploring its role in

psychiatric and neurodegenerative conditions including major depression, addiction,

Tourette syndrome, Alzheimer's disease, and schizophrenia.

Despite decades of clinical use, the technology underlying DBS systems remained

unchanged until the recent introduction of directional leads. Traditional cylindrical

DBS electrodes produce approximately spherical volumes of tissue activation, often

leading to the unintended stimulation of adjacent structures. The potential benefits of

directing stimulation between contacts or in specific directions have long been

anticipated (5). Directional DBS leads to feature-segmented electrodes that allow

clinicians to steer the electrical field, improve therapeutic precision, expand the

therapeutic window, and reduce stimulation-induced side effects (6,7). This steering

is particularly beneficial in targeting areas such as the subthalamic nucleus (STN),

where directional stimulation can avoid spreading to the internal capsule (IC) while

enhancing the coverage of functionally relevant pathways.

Accurate knowledge of the final orientation of the segmented contacts is crucial for

individualized stimulation planning (8–11). Although electrodes are inserted with the

intended orientation (typically anterior), intraoperative lead rotation and torsion can

alter the final position. Despite the directional markers in existing commercial

models, accurately determining the orientation of the lead continues to pose

difficulties (12). Consequently, postoperative imaging is essential to confirm lead

orientation. Among various imaging techniques, postoperative computed

tomography (CT) combined with preoperative magnetic resonance imaging (MRI)

and interpretation using commercial software (e.g., Brainlab Elements™) is a

common practice (13). However, CT-based methods may lose accuracy when leads

are positioned at oblique angles (>40°) (14).

Techniques such as the "Iron Sight" method, which exploits the radiographic

appearance of segmented contacts to infer orientation, have demonstrated high

accuracy in phantom models. Reinacher et al. reported a limit of agreement (LoA)

of ±2.44° and an intraclass correlation coefficient (ICC) of 0.9998 using this method

(12). However, its inter-rater reliability in real-world clinical scenarios has not been

sufficiently studied. Furthermore, the agreement between fluoroscopy-based

estimates and software-derived CT orientations has not yet been systematically

quantified.

Aim and Objectives

Aim

97 To evaluate the inter-rater agreement in determining directional DBS lead orientation

98 using the Iron Sight method on postoperative 3D rotational fluoroscopy and to

99 compare these observer-derived estimates with orientation measurements from

100 Brainlab Elements™ software based on postoperative CT imaging.

Objectives

- 102 To assess interobserver agreement between a neurosurgeon and a
  103 radiologist in determining directional DBS lead orientation using the iron-sight
  104 method on postoperative 3D rotational fluoroscopy.

- 105 To compare orientation measurements obtained via the Iron Sight method
  106 with those derived from the Brainlab Elements™ software using postoperative
  107 CT.

- 108 To assess the association of lead trajectory angles with measurement errors
  109 by the above methods

- 110 The reliability and statistical uncertainty of the assessments were quantified
  111 using circular statistics and intraclass correlation.

Methods

Reporting in the article follows the Guidelines for Reporting Reliability and Agreement Studies (GRRAS) to ensure the full disclosure of design and statistical decisions. This prospective exploratory study was conducted at a high-volume tertiary care center between May 2022 and December 2024.

Participants

All consecutive patients who underwent DBS surgery with directional leads (Vercise Cartesia™, Boston Scientific) during the study period were screened for inclusion. The inclusion criteria were as follows.

- Implantation of directional DBS leads (Vercise Cartesia™, Boston Scientific) for movement disorders.
- The participant (or legally authorized representative) must have given consent.

Patients with incomplete imaging, significant motion artifacts, or postoperative complications affecting the lead visualization were excluded.

The study was conducted in accordance with the Declaration of Helsinki and was approved by the Institutional Ethics Committee at our center (approval number EC/1114/2022, dated April 29, 2022). Participants were recruited from among patients who underwent DBS surgery with directional leads (Vercise Cartesia™, Boston Scientific) between May 1, 2022, and December 31, 2024. All patients and caregivers received both verbal and written explanations of the study, and written

consent was obtained from each participant (or their representative) for participation

and secondary analysis of the de-identified data. Both CT and fluoroscopy were

performed during the postoperative period after admission. All surgeries were

conducted under the Stereotactic and Functional Neurosurgery department at our

center by a single neurosurgeon (PD) according to the standard protocol (15). CT

scans were performed in the radiology department, and rotational fluoroscopy was

performed in the Cath lab (see Table A of the Supplementary Table for the

parameters).

Sample Size

Sample size estimation for comparing rater reliability or agreement has not been

thoroughly explored. Fleiss proposed a rule of thumb that recommends 15–20

samples for reliability studies (16). In the case of a continuous outcome, increasing

the sample size beyond 12 per group did not significantly affect the confidence

interval. Reinacher et al. calculated a 95% confidence interval of ±2.44 ° in their

study of phantom models. Using Zou's formula for ICC precision with 95% CI width ≤

0.1, α = 0.05, ICC = 0.85, and 80% assurance, the required sample size was 57

leads for two raters (17). We chose an expected ICC of 0.85 and a 95% CI width of ≤

0.1 because ICC ≥ 0.75 suggests good reliability. Our study included 70 leads that

met this threshold (18). However, we included 70 leads to account for real-world

variability in human data and to ensure robustness in statistical comparisons.

Postoperative Imaging and Orientation Measurement

Rotational Fluoroscopy (Iron Sight Method)

Within 48 hours postoperatively, all patients were brought to Cathlab for 3D rotational fluoroscopy. The acquisition parameters are listed in the Supplementary Table. Images were reconstructed using multiple projections and independently reviewed by two raters.

- A neurosurgeon (RA) experienced in DBS surgery and programming
- A radiologist (CK) trained in neuroimaging

The observers performed measurements independently on anonymized files and had no access to each other's results. Using the Iron Sight method, each observer independently estimated the rotational orientation of each lead relative to the anterior direction, expressed in degrees (–180° to +180°) (Supplementary Data).

Determination of lead orientation on CT scans

Postoperative CT scans were uploaded into Brainlab Elements™ (Lead Localization module). The acquisition parameters are listed in Table A of the Supplementary Tables. The system automatically detected the lead axis and segmented the contact positions based on CT artifact patterns (as described by Hellerbach et al., 2018 (19)), which were then visually confirmed. The lead model was set to Boston Scientific Vercise Cartesia™. The output orientation is recorded as the reference value for each lead. Given that Brainlab Elements™ itself has known limitations at oblique angles, this method comparison should be interpreted as assessing agreement between the two techniques rather than validating one against a true

177 gold standard. No phantom-based ground truth data were available in this clinical

178 setting.

Statistical Analysis

180 All statistical analyses were performed using MATLAB (MathWorks, Natick, MA,

181 USA) and the Circular Statistics Toolbox (20). Statistical significance was set *P-value*

182 *< 0.05.*

Interobserver Agreement

To assess interobserver agreement using the Iron Sight method, orientation angles

were treated as circular variables ranging from –180° to +180°. The circular

difference, or shortest angular distance, was calculated for each lead between the

two raters' measurements. Circular variance was computed to quantify the

dispersion of the ratings, and the circular correlation coefficient (21) was used to

assess the strength of association between the estimates of the neurosurgeon and

radiologist. Because angular dispersion was low, we used ICC (2,1), as defined by

Shrout and Fleiss (1979) (22), which is appropriate for two-way random effects

models assessing absolute agreement between two observers. This ICC formulation

accounts for subject variability, rater variability, and residual errors.

A Bland–Altman analysis (23,24) adapted for circular data (Online Appendix) was

conducted to further explore the agreement. This included calculating the mean

angular difference (i.e., circular bias) and standard deviation of the differences. The

95% limits of agreement (LoA) were computed as the bias ± 1.96 of the standard

deviation. To express the uncertainty in these estimates, 95% confidence intervals

for both the bias and limits of agreement were calculated. Finally, the Preiss-Fisher

procedure was applied to test whether the observed agreement could have occurred

by chance (25).

Method Comparison (Ironsight method vs Brainlab™)

For method comparison between the iron-sight approach and Brainlab Elements™,

the Watson–Williams test was used to compare the circular means derived from the

observers' average ratings with the angles obtained from the Elements™

software(26). The Intraclass Correlation Coefficient (ICC (3,1)) was calculated to

quantify the consistency between the two methods, using a two-way mixed-effects

model. Agreement between the iron-sight method and Brainlab™-derived orientation

angles was also visualized using a Bland–Altman plot. The Preiss-Fisher test was

again applied to confirm whether the observed concordance between the methods

exceeded chance expectations. Before the parametric circular tests, assumptions

were assessed and met, including angular homogeneity and concentration.

Regression Analysis

We performed two univariate linear regression analyses to investigate whether

anatomical trajectory angles influenced the angular measurement differences.

Linear regression of the coronal, sagittal, polar, and azimuthal angles (as explained

by Matias et al. (27)) against the angular difference between the two observers'

measurements on rotational fluoroscopy (i.e., interobserver error). A separate

regression of these angles against the angular differences between rotational

fluoroscopy and Brainlab Elements™ (i.e., method discrepancy).

For each regression model, beta coefficients and the corresponding p-values were

calculated to assess the strength and significance of the association between the

variables. Owing to multiple comparisons, we adjusted the p-values (P_{adj}) for the

false discovery rate using the Benjamini-Hochberg method (28). Statistical

significance was set at $P < 0.05$.

Results

Study Sample

Thirty-five patients were included in the final analysis, yielding 70 directional DBS

leads for inter-observer agreement analysis. For the method comparison analysis,

five leads were excluded because the Brainlab Elements™ software failed to

compute the lead orientation, resulting in 67 leads being analyzed. An overview of

the data, including circular means, ranges, and standard deviations for all

measurements, is presented in Table 1.

Interobserver Agreement: Iron Sight Method on Rotational Fluoroscopy

The orientations estimated by the two observers were visualized using polar

histograms (Figure 1A-D). The Rayleigh test for uniformity demonstrated significant

data clustering ($P < 0.05$), indicating that the angles were not randomly distributed

around the circle, but were concentrated around a common direction.

The mean resultant length of the circular distance between the two observers was

0.9985 with a circular variance of 0.0015, indicating highly consistent

measurements. The circular correlation coefficient was 0.9822, confirming a strong

directional agreement. A Circular Bland–Altman analysis (figure 2A) revealed a

mean bias of 0.68° (95% CI: –0.0678° to 1.4335°). The upper limit of agreement

(LoA) was 6.85° (95% CI: 5.55° to 8.15°) and the lower LoA was –5.49° (95% CI: –

6.79° to 4.19°). There was no evidence of proportional bias ($P > 0.05$), suggesting

that the measurement errors did not change with angle. The Preiss–Fisher test

showed that the standard deviation of differences between two observer

measurements yielded a Z-value of –15.40 (Figure 2B), indicating that the observed

agreement was significantly greater than would be expected by chance. The

Intraclass Correlation Coefficient (ICC (2,1)) was 0.983 (95% CI: 0.973–0.989),

indicating excellent inter-observer reliability (Figure 2C and Table B of the

Supplementary Tables for calculation parameters).

Figure 1E shows that the histogram of the circular distances between observers

exhibited a symmetric distribution, approximating a Gaussian shape. Similarly, the

polar histogram of angular differences was visually compared with the von Mises

distribution (Figure 1F) and showed good conformity.

Univariate linear regression analysis revealed that the coronal angle (β = 0.1803, *P-*

value = 0.0367; *adjusted P$_{adj}$ = 0.073*) and polar angle (β = 0.0792, *P-value =*

0.0320; *adjusted P$_{adj}$ = 0.073*) were associated with greater interobserver

differences. Note that after controlling for the false discovery rate using the

Benjamini–Hochberg procedure, these associations did not remain statistically

significant at the threshold of 0.05. The sagittal angle (β = –0.0588, *P-value =*

0.3212; *adjusted P$_{adj}$ = 0.428*) and azimuthal angle (β = 0.0364, *P-value = 0.6618*;

adjusted P$_{adj}$ = 0.662) showed no significant association with interobserver variation.

Method Comparison: Iron Sight Method vs. Brainlab Elements™

Of the 70 leads, 67 were eligible for method comparison, as the software failed to

estimate the orientation in the three leads. Orientation estimates from both methods

are displayed using polar histograms (figure 3A-D), including circular means. The

Rayleigh test again showed significant nonuniformity ($P < 0.05$), indicating that the

data points were directionally concentrated.

The mean resultant length of the circular distance between the iron-sight method

and the Brainlab™-derived orientation was 0.9922, with a circular variance of

0.0078. The circular correlation coefficient between the two methods was 0.8635.

Bland–Altman analysis (figure 4A) revealed a mean bias of –0.69° (95% CI: –

2.4839° to 1.0993°), with an upper LoA of 13.48° (95% CI: 10.37° to 16.58°) and a

lower LoA of –14.86° (95% CI: –17.97° to –11.76°). No significant proportional bias

was observed ($P > 0.05$), and the Preiss–Fisher Z-value was –10.83 (Figure 4B),

again indicating that agreement between methods was not attributable to chance.

The Intraclass Correlation Coefficient (ICC (3,1)) for the agreement between the Iron

Sight method and Brainlab Elements™ was 0.864 (95% CI: 0.786–0.915) (Figure 4C

and Table C of the Supplementary Tables for calculation parameters), suggesting

good consistency between the two approaches.

Figure 3E shows that the histogram of the circular distances between the methods

approximates a Gaussian distribution. Similarly, the polar histogram of angular

differences showed comparable clustering and was visually comparable to the von

Mises distribution (Figure 3F). The Watson–Williams test yielded a *P-value* of *0.7905*

(>0.05), indicating no statistically significant difference in the circular mean

orientations between the two methods.

None of the lead trajectory angles demonstrated a statistically significant relationship

with the angular difference between the fluoroscopy-derived and Brainlab-derived

orientations. The beta coefficients for the coronal, sagittal, polar, and azimuthal

angles were 0.2546 (*P-value = 0.2374, adjusted P_{adj} = 0.7046*), 0.0315 (*P-value =*

0.8073, adjusted P_{adj} = 0.8073), 0.0217 (*P-value = 0.7976, adjusted P_{adj} = 0.8073*),

and 0.1688 (*P-value = 0.3523, adjusted P_{adj} = 0.7046*), respectively.

Discussion

In this study, we evaluated whether rotational fluoroscopy using the Iron Sight method can serve as a reliable tool for determining the orientation of directional deep brain stimulation (DBS) leads. We also compared these results with postoperative CT-based analysis using Brainlab Elements™. To our knowledge, this is among the first clinical studies to assess the interobserver agreement of the Iron Sight method in a real-world patient population using circular statistical methods and to systematically compare these estimates with those of a commercial CT-based platform (Brainlab Elements™).

Our results demonstrate that two independent observers, a neurosurgeon and a radiologist, achieved excellent agreement when using the Iron Sight method. On average, their measurements differed by less than 1°, and in almost all cases the difference was within ±7°. This indicates that, when properly trained, clinicians can consistently determine lead orientation from rotational fluoroscopy images. This real-world evidence supports previous phantom-based work by Reinacher et al., who reported a similar accuracy of the Iron Sight method with a ±2.44° limit of agreement and ICC of 0.9998 in controlled settings (12).

When comparing the Iron Sight method to Brainlab Elements™, the two approaches showed good overall agreement. The mean difference between methods was less than 1° (−0.69°), and in most cases, results fell within ±15°. Although this range is wider than the differences observed between observers, it corresponds to less than

1.5 mm displacement around the electrode circumference — a margin considered

clinically acceptable for programming.

The broader limits of agreement likely reflect intrinsic differences between the two

methods. The Iron Sight technique relies on direct visualization of segmented

contacts, while the CT-based approach depends on artifact symmetry. The latter is

known to lose accuracy when electrodes are angled steeply (>40° relative to the

scanner axis) (14,29). Nevertheless, under typical implant conditions, both methods

produced broadly comparable estimates, and the Watson–Williams test confirmed

no statistically significant difference in average orientation values.

We also examined whether lead trajectory angles (coronal, sagittal, polar, and

azimuthal) contributed to measurement variability. There was a suggestion that more

lateral or off-vertical leads (coronal and polar angles) increased variability between

observers. However, these associations did not remain significant after correction for

the false discovery rate using the Benjamini-Hochberg method (28). Importantly,

none of the trajectory angles meaningfully affected the agreement between

fluoroscopy- and CT-based estimates. This suggests that both methods are robust

across a wide range of implant trajectories in routine practice.

Orientation accuracy matters because even modest misalignments can influence

therapeutic outcomes. Computational models have shown that a 45° misorientation

can substantially reduce activation of target neurons or increase the risk of

stimulating adjacent structures such as the internal capsule (30). In this context, our

observed differences — within ±7° for observers and ±15° between methods — fall

well below clinically concerning thresholds. These results reassure clinicians that

both rotational fluoroscopy and CT-based software provide sufficiently precise

orientation data for safe and effective programming.

Rotational fluoroscopy has additional practical value. It is widely available, can be

performed quickly after surgery, and does not require specialized software. It may

therefore be especially useful in cases where CT-based analysis is inconclusive,

unavailable, or requires corroboration.

Limitations

This study had several limitations. First, although this was a prospective study, it was

conducted at a single institution, with all surgeries performed by the same

neurosurgeon using a single imaging system and a single lead model (Boston

Scientific Vercise Cartesia™). The generalizability to other hardware platforms or

vendors may be limited. Second, the Iron Sight method, although reproducible in our

hands, may depend on the observer's experience and fluoroscopic image quality;

training and standardization could impact its reliability for broader use. Furthermore,

we compared the Iron Sight method with Brainlab Elements™, which itself may not

represent a perfect ground truth. Our reported agreement therefore reflects method

concordance rather than absolute accuracy.

Conclusion

In conclusion, the iron-sight method, when applied to postoperative rotational fluoroscopy, offers excellent interobserver reliability for determining directional DBS lead orientation and demonstrates good agreement with Brainlab Elements™ CT-based estimations. These findings support the clinical utility of rotational fluoroscopy as a reliable imaging modality for orientation assessment, particularly in settings in which software-based methods are unavailable, fail, or require corroboration. Future work should examine reproducibility across different institutions, hardware platforms, and surgical teams. Studies linking orientation accuracy to clinical outcomes such as programming thresholds, therapeutic windows, and stimulation-related side effects will help determine whether small differences in orientation measurement translate into meaningful differences for patients.

Acknowledgements:

We thank Jaslok Hospital and Research Institute for their generous support in funding and providing space for our intervention activities. We thank Dr. Manish Baldia and Dr. Neha Rai for their assistance in setting up the initial protocol and coordinating its implementation.

Autor Roles:

1. Research project: A. Conception, B. Organization, C. Execution;

2. Statistical Analysis: A. Design, B. Execution, C. Review and Critique;

3. Manuscript Preparation: A. Writing of the first draft, B. Review and Critique;

Doshi P: 1A, 1B, 2A, 2C, 3B

Karnavat C: 1C, 2B, 3B

Agarbattiwala R: 1B, 1C, 2A, 2B, 2C, 3A, 3B

Funding Source: This work was supported by a grant from Jaslok Hospital and

Research Center, Mumbai.

Role of the Funder/Sponsor:

The funding source had no role in the design and conduct of the study; the

collection, management, analysis, and interpretation of the data; the preparation,

review, or approval of the manuscript; or the decision to submit the manuscript for

publication.

Conflict of Interest:

None.

Financial Disclosures for the previous 12 months:

The authors declare that there are no additional disclosures to report.

Ethical Compliance Statement:

• The study was conducted in accordance with the Declaration of Helsinki and

was approved by the Institutional Ethics Committee at our center (approval

number EC/1114/2022, dated April 29, 2022).

• Written consents were obtained from all the patients.

• We confirm that we have read the Journal's position on issues involved in

ethical publication and affirm that this work is consistent with those guidelines.

Data Sharing Statement:

The data supporting this study's findings were anonymized and are available in the

Supplementary Data.

Supplementary Data

Appendix	This document describes the mathematical formulation to calculate bias, limits of agreement (LoA), and confidence intervals in a Bland-Altman analysis for circular data
Supplementary Tables	Table A lists imaging acquisition parameters, Table B outlines the ICC calculation parameters for interobserver agreement in rotational fluoroscopy, Table C details the parameters for assessing agreement between the methods.
Supplementary Data	Angle measurement conventions and approaches by the iron-sight method
Anonymized data	Angular data used for inter-rater agreement for the iron-sight method, comparison between the two methods, and regression analysis

References

1. Benabid AL, Pollak P, Gao D, Hoffmann D, Limousin P, Gay E, et al. Chronic

electrical stimulation of the ventralis intermedius nucleus of the thalamus as a

treatment of movement disorders. J Neurosurg. 1996 Feb;84(2):203–14.

2. Deep-Brain Stimulation of the Subthalamic Nucleus or the Pars Interna of the

Globus Pallidus in Parkinson's Disease. N Engl J Med. 2001 Sept 27;345(13):956–

63.

3. Deuschl G, Schade-Brittinger C, Krack P, Volkmann J, Schäfer H, Bötzel K, et al. A

Randomized Trial of Deep-Brain Stimulation for Parkinson's Disease. N Engl J Med.

2006 Aug 31;355(9):896–908.

4. Vidailhet M, Vercueil L, Houeto JL, Krystkowiak P, Benabid AL, Cornu P, et al.

Bilateral deep-brain stimulation of the globus pallidus in primary generalized

dystonia. N Engl J Med. 2005 Feb 3;352(5):459–67.

5. Butson CR, McIntyre CC. Current steering to control the volume of tissue activated

during deep brain stimulation. Brain Stimulat. 2008 Jan;1(1):7–15.

6. Pollo C, Kaelin-Lang A, Oertel MF, Stieglitz L, Taub E, Fuhr P, et al. Directional deep

brain stimulation: an intraoperative double-blind pilot study. Brain. 2014

July;137(7):2015–26.

7. Steigerwald F, Müller L, Johannes S, Matthies C, Volkmann J. Directional deep brain

stimulation of the subthalamic nucleus: A pilot study using a novel neurostimulation

device. Mov Disord. 2016 Aug;31(8):1240–3.

8. Schlaepfer TE, Bewernick BH, Kayser S, Mädler B, Coenen VA. Rapid effects of

deep brain stimulation for treatment-resistant major depression. Biol Psychiatry.

2013 June 15;73(12):1204–12.

9. Coenen VA, Rijntjes M, Prokop T, Piroth T, Amtage F, Urbach H, et al. One-pass

deep brain stimulation of dentato-rubro-thalamic tract and subthalamic nucleus for

tremor-dominant or equivalent type Parkinson's disease. Acta Neurochir (Wien).

2016 Apr;158(4):773–81.

10. Coenen VA, Allert N, Paus S, Kronenbürger M, Urbach H, Mädler B. Modulation of

the Cerebello-Thalamo-Cortical Network in Thalamic Deep Brain Stimulation for

Tremor: A Diffusion Tensor Imaging Study. Neurosurgery. 2014 Dec;75(6):657–70.

11. Coenen VA, Allert N, Mädler B. A role of diffusion tensor imaging fiber tracking in

deep brain stimulation surgery: DBS of the dentato-rubro-thalamic tract (drt) for the

treatment of therapy-refractory tremor. Acta Neurochir (Wien). 2011

Aug;153(8):1579–85.

12. Reinacher PC, Krüger MT, Coenen VA, Shah M, Roelz R, Jenkner C, et al.

Determining the Orientation of Directional Deep Brain Stimulation Electrodes Using

3D Rotational Fluoroscopy. Am J Neuroradiol. 2017 June;38(6):1111–6.

13. Hunsche S, Neudorfer C, Majdoub FE, Maarouf M, Sauner D. Determining the

Rotational Orientation of Directional Deep Brain Stimulation Leads Employing Flat-

Panel Computed Tomography. Oper Neurosurg. 2019 Apr;16(4):465–70.

14. Sitz A, Hoevels M, Hellerbach A, Gierich A, Luyken K, Dembek TA, et al.

Determining the orientation angle of directional leads for deep brain stimulation

using computed tomography and digital x-ray imaging: A phantom study. Med Phys.

2017 Sept;44(9):4463–73.

15. Doshi PK, Chhaya NA, Bhatt MA. Bilateral subthalamic nucleus stimulation for

Parkinson's disease. Neurol India. 2003 Mar;51(1):43–8.

16. Fleiss JL. The Design and Analysis of Clinical Experiments [Internet]. 1st ed. Wiley;

1999 [cited 2025 May 23]. Available from:

https://onlinelibrary.wiley.com/doi/book/10.1002/9781118032923

17. Zou GY. Sample size formulas for estimating intraclass correlation coefficients with

precision and assurance. Stat Med. 2012 Dec 20;31(29):3972–81.

18. Liljequist D, Elfving B, Skavberg Roaldsen K. Intraclass correlation – A discussion

and demonstration of basic features. Chiacchio F, editor. PLOS ONE. 2019 July

22;14(7):e0219854.

19. Hellerbach A, Dembek TA, Hoevels M, Holz JA, Gierich A, Luyken K, et al. DiODe:

Directional Orientation Detection of Segmented Deep Brain Stimulation Leads: A

Sequential Algorithm Based on CT Imaging. Stereotact Funct Neurosurg.

2018;96(5):335–41.

20. Berens P. **CircStat**: A*MATLAB*Toolbox for Circular Statistics. J Stat Softw [Internet].

2009 [cited 2025 May 23];31(10). Available from: http://www.jstatsoft.org/v31/i10/

21. Fisher NI, Lee AJ. A correlation coefficient for circular data. Biometrika.

1983;70(2):327–32.

22. Shrout PE, Fleiss JL. Intraclass correlations: Uses in assessing rater reliability.

Psychol Bull. 1979;86(2):420–8.

23. Bland JM, Altman DG. Statistical methods for assessing agreement between two

methods of clinical measurement. Lancet Lond Engl. 1986 Feb 8;1(8476):307–10.

24. Gerke O. Reporting Standards for a Bland–Altman Agreement Analysis: A Review of

Methodological Reviews. Diagnostics. 2020 May 22;10(5):334.

25. Preiss D, Fisher J. A Measure of Confidence in Bland–Altman Analysis for the

Interchangeability of Two Methods of Measurement. J Clin Monit Comput. 2008

Aug;22(4):257–9.

26. Fisher NI. Statistical Analysis of Circular Data [Internet]. 1st ed. Cambridge

University Press; 1993 [cited 2025 May 23]. Available from:

https://www.cambridge.org/core/product/identifier/9780511564345/type/book

27. Matias CM, Mehanna R, Cooper SE, Amit A, Lempka SF, Silva D, et al. Correlation

Among Anatomic Landmarks, Location of Subthalamic Deep Brain Stimulation

Electrodes, Stimulation Parameters, and Side Effects During Programming

Monopolar Review. Oper Neurosurg. 2015 Mar;11(1):99–109.

28. Benjamini Y, Hochberg Y. Controlling the False Discovery Rate: A Practical and

Powerful Approach to Multiple Testing. J R Stat Soc Ser B Stat Methodol. 1995 Jan

1;57(1):289–300.

29. Fayed I, Syed M, Gingold E, Alizadeh M, Sharan A, Wu C. A Novel and Simple

Method Using Computed Tomography Streak Artifact to Determine the Orientation of

Directional Deep Brain Stimulation Leads. Neurosurgery. 2023 Nov;93(5):1036–45.

30. Van Dijk KJ, Verhagen R, Bour LJ, Heida C, Veltink PH. Avoiding Internal Capsule

Stimulation With a New Eight-Channel Steering Deep Brain Stimulation Lead.

Neuromodulation Technol Neural Interface. 2018 Aug;21(6):553–61.

Table 1
508
509   Table 1. Descriptive Statistics of Directional DBS Lead Orientations Estimated by

510   Observers and Brainlab Elements™

Measurements	circular mean	Circular Standard

	(angular range) (in degrees)	Deviation (in degrees)
Angles measured by observer 1 (Neurosurgeon) (n = 70)	5.55° (-64° to 42°)	16.89°
Angles measured by observer 2 (Radiologist) (n = 70)	4.89° (-67° to 39°)	16.36°
Angles measured by Brainlab Element™ (n = 67)	8.27° (-29° to 46°)	13.51°
Trajectories Coronal Angle (n = 70)	24.24° (15.64° to 36.45°)	4.33°
Trajectories Sagittal Angle (n = 70)	25.84° (11.31° to 41.15°)	6.38°
Trajectories Axial Angle (n = 70)	43.38° (21.21° to 70.45°)	10.09°
Trajectories Azimuth Angle (n = 70)	33.95° (24.48° to 46.74°)	4.57°

511   Descriptive summary of lead orientation measurements from postoperative rotational

512   fluoroscopy and CT in 70 directional leads implanted at our center (2022–2024).

513   Circular means, ranges, and standard deviations are provided for observer

514   estimates using the iron-sight method and Brainlab Elements™ software.

Figures:

Figure 1. Polar histograms of directional DBS lead orientations measured with the

Iron Sight method from postoperative rotational fluoroscopy at Jaslok Hospital,

Mumbai (2022-2024). Panel A-D: Orientation estimates from each observer (n=70

leads). Panel E: Distribution of angular differences between observers, showing a

symmetric pattern consistent with a Gaussian distribution. Panel F: Polar histogram

of observer differences compared with the von Mises distribution. Clinically,

differences between observers were minimal, with 95% of measurements within –

5.49° to +6.85° (approximately ±7°).

Figure 2. Agreement analysis between neurosurgeon and radiologist estimates of

DBS lead orientation using the Iron Sight method on postoperative fluoroscopy

(2022-2024). Panel A: Bland–Altman plot showing a mean difference of 0.68° with

95% limits of agreement from –5.49° to +6.85°. Panel B: Preiss–Fisher test

confirming agreement was highly unlikely due to chance. Panel C: Intraclass

correlation coefficient (ICC(2,1) = 0.983, 95% CI 0.973–0.989) demonstrating

excellent interobserver reliability. In practice, this means two trained observers

almost always agreed within ~7° when using rotational fluoroscopy.

Figure 3. Comparison of the Iron Sight method and Brainlab Elements™ CT-based

software in patients (Jaslok Hospital, 2022-2024). Panel A-D: Orientation estimates

from both methods for 67 leads. Panel E: Distribution of angular differences

approximating a Gaussian shape. Panel F: Polar histogram of differences compared

with the von Mises distribution. On average, the two methods differed by less than

1°, with 95% of measurements within −14.86° to +13.48° (about ±15°).

Figure 4. Agreement between rotational fluoroscopy (Iron Sight method) and

Brainlab Elements™ software in determining orientation of directional DBS leads (67

leads, Jaslok Hospital, Mumbai, 2022-2024). Panel A: Bland–Altman plot showing a

mean bias of −0.69° and 95% limits of agreement from −14.86° to +13.48°. Panel B:

Preiss–Fisher test confirming concordance was not due to chance. Panel C:

Intraclass correlation coefficient (ICC(3,1) = 0.864, 95% CI 0.786–0.915) indicating

good agreement between the two approaches. For clinicians, this means that both

fluoroscopy and CT-based methods are broadly consistent, with most

measurements agreeing within ~15° — a margin corresponding to less than 1.5 mm

around the electrode circumference.

Figure 2A

Figure 2B

Figure 2C

Figure 3A. Rotational Fluorescopy (RF) method

Figure 3B. RF method circular mean = 0.13

Figure 3C. Brainlab Element™

Figure 3D. Brainlab Element™ circular mean = 0.14

Figure 3E. Histogram with Gaussian fit

Circular difference between methods in Radians

Figure 3F. Angular Histogram of circular differences with von Mises Fit

Figure 4A

Figure 4B

Figure 4C