Postoperative Determination of Directional Deep Brain Stimulation (DBS) Lead Orientation Using Rotational Fluoroscopy: Interobserver Agreement and Comparison with CT-Based Software

Authors:

Paresh K Doshi[*1], M.Ch.; Chandresh Karnavat[2], D.N.B.; Raj V Agarbattiwala[1] M.Ch.

Affiliations:

1. Stereotactic and Functional Neurosurgery Department, Jaslok Hospital and Research Center, Mumbai, Maharashtra, India

2. Radiology Department, Jaslok Hospital and Research Center, Mumbai, Maharashtra, India

*Corresponding Author: Prof. Paresh K Doshi

Address:

Jaslok Hospital and Research Center,

15, G Deshmukh Marg, Pedder Road, Mumbai-40026.

Phone Number: +912266573185

Email: pareshkd@gmail.com

Abstract Word Count: 347

Text Word Count: 3485

Number of References: 30

Number of Figures: 4

Number of Tables: 1

Number of Videos: 0

Keywords: Directional Leads, Deep Brain Stimulation, Circular Statistics, Lead Orientation, Iron Sight

Abstract

Introduction:

Commercial software platforms, such as Brainlab Elements™ utilize postoperative Computed Tomography (CT) to estimate deep brain stimulation (DBS) lead orientation. The iron-sight method on rotational fluoroscopy offers a potentially accurate and accessible alternative, but its interobserver agreement and validity in clinical settings remain unverified.

Objective:

To evaluate inter-observer agreement in estimating lead orientation using the iron-sight method on postoperative rotational fluoroscopy and to compare these estimates with Brainlab Elements™ from postoperative CT.

Methods:

This prospective study included 70 leads in patients who underwent DBS implantation with directional leads at a tertiary care center between 2022 and 2024. The inter-rater agreement of independent measurements obtained by two observers utilizing the iron-sight method was evaluated. Subsequently, these measurements were compared with those derived from the Brainlab Element™ software.

Results:

Bland–Altman analysis of measurements between observers revealed a mean bias of 0.68° (95% CI: –0.0678° to 1.4335°), with limits of agreement from –5.49° (CI: –6.79° to –4.19°) to 6.85° (CI: 5.55° to 8.15°). The Preiss–Fisher test confirmed that the agreement was statistically significant and unlikely to be due to chance. (Z = –15.78, $P < 0.001$).

In the method comparison (n = 67), the mean bias was –0.69° (95% CI: –2.48° to 1.10°), with limits of agreement ranging from –14.86° (95% CI: –17.97° to –11.76°) to 13.48° (95% CI: 10.37° to 16.58°). The Watson–Williams test showed no significant difference in circular means (p = 0.7905).

Exploratory regression suggested that coronal and polar lead angles modestly influenced interobserver variability, whereas trajectory angles did not affect method agreement.

Conclusion:

The iron-sight method is a reliable and reproducible tool for estimating directional lead orientation, with excellent interobserver agreement and good concordance with Elements™ estimates.

Introduction

Deep brain stimulation (DBS) is a recognized treatment for movement disorders like Parkinson's disease, tremors, and dystonia, as well as drug-resistant epilepsy and obsessive-compulsive disorder, the latter still being experimental [1–4]. The clinical scope of DBS continues to expand, with ongoing research exploring its role in psychiatric and neurodegenerative conditions including major depression, addiction, Tourette syndrome, Alzheimer's disease, and schizophrenia.

Despite decades of clinical use, the technology underlying DBS systems remained unchanged until the recent introduction of directional leads. Traditional cylindrical DBS electrodes produce approximately spherical volumes of tissue activation, often leading to the unintended stimulation of adjacent structures. The potential benefits of directing stimulation between contacts or in specific directions have long been anticipated [5]. Directional DBS leads to feature-segmented electrodes that allow clinicians to steer the electrical field, improve therapeutic precision, expand the therapeutic window, and reduce stimulation-induced side effects[6,7]. This steering is particularly beneficial in targeting areas such as the subthalamic nucleus (STN), where directional stimulation can avoid spreading to the internal capsule (IC) while enhancing the coverage of functionally relevant pathways.

Accurate knowledge of the final orientation of the segmented contacts is crucial for individualized stimulation planning [8–11]. Although electrodes are inserted with the intended orientation (typically anterior), intraoperative lead rotation and torsion can alter the final position. Despite the directional markers in existing commercial models, accurately determining the orientation of the lead continues to pose difficulties[12]. Consequently, postoperative imaging is essential to confirm lead orientation. Among various imaging techniques, postoperative computed tomography (CT) combined with preoperative magnetic resonance imaging (MRI) and interpretation using commercial software (e.g., Brainlab Elements™) is a common practice[13]. However, CT-based methods may lose accuracy when leads are positioned at oblique angles (>40°)[14].

Techniques such as the "Iron Sight" method, which exploits the radiographic appearance of segmented contacts to infer orientation, have demonstrated high accuracy in phantom models. Reinacher et al. reported a limit of agreement (LoA) of ±2.44° and an intraclass correlation coefficient (ICC) of 0.9998 using this method[12]. However, its inter-rater reliability in real-world clinical scenarios has not been sufficiently studied. Furthermore, the agreement between fluoroscopy-based estimates and software-derived CT orientations has not yet been systematically quantified.

Aim and Objectives

Aim

To evaluate the inter-rater agreement in determining directional DBS lead orientation using the Iron Sight method on postoperative 3D rotational fluoroscopy and to compare these observer-derived estimates with orientation measurements from Brainlab Elements™ software based on postoperative CT imaging.

Objectives

- To assess interobserver agreement between a neurosurgeon and a radiologist in determining directional DBS lead orientation using the iron-sight method on postoperative 3D rotational fluoroscopy.

- To compare orientation measurements obtained via the Iron Sight method with those derived from the Brainlab Elements™ software using postoperative CT.

- To assess the association of lead trajectory angles with measurement errors by the above methods

- The reliability and statistical uncertainty of the assessments were quantified using circular statistics and intraclass correlation.

Methods

Reporting in the article follows the Guidelines for Reporting Reliability and Agreement Studies (GRRAS) to ensure the full disclosure of design and statistical decisions. This prospective exploratory study was conducted at a high-volume tertiary care center between May 2022 and December 2024.

Participants

All consecutive patients who underwent DBS surgery with directional leads (Vercise Cartesia™, Boston Scientific) during the study period were screened for inclusion. The inclusion criteria were as follows.

- Implantation of directional DBS leads (Vercise Cartesia™, Boston Scientific) for movement disorders.

- The participant (or legally authorized representative) must have given consent.

Patients with incomplete imaging, significant motion artifacts, or postoperative complications affecting the lead visualization were excluded.

The study was conducted in accordance with the Declaration of Helsinki and was approved by the Institutional Ethics Committee at our center (approval number EC/1114/2022, dated April 29, 2022). Participants were recruited from among patients who underwent DBS surgery with directional leads (Vercise Cartesia™, Boston Scientific) between May 1, 2022, and December 31, 2024. All patients and caregivers received both verbal and written explanations of the study, and written consent was obtained from each participant (or their representative) for participation and secondary analysis of the de-identified data. Both CT and fluoroscopy were performed during the postoperative period after admission. All surgeries were conducted under the Stereotactic and Functional Neurosurgery department at our center by a single neurosurgeon (XX) according to the standard protocol[15]. CT scans were performed in the radiology department, and rotational fluoroscopy was performed in the Cath lab (see Table A of the Supplementary Table for the parameters).

Sample Size

Sample size estimation for comparing rater reliability or agreement has not been thoroughly explored. Fleiss proposed a rule of thumb that recommends 15–20 samples for reliability studies[16]. In the case of a continuous outcome, increasing the sample size beyond 12 per group did not significantly affect the confidence

interval. Reinacher et al. calculated a 95% confidence interval of ±2.44 ° in their study of phantom models. Using Zou's formula for ICC precision with 95% CI width ≤ 0.1, α = 0.05, ICC = 0.85, and 80% assurance, the required sample size was 57 leads for two raters[17]. We chose an expected ICC of 0.85 and a 95% CI width of ≤ 0.1 because ICC ≥ 0.75 suggests good reliability. Our study included 70 leads that met this threshold[18]. However, we included 70 leads to account for real-world variability in human data and to ensure robustness in statistical comparisons.

Postoperative Imaging and Orientation Measurement

Rotational Fluoroscopy (Iron Sight Method)

Within 48 hours postoperatively, all patients were brought to Cathlab for 3D rotational fluoroscopy. The acquisition parameters are listed in the Supplementary Table. Images were reconstructed using multiple projections and independently reviewed by two raters.

- A neurosurgeon (RA) experienced in DBS surgery and programming
- A radiologist (CK) trained in neuroimaging

The observers performed measurements independently on anonymized files and had no access to each other's results. Using the Iron Sight method, each observer independently estimated the rotational orientation of each lead relative to the anterior direction, expressed in degrees (−180° to +180°). (Supplementary Data).

Determination of lead orientation on CT scans

Postoperative CT scans were uploaded into Brainlab Elements™ (Lead Localization module). The acquisition parameters are listed in Table A of the Supplementary

Tables. The system automatically detected the lead axis and segmented the contact positions based on CT artifact patterns (as described by Hellerbach et al., 2018[19]), which were then visually confirmed. The lead model was set to Boston Scientific Vercise Cartesia™. The output orientation is recorded as the reference value for each lead. Given that Brainlab Elements™ itself has known limitations at oblique angles, this method comparison should be interpreted as assessing agreement between the two techniques rather than validating one against a true gold standard. No phantom-based ground truth data were available in this clinical setting.

Statistical Analysis

All statistical analyses were performed using MATLAB (MathWorks, Natick, MA, USA) and the Circular Statistics Toolbox[20]. Statistical significance was set *P-value < 0.05*.

Interobserver Agreement

To assess interobserver agreement using the Iron Sight method, orientation angles were treated as circular variables ranging from −180° to +180°. The circular difference, or shortest angular distance, was calculated for each lead between the two raters' measurements. Circular variance was computed to quantify the dispersion of the ratings, and the circular correlation coefficient[21] was used to assess the strength of association between the estimates of the neurosurgeon and radiologist. Because angular dispersion was low, we used ICC (2,1), as defined by Shrout and Fleiss (1979)[22], which is appropriate for two-way random effects

models assessing absolute agreement between two observers. This ICC formulation accounts for subject variability, rater variability, and residual errors.

A Bland–Altman analysis[23,24] adapted for circular data (Online Appendix) was conducted to further explore the agreement. This included calculating the mean angular difference (i.e., circular bias) and standard deviation of the differences. The 95% limits of agreement (LoA) were computed as the bias ± 1.96 of the standard deviation. To express the uncertainty in these estimates, 95% confidence intervals for both the bias and limits of agreement were calculated. Finally, the Preiss-Fisher procedure was applied to test whether the observed agreement could have occurred by chance[25].

Method Comparison (Ironsight method vs Brainlab™)

For method comparison between the iron-sight approach and Brainlab Elements™, the Watson–Williams test was used to compare the circular means derived from the observers' average ratings with the angles obtained from the Elements™ software[26]. The Intraclass Correlation Coefficient (ICC (3,1)) was calculated to quantify the consistency between the two methods, using a two-way mixed-effects model. Agreement between the iron-sight method and Brainlab™-derived orientation angles was also visualized using a Bland–Altman plot. The Preiss-Fisher test was again applied to confirm whether the observed concordance between the methods exceeded chance expectations. Before the parametric circular tests, assumptions were assessed and met, including angular homogeneity and concentration.

Regression Analysis

We performed two univariate linear regression analyses to investigate whether anatomical trajectory angles influenced the angular measurement differences. Linear regression of the coronal, sagittal, polar, and azimuthal angles (as explained by Matias et al.[27]) against the angular difference between the two observers' measurements on rotational fluoroscopy (i.e., interobserver error). A separate regression of these angles against the angular differences between rotational fluoroscopy and Brainlab Elements™ (i.e., method discrepancy).

For each regression model, beta coefficients and the corresponding p-values were calculated to assess the strength and significance of the association between the variables. Owing to multiple comparisons, we adjusted the p-values (P_{adj}) for the false discovery rate using the Benjamini-Hochberg method[28]. Statistical significance was set at $P < 0.05$.

Results

Study Sample

Thirty-five patients were included in the final analysis, yielding 70 directional DBS leads for inter-observer agreement analysis. For the method comparison analysis, five leads were excluded because the Brainlab Elements™ software failed to compute the lead orientation, resulting in 67 leads being analyzed. An overview of

the data, including circular means, ranges, and standard deviations for all measurements, is presented in Table 1.

Interobserver Agreement: Iron Sight Method on Rotational Fluoroscopy

The orientations estimated by the two observers were visualized using polar histograms (Figure 1A-D). The Rayleigh test for uniformity demonstrated significant data clustering ($P < 0.05$), indicating that the angles were not randomly distributed around the circle, but were concentrated around a common direction.

The mean resultant length of the circular distance between the two observers was 0.9985 with a circular variance of 0.0015, indicating highly consistent measurements. The circular correlation coefficient was 0.9822, confirming a strong directional agreement. A Circular Bland–Altman analysis (figure 2A) revealed a mean bias of 0.68° (95% CI: –0.0678° to 1.4335°). The upper limit of agreement (LoA) was 6.85° (95% CI: 5.55° to 8.15°) and the lower LoA was –5.49° (95% CI: –6.79° to 4.19°). There was no evidence of proportional bias ($P > 0.05$), suggesting that the measurement errors did not change with angle. The Preiss–Fisher test showed that the standard deviation of differences between two observer measurements yielded a Z-value of –15.40 (Figure 2B), indicating that the observed agreement was significantly greater than would be expected by chance. The Intraclass Correlation Coefficient (ICC (2,1)) was 0.983 (95% CI: 0.973–0.989), indicating excellent inter-observer reliability (Figure 2C and Table B of the Supplementary Tables for calculation parameters).

Figure 1E shows that the histogram of the circular distances between observers exhibited a symmetric distribution, approximating a Gaussian shape. Similarly, the polar histogram of angular differences was visually compared with the von Mises distribution (Figure 1F) and showed good conformity.

Univariate linear regression analysis revealed that the coronal angle (β = 0.1803, *P-value = 0.0367*; *adjusted P_{adj} = 0.073*) and polar angle (β = 0.0792, *P-value = 0.0320*; *adjusted P_{adj} = 0.073*) were associated with greater interobserver differences. Note that after controlling for the false discovery rate using the Benjamini–Hochberg procedure, these associations did not remain statistically significant at the threshold of 0.05. The sagittal angle (β = –0.0588, *P-value = 0.3212*; *adjusted P_{adj} = 0.428*) and azimuthal angle (β = 0.0364, *P-value = 0.6618*; *adjusted P_{adj} = 0.662*) showed no significant association with interobserver variation.

Method Comparison: Iron Sight Method vs. Brainlab Elements™

Of the 70 leads, 67 were eligible for method comparison, as the software failed to estimate the orientation in the three leads. Orientation estimates from both methods are displayed using polar histograms (figure 3A-D), including circular means. The Rayleigh test again showed significant nonuniformity (*P < 0.05*), indicating that the data points were directionally concentrated.

The mean resultant length of the circular distance between the iron-sight method and the Brainlab™-derived orientation was 0.9922, with a circular variance of 0.0078. The circular correlation coefficient between the two methods was 0.8635. Bland–Altman analysis (figure 4A) revealed a mean bias of –0.69° (95% CI: –2.4839° to 1.0993°), with an upper LoA of 13.48° (95% CI: 10.37° to 16.58°) and a lower LoA of –14.86° (95% CI: –17.97° to –11.76°). No significant proportional bias was observed ($P > 0.05$), and the Preiss–Fisher Z-value was –10.83 (Figure 4B), again indicating that agreement between methods was not attributable to chance. The Intraclass Correlation Coefficient (ICC (3,1)) for the agreement between the Iron Sight method and Brainlab Elements™ was 0.864 (95% CI: 0.786–0.915) (Figure 4C and Table C of the Supplementary Tables for calculation parameters), suggesting good consistency between the two approaches.

Figure 3E shows that the histogram of the circular distances between the methods approximates a Gaussian distribution. Similarly, the polar histogram of angular differences showed comparable clustering and was visually comparable to the von Mises distribution (Figure 3F). The Watson–Williams test yielded a *P-value* of *0.7905 (>0.05)*, indicating no statistically significant difference in the circular mean orientations between the two methods.

None of the lead trajectory angles demonstrated a statistically significant relationship with the angular difference between the fluoroscopy-derived and Brainlab-derived orientations. The beta coefficients for the coronal, sagittal, polar, and azimuthal

angles were 0.2546 (*P-value = 0.2374, adjusted P_{adj} = 0.7046*), 0.0315 (*P-value = 0.8073, adjusted P_{adj} = 0.8073*), 0.0217 (*P-value = 0.7976, adjusted P_{adj} = 0.8073*), and 0.1688 (*P-value = 0.3523, adjusted P_{adj} = 0.7046*), respectively.

Discussion

This study evaluated the reliability of the Iron Sight method for determining the orientation of directional DBS leads using postoperative rotational fluoroscopy and compared it with CT-based orientation estimates from the Brainlab Elements™ software. To our knowledge, this is among the first clinical studies to assess the interobserver agreement of the Iron Sight method in a real-world patient population using circular statistical methods and to systematically compare these estimates with those of a commercial CT-based platform (Brainlab Elements™).

The Iron Sight method demonstrated excellent interobserver reliability, with an intraclass correlation coefficient (ICC (2,1)) of 0.983, circular correlation of 0.9822, and circular variance of only 0.0015. The mean angular bias between observers was less than 1°, with narrow limits of agreement and no evidence of proportional bias. These findings suggest that trained clinicians can consistently estimate lead orientation using rotational fluoroscopy with the iron-sight pattern. This real-world evidence supports previous phantom-based work by Reinacher et al., who reported a similar accuracy of the Iron Sight method with a ±2.44° limit of agreement and ICC of 0.9998 in controlled settings[12].

When comparing the Iron Sight method-based estimates to orientation values derived from Brainlab Elements™, we observed good agreement (ICC (3,1) = 0.864), although the limits of agreement were broader (±13–14°). This deviation may reflect intrinsic differences in the measurement principles. While the Iron Sight method relies on direct visual alignment of segmented electrode gaps, Element™ estimates depend on artifact symmetry within CT slices—a technique known to lose accuracy when leads are inserted at angles exceeding 40° relative to the scanner axis[14,29]. Nevertheless, the mean bias between methods was minimal (–0.69°), and the Watson–Williams test confirmed no statistically significant difference in circular means (p = 0.7905), suggesting that both methods produce broadly comparable estimates under optimal imaging conditions.

We examined whether the lead trajectory, defined by coronal, sagittal, polar, and azimuthal angles, affected interobserver variability in orientation estimation. Coronal and polar angles showed significant associations with observer differences in univariate regression; however, these associations did not remain significant after correction for the false discovery rate using the Benjamini-Hochberg method[28]. This suggests that greater lateral deviation or off-vertical tilt may increase variability in interpreting the iron-sight artifact.

No trajectory angle was significantly associated with discrepancies between fluoroscopy- and CT-based orientation estimates, indicating that both methods were relatively robust to moderate angular deviations in clinical practice. Although CT-

based algorithms may lose accuracy at oblique angles above 40°[14], our findings support the reliability of rotational fluoroscopy across a range of implant trajectories and its potential role as a complementary tool in anatomically challenging cases.

In the finite element model and multicompartment neuron simulations by van Dijk et al.[30], displacement of the DBS lead by only 1 mm from the center of the STN resulted in a substantial reduction in STN cell activation and an increased risk of IC fiber stimulation when using conventional cylindrical leads. In contrast, the directional lead could partially compensate for this 1 mm misplacement, particularly when oriented such that one or more contacts faced away from the IC.

Importantly, the study also found that orientation played a critical role; with the lead placed at the center of the STN, the percentage of activated STN cells varied from 53.7% to 75.1% across different rotational orientations of the same lead. This implies that a misorientation of ~45° could meaningfully reduce efficacy or increase the risk of side effects. In contrast, our interobserver error of <7° and method agreement error of ~15° are small relative to the 45° increments between the modeled orientations, suggesting that the observed variation in our study falls within a range where the clinical impact may be limited or manageable through programming.

The clinical relevance of the orientation accuracy in directional leads can be better appreciated by mapping our observed angular error margins onto the spatial tolerances reported in computational modeling studies. In our study, the inter-

observer agreement showed LoA of ±6.85°, and the method comparison with Brainlab Elements™ yielded an LoA of ±14.86°. Assuming a 6 mm radial distance from the lead axis (consistent with the modeling parameters used by van Dijk et al.), a ±14.86° angular discrepancy corresponds to a circumferential displacement of approximately 1.56 mm on the surface of the contact ring. Similarly, a ±6.85° error corresponds to a displacement of roughly 0.72 mm. Our results suggest that both observer-derived (Iron Sight) and CT-based (Brainlab Elements™) orientation estimates fall within the clinically acceptable error margins, as modeled by van Dijk et al. This supports the practical utility of both methods in guiding directional DBS programming, particularly in cases where orientation-sensitive effects, such as capsular side effects, are a concern.

Limitations

This study had several limitations. First, although this was a prospective study, it was conducted at a single institution, with all surgeries performed by the same neurosurgeon using a single imaging system and a single lead model (Boston Scientific Vercise Cartesia™). The generalizability to other hardware platforms or vendors may be limited. Second, the Iron Sight method, although reproducible in our hands, may depend on the observer's experience and fluoroscopic image quality; training and standardization could impact its reliability for broader use. Finally, our method comparison used Brainlab Elements™ as a reference, which itself may have errors under certain conditions. Thus, the reported bias and LoA should be interpreted as reflecting method concordance rather than ground-truth accuracy.

Conclusion

In conclusion, the iron-sight method, when applied to postoperative rotational fluoroscopy, offers excellent interobserver reliability for determining directional DBS lead orientation and demonstrates good agreement with Brainlab Elements™ CT-based estimations. These findings support the clinical utility of rotational fluoroscopy as a reliable imaging modality for orientation assessment, particularly in settings in which software-based methods are unavailable, fail, or require corroboration. Future research should examine the potential correlation between iron-sight-based orientation estimates and stimulation thresholds or side effect profiles. Furthermore, subsequent studies should evaluate reproducibility across various institutions and platforms as well as assess the influence of precise orientation estimation on the long-term outcomes of deep brain stimulation.

Acknowledgements:

This section has been removed to anonymize the author and center information.

Autor Roles:

Conceptualization, Project Administration, Supervision, and Funding Acquisition: Doshi P.

Writing - Original Draft, Formal Analysis, Validation, Software, and Visualisation: Agarbattiwala R.

Data Curation, Methodology, Investigations, Resources, and Writing- Review & Editing: All Authors.

Funding Source: This work was supported by a grant from Jaslok Hospital and Research Center, Mumbai.

Role of the Funder/Sponsor:

The funding source had no role in the design and conduct of the study, the collection, management, analysis, and interpretation of the data, the preparation, review, or approval of the manuscript, or the decision to submit the manuscript for publication.

Conflict of Interest:

None.

Data Sharing Statement:

The data supporting this study's findings were anonymized and are available in the Supplementary Data (Anonymized Data.xlsx).

Supplementary Data

Appendix	This document describes the mathematical formulation to calculate bias, limits of agreement (LoA), and confidence intervals in a Bland-Altman analysis for circular data
Supplementary Tables	Table A lists imaging acquisition parameters, Table B outlines the ICC calculation parameters for interobserver agreement in rotational fluoroscopy, Table C details the parameters for assessing agreement between the methods.
Supplementary Data	Angle measurement conventions and approaches by the iron-sight method
Anonymized data	Angular data used for inter-rater agreement for the iron-sight method, comparison between the two methods, and regression analysis

References

[1]   Benabid A L, Pollak P, Gao D, Hoffmann D, Limousin P, Gay E, Payen I and Benazzouz A 1996 Chronic electrical stimulation of the ventralis intermedius

nucleus of the thalamus as a treatment of movement disorders *J. Neurosurg.* **84** 203–14

[2] Anon 2001 Deep-Brain Stimulation of the Subthalamic Nucleus or the Pars Interna of the Globus Pallidus in Parkinson's Disease *N. Engl. J. Med.* **345** 956–63

[3] Deuschl G, Schade-Brittinger C, Krack P, Volkmann J, Schäfer H, Bötzel K, Daniels C, Deutschländer A, Dillmann U, Eisner W, Gruber D, Hamel W, Herzog J, Hilker R, Klebe S, Kloß M, Koy J, Krause M, Kupsch A, Lorenz D, Lorenzl S, Mehdorn H M, Moringlane J R, Oertel W, Pinsker M O, Reichmann H, Reuß A, Schneider G-H, Schnitzler A, Steude U, Sturm V, Timmermann L, Tronnier V, Trottenberg T, Wojtecki L, Wolf E, Poewe W and Voges J 2006 A Randomized Trial of Deep-Brain Stimulation for Parkinson's Disease *N. Engl. J. Med.* **355** 896–908

[4] Vidailhet M, Vercueil L, Houeto J-L, Krystkowiak P, Benabid A-L, Cornu P, Lagrange C, Tézenas du Montcel S, Dormont D, Grand S, Blond S, Detante O, Pillon B, Ardouin C, Agid Y, Destée A, Pollak P and French Stimulation du Pallidum Interne dans la Dystonie (SPIDY) Study Group 2005 Bilateral deep-brain stimulation of the globus pallidus in primary generalized dystonia *N. Engl. J. Med.* **352** 459–67

[5] Butson C R and McIntyre C C 2008 Current steering to control the volume of tissue activated during deep brain stimulation *Brain Stimulat.* **1** 7–15

[6] Pollo C, Kaelin-Lang A, Oertel M F, Stieglitz L, Taub E, Fuhr P, Lozano A M, Raabe A and Schüpbach M 2014 Directional deep brain stimulation: an intraoperative double-blind pilot study *Brain* **137** 2015–26

[7] Steigerwald F, Müller L, Johannes S, Matthies C and Volkmann J 2016 Directional deep brain stimulation of the subthalamic nucleus: A pilot study using a novel neurostimulation device *Mov. Disord.* **31** 1240–3

[8] Schlaepfer T E, Bewernick B H, Kayser S, Mädler B and Coenen V A 2013 Rapid effects of deep brain stimulation for treatment-resistant major depression *Biol. Psychiatry* **73** 1204–12

[9] Coenen V A, Rijntjes M, Prokop T, Piroth T, Amtage F, Urbach H and Reinacher P C 2016 One-pass deep brain stimulation of dentato-rubro-thalamic tract and subthalamic nucleus for tremor-dominant or equivalent type Parkinson's disease *Acta Neurochir. (Wien)* **158** 773–81

[10] Coenen V A, Allert N, Paus S, Kronenbürger M, Urbach H and Mädler B 2014 Modulation of the Cerebello-Thalamo-Cortical Network in Thalamic Deep Brain Stimulation for Tremor: A Diffusion Tensor Imaging Study *Neurosurgery* **75** 657–70

[11] Coenen V A, Allert N and Mädler B 2011 A role of diffusion tensor imaging fiber tracking in deep brain stimulation surgery: DBS of the dentato-rubro-thalamic tract

(drt) for the treatment of therapy-refractory tremor *Acta Neurochir. (Wien)* **153** 1579–85

[12] Reinacher P C, Krüger M T, Coenen V A, Shah M, Roelz R, Jenkner C and Egger K 2017 Determining the Orientation of Directional Deep Brain Stimulation Electrodes Using 3D Rotational Fluoroscopy *Am. J. Neuroradiol.* **38** 1111–6

[13] Hunsche S, Neudorfer C, Majdoub F E, Maarouf M and Sauner D 2019 Determining the Rotational Orientation of Directional Deep Brain Stimulation Leads Employing Flat-Panel Computed Tomography *Oper. Neurosurg.* **16** 465–70

[14] Sitz A, Hoevels M, Hellerbach A, Gierich A, Luyken K, Dembek T A, Klehr M, Wirths J, Visser□Vandewalle V and Treuer H 2017 Determining the orientation angle of directional leads for deep brain stimulation using computed tomography and digital x□ray imaging: A phantom study *Med. Phys.* **44** 4463–73

[15] Doshi P K, Chhaya N A and Bhatt M A 2003 Bilateral subthalamic nucleus stimulation for Parkinson's disease *Neurol. India* **51** 43–8

[16] Fleiss J L 1999 *The Design and Analysis of Clinical Experiments* (Wiley)

[17] Zou G Y 2012 Sample size formulas for estimating intraclass correlation coefficients with precision and assurance *Stat. Med.* **31** 3972–81

[18] Liljequist D, Elfving B and Skavberg Roaldsen K 2019 Intraclass correlation – A discussion and demonstration of basic features ed F Chiacchio *PLOS ONE* **14** e0219854

[19] Hellerbach A, Dembek T A, Hoevels M, Holz J A, Gierich A, Luyken K, Barbe M T, Wirths J, Visser-Vandewalle V and Treuer H 2018 DiODe: Directional Orientation Detection of Segmented Deep Brain Stimulation Leads: A Sequential Algorithm Based on CT Imaging *Stereotact. Funct. Neurosurg.* **96** 335–41

[20] Berens P 2009 **CircStat**: A *MATLAB* Toolbox for Circular Statistics *J. Stat. Softw.* **31**

[21] Fisher N I and Lee A J 1983 A correlation coefficient for circular data *Biometrika* **70** 327–32

[22] Shrout P E and Fleiss J L 1979 Intraclass correlations: Uses in assessing rater reliability. *Psychol. Bull.* **86** 420–8

[23] Bland J M and Altman D G 1986 Statistical methods for assessing agreement between two methods of clinical measurement *Lancet Lond. Engl.* **1** 307–10

[24] Gerke O 2020 Reporting Standards for a Bland–Altman Agreement Analysis: A Review of Methodological Reviews *Diagnostics* **10** 334

[25] Preiss D and Fisher J 2008 A Measure of Confidence in Bland–Altman Analysis for the Interchangeability of Two Methods of Measurement *J. Clin. Monit. Comput.* **22** 257–9

[26] Fisher N I 1993 *Statistical Analysis of Circular Data* (Cambridge University Press)

[27] Matias C M, Mehanna R, Cooper S E, Amit A, Lempka S F, Silva D, Carlotti C G, Butler R S and Machado A G 2015 Correlation Among Anatomic Landmarks, Location of Subthalamic Deep Brain Stimulation Electrodes, Stimulation Parameters, and Side Effects During Programming Monopolar Review *Oper. Neurosurg.* **11** 99–109

[28] Benjamini Y and Hochberg Y 1995 Controlling the False Discovery Rate: A Practical and Powerful Approach to Multiple Testing *J. R. Stat. Soc. Ser. B Stat. Methodol.* **57** 289–300

[29] Fayed I, Syed M, Gingold E, Alizadeh M, Sharan A and Wu C 2023 A Novel and Simple Method Using Computed Tomography Streak Artifact to Determine the Orientation of Directional Deep Brain Stimulation Leads *Neurosurgery* **93** 1036–45

[30] Van Dijk K J, Verhagen R, Bour L J, Heida C and Veltink P H 2018 Avoiding Internal Capsule Stimulation With a New Eight-Channel Steering Deep Brain Stimulation Lead *Neuromodulation Technol. Neural Interface* **21** 553–61

Table 1

Table 1. Descriptive Statistics of Directional DBS Lead Orientations Estimated by Observers and Brainlab Elements™

Measurements	circular mean (angular range) (in degrees)	Circular Standard Deviation (in degrees)
Angles measured by observer 1 (Neurosurgeon) (n = 70)	5.55° (-64° to 42°)	16.89°
Angles measured by observer 2 (Radiologist) (n = 70)	4.89° (-67° to 39°)	16.36°
Angles measured by Brainlab Element™ (n = 67)	8.27° (-29° to 46°)	13.51°
Trajectories Coronal Angle (n = 70)	24.24° (15.64° to 36.45°)	4.33°
Trajectories Sagittal Angle (n = 70)	25.84° (11.31° to 41.15°)	6.38°

Trajectories Axial Angle (n = 70)	43.38° (21.21° to 70.45°)	10.09°
Trajectories Azimuth Angle (n = 70)	33.95° (24.48° to 46.74°)	4.57°

Descriptive summary of lead orientation measurements from postoperative rotational fluoroscopy and CT in 70 directional leads implanted at our center (2022–2024). Circular means, ranges, and standard deviations are provided for observer estimates using the iron-sight method and Brainlab Elements™ software.

Figures:

Figure 1. Polar histograms and distribution plots of directional DBS lead orientations estimated via the Iron Sight method from postoperative rotational fluoroscopy at Jaslok Hospital, Mumbai (2022-2024). A-D: Observer-specific orientation estimates (n=70 leads). E: Histogram of interobserver angular differences with Gaussian Fit. F: Polar histogram of interobserver angular differences fitted with von Mises distribution.

Figure 2. Agreement analysis between neurosurgeon and radiologist estimates of DBS lead orientation using the Iron Sight method on postoperative fluoroscopy (2022-2024). A: Bland-Altman plot. B: Preiss-Fisher significance test of agreement. C: Intraclass correlation coefficient (ICC (2,1)) for interobserver agreement in 70 leads implanted at Jaslok Hospital, Mumbai

Figure 3. Polar histograms of DBS lead orientation estimates obtained from Iron Sight method and Brainlab Elements™ CT-based software in patients (Jaslok Hospital, 2022-2024). A-D: Observer average vs. software estimates (n=67 leads).

E: Histogram of method-wise circular differences. F: Polar histogram showing method discrepancy fitted with von Mises distribution

Figure 4. Agreement between rotational fluoroscopy (Iron Sight method) and Brainlab Elements™ software in determining orientation of directional DBS leads (67 leads, Jaslok Hospital, Mumbai, 2022-2024). A: Bland-Altman plot. B: Preiss-Fisher test for agreement. C: ICC (3,1) for inter-method reliability.

Figure 4A

Figure 4B

Figure 4C

Figure 3A. Rotational Fluorescopy (RF) method

Figure 3B. RF method circular mean = 0.13

Figure 3C. Brainlab Element™

Figure 3D. Brainlab Element™ circular mean = 0.14

Figure 3E. Histogram with Gaussian fit

Figure 3F. Angular Histogram of circular differences with von Mises Fit

Figure 2A

Figure 2B

Figure 2C