Title: Should antigen-antibody rapid diagnostic tests be used to detect acute HIV infection? A
systematic review and meta-analysis of diagnostic performance

Authors: Emmanuel Fajardo[1], Céline Lastrucci[1], Pascal Jolivet[1], Magdalena DiChiara[1], Carlota
Baptista da Silva[1], Busi Msimanga[1], Anita Sands[2], Cheryl Johnson[1][§]

Authors' Affiliations
[1]Global HIV, Tuberculosis, Hepatitis, and STI Department, World Health Organization, Geneva,
Switzerland
[2]Regulation and Prequalification, World Health Organization, Geneva, Switzerland
10

[§]**Corresponding author:**

Cheryl Johnson
Global HIV, Tuberculosis, Hepatitis, and STI Department
World Health Organization (WHO)
Avenue Appia 20
1211 Geneva
Switzerland
Phone number: (+47) 787 489 404
Email. johnsonc@who.int
20

E-mail addresses of authors
Emmanuel Fajardo: fajardo.emmanuel@gmail.com ORCID: 0000-0002-7561-3368
Céline Lastrucci: lastruccic@who.int ORCID:
Pascal Jolivet: jolivetp@who.int ORCID:
Magdalena DiChiara: barrdichiaram@who.int ORCID:
Carlota Baptista da Silva: carlota.baptista.silva@gmail.com ORCID:0000-0003-2782-4184
Busi Msimanga: mmisangaradebeb@who.int ORCID:
Anita Sands: sandsa@who.int ORCID: 0000-0002-8424-754X
Cheryl Johnson: johnsonc@who.int ORCID: 0000-0001-5499-5523
30

31   **Key words:** acute HIV infection; p24 antigen; Ag/Ab assays; rapid diagnostic test; diagnostic
32   accuracy; feasibility; acceptability; cost-effectiveness; implementation; systematic review**.**

Abstract

Introduction: Prompt HIV diagnosis is essential to enabling effective prevention and treatment, thereby interrupting transmission and sustain epidemic control. Strategies that could diagnose individuals during acute HIV infection (AHI), while leveraging standard systems that deliver same visit results, have potential to further enhance impact. Antibody-only (Ab) rapid diagnostic tests (RDTs), which are unable to detect AHI, remain the global standard, particularly in resource-limited settings. Antigen-antibody (Ag/Ab) RDTs have been available for more than a decade, however, their use has been largely limited to high-income settings and their ability to detect AHI in resource-limited settings remains unclear.

Methods: We systematically searched six databases through November 2024 for studies evaluating Ag/Ab RDTs vs laboratory reference standards in individuals aged ≥18 months. Outcomes included diagnostic accuracy, patient-level outcomes, and implementation. Risk of bias was assessed with QUADAS-2, JBI Critical Appraisal and ROBINS-I tools. Diagnostic accuracy was meta-analyzed with a bivariate random-effect model; other outcomes were synthesized narratively.

Results: From 1,366 screened records, 53 studies from 24 countries were included. The pooled sensitivity of Ag/Ab RDTs for AHI was only 48% (95% CI: 34–62) with specificity of 97% (95% CI: 84–100). Analytical sensitivity (detection of p24 antigen) was 31% and antibody detection was 15%. Ag/Ab RDT sensitivity was higher in studies using serum/plasma specimens (66%) or in laboratory setting (54%) and compared to studies using fingerstick/capillary or venous blood (33–35%). Among oral PrEP users, Ag/Ab RDTs sensitivity was 25% (95% CI, 13–44) and specificity of 94% (95% CI, 64–99). HIV positivity varied widely across population. Uptake, feasibility and linkage were generally favourable, although challenges in resource-limited settings were noted.

Discussion: Ag/Ab RDTs appear to have limited ability to detect AHI, missing more than half of AHI cases, especially in resource-limited-settings and when using fingerstick/whole blood samples and implemented within PrEP programmes. Further research should evaluate cost-effectiveness, feasibility, and implementation under real-world conditions in resource-limited contexts.

Conclusions: Detection of AHI using Ag/Ab RDTs remains a challenge. Programmes should continue to prioritize maintaining flexible, low-cost testing strategies aligned with WHO guidance to maximize access and impact, based on their setting and context.

PROSPERO Number: CRD42024603346

1 INTRODUCTION

Prompt and accurate HIV diagnosis is fundamental to achieving global goals to end HIV as a public health threat by 2030 [1]. HIV testing remains the essential gateway to prevention and treatment, including emerging long-acting products; ensuring that testing is accessible, reliable and responsive to evolving needs is therefore critical [2–4]. Early diagnosis, particularly during acute HIV infection (AHI), could offer an opportunity to enhance both individual and public health benefits. Immediate treatment initiation during AHI could reduce transmission and improve clinical outcomes [5].

.

AHI, defined by the highly infectious phase of early infection, typically within two-weeks after HIV-exposure, and may present with nonspecific or flu-like symptoms [6]. At the individual-level, HIV diagnosis during this brief window remains challenging, with recent data showing that due to declining testing volumes, and late-presentation, that the gap between HIV acquisition and diagnosis has doubled since 2017 [7]. Technologically, antibody (Ab)-only rapid diagnostic test (RDTs)—the global standard for routine HIV testing cannot reliably detect AHI and is known to miss early cases [8]. To date, reliable detection of HIV during AHI has required nucleic acid testing (NAT) techniques, or other laboratory-based Ag/Ab assays, which are costly do not provide same visit results, and are not routinely available in many settings.

While some programmatic approaches, such as outreach testing and AHI screening tools have shown to be beneficial in research settings, their real-world effectiveness and feasibility may vary [9]. Renewed attention has been drawn to the implications of missed AHI recently with efforts to scale-up pre-exposure prophylaxis (PrEP), including long-acting PrEP formulations, and addressing concerns about potential risks of HIV drug resistance and onward transmission.

HIV-1/2 p24 Ag/Ab RDTs (hereafter referred to Ag/Ab RDTs) that detect AHI could represent a potential innovation in programmatic significance as they are more compatible with the World Health Organization (WHO) guidance supporting decentralised, affordable and accessible testing with same visit diagnosis. This review examines the clinical utility of Ag/Ab RDTs, with a focus on diagnostic performance, client-level outcomes (including uptake, positivity, linkage, feasibility and acceptability, preferences, social harm) as well as systems-level evidence on resource implications and both facilitators and barriers to implementation. The analysis builds on a previous scoping review [5], and other systematic reviews [10,11] to assess whether Ag/Ab RDTs can strengthen service delivery and enhance public health impact.

2 METHODS

We registered the protocol in PROSPERO (CRD42024603346) and conducted the review according to PRISMA guidelines (Appendix S1) and the Cochrane Handbook for Diagnostic Test Accuracy Reviews [12,13]. Ethical review and informed consent were not required for this study, as it did not involve the collection or analysis of individual human subject data.

2.1 Eligibility criteria

We included studies assessing Ag/Ab RDTs for AHI in populations aged ≥18 months, regardless of antiretroviral exposure or diagnostic algorithm use. Eligible outcomes included diagnostic accuracy, uptake, positivity, linkage to care, feasibility, acceptability, values and preferences, resource use, social harms and barriers and facilitators to implementation. Reviews, case reports, manufacturer documents, opinion papers, and proof-of-concept studies were excluded.

2.2 Search strategy and selection

We searched ClinicalTrials.gov, Cochrane Library, PubMed, Medline, Scopus, and Web of Science (October–November 2024) using database-specific combinations of key words and MeSH terms (e.g., "acute HIV," "antigen-antibody," "rapid diagnostic tests"). Searches were restricted to January 2010–November 2024 to reflect commercial test availability. No restrictions on language, country or setting were applied; non-English studies were translated with DeepL. Additional records were identified through reference screening. Full search strategies are provided in (Appendix S2). Two independent reviewers (EF, CL) independently screened titles/abstracts and extracted data using standardized forms. Discrepancies were resolved by consensus or third-party adjudication.

2.3 Risk of bias assessment

Diagnostic accuracy studies were appraised with QUADAS-2, evaluating the risk of bias and applicability across four domains [14]. Patient-level impact and implementation studies were assessed with the appropriate Johanna Brigg's Institute (JBI) Critical Appraisal Tool and ROBINS-I tool [15]. Results were summarized in tables and figures.

2.4 Data analysis
2.4.1 Diagnostic accuracy

We included studies comparing Ag/Ab RDTs to serological and/or molecular laboratory-based

assays. AHI was defined as a positive p24 antigen or HIV RNA test with seronegative or

discordant antibody results. We excluded studies using nonclinical specimens or lacking raw

data to calculate sensitivity and specificity.

Accuracy measures were pooled by reference standard, with sensitivity and specificity reported

alongside 95% CIs. Inconclusive results were excluded from pooled analysis but reported

separately. Clinical utility of AHI detection was assessed as the relative increase in case

detection when comparing Ag/Ab RDTs with antibody-only RDTs. Reference standards were

categorized as NAT-based or p24 Ag-based for subgroup analysis.

Analyses were conducted in StataNow 19.5 (StataCorp, College Station, TX). Stata "metadta"

was used to produce pooled analysis and forest plots [16]. Bayesian re-analysis was performed

in R to verify robustness.

147   2.4.2 Patient-level and implementation outcomes

In addition to evaluating the clinical utility of AHI detection, we analyzed patient-level outcomes

descriptively, including test uptake, overall positivity and antigen-only positivity, linkage to

confirmatory testing, linkage to ART initiation, and turnaround time to treatment. These

outcomes are reported as percentages where applicable. Broader implementation outcomes –

feasibility, acceptability, values/preferences, social harms, resource use and barriers and

facilitators to implementation –were summarized narratively.

155   3 RESULTS

The search identified 2,288 records; after removing duplicates, 1,366 unique records were

screened, and 53 studies were included (Figure 1). Of these, 31 were diagnostic test accuracy

studies [17–47], 12 evaluated patient-level impact [43, 48, 50–52, 54, 60, 62–64, 67, 68], 14

addressed feasibility [43, 48, 50–52, 54, 59–60, 62–64, 66–68], nine explored values and

preferences [49, 51, 53–56, 61, 68, 69], three examined resource use [57, 58, 65], and 19

reported barriers and facilitators to implementation. Study categories were not mutually

exclusive. A detailed summary of outcomes reported by each study is available in (Table S1).

Figure 1. PRISMA flowchart of included and excluded records.

3.1 Study characteristics

Thirty-one studies reported on diagnostic accuracy and characteristics are summarized in Table 1. The 31 studies, spanning all five WHO regions were primarily in health facilities (53% laboratories, 33% clinics). Populations included individuals at high ongoing HIV risk (40%), those previously diagnosed with HIV (37%), PrEP users (10%), hospital patients (7%), the general population (3%) and women (3%). All studies evaluated adults; none assessed infants or children. Only six studies (16%) used fingerstick or venous blood with most relying on composite reference standards to define infection status. Across studies, a total of 7,184 specimens were analyzed, with sample sizes ranging from 5 to 1,494; nearly two-thirds (63%) involved fewer than 100 specimens, typically in laboratory-based case-control studies. Details on RDT brands and testing operators are provided in (Table S2).

An additional 23 studies reported on client-level and implementation outcomes [48–69], Further details and study characteristics are provided in (Table S3).

3.2 Risk of bias assessment

Study quality was assesses using QUADAS-2 (Table S4). Most diagnostic accuracy studies were judged to have a low risk of bias for the index test, reference standard, and flow/timing domains. However, risk of bias was frequently high for patient selection, reflecting the predominance of case-control designs and non-consecutive sampling. Concerns regarding applicability were generally low across domains. See (Table S4) for full assessment. Study quality for thirteen non-diagnostic accuracy studies was evaluated using the JBI Critical Appraisal Tool, and three studies were assessed with the ROBINS-I tool. All cross-sectional studies demonstrated a low risk of bias (Table S5). Among the three cohort studies, two had a low risk of bias, while one was judged to have a high risk due to missing data, outcome measurement issues, and selective reporting (Table S6).

Table 1. Key characteristics of diagnostic accuracy studies.

Study	Country and setting	Population	Sample size	Specimen type and reference assays
Beelaert 2010 [17]	Belgium, lab	HIV+/HIV-	239	Serum; p24 Ag
Fox 2011 [18]	UK, lab	HIV+	26	Serum; p24 Ag
Naylor 2011 [19]	UK, hospital	HIV hospital outpatients	5	Fingerstick; p24 Ag
Kilembe 2012 [20]	Rwanda, Zambia	At-risk of HIV	59	Plasma; p24 Ag
Laperche 2012 [21]	France, lab	HIV+	20	Plasma; p24 Ag
Patel 2012 [22]	USA, clinics	At-risk of HIV	33	Plasma; NAT
Rosenberg 2012 [23]	Malawi, clinics	At-risk of HIV	846	Venous blood; NAT
Faraoni 2013 [24]	Italy, hospital	At-risk of HIV	141	Serum; NAT
Brauer 2013 [25]	SA, lab	HIV+	36	Serum; p24 Ag
Kahawata 2013 [26]	Japan, lab	Hospital patients	25	Plasma; NAT, p24 Ag/Ab EIA, WB
Pilcher 2013 [27]	USA, lab	At-risk of HIV	138	Plasma; NAT
Conway 2014 [28]	Australia, clinics	At-risk of HIV	3154	Fingerstick; NAT, p24 Ag
Duong 2014 [29]	Eswatini, at-home testing	General population	12370	Fingerstick; NAT
Ottiger 2015	Switzerland, lab	HIV+	21	Serum; NAT

[30]				
Stekler 2016 [31]	USA, clinic	At-risk of HIV	1494	Whole blood; NAT
Smith 2016 [32]	UK, clinic	At-risk of HIV	21	Serum; p24 Ag
Delaugerre 2017 [33]	France, clinic	PrEP users	13	Plasma; WB
Fitzgerald 2017 [34]	UK, lab	HIV+/HIV-	69	Serum; p24 Ag
Fransen et al 2017 [35]	Belgium, lab	PrEP users	64	Plasma; NAT
Livant 2017 [36]	SA, Uganda, Zim, clinics	PrEP users	215	Plasma; NAT
Masciotra 2017 [37]	USA, lab,	At-risk of HIV	112	Plasma; NAT
Stafylis 2017 [38]	USA, lab,	HIV+/HIV-	49	Serum; NAT, Ab-only EIA, p24 Ag/Ab EIA
Parker 2018 [39]	USA, lab	HIV (+)	125	Serum; NAT, Ab-only EIA, p24 Ag/Ab EIA
van Tienen 2018 [40]	Netherlands, lab	At-risk of HIV	23	Serum; p24 Ag
Chavez 2020 [41]	USA, clinic	At-risk of HIV	878	Venous blood; NAT
Wratil 2020 [42]	Germany, lab	HIV+/HIV-	21	Plasma; NAT, p24 Ag, p24 Ag/Ab EIA
Kerschberger 2021 [43]	Eswatini, outpatient centre	At-risk of HIV	731 745	Plasma, Venous blood; NAT
Sirivichayakul 2021 [44]	Thailand, lab	HIV+/HIV-	50	Plasma; NAT, p24 Ag/Ab EIA
Guiraud 2024 [45]	France, lab	Hospital patients	32	Serum; WB
Manjate 2024 [46]	Mozambique, outpatient centre	Women	762	Serum; p24 Ag
Ciglenecki 2025 [47]	Eswatini, HIV testing sites	At-risk of HIV	1163	Fingerstick; NAT

Abbreviations: Ag/Ab: antigen/antibody; EIA: enzyme immunoassay; NAT: nucleic acid testing; PrEP: pre-exposure prophylaxis; SA: South Africa; UK: United Kingdom; USA: United States of America; WB: Western Blot; Zim: Zimbabwe

3.3 Diagnostic accuracy

A meta-analysis of 31 studies, representing 38 reports, had a pooled diagnostic sensitivity and specificity of Ag/Ab RDTs for AHI of 48% (95% CI: 34–62) and 97% (95% CI: 84–100), respectively, with considerable heterogeneity across studies (Figure 2). Pooled sensitivity and

specificity for p24 antigen detection varied widely (31%, 95% CI: 19–45 to 99%, 95% CI: 97–

100), with considerable heterogeneity across studies (Figure 3).

A   **B**

Figure 2. Diagnostic accuracy of Ag/Ab RDTs. (A) Forest plot of sensitivity and specificity
from individual studies with 95% confidence intervals; pooled estimates are shown as red
diamonds. (B). Summary receiver operating characteristics (SROC) curve with 95% confidence
and prediction regions.

A   **B**

Figure 3. Diagnostic accuracy of Ag/Ab RDTs (p24 Ag detection). (A) Forest plot of
sensitivity and specificity from individual studies with 95% confidence intervals; pooled estimates
are shown as red diamonds. (B). Summary receiver operating characteristics (SROC) curve with
95% confidence and prediction regions.

The pooled estimates for HIV-antibody detection also ranged widely (15% sensitivity, 95% CI:

10–24 to 99% specificity, 95% CI: 96–100), with substantial heterogeneity (Figure 4).

A   **B**

Figure 4. Diagnostic accuracy of Ag/Ab RDTs (antibody detection). (A) Forest plot of sensitivity and specificity from individual studies with 95% confidence intervals; pooled estimates are shown as red diamonds. (B). Summary receiver operating characteristics (SROC) curve with 95% confidence and prediction regions.

In subgroup analysis (Table 2), sensitivity was higher in case-control studies (61%), European countries (70%), serum-based testing (66%), laboratory settings (54%), when using p24 Ag reference assays (58%), and among people living with HIV who had previously been diagnosed (70%). The accuracy of Ag/Ab RDTs in capillary/fingerstick and venous whole blood specimens was lower compared to serum and plasma, with sensitivities of 35% and 33%, respectively, while specificities remained above 99%. Among oral PrEP users, sensitivity and specificity dropped to 25% (95% CI, 13–44) and 94% (95% CI, 64–99).

Table 2. Subgroup analysis of pooled accuracy estimates for Ag/Ab RDTs.

WHO region	p24 Ag/Ab combined		p24 Ag detection		Ab detection	
	Sensitivity (95% CI)	Specificity (95% CI)	Sensitivity (95% CI)	Specificity (95% CI)	Sensitivity (95% CI)	Specificity (95% CI)
African	17% (5–44)	100% (99–100)	12% (8–19)	100% (99–100)	13% (6–31)	100% (99–100)
The Americas	39% (14–71)	99% (92–100)	21% (11–37)	99% (97–100)	26% (12–47)	99% (99–100)
European	70% (56–81)	98% (83–100)	61% (44–75)	95% (83–99)	14% (9–26)	95% (77–99)
South-East Asia	–	–	–	–	–	–
Western Pacific	–	–	–	–	–	–
Study setting						
Laboratory	54% (38–69)	97% (92–99)	34% (20–50)	98% (92–99)	19% (12–28)	98% (92–100)
Non-laboratory	40% (14–73)	98% (93–100)	34% (19–53)	98% (90–99)	13% (5–28)	98% (92–100)

Study design						
Case-control	61%	97%	42%	95%	21%	96%
	(50–72)	(89–99)	(29–57)	(86–98)	(14–30)	(85–99)
Cross-sectional	9%	100%	17%	100%	10%	100%
	(0–69)	(98–100)	(7–37)	(98–100)	(3–29)	(99–100)
Population						
HIV-infected	70%	97%	60%	96%	17%	96%
	(59–79)	(78–100)	(43–74)	(82–99)	(9–29)	(77–99)
At-risk of HIV	29%	99%	18%	99%	18%	100%
	(11–57)	(98–100)	(10–32)	(97–100)	(9–31)	(99–100)
PrEP users	25%	94%	21%	97%	12%	96%
	(13–44)	(64–99)	(9–43)	(44–100)	(7–22)	(78–99)
Specimen type						
Serum	66%	99%	46%	97%	17%	95%
	(51–78)	(87–100)	(26–68)	(85–99)	(8–30)	(82–99)
Plasma	43%	98%	24%	98%	23%	96%
	(23–67)	(87–100)	(13–39)	(84–100)	(13–38)	(82–99)
Venous blood	33%	99%	17%	99%	19%	99%
	(13–63)	(98–99)	(8–33)	(98–100)	(5–50)	(99–100)
Fingerstick blood	35%	99%	24%	99%	1%	99%
	(1–96)	(93–99)	(4–71)	(95–100)	(0–66)	(99–100)
Reference standard						
p24 Ag	58%	100%	45%	99%	11%	99%
	(33–79)	(77–100)	(17–77)	(86–100)	(7–16)	(81–100)
NAT	35%	99%	21%	99%	15%	99%
	(14–64)	(98–100)	(9–42)	(98–100)	(6–31)	(99–100)

237 Overall, there was a high-level of variation in the total number of additional AHI cases identified

238 by the antibody component of Ag/Ab RDTs compared with Ab-only RDTs [22,27,33,35,36,45]

239 (Figure 5).

241 **Figure 5. Relative increase in AHI cases detected by the antibody component of Ag/Ab**
242 **RDTs compared with Ab-only RDTs.**

3.4 Client-level impact outcomes

3.4.1 Uptake and positivity

246 Four observational studies assessed uptake of Ag/Ab RDTs, with high with rates reported

247 among vulnerable and key populations in Belgium (78%), Argentina (100%) and Spain (100%)

[52,62,67], but low initial uptake (14.3%) in primary healthcare clinics in MSM in Australia, which improved over time [51]. Twelve studies reported wide variation in overall test positivity rates (mostly antibody-based) ranging from 0.77% in general populations in the USA [63], 1.3% – 7.7% among at-risk individuals and key populations [43,48,50–52,54,64,67], and 8.1% in antenatal clinics in Kenya [68], up to 19.6% among transgender women in Argentina [62] and high rates in key populations using partner notification services in Vietnam [60]. Nine studies documented p24 antigen-only reactivity: four reported 0% (UK, Australia, Argentina and Spain [48,51,62,67], while positivity reached 7% in the USA [63], 10% in pregnant women in Kenya [68], 11% among key populations in Vietnam [60], 20% in at-risk individuals in Eswatini [43], and 31% among MSM in a community-based clinic in Spain [64].

3.4.2 Linkage to confirmatory testing and treatment

Across studies linkage to confirmatory testing were generally high, as with other HIV testing studies. Ab-reactive results (established HIV) achieved 99%–100% in seven studies [48,50–52,54,62,67]. In four studies, all Ag-reactive cases (100%) were reportedly confirmed, either later on site or through referrals off-site [43,60,63,64].

Turnaround times for confirmatory testing varied from 1–7 day in high-income contexts only [55,56,63]. Treatment initiation after confirmed diagnosis was high, with rates of 86-100% [48,50,52,54,63,64,67], primarily in high-income countries. However, in resource-limited settings, such as Eswatini, delays were common with only 72% initiating treatment within one month, 37% initiating within one week and 13% initiating same day [43]. In Vietnam, barriers such as patient readiness and delays in confirmatory testing also hindered timely treatment [60].

3.4.3 Additional implementation outcomes

Nineteen studies documented facilitators and barriers to implementation [43, 48–57, 59–64, 66–68]. Facilitators included user-friendliness, convenience, and rapid results. high acceptability among clients and healthcare providers due to privacy, confidentiality, and anonymity. Low-cost testing with demonstrated cost-effectiveness [57–59, 67, 68], including willingness to pay if unfunded [55]. Barriers included interpretation challenges for inexperienced operators, particularly with three-test-line results, highlighting the need for robust training [63]. A comprehensive list of facilitators and barriers is provided in Table 3.

Fourteen studies suggested Ag/Ab RDTs could be acceptable and feasible to implement, but
there was considerable variability by setting, test provider, and population. Reported social
harms were rare, though false positives occurred, especially during early implementation,
highlighting the importance of training and quality assurance. Resource use information was
limited to high-income settings, but suggested Ag/Ab RDTs could be cost-effective compared to
alternatives, particularly in higher prevalence or high-risk populations. Key facilitators to success
and buy-in included implementation that was user-friendliness, discreet, able to track reactive
results in real time; however, barriers to implementation included challenges due to limited
training, operator inexperience, lack of interchangeability with other products, stigma, legal and
financial constraints, and confirmatory testing delays. Further details on implementation related
outcomes are provided in (Appendix S3).
292

 Table 3. Studies reporting barriers and facilitators to implementation

Study ID	Setting	Population	Testing modality, test operator & sample type	Facilitators	Barriers
Taegmeyer 2011 [48]	UK, community outreach & hospital-based STI clinics	PWID, MSM, SW, asylum seekers	Community-based; nurses, health assistants; sample type not reported	Increased access to underserved groups; rapid results enabled prevention messaging	False positives in low-prevalence settings; ~1.7% invalid results; limited added value over Ab-only tests
O'Neil 2012 [49]	USA, STI clinics	MSM, TG	Facility-based; counsellors; fingerstick	High trust in fingerstick and venous samples; NAT viewed as accurate reference	Concerns about early diagnosis; low trust in oral RDTs; Ag/Ab RDT not yet approved
Melbroek 2013 [50]	Spain, community centre	MSM	Community-based; counsellors; venous blood	Efficient detection; strong linkage to care (90%)	Delayed ART start due to 7-day confirmation of reactive results
Eu 2014 [51]	Australia, primary care clinic	MSM	Facility-based; operator and sample type not reported	Easy clinical integration: higher HIV positivity than lab tests (4.1% vs 1.3%)	Additional patient cost (AUS$20/test)
Wouters 2014 [52]	Belgium, outpatient centre	MSM, Southern African migrants; others	Facility-based; nurses; whole blood	Successful nurse-led implementation; higher positivity in risk groups; uptake rose (76% to 87%)	False positives; confirmation delays causing anxiety; early interpretation errors; need for stronger training
Yang 2014 [53]	Australia, online survey	MSM	Not applicable	Preference for rapid and non-clinic testing (home most preferred); government coverage improves access	Barriers: low risk perception, stigma, fear of rejection, inconvenient clinic hours
Mayer 2014 [54]	USA, home based pilot	MSM	Self-testing; client; fingerstick	High acceptability and feasibility using biweekly Ag/Ab self-tests	Some users found testing too complex
Conway 2015 [55,56]	Australia, STI clinics	MSM, TG Doctors, nurses	Facility-based; doctors, nurses; fingerstick	High acceptability among clients/ providers; willingness to pay if unfunded	Accuracy concerns; 20-min result delayed consultations
Hoenigl 2016 [57]	USA, community centre	MSM	Community-based; operator and sample type not reported	Cost-effective AHI screening in high-risk groups, lower cost vs lab testing	High costs; need for venipuncture; limited sensitivity during AHI

Study ID	Setting	Population	Testing modality, test operator & sample type	Facilitators	Barriers
Alivayeva 2019 [59]	Kazakhstan, NGO clinic	PLHIV, HTS clients, providers	Community-based; lab techs; erum	Feasible with funding, policy support, validated algorithm, and training	Stigma; discrimination; need legal and social protection
Green 2019 [60]	Vietnam, community centres	MSM, PLHIV, partners	Community-based; community providers; whole blood	Feasible within community PNS using risk scoring and rapid linkage	Weak referral systems; transport delays; confirmatory testing slowed diagnosis
Mullens 2019 [61]	Australia, mobile van	MSM	Community-based; peer testers; no sample type reported	Peer-led testing highly acceptable; discrete, accessible, nonintrusive; online promotion effective	Confidentiality concerns; logistical/safety challenges (equipment, hours, site selection, peer role limits)
Frola 2020 [62]	Argentina, home-based	Transgender women	Home-based; lab techs; venous blood	Peer-supported home testing feasible; effective for case finding	Resource-intensive; required highly skilled multidisciplinary team
Martin 2020 [63]	USA, CLIA-waived sites	General population	Facility-based; operators not reported; fingerstick	Feasible in 2-test algorithm (92% true positives); reduced delays vs lab testing	Very few AHI cases (1/310.785); limited efficiency; operator errors; poor follow-up
Kerschberger 2021 [43]	Eswatini, outpatient clinic	High-risk individuals	Facility-based; lab techs; plasma and venous whole blood	Targeted screening with risk score and Ag/Ab HIV RDT effective for AHI detection	Few same-day ART starts; limited non-lab evaluation; possible misclassification; no cost-effectiveness data
Saz 2021 [64]	Spain, community centre	MSM	Community-based; operator not reported; whole blood	Combined with point-of-care viral load enabled same-day diagnosis/treatment (13 cases)	High per-case cost (EUR 13.649/AHI case); need simplified confirmatory algorithm
Ferrer 2022 [66]	Chile, university hospital	General population	Community-based; operators not reported; whole blood	Validated 6 RDTs before rollout; feasible and accepted; supported confidentiality and policy change	No AHI detected in low-prevalence setting; no NAT confirmation; limited routine utility
Muñoz-Chimeno 2023 [67]	Spain, mobile unit	PWUD, migrants, homeless	Community-based multiplex (HIV, HCV, HBV); nurses; fingerstick	Feasible, acceptable; enabled same-day referral	Requires robust training and supervision

Study ID	Setting	Population	Testing modality, test operator & sample type	Facilitators	Barriers
Ochwoto 2024 [68]	Kenya, primary healthcare	Pregnant women	Facility-based multiplex (HIV, Syphilis, HBV, Malaria); ANC nurses, midwives and lab techs; fingerstick	Improved coverage, data management, and quality; potentially cost-effective	Increased workload; need for training and supervision before rollout

Abbreviations: ANC: Antenatal Care; CLIA: Clinical Laboratory Improvement Amendments; HTC: HIV Testing and Counselling; KP: Key Populations; MSM: Men who have Sex with Men; NAT: Nucleic Acid Testing; PLHIV: People Living with HIV; PNS: Partner Notification Services; PWUD: People Who Use Drugs; PWID: People Who Inject Drugs; PrEP: Pre-Exposure Prophylaxis; STI: Sexually Transmitted Infection; TG: Transgender people; UK: United Kingdom; USA: United States of America

4 DISCUSSION

The findings from this review, which comprehensively evaluated the diagnostic accuracy of Ag/Ab RDTs across multiple domains, highlights that Ag/Ab RDTs have a low to moderate ability to detect AHI (48%, 95% CI: 34–62) and that results per country, context and population may vary widely. Performance using capillary/fingerstick or venous blood samples, and outside of laboratory settings, had low sensitivity during AHI; 35% and 33%, respectively. Based on current data summarized by this review, Ag/Ab RDTs appear to have several limitations and variable performance in terms of their ability to detect AHI within routine HIV testing in low-and middle-income countries, particularly those relying on decentralised routine testing outside laboratories [23,28,29,31,41,47]. These findings are consistent with earlier studies and reviews that compared HIV Ag/Ab RDTs with Ab-only RDTs [22,27,33,35,36,45].

This review also highlighted that while there is interest in the role Ag/Ab RDTs could play in PrEP implementation, test performance during AHI among PrEP users was poor, with pooled sensitivity and specificity of 23% (95% CI: 16–33) and 87% (95% CI: 66–96), respectively. This remains lower than established performance of laboratory-based Ag/Ab assays and NAT. Encouragingly, however, in the most recent update to the WHO guidelines evidence showed that standard Ab-only RDTs, as well as self-tests, are sufficient for PrEP and LAI-PrEP which is important to continue access to these important prevention tools in resource-limited sites [70,71]. A recent modelling study highlighted that existing Ab-only RDTs, as well as self-tests, with even yearly testing for LAI-PrEP would not have a significant impact in incidence or drug resistance [72,73]. Similarly, results from Brazil's implementation of CAB-LA show self-tests performed well and were sufficiently accurate [74].

Despite the limitations related to AHI detection in a routine decentralized testing service, quality-assured Ag/Ab RDTs may have benefits, as indicated in select studies, when implemented in settings with limited access to sensitive assays for blood screening and safety [75] or as part of highly targeted case-finding and outreach efforts. Implementation studies highlighted that given available systems and resources, Ag/Ab RDTs, could offer value and be cost-effective if highly targeted [65]. Even with these potential possibilities, however, efforts to place these tools and achieve impact could be challenging. With resources, and staffing, declining, there are challenges with introducing tests that may also increase the cost of training and limit the ability to have flexible or interchangeable algorithms. Any country or programme considering Ag/Ab RDTs will need to ensure they have the resources to support the potentially higher costs, as well as

transition costs, given shrinking testing capacity globally. WHO's tool kit, low-cost diagnostic
strategy and prioritization guide are important resources that may be helpful to consider in this
uncertain environment as the promote low cost and flexible product selection based on local
needs [76].

Cost-effectiveness analysis, as well as budget impact analysis, with real implementation data
was not available in any resource-limited setting and is a data gap that should be addressed.
Understanding the benefits, and costs, of adopting other assays such as multiplex tests need to
be considered to ensure investment is going toward interventions that have greatest value for
money, and which will have substantial public health impact.

This review has several strengths: it integrates pooled diagnostic accuracy estimates with
patient-level and implementation outcomes, drawing from 53 studies across 25 countries. Its
breadth enhances generalizability and relevance, while identifying implementation barriers and
research gaps. By breaking down evidence by geography, population and sample type additional
insights for real world settings can also be derived.

Despite these strengths, a critical limitation was the publication bias with most evidence being
classified as low certainty due to the large number of case-control studies included [77]. First,
most available evidence originates from high-resource settings, leaving important gaps in low-
and middle-income countries. High-quality, real-world evaluations on the public health impact,
and clinical utility, of Ag/Ab HIV RDTs in routine HIV testing services in resource-limited settings
remain important research gaps to fill. Additional gaps identified also included the inability to
evaluate test performance in children and the high-level of heterogeneity in how studies defined
and measured AHI. Further, while PrEP users were included, because of the timing of the review
and the diversity of products entering the market, our results may also not be fully representative
of all forms of PrEP being implemented. Although we extracted, standardized and recalculated
data as much as possible there is a risk that we included individuals without AHI in our
estimates, or missed studies due to reporting not being sufficiently disaggregated. The review
also focuses on detection of AHI and our findings focus on the additional benefits, and do not
provide a full review of all aspects of diagnostic performance. Changes in the commercially
available products may also affect results as this review is focused on the performance of a
broad class of Ag/Ab rapid tests, and how they may perform and how their general use may
affect programmes, and is not focused on specific brands.

Lastly, while we discuss the importance of data on cost-effectiveness and resource use, there
was insufficient data available, specifically from resource-limited settings. This is critical moment
with substantial uncertainties on future investment and funding for global health. Costs are
shifting and if Ag/Ab RDTs costs decrease and become similar or lower cost when compared to
alternative Ab-only RDTs this would be an important consideration. As with all technologies,
before introducing new products, it remains critical that countries and funders understand the
short-, medium- and long-term cost implications of adopting Ag/Ab RDTs.
375
More data on the real-world feasibility of using Ag/Ab RDTs in routine testing settings with
minimal resource requirements, including client level-impact, particularly studies evaluating the
impact of streamlined pathways for rapid confirmation of AHI (including standardised case
definition) and ART initiation, would be beneficial. Additional economic analysis along with
values and preferences and social harms would also strengthen the evidence base and address
knowledge gaps.
382

5 CONCLUSIONS

384 Prompt and accurate HIV diagnosis remains essential to ending the epidemic. HIV RDTs remain
385 critical to the response. The added value of Ag/Ab RDTs remains unclear due to limited and
386 variable sensitivity during AHI—particularly when using fingerstick/capillary blood specimens.
387
388 WHO guidance remains the same and there is no evidence suggesting Ag/Ab RDTs should be
389 prioritized within routine testing services, including for PrEP delivery. Rather, given the shortages
390 in staff and funding it is important countries embrace WHO's guidance on ensuring country
391 flexibilities and work toward adopting low-cost commodities to maintain testing coverage where it
392 is needed most.
393
394 Programmes in resource-limited settings must continue to carefully balance the risks and
395 benefits of adopting different technologies which may increase complexity or costs, and factor in
396 potential competing priorities (e.g. multiplex testing and self-testing.
397

COMPETING INTERESTS

399 The authors have declared that no competing interests exist.
400
401

AUTHORS' CONTRIBUTIONS

Conceptualization: CJ, AS, EF, CL; Methodology: EF, CJ; Formal analysis: EF, CL; Data curation: EF; Investigation: EF, CL, CJ, AS; Writing – original draft: EF, CJ; Writing – review & editing: EF, CL, PJ, MD, CBS, BS, AS, CJ; Visualization: EF; Supervision: CJ, MD; Project administration: CJ, MD; Funding acquisition: CJ

ACKNOWLEDGEMENTS

We sincerely thank Olanrewaju Ayodele Edun for independently validating our results by rerunning the bivariate random-effect models in R using the Bayesian approach for diagnostic accuracy meta-analysis, and for his invaluable statistical guidance throughout this study.

FUNDING

Funding for this review was provided under the WHO-Gates Grant INV-053439 "Realizing the future of HIV testing, linkage and re-engagement". The funder of the study had no role in study design, data analysis, interpretation, or writing of the report. The corresponding author had full access to all the data in the study and had final responsibility for the decision to submit for publication.

DISCLAIMER

The opinion and views expressed in this article are those of the authors and do not necessarily reflect the official position, policies, or decisions of the World Health Organization.

DATA AVAILABILITY STATEMENT

All data underlying the findings of this study are included in the article and its supporting information files. Extracted datasets and analysis files are available from the corresponding author upon reasonable request.

SUPPORTING INFORMATION

Additional information may be found under the Supporting Information tab for this article.

Appendix S1. PRISMA Checklist

Appendix S2. Full search strategies

Appendix S3. Feasibility, acceptability, values and preferences, social harms, and resource use.

REFERENCES

1. Global health sector strategies on, respectively, HIV, viral hepatitis and sexually transmitted
infections for the period 2022-2030. Geneva: World Health Organization; 2022. Available:
https://iris.who.int/bitstream/handle/10665/360348/9789240053779-eng.pdf?sequence=1
2. Consolidated guidelines on differentiated HIV testing services. Geneva: World Health
Organization; 2024. Available: https://www.who.int/publications/i/item/9789240096394
3. Guidelines on long-acting injectable cabotegravir for HIV prevention. Geneva: World Health
Organization; 2022. Available: https://www.who.int/publications/i/item/9789240054097
4. Guidelines on lenacapavir for HIV prevention and testing strategies for long-acting injectable
pre-exposure prophylaxis (PrEP). Geneva: World Health Organization; 2025. Available from:
https://iris.who.int/bitstream/handle/10665/381892/9789240111608-eng.pdf?sequence=1
5. Rutstein SE, Ananworanich J, Fidler S, et al. Clinical and public health implications of acute
and early HIV detection and treatment: a scoping review. J Int AIDS Soc. 2017 Jun
28;20(1):21579.
6. Bottone PD, Bartlett AH. Diagnosing Acute HIV Infection. *Pediatr Ann.* 2017;46(2):e47-e50.
7. Hall M, Golubchik T, Bonsall D, Abeler-Dörner L, Limbada M, Kosloff B, et al; HPTN 071
(PopART) Phylogenetics protocol team; PANGEA consortium. Demographics of sources of
HIV-1 transmission in Zambia: a molecular epidemiology analysis in the HPTN 071 PopART
study. Lancet Microbe. 2024 Jan;5(1):e62-e71.
8. Adetunji AA, Adewumi MO, Michael OS, Fayemiwo SA, Ogunniyi A, Taiwo BO. Rapid HIV
Antigen-Antibody Assays and Detection of Acute HIV Infection in Sub-Saharan Africa. *Am J*
Trop Med Hyg. 2019;101(2):285-286.
9. Ong JJ, Coulthard K, Quinn C, Tang MJ, Huynh T, Jamil MS, Baggaley R, Johnson C. Risk-
Based Screening Tools to Optimise HIV Testing Services: a Systematic Review. Curr
HIV/AIDS Rep. 2022 Apr;19(2):154-165.
10. Lewis JM, Macpherson P, Adams ER, Ochodo E, Sands A, Taegtmeyer M. Field accuracy of
fourth-generation rapid diagnostic tests for acute HIV-1: a systematic review. *AIDS.*
2015;29(18):2465-2471.
11. Smallwood M, Vijh R, Nauche B, Lebouché B, Joseph L, Pant Pai N. Evaluation of a rapid
point of care test for detecting acute and established HIV infection and examining the role of
study quality on diagnostic accuracy: A Bayesian meta-Analysis. *PLoS One.*
2016;11(2):e0149592.

12. McInnes MDF, Moher D, Thombs BD, McGrath TA, Bossuyt PM, et al. Preferred Reporting
Items for a Systematic Review and Meta-analysis of Diagnostic Test Accuracy Studies: The
PRISMA-DTA Statement [published correction appears in JAMA. 2019 Nov 26;322(20):2026.
13. Deeks JJ, Bossuyt PM, Leeflang MM, Takwoingi Y (editors). Cochrane Handbook for
Systematic Reviews of Diagnostic Test Accuracy. Version 2.0 (updated July 2023).
Cochrane, 2023. Available: https://training.cochrane.org/handbook-diagnostic-test-
accuracy/current.
14. Whiting PF, Rutjes AWS, Westwood ME, Mallett S, Deeks JJ, Reitsma JB, et al. QUADAS-2:
a revised tool for the quality assessment of diagnostic accuracy studies. Ann Intern Med.
2011 Oct 18; 155(8):529–36.
15. Sterne JAC, Hernan MA, Reeves BC, Savović J, Berkman ND, Viswanathan M, et al.
ROBINS-I: a tool for assessing risk of bias in non-randomised studies of interventions. BMJ.
2016 Oct 12; 355:i4919.
16. Nyaga VN, Arbyn M. Metadta: a Stata command for meta-analysis and meta-regression of
diagnostic test accuracy data - a tutorial. Arch Public Health. 2022 Mar 29;80(1):95. Erratum
in: Arch Public Health. 2022 Sep 27;80(1):216
17. Beelaert G, Fransen K. Evaluation of a rapid and simple fourth-generation HIV screening
assay for qualitative detection of HIV p24 antigen and/or antibodies to HIV-1 and HIV-2. J
Virol Methods. 2010 Sep;168(1-2):218-22.
18. Fox J, Dunn H, O'Shea S. Low rates of p24 antigen detection using a fourth-generation point
of care HIV test. Sex Transm Infect. 2011 Mar;87(2):178-9.
19. Naylor E, Axten D, Makia F, Tong C, White J, Fox J. Fourth generation point of care testing
for HIV: validation in an HIV-positive population. Sex Transm Infect. 2011 Jun;87(4):311.
20. Kilembe W, Keeling M, Karita E, Lakhi S, Chetty P, Price MA, Makkan H, Latka M, Likoti M,
Ilukui K, Hurlston M, Allen S, Stevens G, Hunter E. Failure of a novel, rapid antigen and
antibody combination test to detect antigen-positive HIV infection in African adults with early
HIV infection. PLoS One. 2012;7(6):e37154.
21. Laperche S, Leballais L, Ly TD, Plantier JC. Failures in the detection of HIV p24 antigen with
the Determine HIV-1/2 Ag/Ab Combo rapid test. J Infect Dis. 2012 Dec 15;206(12):1946-7;
author reply 1949-50.
22. Patel P, Bennett B, Sullivan T, Parker MM, Heffelfinger JD, Sullivan PS; CDC AHI Study
Group. Rapid HIV screening: missed opportunities for HIV diagnosis and prevention. J Clin
Virol. 2012 May;54(1):42-7.
23. Rosenberg NE, Kamanga G, Phiri S, Nsona D, Pettifor A, Rutstein SE, Kamwendo D,
Hoffman IF, Keating M, Brown LB, Ndalama B, Fiscus SA, Congdon S, Cohen MS, Miller
WC. Detection of acute HIV infection: a field evaluation of the determine® HIV-1/2 Ag/Ab
combo test. J Infect Dis. 2012 Feb 15;205(4):528-34.
24. Faraoni S, Rocchetti A, Gotta F, Ruggiero T, Orofino G, Bonora S, Ghisetti V. Evaluation of a
rapid antigen and antibody combination test in acute HIV infection. J Clin Virol. 2013
May;57(1):84-7.
25. Brauer M, De Villiers JC, Mayaphi SH. Evaluation of the Determine™ fourth generation HIV
rapid assay. J Virol Methods. 2013 Apr;189(1):180-3.
26. Kawahata T, Nagashima M, Sadamasu K, Kojima Y, Mori H. [Evaluation of an
immunochromatographic fourth generation test for the rapid diagnosis of acute HIV
infection]. Kansenshogaku Zasshi. 2013 Jul;87(4):431-4. Japanese

27. Pilcher CD, Louie B, Facente S, Keating S, Hackett J Jr, Vallari A, Hall C, Dowling T, Busch
MP, Klausner JD, Hecht FM, Liska S, Pandori MW. Performance of rapid point-of-care and
laboratory tests for acute and established HIV infection in San Francisco. PLoS One. 2013
Dec 12;8(12):e80629.
28. Conway DP, Holt M, McNulty A, Couldwell DL, Smith DE, Davies SC, Cunningham P, Keen
P, Guy R; Sydney Rapid HIV Test Study. Multi-centre evaluation of the Determine HIV
Combo assay when used for point of care testing in a high risk clinic-based population. PLoS
One. 2014 Apr 8;9(4):e94062.
29. Duong YT, Mavengere Y, Patel H, Moore C, Manjengwa J, Sibandze D, Rasberry C, Mlambo
C, Li Z, Emel L, Bock N, Moore J, Nkambule R, Justman J, Reed J, Bicego G, Ellenberger
DL, Nkengasong JN, Parekh BS. Poor performance of the determine HIV-1/2 Ag/Ab combo
fourth-generation rapid test for detection of acute infections in a National Household Survey
in Swaziland. J Clin Microbiol. 2014 Oct;52(10):3743-8.
30. Ottiger C, Huber AR. Comparison of the new Alere HIV Combo with Alere Determine HIV-1/2
Ag/Ab Combo in actute and primo and established HIV infections. Annals of Clinical and
Laboratory Research. 2015 Vol 3 No 3:22.
31. Stekler JD, Ure G, O'Neal JD, Lane A, Swanson F, Maenza J, Stevens C, Coombs RW,
Dragavon J, Swenson PD, Golden MR. Performance of Determine Combo and other point-
of-care HIV tests among Seattle MSM. J Clin Virol. 2016 Mar;76:8-13.
32. Smit E, Atherton C, Osman H. Determine HIV-1/2 Ab/Ag Combo lacks sensitivity in detecting
early HIV infections. European Society for Clinical Virology winter meeting. London, 13–15
January 2011.
33. Delaugerre C, Antoni G, Mahjoub N, Pialoux G, Cua E, Pasquet A, Hall N, Tremblay C, Cotte
L, Capitant C, Chaix ML, Meyer L, Molina JM; IPERGAY Study Group. Assessment of HIV
Screening Tests for Use in Preexposure Prophylaxis Programs. J Infect Dis. 2017 Aug
1;216(3):382-386.
34. Fitzgerald N, Cross M, O'Shea S, Fox J. Diagnosing acute HIV infection at point of care: a
retrospective analysis of the sensitivity and specificity of a fourth-generation point-of-care
test for detection of HIV core protein p24. Sex Transm Infect. 2017 Mar;93(2):100-101.
35. Fransen K, de Baetselier I, Rammutla E, Ahmed K, Owino F, Agingu W, Venter G, Deese J,
Van Damme L, Crucitti T; FEMPrEP study group. Performance of serological and molecular
tests within acute HIV infection. J Clin Virol. 2017 Aug;93:81-84.
36. Livant E, Heaps A, Kelly C, Maharaj R, Samsunder N, Nhlangulela L, Karugaba P, Panchia
R, Marrazzo J, Chirenje ZM, Parikh UM; VOICE Study Team. The fourth generation
Alere^{TM} HIV Combo rapid test improves detection of acute infection in MTN-003 (VOICE)
samples. J Clin Virol. 2017 Sep;94:15-21.
37. Masciotra S, Luo W, Westheimer E, Cohen SE, Gay CL, Hall L, Pan Y, Peters PJ, Owen SM.
Performance evaluation of the FDA-approved Determine™ HIV-1/2 Ag/Ab Combo assay
using plasma and whole blood specimens. J Clin Virol. 2017 Jun;91:95-100.
38. Stafylis C, Klausner JD. Evaluation of two 4th generation point-of-care assays for the
detection of Human Immunodeficiency Virus infection. PLoS One. 2017 Aug
28;12(8):e0183944.
39. Parker MM, Bennett SB, Sullivan TJ, Fordan S, Wesolowski LG, Wroblewski K, Gaynor AM.
Performance of the Alere Determine™ HIV-1/2 Ag/Ab Combo Rapid Test with algorithm-
defined acute HIV-1 infection specimens. J Clin Virol. 2018 Jul;104:89-91.

40. van Tienen C, Rugebregt S, Scherbeijn S, Götz H, Geurts van Kessel C. The performance of
the Alere HIV combo point-of-care test on stored serum samples; useful for detection of early
HIV-1 infections? Sex Transm Infect. 2018 Aug;94(5):331-333.
41. Chavez PR, Bradley HM, Wesolowski LG, Violette LR, Katz DA, Niemann LA, McMahan VM,
McDougal S, Cornelius-Hudson AM, Ethridge SF, Stekler JD, Delaney KP. Performance
evaluation of four point-of-care HIV tests using unprocessed specimens. J Clin Virol. 2020
Mar;124:104282
42. Wratil PR, Rabenau HF, Eberle J, Stern M, Münchhoff M, Friedrichs I, Stürmer M, Berger A,
Kuttner-May S, Münstermann D, Lucht A, Meixenberger K, Bannert N, Keppler OT.
Comparative multi-assay evaluation of Determine™ HIV-1/2 Ag/Ab Combo rapid diagnostic
tests in acute and chronic HIV infection. Med Microbiol Immunol. 2020 Apr;209(2):139-150.
43. Kerschberger B, Aung A, Mpala Q, Ntshalintshali N, Mamba C, Schomaker M, Tombo ML,
Maphalala G, Sibandze D, Dube L, Kashangura R, Mthethwa-Hleza S, Telnov A, Tour R,
Gonzalez A, Calmy A, Ciglenecki I. Predicting, Diagnosing, and Treating Acute and Early HIV
Infection in a Public Sector Facility in Eswatini. J Acquir Immune Defic Syndr. 2021 Dec
15;88(5):506-517.
44. Sirivichayakul S, Pankam T, Pattanachaiwit S, Phancharoen K, Barisri N, Areeyolwattana S,
Phanuphak N, Phanuphak P. Ability of Alere™ HIV Combo to diagnose acute HIV infection is
based mainly on HIV-1 p24 antigen detection. Asian Pac J Allergy Immunol. 2021 Oct 31.
45. Guiraud V, Naizet A, Khan H, Benhafoun G, Hernandez P, Piccin L, Pichon A, Leng AL,
Yousfi L, Gautheret-Dejean A. Fourth-Generation HIV Rapid Tests: Enhanced Sensitivity and
Reduced Diagnostic Window for HIV-1 Primary Infection Screening. J Med Virol. 2024
Nov;96(11):e70044.
46. Manjate A, Nilsson C, Axelsson M, Lindroth S, Sirbu D, Sacarlal J, Andersson S, Unemo M.
Laboratory-based evaluation of the 4th-generation AlereTM HIV Combo rapid point-of-care
test. PLoS One. 2024 Feb 23;19(2):e0298912.
47. Ciglenecki I, Ntshalintshali N, Mukooza E, Lekelem S, Mavimbela M, Dlamini S, Dube L,
Mabuza N, Haile M, Ellman T, Flores A, Keiser O, Matse S, de la Tour R, Calmy A,
Kerschberger B. Low sensitivity of the fourth-generation antigen/antibody HIV rapid
diagnostic test Determine™ HIV Early Detect for detection of acute HIV infection at the point
of care in rural Eswatini: a diagnostic accuracy study. *J Int AIDS Soc*. 2025;28(7):e26517.
48. Taegtmeyer M, MacPherson P, Jones K, Hopkins M, Moorcroft J, Lalloo DG, Chawla A.
Programmatic evaluation of a combined antigen and antibody test for rapid HIV diagnosis in
a community and sexual health clinic screening programme. PLoS One. 2011;6(11):e28019.
49. O'Neal JD, Golden MR, Branson BM, Stekler JD. HIV nucleic acid amplification testing
versus rapid testing: it is worth the wait. Testing preferences of men who have sex with men.
J Acquir Immune Defic Syndr. 2012 Aug 1;60(4):e117-20.
50. Meulbroek M, Ditzel E, Saz J, Taboada H, Pérez F, Pérez A, Carrillo A, Font G, Marazzi G,
Uya J, Cabrero J, Ingrami M, Marín R, Coll J, Pujol F. BCN Checkpoint, a community-based
centre for men who have sex with men in Barcelona, Catalonia, Spain, shows high efficiency
in HIV detection and linkage to care. HIV Med. 2013 Oct;14 Suppl 3:25-8.
51. Eu B, Roth N, Stoové M, O'Reilly M, Clarke E. Rapid HIV testing increases the rate of HIV
detection in men who have sex with men: using rapid HIV testing in a primary care clinic. Sex
Health. 2014 Mar;11(1):89-90.

52. Wouters K, Fransen K, Beelaert G, Kenyon C, Platteau T, Van Ghyseghem C, Collier I,
Buyze J, Florence E. Use of rapid HIV testing in a low threshold centre in Antwerp, Belgium,
2007-2012. Int J STD AIDS. 2014 Nov;25(13):936-42.
53. Yang M, Prestage G, Maycock B, Brown G, de Wit J, McKechnie M, Guy R, Keen P, Fairley
CK, Zablotska IB. The acceptability of different HIV testing approaches: cross-sectional study
among GMSM in Australia. Sex Transm Infect. 2014 Dec;90(8):592-5.
54. Mayer K, Tomassilli J, Gelman M, Altfeld M, Allen T. HIV self-testing among men who have
sex with men (MSM): acceptability of biweekly finger pricks to perform the alere
determineTM HIV-1/2 Ag/Ab assay; AIDS 2014 - XX International AIDS Conference; July
2014; Melbourne, Australia.
55. Conway DP, Guy R, Davies SC, Couldwell DL, McNulty A, Smith DE, Keen P, Cunningham P,
Holt M; Sydney Rapid HIV Test Study. Rapid HIV Testing Is Highly Acceptable and Preferred
among High-Risk Gay And Bisexual Men after Implementation in Sydney Sexual Health
Clinics. PLoS One. 2015 Apr 21;10(4):e0123814.
56. Conway DP, Guy R, McNulty A, Couldwell DL, Davies SC, Smith DE, Keen P, Cunningham P,
Holt M; Sydney Rapid HIV Test Study. Effect of testing experience and profession on
provider acceptability of rapid HIV testing after implementation in public sexual health clinics
in Sydney. HIV Med. 2015 May;16(5):280-7.
57. Hoenigl M, Graff-Zivin J, Little SJ. Costs per Diagnosis of Acute HIV Infection in Community-
based Screening Strategies: A Comparative Analysis of Four Screening Algorithms. Clin
Infect Dis. 2016 Feb 15;62(4):501-511.
58. Hoenigl M, Chaillon A, Mehta SR, Smith DM, Graff-Zivin J, Little SJ. Screening for acute HIV
infection in community-based settings: Cost-effectiveness and impact on transmissions. J
Infect. 2016 Nov;73(5):476-484.
59. Alibayeva KO, Saparbekov MK, Baiserkin BS, Abishev AT, Tazhibaeva GH, Kasymbekova
SZ. Study of the possibility of introduction of Kazakhstan NGO-based rapid HIV testing
procedures. HIV AIDS (Auckl). 2019 Sep 11;11:219-227.
60. Green Kimberley. 2019. Detecting acute HIV infection to accelerate epidemic control. 2019.
PATH, USAID, PEPFAR. Available at:
https://www.whitehatcom.com/Alere/K_Green_HIV_4th_gen_V2_112019.pdf
61. Mullens AB, Duyker J, Brownlow C, Lemoire J, Daken K, Gow J. Point-of-care testing
(POCT) for HIV/STI targeting MSM in regional Australia at community 'beat' locations. BMC
Health Serv Res. 2019 Feb 2;19(1):93.
62. Frola CE, Zalazar V, Cardozo N, Vázquez ML, Arístegui I, Lucas M, Gun A, Cahn P, Sued O.
Home-based HIV testing: Using different strategies among transgender women in Argentina.
PLoS One. 2020 Mar 19;15(3):e0230429.
63. Martin EG, Cornett J, Mohammed DY, Salaru G. Rapid Testing Algorithm Performance in a
Low-Prevalence Environment. Sex Transm Dis. 2020 May;47(5S Suppl 1):S35-S40.
64. Saz J, Dalmau-Bueno A, Meulbroek M, Pujol F, Coll J, Herraiz-Tomey Á, Pérez F, Marazzi G,
Taboada H, Culqui DR, Caylà JA; BCN Checkpoint Working Group. Use of fourth-generation
rapid combined antigen and antibody diagnostic tests for the detection of acute HIV infection
in a community centre for men who have sex with men, between 2016 and 2019. PLoS One.
2021 Jul 27;16(7):e0255065. doi: 10.1371/journal.pone.0255065. Erratum in: PLoS One.
2021 Oct 11;16(10):e0258613.

65. Williams OD, Dean JA, Crothers A, Gilks CF, Gow J. Economic evaluation of alternative testing regimes and settings to detect undiagnosed HIV in Australia. BMC Health Serv Res. 2021 Jan 7;21(1):30. doi: 10.1186/s12913-020-06040-5. PMID: 33413386; PMCID: PMC7789789.

66. Ferrer P, Bastias C, Beltrán C, Afani A. Diagnosis of HIV infection using mass community rapid testing in Santiago, Chile. J Clin Virol Plus. 2022;2:100064.

67. Muñoz-Chimeno M, Valencia J, Rodriguez-Recio A, Cuevas G, Garcia-Lugo A, Manzano S, Rodriguez-Paredes V, Fernandez B, Morago L, Casado C, Avellón A, Ryan P. HCV, HIV AND HBV rapid test diagnosis in non-clinical outreach settings can be as accurate as conventional laboratory tests. Sci Rep. 2023 May 9;13(1):7554.

68. Ochwoto M, Matiang'i M, Machuki Onchieku N, Ndoria S, Matoke L, Otinga M, Zablon J, Mathebula E, Matoke-Muhia D. The feasibility and impact of deploying a four-tests panel at antenatal care in primary health care facilities of a developing country, Kenya. Front Public Health. 2024 Nov 25;12:1399612.

69. Rodriguez NM, Balian L, Kataki I, Tolliver C, Rivera-De Jesus J, Linnes JC. Stakeholder-engaged development of a rapid test for detection of acute HIV infection. Res Sq [Preprint]. 2024 Apr 19:rs.3.rs-4243639.

70. Kiptinness C, Kuo AP, Reedy AM, Johnson CC, Ngure K, Wagner AD, Ortblad KF. Examining the Use of HIV Self-Testing to Support PrEP Delivery: a Systematic Literature Review. Curr HIV/AIDS Rep. 2022 Oct;19(5):394-408.

71. Tieosapjaroen W, Williams E, Johnson CC, Baptista Da Silva C, Barr-DiChiara M, Ong JJ. Web Annex D. Should HIV testing services with rapid diagnostic tests and/or self-tests be used for initiation or continuation of injectable long-acting PrEP? In: Guidelines on lenacapavir for HIV prevention and testing strategies for long-acting injectable pre-exposure prophylaxis (PrEP). Geneva: World Health Organization; 2025. Available: https://doi.org/10.2471/B09479

72. van de Vijver D, van Kampen J, Wongsa A, Janamnuaysook R, Mesplede T, Matrajt L et al. Evaluating HIV testing strategies to reduce drug resistance during CAB-LA rollout in Thailand. A modelling study. 13th International AIDS Society Conference on HIV Science; 13–17 July; Kigali, Rwanda, 2025. Available: https://www.abstract-archive.org/Abstract/Share/91175

73. Cox SN, Wu L, Wittenauer R, Clark S, Roberts DA, Nwogu IB, Vitruk O, Kuo AP, Johnson C, Jamil MS, Sands A, Schaefer R, Kisia C, Baggaley R, Stekler JD, Akullian A, Sharma M. Impact of HIV self-testing for oral pre-exposure prophylaxis scale-up on drug resistance and HIV outcomes in western Kenya: a modelling study. Lancet HIV. 2024 Mar;11(3):e167-e175.

74. Oliveira Leite, L. Magno, F. Soares, D. Zeballos, L. Dezanet, O. Ferreira, M. Westin, D. Greco, A. Grangeiro, I. Dourado, The PrEP15-19 Choices study group. Equivalent performance of HIV Oral Fluid Self-Testing and Rapid Testing compared to Nucleic Acid Amplification Test in Screening Adolescents for Long-Acting Injectable Cabotegravir in Brazil. 13th International AIDS Society Conference on HIV Science; 13–17 July; Kigali, Rwanda, 2025. Available: https://www.abstract-archive.org/Abstract/Share/92394

75. Vermeulen M, van Schalkwyk C, Jacobs G, et al. The Impact of Early Antiretroviral Treatment (ART) for HIV on the Sensitivity of the Latest Generation of Blood Screening and Point of Care Assays. *Viruses*. 2022;14(7):1426.

696   76. Toolkit to optimize HIV testing algorithms. Geneva World Health Organization; 2021.
697   Available: https://www.who.int/tools/optimizing-hiv-testing-algorithms-toolkit.
698   77. Smallwood M, Pant Pai N. Improving the Quality of Diagnostic Studies Evaluating Point of
699   Care Tests for Acute HIV Infections: Problems and Recommendations. *Diagnostics (Basel)*.
700   2017;7(1):13. Published 2017 Mar 4.